



666 - Afinion Glycohemoglobin;998 - CVS Afinion Glycohemoglobin	AFF-2																								
Glycohemoglobin	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Afinion Glycohemoglobin	163	8.22	0.50	6.70-9.70	6.10																				
747 - Auto Diff 7 (5-Part Diff)	BCX-6					BCX-7					BCX-8					BCX-9					BCX-10				
White Blood Cell Count	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Beckman Coulter ACT 5 diff (v 1.9)	13	7.18	0.26	6.10-8.30	3.61	13	3.85	0.13	3.30-4.50	3.33	13	9.82	0.33	8.30-11.30	3.37	12	17.33	0.55	14.70-19.90	3.18	12	4.80	0.17	4.10-5.50	3.54
PG: Horiba ABX Diag Pentra 60	16	7.21	0.25	6.10-8.30	3.47	16	3.86	0.13	3.30-4.50	3.33	16	9.92	0.27	8.40-11.40	2.73	16	17.47	0.52	14.90-20.10	2.97	16	4.79	0.16	4.10-5.50	3.33
SG: Select Beckman Coulter ACT Instruments (5-part)	16	7.14	0.24	6.00-8.20	3.38	16	3.84	0.13	3.20-4.40	3.42	16	9.75	0.31	8.30-11.30	3.16	15	17.31	0.47	14.70-19.90	2.77	15	4.79	0.15	4.10-5.50	3.13
SG: Select Horiba ABX Instruments (5-part Diff)	21	7.18	0.24	6.10-8.30	3.33	21	3.85	0.12	3.20-4.40	3.16	21	9.89	0.26	8.40-11.40	2.63	21	17.44	0.51	14.80-20.00	2.93	21	4.80	0.14	4.10-5.50	2.92
SG: International Horiba Instruments	24	7.18	0.22	6.10-8.30	3.06	24	3.84	0.13	3.20-4.40	3.16	24	9.90	0.25	8.40-11.40	2.53	24	17.44	0.48	14.80-20.00	2.76	24	4.81	0.14	4.10-5.50	2.92
Lymphocyte	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Beckman Coulter ACT 5 diff (v 1.9)	13	29.75	2.64	21.90-37.70	8.86	13	24.08	2.19	17.50-30.70	9.09	13	23.18	2.09	16.90-29.50	9.01	13	22.05	4.64	8.10-35.90	21.09	13	35.91	5.98	18.00-53.80	16.66
PG: Horiba ABX Diag Pentra 60	16	32.87	2.33	25.90-39.90	7.08	16	26.83	2.39	19.60-34.00	8.92	16	25.68	2.39	18.50-32.90	9.30	16	22.28	3.14	12.90-31.70	14.08	16	39.85	3.14	30.40-49.20	7.89
SG: Select Beckman Coulter ACT Instruments (5-part)	16	29.21	2.63	21.30-37.10	9.01	16	23.34	2.52	15.70-30.90	10.82	16	22.89	1.98	17.00-28.80	8.65	16	20.92	4.59	7.10-34.70	21.96	16	35.94	5.36	19.80-52.00	14.93
SG: Select Horiba ABX Instruments (5-part Diff)	21	32.45	2.26	25.60-39.20	6.98	21	26.93	2.60	19.10-34.70	9.67	21	25.15	2.41	18.00-32.40	9.56	21	21.89	2.87	13.30-30.50	13.11	21	39.90	3.14	30.50-49.30	7.87
SG: International Horiba Instruments	24	33.28	3.24	23.60-43.00	9.73	24	27.55	3.09	18.20-36.80	11.24	24	25.88	2.99	16.90-34.90	11.54	24	23.58	5.43	7.30-39.90	23.01	24	40.82	4.00	28.80-52.80	9.80
Monocyte	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Beckman Coulter ACT 5 diff (v 1.9)	13	1.24	0.56	0.00-2.92	45.24	13	0.79	0.45	0.00-2.15	57.22	13	0.85	0.38	0.00-2.00	44.94	13	2.23	1.07	0.00-5.43	47.80	13	0.97	0.67	0.00-2.97	68.87
PG: Horiba ABX Diag Pentra 60	16	1.83	0.61	0.00-3.67	33.55	16	1.45	0.47	0.03-2.88	32.76	16	1.19	0.53	0.00-2.77	44.37	16	3.01	1.08	0.00-6.24	35.78	16	1.66	0.69	0.00-3.71	41.27
SG: Select Beckman Coulter ACT Instruments (5-part)	16	1.11	0.55	0.00-2.77	49.91	16	0.76	0.42	0.00-2.01	55.00	16	0.78	0.42	0.00-2.05	54.49	16	1.94	1.10	0.00-5.23	56.55	16	0.91	0.62	0.00-2.76	67.91
SG: Select Horiba ABX Instruments (5-part Diff)	21	1.92	0.59	0.16-3.68	30.57	21	1.44	0.47	0.04-2.84	32.50	21	1.27	0.66	0.00-3.24	51.73	21	3.32	1.22	0.00-6.99	36.81	21	1.57	0.70	0.00-3.68	44.78
SG: International Horiba Instruments	24	1.80	0.63	0.00-3.69	35.06	24	1.39	0.47	0.00-2.79	33.45	24	1.27	0.62	0.00-3.12	48.50	24	3.14	1.27	0.00-6.95	40.45	24	1.47	0.71	0.00-3.61	48.57
Granulocyte	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Beckman Coulter ACT 5 diff (v 1.9)	13	50.45	3.60	39.60-61.20	7.14	13	54.58	2.59	46.80-62.40	4.74	13	54.66	3.69	43.60-65.80	6.75	13	59.01	4.60	45.20-72.80	7.80	13	44.64	6.38	25.50-63.70	14.30
PG: Horiba ABX Diag Pentra 60	15	54.31	3.71	43.10-65.50	6.85	16	59.03	3.01	50.00-68.00	5.10	16	59.42	3.61	48.60-70.20	6.08	16	67.86	3.25	58.10-77.70	4.80	16	44.61	2.27	37.80-51.40	5.09
SG: Select Beckman Coulter ACT Instruments (5-part)	16	50.95	3.88	39.30-62.50	7.62	16	54.78	2.44	47.50-62.10	4.45	16	54.61	3.99	42.60-66.60	7.31	16	60.32	4.78	46.00-74.60	7.93	16	44.59	5.74	27.40-61.80	12.87
SG: Select Horiba ABX Instruments (5-part Diff)	20	56.13	7.95	32.30-80.00	14.17	21	59.13	2.98	50.20-68.00	5.04	21	59.37	3.37	49.30-69.50	5.67	21	67.89	2.91	59.20-76.60	4.29	21	44.78	2.33	37.80-51.80	5.20
SG: International Horiba Instruments	23	55.16	7.86	31.60-78.80	14.24	24	58.26	3.69	47.20-69.40	6.33	24	58.53	3.91	46.80-70.20	6.68	24	66.55	4.61	52.80-80.40	6.92	24	44.03	3.13	34.60-53.40	7.11
Eosinophil	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Beckman Coulter ACT 5 diff (v 1.9)	13	10.92	3.32	0.90-20.90	30.46	13	12.91	2.16	6.40-19.40	16.74	13	15.06	4.60	1.30-28.90	30.46	13	6.45	2.07	0.20-12.60	32.34	13	12.99	3.47	2.60-23.40	26.69
PG: Horiba ABX Diag Pentra 60	16	10.71	2.95	1.80-19.60	27.57	16	12.67	1.70	7.60-17.80	13.39	16	13.39	2.59	5.60-21.20	19.33	16	6.35	0.58	4.70-8.10	9.06	16	13.63	1.69	8.50-18.70	12.43
SG: Select Beckman Coulter ACT Instruments (5-part)	16	11.04	3.16	1.50-20.50	28.73	16	13.21	2.30	6.30-20.10	17.42	16	15.38	4.47	2.00-28.80	29.03	16	6.32	1.84	0.80-11.80	29.21	16	12.87	3.20	3.30-22.50	24.81
SG: Select Horiba ABX Instruments (5-part Diff)	21	10.71	2.76	2.40-19.00	25.79	21	12.38	1.85	6.80-18.00	14.92	21	13.90	3.03	4.80-23.00	21.80	21	6.40	0.60	4.60-8.20	9.38	21	13.51	1.65	8.60-18.40	12.22
SG: International Horiba Instruments	24	10.73	2.64	2.80-18.60	24.67	24	12.65	1.95	6.80-18.50	15.35	24	14.02	2.90	5.30-22.70	20.71	24	6.23	0.76	3.90-8.50	12.26	24	13.45	1.66	8.40-18.40	12.39
Basophil	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Beckman Coulter ACT 5 diff (v 1.9)	14	7.82	0.47	6.41-9.23	6.00	14	7.78	0.37	6.68-8.88	4.71	14	6.26	0.52	4.69-7.82	8.33	14	10.19	1.93	4.41-15.98	18.92	14	5.41	1.62	0.55-10.28	29.95
PG: Horiba ABX Diag Pentra 60	16	0.39	0.03	0.32-0.47	6.35	16	0.40	0.00	0.40-0.40	0.00	16	0.30	0.00	0.30-0.30	0.00	16	0.50	0.05	0.34-0.66	10.32	16	0.25	0.05	0.10-0.41	20.64
SG: Select Beckman Coulter ACT Instruments (5-part)	17	7.81	0.46	6.42-9.19	5.92	17	8.05	0.78	5.71-10.40	9.71	17	6.31	0.53	4.73-7.89	8.35	17	10.37	1.72	5.21-15.54	16.60	17	5.34	1.47	0.93-9.75	27.53
SG: Select Horiba ABX Instruments (5-part Diff)	21	0.40	0.02	0.33-0.46	5.52	21	0.40	0.00	0.40-0.40	0.00	21	0.30	0.00	0.30-0.30	0.00	21	0.50	0.05	0.35-0.64	10.06	21	0.25	0.05	0.09-0.40	20.65
SG: International Horiba Instruments	24	0.40	0.02	0.34-0.46	5.15	24	0.40	0.00	0.40-0.40	0.00	24	0.30	0.00	0.30-0.30	0.00	24	0.50	0.05	0.36-0.64	9.35	24	0.25	0.05	0.09-0.40	20.69
Red Blood Cell Count	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Beckman Coulter ACT 5 diff (v 1.9)	13	4.60	0.06	4.32-4.88	1.28	13	3.88	0.04	3.65-4.11	1.11	13	2.58	0.03	2.43-2.73	1.09	13	5.00	0.09	4.70-5.30	1.82	13	4.87	0.07	4.58-5.16	1.46
PG: Horiba ABX Diag Pentra 60	16	4.58	0.05	4.31-4.85	1.07	16	3.87	0.05	3.64-4.10	1.32	16	2.58	0.04	2.43-2.73	1.36	16	5.03	0.04	4.73-5.33	0.70	16	4.84	0.06	4.55-5.13	1.28
SG: Select Beckman Coulter ACT Instruments (5-part)	16	4.59	0.07	4.31-4.87	1.48	16	3.88	0.06	3.65-4.11	1.60	16	2.59	0.03	2.43-2.75	1.16	16	5.00	0.10	4.70-5.30	2.00	16	4.86	0.08	4.57-5.15	1.73
SG: Select Horiba ABX Instruments (5-part Diff)	21	4.58	0.05	4.31-4.85	1.14	21	3.87	0.05	3.64-4.10	1.21	21	2.58	0.03	2.43-2.73	1.32	21	5.01	0.05	4.71-5.31	1.00	21	4.84	0.07	4.55-5.13	1.36
SG: International Horiba Instruments	24	4.58	0.06	4.31-4.85	1.31</																				



Test Name	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Beckman Coulter Act Instruments (5-part)	16	12.41	0.32	11.40-13.40	2.58	16	13.74	0.40	12.50-14.90	2.92	16	12.77	0.41	11.60-14.00	3.20	16	11.79	0.54	10.20-13.40	4.58	16	12.63	0.45	11.30-13.90	3.49
SG: Select Horiba ABX Instruments (5-part Diff)	21	12.40	0.34	11.40-13.40	2.74	21	13.77	0.36	12.70-14.90	2.61	21	12.69	0.36	11.60-13.80	2.83	21	11.60	0.28	10.80-12.40	2.33	21	12.51	0.38	11.40-13.60	3.04
SG: International Horiba Instruments	24	12.47	0.36	11.40-13.60	2.88	24	13.85	0.39	12.60-15.00	2.83	24	12.75	0.38	11.70-13.90	2.97	24	11.73	0.43	10.40-13.00	3.68	24	12.61	0.46	11.20-14.00	3.65
Platelet Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman Coulter Act 5 diff (v 1.9)	13	258.69	11.74	194.00-324.00	4.52	13	112.54	5.16	85.00-141.00	4.60	13	78.92	6.41	59.00-99.00	8.10	13	479.00	100.42	359.00-599.00	20.96	12	148.00	6.54	111.00-185.00	4.39
PG: Horiba ABX Diag Pentra 60	16	253.31	8.82	190.00-316.00	3.48	16	111.38	5.26	83.00-139.00	4.77	16	76.00	4.91	57.00-95.00	6.45	16	498.44	6.10	374.00-623.00	1.22	16	147.06	3.84	110.00-184.00	2.59
SG: Select Beckman Coulter Act Instruments (5-part)	16	258.18	10.34	194.00-323.00	3.99	16	112.76	4.55	85.00-141.00	3.98	16	79.76	5.94	60.00-100.00	7.38	16	485.88	88.12	365.00-608.00	18.13	15	148.33	6.22	111.00-185.00	4.19
SG: Select Horiba ABX Instruments (5-part Diff)	21	252.67	8.09	190.00-316.00	3.20	21	110.76	5.73	83.00-139.00	5.14	21	76.62	5.08	58.00-96.00	6.62	21	499.95	6.56	375.00-625.00	1.32	21	148.00	5.43	111.00-185.00	3.65
SG: International Horiba Instruments	24	252.79	7.76	190.00-316.00	3.08	24	110.25	5.52	83.00-138.00	5.00	24	76.50	4.88	58.00-96.00	6.36	24	499.13	7.13	374.00-624.00	1.42	24	147.83	5.17	111.00-185.00	3.51
702 - Blood Gases	BG-6					BG-7					BG-8					BG-9					BG-10				
Blood Gas - pCO2	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Non - i-STAT Blood Gas Instruments	5	69.20	3.43	63.00-75.00	4.93	5	59.67	3.02	55.00-65.00	5.00	5	81.40	4.65	75.00-87.00	5.68	5	53.00	1.56	48.00-58.00	3.02	5	40.58	2.14	36.00-46.00	5.12
SG: International Blood Gas Inst	5	66.08	3.41	61.00-71.00	5.15	5	54.94	4.63	50.00-60.00	8.36	5	72.00	4.58	66.00-78.00	6.39	6	50.00	4.96	45.00-55.00	10.00	6	39.38	2.79	34.00-44.00	7.18
Blood Gas - pH	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Non - i-STAT Blood Gas Instruments	5	7.15	0.01	7.11-7.19	0.08	5	7.19	0.01	7.15-7.23	0.19	5	6.95	0.03	6.91-6.99	0.42	5	7.25	0.01	7.21-7.29	0.12	5	7.30	0.01	7.26-7.34	0.07
SG: International Blood Gas Inst	6	7.15	0.01	7.11-7.19	0.10	6	7.20	0.02	7.16-7.24	0.33	6	6.95	0.02	6.91-6.99	0.33	6	7.26	0.02	7.22-7.30	0.28	6	7.30	0.01	7.26-7.34	0.14
Blood Gas - pO2	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Non - i-STAT Blood Gas Instruments	5	76.30	14.68	32.00-120.00	19.34	5	204.15	19.14	147.00-261.00	9.36	5	62.90	23.84	0.00-134.00	37.78	5	180.10	8.46	155.00-206.00	4.72	5	72.07	12.36	35.00-109.00	17.22
SG: International Blood Gas Inst	6	76.53	11.38	43.00-111.00	14.81	6	201.10	15.68	154.00-248.00	7.81	6	59.93	19.13	3.00-117.00	31.83	6	177.90	7.55	155.00-201.00	4.27	6	72.55	9.65	44.00-102.00	13.29
Sodium (NA+)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Osmetech Opti CCA	5	129.40	1.44	125.00-133.00	1.09	5	174.87	2.73	171.00-179.00	1.54	5	129.00	1.00	125.00-133.00	0.78	5	165.40	3.17	161.00-169.00	1.94	5	115.90	1.01	112.00-120.00	0.86
Potassium (K+)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Osmetech Opti CCA	5	2.62	0.03	2.10-3.10	1.15	5	7.31	0.18	6.80-7.80	2.47	5	3.21	0.02	2.70-3.70	0.63	5	4.71	0.42	4.20-5.20	8.94	5	1.66	0.10	1.20-2.20	5.88
714 - Complete Cardiac Markers; 715 - Basic Cardiac Markers; 711 - BNP; 720 - Cardiac Markers I; 740 - D-Dimer; 689 - Cardiac Profile	CDM-6N					CDM-7N					CDM-8N					CDM-9N					CDM-10N				
BNP	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat	30	415.47	31.12	322.10-508.90	7.49	31	1447.39	115.96	1099.50-1795.30	8.01	24	55.13	7.47	32.70-77.50	13.56	24	792.92	70.13	582.50-1003.30	8.84	24	2442.63	248.81	1696.20-3189.00	10.19
PG: AlereTriage	106	385.88	50.19	235.30-536.50	13.01	13	1236.92	146.99	795.90-1677.90	11.88	9	59.56	5.38	43.50-75.70	9.03	9	752.78	97.70	459.70-1045.90	12.98	10	1915.00	260.18	1134.50-2695.50	13.59
SG: Select Cardiac Marker Insts.	34	412.61	36.07	304.40-520.80	8.74	31	1447.39	115.96	1099.50-1795.30	8.01	24	55.13	7.47	32.70-77.50	13.56	24	792.92	70.13	582.50-1003.30	8.84	24	2442.63	248.81	1696.20-3189.00	10.19
SG: Select Abbott Instruments - Cardiac Markers	32	414.95	35.79	307.50-522.30	8.63	31	1447.39	115.96	1099.50-1795.30	8.01	24	55.13	7.47	32.70-77.50	13.56	24	792.92	70.13	582.50-1003.30	8.84	24	2442.63	248.81	1696.20-3189.00	10.19
SG: International Abbott i-Stat Instruments	30	415.47	31.12	322.10-508.90	7.49	31	1447.39	115.96	1099.50-1795.30	8.01	24	55.13	7.47	32.70-77.50	13.56	24	792.92	70.13	582.50-1003.30	8.84	24	2442.63	248.81	1696.20-3189.00	10.19
AG: All Group	140	392.32	48.43	247.00-537.60	12.35	44	1385.20	157.63	912.30-1858.10	11.38	33	56.33	7.17	34.80-77.80	12.74	33	781.97	79.06	544.80-1019.20	10.11	33	2309.79	327.77	1326.50-3293.10	14.19
CK-MB (Quantitative)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: AlereTriage	101	7.50	1.31	3.60-11.40	17.47	106	21.92	4.19	9.30-34.50	19.13	105	2.70	0.60	0.90-4.50	22.22	104	12.49	2.70	4.40-20.60	21.60	106	39.61	7.83	16.10-63.10	19.77
SG: Select Cardiac Marker Insts.	5	17.08	2.20	10.50-23.70	12.87	6	54.15	6.06	36.00-72.40	11.18	6	6.60	0.85	4.00-9.10	12.88	6	29.88	3.25	20.10-39.60	10.87	6	102.00	10.50	70.50-133.50	10.29
SG: International OCD Vitros Instruments	5	12.70	0.00	12.70-12.70	0.00	5	37.30	0.00	37.30-37.30	0.00	5	5.30	0.00	5.30-5.30	0.00	5	21.50	0.00	21.50-21.50	0.00	5	68.80	0.00	68.80-68.80	0.00
D-dimer (Quantitative)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: AlereTriage	127	810.81	78.81	575.00-1047.00	9.72	38	1364.32	163.61	873.50-1855.10	11.99	32	665.34	50.88	512.00-818.00	7.65	32	999.47	122.18	632.00-1366.00	12.23	32	2096.25	240.49	1375.00-2818.00	11.47
AG: All Group	135	820.51	108.39	496.00-1146.00	13.21	44	1380.23	188.72	814.00-1946.40	13.67	38	666.69	57.19	495.00-839.00	8.58	38	997.49	125.08	622.00-1372.00	12.55	38	2118.63	287.08	1258.00-2980.00	13.55
Myoglobin (Quantitative)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: AlereTriage	31	93.93	18.95	37.10-150.70	20.17	34	238.57	45.33	102.60-374.60	19.00	33	35.04	7.98	11.10-58.90	22.80	32	146.42	28.00	62						



Calcium																									
	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Roche Cobas Integra / Roche Reagent																									
PG: OCD Vitros 250,350,400,500,750,950																									
SG: Select Roche Instruments / Roche Reagents Groups																									
SG: Abaxis Piccolo Instruments																									
SG: Select Abbott Architect Instruments																									
SG: International OCD Vitros Instruments																									
SG: International Siemens Instruments																									
SG: International Abaxis Piccolo Instruments																									
AG: All Group																									
Creatinine																									
PG: Abaxis Piccolo (Non-Waived)																									
PG: Abaxis Piccolo (Waived)																									
PG: Abbott Architect																									
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent																									
PG: Clinical Data Envoy 500																									
PG: Siemens Dimension/AR/ES/RxL/Xpand																									
PG: Horiba ABX Pentra 400																									
PG: Beckman (Olympus) AU																									
PG: Roche Cobas Integra / Roche Reagent																									
PG: OCD Vitros 250,350,400,500,750,950																									
SG: Select Roche Instruments / Roche Reagents Groups																									
SG: Abaxis Piccolo Instruments																									
SG: Select Abbott Architect Instruments																									
SG: International OCD Vitros Instruments																									
SG: International Abaxis Piccolo Instruments																									
AG: All Group																									
Glucose																									
PG: Abaxis Piccolo (Non-Waived)																									
PG: Abaxis Piccolo (Waived)																									
PG: Abbott Architect																									
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent																									
PG: Clinical Data Envoy 500																									
PG: Siemens Dimension/AR/ES/RxL/Xpand																									
PG: Horiba ABX Pentra 400																									
PG: Beckman (Olympus) AU																									
PG: Roche Cobas Integra / Roche Reagent																									
PG: OCD Vitros 250,350,400,500,750,950																									
SG: Select Roche Instruments / Roche Reagents Groups																									
SG: Abaxis Piccolo Instruments																									
PG: Arkray Spotchem EZ (Waived)																									
SG: Select Abbott Architect Instruments																									
SG: International OCD Vitros Instruments																									
SG: International Siemens Instruments / IFCC																									
SG: International Abaxis Piccolo Instruments																									
AG: All Group																									
Urea Nitrogen (BUN)																									
PG: Abaxis Piccolo (Non-Waived)																									
PG: Abaxis Piccolo (Waived)																									
PG: Abbott Architect																									
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent																									
PG: Clinical Data Envoy 500																									



PG: Siemens Dimension/AR/ES/RxL/Xpand	23	36.92	1.08	34.00-40.00	2.97	23	22.52	0.71	21.00-25.00	3.04	23	11.52	0.59	10.00-14.00	5.00	23	33.52	1.19	31.00-37.00	3.53	23	43.76	1.09	40.00-48.00	2.50	
PG: Horiba ABX Pentra 400	7	35.24	0.54	32.00-38.00	1.43	7	21.37	0.46	19.00-23.00	2.38	7	10.99	0.69	9.00-13.00	6.36	7	32.04	0.82	29.00-35.00	2.50	7	41.96	0.73	38.00-46.00	1.67	
PG: Beckman (Olympus) AU	14	36.67	2.09	34.00-40.00	5.68	15	22.56	1.21	21.00-25.00	5.22	15	11.38	0.62	9.00-13.00	5.45	14	33.47	1.73	30.00-36.00	5.15	15	43.94	2.57	40.00-48.00	5.91	
PG: Roche Cobas Integra / Roche Reagent	7	35.00	1.63	32.00-38.00	4.57	7	21.57	0.79	20.00-24.00	3.64	7	10.57	0.54	9.00-13.00	4.55	7	32.29	1.25	29.00-35.00	4.06	7	43.14	1.57	39.00-47.00	3.72	
PG: OCD Vitros 250,350,400,500,750,950	16	32.41	0.87	29.00-35.00	2.81	16	18.88	0.72	17.00-21.00	3.68	16	9.38	0.50	7.00-11.00	5.56	16	29.19	0.83	26.00-32.00	2.76	16	39.50	0.82	36.00-44.00	2.00	
SG: Select Roche Instruments / Roche Reagents Groups	8	34.88	1.55	32.00-38.00	4.57	8	21.50	0.76	20.00-24.00	3.64	8	10.75	0.71	9.00-13.00	6.36	8	32.25	1.17	29.00-35.00	3.75	8	42.38	2.62	38.00-46.00	6.19	
SG: Abaxis Piccolo Instruments	65	32.77	3.19	30.00-36.00	9.70	9	20.00	0.50	18.00-22.00	2.50	9	10.78	0.44	9.00-13.00	3.64	9	29.56	0.53	27.00-33.00	1.67	9	39.44	1.01	35.00-43.00	2.56	
SG: Select Abbott Architect Instruments	18	36.71	0.96	34.00-40.00	2.70	18	22.21	0.79	20.00-24.00	3.64	18	11.16	0.50	9.00-13.00	4.55	18	33.53	0.77	31.00-37.00	2.35	18	44.47	1.02	40.00-48.00	2.27	
SG: International OCD Vitros Instruments	19	32.29	0.84	29.00-35.00	2.50	19	18.85	0.67	17.00-21.00	3.68	19	9.40	0.50	7.00-11.00	5.56	19	29.05	0.83	26.00-32.00	2.76	19	39.45	0.76	35.00-43.00	2.05	
SG: International Abaxis Piccolo Instruments	69	32.79	3.10	30.00-36.00	9.39	13	20.08	0.44	18.00-22.00	2.00	13	10.87	0.41	9.00-13.00	3.64	13	29.67	0.57	27.00-33.00	2.00	13	39.38	1.13	35.00-43.00	2.82	
AG: All Group	198	34.69	2.78	32.00-38.00	8.00	144	21.65	1.47	20.00-24.00	6.82	144	11.06	0.90	9.00-13.00	8.18	142	32.35	1.96	29.00-35.00	6.25	144	42.41	2.86	38.00-46.00	6.90	
Albumin	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	
PG: Abaxis Piccolo (Non-Waived) / BCP	9	3.46	0.10	3.10-3.90	2.86	9	2.48	0.04	2.30-2.80	1.60	9	1.69	0.03	1.50-1.90	1.76	9	3.23	0.10	2.90-3.50	3.13	9	3.92	0.13	3.50-4.30	3.33	
PG: Abaxis Piccolo (Waived) / BCP	58	3.41	0.11	3.10-3.70	3.24																					
PG: Abbott Architect / BCP	8	3.48	0.05	3.10-3.90	1.43	8	2.48	0.05	2.30-2.80	2.00	8	1.65	0.05	1.50-1.90	2.94	8	3.26	0.05	3.00-3.60	1.52	8	3.98	0.05	3.60-4.40	1.25	
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	37	4.57	0.17	4.10-5.10	3.70	37	3.32	0.12	3.00-3.60	3.64	36	2.32	0.08	2.10-2.50	3.48	36	4.30	0.14	3.90-4.70	3.49	36	5.15	0.19	4.70-5.70	3.65	
PG: Clinical Data Envoy 500 / BCG	10	4.61	0.09	4.10-5.10	1.96	10	3.36	0.07	3.10-3.70	2.06	10	2.39	0.05	2.20-2.60	2.08	10	4.31	0.12	3.90-4.70	2.79	10	5.05	0.12	4.60-5.60	2.35	
PG: Siemens Dimension/AR/ES/RxL/Xpand / BCG	13	3.64	0.09	3.20-4.00	2.50	13	2.53	0.10	2.30-2.80	4.00	12	1.65	0.08	1.40-1.80	5.00	13	3.39	0.12	3.10-3.70	3.53	13	4.17	0.15	3.80-4.60	3.57	
PG: Siemens Dimension/AR/ES/RxL/Xpand / BCP	10	3.64	0.08	3.20-4.00	2.22	10	2.54	0.05	2.30-2.80	2.00	10	1.65	0.05	1.50-1.90	2.94	10	3.35	0.05	3.10-3.70	1.47	10	4.15	0.05	3.80-4.60	1.19	
PG: Horiba ABX Pentra 400 / BCG	8	4.43	0.10	4.00-4.80	2.27	8	3.31	0.10	3.00-3.60	3.03	8	2.31	0.06	2.10-2.50	2.61	8	4.20	0.11	3.80-4.60	2.62	8	4.96	0.12	4.50-5.50	2.40	
PG: Beckman (Olympus) AU / BCG	15	4.61	0.10	4.10-5.10	2.17	15	3.30	0.10	3.00-3.60	3.03	15	2.19	0.09	2.00-2.40	4.09	14	4.33	0.09	3.90-4.70	2.09	15	5.21	0.11	4.70-5.70	2.12	
PG: Roche Cobas Integra / BCG / Roche Reagent	6	4.90	0.09	4.40-5.40	1.84	7	3.64	0.08	3.20-4.00	2.22	7	2.49	0.07	2.30-2.80	2.80	7	4.64	0.10	4.10-5.10	2.17	7	5.49	0.07	5.00-6.00	1.27	
PG: OCD Vitros 250,350,400,500,750,950 / BCG	16	4.87	0.14	4.40-5.40	2.86	16	3.20	0.11	2.90-3.50	3.44	16	1.97	0.06	1.80-2.20	3.00	15	4.62	0.16	4.10-5.10	3.48	16	5.61	0.25	5.00-6.20	4.46	
SG: Select Roche Instruments / BCG / Roche Reagents Groups	7	4.86	0.14	4.40-5.40	2.86	8	3.61	0.11	3.20-4.00	3.06	8	2.45	0.12	2.30-2.80	4.80	8	4.60	0.15	4.10-5.10	3.26	8	5.48	0.07	5.00-6.00	1.27	
SG: Abaxis Piccolo Instruments / BCP	67	3.41	0.11	3.10-3.70	3.24	10	2.48	0.04	2.30-2.80	1.60	10	1.70	0.05	1.50-1.90	2.94	10	3.23	0.10	2.90-3.50	2.81	10	3.92	0.12	3.50-4.30	3.08	
SG: Select Abbott Architect Instruments / BCG	8	4.48	0.49	4.00-5.00	10.89	8	3.16	0.38	2.90-3.50	11.56	8	2.14	0.26	1.90-2.30	12.38	8	4.17	0.49	3.80-4.60	11.67	8	5.01	0.57	4.50-5.50	11.60	
SG: Select Abbott Architect Instruments / BCP	10	3.48	0.04	3.10-3.90	1.14	10	2.48	0.04	2.30-2.80	1.60	10	1.66	0.05	1.50-1.90	2.94	10	3.27	0.05	3.00-3.60	1.52	10	3.98	0.04	3.60-4.40	1.00	
SG: International OCD Vitros Instruments / BCG	19	4.87	0.14	4.40-5.40	2.86	19	3.19	0.10	2.90-3.50	3.13	19	1.97	0.06	1.80-2.20	3.00	18	4.60	0.15	4.10-5.10	3.26	19	5.63	0.27	5.00-6.20	4.82	
SG: International Abaxis Piccolo Instruments / BCP	71	3.42	0.11	3.10-3.70	2.94	14	2.50	0.06	2.30-2.80	2.40	14	1.72	0.07	1.50-1.90	4.12	14	3.23	0.08	2.90-3.50	2.50	14	3.92	0.11	3.50-4.30	2.82	
SG: Select Gen Chem Instruments / BCG	42	4.55	0.25	4.10-5.10	5.43	42	3.29	0.19	3.00-3.60	5.76	42	2.26	0.16	2.10-2.50	6.96	41	4.27	0.24	3.90-4.70	5.58	42	5.09	0.28	4.60-5.60	5.49	
MG: BCG	118	4.50	0.39	4.00-5.00	8.67	119	3.20	0.30	2.90-3.50	9.38	110	2.20	0.21	2.00-2.40	9.55	116	4.30	0.38	3.90-4.70	8.84	109	5.20	0.34	4.70-5.70	6.54	
MG: BCP	92	3.50	0.13	3.10-3.90	3.71	35	2.50	0.06	2.30-2.80	2.40	35	1.70	0.07	1.50-1.90	4.12	35	3.30	0.10	3.00-3.60	3.03	35	4.00	0.14	3.60-4.40	3.50	
Alkaline Phosphatase	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	
PG: Abaxis Piccolo (Non-Waived)	9	206.89	10.91	145.00-269.00	5.27	9	132.56	6.39	93.00-173.00	4.81	9	61.00	3.81	43.00-79.00	6.23	9	190.78	10.24	134.00-248.00	5.34	9	241.22	12.77	169.00-313.00	5.31	
PG: Abaxis Piccolo (Waived)	59	208.25	10.30	146.00-270.00	4.95																					
PG: Abbott Architect	13	308.07	13.41	216.00-400.00	4.35	13	182.31	10.87	127.00-237.00	5.99	13	78.92	5.39	55.00-103.00	6.84	13	281.15	14.77	197.00-365.00	5.27	13	366.31	16.29	256.00-476.00	4.45	
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	35	282.03	17.94	197.00-367.00	6.35	35	169.29	10.99	118.00-220.00	6.51	34	73.46	4.64	51.00-95.00	6.30	34	261.03	26.24	183.00-339.00	10.04	34	337.41	19.52	236.00-438.00	5.79	
PG: Clinical Data Envoy 500	9	252.45	9.10	176.00-328.00	3.61	9	150.81	10.63	106.00-196.00	7.02	9	66.42	7.71	46.00-86.00	11.67	9	234.65	16.00	165.00-306.00	6.81	9	305.03	17.54	214.00-397.00	5.74	
PG: Siemens Dimension/AR/ES/RxL/Xpand	16	324.24	13.98	227.00-421.00	4.32	16	191.18	9.28	134.00-248.00	4.87	16	82.35	4.55	57.00-107.00	5.61	16	295.71	12.30	207.00-385.00	4.16	16	387.88	14.08	272.00-504.00	3.63	
PG: Siemens Dimension/AR/ES/RxL/Xpand / IFCC	7	303.25	5.39	212.00-394.00	1.78	7	177.00	3.29	124.00-230.00	1.86	7	76.75	2.92	54.00-100.00	3.77	7	278.25	4.65	195.00-361.00	1.69	7	370.00	6.26	259.00-481.00	1.70	
PG: Horiba ABX Pentra 400	8	308.00	15.26	216.00-400.00	4.97	8	182.50	9.58	128.00-238.00	5.25	8	79.50	5.40	56.00-104.00	6.75	8	281.13	13.90	197.00-365.00	4.95	8	364.88	18.85	256.00-475.00	5.15	
PG: Beckman (Olympus) AU	15	247.81	24.20	174.00-322.00	9.76	15	145.50	14.5																		

Test Name	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
MG: IFCC	14	207.00	17.00	166.00-248.00	8.21	14	120.00	5.70	96.00-144.00	4.75	14	51.00	2.70	41.00-61.00	5.29	14	191.00	8.40	153.00-229.00	4.40	14	257.00	11.20	206.00-308.00	4.36
AG: All Group	204	181.58	21.34	146.00-218.00	11.70	148	106.96	10.98	86.00-128.00	10.28	138	45.57	6.67	37.00-55.00	14.57	147	169.68	17.12	136.00-204.00	10.06	145	225.55	25.99	181.00-271.00	11.50
AST/SGOT	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	9	167.56	6.88	134.00-202.00	4.11	9	107.89	5.51	86.00-130.00	5.09	9	50.44	1.94	40.00-60.00	3.80	9	154.78	10.84	124.00-186.00	6.97	9	196.78	9.47	158.00-236.00	4.82
PG: Abaxis Piccolo (Waived)	58	167.22	5.86	134.00-200.00	3.53																				
PG: Abbott Architect	13	171.40	3.56	137.00-205.00	2.11	13	107.08	2.63	86.00-128.00	2.43	13	46.31	1.03	37.00-55.00	2.17	13	161.69	7.33	130.00-194.00	4.51	13	202.00	7.11	162.00-242.00	3.51
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	35	152.55	5.52	122.00-184.00	3.59	35	97.47	4.11	78.00-116.00	4.23	34	40.49	2.38	32.00-48.00	6.00	34	141.05	5.27	113.00-169.00	3.76	34	180.97	5.42	145.00-217.00	2.98
PG: Clinical Data Envoy 500	8	179.31	5.46	143.00-215.00	3.07	8	112.03	3.25	90.00-134.00	2.86	8	49.48	1.11	39.00-59.00	2.24	8	169.23	7.88	135.00-203.00	4.67	8	214.44	7.51	171.00-257.00	3.50
PG: Siemens Dimension/AR/ES/RxL/Xpand	23	213.96	7.73	171.00-257.00	3.60	23	122.04	5.09	98.00-146.00	4.18	23	49.92	2.86	40.00-60.00	5.80	23	193.84	8.23	155.00-233.00	4.23	23	260.12	10.27	208.00-312.00	3.96
PG: Horiba ABX Pentra 400	7	174.57	4.89	140.00-210.00	2.80	7	108.71	3.04	87.00-131.00	2.75	7	47.29	1.25	38.00-56.00	2.77	7	161.71	6.92	130.00-194.00	4.26	7	203.57	7.09	163.00-245.00	3.48
PG: Beckman (Olympus) AU	15	153.50	8.40	123.00-185.00	5.45	15	95.06	5.53	76.00-114.00	5.79	15	41.75	2.41	34.00-50.00	5.71	15	142.81	8.20	114.00-172.00	5.73	15	179.00	10.94	143.00-215.00	6.09
PG: Roche Cobas Integra / Roche Reagent	7	165.29	6.55	132.00-198.00	4.00	7	107.14	3.72	86.00-128.00	3.46	7	46.29	2.06	37.00-55.00	4.57	7	158.57	6.40	127.00-191.00	4.03	7	195.71	5.79	157.00-235.00	2.96
PG: OCD Vitros 250,350,400,500,750,950	16	195.00	7.62	156.00-234.00	3.90	16	117.71	4.25	94.00-142.00	3.64	16	50.59	1.87	41.00-61.00	3.73	16	181.94	6.64	146.00-218.00	3.63	16	236.59	9.47	190.00-284.00	4.01
SG: Select Roche Instruments / Roche Reagents Groups	8	164.00	7.07	131.00-197.00	4.33	8	106.38	4.07	85.00-127.00	3.87	8	46.00	2.07	37.00-55.00	4.57	8	156.38	8.59	125.00-187.00	5.51	8	193.75	7.72	155.00-233.00	3.97
SG: Abaxis Piccolo Instruments	67	167.26	5.95	134.00-200.00	3.59	10	107.20	5.63	86.00-128.00	5.23	10	50.30	1.89	40.00-60.00	3.80	10	154.30	10.33	123.00-185.00	6.69	10	195.50	9.80	157.00-235.00	5.00
SG: Select SpotChem Instruments	17	192.44	13.13	154.00-230.00	6.82	17	116.22	7.53	93.00-139.00	6.47	17	50.06	2.90	40.00-60.00	5.80	17	179.39	12.60	143.00-215.00	7.04	17	233.00	17.78	186.00-280.00	7.64
SG: Select Abbott Architect Instruments	18	170.86	4.00	137.00-205.00	2.34	18	107.11	2.54	86.00-128.00	2.34	18	46.32	1.00	37.00-55.00	2.17	18	159.63	7.23	128.00-192.00	4.50	18	201.21	6.77	161.00-241.00	3.38
SG: International OCD Vitros Instruments	19	195.00	7.11	156.00-234.00	3.64	19	116.90	4.32	94.00-140.00	3.68	19	50.52	1.72	41.00-61.00	3.33	19	180.38	7.57	144.00-216.00	4.22	19	236.57	10.03	190.00-284.00	4.22
SG: International Abaxis Piccolo Instruments	68	167.51	6.24	134.00-202.00	3.69	11	107.64	5.54	86.00-130.00	5.09	11	50.55	1.97	41.00-61.00	3.92	11	154.82	9.95	124.00-186.00	6.39	11	196.00	9.45	157.00-235.00	4.80
AG: All Group	200	172.29	19.91	138.00-206.00	11.57	143	107.35	10.43	86.00-128.00	9.72	142	45.97	4.57	37.00-55.00	10.00	142	161.52	20.59	130.00-194.00	12.72	142	207.71	30.27	166.00-250.00	14.57
Bilirubin Total	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	9	2.92	0.25	2.30-3.50	8.62	9	1.62	0.15	1.20-2.00	9.38	9	0.60	0.07	0.20-1.00	11.67	9	2.66	0.27	2.20-3.20	10.00	9	3.50	0.35	2.80-4.20	10.00
PG: Abaxis Piccolo (Waived)	59	2.85	0.31	2.20-3.50	10.69																				
PG: Abbott Architect	13	2.72	0.31	2.20-3.20	11.48	13	1.49	0.17	1.10-1.90	12.00	13	0.42	0.06	0.00-0.80	15.00	13	2.50	0.30	2.00-3.00	12.00	13	3.25	0.46	2.60-3.80	14.38
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	34	3.44	0.31	2.70-4.10	9.12	34	1.90	0.17	1.50-2.30	8.95	33	0.61	0.07	0.20-1.00	10.00	33	3.12	0.26	2.50-3.70	8.39	33	4.15	0.36	3.40-5.00	8.57
PG: Clinical Data Envoy 500	9	2.81	0.26	2.20-3.40	9.29	9	1.48	0.17	1.10-1.90	11.33	8	0.42	0.06	0.00-0.80	15.00	9	2.55	0.22	2.10-3.10	8.46	9	3.44	0.33	2.70-4.10	9.71
PG: Siemens Dimension/AR/ES/RxL/Xpand	17	2.93	0.30	2.30-3.50	10.34	17	1.57	0.19	1.20-2.00	11.88	15	0.41	0.05	0.00-0.80	12.50	17	2.67	0.29	2.20-3.20	10.74	17	3.48	0.36	2.80-4.20	10.57
PG: Siemens Dimension/AR/ES/RxL/Xpand / IFCC	5	3.19	0.05	2.60-3.80	1.56	5	1.72	0.07	1.30-2.10	4.71	5	0.50	0.00	0.10-0.90	0.00	5	2.86	0.06	2.30-3.50	2.07	5	3.88	0.11	3.10-4.70	2.82
PG: Horiba ABX Pentra 400	8	2.99	0.32	2.40-3.60	10.67	8	1.60	0.14	1.20-2.00	8.75	8	0.48	0.05	0.10-0.90	10.00	8	2.68	0.24	2.20-3.20	8.89	8	3.63	0.43	2.90-4.30	11.94
PG: Beckman (Olympus) AU	15	3.26	0.21	2.60-4.00	6.36	15	1.79	0.12	1.40-2.20	6.67	15	0.56	0.05	0.20-1.00	8.33	15	3.00	0.20	2.40-3.60	6.67	15	3.91	0.25	3.10-4.70	6.41
PG: Roche Cobas Integra / Roche Reagent	7	2.74	0.13	2.20-3.20	4.81	7	1.47	0.08	1.10-1.90	5.33	7	0.40	0.06	0.00-0.80	15.00	7	2.56	0.08	2.10-3.10	3.08	7	3.30	0.16	2.60-4.00	4.85
PG: OCD Vitros 250,350,400,500,750,950	15	3.16	0.36	2.60-3.80	11.56	15	1.74	0.14	1.30-2.10	8.24	15	0.49	0.08	0.10-0.90	16.00	15	2.86	0.29	2.30-3.50	10.00	15	3.68	0.39	3.00-4.40	10.54
SG: Select Roche Instruments / Roche Reagents Groups	8	2.76	0.13	2.20-3.40	4.64	8	1.50	0.11	1.10-1.90	7.33	8	0.43	0.09	0.00-0.80	22.50	8	2.59	0.11	2.10-3.10	4.23	8	3.34	0.19	2.60-4.00	5.45
SG: Abaxis Piccolo Instruments	68	2.86	0.30	2.30-3.50	10.34	10	1.62	0.14	1.20-2.00	8.75	10	0.60	0.07	0.20-1.00	11.67	10	2.66	0.26	2.20-3.20	9.26	10	3.50	0.33	2.80-4.20	9.43
SG: Select SpotChem Instruments	16	3.18	0.37	2.60-3.80	11.56	16	1.74	0.14	1.30-2.10	8.24	16	0.49	0.08	0.10-0.90	16.00	16	2.88	0.29	2.30-3.50	10.00	16	3.72	0.41	3.00-4.40	11.08
SG: Select Abbott Architect Instruments	18	2.71	0.30	2.20-3.20	11.11	18	1.49	0.16	1.10-1.90	10.67	18	0.43	0.06	0.00-0.80	15.00	18	2.48	0.28	2.00-3.00	11.20	18	3.24	0.43	2.60-3.80	13.44
SG: International OCD Vitros Instruments	18	3.06	0.44	2.50-3.70	14.19	18	1.69	0.21	1.30-2.10	12.35	18	0.48	0.10	0.10-0.90	20.00	18	2.76	0.39	2.20-3.40	13.93	18	3.56	0.50	2.90-4.30	13.89
SG: International Abaxis Piccolo Instruments	69	2.86	0.30	2.30-3.50	10.34	11	1.61	0.14	1.20-2.00	8.75	11	0.60	0.06	0.20-1.00	10.00	11	2.65	0.25	2.10-3.10	9.62	11	3.48	0.32	2.80-4.20	9.14
AG: All Group	200	3.02	0.38	2.40-3.60	12.67	142	1.68	0.22	1.30-2.10	12.94	138	0.51	0.10	0.10-0.90	20.00	141	2.81	0.34	2.20-3.40	12.14	141	3.70	0.47	3.00-4.40	12.70
Total Protein	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	8	8.09	0.18	7.30-8.90	2.22	8	5.58	0.05	5.00-6.20	0.89	8	3.71	0.06	3.30-4.10	1.62	8	7.46	0.09	6.80-8.30	1.20	8	9.54	0.37	8.60-10.40	3.89
PG: Abaxis Piccolo (Waived)	57	8.11	0.16	7.30-8.90	1.98																				
PG: Abbott Architect	13	7.97	0.12	7.20-8.80	1.50	13	5.39	0.10	4.90-5.90	1.85	13	3.51	0.25	3.10-3.90	7.14	13	7.39	0.11	6.70-8.10	1.49	13	9.24	0.13	8.30-10.10	1.41
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	35	8.08	0.53	7.30-8.90	6.54	35	5.67	0.22	5.10-6.30	3.86	34	3.69	0.20	3.30-4.10	5.41	34	7.61	0.31	6.80-8.40	4.08	34	9.36	0.35	8.50-10.30	3.72
PG: Clinical Data Envoy 500	9	7.94	0.41	7.10-8.70	5.19	9	5.50	0.29	5.00-6.00	5.27	9	3.59	0.22	3.20-4.00	6.11	9	7.45	0.32							



Method	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: Cholestech (S)	6	209.14	7.22	188.00-230.00	3.44	6	160.71	6.70	145.00-177.00	4.16	6	123.71	6.18	112.00-136.00	5.00	5	205.00	17.54	185.00-226.00	8.54	6	231.43	11.16	208.00-254.00	4.85	
PG: Clinical Data Envoy 500	8	199.50	5.60	180.00-220.00	2.80	8	158.93	2.47	143.00-175.00	1.57	8	124.78	6.96	113.00-138.00	5.60	8	192.16	6.92	173.00-211.00	3.59	8	225.81	6.52	203.00-249.00	2.88	
PG: Siemens Dimension/AR/ES/RxL/Xpand	22	196.79	4.66	177.00-217.00	2.39	22	156.17	4.49	140.00-172.00	2.88	22	125.21	3.28	113.00-138.00	2.64	22	187.42	4.65	168.00-206.00	2.51	22	217.92	5.99	196.00-240.00	2.75	
PG: Horiba ABX Pentra 400	7	202.00	8.16	182.00-222.00	4.06	7	156.86	5.15	141.00-173.00	3.25	7	124.71	3.64	113.00-138.00	2.88	7	190.71	7.20	172.00-210.00	3.77	7	221.43	8.12	199.00-243.00	3.67	
PG: Beckman (Olympus) AU	14	196.07	7.40	176.00-216.00	3.78	14	153.73	6.39	139.00-169.00	4.16	14	122.20	6.04	110.00-134.00	4.92	14	186.07	8.43	167.00-205.00	4.52	14	215.07	9.15	194.00-237.00	4.28	
PG: Roche Cobas Integra / Roche Reagent	6	198.50	1.23	179.00-219.00	0.60	6	159.33	2.42	143.00-175.00	1.51	6	126.00	2.68	113.00-139.00	2.14	6	190.33	3.56	171.00-209.00	1.89	6	217.83	2.32	196.00-240.00	1.06	
PG: OCD Vitros 250,350,400,500,750,950	13	201.43	6.01	181.00-221.00	2.99	13	153.92	4.39	139.00-169.00	2.86	13	116.23	2.65	104.00-128.00	2.33	13	192.62	5.16	174.00-212.00	2.69	13	225.92	4.42	203.00-249.00	1.95	
SG: Select Roche Instruments / Roche Reagents Groups	8	197.25	2.55	177.00-217.00	1.27	8	158.00	3.21	142.00-174.00	2.03	8	124.75	3.41	113.00-138.00	2.72	8	189.63	3.29	171.00-209.00	1.74	8	217.38	2.20	195.00-239.00	1.01	
SG: Abaxis Piccolo Instruments	12	195.58	2.91	176.00-216.00	1.48	5	158.25	3.69	142.00-174.00	2.34	5	125.00	1.16	113.00-138.00	0.96	5	189.50	2.65	171.00-209.00	1.37	5	215.00	3.56	194.00-237.00	1.67	
SG: Select Abbott Architect Instruments	15	205.44	1.38	185.00-226.00	0.68	15	160.56	1.21	145.00-177.00	0.75	15	127.56	1.21	115.00-141.00	0.94	15	195.81	3.47	176.00-216.00	1.79	14	228.07	1.87	205.00-251.00	0.83	
SG: International OCD Vitros Instruments	16	201.94	5.70	182.00-222.00	2.82	16	154.00	3.90	139.00-169.00	2.53	16	116.18	2.53	104.00-128.00	2.16	16	192.82	4.93	174.00-212.00	2.54	16	226.76	5.68	204.00-250.00	2.51	
SG: International Abaxis Piccolo Instruments	13	194.85	3.85	176.00-215.00	1.95	5	156.40	5.22	140.00-172.00	3.33	5	124.80	1.10	113.00-138.00	0.88	5	188.00	4.06	169.00-207.00	2.18	5	212.40	6.58	191.00-233.00	3.11	
AG: All Group	142	200.69	7.43	181.00-221.00	3.68	134	158.10	5.97	142.00-174.00	3.80	134	125.22	6.03	113.00-138.00	4.80	133	192.20	9.61	173.00-211.00	5.00	133	222.74	9.60	201.00-245.00	4.30	
Cholesterol HDL	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: Abaxis Piccolo (Non-Waived) / Direct Method	5	63.00	3.74	44.00-82.00	5.87	5	58.75	3.30	41.00-77.00	5.59	5	47.50	3.11	34.00-62.00	6.46	5	68.00	6.58	48.00-88.00	9.71	5	74.50	12.15	53.00-98.00	16.27	
PG: Abaxis Piccolo (Waived) / Direct Method	9	73.11	11.31	51.00-95.00	15.48																					
PG: Abbott Architect / Direct Method	10	81.67	2.61	57.00-107.00	3.17	10	68.70	2.45	48.00-90.00	3.62	10	58.60	2.76	41.00-77.00	4.75	10	78.90	3.04	55.00-103.00	3.80	10	87.80	3.99	62.00-114.00	4.55	
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Direct Method / Alfa Wasserman Reagent	32	80.30	4.75	56.00-104.00	6.00	32	67.79	4.20	48.00-88.00	6.18	32	58.18	4.10	41.00-75.00	7.07	32	77.09	5.28	54.00-100.00	6.88	32	84.64	5.46	60.00-111.00	6.47	
PG: Cholestech (S) / Direct Method	6	72.71	10.06	51.00-95.00	13.84	6	59.29	5.99	41.00-77.00	10.17	6	41.71	5.09	29.00-55.00	12.14	5	76.67	12.48	54.00-100.00	16.23	6	73.43	10.77	51.00-95.00	14.79	
PG: Clinical Data Envoy 500 / Direct Method	8	86.22	5.64	60.00-112.00	6.51	8	74.27	5.29	52.00-96.00	7.16	8	61.98	5.28	43.00-81.00	8.55	8	85.33	5.98	60.00-111.00	7.06	8	96.35	5.30	67.00-125.00	5.52	
PG: Siemens Dimension/AR/ES/RxL/Xpand / Direct Method	21	69.24	13.77	48.00-90.00	20.00	20	71.68	2.88	50.00-94.00	4.03	21	69.13	2.47	48.00-90.00	3.62	21	72.91	3.10	51.00-95.00	4.25	21	72.35	3.26	50.00-94.00	4.58	
PG: Horiba ABX Pentra 400 / Direct Method	6	76.33	6.62	53.00-99.00	8.68	6	64.50	6.66	46.00-85.00	10.31	6	54.00	5.29	38.00-70.00	9.81	6	73.83	7.03	52.00-96.00	9.46	6	80.83	5.78	57.00-105.00	7.16	
PG: Beckman (Olympus) AU / Direct Method	14	84.13	6.46	59.00-109.00	7.74	14	69.33	5.90	48.00-90.00	8.55	14	58.27	5.24	41.00-75.00	8.97	14	81.67	7.53	57.00-107.00	9.15	14	89.53	8.09	63.00-117.00	9.00	
PG: Roche Cobas Integra / Direct Method / Roche Reagent	6	71.67	2.73	50.00-94.00	3.75	6	72.83	2.93	51.00-95.00	3.97	6	69.83	1.60	49.00-91.00	2.29	6	73.17	2.48	51.00-95.00	3.42	6	71.17	3.06	50.00-92.00	4.37	
PG: OCD Vitros 250,350,400,500,750,950 / Direct Method	12	90.54	8.00	64.00-118.00	8.79	12	71.83	7.15	50.00-94.00	9.86	12	60.75	6.45	43.00-79.00	10.66	12	87.42	7.80	61.00-113.00	8.97	12	98.00	12.45	69.00-127.00	12.76	
SG: Select Roche Instruments / Direct Method / Roche Reagents Groups	8	71.00	2.83	50.00-92.00	3.94	8	71.00	4.41	50.00-92.00	6.20	8	66.75	7.42	47.00-87.00	11.04	8	72.13	2.85	50.00-94.00	4.03	8	71.13	2.80	50.00-92.00	3.94	
SG: Abaxis Piccolo Instruments / Direct Method	13	70.00	10.60	49.00-91.00	15.14																					
SG: Select Abbott Architect Instruments / Direct Method	15	81.94	3.06	57.00-107.00	3.78	15	68.69	2.94	48.00-90.00	4.20	15	58.75	2.75	41.00-77.00	4.58	15	79.00	3.58	55.00-103.00	4.56	15	87.88	3.96	62.00-114.00	4.55	
SG: International OCD Vitros Instruments / Direct Method	15	89.65	8.37	63.00-117.00	9.33	15	72.00	6.43	50.00-94.00	8.89	15	60.94	5.55	43.00-79.00	9.18	15	86.94	7.32	61.00-113.00	8.39	15	96.81	12.13	68.00-126.00	12.47	
SG: International Abaxis Piccolo Instruments / Direct Method	14	69.57	10.31	49.00-91.00	14.71	5	58.00	3.32	41.00-75.00	5.69	5	45.80	4.66	32.00-60.00	10.22	5	65.00	8.80	46.00-85.00	13.54	5	77.20	12.13	54.00-100.00	15.71	
MG: Direct Method	137	78.00	10.40	55.00-101.00	13.33	127	69.00	5.90	48.00-90.00	8.55	128	60.00	7.80	42.00-78.00	13.00	127	78.00	7.50	55.00-101.00	9.62	128	84.00	10.90	59.00-109.00	12.98	
MG: Indirect Method	5	74.00	10.50	52.00-96.00	14.19	5	68.00	5.70	48.00-88.00	8.38	5	59.00	5.40	41.00-77.00	9.15	5	72.00	7.70	50.00-94.00	10.69	5	83.00	12.70	58.00-108.00	15.30	
Cholesterol LDL	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: Abaxis Piccolo (Non-Waived) / Calc	5	96.25	2.36	89.00-103.00	2.50	5	71.75	2.87	63.00-81.00	4.03	5	58.00	3.74	47.00-69.00	6.38	5	87.00	7.35	65.00-109.00	8.39	5	98.00	10.23	67.00-129.00	10.41	
PG: Abbott Architect / Direct Method	7	69.33	1.66	64.00-74.00	2.46	7	46.57	0.98	44.00-50.00	2.13	7	29.71	0.76	28.00-32.00	2.67	7	64.29	1.98	58.00-70.00	3.13	7	80.29	2.50	73.00-88.00	3.13	
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Calc / Alfa Wasserman Reagent	23	82.48	15.32	36.00-128.00	18.66	23	62.21	12.35	25.00-99.00	20.00	22	47.98	8.05	24.00-72.00	16.88	23	77.44	14.61	33.00-121.00	18.96	23	92.65	17.37	41.00-145.00	18.71	
PG: Cholestech (S) / Calc	5	97.50	14.53	55.00-142.00	14.80	5	70.50	3.00	62.00-80.00	4.23	5	63.50	8.70	38.00-90.00	13.59	5	94.25	5.68	77.00-111.00	6.06	5	116.50	13.30	77.00-157.00	11.37	
PG: Clinical Data Envoy 500 / Calc	5	73.22	9.51	45.00-102.00	13.01	5	52.56	8.61	27.00-79.00	16.23	5	38.24	6.68	18.00-58.00	17.63	5	67.06	8.27	42.00-92.00	12.39	5	85.28	8.65	59.00-111.00	10.12	
PG: Siemens Dimension/AR/ES/RxL/Xpand / Calc	8	89.02	6.39	70.00-108.00	7.19	8	60.96	5.99	43.00-79.00	9.84	8	38.60	2.96	30.00-48.00	7.69	8	80.40	6.16	61.00-99.00	7.75	8	103.33	10.10	73.00-133.00	9.81	
PG: Siemens Dimension/AR/ES/RxL/Xpand / Direct Method	10	78.64	4.32																							



AG: All Group		142	170.88	8.52	128.00-214.00	4.97	133	127.08	5.99	95.00-159.00	4.72	133	93.34	5.72	70.00-116.00	6.13	132	160.73	8.39	121.00-201.00	5.22	133	191.99	11.18	144.00-240.00	5.83
Amylase	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: Abaxis Piccolo (Waived)		36	132.65	7.44	93.00-173.00	5.56																				
PG: Abbott Architect		6	155.00	1.41	109.00-202.00	0.90	6	90.50	1.05	64.00-118.00	1.10	6	40.67	1.03	29.00-53.00	2.44	6	141.17	1.47	99.00-183.00	1.06	6	187.17	2.04	131.00-243.00	1.07
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent		12	146.43	5.98	102.00-190.00	4.11	12	86.50	3.44	61.00-113.00	3.91	11	40.23	2.56	28.00-52.00	6.50	11	133.46	6.54	93.00-173.00	4.89	11	176.92	10.76	124.00-230.00	6.10
PG: Siemens Dimension/AR/ES/RxL/Xpand		8	152.20	2.25	106.00-198.00	1.51	8	89.00	1.76	62.00-116.00	2.02	8	39.10	1.60	27.00-51.00	4.10	8	138.80	2.25	97.00-181.00	1.65	8	183.70	2.36	129.00-239.00	1.30
PG: Beckman (Olympus) AU		7	112.00	5.03	78.00-146.00	4.46	7	65.71	2.29	46.00-86.00	3.48	7	29.29	1.38	20.00-38.00	4.83	7	102.00	4.55	71.00-133.00	4.41	7	135.14	5.70	95.00-176.00	4.22
PG: OCD Vitros 250,350,400,500,750,950		6	68.71	6.13	48.00-90.00	8.84	5	35.17	6.27	25.00-46.00	18.00	6	30.00	0.00	21.00-39.00	0.00	6	64.29	6.08	45.00-83.00	9.53	6	88.71	7.04	62.00-116.00	7.87
SG: Abaxis Piccolo Instruments		38	132.64	7.28	93.00-173.00	5.49																				
SG: Select Abbott Architect Instruments		9	155.08	2.06	109.00-202.00	1.35	9	90.50	1.58	64.00-118.00	1.76	9	40.50	1.08	29.00-53.00	2.68	9	141.00	2.45	99.00-183.00	1.70	9	187.50	2.80	132.00-244.00	1.49
SG: International OCD Vitros Instruments		8	66.30	6.43	46.00-86.00	9.70	7	35.22	5.47	25.00-46.00	15.71	8	30.00	0.00	21.00-39.00	0.00	8	63.10	5.32	44.00-82.00	8.41	8	87.00	6.38	61.00-113.00	7.36
SG: International Abaxis Piccolo Instruments		39	132.45	7.29	92.00-172.00	5.53																				
SG: Select Gen Chem Instruments		19	140.80	20.59	99.00-183.00	14.61	19	81.74	12.36	57.00-107.00	15.12	19	36.29	5.45	25.00-47.00	15.28	19	126.82	19.00	89.00-165.00	14.96	19	168.53	25.71	118.00-220.00	15.21
AG: All Group		79	139.39	13.50	97.00-181.00	9.71	45	83.62	10.14	59.00-109.00	12.02	49	37.03	4.75	26.00-48.00	12.97	45	128.17	18.03	90.00-166.00	14.06	44	171.64	21.46	120.00-224.00	12.50
Creatine Kinase	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: Abaxis Piccolo (Waived)		26	308.37	28.58	216.00-400.00	9.29																				
PG: Abbott Architect		5	323.43	29.05	226.00-420.00	9.01	5	175.60	15.47	123.00-229.00	8.81	5	62.60	4.28	44.00-82.00	6.83	5	290.80	30.23	204.00-378.00	10.38	5	389.80	43.60	273.00-507.00	11.18
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent		12	287.67	7.77	202.00-374.00	2.71	12	163.17	5.41	114.00-212.00	3.31	11	62.36	2.94	43.00-81.00	4.68	11	261.55	7.75	183.00-341.00	2.98	11	347.82	12.62	244.00-452.00	3.62
PG: Siemens Dimension/AR/ES/RxL/Xpand		5	319.25	3.77	223.00-415.00	1.19	5	173.75	1.71	122.00-226.00	0.98	5	60.50	0.58	43.00-79.00	0.98	5	290.75	6.70	204.00-378.00	2.30	5	390.25	9.29	273.00-507.00	2.38
PG: Siemens Dimension/AR/ES/RxL/Xpand / IFCC		5	305.00	4.00	214.00-397.00	1.31	5	169.00	6.21	118.00-220.00	3.67	5	63.20	5.54	44.00-82.00	8.73	5	274.00	7.71	192.00-356.00	2.81	5	368.40	8.32	258.00-478.00	2.26
PG: Beckman (Olympus) AU		7	276.57	14.55	194.00-360.00	5.23	7	151.43	6.60	106.00-196.00	4.37	7	53.71	2.36	38.00-70.00	4.44	7	248.71	10.83	174.00-324.00	4.34	7	338.14	14.95	237.00-439.00	4.41
PG: Roche Cobas Integra / Roche Reagent		6	323.00	12.28	226.00-420.00	3.81	6	179.33	7.47	125.00-233.00	4.19	6	65.17	4.02	46.00-85.00	6.15	6	293.00	9.90	205.00-381.00	3.38	6	387.83	9.87	272.00-504.00	2.55
PG: OCD Vitros 250,350,400,500,750,950		5	258.40	18.02	181.00-335.00	6.98	5	160.50	13.03	113.00-209.00	8.07	5	59.50	6.95	42.00-78.00	11.67	5	243.25	14.32	170.00-316.00	5.88	5	302.00	19.55	211.00-393.00	6.46
SG: Abaxis Piccolo Instruments		27	308.93	28.20	216.00-402.00	9.13	5	191.00	0.00	134.00-248.00	0.00	5	88.00	0.00	62.00-114.00	0.00	5	290.00	0.00	203.00-377.00	0.00	5	388.00	0.00	272.00-504.00	0.00
SG: Select Abbott Architect Instruments		9	326.50	22.22	229.00-425.00	6.79	9	178.30	11.03	125.00-231.00	6.18	9	63.60	3.34	45.00-83.00	5.16	9	293.80	20.73	206.00-382.00	7.04	9	398.40	31.00	279.00-517.00	7.79
SG: International OCD Vitros Instruments		7	258.00	18.28	181.00-335.00	7.09	7	155.88	11.87	109.00-203.00	7.63	7	58.75	5.12	41.00-77.00	8.64	7	239.88	14.60	168.00-312.00	6.08	7	299.88	23.44	210.00-390.00	7.80
SG: International Abaxis Piccolo Instruments		30	313.45	30.40	219.00-407.00	9.71	5	202.50	11.45	142.00-264.00	5.62	5	94.00	8.29	66.00-122.00	8.83	5	309.25	16.76	216.00-402.00	5.44	5	408.50	18.38	286.00-532.00	4.50
AG: All Group		80	303.45	27.37	212.00-394.00	9.04	55	168.98	13.41	118.00-220.00	7.93	53	62.61	7.97	44.00-82.00	12.70	53	273.07	22.12	191.00-355.00	8.10	53	365.25	34.24	256.00-475.00	9.37
GGT	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: Abaxis Piccolo (Waived)		36	137.65	3.71	127.00-149.00	2.68																				
PG: Abbott Architect		5	156.71	4.57	143.00-171.00	2.93	5	90.00	3.00	81.00-99.00	3.33	5	39.60	0.89	37.00-43.00	2.25	5	142.20	5.07	127.00-157.00	3.59	5	187.60	6.03	170.00-206.00	3.19
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent		6	109.33	3.93	97.00-121.00	3.58	6	66.50	2.26	60.00-74.00	3.43	6	32.17	0.75	30.00-34.00	2.50	6	98.67	3.50	89.00-110.00	3.54	6	130.00	4.43	117.00-143.00	3.38
PG: Siemens Dimension/AR/ES/RxL/Xpand		5	176.67	1.03	174.00-180.00	0.56	5	106.00	1.10	103.00-109.00	1.04	5	51.00	1.67	46.00-56.00	3.33	5	162.50	2.59	155.00-171.00	1.60	5	210.67	1.97	205.00-217.00	0.95
PG: Beckman (Olympus) AU		6	122.67	7.29	101.00-145.00	5.93	6	71.67	4.63	58.00-86.00	6.39	6	31.67	1.86	26.00-38.00	5.94	6	110.83	6.43	92.00-130.00	5.77	6	147.33	8.64	121.00-173.00	5.85
SG: Abaxis Piccolo Instruments		38	137.33	3.87	125.00-149.00	2.85																				
SG: Select Abbott Architect Instruments		6	156.50	4.28	144.00-170.00	2.74	6	90.00	2.68	82.00-98.00	3.00	6	39.83	0.98	37.00-43.00	2.50	6	142.17	4.54	129.00-156.00	3.17	6	187.67	5.39	172.00-204.00	2.87
PG: Intl OCD Vitros 250,350,400,500,750,950		5	200.75	7.63	178.00-224.00	3.78																				
AG: All Group		66	140.73	18.49	86.00-197.00	13.12	30	85.45	16.84	35.00-135.00	19.76	30	39.23	8.13	15.00-63.00	20.77	30	130.97	25.72	54.00-208.00	19.62	30	171.84	31.97	76.00-268.00	18.60
LDH	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: Abbott Architect		5	467.25	8.50	374.00-560.00	1.82	5	269.00	6.05	215.00-323.00	2.27	5	118.75	6.55	95.00-143.00	5.55	5	424.50	7.94	340.00-510.00	1.86	5	563.75	14.80	451.00-677.00	2.62
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent		5	330.50	20.42	265.00-397.00	6.16	5	203.50	6.46	163.00-245.00	3.19	5	97.00	3.56	78.00-116.00	3.71	5	312.75	13.15	250.00-376.00	4.19	5	401.25	11.30	321.00-481.00	2.82
PG: OCD Vitros 250,350,400,500,750,950		5	1506.00	0.00	1205.00-1807.00	0.00	5	920.00	0.00	736.00-1104.00	0.00	5	421.00	0.00	337.00-505.00	0.00	5	1454.00	0.00	1163.00-1745.00	0.00	5	1879.00	0.00	1503.00-2255.00	0.00
SG: International OCD Vitros Instruments		5	1468.80	27.19	1175.00-1763.00	1.85	5	886.20	22.02	709.00-1063.00	2.48	5	408.00	7.55	326.00-490.00	1.84	5	1367.20								



Magnesium	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Abbott Architect	8	3.68	0.10	2.80-4.60	2.70	8	2.54	0.12	1.90-3.10	4.80	8	1.63	0.09	1.20-2.00	5.63	8	3.44	0.07	2.50-4.30	2.06	8	4.21	0.15	3.20-5.30	3.57
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	13	3.32	0.34	2.50-4.10	10.30	13	2.23	0.27	1.70-2.80	12.27	12	1.59	0.14	1.20-2.00	8.75	13	3.13	0.41	2.30-3.90	13.23	13	3.86	0.45	2.90-4.90	11.54
PG: Clinical Data Envoy 500	5	3.57	0.15	2.70-4.50	4.17	5	2.60	0.16	2.00-3.30	5.77	5	1.67	0.10	1.30-2.10	5.88	5	3.37	0.27	2.50-4.30	7.94	5	4.32	0.24	3.20-5.40	5.58
PG: Siemens Dimension/AR/ES/RxL/Xpand	6	3.73	0.15	2.80-4.60	4.05	6	2.57	0.14	2.00-3.30	5.38	6	1.66	0.13	1.30-2.10	7.65	5	3.48	0.17	2.60-4.40	4.86	6	4.30	0.20	3.20-5.40	4.65
PG: Beckman (Olympus) AU	7	3.61	0.11	2.70-4.50	3.06	7	2.54	0.08	1.90-3.10	3.20	7	1.70	0.06	1.30-2.10	3.53	7	3.39	0.15	2.50-4.30	4.41	7	4.19	0.15	3.20-5.30	3.57
PG: OCD Vitros 250,350,400,500,750,950	5	3.90	0.08	2.90-4.90	2.05	5	2.77	0.06	2.10-3.50	2.14	5	1.87	0.06	1.40-2.40	3.16	5	3.63	0.15	2.70-4.50	4.17	5	4.43	0.15	3.30-5.50	3.41
SG: Select Abbott Architect Instruments	10	3.71	0.13	2.80-4.60	3.51	10	2.57	0.12	2.00-3.30	4.62	10	1.64	0.08	1.20-2.00	5.00	10	3.47	0.09	2.60-4.40	2.57	10	4.24	0.15	3.20-5.30	3.57
SG: International OCD Vitros Instruments	5	3.91	0.09	2.90-4.90	2.31	5	2.77	0.05	2.10-3.50	1.79	5	1.88	0.04	1.40-2.40	2.11	5	3.67	0.10	2.80-4.60	2.70	5	4.48	0.13	3.40-5.60	2.89
SG: International Abaxis Piccolo Instruments	5	3.88	0.03	2.90-4.90	0.77	5	2.64	0.04	2.00-3.30	1.54	5	1.69	0.02	1.30-2.10	1.18	5	3.60	0.06	2.70-4.50	1.67	5	4.52	0.12	3.40-5.60	2.44
AG: All Group	48	3.58	0.27	2.70-4.50	7.50	48	2.48	0.24	1.90-3.10	9.60	47	1.67	0.17	1.30-2.10	10.00	47	3.35	0.29	2.50-4.10	8.79	48	4.14	0.34	3.10-5.10	8.29
Uric Acid	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Abaxis Piccolo (Waived)	7	7.73	0.17	6.40-9.00	2.21																				
PG: Abbott Architect	9	9.43	0.09	7.80-11.00	0.96	9	5.82	0.12	4.80-6.80	2.07	9	3.03	0.12	2.50-3.50	4.00	9	8.66	0.13	7.20-10.20	1.49	9	11.16	0.11	9.30-13.10	0.98
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	23	8.58	0.37	7.10-10.10	4.30	23	5.59	0.38	4.60-6.60	6.79	22	3.54	0.24	2.90-4.10	6.86	21	7.88	0.39	6.60-9.20	4.94	22	10.19	0.43	8.50-11.90	4.22
PG: Clinical Data Envoy 500	6	8.86	0.09	7.40-10.40	1.01	6	5.40	0.08	4.50-6.30	1.48	6	2.80	0.05	2.30-3.30	1.79	6	8.13	0.09	6.70-9.50	1.11	6	10.56	0.09	8.80-12.40	0.85
PG: Siemens Dimension/AR/ES/RxL/Xpand	14	8.39	0.12	7.00-9.80	1.31	14	5.15	0.08	4.30-6.10	1.54	14	2.64	0.10	2.20-3.00	3.85	14	7.68	0.14	6.40-9.00	1.82	14	9.98	0.23	8.30-11.70	2.30
PG: Horiba ABX Pentra 400	5	8.18	0.25	6.80-9.60	3.05	5	5.13	0.10	4.20-6.00	1.96	5	2.60	0.08	2.20-3.00	3.08	5	7.63	0.13	6.30-8.90	1.71	5	9.73	0.33	8.10-11.30	3.40
PG: Beckman (Olympus) AU	8	8.66	0.31	7.20-10.20	3.56	8	5.38	0.18	4.50-6.30	3.33	8	2.78	0.09	2.30-3.30	3.21	8	7.96	0.25	6.60-9.40	3.13	8	10.33	0.37	8.50-12.10	3.59
PG: Roche Cobas Integra / Roche Reagent	5	8.62	0.24	7.10-10.10	2.79	5	5.30	0.12	4.40-6.20	2.26	5	2.66	0.09	2.20-3.20	3.33	5	7.86	0.21	6.60-9.20	2.66	5	10.12	0.26	8.40-11.80	2.57
PG: OCD Vitros 250,350,400,500,750,950	7	8.60	0.13	7.10-10.10	1.51	7	5.26	0.10	4.40-6.20	1.89	7	2.61	0.09	2.20-3.00	3.46	7	7.86	0.16	6.60-9.20	2.03	7	10.33	0.10	8.50-12.10	0.97
SG: Abaxis Piccolo Instruments	8	7.68	0.22	6.40-9.00	2.86																				
SG: Select Abbott Architect Instruments	13	9.43	0.11	7.80-11.00	1.17	13	5.84	0.10	4.80-6.80	1.72	13	3.03	0.10	2.50-3.50	3.33	13	8.65	0.13	7.20-10.20	1.49	13	11.17	0.14	9.30-13.10	1.25
SG: International OCD Vitros Instruments	10	8.63	0.14	7.10-10.10	1.63	10	5.27	0.09	4.40-6.20	1.70	10	2.63	0.09	2.20-3.00	3.46	10	7.88	0.15	6.60-9.20	1.90	10	10.41	0.14	8.60-12.20	1.35
AG: All Group	92	8.63	0.50	7.10-10.10	5.81	85	5.43	0.34	4.50-6.30	6.30	84	2.95	0.44	2.50-3.50	14.67	83	7.96	0.41	6.60-9.40	5.13	84	10.31	0.55	8.50-12.10	5.34
Iron	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Abbott Architect	8	180.40	7.63	144.00-216.00	4.22	8	113.63	4.72	91.00-137.00	4.12	8	61.38	3.20	49.00-73.00	5.25	8	166.88	6.69	134.00-200.00	4.01	8	216.38	8.38	173.00-259.00	3.89
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	11	170.08	4.87	136.00-204.00	2.88	11	106.00	2.92	85.00-127.00	2.74	11	56.46	3.26	45.00-67.00	5.89	11	154.77	4.17	124.00-186.00	2.71	11	202.15	2.58	162.00-242.00	1.29
PG: Clinical Data Envoy 500	6	170.24	7.39	136.00-204.00	4.35	6	105.65	3.78	85.00-127.00	3.58	6	57.96	1.69	46.00-70.00	2.93	6	155.84	5.30	125.00-187.00	3.40	6	204.20	8.66	163.00-245.00	4.26
PG: Siemens Dimension/AR/ES/RxL/Xpand	5	177.83	3.25	142.00-214.00	1.85	5	112.33	1.63	90.00-134.00	1.43	5	60.50	1.52	49.00-73.00	2.46	5	163.33	2.25	130.00-196.00	1.41	5	211.50	3.27	170.00-254.00	1.56
PG: Beckman (Olympus) AU	5	200.20	5.54	160.00-240.00	2.75	5	122.40	3.65	98.00-146.00	2.95	5	64.60	1.34	52.00-78.00	2.00	5	181.60	4.83	146.00-218.00	2.64	5	238.40	5.68	190.00-286.00	2.39
PG: Roche Cobas Integra / Roche Reagent	5	176.33	0.58	141.00-211.00	0.34	5	110.00	2.65	88.00-132.00	2.36	5	58.33	3.79	46.00-70.00	6.55	5	165.00	3.46	132.00-198.00	2.12	5	210.00	6.93	168.00-252.00	3.29
SG: Select Abbott Architect Instruments	11	181.00	7.20	145.00-217.00	3.98	11	113.67	4.64	91.00-137.00	4.04	11	60.92	3.37	49.00-73.00	5.57	11	166.83	6.06	134.00-200.00	3.65	11	216.42	7.72	173.00-259.00	3.56
AG: All Group	46	179.80	14.65	144.00-216.00	8.17	46	111.66	8.12	90.00-134.00	7.23	46	59.71	4.58	48.00-72.00	7.67	46	165.05	13.51	132.00-198.00	8.18	46	214.78	17.57	172.00-258.00	8.19
TIBC	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	6	345.63	11.66	310.60-380.60	3.37	6	266.75	9.27	239.00-294.60	3.47	6	200.88	12.25	164.20-237.70	6.10	6	331.13	13.03	292.00-370.20	3.94	6	384.25	17.38	332.00-436.40	4.52
PG: Siemens Dimension/AR/ES/RxL/Xpand	5	328.83	8.01	304.80-352.80	2.44	5	210.17	6.74	190.00-230.40	3.21	5	118.00	4.86	103.40-132.60	4.12	5	300.83	4.40	287.60-314.00	1.46	5	383.00	5.29	367.10-398.90	1.38
PG: Intl OCD Vitros 250,350,400,500,750,950	5	586.00	55.30	420.10-751.90	9.44	5	390.75	29.62	301.90-479.70	7.58	5	203.00	21.26	139.20-266.80	10.47	5	569.00	41.01	446.00-692.00	7.21	5	706.50	65.24	510.80-902.20	9.23
SG: International OCD Vitros Instruments	5	586.00	55.30	420.10-751.90	9.44	5	390.75	29.62	301.90-479.70	7.58	5	203.00	21.26	139.20-266.80	10.47	5	569.00	41.01	446.00-692.00	7.21	5	706.50	65.24	510.80-902.20	9.23
AG: All Group	24	359.41	95.16	73.90-644.90	26.48	24	246.43	68.30	41.50-451.30	27.72	24	159.50	53.32	0.00-319.50	33.43	24	332.04	89.88	62.40-601.60	27.07	24	409.57	112.40	72.40-746.80	27.44
TSH	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Abbott Architect	14	7.89	0.50	6.38-9.40	6.39	14	5.23	0.37	4.11-6.35	7.13	14	2.82	0.20	2.21-3.43	7.23	14	7.26	0.58	5.52-9.00	8.00	14	8.67	0.59	6.90-10.44	6.82
PG: Beckman ACCESS	20	7.91	0.37	6.79-9.03	4.70	20	5.18	0.22	4.51-5.85	4.32	20	2.76	0.14	2.35-3.17	4.93	20	7.20	0.30	6.29-8.11	4.22	20	8.58	0.38	7.45-9.71	4.41
PG: bioMerieux Vitek, Mini Vidas	21	9.23	0.53	7.65-10.81	5.70	21	5.99	0.33	5.01-6.97	5.48	21	3.11	0.21	2.49-3.73	6.66	21	8.48	0.52	6.91-10.05	6.17	21	10.47	0.65	8.51-12.43	6.23
PG: Siemens Dimension/AR/ES/RxL/Xpand	16	3.67	0.36	2.60-4.74	9.73	16	2.38	0.24	1.67-3.09	9.96	16	1.24	0.12	0.89-1.59	9.35	16	3.41	0.34	2.38-4.44	10.03	16	4.41	0.47	3.00-5.82	10.66
PG: Qualigen FastPak Thyroid	28	6.68	0.96	3.81-9.55	14.31	28	4.44	0.66	2.45-6.43	14.95	28	2.22	0.34	1.20-3.24	15.27	28	6.23	0.85	3.69-8.77	13.61	28	7.64	0.95	4.79-10.49	12.43
PG: TOSOH AIA	19	10.47	0.85	7.93-13.01	8.08	20	6.78	0.54	5.17-8.39	7.92	20	3.62	0.23	2.93-4.31	6.33	20	9.58	0.63	7.68-11.48	6.62	20	11.78	0.82	9.33-14.23	6.93
SG: Select Beckman Coulter Instruments (Chem)	21	7.90	0.36	6.81-8.99	4.61	21	5.18	0.22	4.52-5.84	4.23	21	2.76	0.14	2.35-3.17	4.89	21	7.19	0.30	6.29-8.09	4.17	21	8			



AG: All Group	103	3.84	0.81	1.39-6.29	21.23	103	3.26	0.60	1.46-5.07	18.46	103	2.74	0.58	0.99-4.50	21.34	103	3.70	0.78	1.36-6.03	21.04	103	4.08	0.93	1.29-6.86	22.78
Thyroxine (T4)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman ACCESS	6	13.80	0.91	11.00-16.60	6.59	6	10.87	0.72	8.70-13.10	6.61	6	7.93	0.73	6.30-9.50	9.24	6	13.18	1.30	10.60-15.80	9.85	6	15.48	1.45	12.40-18.60	9.35
PG: Siemens Dimension/AR/ES/RxL/Xpand	5	15.18	0.70	12.20-18.20	4.61	5	10.35	0.48	8.30-12.50	4.62	5	6.55	0.53	5.30-7.90	8.03	5	14.15	0.70	11.40-17.00	4.93	5	17.45	0.48	14.00-21.00	2.74
PG: TOSOH AIA	7	13.87	0.76	11.10-16.70	5.47	7	10.17	0.72	8.20-12.20	7.06	7	7.29	0.39	5.80-8.80	5.34	7	13.47	0.62	10.80-16.20	4.59	7	15.80	1.61	12.60-19.00	10.19
SG: Select Beckman Coulter Instruments (Chem)	7	13.83	0.84	11.00-16.60	6.09	7	10.72	0.77	8.60-12.80	7.20	7	7.74	0.83	6.20-9.20	10.91	7	13.26	1.20	10.60-16.00	9.02	7	15.58	1.35	12.50-18.70	8.65
SG: Select Abbott Architect Instruments	6	16.84	1.08	13.40-20.20	6.43	6	11.32	0.53	9.00-13.60	4.69	6	6.80	0.34	5.40-8.20	4.85	6	15.47	0.73	12.40-18.60	4.71	6	19.38	1.52	15.50-23.30	7.89
AG: All Group	31	14.55	1.89	11.60-17.40	13.03	31	10.50	0.76	8.40-12.60	7.24	31	7.17	0.75	5.80-8.60	10.42	31	13.69	1.47	11.00-16.40	10.73	31	16.47	2.43	13.20-19.80	14.73
Triiodothyronine (T3)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman ACCESS	6	2.28	0.14	1.90-2.70	6.09	6	1.58	0.15	1.20-2.00	9.38	6	1.18	0.13	0.80-1.60	10.00	6	2.10	0.16	1.60-2.60	7.62	6	2.76	0.20	2.20-3.40	7.14
PG: TOSOH AIA	5	4.40	0.15	4.00-4.90	3.41	5	3.49	0.02	3.40-3.60	0.57	5	2.84	0.05	2.60-2.90	1.79	5	4.18	0.11	3.90-4.50	2.62	5	4.74	0.05	4.50-4.90	1.06
SG: Select Beckman Coulter Instruments (Chem)	8	2.33	0.18	1.80-2.80	7.83	8	1.61	0.14	1.20-2.00	8.75	8	1.20	0.11	0.90-1.50	9.17	8	2.11	0.17	1.60-2.60	8.10	8	2.75	0.20	2.10-3.30	7.41
SG: Select Abbott Architect Instruments	5	1.55	0.13	1.20-2.00	7.50	5	1.28	0.16	0.80-1.80	12.31	5	1.22	0.21	0.60-1.80	17.50	5	1.50	0.07	1.30-1.70	4.67	5	1.63	0.13	1.20-2.00	7.50
AG: All Group	18	1.95	0.51	0.40-3.40	26.84	18	1.48	0.27	0.70-2.30	18.00	18	1.17	0.18	0.70-1.70	15.00	18	1.81	0.41	0.60-3.00	22.78	18	2.21	0.65	0.30-4.20	29.55
T3 Uptake	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: TOSOH AIA	5	0.40	0.01	0.36-0.44	3.14	5	0.44	0.02	0.39-0.49	3.90	5	0.49	0.01	0.46-0.53	2.47	5	0.41	0.01	0.38-0.43	2.05	5	0.39	0.01	0.36-0.41	1.85
SG: Select Beckman Coulter Instruments (Chem)	5	0.46	0.05	0.30-0.62	11.63	5	0.51	0.07	0.30-0.72	13.68	5	0.54	0.07	0.33-0.74	12.99	5	0.48	0.06	0.30-0.66	12.25	5	0.44	0.03	0.34-0.54	7.67
PG: Abbott Architect i1000 SR	5	0.01	0.00	0.01-0.01	0.00	5	0.00	0.00	0.00-0.00	0.00	5	0.00	0.00	0.00-0.00	0.00	5	0.01	0.00	0.01-0.01	0.00	5	0.01	0.00	0.01-0.01	0.00
AG: All Group	11	0.41	0.05	0.26-0.57	12.26	11	0.45	0.06	0.28-0.63	12.62	11	0.49	0.06	0.31-0.67	12.02	11	0.42	0.06	0.26-0.59	13.06	11	0.39	0.07	0.19-0.59	17.13
Free T3	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Architect	5	5.00	0.16	4.50-5.50	3.20	5	4.40	0.17	3.90-4.90	3.86	5	4.26	0.27	3.50-5.10	6.28	5	4.88	0.29	4.00-5.80	5.92	5	5.50	0.25	4.80-6.20	4.36
PG: Beckman ACCESS	5	5.30	0.15	4.80-5.80	2.83	5	4.45	0.12	4.00-4.80	2.73	5	3.60	0.14	3.20-4.00	3.89	5	5.08	0.17	4.60-5.60	3.33	5	5.50	0.20	4.90-6.10	3.64
PG: TOSOH AIA	5	8.32	2.35	1.30-15.40	28.31	5	7.79	0.08	7.60-8.00	1.03	5	6.39	0.09	6.10-6.70	1.41	5	9.34	0.16	8.80-9.80	1.72	5	10.46	0.67	8.50-12.50	6.38
SG: Select Beckman Coulter Instruments (Chem)	6	5.29	0.14	4.90-5.70	2.64	6	4.41	0.14	4.00-4.90	3.41	6	3.60	0.13	3.20-4.00	3.33	6	5.03	0.19	4.40-5.60	3.80	6	5.49	0.18	5.00-6.00	3.27
SG: Select Abbott Architect Instruments	10	5.04	0.19	4.40-5.60	3.80	10	4.44	0.21	3.80-5.00	4.77	10	4.18	0.28	3.40-5.00	6.67	10	4.91	0.28	4.10-5.70	5.71	10	5.87	1.17	2.40-9.40	19.83
PG: Intl OCD Vitros ECI/ECIQ	5	17.83	1.16	14.40-21.30	6.46	5	16.23	1.07	13.00-19.40	6.60	5	14.47	1.20	10.90-18.10	8.28	5	17.33	1.33	13.30-21.30	7.69	5	18.40	1.25	14.60-22.10	6.79
AG: All Group	20	5.20	0.57	3.50-6.90	10.96	20	4.51	0.47	3.10-5.90	10.44	20	3.97	0.49	2.50-5.50	12.25	20	5.06	0.56	3.40-6.80	10.98	20	5.83	0.99	2.80-8.80	17.07
Cortisol	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	10	26.36	2.69	19.80-33.00	10.19	10	17.40	1.43	13.00-21.80	8.22	10	8.72	0.70	6.50-10.90	8.05	10	24.57	2.12	18.50-30.80	8.62	10	31.46	2.64	23.60-39.40	8.38
UIBC	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	5	230.00	12.12	193.60-266.40	5.27	5	158.00	10.58	126.30-189.70	6.70	5	104.00	1.00	101.00-107.00	0.96	5	215.33	7.51	192.80-237.80	3.49	5	266.00	8.89	239.30-292.70	3.34
Digoxin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	5	2.18	0.21	1.80-2.60	9.55	5	1.42	0.16	1.10-1.70	10.71	5	1.01	0.15	0.80-1.20	15.00	5	1.98	0.22	1.60-2.40	11.00	5	2.45	0.42	1.90-2.90	17.50
Phenytoin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	5	25.28	2.28	19.00-31.60	9.01	5	15.54	1.76	11.60-19.40	11.35	5	9.20	1.14	6.90-11.50	12.39	5	23.10	2.49	17.30-28.90	10.78	5	29.40	4.39	22.00-36.80	14.93
Valproic Acid	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	5	116.92	7.55	88.00-146.00	6.50	5	81.99	4.46	62.00-103.00	5.49	5	55.85	4.19	42.00-70.00	7.50	5	108.06	4.65	81.00-135.00	4.35	5	131.83	9.80	99.00-165.00	7.42
746 - Auto Diff 6 (5-Part Diff)	CL-6					CL-7					CL-8					CL-9					CL-10				
White Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Abbott Cell Dyn Instruments (5-part Diff)	8	21.66	0.46	18.40-25.00	2.12	8	3.28	0.09	2.80-3.80	2.73	8	8.09	0.16	6.90-9.30	1.98	8	8.11	0.18	6.90-9.30	2.22	8	21.45	0.46	18.30-24.70	2.14
PG: Abbott Cell-Dyn Ruby	7	21.71	0.47	18.40-25.00	2.17	7	3.29	0.09	2.80-3.80	2.73	7	8.07	0.16	6.90-9.30	1.98	7	8.10	0.19	6.90-9.30	2.35	7	21.44	0.50	18.20-24.60	2.34
Lymphocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Abbott Cell Dyn Instruments (5-part Diff)	8	17.96	0.77	15.70-20.30	4.28	8	38.60	1.66	33.60-43.60	4.30	8	28.08	0.68	26.10-30.10	2.42	8	27.67	0.60	25.90-29.50	2.17	8	17.90	0.89	15.20-20.60	4.97
PG: Abbott Cell-Dyn Ruby	7	17.81	0.70	15.70-19.90	3.93	7	38.39	1.67	33.40-43.40	4.35	7	28.09	0.73	25.90-30.30	2.60	7	27.69	0.65	25.80-29.60	2.35	7	18.00	0.92	15.30-20.70	5.06
Monocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Abbott Cell Dyn Instruments (5-part Diff)	8	4.60	0.41	3.37-5.83	8.93	8	6.80	1.35	2.76-10.84	19.79	8	6.64	0.35	5.57-7.71	5.35	8	6.50	0.65	4.56-8.44	9.95	8	4.84	0.28	4.01-5.67	5.74
PG: Abbott Cell-Dyn Ruby	7	4.55	0.42	3.28-5.82	9.32	7	7.09	1.13	3.70-10.48	15.94	7														



SG: Select Abbott Cell Dyn Instruments (5-part Diff)	8	11.76	0.27	11.00-12.60	2.29	8	14.86	0.51	13.40-16.40	3.42	8	12.20	0.30	11.30-13.10	2.46	8	12.13	0.34	11.10-13.10	2.81	8	11.68	0.20	11.10-12.30	1.79	
PG: Abbott Cell-Dyn Ruby	7	11.83	0.21	11.20-12.40	1.78	7	14.97	0.44	13.70-16.30	2.93	7	12.24	0.29	11.30-13.10	2.38	7	12.20	0.28	11.40-13.00	2.30	7	11.71	0.19	11.10-12.30	1.62	
Platelet Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
SG: Select Abbott Cell Dyn Instruments (5-part Diff)	8	539.50	19.15	405.00-675.00	3.56	8	88.00	6.55	66.00-110.00	7.39	8	286.38	16.45	215.00-358.00	5.73	8	287.63	15.64	216.00-360.00	5.42	8	537.25	35.54	403.00-671.00	6.61	
PG: Abbott Cell-Dyn Ruby	7	544.57	13.71	409.00-681.00	2.51	7	89.57	5.19	68.00-113.00	5.78	7	289.29	15.38	217.00-361.00	5.33	7	291.29	12.66	218.00-364.00	4.36	7	545.57	28.76	410.00-683.00	5.27	
683 - ASO/CRP:684 - CRP (Quant & Qual)	CR-2N					CR-4N																				
CRP Quantitative	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
600 - Cholestech LDX/CardioChek Pkg	CT-2																									
Cholesterol Total (Waived)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: Cholestech LDX (Waived)	146	214.99	12.45	178.00-253.00	5.81																					
PG: Polymer CardioChek	29	221.62	31.99	126.00-318.00	14.41																					
Glucose (Waived)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: Cholestech LDX (Waived)	29	213.66	10.48	183.00-246.00	4.91																					
AG: All Group	30	214.40	11.07	181.00-247.00	5.19																					
Triglycerides (Waived)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: Cholestech LDX (Waived)	132	180.56	7.88	157.00-205.00	4.36																					
PG: Polymer CardioChek	26	176.23	23.53	106.00-247.00	13.35																					
Cholesterol HDL (Waived)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: Cholestech LDX (Waived)	138	90.69	7.54	69.00-114.00	8.24																					
PG: Polymer CardioChek / Direct Method	29	101.62	12.17	65.00-139.00	11.96																					
Cholesterol LDL (Waived)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: Cholestech LDX (Waived)	53	88.62	16.57	39.00-139.00	18.65																					
PG: Polymer CardioChek / Direct Method	8	105.88	21.11	43.00-169.00	19.91																					
763 - CoaguChek XS Plus/XS Pro (2 specimens);764 - CoaguChek XS Plus/XS Pro (5 specimens)	CXP-6					CXP-7					CXP-8					CXP-9					CXP-10					
Prottime - Whole Blood	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: CoaguChek XS Plus	16	14.44	0.36	12.20-16.60	2.50	16	31.70	1.56	26.90-36.50	4.92	13	14.40	0.24	12.20-16.60	1.67	13	31.16	1.59	26.50-35.90	5.10	13	14.43	0.30	12.20-16.60	2.08	
PG: CoaguChek XS Pro	12	14.21	0.37	12.10-16.30	2.61	12	32.01	1.43	27.20-36.80	4.47	11	14.31	0.54	12.20-16.40	3.78	11	31.81	0.98	27.00-36.60	3.08	11	14.15	0.40	12.00-16.20	2.84	
Prottime W Blood - Waived	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: CoaguChek XS Plus	15	14.14	0.72	11.90-16.30	5.11	14	32.08	1.52	27.50-36.70	4.74																
INR	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: CoaguChek XS Plus	83	1.20	0.09	0.90-1.50	7.50	82	2.64	0.17	2.10-3.10	6.54	56	1.20	0.04	1.10-1.30	3.33	56	2.62	0.16	2.20-3.00	5.77	56	1.20	0.03	1.10-1.30	2.50	
PG: CoaguChek XS Pro	12	1.18	0.04	1.10-1.30	3.33	12	2.66	0.12	2.30-3.10	4.44	11	1.25	0.13	0.80-1.60	10.83	11	2.65	0.09	2.40-3.00	3.33	11	1.17	0.05	1.00-1.40	4.17	
745 - Auto Diff 5 (5-Part Diff)	DIF-6					DIF-7					DIF-8					DIF-9					DIF-10					
White Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
SG: Select Beckman Coulter Instruments (5-part Diff)	5	21.40	0.58	18.20-24.60	2.71	5	3.88	0.13	3.30-4.50	3.33	5	9.64	0.15	8.20-11.00	1.56	5	9.58	0.18	8.20-11.00	1.88	5	21.80	0.46	18.50-25.10	2.11	
Lymphocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
SG: Select Beckman Coulter Instruments (5-part Diff)	5	21.48	1.76	16.20-26.80	8.19	5	41.58	0.65	39.60-43.60	1.56	5	30.02	0.37	28.90-31.10	1.23	5	30.04	0.87	27.40-32.60	2.90	5	20.86	1.08	17.70-24.10	5.17	
Monocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
SG: Select Beckman Coulter Instruments (5-part Diff)	5	5.82	2.78	0.00-14.17	47.82	5	1.64	0.48	0.19-3.09	29.45	5	4.60	0.48	3.14-6.05	10.54	5	4.48	0.82	2.03-6.93	18.24	5	6.52	3.11	0.00-15.83	47.62	
Neutrophil	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
SG: Select Beckman Coulter Instruments (5-part Diff)	5	68.04	3.90	56.30-79.70	5.74	5	49.16	0.63	47.30-51.10	1.28	5	59.24	0.81	56.80-61.60	1.37	5	59.20	0.81	56.80-61.60	1.37	5	67.62	3.45	57.20-77.90	5.10	
Eosinophil	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
SG: Select Beckman Coulter Instruments (5-part Diff)	5	4.30	0.58	2.60-6.00	13.49	5	7.62	0.44	6.30-8.90	5.79	5	6.08	0.69	4.00-8.20	11.31	5	6.22	0.88	3.60-8.80	14.19	5	4.58	0.54	3.00-6.20	11.74	
Basophil	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
SG: Select Beckman Coulter Instruments (5-part Diff)	5	0.45	0.31	0.00-1.38	69.11						5	0.15	0.07	0.00-0.36	47.33	5	0.15	0.07	0.00-0.36	47.33	5	0.53	0.47	0.00-1.95	89.06	
Red Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
SG: Select Beckman Coulter Instruments (5-part Diff)	5	5.51	0.10	5.18-5.84	1.85	5	2.59	0.04	2.43-2.75	1.62	5	4.25	0.07	4.00-4.50	1.74	5	4.25	0.07	4.00-4.50	1.72	5	5.49	0.06	5.16-5.82	1.02	
Hemoglobin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
SG: Select Beckman Coulter Instruments (5-part Diff)	5	16.24	0.15	15.10-17.30	0.93	5	6.36	0.09	6.00-6.80	1.41	5	11.92	0.15	11.10-12.70	1.26	5	11.88	0.19	11.10-12.70	1.60	5					



AG: All Group	762	18.53	0.52	17.20-19.80	2.81	761	6.05	0.23	5.70-6.50	3.77	763	13.22	0.28	12.30-14.10	2.12	763	13.23	0.32	12.30-14.10	2.42	762	18.48	0.90	17.20-19.80	4.86
Lymphocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn 1700	19	11.93	0.63	10.05-13.81	5.26	19	49.98	3.06	40.81-59.15	6.12	19	26.30	1.32	22.35-30.25	5.00	19	26.10	1.44	21.78-30.42	5.52	19	12.04	0.82	9.59-14.49	6.78
PG: Abbott Cell-Dyn 1800	52	10.74	0.62	8.88-12.60	5.78	53	45.13	2.44	37.82-52.44	5.40	53	23.43	1.48	18.98-27.88	6.33	53	23.25	1.41	19.03-27.47	6.05	52	10.72	0.63	8.82-12.62	5.90
PG: CDS Medonic 620	6	14.45	0.35	13.40-15.50	2.43	6	58.00	2.74	49.78-66.22	4.72	6	29.95	1.30	26.04-33.87	4.36	6	31.37	1.66	26.39-36.35	5.29	6	13.92	0.87	11.32-16.52	6.22
PG: Beckman Coulter AcT Diff/Diff 2	221	14.07	0.97	11.16-16.98	6.89	218	59.89	1.62	55.03-64.75	2.71	221	30.53	2.11	24.20-36.86	6.91	221	30.55	2.11	24.22-36.88	6.90	220	14.10	0.92	11.35-16.85	6.50
PG: Horiba ABX Diag. Micros/45/60	72	11.74	0.77	9.42-14.06	6.58	72	43.96	3.48	33.52-54.40	7.91	73	23.26	1.85	17.70-28.82	7.97	72	23.05	1.53	18.45-27.65	6.66	72	11.63	0.73	9.44-13.82	6.28
PG: None	5	10.90	0.00	10.90-10.90	0.00	5	42.30	0.00	42.30-42.30	0.00	5	22.40	0.00	22.40-22.40	0.00	5	21.80	0.00	21.80-21.80	0.00	5	11.10	0.00	11.10-11.10	0.00
SG: Select Abbott Cell Dyn Instruments (3-part Diff)	226	12.82	1.43	8.53-17.11	11.15	233	49.86	3.30	39.97-59.75	6.61	230	25.72	1.75	20.46-30.98	6.82	232	25.76	1.82	20.29-31.23	7.07	224	12.74	1.44	8.42-17.06	11.30
SG: Select Beckman Coulter Instruments (3-part Diff)	223	14.05	1.00	11.04-17.06	7.13	220	59.87	1.72	54.70-65.04	2.88	223	30.50	2.14	24.07-36.93	7.02	223	30.54	2.14	24.12-36.96	7.01	222	14.08	0.95	11.22-16.94	6.78
PG: Drew Scientific D3	24	13.77	0.93	10.97-16.57	6.78	25	54.67	2.27	47.85-61.49	4.16	25	27.69	1.21	24.08-31.30	4.35	25	27.75	1.73	22.56-32.94	6.23	24	13.37	0.69	11.30-15.44	5.16
PG: Abbott Cell-Dyn Emerald	154	13.58	0.87	10.98-16.18	6.38	160	51.30	1.93	45.52-57.08	3.75	157	26.36	1.19	22.78-29.94	4.52	159	26.48	1.17	22.98-29.98	4.40	152	13.47	0.95	10.62-16.32	7.05
PG: CDS M Series	186	13.99	0.95	11.14-16.84	6.79	184	58.59	2.05	52.44-64.74	3.50	186	29.21	2.22	22.55-35.87	7.60	186	29.26	2.21	22.62-35.90	7.56	184	14.04	0.56	12.35-15.73	4.00
SG: CDS Hematology Instruments	192	14.00	0.94	11.18-16.82	6.72	190	58.57	2.07	52.36-64.78	3.53	192	29.23	2.20	22.64-35.82	7.52	192	29.32	2.22	22.65-35.99	7.59	190	14.04	0.57	12.33-15.75	4.07
AG: All Group	739	13.42	1.36	9.34-17.50	10.13	742	54.63	6.08	36.39-72.87	11.13	745	27.87	3.16	18.39-37.35	11.33	746	27.90	3.20	18.29-37.51	11.48	734	13.39	1.32	9.42-17.36	9.89
MCV	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn 1700	20	91.76	1.94	86.00-97.60	2.11	20	72.99	1.45	68.70-77.30	1.99	20	83.13	1.72	77.90-88.30	2.07	20	83.21	1.68	78.20-88.20	2.02	20	91.85	2.02	85.80-98.00	2.20
PG: Abbott Cell-Dyn 1800	52	96.13	2.19	89.50-102.70	2.28	53	75.43	1.66	70.40-80.40	2.20	53	86.37	1.76	81.10-91.70	2.04	53	86.34	1.76	81.00-91.60	2.04	52	95.83	2.02	89.70-101.90	2.11
PG: CDS Medonic 620	6	88.60	3.27	78.80-98.40	3.69	6	69.80	2.25	63.00-76.50	3.22	6	79.03	2.61	71.20-86.80	3.30	6	79.40	2.53	71.80-87.00	3.19	6	88.58	3.06	79.40-97.80	3.45
PG: Beckman Coulter AcT Diff/Diff 2	228	90.76	5.49	74.30-107.30	6.05	227	73.62	3.41	63.40-83.80	4.63	227	82.77	1.05	79.60-86.00	1.27	227	82.50	3.86	70.90-94.10	4.68	227	90.89	4.06	78.70-103.10	4.47
PG: Horiba ABX Diag. Micros/45/60	66	89.03	0.99	86.00-92.00	1.11	66	70.17	2.60	62.40-78.00	3.70	66	80.11	0.84	77.60-82.60	1.05	66	80.02	0.79	77.60-82.40	0.99	66	88.89	0.86	86.30-91.50	0.97
PG: None	5	89.00	0.00	89.00-89.00	0.00	5	70.00	0.00	70.00-70.00	0.00	5	80.00	0.00	80.00-80.00	0.00	5	80.00	0.00	80.00-80.00	0.00	5	89.00	0.00	89.00-89.00	0.00
SG: Select Abbott Cell Dyn Instruments (3-part Diff)	234	96.91	2.50	89.40-104.40	2.58	236	77.74	2.77	69.40-86.00	3.56	235	87.53	2.18	81.00-94.00	2.49	235	87.31	4.19	74.70-99.90	4.80	234	96.81	2.40	89.60-104.00	2.48
SG: Select Beckman Coulter Instruments (3-part Diff)	229	90.76	5.48	74.40-107.20	6.04	228	73.62	3.40	63.40-83.80	4.62	228	82.77	1.04	79.70-85.90	1.26	228	82.50	3.85	71.00-94.00	4.67	228	90.88	4.05	78.80-103.10	4.46
PG: Drew Scientific D3	25	93.87	1.18	90.40-97.40	1.26	25	76.64	0.94	73.80-79.40	1.23	25	85.31	1.00	82.30-88.30	1.17	25	85.43	1.05	82.30-88.60	1.23	25	94.07	1.19	90.50-97.70	1.26
PG: Abbott Cell-Dyn Emerald	161	97.85	1.51	93.30-102.30	1.54	162	79.08	1.87	73.50-84.70	2.36	161	88.49	1.23	84.80-92.20	1.38	161	88.16	4.54	74.60-101.80	5.15	161	97.78	1.38	93.70-101.90	1.41
PG: CDS M Series	186	90.87	1.46	86.50-95.30	1.61	186	71.11	1.80	65.70-76.50	2.53	186	80.76	1.20	77.20-84.40	1.49	186	80.44	5.46	64.00-96.80	6.79	186	90.72	2.01	84.70-96.70	2.22
SG: CDS Hematology Instruments	192	90.80	1.58	86.10-95.50	1.74	192	71.07	1.82	65.60-76.60	2.56	192	80.71	1.29	76.80-84.60	1.60	192	80.41	5.39	64.20-96.60	6.70	192	90.65	2.07	84.50-96.90	2.28
AG: All Group	748	92.70	4.57	79.00-106.40	4.93	749	74.11	3.96	62.20-86.00	5.34	748	83.64	3.30	73.70-93.50	3.95	748	83.41	5.12	68.00-98.80	6.14	747	92.67	4.10	80.40-105.00	4.42
Monocyte/Mixed	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn 1700	19	5.86	0.58	4.20-7.60	9.83	19	14.03	1.97	8.10-19.90	14.07	19	9.16	0.88	6.60-11.80	9.57	19	9.20	0.70	7.10-11.30	7.72	19	5.85	0.48	4.40-7.20	8.28
PG: Abbott Cell-Dyn 1800	51	6.80	0.45	5.40-8.20	6.62	51	18.85	1.57	14.10-23.50	8.35	51	11.10	0.71	9.00-13.20	6.40	51	11.17	0.65	9.30-13.10	5.80	51	6.67	0.41	5.50-7.90	6.12
PG: CDS Medonic 620	6	6.23	1.59	1.40-11.00	25.65	6	9.58	1.81	4.20-15.00	18.85	6	9.78	0.89	7.10-12.50	9.08	6	9.42	1.03	6.30-12.50	10.96	6	5.78	2.11	0.00-12.10	36.38
PG: Beckman Coulter AcT Diff/Diff 2	220	5.04	0.57	3.30-6.70	11.40	220	7.26	1.40	3.10-11.50	19.18	220	6.57	0.89	3.90-9.30	13.48	220	6.58	0.74	4.40-8.80	11.21	219	5.02	0.57	3.30-6.70	11.40
PG: Horiba ABX Diag. Micros/45/60	73	3.63	0.30	2.70-4.50	8.33	74	18.08	2.80	9.70-26.50	15.47	74	8.46	0.79	6.10-10.90	9.29	74	8.57	0.81	6.20-11.00	9.42	74	3.64	0.23	2.90-4.30	6.39
PG: None	5	3.50	0.00	3.50-3.50	0.00	5	19.80	0.00	19.80-19.80	0.00	5	8.40	0.00	8.40-8.40	0.00	5	8.80	0.00	8.80-8.80	0.00	5	3.50	0.00	3.50-3.50	0.00
SG: Select Beckman Coulter Instruments (3-part Diff)	221	5.04	0.57	3.30-6.70	11.40	221	7.24	1.41	3.00-11.40	19.58	221	6.57	0.90	3.90-9.30	13.48	221	6.57	0.75	4.30-8.80	11.36	220	5.02	0.58	3.30-6.70	11.60
PG: Drew Scientific D3	25	3.20	0.20	2.60-3.80	6.25	25	10.87	1.13	7.50-14.30	10.37	25	6.50	0.53	4.90-8.10	8.15	25	6.51	0.52	4.90-8.10	8.00	25	3.20	0.15	2.80-3.70	4.69
PG: Abbott Cell-Dyn Emerald	160	2.99	0.28	2.20-3.80	9.33	162	12.20	1.81	6.80-17.60	14.84	158	6.56	0.67	4.60-8.60	10.15	160	6.52	0.79	4.10-8.90	12.15	160	3.03	0.29	2.10-3.90	9.67
PG: CDS M Series	186	6.34	0.79	3.90-8.70	12.54	183	10.41	2.06	4.20-16.60	19.81	186	8.68	1.20	5.10-12.30	13.68	186	8.55	1.14	5.20-12.00	13.26	186	6.36	0.86	3.80-9.00	13.44
SG: CDS Hematology Instruments	192	6.34	0.82	3.80-8.80	13.02	189	10.39	2.06	4.20-16.60	19.81	192	8.72	1.20	5.10-12.30	13.79	192	8.58	1.15	5.20-12.00	13.37	192	6.34	0.92	3.50-9.10	14.60
AG: All Group	735	4.84	1.46	0.40-9.20	30.																				



AG: All Group	732	13.60	1.16	10.20-17.10	8.46	731	15.27	1.33	11.30-19.30	8.69	731	14.54	1.46	10.10-18.90	10.07	731	14.58	1.38	10.50-18.70	9.45	730	13.55	1.25	9.80-17.40	9.19	
Red Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: Abbott Cell-Dyn 1700	20	5.81	0.14	5.46-6.16	2.46	20	2.47	0.07	2.32-2.62	2.71	20	4.66	0.12	4.38-4.94	2.51	20	4.65	0.14	4.37-4.93	2.97	20	5.81	0.20	5.46-6.16	3.51	
PG: Abbott Cell-Dyn 1800	52	5.67	0.13	5.33-6.01	2.29	53	2.52	0.07	2.37-2.67	2.74	53	4.59	0.12	4.31-4.87	2.51	53	4.58	0.17	4.31-4.85	3.65	53	5.67	0.16	5.33-6.01	2.73	
PG: CDS Medonic 620	8	5.80	0.06	5.45-6.15	1.09	8	2.35	0.06	2.21-2.49	2.43	8	4.58	0.13	4.31-4.85	2.75	8	4.55	0.07	4.28-4.82	1.45	8	5.84	0.13	5.49-6.19	2.16	
PG: Beckman Coulter AcT Diff/Diff 2	231	5.95	0.22	5.59-6.31	3.70	230	2.46	0.10	2.31-2.61	4.02	230	4.68	0.11	4.40-4.96	2.33	230	4.69	0.12	4.41-4.97	2.54	230	5.90	0.40	5.55-6.25	6.75	
PG: Horiba ABX Diag. Micros/45/60	72	5.78	0.20	5.43-6.13	3.51	73	2.35	0.05	2.21-2.49	2.30	73	4.54	0.12	4.27-4.81	2.73	74	4.53	0.15	4.26-4.80	3.22	74	5.78	0.27	5.43-6.13	4.74	
PG: None	5	5.62	0.00	5.28-5.96	0.00	5	2.33	0.00	2.19-2.47	0.00	5	4.52	0.00	4.25-4.79	0.00	5	4.44	0.00	4.17-4.71	0.00	5	5.66	0.00	5.32-6.00	0.00	
SG: Select Abbott Cell Dyn Instruments (3-part Diff)	235	5.51	0.26	5.18-5.84	4.63						234	4.40	0.19	4.14-4.66	4.39	236	4.40	0.21	4.14-4.66	4.84	236	5.51	0.25	5.18-5.84	4.59	
SG: Select Beckman Coulter Instruments (3-part Diff)	233	5.95	0.22	5.59-6.31	3.70	232	2.46	0.10	2.31-2.61	4.02	232	4.68	0.11	4.40-4.96	2.33	232	4.68	0.12	4.40-4.96	2.54	232	5.90	0.40	5.55-6.25	6.73	
PG: Drew Scientific D3	25	5.70	0.20	5.36-6.04	3.42	25	2.39	0.08	2.25-2.53	3.51	25	4.52	0.18	4.25-4.79	3.89	25	4.51	0.29	4.24-4.78	6.36	25	5.64	0.32	5.30-5.98	5.59	
PG: Abbott Cell-Dyn Emerald	162	5.42	0.24	5.09-5.75	4.43	161	2.28	0.08	2.14-2.42	3.68	160	4.31	0.14	4.05-4.57	3.23	162	4.32	0.17	4.06-4.58	4.00	162	5.42	0.23	5.09-5.75	4.21	
PG: CDS M Series	187	5.87	0.17	5.52-6.22	2.88	186	2.35	0.06	2.21-2.49	2.51	187	4.55	0.20	4.28-4.82	4.29	187	4.56	0.08	4.29-4.83	1.73	187	5.84	0.28	5.49-6.19	4.81	
SG: CDS Hematology Instruments	195	5.87	0.17	5.52-6.22	2.83	194	2.35	0.06	2.21-2.49	2.47	195	4.55	0.19	4.28-4.82	4.22	195	4.56	0.08	4.29-4.83	1.73	195	5.84	0.28	5.49-6.19	4.74	
AG: All Group	762	5.77	0.28	5.42-6.12	4.94	753	2.38	0.11	2.24-2.52	4.54	761	4.54	0.20	4.27-4.81	4.34	764	4.54	0.19	4.27-4.81	4.25	764	5.74	0.35	5.40-6.08	6.13	
White Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: Abbott Cell-Dyn 1700	20	21.00	1.07	17.90-24.10	5.10	20	2.17	0.12	1.90-2.50	5.45	20	8.55	0.36	7.20-9.80	4.24	20	8.57	0.39	7.30-9.90	4.53	20	20.90	1.09	17.80-24.00	5.22	
PG: Abbott Cell-Dyn 1800	53	20.89	1.16	17.80-24.00	5.55	53	2.03	0.12	1.70-2.30	6.00	53	8.15	0.46	7.00-9.40	5.61	53	8.16	0.51	7.00-9.40	6.22	53	20.97	1.20	17.90-24.10	5.67	
PG: CDS Medonic 620	8	21.51	0.40	18.30-24.70	1.86	8	1.95	0.14	1.70-2.30	7.00	8	8.41	0.37	7.10-9.70	4.40	8	8.25	0.38	7.10-9.50	4.58	8	22.09	1.29	18.80-25.40	5.84	
PG: Beckman Coulter AcT Diff/Diff 2	231	21.77	1.19	18.50-25.10	5.46	230	2.20	0.13	1.90-2.50	5.91	230	8.61	0.36	7.30-9.90	4.19	230	8.63	0.29	7.30-9.90	3.37	230	21.64	1.46	18.40-24.80	6.76	
PG: Horiba ABX Diag. Micros/45/60	73	20.74	1.73	17.60-23.80	8.36	73	1.92	0.10	1.60-2.20	5.26	74	7.97	0.25	6.80-9.20	3.13	74	7.94	0.27	6.70-9.10	3.42	74	20.31	1.26	17.30-23.30	6.21	
PG: None	5	20.00	0.00	17.00-23.00	0.00	5	1.90	0.00	1.60-2.20	0.00	5	7.90	0.00	6.70-9.10	0.00	5	7.70	0.00	6.50-8.90	0.00	5	20.00	0.00	17.00-23.00	0.00	
SG: Select Abbott Cell Dyn Instruments (3-part Diff)	236	20.00	1.07	17.00-23.00	5.35	236	2.05	0.12	1.80-2.40	5.71	236	8.08	0.37	6.90-9.30	4.57	236	8.09	0.37	6.90-9.30	4.57	236	19.97	1.07	17.00-23.00	5.35	
SG: Select Beckman Coulter Instruments (3-part Diff)	233	21.76	1.19	18.50-25.10	5.46	232	2.20	0.13	1.90-2.50	5.91	232	8.61	0.35	7.30-9.90	4.07	232	8.63	0.29	7.30-9.90	3.37	232	21.64	1.46	18.40-24.80	6.76	
PG: Drew Scientific D3	25	20.35	0.72	17.30-23.50	3.53	25	2.11	0.15	1.80-2.40	7.14	25	8.02	0.90	6.80-9.20	11.25	25	8.12	0.58	6.90-9.30	7.16	25	20.00	1.10	17.00-23.00	5.50	
PG: Abbott Cell-Dyn Emerald	162	19.60	0.76	16.70-22.50	3.88	162	2.04	0.11	1.70-2.30	5.50	162	7.99	0.27	6.80-9.20	3.38	162	8.01	0.25	6.80-9.20	3.13	162	19.54	0.68	16.60-22.40	3.49	
PG: CDS M Series	187	21.03	0.85	17.90-24.10	4.05	186	1.82	0.09	1.50-2.10	5.00	187	8.01	0.23	6.80-9.20	2.88	187	8.04	0.24	6.80-9.20	3.00	187	20.82	1.52	17.70-23.90	7.31	
SG: CDS Hematology Instruments	195	21.05	0.84	17.90-24.10	4.00	194	1.82	0.10	1.50-2.10	5.56	195	8.03	0.25	6.80-9.20	3.13	195	8.04	0.25	6.80-9.20	3.13	195	20.87	1.53	17.80-24.00	7.32	
SG: Select ABX Micros Instruments	73	20.74	1.73	17.60-23.80	8.36	73	1.92	0.10	1.60-2.20	5.26	74	7.97	0.25	6.80-9.20	3.13	74	7.94	0.27	6.70-9.10	3.42	74	20.31	1.26	17.30-23.30	6.21	
AG: All Group	764	20.88	1.33	17.80-24.00	6.36	762	2.03	0.19	1.70-2.30	9.50	764	8.21	0.44	7.00-9.40	5.37	764	8.23	0.41	7.00-9.40	5.12	764	20.74	1.50	17.60-23.80	7.25	
Hematocrit (Waived)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: Microhematocrit, Spun	22	47.05	5.45	30.60-63.40	11.60	12	16.00	1.41	11.80-20.20	8.81																
PG: ST Hematastat	5	48.00	2.55	40.40-55.60	5.31																					
AG: All Group	30	48.40	3.33	38.40-58.40	6.88																					
Hemoglobin (Waived)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: HemoCue	50	17.32	0.81	14.90-19.70	4.68																					
PG: HemoCue 201	95	17.26	1.46	12.90-21.70	8.44																					
PG: Alere HemoPoint H2	35	17.46	2.07	11.30-23.70	11.83																					
SG: Select Waived Hemoglobin Instruments	182	17.28	1.50	12.80-21.80	8.67																					
AG: All Group	184	17.29	1.50	12.80-21.80	8.67	5	6.28	0.16	5.80-6.80	2.54																
608 - Hemocue Glucose/Hgb;628 - AFC Waived/PPM	HQ-2																									
Glucose Whole Blood (Hemocue)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: HemoCue	38	105.85	10.85	74.00-138.00	10.19																					
PG: HemoCue 201	136	106.72	8.09	83.00-131.00	7.57																					
Hemoglobin (Hemocue)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: HemoCue	27	12.63	1.17	9.10-16.10	9.29																					
PG: HemoCue 201	73	13.17	0.66	11.30-15.10	4.92</																					



PG: Abbott i-Stat / Thermal	8	22.88	0.64	21.00-24.80	2.79	8	22.25	0.71	20.20-24.40	3.18	8	17.63	0.74	15.40-19.80	4.20	8	22.38	0.92	19.60-25.20	4.11	8	18.63	0.92	15.80-21.40	4.95
SG: International Abbott i-Stat Instruments / Thermal	41	22.88	1.19	19.30-26.50	5.20	13	22.38	0.65	20.40-24.30	2.90	13	17.69	0.63	15.80-19.60	3.56	13	22.54	0.78	20.20-24.80	3.47	13	18.85	0.80	16.40-21.20	4.26
PG: Abbott i-Stat (Waived) / Thermal	28	22.83	1.37	18.70-26.90	5.96																				
SG: Select Abbott i-Stat Instruments / Thermal	36	22.84	1.24	19.10-26.50	5.44	8	22.25	0.71	20.20-24.40	3.18	8	17.63	0.74	15.40-19.80	4.20	8	22.38	0.92	19.60-25.20	4.11	8	18.63	0.92	15.80-21.40	4.95
Creatinine	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat / Thermal	18	3.85	0.10	3.30-4.50	2.56	17	2.68	0.10	2.30-3.10	3.70	17	6.59	0.30	5.60-7.60	4.55	17	4.78	0.16	4.10-5.50	3.33	16	0.34	0.05	0.00-0.60	16.67
SG: International Abbott i-Stat Instruments / Thermal	55	3.82	0.14	3.20-4.40	3.42	23	2.69	0.09	2.30-3.10	3.33	23	6.63	0.28	5.60-7.60	4.24	23	4.76	0.16	4.10-5.50	3.33	22	0.33	0.05	0.00-0.60	16.67
PG: Abbott i-Stat (Waived) / Thermal	31	3.81	0.15	3.20-4.40	3.95																				
SG: Select Abbott i-Stat Instruments / Thermal	49	3.82	0.14	3.20-4.40	3.68	17	2.68	0.10	2.30-3.10	3.70	17	6.59	0.30	5.60-7.60	4.55	17	4.78	0.16	4.10-5.50	3.33	16	0.34	0.05	0.00-0.60	16.67
Glucose	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat / Thermal	10	186.60	1.43	167.90-205.30	0.77	10	241.50	2.01	217.30-265.60	0.83	10	37.80	0.79	32.00-44.00	2.11	10	193.50	1.43	174.20-212.80	0.74	10	50.80	0.63	45.00-57.00	1.24
SG: International Abbott i-Stat Instruments / Thermal	44	186.44	1.26	167.80-205.00	0.70	15	241.25	1.74	217.10-265.30	0.72	15	37.69	0.78	32.00-44.00	2.11	15	193.27	1.38	174.00-212.60	0.73	15	50.69	0.53	45.00-57.00	1.05
PG: Abbott i-Stat (Waived) / Thermal	29	186.40	1.28	167.80-205.00	0.70																				
SG: Select Abbott i-Stat Instruments / Thermal	39	186.45	1.30	167.80-205.00	0.70	10	241.50	2.01	217.30-265.60	0.83	10	37.80	0.79	32.00-44.00	2.11	10	193.50	1.43	174.20-212.80	0.74	10	50.80	0.63	45.00-57.00	1.24
Hematocrit	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat / Thermal	6	20.50	0.55	19.30-21.70	2.68	6	30.00	0.63	28.20-31.80	2.10	6	20.33	0.52	19.10-21.50	2.56	6	32.83	0.75	30.80-34.80	2.29	6	13.00	0.00	12.20-13.80	0.00
SG: International Abbott i-Stat Instruments / Thermal	18	20.28	0.46	19.10-21.50	2.27	11	29.91	0.54	28.10-31.70	1.81	11	20.18	0.41	19.00-21.40	1.98	11	32.55	0.69	30.60-34.50	2.12	11	12.82	0.41	12.00-13.60	3.13
PG: Abbott i-Stat (Waived) / Thermal	7	20.14	0.38	18.90-21.30	1.89																				
SG: Select Abbott i-Stat Instruments / Thermal	13	20.31	0.48	19.10-21.50	2.36	6	30.00	0.63	28.20-31.80	2.10	6	20.33	0.52	19.10-21.50	2.56	6	32.83	0.75	30.80-34.80	2.29	6	13.00	0.00	12.20-13.80	0.00
Hemoglobin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat / Thermal	5	6.96	0.15	6.50-7.50	2.14	5	10.16	0.09	9.50-10.90	0.88	5	6.90	0.14	6.40-7.40	2.03	5	11.18	0.27	10.40-12.00	2.41	5	4.32	0.18	4.00-4.60	4.19
SG: International Abbott i-Stat Instruments / Thermal	17	6.88	0.13	6.40-7.40	1.88	10	10.15	0.11	9.50-10.90	1.08	10	6.85	0.11	6.40-7.40	1.59	10	11.07	0.23	10.30-11.90	2.07	10	4.30	0.16	4.00-4.60	3.72
PG: Abbott i-Stat (Waived) / Thermal	7	6.84	0.11	6.30-7.30	1.62																				
SG: Select Abbott i-Stat Instruments / Thermal	12	6.89	0.14	6.40-7.40	2.03	5	10.16	0.09	9.50-10.90	0.88	5	6.90	0.14	6.40-7.40	2.03	5	11.18	0.27	10.40-12.00	2.41	5	4.32	0.18	4.00-4.60	4.19
Ionized Calcium	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat / Thermal	6	2.11	0.04	1.98-2.24	2.04	6	0.60	0.01	0.59-0.62	0.83	6	2.32	0.05	2.18-2.46	2.03	6	1.02	0.01	0.98-1.06	1.18	6	0.56	0.02	0.50-0.62	3.39
SG: International Abbott i-Stat Instruments / Thermal	23	2.12	0.03	2.03-2.21	1.46	11	0.60	0.01	0.59-0.62	0.83	11	2.32	0.04	2.20-2.44	1.77	11	1.02	0.01	0.99-1.05	1.08	11	0.56	0.01	0.52-0.60	2.50
PG: Abbott i-Stat (Waived) / Thermal	12	2.11	0.03	2.03-2.19	1.28																				
SG: Select Abbott i-Stat Instruments / Thermal	18	2.11	0.03	2.01-2.21	1.52	6	0.60	0.01	0.59-0.62	0.83	6	2.32	0.05	2.18-2.46	2.03	6	1.02	0.01	0.98-1.06	1.18	6	0.56	0.02	0.50-0.62	3.39
Potassium (K+)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat / Thermal	13	2.72	0.04	2.20-3.20	1.48	13	6.62	0.04	6.10-7.10	0.61	13	3.22	0.04	2.70-3.70	1.25	13	4.51	0.03	4.00-5.00	0.67	13	2.03	0.05	1.50-2.50	2.50
SG: International Abbott i-Stat Instruments / Thermal	47	2.73	0.05	2.20-3.20	1.85	18	6.63	0.05	6.10-7.10	0.76	18	3.22	0.04	2.70-3.70	1.25	18	4.52	0.04	4.00-5.00	0.89	18	2.04	0.05	1.50-2.50	2.50
PG: Abbott i-Stat (Waived) / Thermal	29	2.73	0.05	2.20-3.20	1.85																				
SG: Select Abbott i-Stat Instruments / Thermal	42	2.73	0.04	2.20-3.20	1.85	13	6.62	0.04	6.10-7.10	0.61	13	3.22	0.04	2.70-3.70	1.25	13	4.51	0.03	4.00-5.00	0.67	13	2.03	0.05	1.50-2.50	2.50
Sodium (NA+)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat / Thermal	10	125.40	0.70	121.40-129.40	0.56	10	167.70	1.06	164.00-172.00	0.65	10	122.50	0.53	119.00-127.00	0.43	10	153.80	0.79	150.00-158.00	0.52	10	112.80	0.63	109.00-117.00	0.56
SG: International Abbott i-Stat Instruments / Thermal	44	125.44	0.66	121.40-129.40	0.56	15	167.73	0.88	164.00-172.00	0.54	15	122.53	0.52	119.00-127.00	0.42	15	153.87	0.74	150.00-158.00	0.48	15	112.67	0.62	109.00-117.00	0.55
PG: Abbott i-Stat (Waived) / Thermal	29	125.50	0.68	121.50-129.50	0.56																				
SG: Select Abbott i-Stat Instruments / Thermal	39	125.47	0.68	121.50-129.50	0.56	10	167.70	1.06	164.00-172.00	0.65	10	122.50	0.53	119.00-127.00	0.43	10	153.80	0.79	150.00-158.00	0.52	10	112.80	0.63	109.00-117.00	0.56
Urea Nitrogen (BUN)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat / Thermal	9	67.56	1.51	61.50-73.70	2.23	9	46.89	0.93	42.70-51.10	1.98	9	6.56	0.53	4.60-8.60	8.03	9	58.11	1.27	52.90-63.30	2.19	9	8.56	0.53	6.60-10.60	6.16
SG: International Abbott i-Stat Instruments / Thermal	43	67.09	1.56	61.10-73.10	2.32	14	47.21	0.88	43.00-51.40	1.86	14	6.48	0.49	4.50-8.50	7.54	14	58.09	1.06	52.90-63.30	1.82	14	8.52	0.42	6.50-10.50	4.94
PG: Abbott i-Stat (Waived) / Thermal	29	66.77	1.59	60.80-72.80	2.38																				
SG: Select Abbott i-Stat Instruments / Thermal	38	66.95	1.59	60.90-72.90	2.38	9	46.89	0.93	42.70-51.10	1.98	9	6.56	0.53	4.60-8.60	8.03	9	58.11	1.27	52.90-63.30	2.19	9	8.56	0.53	6.60-10.60	6.16
HCO3	Labs	Mean	SD	Range	CV.%</																				



AG: All Group	50	614.30	91.19	340.00-888.00	14.85																				
Progesterone	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	12	11.61	2.72	3.40-19.80	23.45																				
PSA	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Architect	5	3.51	0.12	3.14-3.88	3.50																				
PG: Beckman ACCESS	20	5.06	0.17	4.54-5.58	3.40																				
PG: bioMerieux Vitex, Mini Vidas	16	4.75	0.32	3.78-5.72	6.78																				
PG: Siemens Dimension/AR/ES/RxL/Expand	14	3.76	0.20	3.15-4.37	5.43																				
PG: Qualigen FastPack PSA	18	4.26	0.56	2.56-5.96	13.26																				
PG: TOSOH AIA	17	4.13	0.43	2.83-5.43	10.46																				
SG: Select Beckman Coulter Instruments (Chem)	21	5.06	0.17	4.56-5.56	3.32																				
PG: Roche Cobas e 411	5	4.77	0.35	3.73-5.81	7.25																				
PG: Abbott Architect i1000 SR	17	3.71	0.24	3.00-4.42	6.39																				
SG: Select Abbott Architect Instruments	27	3.69	0.22	3.02-4.36	6.04																				
PG: NanoEnTek Frend System	7	6.29	0.56	4.63-7.96	8.82																				
AG: All Group	130	4.38	0.76	2.10-6.66	17.33																				
Testosterone	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman ACCESS	15	4.34	0.18	3.80-4.80	4.19																				
PG: Siemens Immulite 1000	5	4.24	0.31	3.30-5.10	7.38																				
PG: Qualigen FastPak Testosterone	10	3.62	0.45	2.30-5.00	12.50																				
PG: TOSOH AIA	8	5.79	0.46	4.40-7.20	7.93																				
SG: Select Beckman Coulter Instruments (Chem)	16	4.33	0.19	3.70-4.90	4.42																				
SG: Select Abbott Architect Instruments	15	5.28	1.39	1.10-9.50	26.23																				
AG: All Group	59	4.72	1.14	1.30-8.10	24.26																				
Transferrin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	5	247.75	5.19	232.20-263.40	2.09																				
685 - Special Chemistry II	SPC-3					SPC-4																			
Insulin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	8	70.82	18.35	15.70-125.80	25.92	8	116.71	35.16	11.20-222.20	30.13															
PTH	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	11	17.19	3.25	7.40-26.90	18.90	10	25.59	4.28	12.80-38.40	16.72															
Vitamin D	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman ACCESS	13	27.91	3.72	16.76-39.06	13.31	13	47.70	7.07	26.51-68.90	14.81															
PG: TOSOH AIA	7	35.10	6.39	15.93-54.27	18.21	7	50.89	8.18	26.34-75.44	16.08															
SG: Select Beckman Coulter Instruments (Chem)	14	27.94	3.57	17.22-38.66	12.78	14	48.01	6.89	27.35-68.67	14.34															
SG: Select Roche Instruments	5	15.96	1.19	12.40-19.52	7.44	5	27.24	1.25	23.49-30.99	4.59															
PG: Roche Cobas e 411	5	15.96	1.19	12.40-19.52	7.44	5	27.24	1.25	23.49-30.99	4.59															
PG: Abbott Architect i1000 SR	19	26.07	2.81	17.63-34.51	10.79	19	39.91	4.10	27.62-52.20	10.27															
SG: Select Abbott Architect Instruments	27	26.26	2.45	18.90-33.62	9.34	27	40.07	3.50	29.59-50.55	8.72															
PG: Qualigen FastPack Vitamin D	14	30.45	4.89	15.77-45.13	16.07	16	38.84	7.38	16.71-60.97	19.00															
AG: All Group	68	28.75	5.40	12.55-44.95	18.79	70	43.19	7.84	19.68-66.70	18.15															
752 - Auto Diff 8 (5-part Diff)	SXT-6					SXT-7					SXT-8					SXT-9					SXT-10				
White Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex XS-1000i	5	21.11	0.37	17.90-24.30	1.75	5	3.88	0.08	3.30-4.50	2.05	5	7.77	0.19	6.60-9.00	2.44	5	7.64	0.15	6.50-8.70	1.97	5	20.91	0.46	17.80-24.00	2.20
SG: Select Sysmex Instruments (5-part Diff)	7	20.74	0.73	17.60-23.80	3.53	7	3.83	0.13	3.20-4.40	3.42	7	7.62	0.33	6.50-8.70	4.34	7	7.54	0.21	6.40-8.60	2.80	7	20.67	0.62	17.60-23.80	3.00
Lymphocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex XS-1000i	5	19.14	0.30	18.20-20.00	1.57	5	15.18	0.71	13.10-17.30	4.67	5	25.90	0.22	25.20-26.60	0.85	5	26.44	0.59	24.60-28.20	2.23	5	18.68	0.47	17.30-20.10	2.51
SG: Select Sysmex Instruments (5-part Diff)	7	18.09	1.82	12.60-23.60	10.06	7	13.90	2.29	7.00-20.80	16.47	7	25.04	2.11	18.70-31.30	8.40	7	25.43	2.38	18.30-32.50	9.37	7	17.54	2.01	11.50-23.50	11.49
Neutrophil	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex XS-1000i	5	55.12	0.91	52.40-57.80	1.65	5	58.74	0.74	56.50-60.90	1.26	5	50.28	1.05	47.10-53.40	2.09	5	50.52	0.72	48.30-52.70	1.43	5	56.36	0.25	55.60-57.10	0.44
SG: Select Sysmex Instruments (5-part Diff)	7	57.83	4.72	43.60-72.00	8.17	7	62.69	6.78	42.40-83.00	10.81	7	52.66	4.15	40.30-65.20	7.87	7	52.91	4.25	40.10-65.70	8.03	7	58.40	3.69	47.30-69.50	6.32
Mono/Mixed	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex XS-1000i	5	1.08	0.29	0.20-2.00	26.36	5	0.86	0.22	0.20-1.60	24.44	5	1.74	0.24	1.00-2.40	14.12	5	1.54	0.32	0.50-2.50	21.33	5	0.94	0.18	0.40-1.40	20.00
PG: Sysmex XN Series - XN-10 XN-20	5	6.40	0.42	5.10-7.70	6.56	5	4.30	0.28	3.50-5.10	6.51	5	5.70	3.39	0.00-15.90	59.47	5	4.55	2.19	0.00-11.20	47.61	5	8.10	1.84	2.60-13.60	22.72
Eosinophil	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean													



Test Name	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %
SG: Select Sysmex Instruments (5-part Diff)	7	52.76	0.96	49.60-56.00	1.82	7	19.19	0.33	18.00-20.40	1.72	7	40.79	0.60	38.40-43.20	1.47	7	40.70	0.38	38.30-43.10	0.93	7	52.90	0.97	49.70-56.10	1.83
MCV	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %
PG: Sysmex XS-1000i	5	95.18	1.09	91.90-98.50	1.14	5	80.24	0.72	78.00-82.40	0.90	5	89.26	0.79	86.90-91.70	0.88	5	89.06	0.79	86.70-91.50	0.89	5	94.96	0.85	92.50-97.50	0.89
SG: Select Sysmex Instruments (5-part Diff)	6	94.10	1.88	88.50-99.70	2.00	7	79.61	1.24	75.90-83.30	1.56	7	88.70	1.16	85.20-92.20	1.31	7	88.51	1.15	85.00-92.00	1.30	7	94.20	1.48	89.80-98.60	1.57
Platelet Count	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %
PG: Sysmex XS-1000i	5	423.00	14.75	317.00-529.00	3.48	5	56.40	3.85	42.00-70.00	6.79	5	212.60	3.91	160.00-266.00	1.83	5	211.60	3.65	159.00-265.00	1.70	5	421.60	9.34	317.00-528.00	2.20
SG: Select Sysmex Instruments (5-part Diff)	7	426.43	14.58	320.00-533.00	3.43	7	55.71	3.35	42.00-70.00	6.07	7	213.43	3.60	160.00-266.00	1.69	7	211.14	6.04	158.00-264.00	2.84	7	424.57	9.61	319.00-531.00	2.26
RDW	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %
PG: Sysmex XS-1000i	5	15.04	0.11	14.70-15.30	0.73	5	16.64	0.20	16.00-17.20	1.14	5	15.70	0.21	15.10-16.30	1.34	5	15.72	0.11	15.40-16.00	0.70	5	15.20	0.12	14.80-15.60	0.79
SG: Select Sysmex Instruments (5-part Diff)	7	14.73	0.54	13.10-16.30	3.67	7	16.49	0.33	15.50-17.50	2.00	7	15.47	0.43	14.20-16.80	2.77	7	15.44	0.48	14.00-16.80	3.12	7	14.84	0.64	12.90-16.70	4.26
613 - Auto Diff I	SYX-6					SYX-7					SYX-8					SYX-9					SYX-10				
White Blood Cell Count	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %
PG: Sysmex K-800,1000,4500,KX-21	17	20.59	0.41	17.50-23.70	1.99	17	2.66	0.09	2.30-3.10	3.33	17	8.21	0.17	7.00-9.40	2.07	17	8.18	0.17	7.00-9.40	2.07	17	20.50	0.40	17.40-23.60	1.95
PG: Sysmex poch-H100i	21	20.42	0.46	17.30-23.50	2.25	22	2.61	0.08	2.20-3.00	3.08	22	8.16	0.25	7.00-9.40	3.05	22	8.15	0.21	6.90-9.30	2.59	22	20.38	0.49	17.30-23.50	2.40
SG: Select Sysmex Instruments (3-part Diff)	87	21.01	0.68	17.90-24.10	3.24	88	2.70	0.12	2.30-3.10	4.44	88	8.37	0.28	7.10-9.70	3.33	88	8.37	0.30	7.10-9.70	3.57	88	20.97	0.74	17.90-24.10	3.52
PG: Sysmex XP-300	49	21.38	0.58	18.20-24.60	2.71	49	2.75	0.11	2.30-3.10	4.07	49	8.52	0.23	7.20-9.80	2.71	49	8.53	0.26	7.20-9.80	3.06	49	21.38	0.65	18.20-24.60	3.04
SG: International Sysmex Instruments	25	20.48	0.44	17.40-23.60	2.15	26	2.62	0.08	2.20-3.00	3.08	26	8.17	0.24	7.00-9.40	2.93	26	8.17	0.20	7.00-9.40	2.44	26	20.45	0.48	17.30-23.50	2.35
Neutrophil	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %
PG: Sysmex K-800,1000,4500,KX-21	17	26.86	0.75	24.60-29.10	2.79	17	70.54	1.32	66.50-74.50	1.87	17	52.11	1.17	48.60-55.60	2.25	17	52.31	1.07	49.10-55.50	2.05	17	26.84	0.35	25.70-27.90	1.34
PG: Sysmex poch-H100i	20	28.57	0.86	26.00-31.20	3.01	21	73.23	1.54	68.60-77.80	2.10	21	54.53	1.05	51.40-57.60	1.93	21	54.52	1.08	51.30-57.70	1.98	21	28.06	0.88	25.50-30.70	3.13
SG: Select Sysmex Instruments (3-part Diff)	84	27.34	1.04	24.20-30.40	3.81	85	70.71	2.08	64.50-76.90	2.94	85	52.73	1.45	48.40-57.10	2.75	85	52.89	1.44	48.60-57.20	2.72	84	27.22	0.93	24.40-30.00	3.42
PG: Sysmex XP-300	47	27.01	0.80	24.60-29.40	2.96	47	69.72	1.55	65.00-74.40	2.22	47	52.18	1.01	49.20-55.20	1.93	47	52.39	1.15	48.90-55.90	2.19	46	26.98	0.88	24.40-29.60	3.26
SG: International Sysmex Instruments	24	28.45	0.89	25.80-31.20	3.12	25	73.32	1.46	68.90-77.70	1.99	25	54.38	1.15	50.90-57.90	2.11	25	54.48	0.99	51.50-57.50	1.82	25	28.02	0.86	25.40-30.60	3.07
Red Blood Cell Count	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %
PG: Sysmex K-800,1000,4500,KX-21	17	5.74	0.06	5.40-6.08	1.01	17	2.49	0.03	2.34-2.64	1.12	17	4.17	0.04	3.92-4.42	1.03	17	4.19	0.05	3.94-4.44	1.10	17	5.75	0.06	5.41-6.09	1.10
PG: Sysmex poch-H100i	21	5.80	0.10	5.45-6.15	1.79	22	2.51	0.03	2.36-2.66	1.35	22	4.24	0.06	3.99-4.49	1.51	22	4.22	0.07	3.97-4.47	1.61	22	5.80	0.12	5.45-6.15	2.03
SG: Select Sysmex Instruments (3-part Diff)	87	5.75	0.08	5.41-6.09	1.46	88	2.50	0.04	2.35-2.65	1.68	88	4.19	0.06	3.94-4.44	1.34	88	4.19	0.05	3.94-4.44	1.29	88	5.75	0.09	5.41-6.09	1.63
PG: Sysmex XP-300	49	5.73	0.07	5.39-6.07	1.29	49	2.49	0.05	2.34-2.64	1.93	49	4.17	0.04	3.92-4.42	0.98	49	4.17	0.04	3.92-4.42	0.98	49	5.73	0.08	5.39-6.07	1.45
SG: International Sysmex Instruments	25	5.80	0.10	5.45-6.15	1.78	26	2.51	0.04	2.36-2.66	1.59	26	4.23	0.07	3.98-4.48	1.73	26	4.22	0.07	3.97-4.47	1.61	26	5.80	0.12	5.45-6.15	1.98
Hemoglobin	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %
PG: Sysmex K-800,1000,4500,KX-21	17	18.26	0.28	17.00-19.60	1.58	17	6.34	0.17	5.90-6.70	2.70	17	12.19	0.17	11.30-13.10	1.39	17	12.19	0.17	11.30-13.10	1.39	17	18.29	0.27	17.00-19.60	1.48
PG: Sysmex poch-H100i	21	18.25	0.22	17.00-19.60	1.20	22	6.19	0.09	5.80-6.60	1.45	22	12.05	0.16	11.20-12.80	1.33	22	12.03	0.14	11.20-12.80	1.17	22	18.12	0.30	16.80-19.40	1.66
SG: Select Sysmex Instruments (3-part Diff)	87	18.23	0.28	16.90-19.50	1.54	88	6.22	0.14	5.80-6.60	2.26	88	12.14	0.17	11.30-12.90	1.40	88	12.16	0.19	11.30-13.10	1.56	87	18.24	0.34	16.90-19.50	1.87
PG: Sysmex XP-300	49	18.22	0.31	16.90-19.50	1.70	49	6.20	0.13	5.80-6.60	2.10	49	12.16	0.17	11.30-13.10	1.39	49	12.20	0.19	11.30-13.10	1.56	48	18.27	0.37	17.00-19.60	2.02
SG: International Sysmex Instruments	25	18.24	0.24	16.90-19.50	1.32	26	6.20	0.09	5.80-6.60	1.45	26	12.05	0.16	11.30-12.90	1.32	26	12.05	0.14	11.20-12.80	1.17	26	18.13	0.30	16.80-19.40	1.66
Lymphocyte	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %
PG: Sysmex K-800,1000,4500,KX-21	17	60.44	0.68	58.40-62.40	1.11	17	12.14	0.99	9.10-15.10	8.18	17	29.94	0.80	27.50-32.30	2.68	17	30.15	0.79	27.80-32.60	2.62	17	60.43	0.50	58.90-61.90	0.83
PG: Sysmex poch-H100i	20	59.91	0.73	57.70-62.10	1.22	19	11.13	0.93	8.30-13.90	8.38	21	28.96	0.98	26.10-31.90	3.38	21	29.21	0.78	26.90-31.50	2.67	21	60.03	0.59	58.20-61.80	0.98
SG: Select Sysmex Instruments (3-part Diff)	85	60.13	0.78	57.80-62.40	1.30	81	12.01	1.13	8.60-15.40	9.42	86	29.48	0.97	26.60-32.40	3.29	86	29.66	0.95	26.80-32.60	3.23	86	60.21	0.65	58.30-62.20	1.08
PG: Sysmex XP-300	48	60.10	0.82	57.60-62.60	1.36	45	12.31	1.09	9.00-15.60	8.86	48	29.54	0.93	26.70-32.30	3.15	48	29.68	1.00	26.70-32.70	3.37	48	60.22	0.70	58.10-62.30	1.16
SG: International Sysmex Instruments	24	60.00	0.79	57.60-62.40	1.32	23	11.03	1.12	7.60-14.40	10.18	25	29.03	0.93	26.20-31.80	3.21	25	29.21	0.74	27.00-31.40	2.53	25	60.13	0.59	58.30-61.90	0.98
Hematocrit	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %	Labs	Mean	SD	Range	CV, %
PG: Sysmex K-800,1000,4500,KX-21	17	48.86	0.62	46.00-51.80	1.27	17	17.21	0.20	16.20-18.20	1.16	17	32.15	0.42	30.30-34.10	1.30	17	32.32	0.59	30.40-34.20	1.83	17	48.98	0.64	46.10-51.90	1.31
PG: Sysmex poch-H100i	21	50.72	1.48	47.70-53.70	2.92	22	18.12	0.44	17.00-19.20	2.43	22	34.10	0.89	32.10-36.10	2.61	22	33.96	0.97	32.00-36.00	2.85	22	50.71	1.47	47.70-53.70	2.90
SG: Select Sysmex Instruments (3-part Diff)	87	49.41	1.24	46.40-52.40	2.51	88	17.51	0.47	16.40-18.60	2.69	88	32.77	0.95	30.80-34.80	2.90	88	32.79	0.94	30.80-34.80	2.87	88	49.45	1.25	46.50-52.50	2.55
PG: Sysmex XP-300	49	49.07	0.87	46.20-52.00	1.77	49	17.36	0.29	16.40-18.40	1.67	49	32.41	0.46	30.50-34.30	1.42	49	32.45	0.53	30.50-34.30	1.64	49	49.07	0.92	46.20-52.00	1.87
SG: International Sysmex Instruments	25	50.63	1.44	47.60-53.60	2.85	26	18.07	0.47	17.00-19.20	2.60	26	33.97	0.94	32.00-36.00	2.76	26	33.83	0.99	31.80-35.80	2.93</					



Test Name	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%					
SG: Select Sysmex Instruments (3-part Diff)	82	12.06	1.89	6.40-17.80	15.62	83	12.44	1.83	6.90-17.90	14.76	83	12.06	1.85	6.50-17.60	15.29	83	12.07	1.86	6.50-17.70	15.37	83	12.18	1.93	6.40-18.00	15.82					
PG: Sysmex XP-300	48	10.98	0.26	10.20-11.80	2.36	48	11.40	0.23	10.70-12.10	2.02	48	10.99	0.32	10.00-12.00	2.91	48	10.98	0.33	10.00-12.00	3.00	48	11.04	0.26	10.20-11.80	2.36					
SG: International Sysmex Instruments	22	15.60	0.47	14.20-17.00	3.01	23	15.84	0.17	15.30-16.30	1.08	23	15.42	0.25	14.70-16.10	1.62	23	15.47	0.25	14.80-16.30	1.61	23	15.71	0.50	14.20-17.20	3.18					
708 - Tumor Markers																														
CA 125						TM-6					TM-7					TM-8					TM-9					TM-10				
PG: TOSOH AIA	5	95.40	1.56	90.70-100.10	1.64	5	396.40	1.38	392.30-400.50	0.35	5	99.00	2.94	90.20-107.80	2.97	5	397.23	11.89	361.50-432.90	2.99	5	400.00	6.93	379.20-420.80	1.73					
PG: OCD Vitros ECI/ECIQ	5	27.13	2.67	19.10-35.10	9.85	5	116.43	19.33	58.40-174.40	16.61	5	25.90	2.04	19.80-32.00	7.88	5	132.33	3.21	122.70-141.90	2.43	5	129.67	3.06	120.50-138.90	2.36					
SG: International OCD Vitros Instruments	5	26.48	2.54	18.90-34.10	9.58	5	121.32	18.57	65.60-177.00	15.31	5	25.88	1.67	20.90-30.90	6.45	5	133.00	2.94	124.20-141.80	2.21	5	130.25	2.75	122.10-138.60	2.11					
PSA																														
AG: All Group	7	12.45	1.45	8.11-16.79	11.61	7	12.98	1.60	8.17-17.79	12.35	7	12.61	1.35	8.57-16.65	10.68	7	12.80	1.77	7.49-18.11	13.83	7	13.14	1.64	8.22-18.06	12.48					
705 - Urine Microalbumin & Creatinine																														
Urine Creatinine																														
UM-2																														
PG: Siemens Clinitek 50	13	276.92	43.85	145.00-409.00	15.85																									
PG: Siemens Clinitek Status	28	259.11	52.42	102.00-416.00	20.23																									
PG: Bayer DCA 2000	5	195.04	6.89	174.00-216.00	3.54																									
PG: Siemens Dimension/AR/ES/RxL/Xpand	7	164.44	9.50	136.00-193.00	5.79																									
PG: Beckman (Olympus) AU	6	172.92	17.16	121.00-225.00	9.94																									
SG: Select Beckman Coulter Instruments (Chem)	7	175.36	16.95	124.00-226.00	9.66																									
PG: Siemens (Bayer) DCA Vantage	17	186.61	11.09	154.00-220.00	5.94																									
SG: Select Siemens (Bayer) DCA Instruments	22	188.53	10.76	157.00-221.00	5.71																									
PG: Afinion ACR	9	168.04	3.63	157.00-179.00	2.14																									
AG: All Group	90	218.38	56.28	49.00-387.00	25.83																									
Urine Microalbumin																														
UM-2																														
PG: Siemens Clinitek 50	11	10.00	0.00	10.00-10.00	0.00																									
PG: Siemens Clinitek Status	23	10.00	0.00	10.00-10.00	0.00																									
PG: Bayer DCA 2000	6	27.98	1.56	23.00-33.00	5.71																									
PG: Siemens Dimension/AR/ES/RxL/Xpand	7	25.50	2.33	19.00-33.00	8.85																									
PG: Beckman (Olympus) AU	7	26.47	4.02	14.00-38.00	15.38																									
SG: Select Beckman Coulter Instruments (Chem)	8	26.54	3.73	16.00-38.00	13.70																									
SG: Semi-quant Methods - Ur Microalbumin-Creat	36	10.43	2.57	2.00-18.00	26.00																									
PG: Siemens (Bayer) DCA Vantage	18	25.10	3.83	14.00-36.00	15.20																									
SG: Select Siemens (Bayer) DCA Instruments	24	25.82	3.60	15.00-37.00	13.85																									
PG: Afinion ACR	9	24.33	2.30	17.00-31.00	9.58																									
SG: Select Abbott Architect Instruments	5	28.04	0.71	26.00-30.00	2.50																									
AG: All Group	88	19.20	8.29	0.00-44.00	43.68																									
728 - Urine Chemistry																														
Total Protein - Urine						URC-3					URC-4																			
AG: All Group	5	106.03	2.28	99.20-112.80	2.15	5	11.83	3.81	0.40-23.20	32.29																				
Creatinine - Urine																														
AG: All Group	5	178.18	18.23	123.50-232.90	10.23	5	64.43	10.23	33.70-95.10	15.89																				
601 - Waived/PPM Tests II;609 - Waived/PPM Tests I;664 - Whole Blood Glucose;669 - Whole Blood Glucose																														
Glucose Wn Blood (Waived)						WBG-6					WBG-7					WBG-8					WBG-9					WBG-10				
PG: Abbott Freestyle Lite	12	185.67	19.11	129.00-243.00	10.27																									
PG: Bayer Ascensia Contour	108	166.32	22.21	99.00-233.00	13.37	19	59.05	6.42	40.00-78.00	10.85	5	320.40	55.62	153.00-487.00	17.38	5	93.60	12.74	56.00-132.00	13.51	5	81.40	11.19	47.00-115.00	13.83					
PG: Bayer Ascensia Elite XL	8	165.63	13.33	126.00-206.00	8.01																									
PG: Lifescan One Touch II	5	232.00	43.84	101.00-363.00	18.88																									
PG: Lifescan One Touch Ultra	65	217.30	43.41	87.00-347.00	20.00	9	79.33	11.62	44.00-114.00	14.68																				
PG: Lifescan One Touch, others	7	227.14	27.23	145.00-309.00	11.98																									
PG: McKesson/SunMark True Track	23	523.17	14.34	480.00-566.00	2.73																									
PG: Medisense Precis. Xtra/Optium	5	229.60	10.99	197.00-263.00	4.78																									
PG: Roche Accu-Chek Advantage/Comp/GTS	34	194.88	34.78	91.00-299.00	17.85																									
PG: Roche AccuChek Aviva	49	196.84	8.75	171.00-223.00	4.42																									
PG: Roche Accu-Chek II	6	181.67	14.86	137.00-227.00	8.19																									
PG: Roche Accu-Chek, others	18	195.72	23.23	126.00-266.00	11.84																									
SG: Select Lifescan One Touch Instruments	79	218.65	41.65	94.00-344.00	19.00	10	77.40	12.55	40.00-115.00	16.23																				
SG: Select Abbott Freestyle Instruments	17	187.76	17.91	134.00-242.00	9.52																									
PG: Roche AccuChek Inform	67	194.72	5.78	178.00-212.00	2.97	36	75.22	1.69	70.00-80.00	2.27	32	309.22	43.76	178.00-440.00	14.17	32	111.69	2.76	104.00-120.00	2.50	32	96.31	1.93	90.00-102.00	1.98					
PG: Consult Diag Quintet Blood Glucose	27	223.25	35.83	116.00-330.00	16.05	8	75.88	8.90	49.00-103.00	11.71																				
PG: Medisense Precision Xceed Pro	26	216.89	24.42	144.00-290.00	11.24	5	80.40	6.23	61.00-99.00	7.75																				
SG: Select Medisense Instruments	36	216.68	25.14	142.00-292.00	11.57	5	80.40	6.23	61.00-99.00	7.75																				
PG: Home Diagnostics True Result	5	126.75	25.18	51.00-203.00	19.84																									
SG: Select Roche WBG Meters	174	195.00	18.16	140.00-250.00	9.33	48	74.04	5.18	58.00-90.00	7.03	34	309.76	42.51	183.00-438.00	13.71	34	111.88	3.05	103.00-121.00	2.77	34	96.35	1.94	90.00-102.00	1.98					
SG: Select Bayer WBG Meters	127	169.40	28.11	85.00-253.00	16.63	22	61.00	8.22	36.00-86.00	13.44	6	327.17	52.43	170.00-484.00	16.02	6	96.67	13.65	56.00-138.00	14.02	6	84.50	12.57	47.00-123.00	14.82					
PG: Nova Biomedical StatStrip Xpress	5	182.20	14.45	139.00-225.00	7.91																									
SG: Select True/True Track WBG Instruments	25	523.36	13.78	482.00-564.00	2.64	5																								



American Academy of Family Physicians
 11400 Tomahawk Creek Parkway
 Leawood, KS 66211-2672
 1-800-274-7911
 www.aafp.org/pt pt@aafp.org

SG: Select True Track/TRUE Balance WBG Instruments	23	523.17	14.34	480.00-566.00	2.73																							
PG: Nipro TRUEresult	9	137.67	12.82	100.00-176.00	9.28																							
PG: Roche AccuChek Performa	89	198.18	7.03	177.00-219.00	3.54	8	76.13	2.53	69.00-84.00	3.29																		
SG: Select TrueResult Meters	18	130.72	18.20	76.00-186.00	13.89																							
AG: All Group	607	195.79	36.28	87.00-305.00	18.52	103	72.83	10.35	42.00-104.00	14.25	50	328.50	65.37	133.00-525.00	19.88	50	110.84	13.27	71.00-151.00	11.98	50	94.66	9.84	66.00-124.00	10.32			



789 - Bacterial Antigen Detection	Labs	Response	Number	Percent
C. difficile Toxin Detect				
AD-6				
* Denotes Correct Response				
AG: All Methods	3			
		*Negative	3	100.00%
Cepheid GeneXpert	1	*Negative	1	100.00%
Alere C diff Quik Chek Complete	2	*Negative	2	100.00%
AD-7				
* Denotes Correct Response				
AG: All Methods	3			
		*Positive	3	100.00%
Cepheid GeneXpert	1	*Positive	1	100.00%
Alere C diff Quik Chek Complete	2	*Positive	2	100.00%
AD-8				
* Denotes Correct Response				
AG: All Methods	3			
		*Negative	3	100.00%
Cepheid GeneXpert	1	*Negative	1	100.00%
Alere C diff Quik Chek Complete	2	*Negative	2	100.00%
AD-9				
* Denotes Correct Response				
AG: All Methods	3			
		*Positive	3	100.00%
Cepheid GeneXpert	1	*Positive	1	100.00%
Alere C diff Quik Chek Complete	2	*Positive	2	100.00%
AD-10				
* Denotes Correct Response				
AG: All Methods	3			
		*Positive	3	100.00%
Cepheid GeneXpert	1	*Positive	1	100.00%
Alere C diff Quik Chek Complete	2	*Positive	2	100.00%
Strep B Ag Detect				
AD-6				
* Denotes Correct Response				
AG: All Methods	1			
		*Positive	1	100.00%
Cepheid GeneXpert	1	*Positive	1	100.00%
AD-7				
* Denotes Correct Response				
AG: All Methods	1			
		*Negative	1	100.00%
Cepheid GeneXpert	1	*Negative	1	100.00%
AD-8				
* Denotes Correct Response				
AG: All Methods	1			
		*Positive	1	100.00%
Cepheid GeneXpert	1	*Positive	1	100.00%
AD-9				
* Denotes Correct Response				
AG: All Methods	1			
		*Negative	1	100.00%
Cepheid GeneXpert	1	*Negative	1	100.00%
AD-10				
* Denotes Correct Response				
AG: All Methods	1			
		*Positive	1	100.00%
Cepheid GeneXpert	1	*Positive	1	100.00%
643 - ANA (Latex);760 - ANA (ELISA);757 - Special Immunology	Labs	Response	Number	Percent
ANA (ELISA) (International)				
AE-6				
* Denotes Correct Response				
PG: Dynex DS2	1			
		*Positive	1	100.00%
AE-7				
* Denotes Correct Response				
PG: Dynex DS2	1			
		*Positive	1	100.00%
AE-8				
* Denotes Correct Response				
PG: Dynex DS2	1			
		*Negative	1	100.00%
AE-9				
* Denotes Correct Response				
PG: Dynex DS2	1			
		*Negative	1	100.00%
AE-10				
* Denotes Correct Response				
PG: Dynex DS2	2			
		*Positive	1	50.00%
		Negative	1	50.00%



671 - Total IgE;674 - Diagnostic Allergy;765 - Allergen Testing (1 specimen)	Labs	Response	Number	Percent
Aspergillus fumigatus Alrqn				
* Denotes Correct Response				
AL-6(Q)				
AG: All Methods	7			
		*1	4	57.14%
		*1/0	2	28.57%
		3	1	14.29%
Hitachi CLA-1	6	*1	3	50.00%
Hycor Hytec 288	1	*1	1	100.00%
Hitachi CLA-1	6	*1/0	2	33.33%
Hitachi CLA-1		3	1	16.67%
Bermuda Grass Alrqn				
* Denotes Correct Response				
AL-6(Q)				
AG: All Methods	10			
		*2	5	50.00%
		*3	4	40.00%
		4	1	10.00%
Hitachi CLA-1	8	*2	3	37.50%
Hycor Hytec 288	1	*2	1	100.00%
Phadia ImmunoCAP 100/250/1000-kUA/L	1	*2	1	100.00%
Hitachi CLA-1	8	*3	4	50.00%
Hitachi CLA-1		4	1	12.50%
Cat Epithelium Alrqn				
* Denotes Correct Response				
AL-6(Q)				
AG: All Methods	8			
		*3	2	25.00%
		*4	5	62.50%
		2	1	12.50%
Hitachi CLA-1	6	*3	1	16.67%
Hycor Hytec 288	1	*3	1	100.00%
Hitachi CLA-1	6	*4	5	83.33%
Phadia ImmunoCAP 100/250/1000-kUA/L	1	2	1	100.00%
Common Ragweed Alrqn				
* Denotes Correct Response				
AL-6(Q)				
AG: All Methods	2			
		*2	2	100.00%
Hitachi CLA-1	1	*2	1	100.00%
Phadia ImmunoCAP 100/250/1000-kUA/L	1	*2	1	100.00%
Dog Dander Alrqn				
* Denotes Correct Response				
AL-6(Q)				
AG: All Methods	8			
		*2	2	25.00%
		*3	2	25.00%
		*4	3	37.50%
		0	1	12.50%
Hitachi CLA-1	6	*2	1	16.67%
Phadia ImmunoCAP 100/250/1000-kUA/L	1	*2	1	100.00%
Hitachi CLA-1	6	*3	2	33.33%
Hitachi CLA-1		*4	3	50.00%
Hycor Hytec 288	1	0	1	100.00%
Dog Epithelium Alrqn				
* Denotes Correct Response				
AL-6(Q)				
AG: All Methods	3			
		*2	1	33.33%
		*3	1	33.33%
		*4	1	33.33%
Hitachi CLA-1	3	*2	1	33.33%
Hitachi CLA-1		*3	1	33.33%
Hitachi CLA-1		*4	1	33.33%
English Plantain Alrqn				
* Denotes Correct Response				
AL-6(Q)				
AG: All Methods	5			
		*1	1	20.00%
		*2	3	60.00%
		4	1	20.00%
Hitachi CLA-1	5	*1	1	20.00%
Hitachi CLA-1		*2	3	60.00%



Hitachi CLA-1		4	1	20.00%
House Dust Mite Alrgn				
AL-6(Q)				
* Denotes Correct Response				
AG: All Methods	8			
		*2	4	50.00%
		*3	2	25.00%
		*4	1	12.50%
		1	1	12.50%
Hitachi CLA-1	7	*2	3	42.86%
Phadia ImmunoCAP 100/250/1000-kUA/L	1	*2	1	100.00%
Hitachi CLA-1	7	*3	2	28.57%
Hitachi CLA-1		*4	1	14.29%
Hitachi CLA-1		1	1	14.29%
Box Elder, Maple Alrgn				
AL-6(Q)				
* Denotes Correct Response				
AG: All Methods	6			
		*1	2	33.33%
		*1/0	2	33.33%
		*2	1	16.67%
		3	1	16.67%
Hitachi CLA-1	5	*1	2	40.00%
Hitachi CLA-1		*1/0	2	40.00%
Hycor Hytec 288	1	*2	1	100.00%
Hitachi CLA-1	5	3	1	20.00%
Oak Alrgn				
AL-6(Q)				
* Denotes Correct Response				
AG: All Methods	7			
		*2	5	71.43%
		*3	2	28.57%
Hitachi CLA-1	5	*2	3	60.00%
Hycor Hytec 288	1	*2	1	100.00%
Phadia ImmunoCAP 100/250/1000-kUA/L	1	*2	1	100.00%
Hitachi CLA-1	5	*3	2	40.00%
Peanut (Arachis hyogaea) Alrgn				
AL-6(Q)				
* Denotes Correct Response				
AG: All Methods	8			
		*3	4	50.00%
		*4	3	37.50%
		2	1	12.50%
Hitachi CLA-1	6	*3	3	50.00%
Hycor Hytec 288	1	*3	1	100.00%
Hitachi CLA-1	6	*4	3	50.00%
Phadia ImmunoCAP 100/250/1000-kUA/L	1	2	1	100.00%
Russian Thistle Alrgn				
AL-6(Q)				
* Denotes Correct Response				
AG: All Methods	2			
		*2	2	100.00%
Hycor Hytec 288	1	*2	1	100.00%
Phadia ImmunoCAP 100/250/1000-kUA/L	1	*2	1	100.00%
Soybean Alrgn				
AL-6(Q)				
* Denotes Correct Response				
AG: All Methods	8			
		*1	3	37.50%
		*2	5	62.50%
Hitachi CLA-1	6	*1	2	33.33%
Phadia ImmunoCAP 100/250/1000-kUA/L	1	*1	1	100.00%
Hitachi CLA-1	6	*2	4	66.67%
Hycor Hytec 288	1	*2	1	100.00%
Timothy Grass Alrgn				
AL-6(Q)				
* Denotes Correct Response				
AG: All Methods	8			
		*3	1	12.50%
		*4	7	87.50%
Phadia ImmunoCAP 100/250/1000-kUA/L	1	*3	1	100.00%
Hitachi CLA-1	7	*4	7	100.00%
Alternaria alternata Alrgn				
AL-6(Q)				
* Denotes Correct Response				
AG: All Methods	10			



		*1	8	80.00%
		*2	1	10.00%
		4	1	10.00%
Hitachi CLA-1	8	*1	6	75.00%
Hycor Hytec 288	1	*1	1	100.00%
Phadia ImmunoCAP 100/250/1000-kUA/L	1	*1	1	100.00%
Hitachi CLA-1	8	*2	1	12.50%
Hitachi CLA-1		4	1	12.50%
Almond Alrgn				
AL-6(Q)		* Denotes Correct Response		
AG: All Methods	4			
		*1	3	75.00%
		*2	1	25.00%
Hitachi CLA-1	4	*1	3	75.00%
Hitachi CLA-1		*2	1	25.00%
Apple Alrgn				
AL-6(Q)		* Denotes Correct Response		
AG: All Methods	4			
		*0	2	50.00%
		*1	2	50.00%
Hitachi CLA-1	4	*0	2	50.00%
Hitachi CLA-1		*1	2	50.00%
Ash - White Alrgn				
AL-6(Q)		* Denotes Correct Response		
AG: All Methods	6			
		*0	1	16.67%
		*1/0	3	50.00%
		*2	1	16.67%
		3	1	16.67%
Hitachi CLA-1	5	*0	1	20.00%
Hitachi CLA-1		*1/0	3	60.00%
Hycor Hytec 288	1	*2	1	100.00%
Hitachi CLA-1	5	3	1	20.00%
Barley Alrgn				
AL-6(Q)		* Denotes Correct Response		
AG: All Methods	5			
		*0	2	40.00%
		*1	2	40.00%
		2	1	20.00%
Hitachi CLA-1	4	*0	2	50.00%
Hitachi CLA-1		*1	2	50.00%
Hycor Hytec 288	1	2	1	100.00%
Elm - White Alrgn				
AL-6(Q)		* Denotes Correct Response		
AG: All Methods	6			
		*1	2	33.33%
		*1/0	3	50.00%
		4	1	16.67%
Hitachi CLA-1	6	*1	2	33.33%
Hitachi CLA-1		*1/0	3	50.00%
Hitachi CLA-1		4	1	16.67%
Beef Alrgn				
AL-6(Q)		* Denotes Correct Response		
AG: All Methods	4			
		*0	2	50.00%
		*1	2	50.00%
Hitachi CLA-1	4	*0	2	50.00%
Hitachi CLA-1		*1	2	50.00%
Birch/Alder Mix Alrgn				
AL-6(Q)		* Denotes Correct Response		
AG: All Methods	6			
		*1	5	83.33%
		4	1	16.67%
Hitachi CLA-1	6	*1	5	83.33%
Hitachi CLA-1		4	1	16.67%
Cedar - Mountain Alrgn				
AL-6(Q)		* Denotes Correct Response		
AG: All Methods	5			



		*1	1	20.00%
		*1/0	3	60.00%
		3	1	20.00%
Hitachi CLA-1	5	*1	1	20.00%
Hitachi CLA-1		*1/0	3	60.00%
Hitachi CLA-1		3	1	20.00%
Chicken Alrgn				
AL-6(Q)				
* Denotes Correct Response				
AG: All Methods	4			
		*0	3	75.00%
		*1/0	1	25.00%
Hitachi CLA-1	4	*0	3	75.00%
Hitachi CLA-1		*1/0	1	25.00%
Chocolate Alrgn				
AL-6(Q)				
* Denotes Correct Response				
AG: All Methods	4			
		*0	3	75.00%
		*1/0	1	25.00%
Hitachi CLA-1	4	*0	3	75.00%
Hitachi CLA-1		*1/0	1	25.00%
Cladosporium Alrgn				
AL-6(Q)				
* Denotes Correct Response				
AG: All Methods	7			
		*0	1	14.29%
		*1	3	42.86%
		*1/0	2	28.57%
		3	1	14.29%
Hycor Hytec 288	1	*0	1	100.00%
Hitachi CLA-1	6	*1	3	50.00%
Hitachi CLA-1		*1/0	2	33.33%
Hitachi CLA-1		3	1	16.67%
Cocklebur Alrgn				
AL-6(Q)				
* Denotes Correct Response				
AG: All Methods	6			
		*1	3	50.00%
		*1/0	2	33.33%
		3	1	16.67%
Hitachi CLA-1	6	*1	3	50.00%
Hitachi CLA-1		*1/0	2	33.33%
Hitachi CLA-1		3	1	16.67%
Oat Alrgn				
AL-6(Q)				
* Denotes Correct Response				
AG: All Methods	4			
		*0	1	25.00%
		*1	2	50.00%
		*1/0	1	25.00%
Hitachi CLA-1	4	*0	1	25.00%
Hitachi CLA-1		*1	2	50.00%
Hitachi CLA-1		*1/0	1	25.00%
Corn Alrgn				
AL-6(Q)				
* Denotes Correct Response				
AG: All Methods	6			
		*0	1	16.67%
		*1	4	66.67%
		*2	1	16.67%
Hitachi CLA-1	5	*0	1	20.00%
Hitachi CLA-1		*1	4	80.00%
Hycor Hytec 288	1	*2	1	100.00%
Egg - whole Alrgn				
AL-6(Q)				
* Denotes Correct Response				
AG: All Methods	6			
		*2	2	33.33%
		*3	3	50.00%
		1	1	16.67%
Hitachi CLA-1	5	*2	1	20.00%
Hycor Hytec 288	1	*2	1	100.00%
Hitachi CLA-1	5	*3	3	60.00%
Hitachi CLA-1		1	1	20.00%



AL-6(Q)		Garlic Alrgn		
AG: All Methods		* Denotes Correct Response		
	4			
		*1	2	50.00%
		*1/0	1	25.00%
		*2	1	25.00%
Hitachi CLA-1	4	*1	2	50.00%
Hitachi CLA-1		*1/0	1	25.00%
Hitachi CLA-1		*2	1	25.00%
AL-6(Q)		Lambs Quarters Alrgn		
AG: All Methods		* Denotes Correct Response		
	7			
		*1	2	28.57%
		*2	3	42.86%
		*3	2	28.57%
Hitachi CLA-1	6	*1	2	33.33%
Hitachi CLA-1		*2	3	50.00%
Hitachi CLA-1		*3	1	16.67%
Hycor Hytec 288	1	*3	1	100.00%
AL-6(Q)		Marshelder - Rough Alrgn		
AG: All Methods		* Denotes Correct Response		
	6			
		*1	3	50.00%
		*2	2	33.33%
		4	1	16.67%
Hitachi CLA-1	5	*1	3	60.00%
Hitachi CLA-1		*2	1	20.00%
Hycor Hytec 288	1	*2	1	100.00%
Hitachi CLA-1	5	4	1	20.00%
AL-6(Q)		Milk Alrgn		
AG: All Methods		* Denotes Correct Response		
	5			
		*1	2	40.00%
		*2	3	60.00%
Hitachi CLA-1	5	*1	2	40.00%
Hitachi CLA-1		*2	3	60.00%
AL-6(Q)		Mite D Farinae Alrgn		
AG: All Methods		* Denotes Correct Response		
	7			
		*4	7	100.00%
Hitachi CLA-1	7	*4	7	100.00%
AL-6(Q)		Mulberry Mix Alrgn		
AG: All Methods		* Denotes Correct Response		
	6			
		*0	3	50.00%
		*1	1	16.67%
		*1/0	1	16.67%
		3	1	16.67%
Hitachi CLA-1	6	*0	3	50.00%
Hitachi CLA-1		*1	1	16.67%
Hitachi CLA-1		*1/0	1	16.67%
Hitachi CLA-1		3	1	16.67%
AL-6(Q)		Onion Mix Alrgn		
AG: All Methods		* Denotes Correct Response		
	4			
		*0	2	50.00%
		*1/0	2	50.00%
Hitachi CLA-1	4	*0	2	50.00%
Hitachi CLA-1		*1/0	2	50.00%
AL-6(Q)		Orange Alrgn		
AG: All Methods		* Denotes Correct Response		
	6			
		*1	3	50.00%
		*2	3	50.00%
Hitachi CLA-1	5	*1	3	60.00%
Hitachi CLA-1		*2	2	40.00%
Hycor Hytec 288	1	*2	1	100.00%
AL-6(Q)		Pigweed Alrgn		



AL-6(Q)		* Denotes Correct Response		
AG: All Methods	6			
		*1	3	50.00%
		*1/0	2	33.33%
		*2	1	16.67%
Hitachi CLA-1	6	*1	3	50.00%
Hitachi CLA-1		*1/0	2	33.33%
Hitachi CLA-1		*2	1	16.67%
AL-6(Q)		Pine Mix Alrqn		
AL-6(Q)		* Denotes Correct Response		
AG: All Methods	6			
		*0	1	16.67%
		*1	1	16.67%
		*1/0	3	50.00%
		3	1	16.67%
Hitachi CLA-1	6	*0	1	16.67%
Hitachi CLA-1		*1	1	16.67%
Hitachi CLA-1		*1/0	3	50.00%
Hitachi CLA-1		3	1	16.67%
AL-6(Q)		Pork Alrqn		
AL-6(Q)		* Denotes Correct Response		
AG: All Methods	5			
		*1	4	80.00%
		*2	1	20.00%
Hitachi CLA-1	4	*1	3	75.00%
Hycor Hytec 288	1	*1	1	100.00%
Hitachi CLA-1	4	*2	1	25.00%
AL-6(Q)		Potato Alrqn		
AL-6(Q)		* Denotes Correct Response		
AG: All Methods	4			
		*1	2	50.00%
		*2	2	50.00%
Hitachi CLA-1	4	*1	2	50.00%
Hitachi CLA-1		*2	2	50.00%
AL-6(Q)		Privet Alrqn		
AL-6(Q)		* Denotes Correct Response		
AG: All Methods	6			
		*0	1	16.67%
		*1	1	16.67%
		*1/0	3	50.00%
		2	1	16.67%
Hitachi CLA-1	6	*0	1	16.67%
Hitachi CLA-1		*1	1	16.67%
Hitachi CLA-1		*1/0	3	50.00%
Hitachi CLA-1		2	1	16.67%
AL-6(Q)		Ragweed - Short Alrqn		
AL-6(Q)		* Denotes Correct Response		
AG: All Methods	7			
		*2	3	42.86%
		*3	3	42.86%
		4	1	14.29%
Hitachi CLA-1	6	*2	3	50.00%
Hitachi CLA-1		*3	2	33.33%
Hycor Hytec 288	1	*3	1	100.00%
Hitachi CLA-1	6	4	1	16.67%
AL-6(Q)		Rice Alrqn		
AL-6(Q)		* Denotes Correct Response		
AG: All Methods	6			
		*1	2	33.33%
		*2	2	33.33%
		*3	1	16.67%
		1/0	1	16.67%
Hitachi CLA-1	5	*1	2	40.00%
Hitachi CLA-1		*2	1	20.00%
Hycor Hytec 288	1	*2	1	100.00%
Hitachi CLA-1	5	*3	1	20.00%
Hitachi CLA-1		1/0	1	20.00%
AL-6(Q)		Sheep Sorrel Alrqn		



AL-6(Q)		* Denotes Correct Response		
AG: All Methods	6			
		*2	1	16.67%
		*3	4	66.67%
		*4	1	16.67%
Hitachi CLA-1	6	*2	1	16.67%
Hitachi CLA-1		*3	4	66.67%
Hitachi CLA-1		*4	1	16.67%
AL-6(Q)		Shellfish Mix Alrqn		
AL-6(Q)		* Denotes Correct Response		
AG: All Methods	5			
		*2	2	40.00%
		*3	2	40.00%
		*4	1	20.00%
Hitachi CLA-1	5	*2	2	40.00%
Hitachi CLA-1		*3	2	40.00%
Hitachi CLA-1		*4	1	20.00%
AL-6(Q)		Sycamore - American Alrqn		
AL-6(Q)		* Denotes Correct Response		
AG: All Methods	6			
		*0	3	50.00%
		*1/0	2	33.33%
		2	1	16.67%
Hitachi CLA-1	6	*0	3	50.00%
Hitachi CLA-1		*1/0	2	33.33%
Hitachi CLA-1		2	1	16.67%
AL-6(Q)		Tomato Alrqn		
AL-6(Q)		* Denotes Correct Response		
AG: All Methods	6			
		*1	1	16.67%
		*2	5	83.33%
Hitachi CLA-1	5	*1	1	20.00%
Hitachi CLA-1		*2	4	80.00%
Hycor Hytec 288	1	*2	1	100.00%
AL-6(Q)		Tuna Alrqn		
AL-6(Q)		* Denotes Correct Response		
AG: All Methods	4			
		*0	1	25.00%
		*1/0	3	75.00%
Hitachi CLA-1	4	*0	1	25.00%
Hitachi CLA-1		*1/0	3	75.00%
AL-6(Q)		Vegetable Mix Alrqn		
AL-6(Q)		* Denotes Correct Response		
AG: All Methods	4			
		*1	4	100.00%
Hitachi CLA-1	4	*1	4	100.00%
AL-6(Q)		Walnut/Hickory/Pecan Mix Alrqn		
AL-6(Q)		* Denotes Correct Response		
AG: All Methods	6			
		*1	3	50.00%
		*1/0	2	33.33%
		4	1	16.67%
Hitachi CLA-1	6	*1	3	50.00%
Hitachi CLA-1		*1/0	2	33.33%
Hitachi CLA-1		4	1	16.67%
AL-6(Q)		Waterhemp Alrqn		
AL-6(Q)		* Denotes Correct Response		
AG: All Methods	4			
		*3	3	75.00%
		*4	1	25.00%
Hitachi CLA-1	4	*3	3	75.00%
Hitachi CLA-1		*4	1	25.00%
AL-6(Q)		Wheat Alrqn		
AL-6(Q)		* Denotes Correct Response		
AG: All Methods	5			
		*1	3	60.00%
		*2	2	40.00%
Hitachi CLA-1	4	*1	3	75.00%



Hitachi CLA-1		*2	1	25.00%
Hycor HYTEC 288	1	*2	1	100.00%
Yeast - Bakers Alrgn				
AL-6(Q) * Denotes Correct Response				
AG: All Methods	5			
		*0	4	80.00%
		*1/0	1	20.00%
Hitachi CLA-1	5	*0	4	80.00%
Hitachi CLA-1		*1/0	1	20.00%
Acacia Alrgn				
AL-6(Q) * Denotes Correct Response				
AG: All Methods	2			
		*1/0	1	50.00%
		*2	1	50.00%
Hitachi CLA-1	2	*1/0	1	50.00%
Hitachi CLA-1		*2	1	50.00%
Bahia Grass Alrgn				
AL-6(Q) * Denotes Correct Response				
AG: All Methods	2			
		*3	1	50.00%
		*4	1	50.00%
Hitachi CLA-1	2	*3	1	50.00%
Hitachi CLA-1		*4	1	50.00%
Candida Alrgn				
AL-6(Q) * Denotes Correct Response				
AG: All Methods	2			
		*1/0	1	50.00%
		*2	1	50.00%
Hitachi CLA-1	2	*1/0	1	50.00%
Hitachi CLA-1		*2	1	50.00%
Cat Dander Alrgn				
AL-6(Q) * Denotes Correct Response				
AG: All Methods	2			
		*4	2	100.00%
Hitachi CLA-1	2	*4	2	100.00%
683 - ASO/CRP;759 - Complete Immunology;729 - AFC Immunology				
	Labs	Response	Number	Percent
ASO (Qual) (International)				
AS-6 * Denotes Correct Response				
PG: Wampole ColorCard ASO	1			
		*Negative	1	100.00%
AS-7 * Denotes Correct Response				
PG: Wampole ColorCard ASO	1			
		*Negative	1	100.00%
AS-8 * Denotes Correct Response				
PG: Wampole ColorCard ASO	1			
		*Positive	1	100.00%
AS-9 * Denotes Correct Response				
PG: Wampole ColorCard ASO	1			
		*Positive	1	100.00%
AS-10 * Denotes Correct Response				
PG: Wampole ColorCard ASO	1			
		*Positive	1	100.00%
ASO (Semi-Quant)				
AS-6 * Denotes Correct Response				
PG: Fisher HC Sure-Vue ASO	1			
AS-7 * Denotes Correct Response				
PG: Fisher HC Sure-Vue ASO	1			
AS-8 * Denotes Correct Response				
PG: Fisher HC Sure-Vue ASO	1			
AS-9 * Denotes Correct Response				
PG: Fisher HC Sure-Vue ASO	1			
AS-10 * Denotes Correct Response				
PG: Fisher HC Sure-Vue ASO	1			
ASO (Qual)				
AS-6 * Denotes Correct Response				
PG: Fisher HC Sure-Vue ASO	1			



			Number	Percent
		*Negative	1	100.00%
AS-7		* Denotes Correct Response		
PG: Fisher HC Sure-Vue ASO	1			
		*Negative	1	100.00%
AS-8		* Denotes Correct Response		
PG: Fisher HC Sure-Vue ASO	1			
		*Positive	1	100.00%
AS-9		* Denotes Correct Response		
PG: Fisher HC Sure-Vue ASO	1			
		*Positive	1	100.00%
AS-10		* Denotes Correct Response		
PG: Fisher HC Sure-Vue ASO	1			
		*Positive	1	100.00%
782 - Complete Bacteriology	Labs	Response	Number	Percent
		Ear Culture		
BA-4		* Denotes Correct Response		
AG: All Methods	1			
		*Moraxella catarrhalis (Branhamella)	1	100.00%
		Staphylococcus, coagulase negative	1	100.00%
Cultures	1	*Moraxella catarrhalis (Branhamella)	1	100.00%
Cultures		Staphylococcus, coagulase negative	1	100.00%
		CSF Culture		
BA-6		* Denotes Correct Response		
AG: All Methods	1			
		*Haemophilus species	1	100.00%
Cultures	1	*Haemophilus species	1	100.00%
		Wound Culture		
BA-5		* Denotes Correct Response		
AG: All Methods	2			
		*Staphylococcus aureus	1	50.00%
		*Staphylococcus, coagulase positive	1	50.00%
Cultures	2	*Staphylococcus aureus	1	50.00%
Cultures		*Staphylococcus, coagulase positive	1	50.00%
650 - Blood Bank; 736 - ABO/Rh Only	Labs	Response	Number	Percent
		ABO Group		
BB-6		* Denotes Correct Response		
AG: All Methods	1			
		*Group B	1	100.00%
Blood Bank Testing	1	*Group B	1	100.00%
BB-7		* Denotes Correct Response		
AG: All Methods	1			
		*Group O	1	100.00%
Blood Bank Testing	1	*Group O	1	100.00%
BB-8		* Denotes Correct Response		
AG: All Methods	1			
		*Group A	1	100.00%
Blood Bank Testing	1	*Group A	1	100.00%
BB-9		* Denotes Correct Response		
AG: All Methods	1			
		*Group B	1	100.00%
Blood Bank Testing	1	*Group B	1	100.00%
BB-10		* Denotes Correct Response		
AG: All Methods	1			
		*Group A	1	100.00%
Blood Bank Testing	1	*Group A	1	100.00%
		Rh Type		
BB-6		* Denotes Correct Response		
AG: All Methods	9			
		*Rh Positive	9	100.00%
Blood Bank Testing	9	*Rh Positive	9	100.00%
BB-7		* Denotes Correct Response		
AG: All Methods	9			
		*Rh Negative	9	100.00%
Blood Bank Testing	9	*Rh Negative	9	100.00%
BB-8		* Denotes Correct Response		
AG: All Methods	9			
		*Rh Positive	9	100.00%



Blood Bank Testing	9	*Rh Positive	9	100.00%
BB-9		* Denotes Correct Response		
AG: All Methods	9			
		*Rh Positive	9	100.00%
Blood Bank Testing	9	*Rh Positive	9	100.00%
BB-10		* Denotes Correct Response		
AG: All Methods	9			
		*Rh Negative	9	100.00%
Blood Bank Testing	9	*Rh Negative	9	100.00%
661 - Urine Colony Count;784 - Presump ID/Colony Count - Urine;785 - Presump ID/Colony Count-Urine & Strep Combo;775 - Urine Culture;778 - Urine & Throat Cultures;779 - Urine & Genital Cultures;781 - Urine, Throat & Genital Cultures;782 - Complete Bacteri	Labs	Response	Number	Percent
		Colony Count		
CC-6		* Denotes Correct Response		
AG: All Methods	31			
		*Greater than 100,000	25	80.65%
		10,000 - 100,000	5	16.13%
		No growth	1	3.23%
Calibrated Loop - Manual	8	*Greater than 100,000	7	87.50%
Uri-Chek (Troy Bio)	2	*Greater than 100,000	1	50.00%
Uricult (Orion)	19	*Greater than 100,000	15	78.95%
Uri-Three (Culture Kits)	1	*Greater than 100,000	1	100.00%
Uricult (Lifesign)	1	*Greater than 100,000	1	100.00%
Calibrated Loop - Manual	8	10,000 - 100,000	1	12.50%
Uri-Chek (Troy Bio)	2	10,000 - 100,000	1	50.00%
Uricult (Orion)	19	10,000 - 100,000	3	15.79%
Uricult (Orion)		No growth	1	5.26%
784 - Presump ID/Colony Count - Urine;785 - Presump ID/Colony Count-Urine & Strep Combo	Labs	Response	Number	Percent
		Colony Count		
CC-6		* Denotes Correct Response		
AG: All Methods	31			
		*Greater than 100,000	25	80.65%
		10,000 - 100,000	5	16.13%
		No growth	1	3.23%
Calibrated Loop - Manual	8	*Greater than 100,000	7	87.50%
Uri-Chek (Troy Bio)	2	*Greater than 100,000	1	50.00%
Uricult (Orion)	19	*Greater than 100,000	15	78.95%
Uri-Three (Culture Kits)	1	*Greater than 100,000	1	100.00%
Uricult (Lifesign)	1	*Greater than 100,000	1	100.00%
Calibrated Loop - Manual	8	10,000 - 100,000	1	12.50%
Uri-Chek (Troy Bio)	2	10,000 - 100,000	1	50.00%
Uricult (Orion)	19	10,000 - 100,000	3	15.79%
Uricult (Orion)		No growth	1	5.26%
CC-7		* Denotes Correct Response		
AG: All Methods	6			
		*10,000 - 100,000	2	33.33%
		*Greater than 100,000	3	50.00%
		Less than 10,000	1	16.67%
Uri-Chek (Troy Bio)	1	*10,000 - 100,000	1	100.00%
Uricult (Orion)	2	*10,000 - 100,000	1	50.00%
Calibrated Loop - Manual	3	*Greater than 100,000	2	66.67%
Uricult (Orion)	2	*Greater than 100,000	1	50.00%
Calibrated Loop - Manual	3	Less than 10,000	1	33.33%
CC-8		* Denotes Correct Response		
AG: All Methods	6			
		*10,000 - 100,000	2	33.33%
		*Greater than 100,000	2	33.33%
		Less than 10,000	1	16.67%
		No growth	1	16.67%
Calibrated Loop - Manual	3	*10,000 - 100,000	1	33.33%
Uricult (Orion)	2	*10,000 - 100,000	1	50.00%
Calibrated Loop - Manual	3	*Greater than 100,000	1	33.33%
Uricult (Orion)	2	*Greater than 100,000	1	50.00%
Uri-Chek (Troy Bio)	1	Less than 10,000	1	100.00%
Calibrated Loop - Manual	3	No growth	1	33.33%
CC-9		* Denotes Correct Response		



AG: All Methods	6			
		*Greater than 100,000	6	100.00%
Calibrated Loop - Manual	3	*Greater than 100,000	3	100.00%
Uri-Chek (Troy Bio)	1	*Greater than 100,000	1	100.00%
Uricult (Orion)	2	*Greater than 100,000	2	100.00%
CC-10		* Denotes Correct Response		
AG: All Methods	6			
		*10,000 - 100,000	4	66.67%
		*Greater than 100,000	2	33.33%
Calibrated Loop - Manual	3	*10,000 - 100,000	2	66.67%
Uri-Chek (Troy Bio)	1	*10,000 - 100,000	1	100.00%
Uricult (Orion)	2	*10,000 - 100,000	1	50.00%
Calibrated Loop - Manual	3	*Greater than 100,000	1	33.33%
Uricult (Orion)	2	*Greater than 100,000	1	50.00%
		Urine Presump Id		
CC-6		* Denotes Correct Response		
AG: All Methods	3			
		*Presumptive Klebsiella sp.	2	66.67%
		Presumptive E. Coli	1	33.33%
Calibrated Loop - Manual	2	*Presumptive Klebsiella sp.	2	100.00%
Uricult (Orion)	1	Presumptive E. Coli	1	100.00%
CC-7		* Denotes Correct Response		
AG: All Methods	3			
		*Gram Positive Organism	1	33.33%
		*Presumptive Enterococcus sp.	1	33.33%
		Presumptive Proteus sp.	1	33.33%
Uricult (Orion)	1	*Gram Positive Organism	1	100.00%
Calibrated Loop - Manual	2	*Presumptive Enterococcus sp.	1	50.00%
Calibrated Loop - Manual		Presumptive Proteus sp.	1	50.00%
CC-8		* Denotes Correct Response		
AG: All Methods	3			
		*Gram Negative Organsim	1	33.33%
		*Presumptive Proteus sp.	2	66.67%
Uricult (Orion)	1	*Gram Negative Organsim	1	100.00%
Calibrated Loop - Manual	2	*Presumptive Proteus sp.	2	100.00%
CC-9		* Denotes Correct Response		
AG: All Methods	3			
		*Presumptive E. Coli	3	100.00%
Calibrated Loop - Manual	2	*Presumptive E. Coli	2	100.00%
Uricult (Orion)	1	*Presumptive E. Coli	1	100.00%
CC-10		* Denotes Correct Response		
AG: All Methods	3			
		*Gram Negative Organsim	1	33.33%
		*Presumptive Enterobacter sp.	1	33.33%
		*Presumptive Klebsiella sp.	1	33.33%
Uricult (Orion)	1	*Gram Negative Organsim	1	100.00%
Calibrated Loop - Manual	2	*Presumptive Enterobacter sp.	1	50.00%
Calibrated Loop - Manual		*Presumptive Klebsiella sp.	1	50.00%
		Susceptibility Amikacin		
CC-6		* Denotes Correct Response		
AG: All Methods	2			
		*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
		Susceptibility Amox/Clavulanate		
CC-6		* Denotes Correct Response		
AG: All Methods	6			
		*Sensitive	6	100.00%
Susceptibility	6	*Sensitive	6	100.00%
		Susceptibility Ampicillin		
CC-6		* Denotes Correct Response		
AG: All Methods	8			
		*Resistant	7	87.50%
		Sensitive	1	12.50%
Susceptibility	8	*Resistant	7	87.50%
Susceptibility		Sensitive	1	12.50%
		Susceptibility Ampicillin/Sulbactam		
CC-6		* Denotes Correct Response		
AG: All Methods	2			



		*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
		Susceptibility Aztreonam		
CC-6		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
		Susceptibility Cefazolin		
CC-6		* Denotes Correct Response		
AG: All Methods	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
		Susceptibility Cefepime		
CC-6		* Denotes Correct Response		
AG: All Methods	2			
		*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
		Susceptibility Cefotaxime		
CC-6		* Denotes Correct Response		
AG: All Methods	2			
		*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
		Susceptibility Cefpodoxime		
CC-6		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
		Susceptibility Ceftazidime		
CC-6		* Denotes Correct Response		
AG: All Methods	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
		Susceptibility Ceftriaxone		
CC-6		* Denotes Correct Response		
AG: All Methods	6			
		*Sensitive	6	100.00%
Susceptibility	6	*Sensitive	6	100.00%
		Susceptibility Cefuroxime		
CC-6		* Denotes Correct Response		
AG: All Methods	4			
		*Sensitive	4	100.00%
Susceptibility	4	*Sensitive	4	100.00%
		Susceptibility Cephalothin		
CC-6		* Denotes Correct Response		
AG: All Methods	6			
		*Sensitive	6	100.00%
Susceptibility	6	*Sensitive	6	100.00%
		Susceptibility Ciprofloxacin		
CC-6		* Denotes Correct Response		
AG: All Methods	10			
		*Sensitive	10	100.00%
Susceptibility	10	*Sensitive	10	100.00%
		Susceptibility Gentamicin		
CC-6		* Denotes Correct Response		
AG: All Methods	6			
		*Sensitive	6	100.00%
Susceptibility	6	*Sensitive	6	100.00%
		Susceptibility Imipenem		
CC-6		* Denotes Correct Response		
AG: All Methods	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
		Susceptibility Levofloxacin		
CC-6		* Denotes Correct Response		
AG: All Methods	6			
		*Sensitive	6	100.00%
Susceptibility	6	*Sensitive	6	100.00%



CC-6		Susceptibility Meropenem		
AG: All Methods		* Denotes Correct Response		
	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
CC-6		Susceptibility Minocycline		
AG: All Methods		* Denotes Correct Response		
	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
CC-6		Susceptibility Nitrofurantoin		
AG: All Methods		* Denotes Correct Response		
	9			
		*Intermediate	4	44.44%
		*Sensitive	5	55.56%
Susceptibility	9	*Intermediate	4	44.44%
Susceptibility		*Sensitive	5	55.56%
CC-6		Susceptibility Norfloxacin/Noroxin		
AG: All Methods		* Denotes Correct Response		
	2			
		*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
CC-6		Susceptibility Piperacillin/Tazobactam		
AG: All Methods		* Denotes Correct Response		
	2			
		*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
CC-6		Susceptibility Sulfamethoxazole		
AG: All Methods		* Denotes Correct Response		
	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
CC-6		Susceptibility Tetracycline		
AG: All Methods		* Denotes Correct Response		
	6			
		*Sensitive	5	83.33%
		Resistant	1	16.67%
Susceptibility	6	*Sensitive	5	83.33%
Susceptibility		Resistant	1	16.67%
CC-6		Susceptibility Ticarcillin/Clavulanate		
AG: All Methods		* Denotes Correct Response		
	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
CC-6		Susceptibility Tobramycin		
AG: All Methods		* Denotes Correct Response		
	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
CC-6		Susceptibility Trimethoprim		
AG: All Methods		* Denotes Correct Response		
	2			
		*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
CC-6		Susceptibility Trimeth./Sulfameth.		
AG: All Methods		* Denotes Correct Response		
	8			
		*Sensitive	8	100.00%
Susceptibility	8	*Sensitive	8	100.00%
CC-6		Susceptibility Ertapenem		
AG: All Methods		* Denotes Correct Response		
	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
714 - Complete Cardiac Markers;715 - Basic Cardiac Markers;720 - Cardiac Markers I	Labs	Response	Number	Percent
		Myoglobin (Qualitative)		



CDM-6		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	1	50.00%
		Negative	1	50.00%
Princeton Biomedical Corp. Lifesign MI	2	*Positive	1	50.00%
Princeton Biomedical Corp. Lifesign MI		Negative	1	50.00%
CDM-7		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Princeton Biomedical Corp. Lifesign MI	2	*Positive	2	100.00%
CDM-8		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Princeton Biomedical Corp. Lifesign MI	2	*Negative	2	100.00%
CDM-9		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	1	50.00%
		Negative	1	50.00%
Princeton Biomedical Corp. Lifesign MI	2	*Positive	1	50.00%
Princeton Biomedical Corp. Lifesign MI		Negative	1	50.00%
CDM-10		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Princeton Biomedical Corp. Lifesign MI	2	*Positive	2	100.00%
CK-MB (Qualitative)				
CDM-6		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Princeton Biomedical Corp. Lifesign MI	2	*Positive	2	100.00%
CDM-7		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Princeton Biomedical Corp. Lifesign MI	2	*Positive	2	100.00%
CDM-8		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Princeton Biomedical Corp. Lifesign MI	2	*Positive	2	100.00%
CDM-9		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Princeton Biomedical Corp. Lifesign MI	2	*Positive	2	100.00%
CDM-10		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Princeton Biomedical Corp. Lifesign MI	2	*Positive	2	100.00%
Troponin I (Qualitative)				
CDM-6		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Princeton Biomedical Corp. Lifesign MI	2	*Positive	2	100.00%
CDM-7		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Princeton Biomedical Corp. Lifesign MI	2	*Positive	2	100.00%
CDM-8		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Princeton Biomedical Corp. Lifesign MI	2	*Negative	2	100.00%
CDM-9		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Princeton Biomedical Corp. Lifesign MI	2	*Positive	2	100.00%
CDM-10		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Princeton Biomedical Corp. Lifesign MI	2	*Positive	2	100.00%



639 - Complete Urinalysis Package;638 - Urine Sediment;641 - Clinical Microscopy;601 - Waived/PPM Tests II;609 - Waived/PPM Tests I;646 - Clinical Microscopy Extra;628 - AFC Waived/PPM	Labs	Response	Number	Percent
Urine Sediment Id				
CM-10				
* Denotes Correct Response				
AG: All Methods	820			
		*Ammonium magnesium phosphate (Triple phosphate)	700	85.37%
		Ammonium biurate crystal	2	0.24%
		Amorphous urates/phosphates	16	1.95%
		Calcium oxalate crystal	17	2.07%
		Calcium phosphate crystal	34	4.15%
		Cholesterol crystal	4	0.49%
		Granular cast	1	0.12%
		Hippuric acid	8	0.98%
		Hyaline cast	2	0.24%
		Leucine crystal	1	0.12%
		Red blood cell (RBC) (Erythrocyte)	2	0.24%
		RTE with fat globules (oval fat bodies)	1	0.12%
		Sulfonamide crystal	1	0.12%
		Uric acid crystal	30	3.66%
		Waxy cast	1	0.12%
Microscopy	820	*Ammonium magnesium phosphate (Triple phosphate)	700	85.37%
Microscopy		Ammonium biurate crystal	2	0.24%
Microscopy		Amorphous urates/phosphates	16	1.95%
Microscopy		Calcium oxalate crystal	17	2.07%
Microscopy		Calcium phosphate crystal	34	4.15%
Microscopy		Cholesterol crystal	4	0.49%
Microscopy		Granular cast	1	0.12%
Microscopy		Hippuric acid	8	0.98%
Microscopy		Hyaline cast	2	0.24%
Microscopy		Leucine crystal	1	0.12%
Microscopy		Red blood cell (RBC) (Erythrocyte)	2	0.24%
Microscopy		RTE with fat globules (oval fat bodies)	1	0.12%
Microscopy		Sulfonamide crystal	1	0.12%
Microscopy		Uric acid crystal	30	3.66%
Microscopy		Waxy cast	1	0.12%
CM-11				
* Denotes Correct Response				
AG: All Methods	818			
		*Amorphous urates/phosphates	601	73.47%
		*Ident Unknown - Referred	43	5.26%
		Ammonium biurate crystal	9	1.10%
		Ammonium magnesium phosphate (Triple phosphate)	19	2.32%
		Bacteria	47	5.75%
		Bilirubin crystal	1	0.12%
		Calcium carbonate	11	1.34%
		Calcium oxalate crystal	2	0.24%
		Cellular (WBC) cast	1	0.12%
		Fat droplets or globules	4	0.49%
		Fiber (exogenous)/fecal contamination	12	1.47%
		Granular cast	1	0.12%
		Hippuric acid	1	0.12%
		Mucous strands	3	0.37%
		Pollen grain	2	0.24%
		RBC/Blood/Hgb cast	1	0.12%
		Red blood cell (RBC) (Erythrocyte)	12	1.47%
		Red Blood Cell (RBC) (Erythrocyte)	1	0.12%
		Renal tubular epithelial (RTE)	6	0.73%
		RTE with fat globules (oval fat bodies)	1	0.12%
		Squamous epithelial cell	5	0.61%
		Starch granules	5	0.61%
		Transitional epithelial cell (urothelial)	8	0.98%
		Uric acid crystal	2	0.24%
		White blood cell (WBC) (Leukocyte)	17	2.08%
		Yeast/fungi	3	0.37%
Microscopy	818	*Amorphous urates/phosphates	601	73.47%
Microscopy		*Ident Unknown - Referred	43	5.26%
Microscopy		Ammonium biurate crystal	9	1.10%
Microscopy		Ammonium magnesium phosphate (Triple phosphate)	19	2.32%



Microscopy		Bacteria	47	5.75%
Microscopy		Bilirubin crystal	1	0.12%
Microscopy		Calcium carbonate	11	1.34%
Microscopy		Calcium oxalate crystal	2	0.24%
Microscopy		Cellular (WBC) cast	1	0.12%
Microscopy		Fat droplets or globules	4	0.49%
Microscopy		Fiber (exogenous)/fecal contamination	12	1.47%
Microscopy		Granular cast	1	0.12%
Microscopy		Hippuric acid	1	0.12%
Microscopy		Mucous strands	3	0.37%
Microscopy		Pollen grain	2	0.24%
Microscopy		RBC/Blood/Hgb cast	1	0.12%
Microscopy		Red blood cell (RBC) (Erythrocyte)	12	1.47%
Microscopy		Red Blood Cell (RBC) (Erythrocyte)	1	0.12%
Microscopy		Renal tubular epithelial (RTE)	6	0.73%
Microscopy		RTE with fat globules (oval fat bodies)	1	0.12%
Microscopy		Squamous epithelial cell	5	0.61%
Microscopy		Starch granules	5	0.61%
Microscopy		Transitional epithelial cell (urothelial)	8	0.98%
Microscopy		Uric acid crystal	2	0.24%
Microscopy		White blood cell (WBC) (Leukocyte)	17	2.08%
Microscopy		Yeast/fungi	3	0.37%
CM-12		* Denotes Correct Response		
AG: All Methods	818			
		*Starch granule	1	0.12%
		*Starch granules	558	68.22%
		Ammonium biurate crystal	2	0.24%
		Ammonium magnesium phosphate (Triple phosphate)	1	0.12%
		Amorphous urates/phosphates	2	0.24%
		Calcium carbonate	3	0.37%
		Calcium oxalate crystal	133	16.26%
		Cellular (RTE) cast	1	0.12%
		Cholesterol crystal	8	0.98%
		Cystine crystal	15	1.83%
		Fat droplets or globules	15	1.83%
		Granular cast	2	0.24%
		Hippuric acid	3	0.37%
		Hyaline cast	2	0.24%
		Ident Unknown - Referred	5	0.61%
		Leucine crystal	10	1.22%
		Mucous strands	1	0.12%
		Red blood cell (RBC) (Erythrocyte)	2	0.24%
		Red Blood Cell (RBC) (Erythrocyte)	1	0.12%
		Renal tubular epithelial (RTE)	2	0.24%
		RTE with fat globules (oval fat bodies)	13	1.59%
		Squamous epithelial cell	14	1.71%
		Squamous epithelial cell	1	0.12%
		Transitional epithelial cell (urothelial)	1	0.12%
		Uric acid crystal	20	2.44%
		White blood cell (WBC) (Leukocyte)	2	0.24%
Microscopy	818	*Starch granule	1	0.12%
Microscopy		*Starch granules	558	68.22%
Microscopy		Ammonium biurate crystal	2	0.24%
Microscopy		Ammonium magnesium phosphate (Triple phosphate)	1	0.12%
Microscopy		Amorphous urates/phosphates	2	0.24%
Microscopy		Calcium carbonate	3	0.37%
Microscopy		Calcium oxalate crystal	133	16.26%
Microscopy		Cellular (RTE) cast	1	0.12%
Microscopy		Cholesterol crystal	8	0.98%
Microscopy		Cystine crystal	15	1.83%
Microscopy		Fat droplets or globules	15	1.83%
Microscopy		Granular cast	2	0.24%
Microscopy		Hippuric acid	3	0.37%
Microscopy		Hyaline cast	2	0.24%
Microscopy		Ident Unknown - Referred	5	0.61%
Microscopy		Leucine crystal	10	1.22%
Microscopy		Mucous strands	1	0.12%
Microscopy		Red blood cell (RBC) (Erythrocyte)	2	0.24%
Microscopy		Red Blood Cell (RBC) (Erythrocyte)	1	0.12%



Microscopy		Renal tubular epithelial (RTE)	2	0.24%
Microscopy		RTE with fat globules (oval fat bodies)	13	1.59%
Microscopy		Squamous epithelial cell	14	1.71%
Microscopy		Squamous epithelial cell	1	0.12%
Microscopy		Transitional epithelial cell (urothelial)	1	0.12%
Microscopy		Uric acid crystal	20	2.44%
Microscopy		White blood cell (WBC) (Leukocyte)	2	0.24%
CM-13		* Denotes Correct Response		
AG: All Methods	818			
		*Squamous epithelial cell	770	94.13%
		Calcium oxalate crystal	4	0.49%
		Cholesterol crystal	1	0.12%
		Clue cell	3	0.37%
		Ident Unknown - Referred	1	0.12%
		Pollen grain	1	0.12%
		Renal tubular epithelial (RTE)	4	0.49%
		RTE with fat globules (oval fat bodies)	1	0.12%
		Squamous epithelial cell	1	0.12%
		Starch granules	11	1.34%
		Transitional epithelial cell (urothelial)	16	1.96%
		Trichomonas vaginalis	1	0.12%
		Uric acid crystal	1	0.12%
		White blood cell (WBC) (Leukocyte)	3	0.37%
Microscopy	818	*Squamous epithelial cell	770	94.13%
Microscopy		Calcium oxalate crystal	4	0.49%
Microscopy		Cholesterol crystal	1	0.12%
Microscopy		Clue cell	3	0.37%
Microscopy		Ident Unknown - Referred	1	0.12%
Microscopy		Pollen grain	1	0.12%
Microscopy		Renal tubular epithelial (RTE)	4	0.49%
Microscopy		RTE with fat globules (oval fat bodies)	1	0.12%
Microscopy		Squamous epithelial cell	1	0.12%
Microscopy		Starch granules	11	1.34%
Microscopy		Transitional epithelial cell (urothelial)	16	1.96%
Microscopy		Trichomonas vaginalis	1	0.12%
Microscopy		Uric acid crystal	1	0.12%
Microscopy		White blood cell (WBC) (Leukocyte)	3	0.37%
		Vaginal Wet Prep		
CM-14		* Denotes Correct Response		
AG: All Methods	1055			
		*Trichomonas vaginalis	1021	96.78%
		Bacteria present	1	0.09%
		Clue cell	2	0.19%
		Spermatozoa present	23	2.18%
		Squamous epithelial cell	1	0.09%
		Starch granule	1	0.09%
		Transitional epithelial cell (urothelial)	1	0.09%
		White Blood Cell (WBC) (Leukocyte)	5	0.47%
Microscopy	1055	*Trichomonas vaginalis	1021	96.78%
Microscopy		Bacteria present	1	0.09%
Microscopy		Clue cell	2	0.19%
Microscopy		Spermatozoa present	23	2.18%
Microscopy		Squamous epithelial cell	1	0.09%
Microscopy		Starch granule	1	0.09%
Microscopy		Transitional epithelial cell (urothelial)	1	0.09%
Microscopy		White Blood Cell (WBC) (Leukocyte)	5	0.47%
CM-15		* Denotes Correct Response		
AG: All Methods	1053			
		*Red Blood Cell (RBC) (Erythrocyte)	293	27.83%
		Bacteria present	5	0.47%
		Clue cell	19	1.80%
		Leucine crystal	1	0.09%
		Pollen grain	2	0.19%
		Red blood cell (RBC) (Erythrocyte)	1	0.09%
		Spermatozoa absent	2	0.19%
		Squamous epithelial cell	1	0.09%
		Squamous epithelial cell	8	0.76%
		Starch granule	1	0.09%
		Trichomonas vaginalis	8	0.76%



		White Blood Cell (WBC) (Leukocyte)	695	66.00%
		Yeast/fungi present	17	1.61%
Microscopy	1053	*Red Blood Cell (RBC) (Erythrocyte)	293	27.83%
Microscopy		Bacteria present	5	0.47%
Microscopy		Clue cell	19	1.80%
Microscopy		Leucine crystal	1	0.09%
Microscopy		Pollen grain	2	0.19%
Microscopy		Red blood cell (RBC) (Erythrocyte)	1	0.09%
Microscopy		Spermatozoa absent	2	0.19%
Microscopy		Squamous epithelial cell	1	0.09%
Microscopy		Squamous epithelial cell	8	0.76%
Microscopy		Starch granule	1	0.09%
Microscopy		Trichomonas vaginalis	8	0.76%
Microscopy		White Blood Cell (WBC) (Leukocyte)	695	66.00%
Microscopy		Yeast/fungi present	17	1.61%
		Skin Scraping		
CM-16		* Denotes Correct Response		
AG: All Methods	634			
		*No scabies present	6	0.95%
		*Yeast/fungi present	622	98.11%
		Positive Fern Test	1	0.16%
		Yeast/fungi absent	5	0.79%
Microscopy	634	*No scabies present	6	0.95%
Microscopy		*Yeast/fungi present	622	98.11%
Microscopy		Positive Fern Test	1	0.16%
Microscopy		Yeast/fungi absent	5	0.79%
		Fecal Leukocytes		
CM-18		* Denotes Correct Response		
AG: All Methods	256			
		*No neutrophils are present	214	83.59%
		Eosinophils are present	1	0.39%
		Neutrophils are present	41	16.02%
Microscopy	256	*No neutrophils are present	214	83.59%
Microscopy		Eosinophils are present	1	0.39%
Microscopy		Neutrophils are present	41	16.02%
		Nasal Smear		
CM-17		* Denotes Correct Response		
AG: All Methods	251			
		*Eosinophils are present	223	88.84%
		No eosinophils are present	28	11.16%
Microscopy	251	*Eosinophils are present	223	88.84%
Microscopy		No eosinophils are present	28	11.16%
683 - ASO/CRP;684 - CRP (Quant & Qual)		Labs	Response	Number
			CRP Qualitative	Percent
774 - GenProbe GC/Chlamydia;794 - GC & Chlamydia Antigen Detection (5 specimens);795 - GC & Chlamydia Antigen Detection (2 specimens)		Labs	Response	Number
			Chlamydia - Non Culture Tech	Percent
CY-6		* Denotes Correct Response		
AG: All Methods	8			
		*Positive	8	100.00%
BD Probe Tec ET	1	*Positive	1	100.00%
Gen-Probe	1	*Positive	1	100.00%
Quidel QuickVue Chlamydia	3	*Positive	3	100.00%
Cepheid GeneXpert	2	*Positive	2	100.00%
Hologic Aptima Combo 2	1	*Positive	1	100.00%
CY-7		* Denotes Correct Response		
AG: All Methods	8			
		*Negative	6	75.00%
		Positive	2	25.00%
BD Probe Tec ET	1	*Negative	1	100.00%
Quidel QuickVue Chlamydia	3	*Negative	3	100.00%
Cepheid GeneXpert	2	*Negative	2	100.00%
Gen-Probe	1	Positive	1	100.00%
Hologic Aptima Combo 2	1	Positive	1	100.00%
CY-8		* Denotes Correct Response		
AG: All Methods	8			
		*Positive	8	100.00%



BD Probe Tec ET	1	*Positive	1	100.00%
Gen-Probe	1	*Positive	1	100.00%
Quidel QuickVue Chlamydia	3	*Positive	3	100.00%
Cepheid GeneXpert	2	*Positive	2	100.00%
Hologic Aptima Combo 2	1	*Positive	1	100.00%
CY-9		* Denotes Correct Response		
AG: All Methods	7			
		*Negative	7	100.00%
BD Probe Tec ET	1	*Negative	1	100.00%
Gen-Probe	1	*Negative	1	100.00%
Quidel QuickVue Chlamydia	3	*Negative	3	100.00%
Cepheid GeneXpert	1	*Negative	1	100.00%
Hologic Aptima Combo 2	1	*Negative	1	100.00%
CY-10		* Denotes Correct Response		
AG: All Methods	8			
		*Negative	6	75.00%
		Positive	2	25.00%
BD Probe Tec ET	1	*Negative	1	100.00%
Quidel QuickVue Chlamydia	3	*Negative	3	100.00%
Cepheid GeneXpert	2	*Negative	2	100.00%
Gen-Probe	1	Positive	1	100.00%
Hologic Aptima Combo 2	1	Positive	1	100.00%
		GC - Non Culture Tech		
CY-6		* Denotes Correct Response		
AG: All Methods	5			
		*Negative	5	100.00%
BD Probe Tec ET	1	*Negative	1	100.00%
Gen-Probe	1	*Negative	1	100.00%
Cepheid GeneXpert	2	*Negative	2	100.00%
Hologic Aptima Combo 2	1	*Negative	1	100.00%
CY-7		* Denotes Correct Response		
AG: All Methods	5			
		*Positive	5	100.00%
BD Probe Tec ET	1	*Positive	1	100.00%
Gen-Probe	1	*Positive	1	100.00%
Cepheid GeneXpert	2	*Positive	2	100.00%
Hologic Aptima Combo 2	1	*Positive	1	100.00%
CY-8		* Denotes Correct Response		
AG: All Methods	5			
		*Negative	5	100.00%
BD Probe Tec ET	1	*Negative	1	100.00%
Gen-Probe	1	*Negative	1	100.00%
Cepheid GeneXpert	2	*Negative	2	100.00%
Hologic Aptima Combo 2	1	*Negative	1	100.00%
CY-9		* Denotes Correct Response		
AG: All Methods	4			
		*Negative	4	100.00%
BD Probe Tec ET	1	*Negative	1	100.00%
Gen-Probe	1	*Negative	1	100.00%
Cepheid GeneXpert	1	*Negative	1	100.00%
Hologic Aptima Combo 2	1	*Negative	1	100.00%
CY-10		* Denotes Correct Response		
AG: All Methods	5			
		*Positive	5	100.00%
BD Probe Tec ET	1	*Positive	1	100.00%
Gen-Probe	1	*Positive	1	100.00%
Cepheid GeneXpert	2	*Positive	2	100.00%
Hologic Aptima Combo 2	1	*Positive	1	100.00%
772 - Dermatophyte Screen/Culture	Labs	Response	Number	Percent
		Dermatophyte Culture		
DM-6		* Denotes Correct Response		
AG: All Methods	2			
		*Positive, Dermatophytes Present	2	100.00%
Dermatophytes	2	*Positive, Dermatophytes Present	2	100.00%
DM-7		* Denotes Correct Response		
AG: All Methods	2			
		*Negative, No Dermatophytes Present	2	100.00%
Dermatophytes	2	*Negative, No Dermatophytes Present	2	100.00%



DM-8		* Denotes Correct Response		
AG: All Methods	2			
		*Positive, Dermatophytes Present	1	50.00%
		Negative, No Dermatophytes Present	1	50.00%
Dermatophytes	2			
		*Positive, Dermatophytes Present	1	50.00%
Dermatophytes		Negative, No Dermatophytes Present	1	50.00%
DM-9		* Denotes Correct Response		
AG: All Methods	2			
		*Positive, Dermatophytes Present	2	100.00%
Dermatophytes	2			
		*Positive, Dermatophytes Present	2	100.00%
DM-10		* Denotes Correct Response		
AG: All Methods	3			
		*Microsporum species	1	33.33%
		*Positive, Dermatophytes Present	1	33.33%
		Negative, No Dermatophytes Present	1	33.33%
Dermatophytes	3			
		*Microsporum species	1	33.33%
Dermatophytes		*Positive, Dermatophytes Present	1	33.33%
Dermatophytes		Negative, No Dermatophytes Present	1	33.33%
Dermatophyte Screen				
DM-6		* Denotes Correct Response		
AG: All Methods	9			
		*Positive, Dermatophytes Present	6	66.67%
		Negative, No Dermatophytes Present	3	33.33%
Dermatophytes	9			
		*Positive, Dermatophytes Present	6	66.67%
Dermatophytes		Negative, No Dermatophytes Present	3	33.33%
DM-7		* Denotes Correct Response		
AG: All Methods	9			
		*Negative, No Dermatophytes Present	8	88.89%
		Positive, Dermatophytes Present	1	11.11%
Dermatophytes	9			
		*Negative, No Dermatophytes Present	8	88.89%
Dermatophytes		Positive, Dermatophytes Present	1	11.11%
DM-8		* Denotes Correct Response		
AG: All Methods	9			
		*Positive, Dermatophytes Present	3	33.33%
		Negative, No Dermatophytes Present	6	66.67%
Dermatophytes	9			
		*Positive, Dermatophytes Present	3	33.33%
Dermatophytes		Negative, No Dermatophytes Present	6	66.67%
DM-9		* Denotes Correct Response		
AG: All Methods	9			
		*Positive, Dermatophytes Present	6	66.67%
		Negative, No Dermatophytes Present	3	33.33%
Dermatophytes	9			
		*Positive, Dermatophytes Present	6	66.67%
Dermatophytes		Negative, No Dermatophytes Present	3	33.33%
DM-10		* Denotes Correct Response		
AG: All Methods	9			
		*Positive, Dermatophytes Present	9	100.00%
Dermatophytes	9			
		*Positive, Dermatophytes Present	9	100.00%
777 - Genital Culture:796 - Genital Culture (2 specimens)				
Labs	Response	Number	Percent	
Genital Culture				
GC-6		* Denotes Correct Response		
AG: All Methods	3			
		*Presumptive positive for GC, GR(-) diplococci, oxidase positive organism	2	66.67%
		*Presumptive positive for GC, growth referred for ID/confirmation	1	33.33%
Cultures	3			
		*Presumptive positive for GC, GR(-) diplococci, oxidase positive organism	2	66.67%
Cultures		*Presumptive positive for GC, growth referred for ID/confirmation	1	33.33%
GC-7		* Denotes Correct Response		
AG: All Methods	3			
		*Negative for GC	2	66.67%
		*No Gram Negative diplococci present	1	33.33%
Cultures	3			
		*Negative for GC	2	66.67%
Cultures		*No Gram Negative diplococci present	1	33.33%
GC-8		* Denotes Correct Response		
AG: All Methods	3			
		*Negative for GC	2	66.67%
		*No Gram Negative diplococci present	1	33.33%



Cultures	3	*Negative for GC	2	66.67%
Cultures		*No Gram Negative diplococci present	1	33.33%
GC-9		* Denotes Correct Response		
AG: All Methods	6			
		*Group B Streptococcus present	1	16.67%
		*Negative for GC	1	16.67%
		*No growth, GC selective media	3	50.00%
		*Streptococcus, Group B	1	16.67%
Cultures	6	*Group B Streptococcus present	1	16.67%
Cultures		*Negative for GC	1	16.67%
Cultures		*No growth, GC selective media	3	50.00%
Cultures		*Streptococcus, Group B	1	16.67%
GC-10		* Denotes Correct Response		
AG: All Methods	4			
		*Negative for GC	1	25.00%
		*No growth, GC selective media	3	75.00%
Cultures	4	*Negative for GC	1	25.00%
Cultures		*No growth, GC selective media	3	75.00%
668 - Gram Stain	Labs	Response	Number	Percent
		Gram Stain		
GS-6		* Denotes Correct Response		
AG: All Methods	11			
		*Gram Negative	11	100.00%
Gram Stain	10	*Gram Negative	10	100.00%
Intl Gram Stain	1	*Gram Negative	1	100.00%
GS-7		* Denotes Correct Response		
AG: All Methods	11			
		*Gram Positive	10	90.91%
		Gram Negative	1	9.09%
Gram Stain	10	*Gram Positive	9	90.00%
Intl Gram Stain	1	*Gram Positive	1	100.00%
Gram Stain	10	Gram Negative	1	10.00%
GS-8		* Denotes Correct Response		
AG: All Methods	11			
		*Gram Negative	10	90.91%
		Gram Positive	1	9.09%
Gram Stain	10	*Gram Negative	9	90.00%
Intl Gram Stain	1	*Gram Negative	1	100.00%
Gram Stain	10	Gram Positive	1	10.00%
GS-9		* Denotes Correct Response		
AG: All Methods	11			
		*Gram Negative	11	100.00%
Gram Stain	10	*Gram Negative	10	100.00%
Intl Gram Stain	1	*Gram Negative	1	100.00%
GS-10		* Denotes Correct Response		
AG: All Methods	10			
		*Gram Positive	8	80.00%
		Gram Negative	2	20.00%
Gram Stain	9	*Gram Positive	7	77.78%
Intl Gram Stain	1	*Gram Positive	1	100.00%
Gram Stain	9	Gram Negative	2	22.22%
		Gram Stain Morph		
GS-6		* Denotes Correct Response		
AG: All Methods	9			
		*Rods/Bacilli	9	100.00%
Gram Stain	8	*Rods/Bacilli	8	100.00%
Intl Gram Stain	1	*Rods/Bacilli	1	100.00%
GS-7		* Denotes Correct Response		
AG: All Methods	9			
		*Cocci	3	33.33%
		*Cocci in Chains	5	55.56%
		*Cocci in Pairs	1	11.11%
Gram Stain	8	*Cocci	2	25.00%
Intl Gram Stain	1	*Cocci	1	100.00%
Gram Stain	8	*Cocci in Chains	5	62.50%
Gram Stain		*Cocci in Pairs	1	12.50%
GS-8		* Denotes Correct Response		
AG: All Methods	9			



		*Rods/Bacilli	9	100.00%
Gram Stain	8	*Rods/Bacilli	8	100.00%
Intl Gram Stain	1	*Rods/Bacilli	1	100.00%
GS-9		* Denotes Correct Response		
AG: All Methods	9			
		*Rods/Bacilli	9	100.00%
Gram Stain	8	*Rods/Bacilli	8	100.00%
Intl Gram Stain	1	*Rods/Bacilli	1	100.00%
GS-10		* Denotes Correct Response		
AG: All Methods	9			
		*Rods/Bacilli	9	100.00%
Gram Stain	8	*Rods/Bacilli	8	100.00%
Intl Gram Stain	1	*Rods/Bacilli	1	100.00%
678 - Serum hCG;759 - Complete Immunology;729 - AFC Immunology	Labs	Response	Number	Percent
		HCG - Qual Serum		
HCG-6		* Denotes Correct Response		
AG: All Methods	100			
		*Positive	100	100.00%
Beck Coulter ICON 25	11	*Positive	11	100.00%
Fisher HC Sure-Vue hCG STAT	2	*Positive	2	100.00%
Sekisui Diagnostics OSOM	2	*Positive	2	100.00%
Henry Schein One Step + (hCG)	5	*Positive	5	100.00%
Immunostics Immuno/hCG	2	*Positive	2	100.00%
Inverness Signify HCG	2	*Positive	2	100.00%
LifeSign Status hCG	2	*Positive	2	100.00%
McKesson Medi-Lab hCG - Urine	1	*Positive	1	100.00%
Quidel QuickVue + hCG	1	*Positive	1	100.00%
Quidel QuickVue One-Step hCG Combo (U/S)	36	*Positive	36	100.00%
Select Diag hCG Combo	4	*Positive	4	100.00%
SP Brand (Cardinal) HCG Combo	12	*Positive	12	100.00%
Polymedco Poly stat hCG Serum	1	*Positive	1	100.00%
Stanbio True 20 Plus	1	*Positive	1	100.00%
Consult Diag hCG Combo Cassette	14	*Positive	14	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	1	*Positive	1	100.00%
Medline Pregnancy Combo (serum)	2	*Positive	2	100.00%
Alere hCG Combo Cassette (20/10 mIU/mL)	1	*Positive	1	100.00%
HCG-7		* Denotes Correct Response		
AG: All Methods	100			
		*Positive	100	100.00%
Beck Coulter ICON 25	11	*Positive	11	100.00%
Fisher HC Sure-Vue hCG STAT	2	*Positive	2	100.00%
Sekisui Diagnostics OSOM	2	*Positive	2	100.00%
Henry Schein One Step + (hCG)	5	*Positive	5	100.00%
Immunostics Immuno/hCG	2	*Positive	2	100.00%
Inverness Signify HCG	2	*Positive	2	100.00%
LifeSign Status hCG	2	*Positive	2	100.00%
McKesson Medi-Lab hCG - Urine	1	*Positive	1	100.00%
Quidel QuickVue + hCG	1	*Positive	1	100.00%
Quidel QuickVue One-Step hCG Combo (U/S)	36	*Positive	36	100.00%
Select Diag hCG Combo	4	*Positive	4	100.00%
SP Brand (Cardinal) HCG Combo	12	*Positive	12	100.00%
Polymedco Poly stat hCG Serum	1	*Positive	1	100.00%
Stanbio True 20 Plus	1	*Positive	1	100.00%
Consult Diag hCG Combo Cassette	14	*Positive	14	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	1	*Positive	1	100.00%
Medline Pregnancy Combo (serum)	2	*Positive	2	100.00%
Alere hCG Combo Cassette (20/10 mIU/mL)	1	*Positive	1	100.00%
HCG-8		* Denotes Correct Response		
AG: All Methods	100			
		*Negative	100	100.00%
Beck Coulter ICON 25	11	*Negative	11	100.00%
Fisher HC Sure-Vue hCG STAT	2	*Negative	2	100.00%
Sekisui Diagnostics OSOM	2	*Negative	2	100.00%
Henry Schein One Step + (hCG)	5	*Negative	5	100.00%
Immunostics Immuno/hCG	2	*Negative	2	100.00%
Inverness Signify HCG	2	*Negative	2	100.00%
LifeSign Status hCG	2	*Negative	2	100.00%
McKesson Medi-Lab hCG - Urine	1	*Negative	1	100.00%



Quidel QuickVue + hCG	1	*Negative	1	100.00%
Quidel QuickVue One-Step hCG Combo (U/S)	36	*Negative	36	100.00%
Select Diag hCG Combo	4	*Negative	4	100.00%
SP Brand (Cardinal) HCG Combo	12	*Negative	12	100.00%
Polymedco Poly stat hCG Serum	1	*Negative	1	100.00%
Stanbio True 20 Plus	1	*Negative	1	100.00%
Consult Diag hCG Combo Cassette	14	*Negative	14	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	1	*Negative	1	100.00%
Medline Pregnancy Combo (serum)	2	*Negative	2	100.00%
Alere hCG Combo Cassette (20/10 mIU/mL)	1	*Negative	1	100.00%
HCG-9		* Denotes Correct Response		
AG: All Methods	100			
		*Positive	100	100.00%
Beck Coulter ICON 25	11	*Positive	11	100.00%
Fisher HC Sure-Vue hCG STAT	2	*Positive	2	100.00%
Sekisui Diagnostics OSOM	2	*Positive	2	100.00%
Henry Schein One Step + (hCG)	5	*Positive	5	100.00%
Immunostics Immuno/hCG	2	*Positive	2	100.00%
Inverness Signify HCG	2	*Positive	2	100.00%
LifeSign Status hCG	2	*Positive	2	100.00%
McKesson Medi-Lab hCG - Urine	1	*Positive	1	100.00%
Quidel QuickVue + hCG	1	*Positive	1	100.00%
Quidel QuickVue One-Step hCG Combo (U/S)	36	*Positive	36	100.00%
Select Diag hCG Combo	4	*Positive	4	100.00%
SP Brand (Cardinal) HCG Combo	12	*Positive	12	100.00%
Polymedco Poly stat hCG Serum	1	*Positive	1	100.00%
Stanbio True 20 Plus	1	*Positive	1	100.00%
Consult Diag hCG Combo Cassette	14	*Positive	14	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	1	*Positive	1	100.00%
Medline Pregnancy Combo (serum)	2	*Positive	2	100.00%
Alere hCG Combo Cassette (20/10 mIU/mL)	1	*Positive	1	100.00%
HCG-10		* Denotes Correct Response		
AG: All Methods	100			
		*Positive	99	99.00%
		Negative	1	1.00%
Beck Coulter ICON 25	11	*Positive	11	100.00%
Fisher HC Sure-Vue hCG STAT	2	*Positive	2	100.00%
Sekisui Diagnostics OSOM	2	*Positive	2	100.00%
Henry Schein One Step + (hCG)	5	*Positive	5	100.00%
Immunostics Immuno/hCG	2	*Positive	2	100.00%
Inverness Signify HCG	2	*Positive	2	100.00%
LifeSign Status hCG	2	*Positive	2	100.00%
McKesson Medi-Lab hCG - Urine	1	*Positive	1	100.00%
Quidel QuickVue + hCG	1	*Positive	1	100.00%
Quidel QuickVue One-Step hCG Combo (U/S)	36	*Positive	36	100.00%
Select Diag hCG Combo	4	*Positive	3	75.00%
SP Brand (Cardinal) HCG Combo	12	*Positive	12	100.00%
Polymedco Poly stat hCG Serum	1	*Positive	1	100.00%
Stanbio True 20 Plus	1	*Positive	1	100.00%
Consult Diag hCG Combo Cassette	14	*Positive	14	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	1	*Positive	1	100.00%
Medline Pregnancy Combo (serum)	2	*Positive	2	100.00%
Alere hCG Combo Cassette (20/10 mIU/mL)	1	*Positive	1	100.00%
Select Diag hCG Combo	4	Negative	1	25.00%
HCG-6		HCG - Qual Serum (International)		
		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Henry Schein One Step + (hCG)	1	*Positive	1	100.00%
SAS Ultra HCG	2	*Positive	2	100.00%
HCG-7		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Henry Schein One Step + (hCG)	1	*Positive	1	100.00%
SAS Ultra HCG	2	*Positive	2	100.00%
HCG-8		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Henry Schein One Step + (hCG)	1	*Negative	1	100.00%



SAS Ultra HCG	2	*Negative	2	100.00%
HCG-9		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Henry Schein One Step + (hCG)	1	*Positive	1	100.00%
SAS Ultra HCG	2	*Positive	2	100.00%
HCG-10		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Henry Schein One Step + (hCG)	1	*Positive	1	100.00%
SAS Ultra HCG	2	*Positive	2	100.00%
762 - HIV Antibody (Waived);766 - Anti-HIV (5 specimens);750 - Viral Markers	Labs	Response	Number	Percent
		HIV Antibody (Waived)		
HIV-6		* Denotes Correct Response		
AG: All Methods	78			
		*Negative	64	82.05%
		Positive	14	17.95%
Alere Clearview Complete HIV 1/2 (mod)	11	*Negative	9	81.82%
Alere Clearview HIV 1/2 STAT PAK (WB)	39	*Negative	36	92.31%
Orasure OraQuick Anti-HIV	12	*Negative	12	100.00%
Trinity Biotech Uni-Gold Recombigen HIV	2	*Negative	2	100.00%
Alere Determine HIV Ag/Ab	10	*Negative	1	10.00%
Chembio HIV 1/2 STAT-PAK Cassette	4	*Negative	4	100.00%
Alere Clearview Complete HIV 1/2 (mod)	11	Positive	2	18.18%
Alere Clearview HIV 1/2 STAT PAK (WB)	39	Positive	3	7.69%
Alere Determine HIV Ag/Ab	10	Positive	9	90.00%
		HIV Antibody (International)		
HIV-6		* Denotes Correct Response		
AG: All Methods	4			
		*Negative	4	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
Intl Chembio HIV 1/2 STAT-PAK Dipstick	1	*Negative	1	100.00%
		HIV Antibody		
HIV-6		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	1	33.33%
		Positive	2	66.67%
Alere Clearview Complete HIV 1/2 (mod)	2	*Negative	1	50.00%
Alere Clearview Complete HIV 1/2 (mod)		Positive	1	50.00%
Abbott Architect i1000 SR	1	Positive	1	100.00%
766 - Anti-HIV (5 specimens);750 - Viral Markers	Labs	Response	Number	Percent
		HIV Antibody		
HIV-6		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	1	33.33%
		Positive	2	66.67%
Alere Clearview Complete HIV 1/2 (mod)	2	*Negative	1	50.00%
Alere Clearview Complete HIV 1/2 (mod)		Positive	1	50.00%
Abbott Architect i1000 SR	1	Positive	1	100.00%
HIV-7		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Alere Clearview Complete HIV 1/2 (mod)	1	*Positive	1	100.00%
Abbott Architect i1000 SR	1	*Positive	1	100.00%
HIV-8		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Alere Clearview Complete HIV 1/2 (mod)	1	*Positive	1	100.00%
Abbott Architect i1000 SR	1	*Positive	1	100.00%
HIV-9		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Alere Clearview Complete HIV 1/2 (mod)	1	*Negative	1	100.00%
Abbott Architect i1000 SR	1	*Negative	1	100.00%
HIV-10		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%



Alere Clearview Complete HIV 1/2 (mod)	1	*Negative	1	100.00%
Abbott Architect i1000 SR	1	*Negative	1	100.00%
HIV Antibody (Waived)				
HIV-6				
* Denotes Correct Response				
AG: All Methods	78			
		*Negative	64	82.05%
		Positive	14	17.95%
Alere Clearview Complete HIV 1/2 (mod)	11	*Negative	9	81.82%
Alere Clearview HIV 1/2 STAT PAK (WB)	39	*Negative	36	92.31%
Orasure OraQuick Anti-HIV	12	*Negative	12	100.00%
Trinity Biotech Uni-Gold Recombigen HIV	2	*Negative	2	100.00%
Alere Determine HIV Ag/Ab	10	*Negative	1	10.00%
Chembio HIV 1/2 STAT-PAK Cassette	4	*Negative	4	100.00%
Alere Clearview Complete HIV 1/2 (mod)	11	Positive	2	18.18%
Alere Clearview HIV 1/2 STAT PAK (WB)	39	Positive	3	7.69%
Alere Determine HIV Ag/Ab	10	Positive	9	90.00%
HIV-7				
* Denotes Correct Response				
AG: All Methods	6			
		*Positive	6	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	4	*Positive	4	100.00%
Orasure OraQuick Anti-HIV	1	*Positive	1	100.00%
Trinity Biotech Uni-Gold Recombigen HIV	1	*Positive	1	100.00%
HIV-8				
* Denotes Correct Response				
AG: All Methods	6			
		*Positive	6	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	4	*Positive	4	100.00%
Orasure OraQuick Anti-HIV	1	*Positive	1	100.00%
Trinity Biotech Uni-Gold Recombigen HIV	1	*Positive	1	100.00%
HIV-9				
* Denotes Correct Response				
AG: All Methods	6			
		*Negative	6	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	4	*Negative	4	100.00%
Orasure OraQuick Anti-HIV	1	*Negative	1	100.00%
Trinity Biotech Uni-Gold Recombigen HIV	1	*Negative	1	100.00%
HIV-10				
* Denotes Correct Response				
AG: All Methods	6			
		*Negative	6	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	4	*Negative	4	100.00%
Orasure OraQuick Anti-HIV	1	*Negative	1	100.00%
Trinity Biotech Uni-Gold Recombigen HIV	1	*Negative	1	100.00%
HIV Antibody (International)				
HIV-6				
* Denotes Correct Response				
AG: All Methods	4			
		*Negative	4	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
Intl Chembio HIV 1/2 STAT-PAK Dipstick	1	*Negative	1	100.00%
HIV-7				
* Denotes Correct Response				
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Positive	3	100.00%
HIV-8				
* Denotes Correct Response				
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Positive	3	100.00%
HIV-9				
* Denotes Correct Response				
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
HIV-10				
* Denotes Correct Response				
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
644 - H. pylori:628 - AFC Waived/PPM	Labs	Response	Number	Percent
H. pylori Ab Detect				
HP-2				
* Denotes Correct Response				
AG: All Methods	96			
		*Positive	89	92.71%
		Negative	7	7.29%



Test Name	Labs	Response	Number	Percent
Beck Coulter FlexSure	3	*Positive	3	100.00%
Beck Coulter ICON HP	3	*Positive	3	100.00%
DiagTest Group Clarity H. pylori	2	*Positive	2	100.00%
Instant Technologies i-Screen	2	*Positive	2	100.00%
McKesson Medi-Lab Perf H. pylori	6	*Positive	6	100.00%
Polymedco Poly stat (WB - Waived)	3	*Positive	3	100.00%
Henry Schein One Step+ (H. pylori)	12	*Positive	11	91.67%
Select Diag H. pylori Cassette	4	*Positive	4	100.00%
Quidel QuickVue H. pylori gII (S/P)	7	*Positive	7	100.00%
Quidel QuickVue H. pylori gII (WB)	14	*Positive	13	92.86%
SP Brand (Cardinal) HP Rapid Test	10	*Positive	10	100.00%
Alere Clearview II S/P (mod)	4	*Positive	4	100.00%
Alere Clearview II WB (waived)	4	*Positive	3	75.00%
LifeSign Status H. pylori	2	*Positive	2	100.00%
Instant-View H. pylori	1	*Positive	1	100.00%
Consult Diag (PSS) H pylori	18	*Positive	14	77.78%
Jant Accutest H pylori	1	*Positive	1	100.00%
Henry Schein One Step+ (H. pylori)	12	Negative	1	8.33%
Quidel QuickVue H. pylori gII (WB)	14	Negative	1	7.14%
Alere Clearview II WB (waived)	4	Negative	1	25.00%
Consult Diag (PSS) H pylori	18	Negative	4	22.22%
933 - HIV Antibody/Antigen Combo (2 specimens);934 - HIV Antibody/Antigen Combo (5 specimens)	Labs	Response	Number	Percent
HVAA-6		HIV Antibody (Waived)		
		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	1	50.00%
		Positive	1	50.00%
Alere Determine HIV Ag/Ab	2	*Negative	1	50.00%
Alere Determine HIV Ag/Ab		Positive	1	50.00%
HVAA-7		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Alere Determine HIV Ag/Ab	2	*Positive	2	100.00%
HVAA-6		HIV Antigen (p24) (Waived)		
		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	1	50.00%
		Negative	1	50.00%
Alere Determine HIV Ag/Ab	2	*Positive	1	50.00%
Alere Determine HIV Ag/Ab		Negative	1	50.00%
HVAA-7		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Alere Determine HIV Ag/Ab	2	*Negative	2	100.00%
603 - Blood Cell Identification;610 - Regular Hematology;749 - Auto Diff II w-Blood Cell Identification	Labs	Response	Number	Percent
ID-7		Blood Cell Identification		
		* Denotes Correct Response		
AG: All Methods	75			
		*Eosinophil, any stage	75	100.00%
Blood Cell Microscopy	75	*Eosinophil, any stage	75	100.00%
ID-8		* Denotes Correct Response		
AG: All Methods	75			
		*Monocyte	71	94.67%
		Lymphocyte	1	1.33%
		Lymphocyte, reactive (atypical, variant, Downey)	2	2.67%
		Promonocyte	1	1.33%
Blood Cell Microscopy	75	*Monocyte	71	94.67%
Blood Cell Microscopy		Lymphocyte	1	1.33%
Blood Cell Microscopy		Lymphocyte, reactive (atypical, variant, Downey)	2	2.67%
Blood Cell Microscopy		Promonocyte	1	1.33%
ID-9		* Denotes Correct Response		
AG: All Methods	75			
		*Lymphocyte	75	100.00%
Blood Cell Microscopy	75	*Lymphocyte	75	100.00%
ID-10		* Denotes Correct Response		
AG: All Methods	75			



		*Platelet, normal	74	98.67%
		Other, not listed	1	1.33%
Blood Cell Microscopy	75	*Platelet, normal	74	98.67%
Blood Cell Microscopy		Other, not listed	1	1.33%
ID-11		* Denotes Correct Response		
AG: All Methods	75			
		*Segmented or band neutrophil	69	92.00%
		Hypersegmented neutrophil	4	5.33%
		Seg/band neutrophil with vacuoles	2	2.67%
Blood Cell Microscopy	75	*Segmented or band neutrophil	69	92.00%
Blood Cell Microscopy		Hypersegmented neutrophil	4	5.33%
Blood Cell Microscopy		Seg/band neutrophil with vacuoles	2	2.67%
ID-12		* Denotes Correct Response		
AG: All Methods	75			
		*Red Blood Cell, normal (RBC) (Erythrocyte)	74	98.67%
		Microcyte	1	1.33%
Blood Cell Microscopy	75	*Red Blood Cell, normal (RBC) (Erythrocyte)	74	98.67%
Blood Cell Microscopy		Microcyte	1	1.33%
		Blood Cell Identification (International)		
ID-7		* Denotes Correct Response		
AG: All Methods	1			
		*Eosinophil, any stage	1	100.00%
Intl Blood Cell Microscopy	1	*Eosinophil, any stage	1	100.00%
ID-8		* Denotes Correct Response		
AG: All Methods	1			
		*Monocyte	1	100.00%
Intl Blood Cell Microscopy	1	*Monocyte	1	100.00%
ID-9		* Denotes Correct Response		
AG: All Methods	1			
		*Lymphocyte	1	100.00%
Intl Blood Cell Microscopy	1	*Lymphocyte	1	100.00%
ID-10		* Denotes Correct Response		
AG: All Methods	1			
		*Platelet, normal	1	100.00%
Intl Blood Cell Microscopy	1	*Platelet, normal	1	100.00%
ID-11		* Denotes Correct Response		
AG: All Methods	1			
		*Segmented or band neutrophil	1	100.00%
Intl Blood Cell Microscopy	1	*Segmented or band neutrophil	1	100.00%
ID-12		* Denotes Correct Response		
AG: All Methods	1			
		*Red Blood Cell, normal (RBC) (Erythrocyte)	1	100.00%
Intl Blood Cell Microscopy	1	*Red Blood Cell, normal (RBC) (Erythrocyte)	1	100.00%
609 - Waived/PPM Tests 1;675 - Infectious Mononucleosis;753 - Infectious Mononucleosis (Waived) (1 specimen);754 - Infectious Mono/RF Combo;759 - Complete Immunology;628 - AFC Waived/PPM;729 - AFC Immunology	Labs	Response	Number	Percent
		Infectious Mono (Waived)		
IM-6		* Denotes Correct Response		
AG: All Methods	487			
		*Positive	474	97.33%
		Negative	13	2.67%
Beck Coulter ICON (WB)	1	*Positive	1	100.00%
Alere Acceava (S)	2	*Positive	2	100.00%
Alere Acceava (WB)	52	*Positive	51	98.08%
Sekisui Diagnostics OSOM Mono (WB)	193	*Positive	190	98.45%
Henry Schein One-Step+ Mono	34	*Positive	33	97.06%
Instant Technologies i-Screen (WB)	2	*Positive	2	100.00%
Inverness Signify Mono	4	*Positive	4	100.00%
LifeSign Status Inf Mono	34	*Positive	33	97.06%
McKesson Medi-Lab Perf - Inf Mono	20	*Positive	19	95.00%
Polymedco Poly stat Mono - (WB)	7	*Positive	7	100.00%
Quidel QuickVue+ (WB)	25	*Positive	22	88.00%
Remel A.R.C. Mono Test	1	*Positive	1	100.00%
SP Brand (Cardinal) Rapid Test	7	*Positive	7	100.00%
Alere Clearview Mono Cassette	32	*Positive	30	93.75%
Pulse Scientific	1	*Positive	1	100.00%
Diag Test Group Clarity Mono	4	*Positive	4	100.00%



Fisher HC Sure-Vue Mono	2	*Positive	2	100.00%
Immunostics Inc. ImmunoDetector Mono (WB)	1	*Positive	1	100.00%
Meridian Immunocard STAT! Mono (WB)	5	*Positive	5	100.00%
NDC ProAdvantage Mono	2	*Positive	2	100.00%
Consult Diag Mono Cassette	52	*Positive	51	98.08%
Alere Acceava (WB) Mono Cassette	5	*Positive	5	100.00%
Medline Mono Cassette (WB)	1	*Positive	1	100.00%
Alere Acceava (WB)	52	Negative	1	1.92%
Sekisui Diagnostics OSOM Mono (WB)	193	Negative	3	1.55%
Henry Schein One-Step+ Mono	34	Negative	1	2.94%
LifeSign Status Inf Mono	34	Negative	1	2.94%
McKesson Medi-Lab Perf - Inf Mono	20	Negative	1	5.00%
Quidel QuickVue+ (WB)	25	Negative	3	12.00%
Alere Clearview Mono Cassette	32	Negative	2	6.25%
Consult Diag Mono Cassette	52	Negative	1	1.92%
		Infectious Mononucleosis		
IM-6		* Denotes Correct Response		
AG: All Methods				
	78			
		*Positive	78	100.00%
Beck Coulter ICON (WB)	1	*Positive	1	100.00%
Alere Acceava (WB)	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	33	*Positive	33	100.00%
LifeSign Status Inf Mono	2	*Positive	2	100.00%
Lifesign Unistep	1	*Positive	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	3	*Positive	3	100.00%
Polymedco Poly stat Mono - (WB)	2	*Positive	2	100.00%
Polymedco Poly stat Mono (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (WB)	8	*Positive	8	100.00%
Seradyn Color Slide II	6	*Positive	6	100.00%
Alere Clearview Mono Cassette	10	*Positive	10	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	2	*Positive	2	100.00%
NDC ProAdvantage Mono	1	*Positive	1	100.00%
Consult Diag Mono Cassette	4	*Positive	4	100.00%
675 - Infectious Mononucleosis;754 - Infectious Mono/RF Combo;759 - Complete Immunology;628 - AFC Waived/PPM;729 - AFC Immunology				
	Labs	Response	Number	Percent
		Infectious Mononucleosis		
IM-6		* Denotes Correct Response		
AG: All Methods				
	78			
		*Positive	78	100.00%
Beck Coulter ICON (WB)	1	*Positive	1	100.00%
Alere Acceava (WB)	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	33	*Positive	33	100.00%
LifeSign Status Inf Mono	2	*Positive	2	100.00%
Lifesign Unistep	1	*Positive	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	3	*Positive	3	100.00%
Polymedco Poly stat Mono - (WB)	2	*Positive	2	100.00%
Polymedco Poly stat Mono (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (WB)	8	*Positive	8	100.00%
Seradyn Color Slide II	6	*Positive	6	100.00%
Alere Clearview Mono Cassette	10	*Positive	10	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	2	*Positive	2	100.00%
NDC ProAdvantage Mono	1	*Positive	1	100.00%
Consult Diag Mono Cassette	4	*Positive	4	100.00%
		* Denotes Correct Response		
IM-7		* Denotes Correct Response		
AG: All Methods				
	78			
		*Negative	78	100.00%
Beck Coulter ICON (WB)	1	*Negative	1	100.00%
Alere Acceava (WB)	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	33	*Negative	33	100.00%
LifeSign Status Inf Mono	2	*Negative	2	100.00%
Lifesign Unistep	1	*Negative	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	3	*Negative	3	100.00%
Polymedco Poly stat Mono - (WB)	2	*Negative	2	100.00%
Polymedco Poly stat Mono (S/P)	2	*Negative	2	100.00%
Quidel QuickVue+ (S/P)	2	*Negative	2	100.00%
Quidel QuickVue+ (WB)	8	*Negative	8	100.00%



Seradyn Color Slide II	6	*Negative	6	100.00%
Alere Clearview Mono Cassette	10	*Negative	10	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	2	*Negative	2	100.00%
NDC ProAdvantage Mono	1	*Negative	1	100.00%
Consult Diag Mono Cassette	4	*Negative	4	100.00%
IM-8		* Denotes Correct Response		
AG: All Methods	78			
		*Positive	78	100.00%
Beck Coulter ICON (WB)	1	*Positive	1	100.00%
Alere Acceava (WB)	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	33	*Positive	33	100.00%
LifeSign Status Inf Mono	2	*Positive	2	100.00%
Lifesign Unistep	1	*Positive	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	3	*Positive	3	100.00%
Polymedco Poly stat Mono - (WB)	2	*Positive	2	100.00%
Polymedco Poly stat Mono (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (WB)	8	*Positive	8	100.00%
Seradyn Color Slide II	6	*Positive	6	100.00%
Alere Clearview Mono Cassette	10	*Positive	10	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	2	*Positive	2	100.00%
NDC ProAdvantage Mono	1	*Positive	1	100.00%
Consult Diag Mono Cassette	4	*Positive	4	100.00%
IM-9		* Denotes Correct Response		
AG: All Methods	78			
		*Negative	78	100.00%
Beck Coulter ICON (WB)	1	*Negative	1	100.00%
Alere Acceava (WB)	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	33	*Negative	33	100.00%
LifeSign Status Inf Mono	2	*Negative	2	100.00%
Lifesign Unistep	1	*Negative	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	3	*Negative	3	100.00%
Polymedco Poly stat Mono - (WB)	2	*Negative	2	100.00%
Polymedco Poly stat Mono (S/P)	2	*Negative	2	100.00%
Quidel QuickVue+ (S/P)	2	*Negative	2	100.00%
Quidel QuickVue+ (WB)	8	*Negative	8	100.00%
Seradyn Color Slide II	6	*Negative	6	100.00%
Alere Clearview Mono Cassette	10	*Negative	10	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	2	*Negative	2	100.00%
NDC ProAdvantage Mono	1	*Negative	1	100.00%
Consult Diag Mono Cassette	4	*Negative	4	100.00%
IM-10		* Denotes Correct Response		
AG: All Methods	77			
		*Positive	76	98.70%
		Negative	1	1.30%
Beck Coulter ICON (WB)	1	*Positive	1	100.00%
Alere Acceava (WB)	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	33	*Positive	33	100.00%
LifeSign Status Inf Mono	2	*Positive	2	100.00%
Lifesign Unistep	1	*Positive	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	3	*Positive	3	100.00%
Polymedco Poly stat Mono - (WB)	2	*Positive	2	100.00%
Polymedco Poly stat Mono (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (WB)	8	*Positive	8	100.00%
Seradyn Color Slide II	5	*Positive	5	100.00%
Alere Clearview Mono Cassette	10	*Positive	9	90.00%
Sekisui Diagnostics OSOM Mono Non-Waived	2	*Positive	2	100.00%
NDC ProAdvantage Mono	1	*Positive	1	100.00%
Consult Diag Mono Cassette	4	*Positive	4	100.00%
Alere Clearview Mono Cassette	10	Negative	1	10.00%
		Infectious Mono (Waived)		
IM-6		* Denotes Correct Response		
AG: All Methods	487			
		*Positive	474	97.33%
		Negative	13	2.67%
Beck Coulter ICON (WB)	1	*Positive	1	100.00%
Alere Acceava (S)	2	*Positive	2	100.00%
Alere Acceava (WB)	52	*Positive	51	98.08%



Sekisui Diagnostics OSOM Mono (WB)	193	*Positive	190	98.45%
Henry Schein One-Step+ Mono	34	*Positive	33	97.06%
Instant Technologies i-Screen (WB)	2	*Positive	2	100.00%
Inverness Signify Mono	4	*Positive	4	100.00%
LifeSign Status Inf Mono	34	*Positive	33	97.06%
McKesson Medi-Lab Perf - Inf Mono	20	*Positive	19	95.00%
Polymedco Poly stat Mono - (WB)	7	*Positive	7	100.00%
Quidel QuickVue+ (WB)	25	*Positive	22	88.00%
Remel A.R.C. Mono Test	1	*Positive	1	100.00%
SP Brand (Cardinal) Rapid Test	7	*Positive	7	100.00%
Alere Clearview Mono Cassette	32	*Positive	30	93.75%
Pulse Scientific	1	*Positive	1	100.00%
Diag Test Group Clarity Mono	4	*Positive	4	100.00%
Fisher HC Sure-View Mono	2	*Positive	2	100.00%
Immunostics Inc. ImmunoDetector Mono (WB)	1	*Positive	1	100.00%
Meridian Immunocard STAT! Mono (WB)	5	*Positive	5	100.00%
NDC ProAdvantage Mono	2	*Positive	2	100.00%
Consult Diag Mono Cassette	52	*Positive	51	98.08%
Alere Acceava (WB) Mono Cassette	5	*Positive	5	100.00%
Medline Mono Cassette (WB)	1	*Positive	1	100.00%
Alere Acceava (WB)	52	Negative	1	1.92%
Sekisui Diagnostics OSOM Mono (WB)	193	Negative	3	1.55%
Henry Schein One-Step+ Mono	34	Negative	1	2.94%
LifeSign Status Inf Mono	34	Negative	1	2.94%
McKesson Medi-Lab Perf - Inf Mono	20	Negative	1	5.00%
Quidel QuickVue+ (WB)	25	Negative	3	12.00%
Alere Clearview Mono Cassette	32	Negative	2	6.25%
Consult Diag Mono Cassette	52	Negative	1	1.92%
IM-7		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Diag Test Group Clarity Mono	1	*Negative	1	100.00%
IM-8		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Diag Test Group Clarity Mono	1	*Positive	1	100.00%
IM-9		* Denotes Correct Response		
AG: All Methods	1			
		Negative	1	100.00%
Diag Test Group Clarity Mono	1	Negative	1	100.00%
IM-10		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Diag Test Group Clarity Mono	1	*Positive	1	100.00%
		Infectious Mononucleosis (International)		
IM-6		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Alere Clearview Mono Cassette	1	*Positive	1	100.00%
Intl Oxoid Infectious Mono	1	*Positive	1	100.00%
Fisher HC Sure-View Mono	1	*Positive	1	100.00%
IM-7		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Alere Clearview Mono Cassette	1	*Negative	1	100.00%
Intl Oxoid Infectious Mono	1	*Negative	1	100.00%
Fisher HC Sure-View Mono	1	*Negative	1	100.00%
IM-8		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Alere Clearview Mono Cassette	1	*Positive	1	100.00%
Intl Oxoid Infectious Mono	1	*Positive	1	100.00%
Fisher HC Sure-View Mono	1	*Positive	1	100.00%
IM-9		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Alere Clearview Mono Cassette	1	*Negative	1	100.00%
Intl Oxoid Infectious Mono	1	*Negative	1	100.00%



Test Name	Labs	Response	Number	Percent
Fisher HC Sure-Vue Mono	1	*Negative	1	100.00%
IM-10		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Alere Clearview Mono Cassette	1	*Positive	1	100.00%
Intl Oxoid Infectious Mono	1	*Positive	1	100.00%
Fisher HC Sure-Vue Mono	1	*Positive	1	100.00%
637 - KOH - Glass Slides (Skin Hair Nail)	Labs	Response	Number	Percent
		KOH Preparation		
KOH-3		* Denotes Correct Response		
AG: All Methods	23			
		*No fungal elements/yeast observed	23	100.00%
KOH Glass Slide	23	*No fungal elements/yeast observed	23	100.00%
KOH-4		* Denotes Correct Response		
AG: All Methods	23			
		*Fungal elements/yeast observed	23	100.00%
KOH Glass Slide	23	*Fungal elements/yeast observed	23	100.00%
654 - Lyme Serology	Labs	Response	Number	Percent
		Lyme Serology		
LY-2		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
bioMerieux Mini Vidas	1	*Positive	1	100.00%
758 - Mycoplasma Antibody	Labs	Response	Number	Percent
		Mycoplasma Antibody		
MY-2		* Denotes Correct Response		
AG: All Methods	68			
		*Positive	67	98.53%
		Negative	1	1.47%
Meridian ImmunoCard	68	*Positive	67	98.53%
Meridian ImmunoCard		Negative	1	1.47%
791 - Fecal Occult Blood Add-On;601 - Waived/PPM Tests II;609 - Waived/PPM Tests I;647 - Mini-Waived Package;628 - AFC Waived/PPM	Labs	Response	Number	Percent
		Fecal Occult Blood		
OC-2		* Denotes Correct Response		
AG: All Methods	854			
		*Negative	817	95.67%
		Positive	37	4.33%
Occult Blood	854	*Negative	817	95.67%
Occult Blood		Positive	37	4.33%
962 - S. pneumoniae Antigen (2 specimens);974 - S. pneumoniae Antigen (5 specimens)	Labs	Response	Number	Percent
		Streptococcus pneumoniae Antigen Detection		
PNU-6		* Denotes Correct Response		
AG: All Methods	64			
		*Positive	64	100.00%
Alere BinaxNOW Strep pneumoniae Antigen Card	64	*Positive	64	100.00%
PNU-7		* Denotes Correct Response		
AG: All Methods	64			
		*Negative	64	100.00%
Alere BinaxNOW Strep pneumoniae Antigen Card	64	*Negative	64	100.00%
974 - S. pneumoniae Antigen (5 specimens)	Labs	Response	Number	Percent
		Streptococcus pneumoniae Antigen Detection		
PNU-6		* Denotes Correct Response		
AG: All Methods	64			
		*Positive	64	100.00%
Alere BinaxNOW Strep pneumoniae Antigen Card	64	*Positive	64	100.00%
PNU-7		* Denotes Correct Response		
AG: All Methods	64			
		*Negative	64	100.00%
Alere BinaxNOW Strep pneumoniae Antigen Card	64	*Negative	64	100.00%
PNU-8		* Denotes Correct Response		
AG: All Methods	64			
		*Positive	64	100.00%
Alere BinaxNOW Strep pneumoniae Antigen Card	64	*Positive	64	100.00%
PNU-9		* Denotes Correct Response		
AG: All Methods	64			



		*Negative	62	96.88%
		Positive	2	3.13%
Alere BinaxNOW Strep pneumoniae Antigen Card	64	*Negative	62	96.88%
Alere BinaxNOW Strep pneumoniae Antigen Card		Positive	2	3.13%
PNU-10		* Denotes Correct Response		
AG: All Methods	64			
		*Positive	62	96.88%
		Negative	2	3.13%
Alere BinaxNOW Strep pneumoniae Antigen Card	64	*Positive	62	96.88%
Alere BinaxNOW Strep pneumoniae Antigen Card		Negative	2	3.13%
950 - Post Vasectomy Sperm Analysis	Labs	Response	Number	Percent
		Sperm Count (Post Vasectomy)		
PVAS-3		* Denotes Correct Response		
AG: All Methods	16			
		*Positive for Sperm	15	93.75%
		Negative for Sperm	1	6.25%
Hemocytometer	4	*Positive for Sperm	4	100.00%
Post Vasectomy Microscopy	12	*Positive for Sperm	11	91.67%
Post Vasectomy Microscopy		Negative for Sperm	1	8.33%
PVAS-4		* Denotes Correct Response		
AG: All Methods	16			
		*Positive for Sperm	16	100.00%
Hemocytometer	4	*Positive for Sperm	4	100.00%
Post Vasectomy Microscopy	12	*Positive for Sperm	12	100.00%
754 - Infectious Mono/RF Combo;653 - Rheumatoid Factor;759 - Complete Immunology	Labs	Response	Number	Percent
		Rheumatoid Factor		
RF-6		* Denotes Correct Response		
AG: All Methods	12			
		*N/A (Negative)	12	100.00%
ASI	1	*N/A (Negative)	1	100.00%
Siemens RapiTex	2	*N/A (Negative)	2	100.00%
Seradyn Color Slide	2	*N/A (Negative)	2	100.00%
Wampole ColorCard RF	2	*N/A (Negative)	2	100.00%
Wampole Rheumatex	2	*N/A (Negative)	2	100.00%
Wampole Rheumaton	1	*N/A (Negative)	1	100.00%
Biokit Rhemajet RF	1	*N/A (Negative)	1	100.00%
Remel SeraTest RF-Latex	1	*N/A (Negative)	1	100.00%
RF-7		* Denotes Correct Response		
AG: All Methods	12			
		*N/A (Negative)	12	100.00%
ASI	1	*N/A (Negative)	1	100.00%
Siemens RapiTex	2	*N/A (Negative)	2	100.00%
Seradyn Color Slide	2	*N/A (Negative)	2	100.00%
Wampole ColorCard RF	2	*N/A (Negative)	2	100.00%
Wampole Rheumatex	2	*N/A (Negative)	2	100.00%
Wampole Rheumaton	1	*N/A (Negative)	1	100.00%
Biokit Rhemajet RF	1	*N/A (Negative)	1	100.00%
Remel SeraTest RF-Latex	1	*N/A (Negative)	1	100.00%
RF-8		* Denotes Correct Response		
AG: All Methods	12			
		*Positive	11	91.67%
		N/A (Negative)	1	8.33%
ASI	1	*Positive	1	100.00%
Siemens RapiTex	2	*Positive	2	100.00%
Seradyn Color Slide	2	*Positive	2	100.00%
Wampole ColorCard RF	2	*Positive	2	100.00%
Wampole Rheumatex	2	*Positive	2	100.00%
Biokit Rhemajet RF	1	*Positive	1	100.00%
Remel SeraTest RF-Latex	1	*Positive	1	100.00%
Wampole Rheumaton	1	N/A (Negative)	1	100.00%
RF-9		* Denotes Correct Response		
AG: All Methods	12			
		*Positive	11	91.67%
		N/A (Negative)	1	8.33%
ASI	1	*Positive	1	100.00%
Siemens RapiTex	2	*Positive	2	100.00%
Seradyn Color Slide	2	*Positive	2	100.00%



Wampole ColorCard RF	2	*Positive	2	100.00%
Wampole Rheumatex	2	*Positive	2	100.00%
Biokit Rhemajet RF	1	*Positive	1	100.00%
Remel SeraTest RF-Latex	1	*Positive	1	100.00%
Wampole Rheumaton	1	N/A (Negative)	1	100.00%
RF-10		* Denotes Correct Response		
AG: All Methods	12			
		*N/A (Negative)	12	100.00%
ASI	1	*N/A (Negative)	1	100.00%
Siemens RapiTex	2	*N/A (Negative)	2	100.00%
Seradyn Color Slide	2	*N/A (Negative)	2	100.00%
Wampole ColorCard RF	2	*N/A (Negative)	2	100.00%
Wampole Rheumatex	2	*N/A (Negative)	2	100.00%
Wampole Rheumaton	1	*N/A (Negative)	1	100.00%
Biokit Rhemajet RF	1	*N/A (Negative)	1	100.00%
Remel SeraTest RF-Latex	1	*N/A (Negative)	1	100.00%
		Rheumatoid Factor (International)		
RF-6		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Wampole ColorCard RF	1	*Negative	1	100.00%
Fisher HC Sure-Vue RF	1	*Negative	1	100.00%
Immunostics Rhuemacol/LEX	1	*Negative	1	100.00%
RF-7		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Wampole ColorCard RF	1	*Negative	1	100.00%
Fisher HC Sure-Vue RF	1	*Negative	1	100.00%
Immunostics Rhuemacol/LEX	1	*Negative	1	100.00%
RF-8		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Wampole ColorCard RF	1	*Positive	1	100.00%
Fisher HC Sure-Vue RF	1	*Positive	1	100.00%
Immunostics Rhuemacol/LEX	1	*Positive	1	100.00%
RF-9		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Wampole ColorCard RF	1	*Positive	1	100.00%
Fisher HC Sure-Vue RF	1	*Positive	1	100.00%
Immunostics Rhuemacol/LEX	1	*Positive	1	100.00%
RF-10		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Wampole ColorCard RF	1	*Negative	1	100.00%
Fisher HC Sure-Vue RF	1	*Negative	1	100.00%
Immunostics Rhuemacol/LEX	1	*Negative	1	100.00%
		Rheumatoid Factor (Semi Quant)		
RF-6		* Denotes Correct Response		
PG: Fisher HC Sure-Vue RF (Semi Quant)	1			
		*Negative	1	100.00%
RF-7		* Denotes Correct Response		
PG: Fisher HC Sure-Vue RF (Semi Quant)	1			
		*Negative	1	100.00%
RF-8		* Denotes Correct Response		
PG: Fisher HC Sure-Vue RF (Semi Quant)	1			
		*80 IU/mL	1	100.00%
RF-9		* Denotes Correct Response		
PG: Fisher HC Sure-Vue RF (Semi Quant)	1			
		*80 IU/mL	1	100.00%
RF-10		* Denotes Correct Response		
PG: Fisher HC Sure-Vue RF (Semi Quant)	1			
		*Negative	1	100.00%
645 - Fetal Membrane Rupture	Labs	Response	Number	Percent
		Fetal Membranes Rupture PAMG-1		
ROM-3		* Denotes Correct Response		
AG: All Methods	16			
		*Positive	15	93.75%



		Negative	1	6.25%
AmniSure ROM Test	16	*Positive	15	93.75%
AmniSure ROM Test		Negative	1	6.25%
ROM-4		* Denotes Correct Response		
AG: All Methods	16			
		*Negative	16	100.00%
AmniSure ROM Test	16	*Negative	16	100.00%
601 - Waived/PPM Tests II;609 - Waived/PPM Tests I;785 - Presump ID/Colony Count-Urine & Strep Combo;691 - Group A Strep Antigen (5 specimens);699 - Group A Strep Antigen (1 specimen);783 - Group A Strep Antigen/Add-on (2 specimens);665 - Group A Strep An	Labs	Response	Number	Percent
		Strep A Ag Det. (Waived)		
RS-6		* Denotes Correct Response		
AG: All Methods	865			
		*Negative	853	98.61%
		Positive	12	1.39%
Abbott Signify Strep A - Waived	8	*Negative	8	100.00%
Beck Coulter ICON DS Strep	1	*Negative	1	100.00%
Beck Coulter ICON SC Strep	2	*Negative	2	100.00%
Alere Acceava Strep A Dipstick	93	*Negative	93	100.00%
Diag Test Group Clarity - Strip Strep Ag	1	*Negative	1	100.00%
Fisher HC Sure-Vue Strep A (Waived)	11	*Negative	11	100.00%
Sekisui Diagnostics OSOM	44	*Negative	44	100.00%
Sekisui Diagnostics OSOM Ultra Strep	54	*Negative	53	98.15%
Henry Schein OneStep + Rapid Cass	4	*Negative	4	100.00%
Henry Schein OneStep+ Rapid Dip	65	*Negative	64	98.46%
Immunostics Immuno/Strep A Det	1	*Negative	1	100.00%
Innovacon One Step Strep Ag	1	*Negative	1	100.00%
Instant Technologies i Strep	2	*Negative	2	100.00%
Inverness Signify Strep A Dipstick	7	*Negative	7	100.00%
McKesson Strep A Cas	4	*Negative	4	100.00%
McKesson Strep A Dip	39	*Negative	37	94.87%
McKesson Strep A Test - Twist	2	*Negative	2	100.00%
Permaxim Rediscreen Strep A Rapid	1	*Negative	1	100.00%
Polymedco Poly Stat Strep A (W)	3	*Negative	3	100.00%
PSS Rapid Test Strip Strep A	2	*Negative	2	100.00%
Quidel QuickVue Dipstick Strep	95	*Negative	94	98.95%
Quidel QuickVue In-Line - 25 Test	94	*Negative	88	93.62%
Quidel QuickVue+ Strep A	39	*Negative	38	97.44%
Select Diag Strep A Dip	4	*Negative	4	100.00%
SP (Cardinal) Rapid Strep Cassette	5	*Negative	5	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	72	*Negative	72	100.00%
Stanbio QuStick Strep A	1	*Negative	1	100.00%
Alere Clearview Exact II Cas	1	*Negative	1	100.00%
Alere Clearview Exact II Dip	38	*Negative	38	100.00%
Sekisui Diagnostics OSOM Ultra	5	*Negative	5	100.00%
Consult Diag (PSS) Strep A Dipstick	115	*Negative	115	100.00%
Consult Diag (PSS) Strep A Twist	4	*Negative	4	100.00%
MooreBrand Strep A Dipstick	5	*Negative	5	100.00%
NDC ProAdvantage Strep Strips	1	*Negative	1	100.00%
Jant Accutest Accustrip Strep A	3	*Negative	3	100.00%
Meridian ImmunoCard STAT Strep A	1	*Negative	1	100.00%
LifeSign Bio Strep A Test STRip	2	*Negative	2	100.00%
Quidel Sofia Strep A FIA	3	*Negative	3	100.00%
BD Veritor Strep A	14	*Negative	14	100.00%
Alere Acceava Strep A Dipstick	12	*Negative	12	100.00%
Alere i Strep A	3	*Negative	3	100.00%
Medline Strep A Dipstick	2	*Negative	2	100.00%
Sekisui Diagnostics OSOM Ultra Strep	54	Positive	1	1.85%
Henry Schein OneStep+ Rapid Dip	65	Positive	1	1.54%
McKesson Strep A Dip	39	Positive	2	5.13%
Quidel QuickVue Dipstick Strep	95	Positive	1	1.05%
Quidel QuickVue In-Line - 25 Test	94	Positive	6	6.38%
Quidel QuickVue+ Strep A	39	Positive	1	2.56%
RS-7		* Denotes Correct Response		
AG: All Methods	70			
		*Positive	68	97.14%
		Negative	2	2.86%
Sekisui Diagnostics OSOM	3	*Positive	3	100.00%



Sekisui Diagnostics OSOM Ultra Strep	2	*Positive	2	100.00%
Innovacon One Step Strep Ag	1	*Positive	1	100.00%
Instant Technologies i Strep	1	*Positive	1	100.00%
Quidel QuickVue Dipstick Strep	31	*Positive	31	100.00%
Quidel QuickVue In-Line - 25 Test	4	*Positive	3	75.00%
Quidel QuickVue+ Strep A	7	*Positive	7	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	5	*Positive	5	100.00%
Consult Diag (PSS) Strep A Dipstick	8	*Positive	8	100.00%
Consult Diag (PSS) Strep A Twist	1	*Positive	1	100.00%
Jant Accutest Accustrip Strep A	3	*Positive	3	100.00%
BD Veritor Strep A	1	*Positive	1	100.00%
Alere Acceava Strep A Dipstick	1	*Positive	1	100.00%
Alere i Strep A	1	*Positive	1	100.00%
McKesson Strep A Dip	1	Negative	1	100.00%
Quidel QuickVue In-Line - 25 Test	4	Negative	1	25.00%
		Strep A Aq Det.		
RS-6		* Denotes Correct Response		
AG: All Methods	208			
		*Negative	206	99.04%
		Positive	2	0.96%
Abbott Signify Strep A - Waived	2	*Negative	2	100.00%
BD Check Strep A (Dip)	2	*Negative	2	100.00%
Beck Coulter ICON DS Strep	2	*Negative	2	100.00%
Alere Acceava Strep A Dipstick	17	*Negative	17	100.00%
Diag Test Group Clarity - Strip Strep Aq	1	*Negative	1	100.00%
Fisher HC Sure-Vue Select Strep Aq	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM	23	*Negative	23	100.00%
Sekisui Diagnostics OSOM Ultra Strep	28	*Negative	28	100.00%
Henry Schein OneStep+ Rapid Dip	6	*Negative	6	100.00%
Immunostics Immuno/Strep A Det	2	*Negative	2	100.00%
Inverness Signify Strep A Dipstick	1	*Negative	1	100.00%
McKesson Strep A Cas	2	*Negative	2	100.00%
McKesson Strep A Dip	11	*Negative	11	100.00%
Permaxim Rediscreen Strep A Rapid	1	*Negative	1	100.00%
Quidel QuickVue Dipstick Strep	25	*Negative	25	100.00%
Quidel QuickVue In-Line - 25 Test	14	*Negative	14	100.00%
Quidel QuickVue+ Strep A	40	*Negative	39	97.50%
Select Diag Strep A Dip	1	*Negative	1	100.00%
SP (Cardinal) Rapid Strep Cassette	1	*Negative	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	*Negative	3	100.00%
Stanbio QuStick Strep A	2	*Negative	2	100.00%
Alere Clearview Exact II Dip	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM Ultra	1	*Negative	1	100.00%
Consult Diag (PSS) Strep A Dipstick	12	*Negative	12	100.00%
Consult Diag (PSS) Strep A Twist	1	*Negative	1	100.00%
Quidel Sofia Strep A FIA	5	*Negative	5	100.00%
BD Veritor Strep A	2	*Negative	2	100.00%
PSS Rapid Test Strip Strep A	1	Positive	1	100.00%
Quidel QuickVue+ Strep A	40	Positive	1	2.50%
RS-7		* Denotes Correct Response		
AG: All Methods	203			
		*Positive	194	95.57%
		Negative	9	4.43%
Abbott Signify Strep A - Waived	2	*Positive	2	100.00%
BD Check Strep A (Dip)	2	*Positive	2	100.00%
Beck Coulter ICON DS Strep	2	*Positive	1	50.00%
Alere Acceava Strep A Dipstick	16	*Positive	14	87.50%
Diag Test Group Clarity - Strip Strep Aq	1	*Positive	1	100.00%
Fisher HC Sure-Vue Select Strep Aq	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM	23	*Positive	23	100.00%
Sekisui Diagnostics OSOM Ultra Strep	27	*Positive	26	96.30%
Henry Schein OneStep+ Rapid Dip	6	*Positive	6	100.00%
Immunostics Immuno/Strep A Det	2	*Positive	1	50.00%
Inverness Signify Strep A Dipstick	1	*Positive	1	100.00%
McKesson Strep A Cas	2	*Positive	2	100.00%
McKesson Strep A Dip	11	*Positive	11	100.00%
Permaxim Rediscreen Strep A Rapid	1	*Positive	1	100.00%
Quidel QuickVue Dipstick Strep	25	*Positive	24	96.00%
Quidel QuickVue In-Line - 25 Test	14	*Positive	14	100.00%



Quidel QuickVue+ Strep A	38	*Positive	37	97.37%
Select Diag Strep A Dip	1	*Positive	1	100.00%
SP (Cardinal) Rapid Strep Cassette	1	*Positive	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	*Positive	2	66.67%
Stanbio QuStick Strep A	2	*Positive	2	100.00%
Alere Clearview Exact II Dip	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Ultra	1	*Positive	1	100.00%
Consult Diag (PSS) Strep A Dipstick	12	*Positive	12	100.00%
Consult Diag (PSS) Strep A Twist	1	*Positive	1	100.00%
Quidel Sofia Strep A FIA	4	*Positive	4	100.00%
BD Veritor Strep A	2	*Positive	2	100.00%
Beck Coulter ICON DS Strep	2	Negative	1	50.00%
Alere Aceava Strep A Dipstick	16	Negative	2	12.50%
Sekisui Diagnostics OSOM Ultra Strep	27	Negative	1	3.70%
Immunostics Immuno/Strep A Det	2	Negative	1	50.00%
PSS Rapid Test Strip Strep A	1	Negative	1	100.00%
Quidel QuickVue Dipstick Strep	25	Negative	1	4.00%
Quidel QuickVue+ Strep A	38	Negative	1	2.63%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	Negative	1	33.33%
		Strep A Ag Det. (International)		
RS-6		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Alere Clearview Exact II Cas	1	*Negative	1	100.00%
		RS-7		
		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Alere Clearview Exact II Cas	1	*Positive	1	100.00%
691 - Group A Strep Antigen (5 specimens)		Labs	Response	Number
			Strep A Ag Det.	Percent
			* Denotes Correct Response	
RS-6				
AG: All Methods	208			
		*Negative	206	99.04%
		Positive	2	0.96%
Abbott Signify Strep A - Waived	2	*Negative	2	100.00%
BD Check Strep A (Dip)	2	*Negative	2	100.00%
Beck Coulter ICON DS Strep	2	*Negative	2	100.00%
Alere Aceava Strep A Dipstick	17	*Negative	17	100.00%
Diag Test Group Clarity - Strip Strep Ag	1	*Negative	1	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM	23	*Negative	23	100.00%
Sekisui Diagnostics OSOM Ultra Strep	28	*Negative	28	100.00%
Henry Schein OneStep+ Rapid Dip	6	*Negative	6	100.00%
Immunostics Immuno/Strep A Det	2	*Negative	2	100.00%
Inverness Signify Strep A Dipstick	1	*Negative	1	100.00%
McKesson Strep A Cas	2	*Negative	2	100.00%
McKesson Strep A Dip	11	*Negative	11	100.00%
Permaxim Rediscreen Strep A Rapid	1	*Negative	1	100.00%
Quidel QuickVue Dipstick Strep	25	*Negative	25	100.00%
Quidel QuickVue In-Line - 25 Test	14	*Negative	14	100.00%
Quidel QuickVue+ Strep A	40	*Negative	39	97.50%
Select Diag Strep A Dip	1	*Negative	1	100.00%
SP (Cardinal) Rapid Strep Cassette	1	*Negative	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	*Negative	3	100.00%
Stanbio QuStick Strep A	2	*Negative	2	100.00%
Alere Clearview Exact II Dip	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM Ultra	1	*Negative	1	100.00%
Consult Diag (PSS) Strep A Dipstick	12	*Negative	12	100.00%
Consult Diag (PSS) Strep A Twist	1	*Negative	1	100.00%
Quidel Sofia Strep A FIA	5	*Negative	5	100.00%
BD Veritor Strep A	2	*Negative	2	100.00%
PSS Rapid Test Strip Strep A	1	Positive	1	100.00%
Quidel QuickVue+ Strep A	40	Positive	1	2.50%
		RS-7		
		* Denotes Correct Response		
AG: All Methods	203			
		*Positive	194	95.57%
		Negative	9	4.43%
Abbott Signify Strep A - Waived	2	*Positive	2	100.00%
BD Check Strep A (Dip)	2	*Positive	2	100.00%



Beck Coulter ICON DS Strep	2	*Positive	1	50.00%
Alere Acceava Strep A Dipstick	16	*Positive	14	87.50%
Diag Test Group Clarity - Strip Strep Ag	1	*Positive	1	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM	23	*Positive	23	100.00%
Sekisui Diagnostics OSOM Ultra Strep	27	*Positive	26	96.30%
Henry Schein OneStep+ Rapid Dip	6	*Positive	6	100.00%
Immunostics Immuno/Strep A Det	2	*Positive	1	50.00%
Inverness Signify Strep A Dipstick	1	*Positive	1	100.00%
McKesson Strep A Cas	2	*Positive	2	100.00%
McKesson Strep A Dip	11	*Positive	11	100.00%
Permaxim Rediscreen Strep A Rapid	1	*Positive	1	100.00%
Quidel QuickVue Dipstick Strep	25	*Positive	24	96.00%
Quidel QuickVue In-Line - 25 Test	14	*Positive	14	100.00%
Quidel QuickVue+ Strep A	38	*Positive	37	97.37%
Select Diag Strep A Dip	1	*Positive	1	100.00%
SP (Cardinal) Rapid Strep Cassette	1	*Positive	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	*Positive	2	66.67%
Stanbio QuStick Strep A	2	*Positive	2	100.00%
Alere Clearview Exact II Dip	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Ultra	1	*Positive	1	100.00%
Consult Diag (PSS) Strep A Dipstick	12	*Positive	12	100.00%
Consult Diag (PSS) Strep A Twist	1	*Positive	1	100.00%
Quidel Sofia Strep A FIA	4	*Positive	4	100.00%
BD Veritor Strep A	2	*Positive	2	100.00%
Beck Coulter ICON DS Strep	2	Negative	1	50.00%
Alere Acceava Strep A Dipstick	16	Negative	2	12.50%
Sekisui Diagnostics OSOM Ultra Strep	27	Negative	1	3.70%
Immunostics Immuno/Strep A Det	2	Negative	1	50.00%
PSS Rapid Test Strip Strep A	1	Negative	1	100.00%
Quidel QuickVue Dipstick Strep	25	Negative	1	4.00%
Quidel QuickVue+ Strep A	38	Negative	1	2.63%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	Negative	1	33.33%
RS-8		* Denotes Correct Response		
AG: All Methods	174			
		*Negative	170	97.70%
		Positive	4	2.30%
Abbott Signify Strep A - Waived	2	*Negative	2	100.00%
BD Check Strep A (Dip)	2	*Negative	2	100.00%
Beck Coulter ICON DS Strep	2	*Negative	2	100.00%
Alere Acceava Strep A Dipstick	17	*Negative	16	94.12%
Diag Test Group Clarity - Strip Strep Ag	1	*Negative	1	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM	23	*Negative	23	100.00%
Sekisui Diagnostics OSOM Ultra Strep	23	*Negative	23	100.00%
Henry Schein OneStep+ Rapid Dip	6	*Negative	6	100.00%
Immunostics Immuno/Strep A Det	2	*Negative	2	100.00%
Inverness Signify Strep A Dipstick	1	*Negative	1	100.00%
McKesson Strep A Cas	1	*Negative	1	100.00%
McKesson Strep A Dip	5	*Negative	5	100.00%
Permaxim Rediscreen Strep A Rapid	1	*Negative	1	100.00%
Quidel QuickVue Dipstick Strep	19	*Negative	19	100.00%
Quidel QuickVue In-Line - 25 Test	11	*Negative	11	100.00%
Quidel QuickVue+ Strep A	34	*Negative	32	94.12%
SP (Cardinal) Rapid Strep Cassette	1	*Negative	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	*Negative	3	100.00%
Stanbio QuStick Strep A	1	*Negative	1	100.00%
Alere Clearview Exact II Dip	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM Ultra	1	*Negative	1	100.00%
Consult Diag (PSS) Strep A Dipstick	10	*Negative	10	100.00%
Consult Diag (PSS) Strep A Twist	1	*Negative	1	100.00%
Quidel Sofia Strep A FIA	3	*Negative	3	100.00%
BD Veritor Strep A	1	*Negative	1	100.00%
Alere Acceava Strep A Dipstick	17	Positive	1	5.88%
PSS Rapid Test Strip Strep A	1	Positive	1	100.00%
Quidel QuickVue+ Strep A	34	Positive	2	5.88%
RS-9		* Denotes Correct Response		
AG: All Methods	175			
		*Positive	167	95.43%



		Negative	8	4.57%
Abbott Signify Strep A - Waived	2	*Positive	2	100.00%
BD Check Strep A (Dip)	2	*Positive	2	100.00%
Beck Coulter ICON DS Strep	2	*Positive	1	50.00%
Alere Acceava Strep A Dipstick	17	*Positive	16	94.12%
Diag Test Group Clarity - Strip Strep Ag	1	*Positive	1	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM	23	*Positive	23	100.00%
Sekisui Diagnostics OSOM Ultra Strep	23	*Positive	23	100.00%
Henry Schein OneStep+ Rapid Dip	6	*Positive	6	100.00%
Immunostics Immuno/Strep A Det	2	*Positive	1	50.00%
Inverness Signify Strep A Dipstick	1	*Positive	1	100.00%
McKesson Strep A Cas	1	*Positive	1	100.00%
McKesson Strep A Dip	5	*Positive	5	100.00%
Permaxim Rediscreen Strep A Rapid	1	*Positive	1	100.00%
Quidel QuickVue Dipstick Strep	19	*Positive	18	94.74%
Quidel QuickVue In-Line - 25 Test	11	*Positive	11	100.00%
Quidel QuickVue+ Strep A	34	*Positive	32	94.12%
SP (Cardinal) Rapid Strep Cassette	1	*Positive	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	*Positive	2	66.67%
Stanbio QuStick Strep A	1	*Positive	1	100.00%
Alere Clearview Exact II Dip	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Ultra	1	*Positive	1	100.00%
Consult Diag (PSS) Strep A Dipstick	10	*Positive	10	100.00%
Consult Diag (PSS) Strep A Twist	1	*Positive	1	100.00%
Quidel Sofia Strep A FIA	4	*Positive	4	100.00%
BD Veritor Strep A	1	*Positive	1	100.00%
Beck Coulter ICON DS Strep	2	Negative	1	50.00%
Alere Acceava Strep A Dipstick	17	Negative	1	5.88%
Immunostics Immuno/Strep A Det	2	Negative	1	50.00%
PSS Rapid Test Strip Strep A	1	Negative	1	100.00%
Quidel QuickVue Dipstick Strep	19	Negative	1	5.26%
Quidel QuickVue+ Strep A	34	Negative	2	5.88%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	Negative	1	33.33%
RS-10		* Denotes Correct Response		
AG: All Methods	175			
		*Negative	172	98.29%
		Positive	3	1.71%
Abbott Signify Strep A - Waived	2	*Negative	2	100.00%
BD Check Strep A (Dip)	2	*Negative	2	100.00%
Beck Coulter ICON DS Strep	2	*Negative	2	100.00%
Alere Acceava Strep A Dipstick	17	*Negative	17	100.00%
Diag Test Group Clarity - Strip Strep Ag	1	*Negative	1	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM	23	*Negative	23	100.00%
Sekisui Diagnostics OSOM Ultra Strep	23	*Negative	23	100.00%
Henry Schein OneStep+ Rapid Dip	6	*Negative	6	100.00%
Immunostics Immuno/Strep A Det	2	*Negative	2	100.00%
Inverness Signify Strep A Dipstick	1	*Negative	1	100.00%
McKesson Strep A Cas	1	*Negative	1	100.00%
McKesson Strep A Dip	5	*Negative	5	100.00%
Permaxim Rediscreen Strep A Rapid	1	*Negative	1	100.00%
PSS Rapid Test Strip Strep A	1	*Negative	1	100.00%
Quidel QuickVue Dipstick Strep	19	*Negative	19	100.00%
Quidel QuickVue In-Line - 25 Test	11	*Negative	11	100.00%
Quidel QuickVue+ Strep A	34	*Negative	31	91.18%
SP (Cardinal) Rapid Strep Cassette	1	*Negative	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	*Negative	3	100.00%
Stanbio QuStick Strep A	1	*Negative	1	100.00%
Alere Clearview Exact II Dip	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM Ultra	1	*Negative	1	100.00%
Consult Diag (PSS) Strep A Dipstick	10	*Negative	10	100.00%
Consult Diag (PSS) Strep A Twist	1	*Negative	1	100.00%
Quidel Sofia Strep A FIA	4	*Negative	4	100.00%
BD Veritor Strep A	1	*Negative	1	100.00%
Quidel QuickVue+ Strep A	34	Positive	3	8.82%
RS-6		Strep A Ag Det. (Waived)		
		* Denotes Correct Response		
AG: All Methods	865			



		*Negative	853	98.61%
		Positive	12	1.39%
Abbott Signify Strep A - Waived	8	*Negative	8	100.00%
Beck Coulter ICON DS Strep	1	*Negative	1	100.00%
Beck Coulter ICON SC Strep	2	*Negative	2	100.00%
Alere Acceava Strep A Dipstick	93	*Negative	93	100.00%
Diag Test Group Clarity - Strip Strep Ag	1	*Negative	1	100.00%
Fisher HC Sure-Vue Strep A (Waived)	11	*Negative	11	100.00%
Sekisui Diagnostics OSOM	44	*Negative	44	100.00%
Sekisui Diagnostics OSOM Ultra Strep	54	*Negative	53	98.15%
Henry Schein OneStep + Rapid Cass	4	*Negative	4	100.00%
Henry Schein OneStep+ Rapid Dip	65	*Negative	64	98.46%
Immunostics Immuno/Strep A Det	1	*Negative	1	100.00%
Innovacon One Step Strep Ag	1	*Negative	1	100.00%
Instant Technologies i Strep	2	*Negative	2	100.00%
Inverness Signify Strep A Dipstick	7	*Negative	7	100.00%
McKesson Strep A Cas	4	*Negative	4	100.00%
McKesson Strep A Dip	39	*Negative	37	94.87%
McKesson Strep A Test - Twist	2	*Negative	2	100.00%
Permaxim Rediscreen Strep A Rapid	1	*Negative	1	100.00%
Polymedco Poly Stat Strep A (W)	3	*Negative	3	100.00%
PSS Rapid Test Strip Strep A	2	*Negative	2	100.00%
Quidel QuickVue Dipstick Strep	95	*Negative	94	98.95%
Quidel QuickVue In-Line - 25 Test	94	*Negative	88	93.62%
Quidel QuickVue+ Strep A	39	*Negative	38	97.44%
Select Diag Strep A Dip	4	*Negative	4	100.00%
SP (Cardinal) Rapid Strep Cassette	5	*Negative	5	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	72	*Negative	72	100.00%
Stanbio QuStick Strep A	1	*Negative	1	100.00%
Alere Clearview Exact II Cas	1	*Negative	1	100.00%
Alere Clearview Exact II Dip	38	*Negative	38	100.00%
Sekisui Diagnostics OSOM Ultra	5	*Negative	5	100.00%
Consult Diag (PSS) Strep A Dipstick	115	*Negative	115	100.00%
Consult Diag (PSS) Strep A Twist	4	*Negative	4	100.00%
MooreBrand Strep A Dipstick	5	*Negative	5	100.00%
NDC ProAdvantage Strep Strips	1	*Negative	1	100.00%
Jant Accutest Accustrip Strep A	3	*Negative	3	100.00%
Meridian ImmunoCard STAT Strep A	1	*Negative	1	100.00%
LifeSign Bio Strep A Test Strip	2	*Negative	2	100.00%
Quidel Sofia Strep A FIA	3	*Negative	3	100.00%
BD Veritor Strep A	14	*Negative	14	100.00%
Alere Acceava Strep A Dipstick	12	*Negative	12	100.00%
Alere i Strep A	3	*Negative	3	100.00%
Medline Strep A Dipstick	2	*Negative	2	100.00%
Sekisui Diagnostics OSOM Ultra Strep	54	Positive	1	1.85%
Henry Schein OneStep+ Rapid Dip	65	Positive	1	1.54%
McKesson Strep A Dip	39	Positive	2	5.13%
Quidel QuickVue Dipstick Strep	95	Positive	1	1.05%
Quidel QuickVue In-Line - 25 Test	94	Positive	6	6.38%
Quidel QuickVue+ Strep A	39	Positive	1	2.56%
RS-7		* Denotes Correct Response		
AG: All Methods	70			
		*Positive	68	97.14%
		Negative	2	2.86%
Sekisui Diagnostics OSOM	3	*Positive	3	100.00%
Sekisui Diagnostics OSOM Ultra Strep	2	*Positive	2	100.00%
Innovacon One Step Strep Ag	1	*Positive	1	100.00%
Instant Technologies i Strep	1	*Positive	1	100.00%
Quidel QuickVue Dipstick Strep	31	*Positive	31	100.00%
Quidel QuickVue In-Line - 25 Test	4	*Positive	3	75.00%
Quidel QuickVue+ Strep A	7	*Positive	7	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	5	*Positive	5	100.00%
Consult Diag (PSS) Strep A Dipstick	8	*Positive	8	100.00%
Consult Diag (PSS) Strep A Twist	1	*Positive	1	100.00%
Jant Accutest Accustrip Strep A	3	*Positive	3	100.00%
BD Veritor Strep A	1	*Positive	1	100.00%
Alere Acceava Strep A Dipstick	1	*Positive	1	100.00%
Alere i Strep A	1	*Positive	1	100.00%
McKesson Strep A Dip	1	Negative	1	100.00%



Quidel QuickVue In-Line - 25 Test	4	Negative	1	25.00%
RS-8		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Alere i Strep A	1	*Negative	1	100.00%
RS-9		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Alere i Strep A	1	*Positive	1	100.00%
RS-10		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Alere i Strep A	1	*Negative	1	100.00%
		Strep A Ag Det. (International)		
RS-6		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Alere Clearview Exact II Cas	1	*Negative	1	100.00%
RS-7		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Alere Clearview Exact II Cas	1	*Positive	1	100.00%
RS-8		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Alere Clearview Exact II Cas	1	*Negative	1	100.00%
RS-9		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Alere Clearview Exact II Cas	1	*Positive	1	100.00%
RS-10		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Alere Clearview Exact II Cas	1	*Negative	1	100.00%
997 - CVS Group A Strep - 2 samples	Labs	Response	Number	Percent
		Strep A Ag Det. (Waived)		
RS-6		* Denotes Correct Response		
AG: All Methods	38			
		*Negative	37	97.37%
		Positive	1	2.63%
Sekisui Diagnostics OSOM Ultra Strep	38	*Negative	37	97.37%
Sekisui Diagnostics OSOM Ultra Strep		Positive	1	2.63%
RS-7		* Denotes Correct Response		
AG: All Methods	38			
		*Positive	37	97.37%
		Negative	1	2.63%
Sekisui Diagnostics OSOM Ultra Strep	38	*Positive	37	97.37%
Sekisui Diagnostics OSOM Ultra Strep		Negative	1	2.63%
673 - Rubella:759 - Complete Immunology	Labs	Response	Number	Percent
		Rubella (Qualitative)		
RU-6		* Denotes Correct Response		
PG: bioMerieux Vidas, Mini	1			
		*Positive	1	100.00%
RU-7		* Denotes Correct Response		
PG: bioMerieux Vidas, Mini	1			
		*Negative	1	100.00%
RU-8		* Denotes Correct Response		
PG: bioMerieux Vidas, Mini	1			
		*Negative	1	100.00%
RU-9		* Denotes Correct Response		
PG: bioMerieux Vidas, Mini	1			
		*Negative	1	100.00%
RU-10		* Denotes Correct Response		
PG: bioMerieux Vidas, Mini	1			
		*Positive	1	100.00%
739 - Sickle Cell Screen	Labs	Response	Number	Percent
		Sickle Cell Screen		



SS-2		* Denotes Correct Response		
AG: All Methods	9			
		*Negative	9	100.00%
Pacific Hemostasis Sickle Screen	1	*Negative	1	100.00%
Streck SickleDex	8	*Negative	8	100.00%
679 - Syphilis Serology	Labs	Response	Number	Percent
		RPR		
SY-6		* Denotes Correct Response		
AG: All Methods	10			
		*Nonreactive	10	100.00%
ASI RPR	2	*Nonreactive	2	100.00%
Becton Dickinson BBL	3	*Nonreactive	3	100.00%
New Horizons	1	*Nonreactive	1	100.00%
Stanbio Quicktest	1	*Nonreactive	1	100.00%
Wampole Impact	1	*Nonreactive	1	100.00%
Diagnostic Direct Syphilis Health Check	1	*Nonreactive	1	100.00%
Intl ASI RPR	1	*Nonreactive	1	100.00%
SY-7		* Denotes Correct Response		
AG: All Methods	10			
		*Reactive	10	100.00%
ASI RPR	2	*Reactive	2	100.00%
Becton Dickinson BBL	3	*Reactive	3	100.00%
New Horizons	1	*Reactive	1	100.00%
Stanbio Quicktest	1	*Reactive	1	100.00%
Wampole Impact	1	*Reactive	1	100.00%
Diagnostic Direct Syphilis Health Check	1	*Reactive	1	100.00%
Intl ASI RPR	1	*Reactive	1	100.00%
SY-8		* Denotes Correct Response		
AG: All Methods	10			
		*Nonreactive	10	100.00%
ASI RPR	2	*Nonreactive	2	100.00%
Becton Dickinson BBL	3	*Nonreactive	3	100.00%
New Horizons	1	*Nonreactive	1	100.00%
Stanbio Quicktest	1	*Nonreactive	1	100.00%
Wampole Impact	1	*Nonreactive	1	100.00%
Diagnostic Direct Syphilis Health Check	1	*Nonreactive	1	100.00%
Intl ASI RPR	1	*Nonreactive	1	100.00%
SY-9		* Denotes Correct Response		
AG: All Methods	10			
		*Reactive	9	90.00%
		Weakly Reactive	1	10.00%
ASI RPR	2	*Reactive	2	100.00%
Becton Dickinson BBL	3	*Reactive	3	100.00%
New Horizons	1	*Reactive	1	100.00%
Stanbio Quicktest	1	*Reactive	1	100.00%
Diagnostic Direct Syphilis Health Check	1	*Reactive	1	100.00%
Intl ASI RPR	1	*Reactive	1	100.00%
Wampole Impact	1	Weakly Reactive	1	100.00%
SY-10		* Denotes Correct Response		
AG: All Methods	10			
		*Reactive	10	100.00%
ASI RPR	2	*Reactive	2	100.00%
Becton Dickinson BBL	3	*Reactive	3	100.00%
New Horizons	1	*Reactive	1	100.00%
Stanbio Quicktest	1	*Reactive	1	100.00%
Wampole Impact	1	*Reactive	1	100.00%
Diagnostic Direct Syphilis Health Check	1	*Reactive	1	100.00%
Intl ASI RPR	1	*Reactive	1	100.00%
		RPR/VDRL (titer)		
SY-6		* Denotes Correct Response		
AG: All Methods	2			
		*N/A (Negative)	1	50.00%
		*Nonreactive	1	50.00%
ASI RPR	1	*N/A (Negative)	1	100.00%
New Horizons	1	*Nonreactive	1	100.00%
SY-7		* Denotes Correct Response		
AG: All Methods	2			
		*1:2	2	100.00%



ASI RPR	1	*1:2	1	100.00%
New Horizons	1	*1:2	1	100.00%
SY-8		* Denotes Correct Response		
AG: All Methods	2			
		*N/A (Negative)	1	50.00%
		*Nonreactive	1	50.00%
ASI RPR	1	*N/A (Negative)	1	100.00%
New Horizons	1	*Nonreactive	1	100.00%
SY-9		* Denotes Correct Response		
AG: All Methods	2			
		*1:2	2	100.00%
ASI RPR	1	*1:2	1	100.00%
New Horizons	1	*1:2	1	100.00%
SY-10		* Denotes Correct Response		
AG: All Methods	2			
		*1:2	1	50.00%
		*1:4	1	50.00%
New Horizons	1	*1:2	1	100.00%
ASI RPR	1	*1:4	1	100.00%
		RPR (International)		
SY-6		* Denotes Correct Response		
AG: All Methods	3			
		*Nonreactive	3	100.00%
BD Macro-Vue RPR	1	*Nonreactive	1	100.00%
Fisher HC Sure-Vue RPR	1	*Nonreactive	1	100.00%
Wampole Impact	1	*Nonreactive	1	100.00%
SY-7		* Denotes Correct Response		
AG: All Methods	3			
		*Reactive	3	100.00%
BD Macro-Vue RPR	1	*Reactive	1	100.00%
Fisher HC Sure-Vue RPR	1	*Reactive	1	100.00%
Wampole Impact	1	*Reactive	1	100.00%
SY-8		* Denotes Correct Response		
AG: All Methods	3			
		*Nonreactive	3	100.00%
BD Macro-Vue RPR	1	*Nonreactive	1	100.00%
Fisher HC Sure-Vue RPR	1	*Nonreactive	1	100.00%
Wampole Impact	1	*Nonreactive	1	100.00%
SY-9		* Denotes Correct Response		
AG: All Methods	3			
		*Reactive	2	66.67%
		Weakly Reactive	1	33.33%
BD Macro-Vue RPR	1	*Reactive	1	100.00%
Fisher HC Sure-Vue RPR	1	*Reactive	1	100.00%
Wampole Impact	1	Weakly Reactive	1	100.00%
SY-10		* Denotes Correct Response		
AG: All Methods	3			
		*Reactive	3	100.00%
BD Macro-Vue RPR	1	*Reactive	1	100.00%
Fisher HC Sure-Vue RPR	1	*Reactive	1	100.00%
Wampole Impact	1	*Reactive	1	100.00%
776 - Throat Culture;780 - Throat & Genital Cultures	Labs	Response	Number	Percent
TC-6		* Denotes Correct Response		
AG: All Methods	82			
		*Beta hemolysis, referred for ID/confirmation	2	2.44%
		*Positive for Group A Streptococcus	42	51.22%
		*Presumptive Positive for Group A Streptococcus	35	42.68%
		*Presumptive Streptococcus species	1	1.22%
		Negative for Group A Streptococcus	2	2.44%
Cultures	82	*Beta hemolysis, referred for ID/confirmation	2	2.44%
Cultures		*Positive for Group A Streptococcus	42	51.22%
Cultures		*Presumptive Positive for Group A Streptococcus	35	42.68%
Cultures		*Presumptive Streptococcus species	1	1.22%
Cultures		Negative for Group A Streptococcus	2	2.44%
TC-7		* Denotes Correct Response		
AG: All Methods	82			
		*Beta hemolysis, referred for ID/confirmation	2	2.44%



		*Positive for Group A Streptococcus	44	53.66%
		*Presumptive Positive for Group A Streptococcus	35	42.68%
		*Presumptive Streptococcus species	1	1.22%
Cultures	82	*Beta hemolysis, referred for ID/confirmation	2	2.44%
Cultures		*Positive for Group A Streptococcus	44	53.66%
Cultures		*Presumptive Positive for Group A Streptococcus	35	42.68%
Cultures		*Presumptive Streptococcus species	1	1.22%
TC-8		* Denotes Correct Response		
AG: All Methods	84			
		*Negative for Group A Streptococcus	70	83.33%
		*No beta hemolysis	3	3.57%
		*No Growth - Strep Selective medium	10	11.90%
		Beta hemolysis, referred for ID/confirmation	1	1.19%
Cultures	84	*Negative for Group A Streptococcus	70	83.33%
Cultures		*No beta hemolysis	3	3.57%
Cultures		*No Growth - Strep Selective medium	10	11.90%
Cultures		Beta hemolysis, referred for ID/confirmation	1	1.19%
TC-9		* Denotes Correct Response		
AG: All Methods	106			
		*Beta hemolysis, referred for ID/confirmation	8	7.55%
		*Growth, referred for identification	2	1.89%
		*Negative for Group A Streptococcus	72	67.92%
		*No Growth - Strep Selective medium	7	6.60%
		*Presumptive Streptococcus species	2	1.89%
		*Streptococcus, beta-hemolytic, not Group A	3	2.83%
		Positive for Group A Streptococcus	7	6.60%
		Presumptive Positive for Group A Streptococcus	5	4.72%
Cultures	106	*Beta hemolysis, referred for ID/confirmation	8	7.55%
Cultures		*Growth, referred for identification	2	1.89%
Cultures		*Negative for Group A Streptococcus	72	67.92%
Cultures		*No Growth - Strep Selective medium	7	6.60%
Cultures		*Presumptive Streptococcus species	2	1.89%
Cultures		*Streptococcus, beta-hemolytic, not Group A	3	2.83%
Cultures		Positive for Group A Streptococcus	7	6.60%
Cultures		Presumptive Positive for Group A Streptococcus	5	4.72%
TC-10		* Denotes Correct Response		
AG: All Methods	106			
		*Growth, referred for identification	3	2.83%
		*Negative for Group A Streptococcus	80	75.47%
		*No beta hemolysis	4	3.77%
		*No Growth - Strep Selective medium	17	16.04%
		Positive for Group A Streptococcus	2	1.89%
Cultures	106	*Growth, referred for identification	3	2.83%
Cultures		*Negative for Group A Streptococcus	80	75.47%
Cultures		*No beta hemolysis	4	3.77%
Cultures		*No Growth - Strep Selective medium	17	16.04%
Cultures		Positive for Group A Streptococcus	2	1.89%
640 - Urinalysis;642 - Urinalysis/hCG;639 - Complete Urinalysis Package;601 - Waived/PPM Tests II;609 - Waived/PPM Tests I;647 - Mini-Waived Package;628 - AFC Waived/PPM;999 - CVS Complete Urinalysis	Labs	Response	Number	Percent
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	21			
		*1.020	4	19.05%
		*1.025	16	76.19%
		>=1.030	1	4.76%
UA-2		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	31			
		*1.005	14	45.16%
		*1.010	15	48.39%
		1.015	2	6.45%
UA-2		* Denotes Correct Response		
MG: Method Group	1290			
		*1.010	105	8.14%
		*1.015	182	14.11%
		*1.020	284	22.02%
		*1.025	635	49.22%



		1.000	2	0.16%
		1.005	54	4.19%
		1.030	28	2.17%
UA-2		* Denotes Correct Response		
MG: Method Group	31			
		*1.005	14	45.16%
		*1.010	15	48.39%
		1.015	2	6.45%
UA-2		* Denotes Correct Response		
MG: Plus Sys - Siemens	22			
		*1.020	5	22.73%
		*1.025	16	72.73%
		>=1.030	1	4.55%
		Urine Creatinine (semi-quant)		
UA-2		* Denotes Correct Response		
AG: All Methods	6			
		*200 mg/dL	1	16.67%
		*300 mg/dL	5	83.33%
Siemens Clinitek Status	5	*200 mg/dL	1	20.00%
Siemens Clinitek Status		*300 mg/dL	4	80.00%
Siemens Clinitek Advantus	1	*300 mg/dL	1	100.00%
		Urine Bilirubin		
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	21			
		*2+	13	61.90%
		*3+	7	33.33%
		1+	1	4.76%
UA-2		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	31			
		*3 mg/dL	7	22.58%
		*6 mg/dL	24	77.42%
UA-2		* Denotes Correct Response		
MG: Method Group	1278			
		*Large (3+)	560	43.82%
		*Moderate (2+)	600	46.95%
		Negative	20	1.56%
		Trace/Small (1+)	98	7.67%
UA-2		* Denotes Correct Response		
MG: Method Group	31			
		*3 mg/dL	7	22.58%
		*6 mg/dL	24	77.42%
UA-2		* Denotes Correct Response		
MG: Plus Sys - Siemens	22			
		*2+	14	63.64%
		*3+	7	31.82%
		1+	1	4.55%
		Urine Ketones		
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	21			
		*2+	20	95.24%
		4+	1	4.76%
UA-2		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	31			
		*15 mg/dL	21	67.74%
		*50 mg/dL	10	32.26%
UA-2		* Denotes Correct Response		
MG: Method Group	1301			
		*Large (3+) / 80 mg/dL	119	9.15%
		*Moderate (2+) / 40 mg/dL	1057	81.25%
		*Small (1+) / 15mg/dL	119	9.15%
		>=160 mg/dL	5	0.38%
		Negative	1	0.08%
UA-2		* Denotes Correct Response		
MG: Method Group	31			
		*15 mg/dL	21	67.74%
		*50 mg/dL	10	32.26%



UA-2		* Denotes Correct Response		
MG: Plus Sys - Siemens	22			
		*2+	21	95.45%
		4+	1	4.55%
UA-2		Urine pH		
		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	21			
		*7.0	21	100.00%
UA-2		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	31			
		*7.0	31	100.00%
UA-2		* Denotes Correct Response		
MG: Method Group	1298			
		*6.5	70	5.39%
		*7.0	1049	80.82%
		*7.5	146	11.25%
		5.0 or less	1	0.08%
		6.0	15	1.16%
		8.0 or greater	17	1.31%
UA-2		* Denotes Correct Response		
MG: Method Group	31			
		*7.0	31	100.00%
UA-2		* Denotes Correct Response		
MG: Plus Sys - Siemens	22			
		*7.0	22	100.00%
UA-2		Urine Blood/Hemoglobin		
		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	21			
		*2+	1	4.76%
		*3+	20	95.24%
UA-2		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	31			
		*250 Ery/uL	31	100.00%
UA-2		* Denotes Correct Response		
MG: Method Group	1306			
		*Large (3+)	1214	92.96%
		*Moderate (2+)	79	6.05%
		Negative	5	0.38%
		Small (1+)	4	0.31%
		Trace	4	0.31%
UA-2		* Denotes Correct Response		
MG: Method Group	31			
		*250 Ery/uL	31	100.00%
UA-2		* Denotes Correct Response		
MG: Plus Sys - Siemens	22			
		*2+	1	4.55%
		*3+	21	95.45%
UA-2		Urine Glucose (Reducing Sub)		
		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	21			
		*2+	19	90.48%
		*3+	2	9.52%
UA-2		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	31			
		*>= 1000 mg/dL	30	96.77%
		250 mg/dL	1	3.23%
UA-2		* Denotes Correct Response		
MG: Method Group	1313			
		*>= 1000 mg/dL	332	25.29%
		*250 mg/dL	111	8.45%
		*500 mg/dL	854	65.04%
		<100 mg/dL	5	0.38%
		100 mg/dL	10	0.76%
		Negative	1	0.08%
UA-2		* Denotes Correct Response		



MG: Method Group	31			
		*>= 1000 mg/dL	30	96.77%
		250 mg/dL	1	3.23%
UA-2		* Denotes Correct Response		
MG: Plus Sys - Siemens	22			
		*2+	20	90.91%
		*3+	2	9.09%
UA-2		Urine Protein - Qual		
		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	21			
		*3+	21	100.00%
UA-2		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	31			
		*100 mg/dL	16	51.61%
		*500 mg/dL	14	45.16%
		30 mg/dL	1	3.23%
UA-2		* Denotes Correct Response		
MG: Method Group	1316			
		*>= 1000 (4+)	70	5.32%
		*100 (2+)	161	12.23%
		*300-500 (3+)	1067	81.08%
		30 (1+)	14	1.06%
		Negative	1	0.08%
		Trace	3	0.23%
UA-2		* Denotes Correct Response		
MG: Method Group	31			
		*100 mg/dL	16	51.61%
		*500 mg/dL	14	45.16%
		30 mg/dL	1	3.23%
UA-2		* Denotes Correct Response		
MG: Plus Sys - Siemens	22			
		*3+	22	100.00%
UA-2		Urine Urobilinogen		
		* Denotes Correct Response		
PG: Siemens Clinitek 50 / Plus Sys - Siemens / Siemens Urine Urobilinogen	3			
		*4.0 EU/dL	2	66.67%
		1.0 EU/dL	1	33.33%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	16			
		*16 umol/L	2	12.50%
		*33 umol/L	5	31.25%
		*66 umol/L	7	43.75%
		131 umol/L	1	6.25%
		3.2 umol/L	1	6.25%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Urobilinogen	4			
		*4.0 EU/dL	4	100.00%
UA-2		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	31			
		*4.0 mg/dL	30	96.77%
		1 mg/dL	1	3.23%
UA-2		* Denotes Correct Response		
MG: Method Group	1271			
		*>= 8.0 mg/dL	80	6.29%
		*0.2/Normal	85	6.69%
		*1.0 mg/dL	67	5.27%
		*2.0 mg/dL	338	26.59%
		*4.0 mg/dL	701	55.15%
UA-2		* Denotes Correct Response		
MG: Method Group	31			
		*4.0 mg/dL	30	96.77%
		1 mg/dL	1	3.23%
UA-2		* Denotes Correct Response		
MG: Method Group	23			
		*16 umol/L	2	8.70%



		*33 umol/L	5	21.74%
		*4.0 EU/dL	6	26.09%
		*66 umol/L	7	30.43%
		1.0 EU/dL	1	4.35%
		131 umol/L	1	4.35%
		3.2 umol/L	1	4.35%
Urine Leuk Esterase				
UA-2 * Denotes Correct Response				
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	21			
		*1+	14	66.67%
		*2+	6	28.57%
		Trace	1	4.76%
UA-2 * Denotes Correct Response				
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	31			
		*500 Leu/uL	31	100.00%
UA-2 * Denotes Correct Response				
MG: Method Group	1306			
		*Large (3+)	125	9.57%
		*Moderate (2+)	552	42.27%
		*Small (1+)	580	44.41%
		Negative	10	0.77%
		Trace	39	2.99%
UA-2 * Denotes Correct Response				
MG: Method Group	31			
		*500 Leu/uL	31	100.00%
UA-2 * Denotes Correct Response				
MG: Plus Sys - Siemens	22			
		*1+	15	68.18%
		*2+	6	27.27%
		Trace	1	4.55%
Urine Nitrite				
UA-2 * Denotes Correct Response				
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	21			
		*Positive	21	100.00%
UA-2 * Denotes Correct Response				
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	31			
		*Positive	31	100.00%
UA-2 * Denotes Correct Response				
MG: Method Group	1302			
		*Positive	1288	98.92%
		Negative	14	1.08%
UA-2 * Denotes Correct Response				
MG: Method Group	31			
		*Positive	31	100.00%
UA-2 * Denotes Correct Response				
MG: Plus Sys - Siemens	22			
		*Positive	22	100.00%
Urine Microalbumin				
UA-2 * Denotes Correct Response				
AG: All Methods	140			
		*100 mg/L (3+)	58	41.43%
		*150 mg/L	59	42.14%
		*50 mg/L (2+)	15	10.71%
		10 mg/L (Trace)	3	2.14%
		20/30 mg/L (Small)	2	1.43%
		80 mg/L	1	0.71%
		Negative	2	1.43%
Bayer Albustix	1	*100 mg/L (3+)	1	100.00%
Siemens Clinitek Status	43	*100 mg/L (3+)	1	2.33%
Siemens Multistix - Visual	3	*100 mg/L (3+)	3	100.00%
Roche Chemstrip Micral	73	*100 mg/L (3+)	52	71.23%
Roche Chemstrips - Visual	2	*100 mg/L (3+)	1	50.00%
Siemens Clinitek 10 / 100	3	*150 mg/L	3	100.00%
Siemens Clinitek 50	13	*150 mg/L	12	92.31%
Siemens Clinitek Status	43	*150 mg/L	41	95.35%
Roche Chemstrip Micral	73	*150 mg/L	1	1.37%
Hemocue Albumin 201	2	*150 mg/L	2	100.00%



Roche Chemstrip Micral	73	*50 mg/L (2+)	15	20.55%
Siemens Clinitek 50	13	10 mg/L (Trace)	1	7.69%
Siemens Clinitek Status	43	10 mg/L (Trace)	1	2.33%
Roche Chemstrip Micral	73	10 mg/L (Trace)	1	1.37%
Roche Chemstrip Micral		20/30 mg/L (Small)	2	2.74%
Roche Chemstrips - Visual	2	80 mg/L	1	50.00%
Roche Chemstrip Micral	73	Negative	2	2.74%
		Fluid pH		
UA-2		* Denotes Correct Response		
PG: Nitrazine Paper	9			
		*7.0	7	77.78%
		*7.5	1	11.11%
		6.0	1	11.11%
		Urine hCG		
UA-2		* Denotes Correct Response		
AG: All Methods	1135			
		*Positive	1112	97.97%
		Negative	23	2.03%
Abbott Signify hCG Combo	1	*Positive	1	100.00%
Abbott Signify hCG Urine Dipstick	1	*Positive	1	100.00%
ASI ProPhase Plus	1	*Positive	1	100.00%
Baby Test (Dollar General)	1	*Positive	1	100.00%
Siemens Clinitek Status	11	*Positive	11	100.00%
Beck Coulter ICON 20 HCG	7	*Positive	7	100.00%
Beck Coulter ICON 25 hCG	91	*Positive	91	100.00%
Alere Acceava HCG Basic II	20	*Positive	20	100.00%
Alere Acceava hCG Combo	5	*Positive	5	100.00%
Alere Acceava hCG-Urine	11	*Positive	11	100.00%
Conception Tech One Step hCG	2	*Positive	2	100.00%
Diag Test Group Clarity hCG Plus Urine Cassettes	6	*Positive	6	100.00%
Fisher HC Sure-Vue hCG STAT	5	*Positive	4	80.00%
Fisher HC Sure-Vue Urine hCG	14	*Positive	13	92.86%
Fisher HC Sure-Vue U/S hCG	4	*Positive	4	100.00%
Sekisui Diagnostics OSOM Urine	17	*Positive	17	100.00%
Sekisui Diagnostics OSOM Card Pregnancy	37	*Positive	36	97.30%
Sekisui Diagnostics OSOM hCG Combo	45	*Positive	44	97.78%
Henry Schein One Step + (hCG)	63	*Positive	59	93.65%
Immunostics Detector Combo	3	*Positive	3	100.00%
Immunostics hCG Detector-urine	3	*Positive	3	100.00%
Instant Tech. i Pregnancy	4	*Positive	4	100.00%
Inverness Clearview Combo II	5	*Positive	5	100.00%
Inv Clearview Urine - Cassette	3	*Positive	3	100.00%
Inverness Signify hCG combo	4	*Positive	4	100.00%
Inverness Signify hCG urine	5	*Positive	5	100.00%
LifeSign Status hCG	1	*Positive	1	100.00%
McKesson Medi-Lab hCG - Urine	46	*Positive	46	100.00%
Meridian Immunocard Stat!	1	*Positive	1	100.00%
Permaxim RediScreen	3	*Positive	3	100.00%
Polymedco Poly stat hCG Urine	1	*Positive	1	100.00%
Quidel QuickVue + hCG	31	*Positive	29	93.55%
Quidel QuickVue One-Step hCG Combo U/S	97	*Positive	94	96.91%
Quidel QuickVue One-Step Urine	192	*Positive	186	96.88%
Quidel RapidVue hCG - Urine	10	*Positive	10	100.00%
SAS Ultra HCG	4	*Positive	4	100.00%
Select Diag hCG Cassette (urine)	24	*Positive	23	95.83%
Select Diag hCG Combo	12	*Positive	12	100.00%
Select Diag hCG Dipstick	7	*Positive	7	100.00%
SP (Cardinal) Brand combo	14	*Positive	13	92.86%
SP (Cardinal) Brand-cassette	6	*Positive	6	100.00%
SP Brand (Cardinal) HCG Combo	34	*Positive	34	100.00%
Stanbio Qupid	6	*Positive	6	100.00%
Stanbio QuPID Plus	1	*Positive	1	100.00%
Stanbio QuStick	2	*Positive	2	100.00%
Stanbio TRUE hCG	5	*Positive	5	100.00%
MooreBrand hCG Cassette	6	*Positive	6	100.00%
Nova Scientific Century S/U hCG	2	*Positive	2	100.00%
Diag Test Group Clarity hCG Urine Strips	2	*Positive	2	100.00%
Jant AccuTest	8	*Positive	8	100.00%
Becton Dickinson Precise One-Step	1	*Positive	1	100.00%



Consult Diag (PSS) hCG Urine Cassette	133	*Positive	133	100.00%
CLIAwaived Inc Pregnancy Urine Dip-Strip	6	*Positive	5	83.33%
Germaine AIM Step Combo Pregnancy	1	*Positive	1	100.00%
NDC ProAdvantage Urine hCG Cassette	5	*Positive	5	100.00%
NDC ProAdvantage Urine hCG Dipstick	1	*Positive	1	100.00%
Consult Diag (PSS) hCG Urine Dipstick	20	*Positive	20	100.00%
SP (Cardinal) Brand Dipstick	9	*Positive	9	100.00%
Consult Diag hCG Combo Cassette (Urine)	7	*Positive	7	100.00%
McKesson Medi-Lab Performance hCG Urine Cassette	8	*Positive	7	87.50%
Alere hCG Combo Cassette (25 mIU/mL)	6	*Positive	6	100.00%
Alere hCG Cassette (25 mIU/mL)	3	*Positive	3	100.00%
Alere hCG Combo Cassette (20/10 mIU/mL)	1	*Positive	1	100.00%
Alere hCG Dipstick	5	*Positive	5	100.00%
Medline Pregnancy Urine Dip-Strip	7	*Positive	7	100.00%
Medline Pregnancy Urine Cassette Test	4	*Positive	4	100.00%
McKesson hCG Urine Test Dipstick	7	*Positive	7	100.00%
Medline Pregnancy Combo Cassette (urine)	1	*Positive	1	100.00%
Quidel QuickVue hCG Urine Test	14	*Positive	14	100.00%
MooreBrand hCG Dipstick	4	*Positive	4	100.00%
Good Sense One Step Pregnancy Test	1	*Positive	1	100.00%
Henry Schein One Step + HCG Combo	1	*Positive	1	100.00%
Henry Schein One Step hCG Urine Cassette	3	*Positive	3	100.00%
Medline hCG Test Strip 25 mIU/mL	2	*Positive	2	100.00%
Fisher HC Sure-Vue hCG STAT	5	Negative	1	20.00%
Fisher HC Sure-Vue Urine hCG	14	Negative	1	7.14%
Sekisui Diagnostics OSOM Card Pregnancy	37	Negative	1	2.70%
Sekisui Diagnostics OSOM hCG Combo	45	Negative	1	2.22%
Henry Schein One Step + (hCG)	63	Negative	4	6.35%
Quidel QuickVue + hCG	31	Negative	2	6.45%
Quidel QuickVue One-Step hCG Combo U/S	97	Negative	3	3.09%
Quidel QuickVue One-Step Urine	192	Negative	6	3.13%
Select Diag hCG Cassette (urine)	24	Negative	1	4.17%
SP (Cardinal) Brand combo	14	Negative	1	7.14%
CLIAwaived Inc Pregnancy Urine Dip-Strip	6	Negative	1	16.67%
McKesson Medi-Lab Performance hCG Urine Cassette	8	Negative	1	12.50%
		Urine hCG (non-waived)		
UA-2		* Denotes Correct Response		
PG: Quidel Sofia hCG FIA (urine)	1			
		*Positive	1	100.00%
661 - Urine Colony Count;784 - Presump ID/Colony Count - Urine;785 - Presump ID/Colony Count-Urine & Strep Combo;775 - Urine Culture;778 - Urine & Throat Cultures;779 - Urine & Genital Cultures;781 - Urine, Throat & Genital Cultures;782 - Complete Bacteri	Labs	Response	Number	Percent
		Susceptibility Cefazolin		
UC-6		* Denotes Correct Response		
AG: All Methods	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
		Susceptibility Cefepime		
UC-6		* Denotes Correct Response		
AG: All Methods	2			
		*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
		Susceptibility Cefotaxime		
UC-6		* Denotes Correct Response		
AG: All Methods	2			
		*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
		Urine Culture		
UC-6		* Denotes Correct Response		
AG: All Methods	32			
		*Gram Negative organism	4	12.50%
		*Growth, referred for identification	17	53.13%
		*Klebsiella pneumoniae	4	12.50%
		*Presumptive Klebsiella species	2	6.25%
		*Uricult - Growth (Positive)	5	15.63%
Cultures	32	*Gram Negative organism	4	12.50%
Cultures		*Growth, referred for identification	17	53.13%



Cultures		*Klebsiella pneumoniae	4	12.50%
Cultures		*Presumptive Klebsiella species	2	6.25%
Cultures		*Uricult - Growth (Positive)	5	15.63%
UC-7		* Denotes Correct Response		
AG: All Methods	32			
		*Enterococcus faecalis	2	6.25%
		*Enterococcus species, NOS	1	3.13%
		*Gram Positive organism	3	9.38%
		*Growth, referred for identification	19	59.38%
		*Presumptive Enterococcus species	1	3.13%
		*Uricult - Growth (Positive)	4	12.50%
		Staphylococcus species, NOS	1	3.13%
		Uricult - No Growth (Negative)	1	3.13%
Cultures	32	*Enterococcus faecalis	2	6.25%
Cultures		*Enterococcus species, NOS	1	3.13%
Cultures		*Gram Positive organism	3	9.38%
Cultures		*Growth, referred for identification	19	59.38%
Cultures		*Presumptive Enterococcus species	1	3.13%
Cultures		*Uricult - Growth (Positive)	4	12.50%
Cultures		Staphylococcus species, NOS	1	3.13%
Cultures		Uricult - No Growth (Negative)	1	3.13%
UC-8		* Denotes Correct Response		
AG: All Methods	30			
		*Gram Negative organism	3	10.00%
		*Growth, referred for identification	18	60.00%
		*Presumptive Proteus species	1	3.33%
		*Proteus mirabilis	2	6.67%
		*Uricult - Growth (Positive)	5	16.67%
		No Growth	1	3.33%
Cultures	30	*Gram Negative organism	3	10.00%
Cultures		*Growth, referred for identification	18	60.00%
Cultures		*Presumptive Proteus species	1	3.33%
Cultures		*Proteus mirabilis	2	6.67%
Cultures		*Uricult - Growth (Positive)	5	16.67%
Cultures		No Growth	1	3.33%
UC-9		* Denotes Correct Response		
AG: All Methods	5			
		*Escherichia coli	1	20.00%
		*Gram Negative organism	1	20.00%
		*Presumptive E. coli	1	20.00%
		*Uricult - Growth (Positive)	2	40.00%
Cultures	5	*Escherichia coli	1	20.00%
Cultures		*Gram Negative organism	1	20.00%
Cultures		*Presumptive E. coli	1	20.00%
Cultures		*Uricult - Growth (Positive)	2	40.00%
UC-10		* Denotes Correct Response		
AG: All Methods	5			
		*Gram Negative organism	2	40.00%
		*Uricult - Growth (Positive)	2	40.00%
		Enterobacter species, NOS	1	20.00%
Cultures	5	*Gram Negative organism	2	40.00%
Cultures		*Uricult - Growth (Positive)	2	40.00%
Cultures		Enterobacter species, NOS	1	20.00%
Susceptibility Cefpodoxime		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
Susceptibility Cefazidime		* Denotes Correct Response		
AG: All Methods	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
Susceptibility Ceftriaxone		* Denotes Correct Response		
AG: All Methods	6			
		*Sensitive	6	100.00%
Susceptibility	6	*Sensitive	6	100.00%



UC-6		Susceptibility Cefuroxime		
AG: All Methods		* Denotes Correct Response		
	4			
		*Sensitive	4	100.00%
Susceptibility	4	*Sensitive	4	100.00%
UC-6		Susceptibility Cephalothin		
AG: All Methods		* Denotes Correct Response		
	6			
		*Sensitive	6	100.00%
Susceptibility	6	*Sensitive	6	100.00%
UC-6		Susceptibility Ciprofloxacin		
AG: All Methods		* Denotes Correct Response		
	10			
		*Sensitive	10	100.00%
Susceptibility	10	*Sensitive	10	100.00%
UC-6		Susceptibility Gentamicin		
AG: All Methods		* Denotes Correct Response		
	6			
		*Sensitive	6	100.00%
Susceptibility	6	*Sensitive	6	100.00%
UC-6		Susceptibility Imipenem		
AG: All Methods		* Denotes Correct Response		
	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
UC-6		Susceptibility Levofloxacin		
AG: All Methods		* Denotes Correct Response		
	6			
		*Sensitive	6	100.00%
Susceptibility	6	*Sensitive	6	100.00%
UC-6		Susceptibility Meropenem		
AG: All Methods		* Denotes Correct Response		
	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
UC-6		Susceptibility Minocycline		
AG: All Methods		* Denotes Correct Response		
	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
UC-6		Susceptibility Nitrofurantoin		
AG: All Methods		* Denotes Correct Response		
	9			
		*Intermediate	4	44.44%
		*Sensitive	5	55.56%
Susceptibility	9	*Intermediate	4	44.44%
Susceptibility		*Sensitive	5	55.56%
UC-6		Susceptibility Norfloxacin/Noroxin		
AG: All Methods		* Denotes Correct Response		
	2			
		*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
UC-6		Susceptibility Piperacillin/Tazobactam		
AG: All Methods		* Denotes Correct Response		
	2			
		*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
UC-6		Susceptibility Sulfamethoxazole		
AG: All Methods		* Denotes Correct Response		
	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
UC-6		Susceptibility Tetracycline		
AG: All Methods		* Denotes Correct Response		
	6			
		*Sensitive	5	83.33%
		Resistant	1	16.67%



Susceptibility	6	*Sensitive	5	83.33%
Susceptibility		Resistant	1	16.67%
Susceptibility Ticarcillin/Clavulanate				
UC-6		* Denotes Correct Response		
AG: All Methods	1			
Susceptibility	1	*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
Susceptibility Tobramycin				
UC-6		* Denotes Correct Response		
AG: All Methods	3			
Susceptibility	3	*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
Susceptibility Trimethoprim				
UC-6		* Denotes Correct Response		
AG: All Methods	2			
Susceptibility	2	*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
Susceptibility Trimeth./Sulfameth.				
UC-6		* Denotes Correct Response		
AG: All Methods	8			
Susceptibility	8	*Sensitive	8	100.00%
Susceptibility	8	*Sensitive	8	100.00%
Susceptibility Amikacin				
UC-6		* Denotes Correct Response		
AG: All Methods	2			
Susceptibility	2	*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
Susceptibility Amox/Clavulanate				
UC-6		* Denotes Correct Response		
AG: All Methods	6			
Susceptibility	6	*Sensitive	6	100.00%
Susceptibility	6	*Sensitive	6	100.00%
Susceptibility Ampicillin				
UC-6		* Denotes Correct Response		
AG: All Methods	8			
		*Resistant	7	87.50%
		Sensitive	1	12.50%
Susceptibility	8	*Resistant	7	87.50%
Susceptibility		Sensitive	1	12.50%
Susceptibility Ampicillin/Sulbactam				
UC-6		* Denotes Correct Response		
AG: All Methods	2			
Susceptibility	2	*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
Susceptibility Aztreonam				
UC-6		* Denotes Correct Response		
AG: All Methods	1			
Susceptibility	1	*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
Susceptibility Ertapenem				
UC-6		* Denotes Correct Response		
AG: All Methods	1			
Susceptibility	1	*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
703 - Urine Drug Screen; 713 - Urine Drug Screen (50 ml sample volume)				
UDS-3	Labs	Response	Number	Percent
Barbiturates				
UCS-3		* Denotes Correct Response		
AG: All Methods	109			
		*Negative	109	100.00%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand	1	*Negative	1	100.00%
E-Screen	8	*Negative	8	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	10	*Negative	10	100.00%
Instant Technologies i Screen	4	*Negative	4	100.00%
MedTox Profile II	1	*Negative	1	100.00%



Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	2	*Negative	2	100.00%
MedTox Profile V	3	*Negative	3	100.00%
Branan Med Corp QuickTox	4	*Negative	4	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	5	*Negative	5	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Alere iCassette	26	*Negative	26	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	2	*Negative	2	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	109			
		*Negative	109	100.00%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand	1	*Negative	1	100.00%
E-Screen	8	*Negative	8	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	10	*Negative	10	100.00%
Instant Technologies i Screen	4	*Negative	4	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	2	*Negative	2	100.00%
MedTox Profile V	3	*Negative	3	100.00%
Branan Med Corp QuickTox	4	*Negative	4	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	5	*Negative	5	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Alere iCassette	26	*Negative	26	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	2	*Negative	2	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
UDS-3		Benzodiazepines		
AG: All Methods	111	* Denotes Correct Response		
		*Positive	108	97.30%
		Negative	3	2.70%
AlereTriage	3	*Positive	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand	1	*Positive	1	100.00%
E-Screen	8	*Positive	8	100.00%
Express Diag. Express Check	1	*Positive	1	100.00%
Alere iCup Drugs of Abuse	11	*Positive	11	100.00%
Instant Technologies i Screen	4	*Positive	4	100.00%
Noble Med Split Specimen Cup	2	*Positive	2	100.00%
Beckman (Olympus) AU	1	*Positive	1	100.00%



Innovacon Multi-Clin	10	*Positive	10	100.00%
Tokyo/Carolina Biolis 24i	1	*Positive	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Positive	5	100.00%
ASC One Screen Drug Test Card	2	*Positive	2	100.00%
MedTox Profile V	3	*Positive	3	100.00%
Branan Med Corp QuickTox	4	*Positive	4	100.00%
Jant Accutest Multi Drug Screen	2	*Positive	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Positive	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	5	*Positive	4	80.00%
Mindray BS-200	2	*Positive	2	100.00%
Alere iCassette	26	*Positive	26	100.00%
Roche Hitachi 717	1	*Positive	1	100.00%
US Diagnostics UScreen Cup	2	*Positive	2	100.00%
Abbott Architect c4000	1	*Positive	1	100.00%
Discover Drug Test Card	1	*Positive	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Positive	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Positive	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Positive	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Positive	1	100.00%
MedTox Profile II	1	Negative	1	100.00%
MedTox Profile III ER	1	Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	5	Negative	1	20.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	111			
		*Negative	111	100.00%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand	1	*Negative	1	100.00%
E-Screen	8	*Negative	8	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	11	*Negative	11	100.00%
Instant Technologies i Screen	4	*Negative	4	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Noble Med Split Specimen Cup	2	*Negative	2	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	2	*Negative	2	100.00%
MedTox Profile V	3	*Negative	3	100.00%
Branan Med Corp QuickTox	4	*Negative	4	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	5	*Negative	5	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Alere iCassette	26	*Negative	26	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	2	*Negative	2	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
UDS-3		Amphetamines		
		* Denotes Correct Response		
AG: All Methods	121			
		*Negative	120	99.17%
		Positive	1	0.83%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand	2	*Negative	2	100.00%
E-Screen	10	*Negative	10	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	14	*Negative	14	100.00%
Instant Technologies i Screen	4	*Negative	4	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	2	*Negative	2	100.00%



Beckman (Olympus) AU	1	*Negative	1	100.00%
Redwood Biotech Panel Dip	1	*Negative	1	100.00%
Innovacon Multi-Clin	9	*Negative	9	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	2	*Negative	2	100.00%
MedTox Profile V	3	*Negative	3	100.00%
Branan Med Corp QuickTox	4	*Negative	4	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	5	*Negative	4	80.00%
Mindray BS-200	2	*Negative	2	100.00%
Alere iCassette	28	*Negative	28	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	2	*Negative	2	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immutest Multi-Drug Screen Card	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	5	Positive	1	20.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	121			
		*Positive	110	90.91%
		Negative	11	9.09%
AlereTriage	3	*Positive	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand	2	*Positive	2	100.00%
E-Screen	10	*Positive	3	30.00%
Express Diag. Express Check	1	*Positive	1	100.00%
Alere iCup Drugs of Abuse	14	*Positive	13	92.86%
Instant Technologies i Screen	4	*Positive	4	100.00%
MedTox Profile II	2	*Positive	2	100.00%
Noble Med Split Specimen Cup	2	*Positive	2	100.00%
Beckman (Olympus) AU	1	*Positive	1	100.00%
Redwood Biotech Panel Dip	1	*Positive	1	100.00%
Innovacon Multi-Clin	9	*Positive	9	100.00%
MedTox Profile III ER	1	*Positive	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Positive	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Positive	5	100.00%
ASC One Screen Drug Test Card	2	*Positive	2	100.00%
MedTox Profile V	3	*Positive	3	100.00%
Branan Med Corp QuickTox	4	*Positive	4	100.00%
Jant Accutest Multi Drug Screen	2	*Positive	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Positive	3	75.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	5	*Positive	5	100.00%
Mindray BS-200	2	*Positive	2	100.00%
Alere iCassette	28	*Positive	27	96.43%
Roche Hitachi 717	1	*Positive	1	100.00%
US Diagnostics UScreen Cup	2	*Positive	2	100.00%
Abbott Architect c4000	1	*Positive	1	100.00%
Discover Drug Test Card	1	*Positive	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Positive	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Positive	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Positive	3	100.00%
E-Screen	10	Negative	7	70.00%
Alere iCup Drugs of Abuse	14	Negative	1	7.14%
Amedica Biotech AmediCheck Instant Test Cup	4	Negative	1	25.00%
Alere iCassette	28	Negative	1	3.57%
Alere Immutest Multi-Drug Screen Card	1	Negative	1	100.00%
UDS-3		Cannabinoids (THC)		
		* Denotes Correct Response		
AG: All Methods	129			
		*Positive	124	96.12%
		Negative	5	3.88%
AlereTriage	3	*Positive	3	100.00%



Siemens Dim/AR/ES/RxL/Xpand	2	*Positive	2	100.00%
E-Screen	16	*Positive	16	100.00%
Express Diag. Express Check	1	*Positive	1	100.00%
Alere iCup Drugs of Abuse	14	*Positive	14	100.00%
Instant Technologies i Screen	5	*Positive	5	100.00%
Integrated E-Z Split Key Cup	1	*Positive	1	100.00%
MedTox Profile II	2	*Positive	2	100.00%
Noble Med Split Specimen Cup	2	*Positive	2	100.00%
Beckman (Olympus) AU	1	*Positive	1	100.00%
Redwood Biotech Panel Dip	1	*Positive	1	100.00%
Innovacon Multi-Clin	10	*Positive	9	90.00%
MedTox Profile III ER	1	*Positive	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Positive	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Positive	4	80.00%
ASC One Screen Drug Test Card	2	*Positive	2	100.00%
MedTox Profile V	3	*Positive	3	100.00%
Branan Med Corp QuickTox	4	*Positive	4	100.00%
Jant Accutest Multi Drug Screen	2	*Positive	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Positive	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	5	*Positive	4	80.00%
Mindray BS-200	2	*Positive	2	100.00%
Alere iCassette	28	*Positive	26	92.86%
Roche Hitachi 717	1	*Positive	1	100.00%
US Diagnostics UScreen Cup	2	*Positive	2	100.00%
Abbott Architect c4000	1	*Positive	1	100.00%
Discover Drug Test Card	1	*Positive	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Positive	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Positive	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Positive	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Positive	1	100.00%
Innovacon Multi-Clin	10	Negative	1	10.00%
CLIAwaived Inc Rapid Drug Cup	5	Negative	1	20.00%
American Screening Corp Reveal Multi-Drug Test Cup	5	Negative	1	20.00%
Alere iCassette	28	Negative	2	7.14%
UDS-4		* Denotes Correct Response		
AG: All Methods	129			
		*Negative	129	100.00%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand	2	*Negative	2	100.00%
E-Screen	16	*Negative	16	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	14	*Negative	14	100.00%
Instant Technologies i Screen	5	*Negative	5	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	2	*Negative	2	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Redwood Biotech Panel Dip	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	2	*Negative	2	100.00%
MedTox Profile V	3	*Negative	3	100.00%
Branan Med Corp QuickTox	4	*Negative	4	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	5	*Negative	5	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Alere iCassette	28	*Negative	28	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	2	*Negative	2	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%



AG: All Methods	128	*Negative	128	100.00%
Cocaine Metabolites				
UDS-3				
* Denotes Correct Response				
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand	2	*Negative	2	100.00%
E-Screen	16	*Negative	16	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	14	*Negative	14	100.00%
Instant Technologies i Screen	5	*Negative	5	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	2	*Negative	2	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Redwood Biotech Panel Dip	1	*Negative	1	100.00%
Innovacon Multi-Clin	9	*Negative	9	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	2	*Negative	2	100.00%
MedTox Profile V	3	*Negative	3	100.00%
Branan Med Corp QuickTox	4	*Negative	4	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	5	*Negative	5	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Alere iCassette	28	*Negative	28	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	2	*Negative	2	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
UDS-4				
* Denotes Correct Response				
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand	2	*Negative	2	100.00%
E-Screen	16	*Negative	16	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	14	*Negative	14	100.00%
Instant Technologies i Screen	5	*Negative	5	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	2	*Negative	2	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Redwood Biotech Panel Dip	1	*Negative	1	100.00%
Innovacon Multi-Clin	9	*Negative	9	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	2	*Negative	2	100.00%
MedTox Profile V	3	*Negative	3	100.00%
Branan Med Corp QuickTox	4	*Negative	4	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	5	*Negative	5	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Alere iCassette	28	*Negative	28	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	2	*Negative	2	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%



Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immutest Multi-Drug Screen Card	1	*Negative	1	100.00%
		Ethanol		
UDS-3		* Denotes Correct Response		
AG: All Methods	5			
		*Positive	5	100.00%
Beckman (Olympus) AU	1	*Positive	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Positive	1	100.00%
Mindray BS-200	2	*Positive	2	100.00%
Roche Hitachi 717	1	*Positive	1	100.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	5			
		*Positive	5	100.00%
Beckman (Olympus) AU	1	*Positive	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Positive	1	100.00%
Mindray BS-200	2	*Positive	2	100.00%
Roche Hitachi 717	1	*Positive	1	100.00%
		MDMA (Ecstasy)		
UDS-3		* Denotes Correct Response		
AG: All Methods	77			
		*Negative	77	100.00%
E-Screen	4	*Negative	4	100.00%
Alere iCup Drugs of Abuse	10	*Negative	10	100.00%
Instant Technologies i Screen	1	*Negative	1	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
CLIAwaived Inc Rapid Drug Cup	4	*Negative	4	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	3	*Negative	3	100.00%
Alere iCassette	26	*Negative	26	100.00%
US Diagnostics UScreen Cup	2	*Negative	2	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immutest Multi-Drug Screen Card	1	*Negative	1	100.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	77			
		*Negative	77	100.00%
E-Screen	4	*Negative	4	100.00%
Alere iCup Drugs of Abuse	10	*Negative	10	100.00%
Instant Technologies i Screen	1	*Negative	1	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
CLIAwaived Inc Rapid Drug Cup	4	*Negative	4	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	3	*Negative	3	100.00%
Alere iCassette	26	*Negative	26	100.00%
US Diagnostics UScreen Cup	2	*Negative	2	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immutest Multi-Drug Screen Card	1	*Negative	1	100.00%
		Methadone		
UDS-3		* Denotes Correct Response		
AG: All Methods	71			
		*Negative	71	100.00%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand	1	*Negative	1	100.00%
E-Screen	8	*Negative	8	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	9	*Negative	9	100.00%



Instant Technologies i Screen	3	*Negative	3	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	1	*Negative	1	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	2	*Negative	2	100.00%
MedTox Profile V	3	*Negative	3	100.00%
Branan Med Corp QuickTox	4	*Negative	4	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	5	*Negative	5	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	2	*Negative	2	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	71			
		*Negative	71	100.00%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand	1	*Negative	1	100.00%
E-Screen	8	*Negative	8	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	9	*Negative	9	100.00%
Instant Technologies i Screen	3	*Negative	3	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	1	*Negative	1	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	2	*Negative	2	100.00%
MedTox Profile V	3	*Negative	3	100.00%
Branan Med Corp QuickTox	4	*Negative	4	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	5	*Negative	5	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	2	*Negative	2	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
UDS-3		* Denotes Correct Response		
AG: All Methods	75			
		*Negative	75	100.00%
AlereTriage	3	*Negative	3	100.00%
E-Screen	10	*Negative	10	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	13	*Negative	13	100.00%
Instant Technologies i Screen	4	*Negative	4	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Noble Med Split Specimen Cup	2	*Negative	2	100.00%
Redwood Biotech Panel Dip	1	*Negative	1	100.00%
Innovacon Multi-Clin	1	*Negative	1	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%



CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	2	*Negative	2	100.00%
MedTox Profile V	3	*Negative	3	100.00%
Branan Med Corp QuickTox	4	*Negative	4	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Alere iCassette	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	2	*Negative	2	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immutest Multi-Drug Screen Card	1	*Negative	1	100.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	75			
		*Negative	74	98.67%
		Positive	1	1.33%
AlereTriage	3	*Negative	3	100.00%
E-Screen	10	*Negative	10	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	13	*Negative	13	100.00%
Instant Technologies i Screen	4	*Negative	3	75.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Noble Med Split Specimen Cup	2	*Negative	2	100.00%
Redwood Biotech Panel Dip	1	*Negative	1	100.00%
Innovacon Multi-Clin	1	*Negative	1	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	2	*Negative	2	100.00%
MedTox Profile V	3	*Negative	3	100.00%
Branan Med Corp QuickTox	4	*Negative	4	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Alere iCassette	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	2	*Negative	2	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immutest Multi-Drug Screen Card	1	*Negative	1	100.00%
Instant Technologies i Screen	4	Positive	1	25.00%
		Opiates		
UDS-3		* Denotes Correct Response		
AG: All Methods	127			
		*Negative	127	100.00%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand	2	*Negative	2	100.00%
E-Screen	16	*Negative	16	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	13	*Negative	13	100.00%
Instant Technologies i Screen	5	*Negative	5	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Noble Med Split Specimen Cup	2	*Negative	2	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Redwood Biotech Panel Dip	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	2	*Negative	2	100.00%
MedTox Profile V	3	*Negative	3	100.00%
Branan Med Corp QuickTox	4	*Negative	4	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%



Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	5	*Negative	5	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Alere iCassette	28	*Negative	28	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	2	*Negative	2	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	128			
		*Negative	127	99.22%
		Positive	1	0.78%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand	2	*Negative	2	100.00%
E-Screen	16	*Negative	16	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	14	*Negative	13	92.86%
Instant Technologies i Screen	5	*Negative	5	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Noble Med Split Specimen Cup	2	*Negative	2	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Redwood Biotech Panel Dip	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	2	*Negative	2	100.00%
MedTox Profile V	3	*Negative	3	100.00%
Branan Med Corp QuickTox	4	*Negative	4	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	5	*Negative	5	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Alere iCassette	28	*Negative	28	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	2	*Negative	2	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	14	Positive	1	7.14%
		Oxycodone		
UDS-3		* Denotes Correct Response		
AG: All Methods	88			
		*Negative	86	97.73%
		Positive	2	2.27%
Alere iCup Drugs of Abuse	10	*Negative	10	100.00%
Instant Technologies i Screen	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	2	*Negative	2	100.00%
MedTox Profile V	3	*Negative	3	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	5	*Negative	4	80.00%
Mindray BS-200	2	*Negative	2	100.00%



Alere iCassette	26	*Negative	25	96.15%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	2	*Negative	2	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	2	*Negative	2	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	5	Positive	1	20.00%
Alere iCassette	26	Positive	1	3.85%
UDS-4		* Denotes Correct Response		
AG: All Methods	88			
		*Positive	86	97.73%
		Negative	2	2.27%
Alere iCup Drugs of Abuse	10	*Positive	10	100.00%
Instant Technologies i Screen	2	*Positive	2	100.00%
Noble Med Split Specimen Cup	1	*Positive	1	100.00%
Innovacon Multi-Clin	10	*Positive	10	100.00%
Tokyo/Carolina Biolis 24i	1	*Positive	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Positive	5	100.00%
ASC One Screen Drug Test Card	2	*Positive	2	100.00%
MedTox Profile V	3	*Positive	3	100.00%
Jant Accutest Multi Drug Screen	2	*Positive	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Positive	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	5	*Positive	5	100.00%
Mindray BS-200	2	*Positive	2	100.00%
Alere iCassette	26	*Positive	25	96.15%
Roche Hitachi 717	1	*Positive	1	100.00%
US Diagnostics UScreen Cup	2	*Positive	2	100.00%
Discover Drug Test Card	1	*Positive	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	2	*Positive	2	100.00%
Diag Test Group Clarity DOA Screen	3	*Positive	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Positive	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Positive	1	100.00%
MedTox Profile III ER	1	Negative	1	100.00%
Alere iCassette	26	Negative	1	3.85%
		PCP (Phencyclidine)		
UDS-3		* Denotes Correct Response		
AG: All Methods	114			
		*Positive	109	95.61%
		Negative	5	4.39%
AlereTriage	3	*Positive	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand	2	*Positive	2	100.00%
E-Screen	16	*Positive	16	100.00%
Express Diag. Express Check	1	*Positive	1	100.00%
Alere iCup Drugs of Abuse	9	*Positive	8	88.89%
Instant Technologies i Screen	3	*Positive	2	66.67%
Integrated E-Z Split Key Cup	1	*Positive	1	100.00%
MedTox Profile II	2	*Positive	2	100.00%
Noble Med Split Specimen Cup	1	*Positive	1	100.00%
Beckman (Olympus) AU	1	*Positive	1	100.00%
Redwood Biotech Panel Dip	1	*Positive	1	100.00%
Innovacon Multi-Clin	10	*Positive	9	90.00%
MedTox Profile III ER	1	*Positive	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Positive	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Positive	5	100.00%
ASC One Screen Drug Test Card	2	*Positive	2	100.00%
MedTox Profile V	3	*Positive	3	100.00%
Branan Med Corp QuickTox	4	*Positive	4	100.00%
Jant Accutest Multi Drug Screen	2	*Positive	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	1	*Positive	1	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Positive	3	75.00%
Mindray BS-200	1	*Positive	1	100.00%
Alere iCassette	27	*Positive	26	96.30%
Roche Hitachi 717	1	*Positive	1	100.00%
US Diagnostics UScreen Cup	2	*Positive	2	100.00%
Abbott Architect c4000	1	*Positive	1	100.00%



Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Positive	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Positive	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Positive	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Positive	1	100.00%
Alere iCup Drugs of Abuse	9	Negative	1	11.11%
Instant Technologies i Screen	3	Negative	1	33.33%
Innovacon Multi-Clin	10	Negative	1	10.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	Negative	1	25.00%
Alere iCassette	27	Negative	1	3.70%
UDS-4		* Denotes Correct Response		
AG: All Methods	114			
		*Negative	113	99.12%
		Positive	1	0.88%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand	2	*Negative	2	100.00%
E-Screen	16	*Negative	16	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	9	*Negative	9	100.00%
Instant Technologies i Screen	3	*Negative	3	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Redwood Biotech Panel Dip	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	2	*Negative	2	100.00%
MedTox Profile V	3	*Negative	3	100.00%
Branan Med Corp QuickTox	4	*Negative	4	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	1	*Negative	1	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	1	*Negative	1	100.00%
Alere iCassette	27	*Negative	27	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	2	*Negative	1	50.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	2	Positive	1	50.00%
		Propoxyphene		
UDS-3		* Denotes Correct Response		
AG: All Methods	55			
		*Negative	55	100.00%
E-Screen	4	*Negative	4	100.00%
Alere iCup Drugs of Abuse	5	*Negative	5	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	7	*Negative	7	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	1	*Negative	1	100.00%
MedTox Profile V	3	*Negative	3	100.00%
Branan Med Corp QuickTox	4	*Negative	4	100.00%
Jant Accutest Multi Drug Screen	1	*Negative	1	100.00%
Alere iCassette	26	*Negative	26	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	55			
		*Negative	55	100.00%
E-Screen	4	*Negative	4	100.00%
Alere iCup Drugs of Abuse	5	*Negative	5	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	7	*Negative	7	100.00%



Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	1	*Negative	1	100.00%
MedTox Profile V	3	*Negative	3	100.00%
Branan Med Corp QuickTox	4	*Negative	4	100.00%
Jant Accutest Multi Drug Screen	1	*Negative	1	100.00%
Alere iCassette	26	*Negative	26	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
		Tricyclic Antidepressants		
UDS-3		* Denotes Correct Response		
AG: All Methods	71			
		*Negative	70	98.59%
		Positive	1	1.41%
AlereTriage	3	*Negative	3	100.00%
Alere iCup Drugs of Abuse	9	*Negative	9	100.00%
Instant Technologies i Screen	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Innovacon Multi-Clin	9	*Negative	9	100.00%
CLIAwaived Inc Rapid Drug Cup	3	*Negative	3	100.00%
MedTox Profile V	3	*Negative	3	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	1	*Negative	1	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
Alere iCassette	26	*Negative	25	96.15%
US Diagnostics UScreen Cup	1	*Negative	1	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	2	*Negative	2	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
Alere iCassette	26	Positive	1	3.85%
UDS-4		* Denotes Correct Response		
AG: All Methods	71			
		*Negative	71	100.00%
AlereTriage	3	*Negative	3	100.00%
Alere iCup Drugs of Abuse	9	*Negative	9	100.00%
Instant Technologies i Screen	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Innovacon Multi-Clin	9	*Negative	9	100.00%
CLIAwaived Inc Rapid Drug Cup	3	*Negative	3	100.00%
MedTox Profile V	3	*Negative	3	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	1	*Negative	1	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
Alere iCassette	26	*Negative	26	100.00%
US Diagnostics UScreen Cup	1	*Negative	1	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	2	*Negative	2	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
		Buprenorphine		
UDS-3		* Denotes Correct Response		
AG: All Methods	22			
		*Negative	22	100.00%
Alere iCup Drugs of Abuse	5	*Negative	5	100.00%
Innovacon Multi-Clin	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	2	*Negative	2	100.00%
ASC One Screen Drug Test Card	2	*Negative	2	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	1	*Negative	1	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	22			
		*Negative	22	100.00%
Alere iCup Drugs of Abuse	5	*Negative	5	100.00%



Innovacon Multi-Clin	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	2	*Negative	2	100.00%
ASC One Screen Drug Test Card	2	*Negative	2	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	1	*Negative	1	100.00%
Mindrav BS-200	2	*Negative	2	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
756 - Viral Antibodies	Labs	Response	Number	Percent
VA-2		HSV Ab (Herpes Simplex Virus)		
AG: All Methods	2	* Denotes Correct Response		
		*Negative	2	100.00%
Biokit HSV-II	2	*Negative	2	100.00%
VA-2		Rubella Antibody		
AG: All Methods	1	* Denotes Correct Response		
		*Positive	1	100.00%
bioMerieux Mini Vidas	1	*Positive	1	100.00%
VA-2		Paramyxovirus (mumps) IgG Ab		
AG: All Methods	1	* Denotes Correct Response		
		*Positive	1	100.00%
bioMerieux Mini Vidas	1	*Positive	1	100.00%
770 - Viral Antigen Detection;798 - Viral Antigen Detection (ROT);797 - Viral Antigen Detection II	Labs	Response	Number	Percent
VAD-6		Rotavirus - Ag Detect		
AG: All Methods	9	* Denotes Correct Response		
		*Negative	9	100.00%
Meridian Immunocard Stat! Rotavirus	5	*Negative	5	100.00%
SAS Rota-Test	4	*Negative	4	100.00%
VAD-7		* Denotes Correct Response		
AG: All Methods	9			
		*Negative	9	100.00%
Meridian Immunocard Stat! Rotavirus	5	*Negative	5	100.00%
SAS Rota-Test	4	*Negative	4	100.00%
VAD-8		* Denotes Correct Response		
AG: All Methods	9			
		*Positive	9	100.00%
Meridian Immunocard Stat! Rotavirus	5	*Positive	5	100.00%
SAS Rota-Test	4	*Positive	4	100.00%
VAD-9		* Denotes Correct Response		
AG: All Methods	9			
		*Negative	9	100.00%
Meridian Immunocard Stat! Rotavirus	5	*Negative	5	100.00%
SAS Rota-Test	4	*Negative	4	100.00%
VAD-10		* Denotes Correct Response		
AG: All Methods	9			
		*Negative	9	100.00%
Meridian Immunocard Stat! Rotavirus	5	*Negative	5	100.00%
SAS Rota-Test	4	*Negative	4	100.00%
VAD-6		Influenza A - Ag Detect		
AG: All Methods	53	* Denotes Correct Response		
		*Positive	52	98.11%
		Negative	1	1.89%
Alere BinaxNOW Influenza A+B Card	5	*Positive	5	100.00%
Sekisui Diagnostics OSOM Influenza A & B	15	*Positive	15	100.00%
Quidel QuickVue A+B	20	*Positive	20	100.00%
Quidel Sofia Influenza A+B	6	*Positive	6	100.00%
Alere Influenza A & B	3	*Positive	2	66.67%
BD Veritor Flu A+B	2	*Positive	2	100.00%
Cepheid GeneXpert - Xpert Flu	2	*Positive	2	100.00%
Alere Influenza A & B	3	Negative	1	33.33%
VAD-7		* Denotes Correct Response		



AG: All Methods	53			
		*Positive	53	100.00%
Alere BinaxNOW Influenza A+B Card	5	*Positive	5	100.00%
Sekisui Diagnostics OSOM Influenza A & B	15	*Positive	15	100.00%
Quidel QuickVue A+B	20	*Positive	20	100.00%
Quidel Sofia Influenza A+B	6	*Positive	6	100.00%
Alere Influenza A & B	3	*Positive	3	100.00%
BD Veritor Flu A+B	2	*Positive	2	100.00%
Cepheid GeneXpert - Xpert Flu	2	*Positive	2	100.00%
VAD-8		* Denotes Correct Response		
AG: All Methods	53			
		*Negative	50	94.34%
		Positive	3	5.66%
Alere BinaxNOW Influenza A+B Card	5	*Negative	5	100.00%
Sekisui Diagnostics OSOM Influenza A & B	15	*Negative	15	100.00%
Quidel QuickVue A+B	20	*Negative	17	85.00%
Quidel Sofia Influenza A+B	6	*Negative	6	100.00%
Alere Influenza A & B	3	*Negative	3	100.00%
BD Veritor Flu A+B	2	*Negative	2	100.00%
Cepheid GeneXpert - Xpert Flu	2	*Negative	2	100.00%
Quidel QuickVue A+B	20	Positive	3	15.00%
VAD-9		* Denotes Correct Response		
AG: All Methods	53			
		*Negative	53	100.00%
Alere BinaxNOW Influenza A+B Card	5	*Negative	5	100.00%
Sekisui Diagnostics OSOM Influenza A & B	15	*Negative	15	100.00%
Quidel QuickVue A+B	20	*Negative	20	100.00%
Quidel Sofia Influenza A+B	6	*Negative	6	100.00%
Alere Influenza A & B	3	*Negative	3	100.00%
BD Veritor Flu A+B	2	*Negative	2	100.00%
Cepheid GeneXpert - Xpert Flu	2	*Negative	2	100.00%
VAD-10		* Denotes Correct Response		
AG: All Methods	53			
		*Positive	53	100.00%
Alere BinaxNOW Influenza A+B Card	5	*Positive	5	100.00%
Sekisui Diagnostics OSOM Influenza A & B	15	*Positive	15	100.00%
Quidel QuickVue A+B	20	*Positive	20	100.00%
Quidel Sofia Influenza A+B	6	*Positive	6	100.00%
Alere Influenza A & B	3	*Positive	3	100.00%
BD Veritor Flu A+B	2	*Positive	2	100.00%
Cepheid GeneXpert - Xpert Flu	2	*Positive	2	100.00%
		Influenza A/B - Aq Detect		
VAD-6		* Denotes Correct Response		
AG: All Methods	4			
		*Positive	4	100.00%
Quidel QuickVue	4	*Positive	4	100.00%
VAD-7		* Denotes Correct Response		
AG: All Methods	4			
		*Positive	4	100.00%
Quidel QuickVue	4	*Positive	4	100.00%
VAD-8		* Denotes Correct Response		
AG: All Methods	4			
		*Negative	4	100.00%
Quidel QuickVue	4	*Negative	4	100.00%
VAD-9		* Denotes Correct Response		
AG: All Methods	4			
		*Negative	4	100.00%
Quidel QuickVue	4	*Negative	4	100.00%
VAD-10		* Denotes Correct Response		
AG: All Methods	4			
		*Positive	4	100.00%
Quidel QuickVue	4	*Positive	4	100.00%
		Influenza B - Aq Detect		
VAD-6		* Denotes Correct Response		
AG: All Methods	53			
		*Negative	53	100.00%
Alere BinaxNOW Influenza A+B Card	5	*Negative	5	100.00%
Sekisui Diagnostics OSOM Influenza A & B	15	*Negative	15	100.00%



Quidel QuickVue A+B	20	*Negative	20	100.00%
Quidel Sofia Influenza A+B	6	*Negative	6	100.00%
Alere Influenza A & B	3	*Negative	3	100.00%
BD Veritor Flu A+B	2	*Negative	2	100.00%
Cepheid GeneXpert - Xpert Flu	2	*Negative	2	100.00%
VAD-7		* Denotes Correct Response		
AG: All Methods	53			
		*Negative	53	100.00%
Alere BinaxNOW Influenza A+B Card	5	*Negative	5	100.00%
Sekisui Diagnostics OSOM Influenza A & B	15	*Negative	15	100.00%
Quidel QuickVue A+B	20	*Negative	20	100.00%
Quidel Sofia Influenza A+B	6	*Negative	6	100.00%
Alere Influenza A & B	3	*Negative	3	100.00%
BD Veritor Flu A+B	2	*Negative	2	100.00%
Cepheid GeneXpert - Xpert Flu	2	*Negative	2	100.00%
VAD-8		* Denotes Correct Response		
AG: All Methods	53			
		*Negative	52	98.11%
		Positive	1	1.89%
Alere BinaxNOW Influenza A+B Card	5	*Negative	4	80.00%
Sekisui Diagnostics OSOM Influenza A & B	15	*Negative	15	100.00%
Quidel QuickVue A+B	20	*Negative	20	100.00%
Quidel Sofia Influenza A+B	6	*Negative	6	100.00%
Alere Influenza A & B	3	*Negative	3	100.00%
BD Veritor Flu A+B	2	*Negative	2	100.00%
Cepheid GeneXpert - Xpert Flu	2	*Negative	2	100.00%
Alere BinaxNOW Influenza A+B Card	5	Positive	1	20.00%
VAD-9		* Denotes Correct Response		
AG: All Methods	53			
		*Negative	53	100.00%
Alere BinaxNOW Influenza A+B Card	5	*Negative	5	100.00%
Sekisui Diagnostics OSOM Influenza A & B	15	*Negative	15	100.00%
Quidel QuickVue A+B	20	*Negative	20	100.00%
Quidel Sofia Influenza A+B	6	*Negative	6	100.00%
Alere Influenza A & B	3	*Negative	3	100.00%
BD Veritor Flu A+B	2	*Negative	2	100.00%
Cepheid GeneXpert - Xpert Flu	2	*Negative	2	100.00%
VAD-10		* Denotes Correct Response		
AG: All Methods	53			
		*Negative	53	100.00%
Alere BinaxNOW Influenza A+B Card	5	*Negative	5	100.00%
Sekisui Diagnostics OSOM Influenza A & B	15	*Negative	15	100.00%
Quidel QuickVue A+B	20	*Negative	20	100.00%
Quidel Sofia Influenza A+B	6	*Negative	6	100.00%
Alere Influenza A & B	3	*Negative	3	100.00%
BD Veritor Flu A+B	2	*Negative	2	100.00%
Cepheid GeneXpert - Xpert Flu	2	*Negative	2	100.00%
		RSV - Aq Detect		
VAD-6		* Denotes Correct Response		
AG: All Methods	26			
		*Negative	26	100.00%
BD EZ RSV	1	*Negative	1	100.00%
Alere BinaxNOW RSV	9	*Negative	9	100.00%
Quidel QuickVue RSV	3	*Negative	3	100.00%
Wampole Clearview RSV	4	*Negative	4	100.00%
SAS RSV Alert	1	*Negative	1	100.00%
Quidel QuickVue RSV 10	2	*Negative	2	100.00%
Quidel Sofia RSV	5	*Negative	5	100.00%
Cepheid GeneXpert - Xpert RSV	1	*Negative	1	100.00%
VAD-7		* Denotes Correct Response		
AG: All Methods	26			
		*Positive	26	100.00%
BD EZ RSV	1	*Positive	1	100.00%
Alere BinaxNOW RSV	9	*Positive	9	100.00%
Quidel QuickVue RSV	3	*Positive	3	100.00%
Wampole Clearview RSV	4	*Positive	4	100.00%
SAS RSV Alert	1	*Positive	1	100.00%
Quidel QuickVue RSV 10	2	*Positive	2	100.00%
Quidel Sofia RSV	5	*Positive	5	100.00%



Cepheid GeneXpert - Xpert RSV	1	*Positive	1	100.00%
VAD-8		* Denotes Correct Response		
AG: All Methods	26			
		*Negative	26	100.00%
BD EZ RSV	1	*Negative	1	100.00%
Alere BinaxNOW RSV	9	*Negative	9	100.00%
Quidel QuickVue RSV	3	*Negative	3	100.00%
Wampole Clearview RSV	4	*Negative	4	100.00%
SAS RSV Alert	1	*Negative	1	100.00%
Quidel QuickVue RSV 10	2	*Negative	2	100.00%
Quidel Sofia RSV	5	*Negative	5	100.00%
Cepheid GeneXpert - Xpert RSV	1	*Negative	1	100.00%
VAD-9		* Denotes Correct Response		
AG: All Methods	26			
		*Negative	26	100.00%
BD EZ RSV	1	*Negative	1	100.00%
Alere BinaxNOW RSV	9	*Negative	9	100.00%
Quidel QuickVue RSV	3	*Negative	3	100.00%
Wampole Clearview RSV	4	*Negative	4	100.00%
SAS RSV Alert	1	*Negative	1	100.00%
Quidel QuickVue RSV 10	2	*Negative	2	100.00%
Quidel Sofia RSV	5	*Negative	5	100.00%
Cepheid GeneXpert - Xpert RSV	1	*Negative	1	100.00%
VAD-10		* Denotes Correct Response		
AG: All Methods	26			
		*Negative	26	100.00%
BD EZ RSV	1	*Negative	1	100.00%
Alere BinaxNOW RSV	9	*Negative	9	100.00%
Quidel QuickVue RSV	3	*Negative	3	100.00%
Wampole Clearview RSV	4	*Negative	4	100.00%
SAS RSV Alert	1	*Negative	1	100.00%
Quidel QuickVue RSV 10	2	*Negative	2	100.00%
Quidel Sofia RSV	5	*Negative	5	100.00%
Cepheid GeneXpert - Xpert RSV	1	*Negative	1	100.00%
Adenovirus				
VAD-6		* Denotes Correct Response		
AG: All Methods	15			
		*Positive	14	93.33%
		Negative	1	6.67%
SAS Adeno Test	11	*Positive	11	100.00%
RPS AdenoPlus	4	*Positive	3	75.00%
RPS AdenoPlus		Negative	1	25.00%
VAD-7		* Denotes Correct Response		
AG: All Methods	15			
		*Negative	15	100.00%
SAS Adeno Test	11	*Negative	11	100.00%
RPS AdenoPlus	4	*Negative	4	100.00%
VAD-8		* Denotes Correct Response		
AG: All Methods	15			
		*Negative	15	100.00%
SAS Adeno Test	11	*Negative	11	100.00%
RPS AdenoPlus	4	*Negative	4	100.00%
VAD-9		* Denotes Correct Response		
AG: All Methods	15			
		*Negative	14	93.33%
		Positive	1	6.67%
SAS Adeno Test	11	*Negative	10	90.91%
RPS AdenoPlus	4	*Negative	4	100.00%
SAS Adeno Test	11	Positive	1	9.09%
VAD-10		* Denotes Correct Response		
AG: All Methods	15			
		*Positive	15	100.00%
SAS Adeno Test	11	*Positive	11	100.00%
RPS AdenoPlus	4	*Positive	4	100.00%
932 - Hepatitis C Antibody (Waived);767 - Hepatitis Markers;750 - Viral Markers	Labs	Response	Number	Percent
		HCV (Waived)		
VM-6		* Denotes Correct Response		



AG: All Methods	2			
		*Negative	2	100.00%
OraSure OraQuick HCV Rapid Antibody Test	2	*Negative	2	100.00%
VM-7		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
OraSure OraQuick HCV Rapid Antibody Test	2	*Positive	2	100.00%
VM-6		HBsAg		
		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect i1000 SR	1	*Negative	1	100.00%
VM-7		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect i1000 SR	1	*Negative	1	100.00%
VM-6		HCV		
		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Abbott Architect i1000 SR	1	*Negative	1	100.00%
VM-7		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Abbott Architect Ci4100	1	*Positive	1	100.00%
Abbott Architect i1000 SR	1	*Positive	1	100.00%
767 - Hepatitis Markers;750 - Viral Markers	Labs	Response	Number	Percent
		HCV		
VM-6		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Abbott Architect i1000 SR	1	*Negative	1	100.00%
VM-7		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Abbott Architect Ci4100	1	*Positive	1	100.00%
Abbott Architect i1000 SR	1	*Positive	1	100.00%
VM-8		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
VM-9		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
VM-10		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Abbott Architect Ci4100	1	*Positive	1	100.00%
VM-6		HAV		
		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect i1000 SR	1	*Negative	1	100.00%
VM-7		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect i1000 SR	1	*Negative	1	100.00%
VM-6		Anti-HBc (International)		
		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
VM-7		* Denotes Correct Response		
AG: All Methods	3			



		*Negative	2	66.67%
		Positive	1	33.33%
Intl OCD Vitros ECi/ECiQ	3	*Negative	2	66.67%
Intl OCD Vitros ECi/ECiQ		Positive	1	33.33%
VM-8		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Negative	3	100.00%
VM-9		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Intl OCD Vitros ECi/ECiQ	2	*Positive	2	100.00%
VM-10		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Negative	3	100.00%
		HBsAb (International)		
VM-6		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Negative	3	100.00%
VM-7		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Positive	3	100.00%
VM-8		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Negative	3	100.00%
VM-9		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Negative	3	100.00%
VM-10		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Positive	3	100.00%
		HBsAb		
VM-6		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
VM-7		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Abbott Architect Ci4100	1	*Positive	1	100.00%
VM-8		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
VM-9		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
VM-10		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Abbott Architect Ci4100	1	*Positive	1	100.00%
		HBsAg (International)		
VM-6		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Intl OCD Vitros ECi/ECiQ	2	*Negative	2	100.00%
VM-7		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Negative	3	100.00%



VM-8		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
VM-9		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Intl OCD Vitros ECI/ECIQ	2	*Positive	2	100.00%
VM-10		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Intl OCD Vitros ECI/ECIQ	2	*Negative	2	100.00%
HCV (International)				
VM-6		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
VM-7		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Intl OCD Vitros ECI/ECIQ	2	*Positive	2	100.00%
VM-8		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
VM-9		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
VM-10		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Intl OCD Vitros ECI/ECIQ	2	*Positive	2	100.00%
HBsAg				
VM-6		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect i1000 SR	1	*Negative	1	100.00%
VM-7		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect i1000 SR	1	*Negative	1	100.00%
771 - Affirm VP				
Labs	Response	Number	Percent	
Candida sp. - Non Culture				
VP-6		* Denotes Correct Response		
AG: All Methods	37			
		*Negative	36	97.30%
		Positive	1	2.70%
Affirm VP	37	*Negative	36	97.30%
Affirm VP		Positive	1	2.70%
VP-7		* Denotes Correct Response		
AG: All Methods	37			
		*Negative	35	94.59%
		Positive	2	5.41%
Affirm VP	37	*Negative	35	94.59%
Affirm VP		Positive	2	5.41%
VP-8		* Denotes Correct Response		
AG: All Methods	37			
		*Negative	37	100.00%
Affirm VP	37	*Negative	37	100.00%
VP-9		* Denotes Correct Response		
AG: All Methods	37			
		*Positive	35	94.59%
		Negative	2	5.41%
Affirm VP	37	*Positive	35	94.59%
Affirm VP		Negative	2	5.41%
VP-10		* Denotes Correct Response		



AG: All Methods	37			
		*Positive	35	94.59%
		Negative	2	5.41%
Affirm VP	37	*Positive	35	94.59%
Affirm VP		Negative	2	5.41%
VP-6		G. vaginalis - Non Culture Tech		
		* Denotes Correct Response		
AG: All Methods	37			
		*Negative	37	100.00%
Affirm VP	37	*Negative	37	100.00%
VP-7		* Denotes Correct Response		
AG: All Methods	37			
		*Positive	35	94.59%
		Negative	2	5.41%
Affirm VP	37	*Positive	35	94.59%
Affirm VP		Negative	2	5.41%
VP-8		* Denotes Correct Response		
AG: All Methods	37			
		*Negative	36	97.30%
		Positive	1	2.70%
Affirm VP	37	*Negative	36	97.30%
Affirm VP		Positive	1	2.70%
VP-9		* Denotes Correct Response		
AG: All Methods	37			
		*Negative	34	91.89%
		Positive	3	8.11%
Affirm VP	37	*Negative	34	91.89%
Affirm VP		Positive	3	8.11%
VP-10		* Denotes Correct Response		
AG: All Methods	37			
		*Negative	36	97.30%
		Positive	1	2.70%
Affirm VP	37	*Negative	36	97.30%
Affirm VP		Positive	1	2.70%
VP-6		T. vaginalis - Non Culture		
		* Denotes Correct Response		
AG: All Methods	37			
		*Positive	36	97.30%
		Negative	1	2.70%
Affirm VP	37	*Positive	36	97.30%
Affirm VP		Negative	1	2.70%
VP-7		* Denotes Correct Response		
AG: All Methods	37			
		*Negative	36	97.30%
		Positive	1	2.70%
Affirm VP	37	*Negative	36	97.30%
Affirm VP		Positive	1	2.70%
VP-8		* Denotes Correct Response		
AG: All Methods	37			
		*Negative	37	100.00%
Affirm VP	37	*Negative	37	100.00%
VP-9		* Denotes Correct Response		
AG: All Methods	37			
		*Positive	35	94.59%
		Negative	2	5.41%
Affirm VP	37	*Positive	35	94.59%
Affirm VP		Negative	2	5.41%
VP-10		* Denotes Correct Response		
AG: All Methods	37			
		*Negative	36	97.30%
		Positive	1	2.70%
Affirm VP	37	*Negative	36	97.30%
Affirm VP		Positive	1	2.70%