



666 - Afinion Glycohemoglobin	AFF-2																								
Glycohemoglobin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Afinion Glycohemoglobin	143	8.05	0.27	7.20-8.80	3.38																				
747 - Auto Diff 7 (AcT 5 Diff & Horiba 5-Part Diff)	BCX-6					BCX-7					BCX-8					BCX-9					BCX-10				
White Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman Coulter AcT 5 diff	24	2.70	0.10	2.30-3.10	3.70	24	19.49	0.68	16.60-22.40	3.49	24	6.87	0.25	5.90-7.90	3.62	24	27.15	0.95	23.10-31.30	3.49	24	9.77	0.33	8.30-11.30	3.37
AG: All Group	25	2.70	0.10	2.30-3.10	3.70	25	19.48	0.66	16.60-22.40	3.38	25	6.86	0.25	5.90-7.90	3.62	25	27.16	0.93	23.10-31.30	3.42	25	9.77	0.32	8.30-11.30	3.27
Lymphocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman Coulter AcT 5 diff	23	20.39	3.04	11.30-29.50	14.90	23	14.49	1.77	9.20-19.80	12.21	23	34.66	3.65	23.80-45.70	10.52	24	42.88	2.89	34.20-51.60	6.74	22	29.25	3.44	19.00-39.60	11.74
PG: Horiba Pentra 60C+	5	26.70	0.00	26.70-26.70	0.00	5	18.00	0.00	18.00-18.00	0.00	5	38.20	0.00	38.20-38.20	0.00	5	46.50	0.00	46.50-46.50	0.00	5	37.10	0.00	37.10-37.10	0.00
SG: International Horiba Instruments	5	26.85	1.96	21.00-32.80	7.29	5	22.23	5.15	6.70-37.60	23.20	5	39.92	5.19	24.30-55.50	13.01	5	50.10	11.30	16.20-84.00	22.53	5	41.17	6.14	22.80-59.60	14.90
Monocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman Coulter AcT 5 diff	24	1.14	0.54	0.00-2.75	47.19	24	2.22	1.16	0.00-5.69	52.16	24	0.99	0.37	0.00-2.10	37.27	24	3.30	1.06	0.11-6.49	32.24	24	1.48	0.85	0.00-4.04	57.64
PG: Horiba Pentra 60C+	5	0.50	0.00	0.50-0.50	0.00	5	1.20	0.00	1.20-1.20	0.00	5	0.30	0.00	0.30-0.30	0.00	5	2.30	0.00	2.30-2.30	0.00	5	0.40	0.00	0.40-0.40	0.00
SG: International Horiba Instruments	5	0.90	0.27	0.09-1.71	30.11	5	1.48	0.39	0.32-2.64	26.08	5	2.33	2.97	0.00-11.23	127.34	5	2.18	0.40	1.00-3.37	18.12	5	0.50	0.14	0.08-0.92	28.20
Granulocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman Coulter AcT 5 diff	23	57.94	3.66	46.90-68.90	6.32	23	62.36	2.64	54.50-70.30	4.23	23	50.73	4.13	38.30-63.10	8.15	23	41.00	3.01	32.00-50.00	7.34	23	57.61	4.31	44.70-70.50	7.48
PG: Horiba Pentra 60C+	5	60.30	0.00	60.30-60.30	0.00	5	74.20	0.00	74.20-74.20	0.00	5	50.00	0.00	50.00-50.00	0.00	5	44.80	0.00	44.80-44.80	0.00	5	58.10	0.00	58.10-58.10	0.00
SG: International Horiba Instruments	5	61.90	2.94	53.10-70.70	4.73	5	70.65	4.59	56.90-84.50	6.49	5	48.15	4.49	34.70-61.70	9.32	5	41.05	7.62	18.10-63.90	18.59	5	52.95	4.94	38.20-67.80	9.32
Eosinophil	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman Coulter AcT 5 diff	22	8.27	2.00	2.30-14.30	24.10	23	5.27	0.69	3.20-7.40	13.02	23	10.36	1.55	5.80-15.10	14.90	24	6.17	0.78	3.90-8.50	12.58	23	5.72	1.34	1.70-9.70	23.51
PG: Horiba Pentra 60C+	5	11.90	0.00	11.90-11.90	0.00	5	5.80	0.00	5.80-5.80	0.00	5	11.30	0.00	11.30-11.30	0.00	5	6.10	0.00	6.10-6.10	0.00	5	4.20	0.00	4.20-4.20	0.00
SG: International Horiba Instruments	5	9.80	1.69	4.70-14.90	17.24	5	4.88	1.02	1.80-8.00	20.82	5	9.35	2.33	2.40-16.40	24.79	5	6.38	4.35	0.00-19.40	67.97	5	5.18	2.10	0.00-11.50	40.38
Basophil	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman Coulter AcT 5 diff	23	11.40	0.44	10.07-12.72	3.88	23	15.20	0.41	13.96-16.44	2.71	23	3.23	0.18	2.69-3.77	5.56	23	6.60	0.28	5.76-7.45	4.25	23	4.37	0.20	3.79-4.96	4.48
PG: Horiba Pentra 60C+	5	0.60	0.00	0.60-0.60	0.00	5	0.80	0.00	0.80-0.80	0.00	5	0.20	0.00	0.20-0.20	0.00	5	0.30	0.00	0.30-0.30	0.00	5	0.20	0.00	0.20-0.20	0.00
SG: International Horiba Instruments	5	0.55	0.06	0.38-0.72	10.49	5	0.78	0.05	0.63-0.93	6.45	5	0.25	0.10	0.00-0.55	40.00	5	0.30	0.08	0.06-0.55	27.23	5	0.20	0.00	0.20-0.20	0.00
Red Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman Coulter AcT 5 diff	24	2.37	0.04	2.23-2.51	1.69	24	4.98	0.06	4.68-5.28	1.16	24	3.94	0.04	3.70-4.18	1.12	24	4.94	0.06	4.64-5.24	1.11	24	6.22	0.09	5.85-6.59	1.41
AG: All Group	25	2.37	0.04	2.23-2.51	1.65	25	4.98	0.06	4.68-5.28	1.14	25	3.94	0.04	3.70-4.18	1.09	25	4.95	0.06	4.65-5.25	1.11	25	6.22	0.09	5.85-6.59	1.38
Hemoglobin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman Coulter AcT 5 diff	24	6.16	0.07	5.80-6.60	0.97	24	16.03	0.14	14.90-17.10	0.88	24	12.65	0.11	11.80-13.60	0.87	24	13.90	0.13	12.90-14.90	0.94	24	17.80	0.20	16.60-19.00	1.12
AG: All Group	25	6.16	0.06	5.80-6.60	0.97	25	16.06	0.18	15.00-17.20	1.12	25	12.66	0.12	11.80-13.60	0.94	25	13.92	0.15	12.90-14.90	1.08	25	17.81	0.20	16.60-19.00	1.12
Hematocrit	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman Coulter AcT 5 diff	24	17.54	0.47	16.40-18.60	2.69	24	46.14	0.70	43.30-48.90	1.52	24	36.05	0.57	33.90-38.30	1.58	24	40.48	0.62	38.10-42.90	1.53	24	51.80	0.87	48.70-54.90	1.68
AG: All Group	25	17.54	0.46	16.40-18.60	2.63	25	46.13	0.69	43.30-48.90	1.50	25	36.06	0.56	33.90-38.30	1.55	25	40.49	0							



PG: Intl OCD Vitros 250,350,400,500,750,950	5	13.00	0.00	13.00-13.00	0.00	5	34.50	0.00	34.50-34.50	0.00	5	66.00	0.00	66.00-66.00	0.00	5	5.40	0.00	5.40-5.40	0.00	5	20.40	0.00	20.40-20.40	0.00
PG: OCD Vitros 5600	5	13.90	0.00	13.90-13.90	0.00	5	39.70	0.00	39.70-39.70	0.00	5	74.00	0.00	74.00-74.00	0.00	5	5.50	0.00	5.50-5.50	0.00	5	23.00	0.00	23.00-23.00	0.00
Myoglobin (Quantitative)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: AlereTriage	31	88.84	13.41	48.60-129.00	15.10	31	216.32	27.20	134.70-297.90	12.58	31	360.35	61.68	175.40-545.40	17.11	31	35.53	6.18	17.00-54.00	17.41	31	139.81	19.68	80.80-198.80	14.08
AG: All Group	34	88.65	13.68	47.70-129.70	15.42	34	219.76	31.95	123.90-315.70	14.54	34	370.48	72.97	151.60-589.40	19.70	34	35.37	6.10	17.10-53.70	17.23	34	140.44	20.76	78.10-202.70	14.79
Troponin I (Quantitative)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat	23	4.51	0.74	2.29-6.73	16.38	23	19.63	3.19	10.06-29.21	16.26	22	36.44	5.37	20.34-52.54	14.73	22	0.31	0.04	0.17-0.44	14.54	22	9.52	1.49	5.04-14.00	15.69
PG: AlereTriage	35	0.47	0.11	0.15-0.78	22.51	36	3.77	0.74	1.56-5.98	19.56	36	9.00	1.53	4.42-13.59	16.98	36	0.05	0.00	0.05-0.05	0.00	35	1.33	0.32	0.38-2.27	23.74
PG: TOSOH AIA	5	14.27	1.14	10.84-17.70	8.02	5	48.91	1.46	44.54-53.28	2.98	5	87.98	0.00	87.98-87.98	0.00	5	1.04	0.00	1.04-1.04	0.00	5	26.53	0.00	26.53-26.53	0.00
PG: Intl Abbott i-Stat	5	4430.00	474.76	3005.71-5854.29	10.72	5	20445.0	2443.01	13116.0-27774.0	11.95															
AG: All Group	61	2.10	2.10	0.00-8.39	99.66	62	10.03	8.24	0.00-34.76	82.17	61	19.65	14.43	0.00-62.93	73.43	61	0.15	0.13	0.00-0.53	86.49	60	4.55	4.24	0.00-17.26	93.01
D-dimer (Quantitative)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: AlereTriage	47	265.55	32.86	167.00-364.20	12.37	47	741.47	72.64	523.00-959.00	9.80	46	1193.70	242.72	465.50-1921.90	20.33	45	104.00	8.09	79.70-128.30	7.79	46	450.17	55.37	284.00-616.00	12.31
PG: Beckman AU	5	115.00	0.00	115.00-115.00	0.00	5	305.00	0.00	305.00-305.00	0.00	5	600.00	0.00	600.00-600.00	0.00	5	70.00	0.00	70.00-70.00	0.00	5	185.00	0.00	185.00-185.00	0.00
PG: Roche H 232	5	180.00	25.98	102.10-257.90	14.44	5	388.33	33.29	288.00-488.00	8.58															
AG: All Group	50	263.42	35.16	157.90-368.90	13.38	47	753.02	69.53	545.00-962.00	9.23	44	1239.09	113.22	899.40-1578.80	9.14	43	102.70	5.06	87.50-117.90	4.95	46	450.17	55.37	284.00-616.00	12.31
680 - Coagulation - Plasma	CG-6					CG-7					CG-8					CG-9					CG-10				
Protime - Plasma	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex CA Series / Dade Innovin	5	11.37	0.72	9.70-13.10	6.32	5	10.63	0.40	9.00-12.20	3.77	5	14.17	0.38	12.10-16.30	2.68	5	32.93	2.48	28.00-37.80	7.54	5	23.07	1.53	19.60-26.60	6.62
PG: Diagnostica Stago STA Satellite / Diag STA Neoplastine CL+	16	14.09	0.45	12.00-16.20	3.19	15	13.72	0.86	11.60-15.80	6.35	16	17.08	1.53	14.50-19.70	8.95	14	40.93	2.67	34.80-47.00	6.55	13	28.56	2.18	24.30-32.90	7.62
SG: Select Diagnostica Stago / Diag STA Neoplastine CL+	18	14.01	0.48	11.90-16.10	3.43	17	13.62	0.86	11.60-15.60	6.32	18	16.99	1.46	14.40-19.60	8.59	16	40.88	2.58	34.80-47.00	6.31	15	28.41	2.10	24.10-32.70	7.39
INR	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex CA Series / Dade Innovin	5	1.10	0.10	0.80-1.40	9.09	5	1.07	0.06	0.90-1.30	5.45	5	1.40	0.10	1.10-1.70	7.14	5	3.20	0.36	2.10-4.30	11.25	5	2.27	0.32	1.30-3.30	13.91
PG: Diagnostica Stago STA Satellite	16	1.12	0.05	1.00-1.30	4.55	15	1.08	0.10	0.80-1.40	9.09	16	1.44	0.19	0.80-2.00	13.57	16	4.47	0.86	1.90-7.10	19.11	13	2.81	0.35	1.80-3.80	12.50
SG: All Method Plasma Coagulation	16	1.12	0.05	1.00-1.30	4.55	15	1.08	0.10	0.80-1.40	9.09	16	1.44	0.19	0.80-2.00	13.57	16	4.47	0.86	1.90-7.10	19.11	13	2.81	0.35	1.80-3.80	12.50
SG: Select Diagnostica Stago	18	1.11	0.06	0.90-1.30	5.45	17	1.08	0.10	0.80-1.40	9.09	18	1.44	0.18	0.90-1.90	12.86	18	4.49	0.82	2.00-7.00	18.22	15	2.81	0.33	1.80-3.80	11.79
612 - MultiChem 5;616 - MultiChem 10;617 - MultiChem 15;618 - MultiChem 20;619 - MultiChem 21+;717 - Waived Chemistry;676 - Endocrinology;719 - Quantitative hCG (Serum)	CH-6N					CH-7N					CH-8N					CH-9N					CH-10N				
HCG - Quant	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman ACCESS	5	933.69	20.47	873.00-996.00	2.19	5	1202.59	39.88	1083.00-1323.00	3.32	5	740.29	50.56	588.00-892.00	6.84	5	242.90	2.70	235.00-251.00	1.11	5	1250.10	100.66	948.00-1552.00	8.06
PG: TOSOH AIA	5	1163.60	41.05	1041.00-1287.00	3.53	5	1579.33	139.14	1162.00-1996.00	8.81	5	968.20	76.04	740.00-1196.00	7.85	5	275.00	14.72	231.00-319.00	5.35	5	1657.73	99.83	1359.00-1957.00	6.02
PG: Roche Cobas e Series	5	932.10	39.71	813.00-1051.00	4.26	5	1269.33	45.65	1132.00-1406.00	3.60	5	762.23	27.28	680.00-844.00	3.58	5	235.13	7.72	212.00-258.00	3.28	5	1367.67	46.98	1227.00-1509.00	3.44
PG: Intl Abbott i-Stat	5	792.33	105.10	477.00-1107.00	13.27	5	1074.33	34.44	971.00-1177.00	3.20	5	662.67	20.84	601.00-725.00	3.14	5	240.00	10.54	209.00-272.00	4.38	5	1025.67	21.50	962.00-1091.00	2.10
AG: All Group	15	932.34																							



PG: Beckman AU	18	4.96	0.08	4.50-5.50	1.60	18	5.95	0.12	5.50-6.50	2.00	18	4.51	0.07	4.00-5.00	1.56	18	3.07	0.05	2.60-3.60	1.61	18	6.19	0.16	5.70-6.70	2.58
PG: Roche Cobas Integra	5	5.10	0.00	4.60-5.60	0.00	5	6.06	0.06	5.60-6.60	0.82	5	4.58	0.04	4.10-5.10	0.87	5	3.08	0.04	2.60-3.60	1.29	5	6.32	0.04	5.80-6.80	0.63
PG: OCD Vitros 250,350,400,500,750,950	8	5.28	0.07	4.80-5.80	1.32	8	6.34	0.09	5.80-6.80	1.43	8	4.71	0.06	4.20-5.20	1.28	8	3.16	0.05	2.70-3.70	1.56	8	6.61	0.13	6.10-7.10	1.82
SG: Select Beckman Coulter Instruments (Chem)	22	4.98	0.09	4.50-5.50	1.80	22	5.98	0.12	5.50-6.50	2.00	22	4.52	0.07	4.00-5.00	1.56	22	3.06	0.05	2.60-3.60	1.61	22	6.21	0.15	5.70-6.70	2.42
SG: Select Siemens Instruments	14	5.05	0.14	4.60-5.60	2.75	14	6.12	0.09	5.60-6.60	1.48	14	4.58	0.07	4.10-5.10	1.52	14	3.01	0.04	2.50-3.50	1.33	14	6.34	0.09	5.80-6.80	1.43
SG: Select Roche Instruments	6	5.12	0.04	4.60-5.60	0.78	6	6.10	0.11	5.60-6.60	1.80	6	4.60	0.06	4.10-5.10	1.30	6	3.10	0.06	2.60-3.60	1.94	6	6.35	0.08	5.90-6.90	1.25
SG: Select OCD Vitros Instruments	11	5.25	0.10	4.80-5.80	1.89	11	6.32	0.09	5.80-6.80	1.43	11	4.70	0.06	4.20-5.20	1.28	11	3.15	0.05	2.60-3.60	1.61	11	6.58	0.13	6.10-7.10	1.97
SG: Abaxis Piccolo Instruments	39	5.06	0.25	4.60-5.60	4.90	39	6.05	0.25	5.60-6.60	4.10	9	4.75	0.23	4.20-5.20	4.89	9	3.09	0.11	2.60-3.60	3.55	9	6.40	0.20	5.90-6.90	3.13
SG: Select Abbott Architect Instruments	8	5.11	0.04	4.60-5.60	0.78	8	6.11	0.06	5.60-6.60	0.98	8	4.61	0.06	4.10-5.10	1.30	8	3.10	0.00	2.60-3.60	0.00	8	6.37	0.05	5.90-6.90	0.78
SG: International OCD Vitros Instruments	14	5.26	0.10	4.80-5.80	1.89	14	6.32	0.10	5.80-6.80	1.59	14	4.71	0.08	4.20-5.20	1.70	14	3.15	0.05	2.70-3.70	1.56	14	6.59	0.13	6.10-7.10	1.97
SG: International Siemens Instruments	14	5.05	0.14	4.60-5.60	2.75	14	6.13	0.07	5.60-6.60	1.15	14	4.59	0.05	4.10-5.10	1.09	14	3.01	0.05	2.50-3.50	1.67	14	6.36	0.07	5.90-6.90	0.94
SG: International Abaxis Piccolo Instruments	40	5.09	0.17	4.60-5.60	3.33	43	6.02	0.27	5.50-6.50	4.50	11	4.62	0.21	4.10-5.10	4.57	11	3.08	0.12	2.60-3.60	3.87	12	6.34	0.21	5.80-6.80	3.33
SG: Select Gen Chem Instruments	62	5.11	0.14	4.60-5.60	2.55	62	6.15	0.19	5.70-6.70	3.06	61	4.60	0.10	4.10-5.10	2.17	60	3.05	0.07	2.60-3.60	2.26	59	6.39	0.19	5.90-6.90	2.97
AG: All Group	130	5.11	0.15	4.60-5.60	2.94	132	6.13	0.20	5.60-6.60	3.28	100	4.62	0.10	4.10-5.10	2.17	100	3.06	0.07	2.60-3.60	2.26	99	6.40	0.18	5.90-6.90	2.81
Chloride (CL-)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	9	107.18	2.13	102.00-112.00	1.96	9	112.16	1.60	106.00-118.00	1.43	9	103.54	2.15	99.00-109.00	2.12	9	95.83	3.08	91.00-101.00	3.23	9	111.87	1.63	106.00-118.00	1.43
PG: Abaxis Piccolo (Waived)	30	106.87	2.49	102.00-112.00	2.34	30	110.77	2.03	105.00-117.00	1.80															
PG: Abbott Architect	6	102.67	0.82	98.00-108.00	0.78	6	107.33	1.21	102.00-112.00	1.12	6	100.17	1.47	95.00-105.00	1.50	6	93.33	1.03	88.00-98.00	1.08	6	108.67	0.82	104.00-114.00	0.73
PG: Alfa Wasserman Alera/Axcel	24	104.54	1.54	100.00-110.00	1.43	24	111.13	1.81	105.00-117.00	1.62	24	101.61	2.58	97.00-107.00	2.55	23	91.53	1.74	87.00-97.00	1.85	24	111.60	3.94	106.00-118.00	3.48
PG: Clinical Data Envoy 500	5	106.86	1.75	102.00-112.00	1.59	5	114.10	2.56	108.00-120.00	2.28	5	103.02	1.60	98.00-108.00	1.55	6	94.28	4.36	89.00-99.00	4.68	5	115.18	2.22	109.00-121.00	1.91
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	13	103.23	1.30	98.00-108.00	1.26	13	110.08	1.44	105.00-116.00	1.27	13	100.08	1.19	95.00-105.00	1.20	13	89.69	1.03	86.00-95.00	1.11	13	111.46	1.20	105.00-117.00	1.08
PG: Beckman AU	16	101.25	1.88	96.00-106.00	1.88	17	106.53	1.88	102.00-112.00	1.78	17	98.88	1.73	94.00-104.00	1.72	17	91.35	1.46	86.00-96.00	1.65	17	107.53	2.10	103.00-113.00	1.94
PG: Roche Cobas Integra	5	103.12	1.13	98.00-108.00	1.07	5	108.44	1.13	103.00-113.00	1.02	5	100.18	0.84	95.00-105.00	0.80	5	91.88	0.76	87.00-97.00	0.87	5	110.60	1.52	105.00-117.00	1.35
PG: OCD Vitros 250,350,400,500,750,950	8	102.25	1.39	97.00-107.00	1.37	8	108.50	1.41	104.00-114.00	1.28	8	99.75	1.67	95.00-105.00	1.70	8	92.00	1.07	87.00-97.00	1.20	8	109.88	2.10	105.00-116.00	1.91
SG: Select Beckman Coulter Instruments (Chem)	20	101.50	1.82	97.00-107.00	1.76	21	106.57	1.78	102.00-112.00	1.68	21	99.14	1.68	94.00-104.00	1.72	21	91.62	1.47	87.00-97.00	1.63	21	107.62	1.96	103.00-113.00	1.85
SG: Select Siemens Instruments	14	103.36	1.34	98.00-108.00	1.26	14	110.14	1.41	105.00-116.00	1.27	14	100.36	1.55	95.00-105.00	1.50	14	90.00	1.52	86.00-95.00	1.67	14	111.36	1.22	105.00-117.00	1.08
SG: Select Roche Instruments	6	102.93	1.11	98.00-108.00	1.07	6	108.20	1.17	103.00-113.00	1.11	6	99.98	0.90	95.00-105.00	0.90	6	91.73	0.77	87.00-97.00	0.87	6	110.17	1.72	105.00-116.00	1.55
SG: Select OCD Vitros Instruments	11	102.09	1.22	97.00-107.00	1.18	11	108.09	1.38	103.00-113.00	1.30	11	99.73	1.42	95.00-105.00	1.40	11	91.82	0.98	87.00-97.00	1.09	11	109.45	1.92	104.00-114.00	1.74
SG: Abaxis Piccolo Instruments	39	106.94	2.39	102.00-112.00	2.24	39	111.09	2.01	105.00-117.00	1.80	9	103.54	2.15	99.00-109.00	2.12	9	95.83	3.08	91.00-101.00	3.23	9	111.87	1.63	106.00-118.00	1.43
SG: Select Abbott Architect Instruments	8	102.88	0.83	98.00-108.00	0.78	8	107.50	1.07	103.00-113.00	1.02	8	100.38	1.30	95.00-105.00	1.30	8	93.50	0.93	89.00-99.00	0.96	8	108.75	0.71	104.00-114.00	0.64
SG: International OCD Vitros Instruments	14	101.93	1.14	97.00-107.00	1.08	14	108.00	1.24	103.00-113.00	1.11	14	99.57	1.34	95.00-105.00	1.30	14	91.71	0.91	87.00-97.00	0.98	14	109.43	1.70	104.00-114.00	1.56
SG: International Siemens Instruments	14	103.43	1.45	98.00-108.00	1.46	14	110.29	1.59	105.00-116.00	1.45	14	100.21	1.25	95.00-105.00	1.30	14	89.93	1.33	86.00-95.00	1.44					
SG: International Abaxis Piccolo Instruments	43	106.73	2.41	102.00-112.00	2.24	43	111.01	1.97	105.00-117.00	1.80	13	103.61	1.96	99.00-109.00	1.92	13	95.65	2.61	91.00-101.00	2.71	13	111.29	1.94	105.00-117.00	1.71
SG: Select Gen Chem Instruments	58	103.52	2.37	99.00-109.00	2.31	59	109.35	3.20	104.00-114.00	2.94	59	100.72	2.41	96.00-106.00	2.38	59	92.11	2.15	87.00-97.00	2.28	59	110.21	3.82	105.00-116.00	3.45
AG: All Group	128	104.39	2.75	99.00-109.00	2.69	129	109.80	2.69	105.00-116.00	2.45	99	100.77	2.30	96.00-106.00	2.28	99	92.09	2.42	87.00-97.00	2.61	99	110.43	3.17	105.00-116.00	2.91
Free T4	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Architect	6	3.40	0.24	2.67-4.13	7.11	6	4.19	0.37	3.09-5.29	8.74	6	2.99	0.17	2.46-3.51	5.83	6	1.35	0.10	1.07-1.64	6.99	6	4.39	0.27	3.58-5.20	6.17
PG: Beckman ACCESS	15	3.11	0.34	2.09-4.14	10.98	15	3.58	0.36	2.50-4.66	10.06	15	2.82	0.28	1.98-3.66	9.93	15	1.50	0.18	0.96-2.05	12.15	15	3.70	0.32	2.74-4.66	8.63
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	9	5.38	0.65	3.42-7.33	12.11	9	6.62	1.22	2.96-10.28	18.43	9	4.62	0.45	3.28-5.96	9.69	9	1.85	0.09	1.59-2.11	4.74	9	6.82	1.31	2.88-10.77	19.27
PG: Qualigen FastPak Thyroid	7	5.77	0.41	4.55-6.99	7.05	7	6.00	0.00	6.00-6.00	0.00	7	5.07	0.45	3.71-6.43	8.94	7	2.13	0.19	1.56-2.70	8.88	7	6.00	0.00	6.00-6.00	0.00
PG: TOSOH AIA	9	5.09	0.54	3.47-6.71	10.62	9	6.19	0.65	4.24-8.15	10.53	9	4.46	0.43	3.16-5.76	9.73	9	2.00	0.20	1.40-2.61	10.10	9	6.40	0.68	4.35-8.45	10.68
SG: Select Beckman Coulter Instruments (Chem)	20	3.19	0.35	2.13-4.25	11.05	20	3.68	0.38	2.53-4.83	10.42	20	2.89	0.31	1.97-3.81	10.63	20	1.51	0.18	0.98-2.05	11.74	20	3.81	0.35	2.76-4.85	9.16
SG: Select Siemens Instruments	12	5.10	0.80	2.68-7.50	15.77	12	6.23	1.28	2.37-10.08	20.63	12	4.36	0.63	2.47-6.25	14.47	12	1.85	0.19	1.28-2.40	10.11	12	6.45	1.34	2.43-10.47	20.78
SG: Select Roche Instruments	5	4.12	0.28	3.28-4.96	6.77	5	5.40	0.44	4.07-6.72	8.17	5	3.59	0.25	2.85-4.34	6.92	5	1.70	0.13	1.32-2.08	7.42	5	5.84	0.53	4.25-7.44	9.10
PG: Roche Cobas e Series	5	4.12	0.28	3.28-4.96	6.77	5	5.40	0.44	4.07-6.72	8.17	5	3.59	0.25	2.85-4.34	6.92	5	1.70	0.13	1.32-2.08	7.42	5	5.84	0.53	4.25-7.44	9.10
SG: Select Abbott Architect Instruments	12	3.52	0.22	2.86-4.17	6.21	12	4.28	0.39	3.11-5.46	9.11	12	3.03	0.15	2.59-3.47	4.82	12	1.35	0.07	1.14-1.56	5.21	12	4.54	0.35	3.49-5.58	7.67
PG: Intl OCD Vitros ECi/ECiQ	5	6.99	0.00	6.99-6.99	0.00	5	6.99	0.00	6.99-6.99	0.00	5	6.99	0.00	6.99-6.99	0.00	5	4.94	0.19	4.36-5						



AG: All Group	130	24.35	2.30	17.00-31.00	9.58	130	27.86	2.64	20.00-36.00	9.29	100	22.08	2.73	14.00-30.00	12.27	99	15.73	2.09	10.00-22.00	13.13	100	28.18	3.25	18.00-38.00	11.79
Calcium	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	9	10.56	0.31	9.60-11.60	2.92	9	11.64	0.24	10.60-12.60	2.07	9	9.86	0.11	8.90-10.90	1.11	9	8.11	0.14	7.10-9.10	1.85	9	11.92	0.19	10.90-12.90	1.60
PG: Abaxis Piccolo (Waived)	31	10.58	0.25	9.60-11.60	2.36	31	11.65	0.27	10.60-12.60	2.33															
PG: Abbott Architect	6	10.43	0.17	9.40-11.40	1.73	6	11.73	0.22	10.70-12.70	1.88	6	9.77	0.19	8.80-10.80	1.94	6	7.82	0.15	6.80-8.80	1.92	6	11.95	0.15	11.00-13.00	1.25
PG: Alfa Wasserman Alera/Axcel	25	10.49	0.26	9.50-11.50	2.48	25	11.80	0.24	10.80-12.80	2.03	25	10.15	0.43	9.20-11.20	4.22	24	8.19	0.17	7.20-9.20	2.07	25	11.86	0.88	10.90-12.90	7.39
PG: Clinical Data Envoy 500	6	10.01	0.13	9.00-11.00	1.20	6	11.37	0.32	10.40-12.40	2.81	6	9.34	0.14	8.30-10.30	1.51	6	7.45	0.10	6.40-8.40	1.35	6	11.52	0.17	10.50-12.50	1.48
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	12	10.22	0.30	9.20-11.20	2.94	12	11.58	0.40	10.60-12.60	3.45	12	9.68	0.31	8.70-10.70	3.20	12	7.84	0.27	6.80-8.80	3.46	12	11.81	0.33	10.80-12.80	2.80
PG: Beckman AU	18	10.18	0.27	9.20-11.20	2.65	18	11.40	0.35	10.40-12.40	3.07	18	9.57	0.25	8.60-10.60	2.60	18	7.70	0.20	6.70-8.70	2.60	18	11.69	0.37	10.70-12.70	3.16
PG: Roche Cobas Integra	5	10.36	0.55	9.40-11.40	5.29	5	11.56	0.61	10.60-12.60	5.26	5	9.68	0.53	8.70-10.70	5.46	5	7.84	0.44	6.80-8.80	5.64	5	11.96	0.72	11.00-13.00	6.00
PG: OCD Vitros 250,350,400,500,750,950	8	10.39	0.15	9.40-11.40	1.44	8	11.69	0.24	10.70-12.70	2.05	8	9.79	0.15	8.80-10.80	1.53	8	7.75	0.14	6.80-8.80	1.79	8	12.06	0.25	11.10-13.10	2.07
SG: Select Beckman Coulter Instruments (Chem)	21	10.19	0.26	9.20-11.20	2.55	21	11.41	0.33	10.40-12.40	2.89	21	9.60	0.25	8.60-10.60	2.60	21	7.73	0.20	6.70-8.70	2.60	21	11.71	0.34	10.70-12.70	2.91
SG: Select Siemens Instruments	13	10.22	0.29	9.20-11.20	2.84	13	11.58	0.39	10.60-12.60	3.36	13	9.67	0.30	8.70-10.70	3.09	13	7.82	0.27	6.80-8.80	3.46	13	11.82	0.32	10.80-12.80	2.71
SG: Select Roche Instruments	6	10.43	0.52	9.40-11.40	5.00	6	11.67	0.61	10.70-12.70	5.21	6	9.75	0.50	8.80-10.80	5.10	6	7.88	0.41	6.90-8.90	5.19	6	12.03	0.67	11.00-13.00	5.58
SG: Select OCD Vitros Instruments	11	10.40	0.14	9.40-11.40	1.35	11	11.66	0.21	10.70-12.70	1.79	11	9.80	0.14	8.80-10.80	1.43	11	7.75	0.14	6.80-8.80	1.79	11	12.02	0.25	11.00-13.00	2.08
SG: Abaxis Piccolo Instruments	40	10.58	0.26	9.60-11.60	2.45	40	11.65	0.26	10.60-12.60	2.24	9	9.86	0.11	8.90-10.90	1.11	9	8.11	0.14	7.10-9.10	1.85	9	11.92	0.19	10.90-12.90	1.60
SG: Select Abbott Architect Instruments	8	10.48	0.17	9.50-11.50	1.62	8	11.79	0.21	10.80-12.80	1.78	8	9.83	0.19	8.80-10.80	1.94	8	7.85	0.14	6.90-8.90	1.77	8	12.05	0.23	11.10-13.10	1.90
SG: International OCD Vitros Instruments	14	10.41	0.16	9.40-11.40	1.54	14	11.68	0.21	10.70-12.70	1.79	14	9.81	0.15	8.80-10.80	1.53	14	7.78	0.14	6.80-8.80	1.79	14	12.04	0.24	11.00-13.00	2.00
SG: International Abaxis Piccolo Instruments	44	10.57	0.26	9.60-11.60	2.45	44	11.64	0.26	10.60-12.60	2.24	13	9.87	0.17	8.90-10.90	1.72	13	8.09	0.16	7.10-9.10	1.98	13	11.92	0.19	10.90-12.90	1.60
SG: Select Gen Chem Instruments	60	10.34	0.29	9.30-11.30	2.82	60	11.62	0.34	10.60-12.60	2.93	60	9.83	0.43	8.80-10.80	4.39	59	7.91	0.31	6.90-8.90	3.92	60	11.80	0.62	10.80-12.80	5.25
AG: All Group	130	10.41	0.31	9.40-11.40	2.98	130	11.63	0.32	10.60-12.60	2.76	99	9.81	0.38	8.80-10.80	3.88	98	7.89	0.29	6.90-8.90	3.67	99	11.85	0.53	10.90-12.90	4.45
T3 Uptake	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	7	0.50	0.06	0.34-0.67	11.16	7	0.48	0.08	0.25-0.71	16.08	7	0.51	0.04	0.40-0.61	6.97	7	0.54	0.07	0.34-0.74	12.10	7	0.47	0.09	0.20-0.75	19.58
Creatinine	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	9	2.82	0.14	2.40-3.20	5.00	9	3.88	0.18	3.30-4.50	4.69	9	2.37	0.19	2.01-2.73	8.10	8	1.11	0.09	0.81-1.41	8.18	9	3.97	0.20	3.37-4.57	5.06
PG: Abaxis Piccolo (Waived)	31	2.87	0.17	2.50-3.30	5.86	31	3.81	0.16	3.20-4.40	4.25															
PG: Abbott Architect	6	3.01	0.06	2.50-3.50	2.00	6	4.09	0.08	3.50-4.70	2.00	6	2.50	0.06	2.13-2.88	2.40	6	0.85	0.02	0.55-1.15	2.71	6	4.36	0.07	3.71-5.01	1.67
PG: Alfa Wasserman Alera/Axcel	25	2.80	0.11	2.40-3.20	4.07	25	3.60	0.17	3.10-4.10	4.72	24	2.43	0.10	2.07-2.79	4.24	24	1.09	0.06	0.79-1.39	5.45	24	3.71	0.31	3.15-4.27	8.44
PG: Clinical Data Envoy 500	6	2.89	0.14	2.50-3.30	4.98	6	3.66	0.12	3.10-4.30	3.24	6	2.41	0.14	2.05-2.77	5.83	6	1.04	0.08	0.74-1.34	8.00	6	3.94	0.12	3.35-4.53	2.92
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	7	2.89	0.14	2.50-3.30	4.84	7	3.85	0.17	3.20-4.40	4.29	7	2.41	0.12	2.05-2.77	5.00	7	1.03	0.07	0.73-1.33	7.00	7	4.09	0.18	3.48-4.70	4.39
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL / IFCC	6	2.87	0.05	2.50-3.30	1.74	6	3.88	0.09	3.30-4.50	2.42	6	2.37	0.05	2.00-2.80	2.11	6	1.04	0.03	0.74-1.34	3.00	6	4.09	0.07	3.50-4.70	1.81
PG: Beckman AU	18	2.87	0.19	2.50-3.30	6.66	18	3.76	0.15	3.20-4.40	3.99	18	2.38	0.09	2.02-2.74	3.87	18	0.95	0.05	0.65-1.25	5.56	18	3.98	0.17	3.38-4.58	4.30
PG: Roche Cobas Integra	5	2.74	0.13	2.30-3.10	4.89	5	3.54	0.21	3.00-4.00	6.00	5	2.28	0.13	1.94-2.62	5.70	5	0.94	0.06	0.64-1.24	5.85	5	3.72	0.22	3.16-4.28	5.95
PG: OCD Vitros 250,350,400,500,750,950	8	2.88	0.04	2.50-3.30	1.42	8	3.77	0.05	3.20-4.40	1.32	8	2.45	0.05	2.08-2.82	2.08	8	0.94	0.04	0.64-1.24	4.57	8	3.98	0.06	3.38-4.58	1.58
SG: Select Beckman Coulter Instruments (Chem)	22	2.88	0.17	2.50-3.30	6.04	22	3.78	0.14	3.20-4.40	3.70	22	2.39	0.09	2.03-2.75	3.56	22	0.96	0.06	0.66-1.26	6.04	22	4.01	0.17	3.41-4.61	4.25
SG: Select Siemens Instruments	8	2.90	0.13	2.50-3.30	4.52	8	3.84	0.15	3.20-4.40	3.98	8	2.41	0.11	2.05-2.77	4.69	8	1.02	0.07	0.72-1.32	7.16	8	4.08	0.17	3.47-4.69	4.04
SG: Select Roche Instruments	6	2.72	0.13	2.30-3.10	4.89	6	3.53	0.19	3.00-4.00	5.43	6	2.27	0.12	1.93-2.61	5.33	6	0.93	0.05	0.63-1.23	5.59	6	3.70	0.20	3.15-4.25	5.41
SG: Select OCD Vitros Instruments	11	2.89	0.05	2.50-3.30	1.73	11	3.79	0.06	3.20-4.40	1.45	11	2.44	0.05	2.07-2.81	2.08	11	0.95	0.04	0.65-1.25	4.63	11	3.98	0.06	3.38-4.58	1.50
SG: Abaxis Piccolo Instruments	40	2.86	0.16	2.50-3.30	5.59	40	3.82	0.17	3.20-4.40	4.47	9	2.37	0.19	2.01-2.73	8.10	8	1.11	0.09	0.81-1.41	8.18	9	3.97	0.20	3.37-4.57	5.06
SG: Select Abbott Architect Instruments	8	3.01	0.05	2.50-3.50	1.73	8	4.10	0.07	3.50-4.70	1.73	8	2.50	0.05	2.13-2.88	2.00	8	0.86	0.03	0.56-1.16	3.33	8	4.36	0.07	3.71-5.01	1.70
SG: International OCD Vitros Instruments	14	2.90	0.05	2.50-3.30	1.79	14	3.79	0.05	3.20-4.40	1.32	14	2.45	0.05	2.08-2.82	2.08	14	0.95	0.04	0.65-1.25	5.56	14	3.98	0.06	3.38-4.58	1.50
SG: International Abaxis Piccolo Instruments	44	2.85	0.16	2.50-3.30	5.61	44	3.83	0.16	3.20-4.40	4.28	13	2.41	0.17	2.05-2.77	7.08	12	1.10	0.08	0.80-1.40	7.36	13	3.94	0.17	3.35-4.53	4.37
SG: Select Gen Chem Instruments	61	2.87	0.15	2.50-3.30	5.23	61	3.74	0.21	3.10-4.30	5.72	60	2.42	0.10	2.06-2.78	4.17	60	1.01	0.10	0.71-1.31	10.00	60	3.93	0.31	3.34-4.52	7.95
AG: All Group	132	2.86	0.14	2.50-3.30	5.03	132	3.77	0.19	3.20-4.40	5.12	100	2.41	0.11	2.05-2.77	4.61	99	1.01	0.10	0.70-1.30	9.41	100	3.95	0.27	3.36-4.54	6.92
Thyroxine (T4)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman ACCESS	6	12.85	1.16	10.30-15.50	8.99	6	16.23	1.49	13.00-19.40	9.20	6	10.88	1.57	8.70-13.10	14.40	6	4.18	0.65	3.20-5.20	15.48	6	16.99	1.98	13.60-20.40	11.65
PG: TOSOH AIA	5	11.90	1.06	9.50-14.30	8.91	5	16.60	0.72	13.30-19.90	4.34	5	9.75	1.28	7.80-11.80	13.06	5	3.70	0.38	2.70-4.70	10.27	5	16.65	1.42	13.40-20.00	8.50
SG: Select Beckman Coulter Instruments (Chem)	7	13.10	1.25	10.50-15.70	9.54	7	16.48	1.52	13.20-19.80	9.21	7	11.13	1.57	8.90-13.30	14.14	7	4.21	0.60	3.20-5.20	14.29	7	17.14	1.85	13.70-20.50	10.82
SG: Select Siemens Instruments	5	12.15	0.19	9.80-14.60	1.56	5	15.30	1.68	12.20-18.40	10.98	5	10.03	0.62	8.00-12.00	6.20	5	3.88								



SG: International Abaxis Piccolo Instruments	43	154.40	1.73	139.00-169.00	1.10	43	194.91	2.35	176.00-215.00	1.23	13	134.90	2.13	122.00-149.00	1.56	13	74.65	1.30	68.00-83.00	1.73	13	204.01	2.92	184.00-224.00	1.42
SG: Select Gen Chem Instruments	61	160.95	6.56	145.00-177.00	4.10	61	203.30	8.27	183.00-223.00	4.09	60	139.63	5.68	126.00-154.00	4.07	60	75.73	3.53	68.00-84.00	4.61	60	211.85	11.47	191.00-233.00	5.42
AG: All Group	133	157.41	6.46	141.00-173.00	4.14	133	199.45	7.85	179.00-219.00	3.97	100	137.37	6.20	123.00-151.00	4.53	100	74.31	4.03	67.00-81.00	5.41	100	209.49	10.11	188.00-230.00	4.83
Triiodothyronine (T3)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: TOSOH AIA	5	4.50	0.45	3.10-5.80	10.00	5	5.14	0.20	4.50-5.70	3.92	5	4.11	0.47	2.70-5.50	11.46	5	2.51	0.11	2.20-2.80	4.40	5	5.38	0.51	3.90-6.90	9.44
SG: Select Beckman Coulter Instruments (Chem)	5	1.75	0.19	1.10-2.30	11.18	5	2.04	0.29	1.10-2.90	14.50	5	1.56	0.12	1.20-2.00	7.50	5	1.10	0.13	0.70-1.50	11.82	5	2.19	0.36	1.10-3.30	16.82
PG: OCD Vitros 5600	5	7.81	0.00	7.80-7.80	0.00	5	7.81	0.00	7.80-7.80	0.00	5	7.81	0.00	7.80-7.80	0.00	5	5.85	0.00	5.90-5.90	0.00	5	7.81	0.00	7.80-7.80	0.00
AG: All Group	7	1.63	0.31	0.70-2.50	19.38	7	1.92	0.38	0.80-3.00	20.00	7	1.48	0.23	0.80-2.20	15.33	7	1.03	0.21	0.40-1.60	21.00	7	2.03	0.47	0.60-3.40	23.50
TSH	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Architect	6	4.26	0.17	3.74-4.78	4.06	6	5.68	0.28	4.85-6.50	4.84	6	3.50	0.11	3.16-3.84	3.20	6	1.12	0.05	0.97-1.27	4.55	6	5.90	0.32	4.95-6.85	5.37
PG: Beckman ACCESS	18	4.53	0.32	3.58-5.48	7.00	17	5.96	0.54	4.33-7.59	9.09	18	3.68	0.23	2.98-4.38	6.36	18	1.19	0.07	0.97-1.41	6.05	17	6.40	0.58	4.65-8.15	9.11
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	12	4.44	0.57	2.72-6.16	12.91	12	5.92	0.78	3.59-8.25	13.11	12	3.64	0.50	2.13-5.15	13.85	12	1.22	0.13	0.84-1.60	10.41	12	6.33	0.90	3.63-9.03	14.22
PG: Qualigen FastPak Thyroid	13	4.72	0.41	3.47-5.96	8.79	13	6.38	0.53	4.79-7.97	8.32	13	3.82	0.43	2.53-5.11	11.26	12	1.19	0.11	0.87-1.51	9.08	13	6.77	0.59	5.00-8.54	8.70
PG: TOSOH AIA	13	5.63	0.32	4.68-6.58	5.63	13	7.16	0.32	6.20-8.12	4.47	13	4.67	0.25	3.91-5.43	5.44	13	1.60	0.09	1.35-1.86	5.31	13	7.69	0.32	6.72-8.66	4.21
SG: Select Beckman Coulter Instruments (Chem)	24	4.53	0.29	3.65-5.42	6.51	23	5.96	0.51	4.42-7.50	8.62	24	3.70	0.22	3.04-4.36	5.97	24	1.18	0.07	0.98-1.38	5.76	23	6.44	0.52	4.87-8.01	8.14
SG: Select Siemens Instruments	16	4.42	0.53	2.84-6.00	11.92	16	5.91	0.73	3.72-8.10	12.34	16	3.64	0.51	2.10-5.18	14.07	16	1.24	0.14	0.83-1.65	11.13	16	6.26	0.97	3.36-9.16	15.46
PG: Roche Cobas e Series	7	4.62	0.13	4.22-5.02	2.88	7	5.93	0.17	5.42-6.44	2.85	7	3.92	0.12	3.56-4.28	3.06	7	1.50	0.12	1.16-1.85	7.67	7	6.22	0.19	5.64-6.80	3.12
SG: Select Abbott Architect Instruments	11	4.25	0.23	3.55-4.95	5.51	12	5.51	0.53	3.92-7.10	9.64	11	3.46	0.18	2.93-3.99	5.09	11	1.11	0.07	0.92-1.31	5.86	12	5.87	0.30	4.98-6.76	5.06
PG: NanoEnTek FrenD System	6	2.09	0.25	1.34-2.84	11.96	6	2.84	0.48	1.39-4.29	17.04	6	1.67	0.24	0.95-2.39	14.37	6	0.60	0.12	0.25-0.95	19.33	6	2.92	0.58	1.19-4.65	19.76
PG: Intl OCD Vitros ECi/ECiQ	5	6.14	0.13	5.75-6.53	2.13	5	8.11	0.23	7.42-8.80	2.85	5	5.12	0.14	4.71-5.53	2.66	5	1.66	0.06	1.47-1.85	3.86	5	8.69	0.11	8.37-9.01	1.24
AG: All Group	89	4.76	0.64	2.85-6.67	13.36	89	6.25	0.84	3.73-8.77	13.46	89	3.91	0.56	2.23-5.59	14.32	88	1.29	0.21	0.67-1.91	16.05	89	6.67	0.91	3.94-9.40	12.63
Urea Nitrogen (BUN)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	9	23.11	0.93	21.00-25.00	3.91	9	30.67	1.12	28.00-34.00	3.55	9	19.78	0.97	18.00-22.00	5.00	9	9.11	0.60	7.00-11.00	6.67	9	32.22	1.30	29.00-35.00	4.06
PG: Abaxis Piccolo (Waived)	31	23.43	0.59	21.00-25.00	2.61	31	30.89	0.94	28.00-34.00	2.90															
PG: Abbott Architect	6	25.50	1.05	24.00-28.00	3.85	6	33.67	1.03	31.00-37.00	2.94	6	21.33	0.52	19.00-23.00	2.38	6	9.83	0.41	8.00-12.00	4.00	6	35.50	1.05	33.00-39.00	2.78
PG: Alfa Wasserman Alera/Axcel	25	24.84	0.85	23.00-27.00	3.60	25	32.76	0.97	30.00-36.00	3.03	24	21.29	0.81	19.00-23.00	3.81	24	10.00	0.59	8.00-12.00	6.00	23	34.61	0.99	32.00-38.00	2.86
PG: Clinical Data Envoy 500	6	25.23	1.06	23.00-27.00	4.40	6	32.78	1.28	30.00-36.00	3.94	6	21.45	1.59	19.00-23.00	7.62	6	9.57	1.22	8.00-12.00	12.00	6	35.15	1.86	32.00-38.00	5.43
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	13	25.69	1.11	24.00-28.00	4.23	12	33.83	1.47	31.00-37.00	4.41	13	22.08	0.95	20.00-24.00	4.55	13	10.38	0.77	8.00-12.00	8.00	13	35.92	1.12	33.00-39.00	3.06
PG: Beckman AU	18	25.83	0.92	24.00-28.00	3.46	18	33.61	1.20	31.00-37.00	3.53	17	21.76	0.83	20.00-24.00	3.64	18	10.17	0.38	8.00-12.00	4.00	18	35.83	1.95	33.00-39.00	5.28
PG: Roche Cobas Integra	5	25.40	0.89	23.00-27.00	3.60	5	33.00	1.23	30.00-36.00	3.64	5	21.40	1.34	19.00-23.00	6.19	5	9.80	0.45	8.00-12.00	4.00	5	35.00	1.87	32.00-38.00	5.43
PG: OCD Vitros 250,350,400,500,750,950	8	20.75	0.46	19.00-23.00	2.38	8	27.13	0.64	25.00-29.00	2.22	8	17.75	0.46	16.00-20.00	2.78	8	8.13	0.35	6.00-10.00	5.00	8	28.63	0.52	26.00-32.00	1.72
SG: Select Beckman Coulter Instruments (Chem)	22	26.05	1.00	24.00-28.00	3.85	22	33.77	1.15	31.00-37.00	3.53	21	21.95	0.86	20.00-24.00	4.09	22	10.32	0.48	8.00-12.00	5.00	22	35.95	1.79	33.00-39.00	5.00
SG: Select Siemens Instruments	14	25.79	1.12	24.00-28.00	4.23	13	33.85	1.41	31.00-37.00	4.12	14	22.07	0.92	20.00-24.00	4.09	14	10.43	0.76	8.00-12.00	8.00	14	36.00	1.11	33.00-39.00	3.06
SG: Select Roche Instruments	6	25.50	0.84	24.00-28.00	3.08	6	33.17	1.17	30.00-36.00	3.64	6	21.50	1.23	20.00-24.00	5.45	6	9.83	0.41	8.00-12.00	4.00	6	35.17	1.72	32.00-38.00	4.86
SG: Select OCD Vitros Instruments	11	20.64	0.51	19.00-23.00	2.38	11	27.00	0.63	25.00-29.00	2.22	11	17.64	0.51	16.00-20.00	2.78	11	8.00	0.45	6.00-10.00	5.00	11	28.55	0.52	26.00-32.00	1.72
SG: Abaxis Piccolo Instruments	40	23.36	0.68	21.00-25.00	3.04	40	30.84	0.97	28.00-34.00	3.23	9	19.78	0.97	18.00-22.00	5.00	9	9.11	0.60	7.00-11.00	6.67	9	32.22	1.30	29.00-35.00	4.06
SG: Select Abbott Architect Instruments	8	25.50	0.93	24.00-28.00	3.46	8	33.63	0.92	31.00-37.00	2.65	8	21.50	0.54	20.00-24.00	2.27	8	9.88	0.35	8.00-12.00	4.00	8	35.50	0.93	33.00-39.00	2.50
SG: International OCD Vitros Instruments	14	20.79	0.58	19.00-23.00	2.86	14	27.14	0.66	25.00-29.00	2.59	14	17.71	0.61	16.00-20.00	3.33	14	8.07	0.47	6.00-10.00	6.25	14	28.71	0.61	26.00-32.00	2.07
SG: International Abaxis Piccolo Instruments	44	23.37	0.66	21.00-25.00	3.04	44	30.86	0.94	28.00-34.00	2.90	13	19.88	0.82	18.00-22.00	4.00	13	9.32	0.64	7.00-11.00	6.67	13	32.33	1.11	29.00-35.00	3.44
SG: Select Gen Chem Instruments	61	25.40	1.06	23.00-27.00	4.40	61	33.24	1.15	30.00-36.00	3.33	59	21.57	0.93	20.00-24.00	4.09	60	10.06	0.64	8.00-12.00	6.00	59	35.29	1.51	32.00-38.00	4.29
AG: All Group	132	24.43	1.75	22.00-26.00	7.50	131	32.04	2.19	29.00-35.00	6.88	99	21.04	1.61	19.00-23.00	7.62	100	9.78	0.94	8.00-12.00	9.00	99	34.35	2.64	31.00-37.00	7.65
Albumin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived) / BCP	8	2.64	0.21	2.30-2.90	8.08	8	2.89	0.29	2.60-3.20	10.00	8	2.49	0.19	2.30-2.80	7.60	8	2.05	0.09	1.90-2.30	4.29	8	2.95	0.31	2.70-3.30	10.33
PG: Abaxis Piccolo (Waived) / BCP	36	2.51	0.10	2.30-2.80	4.00	36	2.76	0.10	2.50-3.10	3.57															
PG: Abbott Architect / BCP	6	2.48	0.07	2.30-2.80	3.20	6	2.73	0.10	2.40-3.00	3.70	6	2.33	0.10	2.10-2.50	4.35	6	1.92	0.07	1.70-2.10	4.21	6	2.82	0.08	2.50-3.10	2.86
PG: Alfa Wasserman Alera/Axcel / BCG	25	3.18	0.09	2.90-3.50	2.81	25	3.62	0.09	3.20-4.00	2.50	25	3.02	0.20	2.70-3.30	6.67	24	2.26	0.06	2.10-2.50	2.61	25	3.64	0.34	3.20-4.00	9.44
PG: Clinical Data Envoy 500 / BCG	5	3.37	0.12	3.10-3.70	3.53	5	3.74	0.15	3.30-4.10	4.05	5	3.06	0.12	2.80-3.40	3.55	5	2.40	0.07	2.20-2.60	2.92	5	3.84	0.09	3.40-4.20	2.37
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL / BCG	8	2.50	0.19	2.30-2.80	7.60	8	2.79	0.19	2.50-3.10	6.79	8	2.38	0.16	2.20-2.60	6.67	8	1.90	0.14	1.70-2.10	7.37	8	2.86	0.19	2.60-3.20	6.55
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL / BCP																									



PG: Beckman AU	18	175.50	15.35	123.00-229.00	8.75	18	223.22	19.52	156.00-290.00	8.74	18	148.28	13.44	104.00-192.00	9.05	18	65.78	6.35	46.00-86.00	9.55	18	235.06	21.03	165.00-306.00	8.94
PG: Roche Cobas Integra	5	195.38	3.83	137.00-254.00	1.95	5	248.28	6.77	174.00-322.00	2.74	5	163.42	3.64	114.00-212.00	2.21	5	73.20	1.48	51.00-95.00	2.05	5	262.02	6.90	183.00-341.00	2.63
PG: OCD Vitros 250,350,400,500,750,950	7	156.29	6.24	109.00-203.00	3.97	7	182.57	6.73	128.00-238.00	3.66	7	136.29	7.95	95.00-177.00	5.88	7	70.00	3.74	49.00-91.00	5.29	7	188.00	10.03	132.00-244.00	5.32
SG: Select Beckman Coulter Instruments (Chem)	20	176.50	14.85	124.00-230.00	8.42	20	223.95	18.63	157.00-291.00	8.30	20	148.60	12.83	104.00-194.00	8.59	20	65.95	6.03	46.00-86.00	9.09	20	236.05	20.19	165.00-307.00	8.56
SG: Select Siemens Instruments	8	216.75	7.89	152.00-282.00	3.64	8	277.75	9.16	195.00-361.00	3.31	8	185.63	9.16	130.00-242.00	4.95	8	83.25	4.50	58.00-108.00	5.42	8	295.13	9.37	207.00-384.00	3.19
SG: Select Roche Instruments	6	194.32	4.30	136.00-252.00	2.22	6	247.23	6.57	173.00-321.00	2.67	6	163.18	3.31	114.00-212.00	2.02	6	73.17	1.33	51.00-95.00	1.78	6	261.68	6.23	183.00-341.00	2.37
SG: Select OCD Vitros Instruments	10	155.60	5.78	109.00-203.00	3.72	10	181.80	7.10	127.00-237.00	3.90	10	136.00	7.12	95.00-177.00	5.22	10	69.80	3.62	49.00-91.00	5.14	10	187.40	9.06	131.00-243.00	4.87
SG: Abaxis Piccolo Instruments	40	159.70	9.90	112.00-208.00	6.19	40	195.38	13.76	137.00-254.00	7.08	5	139.80	19.54	98.00-182.00	13.93	5	67.00	8.00	47.00-87.00	11.94	5	213.20	35.63	149.00-277.00	16.71
SG: Select Abbott Architect Instruments	8	203.75	6.47	143.00-265.00	3.19	8	259.50	9.05	182.00-338.00	3.50	8	170.88	5.59	120.00-222.00	3.27	8	76.38	2.33	53.00-99.00	3.03	8	273.75	8.78	192.00-356.00	3.21
SG: International OCD Vitros Instruments	13	157.77	6.72	111.00-205.00	4.24	13	184.77	8.67	130.00-241.00	4.70	13	137.23	7.00	96.00-178.00	5.11	13	69.62	4.13	49.00-91.00	5.86	13	190.54	10.88	134.00-248.00	5.71
SG: International Abaxis Piccolo Instruments	41	159.51	9.85	112.00-208.00	6.19	41	195.10	13.70	137.00-254.00	7.03	6	137.67	18.24	97.00-179.00	13.19	6	66.67	7.20	47.00-87.00	10.75	6	208.50	33.89	146.00-272.00	16.22
MG: IFCC	8	202.00	14.90	141.00-263.00	7.38	8	259.00	19.60	181.00-337.00	7.57	8	170.00	12.20	119.00-221.00	7.18	8	76.00	6.00	53.00-99.00	7.89	8	272.00	18.80	190.00-354.00	6.91
ALT/SGPT	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	7	134.57	4.50	108.00-162.00	3.33	7	179.57	4.50	144.00-216.00	2.50	7	113.00	3.96	90.00-136.00	3.54	7	45.86	4.38	37.00-55.00	9.57	7	189.71	4.07	152.00-228.00	2.16
PG: Abaxis Piccolo (Waived)	35	134.71	2.47	108.00-162.00	1.85	35	180.20	3.07	144.00-216.00	1.72															
PG: Abbott Architect	6	153.83	2.79	123.00-185.00	1.82	6	208.33	3.83	166.00-250.00	1.83	6	127.33	2.50	102.00-152.00	1.97	6	46.67	1.75	38.00-56.00	3.83	6	221.83	3.71	178.00-266.00	1.67
PG: Alfa Wasserman Alera/Axcel	25	122.84	8.38	98.00-148.00	6.83	25	168.20	8.23	134.00-202.00	4.88	24	102.04	7.20	82.00-122.00	7.06	24	35.58	2.21	29.00-43.00	6.11	23	180.61	3.82	145.00-217.00	2.10
PG: Clinical Data Envoy 500	5	150.10	3.75	120.00-180.00	2.47	5	200.32	5.98	160.00-240.00	3.00	5	125.18	2.63	100.00-150.00	2.08	5	46.14	0.95	37.00-55.00	1.96	5	216.20	3.56	173.00-259.00	1.67
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL / IFCC	9	159.44	3.91	127.00-191.00	2.45	9	211.89	4.20	170.00-254.00	1.98	9	132.22	2.91	106.00-158.00	2.20	9	50.33	2.45	40.00-60.00	4.80	9	226.56	3.64	182.00-272.00	1.59
PG: Beckman AU	18	134.28	4.32	107.00-161.00	3.21	18	182.28	5.36	146.00-218.00	2.97	18	110.89	3.61	89.00-133.00	3.24	18	40.67	1.41	33.00-49.00	3.41	18	192.06	5.87	154.00-230.00	3.07
PG: Roche Cobas Integra	5	143.80	10.45	115.00-173.00	7.22	5	195.20	14.13	156.00-234.00	7.23	5	119.40	8.62	95.00-143.00	7.23	5	43.40	3.21	34.00-52.00	7.44	5	205.80	15.58	165.00-247.00	7.57
PG: OCD Vitros 250,350,400,500,750,950	7	158.57	4.12	127.00-191.00	2.58	7	206.00	6.25	165.00-247.00	3.01	7	136.57	3.91	110.00-164.00	2.85	7	63.29	6.16	50.00-76.00	9.84	7	216.71	6.16	174.00-260.00	2.86
SG: Select Beckman Coulter Instruments (Chem)	20	136.10	6.94	109.00-163.00	5.07	20	184.10	7.60	147.00-221.00	4.13	20	112.25	5.43	90.00-134.00	4.82	20	41.65	3.31	34.00-50.00	7.86	20	194.00	8.20	155.00-233.00	4.23
SG: Select Siemens Instruments	5	157.40	6.99	126.00-188.00	4.46	5	212.60	9.63	170.00-256.00	4.51	5	131.40	5.77	105.00-157.00	4.43	5	50.60	1.34	41.00-61.00	2.55	5	223.60	11.63	179.00-269.00	5.18
SG: Select Roche Instruments	6	144.33	9.44	115.00-173.00	6.53	6	195.83	12.73	157.00-235.00	6.48	6	119.83	7.78	96.00-144.00	6.50	6	43.67	2.94	35.00-53.00	6.59	6	207.00	14.24	166.00-248.00	6.86
SG: Select OCD Vitros Instruments	10	153.90	9.21	123.00-185.00	5.97	10	201.90	8.66	162.00-242.00	4.31	10	131.50	9.27	106.00-158.00	7.05	8	62.63	6.00	50.00-76.00	9.52	10	212.30	9.17	170.00-254.00	4.34
SG: Abaxis Piccolo Instruments	42	134.69	2.83	108.00-162.00	2.07	42	180.10	3.29	144.00-216.00	1.83	7	113.00	3.96	90.00-136.00	3.54	7	45.86	4.38	37.00-55.00	9.57	7	189.71	4.07	152.00-228.00	2.16
SG: Select Abbott Architect Instruments	8	154.50	2.67	124.00-186.00	1.74	8	209.50	3.93	168.00-252.00	1.86	8	128.25	2.71	102.00-154.00	2.11	8	46.88	1.55	38.00-56.00	3.40	8	222.88	3.68	178.00-268.00	1.66
SG: International OCD Vitros Instruments	12	154.83	8.63	124.00-186.00	5.55	12	202.75	8.13	162.00-244.00	3.99	12	132.67	8.84	106.00-160.00	6.62	10	63.10	5.88	50.00-76.00	9.37	12	213.58	8.91	171.00-257.00	4.16
SG: International Abaxis Piccolo Instruments	43	134.81	2.91	108.00-162.00	2.15	43	180.23	3.37	144.00-216.00	1.89	8	113.25	3.73	90.00-136.00	3.27	8	46.88	4.97	38.00-56.00	10.64	8	190.88	5.00	153.00-229.00	2.62
MG: IFCC	11	157.00	8.10	126.00-188.00	5.16	11	209.00	9.40	167.00-251.00	4.50	11	130.00	6.60	104.00-156.00	5.08	11	50.00	3.60	40.00-60.00	7.20	11	223.00	10.30	178.00-268.00	4.62
MG: IFCC	11	157.00	8.10	126.00-188.00	5.16	11	209.00	9.40	167.00-251.00	4.50	11	130.00	6.60	104.00-156.00	5.08	11	50.00	3.60	40.00-60.00	7.20	11	223.00	10.30	178.00-268.00	4.62
AG: All Group	133	139.03	13.15	111.00-167.00	9.50	133	186.97	15.90	150.00-224.00	8.50	100	117.69	13.06	94.00-142.00	11.10	96	44.30	8.09	35.00-53.00	18.41	96	201.76	18.07	162.00-242.00	8.96
AST/SGOT	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	7	207.29	7.43	166.00-248.00	3.57	7	245.29	7.20	196.00-294.00	2.94	7	187.57	6.13	150.00-226.00	3.24	7	131.14	5.52	105.00-157.00	4.20	7	253.71	9.88	203.00-305.00	3.90
PG: Abaxis Piccolo (Waived)	35	205.29	5.06	164.00-246.00	2.49	35	244.40	6.26	195.00-293.00	2.58															
PG: Abbott Architect	6	211.17	3.43	169.00-253.00	1.61	6	253.00	5.14	202.00-304.00	2.02	6	191.33	3.50	153.00-229.00	1.83	6	132.00	3.03	106.00-158.00	2.27	6	261.33	5.01	209.00-313.00	1.92
PG: Alfa Wasserman Alera/Axcel	25	191.44	8.08	153.00-229.00	4.24	25	229.56	6.16	184.00-276.00	2.70	25	178.56	13.48	143.00-215.00	7.54	24	122.67	3.42	98.00-148.00	2.76	25	231.44	27.60	185.00-277.00	11.95
PG: Clinical Data Envoy 500	5	211.80	9.73	170.00-254.00	4.58	5	251.80	11.52	202.00-302.00	4.56	5	193.40	8.38	154.00-232.00	4.35	5	133.00	5.96	106.00-160.00	4.51	5	260.20	10.78	208.00-312.00	4.15
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	13	217.15	8.05	174.00-260.00	3.73	13	260.23	8.98	208.00-312.00	3.46	13	198.54	8.00	159.00-239.00	4.02	13	134.15	5.26	107.00-161.00	3.96	13	271.23	9.81	217.00-325.00	3.62
PG: Beckman AU	18	188.33	6.28	150.00-226.00	3.35	18	224.56	7.89	180.00-270.00	3.51	18	170.44	5.87	136.00-204.00	3.47	18	117.56	4.16	94.00-142.00	3.56	18	230.89	7.96	185.00-277.00	3.46
PG: Roche Cobas Integra	5	215.00	10.27	172.00-258.00	4.79	5	258.00	12.08	206.00-310.00	4.69	5	196.00	7.87	157.00-235.00	4.03	5	136.60	6.35	110.00-164.00	4.60	5	266.60	14.10	214.00-320.00	5.28
PG: OCD Vitros 250,350,400,500,750,950	7	223.71	5.47	179.00-269.00	2.46	7	282.00	8.83	226.00-338.00	3.12	7	201.29	4.68	161.00-241.00	2.34	7	131.57	3.26	106.00-158.00	2.50	7	299.43	8.44	239.00-359.00	2.81
SG: Select Beckman Coulter Instruments (Chem)	20	189.90	7.68	152.00-228.00	4.05	20	226.25	9.35	181.00-271.00	4.12	20	171.85	7.19	138.00-206.00	4.19	20	118.55	5.11	95.00-143.00	4.29	20	232.80	9.82	186.00-280.00	4.21
SG: Select Siemens Instruments	14	218.00	8.36	174.00-262.00	3.85	14	260.86																		



SG: Select Abbott Architect Instruments	8	2.94	0.23	2.30-3.50	7.93	8	3.98	0.28	3.20-4.80	7.00	8	2.48	0.22	2.00-3.00	8.80	8	0.96	0.13	0.60-1.40	13.00	8	4.16	0.29	3.40-5.00	6.90
SG: International OCD Vitros Instruments	13	3.27	0.26	2.60-4.00	7.88	13	4.50	0.27	3.60-5.40	6.00	13	2.70	0.22	2.20-3.20	8.15	13	0.85	0.11	0.50-1.30	12.22	13	4.75	0.33	3.80-5.80	6.88
SG: International Abaxis Piccolo Instruments	41	2.88	0.40	2.30-3.50	13.79	41	3.84	0.54	3.00-4.60	14.21	6	2.60	0.17	2.10-3.10	6.54	6	1.07	0.05	0.70-1.50	4.55	6	4.32	0.32	3.40-5.20	7.44
SG: Select Gen Chem Instruments	58	3.28	0.31	2.60-4.00	9.39	58	4.47	0.45	3.60-5.40	10.00	58	2.79	0.42	2.20-3.40	15.00	57	1.07	0.12	0.70-1.50	10.91	57	4.64	0.52	3.70-5.50	11.30
AG: All Group	126	3.10	0.30	2.50-3.70	9.68	127	4.22	0.44	3.40-5.00	10.48	92	2.67	0.28	2.20-3.20	10.74	91	1.01	0.14	0.60-1.40	14.00	91	4.56	0.46	3.70-5.50	10.00
Total Protein	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	5	5.54	0.06	5.00-6.00	0.91	5	6.64	0.06	5.90-7.30	0.76	5	4.98	0.04	4.50-5.50	0.80	5	3.30	0.07	3.00-3.60	2.12	5	6.92	0.11	6.20-7.60	1.59
PG: Abaxis Piccolo (Waived)	34	5.49	0.07	5.00-6.00	1.27	34	6.62	0.08	5.90-7.30	1.21															
PG: Abbott Architect	6	5.37	0.15	4.90-5.90	2.78	6	6.58	0.15	5.90-7.30	2.27	6	4.82	0.16	4.30-5.30	3.33	6	3.00	0.16	2.70-3.30	5.00	6	6.90	0.11	6.20-7.60	1.59
PG: Alfa Wasserman Alera/Axcel	25	5.62	0.12	5.00-6.20	2.14	25	6.82	0.13	6.10-7.50	1.91	25	5.11	0.38	4.60-5.60	7.45	24	3.25	0.07	3.00-3.60	2.12	23	7.07	0.27	6.40-7.80	3.80
PG: Clinical Data Envoy 500	5	5.36	0.48	4.90-5.90	9.07	5	6.74	0.15	6.00-7.40	2.24	5	4.91	0.10	4.40-5.40	2.04	5	3.20	0.07	2.90-3.50	2.19	5	6.97	0.12	6.30-7.70	1.71
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	13	5.58	0.14	5.00-6.20	2.50	13	6.78	0.17	6.10-7.50	2.50	13	5.03	0.13	4.50-5.50	2.60	13	3.27	0.10	3.00-3.60	3.03	13	7.02	0.17	6.30-7.70	2.43
PG: Beckman AU	18	5.31	0.15	4.80-5.80	2.83	18	6.42	0.16	5.80-7.00	2.50	18	4.74	0.14	4.20-5.20	2.98	18	3.07	0.08	2.80-3.40	2.58	18	6.73	0.26	6.00-7.40	3.88
PG: Roche Cobas Integra	5	5.28	0.34	4.80-5.80	6.23	5	6.34	0.42	5.70-6.90	6.67	5	4.70	0.29	4.20-5.20	6.17	5	3.08	0.22	2.80-3.40	7.10	5	6.64	0.45	5.90-7.30	6.82
PG: OCD Vitros 250,350,400,500,750,950	7	5.31	0.09	4.80-5.80	1.70	7	6.43	0.10	5.80-7.00	1.56	7	4.83	0.13	4.30-5.30	2.71	7	3.21	0.04	2.90-3.50	1.25	7	6.70	0.12	6.00-7.40	1.79
SG: Select Beckman Coulter Instruments (Chem)	22	5.33	0.14	4.80-5.80	2.64	22	6.43	0.17	5.80-7.00	2.66	22	4.75	0.14	4.30-5.30	2.92	22	3.09	0.08	2.80-3.40	2.58	22	6.73	0.23	6.00-7.40	3.58
SG: Select Siemens Instruments	14	5.56	0.16	5.00-6.20	2.68	14	6.76	0.20	6.10-7.50	2.79	14	5.01	0.15	4.50-5.50	3.00	14	3.25	0.12	3.00-3.60	3.64	14	7.00	0.19	6.30-7.70	2.71
SG: Select Roche Instruments	6	5.33	0.33	4.80-5.80	6.23	6	6.40	0.41	5.80-7.00	6.25	6	4.75	0.29	4.30-5.30	6.04	6	3.12	0.21	2.80-3.40	6.77	6	6.70	0.43	6.00-7.40	6.42
SG: Select OCD Vitros Instruments	10	5.35	0.12	4.90-5.90	2.22	10	6.44	0.13	5.80-7.00	2.03	10	4.86	0.16	4.40-5.40	3.27	10	3.21	0.09	2.90-3.50	2.81	10	6.70	0.13	6.00-7.40	1.79
SG: Abaxis Piccolo Instruments	39	5.50	0.07	5.00-6.00	1.27	39	6.62	0.07	5.90-7.30	1.06	5	4.98	0.04	4.50-5.50	0.80	5	3.30	0.07	3.00-3.60	2.12	5	6.92	0.11	6.20-7.60	1.59
SG: Select Abbott Architect Instruments	8	5.39	0.14	4.90-5.90	2.59	8	6.60	0.13	5.90-7.30	1.97	8	4.83	0.14	4.30-5.30	2.92	8	2.99	0.14	2.70-3.30	4.67	8	6.91	0.10	6.20-7.60	1.45
SG: International OCD Vitros Instruments	13	5.35	0.11	4.90-5.90	1.85	13	6.43	0.11	5.80-7.00	1.72	13	4.87	0.14	4.40-5.40	2.86	13	3.22	0.08	2.90-3.50	2.50	13	6.68	0.12	6.00-7.40	1.79
SG: International Abaxis Piccolo Instruments	40	5.50	0.07	5.00-6.00	1.27	40	6.62	0.07	5.90-7.30	1.06	6	4.98	0.04	4.50-5.50	0.80	6	3.30	0.06	3.00-3.60	1.82	6	6.92	0.10	6.20-7.60	1.45
SG: Select Gen Chem Instruments	60	5.46	0.23	5.00-6.00	4.18	60	6.64	0.23	5.90-7.30	3.48	60	4.93	0.31	4.40-5.40	6.33	59	3.15	0.13	2.80-3.40	4.19	58	6.91	0.27	6.20-7.60	3.91
AG: All Group	129	5.47	0.19	5.00-6.00	3.45	129	6.62	0.21	5.90-7.30	3.18	95	4.92	0.27	4.40-5.40	5.51	94	3.18	0.14	2.90-3.50	4.06	93	6.89	0.27	6.20-7.60	3.91
Phosphorous	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	5	4.07	0.21	3.50-4.70	5.12	5	4.97	0.06	4.80-5.20	1.20	5	3.63	0.06	3.40-3.80	1.67	5	2.30	0.10	2.00-2.60	4.35	5	5.23	0.15	4.80-5.70	2.88
PG: Alfa Wasserman Alera/Axcel	6	3.50	0.09	3.20-3.80	2.57	6	4.27	0.15	3.80-4.80	3.49	5	3.08	0.13	2.70-3.50	4.19	5	1.94	0.09	1.60-2.20	4.74	5	4.18	0.57	2.50-5.90	13.57
PG: Beckman AU	9	3.59	0.15	3.20-4.00	4.17	9	4.47	0.16	4.00-5.00	3.56	9	3.16	0.13	2.80-3.60	4.06	9	1.86	0.08	1.70-2.10	4.21	9	4.64	0.16	4.10-5.10	3.48
SG: Select Beckman Coulter Instruments (Chem)	12	3.62	0.14	3.20-4.00	4.17	12	4.50	0.15	4.00-5.00	3.33	12	3.18	0.12	2.80-3.60	3.75	12	1.88	0.08	1.70-2.10	4.21	12	4.70	0.17	4.20-5.20	3.62
SG: Select Siemens Instruments	5	3.78	0.05	3.60-3.90	1.32	5	4.58	0.10	4.30-4.90	2.17	5	3.30	0.08	3.10-3.50	2.42	5	2.05	0.06	1.90-2.30	2.86	5	4.85	0.06	4.70-5.10	1.22
SG: Select Roche Instruments	5	3.58	0.15	3.20-4.00	4.17	5	4.45	0.06	4.30-4.70	1.33	5	3.18	0.10	2.90-3.50	3.13	5	1.90	0.08	1.70-2.10	4.21	5	4.58	0.10	4.30-4.90	2.17
SG: Select OCD Vitros Instruments	5	3.60	0.18	3.10-4.10	5.00	5	4.38	0.28	3.60-5.20	6.36	5	3.20	0.18	2.70-3.70	5.63	5	2.03	0.15	1.60-2.50	7.50	5	4.58	0.17	4.10-5.10	3.70
SG: International OCD Vitros Instruments	7	3.70	0.18	3.20-4.20	4.86	7	4.49	0.25	3.80-5.30	5.56	7	3.29	0.17	2.80-3.80	5.15	7	2.10	0.15	1.70-2.50	7.14	7	4.70	0.20	4.10-5.30	4.26
SG: International Abaxis Piccolo Instruments	7	4.10	0.16	3.60-4.60	3.90	7	4.96	0.11	4.70-5.30	2.20	7	3.65	0.09	3.40-4.00	2.43	7	2.41	0.13	2.00-2.80	5.42	7	5.24	0.10	4.90-5.50	1.92
SG: Select Gen Chem Instruments	23	3.60	0.13	3.20-4.00	3.61	23	4.43	0.18	3.90-4.90	4.09	22	3.16	0.12	2.80-3.60	3.75	22	1.90	0.07	1.70-2.10	3.68	22	4.60	0.38	3.50-5.70	8.26
AG: All Group	38	3.65	0.19	3.10-4.30	5.14	38	4.48	0.22	3.80-5.20	4.89	37	3.22	0.17	2.70-3.70	5.31	37	1.96	0.14	1.60-2.40	7.00	37	4.67	0.35	3.70-5.80	7.45
Cholesterol Total	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	5	144.50	2.08	131.00-160.00	1.45	5	176.50	4.51	159.00-195.00	2.54	5	131.50	2.65	119.00-145.00	1.97	5	80.25	1.26	72.00-88.00	1.63	5	182.75	2.75	165.00-201.00	1.53
PG: Alfa Wasserman Alera/Axcel	22	147.91	2.62	133.00-163.00	1.76	22	181.23	4.12	163.00-199.00	2.27	21	131.76	2.36	119.00-145.00	1.82	21	81.86	2.24	74.00-90.00	2.68	21	185.76	12.76	167.00-205.00	6.88
PG: Cholestech LDX (Waived)	5	146.33	4.51	131.00-161.00	3.08	5	186.50	6.86	168.00-206.00	3.69	5	127.50	3.00	115.00-141.00	2.34	5	100.00	0.00	90.00-110.00	0.00	5	190.00	5.48	171.00-209.00	2.89
PG: Clinical Data Envoy 500	5	146.60	4.28	132.00-162.00	2.93	5	182.60	5.18	165.00-201.00	2.84	5	128.80	5.93	116.00-142.00	4.57	5	78.80	2.68	71.00-87.00	3.42	5	187.60	6.03	169.00-207.00	3.19
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	12	138.33	4.75	124.00-152.00	3.48	12	171.25	5.21	154.00-188.00	3.04	12	122.08	3.45	110.00-134.00	2.79	12	72.83	2.98	66.00-80.00	4.11	12	180.08	9.37	162.00-198.00	5.22
PG: Beckman AU	18	140.50	4.25	127.00-155.00	2.98	18	173.11	4.59	156.00-190.00	2.66	18	124.28	3.86	112.00-136.00	3.15	18	76.11	2.49	68.00-84.00	3.29	18	180.61	4.78	163.00-199.00	2.65
PG: Roche Cobas Integra	5	142.20	8.07	128.00-156.00	5.70	5	174.60	9.32	158.00-193.00	5.31	5	125.00	6.86	113.00-138.00	5.52	5	77.40	4.22	69.00-85.00	5.45	5	183.40	9.89	165.00-201.00	5.41
PG: OCD Vitros 250,350,400,500,750,950	7	141.57	3.21	128.00-156.00	2.25	7	177.71	4.35	160.00-196.00	2.42	7	124.43	4.96	112.00-136.00	4.03	7	72.43	3.26	65.00-79.00	4.58	7	185.29	3.04	167.00-204.00	1.62
SG: Select Beckman Coulter Instruments (Chem)	22	141.36	4.37	127.00-155.00	3.12	22	174.68	5.52	158.00-193.00	3.14	22	125.09	4.08	113.00-138.00	3.28	22	76.55	2.54	69.00-85.00	3.25	22	182.23	5.87	164.00-200.00	3.24
SG: Select Siemens Instruments	13	137.85	4.88	124.00-152.00	3.55	13	170.54	5.61	154.00-188.00	3.27	13	121.69	3.59	110.00-134.00	2.95	13	72.69	2.90	66.00-80.00	3.97	13	179.31			



SG: Select Abbott Architect Instruments / Direct Method	5	64.00	2.45	45.00-83.00	3.75	5	79.20	3.11	55.00-103.00	3.92	5	57.00	2.00	40.00-74.00	3.51	5	35.00	2.00	25.00-46.00	5.71	5	83.00	3.94	58.00-108.00	4.70
SG: International OCD Vitros Instruments / Direct Method	13	66.69	2.43	47.00-87.00	3.58	13	87.54	2.85	62.00-114.00	3.18	13	58.08	2.46	41.00-75.00	4.31	13	35.77	1.48	25.00-47.00	4.17	13	92.31	4.46	64.00-120.00	4.89
SG: International Abaxis Piccolo Instruments / Direct Method	8	58.50	3.85	41.00-77.00	6.61	8	77.38	10.60	54.00-100.00	13.77															
SG: Select Gen Chem Instruments / Direct Method	53	64.01	4.03	45.00-83.00	6.25	53	78.75	5.30	55.00-103.00	6.71	53	57.02	4.80	40.00-74.00	8.42	52	34.80	2.57	25.00-46.00	7.43	53	80.82	9.36	57.00-105.00	11.60
MG: Direct Method	149	78.00	16.90	55.00-101.00	21.67	163	96.00	18.30	67.00-125.00	19.06	87	58.00	6.30	41.00-75.00	10.86	86	35.00	4.40	25.00-46.00	12.57	87	85.00	9.70	60.00-111.00	11.41
MG: Direct Method	149	78.00	16.90	55.00-101.00	21.67	163	96.00	18.30	67.00-125.00	19.06	87	58.00	6.30	41.00-75.00	10.86	86	35.00	4.40	25.00-46.00	12.57	87	85.00	9.70	60.00-111.00	11.41
AG: All Group	130	73.23	14.44	51.00-95.00	19.73	124	89.19	16.47	62.00-116.00	18.54	90	58.21	10.74	41.00-75.00	18.45	89	35.35	7.00	25.00-46.00	20.00	90	85.81	15.08	60.00-112.00	17.56
Cholesterol LDL	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived) / Calc	5	49.75	1.71	45.00-55.00	3.40	5	62.50	13.18	23.00-103.00	20.95	5	57.25	11.00	24.00-90.00	19.30	5	34.00	9.20	6.00-62.00	27.06	5	60.00	7.66	37.00-83.00	12.83
PG: Alfa Wasserman Alera/Axcel / Calc	13	51.45	3.46	41.00-62.00	6.86	13	66.75	6.29	48.00-86.00	9.40	13	44.80	3.69	34.00-56.00	8.22	11	22.82	2.14	17.00-29.00	9.13	13	64.95	12.74	27.00-103.00	19.54
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL / Calc	7	44.14	6.91	23.00-65.00	15.68	7	55.71	7.27	34.00-78.00	13.04	7	37.23	7.81	14.00-60.00	21.08	6	21.70	8.10	0.00-46.00	36.82	7	58.17	6.55	38.00-78.00	11.38
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL / Direct Method	5	47.40	8.90	20.00-74.00	18.94	5	62.75	3.50	53.00-74.00	5.56	5	42.60	8.65	17.00-69.00	20.00	5	27.40	4.39	14.00-40.00	16.30	5	60.40	10.90	27.00-93.00	18.17
PG: Beckman AU / Calc	8	42.88	6.53	24.00-63.00	15.12	8	52.75	7.89	29.00-77.00	14.91	8	37.00	5.26	21.00-53.00	14.32	8	21.00	3.89	9.00-33.00	18.57	8	56.13	7.62	33.00-79.00	13.57
PG: Beckman AU / Direct Method	6	40.17	2.32	33.00-47.00	5.75	6	49.17	2.40	42.00-56.00	4.90	6	35.83	1.72	31.00-41.00	4.72	6	22.00	0.63	20.00-24.00	2.73	6	51.77	2.61	44.00-60.00	5.00
PG: OCD Vitros 250,350,400,500,750,950 / Calc	6	41.33	4.41	28.00-54.00	10.73	6	50.50	4.59	37.00-65.00	9.02	6	37.33	5.01	22.00-52.00	13.51	6	15.33	3.62	4.00-26.00	24.00	6	52.50	6.03	35.00-71.00	11.32
SG: Select Beckman Coulter Instruments (Chem) / Calc	9	43.67	6.56	24.00-64.00	15.00	9	54.00	8.28	29.00-79.00	15.37	9	37.56	5.20	22.00-54.00	13.68	9	21.56	4.00	10.00-34.00	18.18	9	57.44	8.16	32.00-82.00	14.39
SG: Select Beckman Coulter Instruments (Chem) / Direct Method	9	41.33	4.18	28.00-54.00	10.24	9	50.89	5.71	34.00-68.00	11.18	9	36.44	2.88	27.00-45.00	8.06	9	21.89	0.93	19.00-25.00	4.09	9	53.51	5.95	36.00-72.00	10.93
SG: Select Siemens Instruments / Calc	8	45.88	8.06	22.00-70.00	17.61	8	54.63	7.41	33.00-77.00	13.45	8	37.70	7.35	16.00-60.00	19.47	7	22.17	7.50	0.00-45.00	34.09	8	58.53	6.15	41.00-77.00	10.34
SG: Select Roche Instruments / Direct Method	5	79.00	4.24	66.00-92.00	5.32	5	98.00	5.66	81.00-115.00	5.82	5	70.00	4.24	57.00-83.00	6.00	5	42.50	2.12	37.00-49.00	4.88	5	101.50	7.78	79.00-125.00	7.65
SG: Select OCD Vitros Instruments / Calc	7	40.71	4.35	28.00-54.00	10.49	7	49.86	4.53	37.00-64.00	9.00	7	37.00	4.66	23.00-51.00	12.70	7	14.43	4.08	2.00-26.00	29.29	7	52.29	5.53	36.00-69.00	10.58
SG: Abaxis Piccolo Instruments / Calc	7	52.71	3.95	41.00-65.00	7.36	7	61.00	10.98	28.00-94.00	18.03	5	57.25	11.00	24.00-90.00	19.30	5	34.00	9.20	6.00-62.00	27.06	5	60.00	7.66	37.00-83.00	12.83
PG: Polymer CardioChek / Calc	9	9.80	4.30	0.00-23.00	43.00	11	27.35	5.11	12.00-42.00	18.89															
SG: International OCD Vitros Instruments / Calc	10	39.60	5.17	24.00-56.00	13.00	10	49.30	6.53	30.00-69.00	13.27	10	35.80	5.18	20.00-52.00	14.44	10	14.10	3.81	3.00-25.00	27.14	10	51.10	7.81	28.00-74.00	15.29
SG: International Abaxis Piccolo Instruments / Calc	8	53.88	4.91	39.00-69.00	9.07	8	58.38	12.59	20.00-96.00	21.72	5	55.80	10.06	26.00-86.00	18.04	5	35.00	8.28	10.00-60.00	23.71	5	60.00	7.66	37.00-83.00	12.83
SG: Select Gen Chem Instruments / Calc	24	48.37	5.95	30.00-66.00	12.29	24	61.53	9.26	34.00-90.00	15.00	24	41.56	5.43	26.00-58.00	12.86	22	22.32	3.01	13.00-31.00	13.64	24	61.81	11.01	29.00-95.00	17.74
SG: Select Gen Chem Instruments / Direct Method	13	43.85	5.52	28.00-61.00	12.50	13	54.08	7.18	32.00-76.00	13.33	13	38.62	4.29	26.00-52.00	11.03	13	23.15	2.15	16.00-30.00	9.57	13	56.51	7.07	36.00-78.00	12.46
MG: Calc	49	48.00	8.30	23.00-73.00	17.29	49	59.00	10.80	27.00-91.00	18.31	46	42.00	9.40	14.00-70.00	22.38	42	23.00	7.50	1.00-46.00	32.61	46	61.00	11.50	27.00-96.00	18.85
MG: Direct Method	18	45.00	6.50	26.00-65.00	14.44	17	56.00	7.40	34.00-78.00	13.21	18	40.00	5.80	23.00-57.00	14.50	18	24.00	3.40	14.00-34.00	14.17	18	58.00	8.20	33.00-83.00	14.14
Triglycerides	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	5	164.50	5.80	124.00-206.00	3.52	5	193.25	8.02	145.00-241.00	4.15	5	150.75	3.77	113.00-189.00	2.52	5	107.25	3.40	80.00-134.00	3.18	5	198.50	6.81	149.00-249.00	3.42
PG: Alfa Wasserman Alera/Axcel	22	159.18	4.87	119.00-199.00	3.08	22	184.50	5.47	139.00-231.00	2.97	22	149.18	11.07	112.00-186.00	7.45	21	108.43	3.19	81.00-135.00	2.96	21	188.95	9.33	142.00-236.00	4.92
PG: Clinical Data Envoy 500	5	152.20	2.77	114.00-190.00	1.84	5	176.80	5.50	133.00-221.00	3.11	5	139.40	1.14	104.00-174.00	0.79	5	102.40	2.19	77.00-128.00	2.16	5	186.00	6.21	140.00-233.00	3.33
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	12	147.92	4.48	111.00-185.00	3.04	12	174.83	4.71	131.00-219.00	2.69	12	134.83	3.24	101.00-169.00	2.37	12	93.58	2.97	71.00-118.00	3.19	12	182.33	8.76	137.00-228.00	4.84
PG: Beckman AU	17	156.71	6.21	118.00-196.00	3.95	17	182.47	6.33	137.00-228.00	3.46	17	142.71	5.02	107.00-179.00	3.50	17	103.18	3.76	77.00-129.00	3.69	17	187.82	6.43	141.00-235.00	3.40
PG: Roche Cobas Integra	5	151.00	7.07	113.00-189.00	4.70	5	173.60	7.83	131.00-218.00	4.48	5	138.20	6.76	104.00-173.00	4.93	5	101.80	5.67	77.00-128.00	5.59	5	179.40	8.44	134.00-224.00	4.69
PG: OCD Vitros 250,350,400,500,750,950	7	169.43	7.25	127.00-211.00	4.32	7	203.00	9.80	152.00-254.00	4.83	7	155.43	5.19	116.00-194.00	3.35	7	108.71	3.64	82.00-136.00	3.30	7	209.43	9.78	157.00-261.00	4.69
SG: Select Beckman Coulter Instruments (Chem)	21	157.33	5.76	118.00-196.00	3.69	21	182.86	5.72	137.00-229.00	3.11	21	142.76	4.73	107.00-179.00	3.29	21	103.10	3.43	77.00-129.00	3.30	21	188.24	6.20	141.00-235.00	3.30
SG: Select Siemens Instruments	13	148.38	4.61	111.00-185.00	3.11	13	175.31	4.82	131.00-219.00	2.74	13	135.31	3.54	101.00-169.00	2.59	13	94.15	3.51	71.00-118.00	3.72	13	182.85	8.59	137.00-229.00	4.70
SG: Select Roche Instruments	6	151.33	6.38	113.00-189.00	4.24	6	174.33	7.23	131.00-218.00	4.14	6	138.67	6.15	104.00-174.00	4.46	6	102.00	5.10	77.00-128.00	5.00	6	180.17	7.78	135.00-225.00	4.33
SG: Select OCD Vitros Instruments	10	170.50	6.22	128.00-214.00	3.63	10	203.50	8.10	153.00-255.00	3.97	10	156.30	4.57	117.00-195.00	2.95	10	109.50	3.50	83.00-138.00	3.18	10	210.20	8.13	158.00-263.00	3.86
SG: Abaxis Piccolo Instruments	7	166.86	5.18	125.00-209.00	3.11	7	195.29	6.24	146.00-244.00	3.18	5	150.75	3.77	113.00-189.00	2.52	5	107.25	3.40	80.00-134.00	3.18	5	198.50	6.81	149.00-249.00	3.42
PG: Polymer CardioChek	67	170.76	16.51	128.00-214.00	9.65	67	195.87	17.55	147.00-245.00	8.98															
SG: Select Abbott Architect Instruments	5	152.80	1.64	115.00-191.00	1.05	5	179.40	2.51	134.00-224.00	1.40	5	139.60	1.82	105.00-175.00	1.29	5	99.80	1.48	75.00-125.00	1.50	5	185.60	1.82	140.00-233.00	0.97
SG: International OCD Vitros Instruments	13	171.62	6.53	129.00-215.00	3.78	13	204.31	8.23	153.00-255.00	4.02	13	156.85	5.00	118.00-196.00	3.18	13	110.54	3.78	83.00-139.00	3.42	13	211.62	8.93	159.00-265.00	4.20
SG: International Abaxis Piccolo Instruments	8	167.50	5.13	126.00-210.00	3.04	8	195.63	5.85	147.00-245.00	3.01	5	151.20	3.42	113.00-189.00	2.25	5	108.20	3.63	81.00-135.00	3.33	5	199.80	6.57	150.00-250.00	3.30
SG: Select Gen Chem Instruments	53	157.19	5.39	118.00-196.00	3.44	53	182.64	5.76	137																

Print Date: 7/29/2019



SG: Select Siemens Instruments	6	113.50	1.38	91.00-137.00	1.23	6	148.00	1.67	118.00-178.00	1.15	6	97.17	0.98	78.00-116.00	1.03	6	45.00	0.63	36.00-54.00	1.33	6	155.00	1.10	124.00-186.00	0.71
SG: Select Abbott Architect Instruments	5	114.75	1.71	92.00-138.00	1.48	5	150.25	1.50	120.00-180.00	1.00	5	98.00	1.83	78.00-118.00	1.84	5	45.50	1.00	37.00-55.00	2.17	5	158.00	1.83	126.00-190.00	1.14
PG: Intl OCD Vitros 250,350,400,500,750,950	5	126.33	6.66	101.00-151.00	5.32	5	169.00	6.08	135.00-203.00	3.61	5	101.67	4.62	82.00-122.00	4.51	5	34.00	2.65	27.00-41.00	7.65	5	180.33	4.62	144.00-216.00	2.56
SG: Select Gen Chem Instruments	21	119.05	7.82	95.00-143.00	6.55	21	155.81	10.82	125.00-187.00	6.92	20	101.20	5.90	81.00-121.00	5.84	21	46.76	3.40	38.00-56.00	7.23	21	163.71	11.16	131.00-197.00	6.83
AG: All Group	31	118.13	6.88	94.00-142.00	5.85	31	155.13	10.27	124.00-186.00	6.65	30	100.40	5.12	80.00-120.00	5.10	31	45.81	3.59	37.00-55.00	7.83	31	163.16	10.88	130.00-196.00	6.69
TIBC	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Alfa Wasserman Alera/Axcel	5	252.33	7.23	230.60-274.00	2.87	5	280.33	5.13	264.90-295.70	1.83	5	231.00	3.61	220.20-241.80	1.56	5	162.67	5.69	145.60-179.80	3.50	5	292.33	8.62	266.40-318.20	2.95
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	5	177.75	6.45	158.50-197.20	3.63	5	231.00	7.96	207.10-254.90	3.45	5	152.00	1.83	146.50-157.50	1.20	5	77.25	4.99	62.30-92.30	6.46	5	238.50	3.87	226.90-250.10	1.62
SG: Select Beckman Coulter Instruments (Chem)	5	233.80	44.05	101.70-365.90	18.84	5	292.00	57.39	119.80-464.20	19.65	5	199.75	39.42	81.50-318.10	19.73	5	114.00	15.92	66.20-161.80	13.96	5	308.00	58.77	131.70-484.30	19.08
SG: Select Siemens Instruments	5	177.75	6.45	158.50-197.20	3.63	5	231.00	7.96	207.10-254.90	3.45	5	152.00	1.83	146.50-157.50	1.20	5	77.25	4.99	62.30-92.30	6.46	5	238.50	3.87	226.90-250.10	1.62
PG: Intl OCD Vitros 250,350,400,500,750,950	5	304.00	45.03	168.90-439.10	14.81	5	407.00	77.00	176.00-638.00	18.92	5	251.67	37.43	139.40-364.00	14.87	5	158.00	29.31	70.10-245.90	18.55	5	448.33	63.26	258.50-638.10	14.11
AG: All Group	15	216.69	42.90	88.00-345.40	19.80	15	263.53	48.18	119.00-408.00	18.28	15	194.52	44.28	61.70-327.30	22.77	16	115.86	34.94	11.10-220.70	30.15	15	275.22	51.48	120.80-429.60	18.71
UIBC	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	5	121.00	9.72	91.80-150.20	8.03	5	147.00	12.06	110.80-183.20	8.20	5	110.40	8.08	86.20-134.60	7.32	5	72.40	5.81	55.00-89.80	8.02	5	152.80	11.65	117.90-187.80	7.62
Digoxin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	5	1.50	0.10	1.20-1.80	6.67	5	1.83	0.15	1.40-2.20	8.33	5	1.33	0.06	1.00-1.60	4.62	5	0.90	0.20	0.70-1.10	22.22	5	1.80	0.27	1.40-2.20	14.44
746 - Auto Diff 6 (Abbott 5-Part Diff)	CL-6					CL-7					CL-8					CL-9					CL-10				
White Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn Ruby	9	2.93	0.14	2.50-3.30	4.83	9	19.50	0.46	16.60-22.40	2.36	8	3.03	0.09	2.50-3.50	3.00	9	19.49	0.50	16.60-22.40	2.56	9	7.29	0.24	6.20-8.40	3.29
Lymphocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn Ruby	9	39.68	4.53	26.10-53.30	11.41	9	17.28	0.92	14.50-20.10	5.32	8	40.06	5.10	24.80-55.40	12.72	9	17.77	0.81	15.40-20.20	4.55	8	24.60	1.74	19.40-29.80	7.07
Monocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn Ruby	9	10.16	4.69	0.00-24.24	46.18	9	3.84	0.39	2.68-5.00	10.03	8	6.73	2.67	0.00-14.75	39.73	9	3.61	0.57	1.91-5.31	15.68	9	4.95	0.82	2.49-7.41	16.57
Granulocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn Ruby	9	45.59	2.05	39.50-51.80	4.50	9	74.21	0.90	71.50-76.90	1.21	8	46.83	1.37	42.70-50.90	2.93	9	73.89	0.69	71.80-76.00	0.93	9	66.53	2.18	60.00-73.00	3.28
Eosinophil	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn Ruby	9	4.30	0.30	3.40-5.20	6.98	8	4.39	0.21	3.80-5.00	4.77	8	4.30	0.32	3.30-5.30	7.44	9	4.51	0.20	3.90-5.10	4.67	9	2.88	0.20	2.30-3.50	7.24
Basophil	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn Ruby	7	0.16	0.19	0.00-0.72	115.63	8	0.21	0.08	0.00-0.44	36.67	6	0.05	0.03	0.00-0.15	66.00	8	0.20	0.10	0.00-0.50	50.00	9	0.20	0.14	0.00-0.63	71.00
Red Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn Ruby	9	2.24	0.06	2.11-2.37	2.63	9	5.22	0.08	4.91-5.53	1.49	8	2.25	0.05	2.12-2.38	2.27	9	5.19	0.07	4.88-5.50	1.43	9	4.47	0.08	4.20-4.74	1.86
Hemoglobin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn Ruby	9	5.72	0.08	5.30-6.10	1.40	9	16.56	0.26	15.40-17.80	1.57	9	5.61	0.15	5.20-6.00	2.68	9	16.48	0.27	15.30-17.70	1.64	9	13.31	0.28	12.40-14.20	2.11
Hematocrit	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs									

Print Date: 7/29/2019



hsCRP	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
AG: All Group	5	0.23	0.03	0.14-0.32	13.04	5	1.04	0.11	0.71-1.37	10.58																
602 - Hemoglobin and Hematocrit - Waived;614 - Auto Diff II (All Other 3-part);610 - Regular Hematology;601 - Waived/PPM Tests II;609 - Waived/PPM Tests I	HD-6					HD-7					HD-8					HD-9					HD-10					
	Granulocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
	PG: Abbott Cell-Dyn 1700	10	30.93	1.74	25.70-36.10	5.63	10	84.84	1.33	80.80-88.80	1.57	10	41.88	2.09	35.60-48.20	4.99	10	80.51	1.52	75.90-85.10	1.89	10	40.69	2.12	34.30-47.10	5.21
	PG: Abbott Cell-Dyn 1800	6	31.72	2.30	24.80-38.60	7.26	6	85.43	1.98	79.50-91.30	2.31	6	40.75	2.36	33.70-47.90	5.78	6	80.30	1.90	74.60-86.00	2.37	6	41.75	2.41	34.60-49.00	5.77
	PG: CDS Medonic 620	9	25.41	1.92	19.60-31.20	7.56	9	80.42	0.66	78.40-82.40	0.82	9	35.96	3.36	25.90-46.10	9.33	9	77.31	0.51	75.80-78.80	0.66	9	34.76	2.34	27.80-41.80	6.70
	PG: Beckman Coulter AcT Diff/Diff 2	154	30.44	2.19	23.80-37.00	7.20	157	82.35	8.75	56.00-108.50	10.63	154	40.43	2.79	32.00-48.80	6.91	157	76.61	10.88	44.00-109.20	14.20	157	40.64	6.00	22.60-58.60	14.78
	PG: Horiba Micros/45/60	53	38.92	1.42	34.60-43.20	3.65	53	88.51	9.27	60.70-116.30	10.47	53	49.38	1.47	45.00-53.80	2.98	53	84.92	0.63	83.00-86.80	0.74	53	49.35	1.42	45.00-53.60	2.88
	PG: Drew Scientific D3	13	35.58	1.75	30.40-40.90	4.92	13	88.79	0.74	86.60-91.00	0.83	13	44.98	1.61	40.20-49.80	3.58	13	83.56	0.65	81.60-85.50	0.78	13	45.49	1.67	40.50-50.50	3.67
	PG: Abbott Cell-Dyn Emerald	87	36.60	3.93	24.80-48.40	10.74	87	87.42	7.45	65.10-109.80	8.52	87	46.31	5.16	30.80-61.80	11.14	87	82.00	10.80	49.60-114.40	13.17	84	46.44	1.28	42.60-50.20	2.76
	PG: CDS M Series	149	26.23	2.25	19.40-33.00	8.59	149	79.53	9.02	52.40-106.60	11.35	149	34.75	4.09	22.40-47.00	11.79	149	75.86	9.27	48.10-103.70	12.21	147	35.43	2.42	28.10-42.70	6.84
SG: Select CDS Instruments	158	26.18	2.23	19.50-32.90	8.51	158	79.58	8.76	53.30-105.90	11.01	158	34.81	4.05	22.60-46.90	11.64	158	75.94	9.01	48.90-102.90	11.87	156	35.39	2.42	28.10-42.70	6.84	
AG: All Group	478	31.24	4.99	16.20-46.20	15.99	485	83.28	8.98	56.40-110.20	10.78	479	40.98	5.67	24.00-58.00	13.83	485	78.57	9.81	49.20-108.00	12.48	476	41.31	5.39	25.10-57.50	13.05	
Hematocrit	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: Abbott Cell-Dyn 1700	11	51.01	1.38	47.90-54.10	2.71	11	38.92	1.15	36.60-41.20	2.96	11	19.65	0.57	18.40-20.80	2.91	11	32.52	1.27	30.60-34.50	3.91	11	35.05	1.02	33.00-37.20	2.91	
PG: Abbott Cell-Dyn 1800	6	51.12	0.52	48.00-54.20	1.02	6	39.52	0.70	37.10-41.90	1.77	6	20.45	0.54	19.20-21.60	2.65	6	33.68	0.39	31.70-35.70	1.16	6	36.03	1.19	33.80-38.20	3.31	
PG: CDS Medonic 620	10	49.65	0.75	46.70-52.70	1.51	10	38.12	0.57	35.80-40.40	1.50	10	18.23	0.35	17.10-19.30	1.92	10	31.84	0.51	29.90-33.70	1.60	10	33.41	0.42	31.40-35.40	1.26	
PG: Beckman Coulter AcT Diff/Diff 2	166	50.55	1.40	47.50-53.50	2.77	166	38.51	1.71	36.20-40.80	4.44	166	18.90	0.85	17.80-20.00	4.50	166	32.66	1.36	30.70-34.70	4.16	166	34.27	2.06	32.20-36.40	6.01	
PG: Horiba Micros/45/60	53	49.86	1.70	46.90-52.90	3.41	53	35.81	0.78	33.70-37.90	2.18	53	17.02	0.54	16.00-18.00	3.18	52	30.47	0.64	28.70-32.30	2.10	53	31.99	0.58	30.10-33.90	1.81	
SG: Select Abbott Cell Dyn Instruments (3-part Diff)	105	51.80	1.56	48.70-54.90	3.01	105	39.27	1.56	36.90-41.70	3.97	105	19.94	1.31	18.70-21.10	6.58	105	33.18	1.19	31.20-35.20	3.58	104	35.38	1.08	33.30-37.50	3.05	
PG: Drew Scientific D3	13	50.23	1.53	47.20-53.20	3.05	13	38.35	0.94	36.10-40.70	2.45	13	19.32	0.62	18.10-20.50	3.21	13	32.69	0.94	30.70-34.70	2.87	13	34.62	1.32	32.50-36.70	3.82	
PG: Abbott Cell-Dyn Emerald	88	51.95	1.60	48.90-55.10	3.08	88	39.29	1.64	36.90-41.70	4.17	88	19.94	1.40	18.70-21.10	7.04	88	33.22	1.19	31.20-35.20	3.58	87	35.37	1.07	33.30-37.50	3.02	
PG: CDS M Series	152	49.61	3.39	46.60-52.60	6.83	152	38.09	1.24	35.80-40.40	3.25	150	18.14	0.50	17.00-19.20	2.71	152	31.70	1.56	29.80-33.60	4.92	152	33.72	2.61	31.70-35.70	7.74	
SG: Select CDS Instruments	162	49.61	3.29	46.60-52.60	6.63	162	38.09	1.20	35.80-40.40	3.15	160	18.14	0.49	17.00-19.20	2.71	162	31.71	1.51	29.80-33.60	4.76	162	33.70	2.53	31.70-35.70	7.51	
AG: All Group	500	50.43	2.38	47.40-53.40	4.72	500	38.24	1.71	35.90-40.50	4.45	456	18.58	0.82	17.50-19.70	4.41	499	32.23	1.54	30.30-34.10	4.78	499	34.08	2.16	32.10-36.10	6.33	
Hemoglobin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: Abbott Cell-Dyn 1700	11	17.05	0.44	15.90-18.30	2.57	11	13.57	0.28	12.60-14.60	2.06	11	6.75	0.22	6.30-7.30	3.24	11	11.21	0.38	10.40-12.00	3.39	11	12.02	0.26	11.20-12.80	2.17	
PG: Abbott Cell-Dyn 1800	6	17.15	0.35	16.00-18.40	2.03	6	13.58	0.29	12.60-14.60	2.13	6	6.83	0.08	6.30-7.30	1.18	6	11.45	0.16	10.70-12.30	1.39	6	12.08	0.13	11.30-12.90	1.07	
PG: CDS Medonic 620	10	17.09	0.20	15.90-18.30	1.17	10	13.51	0.20	12.60-14.40	1.48	10	6.57	0.13	6.10-7.10	1.97	10	11.16	0.15	10.40-12.00	1.34	10	11.92	0.19	11.10-12.70		



PG: Horiba Micros/45/60	53	17.34	1.30	13.40-21.20	7.51	53	2.56	0.19	2.00-3.20	7.31	53	16.51	1.59	11.70-21.30	9.64	53	4.10	0.30	3.20-5.00	7.32	53	15.07	1.29	11.20-19.00	8.54
PG: Drew Scientific D3	13	9.40	0.72	7.20-11.60	7.66	13	2.17	0.33	1.20-3.20	15.00	13	8.82	1.08	5.60-12.00	12.27	13	2.91	0.29	2.00-3.80	10.00	13	8.52	0.90	5.80-11.20	10.59
PG: Abbott Cell-Dyn Emerald	87	9.60	1.74	4.40-14.80	18.13	87	1.76	0.28	1.00-2.60	15.56	83	8.72	1.36	4.60-12.80	15.63	87	2.75	0.44	1.40-4.00	16.30	87	8.33	1.65	3.40-13.30	19.88
PG: CDS M Series	149	10.80	1.65	5.90-15.80	15.28	148	8.60	0.71	6.50-10.70	8.26	149	12.40	2.08	6.20-18.60	16.77	148	7.99	0.90	5.30-10.70	11.25	149	12.07	1.73	6.90-17.30	14.30
SG: Select CDS Instruments	158	10.85	1.63	6.00-15.80	14.95	157	8.61	0.69	6.50-10.70	8.02	158	12.37	2.09	6.10-18.70	16.85	157	7.99	0.88	5.40-10.60	11.00	158	12.08	1.72	6.90-17.30	14.21
AG: All Group	455	11.36	2.40	4.20-18.60	21.05	398	6.87	2.41	0.00-14.10	34.93	454	11.46	2.79	3.10-19.90	24.26	481	6.10	2.21	0.00-12.70	36.23	466	11.07	2.67	3.10-19.10	24.05
Platelet Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn 1700	11	167.91	14.19	126.00-210.00	8.45	11	524.18	55.39	393.00-655.00	10.57	11	69.18	8.04	52.00-86.00	11.59	11	241.45	27.24	181.00-301.00	11.29	11	284.73	24.63	214.00-356.00	8.63
PG: Abbott Cell-Dyn 1800	6	177.17	7.65	133.00-221.00	4.35	6	536.50	12.53	403.00-671.00	2.33	6	69.67	7.53	53.00-88.00	10.71	6	248.67	7.50	187.00-311.00	3.01	6	299.33	18.75	224.00-374.00	6.25
PG: CDS Medonic 620	10	142.90	6.12	107.00-179.00	4.27	10	467.30	13.19	350.00-584.00	2.83	10	72.30	5.31	54.00-90.00	7.36	10	213.60	4.77	161.00-268.00	2.24	10	277.30	11.26	208.00-346.00	4.08
PG: Beckman Coulter AcT Diff/Diff 2	166	163.70	13.63	123.00-205.00	8.29	166	520.50	23.81	391.00-651.00	4.57	166	75.35	6.78	56.00-94.00	9.07	166	242.42	16.68	182.00-303.00	6.90	166	296.10	22.30	222.00-370.00	7.53
PG: Horiba Micros/45/60	53	159.45	10.70	119.00-199.00	6.73	53	496.75	21.89	373.00-621.00	4.41	53	78.26	6.87	59.00-98.00	8.85	53	242.42	10.62	182.00-303.00	4.38	53	294.68	15.98	221.00-369.00	5.42
SG: Select Abbott Cell Dyn Instruments (3-part Diff)	105	160.94	15.50	121.00-201.00	9.63	105	493.69	38.79	371.00-618.00	7.85	104	72.54	12.08	55.00-91.00	16.58	105	238.72	22.31	179.00-299.00	9.33	105	292.42	38.21	219.00-365.00	13.08
PG: Drew Scientific D3	13	163.69	22.27	123.00-205.00	13.60	13	502.23	26.75	377.00-628.00	5.34	13	75.85	8.92	57.00-95.00	11.71	13	239.85	12.58	180.00-300.00	5.25	13	301.62	19.58	227.00-378.00	6.49
PG: Abbott Cell-Dyn Emerald	88	158.97	15.25	119.00-199.00	9.62	88	486.95	33.86	365.00-609.00	6.96	87	73.16	12.71	55.00-91.00	17.40	88	237.70	22.31	179.00-298.00	9.37	88	292.91	40.55	220.00-366.00	13.86
PG: CDS M Series	149	140.56	15.30	106.00-176.00	10.85	152	453.52	52.27	341.00-568.00	11.52	148	69.97	5.10	53.00-88.00	7.29	152	206.93	26.34	155.00-259.00	12.71	152	275.16	44.40	206.00-344.00	16.15
SG: Select CDS Instruments	159	140.71	14.89	106.00-176.00	10.57	162	454.37	50.82	341.00-568.00	11.19	158	70.11	5.13	53.00-88.00	7.29	162	207.34	25.58	155.00-259.00	12.37	162	275.30	43.08	206.00-344.00	15.67
AG: All Group	497	155.29	17.58	116.00-194.00	11.35	500	490.34	46.22	368.00-613.00	9.43	495	73.41	8.28	55.00-91.00	11.37	500	230.13	26.02	173.00-288.00	11.30	500	288.45	34.60	216.00-360.00	12.01
RDW	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn 1700	9	17.88	0.86	15.30-20.50	4.80	9	17.42	0.77	15.10-19.70	4.43	9	14.27	0.61	12.50-16.10	4.27	9	18.00	0.69	15.90-20.10	3.83	9	14.46	0.61	12.70-16.30	4.21
PG: Abbott Cell-Dyn 1800	5	16.86	0.73	14.70-19.10	4.32	5	16.92	0.72	14.70-19.10	4.26	5	13.26	0.69	11.20-15.40	5.19	5	17.10	0.43	15.80-18.40	2.51	5	14.18	0.48	12.80-15.60	3.38
PG: CDS Medonic 620	9	15.02	0.96	12.10-17.90	6.40	9	14.17	0.69	12.10-16.30	4.86	9	11.99	0.59	10.20-13.80	4.92	9	14.37	0.66	12.40-16.40	4.58	9	12.54	0.62	10.60-14.40	4.96
PG: Beckman Coulter AcT Diff/Diff 2	158	15.03	0.82	12.50-17.50	5.47	158	15.09	0.69	13.00-17.20	4.57	158	12.55	0.51	11.10-14.10	4.05	158	15.35	0.73	13.20-17.60	4.74	158	12.65	0.65	10.60-14.60	5.16
PG: Horiba Micros/45/60	45	13.77	0.46	12.40-15.20	3.33	45	13.39	0.44	12.10-14.70	3.28	45	11.88	0.35	10.90-12.90	2.94	45	13.22	0.39	12.00-14.40	2.95	45	11.83	0.36	10.70-12.90	3.05
SG: Select Abbott Cell Dyn Instruments (3-part Diff)	101	15.21	1.10	11.90-18.50	7.24	101	14.92	1.30	11.00-18.80	8.72	101	12.42	0.85	9.90-14.90	6.85	101	15.15	1.18	11.70-18.70	7.76	101	12.72	0.86	10.10-15.30	6.77
PG: Drew Scientific D3	13	15.13	0.50	13.60-16.60	3.31	13	14.52	0.41	13.30-15.70	2.83	13	12.26	0.58	10.60-14.00	4.72	13	14.71	0.52	13.10-16.30	3.54	13	12.65	0.58	10.90-14.30	4.60
PG: Abbott Cell-Dyn Emerald	87	14.84	0.51	13.30-16.30	3.45	87	14.55	0.92	11.80-17.40	6.30	87	12.18	0.57	10.50-13.90	4.67	87	14.75	0.58	12.90-16.50	4.01	87	12.46	0.56	10.80-14.20	4.48
PG: CDS M Series	148	15.05	1.15	11.70-18.60	7.62	148	14.29	1.74	9.10-19.50	12.17	147	11.91	0.78	9.60-14.20	6.55	148	14.26	0.82	11.80-16.80	5.73	148	12.63	1.31	8.70-16.50	10.40
SG: Select CDS Instruments	157	15.05	1.14	11.70-18.50	7.55	157	14.28	1.70	9.20-19.40	11.89	156	11.91	0.77	9.60-14.20	6.47	157	14.27	0.81	11.90-16.70	5.66	157	12.63	1.28	8.80-16.40	10.16
AG: All Group	475	14.97	1.06	11.80-18.20	7.07	475	14.62	1.34	10.60-18.60	9.18	474	12.25	0.74	10.00-14.40	6.07	475	14.74	1.10	11.40-18.00	7.48	475	12.59	0.96	9.70-15.50	7.62
Red Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn 1700	11	6.19	0.15	5.82-6.56	2.37	11	4.68	0.11	4.40-4.96	2.24	11	2.12	0.04	1.99-2.25	2.03	11	4.33	0.17	4.07-4.59	4.00	11	3.76	0.07	3.53-3.99	1.78
PG: Abbott Cell-Dyn 1800	6	5.93	0.16	5.57-6.29	2.70	6	4.53	0.09	4.26-4.80	1.88	6	2.12	0.05	1.99-2.25	2.31	6	4.29	0.09	4.03-4.55	2.05	6	3.72	0.11	3.50-3.94	3.06
PG: CDS Medonic 620	10	6.28	0.07	5.90-6.66	1.11	10	4.65	0.04	4.37-4.93	0.95	10	2.06	0.03	1.94-2.18	1.41	10	4.33	0.05	4.07-4.59	1.25	10	3.74	0.06	3.52-3.96	1.52
PG: Beckman Coulter AcT Diff/Diff 2	166	6.22	0.15	5.85-6.59	2.46	166	4.66	0.20	4.38-4.94	4.31	166	2.07	0.09	1.95-2.19	4.35	166	4.36	0.17	4.10-4.62	4.01	166	3.72	0.23	3.50-3.94	6.18
PG: Horiba Micros/45/60	53	6.21	0.19	5.84-6.58	3.11	53	4.56	0.08	4.29-4.83	1.80	53	1.99	0.06	1.87-2.11	2.96	52	4.29	0.08	4.03-4.55	1.77	53	3.65	0.06	3.43-3.87	1.62
SG: Select Abbott Cell Dyn Instruments (3-part Diff)	105	6.03	0.17	5.67-6.39	2.90	105	4.52	0.18	4.25-4.79	4.03	105	2.04	0.14	1.92-2.16	6.86	105	4.25	0.14	4.00-4.50	3.39	104	3.64	0.12	3.42-3.86	3.41
PG: Drew Scientific D3	13	6.06	0.13	5.70-6.42	2.15	13	4.55	0.07	4.28-4.82	1.49	13	2.04	0.05	1.92-2.16	2.25	13	4.29	0.09	4.03-4.55	2.03	13	3.67	0.09	3.45-3.89	2.32
PG: Abbott Cell-Dyn Emerald	88	6.01	0.17	5.65-6.37	2.80	88	4.50	0.19	4.23-4.77	4.11	88	2.03	0.15	1.91-2.15	7.24	88	4.24	0.14	3.99-4.49	3.30	87	3.61	0.12	3.39-3.83	3.30
PG: CDS M Series	148	6.27	0.17	5.89-6.65	2.74	152	4.61	0.13	4.33-4.89	2.86	150	2.04	0.04	1.92-2.16	1.96	151	4.29	0.17	4.03-4.55	4.08	152	3.76	0.28	3.53-3.99	7.55
SG: Select CDS Instruments	158	6.27	0.17	5.89-6.65	2.66	162	4.61	0.13	4.33-4.89	2.80	160	2.04	0.04	1.92-2.16	1.96	161	4.29	0.17	4.03-4.55	3.99	162	3.76	0.28	3.53-3.99	7.34
AG: All Group	496	6.19	0.19	5.82-6.56	3.07	500	4.60	0.17	4.32-4.88	3.70	498	2.04	0.09	1.92-2.16	4.46	498	4.30	0.16	4.04-4.56	3.79	499	3.71	0.22	3.49-3.93	5.93
White Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn 1700	11	5.48	0.26	4.70-6.30	4.73	11	25.74	1.26	21.80-29.60	4.90	11	2.84	0.13	2.40-3.20	4.64	11	16.36	0.83	13.90-18.90	5.06	11	7.95	0.42	6.70-9.10	5.32
PG: Abbott Cell-Dyn 1800	6	4.92	0.23	4.20-5.60	4.69	6	24.42	1.19	20.70-28.10	4.88	6	2.57	0.15	2.20-3.00	5.77	6	15.23	0.69	12.90-17.50	4.54	6	6.97	0.22	6.00-8.10	3.14
PG: CDS Medonic 620	10	4.93	0.15	4.20-5.60	3.06	10	25.32	0.42	21.50-29.10	1.66	10	2.55	0.10	2.20-3.00	3.85	10	15.57	0.44	13.30-17.90	2.82	10	7.08	0.27	6.00-8.20	3.8



PG: Microhematocrit, Spun	16	45.31	3.83	33.80-56.80	8.43	9	32.89	1.17	29.40-36.40	3.56															
SG: Select Microhematocrit Instruments	25	47.76	5.91	30.10-65.50	12.36	9	32.89	1.17	29.40-36.40	3.56															
AG: All Group	26	47.88	5.83	30.40-65.40	12.17	9	32.89	1.17	29.40-36.40	3.56															
Hemoglobin (Waived)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: HemoCue	40	15.51	2.09	9.20-21.80	13.48																				
PG: HemoCue 201	100	15.94	1.81	10.50-21.30	11.38																				
PG: Alere HemoPoint H2	38	16.87	1.89	11.20-22.60	11.18	5	12.40	0.95	9.60-15.30	7.66															
SG: Select Waived Hemoglobin Instruments	180	16.01	2.00	10.00-22.00	12.50																				
AG: All Group	183	16.02	1.99	10.00-22.00	12.44																				
608 - Hemocue Glucose/Hemoglobin	HQ-2																								
Glucose Whole Blood (Hemocue)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: HemoCue	40	321.93	13.37	282.00-362.00	4.16																				
PG: HemoCue 201	149	319.42	23.19	249.00-389.00	7.27																				
Hemoglobin (Hemocue)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: HemoCue	25	15.30	0.66	13.30-17.30	4.31																				
PG: HemoCue 201	83	15.19	0.40	14.00-16.40	2.63																				
606 - Hemocue WBC	HQW-6					HQW-7					HQW-8					HQW-9					HQW-10				
White Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Hemocue WBC System	11	3.02	0.18	2.50-3.50	6.00	11	7.70	0.37	6.50-8.90	4.81	11	22.57	0.72	19.20-26.00	3.19	11	2.93	0.23	2.50-3.30	7.93	11	7.69	0.34	6.50-8.90	4.42
700 - i-Stat Chemistry;721 - i-Stat Chemistry & Blood Gases;722 - i-Stat Chemistry - Waived;723 - i-Stat Blood Gases	IST-6					IST-7					IST-8					IST-9					IST-10				
Blood Gas - pCO2	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat	8	49.31	2.26	44.00-54.00	4.69	9	70.51	3.52	65.00-77.00	4.93	9	42.42	1.61	37.00-47.00	3.81	9	45.58	1.89	41.00-51.00	4.13	8	49.29	1.03	44.00-54.00	2.04
SG: International Abbott i-Stat Instruments	12	49.38	1.93	44.00-54.00	3.88	13	70.89	3.06	65.00-77.00	4.37	13	42.68	1.46	38.00-48.00	3.49	13	45.32	1.66	40.00-50.00	3.78	12	49.53	0.99	45.00-55.00	2.00
Blood Gas - pH	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat	8	7.32	0.01	7.28-7.36	0.07	9	6.93	0.01	6.89-6.97	0.10	9	7.29	0.01	7.25-7.33	0.12	9	7.31	0.01	7.27-7.35	0.10	8	7.32	0.01	7.28-7.36	0.08
SG: International Abbott i-Stat Instruments	12	7.31	0.01	7.27-7.35	0.07	13	6.92	0.01	6.88-6.96	0.10	13	7.29	0.01	7.25-7.33	0.11	13	7.31	0.01	7.27-7.35	0.10	12	7.31	0.01	7.27-7.35	0.07
Blood Gas - pO2	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat	6	198.50	9.22	171.00-227.00	4.62	8	60.00	3.02	51.00-69.00	5.00	8	78.00	4.04	66.00-90.00	5.13	9	170.56	9.89	141.00-201.00	5.79	9	197.22	12.81	159.00-235.00	6.50
SG: International Abbott i-Stat Instruments	10	201.10	8.70	175.00-227.00	4.33	12	58.33	3.77	47.00-69.00	6.55	12	76.83	3.90	65.00-89.00	5.06	13	171.00	9.05	144.00-198.00	5.26	13	198.46	11.36	164.00-232.00	5.76
Chloride (CL-)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat / Thermal	6	91.83	0.75	87.00-97.00	0.87	6	111.67	0.82	106.10-117.30	0.73	6	79.67	0.52	76.00-84.00	0.65	6	97.17	0.75	92.30-102.10	0.82	6	91.50	0.55	87.00-97.00	0.60
SG: International Abbott i-Stat Instruments / Thermal	25	91.60	0.76	87.00-97.00	0.87	10	112.10	0.88	106.50-117.70	0.80	10	79.60	0.52	76.00-84.00	0.65	10	97.10	0.57	92.20-102.00	0.62	10	91.50	0.53	87.00-97.00	0.58
PG: Abbott i-Stat (Waived) / Thermal	15	91.60	0.83	87.00-97.00	0.91																				
CO2 (Total CO2 Bicarb)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat / Thermal	6	28.17	0.41	27.00-29.40	1.45	7	18.14	1.07	14.90-21.30	5.91	7	23.43	0.98	20.50-26.30	4.19	7	25.14	1.68	20.10-30.10	6.69	7	27.86	1.35	23.80-31.90	4.84
SG: International Abbott i-Stat Instruments / Thermal	25	28.12	1.01	25.10-31.10	3.59	11	18.27	0.91	15.60-21.00	4.92	11	23.55	0.82	21.00-26.00	3.49	11	25.45	1.37	21.40-29.60	5.37	11	27.91	1.14	24.50-31.30	4.09
PG: Abbott i-Stat (Waived) / Thermal	15	28.13	1.25	24.40-31.90	4.45																				
Glucose	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat / Thermal	9	86.22	1.48	77.60-94.80	1.74	9	36.00	0.71	30.00-42.00	1.97	9	49.89	0.78	44.00-56.00	1.60	9	195.22	1.86	176.00-215.00	0.97	9	86.11	1.27	77.00-95.00	1.51
SG: International Abbott i-Stat Instruments / Thermal	28	86.20	1.94	77.60-94.80	2.25	13	35.87	0.76	30.00-42.00	2.22	13	49.65	0.87	44.00-56.00	1.80	13	194.62	1.84	176.00-215.00	0.95	13	85.81	1.22	77.00-95.00	1.42
PG: Abbott i-Stat (Waived) / Thermal	15	86.47	2.33	77.80-95.20	2.69																				
Hematocrit	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%										



PG: Abbott i-Stat	5	24.63	1.15	21.20-28.10	4.67	5	14.33	0.67	12.30-16.30	4.69	5	20.25	0.34	19.30-21.30	1.67	5	22.70	0.77	20.40-25.00	3.39	5	24.80	0.36	23.70-25.90	1.45
SG: International Abbott i-Stat Instruments	7	24.77	0.77	22.50-27.10	3.10	8	14.51	0.52	12.90-16.10	3.59	8	20.54	0.43	19.20-21.80	2.10	8	22.48	0.73	20.30-24.70	3.24	7	25.10	0.44	23.80-26.40	1.75
BEecf	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat / Thermal	5	-2.50	0.71	0.00--1.00	-23.33	5	-18.00	0.82	0.00--16.00	-4.44	5	-6.25	0.50	0.00--5.00	-8.33	5	-3.50	0.58	0.00--2.00	-15.00	5	-1.33	0.58	0.00-1.00	-60.00
sO2	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat	5	100.00	0.00	100.00-100.00	0.00	5	69.50	3.00	61.00-79.00	4.29	5	93.50	1.00	91.00-97.00	1.06	5	99.25	0.50	98.00-101.00	0.51	5	99.67	0.58	98.00-102.00	0.60
Lactate	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat	5	0.72	0.00	0.72-0.72	0.00	5	5.16	0.09	4.88-5.44	1.78	5	6.49	0.08	6.26-6.72	1.20	5	2.34	0.11	2.02-2.66	4.53	5	0.78	0.05	0.63-0.93	6.28
PG: Intl Abbott i-Stat	5	0.79	0.02	0.72-0.86	2.78	5	5.12	0.03	5.04-5.20	0.49	5	6.49	0.04	6.37-6.61	0.60	5	2.30	0.02	2.25-2.35	0.70	5	0.78	0.07	0.58-0.98	8.46
Anion Gap	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat (Waived) / Thermal	8	23.38	1.69	18.00-28.00	7.39																				
SG: Select Abbott i-Stat Instruments / Thermal	9	24.22	2.99	15.00-33.00	12.50																				
Creatinine	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat / Thermal	12	2.47	0.11	2.10-2.90	4.40	12	6.97	0.23	6.00-8.10	3.29	12	0.31	0.05	0.00-0.60	16.67	12	4.64	0.45	3.90-5.30	9.78	12	2.47	0.09	2.10-2.90	3.60
SG: International Abbott i-Stat Instruments / Thermal	35	2.49	0.09	2.10-2.90	3.60	17	6.96	0.25	6.00-8.10	3.57	17	0.32	0.05	0.00-0.60	16.67	17	4.65	0.38	4.00-5.40	8.09	17	2.47	0.08	2.10-2.90	3.20
PG: Abbott i-Stat (Waived) / Thermal	18	2.49	0.05	2.10-2.90	2.00																				
716 - Blood Lead - Waived Methods;709 - Blood Lead	LE-6					LE-7					LE-8					LE-9					LE-10				
Blood Lead (Waived)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: ESA LeadCare II (Waived)	49	5.02	0.73	2.80-7.20	14.60																				
662 - Neonatal Bilirubin (5 specimens);663 - Neonatal Bilirubin Add-On	NB-6					NB-7					NB-8					NB-9					NB-10				
Neonatal Bilirubin (Total)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Reichert (Leica) Unistat Bilirubinometer	25	7.39	0.23	5.90-8.90	3.11	25	16.99	0.39	13.60-20.40	2.29	25	11.68	0.29	9.40-14.00	2.56	25	18.60	1.26	14.90-22.30	6.77					
AG: All Group	33	7.39	0.27	5.90-8.90	3.65	33	16.78	0.66	13.40-20.20	3.93	31	11.48	0.67	9.20-13.80	5.83	31	18.46	1.22	14.80-22.20	6.59	6	0.15	0.06	0.00-0.60	25.00
748 - i-Stat Prottime	PTI-6					PTI-7					PTI-8					PTI-9					PTI-10				
Prottime - Whole Blood	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat Prottime	5	28.58	1.09	24.30-32.90	3.81	5	14.18	1.35	12.10-16.30	9.51	5	14.80	0.58	12.60-17.00	3.92	5	26.74	3.43	22.70-30.70	12.85	5	15.23	0.82	12.90-17.50	5.39
PG: Abbott Coagusense	5	37.70	6.33	32.00-43.40	16.76	5	19.17	7.42	16.30-22.10	38.65	5	23.80	15.16	20.20-27.40	63.70	5	42.23	5.44	35.90-48.50	12.89	5	14.67	2.47	12.50-16.90	16.80
SG: International Abbott i-Stat Instruments	6	28.22	1.32	24.00-32.40	4.68	6	14.12	1.21	12.00-16.20	8.58	6	14.82	0.52	12.60-17.00	3.51	6	27.32	3.38	23.20-31.40	12.38	5	14.92	0.99	12.70-17.10	6.64
SG: International Abbott Coagusense Instruments	5	37.66	5.48	32.00-43.40	14.54	5	18.35	6.28	15.60-21.20	34.13	5	23.80	15.16	20.20-27.40	63.70	5	40.95	5.13	34.90-47.10	12.51	5	14.67	2.47	12.50-16.90	16.80
INR	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat Prottime	6	2.47	0.10	2.20-2.80	4.00	6	1.22	0.12	0.80-1.60	10.00	6	1.28	0.04	1.20-1.40	3.08	6	2.32	0.26	1.50-3.10	11.30	5	1.30	0.07	1.10-1.50	5.38
PG: Abbott Coagusense	5	3.08	0.46	1.80-4.50	14.52	5	1.27	0.06	1.10-1.50	4.62	5	1.30	0.00	1.30-1.30	0.00	5	3.35	0.33	2.40-4.40	9.71	5	1.28	0.15	0.90-1.80	11.54
SG: International Abbott i-Stat Instruments	11	2.48	0.14	2.10-2.90	5.60	11	1.24	0.10	0.90-1.50	8.33	11	1.26	0.05	1.10-1.50	3.85	11	2.42	0.23	1.70-3.10	9.58	10	1.27	0.07	1.10-1.50	5.38
SG: International Abbott Coagusense Instruments	6	3.07	0.41	1.90-4.30	13.23	5	1.28	0.05	1.10-1.50	3.85	5	1.30	0.00	1.30-1.30	0.00	5	3.28	0.33	2.30-4.30	10.00	5	1.28	0.15	0.90-1.80	11.54
754 - Infectious Mono/RF Combo;653 - Rheumatoid Factor	RF-6 N					RF-7 N					RF-8 N					RF-9 N					RF-10 N				
Rheumatoid Factor (Quant)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	5	69.07	10.57	37.40-100.80	15.30	5	9.87	5.00	0.00-24.90	50.61	5	124.97	9.86	95.40-154.60	7.89	5	69.40	9.55	40.80-98.10	13.76	5	10.07	5.00	0.00-25.10	49.50
686 - Special Chemistry;667 - PSA Add-on	SC-3					SC-4																			
CEA	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	6	9.82	0.82	7.36-12.28	8.34	6	19.96	1.56	15.29-24.63	7.81															
DHEA-S	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	5	0.20	0.03	0.10-0.30	15.00	5	0.32	0.03	0.20-0.40	10.00															
Estradiol	Labs</																								



PG: OCD Vitros ECI/ECIQ	5	34.93	1.16	31.40-38.40	3.32	5	68.50	3.21	58.90-78.10	4.69															
AG: All Group	8	20.54	3.35	10.40-30.60	16.34	8	36.85	6.05	18.80-55.00	16.40															
Progesterone	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Beckman Coulter Instruments (Chem)	5	12.93	0.56	11.20-14.60	4.34	5	22.03	1.68	17.00-27.00	7.64															
AG: All Group	11	14.17	3.67	3.20-25.20	25.85	11	24.09	5.69	7.00-41.20	23.61															
Prolactin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	9	16.68	1.81	11.30-22.10	10.84	9	30.18	3.38	20.10-40.30	11.19															
PSA	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman ACCESS	19	7.99	0.40	6.78-9.20	5.03	19	15.35	0.94	12.53-18.17	6.13															
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	11	6.61	0.33	5.62-7.60	5.01	11	13.10	0.56	11.43-14.77	4.25															
PG: Qualigen FastPack PSA	8	9.09	1.18	5.56-12.62	12.93	8	20.26	2.41	13.03-27.49	11.90															
PG: TOSOH AIA	9	6.07	0.28	5.22-6.92	4.68	9	11.35	0.58	9.62-13.08	5.08															
SG: Select Beckman Coulter Instruments (Chem)	24	7.94	0.39	6.77-9.11	4.92	24	15.17	0.97	12.26-18.08	6.39															
SG: Select Siemens Instruments	14	6.56	0.41	5.33-7.79	6.25	14	13.05	0.68	11.01-15.09	5.21															
SG: Select OCD Vitros Instruments	6	3.90	0.18	3.37-4.43	4.56	6	7.23	0.35	6.18-8.28	4.83															
PG: Roche Cobas e Series	5	7.99	0.31	7.07-8.91	3.83	5	15.37	0.32	14.40-16.34	2.11															
SG: Select Abbott Architect Instruments	10	6.12	0.34	5.11-7.13	5.49	10	11.86	0.51	10.32-13.40	4.33															
PG: NanoEnTek Frend System	6	7.32	1.37	3.20-11.44	18.74	6	15.13	3.18	5.60-24.66	20.99															
Vitamin B12	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman ACCESS	9	423.33	34.31	320.00-526.00	8.11	9	704.67	59.34	527.00-883.00	8.41															
SG: Select Beckman Coulter Instruments (Chem)	14	431.71	32.67	334.00-530.00	7.57	14	733.57	64.04	542.00-926.00	8.72															
SG: Select OCD Vitros Instruments	5	539.00	49.49	391.00-688.00	9.18	5	820.60	63.99	629.00-1013.00	7.80															
PG: Roche Cobas e Series	5	514.75	32.61	417.00-613.00	6.33	5	977.00	86.85	717.00-1237.00	8.88															
SG: Select Abbott Architect Instruments	10	521.70	31.94	426.00-618.00	6.11	10	918.20	49.37	770.00-1066.00	5.38															
AG: All Group	39	496.79	61.35	313.00-681.00	12.35	39	844.56	111.29	511.00-1179.00	13.17															
Testosterone	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman ACCESS	14	4.89	0.26	4.10-5.70	5.31	14	8.49	0.47	7.10-9.90	5.53															
SG: Select Beckman Coulter Instruments (Chem)	19	4.77	0.32	3.80-5.80	6.67	19	8.23	0.61	6.40-10.00	7.44															
SG: Select Abbott Architect Instruments	5	5.12	0.11	4.80-5.40	2.16	5	9.49	0.28	8.70-10.30	2.95															
PG: NanoEnTek Frend System	5	10.76	2.98	1.90-19.70	27.59	5	13.66	2.32	6.70-20.70	16.93															
AG: All Group	39	5.37	0.85	2.90-8.00	15.74	39	9.02	1.28	5.20-12.80	14.22															
Transferrin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	5	159.03	9.21	131.40-186.60	5.79	5	196.10	8.01	172.10-220.10	4.08															
685 - Special Chemistry II	SPC-3					SPC-4																			
Insulin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Siemens Immulite Instruments	5	22.00	1.41	17.80-26.20	6.41	5	17.70	1.84	12.20-23.20	10.40															
AG: All Group	5	60.80	4.02	48.70-72.90	6.61	5	35.10	5.39	18.90-51.30	15.36															
PTH	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: TOSOH AIA	5	49.30	0.00	49.30-49.30	0.00	5	16.60	0.00	16.60-16.60	0.00															
SG: Select Beckman Coulter Instruments (Chem)	5	42.03	11.39	7.80-76.20	27.12	5	15.78	7.80	0.00-39.20	49.37															
PG: Abbott Architect i1000 SR	5	80.00	0.00	80.00-80.00	0.00	5	19.60	0.00	19.60-19.60	0.00															
AG: All Group	9	48.20	14.82	3.70-92.70	30.75	9	19.24	8.46	0.00-44.60	44.06															
Vitamin D	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman ACCESS	15	59.80	5.78	42.45-77.15	9.67	15	30.80	3.41	20.57-41.03	11.07															
PG: TOSOH AIA	6	66.85	4.97	51.92-81.77	7.44	6	38.78	3.13	29.38-48.18	8.08															



PG: Sysmex XN Series	18	0.75	0.81	0.00-3.20	101.25	20	0.65	0.16	0.20-1.20	22.86	24	0.70	0.91	0.00-3.40	128.57	24	1.75	2.60	0.00-9.60	144.44	23	1.63	2.31	0.00-8.50	144.38
Eosinophil	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%
PG: Sysmex XS-1000i	14	15.84	0.50	14.30-17.30	3.16	14	14.90	0.57	13.20-16.60	3.83	14	15.99	0.51	14.50-17.50	3.13	14	14.85	0.51	13.40-16.40	3.42	14	13.62	0.69	11.50-15.70	5.07
SG: Select Sysmex Instruments (5-part Diff)	35	15.83	0.61	14.00-17.60	3.86	35	15.26	0.81	12.90-17.70	5.29	37	16.30	1.84	10.80-21.80	11.29	35	15.05	0.68	13.00-17.00	4.53	36	13.66	1.01	10.70-16.70	7.37
PG: Sysmex XN Series	20	15.86	0.68	13.90-17.90	4.28	20	15.63	0.67	13.60-17.60	4.23	22	16.53	2.34	9.50-23.50	14.18	20	15.28	0.60	13.50-17.10	3.92	21	13.74	1.18	10.20-17.20	8.61
Basophil	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%
PG: Sysmex XS-1000i	14	9.86	0.81	7.44-12.28	8.19	14	9.55	0.99	6.59-12.51	10.35	14	9.66	0.79	7.30-12.02	8.15	14	9.33	0.78	7.00-11.66	8.32	14	7.94	0.86	5.37-10.51	10.79
SG: Select Sysmex Instruments (5-part Diff)	33	9.51	0.81	7.07-11.95	8.55	33	9.04	0.91	6.30-11.78	10.09	33	9.64	0.80	7.25-12.03	8.28	33	9.04	0.70	6.94-11.14	7.75	33	7.59	0.78	5.25-9.93	10.26
PG: Sysmex XT-4000i	5	86.10	0.00	86.10-86.10	0.00	5	88.10	0.00	88.10-88.10	0.00	5	87.10	0.00	87.10-87.10	0.00	5	89.70	0.00	89.70-89.70	0.00	5	72.80	0.00	72.80-72.80	0.00
PG: Sysmex XN Series	19	9.26	0.74	7.05-11.47	7.96	19	8.66	0.65	6.72-10.60	7.48	19	9.63	0.83	7.15-12.11	8.59	19	8.83	0.57	7.12-10.54	6.47	19	7.34	0.62	5.48-9.20	8.45
Red Blood Cell Count	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%
PG: Sysmex XS-1000i	14	2.20	0.03	2.07-2.33	1.55	14	5.74	0.06	5.40-6.08	1.01	14	2.21	0.03	2.08-2.34	1.18	14	5.76	0.06	5.41-6.11	0.99	14	4.64	0.04	4.36-4.92	0.84
SG: Select Sysmex Instruments (5-part Diff)	40	2.19	0.03	2.06-2.32	1.55	40	5.80	0.08	5.45-6.15	1.34	40	2.20	0.03	2.07-2.33	1.55	40	5.77	0.20	5.42-6.12	3.52	40	4.69	0.19	4.41-4.97	3.99
PG: Sysmex XN Series	25	2.18	0.03	2.05-2.31	1.33	25	5.84	0.06	5.49-6.19	1.10	25	2.19	0.03	2.06-2.32	1.37	25	5.77	0.26	5.42-6.12	4.42	25	4.72	0.23	4.44-5.00	4.87
Hemoglobin	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%
PG: Sysmex XS-1000i	14	5.66	0.09	5.30-6.10	1.58	14	17.73	0.18	16.50-18.90	1.02	14	5.69	0.10	5.30-6.10	1.58	14	17.73	0.19	16.50-18.90	1.07	14	13.20	0.10	12.30-14.10	0.76
SG: Select Sysmex Instruments (5-part Diff)	40	5.70	0.07	5.30-6.10	1.23	40	17.67	0.14	16.50-18.90	0.79	40	5.72	0.07	5.30-6.10	1.23	40	17.57	0.76	16.40-18.80	4.26	40	13.34	0.68	12.40-14.20	5.11
PG: Sysmex XN Series	25	5.72	0.06	5.30-6.10	1.05	25	17.65	0.10	16.40-18.80	0.57	25	5.74	0.05	5.30-6.10	0.88	25	17.48	0.94	16.30-18.70	5.37	25	13.42	0.85	12.50-14.30	6.34
Hematocrit	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%
PG: Sysmex XS-1000i	14	17.83	0.29	16.70-18.90	1.63	14	53.76	0.67	50.60-57.00	1.25	14	17.90	0.24	16.80-19.00	1.34	14	53.96	0.63	50.80-57.20	1.17	14	39.89	0.46	37.50-42.30	1.15
SG: Select Sysmex Instruments (5-part Diff)	40	17.57	0.47	16.50-18.70	2.67	40	53.45	0.95	50.30-56.70	1.78	40	17.62	0.47	16.50-18.70	2.67	39	53.13	2.46	49.90-56.30	4.63	39	40.06	2.29	37.70-42.50	5.71
PG: Sysmex XN Series	25	17.37	0.40	16.40-18.40	2.30	25	53.19	0.94	50.00-56.40	1.77	25	17.40	0.38	16.40-18.40	2.18	24	52.54	2.95	49.40-55.60	5.62	24	40.09	2.89	37.70-42.50	7.21
MCV	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%
PG: Sysmex XS-1000i	14	81.05	0.79	78.70-83.50	0.97	14	93.74	0.77	91.40-96.00	0.82	14	81.04	0.68	79.00-83.00	0.84	14	93.41	1.37	89.30-97.50	1.47	14	85.97	0.79	83.60-88.40	0.92
SG: Select Sysmex Instruments (5-part Diff)	40	80.20	1.35	76.20-84.30	1.68	40	92.15	1.75	87.00-97.50	1.90	40	80.15	1.35	76.20-84.30	1.68	40	91.90	1.97	86.00-97.80	2.14	40	85.32	1.71	80.20-90.40	2.00
PG: Sysmex XN Series	25	79.62	1.25	75.80-83.30	1.57	25	91.16	1.37	87.10-95.30	1.50	25	79.54	1.24	75.80-83.20	1.56	25	90.94	1.63	86.00-95.80	1.79	25	84.88	1.95	79.10-90.80	2.30
Platelet Count	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%
PG: Sysmex XS-1000i	14	58.71	1.73	44.00-74.00	2.88	14	386.14	7.42	290.00-483.00	1.92	14	57.21	3.26	43.00-71.00	5.79	14	387.50	10.98	291.00-485.00	2.84	14	209.36	7.16	157.00-261.00	3.44
SG: Select Sysmex Instruments (5-part Diff)	40	55.80	3.88	42.00-70.00	6.96	40	404.67	17.56	304.00-506.00	4.35	40	55.50	3.36	42.00-70.00	6.07	40	401.95	36.93	302.00-503.00	9.18	40	217.55	27.85	164.00-273.00	12.80
PG: Sysmex XN Series	25	53.80	3.20	41.00-68.00	5.93	25	415.04	12.60	311.00-519.00	3.04	25	54.24	2.62	41.00-68.00	4.81	25	409.64	44.31	308.00-513.00	10.80	25	222.24	34.21	167.00-278.00	15.41
RDW	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	L									



PG: Sysmex K-800,1000,4500,KX-21	15	16.91	0.36	15.90-17.90	2.13	15	48.07	0.82	45.20-51.00	1.70	15	16.97	0.33	16.00-18.00	1.94	15	47.86	1.35	45.00-50.80	2.82	15	32.61	0.71	30.60-34.60	2.18
PG: Sysmex pochH-100i	25	17.90	0.46	16.80-19.00	2.57	25	50.58	1.31	47.60-53.60	2.59	25	17.78	0.60	16.70-18.90	3.37	25	49.98	3.58	47.00-53.00	7.16	24	34.69	0.90	32.60-36.80	2.59
SG: Select Sysmex Instruments (3-part Diff)	124	17.17	0.51	16.20-18.20	2.97	124	48.43	1.44	45.50-51.30	2.95	123	17.10	0.55	16.10-18.10	3.22	124	48.14	3.48	45.20-51.00	7.23	123	33.05	1.82	31.00-35.00	5.52
PG: Sysmex XP-300	84	16.99	0.31	16.00-18.00	1.82	84	47.86	0.84	45.00-50.80	1.75	83	16.93	0.39	15.90-17.90	2.31	84	47.64	3.55	44.70-50.50	7.46	84	32.66	1.90	30.70-34.70	5.81
SG: International Sysmex Instruments	29	17.89	0.44	16.80-19.00	2.46	29	50.51	1.24	47.50-53.50	2.46	29	17.78	0.57	16.70-18.90	3.20	29	50.01	3.33	47.00-53.00	6.64	28	34.61	0.85	32.50-36.70	2.46
Mono/Mixed	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex K-800,1000,4500,KX-21	15	18.73	1.72	13.50-23.90	9.20	15	14.48	0.99	11.50-17.50	6.83	15	19.08	1.69	14.00-24.20	8.85	15	13.88	0.69	11.80-16.00	4.96	15	16.49	1.21	12.90-20.10	7.33
PG: Sysmex pochH-100i	24	18.11	1.89	12.40-23.80	10.44	24	13.03	0.69	10.90-15.10	5.31	24	18.03	1.70	12.90-23.10	9.44	24	12.59	0.93	9.80-15.40	7.38	24	14.70	1.20	11.10-18.30	8.16
SG: Select Sysmex Instruments (3-part Diff)	121	18.82	1.88	13.20-24.40	10.00	122	13.96	0.94	11.20-16.80	6.71	122	18.55	1.79	13.20-24.00	9.62	122	13.93	1.24	10.20-17.60	8.92	122	15.96	1.37	11.90-20.10	8.56
PG: Sysmex XP-300	82	19.05	1.87	13.40-24.60	9.84	83	14.14	0.82	11.60-16.60	5.82	83	18.61	1.82	13.10-24.10	9.78	83	14.33	1.13	10.90-17.70	7.90	83	16.23	1.23	12.50-19.90	7.59
SG: International Sysmex Instruments	28	18.07	1.89	12.40-23.80	10.44	28	13.06	0.68	11.10-15.10	5.19	28	18.04	1.65	13.10-22.90	9.11	28	12.61	0.87	10.00-15.20	6.90	27	14.76	1.22	11.10-18.50	8.24
MCV	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex K-800,1000,4500,KX-21	15	70.69	1.16	67.20-74.20	1.64	15	85.74	1.06	82.50-88.90	1.24	15	70.68	1.19	67.10-74.30	1.68	15	85.81	1.06	82.60-89.00	1.24	15	78.39	1.22	74.70-82.10	1.56
PG: Sysmex pochH-100i	20	73.86	1.13	70.50-77.30	1.53	20	88.82	1.39	84.60-93.00	1.57	20	73.78	1.18	70.30-77.30	1.60	20	88.57	2.37	81.50-95.70	2.67	20	82.15	1.54	77.60-86.80	1.87
SG: Select Sysmex Instruments (3-part Diff)	118	71.68	1.37	67.60-75.80	1.91	118	86.06	5.22	70.40-101.80	6.06	118	71.67	1.54	67.10-76.30	2.15	118	86.41	2.09	80.10-92.70	2.42	118	79.41	1.82	73.90-84.90	2.29
PG: Sysmex XP-300	83	71.33	0.83	68.80-73.80	1.16	83	85.46	6.00	67.50-103.50	7.02	83	71.34	1.20	67.70-74.90	1.68	83	85.99	1.82	80.50-91.50	2.12	83	78.94	1.30	75.00-82.80	1.65
SG: International Sysmex Instruments	24	73.88	1.10	70.60-77.20	1.49	24	88.92	1.32	84.90-92.90	1.48	24	73.86	1.18	70.40-77.40	1.60	24	88.66	2.19	82.10-95.30	2.47	24	82.10	1.42	77.80-86.40	1.73
Platelet Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex K-800,1000,4500,KX-21	15	67.60	4.53	51.00-85.00	6.62	15	398.87	17.00	299.00-499.00	4.26	15	66.20	4.83	50.00-83.00	7.27	15	399.93	14.44	300.00-500.00	3.60	15	196.67	11.83	148.00-246.00	5.99
PG: Sysmex pochH-100i	25	66.36	4.40	50.00-83.00	6.67	25	381.96	15.02	287.00-478.00	3.93	25	66.00	6.00	50.00-83.00	9.09	25	373.60	44.66	281.00-468.00	11.95	25	199.60	40.36	150.00-250.00	20.20
SG: Select Sysmex Instruments (3-part Diff)	122	66.42	4.49	50.00-83.00	6.82	121	393.31	19.93	295.00-491.00	5.06	122	66.57	11.70	50.00-84.00	17.46	121	390.17	41.91	293.00-488.00	10.74	122	195.91	26.32	147.00-245.00	13.42
PG: Sysmex XP-300	82	66.22	4.53	50.00-83.00	6.82	81	395.78	20.63	297.00-495.00	5.20	82	66.82	13.77	50.00-84.00	20.60	81	393.48	43.41	295.00-491.00	11.04	82	194.65	22.84	146.00-244.00	11.69
SG: International Sysmex Instruments	29	65.72	4.38	50.00-83.00	6.67	29	379.31	16.40	284.00-474.00	4.33	29	65.62	5.66	50.00-83.00	8.64	29	372.41	42.46	279.00-465.00	11.42	28	198.04	38.34	149.00-248.00	19.34
RDW	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex K-800,1000,4500,KX-21	14	10.63	0.31	9.70-11.50	2.92	14	10.60	0.50	9.10-12.10	4.72	14	10.56	0.33	9.60-11.60	3.11	14	10.64	0.51	9.10-12.10	4.81	14	10.09	0.47	8.70-11.50	4.65
PG: Sysmex pochH-100i	18	15.01	0.27	14.20-15.80	1.80	18	14.76	0.48	13.40-16.20	3.24	18	15.03	0.28	14.20-15.80	1.87	18	14.89	0.56	13.20-16.60	3.76	18	14.44	0.41	13.20-15.60	2.85
SG: Select Sysmex Instruments (3-part Diff)	114	11.04	1.75	5.80-16.30	15.91	114	10.94	1.72	5.70-16.10	15.78	114	11.02	1.77	5.70-16.30	16.09	114	10.95	1.77	5.70-16.30	16.09	114	10.42	1.80	5.00-15.80	17.31
PG: Sysmex XP-300	82	10.23	0.23	9.50-10.90	2.25	82	10.16	0.41	9.00-11.40	4.02	82	10.21	0.25	9.40-10.90	2.45	82	10.14	0.38	9.00-11.20	3.76	82	9.59	0.35	8.60-10.60	3.65
SG: International Sysmex Instruments	22	15.00	0.27	14.20-15.80	1.80	22	14.75	0.44	13.50-16.10	2.97	22	15.02	0.29	14.10-15.90	1.93	22	14.86	0.54	13.30-16.50	3.62	22	14.42	0.38	13.30-15.50	2.64
MPV	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Intl																									



601 - Waived/PPM Tests II:609 - Waived/PPM Tests I:664 - Whole Blood Glucose:669 - Whole Blood Glucose	WBG-6					WBG-7					WBG-8					WBG-9					WBG-10				
Glucose Wh Blood (Waived)	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%
PG: Abbott Freestyle Lite	12	350.08	23.62	279.00-421.00	6.74																				
PG: Bayer Ascensia Contour	135	283.85	48.11	140.00-428.00	16.94	15	40.00	6.90	19.00-61.00	17.25															
PG: Bayer Ascensia Elite XL	8	283.00	48.28	138.00-428.00	17.07																				
PG: Lifescan One Touch Ultra	35	365.77	67.36	164.00-568.00	18.42																				
PG: McKesson/SunMark True Track	5	499.00	128.96	112.00-886.00	25.85																				
PG: Roche Accu-Chek Advantage/Comp/GTS	21	332.48	60.01	152.00-512.00	18.07																				
PG: Roche AccuChek Aviva	25	353.56	12.24	317.00-391.00	3.45																				
PG: Roche Accu-Chek II	6	337.17	25.69	260.00-414.00	7.63																				
PG: Roche Accu-Chek, others	15	352.00	17.25	300.00-404.00	4.89																				
SG: Select Lifescan One Touch Instruments	45	361.29	62.30	174.00-548.00	17.26																				
SG: Select Lifescan SureStep Instruments	5	369.00	13.44	329.00-409.00	3.63																				
SG: Select Abbott Freestyle Instruments	15	348.27	23.43	278.00-418.00	6.72																				
PG: Roche AccuChek Inform	72	359.00	32.39	262.00-456.00	9.03	35	58.46	1.46	54.00-63.00	2.59															
PG: Consult Diag Quintet Blood Glucose	15	363.13	51.26	209.00-517.00	14.13	5	40.25	4.57	26.00-54.00	11.50															
PG: Medisense Precision Xceed Pro	24	334.75	28.61	249.00-421.00	8.54																				
SG: Select Medisense Instruments	30	336.87	29.90	247.00-427.00	8.87	5	42.00	2.83	34.00-50.00	6.67															
SG: Select Roche WBG Meters	256	354.37	31.93	258.00-450.00	9.01	51	58.31	1.57	53.00-63.00	2.76	6	111.50	3.27	102.00-122.00	2.95	6	202.33	6.38	183.00-221.00	3.17	6	283.00	8.97	256.00-310.00	3.18
SG: Select Bayer WBG Meters	167	283.96	50.70	132.00-436.00	17.85	22	40.73	6.42	22.00-60.00	15.61	5	91.80	10.78	60.00-124.00	11.74	5	171.60	19.01	115.00-229.00	11.05	5	234.20	29.60	145.00-323.00	12.65
PG: Nova Biomedical StatStrip Xpress	9	312.00	19.63	253.00-371.00	6.28																				
PG: Nova Biomedical StatStrip Glucose meter	5	317.20	8.17	292.00-342.00	2.59																				
SG: Select Nova Biomedical StatStrip Instruments	14	313.86	16.26	265.00-363.00	5.19																				
PG: Bayer Contour Next	18	291.83	26.49	213.00-372.00	9.08	6	42.67	5.79	26.00-60.00	13.49															
PG: Roche AccuChek Performa	117	356.81	27.57	274.00-440.00	7.73	8	58.25	1.83	53.00-63.00	3.10															
SG: Select TrueResult Meters	5	341.00	20.66	279.00-403.00	6.07																				
PG: True Metrix Blood Glucose Meter	10	337.70	25.32	262.00-414.00	7.49																				
PG: McKesson True Metrix Pro Blood Glucose Meter	26	339.88	36.65	230.00-450.00	10.76	5	46.25	2.75	38.00-54.00	6.09															
PG: EvenCare G3 Blood Glucose System	6	334.33	26.58	254.00-414.00	7.96																				
SG: Select EvenCare Whole Blood Glucose	9	338.67	26.08	261.00-417.00	7.70																				
AG: All Group	585	337.07	38.12	223.00-451.00	11.31	99	50.80	8.94	24.00-78.00	17.45	21	101.86	12.18	65.00-139.00	11.96	21	191.90	20.12	132.00-252.00	10.47	21	267.67	30.19	177.00-359.00	11.27
683 - Antistreptolysin O (ASO):684 - C-Reactive Protein	zzzCR-3N																								
CRP Quantitative	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%
AG: All Group	9	2.86	0.31	1.92-3.80	10.91																				



789 - Group B Screen	Labs	Response	Number	Percent
		Strep B Ag Detect		
AD-6		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Cepheid GeneXpert	1	*Negative	1	100.00%
AD-7		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Cepheid GeneXpert	1	*Positive	1	100.00%
AD-8		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Cepheid GeneXpert	1	*Negative	1	100.00%
AD-9		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Cepheid GeneXpert	1	*Positive	1	100.00%
AD-10		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Cepheid GeneXpert	1	*Negative	1	100.00%
760 - ANA (ELISA)	Labs	Response	Number	Percent
		ANA (ELISA) (International)		
AE-6		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Dynex DS2	1	*Positive	1	100.00%
AE-7		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Dynex DS2	1	*Negative	1	100.00%
AE-8		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Dynex DS2	1	*Positive	1	100.00%
AE-9		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Dynex DS2	1	*Positive	1	100.00%
AE-10		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Dynex DS2	1	*Negative	1	100.00%
765 - Allergen Testing	Labs	Response	Number	Percent
		Ash - White Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*0	1	50.00%
		*1/0	1	50.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	3			
		*0	3	100.00%
		Birch/Alder Mix Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*1	2	100.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	4			
		*0	3	75.00%
		*1/0	1	25.00%
		Cedar - Mountain Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*1	2	100.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	4			



		*0	4	100.00%
		Cottonwood - Eastern Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*1	2	100.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	4			
		*0	4	100.00%
		Elm - White Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*1	2	100.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	4			
		0	3	75.00%
		2	1	25.00%
		Mulberry Mix Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	1			
		1	1	100.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	3			
		*0	3	100.00%
		Pine Mix Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	1			
		1/0	1	100.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*0	2	100.00%
		Privet Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	1			
		1/0	1	100.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*0	2	100.00%
		Sycamore - American Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	1			
		1	1	100.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*0	2	100.00%
		Marshelder - Rough Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	1			
		1	1	100.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*0	2	100.00%
		Pigweed Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*0	1	50.00%
		*1	1	50.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	4			
		*0	3	75.00%
		*1/0	1	25.00%
		Ragweed - Short Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*0	1	50.00%
		*1	1	50.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	4			



		*0	2	50.00%
		*1	2	50.00%
		Sheep Sorrel Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*0	1	50.00%
		*1	1	50.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	4			
		*0	3	75.00%
		*1	1	25.00%
		Waterhemp Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	1			
		2	1	100.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*1	2	100.00%
		Mite D Farinae Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*2	1	50.00%
		*4	1	50.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	4			
		*2	1	25.00%
		*3	3	75.00%
		Cladosporium Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*2	1	50.00%
		0	1	50.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	4			
		*0	4	100.00%
		Wheat Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	3			
		*0	1	33.33%
		*1	1	33.33%
		*1/0	1	33.33%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*1	1	50.00%
		*1/0	1	50.00%
		Vegetable Mix Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*1	1	50.00%
		*1/0	1	50.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	1			
		0	1	100.00%
		Tuna Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*1	1	50.00%
		*1/0	1	50.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	1			
		1/0	1	100.00%
		Shellfish Mix Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*3	1	50.00%
		*4	1	50.00%
AL-4		* Denotes Correct Response		



PG: Hitachi CLA-1	1			
		4	1	100.00%
		Rice Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	3			
		0	1	33.33%
		1	1	33.33%
		2	1	33.33%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*0	1	50.00%
		*1	1	50.00%
		Potato Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*1	1	50.00%
		*3	1	50.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	1			
		1	1	100.00%
		Pork Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	3			
		*0	1	33.33%
		*1	1	33.33%
		*2	1	33.33%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		0	1	50.00%
		2	1	50.00%
		Orange Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*2	1	50.00%
		*3	1	50.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	1			
		1	1	100.00%
		Onion Mix Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	3			
		0	1	33.33%
		1	1	33.33%
		2	1	33.33%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*0	1	50.00%
		*1/0	1	50.00%
		Milk Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	3			
		*1	1	33.33%
		*2	1	33.33%
		*3	1	33.33%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*3	1	50.00%
		*4	1	50.00%
		Chicken Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	3			
		0	2	66.67%
		2	1	33.33%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*0	2	100.00%
		Alternaria alternata Alrgn		



AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		0	1	50.00%
		2	1	50.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	4			
		*0	2	50.00%
		*1	1	25.00%
		*1/0	1	25.00%
		Aspergillus fumigatus Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		0	1	50.00%
		2	1	50.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	4			
		*0	4	100.00%
		Bermuda Grass Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	1			
		1	1	100.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	3			
		*0	2	66.67%
		*1/0	1	33.33%
		Cat Epithelium Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	1			
		4	1	100.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	1			
		*4	1	100.00%
		Dog Dander Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	1			
		1	1	100.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	3			
		*1	1	33.33%
		*2	2	66.67%
		Timothy Grass Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*3	1	50.00%
		*4	1	50.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	4			
		*4	4	100.00%
		Beech, American Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	1			
		1	1	100.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*0	2	100.00%
		Box Elder, Maple Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	1			
		0	1	100.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	3			
		*0	3	100.00%
		Peanut (Arachis hyogaea) Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	3			
		*3	1	33.33%
		*4	2	66.67%



AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*4	2	100.00%
		Cat Dander Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	1			
		2	1	100.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	3			
		*3	2	66.67%
		*4	1	33.33%
		Soybean Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	3			
		0	1	33.33%
		2	2	66.67%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*1	1	50.00%
		*2	1	50.00%
		Cockroach Mix Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	1			
		1	1	100.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	3			
		*0	2	66.67%
		*1/0	1	33.33%
		Oak, White Alrgn		
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*1	2	100.00%
		Penicillium Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	1			
		0	1	100.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	3			
		*0	2	66.67%
		*1/0	1	33.33%
		Walnut/Hickory/Pecan Mix Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		0	1	50.00%
		1	1	50.00%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	4			
		*0	4	100.00%
		Yeast - Bakers Alrgn		
AL-3		* Denotes Correct Response		
PG: Hitachi CLA-1	3			
		0	2	66.67%
		2	1	33.33%
AL-4		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*0	2	100.00%
683 - Antistreptolysin O (ASO)	Labs	Response	Number	Percent
		ASO (Qual) (International)		
AS-6		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Wampole ColorCard ASO	1	*Positive	1	100.00%
AS-7		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Wampole ColorCard ASO	1	*Negative	1	100.00%
AS-8		* Denotes Correct Response		



AG: All Methods	1			
		*Negative	1	100.00%
Wampole ColorCard ASO	1	*Negative	1	100.00%
AS-9		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Wampole ColorCard ASO	1	*Negative	1	100.00%
AS-10		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Wampole ColorCard ASO	1	*Positive	1	100.00%
782 - Complete Bacteriology	Labs	Response	Number	Percent
		Wound Culture		
BA-5		* Denotes Correct Response		
AG: All Methods	2			
		*Gram Negative organism	1	50.00%
		*Providencia sturatii	1	50.00%
Cultures	2	*Gram Negative organism	1	50.00%
Cultures		*Providencia sturatii	1	50.00%
975 - Bacterial Vaginosis	Labs	Response	Number	Percent
		Bacterial Vaginosa (Sialidase)		
BV-3		* Denotes Correct Response		
AG: All Methods	100			
		*Negative	98	98.00%
		Positive	2	2.00%
Sekisui Diagnostics OSOM BVBlue	100	*Negative	98	98.00%
Sekisui Diagnostics OSOM BVBlue		Positive	2	2.00%
BV-4		* Denotes Correct Response		
AG: All Methods	100			
		*Positive	91	91.00%
		Negative	9	9.00%
Sekisui Diagnostics OSOM BVBlue	100	*Positive	91	91.00%
Sekisui Diagnostics OSOM BVBlue		Negative	9	9.00%
661 - Urine Colony Count;784 - Presump ID/Colony Count - Urine;785 - Presump ID/Colony Count-Urine & Strep Combo;775 - Urine Culture;778 - Urine & Throat Cultures;779 - Urine & Genital Cultures;781 - Urine, Throat & Genital Cultures;782 - Complete Bacteri	Labs	Response	Number	Percent
		Colony Count		
CC-6		* Denotes Correct Response		
AG: All Methods	31			
		*10,000 - 100,000	10	32.26%
		*Greater than 100,000	21	67.74%
Calibrated Loop - Manual	10	*10,000 - 100,000	1	10.00%
Uri-Chek (Troy Bio)	6	*10,000 - 100,000	2	33.33%
Uricult (Orion)	12	*10,000 - 100,000	5	41.67%
Uri-Three (Culture Kits	1	*10,000 - 100,000	1	100.00%
Uricult (Lifesign)	1	*10,000 - 100,000	1	100.00%
Calibrated Loop - Manual	10	*Greater than 100,000	9	90.00%
Healthlink URI-ID	1	*Greater than 100,000	1	100.00%
Uri-Chek (Troy Bio)	6	*Greater than 100,000	4	66.67%
Uricult (Orion)	12	*Greater than 100,000	7	58.33%
784 - Presump ID/Colony Count - Urine;785 - Presump ID/Colony Count-Urine & Strep Combo	Labs	Response	Number	Percent
		Colony Count		
CC-6		* Denotes Correct Response		
AG: All Methods	31			
		*10,000 - 100,000	10	32.26%
		*Greater than 100,000	21	67.74%
Calibrated Loop - Manual	10	*10,000 - 100,000	1	10.00%
Uri-Chek (Troy Bio)	6	*10,000 - 100,000	2	33.33%
Uricult (Orion)	12	*10,000 - 100,000	5	41.67%
Uri-Three (Culture Kits	1	*10,000 - 100,000	1	100.00%
Uricult (Lifesign)	1	*10,000 - 100,000	1	100.00%
Calibrated Loop - Manual	10	*Greater than 100,000	9	90.00%
Healthlink URI-ID	1	*Greater than 100,000	1	100.00%
Uri-Chek (Troy Bio)	6	*Greater than 100,000	4	66.67%
Uricult (Orion)	12	*Greater than 100,000	7	58.33%
CC-7		* Denotes Correct Response		



AG: All Methods	5			
		*10,000 - 100,000	1	20.00%
		*Greater than 100,000	4	80.00%
Uri-Chek (Troy Bio)	1	*10,000 - 100,000	1	100.00%
Calibrated Loop - Manual	4	*Greater than 100,000	4	100.00%
CC-8		* Denotes Correct Response		
AG: All Methods	5			
		*No growth	5	100.00%
Calibrated Loop - Manual	4	*No growth	4	100.00%
Uri-Chek (Troy Bio)	1	*No growth	1	100.00%
CC-9		* Denotes Correct Response		
AG: All Methods	5			
		*10,000 - 100,000	2	40.00%
		*Greater than 100,000	3	60.00%
Calibrated Loop - Manual	4	*10,000 - 100,000	2	50.00%
Calibrated Loop - Manual		*Greater than 100,000	2	50.00%
Uri-Chek (Troy Bio)	1	*Greater than 100,000	1	100.00%
CC-10		* Denotes Correct Response		
AG: All Methods	5			
		*10,000 - 100,000	3	60.00%
		*Greater than 100,000	2	40.00%
Calibrated Loop - Manual	4	*10,000 - 100,000	2	50.00%
Uri-Chek (Troy Bio)	1	*10,000 - 100,000	1	100.00%
Calibrated Loop - Manual	4	*Greater than 100,000	2	50.00%
		Urine Presump Id		
CC-6		* Denotes Correct Response		
AG: All Methods	3			
		*Gram Positive Organism	1	33.33%
		*Presumptive Enterococcus sp.	1	33.33%
		Presumptive Proteus sp.	1	33.33%
Calibrated Loop - Manual	2	*Gram Positive Organism	1	50.00%
Calibrated Loop - Manual		Presumptive Proteus sp.	1	50.00%
CC-7		* Denotes Correct Response		
AG: All Methods	3			
		*Gram Negative Organsim	1	33.33%
		*Presumptive Proteus sp.	2	66.67%
Calibrated Loop - Manual	2	*Gram Negative Organsim	1	50.00%
Calibrated Loop - Manual		*Presumptive Proteus sp.	1	50.00%
CC-8		* Denotes Correct Response		
AG: All Methods	3			
		*No Growth	3	100.00%
Calibrated Loop - Manual	2	*No Growth	2	100.00%
CC-9		* Denotes Correct Response		
AG: All Methods	3			
		*Gram Negative Organsim	1	33.33%
		*Presumptive Enterobacter sp.	2	66.67%
Calibrated Loop - Manual	2	*Gram Negative Organsim	1	50.00%
Calibrated Loop - Manual		*Presumptive Enterobacter sp.	1	50.00%
CC-10		* Denotes Correct Response		
AG: All Methods	3			
		*Presumptive Pseudomonas sp.	2	66.67%
		Gram Positive Organism	1	33.33%
Calibrated Loop - Manual	2	*Presumptive Pseudomonas sp.	1	50.00%
Calibrated Loop - Manual		Gram Positive Organism	1	50.00%
		Susceptibility Amox/Clavulanate		
CC-6		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
		Susceptibility Ampicillin		
CC-6		* Denotes Correct Response		
AG: All Methods	5			
		*Sensitive	5	100.00%
Susceptibility	5	*Sensitive	5	100.00%
		Susceptibility Cephalothin		
CC-6		* Denotes Correct Response		
AG: All Methods	2			
		*Resistant	2	100.00%



Susceptibility	2	*Resistant	2	100.00%
		Susceptibility Ciprofloxacin		
CC-6		* Denotes Correct Response		
AG: All Methods	6			
		*Intermediate	2	33.33%
		*Sensitive	4	66.67%
Susceptibility	6	*Intermediate	2	33.33%
Susceptibility		*Sensitive	4	66.67%
		Susceptibility Gentamicin		
CC-6		* Denotes Correct Response		
AG: All Methods	1			
		*Resistant	1	100.00%
Susceptibility	1	*Resistant	1	100.00%
		Susceptibility Levofloxacin		
CC-6		* Denotes Correct Response		
AG: All Methods	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
		Susceptibility Linezolid		
CC-6		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
		Susceptibility Minocycline		
CC-6		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
		Susceptibility Penicillin		
CC-6		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
		Susceptibility Tetracycline		
CC-6		* Denotes Correct Response		
AG: All Methods	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
		Susceptibility Trimethoprim		
CC-6		* Denotes Correct Response		
AG: All Methods	1			
		*Resistant	1	100.00%
Susceptibility	1	*Resistant	1	100.00%
		Susceptibility Trimeth./Sulfameth.		
CC-6		* Denotes Correct Response		
AG: All Methods	1			
		*Resistant	1	100.00%
Susceptibility	1	*Resistant	1	100.00%
		Susceptibility Vancomycin		
CC-6		* Denotes Correct Response		
AG: All Methods	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
964 - C. difficile Toxin/Ag (2 specimens);963 - C. difficile Toxin/Ag (5 specimens)	Labs	Response	Number	Percent
		C. difficile Antigen Detection		
CDF-6		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Cepheid GeneXpert - Xpert C. diff	1	*Negative	1	100.00%
CDF-7		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Cepheid GeneXpert - Xpert C. diff	1	*Positive	1	100.00%
795 - Vaginal Pathogens II (2 specimens);794 - Vaginal Pathogens II (5 specimens);977 - Trichomonas Ag Det (Wavied)	Labs	Response	Number	Percent



		T. vaginalis - Non Culture Tech		
CGT-6		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Cepheid GeneXpert	1	*Negative	1	100.00%
BD Max	1	*Negative	1	100.00%
CGT-7		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Cepheid GeneXpert	1	*Positive	1	100.00%
BD Max	1	*Positive	1	100.00%
		Chlamydia - Non Culture Tech		
CGT-6		* Denotes Correct Response		
AG: All Methods	22			
		*Positive	22	100.00%
Quidel QuickVue Chlamydia	2	*Positive	2	100.00%
Cepheid GeneXpert	18	*Positive	18	100.00%
BD Viper XTR	1	*Positive	1	100.00%
BD Max	1	*Positive	1	100.00%
CGT-7		* Denotes Correct Response		
AG: All Methods	22			
		*Positive	22	100.00%
Quidel QuickVue Chlamydia	2	*Positive	2	100.00%
Cepheid GeneXpert	18	*Positive	18	100.00%
BD Viper XTR	1	*Positive	1	100.00%
BD Max	1	*Positive	1	100.00%
		GC - Non Culture Tech		
CGT-6		* Denotes Correct Response		
AG: All Methods	20			
		*Positive	20	100.00%
Cepheid GeneXpert	18	*Positive	18	100.00%
BD Viper XTR	1	*Positive	1	100.00%
BD Max	1	*Positive	1	100.00%
CGT-7		* Denotes Correct Response		
AG: All Methods	20			
		*Negative	20	100.00%
Cepheid GeneXpert	18	*Negative	18	100.00%
BD Viper XTR	1	*Negative	1	100.00%
BD Max	1	*Negative	1	100.00%
794 - Vaginal Pathogens II (5 specimens)	Labs	Response	Number	Percent
		Chlamydia - Non Culture Tech		
CGT-6		* Denotes Correct Response		
AG: All Methods	22			
		*Positive	22	100.00%
Quidel QuickVue Chlamydia	2	*Positive	2	100.00%
Cepheid GeneXpert	18	*Positive	18	100.00%
BD Viper XTR	1	*Positive	1	100.00%
BD Max	1	*Positive	1	100.00%
CGT-7		* Denotes Correct Response		
AG: All Methods	22			
		*Positive	22	100.00%
Quidel QuickVue Chlamydia	2	*Positive	2	100.00%
Cepheid GeneXpert	18	*Positive	18	100.00%
BD Viper XTR	1	*Positive	1	100.00%
BD Max	1	*Positive	1	100.00%
CGT-8		* Denotes Correct Response		
AG: All Methods	21			
		*Negative	21	100.00%
Quidel QuickVue Chlamydia	2	*Negative	2	100.00%
Cepheid GeneXpert	17	*Negative	17	100.00%
BD Viper XTR	1	*Negative	1	100.00%
BD Max	1	*Negative	1	100.00%
CGT-9		* Denotes Correct Response		
AG: All Methods	21			
		*Negative	21	100.00%
Quidel QuickVue Chlamydia	2	*Negative	2	100.00%
Cepheid GeneXpert	17	*Negative	17	100.00%
BD Viper XTR	1	*Negative	1	100.00%



BD Max	1	*Negative	1	100.00%
CGT-10		* Denotes Correct Response		
AG: All Methods	21			
		*Positive	21	100.00%
Quidel QuickVue Chlamydia	2	*Positive	2	100.00%
Cepheid GeneXpert	17	*Positive	17	100.00%
BD Viper XTR	1	*Positive	1	100.00%
BD Max	1	*Positive	1	100.00%
		GC - Non Culture Tech		
CGT-6		* Denotes Correct Response		
AG: All Methods	20			
		*Positive	20	100.00%
Cepheid GeneXpert	18	*Positive	18	100.00%
BD Viper XTR	1	*Positive	1	100.00%
BD Max	1	*Positive	1	100.00%
CGT-7		* Denotes Correct Response		
AG: All Methods	20			
		*Negative	20	100.00%
Cepheid GeneXpert	18	*Negative	18	100.00%
BD Viper XTR	1	*Negative	1	100.00%
BD Max	1	*Negative	1	100.00%
CGT-8		* Denotes Correct Response		
AG: All Methods	19			
		*Positive	19	100.00%
Cepheid GeneXpert	17	*Positive	17	100.00%
BD Viper XTR	1	*Positive	1	100.00%
BD Max	1	*Positive	1	100.00%
CGT-9		* Denotes Correct Response		
AG: All Methods	19			
		*Positive	19	100.00%
Cepheid GeneXpert	17	*Positive	17	100.00%
BD Viper XTR	1	*Positive	1	100.00%
BD Max	1	*Positive	1	100.00%
CGT-10		* Denotes Correct Response		
AG: All Methods	19			
		*Negative	19	100.00%
Cepheid GeneXpert	17	*Negative	17	100.00%
BD Viper XTR	1	*Negative	1	100.00%
BD Max	1	*Negative	1	100.00%
		T. vaginalis - Non Culture Tech		
CGT-6		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Cepheid GeneXpert	1	*Negative	1	100.00%
BD Max	1	*Negative	1	100.00%
CGT-7		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Cepheid GeneXpert	1	*Positive	1	100.00%
BD Max	1	*Positive	1	100.00%
CGT-8		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Cepheid GeneXpert	1	*Positive	1	100.00%
BD Max	1	*Positive	1	100.00%
CGT-9		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
BD Max	1	*Negative	1	100.00%
CGT-10		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Cepheid GeneXpert	1	*Negative	1	100.00%
BD Max	1	*Negative	1	100.00%
		Trichomonas Ag Det (Waived)		
CGT-6		* Denotes Correct Response		
AG: All Methods	97			
		*Negative	86	88.66%



		Positive	11	11.34%
Sekisui Diagnostics OSOM Trichomonas	97	*Negative	86	88.66%
Sekisui Diagnostics OSOM Trichomonas		Positive	11	11.34%
CGT-7		* Denotes Correct Response		
AG: All Methods	97			
		*Positive	90	92.78%
		Negative	7	7.22%
Sekisui Diagnostics OSOM Trichomonas	97	*Positive	90	92.78%
Sekisui Diagnostics OSOM Trichomonas		Negative	7	7.22%
639 - Complete Urinalysis Package;638 - Urine Sediment;641 - Clinical Microscopy;601 - Waived/PPM Tests II;609 - Waived/PPM Tests I	Labs	Response	Number	Percent
		Urine Sediment Id		
CM-10		* Denotes Correct Response		
AG: All Methods	625			
		*Mucous strands	543	86.88%
		Bacteria	9	1.44%
		Fatty (fat globules) cast	1	0.16%
		Fiber (exogenous)/fecal contamination	16	2.56%
		Hyaline cast	5	0.80%
		Ident Unknown - Referred	4	0.64%
		Renal tubular epithelial (RTE)	1	0.16%
		Squamous epithelial cell	5	0.80%
		Transitional epithelial cell (urothelial)	1	0.16%
		Trichomonas vaginalis	1	0.16%
		Waxy cast	1	0.16%
		Yeast/fungi	36	5.76%
		Yeast/fungi present	2	0.32%
Microscopy	625	*Mucous strands	543	86.88%
Microscopy		Bacteria	9	1.44%
Microscopy		Fatty (fat globules) cast	1	0.16%
Microscopy		Fiber (exogenous)/fecal contamination	16	2.56%
Microscopy		Hyaline cast	5	0.80%
Microscopy		Ident Unknown - Referred	4	0.64%
Microscopy		Renal tubular epithelial (RTE)	1	0.16%
Microscopy		Squamous epithelial cell	5	0.80%
Microscopy		Transitional epithelial cell (urothelial)	1	0.16%
Microscopy		Trichomonas vaginalis	1	0.16%
Microscopy		Waxy cast	1	0.16%
Microscopy		Yeast/fungi	36	5.76%
Microscopy		Yeast/fungi present	2	0.32%
CM-11		* Denotes Correct Response		
AG: All Methods	620			
		*Cystine crystal	585	94.35%
		*Ident Unknown - Referred	1	0.16%
		Ammonium magnesium phosphate (Triple phosphate)	1	0.16%
		Calcium oxalate crystal	10	1.61%
		Hippuric acid	3	0.48%
		Uric acid crystal	20	3.23%
Microscopy	620	*Cystine crystal	585	94.35%
Microscopy		*Ident Unknown - Referred	1	0.16%
Microscopy		Ammonium magnesium phosphate (Triple phosphate)	1	0.16%
Microscopy		Calcium oxalate crystal	10	1.61%
Microscopy		Hippuric acid	3	0.48%
Microscopy		Uric acid crystal	20	3.23%
CM-12		* Denotes Correct Response		
AG: All Methods	625			
		*Red blood cell (RBC) (Erythrocyte)	611	97.76%
		Fat droplets or globules	3	0.48%
		RBC/Blood/Hgb cast	2	0.32%
		Renal tubular epithelial (RTE)	5	0.80%
		RTE with fat globules (oval fat bodies)	1	0.16%
		White blood cell (WBC) (Leukocyte)	2	0.32%
		Yeast/fungi	1	0.16%
Microscopy	625	*Red blood cell (RBC) (Erythrocyte)	611	97.76%
Microscopy		Fat droplets or globules	3	0.48%
Microscopy		RBC/Blood/Hgb cast	2	0.32%
Microscopy		Renal tubular epithelial (RTE)	5	0.80%
Microscopy		RTE with fat globules (oval fat bodies)	1	0.16%



Microscopy		White blood cell (WBC) (Leukocyte)	2	0.32%
Microscopy		Yeast/fungi	1	0.16%
CM-13		* Denotes Correct Response		
AG: All Methods	625			
		*White blood cell (WBC) (Leukocyte)	611	97.76%
		Cellular (WBC) cast	1	0.16%
		RBC/Blood/Hgb cast	2	0.32%
		Red blood cell (RBC) (Erythrocyte)	2	0.32%
		Renal tubular epithelial (RTE)	5	0.80%
		RTE with fat globules (oval fat bodies)	2	0.32%
		Transitional epithelial cell (urothelial)	2	0.32%
Microscopy	625	*White blood cell (WBC) (Leukocyte)	611	97.76%
Microscopy		Cellular (WBC) cast	1	0.16%
Microscopy		RBC/Blood/Hgb cast	2	0.32%
Microscopy		Red blood cell (RBC) (Erythrocyte)	2	0.32%
Microscopy		Renal tubular epithelial (RTE)	5	0.80%
Microscopy		RTE with fat globules (oval fat bodies)	2	0.32%
Microscopy		Transitional epithelial cell (urothelial)	2	0.32%
		Vaginal Wet Prep		
CM-14		* Denotes Correct Response		
AG: All Methods	835			
		*Red Blood Cell (RBC) (Erythrocyte)	714	85.51%
		*Spermatozoa absent	2	0.24%
		*Yeast/fungi absent	2	0.24%
		Bacteria present	2	0.24%
		Mucous strand	1	0.12%
		Pollen grain	7	0.84%
		Positive Fern Test	1	0.12%
		Spermatozoa present	1	0.12%
		Starch granule	4	0.48%
		Trichomonas vaginalis	1	0.12%
		White blood cell (WBC) (Leukocyte)	1	0.12%
		White Blood Cell (WBC) (Leukocyte)	68	8.14%
		Yeast/fungi present	31	3.71%
Microscopy	835	*Red Blood Cell (RBC) (Erythrocyte)	714	85.51%
Microscopy		*Spermatozoa absent	2	0.24%
Microscopy		*Yeast/fungi absent	2	0.24%
Microscopy		Bacteria present	2	0.24%
Microscopy		Mucous strand	1	0.12%
Microscopy		Pollen grain	7	0.84%
Microscopy		Positive Fern Test	1	0.12%
Microscopy		Spermatozoa present	1	0.12%
Microscopy		Starch granule	4	0.48%
Microscopy		Trichomonas vaginalis	1	0.12%
Microscopy		White blood cell (WBC) (Leukocyte)	1	0.12%
Microscopy		White Blood Cell (WBC) (Leukocyte)	68	8.14%
Microscopy		Yeast/fungi present	31	3.71%
CM-15		* Denotes Correct Response		
AG: All Methods	833			
		*Ident Unknown - Referred	1	0.12%
		*Spermatozoa absent	5	0.60%
		*Starch granule	693	83.19%
		Bacteria present	1	0.12%
		Calcium oxalate crystal	2	0.24%
		Clue cell	5	0.60%
		Cystine crystal	1	0.12%
		Fat droplets or globules	2	0.24%
		Fiber (exogenous)/fecal contamination	3	0.36%
		Pollen grain	8	0.96%
		Red Blood Cell (RBC) (Erythrocyte)	35	4.20%
		RTE with fat globules (oval fat bodies)	1	0.12%
		Squamous epithelial cell	6	0.72%
		Trichomonas vaginalis	9	1.08%
		White Blood Cell (WBC) (Leukocyte)	60	7.20%
		Yeast/fungi present	1	0.12%
Microscopy	833	*Ident Unknown - Referred	1	0.12%
Microscopy		*Spermatozoa absent	5	0.60%
Microscopy		*Starch granule	693	83.19%
Microscopy		Bacteria present	1	0.12%



Microscopy		Calcium oxalate crystal	2	0.24%
Microscopy		Clue cell	5	0.60%
Microscopy		Cystine crystal	1	0.12%
Microscopy		Fat droplets or globules	2	0.24%
Microscopy		Fiber (exogenous)/fecal contamination	3	0.36%
Microscopy		Pollen grain	8	0.96%
Microscopy		Red Blood Cell (RBC) (Erythrocyte)	35	4.20%
Microscopy		RTE with fat globules (oval fat bodies)	1	0.12%
Microscopy		Squamous epithelial cell	6	0.72%
Microscopy		Trichomonas vaginalis	9	1.08%
Microscopy		White Blood Cell (WBC) (Leukocyte)	60	7.20%
Microscopy		Yeast/fungi present	1	0.12%
		Skin Scraping		
CM-16		* Denotes Correct Response		
AG: All Methods	575			
		*No scabies present	23	4.00%
		*Yeast/fungi absent	527	91.65%
		Scabies present	1	0.17%
		Yeast/fungi present	24	4.17%
Microscopy	575	*No scabies present	23	4.00%
Microscopy		*Yeast/fungi absent	527	91.65%
Microscopy		Scabies present	1	0.17%
Microscopy		Yeast/fungi present	24	4.17%
		Nasal Smear		
CM-17		* Denotes Correct Response		
AG: All Methods	184			
		*Eosinophils are present	181	98.37%
		No eosinophils are present	3	1.63%
Microscopy	184	*Eosinophils are present	181	98.37%
Microscopy		No eosinophils are present	3	1.63%
683 - Antistreptolysin O (ASO);684 - C-Reactive Protein	Labs	Response	Number	Percent
		CRP Qualitative		
CR-3		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Inverness Immunex	1	*Positive	1	100.00%
CR-4		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Inverness Immunex	1	*Negative	1	100.00%
772 - Dermatophyte Screen	Labs	Response	Number	Percent
		Dermatophyte Screen		
DM-6		* Denotes Correct Response		
AG: All Methods	12			
		*Negative, No Dermatophytes Present	10	83.33%
		Positive, Dermatophytes Present	2	16.67%
Dermatophytes	12	*Negative, No Dermatophytes Present	10	83.33%
Dermatophytes		Positive, Dermatophytes Present	2	16.67%
DM-7		* Denotes Correct Response		
AG: All Methods	12			
		*Positive, Dermatophytes Present	12	100.00%
Dermatophytes	12	*Positive, Dermatophytes Present	12	100.00%
DM-8		* Denotes Correct Response		
AG: All Methods	12			
		*Positive, Dermatophytes Present	11	91.67%
		Negative, No Dermatophytes Present	1	8.33%
Dermatophytes	12	*Positive, Dermatophytes Present	11	91.67%
Dermatophytes		Negative, No Dermatophytes Present	1	8.33%
DM-9		* Denotes Correct Response		
AG: All Methods	12			
		*Negative, No Dermatophytes Present	12	100.00%
Dermatophytes	12	*Negative, No Dermatophytes Present	12	100.00%
DM-10		* Denotes Correct Response		
AG: All Methods	12			
		*Negative, No Dermatophytes Present	12	100.00%
Dermatophytes	12	*Negative, No Dermatophytes Present	12	100.00%
777 - Genital Culture;796 - Genital Culture (2 specimens)	Labs	Response	Number	Percent
		Genital Culture		



GC-6		* Denotes Correct Response		
AG: All Methods	3			
		*No growth, GC selective media	3	100.00%
Cultures	3	*No growth, GC selective media	3	100.00%
GC-7		* Denotes Correct Response		
AG: All Methods	3			
		*No growth, GC selective media	3	100.00%
Cultures	3	*No growth, GC selective media	3	100.00%
GC-8		* Denotes Correct Response		
AG: All Methods	3			
		*Negative for GC	2	66.67%
		*No Gram Negative diplococci present	1	33.33%
Cultures	3	*Negative for GC	2	66.67%
Cultures		*No Gram Negative diplococci present	1	33.33%
GC-9		* Denotes Correct Response		
AG: All Methods	5			
		*Positive for GC	2	40.00%
		*Presumptive positive for GC, GR(-) diplococci, oxidase positive organism	1	20.00%
		*Presumptive positive for GC, growth referred for ID/confirmation	1	20.00%
		No growth, GC selective media	1	20.00%
Cultures	5	*Positive for GC	2	40.00%
Cultures		*Presumptive positive for GC, GR(-) diplococci, oxidase positive organism	1	20.00%
Cultures		*Presumptive positive for GC, growth referred for ID/confirmation	1	20.00%
Cultures		No growth, GC selective media	1	20.00%
GC-10		* Denotes Correct Response		
AG: All Methods	4			
		*Negative for GC	2	50.00%
		*No growth, GC selective media	2	50.00%
Cultures	4	*Negative for GC	2	50.00%
Cultures		*No growth, GC selective media	2	50.00%
630 - Gastric Occult Blood & pH	Labs	Response	Number	Percent
		Gastric Occult Blood		
GOB-3		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Beckman Coulter Gastrocult	1	*Negative	1	100.00%
GOB-4		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Beckman Coulter Gastrocult	1	*Positive	1	100.00%
668 - Gram Stain	Labs	Response	Number	Percent
		Gram Stain		
GS-6		* Denotes Correct Response		
AG: All Methods	6			
		*Gram Positive	6	100.00%
Gram Stain	6	*Gram Positive	6	100.00%
GS-7		* Denotes Correct Response		
AG: All Methods	7			
		*Gram Positive	6	85.71%
		Gram Negative	1	14.29%
Gram Stain	7	*Gram Positive	6	85.71%
Gram Stain		Gram Negative	1	14.29%
GS-8		* Denotes Correct Response		
AG: All Methods	7			
		*Gram Negative	7	100.00%
Gram Stain	7	*Gram Negative	7	100.00%
GS-9		* Denotes Correct Response		
AG: All Methods	7			
		*Gram Positive	3	42.86%
		Gram Negative	4	57.14%
Gram Stain	7	*Gram Positive	3	42.86%
Gram Stain		Gram Negative	4	57.14%
GS-10		* Denotes Correct Response		
AG: All Methods	7			
		*Gram Negative	6	85.71%



		Gram Positive	1	14.29%
Gram Stain	7	*Gram Negative	6	85.71%
Gram Stain		Gram Positive	1	14.29%
		Gram Stain Morph		
GS-6		* Denotes Correct Response		
AG: All Methods	6			
		*Cocci	5	83.33%
		*Cocci in Pairs	1	16.67%
Gram Stain	6	*Cocci	5	83.33%
Gram Stain		*Cocci in Pairs	1	16.67%
GS-7		* Denotes Correct Response		
AG: All Methods	7			
		*Rods/Bacilli	7	100.00%
Gram Stain	7	*Rods/Bacilli	7	100.00%
GS-8		* Denotes Correct Response		
AG: All Methods	7			
		*Rods/Bacilli	7	100.00%
Gram Stain	7	*Rods/Bacilli	7	100.00%
GS-9		* Denotes Correct Response		
AG: All Methods	7			
		*Rods/Bacilli	7	100.00%
Gram Stain	7	*Rods/Bacilli	7	100.00%
GS-10		* Denotes Correct Response		
AG: All Methods	7			
		*Cocci	2	28.57%
		*Cocci in Pairs	1	14.29%
		*Diplococci	4	57.14%
Gram Stain	7	*Cocci	2	28.57%
Gram Stain		*Cocci in Pairs	1	14.29%
Gram Stain		*Diplococci	4	57.14%
978 - Gastrointestinal Viral Detection	Labs	Response	Number	Percent
		Rotavirus - Ag Detect		
GVD-6		* Denotes Correct Response		
AG: All Methods	9			
		*Positive	9	100.00%
Meridian Immunocard Stat! Rotavirus	5	*Positive	5	100.00%
SAS Rota-Test	4	*Positive	4	100.00%
GVD-7		* Denotes Correct Response		
AG: All Methods	9			
		*Positive	9	100.00%
Meridian Immunocard Stat! Rotavirus	5	*Positive	5	100.00%
SAS Rota-Test	4	*Positive	4	100.00%
GVD-8		* Denotes Correct Response		
AG: All Methods	9			
		*Negative	9	100.00%
Meridian Immunocard Stat! Rotavirus	5	*Negative	5	100.00%
SAS Rota-Test	4	*Negative	4	100.00%
GVD-9		* Denotes Correct Response		
AG: All Methods	9			
		*Negative	8	88.89%
		Positive	1	11.11%
Meridian Immunocard Stat! Rotavirus	5	*Negative	5	100.00%
SAS Rota-Test	4	*Negative	3	75.00%
SAS Rota-Test		Positive	1	25.00%
GVD-10		* Denotes Correct Response		
AG: All Methods	9			
		*Positive	8	88.89%
		Negative	1	11.11%
Meridian Immunocard Stat! Rotavirus	5	*Positive	5	100.00%
SAS Rota-Test	4	*Positive	3	75.00%
SAS Rota-Test		Negative	1	25.00%
678 - Serum hCG	Labs	Response	Number	Percent
		HCG - Qual Serum		
HCG-6		* Denotes Correct Response		
AG: All Methods	78			
		*Negative	77	98.72%
		Positive	1	1.28%
Beck Coulter ICON 25	7	*Negative	7	100.00%



Fisher HC Sure-Vue hCG STAT	2	*Negative	2	100.00%
Sekisui Diagnostics OSOM	2	*Negative	2	100.00%
Henry Schein One Step + (hCG)	6	*Negative	6	100.00%
Immunostics Immuno/hCG	1	*Negative	1	100.00%
Inverness Signify HCG	2	*Negative	2	100.00%
LifeSign Status hCG	1	*Negative	1	100.00%
McKesson Medi-Lab hCG - Urine	1	*Negative	1	100.00%
Quidel QuickVue + hCG	2	*Negative	2	100.00%
Quidel QuickVue One-Step hCG Combo (U/S)	27	*Negative	27	100.00%
Select Diag hCG Combo	3	*Negative	2	66.67%
SP Brand (Cardinal) HCG Combo	13	*Negative	13	100.00%
Polymedco Poly stat hCG Serum	1	*Negative	1	100.00%
Stanbio True 20 Plus	1	*Negative	1	100.00%
Consult Diag hCG Combo Cassette	8	*Negative	8	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	1	*Negative	1	100.00%
Select Diag hCG Combo	3	Positive	1	33.33%
HCG-7		* Denotes Correct Response		
AG: All Methods	78			
		*Negative	77	98.72%
		Positive	1	1.28%
Beck Coulter ICON 25	7	*Negative	7	100.00%
Fisher HC Sure-Vue hCG STAT	2	*Negative	2	100.00%
Sekisui Diagnostics OSOM	2	*Negative	2	100.00%
Henry Schein One Step + (hCG)	6	*Negative	6	100.00%
Immunostics Immuno/hCG	1	*Negative	1	100.00%
Inverness Signify HCG	2	*Negative	2	100.00%
LifeSign Status hCG	1	*Negative	1	100.00%
McKesson Medi-Lab hCG - Urine	1	*Negative	1	100.00%
Quidel QuickVue + hCG	2	*Negative	2	100.00%
Quidel QuickVue One-Step hCG Combo (U/S)	27	*Negative	27	100.00%
Select Diag hCG Combo	3	*Negative	2	66.67%
SP Brand (Cardinal) HCG Combo	13	*Negative	13	100.00%
Polymedco Poly stat hCG Serum	1	*Negative	1	100.00%
Stanbio True 20 Plus	1	*Negative	1	100.00%
Consult Diag hCG Combo Cassette	8	*Negative	8	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	1	*Negative	1	100.00%
Select Diag hCG Combo	3	Positive	1	33.33%
HCG-8		* Denotes Correct Response		
AG: All Methods	78			
		*Positive	77	98.72%
		Negative	1	1.28%
Beck Coulter ICON 25	7	*Positive	7	100.00%
Fisher HC Sure-Vue hCG STAT	2	*Positive	2	100.00%
Sekisui Diagnostics OSOM	2	*Positive	2	100.00%
Henry Schein One Step + (hCG)	6	*Positive	6	100.00%
Immunostics Immuno/hCG	1	*Positive	1	100.00%
Inverness Signify HCG	2	*Positive	2	100.00%
LifeSign Status hCG	1	*Positive	1	100.00%
McKesson Medi-Lab hCG - Urine	1	*Positive	1	100.00%
Quidel QuickVue + hCG	2	*Positive	2	100.00%
Quidel QuickVue One-Step hCG Combo (U/S)	27	*Positive	27	100.00%
Select Diag hCG Combo	3	*Positive	2	66.67%
SP Brand (Cardinal) HCG Combo	13	*Positive	13	100.00%
Polymedco Poly stat hCG Serum	1	*Positive	1	100.00%
Stanbio True 20 Plus	1	*Positive	1	100.00%
Consult Diag hCG Combo Cassette	8	*Positive	8	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	1	*Positive	1	100.00%
Select Diag hCG Combo	3	Negative	1	33.33%
HCG-9		* Denotes Correct Response		
AG: All Methods	78			
		*Positive	77	98.72%
		Negative	1	1.28%
Beck Coulter ICON 25	7	*Positive	7	100.00%
Fisher HC Sure-Vue hCG STAT	2	*Positive	2	100.00%
Sekisui Diagnostics OSOM	2	*Positive	2	100.00%
Henry Schein One Step + (hCG)	6	*Positive	6	100.00%
Immunostics Immuno/hCG	1	*Positive	1	100.00%
Inverness Signify HCG	2	*Positive	2	100.00%
LifeSign Status hCG	1	*Positive	1	100.00%



McKesson Medi-Lab hCG - Urine	1	*Positive	1	100.00%
Quidel QuickVue + hCG	2	*Positive	2	100.00%
Quidel QuickVue One-Step hCG Combo (U/S)	27	*Positive	27	100.00%
Select Diag hCG Combo	3	*Positive	2	66.67%
SP Brand (Cardinal) HCG Combo	13	*Positive	13	100.00%
Polymedco Poly stat hCG Serum	1	*Positive	1	100.00%
Stanbio True 20 Plus	1	*Positive	1	100.00%
Consult Diag hCG Combo Cassette	8	*Positive	8	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	1	*Positive	1	100.00%
Select Diag hCG Combo	3	Negative	1	33.33%
HCG-10		* Denotes Correct Response		
AG: All Methods	78			
		*Positive	78	100.00%
Beck Coulter ICON 25	7	*Positive	7	100.00%
Fisher HC Sure-Vue hCG STAT	2	*Positive	2	100.00%
Sekisui Diagnostics OSOM	2	*Positive	2	100.00%
Henry Schein One Step + (hCG)	6	*Positive	6	100.00%
Immunostics Immuno/hCG	1	*Positive	1	100.00%
Inverness Signify HCG	2	*Positive	2	100.00%
LifeSign Status hCG	1	*Positive	1	100.00%
McKesson Medi-Lab hCG - Urine	1	*Positive	1	100.00%
Quidel QuickVue + hCG	2	*Positive	2	100.00%
Quidel QuickVue One-Step hCG Combo (U/S)	27	*Positive	27	100.00%
Select Diag hCG Combo	3	*Positive	3	100.00%
SP Brand (Cardinal) HCG Combo	13	*Positive	13	100.00%
Polymedco Poly stat hCG Serum	1	*Positive	1	100.00%
Stanbio True 20 Plus	1	*Positive	1	100.00%
Consult Diag hCG Combo Cassette	8	*Positive	8	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	1	*Positive	1	100.00%
		HCG - Qual Serum (International)		
HCG-6		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Henry Schein One Step + (hCG)	1	*Negative	1	100.00%
SAS Ultra HCG	2	*Negative	2	100.00%
HCG-7		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Henry Schein One Step + (hCG)	1	*Negative	1	100.00%
SAS Ultra HCG	2	*Negative	2	100.00%
HCG-8		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Henry Schein One Step + (hCG)	1	*Positive	1	100.00%
SAS Ultra HCG	2	*Positive	2	100.00%
HCG-9		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Henry Schein One Step + (hCG)	1	*Positive	1	100.00%
SAS Ultra HCG	2	*Positive	2	100.00%
HCG-10		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Henry Schein One Step + (hCG)	1	*Positive	1	100.00%
SAS Ultra HCG	2	*Positive	2	100.00%
933 - HIV Antibody/Antigen Combo (2 specimens);934 - HIV Antibody/Antigen Combo (5 specimens)	Labs	Response	Number	Percent
		HIV Antibody (Waived)		
HIV-6		* Denotes Correct Response		
AG: All Methods	98			
		*Negative	93	94.90%
		Positive	5	5.10%
Alere Clearview HIV 1/2 STAT PAK (WB)	43	*Negative	41	95.35%
Orasure OraQuick Anti-HIV	9	*Negative	8	88.89%
Trinity Biotech Uni-Gold Recombigen HIV	2	*Negative	2	100.00%
Alere Determine HIV Ag/Ab	30	*Negative	29	96.67%
Chembio HIV 1/2 STAT-PAK Dipstick	6	*Negative	6	100.00%
Chembio HIV 1/2 STAT-PAK Cassette	3	*Negative	2	66.67%



bioLytical INSTI HIV Antibody	5	*Negative	5	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	43	Positive	2	4.65%
Orasure OraQuick Anti-HIV	9	Positive	1	11.11%
Alere Determine HIV Ag/Ab	30	Positive	1	3.33%
Chembio HIV 1/2 STAT-PAK Cassette	3	Positive	1	33.33%
HIV-7		* Denotes Correct Response		
AG: All Methods	98			
		*Positive	95	96.94%
		Negative	3	3.06%
Alere Clearview HIV 1/2 STAT PAK (WB)	43	*Positive	41	95.35%
Orasure OraQuick Anti-HIV	9	*Positive	9	100.00%
Trinity Biotech Uni-Gold Recombigen HIV	2	*Positive	2	100.00%
Alere Determine HIV Ag/Ab	30	*Positive	29	96.67%
Chembio HIV 1/2 STAT-PAK Dipstick	6	*Positive	6	100.00%
Chembio HIV 1/2 STAT-PAK Cassette	3	*Positive	3	100.00%
bioLytical INSTI HIV Antibody	5	*Positive	5	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	43	Negative	2	4.65%
Alere Determine HIV Ag/Ab	30	Negative	1	3.33%
HIV Antibody (International)				
HIV-6		* Denotes Correct Response		
AG: All Methods	4			
		*Negative	4	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Negative	3	100.00%
Intl Chembio HIV 1/2 STAT-PAK Dipstick	1	*Negative	1	100.00%
HIV-7		* Denotes Correct Response		
AG: All Methods	4			
		*Positive	4	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Positive	3	100.00%
Intl Chembio HIV 1/2 STAT-PAK Dipstick	1	*Positive	1	100.00%
HIV Antigen (p24) (Waived)				
HIV-6		* Denotes Correct Response		
AG: All Methods	30			
		*Positive	28	93.33%
		Negative	2	6.67%
Alere Determine HIV Ag/Ab	30	*Positive	28	93.33%
Alere Determine HIV Ag/Ab		Negative	2	6.67%
HIV-7		* Denotes Correct Response		
AG: All Methods	30			
		*Negative	28	93.33%
		Positive	2	6.67%
Alere Determine HIV Ag/Ab	30	*Negative	28	93.33%
Alere Determine HIV Ag/Ab		Positive	2	6.67%
934 - HIV Antibody/Antigen Combo (5 specimens)	Labs	Response	Number	Percent
HIV Antibody (Waived)				
HIV-6		* Denotes Correct Response		
AG: All Methods	98			
		*Negative	93	94.90%
		Positive	5	5.10%
Alere Clearview HIV 1/2 STAT PAK (WB)	43	*Negative	41	95.35%
Orasure OraQuick Anti-HIV	9	*Negative	8	88.89%
Trinity Biotech Uni-Gold Recombigen HIV	2	*Negative	2	100.00%
Alere Determine HIV Ag/Ab	30	*Negative	29	96.67%
Chembio HIV 1/2 STAT-PAK Dipstick	6	*Negative	6	100.00%
Chembio HIV 1/2 STAT-PAK Cassette	3	*Negative	2	66.67%
bioLytical INSTI HIV Antibody	5	*Negative	5	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	43	Positive	2	4.65%
Orasure OraQuick Anti-HIV	9	Positive	1	11.11%
Alere Determine HIV Ag/Ab	30	Positive	1	3.33%
Chembio HIV 1/2 STAT-PAK Cassette	3	Positive	1	33.33%
HIV-7		* Denotes Correct Response		
AG: All Methods	98			
		*Positive	95	96.94%
		Negative	3	3.06%
Alere Clearview HIV 1/2 STAT PAK (WB)	43	*Positive	41	95.35%
Orasure OraQuick Anti-HIV	9	*Positive	9	100.00%
Trinity Biotech Uni-Gold Recombigen HIV	2	*Positive	2	100.00%
Alere Determine HIV Ag/Ab	30	*Positive	29	96.67%
Chembio HIV 1/2 STAT-PAK Dipstick	6	*Positive	6	100.00%



Chembio HIV 1/2 STAT-PAK Cassette	3	*Positive	3	100.00%
bioLytical INSTI HIV Antibody	5	*Positive	5	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	43	Negative	2	4.65%
Alere Determine HIV Ag/Ab	30	Negative	1	3.33%
HIV-8		* Denotes Correct Response		
AG: All Methods	8			
		*Positive	8	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	4	*Positive	4	100.00%
Orasure OraQuick Anti-HIV	1	*Positive	1	100.00%
Trinity Biotech Uni-Gold Recombigen HIV	1	*Positive	1	100.00%
Alere Determine HIV Ag/Ab	1	*Positive	1	100.00%
bioLytical INSTI HIV Antibody	1	*Positive	1	100.00%
HIV-9		* Denotes Correct Response		
AG: All Methods	8			
		*Negative	8	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	4	*Negative	4	100.00%
Orasure OraQuick Anti-HIV	1	*Negative	1	100.00%
Trinity Biotech Uni-Gold Recombigen HIV	1	*Negative	1	100.00%
Alere Determine HIV Ag/Ab	1	*Negative	1	100.00%
bioLytical INSTI HIV Antibody	1	*Negative	1	100.00%
HIV-10		* Denotes Correct Response		
AG: All Methods	8			
		*Negative	8	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	4	*Negative	4	100.00%
Orasure OraQuick Anti-HIV	1	*Negative	1	100.00%
Trinity Biotech Uni-Gold Recombigen HIV	1	*Negative	1	100.00%
Alere Determine HIV Ag/Ab	1	*Negative	1	100.00%
bioLytical INSTI HIV Antibody	1	*Negative	1	100.00%
		HIV Antibody (International)		
HIV-6		* Denotes Correct Response		
AG: All Methods	4			
		*Negative	4	100.00%
Intl OCD Vitros ECI/ECiQ	3	*Negative	3	100.00%
Intl Chembio HIV 1/2 STAT-PAK Dipstick	1	*Negative	1	100.00%
HIV-7		* Denotes Correct Response		
AG: All Methods	4			
		*Positive	4	100.00%
Intl OCD Vitros ECI/ECiQ	3	*Positive	3	100.00%
Intl Chembio HIV 1/2 STAT-PAK Dipstick	1	*Positive	1	100.00%
HIV-8		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECI/ECiQ	3	*Positive	3	100.00%
HIV-9		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECiQ	3	*Negative	3	100.00%
HIV-10		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECiQ	3	*Negative	3	100.00%
		HIV Antibody		
HIV-6		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
OCD Vitros 5600	1	*Negative	1	100.00%
HIV-7		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
OCD Vitros 5600	1	*Positive	1	100.00%
HIV-8		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
OCD Vitros 5600	1	*Positive	1	100.00%
HIV-9		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
OCD Vitros 5600	1	*Negative	1	100.00%



HIV-10		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
OCD Vitros 5600	1	*Negative	1	100.00%
		HIV Ab/P24 Ag Combo (International)		
HIV-6		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Intl OCD Vitros ECI/ECiQ	2	*Positive	2	100.00%
HIV-7		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Intl OCD Vitros ECI/ECiQ	2	*Positive	2	100.00%
HIV-8		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Intl OCD Vitros ECI/ECiQ	2	*Positive	2	100.00%
HIV-9		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Intl OCD Vitros ECI/ECiQ	2	*Negative	2	100.00%
HIV-10		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Intl OCD Vitros ECI/ECiQ	2	*Negative	2	100.00%
		HIV Antigen (p24) (Waived)		
HIV-6		* Denotes Correct Response		
AG: All Methods	30			
		*Positive	28	93.33%
		Negative	2	6.67%
Alere Determine HIV Ag/Ab	30	*Positive	28	93.33%
Alere Determine HIV Ag/Ab		Negative	2	6.67%
HIV-7		* Denotes Correct Response		
AG: All Methods	30			
		*Negative	28	93.33%
		Positive	2	6.67%
Alere Determine HIV Ag/Ab	30	*Negative	28	93.33%
Alere Determine HIV Ag/Ab		Positive	2	6.67%
HIV-8		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Alere Determine HIV Ag/Ab	1	*Negative	1	100.00%
HIV-9		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Alere Determine HIV Ag/Ab	1	*Negative	1	100.00%
HIV-10		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Alere Determine HIV Ag/Ab	1	*Negative	1	100.00%
644 - H. pylori	Labs	Response	Number	Percent
		H. pylori Ab Detect		
HP-2		* Denotes Correct Response		
AG: All Methods	63			
		*Positive	58	92.06%
		Negative	5	7.94%
Beck Coulter FlexSure	3	*Positive	3	100.00%
Beck Coulter ICON HP	2	*Positive	2	100.00%
McKesson Medi-Lab Perf H. pylori	6	*Positive	6	100.00%
Henry Schein One Step+ (H. pylori)	11	*Positive	8	72.73%
Select Diag H. pylori Cassette	2	*Positive	2	100.00%
Quidel QuickVue H. pylori gII (S/P)	3	*Positive	3	100.00%
Quidel QuickVue H. pylori gII (WB)	7	*Positive	7	100.00%
SP Brand (Cardinal) HP Rapid Test	8	*Positive	7	87.50%
Alere Clearview II S/P (mod)	2	*Positive	2	100.00%
Alere Clearview II WB (waived)	3	*Positive	3	100.00%
LifeSign Status H. pylori	2	*Positive	2	100.00%
Consult Diag H pylori	12	*Positive	11	91.67%



Jant Accutest H pylori	1	*Positive	1	100.00%
Medline H. pylori Cassette	1	*Positive	1	100.00%
Henry Schein One Step+ (H. pylori)	11	Negative	3	27.27%
SP Brand (Cardinal) HP Rapid Test	8	Negative	1	12.50%
Consult Diag H pylori	12	Negative	1	8.33%
929 - Genital Viral Detection I	Labs	Response	Number	Percent
		HSV 1		
HVD-6		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Hologic Aptima HSV 1 & 2	1	*Negative	1	100.00%
HVD-7		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Hologic Aptima HSV 1 & 2	1	*Negative	1	100.00%
HVD-8		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Hologic Aptima HSV 1 & 2	1	*Negative	1	100.00%
HVD-9		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Hologic Aptima HSV 1 & 2	1	*Positive	1	100.00%
HVD-10		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Hologic Aptima HSV 1 & 2	1	*Negative	1	100.00%
		HSV 2		
HVD-6		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Hologic Aptima HSV 1 & 2	2	*Positive	2	100.00%
HVD-7		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Hologic Aptima HSV 1 & 2	2	*Negative	2	100.00%
HVD-8		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Hologic Aptima HSV 1 & 2	2	*Positive	2	100.00%
HVD-9		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Hologic Aptima HSV 1 & 2	2	*Negative	2	100.00%
HVD-10		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	1	50.00%
		Negative	1	50.00%
Hologic Aptima HSV 1 & 2	2	*Positive	1	50.00%
Hologic Aptima HSV 1 & 2		Negative	1	50.00%
603 - Blood Cell Identification;610 - Regular Hematology	Labs	Response	Number	Percent
		Blood Cell Identification		
ID-7		* Denotes Correct Response		
AG: All Methods	63			
		*Auer rod	58	92.06%
		*Identification Unknown - Referred	2	3.17%
		*Immature/abnormal cell - Referred	2	3.17%
		Blast cell	1	1.59%
Blood Cell Microscopy	63	*Auer rod	58	92.06%
Blood Cell Microscopy		*Identification Unknown - Referred	2	3.17%
Blood Cell Microscopy		*Immature/abnormal cell - Referred	2	3.17%
Blood Cell Microscopy		Blast cell	1	1.59%
ID-8		* Denotes Correct Response		
AG: All Methods	63			
		*Identification Unknown - Referred	1	1.59%
		*Immature/abnormal cell - Referred	1	1.59%
		*Polychromatophilic (non-nucleated) red cell	54	85.71%
		Anisocytosis	1	1.59%



		Macrocyte (oval or round)	3	4.76%
		Reticulocyte (supravital stain)	2	3.17%
		Spherocyte	1	1.59%
Blood Cell Microscopy	63	*Identification Unknown - Referred	1	1.59%
Blood Cell Microscopy		*Immature/abnormal cell - Referred	1	1.59%
Blood Cell Microscopy		*Polychromatophilic (non-nucleated) red cell	54	85.71%
Blood Cell Microscopy		Anisocytosis	1	1.59%
Blood Cell Microscopy		Macrocyte (oval or round)	3	4.76%
Blood Cell Microscopy		Reticulocyte (supravital stain)	2	3.17%
Blood Cell Microscopy		Spherocyte	1	1.59%
ID-9		* Denotes Correct Response		
AG: All Methods	63			
		*Segmented or band neutrophil	61	96.83%
		Hypersegmented neutrophil	2	3.17%
Blood Cell Microscopy	63	*Segmented or band neutrophil	61	96.83%
Blood Cell Microscopy		Hypersegmented neutrophil	2	3.17%
ID-10		* Denotes Correct Response		
AG: All Methods	63			
		*Red Blood Cell, normal (RBC) (Erythrocyte)	41	65.08%
		Anisocytosis	1	1.59%
		Hypochromasia	4	6.35%
		Microcyte	8	12.70%
		Microcyte (with central pallor)	8	12.70%
		Microcyte (with hypochromasia)	1	1.59%
Blood Cell Microscopy	63	*Red Blood Cell, normal (RBC) (Erythrocyte)	41	65.08%
Blood Cell Microscopy		Anisocytosis	1	1.59%
Blood Cell Microscopy		Hypochromasia	4	6.35%
Blood Cell Microscopy		Microcyte	8	12.70%
Blood Cell Microscopy		Microcyte (with central pallor)	8	12.70%
Blood Cell Microscopy		Microcyte (with hypochromasia)	1	1.59%
ID-11		* Denotes Correct Response		
AG: All Methods	63			
		*Lymphocyte	61	96.83%
		Immature/abnormal cell - Referred	1	1.59%
		Lymphocyte, immature	1	1.59%
Blood Cell Microscopy	63	*Lymphocyte	61	96.83%
Blood Cell Microscopy		Immature/abnormal cell - Referred	1	1.59%
Blood Cell Microscopy		Lymphocyte, immature	1	1.59%
ID-12		* Denotes Correct Response		
AG: All Methods	63			
		*Identification Unknown - Referred	2	3.17%
		*Immature/abnormal cell - Referred	8	12.70%
		*Myelocyte cell	30	47.62%
		Basophil, any stage	1	1.59%
		Degenerated neutrophil	1	1.59%
		Lymphocyte, immature	2	3.17%
		Lymphocyte, reactive (atypical, variant, Downey)	3	4.76%
		Metamyelocyte (juvenile)	7	11.11%
		Monoblast	1	1.59%
		Monocyte	2	3.17%
		Prolymphocyte	1	1.59%
		Promyelocyte	4	6.35%
		Seg/band neutrophil with toxic granulation	1	1.59%
Blood Cell Microscopy	63	*Identification Unknown - Referred	2	3.17%
Blood Cell Microscopy		*Immature/abnormal cell - Referred	8	12.70%
Blood Cell Microscopy		*Myelocyte cell	30	47.62%
Blood Cell Microscopy		Basophil, any stage	1	1.59%
Blood Cell Microscopy		Degenerated neutrophil	1	1.59%
Blood Cell Microscopy		Lymphocyte, immature	2	3.17%
Blood Cell Microscopy		Lymphocyte, reactive (atypical, variant, Downey)	3	4.76%
Blood Cell Microscopy		Metamyelocyte (juvenile)	7	11.11%
Blood Cell Microscopy		Monoblast	1	1.59%
Blood Cell Microscopy		Monocyte	2	3.17%
Blood Cell Microscopy		Prolymphocyte	1	1.59%
Blood Cell Microscopy		Promyelocyte	4	6.35%
Blood Cell Microscopy		Seg/band neutrophil with toxic granulation	1	1.59%
		Blood Cell Identification (International)		
ID-7		* Denotes Correct Response		
AG: All Methods	1			



		*Auer rod	1	100.00%
Intl Blood Cell Microscopy	1	*Auer rod	1	100.00%
ID-8		* Denotes Correct Response		
AG: All Methods	1			
		*Polychromatophilic (non-nucleated) red cell	1	100.00%
Intl Blood Cell Microscopy	1	*Polychromatophilic (non-nucleated) red cell	1	100.00%
ID-9		* Denotes Correct Response		
AG: All Methods	1			
		*Segmented or band neutrophil	1	100.00%
Intl Blood Cell Microscopy	1	*Segmented or band neutrophil	1	100.00%
ID-10		* Denotes Correct Response		
AG: All Methods	1			
		*Red Blood Cell, normal (RBC) (Erythrocyte)	1	100.00%
Intl Blood Cell Microscopy	1	*Red Blood Cell, normal (RBC) (Erythrocyte)	1	100.00%
ID-11		* Denotes Correct Response		
AG: All Methods	1			
		*Lymphocyte	1	100.00%
Intl Blood Cell Microscopy	1	*Lymphocyte	1	100.00%
ID-12		* Denotes Correct Response		
AG: All Methods	2			
		*Myelocyte cell	1	50.00%
		Lymphocyte, reactive (atypical, variant, Downey)	1	50.00%
Intl Blood Cell Microscopy	2	*Myelocyte cell	1	50.00%
Intl Blood Cell Microscopy		Lymphocyte, reactive (atypical, variant, Downey)	1	50.00%
609 - Waived/PPM Tests ;675 - Infectious Mononucleosis;753 - Infectious Mononucleosis (Waived) (1 specimen);754 - Infectious Mono/RF Combo	Labs	Response	Number	Percent
		Infectious Mono (Waived)		
IM-6		* Denotes Correct Response		
AG: All Methods	463			
		*Positive	459	99.14%
		Negative	4	0.86%
Alere Acceava (S)	1	*Positive	1	100.00%
Alere Acceava (WB)	53	*Positive	53	100.00%
Sekisui Diagnostics OSOM Mono (WB)	181	*Positive	180	99.45%
Henry Schein One-Step+ Mono	47	*Positive	47	100.00%
Instant Technologies i-Screen (WB)	1	*Positive	1	100.00%
Inverness Signify Mono	4	*Positive	4	100.00%
LifeSign Status Inf Mono (WB)	30	*Positive	30	100.00%
McKesson Medi-Lab Perf - Inf Mono	9	*Positive	8	88.89%
Polymedco Poly stat Mono - (WB)	6	*Positive	6	100.00%
Quidel QuickVue+ (WB)	9	*Positive	9	100.00%
SP Brand (Cardinal) Rapid Test	31	*Positive	31	100.00%
Alere Clearview Mono Cassette	20	*Positive	20	100.00%
Pulse Scientific	2	*Positive	2	100.00%
Diag Test Group Clarity Mono	2	*Positive	2	100.00%
Fisher HC Sure-View Mono	6	*Positive	6	100.00%
Immunostics Inc. ImmunoDetector Mono (WB)	1	*Positive	1	100.00%
Meridian Immunocard STAT! Mono (WB)	6	*Positive	6	100.00%
NDC ProAdvantage Mono	2	*Positive	2	100.00%
Consult Diag Mono Cassette (WB)	37	*Positive	36	97.30%
Alere Acceava (WB) Mono Cassette	15	*Positive	14	93.33%
Sekisui Diagnostics OSOM Mono (WB)	181	Negative	1	0.55%
McKesson Medi-Lab Perf - Inf Mono	9	Negative	1	11.11%
Consult Diag Mono Cassette (WB)	37	Negative	1	2.70%
Alere Acceava (WB) Mono Cassette	15	Negative	1	6.67%
		Infectious Mononucleosis		
IM-6		* Denotes Correct Response		
AG: All Methods	43			
		*Positive	41	95.35%
		Negative	2	4.65%
Beck Coulter ICON (WB)	1	*Positive	1	100.00%
Alere Acceava (WB)	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	9	*Positive	9	100.00%
LifeSign Status Inf Mono (WB)	1	*Positive	1	100.00%
Lifesign Unistep	1	*Positive	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	2	*Positive	2	100.00%
Polymedco Poly stat Mono (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (S/P)	2	*Positive	2	100.00%



Quidel QuickVue+ (WB)	8	*Positive	8	100.00%
Seradyn Color Slide II	3	*Positive	3	100.00%
Alere Clearview Mono Cassette	8	*Positive	7	87.50%
Sekisui Diagnostics OSOM Mono Non-Waived	1	*Positive	1	100.00%
Consult Diag Mono Cassette (WB)	2	*Positive	2	100.00%
LifeSign Status Inf Mono (S/P)	1	*Positive	1	100.00%
Alere Clearview Mono Cassette	8	Negative	1	12.50%
Consult Diag Mono Cassette (Serum)	1	Negative	1	100.00%
675 - Infectious Mononucleosis;754 - Infectious Mono/RF Combo	Labs	Response	Number	Percent
		Infectious Mononucleosis		
IM-6		* Denotes Correct Response		
AG: All Methods	43			
		*Positive	41	95.35%
		Negative	2	4.65%
Beck Coulter ICON (WB)	1	*Positive	1	100.00%
Alere Acceava (WB)	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	9	*Positive	9	100.00%
LifeSign Status Inf Mono (WB)	1	*Positive	1	100.00%
Lifesign Unistep	1	*Positive	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	2	*Positive	2	100.00%
Polymedco Poly stat Mono (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (WB)	8	*Positive	8	100.00%
Seradyn Color Slide II	3	*Positive	3	100.00%
Alere Clearview Mono Cassette	8	*Positive	7	87.50%
Sekisui Diagnostics OSOM Mono Non-Waived	1	*Positive	1	100.00%
Consult Diag Mono Cassette (WB)	2	*Positive	2	100.00%
LifeSign Status Inf Mono (S/P)	1	*Positive	1	100.00%
Alere Clearview Mono Cassette	8	Negative	1	12.50%
Consult Diag Mono Cassette (Serum)	1	Negative	1	100.00%
IM-7		* Denotes Correct Response		
AG: All Methods	43			
		*Positive	42	97.67%
		Negative	1	2.33%
Beck Coulter ICON (WB)	1	*Positive	1	100.00%
Alere Acceava (WB)	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	9	*Positive	9	100.00%
LifeSign Status Inf Mono (WB)	1	*Positive	1	100.00%
Lifesign Unistep	1	*Positive	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	2	*Positive	2	100.00%
Polymedco Poly stat Mono (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (WB)	8	*Positive	8	100.00%
Seradyn Color Slide II	3	*Positive	3	100.00%
Alere Clearview Mono Cassette	8	*Positive	8	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	1	*Positive	1	100.00%
Consult Diag Mono Cassette (WB)	2	*Positive	2	100.00%
LifeSign Status Inf Mono (S/P)	1	*Positive	1	100.00%
Consult Diag Mono Cassette (Serum)	1	Negative	1	100.00%
IM-8		* Denotes Correct Response		
AG: All Methods	43			
		*Negative	42	97.67%
		Positive	1	2.33%
Beck Coulter ICON (WB)	1	*Negative	1	100.00%
Alere Acceava (WB)	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	9	*Negative	9	100.00%
LifeSign Status Inf Mono (WB)	1	*Negative	1	100.00%
Lifesign Unistep	1	*Negative	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	2	*Negative	2	100.00%
Polymedco Poly stat Mono (S/P)	2	*Negative	2	100.00%
Quidel QuickVue+ (S/P)	2	*Negative	2	100.00%
Quidel QuickVue+ (WB)	8	*Negative	8	100.00%
Seradyn Color Slide II	3	*Negative	3	100.00%
Alere Clearview Mono Cassette	8	*Negative	8	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	1	*Negative	1	100.00%
Consult Diag Mono Cassette (WB)	2	*Negative	2	100.00%
LifeSign Status Inf Mono (S/P)	1	*Negative	1	100.00%
Consult Diag Mono Cassette (Serum)	1	Positive	1	100.00%



IM-9		* Denotes Correct Response		
AG: All Methods	43			
		*Negative	42	97.67%
		Positive	1	2.33%
Beck Coulter ICON (WB)	1	*Negative	1	100.00%
Alere Acceava (WB)	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	9	*Negative	9	100.00%
LifeSign Status Inf Mono (WB)	1	*Negative	1	100.00%
Lifesign Unistep	1	*Negative	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	2	*Negative	2	100.00%
Polymedco Poly stat Mono (S/P)	2	*Negative	2	100.00%
Quidel QuickVue+ (S/P)	2	*Negative	2	100.00%
Quidel QuickVue+ (WB)	8	*Negative	8	100.00%
Seradyn Color Slide II	3	*Negative	3	100.00%
Alere Clearview Mono Cassette	8	*Negative	8	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	1	*Negative	1	100.00%
Consult Diag Mono Cassette (WB)	2	*Negative	2	100.00%
LifeSign Status Inf Mono (S/P)	1	*Negative	1	100.00%
Consult Diag Mono Cassette (Serum)	1	Positive	1	100.00%
IM-10		* Denotes Correct Response		
AG: All Methods	43			
		*Positive	43	100.00%
Beck Coulter ICON (WB)	1	*Positive	1	100.00%
Alere Acceava (WB)	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	9	*Positive	9	100.00%
LifeSign Status Inf Mono (WB)	1	*Positive	1	100.00%
Lifesign Unistep	1	*Positive	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	2	*Positive	2	100.00%
Polymedco Poly stat Mono (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (WB)	8	*Positive	8	100.00%
Seradyn Color Slide II	3	*Positive	3	100.00%
Alere Clearview Mono Cassette	8	*Positive	8	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	1	*Positive	1	100.00%
Consult Diag Mono Cassette (WB)	2	*Positive	2	100.00%
Consult Diag Mono Cassette (Serum)	1	*Positive	1	100.00%
LifeSign Status Inf Mono (S/P)	1	*Positive	1	100.00%
		Infectious Mono (Waived)		
IM-6		* Denotes Correct Response		
AG: All Methods	463			
		*Positive	459	99.14%
		Negative	4	0.86%
Alere Acceava (S)	1	*Positive	1	100.00%
Alere Acceava (WB)	53	*Positive	53	100.00%
Sekisui Diagnostics OSOM Mono (WB)	181	*Positive	180	99.45%
Henry Schein One-Step+ Mono	47	*Positive	47	100.00%
Instant Technologies i-Screen (WB)	1	*Positive	1	100.00%
Inverness Signify Mono	4	*Positive	4	100.00%
LifeSign Status Inf Mono (WB)	30	*Positive	30	100.00%
McKesson Medi-Lab Perf - Inf Mono	9	*Positive	8	88.89%
Polymedco Poly stat Mono - (WB)	6	*Positive	6	100.00%
Quidel QuickVue+ (WB)	9	*Positive	9	100.00%
SP Brand (Cardinal) Rapid Test	31	*Positive	31	100.00%
Alere Clearview Mono Cassette	20	*Positive	20	100.00%
Pulse Scientific	2	*Positive	2	100.00%
Diag Test Group Clarity Mono	2	*Positive	2	100.00%
Fisher HC Sure-Vue Mono	6	*Positive	6	100.00%
Immunostics Inc. ImmunoDetector Mono (WB)	1	*Positive	1	100.00%
Meridian Immunocard STAT! Mono (WB)	6	*Positive	6	100.00%
NDC ProAdvantage Mono	2	*Positive	2	100.00%
Consult Diag Mono Cassette (WB)	37	*Positive	36	97.30%
Alere Acceava (WB) Mono Cassette	15	*Positive	14	93.33%
Sekisui Diagnostics OSOM Mono (WB)	181	Negative	1	0.55%
McKesson Medi-Lab Perf - Inf Mono	9	Negative	1	11.11%
Consult Diag Mono Cassette (WB)	37	Negative	1	2.70%
Alere Acceava (WB) Mono Cassette	15	Negative	1	6.67%
		Infectious Mononucleosis (International)		
IM-6		* Denotes Correct Response		
AG: All Methods	2			



		*Positive	2	100.00%
Alere Clearview Mono Cassette	1	*Positive	1	100.00%
Fisher HC Sure-View Mono	1	*Positive	1	100.00%
IM-7		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Alere Clearview Mono Cassette	1	*Positive	1	100.00%
Fisher HC Sure-View Mono	1	*Positive	1	100.00%
IM-8		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Alere Clearview Mono Cassette	1	*Negative	1	100.00%
Fisher HC Sure-View Mono	1	*Negative	1	100.00%
IM-9		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Alere Clearview Mono Cassette	1	*Negative	1	100.00%
Fisher HC Sure-View Mono	1	*Negative	1	100.00%
IM-10		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Alere Clearview Mono Cassette	1	*Positive	1	100.00%
Fisher HC Sure-View Mono	1	*Positive	1	100.00%
637 - KOH - Glass Slides (Skin Hair Nail)	Labs	Response	Number	Percent
		KOH Preparation		
KOH-3		* Denotes Correct Response		
AG: All Methods	19			
		*Fungal elements/yeast observed	19	100.00%
KOH Glass Slide	19	*Fungal elements/yeast observed	19	100.00%
KOH-4		* Denotes Correct Response		
AG: All Methods	19			
		*Fungal elements/yeast observed	12	63.16%
		No fungal elements/yeast observed	7	36.84%
KOH Glass Slide	19	*Fungal elements/yeast observed	12	63.16%
KOH Glass Slide		No fungal elements/yeast observed	7	36.84%
654 - Lyme Serology	Labs	Response	Number	Percent
		Lyme Antibody IgG		
LY-2		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Quidel Sofia/Sofia 2 Lyme FIA	1	*Negative	1	100.00%
		Lyme Antibody IgM		
LY-2		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Quidel Sofia/Sofia 2 Lyme FIA	1	*Negative	1	100.00%
768 - MRSA (Molecular) (2 specimens)	Labs	Response	Number	Percent
		MRSA (Molecular)		
MRSA-6		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Cepheid GeneXpert - Xpert MRSA	1	*Negative	1	100.00%
MRSA-7		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Cepheid GeneXpert - Xpert MRSA	1	*Positive	1	100.00%
758 - Mycoplasma Antibody	Labs	Response	Number	Percent
		Mycoplasma Antibody		
MY-2		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Meridian ImmunoCard	2	*Positive	2	100.00%
791 - Fecal Occult Blood Add-On;601 - Waived/PPM Tests II;609 - Waived/PPM Tests I;647 - Mini-Waived Package	Labs	Response	Number	Percent
		Fecal Occult Blood		
OC-2		* Denotes Correct Response		



PG: Occult Blood / Guaiac methods	652			
		*Negative	608	93.25%
		Positive	44	6.75%
OC-2		* Denotes Correct Response		
PG: Occult Blood / Immunochemical methods	111			
		*Negative	110	99.10%
		Positive	1	0.90%
950 - Post Vasectomy Sperm Analysis	Labs	Response	Number	Percent
		Sperm Count (Post Vasectomy)		
PVAS-3		* Denotes Correct Response		
AG: All Methods	18			
		*Negative for Sperm	18	100.00%
Hemocytometer	7	*Negative for Sperm	7	100.00%
Post Vasectomy Microscopy	11	*Negative for Sperm	11	100.00%
PVAS-4		* Denotes Correct Response		
AG: All Methods	18			
		*Positive for Sperm	18	100.00%
Hemocytometer	7	*Positive for Sperm	7	100.00%
Post Vasectomy Microscopy	11	*Positive for Sperm	11	100.00%
754 - Infectious Mono/RF Combo;653 - Rheumatoid Factor	Labs	Response	Number	Percent
		Rheumatoid Factor		
RF-6		* Denotes Correct Response		
AG: All Methods	11			
		*Positive	10	90.91%
		N/A (Negative)	1	9.09%
ASI	1	*Positive	1	100.00%
Siemens RapiTex	1	*Positive	1	100.00%
Seradyn Color Slide	2	*Positive	2	100.00%
Stanbio	1	*Positive	1	100.00%
Wampole ColorCard RF	2	*Positive	2	100.00%
Wampole Rheumatex	2	*Positive	1	50.00%
Fisher HC Sure-Vue RF	1	*Positive	1	100.00%
Biokit Rhemajet RF	1	*Positive	1	100.00%
Wampole Rheumatex	2	N/A (Negative)	1	50.00%
RF-7		* Denotes Correct Response		
AG: All Methods	11			
		*N/A (Negative)	11	100.00%
ASI	1	*N/A (Negative)	1	100.00%
Siemens RapiTex	1	*N/A (Negative)	1	100.00%
Seradyn Color Slide	2	*N/A (Negative)	2	100.00%
Stanbio	1	*N/A (Negative)	1	100.00%
Wampole ColorCard RF	2	*N/A (Negative)	2	100.00%
Wampole Rheumatex	2	*N/A (Negative)	2	100.00%
Fisher HC Sure-Vue RF	1	*N/A (Negative)	1	100.00%
Biokit Rhemajet RF	1	*N/A (Negative)	1	100.00%
RF-8		* Denotes Correct Response		
AG: All Methods	11			
		*Positive	11	100.00%
ASI	1	*Positive	1	100.00%
Siemens RapiTex	1	*Positive	1	100.00%
Seradyn Color Slide	2	*Positive	2	100.00%
Stanbio	1	*Positive	1	100.00%
Wampole ColorCard RF	2	*Positive	2	100.00%
Wampole Rheumatex	2	*Positive	2	100.00%
Fisher HC Sure-Vue RF	1	*Positive	1	100.00%
Biokit Rhemajet RF	1	*Positive	1	100.00%
RF-9		* Denotes Correct Response		
AG: All Methods	11			
		*Positive	11	100.00%
ASI	1	*Positive	1	100.00%
Siemens RapiTex	1	*Positive	1	100.00%
Seradyn Color Slide	2	*Positive	2	100.00%
Stanbio	1	*Positive	1	100.00%
Wampole ColorCard RF	2	*Positive	2	100.00%
Wampole Rheumatex	2	*Positive	2	100.00%
Fisher HC Sure-Vue RF	1	*Positive	1	100.00%
Biokit Rhemajet RF	1	*Positive	1	100.00%



RF-10		* Denotes Correct Response		
AG: All Methods	11			
		*N/A (Negative)	11	100.00%
ASI	1	*N/A (Negative)	1	100.00%
Siemens RapiTex	1	*N/A (Negative)	1	100.00%
Seradyn Color Slide	2	*N/A (Negative)	2	100.00%
Stanbio	1	*N/A (Negative)	1	100.00%
Wampole ColorCard RF	2	*N/A (Negative)	2	100.00%
Wampole Rheumatex	2	*N/A (Negative)	2	100.00%
Fisher HC Sure-Vue RF	1	*N/A (Negative)	1	100.00%
Biokit Rhemajet RF	1	*N/A (Negative)	1	100.00%
		Rheumatoid Factor (International)		
RF-6		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Fisher HC Sure-Vue RF	2	*Positive	2	100.00%
Immunostics Rheumacol/LEX	1	*Positive	1	100.00%
RF-7		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Fisher HC Sure-Vue RF	2	*Negative	2	100.00%
Immunostics Rheumacol/LEX	1	*Negative	1	100.00%
RF-8		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Fisher HC Sure-Vue RF	2	*Positive	2	100.00%
Immunostics Rheumacol/LEX	1	*Positive	1	100.00%
RF-9		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Fisher HC Sure-Vue RF	2	*Positive	2	100.00%
Immunostics Rheumacol/LEX	1	*Positive	1	100.00%
RF-10		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Fisher HC Sure-Vue RF	2	*Negative	2	100.00%
Immunostics Rheumacol/LEX	1	*Negative	1	100.00%
		Rheumatoid Factor (Semi Quant)		
RF-6		* Denotes Correct Response		
PG: Fisher HC Sure-Vue RF (Semi Quant)	1			
		*80 IU/mL	1	100.00%
RF-7		* Denotes Correct Response		
PG: Fisher HC Sure-Vue RF (Semi Quant)	1			
		*Negative	1	100.00%
RF-8		* Denotes Correct Response		
PG: Fisher HC Sure-Vue RF (Semi Quant)	1			
		*160 IU/mL	1	100.00%
RF-9		* Denotes Correct Response		
PG: Fisher HC Sure-Vue RF (Semi Quant)	1			
		*80 IU/mL	1	100.00%
RF-10		* Denotes Correct Response		
PG: Fisher HC Sure-Vue RF (Semi Quant)	1			
		*Negative	1	100.00%
645 - Fetal Membrane Rupture	Labs	Response	Number	Percent
		Fetal Membranes Rupture PAMG-1		
ROM-3		* Denotes Correct Response		
AG: All Methods	18			
		*Negative	18	100.00%
AmniSure ROM Test	17	*Negative	17	100.00%
Actim PROM	1	*Negative	1	100.00%
ROM-4		* Denotes Correct Response		
AG: All Methods	18			
		*Positive	18	100.00%
AmniSure ROM Test	17	*Positive	17	100.00%
Actim PROM	1	*Positive	1	100.00%



601 - Waived/PPM Tests II;609 - Waived/PPM Tests I;785 - Presump ID/Colony Count-Urine & Strep Combo;691 - Group A Strep Antigen (5 specimens);783 - Group A Strep Antigen/Add-on (2 specimens);665 - Group A Strep Antigen (Waived) (2 specimens)	Labs	Response	Number	Percent
		Strep A Ag Det. (Waived)		
RS-6		* Denotes Correct Response		
AG: All Methods	813			
		*Positive	777	95.57%
		Negative	36	4.43%
Abbott Signify Strep A - Cassette	1	*Positive	1	100.00%
Abbott Signify Strep A - Waived	6	*Positive	6	100.00%
Alere Acceava Strep A Dipstick	112	*Positive	107	95.54%
Fisher HC Sure-Vue Strep A (Waived)	8	*Positive	8	100.00%
Sekisui Diagnostics OSOM	28	*Positive	28	100.00%
Sekisui Diagnostics OSOM Ultra Strep	61	*Positive	57	93.44%
Henry Schein OneStep + Rapid Strep Cass	6	*Positive	5	83.33%
Henry Schein OneStep+ Rapid Strep Dip	77	*Positive	76	98.70%
Innovacon One Step Strep Ag	1	*Positive	1	100.00%
Instant Technologies i Strep	1	*Positive	1	100.00%
Inverness Signify Strep A Dipstick	9	*Positive	9	100.00%
McKesson Strep A Cas	3	*Positive	2	66.67%
McKesson Strep A Dip	34	*Positive	32	94.12%
McKesson Strep A Test - Twist	1	*Positive	1	100.00%
Permaxim Rediscreen Strep A Rapid	1	*Positive	1	100.00%
Polymedco Poly Stat Strep A (W)	1	*Positive	1	100.00%
PSS Rapid Test Strip Strep A	2	*Positive	2	100.00%
Quidel QuickVue Dipstick Strep	84	*Positive	80	95.24%
Quidel QuickVue In-Line - 25 Test	73	*Positive	60	82.19%
Quidel QuickVue+ Strep A	31	*Positive	29	93.55%
Select Diag Strep A Dip	3	*Positive	3	100.00%
SP (Cardinal) Rapid Strep Cassette	4	*Positive	4	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	59	*Positive	58	98.31%
Alere Clearview Exact II Dip	11	*Positive	11	100.00%
Sekisui Diagnostics OSOM Ultra	7	*Positive	7	100.00%
Consult Diag Strep A Dipstick	95	*Positive	93	97.89%
Consult Diag Strep A Twist	1	*Positive	1	100.00%
MooreBrand Strep A Dipstick	3	*Positive	3	100.00%
NDC ProAdvantage Strep Strips	1	*Positive	1	100.00%
Meridian ImmunoCard STAT Strep A	1	*Positive	1	100.00%
LifeSign Bio Strep A Test STRip	1	*Positive	1	100.00%
Quidel Sofia/Sofia 2 Strep A FIA	27	*Positive	27	100.00%
BD Veritor Strep A	35	*Positive	35	100.00%
Alere ID NOW Strep A	16	*Positive	16	100.00%
Medline Strep A Dipstick	4	*Positive	4	100.00%
Roche Cobas Liat Strep A	4	*Positive	4	100.00%
Lifesign Status Strep A Plus	1	*Positive	1	100.00%
Alere Acceava Strep A Dipstick	112	Negative	5	4.46%
Sekisui Diagnostics OSOM Ultra Strep	61	Negative	4	6.56%
Henry Schein OneStep + Rapid Strep Cass	6	Negative	1	16.67%
Henry Schein OneStep+ Rapid Strep Dip	77	Negative	1	1.30%
McKesson Strep A Cas	3	Negative	1	33.33%
McKesson Strep A Dip	34	Negative	2	5.88%
Quidel QuickVue Dipstick Strep	84	Negative	4	4.76%
Quidel QuickVue In-Line - 25 Test	73	Negative	13	17.81%
Quidel QuickVue+ Strep A	31	Negative	2	6.45%
SP Brand (Cardinal) Rapid Strep A Dipstick	59	Negative	1	1.69%
Consult Diag Strep A Dipstick	95	Negative	2	2.11%
RS-7		* Denotes Correct Response		
AG: All Methods	149			
		*Negative	146	97.99%
		Positive	3	2.01%
Abbott Signify Strep A - Cassette	1	*Negative	1	100.00%
Abbott Signify Strep A - Waived	1	*Negative	1	100.00%
Alere Acceava Strep A Dipstick	6	*Negative	6	100.00%
Fisher HC Sure-Vue Strep A (Waived)	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM	5	*Negative	5	100.00%
Sekisui Diagnostics OSOM Ultra Strep	23	*Negative	22	95.65%
Henry Schein OneStep+ Rapid Strep Dip	13	*Negative	13	100.00%
Innovacon One Step Strep Ag	1	*Negative	1	100.00%



McKesson Strep A Dip	6	*Negative	6	100.00%
Quidel QuickVue Dipstick Strep	15	*Negative	14	93.33%
Quidel QuickVue In-Line - 25 Test	25	*Negative	24	96.00%
Quidel QuickVue+ Strep A	4	*Negative	4	100.00%
Select Diag Strep A Dip	1	*Negative	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	13	*Negative	13	100.00%
Sekisui Diagnostics OSOM Ultra	1	*Negative	1	100.00%
Consult Diag Strep A Dipstick	14	*Negative	14	100.00%
MooreBrand Strep A Dipstick	1	*Negative	1	100.00%
BD Veritor Strep A	8	*Negative	8	100.00%
Alere ID NOW Strep A	4	*Negative	4	100.00%
Medline Strep A Dipstick	2	*Negative	2	100.00%
Roche Cobas Liat Strep A	4	*Negative	4	100.00%
Sekisui Diagnostics OSOM Ultra Strep	23	Positive	1	4.35%
Quidel QuickVue Dipstick Strep	15	Positive	1	6.67%
Quidel QuickVue In-Line - 25 Test	25	Positive	1	4.00%
		Strep A Ag Det.		
RS-6		* Denotes Correct Response		
AG: All Methods	175			
		*Positive	174	99.43%
		Negative	1	0.57%
Abbott Signify Strep A - Waived	2	*Positive	2	100.00%
BD Check Strep A (Dip)	2	*Positive	2	100.00%
Beck Coulter ICON DS Strep	1	*Positive	1	100.00%
Alere Acceava Strep A Dipstick	2	*Positive	2	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM	15	*Positive	15	100.00%
Sekisui Diagnostics OSOM Ultra Strep	5	*Positive	5	100.00%
Henry Schein OneStep+ Rapid Strep Dip	7	*Positive	7	100.00%
Immunostics Immuno/Strep A Det	2	*Positive	2	100.00%
Inverness Signify Strep A Dipstick	1	*Positive	1	100.00%
McKesson Strep A Cas	2	*Positive	2	100.00%
McKesson Strep A Dip	9	*Positive	9	100.00%
Permaxim Rediscreen Strep A Rapid	1	*Positive	1	100.00%
PSS Rapid Test Strip Strep A	1	*Positive	1	100.00%
Quidel QuickVue Dipstick Strep	8	*Positive	8	100.00%
Quidel QuickVue In-Line - 25 Test	9	*Positive	9	100.00%
Quidel QuickVue+ Strep A	17	*Positive	16	94.12%
Select Diag Strep A Dip	1	*Positive	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	1	*Positive	1	100.00%
Stanbio QuStick Strep A	2	*Positive	2	100.00%
Consult Diag Strep A Dipstick	18	*Positive	18	100.00%
Meridian ImmunoCard STAT Strep A	1	*Positive	1	100.00%
Quidel Sofia/Sofia 2 Strep A FIA	10	*Positive	10	100.00%
BD Veritor Strep A	6	*Positive	6	100.00%
Quidel Solana	5	*Positive	5	100.00%
Cepheid GeneXpert - Xpert Strep A	46	*Positive	46	100.00%
Quidel QuickVue+ Strep A	17	Negative	1	5.88%
RS-7		* Denotes Correct Response		
AG: All Methods	175			
		*Negative	174	99.43%
		Positive	1	0.57%
Abbott Signify Strep A - Waived	2	*Negative	2	100.00%
BD Check Strep A (Dip)	2	*Negative	2	100.00%
Beck Coulter ICON DS Strep	1	*Negative	1	100.00%
Alere Acceava Strep A Dipstick	2	*Negative	2	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM	15	*Negative	15	100.00%
Sekisui Diagnostics OSOM Ultra Strep	5	*Negative	5	100.00%
Henry Schein OneStep+ Rapid Strep Dip	7	*Negative	6	85.71%
Immunostics Immuno/Strep A Det	2	*Negative	2	100.00%
Inverness Signify Strep A Dipstick	1	*Negative	1	100.00%
McKesson Strep A Cas	2	*Negative	2	100.00%
McKesson Strep A Dip	9	*Negative	9	100.00%
Permaxim Rediscreen Strep A Rapid	1	*Negative	1	100.00%
PSS Rapid Test Strip Strep A	1	*Negative	1	100.00%
Quidel QuickVue Dipstick Strep	8	*Negative	8	100.00%
Quidel QuickVue In-Line - 25 Test	9	*Negative	9	100.00%
Quidel QuickVue+ Strep A	17	*Negative	17	100.00%



Select Diag Strep A Dip	1	*Negative	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	1	*Negative	1	100.00%
Stanbio QuStick Strep A	2	*Negative	2	100.00%
Consult Diag Strep A Dipstick	18	*Negative	18	100.00%
Meridian ImmunoCard STAT Strep A	1	*Negative	1	100.00%
Quidel Sofia/Sofia 2 Strep A FIA	10	*Negative	10	100.00%
BD Veritor Strep A	6	*Negative	6	100.00%
Quidel Solana	5	*Negative	5	100.00%
Cepheid GeneXpert - Xpert Strep A	46	*Negative	46	100.00%
Henry Schein OneStep+ Rapid Strep Dip	7	Positive	1	14.29%
		Strep A Aq Det. (International)		
RS-6		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Alere Clearview Exact II Cas	1	*Positive	1	100.00%
RS-7		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Alere Clearview Exact II Cas	1	*Negative	1	100.00%
691 - Group A Strep Antigen (5 specimens)	Labs	Response	Number	Percent
		Strep A Aq Det.		
RS-6		* Denotes Correct Response		
AG: All Methods	175			
		*Positive	174	99.43%
		Negative	1	0.57%
Abbott Signify Strep A - Waived	2	*Positive	2	100.00%
BD Check Strep A (Dip)	2	*Positive	2	100.00%
Beck Coulter ICON DS Strep	1	*Positive	1	100.00%
Alere Acceava Strep A Dipstick	2	*Positive	2	100.00%
Fisher HC Sure-Vue Select Strep Aq	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM	15	*Positive	15	100.00%
Sekisui Diagnostics OSOM Ultra Strep	5	*Positive	5	100.00%
Henry Schein OneStep+ Rapid Strep Dip	7	*Positive	7	100.00%
Immunostics Immuno/Strep A Det	2	*Positive	2	100.00%
Inverness Signify Strep A Dipstick	1	*Positive	1	100.00%
McKesson Strep A Cas	2	*Positive	2	100.00%
McKesson Strep A Dip	9	*Positive	9	100.00%
Permaxim Rediscreen Strep A Rapid	1	*Positive	1	100.00%
PSS Rapid Test Strip Strep A	1	*Positive	1	100.00%
Quidel QuickVue Dipstick Strep	8	*Positive	8	100.00%
Quidel QuickVue In-Line - 25 Test	9	*Positive	9	100.00%
Quidel QuickVue+ Strep A	17	*Positive	16	94.12%
Select Diag Strep A Dip	1	*Positive	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	1	*Positive	1	100.00%
Stanbio QuStick Strep A	2	*Positive	2	100.00%
Consult Diag Strep A Dipstick	18	*Positive	18	100.00%
Meridian ImmunoCard STAT Strep A	1	*Positive	1	100.00%
Quidel Sofia/Sofia 2 Strep A FIA	10	*Positive	10	100.00%
BD Veritor Strep A	6	*Positive	6	100.00%
Quidel Solana	5	*Positive	5	100.00%
Cepheid GeneXpert - Xpert Strep A	46	*Positive	46	100.00%
Quidel QuickVue+ Strep A	17	Negative	1	5.88%
RS-7		* Denotes Correct Response		
AG: All Methods	175			
		*Negative	174	99.43%
		Positive	1	0.57%
Abbott Signify Strep A - Waived	2	*Negative	2	100.00%
BD Check Strep A (Dip)	2	*Negative	2	100.00%
Beck Coulter ICON DS Strep	1	*Negative	1	100.00%
Alere Acceava Strep A Dipstick	2	*Negative	2	100.00%
Fisher HC Sure-Vue Select Strep Aq	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM	15	*Negative	15	100.00%
Sekisui Diagnostics OSOM Ultra Strep	5	*Negative	5	100.00%
Henry Schein OneStep+ Rapid Strep Dip	7	*Negative	6	85.71%
Immunostics Immuno/Strep A Det	2	*Negative	2	100.00%
Inverness Signify Strep A Dipstick	1	*Negative	1	100.00%
McKesson Strep A Cas	2	*Negative	2	100.00%
McKesson Strep A Dip	9	*Negative	9	100.00%
Permaxim Rediscreen Strep A Rapid	1	*Negative	1	100.00%



PSS Rapid Test Strip Strep A	1	*Negative	1	100.00%
Quidel QuickVue Dipstick Strep	8	*Negative	8	100.00%
Quidel QuickVue In-Line - 25 Test	9	*Negative	9	100.00%
Quidel QuickVue+ Strep A	17	*Negative	17	100.00%
Select Diag Strep A Dip	1	*Negative	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	1	*Negative	1	100.00%
Stanbio QuStick Strep A	2	*Negative	2	100.00%
Consult Diag Strep A Dipstick	18	*Negative	18	100.00%
Meridian ImmunoCard STAT Strep A	1	*Negative	1	100.00%
Quidel Sofia/Sofia 2 Strep A FIA	10	*Negative	10	100.00%
BD Veritor Strep A	6	*Negative	6	100.00%
Quidel Solana	5	*Negative	5	100.00%
Cepheid GeneXpert - Xpert Strep A	46	*Negative	46	100.00%
Henry Schein OneStep+ Rapid Strep Dip	7	Positive	1	14.29%
RS-8		* Denotes Correct Response		
AG: All Methods	117			
		*Negative	116	99.15%
		Positive	1	0.85%
Abbott Signify Strep A - Waived	2	*Negative	2	100.00%
BD Check Strep A (Dip)	2	*Negative	2	100.00%
Beck Coulter ICON DS Strep	1	*Negative	1	100.00%
Alere Acceava Strep A Dipstick	2	*Negative	2	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM	15	*Negative	15	100.00%
Sekisui Diagnostics OSOM Ultra Strep	4	*Negative	4	100.00%
Henry Schein OneStep+ Rapid Strep Dip	7	*Negative	6	85.71%
Immunostics Immuno/Strep A Det	2	*Negative	2	100.00%
Inverness Signify Strep A Dipstick	1	*Negative	1	100.00%
McKesson Strep A Cas	1	*Negative	1	100.00%
McKesson Strep A Dip	4	*Negative	4	100.00%
Permaxim Rediscreen Strep A Rapid	1	*Negative	1	100.00%
PSS Rapid Test Strip Strep A	1	*Negative	1	100.00%
Quidel QuickVue Dipstick Strep	7	*Negative	7	100.00%
Quidel QuickVue In-Line - 25 Test	9	*Negative	9	100.00%
Quidel QuickVue+ Strep A	16	*Negative	16	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	1	*Negative	1	100.00%
Consult Diag Strep A Dipstick	14	*Negative	14	100.00%
Meridian ImmunoCard STAT Strep A	1	*Negative	1	100.00%
Quidel Sofia/Sofia 2 Strep A FIA	5	*Negative	5	100.00%
BD Veritor Strep A	1	*Negative	1	100.00%
Quidel Solana	3	*Negative	3	100.00%
Cepheid GeneXpert - Xpert Strep A	16	*Negative	16	100.00%
Henry Schein OneStep+ Rapid Strep Dip	7	Positive	1	14.29%
RS-9		* Denotes Correct Response		
AG: All Methods	117			
		*Positive	116	99.15%
		Negative	1	0.85%
Abbott Signify Strep A - Waived	2	*Positive	2	100.00%
BD Check Strep A (Dip)	2	*Positive	2	100.00%
Beck Coulter ICON DS Strep	1	*Positive	1	100.00%
Alere Acceava Strep A Dipstick	2	*Positive	2	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM	15	*Positive	15	100.00%
Sekisui Diagnostics OSOM Ultra Strep	4	*Positive	4	100.00%
Henry Schein OneStep+ Rapid Strep Dip	7	*Positive	7	100.00%
Immunostics Immuno/Strep A Det	2	*Positive	2	100.00%
Inverness Signify Strep A Dipstick	1	*Positive	1	100.00%
McKesson Strep A Cas	1	*Positive	1	100.00%
McKesson Strep A Dip	4	*Positive	4	100.00%
Permaxim Rediscreen Strep A Rapid	1	*Positive	1	100.00%
PSS Rapid Test Strip Strep A	1	*Positive	1	100.00%
Quidel QuickVue Dipstick Strep	7	*Positive	7	100.00%
Quidel QuickVue In-Line - 25 Test	9	*Positive	9	100.00%
Quidel QuickVue+ Strep A	16	*Positive	15	93.75%
SP Brand (Cardinal) Rapid Strep A Dipstick	1	*Positive	1	100.00%
Consult Diag Strep A Dipstick	14	*Positive	14	100.00%
Meridian ImmunoCard STAT Strep A	1	*Positive	1	100.00%
Quidel Sofia/Sofia 2 Strep A FIA	5	*Positive	5	100.00%
BD Veritor Strep A	1	*Positive	1	100.00%



Quidel Solana	3	*Positive	3	100.00%
Cepheid GeneXpert - Xpert Strep A	16	*Positive	16	100.00%
Quidel QuickVue+ Strep A	16	Negative	1	6.25%
RS-10		* Denotes Correct Response		
AG: All Methods	117			
		*Negative	114	97.44%
		Positive	3	2.56%
Abbott Signify Strep A - Waived	2	*Negative	2	100.00%
BD Check Strep A (Dip)	2	*Negative	2	100.00%
Beck Coulter ICON DS Strep	1	*Negative	1	100.00%
Alere Acceava Strep A Dipstick	2	*Negative	2	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM	15	*Negative	15	100.00%
Sekisui Diagnostics OSOM Ultra Strep	4	*Negative	4	100.00%
Henry Schein OneStep+ Rapid Strep Dip	7	*Negative	6	85.71%
Immunostics Immuno/Strep A Det	2	*Negative	2	100.00%
Inverness Signify Strep A Dipstick	1	*Negative	1	100.00%
McKesson Strep A Cas	1	*Negative	1	100.00%
McKesson Strep A Dip	4	*Negative	4	100.00%
Permaxim Rediscreen Strep A Rapid	1	*Negative	1	100.00%
PSS Rapid Test Strip Strep A	1	*Negative	1	100.00%
Quidel QuickVue Dipstick Strep	7	*Negative	7	100.00%
Quidel QuickVue In-Line - 25 Test	9	*Negative	9	100.00%
Quidel QuickVue+ Strep A	16	*Negative	15	93.75%
SP Brand (Cardinal) Rapid Strep A Dipstick	1	*Negative	1	100.00%
Consult Diag Strep A Dipstick	14	*Negative	14	100.00%
Quidel Sofia/Sofia 2 Strep A FIA	5	*Negative	5	100.00%
BD Veritor Strep A	1	*Negative	1	100.00%
Quidel Solana	3	*Negative	3	100.00%
Cepheid GeneXpert - Xpert Strep A	16	*Negative	16	100.00%
Henry Schein OneStep+ Rapid Strep Dip	7	Positive	1	14.29%
Quidel QuickVue+ Strep A	16	Positive	1	6.25%
Meridian ImmunoCard STAT Strep A	1	Positive	1	100.00%
		Strep A Ag Det. (Waived)		
RS-6		* Denotes Correct Response		
AG: All Methods	813			
		*Positive	777	95.57%
		Negative	36	4.43%
Abbott Signify Strep A - Cassette	1	*Positive	1	100.00%
Abbott Signify Strep A - Waived	6	*Positive	6	100.00%
Alere Acceava Strep A Dipstick	112	*Positive	107	95.54%
Fisher HC Sure-Vue Strep A (Waived)	8	*Positive	8	100.00%
Sekisui Diagnostics OSOM	28	*Positive	28	100.00%
Sekisui Diagnostics OSOM Ultra Strep	61	*Positive	57	93.44%
Henry Schein OneStep + Rapid Strep Cass	6	*Positive	5	83.33%
Henry Schein OneStep+ Rapid Strep Dip	77	*Positive	76	98.70%
Innovacon One Step Strep Ag	1	*Positive	1	100.00%
Instant Technologies i Strep	1	*Positive	1	100.00%
Inverness Signify Strep A Dipstick	9	*Positive	9	100.00%
McKesson Strep A Cas	3	*Positive	2	66.67%
McKesson Strep A Dip	34	*Positive	32	94.12%
McKesson Strep A Test - Twist	1	*Positive	1	100.00%
Permaxim Rediscreen Strep A Rapid	1	*Positive	1	100.00%
Polymedco Poly Stat Strep A (W)	1	*Positive	1	100.00%
PSS Rapid Test Strip Strep A	2	*Positive	2	100.00%
Quidel QuickVue Dipstick Strep	84	*Positive	80	95.24%
Quidel QuickVue In-Line - 25 Test	73	*Positive	60	82.19%
Quidel QuickVue+ Strep A	31	*Positive	29	93.55%
Select Diag Strep A Dip	3	*Positive	3	100.00%
SP (Cardinal) Rapid Strep Cassette	4	*Positive	4	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	59	*Positive	58	98.31%
Alere Clearview Exact II Dip	11	*Positive	11	100.00%
Sekisui Diagnostics OSOM Ultra	7	*Positive	7	100.00%
Consult Diag Strep A Dipstick	95	*Positive	93	97.89%
Consult Diag Strep A Twist	1	*Positive	1	100.00%
MooreBrand Strep A Dipstick	3	*Positive	3	100.00%
NDC ProAdvantage Strep Strips	1	*Positive	1	100.00%
Meridian ImmunoCard STAT Strep A	1	*Positive	1	100.00%
LifeSign Bio Strep A Test STrip	1	*Positive	1	100.00%



Quidel Sofia/Sofia 2 Strep A FIA	27	*Positive	27	100.00%
BD Veritor Strep A	35	*Positive	35	100.00%
Alere ID NOW Strep A	16	*Positive	16	100.00%
Medline Strep A Dipstick	4	*Positive	4	100.00%
Roche Cobas Liat Strep A	4	*Positive	4	100.00%
Lifesign Status Strep A Plus	1	*Positive	1	100.00%
Alere Acceava Strep A Dipstick	112	Negative	5	4.46%
Sekisui Diagnostics OSOM Ultra Strep	61	Negative	4	6.56%
Henry Schein OneStep + Rapid Strep Cass	6	Negative	1	16.67%
Henry Schein OneStep+ Rapid Strep Dip	77	Negative	1	1.30%
McKesson Strep A Cas	3	Negative	1	33.33%
McKesson Strep A Dip	34	Negative	2	5.88%
Quidel QuickVue Dipstick Strep	84	Negative	4	4.76%
Quidel QuickVue In-Line - 25 Test	73	Negative	13	17.81%
Quidel QuickVue+ Strep A	31	Negative	2	6.45%
SP Brand (Cardinal) Rapid Strep A Dipstick	59	Negative	1	1.69%
Consult Diag Strep A Dipstick	95	Negative	2	2.11%
RS-7		* Denotes Correct Response		
AG: All Methods	149			
		*Negative	146	97.99%
		Positive	3	2.01%
Abbott Signify Strep A - Cassette	1	*Negative	1	100.00%
Abbott Signify Strep A - Waived	1	*Negative	1	100.00%
Alere Acceava Strep A Dipstick	6	*Negative	6	100.00%
Fisher HC Sure-Vue Strep A (Waived)	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM	5	*Negative	5	100.00%
Sekisui Diagnostics OSOM Ultra Strep	23	*Negative	22	95.65%
Henry Schein OneStep+ Rapid Strep Dip	13	*Negative	13	100.00%
Innovacon One Step Strep Ag	1	*Negative	1	100.00%
McKesson Strep A Dip	6	*Negative	6	100.00%
Quidel QuickVue Dipstick Strep	15	*Negative	14	93.33%
Quidel QuickVue In-Line - 25 Test	25	*Negative	24	96.00%
Quidel QuickVue+ Strep A	4	*Negative	4	100.00%
Select Diag Strep A Dip	1	*Negative	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	13	*Negative	13	100.00%
Sekisui Diagnostics OSOM Ultra	1	*Negative	1	100.00%
Consult Diag Strep A Dipstick	14	*Negative	14	100.00%
MooreBrand Strep A Dipstick	1	*Negative	1	100.00%
BD Veritor Strep A	8	*Negative	8	100.00%
Alere ID NOW Strep A	4	*Negative	4	100.00%
Medline Strep A Dipstick	2	*Negative	2	100.00%
Roche Cobas Liat Strep A	4	*Negative	4	100.00%
Sekisui Diagnostics OSOM Ultra Strep	23	Positive	1	4.35%
Quidel QuickVue Dipstick Strep	15	Positive	1	6.67%
Quidel QuickVue In-Line - 25 Test	25	Positive	1	4.00%
RS-8		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Quidel QuickVue In-Line - 25 Test	1	*Negative	1	100.00%
Medline Strep A Dipstick	2	*Negative	2	100.00%
RS-9		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Quidel QuickVue In-Line - 25 Test	1	*Positive	1	100.00%
Medline Strep A Dipstick	2	*Positive	2	100.00%
RS-10		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Quidel QuickVue In-Line - 25 Test	1	*Negative	1	100.00%
Medline Strep A Dipstick	2	*Negative	2	100.00%
		Strep A Ag Det. (International)		
RS-6		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Alere Clearview Exact II Cas	1	*Positive	1	100.00%
RS-7		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Alere Clearview Exact II Cas	1	*Negative	1	100.00%



RS-8		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Alere Clearview Exact II Cas	1	*Negative	1	100.00%
RS-9		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Alere Clearview Exact II Cas	1	*Positive	1	100.00%
RS-10		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Alere Clearview Exact II Cas	1	*Negative	1	100.00%
739 - Sickel Cell Screen	Labs	Response	Number	Percent
		Sickel Cell Screen		
SS-2		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Pacific Hemostasis Sickel Screen	1	*Positive	1	100.00%
Streack SickelDex	1	*Positive	1	100.00%
679 - Syphilis Serology	Labs	Response	Number	Percent
		RPR		
SY-6		* Denotes Correct Response		
AG: All Methods	11			
		*Reactive	9	81.82%
		*Weakly Reactive	2	18.18%
ASI RPR	3	*Reactive	2	66.67%
Becton Dickinson BBL	4	*Reactive	4	100.00%
New Horizons	1	*Reactive	1	100.00%
Wampole Impact	1	*Reactive	1	100.00%
Intl ASI RPR	1	*Reactive	1	100.00%
ASI RPR	3	*Weakly Reactive	1	33.33%
Stanbio Quicktest	1	*Weakly Reactive	1	100.00%
SY-7		* Denotes Correct Response		
AG: All Methods	11			
		*Reactive	11	100.00%
ASI RPR	3	*Reactive	3	100.00%
Becton Dickinson BBL	4	*Reactive	4	100.00%
New Horizons	1	*Reactive	1	100.00%
Stanbio Quicktest	1	*Reactive	1	100.00%
Wampole Impact	1	*Reactive	1	100.00%
Intl ASI RPR	1	*Reactive	1	100.00%
SY-8		* Denotes Correct Response		
AG: All Methods	11			
		*Nonreactive	10	90.91%
		Reactive	1	9.09%
ASI RPR	3	*Nonreactive	3	100.00%
Becton Dickinson BBL	4	*Nonreactive	3	75.00%
New Horizons	1	*Nonreactive	1	100.00%
Stanbio Quicktest	1	*Nonreactive	1	100.00%
Wampole Impact	1	*Nonreactive	1	100.00%
Intl ASI RPR	1	*Nonreactive	1	100.00%
Becton Dickinson BBL	4	Reactive	1	25.00%
SY-9		* Denotes Correct Response		
AG: All Methods	11			
		*Reactive	6	54.55%
		*Weakly Reactive	4	36.36%
		Nonreactive	1	9.09%
ASI RPR	3	*Reactive	1	33.33%
Becton Dickinson BBL	4	*Reactive	3	75.00%
Wampole Impact	1	*Reactive	1	100.00%
Intl ASI RPR	1	*Reactive	1	100.00%
ASI RPR	3	*Weakly Reactive	2	66.67%
New Horizons	1	*Weakly Reactive	1	100.00%
Stanbio Quicktest	1	*Weakly Reactive	1	100.00%
Becton Dickinson BBL	4	Nonreactive	1	25.00%
SY-10		* Denotes Correct Response		
AG: All Methods	11			
		*Nonreactive	11	100.00%



ASI RPR	3	*Nonreactive	3	100.00%
Becton Dickinson BBL	4	*Nonreactive	4	100.00%
New Horizons	1	*Nonreactive	1	100.00%
Stanbio Quicktest	1	*Nonreactive	1	100.00%
Wampole Impact	1	*Nonreactive	1	100.00%
Intl ASI RPR	1	*Nonreactive	1	100.00%
		RPR/VDRL (titer)		
SY-6		* Denotes Correct Response		
AG: All Methods	3			
		*1:1	1	33.33%
		*1:2	2	66.67%
New Horizons	1	*1:1	1	100.00%
ASI RPR	1	*1:2	1	100.00%
Becton Dickinson BBL	1	*1:2	1	100.00%
SY-7		* Denotes Correct Response		
AG: All Methods	3			
		*1:2	1	33.33%
		*1:4	2	66.67%
New Horizons	1	*1:2	1	100.00%
ASI RPR	1	*1:4	1	100.00%
Becton Dickinson BBL	1	*1:4	1	100.00%
SY-8		* Denotes Correct Response		
AG: All Methods	3			
		*N/A (Negative)	3	100.00%
ASI RPR	1	*N/A (Negative)	1	100.00%
Becton Dickinson BBL	1	*N/A (Negative)	1	100.00%
New Horizons	1	*N/A (Negative)	1	100.00%
SY-9		* Denotes Correct Response		
AG: All Methods	3			
		*1:1	1	33.33%
		*1:2	2	66.67%
New Horizons	1	*1:1	1	100.00%
ASI RPR	1	*1:2	1	100.00%
Becton Dickinson BBL	1	*1:2	1	100.00%
SY-10		* Denotes Correct Response		
AG: All Methods	3			
		*N/A (Negative)	3	100.00%
ASI RPR	1	*N/A (Negative)	1	100.00%
Becton Dickinson BBL	1	*N/A (Negative)	1	100.00%
New Horizons	1	*N/A (Negative)	1	100.00%
		RPR (International)		
SY-6		* Denotes Correct Response		
AG: All Methods	3			
		*Reactive	2	66.67%
		*Weakly Reactive	1	33.33%
BD Macro-Vue RPR	1	*Reactive	1	100.00%
Wampole Impact	1	*Reactive	1	100.00%
Fisher HC Sure-Vue RPR	1	*Weakly Reactive	1	100.00%
SY-7		* Denotes Correct Response		
AG: All Methods	3			
		*Reactive	3	100.00%
BD Macro-Vue RPR	1	*Reactive	1	100.00%
Fisher HC Sure-Vue RPR	1	*Reactive	1	100.00%
Wampole Impact	1	*Reactive	1	100.00%
SY-8		* Denotes Correct Response		
AG: All Methods	3			
		*Nonreactive	3	100.00%
BD Macro-Vue RPR	1	*Nonreactive	1	100.00%
Fisher HC Sure-Vue RPR	1	*Nonreactive	1	100.00%
Wampole Impact	1	*Nonreactive	1	100.00%
SY-9		* Denotes Correct Response		
AG: All Methods	3			
		*Reactive	2	66.67%
		*Weakly Reactive	1	33.33%
BD Macro-Vue RPR	1	*Reactive	1	100.00%
Wampole Impact	1	*Reactive	1	100.00%
Fisher HC Sure-Vue RPR	1	*Weakly Reactive	1	100.00%
SY-10		* Denotes Correct Response		



AG: All Methods	3			
		*Nonreactive	3	100.00%
BD Macro-Vue RPR	1	*Nonreactive	1	100.00%
Fisher HC Sure-Vue RPR	1	*Nonreactive	1	100.00%
Wampole Impact	1	*Nonreactive	1	100.00%
776 - Throat Culture;780 - Throat & Genital Cultures	Labs	Response	Number	Percent
		Throat Culture		
TC-6		* Denotes Correct Response		
AG: All Methods	59			
		*Negative for Group A Streptococcus	49	83.05%
		*Neisseria gonorrhoeae	1	1.69%
		*No beta hemolysis	1	1.69%
		*No Growth - Strep Selective medium	8	13.56%
Cultures	59	*Negative for Group A Streptococcus	49	83.05%
Cultures		*Neisseria gonorrhoeae	1	1.69%
Cultures		*No beta hemolysis	1	1.69%
Cultures		*No Growth - Strep Selective medium	8	13.56%
TC-7		* Denotes Correct Response		
AG: All Methods	59			
		*Beta hemolysis, referred for ID/confirmation	2	3.39%
		*Positive for Group A Streptococcus	33	55.93%
		*Presumptive Positive for Group A Streptococcus	22	37.29%
		*Presumptive Streptococcus species	1	1.69%
		*Streptococcus pyogenes	1	1.69%
Cultures	59	*Beta hemolysis, referred for ID/confirmation	2	3.39%
Cultures		*Positive for Group A Streptococcus	33	55.93%
Cultures		*Presumptive Positive for Group A Streptococcus	22	37.29%
Cultures		*Presumptive Streptococcus species	1	1.69%
Cultures		*Streptococcus pyogenes	1	1.69%
TC-8		* Denotes Correct Response		
AG: All Methods	60			
		*Haemophilus influenzae	1	1.67%
		*Negative for Group A Streptococcus	49	81.67%
		*No beta hemolysis	1	1.67%
		*No Growth - Strep Selective medium	3	5.00%
		Beta hemolysis, referred for ID/confirmation	1	1.67%
		Positive for Group A Streptococcus	2	3.33%
		Presumptive Positive for Group A Streptococcus	3	5.00%
Cultures	60	*Haemophilus influenzae	1	1.67%
Cultures		*Negative for Group A Streptococcus	49	81.67%
Cultures		*No beta hemolysis	1	1.67%
Cultures		*No Growth - Strep Selective medium	3	5.00%
Cultures		Beta hemolysis, referred for ID/confirmation	1	1.67%
Cultures		Positive for Group A Streptococcus	2	3.33%
Cultures		Presumptive Positive for Group A Streptococcus	3	5.00%
TC-9		* Denotes Correct Response		
AG: All Methods	73			
		*Beta hemolysis, referred for ID/confirmation	6	8.22%
		*Negative for Group A Streptococcus	54	73.97%
		*Presumptive Streptococcus species	1	1.37%
		*Streptococcus, beta-hemolytic, not Group A	4	5.48%
		No beta hemolysis	1	1.37%
		No Growth - Strep Selective medium	3	4.11%
		Positive for Group A Streptococcus	2	2.74%
		Presumptive Positive for Group A Streptococcus	2	2.74%
Cultures	73	*Beta hemolysis, referred for ID/confirmation	6	8.22%
Cultures		*Negative for Group A Streptococcus	54	73.97%
Cultures		*Presumptive Streptococcus species	1	1.37%
Cultures		*Streptococcus, beta-hemolytic, not Group A	4	5.48%
Cultures		No beta hemolysis	1	1.37%
Cultures		No Growth - Strep Selective medium	3	4.11%
Cultures		Positive for Group A Streptococcus	2	2.74%
Cultures		Presumptive Positive for Group A Streptococcus	2	2.74%
TC-10		* Denotes Correct Response		
AG: All Methods	73			
		*Beta hemolysis, referred for ID/confirmation	4	5.48%
		*Positive for Group A Streptococcus	44	60.27%
		*Presumptive Positive for Group A Streptococcus	23	31.51%
		*Presumptive Streptococcus species	1	1.37%



		*Streptococcus pyogenes	1	1.37%
Cultures	73	*Beta hemolysis, referred for ID/confirmation	4	5.48%
Cultures		*Positive for Group A Streptococcus	44	60.27%
Cultures		*Presumptive Positive for Group A Streptococcus	23	31.51%
Cultures		*Presumptive Streptococcus species	1	1.37%
Cultures		*Streptococcus pyogenes	1	1.37%
640 - Urinalysis;642 - Urinalysis/hCG;639 - Complete Urinalysis Package;601 - Waived/PPM Tests II;609 - Waived/PPM Tests I;647 - Mini-Waived Package	Labs	Response	Number	Percent
		Urine Sp Gravity		
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 10 / 100	12			
		*1.015	5	41.67%
		*1.020	6	50.00%
		*1.025	1	8.33%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 50	63			
		*1.015	10	15.87%
		*1.020	36	57.14%
		*1.025	15	23.81%
		1.030	2	3.17%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 500	5			
		*1.015	1	20.00%
		*1.020	1	20.00%
		*1.025	2	40.00%
		1.030	1	20.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Status	637			
		*1.015	67	10.52%
		*1.020	451	70.80%
		*1.025	108	16.95%
		1.030	11	1.73%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Status / Siemens - Plus Sys / Siemens Urine Dipstick	14			
		*1.015	2	14.29%
		*1.020	8	57.14%
		*1.025	3	21.43%
		>=1.030	1	7.14%
UA-2		* Denotes Correct Response		
PG: Siemens Multistix - Visual	202			
		*1.015	32	15.84%
		*1.020	113	55.94%
		*1.025	49	24.26%
		1.010	1	0.50%
		1.030	7	3.47%
UA-2		* Denotes Correct Response		
PG: Henry Schein Urispec	6			
		*1.015	3	50.00%
		*1.020	3	50.00%
UA-2		* Denotes Correct Response		
PG: Consult Diagnostics Urine Reagent Strips	5			
		*1.020	1	20.00%
		*1.025	3	60.00%
		1.030	1	20.00%
UA-2		* Denotes Correct Response		
PG: Diag Test Group Clarity Urochek	24			
		*1.015	1	4.17%
		*1.020	6	25.00%
		*1.025	13	54.17%
		1.030	4	16.67%
UA-2		* Denotes Correct Response		
PG: Roche Chemstrips - Visual	82			
		*1.015	47	57.32%
		*1.020	26	31.71%
		*1.025	3	3.66%
		1.000	1	1.22%
		1.010	4	4.88%



		1.030	1	1.22%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Advantus	24			
		*1.015	1	4.17%
		*1.020	20	83.33%
		*1.025	3	12.50%
UA-2		* Denotes Correct Response		
PG: Consult Diag Urinalysis Test Strips	8			
		*1.015	1	12.50%
		*1.020	5	62.50%
		*1.025	2	25.00%
UA-2		* Denotes Correct Response		
PG: Consult Diag Urine Analyzer	7			
		*1.020	1	14.29%
		*1.025	5	71.43%
		1.030	1	14.29%
UA-2		* Denotes Correct Response		
PG: 120 Urine Analyzer	44			
		*1.025	37	84.09%
		*1.030	7	15.91%
UA-2		* Denotes Correct Response		
PG: McKesson 10 SG Urine Dipstick	7			
		*1.015	1	14.29%
		*1.020	4	57.14%
		*1.030	2	28.57%
UA-2		* Denotes Correct Response		
SG: Select Siemens Clinitek/Visual	944			
		*1.015	116	12.29%
		*1.020	628	66.53%
		*1.025	178	18.86%
		1.010	1	0.11%
		1.030	21	2.22%
UA-2		* Denotes Correct Response		
SG: Select Siemens Clinitek/Visual / Siemens - Plus Sys / Siemens Urine Dipstick	15			
		*1.015	2	13.33%
		*1.020	8	53.33%
		*1.025	4	26.67%
		>=1.030	1	6.67%
UA-2		* Denotes Correct Response		
SG: Select Roche Urinalysis Inst/Visual	89			
		*1.015	49	55.06%
		*1.020	29	32.58%
		*1.025	3	3.37%
		1.000	1	1.12%
		1.005	1	1.12%
		1.010	4	4.49%
		1.030	2	2.25%
UA-2		* Denotes Correct Response		
SG: Select Consult Diag Urinalysis Inst/Visual	20			
		*1.015	1	5.00%
		*1.020	7	35.00%
		*1.025	10	50.00%
		1.030	2	10.00%
UA-2		* Denotes Correct Response		
SG: Select McKesson Urinalysis Inst/Visual	51			
		*1.015	1	1.96%
		*1.020	4	7.84%
		*1.025	37	72.55%
		*1.030	9	17.65%
UA-2		* Denotes Correct Response		
SG: Select Henry Schein Urinalysis Inst/Visual	9			
		*1.015	4	44.44%
		*1.020	5	55.56%
UA-2		* Denotes Correct Response		
MG: Method Group	1163			
		*1.015	174	14.96%
		*1.020	688	59.16%



		*1.025	251	21.58%
		*1.030	41	3.53%
		1.000	1	0.09%
		1.005	1	0.09%
		1.010	7	0.60%
		1.030	41	3.53%
UA-2		* Denotes Correct Response		
MG: Roche - Conventional	2			
		*1.020	2	100.00%
UA-2		* Denotes Correct Response		
MG: Siemens - Plus Sys	15			
		*1.015	2	13.33%
		*1.020	8	53.33%
		*1.025	4	26.67%
		>=1.030	1	6.67%
		Urine Bilirubin		
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 50	63			
		*Negative	63	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 500	5			
		*Negative	5	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Status	630			
		*Negative	630	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Status / Siemens - Plus Sys / Siemens Urine Dipstick	14			
		*Negative	14	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Multistix - Visual	198			
		*Negative	195	98.48%
		Large (3+)	1	0.51%
		Moderate (2+)	2	1.01%
UA-2		* Denotes Correct Response		
PG: Henry Schein Urispec	6			
		*Negative	6	100.00%
UA-2		* Denotes Correct Response		
PG: Consult Diagnostics Urine Reagent Strips	5			
		*Negative	5	100.00%
UA-2		* Denotes Correct Response		
PG: Diag Test Group Clarity Urochek	23			
		*Negative	23	100.00%
UA-2		* Denotes Correct Response		
PG: Roche Chemstrips - Visual	82			
		*Negative	82	100.00%
UA-2		* Denotes Correct Response		
PG: UriScan Reagent Strips	5			
		*Negative	5	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Advantus	24			
		*Negative	24	100.00%
UA-2		* Denotes Correct Response		
PG: Consult Diag Urinalysis Test Strips	8			
		*Negative	8	100.00%
UA-2		* Denotes Correct Response		
PG: Consult Diag Urine Analyzer	7			
		*Negative	7	100.00%
UA-2		* Denotes Correct Response		
PG: 120 Urine Analyzer	44			
		*Negative	44	100.00%
UA-2		* Denotes Correct Response		
PG: McKesson 10 SG Urine Dipstick	7			
		*Negative	7	100.00%
UA-2		* Denotes Correct Response		
SG: Select Siemens Clinitek/Visual	933			
		*Negative	930	99.68%
		Large (3+)	1	0.11%



		Moderate (2+)	2	0.21%
UA-2		* Denotes Correct Response		
SG: Select Siemens Clinitek/Visual / Siemens - Plus Sys / Siemens Urine Dipstick	15			
		*Negative	15	100.00%
UA-2		* Denotes Correct Response		
SG: Select Roche Urinalysis Inst/Visual	89			
		*Negative	89	100.00%
UA-2		* Denotes Correct Response		
SG: Select Consult Diag Urinalysis Inst/Visual	20			
		*Negative	20	100.00%
UA-2		* Denotes Correct Response		
SG: Select McKesson Urinalysis Inst/Visual	51			
		*Negative	51	100.00%
UA-2		* Denotes Correct Response		
SG: Select Henry Schein Urinalysis Inst/Visual	9			
		*Negative	9	100.00%
UA-2		* Denotes Correct Response		
MG: Method Group	1151			
		*Negative	1148	99.74%
		Large (3+)	1	0.09%
		Moderate (2+)	2	0.17%
UA-2		* Denotes Correct Response		
MG: Roche - Conventional	2			
		*Negative	2	100.00%
UA-2		* Denotes Correct Response		
MG: Siemens - Plus Sys	15			
		*Negative	15	100.00%
		Urine Ketones		
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 10 / 100	12			
		*Large (3+) / 80 mg/dL	2	16.67%
		*Moderate (2+) / 40 mg/dL	9	75.00%
		Negative	1	8.33%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 50	63			
		*Large (3+) / 80 mg/dL	3	4.76%
		*Moderate (2+) / 40 mg/dL	58	92.06%
		Small (1+) / 15mg/dL	2	3.17%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 500	5			
		*Moderate (2+) / 40 mg/dL	5	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Status	637			
		*Large (3+) / 80 mg/dL	42	6.59%
		*Moderate (2+) / 40 mg/dL	579	90.89%
		>=160 mg/dL	1	0.16%
		Negative	5	0.78%
		Small (1+) / 15mg/dL	10	1.57%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Status / Siemens - Plus Sys / Siemens Urine Dipstick	14			
		*1+	2	14.29%
		*2+	11	78.57%
		*3+	1	7.14%
UA-2		* Denotes Correct Response		
PG: Siemens Multistix - Visual	204			
		*Large (3+) / 80 mg/dL	89	43.63%
		*Moderate (2+) / 40 mg/dL	107	52.45%
		Negative	3	1.47%
		Small (1+) / 15mg/dL	5	2.45%
UA-2		* Denotes Correct Response		
PG: Henry Schein Urispec	6			
		*Small (1+) / 15mg/dL	6	100.00%
UA-2		* Denotes Correct Response		
PG: Consult Diagnostics Urine Reagent Strips	5			
		*Moderate (2+) / 40 mg/dL	4	80.00%
		Small (1+) / 15mg/dL	1	20.00%



UA-2		* Denotes Correct Response		
PG: Diag Test Group Clarity Urochek	24			
		*Large (3+) / 80 mg/dL	8	33.33%
		*Moderate (2+) / 40 mg/dL	15	62.50%
		Small (1+) / 15mg/dL	1	4.17%
UA-2		* Denotes Correct Response		
PG: Roche Chemstrips - Visual	84			
		*Large (3+) / 80 mg/dL	44	52.38%
		*Moderate (2+) / 40 mg/dL	36	42.86%
		Negative	3	3.57%
		Small (1+) / 15mg/dL	1	1.19%
UA-2		* Denotes Correct Response		
PG: UriScan Reagent Strips	5			
		*Large (3+) / 80 mg/dL	2	40.00%
		*Moderate (2+) / 40 mg/dL	2	40.00%
		Small (1+) / 15mg/dL	1	20.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Advantus	24			
		*Moderate (2+) / 40 mg/dL	18	75.00%
		*Small (1+) / 15mg/dL	6	25.00%
UA-2		* Denotes Correct Response		
PG: Consult Diag Urinalysis Test Strips	8			
		*Large (3+) / 80 mg/dL	3	37.50%
		*Moderate (2+) / 40 mg/dL	5	62.50%
UA-2		* Denotes Correct Response		
PG: Consult Diag Urine Analyzer	7			
		*Large (3+) / 80 mg/dL	1	14.29%
		*Moderate (2+) / 40 mg/dL	6	85.71%
UA-2		* Denotes Correct Response		
PG: 120 Urine Analyzer	44			
		*Large (3+) / 80 mg/dL	14	31.82%
		*Moderate (2+) / 40 mg/dL	30	68.18%
UA-2		* Denotes Correct Response		
SG: Select Siemens Clinitek/Visual	946			
		*Large (3+) / 80 mg/dL	136	14.38%
		*Moderate (2+) / 40 mg/dL	777	82.14%
		*Small (1+) / 15mg/dL	23	2.43%
		>=160 mg/dL	1	0.11%
		Negative	9	0.95%
		Small (1+) / 15mg/dL	23	2.43%
UA-2		* Denotes Correct Response		
SG: Select Siemens Clinitek/Visual / Siemens - Plus Sys / Siemens Urine Dipstick	15			
		*1+	3	20.00%
		*2+	11	73.33%
		*3+	1	6.67%
UA-2		* Denotes Correct Response		
SG: Select Roche Urinalysis Inst/Visual	91			
		*Large (3+) / 80 mg/dL	47	51.65%
		*Moderate (2+) / 40 mg/dL	40	43.96%
		Negative	3	3.30%
		Small (1+) / 15mg/dL	1	1.10%
UA-2		* Denotes Correct Response		
SG: Select Consult Diag Urinalysis Inst/Visual	20			
		*Large (3+) / 80 mg/dL	4	20.00%
		*Moderate (2+) / 40 mg/dL	15	75.00%
		Small (1+) / 15mg/dL	1	5.00%
UA-2		* Denotes Correct Response		
SG: Select McKesson Urinalysis Inst/Visual	51			
		*Large (3+) / 80 mg/dL	17	33.33%
		*Moderate (2+) / 40 mg/dL	32	62.75%
		Negative	1	1.96%
		Small (1+) / 15mg/dL	1	1.96%
UA-2		* Denotes Correct Response		
SG: Select Henry Schein Urinalysis Inst/Visual	9			
		*Moderate (2+) / 40 mg/dL	2	22.22%
		*Small (1+) / 15mg/dL	7	77.78%
UA-2		* Denotes Correct Response		



MG: Method Group	1168			
		*Large (3+) / 80 mg/dL	219	18.75%
		*Moderate (2+) / 40 mg/dL	896	76.71%
		*Small (1+) / 15mg/dL	38	3.25%
		>=160 mg/dL	1	0.09%
		Negative	13	1.11%
		Small (1+) / 15mg/dL	38	3.25%
		Trace/5 mg/dL	1	0.09%
UA-2		* Denotes Correct Response		
MG: Roche - Conventional	2			
		*15 mg/dL	1	50.00%
		*150 mg/dL	1	50.00%
UA-2		* Denotes Correct Response		
MG: Siemens - Plus Sys	15			
		*1+	3	20.00%
		*2+	11	73.33%
		*3+	1	6.67%
		Urine pH		
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 10 / 100	12			
		*5 or less	9	75.00%
		*5.5	2	16.67%
		6	1	8.33%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 50	63			
		*5 or less	28	44.44%
		*5.5	34	53.97%
		6	1	1.59%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 500	5			
		*5 or less	1	20.00%
		*5.5	2	40.00%
		*6	2	40.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Status	637			
		*5 or less	247	38.78%
		*5.5	379	59.50%
		6	7	1.10%
		6.5	1	0.16%
		7	3	0.47%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Status / Siemens - Plus Sys / Siemens Urine Dipstick	14			
		*5.0	6	42.86%
		*5.5	8	57.14%
UA-2		* Denotes Correct Response		
PG: Siemens Multistix - Visual	203			
		*5 or less	106	52.22%
		*5.5	40	19.70%
		*6	57	28.08%
UA-2		* Denotes Correct Response		
PG: Diag Test Group Clarity Urochek	24			
		*5 or less	7	29.17%
		*5.5	6	25.00%
		*6	11	45.83%
UA-2		* Denotes Correct Response		
PG: Roche Chemstrips - Visual	84			
		*5 or less	78	92.86%
		*5.5	4	4.76%
		6	1	1.19%
		7	1	1.19%
UA-2		* Denotes Correct Response		
PG: Roche Urisys / Roche - Conventional / Roche Urine Dipstick	2			
		*5	2	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Advantus	24			
		*5.5	14	58.33%
		*6	8	33.33%



		5 or less	2	8.33%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Advantus / Siemens - Plus Sys / Siemens Urine Dipstick	1			
		*5.5	1	100.00%
UA-2		* Denotes Correct Response		
PG: Consult Diag Urinalysis Test Strips	8			
		*5 or less	5	62.50%
		*5.5	1	12.50%
		*6	2	25.00%
UA-2		* Denotes Correct Response		
PG: 120 Urine Analyzer	44			
		*5.5	18	40.91%
		*6	23	52.27%
		5 or less	3	6.82%
UA-2		* Denotes Correct Response		
SG: Select Siemens Clinitek/Visual	946			
		*5 or less	394	41.65%
		*5.5	472	49.89%
		*6	76	8.03%
		5 or less	394	41.65%
		6	76	8.03%
		6.5	1	0.11%
		7	3	0.32%
UA-2		* Denotes Correct Response		
SG: Select Siemens Clinitek/Visual / Siemens - Plus Sys / Siemens Urine Dipstick	15			
		*5.0	6	40.00%
		*5.5	9	60.00%
UA-2		* Denotes Correct Response		
SG: Select Roche Urinalysis Inst/Visual	91			
		*5 or less	83	91.21%
		*5.5	5	5.49%
		*6	2	2.20%
		6	2	2.20%
		7	1	1.10%
UA-2		* Denotes Correct Response		
SG: Select Consult Diag Urinalysis Inst/Visual	20			
		*5 or less	9	45.00%
		*5.5	4	20.00%
		*6	7	35.00%
UA-2		* Denotes Correct Response		
SG: Select McKesson Urinalysis Inst/Visual	51			
		*5 or less	9	17.65%
		*5.5	18	35.29%
		*6	24	47.06%
		5 or less	9	17.65%
UA-2		* Denotes Correct Response		
SG: Select Henry Schein Urinalysis Inst/Visual	9			
		*5 or less	7	77.78%
		*5.5	2	22.22%
UA-2		* Denotes Correct Response		
SG: Select Medi-Lab Performance UA Inst/Visual	3			
		*5 or less	1	33.33%
		*5.5	1	33.33%
		*6	1	33.33%
UA-2		* Denotes Correct Response		
PG: URS - Urine Reagent Strips	5			
		*5 or less	3	60.00%
		*5.5	1	20.00%
		*6	1	20.00%
UA-2		* Denotes Correct Response		
MG: Method Group	1169			
		*5 or less	523	44.74%
		*5.5	515	44.05%
		*6	126	10.78%
		5 or less	523	44.74%
		6	126	10.78%
		6.5	1	0.09%



		7	4	0.34%
UA-2		* Denotes Correct Response		
MG: Roche - Conventional	2			
		*5	2	100.00%
UA-2		* Denotes Correct Response		
MG: Siemens - Plus Sys	15			
		*5.0	6	40.00%
		*5.5	9	60.00%
		Urine Protein - Qual		
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 10 / 100	12			
		*Negative	12	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 50	63			
		*Negative	63	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 500	5			
		*Negative	5	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Status	638			
		*Negative	634	99.37%
		300-500 (3+)	3	0.47%
		Trace	1	0.16%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Status / Siemens - Plus Sys / Siemens Urine Dipstick	14			
		*Negative	14	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Multistix - Visual	205			
		*Negative	196	95.61%
		30 (1+)	1	0.49%
		Trace	8	3.90%
UA-2		* Denotes Correct Response		
PG: Henry Schein Urispec	6			
		*Negative	6	100.00%
UA-2		* Denotes Correct Response		
PG: Consult Diagnostics Urine Reagent Strips	5			
		*Negative	5	100.00%
UA-2		* Denotes Correct Response		
PG: Diag Test Group Clarity Urochek	24			
		*Negative	23	95.83%
		Trace	1	4.17%
UA-2		* Denotes Correct Response		
PG: Roche Chemstrips - Visual	90			
		*Negative	82	91.11%
		100 (2+)	2	2.22%
		300-500 (3+)	1	1.11%
		Trace	5	5.56%
UA-2		* Denotes Correct Response		
PG: UriScan Reagent Strips	6			
		*Negative	6	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Advantus	24			
		*Negative	24	100.00%
UA-2		* Denotes Correct Response		
PG: Consult Diag Urine Analyzer	7			
		*Negative	7	100.00%
UA-2		* Denotes Correct Response		
PG: 120 Urine Analyzer	44			
		*Negative	44	100.00%
UA-2		* Denotes Correct Response		
SG: Select Siemens Clinitek/Visual	949			
		*Negative	936	98.63%
		30 (1+)	1	0.11%
		300-500 (3+)	3	0.32%
		Trace	9	0.95%
UA-2		* Denotes Correct Response		



SG: Select Siemens Clinitek/Visual / Siemens - Plus Sys / Siemens Urine Dipstick	15			
		*Negative	15	100.00%
UA-2		* Denotes Correct Response		
SG: Select Roche Urinalysis Inst/Visual	97			
		*Negative	89	91.75%
		100 (2+)	2	2.06%
		300-500 (3+)	1	1.03%
		Trace	5	5.15%
UA-2		* Denotes Correct Response		
SG: Select Consult Diag Urinalysis Inst/Visual	20			
		*Negative	18	90.00%
		Trace	2	10.00%
UA-2		* Denotes Correct Response		
SG: Select McKesson Urinalysis Inst/Visual	51			
		*Negative	49	96.08%
		Trace	2	3.92%
UA-2		* Denotes Correct Response		
SG: Select Henry Schein Urinalysis Inst/Visual	9			
		*Negative	9	100.00%
UA-2		* Denotes Correct Response		
PG: URS - Urine Reagent Strips	5			
		*Negative	5	100.00%
UA-2		* Denotes Correct Response		
MG: Method Group	1178			
		*Negative	1150	97.62%
		100 (2+)	2	0.17%
		30 (1+)	1	0.08%
		300-500 (3+)	4	0.34%
		Trace	21	1.78%
UA-2		* Denotes Correct Response		
MG: Roche - Conventional	2			
		*Negative	2	100.00%
UA-2		* Denotes Correct Response		
MG: Siemens - Plus Sys	15			
		*Negative	15	100.00%
		Urine Blood/Hemoglobin		
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 10 / 100	12			
		*Negative	12	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 50	63			
		*Negative	62	98.41%
		Trace	1	1.59%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 500	5			
		*Negative	5	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Status	636			
		*Negative	613	96.38%
		Large (3+)	3	0.47%
		Moderate (2+)	3	0.47%
		Trace	17	2.67%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Status / Siemens - Plus Sys / Siemens Urine Dipstick	14			
		*Negative	14	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Multistix - Visual	203			
		*Negative	202	99.51%
		Trace	1	0.49%
UA-2		* Denotes Correct Response		
PG: Henry Schein Urispec	6			
		*Negative	6	100.00%
UA-2		* Denotes Correct Response		
PG: Consult Diagnostics Urine Reagent Strips	5			
		*Negative	5	100.00%
UA-2		* Denotes Correct Response		



PG: Diag Test Group Clarity Urochek	24			
		*Negative	23	95.83%
		Trace	1	4.17%
UA-2		* Denotes Correct Response		
PG: Roche Chemstrips - Visual	90			
		*Negative	86	95.56%
		Large (3+)	1	1.11%
		Trace	3	3.33%
UA-2		* Denotes Correct Response		
PG: UriScan Reagent Strips	6			
		*Negative	5	83.33%
		Trace	1	16.67%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Advantus	24			
		*Negative	24	100.00%
UA-2		* Denotes Correct Response		
PG: Consult Diag Urinalysis Test Strips	8			
		*Negative	8	100.00%
UA-2		* Denotes Correct Response		
PG: Consult Diag Urine Analyzer	7			
		*Negative	7	100.00%
UA-2		* Denotes Correct Response		
PG: 120 Urine Analyzer	44			
		*Negative	44	100.00%
UA-2		* Denotes Correct Response		
PG: McKesson 10 SG Urine Dipstick	7			
		*Negative	7	100.00%
UA-2		* Denotes Correct Response		
SG: Select Siemens Clinitek/Visual	945			
		*Negative	920	97.35%
		Large (3+)	3	0.32%
		Moderate (2+)	3	0.32%
		Trace	19	2.01%
UA-2		* Denotes Correct Response		
SG: Select Siemens Clinitek/Visual / Siemens - Plus Sys / Siemens Urine Dipstick	15			
		*Negative	15	100.00%
UA-2		* Denotes Correct Response		
SG: Select Roche Urinalysis Inst/Visual	97			
		*Negative	93	95.88%
		Large (3+)	1	1.03%
		Trace	3	3.09%
UA-2		* Denotes Correct Response		
SG: Select Consult Diag Urinalysis Inst/Visual	20			
		*Negative	20	100.00%
UA-2		* Denotes Correct Response		
SG: Select McKesson Urinalysis Inst/Visual	51			
		*Negative	51	100.00%
UA-2		* Denotes Correct Response		
SG: Select Henry Schein Urinalysis Inst/Visual	9			
		*Negative	9	100.00%
UA-2		* Denotes Correct Response		
MG: Method Group	1173			
		*Negative	1142	97.36%
		Large (3+)	4	0.34%
		Moderate (2+)	3	0.26%
		Trace	24	2.05%
UA-2		* Denotes Correct Response		
MG: Roche - Conventional	2			
		*Negative	2	100.00%
UA-2		* Denotes Correct Response		
MG: Siemens - Plus Sys	15			
		*Negative	15	100.00%
		Urine Glucose (Reducing Sub)		
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 10 / 100	12			
		*<100 mg/dL	3	25.00%
		*>= 1000 mg/dL	4	33.33%



		*100 mg/dL	1	8.33%
		*250 mg/dL	4	33.33%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 50	63			
		*100 mg/dL	17	26.98%
		*250 mg/dL	43	68.25%
		<100 mg/dL	2	3.17%
		Negative	1	1.59%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 500	5			
		*100 mg/dL	1	20.00%
		*250 mg/dL	3	60.00%
		<100 mg/dL	1	20.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Status	636			
		*<100 mg/dL	31	4.87%
		*100 mg/dL	194	30.50%
		*250 mg/dL	385	60.53%
		>/= 1000 mg/dL	11	1.73%
		Negative	15	2.36%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Status / Siemens - Plus Sys / Siemens Urine Dipstick	14			
		*1+	11	78.57%
		*Trace	3	21.43%
UA-2		* Denotes Correct Response		
PG: Siemens Multistix - Visual	205			
		*<100 mg/dL	19	9.27%
		*100 mg/dL	55	26.83%
		*250 mg/dL	103	50.24%
		>/= 1000 mg/dL	2	0.98%
		500 mg/dL	2	0.98%
		Negative	24	11.71%
UA-2		* Denotes Correct Response		
PG: Henry Schein Urispec	6			
		*500 mg/dL	6	100.00%
UA-2		* Denotes Correct Response		
PG: Consult Diagnostics Urine Reagent Strips	5			
		*100 mg/dL	1	20.00%
		*250 mg/dL	4	80.00%
UA-2		* Denotes Correct Response		
PG: Diag Test Group Clarity Urochek	24			
		*100 mg/dL	6	25.00%
		*250 mg/dL	15	62.50%
		<100 mg/dL	3	12.50%
UA-2		* Denotes Correct Response		
PG: Roche Chemstrips - Visual	90			
		*>/= 1000 mg/dL	80	88.89%
		*250 mg/dL	2	2.22%
		*500 mg/dL	1	1.11%
		100 mg/dL	4	4.44%
		Negative	3	3.33%
UA-2		* Denotes Correct Response		
PG: UriScan Reagent Strips	6			
		*100 mg/dL	1	16.67%
		*250 mg/dL	2	33.33%
		*500 mg/dL	2	33.33%
		>/= 1000 mg/dL	1	16.67%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Advantus	24			
		*100 mg/dL	16	66.67%
		*250 mg/dL	8	33.33%
UA-2		* Denotes Correct Response		
PG: Consult Diag Urinalysis Test Strips	8			
		*100 mg/dL	1	12.50%
		*250 mg/dL	6	75.00%
		<100 mg/dL	1	12.50%
UA-2		* Denotes Correct Response		



PG: Consult Diag Urine Analyzer	7			
		*100 mg/dL	2	28.57%
		*250 mg/dL	5	71.43%
UA-2		* Denotes Correct Response		
PG: 120 Urine Analyzer	44			
		*100 mg/dL	8	18.18%
		*250 mg/dL	30	68.18%
		<100 mg/dL	3	6.82%
		500 mg/dL	2	4.55%
		Negative	1	2.27%
UA-2		* Denotes Correct Response		
PG: McKesson 10 SG Urine Dipstick	7			
		*100 mg/dL	3	42.86%
		*250 mg/dL	3	42.86%
		>/= 1000 mg/dL	1	14.29%
UA-2		* Denotes Correct Response		
SG: Select Siemens Clinitek/Visual	947			
		*<100 mg/dL	56	5.91%
		*>/= 1000 mg/dL	17	1.80%
		*100 mg/dL	285	30.10%
		*250 mg/dL	547	57.76%
		<100 mg/dL	56	5.91%
		>/= 1000 mg/dL	17	1.80%
		500 mg/dL	2	0.21%
		Negative	40	4.22%
UA-2		* Denotes Correct Response		
SG: Select Siemens Clinitek/Visual / Siemens - Plus Sys / Siemens Urine Dipstick	15			
		*1+	11	73.33%
		*Trace	4	26.67%
UA-2		* Denotes Correct Response		
SG: Select Roche Urinalysis Inst/Visual	97			
		*>/= 1000 mg/dL	84	86.60%
		*250 mg/dL	5	5.15%
		*500 mg/dL	1	1.03%
		100 mg/dL	4	4.12%
		Negative	3	3.09%
UA-2		* Denotes Correct Response		
SG: Select Consult Diag Urinalysis Inst/Visual	20			
		*100 mg/dL	4	20.00%
		*250 mg/dL	15	75.00%
		<100 mg/dL	1	5.00%
UA-2		* Denotes Correct Response		
SG: Select McKesson Urinalysis Inst/Visual	51			
		*100 mg/dL	11	21.57%
		*250 mg/dL	33	64.71%
		<100 mg/dL	3	5.88%
		>/= 1000 mg/dL	1	1.96%
		500 mg/dL	2	3.92%
		Negative	1	1.96%
UA-2		* Denotes Correct Response		
SG: Select Henry Schein Urinalysis Inst/Visual	9			
		*>/= 1000 mg/dL	1	11.11%
		*250 mg/dL	1	11.11%
		*500 mg/dL	7	77.78%
		500 mg/dL	7	77.78%
UA-2		* Denotes Correct Response		
SG: Select Medi-Lab Performance UA Inst/Visual	3			
		*250 mg/dL	3	100.00%
UA-2		* Denotes Correct Response		
MG: Method Group	1176			
		*<100 mg/dL	68	5.78%
		*>/= 1000 mg/dL	104	8.84%
		*100 mg/dL	315	26.79%
		*250 mg/dL	628	53.40%
		*500 mg/dL	14	1.19%
		<100 mg/dL	68	5.78%
		>/= 1000 mg/dL	104	8.84%
		100 mg/dL	315	26.79%



		500 mg/dL	14	1.19%
		Negative	47	4.00%
UA-2		* Denotes Correct Response		
MG: Roche - Conventional	2			
		*>= 1000 mg/dL	2	100.00%
UA-2		* Denotes Correct Response		
MG: Siemens - Plus Sys	15			
		*1+	11	73.33%
		*Trace	4	26.67%
		Urine Leuk Esterase		
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 10 / 100	12			
		*Moderate (2+)	5	41.67%
		*Small (1+)	7	58.33%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 50	63			
		*Moderate (2+)	10	15.87%
		*Small (1+)	50	79.37%
		Large (3+)	1	1.59%
		Trace	2	3.17%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 500	5			
		*Moderate (2+)	2	40.00%
		*Small (1+)	3	60.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Status	634			
		*Moderate (2+)	164	25.87%
		*Small (1+)	433	68.30%
		*Trace	28	4.42%
		Large (3+)	4	0.63%
		Negative	5	0.79%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Status / Siemens - Plus Sys / Siemens Urine Dipstick	14			
		*1+	9	64.29%
		*2+	4	28.57%
		Trace	1	7.14%
UA-2		* Denotes Correct Response		
PG: Siemens Multistix - Visual	204			
		*Moderate (2+)	89	43.63%
		*Small (1+)	95	46.57%
		*Trace	15	7.35%
		Large (3+)	4	1.96%
		Negative	1	0.49%
UA-2		* Denotes Correct Response		
PG: Henry Schein Urispec	6			
		*Large (3+)	6	100.00%
UA-2		* Denotes Correct Response		
PG: Consult Diagnostics Urine Reagent Strips	5			
		*Large (3+)	1	20.00%
		*Small (1+)	3	60.00%
		*Trace	1	20.00%
UA-2		* Denotes Correct Response		
PG: Diag Test Group Clarity Urochek	24			
		*Large (3+)	1	4.17%
		*Moderate (2+)	10	41.67%
		*Small (1+)	12	50.00%
		*Trace	1	4.17%
UA-2		* Denotes Correct Response		
PG: Roche Chemstrips - Visual	90			
		*Large (3+)	17	18.89%
		*Moderate (2+)	65	72.22%
		*Small (1+)	4	4.44%
		Negative	2	2.22%
		Trace	2	2.22%
UA-2		* Denotes Correct Response		
PG: UriScan Reagent Strips	6			
		*Large (3+)	1	16.67%



		*Moderate (2+)	3	50.00%
		*Trace	2	33.33%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Advantus	24			
		*Moderate (2+)	4	16.67%
		*Small (1+)	20	83.33%
UA-2		* Denotes Correct Response		
PG: Consult Diag Urinalysis Test Strips	8			
		*Moderate (2+)	2	25.00%
		*Small (1+)	5	62.50%
		Large (3+)	1	12.50%
UA-2		* Denotes Correct Response		
PG: Consult Diag Urine Analyzer	7			
		*Moderate (2+)	3	42.86%
		*Small (1+)	3	42.86%
		Large (3+)	1	14.29%
UA-2		* Denotes Correct Response		
PG: 120 Urine Analyzer	44			
		*Moderate (2+)	20	45.45%
		*Small (1+)	17	38.64%
		*Trace	3	6.82%
		Large (3+)	1	2.27%
		Negative	3	6.82%
UA-2		* Denotes Correct Response		
PG: McKesson 10 SG Urine Dipstick	7			
		*Moderate (2+)	5	71.43%
		*Small (1+)	2	28.57%
UA-2		* Denotes Correct Response		
SG: Select Siemens Clinitek/Visual	943			
		*Moderate (2+)	274	29.06%
		*Small (1+)	609	64.58%
		*Trace	45	4.77%
		Large (3+)	9	0.95%
		Negative	6	0.64%
		Trace	45	4.77%
UA-2		* Denotes Correct Response		
SG: Select Siemens Clinitek/Visual / Siemens - Plus Sys / Siemens Urine Dipstick	15			
		*1+	10	66.67%
		*2+	4	26.67%
		Trace	1	6.67%
UA-2		* Denotes Correct Response		
SG: Select Roche Urinalysis Inst/Visual	97			
		*Large (3+)	19	19.59%
		*Moderate (2+)	68	70.10%
		*Small (1+)	6	6.19%
		Negative	2	2.06%
		Trace	2	2.06%
UA-2		* Denotes Correct Response		
SG: Select Consult Diag Urinalysis Inst/Visual	20			
		*Large (3+)	3	15.00%
		*Moderate (2+)	5	25.00%
		*Small (1+)	11	55.00%
		*Trace	1	5.00%
		Large (3+)	3	15.00%
UA-2		* Denotes Correct Response		
SG: Select McKesson Urinalysis Inst/Visual	51			
		*Moderate (2+)	25	49.02%
		*Small (1+)	19	37.25%
		*Trace	3	5.88%
		Large (3+)	1	1.96%
		Negative	3	5.88%
UA-2		* Denotes Correct Response		
SG: Select Henry Schein Urinalysis Inst/Visual	9			
		*Large (3+)	7	77.78%
		*Small (1+)	2	22.22%
UA-2		* Denotes Correct Response		
SG: Select Medi-Lab Performance UA Inst/Visual	3			



		*Moderate (2+)	1	33.33%
		*Small (1+)	2	66.67%
UA-2		* Denotes Correct Response		
PG: URS - Urine Reagent Strips	4			
		*Large (3+)	1	25.00%
		*Small (1+)	2	50.00%
		*Trace	1	25.00%
UA-2		* Denotes Correct Response		
MG: Method Group	1171			
		*Large (3+)	42	3.59%
		*Moderate (2+)	392	33.48%
		*Small (1+)	668	57.05%
		*Trace	57	4.87%
		Large (3+)	42	3.59%
		Negative	12	1.02%
		Trace	57	4.87%
UA-2		* Denotes Correct Response		
MG: Roche - Conventional	2			
		*500 Leu/uL	1	50.00%
		*75 Leu/uL	1	50.00%
UA-2		* Denotes Correct Response		
MG: Siemens - Plus Sys	15			
		*1+	10	66.67%
		*2+	4	26.67%
		Trace	1	6.67%
		Urine Nitrite		
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 10 / 100	12			
		*Positive	12	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 50	63			
		*Positive	63	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 500	5			
		*Positive	5	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Status	637			
		*Positive	633	99.37%
		Negative	4	0.63%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Status / Siemens - Plus Sys / Siemens Urine Dipstick	14			
		*Positive	14	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Multistix - Visual	203			
		*Positive	202	99.51%
		Negative	1	0.49%
UA-2		* Denotes Correct Response		
PG: Henry Schein Urispec	6			
		*Positive	6	100.00%
UA-2		* Denotes Correct Response		
PG: Consult Diagnostics Urine Reagent Strips	5			
		*Positive	5	100.00%
UA-2		* Denotes Correct Response		
PG: Diag Test Group Clarity Urochek	23			
		*Positive	23	100.00%
UA-2		* Denotes Correct Response		
PG: Roche Chemstrips - Visual	90			
		*Positive	89	98.89%
		Negative	1	1.11%
UA-2		* Denotes Correct Response		
PG: UriScan Reagent Strips	6			
		*Positive	6	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Advantus	24			
		*Positive	24	100.00%
UA-2		* Denotes Correct Response		
PG: Consult Diag Urinalysis Test Strips	8			



		*Positive	8	100.00%
UA-2		* Denotes Correct Response		
PG: Consult Diag Urine Analyzer	7			
		*Positive	7	100.00%
UA-2		* Denotes Correct Response		
PG: 120 Urine Analyzer	44			
		*Positive	44	100.00%
UA-2		* Denotes Correct Response		
PG: McKesson 10 SG Urine Dipstick	7			
		*Positive	7	100.00%
UA-2		* Denotes Correct Response		
SG: Select Siemens Clinitek/Visual	945			
		*Positive	940	99.47%
		Negative	5	0.53%
UA-2		* Denotes Correct Response		
SG: Select Siemens Clinitek/Visual / Siemens - Plus Sys / Siemens Urine Dipstick	15			
		*Positive	15	100.00%
UA-2		* Denotes Correct Response		
SG: Select Roche Urinalysis Inst/Visual	97			
		*Positive	96	98.97%
		Negative	1	1.03%
UA-2		* Denotes Correct Response		
SG: Select Consult Diag Urinalysis Inst/Visual	20			
		*Positive	20	100.00%
UA-2		* Denotes Correct Response		
SG: Select McKesson Urinalysis Inst/Visual	51			
		*Positive	51	100.00%
UA-2		* Denotes Correct Response		
SG: Select Henry Schein Urinalysis Inst/Visual	9			
		*Positive	9	100.00%
UA-2		* Denotes Correct Response		
MG: Method Group	1172			
		*Positive	1165	99.40%
		Negative	7	0.60%
UA-2		* Denotes Correct Response		
MG: Roche - Conventional	2			
		*Positive	2	100.00%
UA-2		* Denotes Correct Response		
MG: Siemens - Plus Sys	15			
		*Positive	15	100.00%
		Urine Urobilinogen		
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 10 / 100	11			
		*0.2/Normal	11	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 50	63			
		*0.2/Normal	63	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 500	5			
		*0.2/Normal	5	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Status	629			
		*0.2/Normal	626	99.52%
		2.0 mg/dL	3	0.48%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Status / Siemens - Plus Sys / Siemens Urine Dipstick	6			
		*3.2 umol/L	6	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Multistix - Visual	192			
		*0.2/Normal	187	97.40%
		1 mg/dL	1	0.52%
		2.0 mg/dL	3	1.56%
		4 mg/dL	1	0.52%
UA-2		* Denotes Correct Response		
PG: Henry Schein Urispec	5			
		*0.2/Normal	5	100.00%



UA-2		* Denotes Correct Response		
PG: Consult Diagnostics Urine Reagent Strips	5			
		*0.2/Normal	5	100.00%
UA-2		* Denotes Correct Response		
PG: Diag Test Group Clarity Urochek	24			
		*0.2/Normal	23	95.83%
		1 mg/dL	1	4.17%
UA-2		* Denotes Correct Response		
PG: Roche Chemstrips - Visual	82			
		*0.2/Normal	82	100.00%
UA-2		* Denotes Correct Response		
PG: UriScan Reagent Strips	5			
		*0.2/Normal	5	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Advantus	24			
		*0.2/Normal	24	100.00%
UA-2		* Denotes Correct Response		
PG: Consult Diag Urinalysis Test Strips	8			
		*0.2/Normal	8	100.00%
UA-2		* Denotes Correct Response		
PG: Consult Diag Urine Analyzer	7			
		*0.2/Normal	7	100.00%
UA-2		* Denotes Correct Response		
PG: 120 Urine Analyzer	43			
		*0.2/Normal	43	100.00%
UA-2		* Denotes Correct Response		
PG: McKesson 10 SG Urine Dipstick	7			
		*0.2/Normal	6	85.71%
		1 mg/dL	1	14.29%
UA-2		* Denotes Correct Response		
SG: Select Siemens Clinitek/Visual	925			
		*0.2/Normal	917	99.14%
		1 mg/dL	1	0.11%
		2.0 mg/dL	6	0.65%
		4 mg/dL	1	0.11%
UA-2		* Denotes Correct Response		
SG: Select Siemens Clinitek/Visual / Siemens - Plus Sys / Siemens Urine Dipstick	7			
		*3.2 umol/L	7	100.00%
UA-2		* Denotes Correct Response		
SG: Select Roche Urinalysis Inst/Visual	89			
		*0.2/Normal	89	100.00%
UA-2		* Denotes Correct Response		
SG: Select Consult Diag Urinalysis Inst/Visual	20			
		*0.2/Normal	20	100.00%
UA-2		* Denotes Correct Response		
SG: Select McKesson Urinalysis Inst/Visual	50			
		*0.2/Normal	49	98.00%
		1 mg/dL	1	2.00%
UA-2		* Denotes Correct Response		
SG: Select Henry Schein Urinalysis Inst/Visual	8			
		*0.2/Normal	8	100.00%
UA-2		* Denotes Correct Response		
MG: Method Group	1142			
		*0.2/Normal	1131	99.04%
		1 mg/dL	3	0.26%
		2.0 mg/dL	6	0.53%
		4 mg/dL	2	0.18%
UA-2		* Denotes Correct Response		
MG: Roche - Conventional	2			
		*Normal	2	100.00%
UA-2		* Denotes Correct Response		
MG: Siemens - Plus Sys	14			
		*0.2 EU/dL	7	50.00%
		*3.2 umol/L	7	50.00%
		Urine Microalbumin		
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 10 / 100	2			



		*10 mg/L (Trace)	2	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek 50	8			
		*10 mg/L (Trace)	8	100.00%
UA-2		* Denotes Correct Response		
PG: Siemens Clinitek Status	40			
		*10 mg/L (Trace)	36	90.00%
		*Negative	1	2.50%
		150 mg/L	1	2.50%
		20/30 mg/L (Small	2	5.00%
UA-2		* Denotes Correct Response		
PG: Siemens Multistix - Visual	3			
		*Negative	3	100.00%
UA-2		* Denotes Correct Response		
PG: Roche Chemstrip Micral	26			
		10 mg/L (Trace)	3	11.54%
		100 mg/L (3+)	1	3.85%
		20/30 mg/L (Small	10	38.46%
		50 mg/L (2+)	2	7.69%
		Negative	10	38.46%
UA-2		* Denotes Correct Response		
PG: Roche Chemstrips - Visual	1			
		20/30 mg/L (Small	1	100.00%
UA-2		* Denotes Correct Response		
PG: Hemocue Albumin 201	1			
		*Negative	1	100.00%
UA-2		* Denotes Correct Response		
SG: Select Siemens Clinitek/Visual	53			
		*10 mg/L (Trace)	46	86.79%
		*Negative	4	7.55%
		150 mg/L	1	1.89%
		20/30 mg/L (Small	2	3.77%
UA-2		* Denotes Correct Response		
SG: Select Roche Urinalysis Inst/Visual	1			
		20/30 mg/L (Small	1	100.00%
UA-2		Fluid pH		
UA-2		* Denotes Correct Response		
PG: Nitrazine Paper	9			
		*5 or less	2	22.22%
		*5.5	7	77.78%
UA-2		Urine hCG		
UA-2		* Denotes Correct Response		
AG: All Methods	1034			
		*Negative	1026	99.23%
		Positive	8	0.77%
ASI ProPhase Plus	1	*Negative	1	100.00%
Siemens Clinitek Status	12	*Negative	12	100.00%
Beck Coulter ICON 20 HCG	7	*Negative	7	100.00%
Beck Coulter ICON 25 hCG	75	*Negative	73	97.33%
Alere Acceava HCG Basic II	24	*Negative	24	100.00%
Alere Acceava hCG Combo	5	*Negative	5	100.00%
Alere Acceava hCG-Urine	20	*Negative	20	100.00%
Conception Tech One Step hCG	1	*Negative	1	100.00%
Diag Test Group Clarity hCG Plus Urine Cassettes	6	*Negative	6	100.00%
Fisher HC Sure-Vue hCG STAT	4	*Negative	4	100.00%
Fisher HC Sure-Vue Urine hCG	19	*Negative	19	100.00%
Fisher HC Sure-Vure U/S hCG	5	*Negative	5	100.00%
Sekisui Diagnostics OSOM Urine	5	*Negative	5	100.00%
Sekisui Diagnostics OSOM Card Pregnancy	30	*Negative	29	96.67%
Sekisui Diagnostics OSOM hCG Combo	16	*Negative	16	100.00%
Henry Schein One Step + (hCG)	55	*Negative	55	100.00%
Immunostics Detector Combo	3	*Negative	3	100.00%
Immunostics hCG Detector-urine	3	*Negative	3	100.00%
Instant Tech. i Pregnancy	3	*Negative	3	100.00%
Inverness Clearview Combo II	1	*Negative	1	100.00%
Inv Clearview Urine - Cassette	3	*Negative	3	100.00%
Inverness Signify hCG combo	2	*Negative	2	100.00%
Inverness Signify hCG urine	4	*Negative	4	100.00%



LifeSign Status hCG	1	*Negative	1	100.00%
McKesson Medi-Lab hCG - Urine	32	*Negative	32	100.00%
Meridian Immunocard Stat!	1	*Negative	1	100.00%
Permaxim RediScreen	3	*Negative	3	100.00%
Quidel QuickVue + hCG	22	*Negative	22	100.00%
Quidel QuickVue One-Step hCG Combo U/S	58	*Negative	58	100.00%
Quidel QuickVue One-Step Urine	138	*Negative	137	99.28%
Quidel RapidVue hCG - Urine	7	*Negative	7	100.00%
SAS Ultra HCG	2	*Negative	2	100.00%
Select Diag hCG Cassette (urine)	19	*Negative	19	100.00%
Select Diag hCG Combo	10	*Negative	9	90.00%
Select Diag hCG Dipstick	7	*Negative	7	100.00%
SP (Cardinal) Brand combo	25	*Negative	23	92.00%
SP (Cardinal) Brand-cassette	29	*Negative	28	96.55%
SP Brand (Cardinal) HCG Combo	59	*Negative	59	100.00%
Stanbio Qupid	5	*Negative	5	100.00%
Stanbio QuStick	1	*Negative	1	100.00%
Stanbio TRUE hCG	5	*Negative	5	100.00%
MooreBrand hCG Cassette	1	*Negative	1	100.00%
Nova Scientific Century S/U hCG	2	*Negative	2	100.00%
Diag Test Group Clarity hCG Urine Strips	1	*Negative	1	100.00%
Jant AccuTest	4	*Negative	4	100.00%
Instant-View Urine Pregnancy	1	*Negative	1	100.00%
Becton Dickinson Precise One-Step	1	*Negative	1	100.00%
Consult Diag hCG Urine Cassette	138	*Negative	138	100.00%
CLIAwaived Inc Pregnancy Urine Dip-Strip	4	*Negative	4	100.00%
Germaine AIM Step Combo Pregnancy	1	*Negative	1	100.00%
NDC ProAdvantage Urine hCG Cassette	4	*Negative	4	100.00%
Consult Diag hCG Urine Dipstick	43	*Negative	43	100.00%
SP (Cardinal) Brand Dipstick	12	*Negative	12	100.00%
Consult Diag hCG Combo Cassette (Urine)	12	*Negative	12	100.00%
McKesson Medi-Lab Performance hCG Urine Cassette	6	*Negative	6	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	2	*Negative	2	100.00%
Alere hCG Cassette (25 mIU/mL)	1	*Negative	1	100.00%
Alere hCG Dipstick	7	*Negative	7	100.00%
Medline Pregnancy Urine Dip-Strip	2	*Negative	2	100.00%
Medline Pregnancy Urine Cassette Test	5	*Negative	5	100.00%
McKesson hCG Urine Test Dipstick	8	*Negative	8	100.00%
Medline Pregnancy Combo Cassette (urine)	1	*Negative	1	100.00%
Quidel QuickVue hCG Urine Test	20	*Negative	20	100.00%
MooreBrand hCG Dipstick	1	*Negative	1	100.00%
Siemens Clinitest hCG	1	*Negative	1	100.00%
Henry Schein One Step + HCG Combo	3	*Negative	3	100.00%
Henry Schein One Step hCG Urine Cassette	22	*Negative	22	100.00%
Medline hCG Test Strip 25 mIU/mL	1	*Negative	1	100.00%
NCS hCG One Step Pregnancy Test	1	*Negative	1	100.00%
MediChoice Urine hCG Cassette	1	*Negative	1	100.00%
Beck Coulter ICON 25 hCG	75	Positive	2	2.67%
Sekisui Diagnostics OSOM Card Pregnancy	30	Positive	1	3.33%
Quidel QuickVue One-Step Urine	138	Positive	1	0.72%
Select Diag hCG Combo	10	Positive	1	10.00%
SP (Cardinal) Brand combo	25	Positive	2	8.00%
SP (Cardinal) Brand-cassette	29	Positive	1	3.45%
		Urine hCG (non-waived)		
UA-2		* Denotes Correct Response		
PG: Quidel Sofia/Sofia 2 hCG (urine)	1			
		*Negative	1	100.00%
661 - Urine Colony Count;784 - Presump ID/Colony Count - Urine;785 - Presump ID/Colony Count-Urine & Strep Combo;775 - Urine Culture;778 - Urine & Throat Cultures;779 - Urine & Genital Cultures;781 - Urine, Throat & Genital Cultures;782 - Complete Bacteri	Labs	Response	Number	Percent
		Susceptibility Cephalothin		
UC-6		* Denotes Correct Response		
AG: All Methods	2			
		*Resistant	2	100.00%
Susceptibility	2	*Resistant	2	100.00%
		Susceptibility Ciprofloxacin		
UC-6		* Denotes Correct Response		



AG: All Methods	6			
		*Intermediate	2	33.33%
		*Sensitive	4	66.67%
Susceptibility	6	*Intermediate	2	33.33%
Susceptibility		*Sensitive	4	66.67%
		Susceptibility Gentamicin		
UC-6		* Denotes Correct Response		
AG: All Methods	1			
		*Resistant	1	100.00%
Susceptibility	1	*Resistant	1	100.00%
		Susceptibility Levofloxacin		
UC-6		* Denotes Correct Response		
AG: All Methods	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
		Susceptibility Linezolid		
UC-6		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
		Susceptibility Minocycline		
UC-6		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
		Susceptibility Penicillin		
UC-6		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
		Susceptibility Tetracycline		
UC-6		* Denotes Correct Response		
AG: All Methods	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
		Susceptibility Trimethoprim		
UC-6		* Denotes Correct Response		
AG: All Methods	1			
		*Resistant	1	100.00%
Susceptibility	1	*Resistant	1	100.00%
		Susceptibility Trimeth./Sulfameth.		
UC-6		* Denotes Correct Response		
AG: All Methods	1			
		*Resistant	1	100.00%
Susceptibility	1	*Resistant	1	100.00%
		Susceptibility Vancomycin		
UC-6		* Denotes Correct Response		
AG: All Methods	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
		Urine Culture		
UC-6		* Denotes Correct Response		
AG: All Methods	29			
		*Enterococcus faecium	2	6.90%
		*Enterococcus species, NOS	1	3.45%
		*Gram Positive organism	3	10.34%
		*Growth, referred for identification	13	44.83%
		*Presumptive Enterococcus species	2	6.90%
		*Uricult - Growth (Positive)	5	17.24%
		No Growth	1	3.45%
		Uricult - No Growth (Negative)	2	6.90%
Cultures	29	*Enterococcus faecium	2	6.90%
Cultures		*Enterococcus species, NOS	1	3.45%
Cultures		*Gram Positive organism	3	10.34%
Cultures		*Growth, referred for identification	13	44.83%
Cultures		*Presumptive Enterococcus species	2	6.90%
Cultures		*Uricult - Growth (Positive)	5	17.24%



Cultures		No Growth	1	3.45%
Cultures		Uricult - No Growth (Negative)	2	6.90%
UC-7		* Denotes Correct Response		
AG: All Methods	29			
		*Gram Negative organism	3	10.34%
		*Growth, referred for identification	13	44.83%
		*Presumptive Proteus species	2	6.90%
		*Proteus species NOS	2	6.90%
		*Proteus vulgaris	1	3.45%
		*Uricult - Growth (Positive)	7	24.14%
		No Growth	1	3.45%
Cultures	29	*Gram Negative organism	3	10.34%
Cultures		*Growth, referred for identification	13	44.83%
Cultures		*Presumptive Proteus species	2	6.90%
Cultures		*Proteus species NOS	2	6.90%
Cultures		*Proteus vulgaris	1	3.45%
Cultures		*Uricult - Growth (Positive)	7	24.14%
Cultures		No Growth	1	3.45%
UC-8		* Denotes Correct Response		
AG: All Methods	28			
		*No Growth	16	57.14%
		*Uricult - No Growth (Negative)	5	17.86%
		Contaminated specimen; >2 organisms present	1	3.57%
		Growth, referred for identification	4	14.29%
		Uricult - Growth (Positive)	2	7.14%
Cultures	28	*No Growth	16	57.14%
Cultures		*Uricult - No Growth (Negative)	5	17.86%
Cultures		Contaminated specimen; >2 organisms present	1	3.57%
Cultures		Growth, referred for identification	4	14.29%
Cultures		Uricult - Growth (Positive)	2	7.14%
UC-9		* Denotes Correct Response		
AG: All Methods	13			
		*Enterobacter aerogenes	1	7.69%
		*Enterobacter species, NOS	1	7.69%
		*Gram Negative organism	3	23.08%
		*Growth, referred for identification	2	15.38%
		*Presumptive Enterobacter species	1	7.69%
		*Uricult - Growth (Positive)	5	38.46%
Cultures	13	*Enterobacter aerogenes	1	7.69%
Cultures		*Enterobacter species, NOS	1	7.69%
Cultures		*Gram Negative organism	3	23.08%
Cultures		*Growth, referred for identification	2	15.38%
Cultures		*Presumptive Enterobacter species	1	7.69%
Cultures		*Uricult - Growth (Positive)	5	38.46%
UC-10		* Denotes Correct Response		
AG: All Methods	13			
		*Gram Negative organism	2	15.38%
		*Growth, referred for identification	3	23.08%
		*Pseudomonas aeruginosa	1	7.69%
		*Pseudomonas species, NOS	1	7.69%
		*Uricult - Growth (Positive)	5	38.46%
		Presumptive Staphylococcus species	1	7.69%
Cultures	13	*Gram Negative organism	2	15.38%
Cultures		*Growth, referred for identification	3	23.08%
Cultures		*Pseudomonas aeruginosa	1	7.69%
Cultures		*Pseudomonas species, NOS	1	7.69%
Cultures		*Uricult - Growth (Positive)	5	38.46%
Cultures		Presumptive Staphylococcus species	1	7.69%
		Susceptibility Amox/Clavulanate		
UC-6		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
		Susceptibility Ampicillin		
UC-6		* Denotes Correct Response		
AG: All Methods	5			
		*Sensitive	5	100.00%
Susceptibility	5	*Sensitive	5	100.00%



703 - Urine Drug Screen:713 - Urine Drug Screen (50 ml sample volume)	Labs	Response	Number	Percent
		Barbiturates		
UDS-3		* Denotes Correct Response		
AG: All Methods	138			
		*Negative	138	100.00%
AlereTriage	4	*Negative	4	100.00%
E-Screen/Alere	26	*Negative	26	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	12	*Negative	12	100.00%
Instant Technologies i Screen	1	*Negative	1	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Beckman AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	5	*Negative	5	100.00%
CLIAwaived Inc Rapid Drug Cup	4	*Negative	4	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Branan Med Corp QuickTox	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Medica Easy RA	2	*Negative	2	100.00%
Alere iCassette	2	*Negative	2	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
Alere USD UScreen Cup	9	*Negative	9	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Hemosure First Sign UDS Dipstick	15	*Negative	15	100.00%
Siemens Viva/ProE System	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	12	*Negative	12	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Abbott Architect c and ci Series	1	*Negative	1	100.00%
Consult Diag Drugs of Abuse Cups	2	*Negative	2	100.00%
Branan Med Corp ToxCup	6	*Negative	6	100.00%
American Screening Discover Plus Cup	1	*Negative	1	100.00%
McKesson Drugs of Abuse Test Cup	1	*Negative	1	100.00%
E-Screen mCup/Alere	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	4	*Negative	4	100.00%
Siemens Atellica	1	*Negative	1	100.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	138			
		*Negative	137	99.28%
		Positive	1	0.72%
AlereTriage	4	*Negative	4	100.00%
E-Screen/Alere	26	*Negative	26	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	12	*Negative	12	100.00%
Instant Technologies i Screen	1	*Negative	1	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Beckman AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	5	*Negative	5	100.00%
CLIAwaived Inc Rapid Drug Cup	4	*Negative	4	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Branan Med Corp QuickTox	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Medica Easy RA	2	*Negative	2	100.00%
Alere iCassette	2	*Negative	2	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
Alere USD UScreen Cup	9	*Negative	9	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%



Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Hemosure First Sign UDS Dipstick	15	*Negative	15	100.00%
Siemens Viva/ProE System	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	12	*Negative	12	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Abbott Architect c and ci Series	1	*Negative	1	100.00%
Consult Diag Drugs of Abuse Cups	2	*Negative	2	100.00%
Branan Med Corp ToxCup	6	*Negative	6	100.00%
American Screening Discover Plus Cup	1	*Negative	1	100.00%
McKesson Drugs of Abuse Test Cup	1	*Negative	1	100.00%
E-Screen mCup/Alere	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	4	*Negative	3	75.00%
Siemens Atellica	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	4	Positive	1	25.00%
		Benzodiazepines		
UDS-3		* Denotes Correct Response		
AG: All Methods	149			
		*Negative	147	98.66%
		Positive	2	1.34%
AlereTriage	4	*Negative	4	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	1	*Negative	1	100.00%
E-Screen/Alere	26	*Negative	26	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	15	*Negative	14	93.33%
Instant Technologies i Screen	1	*Negative	1	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	5	*Negative	5	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Branan Med Corp QuickTox	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Medica Easy RA	2	*Negative	2	100.00%
Alere iCassette	2	*Negative	2	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
Alere USD UScreen Cup	9	*Negative	9	100.00%
ThermoScientific Indiko Plus	1	*Negative	1	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Hemosure First Sign UDS Dipstick	15	*Negative	15	100.00%
Siemens Viva/ProE System	3	*Negative	3	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	12	*Negative	12	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Abbott Architect c and ci Series	1	*Negative	1	100.00%
Consult Diag Drugs of Abuse Cups	2	*Negative	2	100.00%
Branan Med Corp ToxCup	6	*Negative	6	100.00%
McKesson Drugs of Abuse Test Cup	1	*Negative	1	100.00%
E-Screen mCup/Alere	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	5	*Negative	5	100.00%
Siemens Atellica	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	15	Positive	1	6.67%
American Screening Discover Plus Cup	1	Positive	1	100.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	149			
		*Negative	148	99.33%
		Positive	1	0.67%
AlereTriage	4	*Negative	4	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	1	*Negative	1	100.00%
E-Screen/Alere	26	*Negative	26	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%



Alere iCup Drugs of Abuse	15	*Negative	15	100.00%
Instant Technologies i Screen	1	*Negative	1	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	5	*Negative	5	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Branan Med Corp QuickTox	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Medica Easy RA	2	*Negative	2	100.00%
Alere iCassette	2	*Negative	2	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
Alere USD UScreen Cup	9	*Negative	9	100.00%
ThermoScientific Indiko Plus	1	*Negative	1	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Hemosure First Sign UDS Dipstick	15	*Negative	15	100.00%
Siemens Viva/ProE System	3	*Negative	3	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	12	*Negative	12	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Abbott Architect c and ci Series	1	*Negative	1	100.00%
Consult Diag Drugs of Abuse Cups	2	*Negative	2	100.00%
Branan Med Corp ToxCup	6	*Negative	6	100.00%
American Screening Discover Plus Cup	1	*Negative	1	100.00%
McKesson Drugs of Abuse Test Cup	1	*Negative	1	100.00%
E-Screen mCup/Alere	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	5	*Negative	4	80.00%
Siemens Atellica	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	5	Positive	1	20.00%
		Amphetamines		
UDS-3		* Denotes Correct Response		
AG: All Methods	153			
		*Negative	153	100.00%
AlereTriage	4	*Negative	4	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	1	*Negative	1	100.00%
E-Screen/Alere	23	*Negative	23	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	17	*Negative	17	100.00%
Instant Technologies i Screen	1	*Negative	1	100.00%
MedTox Profile II	3	*Negative	3	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	5	*Negative	5	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Branan Med Corp QuickTox	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Medica Easy RA	2	*Negative	2	100.00%
Alere iCassette	5	*Negative	5	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
Alere USD UScreen Cup	9	*Negative	9	100.00%
ThermoScientific Indiko Plus	1	*Negative	1	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Hemosure First Sign UDS Dipstick	15	*Negative	15	100.00%
Siemens Viva/ProE System	4	*Negative	4	100.00%



TFS DRI Assay	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	12	*Negative	12	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Abbott Architect c and ci Series	1	*Negative	1	100.00%
Consult Diag Drugs of Abuse Cups	2	*Negative	2	100.00%
Branan Med Corp ToxCup	6	*Negative	6	100.00%
American Screening Discover Plus Cup	1	*Negative	1	100.00%
McKesson Drugs of Abuse Test Cup	1	*Negative	1	100.00%
E-Screen mCup/Alere	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	5	*Negative	5	100.00%
Siemens Atellica	1	*Negative	1	100.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	153			
		*Positive	149	97.39%
		Negative	4	2.61%
AlereTriage	4	*Positive	4	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	1	*Positive	1	100.00%
E-Screen/Alere	23	*Positive	21	91.30%
Express Diag. Express Check	1	*Positive	1	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Positive	1	100.00%
Alere iCup Drugs of Abuse	17	*Positive	16	94.12%
Instant Technologies i Screen	1	*Positive	1	100.00%
MedTox Profile II	3	*Positive	3	100.00%
Noble Med Split Specimen Cup	1	*Positive	1	100.00%
Beckman AU	1	*Positive	1	100.00%
Innovacon Multi-Clin	5	*Positive	5	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Positive	5	100.00%
ASC One Screen Drug Test Card	1	*Positive	1	100.00%
MedTox Profile V	2	*Positive	2	100.00%
Branan Med Corp QuickTox	1	*Positive	1	100.00%
Jant Accutest Multi Drug Screen	2	*Positive	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Positive	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Positive	4	100.00%
Mindray BS-200	2	*Positive	2	100.00%
Medica Easy RA	2	*Positive	2	100.00%
Alere iCassette	5	*Positive	5	100.00%
Roche Hitachi 717	1	*Positive	1	100.00%
Alere USD UScreen Cup	9	*Positive	9	100.00%
ThermoScientific Indiko Plus	1	*Positive	1	100.00%
Alere iScreen Drugs of Abuse Card	1	*Positive	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Positive	3	100.00%
Hemosure First Sign UDS Dipstick	15	*Positive	15	100.00%
Siemens Viva/ProE System	4	*Positive	4	100.00%
TFS DRI Assay	1	*Positive	1	100.00%
Premier Biotech Bio-Cup	12	*Positive	12	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Positive	1	100.00%
Abbott Architect c and ci Series	1	*Positive	1	100.00%
Consult Diag Drugs of Abuse Cups	2	*Positive	2	100.00%
Branan Med Corp ToxCup	6	*Positive	6	100.00%
McKesson Drugs of Abuse Test Cup	1	*Positive	1	100.00%
E-Screen mCup/Alere	1	*Positive	1	100.00%
Henry Schein OneStep+ DOA	5	*Positive	5	100.00%
Siemens Atellica	1	*Positive	1	100.00%
E-Screen/Alere	23	Negative	2	8.70%
Alere iCup Drugs of Abuse	17	Negative	1	5.88%
American Screening Discover Plus Cup	1	Negative	1	100.00%
		Cannabinoids (THC)		
UDS-3		* Denotes Correct Response		
AG: All Methods	183			
		*Negative	182	99.45%
		Positive	1	0.55%
AlereTriage	4	*Negative	4	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	1	*Negative	1	100.00%
E-Screen/Alere	51	*Negative	51	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	17	*Negative	17	100.00%
Instant Technologies i Screen	2	*Negative	2	100.00%



Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	3	*Negative	3	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	5	*Negative	4	80.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Branan Med Corp QuickTox	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Medica Easy RA	2	*Negative	2	100.00%
Alere iCassette	5	*Negative	5	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
Alere USD UScreen Cup	9	*Negative	9	100.00%
ThermoScientific Indiko Plus	1	*Negative	1	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
CRL Stat Cup	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	15	*Negative	15	100.00%
Siemens Viva/ProE System	3	*Negative	3	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	12	*Negative	12	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Abbott Architect c and ci Series	1	*Negative	1	100.00%
Consult Diag Drugs of Abuse Cups	2	*Negative	2	100.00%
Branan Med Corp ToxCup	6	*Negative	6	100.00%
American Screening Discover Plus Cup	1	*Negative	1	100.00%
McKesson Drugs of Abuse Test Cup	1	*Negative	1	100.00%
E-Screen mCup/Alere	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	5	*Negative	5	100.00%
Siemens Atellica	1	*Negative	1	100.00%
Innovacon Multi-Clin	5	Positive	1	20.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	182			
		*Negative	179	98.35%
		Positive	3	1.65%
AlereTriage	4	*Negative	4	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	1	*Negative	1	100.00%
E-Screen/Alere	50	*Negative	49	98.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	17	*Negative	16	94.12%
Instant Technologies i Screen	2	*Negative	2	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	3	*Negative	3	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	5	*Negative	5	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Branan Med Corp QuickTox	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Medica Easy RA	2	*Negative	2	100.00%
Alere iCassette	5	*Negative	5	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
Alere USD UScreen Cup	9	*Negative	9	100.00%
ThermoScientific Indiko Plus	1	*Negative	1	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
CRL Stat Cup	1	*Negative	1	100.00%



Hemosure First Sign UDS Dipstick	15	*Negative	15	100.00%
Siemens Viva/ProE System	3	*Negative	3	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	12	*Negative	12	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Abbott Architect c and ci Series	1	*Negative	1	100.00%
Consult Diag Drugs of Abuse Cups	2	*Negative	2	100.00%
Branan Med Corp ToxCup	6	*Negative	6	100.00%
American Screening Discover Plus Cup	1	*Negative	1	100.00%
McKesson Drugs of Abuse Test Cup	1	*Negative	1	100.00%
E-Screen mCup/Alere	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	5	*Negative	4	80.00%
Siemens Atellica	1	*Negative	1	100.00%
E-Screen/Alere	50	Positive	1	2.00%
Alere iCup Drugs of Abuse	17	Positive	1	5.88%
Henry Schein OneStep+ DOA	5	Positive	1	20.00%
		Cocaine Metabolites		
UDS-3		* Denotes Correct Response		
AG: All Methods	183			
		*Negative	183	100.00%
AlereTriage	4	*Negative	4	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	1	*Negative	1	100.00%
E-Screen/Alere	51	*Negative	51	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	17	*Negative	17	100.00%
Instant Technologies i Screen	2	*Negative	2	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	3	*Negative	3	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	5	*Negative	5	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Branan Med Corp QuickTox	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Medica Easy RA	2	*Negative	2	100.00%
Alere iCassette	5	*Negative	5	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
Alere USD UScreen Cup	9	*Negative	9	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
CRL Stat Cup	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	15	*Negative	15	100.00%
Siemens Viva/ProE System	4	*Negative	4	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	12	*Negative	12	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Abbott Architect c and ci Series	1	*Negative	1	100.00%
Consult Diag Drugs of Abuse Cups	2	*Negative	2	100.00%
Branan Med Corp ToxCup	6	*Negative	6	100.00%
American Screening Discover Plus Cup	1	*Negative	1	100.00%
McKesson Drugs of Abuse Test Cup	1	*Negative	1	100.00%
E-Screen mCup/Alere	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	5	*Negative	5	100.00%
Siemens Atellica	1	*Negative	1	100.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	181			
		*Negative	178	98.34%
		Positive	3	1.66%
AlereTriage	4	*Negative	4	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	1	*Negative	1	100.00%
E-Screen/Alere	50	*Negative	49	98.00%
Express Diag. Express Check	1	*Negative	1	100.00%



Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	16	*Negative	15	93.75%
Instant Technologies i Screen	2	*Negative	2	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	3	*Negative	3	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	5	*Negative	5	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Branan Med Corp QuickTox	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Medica Easy RA	2	*Negative	2	100.00%
Alere iCassette	5	*Negative	5	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
Alere USD UScreen Cup	9	*Negative	9	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
CRL Stat Cup	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	15	*Negative	15	100.00%
Siemens Viva/ProE System	4	*Negative	4	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	12	*Negative	12	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Abbott Architect c and ci Series	1	*Negative	1	100.00%
Consult Diag Drugs of Abuse Cups	2	*Negative	2	100.00%
Branan Med Corp ToxCup	6	*Negative	6	100.00%
American Screening Discover Plus Cup	1	*Negative	1	100.00%
McKesson Drugs of Abuse Test Cup	1	*Negative	1	100.00%
E-Screen mCup/Alere	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	5	*Negative	4	80.00%
Siemens Atellica	1	*Negative	1	100.00%
E-Screen/Alere	50	Positive	1	2.00%
Alere iCup Drugs of Abuse	16	Positive	1	6.25%
Henry Schein OneStep+ DOA	5	Positive	1	20.00%
		Ethanol		
UDS-3		* Denotes Correct Response		
AG: All Methods	7			
		*Negative	7	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Medica Easy RA	1	*Negative	1	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
Siemens Viva/ProE System	1	*Negative	1	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
Siemens Atellica	1	*Negative	1	100.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	7			
		*Negative	7	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Medica Easy RA	1	*Negative	1	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
Siemens Viva/ProE System	1	*Negative	1	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
Siemens Atellica	1	*Negative	1	100.00%
		Acetaminophen		
UDS-3		* Denotes Correct Response		
AG: All Methods	7			
		*Negative	6	85.71%
		Positive	1	14.29%
AlereTriage	4	*Negative	4	100.00%
Jant Accutest Multi Drug Screen	1	*Negative	1	100.00%
Alere USD UScreen Cup	1	*Negative	1	100.00%
Noble Med Split Specimen Cup	1	Positive	1	100.00%
UDS-4		* Denotes Correct Response		



AG: All Methods	7			
		*Negative	7	100.00%
AlereTriage	4	*Negative	4	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	1	*Negative	1	100.00%
Alere USD UScreen Cup	1	*Negative	1	100.00%
		MDMA (Ecstasy)		
UDS-3		* Denotes Correct Response		
AG: All Methods	96			
		*Negative	96	100.00%
E-Screen/Alere	4	*Negative	4	100.00%
Alere iCup Drugs of Abuse	12	*Negative	12	100.00%
Innovacon Multi-Clin	5	*Negative	5	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
Branan Med Corp QuickTox	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	2	*Negative	2	100.00%
Medica Easy RA	1	*Negative	1	100.00%
Alere iCassette	2	*Negative	2	100.00%
Alere USD UScreen Cup	9	*Negative	9	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Hemosure First Sign UDS Dipstick	15	*Negative	15	100.00%
Premier Biotech Bio-Cup	12	*Negative	12	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Consult Diag Drugs of Abuse Cups	2	*Negative	2	100.00%
Branan Med Corp ToxCup	6	*Negative	6	100.00%
American Screening Discover Plus Cup	1	*Negative	1	100.00%
McKesson Drugs of Abuse Test Cup	1	*Negative	1	100.00%
E-Screen mCup/Alere	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	5	*Negative	5	100.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	96			
		*Positive	93	96.88%
		Negative	3	3.13%
E-Screen/Alere	4	*Positive	3	75.00%
Alere iCup Drugs of Abuse	12	*Positive	12	100.00%
Innovacon Multi-Clin	5	*Positive	5	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Positive	5	100.00%
Branan Med Corp QuickTox	1	*Positive	1	100.00%
Jant Accutest Multi Drug Screen	2	*Positive	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Positive	3	75.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	2	*Positive	2	100.00%
Medica Easy RA	1	*Positive	1	100.00%
Alere iCassette	2	*Positive	2	100.00%
Alere USD UScreen Cup	9	*Positive	9	100.00%
Alere iScreen Drugs of Abuse Card	1	*Positive	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Positive	3	100.00%
Hemosure First Sign UDS Dipstick	15	*Positive	15	100.00%
Premier Biotech Bio-Cup	12	*Positive	12	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Positive	1	100.00%
Consult Diag Drugs of Abuse Cups	2	*Positive	2	100.00%
Branan Med Corp ToxCup	6	*Positive	6	100.00%
McKesson Drugs of Abuse Test Cup	1	*Positive	1	100.00%
E-Screen mCup/Alere	1	*Positive	1	100.00%
Henry Schein OneStep+ DOA	5	*Positive	5	100.00%
E-Screen/Alere	4	Negative	1	25.00%
Amedica Biotech AmediCheck Instant Test Cup	4	Negative	1	25.00%
American Screening Discover Plus Cup	1	Negative	1	100.00%
		Methadone		
UDS-3		* Denotes Correct Response		
AG: All Methods	122			
		*Negative	122	100.00%
AlereTriage	4	*Negative	4	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	1	*Negative	1	100.00%
E-Screen/Alere	26	*Negative	26	100.00%



Express Diag. Express Check	1	*Negative	1	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	14	*Negative	14	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Beckman AU	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Branan Med Corp QuickTox	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Medica Easy RA	2	*Negative	2	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
Alere USD UScreen Cup	9	*Negative	9	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Hemosure First Sign UDS Dipstick	15	*Negative	15	100.00%
Siemens Viva/ProE System	2	*Negative	2	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Abbott Architect c and ci Series	1	*Negative	1	100.00%
Consult Diag Drugs of Abuse Cups	2	*Negative	2	100.00%
Branan Med Corp ToxCup	6	*Negative	6	100.00%
American Screening Discover Plus Cup	1	*Negative	1	100.00%
McKesson Drugs of Abuse Test Cup	1	*Negative	1	100.00%
E-Screen mCup/Alere	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	3	*Negative	3	100.00%
Siemens Atellica	1	*Negative	1	100.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	122			
		*Negative	117	95.90%
		Positive	5	4.10%
AlereTriage	4	*Negative	4	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	1	*Negative	1	100.00%
E-Screen/Alere	26	*Negative	26	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	14	*Negative	14	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Beckman AU	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	4	80.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Branan Med Corp QuickTox	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	3	75.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Medica Easy RA	2	*Negative	2	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
Alere USD UScreen Cup	9	*Negative	8	88.89%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	2	66.67%
Hemosure First Sign UDS Dipstick	15	*Negative	15	100.00%
Siemens Viva/ProE System	2	*Negative	2	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Abbott Architect c and ci Series	1	*Negative	1	100.00%
Consult Diag Drugs of Abuse Cups	2	*Negative	2	100.00%
Branan Med Corp ToxCup	6	*Negative	6	100.00%
American Screening Discover Plus Cup	1	*Negative	1	100.00%
McKesson Drugs of Abuse Test Cup	1	*Negative	1	100.00%
E-Screen mCup/Alere	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	3	*Negative	2	66.67%
Siemens Atellica	1	*Negative	1	100.00%



CLIAwaived Inc Rapid Drug Cup	5	Positive	1	20.00%
Amedica Biotech AmediCheck Instant Test Cup	4	Positive	1	25.00%
Alere USD UScreen Cup	9	Positive	1	11.11%
Noble Medical Insta-Screen Test Cup	3	Positive	1	33.33%
Henry Schein OneStep+ DOA	3	Positive	1	33.33%
		Methamphetamine		
UDS-3		* Denotes Correct Response		
AG: All Methods	154			
		*Negative	154	100.00%
AlereTriage	4	*Negative	4	100.00%
E-Screen/Alere	50	*Negative	50	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	16	*Negative	16	100.00%
Instant Technologies i Screen	1	*Negative	1	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Branan Med Corp QuickTox	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Alere USD UScreen Cup	9	*Negative	9	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
CRL Stat Cup	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	15	*Negative	15	100.00%
Premier Biotech Bio-Cup	12	*Negative	12	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Consult Diag Drugs of Abuse Cups	2	*Negative	2	100.00%
Branan Med Corp ToxCup	6	*Negative	6	100.00%
American Screening Discover Plus Cup	1	*Negative	1	100.00%
McKesson Drugs of Abuse Test Cup	1	*Negative	1	100.00%
E-Screen mCup/Alere	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	4	*Negative	4	100.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	152			
		*Positive	138	90.79%
		Negative	14	9.21%
AlereTriage	4	*Positive	4	100.00%
E-Screen/Alere	48	*Positive	44	91.67%
Express Diag. Express Check	1	*Positive	1	100.00%
Alere iCup Drugs of Abuse	16	*Positive	15	93.75%
Instant Technologies i Screen	1	*Positive	1	100.00%
Integrated E-Z Split Key Cup	1	*Positive	1	100.00%
MedTox Profile II	2	*Positive	2	100.00%
Noble Med Split Specimen Cup	1	*Positive	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Positive	4	80.00%
ASC One Screen Drug Test Card	1	*Positive	1	100.00%
MedTox Profile V	2	*Positive	2	100.00%
Branan Med Corp QuickTox	1	*Positive	1	100.00%
Jant Accutest Multi Drug Screen	2	*Positive	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Positive	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Positive	4	100.00%
Alere USD UScreen Cup	9	*Positive	9	100.00%
Alere iScreen Drugs of Abuse Card	1	*Positive	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Positive	2	66.67%
Hemosure First Sign UDS Dipstick	15	*Positive	11	73.33%
Premier Biotech Bio-Cup	12	*Positive	12	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Positive	1	100.00%
Consult Diag Drugs of Abuse Cups	2	*Positive	2	100.00%
Branan Med Corp ToxCup	6	*Positive	6	100.00%
McKesson Drugs of Abuse Test Cup	1	*Positive	1	100.00%
E-Screen mCup/Alere	1	*Positive	1	100.00%



Henry Schein OneStep+ DOA	4	*Positive	4	100.00%
E-Screen/Alere	48	Negative	4	8.33%
Exp Results Integ Multi-Drug Screen Cup	1	Negative	1	100.00%
Alere iCup Drugs of Abuse	16	Negative	1	6.25%
CLIAwaived Inc Rapid Drug Cup	5	Negative	1	20.00%
Noble Medical Insta-Screen Test Cup	3	Negative	1	33.33%
CRL Stat Cup	1	Negative	1	100.00%
Hemosure First Sign UDS Dipstick	15	Negative	4	26.67%
American Screening Discover Plus Cup	1	Negative	1	100.00%
		Methaqualone		
UDS-3		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Beckman AU	1	*Negative	1	100.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Beckman AU	1	*Negative	1	100.00%
		Opiates		
UDS-3		* Denotes Correct Response		
AG: All Methods	182			
		*Negative	182	100.00%
AlereTriage	4	*Negative	4	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	1	*Negative	1	100.00%
E-Screen/Alere	51	*Negative	51	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	17	*Negative	17	100.00%
Instant Technologies i Screen	2	*Negative	2	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	5	*Negative	5	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Branan Med Corp QuickTox	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Medica Easy RA	2	*Negative	2	100.00%
Alere iCassette	5	*Negative	5	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
Alere USD UScreen Cup	8	*Negative	8	100.00%
ThermoScientific Indiko Plus	1	*Negative	1	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
CRL Stat Cup	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	15	*Negative	15	100.00%
Siemens Viva/ProE System	4	*Negative	4	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	12	*Negative	12	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Abbott Architect c and ci Series	1	*Negative	1	100.00%
Consult Diag Drugs of Abuse Cups	2	*Negative	2	100.00%
Branan Med Corp ToxCup	6	*Negative	6	100.00%
American Screening Discover Plus Cup	1	*Negative	1	100.00%
McKesson Drugs of Abuse Test Cup	1	*Negative	1	100.00%
E-Screen mCup/Alere	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	5	*Negative	5	100.00%
Siemens Atellica	1	*Negative	1	100.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	181			
		*Negative	178	98.34%
		Positive	3	1.66%
AlereTriage	4	*Negative	4	100.00%



Siemens Dim/AR/ES/RxL/Xpand/EXL	1	*Negative	1	100.00%
E-Screen/Alere	50	*Negative	50	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	17	*Negative	16	94.12%
Instant Technologies i Screen	2	*Negative	2	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	5	*Negative	5	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Branan Med Corp QuickTox	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Medica Easy RA	2	*Negative	2	100.00%
Alere iCassette	5	*Negative	5	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
Alere USD UScreen Cup	8	*Negative	8	100.00%
ThermoScientific Indiko Plus	1	*Negative	1	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
CRL Stat Cup	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	15	*Negative	15	100.00%
Siemens Viva/ProE System	4	*Negative	4	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	12	*Negative	12	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Abbott Architect c and ci Series	1	*Negative	1	100.00%
Consult Diag Drugs of Abuse Cups	2	*Negative	2	100.00%
Branan Med Corp ToxCup	6	*Negative	6	100.00%
McKesson Drugs of Abuse Test Cup	1	*Negative	1	100.00%
E-Screen mCup/Alere	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	5	*Negative	4	80.00%
Siemens Atellica	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	17	Positive	1	5.88%
American Screening Discover Plus Cup	1	Positive	1	100.00%
Henry Schein OneStep+ DOA	5	Positive	1	20.00%
		Oxycodone		
UDS-3		* Denotes Correct Response		
AG: All Methods	113			
		*Negative	113	100.00%
E-Screen/Alere	18	*Negative	18	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	14	*Negative	14	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Beckman AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	5	*Negative	5	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Branan Med Corp QuickTox	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Medica Easy RA	2	*Negative	2	100.00%
Alere iCassette	2	*Negative	2	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
Alere USD UScreen Cup	9	*Negative	9	100.00%
Alere iScreen Drugs of Abuse Card	2	*Negative	2	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Siemens Viva/ProE System	2	*Negative	2	100.00%



TFS DRI Assay	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	12	*Negative	12	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Abbott Architect c and ci Series	1	*Negative	1	100.00%
Consult Diag Drugs of Abuse Cups	2	*Negative	2	100.00%
Branan Med Corp ToxCup	6	*Negative	6	100.00%
American Screening Discover Plus Cup	1	*Negative	1	100.00%
McKesson Drugs of Abuse Test Cup	1	*Negative	1	100.00%
E-Screen mCup/Alere	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	4	*Negative	4	100.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	113			
		*Negative	112	99.12%
		Positive	1	0.88%
E-Screen/Alere	18	*Negative	18	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	14	*Negative	14	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Beckman AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	5	*Negative	5	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Branan Med Corp QuickTox	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Medica Easy RA	2	*Negative	2	100.00%
Alere iCassette	2	*Negative	2	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
Alere USD UScreen Cup	9	*Negative	9	100.00%
Alere iScreen Drugs of Abuse Card	2	*Negative	2	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Siemens Viva/ProE System	2	*Negative	2	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	12	*Negative	12	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Abbott Architect c and ci Series	1	*Negative	1	100.00%
Consult Diag Drugs of Abuse Cups	2	*Negative	2	100.00%
Branan Med Corp ToxCup	6	*Negative	6	100.00%
McKesson Drugs of Abuse Test Cup	1	*Negative	1	100.00%
E-Screen mCup/Alere	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	4	*Negative	4	100.00%
American Screening Discover Plus Cup	1	Positive	1	100.00%
		PCP (Phencyclidine)		
UDS-3		* Denotes Correct Response		
AG: All Methods	163			
		*Negative	163	100.00%
AlereTriage	4	*Negative	4	100.00%
E-Screen/Alere	51	*Negative	51	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	12	*Negative	12	100.00%
Instant Technologies i Screen	2	*Negative	2	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	3	*Negative	3	100.00%
Beckman AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	5	*Negative	5	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Branan Med Corp QuickTox	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	1	*Negative	1	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	1	*Negative	1	100.00%



Medica Easy RA	1	*Negative	1	100.00%
Alere iCassette	5	*Negative	5	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
Alere USD UScreen Cup	7	*Negative	7	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
CRL Stat Cup	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	15	*Negative	15	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	12	*Negative	12	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Consult Diag Drugs of Abuse Cups	2	*Negative	2	100.00%
Branan Med Corp ToxCup	6	*Negative	6	100.00%
McKesson Drugs of Abuse Test Cup	1	*Negative	1	100.00%
E-Screen mCup/Alere	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	5	*Negative	5	100.00%
Siemens Atellica	1	*Negative	1	100.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	162			
		*Negative	161	99.38%
		Positive	1	0.62%
AlereTriage	4	*Negative	4	100.00%
E-Screen/Alere	50	*Negative	50	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	12	*Negative	12	100.00%
Instant Technologies i Screen	2	*Negative	2	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	3	*Negative	3	100.00%
Beckman AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	5	*Negative	5	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Branan Med Corp QuickTox	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	1	*Negative	1	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	1	*Negative	1	100.00%
Medica Easy RA	1	*Negative	1	100.00%
Alere iCassette	5	*Negative	5	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
Alere USD UScreen Cup	7	*Negative	7	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
CRL Stat Cup	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	15	*Negative	15	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	12	*Negative	12	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Consult Diag Drugs of Abuse Cups	2	*Negative	2	100.00%
Branan Med Corp ToxCup	6	*Negative	6	100.00%
McKesson Drugs of Abuse Test Cup	1	*Negative	1	100.00%
E-Screen mCup/Alere	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	5	*Negative	4	80.00%
Siemens Atellica	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	5	Positive	1	20.00%
		Propoxyphene		
UDS-3		* Denotes Correct Response		
AG: All Methods	42			
		*Negative	42	100.00%
E-Screen/Alere	4	*Negative	4	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	10	*Negative	10	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Beckman AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	5	*Negative	5	100.00%
CLIAwaived Inc Rapid Drug Cup	1	*Negative	1	100.00%



MedTox Profile V	2	*Negative	2	100.00%
Alere iCassette	2	*Negative	2	100.00%
Consult Diag Drugs of Abuse Cups	2	*Negative	2	100.00%
Branan Med Corp ToxCup	6	*Negative	6	100.00%
McKesson Drugs of Abuse Test Cup	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	5	*Negative	5	100.00%
Siemens Atellica	1	*Negative	1	100.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	42			
			41	97.62%
		Positive	1	2.38%
E-Screen/Alere	4	*Negative	4	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	10	*Negative	10	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Beckman AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	5	*Negative	5	100.00%
CLIAwaived Inc Rapid Drug Cup	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Alere iCassette	2	*Negative	2	100.00%
Consult Diag Drugs of Abuse Cups	2	*Negative	2	100.00%
Branan Med Corp ToxCup	6	*Negative	6	100.00%
McKesson Drugs of Abuse Test Cup	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	5	*Negative	4	80.00%
Siemens Atellica	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	5	Positive	1	20.00%
		Tricyclic Antidepressants		
UDS-3		* Denotes Correct Response		
AG: All Methods	54			
		*Positive	47	87.04%
		Negative	7	12.96%
AlereTriage	4	*Positive	4	100.00%
Alere iCup Drugs of Abuse	11	*Positive	11	100.00%
Innovacon Multi-Clin	4	*Positive	3	75.00%
MedTox Profile V	2	*Positive	2	100.00%
Branan Med Corp QuickTox	1	*Positive	1	100.00%
Jant Accutest Multi Drug Screen	2	*Positive	2	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
Alere iCassette	2	*Positive	2	100.00%
Alere USD UScreen Cup	6	*Positive	6	100.00%
Alere iScreen Drugs of Abuse Card	1	*Positive	1	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Positive	1	100.00%
Consult Diag Drugs of Abuse Cups	2	*Positive	2	100.00%
Branan Med Corp ToxCup	6	*Positive	6	100.00%
McKesson Drugs of Abuse Test Cup	1	*Positive	1	100.00%
Henry Schein OneStep+ DOA	4	*Positive	4	100.00%
Innovacon Multi-Clin	4	Negative	1	25.00%
CLIAwaived Inc Rapid Drug Cup	1	Negative	1	100.00%
Amedica Biotech AmediCheck Instant Test Cup	1	Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	Negative	3	100.00%
American Screening Discover Plus Cup	1	Negative	1	100.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	54			
		*Negative	54	100.00%
AlereTriage	4	*Negative	4	100.00%
Alere iCup Drugs of Abuse	11	*Negative	11	100.00%
Innovacon Multi-Clin	4	*Negative	4	100.00%
CLIAwaived Inc Rapid Drug Cup	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Branan Med Corp QuickTox	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	1	*Negative	1	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
Alere iCassette	2	*Negative	2	100.00%
Alere USD UScreen Cup	6	*Negative	6	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Consult Diag Drugs of Abuse Cups	2	*Negative	2	100.00%



Branan Med Corp ToxCup	6	*Negative	6	100.00%
American Screening Discover Plus Cup	1	*Negative	1	100.00%
McKesson Drugs of Abuse Test Cup	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	4	*Negative	4	100.00%
		Buprenorphine		
UDS-3		* Denotes Correct Response		
PG: Exp Results Integ Multi-Drug Screen Cup	1			
		Negative	1	100.00%
UDS-3		* Denotes Correct Response		
PG: Alere iCup Drugs of Abuse	11			
		Negative	6	54.55%
		Positive	5	45.45%
UDS-3		* Denotes Correct Response		
PG: CLIAwaived Inc Rapid Drug Cup	5			
		*Positive	5	100.00%
UDS-3		* Denotes Correct Response		
PG: ASC One Screen Drug Test Card	1			
		*Positive	1	100.00%
UDS-3		* Denotes Correct Response		
PG: MedTox Profile V	1			
		*Positive	1	100.00%
UDS-3		* Denotes Correct Response		
PG: Jant Accutest Multi Drug Screen	1			
		*Positive	1	100.00%
UDS-3		* Denotes Correct Response		
PG: Mindray BS-200	2			
		*Positive	2	100.00%
UDS-3		* Denotes Correct Response		
PG: Medica Easy RA	2			
		*Positive	2	100.00%
UDS-3		* Denotes Correct Response		
PG: Alere USD UScreen Cup	2			
		*Positive	2	100.00%
UDS-3		* Denotes Correct Response		
PG: Alere iScreen Drugs of Abuse Card	2			
		*Positive	2	100.00%
UDS-3		* Denotes Correct Response		
PG: Noble Medical Insta-Screen Test Cup	3			
		*Positive	3	100.00%
UDS-3		* Denotes Correct Response		
PG: Siemens Viva/ProE System	5			
		*Positive	5	100.00%
UDS-3		* Denotes Correct Response		
PG: TFS DRI Assay	1			
		*Positive	1	100.00%
UDS-3		* Denotes Correct Response		
PG: Premier Biotech Bio-Cup	12			
		*Positive	12	100.00%
UDS-3		* Denotes Correct Response		
PG: Consult Diagnostics Drugs of Abuse Cups	2			
		Negative	1	50.00%
		Positive	1	50.00%
UDS-3		* Denotes Correct Response		
PG: Branan Med Corp ToxCup	6			
		*Negative	6	100.00%
UDS-3		* Denotes Correct Response		
PG: American Screening Discover Plus Cup	1			
		Negative	1	100.00%
UDS-3		* Denotes Correct Response		
PG: McKesson Drugs of Abuse Test Cup	1			
		*Positive	1	100.00%
UDS-3		* Denotes Correct Response		
PG: Henry Schein OneStep+ DOA	4			
		Negative	3	75.00%
		Positive	1	25.00%
UDS-4		* Denotes Correct Response		
PG: Henry Schein OneStep+ DOA	63			



		*Negative	62	98.41%
		Positive	1	1.59%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	11	*Negative	11	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	1	*Negative	1	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Medica Easy RA	2	*Negative	2	100.00%
Alere USD UScreen Cup	2	*Negative	2	100.00%
Alere iScreen Drugs of Abuse Card	2	*Negative	2	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Siemens Viva/ProE System	5	*Negative	5	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	12	*Negative	12	100.00%
Consult Diag Drugs of Abuse Cups	2	*Negative	2	100.00%
Branan Med Corp ToxCup	6	*Negative	6	100.00%
American Screening Discover Plus Cup	1	*Negative	1	100.00%
McKesson Drugs of Abuse Test Cup	1	*Negative	1	100.00%
Henry Schein OneStep+ DOA	4	*Negative	3	75.00%
Henry Schein OneStep+ DOA		Positive	1	25.00%
		6-Monoacetylmorphine (6 AM)		
UDS-3		* Denotes Correct Response		
AG: All Methods	4			
		*Positive	3	75.00%
		Negative	1	25.00%
Siemens Viva/ProE System	4	*Positive	3	75.00%
Siemens Viva/ProE System		Negative	1	25.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	4			
		*Negative	4	100.00%
Siemens Viva/ProE System	4	*Negative	4	100.00%
		EDDP		
UDS-3		* Denotes Correct Response		
AG: All Methods	15			
		*Negative	15	100.00%
Siemens Viva/ProE System	3	*Negative	3	100.00%
Premier Biotech Bio-Cup	12	*Negative	12	100.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	15			
		*Negative	15	100.00%
Siemens Viva/ProE System	3	*Negative	3	100.00%
Premier Biotech Bio-Cup	12	*Negative	12	100.00%
		Hydrocodone		
UDS-3		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect c and ci Series	1	*Negative	1	100.00%
UDS-4		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect c and ci Series	1	*Negative	1	100.00%
756 - HSV Antibodies	Labs	Response	Number	Percent
		HSV-2 Antibody		
VA-3		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Biokit HSV-II	1	*Negative	1	100.00%
VA-4		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Biokit HSV-II	1	*Positive	1	100.00%
770 - Viral Antigen Detection	Labs	Response	Number	Percent
		Influenza A - Aq Detect		
VAD-6		* Denotes Correct Response		
AG: All Methods	50			
		*Positive	50	100.00%



Alere BinaxNOW Influenza A+B Card	5	*Positive	5	100.00%
Sekisui Diagnostics OSOM Influenza A & B	8	*Positive	8	100.00%
Quidel QuickVue A+B	12	*Positive	12	100.00%
Quidel Sofia/Sofia 2 Influenza A+B	7	*Positive	7	100.00%
BD Veritor Flu A+B	5	*Positive	5	100.00%
Henry Schein OneStep + Influenza A&B (distinguish)	2	*Positive	2	100.00%
Cepheid GeneXpert - Xpert Flu	7	*Positive	7	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	1	*Positive	1	100.00%
Consult Diag Influenza A & B	3	*Positive	3	100.00%
VAD-7		* Denotes Correct Response		
AG: All Methods	50			
		*Negative	50	100.00%
Alere BinaxNOW Influenza A+B Card	5	*Negative	5	100.00%
Sekisui Diagnostics OSOM Influenza A & B	8	*Negative	8	100.00%
Quidel QuickVue A+B	12	*Negative	12	100.00%
Quidel Sofia/Sofia 2 Influenza A+B	7	*Negative	7	100.00%
BD Veritor Flu A+B	5	*Negative	5	100.00%
Henry Schein OneStep + Influenza A&B (distinguish)	2	*Negative	2	100.00%
Cepheid GeneXpert - Xpert Flu	7	*Negative	7	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	1	*Negative	1	100.00%
Consult Diag Influenza A & B	3	*Negative	3	100.00%
VAD-8		* Denotes Correct Response		
AG: All Methods	50			
		*Positive	49	98.00%
		Negative	1	2.00%
Alere BinaxNOW Influenza A+B Card	5	*Positive	5	100.00%
Sekisui Diagnostics OSOM Influenza A & B	8	*Positive	8	100.00%
Quidel QuickVue A+B	12	*Positive	12	100.00%
Quidel Sofia/Sofia 2 Influenza A+B	7	*Positive	6	85.71%
BD Veritor Flu A+B	5	*Positive	5	100.00%
Henry Schein OneStep + Influenza A&B (distinguish)	2	*Positive	2	100.00%
Cepheid GeneXpert - Xpert Flu	7	*Positive	7	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	1	*Positive	1	100.00%
Consult Diag Influenza A & B	3	*Positive	3	100.00%
Quidel Sofia/Sofia 2 Influenza A+B	7	Negative	1	14.29%
VAD-9		* Denotes Correct Response		
AG: All Methods	50			
		*Negative	49	98.00%
		Positive	1	2.00%
Alere BinaxNOW Influenza A+B Card	5	*Negative	5	100.00%
Sekisui Diagnostics OSOM Influenza A & B	8	*Negative	8	100.00%
Quidel QuickVue A+B	12	*Negative	12	100.00%
Quidel Sofia/Sofia 2 Influenza A+B	7	*Negative	6	85.71%
BD Veritor Flu A+B	5	*Negative	5	100.00%
Henry Schein OneStep + Influenza A&B (distinguish)	2	*Negative	2	100.00%
Cepheid GeneXpert - Xpert Flu	7	*Negative	7	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	1	*Negative	1	100.00%
Consult Diag Influenza A & B	3	*Negative	3	100.00%
Quidel Sofia/Sofia 2 Influenza A+B	7	Positive	1	14.29%
VAD-10		* Denotes Correct Response		
AG: All Methods	50			
		*Negative	50	100.00%
Alere BinaxNOW Influenza A+B Card	5	*Negative	5	100.00%
Sekisui Diagnostics OSOM Influenza A & B	8	*Negative	8	100.00%
Quidel QuickVue A+B	12	*Negative	12	100.00%
Quidel Sofia/Sofia 2 Influenza A+B	7	*Negative	7	100.00%
BD Veritor Flu A+B	5	*Negative	5	100.00%
Henry Schein OneStep + Influenza A&B (distinguish)	2	*Negative	2	100.00%
Cepheid GeneXpert - Xpert Flu	7	*Negative	7	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	1	*Negative	1	100.00%
Consult Diag Influenza A & B	3	*Negative	3	100.00%
Influenza B - Aq Detect				
VAD-6		* Denotes Correct Response		
AG: All Methods	50			
		*Negative	50	100.00%
Alere BinaxNOW Influenza A+B Card	5	*Negative	5	100.00%
Sekisui Diagnostics OSOM Influenza A & B	8	*Negative	8	100.00%
Quidel QuickVue A+B	12	*Negative	12	100.00%
Quidel Sofia/Sofia 2 Influenza A+B	7	*Negative	7	100.00%



BD Veritor Flu A+B	5	*Negative	5	100.00%
Henry Schein OneStep + Influenza A&B (distinguish)	2	*Negative	2	100.00%
Cepheid GeneXpert - Xpert Flu	7	*Negative	7	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	1	*Negative	1	100.00%
Consult Diag Influenza A & B	3	*Negative	3	100.00%
VAD-7		* Denotes Correct Response		
AG: All Methods	50			
		*Negative	50	100.00%
Alere BinaxNOW Influenza A+B Card	5	*Negative	5	100.00%
Sekisui Diagnostics OSOM Influenza A & B	8	*Negative	8	100.00%
Quidel QuickVue A+B	12	*Negative	12	100.00%
Quidel Sofia/Sofia 2 Influenza A+B	7	*Negative	7	100.00%
BD Veritor Flu A+B	5	*Negative	5	100.00%
Henry Schein OneStep + Influenza A&B (distinguish)	2	*Negative	2	100.00%
Cepheid GeneXpert - Xpert Flu	7	*Negative	7	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	1	*Negative	1	100.00%
Consult Diag Influenza A & B	3	*Negative	3	100.00%
VAD-8		* Denotes Correct Response		
AG: All Methods	50			
		*Negative	49	98.00%
		Positive	1	2.00%
Alere BinaxNOW Influenza A+B Card	5	*Negative	5	100.00%
Sekisui Diagnostics OSOM Influenza A & B	8	*Negative	8	100.00%
Quidel QuickVue A+B	12	*Negative	12	100.00%
Quidel Sofia/Sofia 2 Influenza A+B	7	*Negative	6	85.71%
BD Veritor Flu A+B	5	*Negative	5	100.00%
Henry Schein OneStep + Influenza A&B (distinguish)	2	*Negative	2	100.00%
Cepheid GeneXpert - Xpert Flu	7	*Negative	7	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	1	*Negative	1	100.00%
Consult Diag Influenza A & B	3	*Negative	3	100.00%
Quidel Sofia/Sofia 2 Influenza A+B	7	Positive	1	14.29%
VAD-9		* Denotes Correct Response		
AG: All Methods	50			
		*Positive	45	90.00%
		Negative	5	10.00%
Alere BinaxNOW Influenza A+B Card	5	*Positive	5	100.00%
Sekisui Diagnostics OSOM Influenza A & B	8	*Positive	8	100.00%
Quidel QuickVue A+B	12	*Positive	9	75.00%
Quidel Sofia/Sofia 2 Influenza A+B	7	*Positive	6	85.71%
BD Veritor Flu A+B	5	*Positive	5	100.00%
Henry Schein OneStep + Influenza A&B (distinguish)	2	*Positive	1	50.00%
Cepheid GeneXpert - Xpert Flu	7	*Positive	7	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	1	*Positive	1	100.00%
Consult Diag Influenza A & B	3	*Positive	3	100.00%
Quidel QuickVue A+B	12	Negative	3	25.00%
Quidel Sofia/Sofia 2 Influenza A+B	7	Negative	1	14.29%
Henry Schein OneStep + Influenza A&B (distinguish)	2	Negative	1	50.00%
VAD-10		* Denotes Correct Response		
AG: All Methods	50			
		*Positive	50	100.00%
Alere BinaxNOW Influenza A+B Card	5	*Positive	5	100.00%
Sekisui Diagnostics OSOM Influenza A & B	8	*Positive	8	100.00%
Quidel QuickVue A+B	12	*Positive	12	100.00%
Quidel Sofia/Sofia 2 Influenza A+B	7	*Positive	7	100.00%
BD Veritor Flu A+B	5	*Positive	5	100.00%
Henry Schein OneStep + Influenza A&B (distinguish)	2	*Positive	2	100.00%
Cepheid GeneXpert - Xpert Flu	7	*Positive	7	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	1	*Positive	1	100.00%
Consult Diag Influenza A & B	3	*Positive	3	100.00%
		RSV - Aq Detect		
VAD-6		* Denotes Correct Response		
AG: All Methods	25			
		*Negative	24	96.00%
		Positive	1	4.00%
BD EZ RSV	1	*Negative	1	100.00%
Alere BinaxNOW RSV	9	*Negative	9	100.00%
Quidel QuickVue RSV	4	*Negative	4	100.00%
Wampole Clearview RSV	3	*Negative	3	100.00%
Quidel Sofia/Sofia 2 RSV	3	*Negative	3	100.00%



Cepheid GeneXpert - Xpert RSV	4	*Negative	4	100.00%
Quidel QuickVue RSV 10	1	Positive	1	100.00%
VAD-7		* Denotes Correct Response		
AG: All Methods	25			
		*Positive	24	96.00%
		Negative	1	4.00%
BD EZ RSV	1	*Positive	1	100.00%
Alere BinaxNOW RSV	9	*Positive	9	100.00%
Quidel QuickVue RSV	4	*Positive	4	100.00%
Wampole Clearview RSV	3	*Positive	3	100.00%
Quidel Sofia/Sofia 2 RSV	3	*Positive	3	100.00%
Cepheid GeneXpert - Xpert RSV	4	*Positive	4	100.00%
Quidel QuickVue RSV 10	1	Negative	1	100.00%
VAD-8		* Denotes Correct Response		
AG: All Methods	25			
		*Positive	24	96.00%
		Negative	1	4.00%
BD EZ RSV	1	*Positive	1	100.00%
Alere BinaxNOW RSV	9	*Positive	9	100.00%
Quidel QuickVue RSV	4	*Positive	4	100.00%
Wampole Clearview RSV	3	*Positive	3	100.00%
Quidel Sofia/Sofia 2 RSV	3	*Positive	3	100.00%
Cepheid GeneXpert - Xpert RSV	4	*Positive	4	100.00%
Quidel QuickVue RSV 10	1	Negative	1	100.00%
VAD-9		* Denotes Correct Response		
AG: All Methods	25			
		*Negative	24	96.00%
		Positive	1	4.00%
BD EZ RSV	1	*Negative	1	100.00%
Alere BinaxNOW RSV	9	*Negative	9	100.00%
Quidel QuickVue RSV	4	*Negative	4	100.00%
Wampole Clearview RSV	3	*Negative	3	100.00%
Quidel Sofia/Sofia 2 RSV	3	*Negative	3	100.00%
Cepheid GeneXpert - Xpert RSV	4	*Negative	4	100.00%
Quidel QuickVue RSV 10	1	Positive	1	100.00%
VAD-10		* Denotes Correct Response		
AG: All Methods	25			
		*Negative	25	100.00%
BD EZ RSV	1	*Negative	1	100.00%
Alere BinaxNOW RSV	9	*Negative	9	100.00%
Quidel QuickVue RSV	4	*Negative	4	100.00%
Wampole Clearview RSV	3	*Negative	3	100.00%
Quidel QuickVue RSV 10	1	*Negative	1	100.00%
Quidel Sofia/Sofia 2 RSV	3	*Negative	3	100.00%
Cepheid GeneXpert - Xpert RSV	4	*Negative	4	100.00%
		Adenovirus		
VAD-6		* Denotes Correct Response		
AG: All Methods	10			
		*Positive	10	100.00%
SAS Adeno Test	10	*Positive	10	100.00%
VAD-7		* Denotes Correct Response		
AG: All Methods	10			
		*Negative	10	100.00%
SAS Adeno Test	10	*Negative	10	100.00%
VAD-8		* Denotes Correct Response		
AG: All Methods	10			
		*Negative	10	100.00%
SAS Adeno Test	10	*Negative	10	100.00%
VAD-9		* Denotes Correct Response		
AG: All Methods	10			
		*Negative	10	100.00%
SAS Adeno Test	10	*Negative	10	100.00%
VAD-10		* Denotes Correct Response		
AG: All Methods	10			
		*Positive	10	100.00%
SAS Adeno Test	10	*Positive	10	100.00%
932 - Hepatitis C Antibody (Waived);767 - Hepatitis Markers	Labs	Response	Number	Percent



		HCV (Waived)		
VM-6		* Denotes Correct Response		
AG: All Methods	4			
		*Negative	4	100.00%
OraSure QraQuick HCV Rapid Antibody Test	4	*Negative	4	100.00%
VM-7		* Denotes Correct Response		
AG: All Methods	4			
		*Negative	4	100.00%
OraSure QraQuick HCV Rapid Antibody Test	4	*Negative	4	100.00%
		HBsAg		
VM-6		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Roche Cobas e Series	1	*Positive	1	100.00%
VM-7		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Roche Cobas e Series	1	*Negative	1	100.00%
Abbott Architect ci4100	1	*Negative	1	100.00%
767 - Hepatitis Markers	Labs	Response	Number	Percent
		HBsAg		
VM-6		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Roche Cobas e Series	1	*Positive	1	100.00%
VM-7		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Roche Cobas e Series	1	*Negative	1	100.00%
Abbott Architect ci4100	1	*Negative	1	100.00%
VM-8		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Roche Cobas e Series	1	*Negative	1	100.00%
Abbott Architect ci4100	1	*Negative	1	100.00%
VM-9		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Roche Cobas e Series	1	*Negative	1	100.00%
Abbott Architect ci4100	1	*Negative	1	100.00%
VM-10		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Roche Cobas e Series	1	*Positive	1	100.00%
		HCV		
VM-6		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect ci4100	1	*Negative	1	100.00%
VM-7		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect ci4100	1	*Negative	1	100.00%
VM-8		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect ci4100	1	*Negative	1	100.00%
VM-9		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Abbott Architect ci4100	1	*Positive	1	100.00%
VM-10		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Abbott Architect ci4100	1	*Positive	1	100.00%
		Anti-HBc IgM		
VM-6		* Denotes Correct Response		



AG: All Methods	1			
		*Positive	1	100.00%
Abbott Architect ci4100	1	*Positive	1	100.00%
VM-7		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect ci4100	1	*Negative	1	100.00%
VM-8		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect ci4100	1	*Negative	1	100.00%
VM-9		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect ci4100	1	*Negative	1	100.00%
VM-10		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect ci4100	1	*Negative	1	100.00%
VM-6		HAV IgM		
AG: All Methods	1	* Denotes Correct Response		
		*Negative	1	100.00%
Abbott Architect ci4100	1	*Negative	1	100.00%
VM-7		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect ci4100	1	*Negative	1	100.00%
VM-8		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect ci4100	1	*Negative	1	100.00%
VM-9		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect ci4100	1	*Negative	1	100.00%
VM-10		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect ci4100	1	*Negative	1	100.00%
Anti-HBc Total (International)				
VM-6		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Positive	3	100.00%
VM-7		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Negative	3	100.00%
VM-8		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Positive	3	100.00%
VM-9		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Negative	3	100.00%
VM-10		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Positive	3	100.00%
HBsAb (International)				
VM-6		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Negative	3	100.00%
VM-7		* Denotes Correct Response		



AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Positive	3	100.00%
VM-8		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Positive	3	100.00%
VM-9		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Negative	3	100.00%
VM-10		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Negative	3	100.00%
		HBsAg (International)		
VM-6		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Positive	3	100.00%
VM-7		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Negative	3	100.00%
VM-8		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Negative	3	100.00%
VM-9		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Negative	3	100.00%
VM-10		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Positive	3	100.00%
		HCV (International)		
VM-6		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Negative	3	100.00%
VM-7		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Negative	3	100.00%
VM-8		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Negative	3	100.00%
VM-9		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Positive	3	100.00%
VM-10		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Positive	3	100.00%
771 - Vaginal Pathogens I	Labs	Response	Number	Percent
		Candida sp. - Non Culture		
VP-6		* Denotes Correct Response		
AG: All Methods	48			
		*Negative	47	97.92%
		Positive	1	2.08%
Affirm VP	47	*Negative	46	97.87%
BD Max	1	*Negative	1	100.00%
Affirm VP	47	Positive	1	2.13%
VP-7		* Denotes Correct Response		



AG: All Methods	48			
		*Negative	48	100.00%
Affirm VP	47	*Negative	47	100.00%
BD Max	1	*Negative	1	100.00%
VP-8		* Denotes Correct Response		
AG: All Methods	48			
		*Positive	45	93.75%
		Negative	3	6.25%
Affirm VP	47	*Positive	44	93.62%
BD Max	1	*Positive	1	100.00%
Affirm VP	47	Negative	3	6.38%
VP-9		* Denotes Correct Response		
AG: All Methods	48			
		*Negative	48	100.00%
Affirm VP	47	*Negative	47	100.00%
BD Max	1	*Negative	1	100.00%
VP-10		* Denotes Correct Response		
AG: All Methods	48			
		*Positive	48	100.00%
Affirm VP	47	*Positive	47	100.00%
BD Max	1	*Positive	1	100.00%
		G. vaginalis - Non Culture		
VP-6		* Denotes Correct Response		
AG: All Methods	48			
		*Negative	48	100.00%
Affirm VP	47	*Negative	47	100.00%
BD Max	1	*Negative	1	100.00%
VP-7		* Denotes Correct Response		
AG: All Methods	48			
		*Positive	48	100.00%
Affirm VP	47	*Positive	47	100.00%
BD Max	1	*Positive	1	100.00%
VP-8		* Denotes Correct Response		
AG: All Methods	48			
		*Negative	47	97.92%
		Positive	1	2.08%
Affirm VP	47	*Negative	46	97.87%
BD Max	1	*Negative	1	100.00%
Affirm VP	47	Positive	1	2.13%
VP-9		* Denotes Correct Response		
AG: All Methods	48			
		*Positive	48	100.00%
Affirm VP	47	*Positive	47	100.00%
BD Max	1	*Positive	1	100.00%
VP-10		* Denotes Correct Response		
AG: All Methods	48			
		*Negative	46	95.83%
		Positive	2	4.17%
Affirm VP	47	*Negative	45	95.74%
BD Max	1	*Negative	1	100.00%
Affirm VP	47	Positive	2	4.26%
		T. vaginalis - Non Culture Tech		
VP-6		* Denotes Correct Response		
AG: All Methods	48			
		*Negative	47	97.92%
		Positive	1	2.08%
Affirm VP	47	*Negative	46	97.87%
BD Max	1	*Negative	1	100.00%
Affirm VP	47	Positive	1	2.13%
VP-7		* Denotes Correct Response		
AG: All Methods	48			
		*Positive	48	100.00%
Affirm VP	47	*Positive	47	100.00%
BD Max	1	*Positive	1	100.00%
VP-8		* Denotes Correct Response		
AG: All Methods	48			
		*Positive	46	95.83%
		Negative	2	4.17%



American Academy of Family Physicians
11400 Tomahawk Creek Parkway
Leawood, KS 66211-2672
1-800-274-7911
www.aafp.org/pt pt@aafp.org

Proficiency Testing Event Summary

AAFP PT 2019-B

Affirm VP	47	*Positive	45	95.74%
BD Max	1	*Positive	1	100.00%
Affirm VP	47	Negative	2	4.26%
VP-9		* Denotes Correct Response		
AG: All Methods	48			
		*Negative	48	100.00%
Affirm VP	47	*Negative	47	100.00%
BD Max	1	*Negative	1	100.00%
VP-10		* Denotes Correct Response		
AG: All Methods	48			
		*Negative	47	97.92%
		Positive	1	2.08%
Affirm VP	47	*Negative	46	97.87%
BD Max	1	*Negative	1	100.00%
Affirm VP	47	Positive	1	2.13%
683 - Antistreptolysin O (ASO);684 - C-Reactive Protein	Labs	Response	Number	Percent
		CRP Qualitative		