



666 - Afinion Glycohemoglobin;998 - CVS Afinion Glycohemoglobin		AFF-1																								
Glycohemoglobin	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	
PG: Afinion Glycohemoglobin	163	8.19	0.21	7.60-8.80	2.56																					
747 - Auto Diff 7 (5-Part Diff)		BCX-1					BCX-2					BCX-3					BCX-4					BCX-5				
White Blood Cell Count	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	
PG: Beckman Coulter ACT 5 diff (v 1.9)	14	9.61	0.23	8.20-11.00	2.40	14	19.87	0.55	16.90-22.90	2.76	14	7.31	0.22	6.20-8.40	3.01	14	16.02	0.50	13.60-18.40	3.13	14	4.87	0.16	4.20-5.60	3.27	
PG: Horiba ABX Diag Pentra 60	16	9.66	0.32	8.20-11.20	3.30	16	20.03	0.79	17.00-23.00	3.95	16	7.44	0.54	6.30-8.50	7.30	16	16.05	0.56	13.60-18.40	3.44	16	4.89	0.20	4.20-5.60	4.08	
SG: Select Beckman Coulter ACT Instruments (5-part)	18	9.57	0.26	8.20-11.00	2.71	18	19.76	0.68	16.80-22.80	3.43	18	7.27	0.31	6.20-8.40	4.25	18	15.98	0.53	13.60-18.40	3.31	18	4.84	0.19	4.10-5.50	3.96	
SG: Select Horiba ABX Instruments (5-part Diff)	23	9.63	0.29	8.20-11.00	3.02	23	19.85	0.73	16.90-22.90	3.67	23	7.38	0.49	6.30-8.50	6.62	23	15.93	0.56	13.50-18.30	3.52	23	4.87	0.19	4.20-5.60	3.88	
SG: International Horiba Instruments	26	9.63	0.28	8.20-11.00	2.92	26	19.83	0.69	16.80-22.80	3.48	26	7.38	0.46	6.30-8.50	6.22	26	15.93	0.53	13.50-18.30	3.33	26	4.87	0.18	4.20-5.60	3.67	
Lymphocyte	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	
PG: Beckman Coulter ACT 5 diff (v 1.9)	14	32.45	1.87	26.90-38.10	5.75	14	16.51	2.09	10.20-22.80	12.67	14	32.17	3.65	21.30-43.20	11.34	14	19.06	1.83	13.60-24.60	9.58	14	39.88	2.23	33.20-46.60	5.59	
PG: Horiba ABX Diag Pentra 60	15	37.95	3.69	26.90-49.10	9.71	15	18.31	3.24	8.60-28.00	17.70	16	34.39	4.03	22.30-46.50	11.72	15	20.89	3.28	11.10-30.70	15.69	16	43.95	2.18	37.50-50.50	4.95	
SG: Select Beckman Coulter ACT Instruments (5-part)	18	31.94	2.17	25.40-38.40	6.80	18	16.13	2.03	10.00-22.20	12.61	18	31.25	4.07	19.10-43.50	13.00	18	18.47	2.21	11.90-25.10	11.95	18	39.12	2.71	31.00-47.20	6.93	
SG: Select Horiba ABX Instruments (5-part Diff)	22	37.26	3.51	26.80-47.80	9.41	22	17.44	3.36	7.30-27.50	19.31	23	33.89	4.10	21.60-46.20	12.09	22	20.03	3.31	10.10-29.90	16.55	23	43.38	2.29	36.50-50.30	5.28	
SG: International Horiba Instruments	25	38.27	4.32	25.30-51.30	11.28	24	18.40	4.62	4.50-32.30	25.11	25	34.91	5.28	19.10-50.70	15.13	24	20.87	4.27	8.10-33.70	20.43	26	44.40	3.73	33.20-55.60	8.40	
Monocyte	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	
PG: Beckman Coulter ACT 5 diff (v 1.9)	12	2.50	0.55	0.85-4.15	22.04	12	2.61	1.36	0.00-6.69	52.11	13	1.86	0.84	0.00-4.39	45.32	13	1.00	0.41	0.00-2.25	41.50	13	1.66	0.37	0.56-2.76	22.11	
PG: Horiba ABX Diag Pentra 60	14	1.44	0.41	0.21-2.67	28.54	15	1.33	0.60	0.00-3.13	45.11	14	1.26	0.50	0.00-2.76	39.60	16	0.59	0.34	0.00-1.60	56.95	15	1.33	0.54	0.00-2.96	40.83	
SG: Select Beckman Coulter ACT Instruments (5-part)	16	2.35	0.58	0.61-4.09	24.64	14	2.39	1.05	0.00-5.55	44.14	17	1.96	0.86	0.00-4.53	43.67	17	0.93	0.40	0.00-2.14	43.44	17	1.58	0.37	0.47-2.69	23.48	
SG: Select Horiba ABX Instruments (5-part Diff)	19	1.47	0.44	0.14-2.80	30.07	20	1.21	0.57	0.00-2.93	47.44	20	1.21	0.55	0.00-2.87	45.70	22	0.59	0.32	0.00-1.55	54.07	21	1.22	0.43	0.00-2.50	35.08	
SG: International Horiba Instruments	22	1.45	0.50	0.00-2.94	34.28	23	1.18	0.56	0.00-2.87	47.63	23	1.17	0.53	0.00-2.77	45.64	25	0.58	0.32	0.00-1.53	54.48	24	1.21	0.42	0.00-2.46	34.55	
Granulocyte	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	
PG: Beckman Coulter ACT 5 diff (v 1.9)	14	51.69	2.07	45.50-57.90	4.00	14	69.43	3.35	59.40-79.50	4.83	14	50.89	3.14	41.50-60.30	6.17	14	64.78	2.19	58.20-71.40	3.38	14	42.32	2.00	36.30-48.30	4.75	
PG: Horiba ABX Diag Pentra 60	16	52.76	4.79	38.40-67.20	9.07	16	76.99	4.07	64.80-89.20	5.29	16	57.89	3.65	47.00-68.80	6.30	16	71.46	4.31	58.60-84.40	6.03	16	46.22	1.70	41.10-51.30	3.68	
SG: Select Beckman Coulter ACT Instruments (5-part)	17	52.07	2.15	45.70-58.50	4.13	17	69.58	3.17	60.10-79.10	4.55	17	51.04	3.09	41.70-60.30	6.06	17	65.08	2.37	58.00-72.20	3.64	17	42.77	2.30	35.90-49.70	5.37	
SG: Select Horiba ABX Instruments (5-part Diff)	22	52.95	4.84	38.40-67.40	9.15	22	77.23	5.11	61.90-92.50	6.62	22	57.66	4.65	43.80-71.70	8.06	22	71.68	4.76	57.40-86.00	6.64	22	45.94	2.48	38.50-53.30	5.40	
SG: International Horiba Instruments	25	52.31	4.86	37.70-66.90	9.29	25	75.90	6.08	57.70-94.10	8.01	25	56.33	5.71	39.20-73.40	10.14	25	70.52	5.54	53.90-87.10	7.86	25	45.18	3.51	34.70-55.70	7.77	
Eosinophil	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	
PG: Beckman Coulter ACT 5 diff (v 1.9)	14	6.55	1.10	3.20-9.80	16.92	14	2.95	0.50	1.40-4.40	17.24	14	7.28	2.02	1.20-13.40	27.67	14	7.24	0.79	4.80-9.60	10.97	14	7.53	1.42	3.20-11.80	18.93	
PG: Horiba ABX Diag Pentra 60	14	6.51	0.96	3.60-9.40	14.77	16	2.38	0.47	1.00-3.80	19.58	15	6.03	0.91	3.30-8.70	15.17	15	5.86	1.00	2.90-8.90	16.95	16	7.96	0.97	5.10-10.90	12.13	
SG: Select Beckman Coulter ACT Instruments (5-part)	18	6.48	1.03	3.40-9.60	15.85	18	3.17	0.79	0.80-5.60	24.69	18	7.64	1.93	1.80-13.40	25.39	18	7.27	0.81	4.90-9.70	11.10	18	7.61	1.31	3.60-11.60	17.37	
SG: Select Horiba ABX Instruments (5-part Diff)	21	6.42	0.83	3.90-8.90	12.97	23	2.46	0.68	0.50-4.50	27.20	22	6.00	1.00	3.00-9.00	16.67	22	6.08	0.95	3.30-8.90	15.57	23	8.33	1.12	4.90-11.70	13.49	
SG: International Horiba Instruments	24	6.25	0.96	3.30-9.10	15.48	26	2.42	0.66	0.40-4.40	27.50	25	5.82	1.05	2.60-8.90	18.10	25	6.01	0.93	3.20-8.80	15.50	26	8.13	1.36	4.00-12.20	16.67	
Basophil	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	
PG: Beckman Coulter ACT 5 diff (v 1.9)	14	6.91	0.21	6.27-7.54	3.07	14	8.13	0.72	5.98-10.29	8.83	14	7.64	0.58	5.88-9.40	7.66	14	7.85	0.46	6.49-9.22	5.79	14	8.55	0.61	6.72-10.39	7.14	
PG: Horiba ABX Diag Pentra 60	16	0.35	0.05	0.20-0.51	14.74	16	0.40	0.08	0.16-0.65	20.43	16	0.39	0.03	0.28-0.49	8.81	16	0.41	0.03	0.33-0.48	6.16	16	0.44	0.05	0.29-0.59	11.42	
SG: Select Beckman Coulter ACT Instruments (5-part)	17	6.92	0.20	6.34-7.51	2.83	17	8.28	0.73	6.08-10.47	8.83	17	7.76	0.63	5.88-9.63	8.07	17	7.93	0.48	6.48-9.37	6.07	17	8.66	0.66	6.70-10.63	7.57	
SG: Select Horiba ABX Instruments (5-part Diff)	21	0.35	0.06	0.17-0.53	17.30	21	0.41	0.10	0.13-0.70	23.29	21	0.40	0.04	0.28-0.51	9.72	21	0.42	0.07	0.22-0.62	16.23	21	0.44	0.06	0.26-0.62	13.50	
SG: International Horiba Instruments	24	0.35	0.06	0.17-0.52	16.99	24	0.41	0.09	0.13-0.69	22.77	24	0.39	0.04	0.27-0.51	10.41	24	0.42	0.06	0.23-0.61	15.28	24	0.44	0.06	0.27-0.62	13.21	
Red Blood Cell Count	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	
PG: Beckman Coulter ACT 5 diff (v 1.9)	14	2.56	0.03	2.41-2.71	1.29	14	4.99	0.08	4.69-5.29	1.60	14	6.06	0.09	5.70-6.42	1.50	14	4.28	0.05	4.02-4.54	1.24	14	4.93	0.07	4.63-5.23	1.34	
PG: Horiba ABX Diag Pentra 60	16	2.56	0.04	2.41-2.71	1.45	16	4.99	0.07	4.69-5.29	1.50	16	6.04	0.09	5.68-6.40	1.49	16	4.26	0.08	4.00-4.52	1.90	16	4.90	0.07	4.61-5.19	1.53	
SG: Select Beckman Coulter ACT Instruments (5-part)	18	2.56	0.03	2.41-2.71	1.25	18	5.00	0.07	4.70-5.30	1.50	18	6.06	0.09	5.70-6.42	1.49	18	4.29	0.06	4.03-4.55	1.28	18	4.94	0.07	4.64-5.24	1.32	
SG: Select Horiba ABX Instruments (5-part Diff)	23	2.56	0.04	2.41-2.71	1.52	23	5.00	0.08	4.70-5.30	1.60	23	6.05	0.09	5.69-6.41	1.50	23	4.27	0.08	4.01-4.53	1.97	23	4.91	0.08	4.62-5.20	1.61	
SG: International Horiba Instruments	26	2.56	0.04	2.41-2.71	1.45	2																				



Test Name	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Beckman Coulter AcT Instruments (5-part)	18	11.87	0.28	11.10-12.70	2.27	18	11.64	0.34	10.60-12.60	2.93	18	11.35	0.31	10.40-12.20	2.74	18	12.67	0.46	11.30-14.10	3.62	18	11.75	0.35	10.60-12.80	2.99
SG: Select Horiba ABX Instruments (5-part Diff)	23	11.85	0.44	10.60-13.20	3.70	23	11.50	0.33	10.50-12.50	2.87	23	11.42	0.36	10.30-12.50	3.16	23	12.65	0.42	11.30-13.90	3.33	23	11.75	0.34	10.70-12.70	2.91
SG: International Horiba Instruments	26	11.92	0.48	10.50-13.30	4.03	26	11.62	0.48	10.20-13.00	4.14	26	11.56	0.52	10.00-13.20	4.48	26	12.77	0.54	11.20-14.40	4.22	26	11.87	0.48	10.50-13.30	4.03
Platelet Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman Coulter AcT 5 diff (v 1.9)	14	81.60	5.17	62.00-103.00	6.34	14	465.80	12.51	350.00-583.00	2.68	14	308.60	11.93	232.00-386.00	3.85	14	219.87	8.89	165.00-275.00	4.05	14	154.60	7.66	116.00-194.00	4.97
PG: Horiba ABX Diag Pentra 60	16	78.94	3.21	59.00-99.00	4.05	16	472.63	14.84	355.00-591.00	3.13	16	311.31	10.29	233.00-389.00	3.31	16	222.38	7.95	167.00-278.00	3.56	16	157.19	7.39	118.00-196.00	4.71
SG: Select Beckman Coulter AcT Instruments (5-part)	18	80.89	4.92	61.00-101.00	6.05	18	467.47	13.44	350.00-584.00	2.87	18	307.11	13.87	230.00-384.00	4.53	18	220.21	9.18	165.00-275.00	4.18	18	155.32	7.78	116.00-194.00	5.03
SG: Select Horiba ABX Instruments (5-part Diff)	23	79.30	3.61	59.00-99.00	4.56	23	468.26	18.64	351.00-585.00	3.97	23	309.30	12.04	232.00-386.00	3.88	23	220.48	11.96	165.00-275.00	5.45	23	156.48	8.69	117.00-195.00	5.58
SG: International Horiba Instruments	26	79.15	3.50	59.00-99.00	4.43	26	467.58	18.50	351.00-585.00	3.95	26	309.35	12.03	232.00-386.00	3.88	26	220.04	11.88	165.00-275.00	5.41	26	156.73	8.36	118.00-196.00	5.35
915 - Body Fluid pH	BFPH-1																								
pH - Body Fluid	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Nitrazine Paper	5	7.25	0.35	6.30-8.30	4.79																				
702 - Blood Gases	BG-1																								
Blood Gas - pCO2	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Non - i-STAT Blood Gas Instruments	8	33.31	2.41	28.00-38.00	7.27	8	29.06	2.26	24.00-34.00	7.93	7	69.60	3.26	64.00-76.00	4.71	8	58.64	3.97	54.00-64.00	6.78	8	52.13	3.70	47.00-57.00	7.12
SG: International Blood Gas Inst	9	32.96	2.70	28.00-38.00	8.18	9	28.97	1.95	24.00-34.00	6.55	9	68.82	3.57	63.00-75.00	5.22	8	57.28	3.57	52.00-62.00	6.32	9	51.94	3.17	47.00-57.00	6.15
Blood Gas - pH	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Non - i-STAT Blood Gas Instruments	8	7.47	0.01	7.43-7.51	0.17	8	7.45	0.01	7.41-7.49	0.15	8	7.15	0.01	7.11-7.19	0.21	8	7.18	0.01	7.14-7.22	0.13	8	7.24	0.01	7.20-7.28	0.10
SG: International Blood Gas Inst	9	7.49	0.03	7.45-7.53	0.40	9	7.46	0.02	7.42-7.50	0.31	9	7.15	0.01	7.11-7.19	0.17	9	7.19	0.02	7.15-7.23	0.32	9	7.25	0.02	7.21-7.29	0.29
Blood Gas - pO2	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Non - i-STAT Blood Gas Instruments	7	112.13	4.44	99.00-125.00	3.93	7	147.96	4.15	136.00-160.00	2.77	7	67.86	9.37	40.00-96.00	13.82	7	198.86	4.67	185.00-213.00	2.36	7	180.03	6.97	159.00-201.00	3.89
SG: International Blood Gas Inst	9	113.57	4.79	100.00-128.00	4.21	9	147.06	4.04	135.00-159.00	2.72	7	69.07	8.67	43.00-95.00	12.61	9	197.91	4.46	185.00-212.00	2.27	9	179.04	6.35	160.00-198.00	3.52
714 - Complete Cardiac Markers; 715 - Basic Cardiac Markers; 711 - BNP; 720 - Cardiac Markers I; 740 - D-Dimer; 689 - Cardiac Profile	CDM-1N																								
BNP	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat	42	478.31	44.73	344.10-612.50	9.35	42	2701.69	211.07	2068.50-3334.90	7.81	32	53.47	8.46	28.10-78.90	15.81	35	1727.23	171.96	1211.30-2243.10	9.96	35	906.49	94.96	621.60-1191.40	10.48
PG: AlereTriage	40	433.67	59.08	256.50-610.90	13.62	12	2398.33	271.09	1585.00-3211.60	11.30	7	65.60	12.90	26.90-104.30	19.66	8	1403.75	133.94	1002.00-1805.60	9.54	6	802.33	99.97	502.40-1102.20	12.46
AG: All Group	85	450.73	63.73	259.50-641.90	14.14	55	2643.08	262.60	1855.30-3430.90	9.94	36	54.97	7.97	31.10-78.90	14.49	43	1667.05	207.73	1043.80-2290.20	12.46	41	891.24	101.51	586.70-1195.70	11.39
CK-MB (Quantitative)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat	5	10.20	0.00	10.20-10.20	0.00	5	62.00	0.00	62.00-62.00	0.00	5	3.20	0.00	3.20-3.20	0.00	5	28.70	0.00	28.70-28.70	0.00	5	16.70	0.00	16.70-16.70	0.00
PG: AlereTriage	40	5.56	1.21	2.00-9.20	21.61	40	29.02	7.18	7.50-50.50	24.76	40	1.75	0.35	0.70-2.80	20.59	40	15.69	2.29	8.80-22.60	14.65	40	9.08	1.55	4.40-13.80	17.03
PG: TOSOH AIA	5	19.30	0.00	19.30-19.30	0.00	5	114.90	0.00	114.90-114.90	0.00	5	7.10	0.00	7.10-7.10	0.00	5	63.80	0.00	63.80-63.80	0.00	5	33.70	0.00	33.70-33.70	0.00
PG: Beckman Coulter UniCel Dxl 600 Access IA	5	18.60	0.00	18.60-18.60	0.00	5	115.60	0.00	115.60-115.60	0.00	5	6.60	0.00	6.60-6.60	0.00	5	59.30	0.00	59.30-59.30	0.00	5	34.20	0.00	34.20-34.20	0.00
PG: Intl OCD Vitros 250,350,400,500,750,950	5	13.00	0.00	13.00-13.00	0.00	5	73.10	0.00	73.10-73.10	0.00	5	4.80	0.00	4.80-4.80	0.00	5	39.60	0.00	39.60-39.60	0.00	5	21.30	0.00	21.30-21.30	0.00
D-dimer (Quantitative)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: AlereTriage	74	770.43	80.13	530.00-1010.00	10.40	45	2288.89	209.04	1662.00-2916.00	9.13	40	559.72	53.03	401.00-719.00	9.46	40	1458.00	108.99	1131.00-1785.00	7.48	40	1026.97	61.58	842.00-1212.00	6.00
AG: All Group	81	774.55	146.67	335.00-1215.00	18.94	51	2280.62	454.54	918.00-3645.00	19.93	46	562.31	116.98	211.00-913.00	20.82	46	1450.67	269.22	643.00-2259.00	18.55	46	1040.42	201.83	435.00-1645.00	19.40
Myoglobin (Quantitative)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: AlereTriage	34	79.68	14.47	36.30-123.10	18.16	34	306.79	66.83	106.30-507.30	21.78	34	30.04	4.87	15.40-44.60	16.23	34	188.97	30.06	98.80-279.20	15.90	33	121.35	25.21	45.70-196.90	20.78
AG: All Group	36	80.21	14.41	37.00-123.40	17.97	36	316.86	77.88	83.30-550.50	24.58	36	30.14	4.78	15.80-44.40	15.88	36	192.93	34.25	90.20-295.60	17.76	35	123.02	25.72	45.80-200.20	20.91
Troponin I (Quantitative)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat	46	4.30	0.57	2.58-6.03	13.37	47	36.98	5.81	19.54-54.42	15.72	39	0.42	0.06	0.25-0.58	13.49	39	18.01	2.38	10.86-25.15	13.23	39	8.75	0.88	6.13-11.38	10.01
PG: AlereTriage	45	0.36	0.08	0.12-0.59	21.91	43	7.30	1.37	3.19-11.42	18.80	41	0.05	0.01	0.03-0.07	15.49	38	2.48	0.54							



PG: Beckman (Olympus) AU	17	113.80	1.06	110.00-118.00	0.96	17	141.05	1.05	137.00-145.00	0.78	17	156.15	1.81	152.00-160.00	1.15	17	137.75	0.85	134.00-142.00	0.65	17	126.50	1.00	123.00-131.00	0.79
PG: Roche Cobas Integra / Roche Reagent	8	115.48	1.39	111.00-119.00	1.22	8	144.82	1.42	141.00-149.00	0.97	8	162.11	1.80	158.00-166.00	1.11	8	140.42	1.47	136.00-144.00	1.07	8	127.76	1.29	124.00-132.00	1.02
PG: Roche Cobas Mira / DCL Reagent (Sekisui)	5	120.00	0.00	116.00-124.00	0.00	5	166.00	0.00	162.00-170.00	0.00	5	354.00	0.00	350.00-358.00	0.00	5	159.00	0.00	155.00-163.00	0.00	5	134.00	0.00	130.00-138.00	0.00
PG: OCD Vitros 250,350,400,500,750,950	11	123.36	1.69	119.00-127.00	1.38	8	167.91	2.39	164.00-172.00	1.43	7	201.38	3.07	197.00-205.00	1.54	11	159.86	3.30	156.00-164.00	2.06	11	136.50	1.83	133.00-141.00	1.31
SG: Select Alfa Wasserman Instruments / Alfa Wasserman Reagent Groups	33	118.26	2.02	114.00-122.00	1.69	33	161.38	1.82	157.00-165.00	1.12	32	196.77	3.29	193.00-201.00	1.68	33	155.98	7.20	152.00-160.00	4.62	33	132.47	4.02	128.00-136.00	3.03
SG: Select Beckman Coulter Instruments (Chem)	18	113.96	1.06	110.00-118.00	0.96	18	141.30	1.19	137.00-145.00	0.85	18	156.78	2.37	153.00-161.00	1.53	18	137.86	0.89	134.00-142.00	0.65	18	126.52	0.98	123.00-131.00	0.79
SG: Select Roche Instruments / Roche Reagent	9	115.54	1.31	112.00-120.00	1.12	9	144.84	1.33	141.00-149.00	0.90	9	161.88	1.83	158.00-166.00	1.11	9	140.37	1.38	136.00-144.00	1.00	9	127.90	1.28	124.00-132.00	1.02
SG: Select OCD Vitros Instruments	12	123.20	1.74	119.00-127.00	1.38	9	167.75	2.34	164.00-172.00	1.37	8	201.78	3.11	198.00-206.00	1.53	12	159.80	3.19	156.00-164.00	2.00	12	136.33	1.88	132.00-140.00	1.40
SG: Abaxis Piccolo Instruments	79	123.44	1.76	119.00-127.00	1.46	17	149.11	2.23	145.00-153.00	1.48	17	167.32	2.14	163.00-171.00	1.26	17	144.95	1.87	141.00-149.00	1.31	17	131.47	1.98	127.00-135.00	1.53
PG: Intl Siemens RAPIDPoint 400/405 System	5	126.00	1.41	122.00-130.00	1.11	5	158.50	3.54	155.00-163.00	2.20	5	187.00	4.24	183.00-191.00	2.25	5	149.50	3.54	146.00-154.00	2.33	5	130.50	3.54	127.00-135.00	2.67
SG: Select Abbott Architect Instruments	9	115.00	1.12	111.00-119.00	0.96	9	143.78	0.97	140.00-148.00	0.69	9	160.44	2.01	156.00-164.00	1.25	9	139.89	0.78	136.00-144.00	0.57	9	127.67	0.71	124.00-132.00	0.55
SG: Select Electrolyte Instruments 1	7	119.14	1.22	115.00-123.00	1.01	7	159.14	2.12	155.00-163.00	1.32	7	191.57	3.26	188.00-196.00	1.72	7	152.14	1.86	148.00-156.00	1.25	7	132.29	1.50	128.00-136.00	1.14
SG: International OCD Vitros Instruments	14	123.29	1.79	119.00-127.00	1.46	11	167.43	2.93	163.00-171.00	1.74	10	202.00	3.10	198.00-206.00	1.53	14	159.76	3.19	156.00-164.00	2.00	14	136.41	1.91	132.00-140.00	1.40
SG: International Abaxis Piccolo Instruments	79	123.46	1.78	119.00-127.00	1.46	17	148.88	2.26	145.00-153.00	1.54	17	167.12	2.18	163.00-171.00	1.32	17	144.71	1.83	141.00-149.00	1.24	17	131.41	2.09	127.00-135.00	1.60
Potassium (K+)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	15	3.36	0.13	2.90-3.90	3.82	16	5.53	0.17	5.00-6.00	3.09	16	7.46	0.25	7.00-8.00	3.33	16	5.02	0.17	4.50-5.50	3.40	15	3.48	0.16	3.00-4.00	4.57
PG: Abaxis Piccolo (Waived)	62	3.44	0.14	2.90-3.90	4.12																				
PG: Abbott Architect	7	2.87	0.05	2.40-3.40	1.72	7	5.49	0.04	5.00-6.00	0.73	7	7.47	0.05	7.00-8.00	0.67	7	5.00	0.00	4.50-5.50	0.00	7	3.57	0.05	3.10-4.10	1.39
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	32	2.93	0.10	2.40-3.40	3.45	32	5.96	0.10	5.50-6.50	1.67	31	8.72	0.23	8.20-9.20	2.64	31	5.44	0.08	4.90-5.90	1.48	31	3.63	0.07	3.10-4.10	1.94
PG: Clinical Data Envoy 500	9	2.62	0.16	2.10-3.10	6.15	9	5.39	0.18	4.90-5.90	3.33	9	7.21	0.22	6.70-7.70	3.06	9	4.79	0.21	4.30-5.30	4.38	9	3.49	0.13	3.00-4.00	3.71
PG: Siemens Dimension/AR/ES/RxL/Xpand	19	2.78	0.05	2.30-3.30	1.79	19	5.47	0.07	5.00-6.00	1.27	19	7.43	0.09	6.90-7.90	1.22	19	4.96	0.08	4.50-5.50	1.60	19	3.50	0.05	3.00-4.00	1.43
PG: Horiba ABX Pentra 400	6	2.85	0.08	2.40-3.40	2.76	6	5.77	0.10	5.30-6.30	1.72	6	8.30	0.22	7.80-8.80	2.65	6	5.20	0.09	4.70-5.70	1.73	6	3.62	0.07	3.10-4.10	2.22
PG: Beckman (Olympus) AU	17	2.83	0.06	2.30-3.30	2.14	17	5.32	0.07	4.80-5.80	1.32	17	7.18	0.13	6.70-7.70	1.81	17	4.88	0.04	4.40-5.40	0.82	17	3.49	0.03	3.00-4.00	0.86
PG: Roche Cobas Integra / Roche Reagent	8	2.83	0.09	2.30-3.30	3.21	8	5.51	0.06	5.00-6.00	1.09	8	7.53	0.09	7.00-8.00	1.20	8	5.03	0.07	4.50-5.50	1.40	8	3.56	0.05	3.10-4.10	1.39
PG: OCD Vitros 250,350,400,500,750,950	11	2.93	0.05	2.40-3.40	1.72	11	6.54	0.08	6.00-7.00	1.23	11	9.96	0.14	9.50-10.50	1.40	11	5.81	0.08	5.30-6.30	1.38	11	3.78	0.06	3.30-4.30	1.58
SG: Select Alfa Wasserman Instruments / Alfa Wasserman Reagent Groups	33	2.93	0.10	2.40-3.40	3.45	33	5.96	0.10	5.50-6.50	1.67	32	8.72	0.23	8.20-9.20	2.64	32	5.43	0.08	4.90-5.90	1.48	32	3.63	0.07	3.10-4.10	1.94
SG: Select Beckman Coulter Instruments (Chem)	18	2.81	0.07	2.30-3.30	2.50	18	5.33	0.07	4.80-5.80	1.32	18	7.20	0.15	6.70-7.70	2.08	18	4.88	0.04	4.40-5.40	0.82	18	3.49	0.04	3.00-4.00	1.14
SG: Select Roche Instruments / Roche Reagent	9	2.83	0.09	2.30-3.30	3.21	9	5.52	0.07	5.00-6.00	1.27	9	7.53	0.09	7.00-8.00	1.20	9	5.03	0.07	4.50-5.50	1.40	9	3.56	0.05	3.10-4.10	1.39
SG: Select Roche Instruments / Roche Reagents Groups	10	2.85	0.10	2.40-3.40	3.45	10	5.53	0.07	5.00-6.00	1.27	10	7.55	0.10	7.00-8.00	1.33	10	5.03	0.07	4.50-5.50	1.40	10	3.57	0.05	3.10-4.10	1.39
SG: Select OCD Vitros Instruments	12	2.93	0.05	2.40-3.40	1.72	12	6.53	0.09	6.00-7.00	1.38	12	9.96	0.14	9.50-10.50	1.40	12	5.81	0.08	5.30-6.30	1.38	12	3.77	0.06	3.30-4.30	1.58
SG: Abaxis Piccolo Instruments	77	3.42	0.14	2.90-3.90	4.12	17	5.52	0.17	5.00-6.00	3.09	17	7.46	0.24	7.00-8.00	3.20	17	5.02	0.16	4.50-5.50	3.20	16	3.48	0.15	3.00-4.00	4.29
SG: Select Abbott Architect Instruments	9	2.88	0.04	2.40-3.40	1.38	9	5.49	0.03	5.00-6.00	0.55	9	7.46	0.07	7.00-8.00	0.93	9	4.99	0.03	4.50-5.50	0.60	9	3.58	0.04	3.10-4.10	1.11
SG: International OCD Vitros Instruments	14	2.93	0.05	2.40-3.40	1.72	14	6.54	0.09	6.00-7.00	1.38	14	9.95	0.14	9.50-10.50	1.40	14	5.82	0.07	5.30-6.30	1.21	14	3.78	0.06	3.30-4.30	1.58
SG: International Abaxis Piccolo Instruments	77	3.42	0.14	2.90-3.90	4.12	17	5.52	0.17	5.00-6.00	3.09	17	7.46	0.24	7.00-8.00	3.20	17	5.02	0.16	4.50-5.50	3.20	16	3.48	0.15	3.00-4.00	4.29
SG: Select Gen Chem Instruments	43	2.79	0.13	2.30-3.30	4.64	43	5.44	0.18	4.90-5.90	3.33	40	7.45	0.41	6.90-7.90	5.54	43	4.93	0.16	4.40-5.40	3.27	43	3.52	0.08	3.00-4.00	2.29
AG: All Group	190	3.06	0.29	2.60-3.60	9.68	125	5.62	0.28	5.10-6.10	5.00						127	5.15	0.28	4.70-5.70	5.38	132	3.57	0.12	3.10-4.10	3.33
Chloride (CL-)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	16	91.28	2.14	86.00-96.00	2.31	16	105.83	2.23	101.00-111.00	2.08	16	115.44	1.76	109.00-121.00	1.57	16	103.67	2.38	99.00-109.00	2.31	16	96.50	2.41	92.00-102.00	2.47
PG: Abaxis Piccolo (Waived)	62	91.98	2.00	87.00-97.00	2.17																				
PG: Abbott Architect	7	90.43	1.27	86.00-95.00	1.44	7	105.86	1.07	101.00-111.00	1.04	7	116.29	1.25	110.00-122.00	1.12	7	103.71	1.60	99.00-109.00	1.54	7	95.86	1.57	91.00-101.00	1.67
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	31	88.34	1.97	84.00-92.00	2.27	31	116.66	1.29	111.00-123.00	1.11	30	140.83	2.15	134.00-148.00	1.49	31	113.36	5.12	107.00-119.00	4.51	31	97.99	2.93	93.00-103.00	2.96
PG: Clinical Data Envoy 500	9	88.22	2.68	84.00-92.00	3																				



PG: Roche Cobas Integra / Roche Reagent	8	16.41	1.59	11.00-21.00	10.00	8	25.50	1.60	20.00-30.00	6.40	8	32.33	1.84	27.00-37.00	5.63	8	23.67	2.12	18.00-30.00	8.75	8	17.59	2.19	11.00-25.00	12.22
PG: OCD Vitros 250,350,400,500,750,950	11	17.79	1.63	13.00-23.00	8.89	11	24.57	1.95	19.00-31.00	7.60	11	32.21	1.58	27.00-37.00	5.00	11	22.29	2.49	15.00-30.00	11.36	11	15.79	2.36	9.00-23.00	15.00
SG: Select Beckman Coulter Instruments (Chem)	17	18.05	1.73	13.00-23.00	9.44	17	27.50	2.87	19.00-37.00	10.36	17	36.86	2.17	30.00-44.00	5.95	17	25.57	2.38	19.00-33.00	9.23	17	18.60	2.35	12.00-26.00	12.11
SG: Select Roche Instruments / Roche Reagent	9	16.03	1.87	10.00-22.00	11.88	9	24.78	2.63	17.00-33.00	10.40	9	31.63	2.72	24.00-40.00	8.44	9	23.26	2.33	16.00-30.00	10.00	9	17.30	2.23	10.00-24.00	12.94
SG: Select Roche Instruments / Roche Reagents Groups	10	15.43	2.60	7.00-23.00	17.33	10	23.80	3.97	12.00-36.00	16.67	10	31.77	2.60	24.00-40.00	8.13	10	22.14	4.18	9.00-35.00	19.09	10	16.47	3.36	6.00-26.00	21.25
SG: Select OCD Vitros Instruments	12	17.73	1.58	13.00-23.00	8.89	12	24.60	1.88	19.00-31.00	7.60	12	32.40	1.68	27.00-37.00	5.31	12	22.33	2.41	15.00-29.00	10.91	12	15.60	2.38	9.00-23.00	15.00
SG: Abaxis Piccolo Instruments	78	17.42	1.23	13.00-21.00	7.06	16	26.83	0.71	25.00-29.00	2.59	16	32.61	0.92	30.00-36.00	2.73	16	24.94	1.00	22.00-28.00	4.00	16	17.33	0.77	15.00-19.00	4.71
SG: Select Abbott Architect Instruments	9	16.44	0.88	13.00-19.00	5.63	9	26.33	1.12	23.00-29.00	4.23	9	35.11	1.76	30.00-40.00	5.14	9	23.78	1.72	19.00-29.00	7.08	9	17.56	1.24	14.00-22.00	6.67
SG: International OCD Vitros Instruments	13	17.69	1.54	14.00-23.00	8.33	13	24.44	1.86	18.00-30.00	7.92	13	32.06	1.69	27.00-37.00	5.31	13	21.88	2.60	14.00-30.00	11.82	13	15.50	2.34	9.00-23.00	14.38
AG: All Group	190	17.42	1.99	11.00-23.00	11.76	128	26.71	2.94	18.00-36.00	10.74	128	34.17	3.66	23.00-45.00	10.88	128	24.80	3.12	16.00-34.00	12.40	127	18.01	2.64	10.00-26.00	14.44
Calcium	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Abaxis Piccolo (Non-Waived)	15	5.89	0.20	4.90-6.90	3.39	15	10.59	0.17	9.60-11.60	1.60	15	12.46	0.21	11.50-13.50	1.68	15	10.14	0.22	9.10-11.10	2.18	15	8.42	0.22	7.40-9.40	2.62
PG: Abaxis Piccolo (Waived)	63	5.77	0.20	4.80-6.80	3.45																				
PG: Abbott Architect	7	5.74	0.13	4.70-6.70	2.28	7	10.93	0.18	9.90-11.90	1.65	7	13.47	0.21	12.50-14.50	1.56	7	10.29	0.26	9.30-11.30	2.52	7	8.33	0.15	7.30-9.30	1.81
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	31	5.79	0.18	4.80-6.80	3.10	32	10.83	0.24	9.80-11.80	2.22	32	12.89	0.67	11.90-13.90	5.19	32	10.47	0.62	9.50-11.50	5.90	32	8.73	0.44	7.70-9.70	5.06
PG: Clinical Data Envoy 500	9	5.61	0.11	4.60-6.60	1.96	9	10.59	0.14	9.60-11.60	1.32	9	13.02	0.20	12.00-14.00	1.54	9	10.02	0.18	9.00-11.00	1.80	9	8.20	0.14	7.20-9.20	1.71
PG: Siemens Dimension/AR/ES/RxL/Xpand	17	5.29	0.13	4.30-6.30	2.45	17	10.48	0.20	9.50-11.50	1.90	17	12.92	0.19	11.90-13.90	1.47	17	9.87	0.17	8.90-10.90	1.72	17	8.13	0.15	7.10-9.10	1.85
PG: Horiba ABX Pentra 400	6	5.47	0.17	4.50-6.50	3.27	6	10.87	0.22	9.90-11.90	2.02	6	13.27	0.31	12.30-14.30	2.33	6	10.25	0.21	9.30-11.30	2.04	6	8.60	0.70	7.60-9.60	8.14
PG: Beckman (Olympus) AU	16	5.87	0.12	4.90-6.90	2.03	17	11.05	0.34	10.00-12.00	3.09	17	13.60	0.37	12.60-14.60	2.72	17	10.46	0.36	9.50-11.50	3.43	17	8.62	0.58	7.60-9.60	6.74
PG: Roche Cobas Integra / Roche Reagent	8	5.52	0.08	4.50-6.50	1.45	8	10.81	0.13	9.80-11.80	1.20	8	13.22	0.28	12.20-14.20	2.12	8	10.46	0.66	9.50-11.50	6.29	8	8.39	0.16	7.40-9.40	1.90
PG: OCD Vitros 250,350,400,500,750,950	11	5.31	0.13	4.30-6.30	2.45	11	10.89	0.22	9.90-11.90	2.02	11	13.43	0.22	12.40-14.40	1.64	11	10.20	0.16	9.20-11.20	1.57	11	8.33	0.18	7.30-9.30	2.17
SG: Select Beckman Coulter Instruments (Chem)	17	5.86	0.12	4.90-6.90	2.03	18	10.99	0.35	10.00-12.00	3.18	18	13.50	0.42	12.50-14.50	3.11	18	10.41	0.37	9.40-11.40	3.56	18	8.60	0.57	7.60-9.60	6.63
SG: Select Roche Instruments / Roche Reagent	9	5.54	0.10	4.50-6.50	1.82	9	10.80	0.12	9.80-11.80	1.11	9	13.24	0.27	12.20-14.20	2.05	9	10.44	0.62	9.40-11.40	5.96	9	8.40	0.16	7.40-9.40	1.90
SG: Select Roche Instruments / Roche Reagents Groups	10	5.59	0.17	4.60-6.60	3.04	10	10.79	0.12	9.80-11.80	1.11	10	13.10	0.49	12.10-14.10	3.74	10	10.39	0.61	9.40-11.40	5.87	10	8.40	0.15	7.40-9.40	1.79
SG: Select OCD Vitros Instruments	12	5.29	0.14	4.30-6.30	2.64	12	10.85	0.26	9.90-11.90	2.39	12	13.37	0.32	12.40-14.40	2.39	12	10.17	0.20	9.20-11.20	1.96	12	8.31	0.18	7.30-9.30	2.17
SG: Abaxis Piccolo Instruments	78	5.80	0.20	4.80-6.80	3.45	16	10.59	0.17	9.60-11.60	1.60	16	12.48	0.21	11.50-13.50	1.68	16	10.13	0.22	9.10-11.10	2.18	16	8.41	0.21	7.40-9.40	2.50
SG: Select Abbott Architect Instruments	9	5.73	0.11	4.70-6.70	1.93	9	10.90	0.17	9.90-11.90	1.56	9	13.40	0.23	12.40-14.40	1.72	9	10.24	0.25	9.20-11.20	2.45	9	8.31	0.14	7.30-9.30	1.69
SG: International OCD Vitros Instruments	14	5.30	0.13	4.30-6.30	2.45	14	10.87	0.23	9.90-11.90	2.11	14	13.38	0.24	12.40-14.40	1.79	14	10.24	0.17	9.20-11.20	1.67	14	8.35	0.17	7.30-9.30	2.05
SG: Select Gen Chem Instruments	41	5.74	0.19	4.70-6.70	3.33	42	10.88	0.31	9.90-11.90	2.84	42	13.36	0.38	12.40-14.40	2.84	42	10.28	0.33	9.30-11.30	3.20	42	8.46	0.49	7.50-9.50	5.76
AG: All Group	189	5.69	0.26	4.70-6.70	4.56	129	10.77	0.28	9.80-11.80	2.69	129	13.06	0.53	12.10-14.10	4.05	129	10.25	0.45	9.30-11.30	4.37	129	8.46	0.41	7.50-9.50	4.82
Creatinine	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Abaxis Piccolo (Non-Waived)	14	0.57	0.09	0.27-0.87	15.44	16	3.44	0.14	2.92-3.96	4.16	16	5.30	0.14	4.50-6.10	2.72	16	2.90	0.10	2.46-3.34	3.55	16	1.51	0.10	1.21-1.81	6.67
PG: Abaxis Piccolo (Waived)	59	0.62	0.11	0.32-0.92	18.33																				
PG: Abbott Architect	7	0.68	0.01	0.38-0.98	1.62	7	3.54	0.10	3.01-4.07	2.86	7	5.69	0.15	4.80-6.60	2.67	7	3.00	0.07	2.55-3.45	2.37	7	1.39	0.03	1.09-1.69	2.23
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	33	0.69	0.05	0.40-1.00	7.54	33	3.05	0.14	2.59-3.51	4.67	32	4.33	0.20	3.70-4.90	4.65	32	2.69	0.10	2.29-3.09	3.72	32	1.52	0.07	1.22-1.82	4.67
SG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent Groups	33	0.69	0.05	0.40-1.00	7.54	33	3.05	0.14	2.59-3.51	4.67	32	4.33	0.20	3.70-4.90	4.65	32	2.69	0.10	2.29-3.09	3.72	32	1.52	0.07	1.22-1.82	4.67
PG: Clinical Data Envoy 500	9	0.50	0.04	0.20-0.80	8.40	9	3.15	0.09	2.68-3.62	2.90	9	4.78	0.13	4.10-5.50	2.71	9	2.75	0.09	2.34-3.16	3.33	9	1.49	0.07	1.19-1.79	4.70
PG: Siemens Dimension/AR/ES/RxL/Xpand	12	0.54	0.04	0.24-0.84	8.00	13	3.33	0.12	2.80-3.80	3.66	14	5.15	0.21	4.30-5.90	4.12	14	2.85	0.14	2.42-3.28	4.91	14	1.47	0.11	1.17-1.77	7.33
PG: Horiba ABX Pentra 400	5	0.49	0.02	0.19-0.79	4.00	6	3.15	0.11	2.60-3.60	3.55	6	4.77	0.14	4.10-5.50	2.92	6	2.67	0.08	2.27-3.07	2.96	6	1.38	0.07	1.08-1.68	5.71
PG: Beckman (Olympus) AU	16	0.53	0.04	0.20-0.80	10.00	17	3.22	0.13	2.70-3.70	4.07	17	4.85	0.30	4.10-5.50	6.25	17	2.74	0.23	2.33-3.15	8.21	17	1.40	0.12	1.10-1.70	8.86
PG: Roche Cobas Integra / Roche Reagent	8	0.57	0.04	0.30-0.90	7.72	8	3.03	0.11	2.58-3.48	3.70	8	4.43	0.16	3.70-5.10	3.66	8	2.62	0.12	2.23-3.01	4.62	8	1.39	0.06	1.09-1.69	4.29
PG: OCD Vitros 250,350,400,500,750,950	10	0.49	0.03	0.19-0.79	6.00	11	2.66	0.07	2.26-3.06	2.78	11	3.77	0.11	3.20-4.40	2.92	11	2.30	0.07	1.95-2.64						



PG: Intl Siemens RAPIDPoint 400/405 System	5	44.12	1.24	38.00-50.00	2.73	5	137.98	1.45	124.00-152.00	1.01	5	191.91	1.28	173.00-211.00	0.68	5	122.27	0.38	110.00-134.00	0.33	5	77.74	0.36	70.00-86.00	0.51
SG: Select SpotChem Instruments	12	48.27	2.05	42.00-54.00	4.38	12	152.13	9.72	137.00-167.00	6.38	10	232.77	3.35	210.00-256.00	1.42	12	134.40	9.19	121.00-147.00	6.87	12	83.07	4.51	75.00-91.00	5.42
SG: Select Abbott Architect Instruments	9	47.00	1.12	41.00-53.00	2.34	9	155.67	2.50	140.00-172.00	1.60	9	227.89	3.52	205.00-251.00	1.54	9	137.67	2.18	124.00-152.00	1.59	9	86.56	1.94	78.00-96.00	2.18
SG: International OCD Vitros Instruments	14	48.76	1.86	43.00-55.00	3.88	14	152.12	8.87	137.00-167.00	5.86	14	230.06	12.22	207.00-253.00	5.30	14	133.88	7.71	121.00-147.00	5.75	14	83.06	4.05	75.00-91.00	4.94
SG: Select Gen Chem Instruments	41	49.83	2.39	44.00-56.00	4.80	42	162.66	6.39	147.00-179.00	3.93	42	234.83	8.44	212.00-259.00	3.57	42	144.46	6.13	130.00-158.00	4.24	42	89.93	4.07	81.00-99.00	4.56
AG: All Group	193	51.52	2.55	46.00-58.00	4.81	133	163.31	6.93	147.00-179.00	4.23	130	235.18	8.56	212.00-259.00	3.66	131	144.89	6.30	131.00-160.00	4.34	131	90.78	4.58	82.00-100.00	5.05
Urea Nitrogen (BUN)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	14	6.61	0.55	5.00-9.00	7.14	16	26.95	0.81	25.00-29.00	2.96	16	41.00	0.97	37.00-45.00	2.44	16	23.58	0.98	22.00-26.00	4.17	16	14.00	0.59	12.00-16.00	4.29
PG: Abaxis Piccolo (Waived)	61	6.49	0.59	4.00-8.00	10.00																				
PG: Abbott Architect	7	6.43	0.54	4.00-8.00	8.33	7	30.29	0.76	27.00-33.00	2.67	7	45.86	1.35	42.00-50.00	2.83	7	25.86	0.38	24.00-28.00	1.54	7	14.71	0.49	13.00-17.00	3.33
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	32	5.78	0.42	4.00-8.00	6.67	32	29.32	1.53	26.00-32.00	5.17	30	44.41	2.35	40.00-48.00	5.23	31	25.86	1.29	24.00-28.00	5.00	31	14.77	0.84	13.00-17.00	5.33
PG: Clinical Data Envoy 500	8	5.75	0.71	4.00-8.00	11.67	8	29.25	0.71	26.00-32.00	2.41	8	43.88	1.13	40.00-48.00	2.50	8	26.13	0.35	24.00-28.00	1.54	8	15.13	0.64	13.00-17.00	4.00
PG: Siemens Dimension/AR/ES/RxL/Xpand	19	5.90	0.54	4.00-8.00	8.33	19	30.29	1.42	27.00-33.00	4.67	19	45.90	2.14	42.00-50.00	4.57	19	26.38	1.28	24.00-28.00	5.00	19	14.90	0.89	13.00-17.00	6.00
PG: Horiba ABX Pentra 400	5	5.58	0.53	4.00-8.00	8.33	5	29.56	0.61	27.00-33.00	2.00	5	44.26	1.24	40.00-48.00	2.73	5	25.00	0.71	23.00-27.00	2.80	5	14.22	0.44	12.00-16.00	2.86
PG: Beckman (Olympus) AU	16	6.05	0.62	4.00-8.00	10.00	17	30.15	1.63	27.00-33.00	5.33	17	45.25	2.57	41.00-49.00	5.78	17	26.25	1.77	24.00-28.00	6.92	17	15.10	1.07	13.00-17.00	7.33
PG: Roche Cobas Integra / Roche Reagent	8	5.46	0.50	3.00-7.00	10.00	8	29.20	0.85	26.00-32.00	2.76	8	43.49	1.69	39.00-47.00	3.95	8	24.87	0.83	23.00-27.00	3.20	8	14.02	0.54	12.00-16.00	3.57
PG: OCD Vitros 250,350,400,500,750,950	11	4.86	0.36	3.00-7.00	8.00	11	25.93	0.47	24.00-28.00	1.92	11	41.93	0.92	38.00-46.00	2.14	11	22.36	0.63	20.00-24.00	2.73	11	12.43	0.51	10.00-14.00	4.17
SG: Select Beckman Coulter Instruments (Chem)	17	6.18	0.66	4.00-8.00	11.67	18	30.26	1.54	27.00-33.00	5.00	18	45.48	2.47	41.00-49.00	5.56	18	26.32	1.70	24.00-28.00	6.54	18	15.10	1.04	13.00-17.00	6.67
SG: Select Roche Instruments / Roche Reagent	9	5.41	0.50	3.00-7.00	10.00	9	29.28	0.84	26.00-32.00	2.76	9	43.66	1.66	40.00-48.00	3.86	9	24.89	0.78	23.00-27.00	3.20	9	14.13	0.60	12.00-16.00	4.29
SG: Select Roche Instruments / Roche Reagents Groups	10	5.47	0.50	3.00-7.00	10.00	10	29.16	0.89	26.00-32.00	3.10	10	43.29	1.94	39.00-47.00	4.42	10	24.80	0.79	23.00-27.00	3.20	10	14.22	0.63	12.00-16.00	4.29
SG: Select OCD Vitros Instruments	12	4.80	0.41	3.00-7.00	8.00	12	25.80	0.68	24.00-28.00	2.69	12	41.73	1.16	38.00-46.00	2.86	12	22.27	0.70	20.00-24.00	3.18	12	12.33	0.62	10.00-14.00	5.00
SG: Abaxis Piccolo Instruments	75	6.51	0.58	5.00-9.00	8.57	17	26.95	0.78	25.00-29.00	2.96	17	41.00	0.95	37.00-45.00	2.20	17	23.55	0.96	22.00-26.00	4.17	17	14.00	0.58	12.00-16.00	4.29
PG: Arkray Spotchem EZ (Waived)	5	6.60	0.00	5.00-9.00	0.00	5	18.20	0.00	16.00-20.00	0.00	5	9.90	0.00	8.00-12.00	0.00	5	18.20	0.00	16.00-20.00	0.00	5	12.40	0.00	10.00-14.00	0.00
SG: Select Abbott Architect Instruments	9	6.33	0.50	4.00-8.00	8.33	9	30.33	0.71	27.00-33.00	2.33	9	45.78	1.39	42.00-50.00	3.04	9	25.89	0.33	24.00-28.00	1.15	9	14.67	0.50	13.00-17.00	3.33
SG: International OCD Vitros Instruments	14	4.82	0.39	3.00-7.00	8.00	14	25.88	0.48	24.00-28.00	1.92	14	41.76	0.97	38.00-46.00	2.38	14	22.35	0.61	20.00-24.00	2.73	14	12.41	0.51	10.00-14.00	4.17
SG: Select Gen Chem Instruments	39	6.07	0.66	4.00-8.00	11.67	40	30.02	1.26	27.00-33.00	4.33	40	45.12	2.06	41.00-49.00	4.67	40	26.05	1.29	24.00-28.00	5.00	40	14.91	0.86	13.00-17.00	6.00
AG: All Group	188	6.04	0.73	4.00-8.00	11.67	131	28.92	2.11	26.00-32.00	7.24	129	43.81	3.80	40.00-48.00	8.64	130	25.19	1.86	23.00-27.00	7.60	130	14.42	1.12	12.00-16.00	7.86
Albumin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived) / BCP	14	4.54	0.10	4.00-5.00	2.22	14	3.44	0.14	3.10-3.70	4.12	14	4.11	0.13	3.70-4.50	3.17	14	3.24	0.09	2.90-3.50	2.81	14	2.69	0.09	2.40-3.00	3.33
PG: Abaxis Piccolo (Waived) / BCP	64	4.53	0.14	4.00-5.00	3.11																				
PG: Abbott Architect / BCP	7	4.40	0.14	4.00-4.80	3.18	7	3.50	0.10	3.10-3.90	2.86	7	4.30	0.14	3.90-4.70	3.26	7	3.29	0.11	3.00-3.60	3.33	7	2.64	0.05	2.30-2.90	1.92
PG: Alfa Was. Ace/NexCT/Alera/Axcel / BCG / Alfa Wasserman Reagent	33	4.26	0.10	3.90-4.70	2.33	33	4.21	0.17	3.80-4.60	4.05	33	5.33	0.42	4.80-5.80	7.92	33	3.94	0.28	3.50-4.30	7.44	33	2.97	0.18	2.70-3.30	6.00
PG: Clinical Data Envoy 500 / BCG	9	4.24	0.15	3.80-4.60	3.57	9	4.24	0.11	3.80-4.60	2.62	9	5.29	0.20	4.80-5.80	3.77	9	3.92	0.12	3.50-4.30	3.08	9	2.97	0.10	2.70-3.30	3.33
PG: Siemens Dimension/AR/ES/RxL/Xpand / BCG	11	4.52	0.16	4.00-5.00	3.56	11	3.57	0.13	3.20-4.00	3.61	11	4.47	0.17	4.00-5.00	3.78	11	3.33	0.10	3.00-3.60	3.03	11	2.59	0.12	2.30-2.90	4.62
PG: Siemens Dimension/AR/ES/RxL/Xpand / BCP	8	4.46	0.07	4.00-5.00	1.56	8	3.51	0.06	3.10-3.90	1.71	8	4.40	0.05	4.00-4.80	1.14	8	3.30	0.05	3.00-3.60	1.52	8	2.59	0.04	2.30-2.90	1.54
PG: Horiba ABX Pentra 400 / BCG	6	4.27	0.17	3.90-4.70	4.19	6	4.17	0.16	3.80-4.60	3.81	5	5.22	0.13	4.70-5.70	2.50	6	3.88	0.12	3.50-4.30	3.08	6	2.97	0.10	2.70-3.30	3.33
PG: Beckman (Olympus) AU / BCG	17	4.40	0.12	4.00-4.80	2.73	17	4.36	0.24	4.00-4.80	5.45	17	5.55	0.19	5.00-6.20	3.39	17	3.97	0.08	3.60-4.40	2.00	17	2.89	0.06	2.60-3.20	2.07
PG: Roche Cobas Integra / BCG / Roche Reagent	8	4.57	0.11	4.10-5.10	2.39	8	4.53	0.07	4.00-5.00	1.56	8	5.77	0.13	5.20-6.40	2.24	8	4.18	0.09	3.80-4.60	2.14	8	3.08	0.05	2.80-3.40	1.61
PG: OCD Vitros 250,350,400,500,750,950 / BCG	11	4.44	0.20	4.00-4.80	4.55	11	4.32	0.17	3.90-4.70	3.95	11	5.65	0.23	5.10-6.30	4.04	11	3.90	0.20	3.50-4.30	5.13	11	2.74	0.15	2.40-3.00	5.56
SG: Select Beckman Coulter Instruments (Chem) / BCG	17	4.40	0.12	4.00-4.80	2.73	17	4.36	0.24	4.00-4.80	5.45	17	5.55	0.19	5.00-6.20	3.39	17	3.97	0.08	3.60-4.40	2.00	17	2.89	0.06	2.60-3.20	2.07
SG: Select Roche Instruments / BCG / Roche Reagent	9	4.58	0.11	4.10-5.10	2.39	9	4.53	0.07	4.00-5.00	1.56	9	5.79	0.14	5.20-6.40	2.41	9	4.19	0.09	3.80-4.60	2.14	9	3.09	0.06	2.80-3.40	1.94
SG: Select Roche Instruments / BCG / Roche Reagents Groups	10	4.57	0.10	4.10-5.10	2.17	10	4.51	0.10	4.00-5																



SG: International Abaxis Piccolo Instruments	75	42.92	3.84	30.00-56.00	8.84	13	106.00	6.82	74.00-138.00	6.42	13	153.69	9.89	108.00-200.00	6.43	13	92.85	6.88	65.00-121.00	7.42	13	48.62	4.91	34.00-64.00	10.00
ALT/SGPT	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	11	23.00	3.44	18.00-28.00	14.78	12	155.33	6.02	124.00-186.00	3.87	12	240.83	7.17	193.00-289.00	2.99	12	131.42	4.48	105.00-157.00	3.44	12	68.75	3.42	55.00-83.00	4.93
PG: Abaxis Piccolo (Waived)	62	21.05	2.86	17.00-25.00	13.81																				
PG: Abbott Architect	7	13.43	1.62	10.00-16.00	12.31	7	179.29	3.04	143.00-215.00	1.68	7	289.43	3.41	231.00-347.00	1.18	7	150.57	3.15	121.00-181.00	2.12	7	72.29	2.06	58.00-86.00	2.92
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	25	7.64	1.34	6.00-10.00	16.25	31	140.42	12.24	112.00-168.00	8.71	31	223.60	30.60	179.00-269.00	13.66	30	121.21	3.88	97.00-145.00	3.22	30	55.38	2.61	44.00-66.00	4.73
SG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent Groups	31	7.56	2.04	6.00-10.00	25.00	32	140.97	12.54	113.00-169.00	8.87	32	224.22	30.39	179.00-269.00	13.57	31	121.57	4.39	98.00-146.00	3.61	31	55.71	3.23	45.00-67.00	5.71
PG: Clinical Data Envoy 500	8	16.38	0.52	13.00-19.00	3.13	8	174.75	6.88	140.00-210.00	3.94	8	278.50	18.82	223.00-335.00	6.74	8	149.13	5.64	119.00-179.00	3.76	8	74.00	3.34	59.00-89.00	4.46
PG: Siemens Dimension/AR/ES/RxL/Xpand	8	21.56	2.19	18.00-26.00	10.00	8	183.56	3.28	147.00-221.00	1.79	8	287.78	6.22	230.00-346.00	2.15	8	154.44	3.17	123.00-185.00	2.08	8	76.56	1.59	62.00-92.00	2.08
PG: Siemens Dimension/AR/ES/RxL/Xpand / IFCC	10	21.27	2.05	17.00-25.00	10.00	11	184.00	13.65	147.00-221.00	7.39	11	286.33	5.96	229.00-343.00	2.10	11	153.42	4.54	122.00-184.00	2.94	11	75.92	2.31	61.00-91.00	3.03
PG: Beckman (Olympus) AU	17	11.80	0.77	10.00-14.00	6.67	17	155.55	5.19	125.00-187.00	3.33	17	250.25	8.17	200.00-300.00	3.28	17	130.45	4.65	104.00-156.00	3.62	17	62.40	2.09	50.00-74.00	3.39
PG: Roche Cobas Integra / Roche Reagent	8	13.64	0.50	11.00-17.00	3.57	8	168.94	1.48	135.00-203.00	0.89	8	268.79	1.82	215.00-323.00	0.67	8	142.65	0.93	114.00-172.00	0.63	8	69.11	0.36	55.00-83.00	0.58
SG: Roche Cobas Integra / Roche Reagents Groups	8	13.64	0.50	11.00-17.00	3.57																				
PG: OCD Vitros 250,350,400,500,750,950	11	28.29	3.67	22.00-34.00	13.21	11	168.00	5.02	134.00-202.00	2.98	11	257.71	7.50	206.00-310.00	2.91	11	146.14	4.05	117.00-175.00	2.81	11	83.50	3.25	67.00-101.00	3.93
SG: Select Beckman Coulter Instruments (Chem)	17	11.80	0.77	10.00-14.00	6.67	18	157.96	7.97	126.00-190.00	5.06	18	253.48	11.41	202.00-304.00	4.51	18	132.05	6.80	106.00-158.00	5.15	18	63.05	3.60	50.00-76.00	5.71
SG: Select Roche Instruments / Roche Reagent	9	13.68	0.48	11.00-17.00	3.57	9	168.95	1.39	135.00-203.00	0.83	9	270.82	6.30	217.00-325.00	2.32	9	142.47	1.03	114.00-170.00	0.70	9	68.99	0.50	55.00-83.00	0.72
SG: Select Roche Instruments / Roche Reagents Groups	10	14.01	1.15	11.00-17.00	7.86																				
SG: Select OCD Vitros Instruments	12	28.07	3.63	22.00-34.00	12.86	12	168.20	4.90	134.00-202.00	2.92	12	257.60	7.24	206.00-310.00	2.79	12	146.20	3.91	117.00-175.00	2.67	12	83.33	3.20	66.00-100.00	3.86
SG: Abaxis Piccolo Instruments	73	21.34	3.01	17.00-25.00	14.29	13	155.23	5.78	124.00-186.00	3.74	13	240.77	6.87	193.00-289.00	2.86	13	131.23	4.34	105.00-157.00	3.28	13	68.62	3.31	55.00-83.00	4.78
SG: Select SpotChem Instruments	11	28.29	3.67	22.00-34.00	13.21	12	168.27	4.96	134.00-202.00	2.98	12	256.91	7.86	206.00-308.00	3.07	12	146.11	3.91	117.00-175.00	2.67	12	81.97	6.69	66.00-98.00	8.17
SG: Select Abbott Architect Instruments	9	13.33	1.41	10.00-16.00	10.77	9	178.22	3.60	142.00-214.00	2.02	9	287.22	5.95	230.00-344.00	2.09	9	149.56	3.71	120.00-180.00	2.47	9	71.78	2.11	58.00-86.00	2.92
SG: International OCD Vitros Instruments	14	28.18	3.50	22.00-34.00	12.50	14	168.71	5.03	135.00-203.00	2.96	14	257.47	6.87	206.00-308.00	2.68	14	146.71	4.00	118.00-176.00	2.72	14	84.29	3.88	67.00-101.00	4.64
SG: International Abaxis Piccolo Instruments	74	21.22	3.18	17.00-25.00	15.24																				
SG: Select Gen Chem Instruments	36	13.38	2.21	10.00-16.00	16.92	39	167.93	12.94	134.00-202.00	7.68	39	269.07	20.86	215.00-323.00	7.77	39	141.58	11.87	114.00-170.00	8.38	39	68.43	6.44	54.00-82.00	9.41
MG: IFCC	12	21.00	2.60	17.00-25.00	12.38	13	185.00	12.80	148.00-222.00	6.92	13	289.00	8.70	231.00-347.00	3.01	13	154.00	4.50	123.00-185.00	2.92	13	76.00	2.20	61.00-91.00	2.89
AST/SGOT	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	12	28.42	2.19	22.00-34.00	7.86	12	220.33	7.52	176.00-264.00	3.41	12	338.42	11.64	270.00-406.00	3.43	12	188.33	6.46	150.00-226.00	3.46	11	97.36	2.91	78.00-116.00	2.99
PG: Abaxis Piccolo (Waived)	63	28.21	1.81	22.00-34.00	6.43																				
PG: Abbott Architect	7	22.43	0.98	18.00-26.00	4.55	7	220.43	1.72	176.00-264.00	0.77	7	344.71	2.43	276.00-414.00	0.70	7	187.86	1.22	150.00-226.00	0.64	7	94.43	1.62	75.00-113.00	1.70
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	28	16.30	2.02	13.00-19.00	12.50	32	198.24	6.39	158.00-238.00	3.23	32	299.78	38.56	240.00-360.00	12.87	32	174.47	25.48	139.00-209.00	14.66	31	83.29	4.41	66.00-100.00	5.30
PG: Clinical Data Envoy 500	8	24.25	2.05	19.00-29.00	8.75	8	224.00	6.02	179.00-269.00	2.68	8	345.13	10.29	276.00-414.00	2.99	8	191.13	6.27	153.00-229.00	3.30	8	98.38	4.07	78.00-118.00	4.18
PG: Siemens Dimension/AR/ES/RxL/Xpand	19	21.00	2.26	17.00-25.00	10.95	19	223.10	8.48	178.00-268.00	3.81	19	349.48	12.75	279.00-419.00	3.67	19	190.10	7.25	152.00-228.00	3.79	19	93.10	3.94	74.00-112.00	4.19
PG: Beckman (Olympus) AU	17	21.10	0.85	17.00-25.00	4.29	17	195.95	8.71	157.00-235.00	4.44	17	304.40	13.47	243.00-365.00	4.44	17	167.15	7.13	134.00-200.00	4.25	17	84.50	3.53	68.00-102.00	4.12
PG: Roche Cobas Integra / Roche Reagent	8	22.16	0.35	18.00-26.00	1.82	8	219.99	4.43	176.00-264.00	2.00	8	326.95	37.53	262.00-392.00	11.47	8	187.94	3.47	150.00-226.00	1.86	8	95.43	1.54	76.00-114.00	1.58
PG: OCD Vitros 250,350,400,500,750,950	11	23.29	1.54	18.00-28.00	6.52	11	235.00	11.16	188.00-282.00	4.77	11	407.79	21.84	326.00-490.00	5.34	11	196.71	8.60	158.00-236.00	4.37	11	96.29	4.10	77.00-115.00	4.27
SG: Select Beckman Coulter Instruments (Chem)	17	21.10	0.85	17.00-25.00	4.29	18	198.57	10.64	159.00-239.00	5.33	18	308.13	15.93	246.00-370.00	5.16	18	168.77	8.58	135.00-203.00	5.09	18	85.05	4.26	68.00-102.00	5.06
SG: Select Roche Instruments / Roche Reagent	9	22.14	0.33	18.00-26.00	1.36	9	220.44	4.35	176.00-264.00	2.00	9	329.17	35.73	263.00-395.00	10.85	9	187.95	3.25	150.00-226.00	1.76	9	95.38	1.45	76.00-114.00	1.47
SG: Select OCD Vitros Instruments	12	23.20	1.52	18.00-28.00	6.52	12	234.60	10.86	188.00-282.00	4.64	12	406.20	21.93	325.00-487.00	5.39	12	195.60	9.34	157.00-235.00	4.74	12	95.93	4.18	77.00-115.00	4.38
SG: Abaxis Piccolo Instruments	75	28.24	1.86	22.00-34.00	6.79	13	220.46	7.22	176.00-264.00	3.27	13	338.38	11.15	270.00-406.00	3.28	13	188.15	6.22	150.00-226.00	3.30	12	97.42	2.78	78.00-116.00	2.89
SG: Select SpotChem Instruments	12	22.73	2.60	18.00-28.00	11.30	12	235.00	10.75	188.00-282.00	4.60	12	400.60	34.89	321.00-481.00	8.70	12	196.93	8.33	158.00-236.00	4.21	12	96.20	3.97	77.00-115.00	4.17
SG: Select Abbott Architect Instruments	9	22.22	0.97	18.00-26.00	4.55	9	219.44	3.88	175.00-263.00	1.78	9	341.89	9.60	274.00-410.00	2.81	9	186.44	4.07	149.00-223.00	2.20	9	93.67	2.55	75.00-113.00	2.66
SG: International OCD Vitros Instruments	14	23.24	1.44	18.00-28.00	6.09	14	235.18	10.07	188.00-282.00	4.30	14	405.53	21.48	325.00-487.00	5										



PG: Clinical Data Envoy 500	8	6.21	0.20	5.60-6.80	3.23	8	6.99	0.35	6.30-7.70	5.00	8	9.64	0.33	8.60-10.60	3.33	8	6.34	0.23	5.70-6.90	3.65	8	4.35	0.16	4.00-4.80	3.64
PG: Siemens Dimension/AR/ES/RxL/Xpand	19	6.62	0.18	5.90-7.30	2.73	19	7.28	0.21	6.60-8.00	2.88	19	10.00	0.28	9.00-11.00	2.80	19	6.60	0.18	5.90-7.30	2.73	19	4.57	0.14	4.10-5.10	3.04
PG: Horiba ABX Pentra 400	6	6.00	0.09	5.40-6.60	1.50	6	6.70	0.16	6.00-7.40	2.24	6	9.23	0.16	8.30-10.10	1.74	6	6.05	0.14	5.50-6.70	2.30	6	4.18	0.10	3.80-4.60	2.38
PG: Beckman (Olympus) AU	17	6.32	0.11	5.70-6.90	1.75	17	6.89	0.13	6.20-7.60	1.88	17	9.43	0.19	8.50-10.30	2.02	17	6.25	0.12	5.70-6.90	1.90	17	4.28	0.10	3.90-4.70	2.33
PG: Roche Cobas Integra / Roche Reagent	8	6.17	0.16	5.60-6.80	2.58	8	6.66	0.28	6.00-7.40	4.18	8	9.11	0.15	8.20-10.00	1.65	8	6.12	0.15	5.50-6.70	2.46	8	4.21	0.09	3.80-4.60	2.14
PG: OCD Vitros 250,350,400,500,750,950	11	6.06	0.20	5.50-6.70	3.11	11	6.67	0.23	6.00-7.40	3.43	11	9.15	0.42	8.30-10.10	4.57	11	6.10	0.20	5.50-6.70	3.28	11	4.26	0.19	3.90-4.70	4.42
SG: Select Beckman Coulter Instruments (Chem)	18	6.35	0.14	5.80-7.00	2.19	18	6.90	0.13	6.20-7.60	1.88	18	9.47	0.20	8.60-10.40	2.11	18	6.27	0.14	5.70-6.90	2.22	18	4.28	0.10	3.90-4.70	2.33
SG: Select Roche Instruments / Roche Reagent	9	6.19	0.16	5.60-6.80	2.58	9	6.68	0.26	6.00-7.40	3.88	9	9.16	0.19	8.30-10.10	2.07	9	6.14	0.15	5.50-6.70	2.46	9	4.21	0.09	3.80-4.60	2.14
SG: Select Roche Instruments / Roche Reagents Groups	10	6.19	0.15	5.60-6.80	2.42	10	6.71	0.27	6.00-7.40	4.03	10	9.19	0.21	8.30-10.10	2.28	10	6.15	0.15	5.60-6.80	2.42	10	4.22	0.09	3.80-4.60	2.14
SG: Select OCD Vitros Instruments	12	6.06	0.19	5.50-6.70	3.11	12	6.67	0.22	6.00-7.40	3.28	12	9.15	0.41	8.20-10.00	4.51	12	6.10	0.19	5.50-6.70	3.11	12	4.27	0.18	3.90-4.70	4.19
SG: Abaxis Piccolo Instruments	74	6.47	0.34	5.80-7.20	5.23	12	6.95	0.07	6.30-7.70	1.00	12	9.60	0.22	8.60-10.60	2.29	12	6.33	0.08	5.70-6.90	1.27	12	4.43	0.09	4.00-4.80	2.05
SG: Select Abbott Architect Instruments	9	6.34	0.15	5.70-6.90	2.38	9	7.07	0.17	6.40-7.80	2.39	9	9.87	0.23	8.90-10.90	2.32	9	6.32	0.15	5.70-6.90	2.38	9	4.22	0.14	3.80-4.60	3.33
SG: International OCD Vitros Instruments	14	6.08	0.19	5.50-6.70	3.11	14	6.68	0.21	6.00-7.40	3.13	14	9.14	0.39	8.20-10.00	4.29	14	6.12	0.18	5.50-6.70	2.95	14	4.28	0.17	3.90-4.70	3.95
AG: All Group	187	6.39	0.29	5.80-7.00	4.53	125	6.99	0.29	6.30-7.70	4.14	125	9.57	0.58	8.60-10.60	6.04	125	6.38	0.41	5.80-7.00	6.41	125	4.38	0.26	4.00-4.80	5.91
Phosphorous	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	5	2.13	0.12	1.70-2.50	5.71	5	4.51	0.15	4.00-5.00	3.33	5	6.03	0.15	5.50-6.50	2.50	5	4.10	0.13	3.70-4.50	3.17	5	3.04	0.21	2.40-3.60	7.00
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	9	1.69	0.13	1.30-2.10	7.65	9	3.86	0.17	3.40-4.40	4.36	9	5.14	0.31	4.20-6.00	6.08	9	3.51	0.16	3.00-4.00	4.57	9	2.50	0.10	2.20-2.80	4.00
AG: All Group	49	1.69	0.23	1.00-2.40	13.53	48	4.01	0.25	3.30-4.80	6.25	48	5.46	0.33	4.50-6.50	6.00	48	3.63	0.24	2.90-4.30	6.67	48	2.55	0.24	1.80-3.20	9.60
Cholesterol Total	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	5	38.60	2.41	35.00-43.00	6.15	6	174.67	2.42	158.00-193.00	1.37	6	220.67	3.33	199.00-243.00	1.49	6	162.83	2.48	147.00-179.00	1.53	6	123.33	1.63	111.00-135.00	1.30
PG: Abaxis Piccolo (Waived)	7	40.14	2.67	36.00-44.00	6.75																				
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	28	46.00	2.00	41.00-51.00	4.35	28	182.13	4.47	164.00-200.00	2.47	27	236.52	6.39	213.00-261.00	2.70	27	168.86	4.00	152.00-186.00	2.37	28	130.63	9.23	118.00-144.00	7.02
PG: Clinical Data Envoy 500	7	44.29	3.25	40.00-48.00	7.50	7	180.57	6.60	163.00-199.00	3.65	7	242.00	6.48	218.00-266.00	2.69	7	169.86	7.29	153.00-187.00	4.29	7	125.71	3.59	113.00-139.00	2.86
PG: Siemens Dimension/AR/ES/RxL/Xpand	16	44.28	3.39	40.00-48.00	7.73	18	178.55	5.32	161.00-197.00	2.96	18	233.65	7.55	211.00-257.00	3.21	18	165.20	4.49	149.00-182.00	2.73	18	124.50	3.25	113.00-138.00	2.64
PG: Horiba ABX Pentra 400	6	43.67	2.73	40.00-48.00	6.14	6	183.00	2.00	165.00-201.00	1.09	6	241.83	7.91	218.00-266.00	3.26	6	170.50	4.72	154.00-188.00	2.75	6	127.50	4.42	115.00-141.00	3.44
PG: Beckman (Olympus) AU	16	41.79	1.08	38.00-46.00	2.62	16	176.58	4.17	159.00-195.00	2.37	16	230.89	6.16	208.00-254.00	2.68	16	162.84	4.02	147.00-179.00	2.45	16	122.74	2.79	111.00-135.00	2.28
PG: Roche Cobas Integra / Roche Reagent	7	43.11	1.66	39.00-47.00	3.95	7	178.97	2.60	161.00-197.00	1.45	7	231.27	3.26	208.00-254.00	1.43	7	165.56	3.74	149.00-183.00	2.23	7	125.22	2.83	113.00-138.00	2.24
PG: OCD Vitros 250,350,400,500,750,950	9	45.83	1.95	41.00-51.00	4.13	9	182.00	5.58	164.00-200.00	3.08	9	238.33	5.37	214.00-262.00	2.27	9	165.17	4.30	149.00-182.00	2.61	9	116.08	2.94	104.00-128.00	2.50
SG: Select Beckman Coulter Instruments (Chem)	17	41.95	1.09	38.00-46.00	2.62	17	177.86	5.08	160.00-196.00	2.87	17	232.82	7.56	210.00-256.00	3.26	17	163.71	4.70	148.00-180.00	2.87	17	122.95	2.87	111.00-135.00	2.36
SG: Select Roche Instruments / Roche Reagent	8	43.72	2.31	40.00-48.00	5.23	8	179.35	2.63	161.00-197.00	1.45	8	232.98	5.72	210.00-256.00	2.45	8	166.12	3.80	149.00-183.00	2.29	8	125.69	2.94	113.00-139.00	2.30
SG: Select Roche Instruments / Roche Reagents Groups	9	43.53	2.24	40.00-48.00	5.00	9	178.98	2.70	161.00-197.00	1.51	9	232.65	5.44	210.00-256.00	2.32	9	165.66	3.81	149.00-183.00	2.29	9	125.39	2.89	113.00-138.00	2.32
SG: Abaxis Piccolo Instruments	12	39.50	2.58	36.00-44.00	6.50	7	175.00	2.38	158.00-193.00	1.37	7	221.00	3.16	199.00-243.00	1.45	7	162.14	2.91	146.00-178.00	1.79	7	123.29	1.50	111.00-135.00	1.22
SG: Select SpotChem Instruments	10	46.15	2.19	41.00-51.00	4.78	10	181.31	5.89	163.00-199.00	3.26	10	236.69	7.84	213.00-261.00	3.29	10	164.46	4.84	148.00-180.00	2.93	10	116.23	2.86	104.00-128.00	2.50
SG: Select Abbott Architect Instruments	7	45.00	2.00	41.00-50.00	4.44	7	181.57	3.41	164.00-200.00	1.87	7	240.57	3.50	217.00-265.00	1.45	7	167.43	3.21	150.00-184.00	1.92	7	124.57	2.99	113.00-138.00	2.40
SG: International OCD Vitros Instruments	12	45.67	1.76	41.00-51.00	3.91	12	182.53	5.37	165.00-201.00	2.95	12	240.07	6.10	216.00-264.00	2.54	12	165.27	4.20	149.00-182.00	2.55	12	117.00	3.68	105.00-129.00	3.16
SG: Select Gen Chem Instruments	37	43.10	2.29	39.00-47.00	5.35	37	179.67	5.09	162.00-198.00	2.83	37	236.93	8.00	213.00-261.00	3.38	37	166.39	5.65	149.00-183.00	3.43	37	124.40	3.57	112.00-136.00	2.90
AG: All Group	117	46.07	10.88	41.00-51.00	23.70	114	180.27	5.41	162.00-198.00	3.00	114	235.94	9.41	212.00-260.00	3.98	114	166.51	5.30	150.00-184.00	3.17	115	124.99	6.74	113.00-138.00	5.36
Cholesterol HDL	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Direct Method / Alfa Wasserman Reagent	27	18.87	1.76	13.00-25.00	9.47	27	69.90	3.66	49.00-91.00	5.29	27	84.10	7.86	59.00-109.00	9.40	27	66.14	6.30	46.00-86.00	9.55	27	54.03	4.63	38.00-70.00	8.52
PG: Cholestech (S) / Direct Method	5	41.33	0.58	29.00-53.00	1.46																				
PG: Clinical Data Envoy 500 / Direct Method	7	25.00	3.56	18.00-33.00	14.40	7	81.86	9.58	57.00-107.00	11.71	7	102.29	10.31	71.00-133.00	10.10	7	78.86	9.55	55.00-103.00	12.03	7	65.57	7.74	46.00-86.00	11.67
PG: Siemens																									



Test Name	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Siemens Dimension/AR/ES/RxL/Xpand	6	1.19	0.07	0.90-1.50	5.83	6	3.43	0.10	2.50-4.30	2.94	6	4.86	0.08	3.70-6.10	1.63	6	3.00	0.08	2.30-3.80	2.67	6	1.93	0.05	1.40-2.40	2.63
PG: Beckman (Olympus) AU	10	1.23	0.07	0.90-1.50	5.83	10	3.40	0.09	2.50-4.30	2.65	10	4.81	0.12	3.60-6.00	2.50	10	3.03	0.11	2.30-3.80	3.67	10	1.98	0.06	1.50-2.50	3.00
SG: Select Beckman Coulter Instruments (Chem)	11	1.23	0.06	0.90-1.50	5.00	11	3.36	0.11	2.50-4.30	3.24	11	4.73	0.20	3.50-5.90	4.26	11	3.01	0.11	2.30-3.80	3.67	11	1.98	0.06	1.50-2.50	3.00
SG: Abaxis Piccolo Instruments	5	1.20	0.04	0.90-1.50	3.33	5	3.42	0.04	2.50-4.30	1.18	5	5.04	0.11	3.80-6.30	2.20	5	3.05	0.11	2.30-3.80	3.67	5	1.93	0.02	1.40-2.40	1.05
AG: All Group	44	1.19	0.13	0.90-1.50	10.00	44	3.33	0.16	2.50-4.10	4.85	44	4.76	0.22	3.60-6.00	4.58	44	3.00	0.13	2.30-3.80	4.33	44	1.92	0.12	1.40-2.40	6.32
Uric Acid	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Abaxis Piccolo (Waived)	7	1.06	0.12	0.90-1.30	10.91																				
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	19	2.93	0.22	2.40-3.40	7.59	20	7.71	0.46	6.40-9.00	5.97	19	11.45	0.52	9.50-13.30	4.47	19	6.74	0.35	5.60-7.80	5.22	19	4.29	0.27	3.60-5.00	6.28
PG: Clinical Data Envoy 500	5	1.92	0.08	1.60-2.20	4.21	5	8.10	0.35	6.70-9.50	4.32	5	12.26	0.47	10.20-14.40	3.82	5	7.10	0.28	5.90-8.30	3.94	5	4.00	0.17	3.30-4.70	4.25
PG: Siemens Dimension/AR/ES/RxL/Xpand	14	1.69	0.11	1.40-2.00	6.47	14	7.49	0.30	6.20-8.80	4.00	14	11.21	0.39	9.30-13.10	3.48	14	6.53	0.20	5.40-7.60	3.08	14	3.65	0.16	3.10-4.30	4.05
PG: Beckman (Olympus) AU	11	1.95	0.10	1.60-2.20	5.26	11	7.92	0.32	6.60-9.20	4.05	11	11.86	0.47	9.90-13.90	3.95	11	6.98	0.30	5.80-8.20	4.29	11	4.00	0.20	3.30-4.70	5.00
PG: OCD Vitros 250,350,400,500,750,950	6	1.91	0.13	1.60-2.20	6.32	7	7.49	0.41	6.20-8.80	5.47	7	11.66	0.47	9.70-13.70	4.02	7	6.51	0.34	5.40-7.60	5.23	7	3.70	0.26	3.10-4.30	7.03
SG: Select Beckman Coulter Instruments (Chem)	12	1.92	0.11	1.60-2.20	5.26	12	7.83	0.36	6.50-9.10	4.62	12	11.70	0.55	9.70-13.70	4.70	12	6.91	0.32	5.70-8.10	4.64	12	3.98	0.20	3.30-4.70	5.00
SG: Select Roche Instruments / Roche Reagent	6	1.83	0.08	1.50-2.10	4.44	6	7.72	0.31	6.40-9.00	4.03	6	11.45	0.39	9.50-13.50	3.39	6	6.73	0.23	5.60-7.80	3.43	6	3.78	0.16	3.20-4.40	4.21
SG: Select OCD Vitros Instruments	7	1.93	0.13	1.60-2.20	6.84	8	7.51	0.40	6.20-8.80	5.33	8	11.66	0.45	9.70-13.70	3.85	8	6.54	0.34	5.40-7.60	5.23	8	3.72	0.26	3.10-4.30	7.03
SG: Abaxis Piccolo Instruments	7	1.02	0.06	0.80-1.20	6.00																				
SG: Select SpotChem Instruments	8	1.90	0.12	1.60-2.20	6.32	9	7.44	0.43	6.10-8.70	5.81	9	11.32	0.91	9.40-13.20	8.05	9	6.48	0.32	5.40-7.60	4.92	9	3.71	0.24	3.10-4.30	6.49
SG: Select Abbott Architect Instruments	6	1.93	0.08	1.60-2.20	4.21	6	8.33	0.10	6.90-9.70	1.20	6	12.60	0.11	10.50-14.70	0.87	6	7.27	0.12	6.10-8.50	1.64	6	4.10	0.13	3.40-4.80	3.17
SG: International OCD Vitros Instruments	9	1.92	0.13	1.60-2.20	6.84	10	7.53	0.37	6.20-8.80	4.93	10	11.71	0.42	9.70-13.70	3.59	10	6.55	0.31	5.50-7.70	4.70	10	3.72	0.23	3.10-4.30	6.49
SG: Select Gen Chem Instruments	27	1.89	0.13	1.60-2.20	6.32	27	7.92	0.41	6.60-9.20	5.19	27	11.88	0.67	9.90-13.90	5.63	27	6.95	0.37	5.80-8.20	5.29	27	3.96	0.22	3.30-4.70	5.50
AG: All Group	62	1.92	0.29	1.60-2.20	15.26	78	7.70	0.44	6.40-9.00	5.71	77	11.55	0.66	9.50-13.50	5.74	77	6.73	0.38	5.60-7.80	5.67	77	3.91	0.34	3.20-4.60	8.72
Iron	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	9	227.20	7.07	182.00-272.00	3.13	9	134.20	4.32	107.00-161.00	3.21	9	202.56	5.57	162.00-244.00	2.76	9	118.78	5.56	95.00-143.00	4.71	9	66.44	6.50	53.00-79.00	9.85
PG: Clinical Data Envoy 500	5	225.20	9.12	180.00-270.00	4.04	5	137.20	4.44	110.00-164.00	3.21	5	210.60	7.34	169.00-253.00	3.46	5	122.00	4.18	98.00-146.00	3.44	5	68.60	3.21	55.00-83.00	4.64
PG: Siemens Dimension/AR/ES/RxL/Xpand	7	232.88	2.42	186.00-280.00	1.03	7	145.75	2.05	117.00-175.00	1.44	7	220.50	3.89	177.00-265.00	1.76	7	127.38	2.56	102.00-152.00	2.05	7	71.75	1.28	58.00-86.00	1.81
PG: Beckman (Olympus) AU	9	252.17	2.62	202.00-302.00	1.03	9	161.50	1.62	130.00-194.00	0.99	9	248.25	2.83	198.00-298.00	1.13	9	138.92	3.96	111.00-167.00	2.88	9	77.08	2.27	62.00-92.00	2.99
SG: Select Beckman Coulter Instruments (Chem)	10	247.93	9.07	198.00-298.00	3.67	10	158.40	6.58	126.00-190.00	4.18	10	243.20	10.75	194.00-292.00	4.44	10	137.21	5.66	110.00-164.00	4.16	10	76.46	3.13	61.00-91.00	4.08
AG: All Group	40	235.40	11.59	188.00-282.00	4.94	40	146.38	11.06	117.00-175.00	7.60	40	222.37	18.00	178.00-266.00	8.11	40	127.71	8.74	102.00-154.00	6.80	39	71.79	5.26	58.00-86.00	7.36
TIBC	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Siemens Dimension/AR/ES/RxL/Xpand	6	322.71	14.20	280.10-365.30	4.40	6	270.86	11.29	237.00-304.80	4.17	6	407.00	15.86	359.40-454.60	3.90	6	239.29	8.42	214.00-264.60	3.52	6	136.43	6.29	117.50-155.30	4.61
PG: Beckman (Olympus) AU	6	378.00	27.32	296.00-460.00	7.23	6	417.55	80.07	177.30-657.70	19.18	5	639.34	139.96	219.40-1059.20	21.89	7	327.00	96.80	36.60-617.40	29.60	7	190.54	53.55	29.90-351.10	28.11
SG: Select Beckman Coulter Instruments (Chem)	7	378.27	22.87	309.70-446.90	6.05	7	357.95	122.10	0.00-724.20	34.12	6	530.94	208.63	0.00-1156.80	39.30	8	303.67	103.14	0.00-613.10	33.96	8	184.22	54.96	19.30-349.10	29.84
AG: All Group	21	361.25	36.13	252.90-469.70	10.00	21	318.95	90.95	46.10-591.90	28.51	20	457.09	149.64	8.20-906.00	32.74	22	278.48	76.59	48.70-508.30	27.50	22	172.60	45.45	36.20-309.00	26.33
TSH	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Abbott Architect	5	0.02	0.00	0.01-0.03	20.00	8	5.58	0.35	4.53-6.63	6.29	8	7.51	0.41	6.28-8.74	5.46	8	5.12	0.42	3.85-6.39	8.28	8	3.03	0.26	2.25-3.81	8.61
PG: Beckman ACCESS	20	0.02	0.00	0.01-0.03	20.00	23	6.19	0.60	4.39-7.99	9.68	20	8.19	0.68	6.16-10.22	8.27	21	5.38	0.40	4.17-6.59	7.49	21	3.18	0.22	2.53-3.83	6.82
PG: Siemens Dimension/AR/ES/RxL/Xpand	9	0.01	0.01	0.00-0.03	60.00	14	5.57	0.61	3.73-7.42	11.04	14	8.11	1.02	5.04-11.18	12.63	13	4.93	0.48	3.49-6.37	9.74	14	2.83	0.25	2.07-3.59	8.94
PG: Qualigen FastPak Thyroid	27	0.13	0.01	0.10-0.16	8.46	27	6.39	0.52	4.82-7.96	8.20	27	8.92	0.66	6.96-10.88	7.34	27	5.64	0.50	4.14-7.14	8.87	27	3.21	0.35	2.17-4.25	10.84
PG: TOSOH AIA	13	0.04	0.02	0.00-0.10	47.50	25	8.00	0.44	6.67-9.33	5.55	25	10.81	0.75	8.57-13.05	6.90	25	7.14	0.34	6.11-8.17	4.80	25	4.20	0.25	3.47-4.94	5.83
SG: Select Beckman Coulter Instruments (Chem)	23	0.02	0.00	0.01-0.03	20.00	25	6.28	0.36	5.19-7.37	5.80	23	8.19	0.62	6.34-10.04	7.52	24	5.41	0.40	4.21-6.61	7.41	24	3.17	0.23	2.48-3.86	7.29
PG: Abbott Architect i1000 SR	6	0.02	0.00	0.01-0.03	10.00	9	5.71	0.23	5.01-6.41	4.10	9	7.61	0.38	6.48-8.74	4.94	9	5.14	0.20	4.53-5.75	3.97	9	3.15	0.11	2.81-3.49	3.59
SG: Select Abbott Architect Instruments	13	0.02																							



AG: All Group	9	0.27	0.07	0.05-0.49	27.37	10	0.47	0.06	0.29-0.65	12.79	10	0.43	0.06	0.25-0.62	14.37	10	0.46	0.06	0.29-0.64	12.49	10	0.50	0.05	0.34-0.66	10.61
Free T3	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman ACCESS	5	0.40	0.21	0.00-1.00	52.50	8	4.80	0.19	4.20-5.40	3.96	8	5.29	0.20	4.70-5.90	3.77	8	4.74	0.20	4.10-5.30	4.26	8	4.11	0.25	3.30-4.80	6.10
SG: Select Beckman Coulter Instruments (Chem)	7	0.51	0.27	0.00-1.30	54.00	10	4.80	0.22	4.10-5.50	4.58	10	5.31	0.19	4.70-5.90	3.58	10	4.74	0.19	4.10-5.30	4.04	10	4.07	0.25	3.30-4.80	6.10
SG: Select Abbott Architect Instruments	5	1.00	0.00	1.00-1.00	0.00	5	5.26	0.06	5.10-5.50	0.94	5	5.50	0.14	5.10-5.90	2.55	5	5.50	0.37	4.40-6.60	6.73	5	4.98	0.36	3.90-6.10	7.20
SG: International OCD Vitros Instruments	5	1.24	0.05	1.00-1.40	4.17	5	16.30	0.44	15.00-17.60	2.70	5	16.93	0.49	15.40-18.40	2.90	5	16.17	0.67	14.20-18.20	4.14	5	15.07	0.67	13.10-17.10	4.44
AG: All Group	13	0.81	0.41	0.00-2.00	51.25	16	4.98	0.41	3.80-6.20	8.20	16	5.47	0.35	4.50-6.50	6.36	16	5.03	0.50	3.50-6.50	10.00	16	4.34	0.57	2.60-6.00	13.26
UIBC	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	6	101.67	15.16	56.20-147.20	14.91	6	204.67	8.48	179.30-230.10	4.14	6	298.67	12.82	260.20-337.20	4.29	6	178.33	13.11	139.00-217.60	7.35	6	107.83	16.15	59.40-156.30	14.98
Digoxin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	5	0.88	0.22	0.70-1.10	24.44	5	2.20	0.11	1.80-2.60	5.00	5	3.23	0.29	2.60-3.80	9.06	5	2.05	0.14	1.70-2.50	6.67	5	1.33	0.13	1.00-1.60	10.00
Phenytoin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	5	7.15	0.24	5.40-9.00	3.33	5	21.85	2.10	16.40-27.40	9.59	5	31.78	2.97	23.90-39.80	9.34	5	19.42	1.90	14.50-24.30	9.79	5	11.75	0.80	8.90-14.80	6.78
746 - Auto Diff 6 (5-Part Diff)																									
White Blood Cell Count																									
AG: All Group	6	21.75	0.84	18.50-25.10	3.85	6	8.17	0.42	7.00-9.40	5.12	6	3.48	0.19	3.00-4.00	5.43	6	21.68	0.33	18.40-25.00	1.52	6	8.19	0.41	7.00-9.40	5.12
Lymphocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	5	15.32	1.01	12.30-18.30	6.60	6	23.72	1.84	18.20-29.20	7.76	6	40.38	3.34	30.40-50.40	8.27	5	14.90	1.21	11.30-18.50	8.12	6	24.37	2.04	18.30-30.50	8.36
Monocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	5	2.21	0.24	1.50-2.92	10.68	6	5.73	1.09	2.46-9.00	19.04	6	6.83	1.40	2.63-11.03	20.50	5	2.57	0.68	0.53-4.61	26.46	6	5.19	0.87	2.58-7.80	16.76
Granulocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	6	74.63	5.21	59.00-90.20	6.98	6	64.20	1.69	59.10-69.30	2.63	6	47.52	3.16	38.00-57.00	6.65	6	74.58	5.72	57.40-91.80	7.68	6	63.98	1.33	60.00-68.00	2.08
Eosinophil	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	6	4.70	1.01	1.70-7.70	21.49	6	5.96	1.20	2.40-9.60	19.83	6	4.61	0.94	1.80-7.40	20.43	6	4.64	1.06	1.40-7.80	23.04	6	6.11	1.37	2.00-10.20	22.46
Basophil	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	5	0.45	0.14	0.02-0.88	31.78	6	0.37	0.11	0.05-0.69	28.92	6	0.66	0.38	0.00-1.81	58.03	6	0.50	0.28	0.00-1.35	56.40	6	0.35	0.19	0.00-0.93	54.86
Red Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	6	5.22	0.14	4.91-5.53	2.59	6	4.64	0.12	4.36-4.92	2.67	6	2.12	0.04	1.99-2.25	2.08	6	5.13	0.17	4.82-5.44	3.24	6	4.61	0.12	4.33-4.89	2.60
Hemoglobin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	6	16.43	0.54	15.30-17.50	3.29	6	13.23	0.37	12.30-14.10	2.80	6	5.60	0.13	5.20-6.00	2.32	6	16.27	0.46	15.20-17.40	2.76	6	13.12	0.37	12.20-14.00	2.82
Hematocrit	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	6	45.00	2.25	42.30-47.70	5.00	6	37.57	1.66	35.30-39.90	4.41	6	16.08	1.08	15.10-17.10	6.71	6	44.35	2.31	41.70-47.10	5.18	6	37.33	1.60	35.10-39.50	4.29
MCV	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	6	86.28	6.23	67.60-105.00	7.22	6	81.07	4.92	66.30-95.90	6.07	6	75.97	4.53	62.40-89.60	5.96	6	86.53	6.56	66.80-106.20	7.58	6	81.00	5.01	66.00-96.00	6.19
RDW	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	6	11.33	1.29	7.40-15.20	11.42	6	13.35	1.65	8.50-18.40	12.31	6	15.02	1.41	10.80-19.20	9.33	6	11.53	1.42	7.20-15.80	12.35	6	13.30	1.59	8.50-18.10	11.95
Platelet Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	6	504.83	27.27	379.00-631.00	5.41	6	287.33	26.14	215.00-359.00	9.09	6	82.57	8.62	62.00-104.00	10.36	6	491.50	24.19	369.00-615.00	4.92	6	286.00	23.66	215.00-358.00	8.29
659 - Blood Oximetry																									
Hemoglobin, total																									
AG: All Group	5	13.55	0.19	12.60-14.60	1.40	5	7.75	0.13	7.30-8.30	1.67	5	13.45	0.27	12.60-14.40	1.93	5	7.78	0.15	7.30-8.30	1.92	5	16.03	0.21	14.90-17.10	1.31
Hematocrit, calculated	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	5	41.77	0.61	39.30-44.30	1.46	5	24.00	0.40	22.60-25.40	1.67	5	41.37	0.97	38.90-43.90	2.34	5	24.33	0.38	22.80-25.80	1.56	5	48.73	0.38	45.80-51.60	0.78
683 - ASO/CRP:684 - CRP (Quant & Qual)																									
CRP Quantitative																									
AG: All Group	12	3.16	0.50	1.67-4.65	15.76	11	1.61	0.19	1.03-2.19	12.05															
600 - Cholestech LDX/CardioChek Pkg																									
CholesterolTotal (Waived)																									
PG: Cholestech LDX (Waived)	128	100.00	0.00	100.00-100.00	0.00																				
PG: Polymer CardioChek	63	100.79	6.30	82.00-120.00	6.24																				
Glucose (Waived)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Cholestech LDX (Waived)	20	50.25	2.57	42.00-58.00	5.20																				
AG: All Group	21	50.33	2.54	43.00-58.00	5.00																				
Triglycerides (Waived)	Labs	Mean	SD	Range	CV																				



Method	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Beckman Coulter Instruments (3-part Diff)	218	79.25	7.24	57.60-101.00	9.13	218	61.77	1.22	58.10-65.50	1.97	218	32.55	3.52	22.00-43.20	10.80	218	79.87	4.33	66.90-92.90	5.42	218	61.09	5.53	44.50-77.70	9.05
PG: Drew Scientific D3	24	82.02	1.12	78.60-85.40	1.37	24	62.83	0.90	60.10-65.50	1.43	24	33.65	1.50	29.20-38.20	4.42	24	82.59	0.97	79.70-85.50	1.17	24	62.74	1.13	59.30-66.10	1.80
PG: Abbott Cell-Dyn Emerald	144	81.98	7.94	58.20-105.80	9.68	144	62.98	7.05	41.90-84.20	11.19	140	35.21	1.61	30.40-40.00	4.57	144	81.85	8.16	57.30-106.30	9.98	144	62.90	7.04	41.80-84.00	11.19
PG: CDS M Series	143	77.57	9.15	50.10-105.00	11.79	143	62.76	5.28	47.00-78.60	8.41	140	31.28	3.76	20.00-42.60	12.01	143	78.58	3.45	68.30-88.90	4.39	143	63.26	2.52	55.70-70.90	3.98
SG: CDS Hematology Instruments	153	77.67	8.87	51.10-104.30	11.42	153	62.68	5.16	47.20-78.20	8.23	150	31.26	3.79	19.90-42.70	12.11	153	78.65	3.36	68.50-88.70	4.27	153	63.12	2.61	55.30-70.90	4.14
AG: All Group	678	80.32	7.26	58.50-102.10	9.04	678	62.80	4.37	49.70-75.90	6.96	671	33.40	3.36	23.30-43.50	10.06	678	80.82	5.10	65.50-96.10	6.31	678	62.62	4.88	48.00-77.20	7.80
Hematocrit	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn 1700	19	51.18	1.62	48.10-54.30	3.16	19	38.77	1.09	36.50-41.10	2.81	19	17.82	0.71	16.70-18.90	3.99	19	50.83	1.61	47.80-53.80	3.17	19	38.40	1.04	36.10-40.70	2.71
PG: Abbott Cell-Dyn 1800	40	52.67	1.87	49.50-55.90	3.55	40	39.57	1.67	37.20-42.00	4.22	40	18.77	0.38	17.70-19.90	2.02	40	51.90	1.24	48.80-55.00	2.39	40	39.19	1.07	36.80-41.60	2.73
PG: CDS Medonic 620	12	50.57	1.86	47.60-53.60	3.68	12	36.60	1.08	34.40-38.80	2.92	12	16.53	0.51	15.50-17.50	3.09	12	50.26	1.96	47.30-53.30	3.90	12	36.63	0.99	34.40-38.80	2.70
PG: Beckman Coulter AcT Diff/Diff 2	226	51.82	2.18	48.70-54.90	4.21	226	38.93	2.64	36.60-41.20	6.79	226	18.15	0.52	17.00-19.20	2.87	226	51.54	1.82	48.40-54.60	3.53	226	38.78	1.08	36.50-41.10	2.76
PG: Horiba ABX Diag. Micros/45/60	80	49.42	1.86	46.40-52.40	3.77	80	36.89	1.64	34.70-39.10	4.44	80	16.44	0.46	15.40-17.40	2.80	80	49.42	1.73	46.40-52.40	3.50	80	36.68	1.05	34.50-38.90	2.86
SG: Select Abbott Cell Dyn Instruments (3-part Diff)	205	52.37	3.19	49.30-55.50	6.09	205	39.41	1.30	37.00-41.80	3.30	198	18.61	0.59	17.50-19.70	3.17	203	52.09	1.88	49.00-55.20	3.61	205	39.36	1.23	37.00-41.80	3.12
SG: Select Beckman Coulter Instruments (3-part Diff)	227	51.82	2.17	48.70-54.90	4.19	227	38.93	2.64	36.60-41.20	6.79	227	18.15	0.52	17.00-19.20	2.87	227	51.54	1.82	48.40-54.60	3.53	227	38.78	1.07	36.50-41.10	2.76
SG: Select Drew Scientific Instruments	24	51.70	2.65	48.60-54.80	5.13	24	38.74	0.95	36.40-41.00	2.45	24	18.29	0.45	17.20-19.40	2.46	24	51.73	1.97	48.60-54.80	3.81	24	38.56	1.09	36.30-40.90	2.82
PG: Drew Scientific D3	24	51.70	2.65	48.60-54.80	5.13	24	38.74	0.95	36.40-41.00	2.45	24	18.29	0.45	17.20-19.40	2.46	24	51.73	1.97	48.60-54.80	3.81	24	38.56	1.09	36.30-40.90	2.82
PG: Abbott Cell-Dyn Emerald	145	52.44	3.60	49.30-55.50	6.87	145	39.44	1.19	37.00-41.80	3.02	138	18.67	0.54	17.60-19.80	2.89	143	52.31	2.00	49.20-55.40	3.82	145	39.53	1.23	37.10-41.90	3.11
PG: CDS M Series	144	51.83	1.21	48.70-54.90	2.34	144	37.37	1.45	35.20-39.60	3.88	141	16.87	0.40	15.90-17.90	2.37	144	51.47	3.61	48.40-54.60	7.01	144	37.51	1.99	35.30-39.80	5.31
SG: CDS Hematology Instruments	156	51.74	1.31	48.60-54.80	2.53	156	37.31	1.44	35.10-39.50	3.86	153	16.85	0.42	15.80-17.80	2.50	156	51.38	3.52	48.30-54.50	6.85	156	37.44	1.95	35.20-39.60	5.21
AG: All Group	693	51.68	2.51	48.60-54.80	4.85	693	38.46	2.12	36.20-40.80	5.51	609	17.67	0.83	16.60-18.80	4.69	691	51.43	2.45	48.30-54.50	4.77	693	38.40	1.65	36.10-40.70	4.30
Hemoglobin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn 1700	19	18.17	0.37	16.90-19.50	2.03	19	13.69	0.33	12.70-14.70	2.41	19	6.44	0.20	6.00-6.80	3.13	19	18.18	0.40	16.90-19.50	2.20	19	13.69	0.42	12.70-14.70	3.07
PG: Abbott Cell-Dyn 1800	40	18.28	0.67	17.00-19.60	3.66	40	13.59	0.51	12.60-14.60	3.75	40	6.34	0.17	5.90-6.70	2.86	40	18.24	0.42	16.90-19.50	2.31	40	13.58	0.32	12.60-14.60	2.35
PG: CDS Medonic 620	12	18.04	0.30	16.70-19.30	1.67	12	13.38	0.22	12.50-14.30	1.64	12	6.15	0.08	5.80-6.60	1.29	12	18.03	0.38	16.70-19.30	2.11	12	13.39	0.17	12.50-14.30	1.27
PG: Beckman Coulter AcT Diff/Diff 2	226	17.98	0.71	16.70-19.30	3.94	226	13.40	0.85	12.50-14.30	6.34	226	6.13	0.14	5.70-6.50	2.30	226	18.00	0.52	16.70-19.30	2.83	226	13.39	0.29	12.50-14.30	2.16
PG: Horiba ABX Diag. Micros/45/60	79	17.86	0.61	16.60-19.20	3.41	79	13.41	0.57	12.50-14.30	4.25	79	6.13	0.16	5.70-6.50	2.62	79	17.93	0.54	16.60-19.20	3.02	79	13.40	0.35	12.50-14.30	2.61
SG: Select Abbott Cell Dyn Instruments (3-part Diff)	205	18.18	0.46	16.90-19.50	2.53	205	13.52	0.35	12.60-14.40	2.59	203	6.20	0.21	5.80-6.60	3.39	205	18.17	0.50	16.90-19.50	2.75	205	13.54	0.31	12.60-14.40	2.30
SG: Select Beckman Coulter Instruments (3-part Diff)	227	17.98	0.71	16.70-19.30	3.94	227	13.40	0.85	12.50-14.30	6.34	227	6.13	0.14	5.70-6.50	2.30	227	17.99	0.52	16.70-19.30	2.83	227	13.39	0.29	12.50-14.30	2.16
SG: Select Drew Scientific Instruments	24	17.99	0.85	16.70-19.30	4.72	24	13.45	0.29	12.60-14.40	2.15	24	6.24	0.18	5.80-6.60	2.90	24	18.07	0.68	16.80-19.40	3.76	24	13.44	0.34	12.50-14.30	2.54
PG: Drew Scientific D3	24	17.99	0.85	16.70-19.30	4.72	24	13.45	0.29	12.60-14.40	2.15	24	6.24	0.18	5.80-6.60	2.90	24	18.07	0.68	16.80-19.40	3.76	24	13.44	0.34	12.50-14.30	2.54
PG: Abbott Cell-Dyn Emerald	145	18.15	0.40	16.90-19.50	2.20	145	13.47	0.29	12.60-14.40	2.15	143	6.12	0.17	5.70-6.50	2.79	145	18.15	0.53	16.80-19.40	2.93	145	13.50	0.29	12.60-14.40	2.15
PG: CDS M Series	144	18.22	0.28	16.90-19.50	1.54	144	13.49	0.46	12.60-14.40	3.41	142	6.21	0.09	5.80-6.60	1.45	144	18.22	1.19	16.90-19.50	6.54	144	13.56	0.63	12.60-14.60	4.63
SG: CDS Hematology Instruments	156	18.20	0.28	16.90-19.50	1.54	156	13.48	0.44	12.60-14.40	3.26	154	6.20	0.09	5.80-6.60	1.45	156	18.20	1.15	16.90-19.50	6.32	156	13.55	0.61	12.60-14.40	4.52
AG: All Group	692	18.07	0.57	16.80-19.40	3.15	692	13.46	0.60	12.60-14.40	4.44	688	6.17	0.16	5.80-6.60	2.58	692	18.09	0.72	16.80-19.40	3.98	692	13.47	0.40	12.60-14.40	2.96
Lymphocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn 1700	18	12.88	0.46	11.51-14.25	3.56	18	29.34	1.28	25.49-33.19	4.37	18	52.86	2.64	44.93-60.79	5.00	18	12.76	0.49	11.28-14.24	3.86	18	29.62	1.18	26.09-33.15	3.97
PG: Abbott Cell-Dyn 1800	40	11.95	0.67	9.93-13.97	5.62	40	27.56	1.45	23.22-31.90	5.25	40	50.65	3.22	41.00-60.30	6.35	40	11.95	0.72	9.79-14.11	6.03	40	27.72	1.58	22.98-32.46	5.70
PG: CDS Medonic 620	10	13.96	0.91	11.24-16.68	6.50	10	29.45	2.67	21.45-37.45	9.05	10	57.22	3.00	48.21-66.23	5.25	9	13.60	1.32	9.64-17.56	9.71	10	29.57	2.26	22.80-36.34	7.63
PG: Beckman Coulter AcT Diff/Diff 2	218	14.91	0.95	12.06-17.76	6.38	216	33.94	1.03	30.85-37.03	3.04	218	60.07	1.66	55.09-65.05	2.76	218	14.97	0.52	13.42-16.52	3.45	218	33.90	2.33	26.90-40.90	6.88
PG: Horiba ABX Diag. Micros/45/60	77	13.47	0.58	11.72-15.2																					



Test Name	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%
SG: Select Beckman Coulter Instruments (3-part Diff)	219	4.89	0.67	2.90-6.90	13.67	219	4.28	0.77	2.00-6.60	17.91	219	7.02	1.26	3.20-10.80	18.00	219	4.85	0.58	3.10-6.50	12.08	219	4.30	0.79	1.90-6.70	18.37
SG: Select Drew Scientific Instruments	24	2.34	0.18	1.80-2.80	7.83	24	4.37	0.34	3.40-5.40	7.73	24	7.07	0.96	4.20-10.00	13.52	24	2.31	0.14	1.90-2.70	5.65	24	4.45	0.48	3.10-5.90	10.67
PG: Drew Scientific D3	24	2.34	0.18	1.80-2.80	7.83	24	4.37	0.34	3.40-5.40	7.73	24	7.07	0.96	4.20-10.00	13.52	24	2.31	0.14	1.90-2.70	5.65	24	4.45	0.48	3.10-5.90	10.67
PG: Abbott Cell-Dyn Emerald	144	1.85	0.25	1.10-2.60	13.16	144	3.80	0.56	2.10-5.50	14.74	144	6.87	1.30	3.00-10.80	18.84	144	1.89	0.31	1.00-2.80	16.32	144	3.75	0.54	2.20-5.40	14.21
PG: CDS M Series	143	8.18	0.94	5.40-11.00	11.46	143	8.94	1.43	4.60-13.20	16.07	129	12.32	2.31	5.40-19.20	18.78	143	8.27	0.85	5.80-10.90	10.24	143	8.92	1.48	4.50-13.30	16.63
SG: CDS Hematology Instruments	151	8.16	0.94	5.40-11.00	11.46	153	8.95	1.40	4.70-13.10	15.73	138	12.32	2.31	5.40-19.20	18.78	150	8.25	0.84	5.80-10.80	10.12	153	8.95	1.47	4.60-13.40	16.22
AG: All Group	667	4.67	2.38	0.00-11.80	50.64	670	5.83	2.35	0.00-12.90	40.52	649	9.23	3.13	0.00-18.60	34.02	670	4.68	2.38	0.00-11.80	50.64	671	5.81	2.36	0.00-12.90	40.69
Platelet Count	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%
PG: Abbott Cell-Dyn 1700	19	562.00	50.29	422.00-703.00	8.95	19	270.47	17.62	203.00-338.00	6.52	19	63.05	5.73	47.00-79.00	9.05	19	562.11	52.54	422.00-703.00	9.34	19	267.05	18.89	200.00-334.00	7.08
PG: Abbott Cell-Dyn 1800	40	568.17	32.12	426.00-710.00	5.65	40	266.12	17.71	200.00-333.00	6.65	40	65.60	4.87	50.00-83.00	7.42	40	566.17	30.79	425.00-708.00	5.44	40	266.23	16.46	200.00-333.00	6.20
PG: CDS Medonic 620	12	496.17	29.13	372.00-620.00	5.87	12	239.58	13.73	180.00-300.00	5.71	12	62.50	5.70	47.00-79.00	9.05	12	488.00	32.13	366.00-610.00	6.58	12	245.50	13.22	185.00-308.00	5.37
PG: Beckman Coulter AcT Diff/Diff 2	226	542.96	26.78	407.00-679.00	4.94	226	258.98	12.97	194.00-324.00	5.02	226	68.02	5.70	51.00-85.00	8.38	226	543.65	23.11	408.00-680.00	4.25	226	257.27	12.72	193.00-321.00	4.94
PG: Horiba ABX Diag. Micros/45/60	80	526.58	35.34	395.00-659.00	6.70	80	267.59	33.87	201.00-335.00	12.65	80	73.95	5.91	56.00-93.00	7.97	80	527.45	21.35	395.00-659.00	4.06	80	266.42	12.69	200.00-333.00	4.77
SG: Select Abbott Cell Dyn Instruments (3-part Diff)	205	535.86	42.30	402.00-670.00	7.89	205	265.66	18.15	200.00-333.00	6.84	202	70.61	11.36	53.00-89.00	16.06	205	534.73	42.93	401.00-669.00	8.02	205	266.52	18.90	200.00-334.00	7.08
SG: Select Beckman Coulter Instruments (3-part Diff)	227	542.95	26.72	407.00-679.00	4.92	227	258.93	12.96	194.00-324.00	5.02	227	68.00	5.70	51.00-85.00	8.38	227	543.58	23.08	408.00-680.00	4.25	227	257.26	12.69	193.00-321.00	4.94
SG: Select Drew Scientific Instruments	24	545.58	34.35	410.00-683.00	6.30	24	274.17	13.25	206.00-343.00	4.82	24	79.75	9.47	60.00-100.00	11.88	24	543.33	34.22	407.00-679.00	6.30	24	268.17	12.31	201.00-335.00	4.59
PG: Drew Scientific D3	24	545.58	34.35	410.00-683.00	6.30	24	274.17	13.25	206.00-343.00	4.82	24	79.75	9.47	60.00-100.00	11.88	24	543.33	34.22	407.00-679.00	6.30	24	268.17	12.31	201.00-335.00	4.59
PG: Abbott Cell-Dyn Emerald	145	522.89	36.79	392.00-654.00	7.04	145	264.77	18.34	199.00-331.00	6.91	142	73.05	12.34	55.00-91.00	16.85	145	521.89	37.69	392.00-653.00	7.22	145	266.43	19.65	200.00-333.00	7.37
PG: CDS M Series	144	482.43	18.44	362.00-603.00	3.82	144	237.16	10.40	178.00-296.00	4.39	142	61.65	4.24	47.00-78.00	6.77	144	478.79	39.18	359.00-599.00	8.18	144	237.62	23.99	179.00-298.00	10.08
SG: CDS Hematology Instruments	156	483.47	19.67	362.00-604.00	4.08	156	237.34	10.65	178.00-296.00	4.51	154	61.71	4.35	47.00-78.00	7.10	156	479.49	38.68	359.00-599.00	8.08	156	238.22	23.41	179.00-298.00	9.83
AG: All Group	693	525.53	39.70	395.00-658.00	7.55	693	257.56	21.17	194.00-323.00	8.22	688	68.46	8.92	51.00-85.00	13.09	693	524.50	42.06	393.00-655.00	8.03	693	257.11	20.66	193.00-321.00	8.05
RDW	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%
PG: Abbott Cell-Dyn 1700	16	15.58	0.69	13.50-17.70	4.42	16	18.17	0.79	15.80-20.60	4.34	16	17.87	0.68	15.90-19.90	3.80	16	15.58	0.64	13.70-17.50	4.10	16	18.34	0.74	16.10-20.50	4.04
PG: Abbott Cell-Dyn 1800	40	14.40	0.47	13.00-15.80	3.26	40	17.20	0.71	15.00-19.40	4.19	40	17.43	0.59	15.60-19.20	3.39	40	14.48	0.52	12.90-16.10	3.59	40	17.13	0.60	15.30-18.90	3.51
PG: CDS Medonic 620	10	13.43	0.72	11.20-15.60	5.37	10	14.98	0.77	12.70-17.30	5.13	10	14.00	0.73	11.80-16.20	5.21	10	13.51	0.80	11.10-15.90	5.93	10	14.91	0.99	11.90-17.90	6.64
PG: Beckman Coulter AcT Diff/Diff 2	216	13.29	0.25	12.60-14.10	1.88	217	15.50	0.52	13.90-17.10	3.35	217	15.83	0.30	14.90-16.70	1.90	217	13.29	0.25	12.60-14.10	1.88	217	15.49	0.56	13.80-17.20	3.61
PG: Horiba ABX Diag. Micros/45/60	69	12.32	0.39	11.10-13.50	3.17	69	13.96	0.42	12.70-15.30	3.00	69	13.94	0.37	12.80-15.00	2.66	69	12.33	0.34	11.30-13.30	2.76	69	14.04	0.39	12.80-15.20	2.79
SG: Select Abbott Cell Dyn Instruments (3-part Diff)	199	13.43	1.12	10.00-16.80	8.36	199	15.74	1.27	11.90-19.50	8.09	199	15.68	1.32	11.70-19.70	8.41	199	13.36	1.09	10.20-16.60	8.06	199	15.80	1.23	12.10-19.50	7.78
SG: Select Beckman Coulter Instruments (3-part Diff)	217	13.29	0.25	12.60-14.10	1.88	218	15.50	0.52	13.90-17.10	3.35	218	15.83	0.30	14.90-16.70	1.90	218	13.29	0.25	12.60-14.10	1.88	218	15.49	0.56	13.80-17.20	3.61
SG: Select Drew Scientific Instruments	24	13.16	0.75	10.90-15.40	5.68	24	15.28	0.73	13.10-17.50	4.77	24	15.11	0.86	12.50-17.70	5.70	24	13.23	0.69	11.10-15.30	5.23	24	15.20	1.33	11.20-19.20	8.75
PG: Drew Scientific D3	24	13.16	0.75	10.90-15.40	5.68	24	15.28	0.73	13.10-17.50	4.77	24	15.11	0.86	12.50-17.70	5.70	24	13.23	0.69	11.10-15.30	5.23	24	15.20	1.33	11.20-19.20	8.75
PG: Abbott Cell-Dyn Emerald	143	12.91	0.75	10.70-15.20	5.81	143	15.05	0.52	13.40-16.60	3.47	143	14.95	0.59	13.20-16.80	3.93	143	12.80	0.55	11.20-14.50	4.30	143	15.14	0.52	13.50-16.70	3.44
PG: CDS M Series	142	13.40	1.19	9.80-17.00	8.88	142	14.69	1.28	10.90-18.50	8.71	142	13.94	1.23	10.20-17.60	8.85	142	13.36	1.36	9.30-17.50	10.15	142	14.71	1.29	10.80-18.60	8.78
SG: CDS Hematology Instruments	152	13.41	1.16	9.90-16.90	8.66	152	14.70	1.25	10.90-18.40	8.50	152	13.95	1.20	10.30-17.50	8.63	152	13.37	1.33	9.40-17.40	9.93	152	14.73	1.27	10.90-18.50	8.64
AG: All Group	662	13.26	0.93	10.50-16.10	6.99	663	15.23	1.16	11.70-18.70	7.63	663	15.14	1.30	11.20-19.00	8.61	663	13.23	0.96	10.30-16.10	7.27	663	15.25	1.17	11.80-18.80	7.65
Red Blood Cell Count	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%
PG: Abbott Cell-Dyn 1700	19	5.54	0.14	5.21-5.87	2.51	19	4.76	0.11	4.47-5.05	2.35	19	2.43	0.08	2.28-2.58	3.21	19	5.51	0.18	5.18-5.84	3.23	19	4.69	0.09	4.41-4.97	1.86
PG: Abbott Cell-Dyn 1800	40	5.47	0.21	5.14-5.80	3.80	40	4.67	0.19	4.39-4.95	4.11	40	2.48	0.06	2.33-2.63	2.46	40	5.41	0.13	5.09-5.73	2.31	40	4.63	0.13	4.35-4.91	2.83
PG: CDS Medonic 620	12	5.69	0.13	5.35-6.03	2.20	12	4.75	0.07	4.46-5.04	1.49	12	2.37	0.04	2.23-2.51	1.81	12	5.68	0.11	5.34-6.02	2.01	12				



PG: Abbott i-Stat	6	24.58	1.53	20.00-29.20	6.22	6	20.63	0.91	17.90-23.30	4.42	6	22.30	0.70	20.20-24.40	3.14	6	21.15	0.61	19.40-23.00	2.88	6	21.10	0.68	19.10-23.10	3.22
SG: International Abbott i-Stat Instruments	10	24.44	1.18	20.90-27.90	4.84	10	20.41	0.76	18.10-22.70	3.73	10	22.06	0.74	19.90-24.30	3.35	10	20.91	0.65	18.90-22.80	3.11	10	21.09	0.54	19.50-22.70	2.56
BEecf	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat / Thermal	5	2.80	0.45	2.00-4.00	13.33	6	-2.17	1.17	0.00-2.00	-60.00	6	-5.83	0.75	0.00-4.00	-13.33	6	-6.33	0.82	0.00-4.00	-13.33	6	-5.50	1.05	0.00-3.00	-16.67
sO2	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat	6	99.00	0.00	99.00-99.00	0.00	6	99.00	0.00	99.00-99.00	0.00	6	90.50	1.38	87.00-95.00	1.54	6	99.00	0.00	99.00-99.00	0.00	6	99.00	0.00	99.00-99.00	0.00
Lactate	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat	5	0.97	0.06	0.80-1.14	5.77	5	1.51	0.04	1.38-1.64	2.91	5	4.64	0.08	4.40-4.88	1.70	5	3.34	0.05	3.18-3.50	1.59	5	2.27	0.06	2.08-2.46	2.73
SG: International Abbott i-Stat Instruments	9	0.95	0.05	0.81-1.09	5.05	9	1.48	0.04	1.35-1.61	2.84	9	4.62	0.07	4.41-4.83	1.54	9	3.32	0.05	3.17-3.47	1.51	9	2.24	0.06	2.07-2.41	2.50
Anion Gap	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Abbott i-Stat Instruments / Thermal	10	41.50	1.18	38.00-46.00	2.86	5	48.50	0.71	47.00-51.00	1.43	5	31.00	0.00	31.00-31.00	0.00	5	34.50	0.71	33.00-37.00	2.00	5	37.50	2.12	32.00-44.00	5.53
716 - Blood Lead - Waived Methods;709 - Blood Lead	LE-1					LE-2					LE-3					LE-4					LE-5				
Blood Lead	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select ESA LeadCare Instruments	5	44.35	2.38	39.90-48.70	5.37	6	4.33	1.14	0.30-8.30	26.51	5	14.48	3.18	10.50-18.50	21.93	5	42.14	3.35	37.90-46.30	7.96	6	4.62	1.79	0.60-8.60	38.91
PG: Magellan Diag LeadCare Ultra	5	31.00	0.00	27.00-35.00	0.00	5	3.00	0.00	0.00-7.00	0.00	5	10.00	0.00	6.00-14.00	0.00	5	37.00	0.00	33.00-41.00	0.00	5	2.00	0.00	0.00-6.00	0.00
Blood Lead (Waived)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: ESA LeadCare II (Waived)	41	43.20	5.16	27.70-58.70	11.94																				
662 - Neonatal Bilirubin (5 specimens);663 - Neonatal Bilirubin (2 specimens)	NB-1					NB-2					NB-3					NB-4					NB-5				
Neonatal Bilirubin (Total)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Reichert (Leica) Unistat Bilirubinometer	9	12.16	0.53	9.80-14.60	4.34	5	0.10	0.00	0.00-0.50	0.00	9	19.72	0.59	15.80-23.60	3.05	9	5.91	0.26	4.70-7.10	4.41	9	17.28	0.44	13.80-20.80	2.54
AG: All Group	18	11.94	0.49	9.50-14.30	4.12	10	0.22	0.11	0.00-0.60	55.00	15	19.27	0.83	15.40-23.20	4.30	15	6.11	0.34	4.90-7.30	5.57	15	16.74	0.86	13.40-20.00	5.15
748 - i-Stat Prottime	PTI-1					PTI-2					PTI-3					PTI-4					PTI-5				
Prottime - Whole Blood	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat Prottime	9	14.37	0.83	12.20-16.60	5.76	9	29.17	5.24	24.80-33.60	17.95	9	15.09	3.07	12.80-17.40	20.33	9	30.57	7.27	26.00-35.20	23.76	9	13.96	0.56	11.90-16.10	4.00
SG: International Abbott i-Stat Instruments	10	14.36	0.78	12.20-16.60	5.42	10	29.25	4.94	24.90-33.70	16.86	10	14.85	2.99	12.60-17.00	20.20	10	30.27	6.92	25.80-34.80	22.84	10	13.94	0.53	11.80-16.00	3.81
AG: All Group	15	14.89	2.36	12.70-17.10	15.84																				
INR	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat Prottime	10	1.21	0.07	1.00-1.40	5.83	10	2.47	0.39	1.30-3.70	15.60	10	1.26	0.24	0.60-2.00	18.46	10	2.61	0.52	1.00-4.20	20.00	10	1.17	0.05	1.00-1.40	4.17
PG: Abbott CoaguSense	7	1.33	0.30	0.40-2.20	23.08	6	3.52	0.19	2.90-4.10	5.43	7	1.30	0.25	0.60-2.00	19.23	6	3.42	0.16	2.90-3.90	4.71	7	1.16	0.17	0.70-1.70	14.17
SG: International Abbott i-Stat Instruments	15	1.21	0.06	1.00-1.40	5.00	15	2.49	0.32	1.50-3.50	12.80	15	1.23	0.20	0.60-1.80	16.67	15	2.57	0.42	1.30-3.90	16.54	15	1.17	0.05	1.00-1.40	4.17
SG: International Abbott CoaguSense Instruments	7	1.33	0.30	0.40-2.20	23.08	7	3.90	1.03	0.80-7.00	26.41	7	1.30	0.25	0.60-2.00	19.23	7	3.76	0.91	1.10-6.50	23.95	7	1.16	0.17	0.70-1.70	14.17
611 - QBC Hematology	QBC-1					QBC-2					QBC-3					QBC-4					QBC-5				
Hemoglobin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: QBC Star - QBC Star Tubes	8	11.59	0.33	10.80-12.40	2.84	9	7.52	0.28	7.00-8.00	3.73	8	11.73	0.61	10.90-12.50	5.21	8	9.80	0.36	9.10-10.50	3.67	8	11.29	0.53	10.50-12.10	4.69
AG: All Group	15	11.37	0.51	10.60-12.20	4.47	16	7.48	0.29	7.00-8.00	3.87	15	11.41	0.62	10.60-12.20	5.44	15	9.67	0.33	9.00-10.40	3.40	15	11.06	0.51	10.30-11.90	4.59
Hematocrit	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: QBC Star - QBC Star Tubes	9	34.98	1.37	32.90-37.10	3.91	8	22.89	0.89	21.50-24.30	3.89	8	35.14	1.26	33.00-37.20	3.59	8	30.04	1.35	28.20-31.80	4.50	8	34.34	1.55	32.20-36.40	4.52
AG: All Group	16	34.96	1.35	32.90-37.10	3.86	15	22.77	0.88	21.40-24.20	3.86	15	34.93	1.47	32.80-37.00	4.21	15	29.40	1.22	27.60-31.20	4.15	15	33.60	1.59	31.60-35.60	4.73
White Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: QBC Star - QBC Star Tubes	9	27.53	2.72	23.40-31.60	9.89	10	8.34	0.60	7.10-9.50	7.23	8	27.08	2.30	23.00-31.20	8.49	10	9.00	0.83	7.70-10.30	9.22	10	18.10	2.02	15.40-20.80	11.16
SG: Select QBC - Cap and Ven	5	16.10	1.68	13.70-18.50	10.43	5	6.18	0.72	5.30-7.10	11.61	5	16.44	1.98	13.90-18.90	12.07	5	6.70	0.80	5.70-7.70	11.94	5	12.26	0.85	10.50-14.10	6.91
SG: Select QBC Instruments	11	26.90	2.84	22.90-30.90	10.56	12	8.18	0.72	7.00-9.40	8.78	10	26.82	2.17	22.80-30.80	8.10	11	8.96	0.80	7.70-10.30	8.89	10	17.78	1.55	15.10-20.50	8.71
Granulocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: QBC Star - QBC Star Tubes	9	45.78	6.16	27.00-65.00	13.48	9	62.33	1.41	58.00-66.00	2.26	9	46.56	4.69	33.00-61.00	10.00	9	55.67	2.50	49.00-64.00	4.46	9	48.11	2.71	40.00-56.00	5.63
AG: All Group	16	47.50	6.35	29.00-67.00	13.13																				



AG: All Group	35	19.53	3.34	9.50-29.50	17.13																				
Folate	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman ACCESS	9	0.41	0.14	0.00-0.80	35.00																				
PG: TOSOH AIA	6	1.15	0.19	0.60-1.80	15.83																				
SG: Select Roche Instruments	5	1.90	0.22	1.20-2.60	11.58																				
SG: Select Abbott Architect Instruments	7	0.41	0.12	0.00-0.80	30.00																				
AG: All Group	23	0.50	0.28	0.00-1.30	56.00																				
FSH	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: TOSOH AIA	5	13.56	0.42	12.30-14.90	3.09																				
AG: All Group	12	10.01	1.91	4.30-15.70	19.10																				
LH	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	14	3.84	0.68	1.80-5.80	17.89																				
Progesterone	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	11	2.30	0.59	0.50-4.10	25.65																				
Prolactin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	11	2.78	0.77	0.50-5.10	27.50																				
PSA	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman ACCESS	21	0.03	0.01	0.01-0.05	20.00																				
PG: Siemens Dimension/AR/ES/RxL/Xpand	11	0.10	0.03	0.00-0.20	33.00																				
PG: Qualigen FastPack PSA	17	0.04	0.00	0.04-0.04	0.00																				
PG: TOSOH AIA	14	0.05	0.02	0.00-0.11	38.00																				
SG: Select Beckman Coulter Instruments (Chem)	23	0.03	0.01	0.02-0.05	16.67																				
SG: Select Sp Chem II	13	0.06	0.01	0.02-0.11	25.00																				
PG: Abbott Architect i1000 SR	7	0.01	0.00	0.01-0.01	10.00																				
SG: Select Abbott Architect Instruments	13	0.01	0.00	0.01-0.01	10.00																				
SG: Select Gen Chem Instruments	13	0.01	0.00	0.01-0.01	10.00																				
AG: All Group	88	0.04	0.03	0.00-0.12	65.00																				
Testosterone	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman ACCESS	13	0.89	0.10	0.60-1.20	11.11																				
PG: Qualigen FastPak Testosterone	9	0.83	0.14	0.40-1.20	17.50																				
PG: TOSOH AIA	8	1.13	0.10	0.80-1.40	9.09																				
SG: Select Beckman Coulter Instruments (Chem)	15	0.89	0.09	0.60-1.20	10.00																				
SG: Select Roche Instruments	5	0.94	0.11	0.60-1.20	12.22																				
PG: Abbott Architect i1000 SR	6	0.94	0.07	0.70-1.10	7.78																				
SG: Select Abbott Architect Instruments	9	0.91	0.10	0.60-1.20	11.11																				
AG: All Group	53	0.94	0.16	0.40-1.40	17.78																				
Vitamin B12	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman ACCESS	16	137.56	8.12	114.00-162.00	5.87																				
PG: TOSOH AIA	6	208.50	23.65	138.00-280.00	11.34																				
SG: Select Beckman Coulter Instruments (Chem)	18	138.28	7.92	114.00-162.00	5.72																				
SG: Select Roche Instruments	5	181.30	23.98	109.00-253.00	13.26																				
SG: Select Sp Chem I	8	209.75	16.53	161.00-260.00	7.86																				
PG: Abbott Architect i1000 SR	8	202.33	34.88	97.00-307.00	17.28																				
SG: Select Abbott Architect Instruments	14	206.27	30.88	113.00-299.00	15.00																				
AG: All Group	49	178.79	36.45	70.00-288.00	20.34																				
685 - Special Chemistry II	SPC-1					SPC-2																			
Insulin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	7	110.18	21.25	46.50-173.90	19.28	8	12.80	1.90	7.10-18.50	14.84															
PTH	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	9	40.16	11.23	6.50-73.90	27.94	10	5.58	1.54	1.00-10.20	27.50															
Vitamin D	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman ACCESS	17	100.98	9.79	71.59-130.37	9.70	16	12.56	2.38	5.42-19.70	18.94															
PG: TOSOH AIA	7	118.37	2.20	111.76-124.98	1.86	8	27.84	1.53	23.26-32.42	5.49															
SG: Select Beckman Coulter Instruments (Chem)	19	102.39	10.32	71.43-133.35	10.08	17	11.95	2.22	5.29-18.61	18.57															
SG: Select Roche Instruments	6	55.55	7.40	33.36-77.74	13.32	6	7.40	1.63	2.53-12.28	21.96															
PG: Roche Cobas e 411	6	55.55	7.40	33.36-77.74	13.32	6	7.40	1.63	2.53-12.28	21.96															
PG: Abbott Architect i1000 SR	12	94.81	4.13	82.43-107.19	4.35	12	16.40	1.94	10.58-22.22	11.83															
SG: Select Abbott Architect Instruments	18	95.21	3.37	85.10-105.32	3.54	18	16.55	1.76	11.26-21.84	10.66															
PG: Qualigen FastPack Vitamin D	10	78.80	9.60	50.01-107.59	12.18	9	21.61	5.29	5.73-37.49	24.49															
SG: Select Gen Chem Instruments	19	96.93	8.19	72.38-121.49	8.44	18	16.55	1.76	11.26-21.84	10.66															
AG: All Group	59	95.64	16.85	45.10-146.18	17.62	46	15.74	4.55	2.08-29.40	28.92															
752 - Auto Diff 8 (5-part Diff)	SXT-1					SXT-2					SXT-3					SXT-4					SXT-5				
White Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Sysmex Instruments (5-part Diff)	13	20.38	0.89	17.30-23.50	4.36	13	7.25	0.56	6.20-8.40	7.53	13	3.79	0.11	3.20-4.40	2.89	13	20.45	0.96	17.40-23.60	4.68	13	7.15	0.67	6.00-8.20	9.44
Lymphocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Sysmex Instruments (5-part Diff)	13	19.33	3.90	7.60-31.00	20.21	13	27.51	4.11	15.20-39.80	14.95	13	11.52	4.39	0.00-24.											



Mono/Mixed	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Sysmex Instruments (5-part Diff)	6	6.20	0.25	5.50-7.00	4.03	5	6.62	1.06	3.40-9.80	16.06	6	5.20	0.38	4.10-6.30	7.31	6	6.08	0.27	5.30-6.90	4.43	5	7.48	0.15	7.00-8.00	2.00
Eosinophil	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Sysmex Instruments (5-part Diff)	11	13.16	1.28	9.40-17.00	9.70	13	14.08	1.59	9.30-18.90	11.28	13	20.42	5.67	3.40-37.40	27.79	11	13.35	1.33	9.40-17.40	9.85	13	13.85	1.80	8.50-19.30	12.95
Basophil	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Sysmex Instruments (5-part Diff)	7	8.34	1.73	3.14-13.54	20.78	7	7.33	1.36	3.25-11.41	18.57	6	9.83	0.56	8.16-11.50	5.67	7	8.36	1.77	3.04-13.68	21.23	7	7.14	1.18	3.61-10.67	16.47
Red Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Sysmex Instruments (5-part Diff)	13	5.57	0.05	5.24-5.90	0.90	12	4.62	0.04	4.34-4.90	0.93	13	2.33	0.03	2.19-2.47	1.46	13	5.58	0.07	5.25-5.91	1.18	13	4.61	0.05	4.33-4.89	1.04
Hemoglobin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Sysmex Instruments (5-part Diff)	13	17.52	0.17	16.30-18.70	0.97	13	13.39	0.12	12.50-14.30	0.90	13	5.82	0.08	5.40-6.20	1.38	13	17.60	0.17	16.40-18.80	0.97	13	13.40	0.11	12.50-14.30	0.82
Hematocrit	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Sysmex Instruments (5-part Diff)	13	51.25	1.30	48.20-54.40	2.53	13	39.32	0.86	36.90-41.70	2.19	13	17.85	0.37	16.70-18.90	2.08	13	51.32	1.35	48.20-54.40	2.63	13	39.20	0.78	36.80-41.60	1.99
MCV	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Sysmex Instruments (5-part Diff)	13	91.95	1.91	86.30-97.70	2.08	13	85.15	1.35	81.10-89.10	1.57	13	76.56	1.11	73.30-79.90	1.45	13	91.89	1.89	86.20-97.60	2.06	13	85.01	1.30	81.10-88.90	1.53
Platelet Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Sysmex Instruments (5-part Diff)	13	426.54	6.17	320.00-534.00	1.45	13	225.23	7.50	169.00-281.00	3.33	13	65.69	4.31	50.00-83.00	6.52	13	427.08	11.66	320.00-534.00	2.74	13	222.38	7.16	167.00-278.00	3.24
RDW	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Sysmex Instruments (5-part Diff)	13	14.10	0.21	13.50-14.70	1.49	13	15.05	0.22	14.40-15.80	1.46	13	15.99	0.38	14.90-17.10	2.38	13	14.08	0.16	13.60-14.60	1.13	13	15.05	0.22	14.40-15.80	1.46
613 - Auto Diff I	SYX-1					SYX-2					SYX-3					SYX-4					SYX-5				
White Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex K-800,1000,4500,KX-21	21	19.91	0.43	16.90-22.90	2.16	21	7.92	0.14	6.70-9.10	1.77	21	2.77	0.06	2.40-3.20	2.14	21	19.82	0.42	16.80-22.80	2.12	21	7.92	0.11	6.70-9.10	1.39
PG: Sysmex pocH-100i	26	19.91	0.71	16.90-22.90	3.57	26	7.99	0.32	6.80-9.20	4.00	26	2.73	0.11	2.30-3.10	4.07	26	19.90	0.62	16.90-22.90	3.12	26	7.97	0.22	6.80-9.20	2.75
PG: Sysmex XP-300	57	20.97	0.55	17.90-24.10	2.62	57	8.44	0.24	7.10-9.70	2.86	57	2.87	0.13	2.50-3.30	4.48	57	20.90	0.52	17.80-24.00	2.49	56	8.44	0.23	7.10-9.70	2.74
SG: International Sysmex Instruments	31	20.15	1.45	17.20-23.20	7.18	31	7.98	0.31	6.80-9.20	3.88	31	2.73	0.11	2.30-3.10	4.07	31	20.01	0.75	17.00-23.00	3.75	31	7.97	0.22	6.80-9.20	2.75
Neutrophil	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex K-800,1000,4500,KX-21	21	26.00	0.68	24.00-28.00	2.62	21	51.86	1.06	48.70-55.10	2.04	21	69.73	1.36	65.60-73.80	1.95	21	26.08	0.63	24.20-28.00	2.41	21	51.39	0.97	48.50-54.30	1.89
PG: Sysmex pocH-100i	25	26.90	1.07	23.70-30.10	3.98	24	53.65	1.57	48.90-58.30	2.93	24	72.10	1.53	67.50-76.70	2.12	25	26.67	0.77	24.40-29.00	2.88	24	53.87	1.17	50.40-57.40	2.17
PG: Sysmex XP-300	55	26.24	0.80	23.80-28.60	3.02	55	51.70	1.20	48.10-55.30	2.32	55	69.57	1.79	64.20-75.00	2.57	55	26.15	0.73	23.90-28.30	2.80	53	51.67	1.13	48.30-55.10	2.19
SG: International Sysmex Instruments	30	26.83	1.00	23.80-29.80	3.73	29	53.70	1.45	49.40-58.10	2.70	29	72.26	1.70	67.20-77.40	2.35	30	26.80	0.89	24.10-29.50	3.32	29	53.85	1.18	50.30-57.30	2.19
Red Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex K-800,1000,4500,KX-21	21	5.67	0.10	5.33-6.01	1.76	21	4.20	0.04	3.95-4.45	0.86	21	2.39	0.03	2.25-2.53	1.30	21	5.64	0.04	5.30-5.98	0.62	21	4.20	0.04	3.95-4.45	0.88
PG: Sysmex pocH-100i	26	5.69	0.10	5.35-6.03	1.76	26	4.31	0.12	4.05-4.57	2.69	26	2.41	0.04	2.27-2.55	1.83	26	5.72	0.11	5.38-6.06	1.99	26	4.27	0.07	4.01-4.53	1.59
PG: Sysmex XP-300	56	5.65	0.13	5.31-5.99	2.30	56	4.22	0.04	3.97-4.47	1.02	56	2.38	0.03	2.24-2.52	1.13	56	5.64	0.09	5.30-5.98	1.58	56	4.23	0.21	3.98-4.48	4.87
SG: International Sysmex Instruments	30	5.71	0.11	5.37-6.05	1.84	31	4.31	0.11	4.05-4.57	2.58	31	2.40	0.04	2.26-2.54	1.79	31	5.74	0.15	5.40-6.08	2.68	31	4.27	0.07	4.01-4.53	1.57
Hemoglobin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex K-800,1000,4500,KX-21	21	17.98	0.37	16.70-19.30	2.06	21	12.09	0.16	11.30-12.90	1.32	21	6.06	0.12	5.70-6.50	1.97	21	17.96	0.21	16.70-19.30	1.17	21	12.15	0.15	11.30-12.90	1.24
PG: Sysmex pocH-100i	26	18.17	0.34	16.90-19.50	1.87	26	12.18	0.38	11.30-13.10	3.11	26	6.02	0.10	5.60-6.40	1.67	26	18.03	0.27	16.70-19.30	1.50	26	12.11	0.20	11.30-12.90	1.65
PG: Sysmex XP-300	57	18.01	0.39	16.70-19.30	2.17	57	12.17	0.13	11.30-13.10	1.07	57	6.01	0.12	5.60-6.40	2.00	57	18.01	0.26	16.70-19.30	1.44	57	12.28	0.80	11.40-13.20	6.50
SG: International Sysmex Instruments	30	18.16	0.34	16.90-19.50	1.92	31	12.16	0.35	11.30-13.10	2.95	31	6.02	0.10	5.60-6.40	1.67	31	18.09	0.48	16.80-19.40	2.65	31	12.10	0.20	11.30-12.90	1.65
Lymphocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex K-800,1000,4500,KX-21	21	59.13	0.55	57.50-60.80	0.93	21	31.22	0.80	28.80-33.60	2.56	21	12.22	0.96	9.30-15.10	7.87	21	59.40	0.79	57.00-61.80	1.33	21	31.19	0.81	28.80-33.60	2.60
PG: Sysmex pocH-100i	25	59.77	0.60	58.00-61.60	1.00	25	30.36	1.05	27.30-33.50	3.45	23	11.13	0.69	9.00-13.20	6.22	25	59.78	0.61	58.00-61.60	1.02	25	30.25	1.11	27.00-33.60	3.66
PG: Sysmex XP-300	56	59.06	0.66	57.10-61.10	1.12	56	31.05	0.97	28.10-33.90	3.13	52	12.51	1.09	9.20-15.80	8.72	56	59.09	0.66	57.10-61.10	1.12	55	31.21	1.16	27.70-34.70	3.72
SG: International Sysmex Instruments	30	59.80	0.58	58.10-61.50	0.97	30	30.27	1.01	27.30-33.30	3.33	28	11.14	0.78	8.80-13.40	7.03	30	59.69	0.67	57.70-61.70	1.12	30	30.23	1.07	27.00-33.40	3.54
Hematocrit	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex K-800,1000,4500,KX-21	21	47.94	1.18	45.00-50.80	2.																				



789 - Bacterial Antigen Detection	Labs	Response	Number	Percent
		C. difficile Toxin Detect		
AD-1		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Alere C diff Quik Chek Complete	2	*Negative	2	100.00%
AD-2		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Alere C diff Quik Chek Complete	2	*Negative	2	100.00%
AD-3		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Alere C diff Quik Chek Complete	2	*Positive	2	100.00%
AD-4		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Alere C diff Quik Chek Complete	2	*Positive	2	100.00%
AD-5		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Alere C diff Quik Chek Complete	2	*Negative	2	100.00%
643 - ANA (Latex);760 - ANA (ELISA);757 - Special Immunology	Labs	Response	Number	Percent
		ANA (ELISA) (International)		
AE-1		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Dynex DS2	1	*Positive	1	100.00%
AE-2		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Dynex DS2	1	*Negative	1	100.00%
AE-3		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Dynex DS2	1	*Positive	1	100.00%
AE-4		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Dynex DS2	1	*Positive	1	100.00%
AE-5		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Dynex DS2	1	*Negative	1	100.00%
671 - Total IgE;674 - Diagnostic Allergy;765 - Allergen Testing (1 specimen)	Labs	Response	Number	Percent
		Aspergillus fumigatus Alrgn		
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	5			
		*0	2	40.00%
		*1/0	3	60.00%
Hitachi CLA-1	5	*0	2	40.00%
Hitachi CLA-1		*1/0	3	60.00%
		Bermuda Grass Alrgn		
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	7			
		*0	1	14.29%
		*1	3	42.86%
		*1/0	2	28.57%
		4	1	14.29%
Hitachi CLA-1	7	*0	1	14.29%
Hitachi CLA-1		*1	3	42.86%
Hitachi CLA-1		*1/0	2	28.57%
Hitachi CLA-1		4	1	14.29%
		Cat Epithelium Alrgn		
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	5			



		*3	2	40.00%
		*4	2	40.00%
		0	1	20.00%
Hitachi CLA-1	5	*3	2	40.00%
Hitachi CLA-1		*4	2	40.00%
Hitachi CLA-1		0	1	20.00%
Dog Dander Alrgn				
AL-1(Q)				
* Denotes Correct Response				
AG: All Methods	5			
		*4	4	80.00%
		0	1	20.00%
Hitachi CLA-1	5	*4	4	80.00%
Hitachi CLA-1		0	1	20.00%
Dog Epithelium Alrgn				
AL-1(Q)				
* Denotes Correct Response				
AG: All Methods	3			
		*4	3	100.00%
Hitachi CLA-1	3	*4	3	100.00%
English Plantain Alrgn				
AL-1(Q)				
* Denotes Correct Response				
AG: All Methods	4			
		*0	1	25.00%
		*1/0	3	75.00%
Hitachi CLA-1	4	*0	1	25.00%
Hitachi CLA-1		*1/0	3	75.00%
House Dust Mite Alrgn				
AL-1(Q)				
* Denotes Correct Response				
AG: All Methods	6			
		*1	1	16.67%
		*2	3	50.00%
		*3	1	16.67%
		0	1	16.67%
Hitachi CLA-1	6	*1	1	16.67%
Hitachi CLA-1		*2	3	50.00%
Hitachi CLA-1		*3	1	16.67%
Hitachi CLA-1		0	1	16.67%
Box Elder, Maple Alrgn				
AL-1(Q)				
* Denotes Correct Response				
AG: All Methods	4			
		*0	4	100.00%
Hitachi CLA-1	4	*0	4	100.00%
Oak Alrgn				
AL-1(Q)				
* Denotes Correct Response				
AG: All Methods	4			
		*1	1	25.00%
		*1/0	1	25.00%
		*2	2	50.00%
Hitachi CLA-1	4	*1	1	25.00%
Hitachi CLA-1		*1/0	1	25.00%
Hitachi CLA-1		*2	2	50.00%
Peanut (Arachis hypogaea) Alrgn				
AL-1(Q)				
* Denotes Correct Response				
AG: All Methods	5			
		*4	5	100.00%
Hitachi CLA-1	5	*4	5	100.00%
Soybean Alrgn				
AL-1(Q)				
* Denotes Correct Response				
AG: All Methods	5			
		*1	1	20.00%
		*2	3	60.00%
		0	1	20.00%
Hitachi CLA-1	5	*1	1	20.00%
Hitachi CLA-1		*2	3	60.00%
Hitachi CLA-1		0	1	20.00%
Timothy Grass Alrgn				
AL-1(Q)				
* Denotes Correct Response				
AG: All Methods	6			
		*4	6	100.00%



Hitachi CLA-1	6	*4	6	100.00%
Alternaria alternata Alrgn				
AL-1(Q) * Denotes Correct Response				
AG: All Methods	7			
		*0	3	42.86%
		*1	4	57.14%
Hitachi CLA-1	7	*0	3	42.86%
Hitachi CLA-1		*1	4	57.14%
Almond Alrgn				
AL-1(Q) * Denotes Correct Response				
AG: All Methods	3			
		*0	2	66.67%
		*1/0	1	33.33%
Hitachi CLA-1	3	*0	2	66.67%
Hitachi CLA-1		*1/0	1	33.33%
Apple Alrgn				
AL-1(Q) * Denotes Correct Response				
AG: All Methods	3			
		*0	3	100.00%
Hitachi CLA-1	3	*0	3	100.00%
Ash - White Alrgn				
AL-1(Q) * Denotes Correct Response				
AG: All Methods	4			
		*0	3	75.00%
		*1/0	1	25.00%
Hitachi CLA-1	4	*0	3	75.00%
Hitachi CLA-1		*1/0	1	25.00%
Barley Alrgn				
AL-1(Q) * Denotes Correct Response				
AG: All Methods	3			
		*0	3	100.00%
Hitachi CLA-1	3	*0	3	100.00%
Elm - White Alrgn				
AL-1(Q) * Denotes Correct Response				
AG: All Methods	5			
		*0	5	100.00%
Hitachi CLA-1	5	*0	5	100.00%
Bean - White Alrgn				
AL-1(Q) * Denotes Correct Response				
AG: All Methods	3			
		*0	3	100.00%
Hitachi CLA-1	3	*0	3	100.00%
Beef Alrgn				
AL-1(Q) * Denotes Correct Response				
AG: All Methods	3			
		*0	3	100.00%
Hitachi CLA-1	3	*0	3	100.00%
Birch/Alder Mix Alrgn				
AL-1(Q) * Denotes Correct Response				
AG: All Methods	5			
		*0	1	20.00%
		*1	1	20.00%
		*1/0	3	60.00%
Hitachi CLA-1	5	*0	1	20.00%
Hitachi CLA-1		*1	1	20.00%
Hitachi CLA-1		*1/0	3	60.00%
Cedar - Mountain Alrgn				
AL-1(Q) * Denotes Correct Response				
AG: All Methods	4			
		*0	4	100.00%
Hitachi CLA-1	4	*0	4	100.00%
Chicken Alrgn				
AL-1(Q) * Denotes Correct Response				
AG: All Methods	3			
		*0	3	100.00%
Hitachi CLA-1	3	*0	3	100.00%
Chocolate Alrgn				



AL-1(Q)		* Denotes Correct Response		
AG: All Methods	3			
		*0	3	100.00%
Hitachi CLA-1	3	*0	3	100.00%
AL-1(Q)		Cladosporium Alrgn		
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	5			
		*0	2	40.00%
		*1	1	20.00%
		*1/0	2	40.00%
Hitachi CLA-1	5	*0	2	40.00%
Hitachi CLA-1		*1	1	20.00%
Hitachi CLA-1		*1/0	2	40.00%
AL-1(Q)		Cocklebur Alrgn		
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	5			
		*0	4	80.00%
		*1/0	1	20.00%
Hitachi CLA-1	5	*0	4	80.00%
Hitachi CLA-1		*1/0	1	20.00%
AL-1(Q)		Oat Alrgn		
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	3			
		*0	1	33.33%
		*1	1	33.33%
		*1/0	1	33.33%
Hitachi CLA-1	3	*0	1	33.33%
Hitachi CLA-1		*1	1	33.33%
Hitachi CLA-1		*1/0	1	33.33%
AL-1(Q)		Corn Alrgn		
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	4			
		*0	2	50.00%
		*1	1	25.00%
		*1/0	1	25.00%
Hitachi CLA-1	4	*0	2	50.00%
Hitachi CLA-1		*1	1	25.00%
Hitachi CLA-1		*1/0	1	25.00%
AL-1(Q)		Cottonwood - Eastern Alrgn		
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	4			
		*0	4	100.00%
Hitachi CLA-1	4	*0	4	100.00%
AL-1(Q)		Garlic Alrgn		
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	3			
		*0	1	33.33%
		*1/0	2	66.67%
Hitachi CLA-1	3	*0	1	33.33%
Hitachi CLA-1		*1/0	2	66.67%
AL-1(Q)		Lambs Quarters Alrgn		
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	5			
		*0	2	40.00%
		*1	1	20.00%
		*1/0	2	40.00%
Hitachi CLA-1	5	*0	2	40.00%
Hitachi CLA-1		*1	1	20.00%
Hitachi CLA-1		*1/0	2	40.00%
AL-1(Q)		Marshelder - Rough Alrgn		
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	4			
		*0	2	50.00%
		*1	1	25.00%
		*1/0	1	25.00%
Hitachi CLA-1	4	*0	2	50.00%
Hitachi CLA-1		*1	1	25.00%
Hitachi CLA-1		*1/0	1	25.00%



		Mite D Farinae Alrqn		
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	6			
		*4	5	83.33%
		0	1	16.67%
Hitachi CLA-1	6	*4	5	83.33%
Hitachi CLA-1		0	1	16.67%
		Mulberry Mix Alrqn		
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	5			
		*0	4	80.00%
		*1/0	1	20.00%
Hitachi CLA-1	5	*0	4	80.00%
Hitachi CLA-1		*1/0	1	20.00%
		Onion Mix Alrqn		
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	3			
		*0	2	66.67%
		*1/0	1	33.33%
Hitachi CLA-1	3	*0	2	66.67%
Hitachi CLA-1		*1/0	1	33.33%
		Orange Alrqn		
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	4			
		*0	1	25.00%
		*1	3	75.00%
Hitachi CLA-1	4	*0	1	25.00%
Hitachi CLA-1		*1	3	75.00%
		Pigweed Alrqn		
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	5			
		*0	3	60.00%
		*1/0	2	40.00%
Hitachi CLA-1	5	*0	3	60.00%
Hitachi CLA-1		*1/0	2	40.00%
		Pine Mix Alrqn		
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	5			
		*0	5	100.00%
Hitachi CLA-1	5	*0	5	100.00%
		Potato Alrqn		
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	3			
		*0	1	33.33%
		*1	1	33.33%
		*1/0	1	33.33%
Hitachi CLA-1	3	*0	1	33.33%
Hitachi CLA-1		*1	1	33.33%
Hitachi CLA-1		*1/0	1	33.33%
		Privet Alrqn		
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	5			
		*0	5	100.00%
Hitachi CLA-1	5	*0	5	100.00%
		Ragweed - Short Alrqn		
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	5			
		*0	1	20.00%
		*1	2	40.00%
		*1/0	2	40.00%
Hitachi CLA-1	5	*0	1	20.00%
Hitachi CLA-1		*1	2	40.00%
Hitachi CLA-1		*1/0	2	40.00%
		Sheep Sorrel Alrqn		
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	5			
		*1	1	20.00%



		*2	3	60.00%
		1/0	1	20.00%
Hitachi CLA-1	5	*1	1	20.00%
Hitachi CLA-1		*2	3	60.00%
Hitachi CLA-1		1/0	1	20.00%
Sycamore - American Alrgn				
AL-1(Q)				
* Denotes Correct Response				
AG: All Methods	5			
		*0	5	100.00%
Hitachi CLA-1	5	*0	5	100.00%
Tomato Alrgn				
AL-1(Q)				
* Denotes Correct Response				
AG: All Methods	4			
		*0	3	75.00%
		*1	1	25.00%
Hitachi CLA-1	4	*0	3	75.00%
Hitachi CLA-1		*1	1	25.00%
Tuna Alrgn				
AL-1(Q)				
* Denotes Correct Response				
AG: All Methods	3			
		*0	3	100.00%
Hitachi CLA-1	3	*0	3	100.00%
Vegetable Mix Alrgn				
AL-1(Q)				
* Denotes Correct Response				
AG: All Methods	3			
		*0	2	66.67%
		*1/0	1	33.33%
Hitachi CLA-1	3	*0	2	66.67%
Hitachi CLA-1		*1/0	1	33.33%
Walnut/Hickory/Pecan Mix Alrgn				
AL-1(Q)				
* Denotes Correct Response				
AG: All Methods	5			
		*0	3	60.00%
		*1/0	2	40.00%
Hitachi CLA-1	5	*0	3	60.00%
Hitachi CLA-1		*1/0	2	40.00%
Waterhemp Alrgn				
AL-1(Q)				
* Denotes Correct Response				
AG: All Methods	3			
		*2	2	66.67%
		*3	1	33.33%
Hitachi CLA-1	3	*2	2	66.67%
Hitachi CLA-1		*3	1	33.33%
Yeast - Bakers Alrgn				
AL-1(Q)				
* Denotes Correct Response				
AG: All Methods	4			
		*0	3	75.00%
		*1	1	25.00%
Hitachi CLA-1	4	*0	3	75.00%
Hitachi CLA-1		*1	1	25.00%
Acacia Alrgn				
AL-1(Q)				
* Denotes Correct Response				
AG: All Methods	2			
		*0	2	100.00%
Hitachi CLA-1	2	*0	2	100.00%
Bahia Grass Alrgn				
AL-1(Q)				
* Denotes Correct Response				
AG: All Methods	2			
		*1	1	50.00%
		*2	1	50.00%
Hitachi CLA-1	2	*1	1	50.00%
Hitachi CLA-1		*2	1	50.00%
Beech, American Alrgn				
AL-1(Q)				
* Denotes Correct Response				
AG: All Methods	2			
		*0	2	100.00%
Hitachi CLA-1	2	*0	2	100.00%



		Candida Alrgn		
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	2			
		*0	2	100.00%
Hitachi CLA-1	2	*0	2	100.00%
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	2			
		*4	2	100.00%
Hitachi CLA-1	2	*4	2	100.00%
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	3			
		*1/0	3	100.00%
Hitachi CLA-1	3	*1/0	3	100.00%
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	2			
		*0	2	100.00%
Hitachi CLA-1	2	*0	2	100.00%
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	2			
		*1/0	1	50.00%
		*2	1	50.00%
Hitachi CLA-1	2	*1/0	1	50.00%
Hitachi CLA-1		*2	1	50.00%
AL-1(Q)		* Denotes Correct Response		
AG: All Methods	2			
		*0	1	50.00%
		*1/0	1	50.00%
Hitachi CLA-1	2	*0	1	50.00%
Hitachi CLA-1		*1/0	1	50.00%
683 - ASO/CRP;759 - Complete Immunology;729 - AFC Immunology		Labs	Response	Number
AS-1		ASO (Qual) (International)		
AG: All Methods		* Denotes Correct Response		
	1			
		*Negative	1	100.00%
Wampole ColorCard ASO	1	*Negative	1	100.00%
AS-2		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Wampole ColorCard ASO	1	*Positive	1	100.00%
AS-3		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Wampole ColorCard ASO	1	*Positive	1	100.00%
AS-4		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Wampole ColorCard ASO	1	*Positive	1	100.00%
AS-5		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Wampole ColorCard ASO	1	*Negative	1	100.00%
782 - Complete Bacteriology		Labs	Response	Number
BA-2		Blood Culture		
AG: All Methods		* Denotes Correct Response		
	1			
		*Streptococcus pneumoniae	1	100.00%
Cultures	1	*Streptococcus pneumoniae	1	100.00%
BA-3		Sputum Culture		
AG: All Methods		* Denotes Correct Response		
	1			
		*Haemophilus species	1	100.00%



Category	Labs	Response	Number	Percent
Cultures	1	*Haemophilus species	1	100.00%
Wound Culture				
BA-1		* Denotes Correct Response		
AG: All Methods	3			
		*Staphylococcus aureus	2	66.67%
		*Staphylococcus epidermidis	1	33.33%
		*Staphylococcus, coagulase positive	1	33.33%
		No Secondary Organism Present	2	66.67%
Cultures	3	*Staphylococcus aureus	2	66.67%
Cultures		*Staphylococcus epidermidis	1	33.33%
Cultures		*Staphylococcus, coagulase positive	1	33.33%
Cultures		No Secondary Organism Present	2	66.67%
650 - Blood Bank;736 - ABO/Rh Only	Labs	Response	Number	Percent
ABO Group				
BB-1		* Denotes Correct Response		
AG: All Methods	1			
		*Group A	1	100.00%
Blood Bank Testing	1	*Group A	1	100.00%
BB-2		* Denotes Correct Response		
AG: All Methods	1			
		*Group O	1	100.00%
Blood Bank Testing	1	*Group O	1	100.00%
BB-3		* Denotes Correct Response		
AG: All Methods	1			
		*Group A	1	100.00%
Blood Bank Testing	1	*Group A	1	100.00%
BB-4		* Denotes Correct Response		
AG: All Methods	1			
		*Group O	1	100.00%
Blood Bank Testing	1	*Group O	1	100.00%
BB-5		* Denotes Correct Response		
AG: All Methods	1			
		*Group B	1	100.00%
Blood Bank Testing	1	*Group B	1	100.00%
Rh Type				
BB-1		* Denotes Correct Response		
AG: All Methods	6			
		*Rh Positive	6	100.00%
Blood Bank Testing	6	*Rh Positive	6	100.00%
BB-2		* Denotes Correct Response		
AG: All Methods	6			
		*Rh Negative	5	83.33%
		Rh Positive	1	16.67%
Blood Bank Testing	6	*Rh Negative	5	83.33%
Blood Bank Testing		Rh Positive	1	16.67%
BB-3		* Denotes Correct Response		
AG: All Methods	6			
		*Rh Positive	5	83.33%
		Rh Negative	1	16.67%
Blood Bank Testing	6	*Rh Positive	5	83.33%
Blood Bank Testing		Rh Negative	1	16.67%
BB-4		* Denotes Correct Response		
AG: All Methods	6			
		*Rh Negative	5	83.33%
		Rh Positive	1	16.67%
Blood Bank Testing	6	*Rh Negative	5	83.33%
Blood Bank Testing		Rh Positive	1	16.67%
BB-5		* Denotes Correct Response		
AG: All Methods	6			
		*Rh Positive	5	83.33%
		Rh Negative	1	16.67%
Blood Bank Testing	6	*Rh Positive	5	83.33%
Blood Bank Testing		Rh Negative	1	16.67%
975 - Bacterial Vaginosis	Labs	Response	Number	Percent
Bacterial Vaginosa (Sialidase)				
BV-1		* Denotes Correct Response		
AG: All Methods	73			
		*Positive	59	80.82%



		Negative	14	19.18%
Sekisui Diagnostics OSOM BVBlue	73	*Positive	59	80.82%
Sekisui Diagnostics OSOM BVBlue		Negative	14	19.18%
BV-2		* Denotes Correct Response		
AG: All Methods	73			
		*Positive	58	79.45%
		Negative	15	20.55%
Sekisui Diagnostics OSOM BVBlue	73	*Positive	58	79.45%
Sekisui Diagnostics OSOM BVBlue		Negative	15	20.55%
661 - Urine Colony Count;784 - Presump ID/Colony Count - Urine;785 - Presump ID/Colony Count-Urine & Strep Combo;775 - Urine Culture;778 - Urine & Throat Cultures;779 - Urine & Genital Cultures;781 - Urine, Throat & Genital Cultures;782 - Complete Bacteri	Labs	Response	Number	Percent
		Colony Count		
CC-1		* Denotes Correct Response		
AG: All Methods	32			
		*Greater than 100,000	32	100.00%
Calibrated Loop - Manual	9	*Greater than 100,000	9	100.00%
Uri-Chek (Troy Bio)	5	*Greater than 100,000	5	100.00%
Uricult (Orion)	16	*Greater than 100,000	16	100.00%
Uri-Three (Culture Kits)	1	*Greater than 100,000	1	100.00%
Uricult (Lifesign)	1	*Greater than 100,000	1	100.00%
784 - Presump ID/Colony Count - Urine;785 - Presump ID/Colony Count-Urine & Strep Combo	Labs	Response	Number	Percent
		Colony Count		
CC-1		* Denotes Correct Response		
AG: All Methods	32			
		*Greater than 100,000	32	100.00%
Calibrated Loop - Manual	9	*Greater than 100,000	9	100.00%
Uri-Chek (Troy Bio)	5	*Greater than 100,000	5	100.00%
Uricult (Orion)	16	*Greater than 100,000	16	100.00%
Uri-Three (Culture Kits)	1	*Greater than 100,000	1	100.00%
Uricult (Lifesign)	1	*Greater than 100,000	1	100.00%
CC-2		* Denotes Correct Response		
AG: All Methods	6			
		*10,000 - 100,000	1	16.67%
		*Less than 10,000	5	83.33%
Uricult (Orion)	1	*10,000 - 100,000	1	100.00%
Calibrated Loop - Manual	4	*Less than 10,000	4	100.00%
Uri-Chek (Troy Bio)	1	*Less than 10,000	1	100.00%
CC-3		* Denotes Correct Response		
AG: All Methods	6			
		*10,000 - 100,000	2	33.33%
		*Greater than 100,000	4	66.67%
Calibrated Loop - Manual	4	*10,000 - 100,000	2	50.00%
Calibrated Loop - Manual		*Greater than 100,000	2	50.00%
Uri-Chek (Troy Bio)	1	*Greater than 100,000	1	100.00%
Uricult (Orion)	1	*Greater than 100,000	1	100.00%
CC-4		* Denotes Correct Response		
AG: All Methods	6			
		*10,000 - 100,000	4	66.67%
		Less than 10,000	2	33.33%
Calibrated Loop - Manual	4	*10,000 - 100,000	3	75.00%
Uricult (Orion)	1	*10,000 - 100,000	1	100.00%
Calibrated Loop - Manual	4	Less than 10,000	1	25.00%
Uri-Chek (Troy Bio)	1	Less than 10,000	1	100.00%
CC-5		* Denotes Correct Response		
AG: All Methods	6			
		*10,000 - 100,000	1	16.67%
		Greater than 100,000	1	16.67%
		Less than 10,000	3	50.00%
		No growth	1	16.67%
Calibrated Loop - Manual	4	*10,000 - 100,000	1	25.00%
Calibrated Loop - Manual		Greater than 100,000	1	25.00%
Calibrated Loop - Manual		Less than 10,000	1	25.00%
Uri-Chek (Troy Bio)	1	Less than 10,000	1	100.00%
Uricult (Orion)	1	Less than 10,000	1	100.00%
Calibrated Loop - Manual	4	No growth	1	25.00%



Urine Presump Id			
CC-1		* Denotes Correct Response	
AG: All Methods	3		
		*Presumptive Enterobacter sp.	1 33.33%
		*Presumptive Klebsiella sp.	2 66.67%
Calibrated Loop - Manual	2	*Presumptive Enterobacter sp.	1 50.00%
Calibrated Loop - Manual		*Presumptive Klebsiella sp.	1 50.00%
Uricult (Orion)	1	*Presumptive Klebsiella sp.	1 100.00%
CC-2		* Denotes Correct Response	
AG: All Methods	3		
		*Presumptive E. Coli	3 100.00%
Calibrated Loop - Manual	2	*Presumptive E. Coli	2 100.00%
Uricult (Orion)	1	*Presumptive E. Coli	1 100.00%
CC-3		* Denotes Correct Response	
AG: All Methods	3		
		*Gram Positive Organism	1 33.33%
		*Presumptive Enterococcus sp.	1 33.33%
		Presumptive Proteus sp.	1 33.33%
Uricult (Orion)	1	*Gram Positive Organism	1 100.00%
Calibrated Loop - Manual	2	*Presumptive Enterococcus sp.	1 50.00%
Calibrated Loop - Manual		Presumptive Proteus sp.	1 50.00%
CC-4		* Denotes Correct Response	
AG: All Methods	3		
		*Presumptive Pseudomonas sp.	2 66.67%
		Presumptive Proteus sp.	1 33.33%
Calibrated Loop - Manual	2	*Presumptive Pseudomonas sp.	2 100.00%
Uricult (Orion)	1	Presumptive Proteus sp.	1 100.00%
CC-5		* Denotes Correct Response	
AG: All Methods	3		
		*Presumptive Proteus sp.	1 33.33%
		Gram Positive Organism	1 33.33%
		Presumptive Staphylococcus sp.	1 33.33%
Calibrated Loop - Manual	2	*Presumptive Proteus sp.	1 50.00%
Uricult (Orion)	1	Gram Positive Organism	1 100.00%
Calibrated Loop - Manual	2	Presumptive Staphylococcus sp.	1 50.00%
Susceptibility Amikacin			
CC-1		* Denotes Correct Response	
AG: All Methods	2		
		*Sensitive	2 100.00%
Susceptibility	2	*Sensitive	2 100.00%
Susceptibility Amox/Clavulanate			
CC-1		* Denotes Correct Response	
AG: All Methods	6		
		*Resistant	6 100.00%
Susceptibility	6	*Resistant	6 100.00%
Susceptibility Ampicillin			
CC-1		* Denotes Correct Response	
AG: All Methods	8		
		*Resistant	7 87.50%
		Intermediate	1 12.50%
Susceptibility	8	*Resistant	7 87.50%
Susceptibility		Intermediate	1 12.50%
Susceptibility Aztreonam			
CC-1		* Denotes Correct Response	
AG: All Methods	1		
		*Sensitive	1 100.00%
Susceptibility	1	*Sensitive	1 100.00%
Susceptibility Cefazolin			
CC-1		* Denotes Correct Response	
AG: All Methods	2		
		*Resistant	2 100.00%
Susceptibility	2	*Resistant	2 100.00%
Susceptibility Cefepime			
CC-1		* Denotes Correct Response	
AG: All Methods	3		
		*Sensitive	3 100.00%
Susceptibility	3	*Sensitive	3 100.00%



Susceptibility Cefotaxime				
CC-1		* Denotes Correct Response		
AG: All Methods	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
Susceptibility Cefpodoxime				
CC-1		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
Susceptibility Cefazidime				
CC-1		* Denotes Correct Response		
AG: All Methods	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
Susceptibility Ceftriaxone				
CC-1		* Denotes Correct Response		
AG: All Methods	6			
		*Sensitive	6	100.00%
Susceptibility	6	*Sensitive	6	100.00%
Susceptibility Cephalothin				
CC-1		* Denotes Correct Response		
AG: All Methods	6			
		*Resistant	6	100.00%
Susceptibility	6	*Resistant	6	100.00%
Susceptibility Ciprofloxacin				
CC-1		* Denotes Correct Response		
AG: All Methods	10			
		*Sensitive	10	100.00%
Susceptibility	10	*Sensitive	10	100.00%
Susceptibility Gentamicin				
CC-1		* Denotes Correct Response		
AG: All Methods	6			
		*Sensitive	6	100.00%
Susceptibility	6	*Sensitive	6	100.00%
Susceptibility Imipenem				
CC-1		* Denotes Correct Response		
AG: All Methods	4			
		*Sensitive	4	100.00%
Susceptibility	4	*Sensitive	4	100.00%
Susceptibility Levofloxacin				
CC-1		* Denotes Correct Response		
AG: All Methods	6			
		*Sensitive	6	100.00%
Susceptibility	6	*Sensitive	6	100.00%
Susceptibility Norfloxacin/Noroxin				
CC-1		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
Susceptibility Piperacillin/Tazobactam				
CC-1		* Denotes Correct Response		
AG: All Methods	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
Susceptibility Sulfamethoxazole				
CC-1		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
Susceptibility Tetracycline				
CC-1		* Denotes Correct Response		
AG: All Methods	6			
		*Intermediate	2	33.33%
		*Sensitive	4	66.67%
Susceptibility	6	*Intermediate	2	33.33%
Susceptibility		*Sensitive	4	66.67%



Susceptibility Ticarcillin/Clavulanate				
CC-1		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
Susceptibility Tobramycin				
CC-1		* Denotes Correct Response		
AG: All Methods	4			
		*Sensitive	4	100.00%
Susceptibility	4	*Sensitive	4	100.00%
Susceptibility Trimethoprim				
CC-1		* Denotes Correct Response		
AG: All Methods	2			
		*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
Susceptibility Trimeth./Sulfameth.				
CC-1		* Denotes Correct Response		
AG: All Methods	8			
		*Sensitive	8	100.00%
Susceptibility	8	*Sensitive	8	100.00%
Susceptibility Ertapenem				
CC-1		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
714 - Complete Cardiac Markers;715 - Basic Cardiac Markers;720 - Cardiac Markers I;740 - D-Dimer	Labs	Response	Number	Percent
Myoglobin (Qualitative)				
CDM-1		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI	2			
		*Positive	2	100.00%
CDM-2		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI	2			
		*Positive	2	100.00%
CDM-3		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI	2			
		*Negative	2	100.00%
CDM-4		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI	2			
		*Positive	2	100.00%
CDM-5		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI	2			
		*Positive	2	100.00%
CK-MB (Qualitative)				
CDM-1		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI	2			
		*Positive	2	100.00%
CDM-2		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI	2			
		*Positive	2	100.00%
CDM-3		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI	2			
		*Negative	1	50.00%
		Positive	1	50.00%
CDM-4		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI	2			
		*Positive	2	100.00%
CDM-5		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI	2			
		*Positive	2	100.00%
Troponin I (Qualitative)				
CDM-1		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI	2			
		*Positive	2	100.00%
CDM-2		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI	2			
		*Positive	2	100.00%



CDM-3		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI	2			
		*Negative	1	50.00%
		Positive	1	50.00%
CDM-4		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI	2			
		*Positive	2	100.00%
CDM-5		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI	2			
		*Positive	2	100.00%
612 - MultiChem 5;616 - MultiChem 10;617 - MultiChem 15;618 - MultiChem 20;619 - MultiChem 21+;676 - Endocrinology	Labs	Response	Number	Percent
		TSH (Semi-Quant)		
CH-1		* Denotes Correct Response		
PG: Biomedical Thyro Test	2			
		*≤ 5	1	50.00%
		>5	1	50.00%
CH-2		* Denotes Correct Response		
PG: Biomedical Thyro Test	1			
		*>5	1	100.00%
CH-3		* Denotes Correct Response		
PG: Biomedical Thyro Test	1			
		*>5	1	100.00%
CH-4		* Denotes Correct Response		
PG: Biomedical Thyro Test	1			
		*>5	1	100.00%
CH-5		* Denotes Correct Response		
PG: Biomedical Thyro Test	2			
		*≤ 5	1	50.00%
		>5	1	50.00%
639 - Complete Urinalysis Package;638 - Urine Sediment;641 - Clinical Microscopy;601 - Waived/PPM Tests II;609 - Waived/PPM Tests I;646 - Clinical Microscopy Extra;628 - AFC Waived/PPM	Labs	Response	Number	Percent
		Urine Sediment Id		
CM-1		* Denotes Correct Response		
AG: All Methods	718			
		*Cystine crystal	640	89.14%
		Ammonium magnesium phosphate (Triple phosphate)	2	0.28%
		Calcium carbonate	1	0.14%
		Calcium oxalate crystal	9	1.25%
		Calcium phosphate crystal	1	0.14%
		Cystine crystal	640	89.14%
		Fiber (exogenous)/fecal contamination	1	0.14%
		Hippuric acid	4	0.56%
		Hyaline cast	1	0.14%
		Ident Unknown - Referred	2	0.28%
		Mucous strands	11	1.53%
		Renal tubular epithelial (RTE)	1	0.14%
		Squamous epithelial cell	4	0.56%
		Sulfonamide crystal	1	0.14%
		Transitional epithelial cell (urothelial)	1	0.14%
		Uric acid crystal	39	5.43%
		Yeast/fungi	1	0.14%
Microscopy	718	*Cystine crystal	640	89.14%
Microscopy		Ammonium magnesium phosphate (Triple phosphate)	2	0.28%
Microscopy		Calcium carbonate	1	0.14%
Microscopy		Calcium oxalate crystal	9	1.25%
Microscopy		Calcium phosphate crystal	1	0.14%
Microscopy		Cystine crystal	640	89.14%
Microscopy		Fiber (exogenous)/fecal contamination	1	0.14%
Microscopy		Hippuric acid	4	0.56%
Microscopy		Hyaline cast	1	0.14%
Microscopy		Ident Unknown - Referred	2	0.28%
Microscopy		Mucous strands	11	1.53%
Microscopy		Renal tubular epithelial (RTE)	1	0.14%
Microscopy		Squamous epithelial cell	4	0.56%
Microscopy		Sulfonamide crystal	1	0.14%



Microscopy		Transitional epithelial cell (urothelial)	1	0.14%
Microscopy		Uric acid crystal	39	5.43%
Microscopy		Yeast/fungi	1	0.14%
CM-2		* Denotes Correct Response		
AG: All Methods	720			
		*Mucous strands	564	78.33%
		Bacteria	2	0.28%
		Bilirubin crystal	1	0.14%
		Cholesterol crystal	1	0.14%
		Clue cell	1	0.14%
		Cystine crystal	10	1.39%
		Fatty (fat globules) cast	1	0.14%
		Fiber (exogenous)/fecal contamination	41	5.69%
		Hyaline cast	5	0.69%
		Ident Unknown - Referred	3	0.42%
		Renal tubular epithelial (RTE)	1	0.14%
		Squamous epithelial cell	7	0.97%
		Sulfonamide crystal	1	0.14%
		Tyrosine crystal	1	0.14%
		Uric acid crystal	1	0.14%
		White blood cell (WBC) (Leukocyte)	1	0.14%
		Yeast/fungi	77	10.69%
		Yeast/fungi present	2	0.28%
Microscopy	720	*Mucous strands	564	78.33%
Microscopy		Bacteria	2	0.28%
Microscopy		Bilirubin crystal	1	0.14%
Microscopy		Cholesterol crystal	1	0.14%
Microscopy		Clue cell	1	0.14%
Microscopy		Cystine crystal	10	1.39%
Microscopy		Fatty (fat globules) cast	1	0.14%
Microscopy		Fiber (exogenous)/fecal contamination	41	5.69%
Microscopy		Hyaline cast	5	0.69%
Microscopy		Ident Unknown - Referred	3	0.42%
Microscopy		Renal tubular epithelial (RTE)	1	0.14%
Microscopy		Squamous epithelial cell	7	0.97%
Microscopy		Sulfonamide crystal	1	0.14%
Microscopy		Tyrosine crystal	1	0.14%
Microscopy		Uric acid crystal	1	0.14%
Microscopy		White blood cell (WBC) (Leukocyte)	1	0.14%
Microscopy		Yeast/fungi	77	10.69%
Microscopy		Yeast/fungi present	2	0.28%
CM-3		* Denotes Correct Response		
AG: All Methods	720			
		*Spermatozoa	688	95.56%
		Ammonium magnesium phosphate (Triple phosphate)	1	0.14%
		Bacteria	1	0.14%
		Ident Unknown - Referred	1	0.14%
		Mucous strands	1	0.14%
		Red blood cell (RBC) (Erythrocyte)	28	3.89%
Microscopy	720	*Spermatozoa	688	95.56%
Microscopy		Ammonium magnesium phosphate (Triple phosphate)	1	0.14%
Microscopy		Bacteria	1	0.14%
Microscopy		Ident Unknown - Referred	1	0.14%
Microscopy		Mucous strands	1	0.14%
Microscopy		Red blood cell (RBC) (Erythrocyte)	28	3.89%
CM-4		* Denotes Correct Response		
AG: All Methods	720			
		*Red blood cell (RBC) (Erythrocyte)	636	88.33%
		Ammonium magnesium phosphate (Triple phosphate)	1	0.14%
		Amorphous urates/phosphates	1	0.14%
		Bacteria	2	0.28%
		Calcium carbonate	1	0.14%
		Calcium oxalate crystal	1	0.14%
		Fat droplets or globules	1	0.14%
		Ident Unknown - Referred	3	0.42%
		Pollen grain	2	0.28%
		RBC/Blood/Hgb cast	6	0.83%
		Renal tubular epithelial (RTE)	2	0.28%
		Spermatozoa	29	4.03%



		Starch granule	1	0.14%
		Starch granules	23	3.19%
		Uric acid crystal	1	0.14%
		White blood cell (WBC) (Leukocyte)	6	0.83%
		Yeast/fungi	4	0.56%
Microscopy	720	*Red blood cell (RBC) (Erythrocyte)	636	88.33%
Microscopy		Ammonium magnesium phosphate (Triple phosphate)	1	0.14%
Microscopy		Amorphous urates/phosphates	1	0.14%
Microscopy		Bacteria	2	0.28%
Microscopy		Calcium carbonate	1	0.14%
Microscopy		Calcium oxalate crystal	1	0.14%
Microscopy		Fat droplets or globules	1	0.14%
Microscopy		Ident Unknown - Referred	3	0.42%
Microscopy		Pollen grain	2	0.28%
Microscopy		RBC/Blood/Hgb cast	6	0.83%
Microscopy		Renal tubular epithelial (RTE)	2	0.28%
Microscopy		Spermatozoa	29	4.03%
Microscopy		Starch granule	1	0.14%
Microscopy		Starch granules	23	3.19%
Microscopy		Uric acid crystal	1	0.14%
Microscopy		White blood cell (WBC) (Leukocyte)	6	0.83%
Microscopy		Yeast/fungi	4	0.56%
		Vaginal Wet Prep		
CM-5		* Denotes Correct Response		
AG: All Methods	960			
		*Yeast/fungi present	747	77.81%
		Bacteria present	10	1.04%
		Clue cell	4	0.42%
		Red blood cell (RBC) (Erythrocyte)	3	0.31%
		Red Blood Cell (RBC) (Erythrocyte)	117	12.19%
		Squamous epithelial cell	3	0.31%
		Starch granule	4	0.42%
		Trichomonas vaginalis	1	0.10%
		White blood cell (WBC) (Leukocyte)	1	0.10%
		White Blood Cell (WBC) (Leukocyte)	63	6.56%
		Yeast/fungi absent	7	0.73%
Microscopy	960	*Yeast/fungi present	747	77.81%
Microscopy		Bacteria present	10	1.04%
Microscopy		Clue cell	4	0.42%
Microscopy		Red blood cell (RBC) (Erythrocyte)	3	0.31%
Microscopy		Red Blood Cell (RBC) (Erythrocyte)	117	12.19%
Microscopy		Squamous epithelial cell	3	0.31%
Microscopy		Starch granule	4	0.42%
Microscopy		Trichomonas vaginalis	1	0.10%
Microscopy		White blood cell (WBC) (Leukocyte)	1	0.10%
Microscopy		White Blood Cell (WBC) (Leukocyte)	63	6.56%
Microscopy		Yeast/fungi absent	7	0.73%
		CM-6		
AG: All Methods		* Denotes Correct Response		
	960			
		*Squamous epithelial cell	878	91.46%
		Bacteria absent	2	0.21%
		Bacteria present	2	0.21%
		Clue cell	66	6.88%
		White Blood Cell (WBC) (Leukocyte)	5	0.52%
		Yeast/fungi present	7	0.73%
Microscopy	960	*Squamous epithelial cell	878	91.46%
Microscopy		Bacteria absent	2	0.21%
Microscopy		Bacteria present	2	0.21%
Microscopy		Clue cell	66	6.88%
Microscopy		White Blood Cell (WBC) (Leukocyte)	5	0.52%
Microscopy		Yeast/fungi present	7	0.73%
		Skin Scraping		
CM-7		* Denotes Correct Response		
AG: All Methods	624			
		*No scabies present	18	2.88%
		*Yeast/fungi absent	534	85.58%
		Scabies present	14	2.24%
		Squamous epithelial cell	1	0.16%
		Yeast/fungi absent	1	0.16%



		Yeast/fungi present	56	8.97%
Microscopy	624	*No scabies present	18	2.88%
Microscopy		*Yeast/fungi absent	534	85.58%
Microscopy		Scabies present	14	2.24%
Microscopy		Squamous epithelial cell	1	0.16%
Microscopy		Yeast/fungi absent	1	0.16%
Microscopy		Yeast/fungi present	56	8.97%
Pinworm Prep				
CM-8		* Denotes Correct Response		
AG: All Methods	329			
		*Pinworm/pinworm eggs present	327	99.39%
		No pinworm/pinworm eggs present	2	0.61%
Microscopy	329	*Pinworm/pinworm eggs present	327	99.39%
Microscopy		No pinworm/pinworm eggs present	2	0.61%
Fern Test				
CM-9		* Denotes Correct Response		
AG: All Methods	405			
		*Positive Fern Test	405	100.00%
Microscopy	405	*Positive Fern Test	405	100.00%
683 - ASO/CRP;684 - CRP (Quant & Qual)	Labs	Response	Number	Percent
CRP Qualitative				
CR-1		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Wampole	1	*Positive	1	100.00%
684 - CRP (Quant & Qual)	Labs	Response	Number	Percent
CRP Qualitative				
CR-1		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Wampole	1	*Positive	1	100.00%
CR-2		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Wampole	1	*Positive	1	100.00%
774 - GenProbe GC/Chlamydia;794 - GC & Chlamydia Antigen Detection (5 specimens);795 - GC & Chlamydia Antigen Detection (2 specimens)	Labs	Response	Number	Percent
Chlamydia - Non Culture Tech				
CY-1		* Denotes Correct Response		
AG: All Methods	8			
		*Positive	8	100.00%
BD Probe Tec ET	1	*Positive	1	100.00%
Gen-Probe	1	*Positive	1	100.00%
Quidel QuickVue Chlamydia	2	*Positive	2	100.00%
Cepheid GeneXpert	3	*Positive	3	100.00%
BD Viper XTR CT/GC	1	*Positive	1	100.00%
CY-2		* Denotes Correct Response		
AG: All Methods	8			
		*Negative	8	100.00%
BD Probe Tec ET	1	*Negative	1	100.00%
Gen-Probe	1	*Negative	1	100.00%
Quidel QuickVue Chlamydia	2	*Negative	2	100.00%
Cepheid GeneXpert	3	*Negative	3	100.00%
BD Viper XTR CT/GC	1	*Negative	1	100.00%
CY-3		* Denotes Correct Response		
AG: All Methods	8			
		*Positive	8	100.00%
BD Probe Tec ET	1	*Positive	1	100.00%
Gen-Probe	1	*Positive	1	100.00%
Quidel QuickVue Chlamydia	2	*Positive	2	100.00%
Cepheid GeneXpert	3	*Positive	3	100.00%
BD Viper XTR CT/GC	1	*Positive	1	100.00%
CY-4		* Denotes Correct Response		
AG: All Methods	8			
		*Negative	7	87.50%
		Positive	1	12.50%
BD Probe Tec ET	1	*Negative	1	100.00%



Gen-Probe	1	*Negative	1	100.00%
Quidel QuickVue Chlamydia	2	*Negative	2	100.00%
Cepheid GeneXpert	3	*Negative	2	66.67%
BD Viper XTR CT/GC	1	*Negative	1	100.00%
Cepheid GeneXpert	3	Positive	1	33.33%
CY-5		* Denotes Correct Response		
AG: All Methods	8			
		*Negative	8	100.00%
BD Probe Tec ET	1	*Negative	1	100.00%
Gen-Probe	1	*Negative	1	100.00%
Quidel QuickVue Chlamydia	2	*Negative	2	100.00%
Cepheid GeneXpert	3	*Negative	3	100.00%
BD Viper XTR CT/GC	1	*Negative	1	100.00%
		GC - Non Culture Tech		
CY-1		* Denotes Correct Response		
AG: All Methods	6			
		*Negative	6	100.00%
BD Probe Tec ET	1	*Negative	1	100.00%
Gen-Probe	1	*Negative	1	100.00%
Cepheid GeneXpert	3	*Negative	3	100.00%
BD Viper XTR CT/GC	1	*Negative	1	100.00%
CY-2		* Denotes Correct Response		
AG: All Methods	6			
		*Positive	5	83.33%
		Negative	1	16.67%
BD Probe Tec ET	1	*Positive	1	100.00%
Cepheid GeneXpert	3	*Positive	3	100.00%
BD Viper XTR CT/GC	1	*Positive	1	100.00%
Gen-Probe	1	Negative	1	100.00%
CY-3		* Denotes Correct Response		
AG: All Methods	6			
		*Negative	6	100.00%
BD Probe Tec ET	1	*Negative	1	100.00%
Gen-Probe	1	*Negative	1	100.00%
Cepheid GeneXpert	3	*Negative	3	100.00%
BD Viper XTR CT/GC	1	*Negative	1	100.00%
CY-4		* Denotes Correct Response		
AG: All Methods	6			
		*Negative	6	100.00%
BD Probe Tec ET	1	*Negative	1	100.00%
Gen-Probe	1	*Negative	1	100.00%
Cepheid GeneXpert	3	*Negative	3	100.00%
BD Viper XTR CT/GC	1	*Negative	1	100.00%
CY-5		* Denotes Correct Response		
AG: All Methods	6			
		*Positive	6	100.00%
BD Probe Tec ET	1	*Positive	1	100.00%
Gen-Probe	1	*Positive	1	100.00%
Cepheid GeneXpert	3	*Positive	3	100.00%
BD Viper XTR CT/GC	1	*Positive	1	100.00%
772 - Dermatophyte Screen/Culture	Labs	Response	Number	Percent
		Dermatophyte Screen		
DM-1		* Denotes Correct Response		
AG: All Methods	7			
		*Negative, No Dermatophytes Present	6	85.71%
		Positive, Dermatophytes Present	1	14.29%
Dermatophytes	7	*Negative, No Dermatophytes Present	6	85.71%
Dermatophytes		Positive, Dermatophytes Present	1	14.29%
DM-2		* Denotes Correct Response		
AG: All Methods	7			
		*Negative, No Dermatophytes Present	5	71.43%
		Positive, Dermatophytes Present	2	28.57%
Dermatophytes	7	*Negative, No Dermatophytes Present	5	71.43%
Dermatophytes		Positive, Dermatophytes Present	2	28.57%
DM-3		* Denotes Correct Response		
AG: All Methods	7			
		*Positive, Dermatophytes Present	6	85.71%
		Negative, No Dermatophytes Present	1	14.29%



Dermatophytes	7	*Positive, Dermatophytes Present	6	85.71%
Dermatophytes		Negative, No Dermatophytes Present	1	14.29%
DM-4		* Denotes Correct Response		
AG: All Methods	7			
		*Positive, Dermatophytes Present	4	57.14%
		Negative, No Dermatophytes Present	3	42.86%
Dermatophytes	7	*Positive, Dermatophytes Present	4	57.14%
Dermatophytes		Negative, No Dermatophytes Present	3	42.86%
DM-5		* Denotes Correct Response		
AG: All Methods	7			
		*Negative, No Dermatophytes Present	6	85.71%
		Positive, Dermatophytes Present	1	14.29%
Dermatophytes	7	*Negative, No Dermatophytes Present	6	85.71%
Dermatophytes		Positive, Dermatophytes Present	1	14.29%
777 - Genital Culture:796 - Genital Culture (2 specimens)	Labs	Response	Number	Percent
		Genital Culture		
GC-1		* Denotes Correct Response		
AG: All Methods	3			
		*Negative for GC	2	66.67%
		*No growth, GC selective media	1	33.33%
Cultures	3	*Negative for GC	2	66.67%
Cultures		*No growth, GC selective media	1	33.33%
GC-2		* Denotes Correct Response		
AG: All Methods	3			
		*Presumptive positive for GC, GR(-) diplococci, oxidase positive organism	3	100.00%
Cultures	3	*Presumptive positive for GC, GR(-) diplococci, oxidase positive organism	3	100.00%
GC-3		* Denotes Correct Response		
AG: All Methods	3			
		*No growth, GC selective media	3	100.00%
Cultures	3	*No growth, GC selective media	3	100.00%
GC-4		* Denotes Correct Response		
AG: All Methods	6			
		*Group B Streptococcus not present	1	16.67%
		*Presumptive positive for GC, GR(-) diplococci, oxidase positive organism	4	66.67%
		Negative for GC	1	16.67%
Cultures	6	*Group B Streptococcus not present	1	16.67%
Cultures		*Presumptive positive for GC, GR(-) diplococci, oxidase positive organism	4	66.67%
Cultures		Negative for GC	1	16.67%
GC-5		* Denotes Correct Response		
AG: All Methods	4			
		*Negative for GC	1	25.00%
		*No growth, GC selective media	3	75.00%
Cultures	4	*Negative for GC	1	25.00%
Cultures		*No growth, GC selective media	3	75.00%
668 - Gram Stain	Labs	Response	Number	Percent
		Gram Stain		
GS-1		* Denotes Correct Response		
AG: All Methods	9			
		*Gram Negative	8	88.89%
		Gram Positive	1	11.11%
Gram Stain	9	*Gram Negative	8	88.89%
Gram Stain		Gram Positive	1	11.11%
GS-2		* Denotes Correct Response		
AG: All Methods	9			
		*Gram Positive	8	88.89%
		Gram Negative	1	11.11%
Gram Stain	9	*Gram Positive	8	88.89%
Gram Stain		Gram Negative	1	11.11%
GS-3		* Denotes Correct Response		
AG: All Methods	9			
		*Gram Negative	9	100.00%
Gram Stain	9	*Gram Negative	9	100.00%
GS-4		* Denotes Correct Response		
AG: All Methods	9			
		*Gram Positive	9	100.00%



Gram Stain	9	*Gram Positive	9	100.00%
GS-5		* Denotes Correct Response		
AG: All Methods	9			
		*Gram Positive	8	88.89%
		Gram Negative	1	11.11%
Gram Stain	9	*Gram Positive	8	88.89%
Gram Stain		Gram Negative	1	11.11%
GS-1		Gram Stain Morph		
		* Denotes Correct Response		
AG: All Methods	8			
		*Rods/Bacilli	8	100.00%
Gram Stain	8	*Rods/Bacilli	8	100.00%
GS-2		* Denotes Correct Response		
AG: All Methods	8			
		*Cocci	4	50.00%
		*Cocci in Chains	4	50.00%
Gram Stain	8	*Cocci	4	50.00%
Gram Stain		*Cocci in Chains	4	50.00%
GS-3		* Denotes Correct Response		
AG: All Methods	8			
		*Rods/Bacilli	8	100.00%
Gram Stain	8	*Rods/Bacilli	8	100.00%
GS-4		* Denotes Correct Response		
AG: All Methods	8			
		*Cocci	7	87.50%
		*Cocci in Pairs	1	12.50%
Gram Stain	8	*Cocci	7	87.50%
Gram Stain		*Cocci in Pairs	1	12.50%
GS-5		* Denotes Correct Response		
AG: All Methods	8			
		*Rods/Bacilli	8	100.00%
Gram Stain	8	*Rods/Bacilli	8	100.00%
678 - Serum hCG;759 - Complete Immunology;729 - AFC Immunology	Labs	Response	Number	Percent
		HCG - Qual Serum		
HCG-1		* Denotes Correct Response		
AG: All Methods	93			
		*Negative	93	100.00%
Beck Coulter ICON 25	7	*Negative	7	100.00%
Fisher HC Sure-Vue hCG STAT	2	*Negative	2	100.00%
Sekisui Diagnostics OSOM	3	*Negative	3	100.00%
Henry Schein One Step + (hCG)	5	*Negative	5	100.00%
Immunostics Immuno/hCG	2	*Negative	2	100.00%
Inverness Signify HCG	2	*Negative	2	100.00%
LifeSign Status hCG	2	*Negative	2	100.00%
McKesson Medi-Lab hCG - Urine	1	*Negative	1	100.00%
Quidel QuickVue + hCG	1	*Negative	1	100.00%
Quidel QuickVue One-Step hCG Combo (U/S)	33	*Negative	33	100.00%
Select Diag hCG Combo	3	*Negative	3	100.00%
SP Brand (Cardinal) HCG Combo	10	*Negative	10	100.00%
Wampole Clearview	1	*Negative	1	100.00%
Polymedco Poly stat hCG Serum	1	*Negative	1	100.00%
Stanbio True 20 Plus	1	*Negative	1	100.00%
Consult Diag hCG Combo Cassette	12	*Negative	12	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	1	*Negative	1	100.00%
Medline Pregnancy Combo (serum)	1	*Negative	1	100.00%
Alere hCG Combo Cassette (20/10 mIU/mL)	1	*Negative	1	100.00%
McKesson Medi-Lab hCG - Serum	4	*Negative	4	100.00%
HCG-2		* Denotes Correct Response		
AG: All Methods	93			
		*Positive	93	100.00%
Beck Coulter ICON 25	7	*Positive	7	100.00%
Fisher HC Sure-Vue hCG STAT	2	*Positive	2	100.00%
Sekisui Diagnostics OSOM	3	*Positive	3	100.00%
Henry Schein One Step + (hCG)	5	*Positive	5	100.00%
Immunostics Immuno/hCG	2	*Positive	2	100.00%
Inverness Signify HCG	2	*Positive	2	100.00%
LifeSign Status hCG	2	*Positive	2	100.00%



McKesson Medi-Lab hCG - Urine	1	*Positive	1	100.00%
Quidel QuickVue + hCG	1	*Positive	1	100.00%
Quidel QuickVue One-Step hCG Combo (U/S)	33	*Positive	33	100.00%
Select Diag hCG Combo	3	*Positive	3	100.00%
SP Brand (Cardinal) HCG Combo	10	*Positive	10	100.00%
Wampole Clearview	1	*Positive	1	100.00%
Polymedco Poly stat hCG Serum	1	*Positive	1	100.00%
Stanbio True 20 Plus	1	*Positive	1	100.00%
Consult Diag hCG Combo Cassette	12	*Positive	12	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	1	*Positive	1	100.00%
Medline Pregnancy Combo (serum)	1	*Positive	1	100.00%
Alere hCG Combo Cassette (20/10 mIU/mL)	1	*Positive	1	100.00%
McKesson Medi-Lab hCG - Serum	4	*Positive	4	100.00%
HCG-3		* Denotes Correct Response		
AG: All Methods	93			
		*Negative	93	100.00%
Beck Coulter ICON 25	7	*Negative	7	100.00%
Fisher HC Sure-Vue hCG STAT	2	*Negative	2	100.00%
Sekisui Diagnostics OSOM	3	*Negative	3	100.00%
Henry Schein One Step + (hCG)	5	*Negative	5	100.00%
Immunostics Immuno/hCG	2	*Negative	2	100.00%
Inverness Signify HCG	2	*Negative	2	100.00%
LifeSign Status hCG	2	*Negative	2	100.00%
McKesson Medi-Lab hCG - Urine	1	*Negative	1	100.00%
Quidel QuickVue + hCG	1	*Negative	1	100.00%
Quidel QuickVue One-Step hCG Combo (U/S)	33	*Negative	33	100.00%
Select Diag hCG Combo	3	*Negative	3	100.00%
SP Brand (Cardinal) HCG Combo	10	*Negative	10	100.00%
Wampole Clearview	1	*Negative	1	100.00%
Polymedco Poly stat hCG Serum	1	*Negative	1	100.00%
Stanbio True 20 Plus	1	*Negative	1	100.00%
Consult Diag hCG Combo Cassette	12	*Negative	12	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	1	*Negative	1	100.00%
Medline Pregnancy Combo (serum)	1	*Negative	1	100.00%
Alere hCG Combo Cassette (20/10 mIU/mL)	1	*Negative	1	100.00%
McKesson Medi-Lab hCG - Serum	4	*Negative	4	100.00%
HCG-4		* Denotes Correct Response		
AG: All Methods	93			
		*Positive	93	100.00%
Beck Coulter ICON 25	7	*Positive	7	100.00%
Fisher HC Sure-Vue hCG STAT	2	*Positive	2	100.00%
Sekisui Diagnostics OSOM	3	*Positive	3	100.00%
Henry Schein One Step + (hCG)	5	*Positive	5	100.00%
Immunostics Immuno/hCG	2	*Positive	2	100.00%
Inverness Signify HCG	2	*Positive	2	100.00%
LifeSign Status hCG	2	*Positive	2	100.00%
McKesson Medi-Lab hCG - Urine	1	*Positive	1	100.00%
Quidel QuickVue + hCG	1	*Positive	1	100.00%
Quidel QuickVue One-Step hCG Combo (U/S)	33	*Positive	33	100.00%
Select Diag hCG Combo	3	*Positive	3	100.00%
SP Brand (Cardinal) HCG Combo	10	*Positive	10	100.00%
Wampole Clearview	1	*Positive	1	100.00%
Polymedco Poly stat hCG Serum	1	*Positive	1	100.00%
Stanbio True 20 Plus	1	*Positive	1	100.00%
Consult Diag hCG Combo Cassette	12	*Positive	12	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	1	*Positive	1	100.00%
Medline Pregnancy Combo (serum)	1	*Positive	1	100.00%
Alere hCG Combo Cassette (20/10 mIU/mL)	1	*Positive	1	100.00%
McKesson Medi-Lab hCG - Serum	4	*Positive	4	100.00%
HCG-5		* Denotes Correct Response		
AG: All Methods	93			
		*Positive	92	98.92%
		Negative	1	1.08%
Beck Coulter ICON 25	7	*Positive	7	100.00%
Fisher HC Sure-Vue hCG STAT	2	*Positive	2	100.00%
Sekisui Diagnostics OSOM	3	*Positive	3	100.00%
Henry Schein One Step + (hCG)	5	*Positive	5	100.00%
Immunostics Immuno/hCG	2	*Positive	2	100.00%
Inverness Signify HCG	2	*Positive	2	100.00%



LifeSign Status hCG	2	*Positive	2	100.00%
McKesson Medi-Lab hCG - Urine	1	*Positive	1	100.00%
Quidel QuickVue + hCG	1	*Positive	1	100.00%
Quidel QuickVue One-Step hCG Combo (U/S)	33	*Positive	33	100.00%
Select Diag hCG Combo	3	*Positive	3	100.00%
SP Brand (Cardinal) HCG Combo	10	*Positive	9	90.00%
Wampole Clearview	1	*Positive	1	100.00%
Polymedco Poly stat hCG Serum	1	*Positive	1	100.00%
Stanbio True 20 Plus	1	*Positive	1	100.00%
Consult Diag hCG Combo Cassette	12	*Positive	12	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	1	*Positive	1	100.00%
Medline Pregnancy Combo (serum)	1	*Positive	1	100.00%
Alere hCG Combo Cassette (20/10 mIU/mL)	1	*Positive	1	100.00%
McKesson Medi-Lab hCG - Serum	4	*Positive	4	100.00%
SP Brand (Cardinal) HCG Combo	10	Negative	1	10.00%
HCG - Qual Serum (International)				
HCG-1				
* Denotes Correct Response				
AG: All Methods	3			
		*Negative	3	100.00%
Henry Schein One Step + (hCG)	1	*Negative	1	100.00%
SAS Ultra HCG	2	*Negative	2	100.00%
HCG-2				
* Denotes Correct Response				
AG: All Methods	3			
		*Positive	3	100.00%
Henry Schein One Step + (hCG)	1	*Positive	1	100.00%
SAS Ultra HCG	2	*Positive	2	100.00%
HCG-3				
* Denotes Correct Response				
AG: All Methods	3			
		*Negative	3	100.00%
Henry Schein One Step + (hCG)	1	*Negative	1	100.00%
SAS Ultra HCG	2	*Negative	2	100.00%
HCG-4				
* Denotes Correct Response				
AG: All Methods	3			
		*Positive	3	100.00%
Henry Schein One Step + (hCG)	1	*Positive	1	100.00%
SAS Ultra HCG	2	*Positive	2	100.00%
HCG-5				
* Denotes Correct Response				
AG: All Methods	3			
		*Positive	3	100.00%
Henry Schein One Step + (hCG)	1	*Positive	1	100.00%
SAS Ultra HCG	2	*Positive	2	100.00%
762 - HIV Antibody (Waived);766 - HIV Antibody (5 specimens);750 - Viral Markers	Labs	Response	Number	Percent
HIV Antibody (Waived)				
HIV-1				
* Denotes Correct Response				
AG: All Methods	82			
		*Positive	82	100.00%
Alere Clearview Complete HIV 1/2 (mod)	10	*Positive	10	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	45	*Positive	45	100.00%
Orasure OraQuick Anti-HIV	13	*Positive	13	100.00%
Trinity Biotech Uni-Gold Recombigen HIV	2	*Positive	2	100.00%
Chembio HIV 1/2 STAT-PAK Dipstick	1	*Positive	1	100.00%
Chembio HIV 1/2 STAT-PAK Cassette	8	*Positive	8	100.00%
bioLytical INSTI HIV Antibody	2	*Positive	2	100.00%
HIV Antibody (International)				
HIV-1				
* Denotes Correct Response				
AG: All Methods	4			
		*Positive	4	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Positive	3	100.00%
Intl Chembio HIV 1/2 STAT-PAK Dipstick	1	*Positive	1	100.00%
766 - HIV Antibody (5 specimens);750 - Viral Markers	Labs	Response	Number	Percent
HIV Antibody (Waived)				
HIV-1				
* Denotes Correct Response				
AG: All Methods	82			
		*Positive	82	100.00%
Alere Clearview Complete HIV 1/2 (mod)	10	*Positive	10	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	45	*Positive	45	100.00%
Orasure OraQuick Anti-HIV	13	*Positive	13	100.00%



Trinity Biotech Uni-Gold Recombigen HIV	2	*Positive	2	100.00%
Chembio HIV 1/2 STAT-PAK Dipstick	1	*Positive	1	100.00%
Chembio HIV 1/2 STAT-PAK Cassette	8	*Positive	8	100.00%
bioLytical INSTI HIV Antibody	2	*Positive	2	100.00%
HIV-2		* Denotes Correct Response		
AG: All Methods	7			
		*Positive	7	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	4	*Positive	4	100.00%
Orasure OraQuick Anti-HIV	1	*Positive	1	100.00%
Trinity Biotech Uni-Gold Recombigen HIV	1	*Positive	1	100.00%
bioLytical INSTI HIV Antibody	1	*Positive	1	100.00%
HIV-3		* Denotes Correct Response		
AG: All Methods	7			
		*Negative	7	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	4	*Negative	4	100.00%
Orasure OraQuick Anti-HIV	1	*Negative	1	100.00%
Trinity Biotech Uni-Gold Recombigen HIV	1	*Negative	1	100.00%
bioLytical INSTI HIV Antibody	1	*Negative	1	100.00%
HIV-4		* Denotes Correct Response		
AG: All Methods	7			
		*Positive	7	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	4	*Positive	4	100.00%
Orasure OraQuick Anti-HIV	1	*Positive	1	100.00%
Trinity Biotech Uni-Gold Recombigen HIV	1	*Positive	1	100.00%
bioLytical INSTI HIV Antibody	1	*Positive	1	100.00%
HIV-5		* Denotes Correct Response		
AG: All Methods	7			
		*Negative	7	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	4	*Negative	4	100.00%
Orasure OraQuick Anti-HIV	1	*Negative	1	100.00%
Trinity Biotech Uni-Gold Recombigen HIV	1	*Negative	1	100.00%
bioLytical INSTI HIV Antibody	1	*Negative	1	100.00%
		HIV Antibody (International)		
HIV-1		* Denotes Correct Response		
AG: All Methods	4			
		*Positive	4	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Positive	3	100.00%
Intl Chembio HIV 1/2 STAT-PAK Dipstick	1	*Positive	1	100.00%
HIV-2		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Positive	3	100.00%
HIV-3		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
HIV-4		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Positive	3	100.00%
HIV-5		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
644 - H. pylori:628 - AFC Waived/PPM	Labs	Response	Number	Percent
		H. pylori Ab Detect		
HP-1		* Denotes Correct Response		
AG: All Methods	75			
		*Negative	74	98.67%
		Positive	1	1.33%
Beck Coulter FlexSure	3	*Negative	3	100.00%
Beck Coulter ICON HP	2	*Negative	2	100.00%
Instant Technologies i-Screen	1	*Negative	1	100.00%
McKesson Medi-Lab Perf H. pylori	6	*Negative	6	100.00%
Polymedco Poly stat (WB - Waived)	3	*Negative	3	100.00%
Henry Schein One Step+ (H. pylori)	6	*Negative	6	100.00%
Select Diag H. pylori Cassette	3	*Negative	3	100.00%
Quidel QuickVue H. pylori gII (S/P)	5	*Negative	5	100.00%



Test Name	Number	Response	Number	Percent
Quidel QuickVue H. pylori gII (WB)	11	*Negative	11	100.00%
SP Brand (Cardinal) HP Rapid Test	10	*Negative	10	100.00%
Alere Clearview II S/P (mod)	2	*Negative	2	100.00%
Alere Clearview II WB (waived)	4	*Negative	4	100.00%
LifeSign Status H. pylori	2	*Negative	2	100.00%
Instant-View H. pylori	1	*Negative	1	100.00%
Consult Diag (PSS) H pylori	15	*Negative	15	100.00%
Jant Accutest H pylori	1	Positive	1	100.00%
933 - HIV Antibody/Antigen Combo (2 specimens);934 - HIV Antibody/Antigen Combo (5 specimens)	Labs	Response	Number	Percent
		HIV Antibody (Waived)		
HVAA-1		* Denotes Correct Response		
AG: All Methods	27			
		*Negative	26	96.30%
		Positive	1	3.70%
Alere Determine HIV Ag/Ab	27	*Negative	26	96.30%
Alere Determine HIV Ag/Ab		Positive	1	3.70%
HVAA-2		* Denotes Correct Response		
AG: All Methods	25			
		*Positive	23	92.00%
		Negative	2	8.00%
Alere Determine HIV Ag/Ab	25	*Positive	23	92.00%
Alere Determine HIV Ag/Ab		Negative	2	8.00%
		HIV Antigen (p24) (Waived)		
HVAA-1		* Denotes Correct Response		
AG: All Methods	26			
		*Negative	26	100.00%
Alere Determine HIV Ag/Ab	26	*Negative	26	100.00%
HVAA-2		* Denotes Correct Response		
AG: All Methods	26			
		*Negative	18	69.23%
		Positive	8	30.77%
Alere Determine HIV Ag/Ab	26	*Negative	18	69.23%
Alere Determine HIV Ag/Ab		Positive	8	30.77%
603 - Blood Cell Identification;610 - Regular Hematology;749 - Auto Diff II w-Blood Cell Identification	Labs	Response	Number	Percent
		Blood Cell Identification		
ID-1		* Denotes Correct Response		
AG: All Methods	68			
		*Platelet, normal	67	98.53%
		Platelet, abnormal morphology	1	1.47%
Blood Cell Microscopy	68	*Platelet, normal	67	98.53%
Blood Cell Microscopy		Platelet, abnormal morphology	1	1.47%
ID-2		* Denotes Correct Response		
AG: All Methods	68			
		*Segmented or band neutrophil	60	88.24%
		Hypersegmented neutrophil	8	11.76%
Blood Cell Microscopy	68	*Segmented or band neutrophil	60	88.24%
Blood Cell Microscopy		Hypersegmented neutrophil	8	11.76%
ID-3		* Denotes Correct Response		
AG: All Methods	68			
		*Basket cell/smudge cell	67	98.53%
		Other, not listed	1	1.47%
Blood Cell Microscopy	68	*Basket cell/smudge cell	67	98.53%
Blood Cell Microscopy		Other, not listed	1	1.47%
ID-4		* Denotes Correct Response		
AG: All Methods	68			
		*Red Blood Cell, normal (RBC) (Erythrocyte)	63	92.65%
		Microcyte	5	7.35%
Blood Cell Microscopy	68	*Red Blood Cell, normal (RBC) (Erythrocyte)	63	92.65%
Blood Cell Microscopy		Microcyte	5	7.35%
ID-5		* Denotes Correct Response		
AG: All Methods	67			
		*Lymphocyte	57	85.07%
		Immature/abnormal cell - Referred	1	1.49%
		Lymphoblast	1	1.49%
		Lymphocyte, immature	1	1.49%
		Lymphocyte, reactive (atypical, variant, Downey)	5	7.46%



		Monocyte	1	1.49%
		Prolymphocyte	1	1.49%
Blood Cell Microscopy	67	*Lymphocyte	57	85.07%
Blood Cell Microscopy		Immature/abnormal cell - Referred	1	1.49%
Blood Cell Microscopy		Lymphoblast	1	1.49%
Blood Cell Microscopy		Lymphocyte, immature	1	1.49%
Blood Cell Microscopy		Lymphocyte, reactive (atypical, variant, Downey)	5	7.46%
Blood Cell Microscopy		Monocyte	1	1.49%
Blood Cell Microscopy		Prolymphocyte	1	1.49%
ID-6		* Denotes Correct Response		
AG: All Methods	65			
		*Identification Unknown - Referred	1	1.54%
		*Immature/abnormal cell - Referred	11	16.92%
		*Prolymphocyte	13	20.00%
		Blast cell	14	21.54%
		Lymphoblast	15	23.08%
		Lymphocyte, immature	7	10.77%
		Metamyelocyte (juvenile)	1	1.54%
		Monoblast	2	3.08%
		Myelocyte cell	1	1.54%
Blood Cell Microscopy	65	*Identification Unknown - Referred	1	1.54%
Blood Cell Microscopy		*Immature/abnormal cell - Referred	11	16.92%
Blood Cell Microscopy		*Prolymphocyte	13	20.00%
Blood Cell Microscopy		Blast cell	14	21.54%
Blood Cell Microscopy		Lymphoblast	15	23.08%
Blood Cell Microscopy		Lymphocyte, immature	7	10.77%
Blood Cell Microscopy		Metamyelocyte (juvenile)	1	1.54%
Blood Cell Microscopy		Monoblast	2	3.08%
Blood Cell Microscopy		Myelocyte cell	1	1.54%
		Blood Cell Identification (International)		
ID-1		* Denotes Correct Response		
AG: All Methods	1			
		*Platelet, normal	1	100.00%
Intl Blood Cell Microscopy	1	*Platelet, normal	1	100.00%
ID-2		* Denotes Correct Response		
AG: All Methods	1			
		*Segmented or band neutrophil	1	100.00%
Intl Blood Cell Microscopy	1	*Segmented or band neutrophil	1	100.00%
ID-3		* Denotes Correct Response		
AG: All Methods	1			
		*Basket cell/smudge cell	1	100.00%
Intl Blood Cell Microscopy	1	*Basket cell/smudge cell	1	100.00%
ID-4		* Denotes Correct Response		
AG: All Methods	1			
		*Red Blood Cell, normal (RBC) (Erythrocyte)	1	100.00%
Intl Blood Cell Microscopy	1	*Red Blood Cell, normal (RBC) (Erythrocyte)	1	100.00%
ID-5		* Denotes Correct Response		
AG: All Methods	1			
		*Lymphocyte	1	100.00%
Intl Blood Cell Microscopy	1	*Lymphocyte	1	100.00%
ID-6		* Denotes Correct Response		
AG: All Methods	2			
		*Lymphocyte, immature	1	50.00%
		Lymphoblast	1	50.00%
Intl Blood Cell Microscopy	2	*Lymphocyte, immature	1	50.00%
Intl Blood Cell Microscopy		Lymphoblast	1	50.00%
609 - Waived/PPM Tests 1;675 - Infectious Mononucleosis;753 - Infectious Mononucleosis (Waived) (1 specimens);754 - Infectious Mono/RF Combo;759 - Complete Immunology;628 - AFC Waived/PPM;729 - AFC Immunology	Labs	Response	Number	Percent
		Infectious Mono (Waived)		
IM-1		* Denotes Correct Response		
AG: All Methods	504			
		*Negative	503	99.80%
		Positive	1	0.20%
Alere Acceava (S)	1	*Negative	1	100.00%
Alere Acceava (WB)	53	*Negative	53	100.00%
Sekisui Diagnostics OSOM Mono (WB)	220	*Negative	219	99.55%



Henry Schein One-Step+ Mono	25	*Negative	25	100.00%
Instant Technologies i-Screen (WB)	2	*Negative	2	100.00%
Inverness Signify Mono	3	*Negative	3	100.00%
LifeSign Status Inf Mono	34	*Negative	34	100.00%
McKesson Medi-Lab Perf - Inf Mono	18	*Negative	18	100.00%
Polymedco Poly stat Mono - (WB)	6	*Negative	6	100.00%
Quidel QuickVue+ (WB)	19	*Negative	19	100.00%
SP Brand (Cardinal) Rapid Test	7	*Negative	7	100.00%
Alere Clearview Mono Cassette	25	*Negative	25	100.00%
Pulse Scientific	1	*Negative	1	100.00%
Diag Test Group Clarity Mono	3	*Negative	3	100.00%
Fisher HC Sure-View Mono	4	*Negative	4	100.00%
Immunostics Inc. ImmunoDetector Mono (WB)	4	*Negative	4	100.00%
Meridian Immunocard STAT! Mono (WB)	5	*Negative	5	100.00%
NDC ProAdvantage Mono	2	*Negative	2	100.00%
Consult Diag Mono Cassette	65	*Negative	65	100.00%
Alere Acceava (WB) Mono Cassette	7	*Negative	7	100.00%
Sekisui Diagnostics OSOM Mono (WB)	220	Positive	1	0.45%
Infectious Mononucleosis				
IM-1				
* Denotes Correct Response				
AG: All Methods	79			
		*Negative	79	100.00%
Beck Coulter ICON (WB)	1	*Negative	1	100.00%
Alere Acceava (WB)	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	36	*Negative	36	100.00%
Henry Schein One-Step+ Mono	1	*Negative	1	100.00%
LifeSign Status Inf Mono	2	*Negative	2	100.00%
Lifesign Unistep	1	*Negative	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	3	*Negative	3	100.00%
Polymedco Poly stat Mono - (WB)	1	*Negative	1	100.00%
Polymedco Poly stat Mono (S/P)	2	*Negative	2	100.00%
Quidel QuickVue+ (S/P)	2	*Negative	2	100.00%
Quidel QuickVue+ (WB)	8	*Negative	8	100.00%
Seradyn Color Slide II	4	*Negative	4	100.00%
SP Brand (Cardinal) Rapid Test	1	*Negative	1	100.00%
Alere Clearview Mono Cassette	8	*Negative	8	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	3	*Negative	3	100.00%
NDC ProAdvantage Mono	1	*Negative	1	100.00%
Consult Diag Mono Cassette	4	*Negative	4	100.00%
675 - Infectious Mononucleosis;754 - Infectious Mono/RF Combo;759 - Complete Immunology;628 - AFC Waived/PPM;729 - AFC Immunology	Labs	Response	Number	Percent
Infectious Mononucleosis				
IM-1				
* Denotes Correct Response				
AG: All Methods	79			
		*Negative	79	100.00%
Beck Coulter ICON (WB)	1	*Negative	1	100.00%
Alere Acceava (WB)	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	36	*Negative	36	100.00%
Henry Schein One-Step+ Mono	1	*Negative	1	100.00%
LifeSign Status Inf Mono	2	*Negative	2	100.00%
Lifesign Unistep	1	*Negative	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	3	*Negative	3	100.00%
Polymedco Poly stat Mono - (WB)	1	*Negative	1	100.00%
Polymedco Poly stat Mono (S/P)	2	*Negative	2	100.00%
Quidel QuickVue+ (S/P)	2	*Negative	2	100.00%
Quidel QuickVue+ (WB)	8	*Negative	8	100.00%
Seradyn Color Slide II	4	*Negative	4	100.00%
SP Brand (Cardinal) Rapid Test	1	*Negative	1	100.00%
Alere Clearview Mono Cassette	8	*Negative	8	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	3	*Negative	3	100.00%
NDC ProAdvantage Mono	1	*Negative	1	100.00%
Consult Diag Mono Cassette	4	*Negative	4	100.00%
IM-2				
* Denotes Correct Response				
AG: All Methods	79			
		*Positive	79	100.00%
Beck Coulter ICON (WB)	1	*Positive	1	100.00%
Alere Acceava (WB)	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	36	*Positive	36	100.00%



Henry Schein One-Step+ Mono	1	*Positive	1	100.00%
LifeSign Status Inf Mono	2	*Positive	2	100.00%
Lifesign Unistep	1	*Positive	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	3	*Positive	3	100.00%
Polymedco Poly stat Mono - (WB)	1	*Positive	1	100.00%
Polymedco Poly stat Mono (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (WB)	8	*Positive	8	100.00%
Seradyn Color Slide II	4	*Positive	4	100.00%
SP Brand (Cardinal) Rapid Test	1	*Positive	1	100.00%
Alere Clearview Mono Cassette	8	*Positive	8	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	3	*Positive	3	100.00%
NDC ProAdvantage Mono	1	*Positive	1	100.00%
Consult Diag Mono Cassette	4	*Positive	4	100.00%
IM-3		* Denotes Correct Response		
AG: All Methods	79			
		*Positive	79	100.00%
Beck Coulter ICON (WB)	1	*Positive	1	100.00%
Alere Acceava (WB)	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	36	*Positive	36	100.00%
Henry Schein One-Step+ Mono	1	*Positive	1	100.00%
LifeSign Status Inf Mono	2	*Positive	2	100.00%
Lifesign Unistep	1	*Positive	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	3	*Positive	3	100.00%
Polymedco Poly stat Mono - (WB)	1	*Positive	1	100.00%
Polymedco Poly stat Mono (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (WB)	8	*Positive	8	100.00%
Seradyn Color Slide II	4	*Positive	4	100.00%
SP Brand (Cardinal) Rapid Test	1	*Positive	1	100.00%
Alere Clearview Mono Cassette	8	*Positive	8	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	3	*Positive	3	100.00%
NDC ProAdvantage Mono	1	*Positive	1	100.00%
Consult Diag Mono Cassette	4	*Positive	4	100.00%
IM-4		* Denotes Correct Response		
AG: All Methods	79			
		*Negative	79	100.00%
Beck Coulter ICON (WB)	1	*Negative	1	100.00%
Alere Acceava (WB)	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	36	*Negative	36	100.00%
Henry Schein One-Step+ Mono	1	*Negative	1	100.00%
LifeSign Status Inf Mono	2	*Negative	2	100.00%
Lifesign Unistep	1	*Negative	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	3	*Negative	3	100.00%
Polymedco Poly stat Mono - (WB)	1	*Negative	1	100.00%
Polymedco Poly stat Mono (S/P)	2	*Negative	2	100.00%
Quidel QuickVue+ (S/P)	2	*Negative	2	100.00%
Quidel QuickVue+ (WB)	8	*Negative	8	100.00%
Seradyn Color Slide II	4	*Negative	4	100.00%
SP Brand (Cardinal) Rapid Test	1	*Negative	1	100.00%
Alere Clearview Mono Cassette	8	*Negative	8	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	3	*Negative	3	100.00%
NDC ProAdvantage Mono	1	*Negative	1	100.00%
Consult Diag Mono Cassette	4	*Negative	4	100.00%
IM-5		* Denotes Correct Response		
AG: All Methods	79			
		*Positive	79	100.00%
Beck Coulter ICON (WB)	1	*Positive	1	100.00%
Alere Acceava (WB)	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	36	*Positive	36	100.00%
Henry Schein One-Step+ Mono	1	*Positive	1	100.00%
LifeSign Status Inf Mono	2	*Positive	2	100.00%
Lifesign Unistep	1	*Positive	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	3	*Positive	3	100.00%
Polymedco Poly stat Mono - (WB)	1	*Positive	1	100.00%
Polymedco Poly stat Mono (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (WB)	8	*Positive	8	100.00%
Seradyn Color Slide II	4	*Positive	4	100.00%



SP Brand (Cardinal) Rapid Test	1	*Positive	1	100.00%
Alere Clearview Mono Cassette	8	*Positive	8	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	3	*Positive	3	100.00%
NDC ProAdvantage Mono	1	*Positive	1	100.00%
Consult Diag Mono Cassette	4	*Positive	4	100.00%
Infectious Mono (Waived)				
IM-1		* Denotes Correct Response		
AG: All Methods	504			
		*Negative	503	99.80%
		Positive	1	0.20%
Alere Acceava (S)	1	*Negative	1	100.00%
Alere Acceava (WB)	53	*Negative	53	100.00%
Sekisui Diagnostics OSOM Mono (WB)	220	*Negative	219	99.55%
Henry Schein One-Step+ Mono	25	*Negative	25	100.00%
Instant Technologies i-Screen (WB)	2	*Negative	2	100.00%
Inverness Signify Mono	3	*Negative	3	100.00%
LifeSign Status Inf Mono	34	*Negative	34	100.00%
McKesson Medi-Lab Perf - Inf Mono	18	*Negative	18	100.00%
Polymedco Poly stat Mono - (WB)	6	*Negative	6	100.00%
Quidel QuickVue+ (WB)	19	*Negative	19	100.00%
SP Brand (Cardinal) Rapid Test	7	*Negative	7	100.00%
Alere Clearview Mono Cassette	25	*Negative	25	100.00%
Pulse Scientific	1	*Negative	1	100.00%
Diag Test Group Clarity Mono	3	*Negative	3	100.00%
Fisher HC Sure-View Mono	4	*Negative	4	100.00%
Immunostics Inc. ImmunoDetector Mono (WB)	4	*Negative	4	100.00%
Meridian Immunocard STAT! Mono (WB)	5	*Negative	5	100.00%
NDC ProAdvantage Mono	2	*Negative	2	100.00%
Consult Diag Mono Cassette	65	*Negative	65	100.00%
Alere Acceava (WB) Mono Cassette	7	*Negative	7	100.00%
Sekisui Diagnostics OSOM Mono (WB)	220	Positive	1	0.45%
Infectious Mononucleosis (International)				
IM-1		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Alere Clearview Mono Cassette	1	*Negative	1	100.00%
Fisher HC Sure-View Mono	1	*Negative	1	100.00%
IM-2		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Alere Clearview Mono Cassette	1	*Positive	1	100.00%
Fisher HC Sure-View Mono	1	*Positive	1	100.00%
IM-3		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Alere Clearview Mono Cassette	1	*Positive	1	100.00%
Fisher HC Sure-View Mono	1	*Positive	1	100.00%
IM-4		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Alere Clearview Mono Cassette	1	*Negative	1	100.00%
Fisher HC Sure-View Mono	1	*Negative	1	100.00%
IM-5		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Alere Clearview Mono Cassette	1	*Positive	1	100.00%
Fisher HC Sure-View Mono	1	*Positive	1	100.00%
637 - KOH - Glass Slides (Skin Hair Nail)	Labs	Response	Number	Percent
KOH Preparation				
KOH-1		* Denotes Correct Response		
AG: All Methods	18			
		*No fungal elements/yeast observed	17	94.44%
		Fungal elements/yeast observed	1	5.56%
KOH Glass Slide	18	*No fungal elements/yeast observed	17	94.44%
KOH Glass Slide		Fungal elements/yeast observed	1	5.56%
KOH-2		* Denotes Correct Response		
AG: All Methods	18			
		*Fungal elements/yeast observed	18	100.00%



Test Name	Labs	Response	Number	Percent
KOH Glass Slide	18	*Fungal elements/yeast observed	18	100.00%
758 - Mycoplasma Antibody	Labs	Response	Number	Percent
		Mycoplasma Antibody		
MY-1		* Denotes Correct Response		
AG: All Methods	5			
		*Negative	5	100.00%
Meridian ImmunoCard	5	*Negative	5	100.00%
791 - Fecal Occult Blood Add-On;601 - Waived/PPM Tests II;609 - Waived/PPM Tests I;647 - Mini-Waived Package;628 - AFC Waived/PPM	Labs	Response	Number	Percent
		Fecal Occult Blood		
OC-1		* Denotes Correct Response		
MG: Guaiac methods	787			
		*Positive	767	97.46%
		Negative	20	2.54%
OC-1		* Denotes Correct Response		
MG: Immunochemical methods	118			
		*Positive	107	90.68%
		Negative	11	9.32%
950 - Post Vasectomy Sperm Analysis	Labs	Response	Number	Percent
		Sperm Count (Post Vasectomy)		
PVAS-1		* Denotes Correct Response		
AG: All Methods	16			
		*Positive for Sperm	16	100.00%
Hemocytometer	4	*Positive for Sperm	4	100.00%
Post Vasectomy Microscopy	12	*Positive for Sperm	12	100.00%
PVAS-2		* Denotes Correct Response		
AG: All Methods	16			
		*Negative for Sperm	16	100.00%
Hemocytometer	4	*Negative for Sperm	4	100.00%
Post Vasectomy Microscopy	12	*Negative for Sperm	12	100.00%
754 - Infectious Mono/RF Combo;653 - Rheumatoid Factor;759 - Complete Immunology	Labs	Response	Number	Percent
		Rheumatoid Factor		
RF-1		* Denotes Correct Response		
AG: All Methods	12			
		*Positive	12	100.00%
ASI	1	*Positive	1	100.00%
Siemens RapiTex	2	*Positive	2	100.00%
Seradyn Color Slide	2	*Positive	2	100.00%
Wampole ColorCard RF	2	*Positive	2	100.00%
Wampole Rheumatex	2	*Positive	2	100.00%
Biokit Rhemajet RF	2	*Positive	2	100.00%
Remel SeraTest RF-Latex	1	*Positive	1	100.00%
RF-2		* Denotes Correct Response		
AG: All Methods	12			
		*N/A (Negative)	11	91.67%
		Positive	1	8.33%
ASI	1	*N/A (Negative)	1	100.00%
Siemens RapiTex	2	*N/A (Negative)	2	100.00%
Seradyn Color Slide	2	*N/A (Negative)	2	100.00%
Wampole ColorCard RF	2	*N/A (Negative)	2	100.00%
Wampole Rheumatex	2	*N/A (Negative)	2	100.00%
Biokit Rhemajet RF	2	*N/A (Negative)	1	50.00%
Remel SeraTest RF-Latex	1	*N/A (Negative)	1	100.00%
Biokit Rhemajet RF	2	Positive	1	50.00%
RF-3		* Denotes Correct Response		
AG: All Methods	12			
		*Positive	10	83.33%
		N/A (Negative)	2	16.67%
ASI	1	*Positive	1	100.00%
Siemens RapiTex	2	*Positive	2	100.00%
Seradyn Color Slide	2	*Positive	2	100.00%
Wampole Rheumatex	2	*Positive	2	100.00%
Biokit Rhemajet RF	2	*Positive	2	100.00%
Remel SeraTest RF-Latex	1	*Positive	1	100.00%
Wampole ColorCard RF	2	N/A (Negative)	2	100.00%
RF-4		* Denotes Correct Response		



AG: All Methods	12			
		*Positive	12	100.00%
ASI	1	*Positive	1	100.00%
Siemens RapiTex	2	*Positive	2	100.00%
Seradyn Color Slide	2	*Positive	2	100.00%
Wampole ColorCard RF	2	*Positive	2	100.00%
Wampole Rheumatex	2	*Positive	2	100.00%
Biokit Rhemajet RF	2	*Positive	2	100.00%
Remel SeraTest RF-Latex	1	*Positive	1	100.00%
RF-5		* Denotes Correct Response		
AG: All Methods	12			
		*N/A (Negative)	11	91.67%
		Positive	1	8.33%
ASI	1	*N/A (Negative)	1	100.00%
Siemens RapiTex	2	*N/A (Negative)	2	100.00%
Seradyn Color Slide	2	*N/A (Negative)	2	100.00%
Wampole ColorCard RF	2	*N/A (Negative)	2	100.00%
Wampole Rheumatex	2	*N/A (Negative)	2	100.00%
Biokit Rhemajet RF	2	*N/A (Negative)	1	50.00%
Remel SeraTest RF-Latex	1	*N/A (Negative)	1	100.00%
Biokit Rhemajet RF	2	Positive	1	50.00%
		Rheumatoid Factor (International)		
RF-1		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Wampole ColorCard RF	1	*Positive	1	100.00%
Fisher HC Sure-Vue RF	1	*Positive	1	100.00%
Immunostics Rhuemacol/LEX	1	*Positive	1	100.00%
RF-2		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Wampole ColorCard RF	1	*Negative	1	100.00%
Fisher HC Sure-Vue RF	1	*Negative	1	100.00%
Immunostics Rhuemacol/LEX	1	*Negative	1	100.00%
RF-3		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Wampole ColorCard RF	1	*Positive	1	100.00%
Fisher HC Sure-Vue RF	1	*Positive	1	100.00%
Immunostics Rhuemacol/LEX	1	*Positive	1	100.00%
RF-4		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Wampole ColorCard RF	1	*Positive	1	100.00%
Fisher HC Sure-Vue RF	1	*Positive	1	100.00%
Immunostics Rhuemacol/LEX	1	*Positive	1	100.00%
RF-5		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Wampole ColorCard RF	1	*Negative	1	100.00%
Fisher HC Sure-Vue RF	1	*Negative	1	100.00%
Immunostics Rhuemacol/LEX	1	*Negative	1	100.00%
		Rheumatoid Factor (Semi Quant)		
RF-1		* Denotes Correct Response		
AG: All Methods	1			
		*160 IU/mL	1	100.00%
Fisher HC Sure-Vue RF (Semi Quant)	1	*160 IU/mL	1	100.00%
RF-2		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Fisher HC Sure-Vue RF (Semi Quant)	1	*Negative	1	100.00%
RF-3		* Denotes Correct Response		
AG: All Methods	1			
		*160 IU/mL	1	100.00%
Fisher HC Sure-Vue RF (Semi Quant)	1	*160 IU/mL	1	100.00%
RF-4		* Denotes Correct Response		
AG: All Methods	1			
		*320 IU/mL	1	100.00%



Test Name	Labs	Response	Number	Percent
Fisher HC Sure-Vue RF (Semi Quant)	1	*320 IU/mL	1	100.00%
RF-5		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Fisher HC Sure-Vue RF (Semi Quant)	1	*Negative	1	100.00%
645 - Fetal Membrane Rupture	Labs	Response	Number	Percent
		Fetal Membranes Rupture PAMG-1		
ROM-1		* Denotes Correct Response		
AG: All Methods	18			
		*Positive	17	94.44%
		Negative	1	5.56%
AmniSure ROM Test	18	*Positive	17	94.44%
AmniSure ROM Test		Negative	1	5.56%
ROM-2		* Denotes Correct Response		
AG: All Methods	18			
		*Negative	17	94.44%
		Positive	1	5.56%
AmniSure ROM Test	18	*Negative	17	94.44%
AmniSure ROM Test		Positive	1	5.56%
601 - Waived/PPM Tests II;609 - Waived/PPM Tests I;785 - Presump ID/Colony Count-Urine & Strep Combo;691 - Group A Strep Antigen (5 specimens);699 - Group A Strep Antigen (1 specimen);783 - Group A Strep Antigen/Add-on (2 specimens);665 - Group A Strep An	Labs	Response	Number	Percent
		Strep A Ag Det. (Waived)		
RS-1		* Denotes Correct Response		
AG: All Methods	927			
		*Positive	792	85.44%
		Negative	135	14.56%
Abbott Signify Strep A - Waived	7	*Positive	5	71.43%
Beck Coulter ICON DS Strep	1	*Positive	1	100.00%
Beck Coulter ICON SC Strep	1	*Positive	1	100.00%
Alere Acceava Strep A Dipstick	100	*Positive	89	89.00%
Diag Test Group Clarity - Strip Strep Ag	1	*Positive	1	100.00%
Fisher HC Sure-Vue Strep A (Waived)	7	*Positive	6	85.71%
Sekisui Diagnostics OSOM	44	*Positive	42	95.45%
Sekisui Diagnostics OSOM Ultra Strep	52	*Positive	50	96.15%
Henry Schein OneStep + Rapid Cass	3	*Positive	2	66.67%
Henry Schein OneStep+ Rapid Dip	48	*Positive	33	68.75%
Innovacon One Step Strep Ag	1	*Positive	1	100.00%
Instant Technologies i Strep	1	*Positive	1	100.00%
Inverness Signify Strep A Dipstick	6	*Positive	5	83.33%
McKesson Strep A Cas	4	*Positive	3	75.00%
McKesson Strep A Dip	46	*Positive	40	86.96%
McKesson Strep A Test - Twist	2	*Positive	1	50.00%
Permaxim Rediscreen Strep A Rapid	1	*Positive	1	100.00%
Polymedco Poly Stat Strep A (W)	3	*Positive	2	66.67%
PSS Rapid Test Strip Strep A	2	*Positive	2	100.00%
Quidel QuickVue Dipstick Strep	97	*Positive	93	95.88%
Quidel QuickVue In-Line - 25 Test	108	*Positive	89	82.41%
Quidel QuickVue+ Strep A	34	*Positive	24	70.59%
Select Diag Strep A Dip	4	*Positive	2	50.00%
SP (Cardinal) Rapid Strep Cassette	6	*Positive	4	66.67%
SP Brand (Cardinal) Rapid Strep A Dipstick	108	*Positive	87	80.56%
Stanbio QuStick Strep A	1	*Positive	1	100.00%
Alere Clearview Exact II Dip	40	*Positive	33	82.50%
Sekisui Diagnostics OSOM Ultra	5	*Positive	5	100.00%
Consult Diag (PSS) Strep A Dipstick	105	*Positive	87	82.86%
Consult Diag (PSS) Strep A Twist	4	*Positive	4	100.00%
MooreBrand Strep A Dipstick	5	*Positive	3	60.00%
NDC ProAdvantage Strep Strips	2	*Positive	1	50.00%
Jant Accutest Accustrip Strep A	3	*Positive	3	100.00%
Meridian ImmunoCard STAT Strep A	1	*Positive	1	100.00%
Quidel Sofia Strep A FIA	6	*Positive	6	100.00%
BD Veritor Strep A	48	*Positive	47	97.92%
Alere Acceava DO NOT USE	10	*Positive	7	70.00%
Alere i Strep A	6	*Positive	6	100.00%
Roche Cobas Liat Strep A	2	*Positive	2	100.00%
Abbott Signify Strep A - Waived	7	Negative	2	28.57%



Alere Acceava Strep A Dipstick	100	Negative	11	11.00%
Fisher HC Sure-Vue Strep A (Waived)	7	Negative	1	14.29%
Sekisui Diagnostics OSOM	44	Negative	2	4.55%
Sekisui Diagnostics OSOM Ultra Strep	52	Negative	2	3.85%
Henry Schein OneStep + Rapid Cass	3	Negative	1	33.33%
Henry Schein OneStep+ Rapid Dip	48	Negative	15	31.25%
Inverness Signify Strep A Dipstick	6	Negative	1	16.67%
McKesson Strep A Cas	4	Negative	1	25.00%
McKesson Strep A Dip	46	Negative	6	13.04%
McKesson Strep A Test - Twist	2	Negative	1	50.00%
Polymedco Poly Stat Strep A (W)	3	Negative	1	33.33%
Quidel QuickVue Dipstick Strep	97	Negative	4	4.12%
Quidel QuickVue In-Line - 25 Test	108	Negative	19	17.59%
Quidel QuickVue+ Strep A	34	Negative	10	29.41%
Select Diag Strep A Dip	4	Negative	2	50.00%
SP (Cardinal) Rapid Strep Cassette	6	Negative	2	33.33%
SP Brand (Cardinal) Rapid Strep A Dipstick	108	Negative	21	19.44%
Alere Clearview Exact II Dip	40	Negative	7	17.50%
Consult Diag (PSS) Strep A Dipstick	105	Negative	18	17.14%
MooreBrand Strep A Dipstick	5	Negative	2	40.00%
NDC ProAdvantage Strep Strips	2	Negative	1	50.00%
LifeSign Bio Strep A Test STRIP	1	Negative	1	100.00%
BD Veritor Strep A	48	Negative	1	2.08%
Alere Acceava DO NOT USE	10	Negative	3	30.00%
RS-2		* Denotes Correct Response		
AG: All Methods	36			
		*Negative	36	100.00%
Sekisui Diagnostics OSOM	2	*Negative	2	100.00%
Sekisui Diagnostics OSOM Ultra Strep	4	*Negative	4	100.00%
Innovacon One Step Strep Ag	1	*Negative	1	100.00%
McKesson Strep A Dip	1	*Negative	1	100.00%
Quidel QuickVue Dipstick Strep	3	*Negative	3	100.00%
Quidel QuickVue In-Line - 25 Test	1	*Negative	1	100.00%
Quidel QuickVue+ Strep A	1	*Negative	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	7	*Negative	7	100.00%
Consult Diag (PSS) Strep A Dipstick	7	*Negative	7	100.00%
Consult Diag (PSS) Strep A Twist	1	*Negative	1	100.00%
MooreBrand Strep A Dipstick	1	*Negative	1	100.00%
Jant Accutest Accustrip Strep A	3	*Negative	3	100.00%
Alere Acceava DO NOT USE	1	*Negative	1	100.00%
Alere i Strep A	3	*Negative	3	100.00%
		Strep A Ag Det.		
RS-1		* Denotes Correct Response		
AG: All Methods	209			
		*Positive	206	98.56%
		Negative	3	1.44%
Abbott Signify Strep A - Waived	2	*Positive	2	100.00%
BD Check Strep A (Dip)	2	*Positive	2	100.00%
Beck Coulter ICON DS Strep	2	*Positive	2	100.00%
Alere Acceava Strep A Dipstick	17	*Positive	17	100.00%
Diag Test Group Clarity - Strip Strep Ag	1	*Positive	1	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM	25	*Positive	25	100.00%
Sekisui Diagnostics OSOM Ultra Strep	29	*Positive	29	100.00%
Henry Schein OneStep+ Rapid Dip	10	*Positive	8	80.00%
Immunostics Immuno/Strep A Det	2	*Positive	2	100.00%
Inverness Signify Strep A Dipstick	1	*Positive	1	100.00%
McKesson Strep A Cas	2	*Positive	2	100.00%
McKesson Strep A Dip	10	*Positive	10	100.00%
Quidel QuickVue Dipstick Strep	22	*Positive	22	100.00%
Quidel QuickVue In-Line - 25 Test	15	*Positive	14	93.33%
Quidel QuickVue+ Strep A	31	*Positive	31	100.00%
Select Diag Strep A Dip	1	*Positive	1	100.00%
SP (Cardinal) Rapid Strep Cassette	1	*Positive	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	*Positive	3	100.00%
Stanbio QuStick Strep A	1	*Positive	1	100.00%
Alere Clearview Exact II Dip	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Ultra	1	*Positive	1	100.00%
Consult Diag (PSS) Strep A Dipstick	15	*Positive	15	100.00%



Quidel Sofia Strep A FIA	5	*Positive	5	100.00%
BD Veritor Strep A	6	*Positive	6	100.00%
Quidel Solana	3	*Positive	3	100.00%
Henry Schein OneStep+ Rapid Dip	10	Negative	2	20.00%
Quidel QuickVue In-Line - 25 Test	15	Negative	1	6.67%
RS-2		* Denotes Correct Response		
AG: All Methods	209			
		*Negative	209	100.00%
Abbott Signify Strep A - Waived	2	*Negative	2	100.00%
BD Check Strep A (Dip)	2	*Negative	2	100.00%
Beck Coulter ICON DS Strep	2	*Negative	2	100.00%
Alere Acceava Strep A Dipstick	17	*Negative	17	100.00%
Diag Test Group Clarity - Strip Strep Ag	1	*Negative	1	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM	25	*Negative	25	100.00%
Sekisui Diagnostics OSOM Ultra Strep	29	*Negative	29	100.00%
Henry Schein OneStep+ Rapid Dip	10	*Negative	10	100.00%
Immunostics Immuno/Strep A Det	2	*Negative	2	100.00%
Inverness Signify Strep A Dipstick	1	*Negative	1	100.00%
McKesson Strep A Cas	2	*Negative	2	100.00%
McKesson Strep A Dip	10	*Negative	10	100.00%
Quidel QuickVue Dipstick Strep	22	*Negative	22	100.00%
Quidel QuickVue In-Line - 25 Test	15	*Negative	15	100.00%
Quidel QuickVue+ Strep A	31	*Negative	31	100.00%
Select Diag Strep A Dip	1	*Negative	1	100.00%
SP (Cardinal) Rapid Strep Cassette	1	*Negative	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	*Negative	3	100.00%
Stanbio QuStick Strep A	1	*Negative	1	100.00%
Alere Clearview Exact II Dip	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM Ultra	1	*Negative	1	100.00%
Consult Diag (PSS) Strep A Dipstick	15	*Negative	15	100.00%
Quidel Sofia Strep A FIA	5	*Negative	5	100.00%
BD Veritor Strep A	6	*Negative	6	100.00%
Quidel Solana	3	*Negative	3	100.00%
		Strep A Ag Det. (International)		
RS-1		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Alere Clearview Exact II Cas	1	*Positive	1	100.00%
RS-2		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Alere Clearview Exact II Cas	1	*Negative	1	100.00%
691 - Group A Strep Antigen (5 specimens)	Labs	Response	Number	Percent
		Strep A Ag Det.		
RS-1		* Denotes Correct Response		
AG: All Methods	209			
		*Positive	206	98.56%
		Negative	3	1.44%
Abbott Signify Strep A - Waived	2	*Positive	2	100.00%
BD Check Strep A (Dip)	2	*Positive	2	100.00%
Beck Coulter ICON DS Strep	2	*Positive	2	100.00%
Alere Acceava Strep A Dipstick	17	*Positive	17	100.00%
Diag Test Group Clarity - Strip Strep Ag	1	*Positive	1	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM	25	*Positive	25	100.00%
Sekisui Diagnostics OSOM Ultra Strep	29	*Positive	29	100.00%
Henry Schein OneStep+ Rapid Dip	10	*Positive	8	80.00%
Immunostics Immuno/Strep A Det	2	*Positive	2	100.00%
Inverness Signify Strep A Dipstick	1	*Positive	1	100.00%
McKesson Strep A Cas	2	*Positive	2	100.00%
McKesson Strep A Dip	10	*Positive	10	100.00%
Quidel QuickVue Dipstick Strep	22	*Positive	22	100.00%
Quidel QuickVue In-Line - 25 Test	15	*Positive	14	93.33%
Quidel QuickVue+ Strep A	31	*Positive	31	100.00%
Select Diag Strep A Dip	1	*Positive	1	100.00%
SP (Cardinal) Rapid Strep Cassette	1	*Positive	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	*Positive	3	100.00%
Stanbio QuStick Strep A	1	*Positive	1	100.00%



Alere Clearview Exact II Dip	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Ultra	1	*Positive	1	100.00%
Consult Diag (PSS) Strep A Dipstick	15	*Positive	15	100.00%
Quidel Sofia Strep A FIA	5	*Positive	5	100.00%
BD Veritor Strep A	6	*Positive	6	100.00%
Quidel Solana	3	*Positive	3	100.00%
Henry Schein OneStep+ Rapid Dip	10	Negative	2	20.00%
Quidel QuickVue In-Line - 25 Test	15	Negative	1	6.67%
RS-2		* Denotes Correct Response		
AG: All Methods	209			
		*Negative	209	100.00%
Abbott Signify Strep A - Waived	2	*Negative	2	100.00%
BD Check Strep A (Dip)	2	*Negative	2	100.00%
Beck Coulter ICON DS Strep	2	*Negative	2	100.00%
Alere Acceava Strep A Dipstick	17	*Negative	17	100.00%
Diag Test Group Clarity - Strip Strep Ag	1	*Negative	1	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM	25	*Negative	25	100.00%
Sekisui Diagnostics OSOM Ultra Strep	29	*Negative	29	100.00%
Henry Schein OneStep+ Rapid Dip	10	*Negative	10	100.00%
Immunostics Immuno/Strep A Det	2	*Negative	2	100.00%
Inverness Signify Strep A Dipstick	1	*Negative	1	100.00%
McKesson Strep A Cas	2	*Negative	2	100.00%
McKesson Strep A Dip	10	*Negative	10	100.00%
Quidel QuickVue Dipstick Strep	22	*Negative	22	100.00%
Quidel QuickVue In-Line - 25 Test	15	*Negative	15	100.00%
Quidel QuickVue+ Strep A	31	*Negative	31	100.00%
Select Diag Strep A Dip	1	*Negative	1	100.00%
SP (Cardinal) Rapid Strep Cassette	1	*Negative	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	*Negative	3	100.00%
Stanbio QuStick Strep A	1	*Negative	1	100.00%
Alere Clearview Exact II Dip	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM Ultra	1	*Negative	1	100.00%
Consult Diag (PSS) Strep A Dipstick	15	*Negative	15	100.00%
Quidel Sofia Strep A FIA	5	*Negative	5	100.00%
BD Veritor Strep A	6	*Negative	6	100.00%
Quidel Solana	3	*Negative	3	100.00%
RS-3		* Denotes Correct Response		
AG: All Methods	178			
		*Positive	169	94.94%
		Negative	9	5.06%
Abbott Signify Strep A - Waived	2	*Positive	2	100.00%
BD Check Strep A (Dip)	2	*Positive	2	100.00%
Beck Coulter ICON DS Strep	2	*Positive	2	100.00%
Alere Acceava Strep A Dipstick	17	*Positive	16	94.12%
Diag Test Group Clarity - Strip Strep Ag	1	*Positive	1	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM	25	*Positive	25	100.00%
Sekisui Diagnostics OSOM Ultra Strep	25	*Positive	25	100.00%
Henry Schein OneStep+ Rapid Dip	10	*Positive	8	80.00%
Immunostics Immuno/Strep A Det	2	*Positive	2	100.00%
Inverness Signify Strep A Dipstick	1	*Positive	1	100.00%
McKesson Strep A Cas	1	*Positive	1	100.00%
McKesson Strep A Dip	5	*Positive	4	80.00%
Quidel QuickVue Dipstick Strep	18	*Positive	16	88.89%
Quidel QuickVue In-Line - 25 Test	13	*Positive	12	92.31%
Quidel QuickVue+ Strep A	29	*Positive	29	100.00%
SP (Cardinal) Rapid Strep Cassette	1	*Positive	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	*Positive	2	66.67%
Alere Clearview Exact II Dip	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Ultra	1	*Positive	1	100.00%
Consult Diag (PSS) Strep A Dipstick	12	*Positive	11	91.67%
Quidel Sofia Strep A FIA	4	*Positive	4	100.00%
BD Veritor Strep A	1	*Positive	1	100.00%
Quidel Solana	1	*Positive	1	100.00%
Alere Acceava Strep A Dipstick	17	Negative	1	5.88%
Henry Schein OneStep+ Rapid Dip	10	Negative	2	20.00%
McKesson Strep A Dip	5	Negative	1	20.00%
Quidel QuickVue Dipstick Strep	18	Negative	2	11.11%



Quidel QuickVue In-Line - 25 Test	13	Negative	1	7.69%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	Negative	1	33.33%
Consult Diag (PSS) Strep A Dipstick	12	Negative	1	8.33%
RS-4		* Denotes Correct Response		
AG: All Methods	178			
		*Negative	177	99.44%
		Positive	1	0.56%
Abbott Signify Strep A - Waived	2	*Negative	2	100.00%
BD Check Strep A (Dip)	2	*Negative	2	100.00%
Beck Coulter ICON DS Strep	2	*Negative	2	100.00%
Alere Acceava Strep A Dipstick	17	*Negative	17	100.00%
Diag Test Group Clarity - Strip Strep Ag	1	*Negative	1	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM	25	*Negative	25	100.00%
Sekisui Diagnostics OSOM Ultra Strep	25	*Negative	25	100.00%
Henry Schein OneStep+ Rapid Dip	10	*Negative	10	100.00%
Immunostics Immuno/Strep A Det	2	*Negative	2	100.00%
Inverness Signify Strep A Dipstick	1	*Negative	1	100.00%
McKesson Strep A Cas	1	*Negative	1	100.00%
McKesson Strep A Dip	5	*Negative	5	100.00%
Quidel QuickVue Dipstick Strep	18	*Negative	17	94.44%
Quidel QuickVue In-Line - 25 Test	13	*Negative	13	100.00%
Quidel QuickVue+ Strep A	29	*Negative	29	100.00%
SP (Cardinal) Rapid Strep Cassette	1	*Negative	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	*Negative	3	100.00%
Alere Clearview Exact II Dip	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM Ultra	1	*Negative	1	100.00%
Consult Diag (PSS) Strep A Dipstick	12	*Negative	12	100.00%
Quidel Sofia Strep A FIA	4	*Negative	4	100.00%
BD Veritor Strep A	1	*Negative	1	100.00%
Quidel Solana	1	*Negative	1	100.00%
Quidel QuickVue Dipstick Strep	18	Positive	1	5.56%
RS-5		* Denotes Correct Response		
AG: All Methods	178			
		*Positive	168	94.38%
		Negative	10	5.62%
Abbott Signify Strep A - Waived	2	*Positive	1	50.00%
BD Check Strep A (Dip)	2	*Positive	2	100.00%
Beck Coulter ICON DS Strep	2	*Positive	1	50.00%
Alere Acceava Strep A Dipstick	17	*Positive	15	88.24%
Diag Test Group Clarity - Strip Strep Ag	1	*Positive	1	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM	25	*Positive	24	96.00%
Sekisui Diagnostics OSOM Ultra Strep	25	*Positive	25	100.00%
Henry Schein OneStep+ Rapid Dip	10	*Positive	8	80.00%
Immunostics Immuno/Strep A Det	2	*Positive	2	100.00%
Inverness Signify Strep A Dipstick	1	*Positive	1	100.00%
McKesson Strep A Cas	1	*Positive	1	100.00%
McKesson Strep A Dip	5	*Positive	4	80.00%
Quidel QuickVue Dipstick Strep	18	*Positive	18	100.00%
Quidel QuickVue In-Line - 25 Test	13	*Positive	13	100.00%
Quidel QuickVue+ Strep A	29	*Positive	28	96.55%
SP (Cardinal) Rapid Strep Cassette	1	*Positive	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	*Positive	3	100.00%
Alere Clearview Exact II Dip	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Ultra	1	*Positive	1	100.00%
Consult Diag (PSS) Strep A Dipstick	12	*Positive	11	91.67%
Quidel Sofia Strep A FIA	4	*Positive	4	100.00%
BD Veritor Strep A	1	*Positive	1	100.00%
Quidel Solana	1	*Positive	1	100.00%
Abbott Signify Strep A - Waived	2	Negative	1	50.00%
Beck Coulter ICON DS Strep	2	Negative	1	50.00%
Alere Acceava Strep A Dipstick	17	Negative	2	11.76%
Sekisui Diagnostics OSOM	25	Negative	1	4.00%
Henry Schein OneStep+ Rapid Dip	10	Negative	2	20.00%
McKesson Strep A Dip	5	Negative	1	20.00%
Quidel QuickVue+ Strep A	29	Negative	1	3.45%
Consult Diag (PSS) Strep A Dipstick	12	Negative	1	8.33%
		Strep A Ag Det. (Waived)		



RS-1		* Denotes Correct Response		
AG: All Methods	927			
		*Positive	792	85.44%
		Negative	135	14.56%
Abbott Signify Strep A - Waived	7	*Positive	5	71.43%
Beck Coulter ICON DS Strep	1	*Positive	1	100.00%
Beck Coulter ICON SC Strep	1	*Positive	1	100.00%
Alere Acceava Strep A Dipstick	100	*Positive	89	89.00%
Diag Test Group Clarity - Strip Strep Ag	1	*Positive	1	100.00%
Fisher HC Sure-Vue Strep A (Waived)	7	*Positive	6	85.71%
Sekisui Diagnostics OSOM	44	*Positive	42	95.45%
Sekisui Diagnostics OSOM Ultra Strep	52	*Positive	50	96.15%
Henry Schein OneStep + Rapid Cass	3	*Positive	2	66.67%
Henry Schein OneStep+ Rapid Dip	48	*Positive	33	68.75%
Innovacon One Step Strep Ag	1	*Positive	1	100.00%
Instant Technologies i Strep	1	*Positive	1	100.00%
Inverness Signify Strep A Dipstick	6	*Positive	5	83.33%
McKesson Strep A Cas	4	*Positive	3	75.00%
McKesson Strep A Dip	46	*Positive	40	86.96%
McKesson Strep A Test - Twist	2	*Positive	1	50.00%
Permaxim Rediscreen Strep A Rapid	1	*Positive	1	100.00%
Polymedco Poly Stat Strep A (W)	3	*Positive	2	66.67%
PSS Rapid Test Strip Strep A	2	*Positive	2	100.00%
Quidel QuickVue Dipstick Strep	97	*Positive	93	95.88%
Quidel QuickVue In-Line - 25 Test	108	*Positive	89	82.41%
Quidel QuickVue+ Strep A	34	*Positive	24	70.59%
Select Diag Strep A Dip	4	*Positive	2	50.00%
SP (Cardinal) Rapid Strep Cassette	6	*Positive	4	66.67%
SP Brand (Cardinal) Rapid Strep A Dipstick	108	*Positive	87	80.56%
Stanbio QuStick Strep A	1	*Positive	1	100.00%
Alere Clearview Exact II Dip	40	*Positive	33	82.50%
Sekisui Diagnostics OSOM Ultra	5	*Positive	5	100.00%
Consult Diag (PSS) Strep A Dipstick	105	*Positive	87	82.86%
Consult Diag (PSS) Strep A Twist	4	*Positive	4	100.00%
MooreBrand Strep A Dipstick	5	*Positive	3	60.00%
NDC ProAdvantage Strep Strips	2	*Positive	1	50.00%
Jant Accutest Accustrip Strep A	3	*Positive	3	100.00%
Meridian ImmunoCard STAT Strep A	1	*Positive	1	100.00%
Quidel Sofia Strep A FIA	6	*Positive	6	100.00%
BD Veritor Strep A	48	*Positive	47	97.92%
Alere Acceava DO NOT USE	10	*Positive	7	70.00%
Alere i Strep A	6	*Positive	6	100.00%
Roche Cobas Liat Strep A	2	*Positive	2	100.00%
Abbott Signify Strep A - Waived	7	Negative	2	28.57%
Alere Acceava Strep A Dipstick	100	Negative	11	11.00%
Fisher HC Sure-Vue Strep A (Waived)	7	Negative	1	14.29%
Sekisui Diagnostics OSOM	44	Negative	2	4.55%
Sekisui Diagnostics OSOM Ultra Strep	52	Negative	2	3.85%
Henry Schein OneStep + Rapid Cass	3	Negative	1	33.33%
Henry Schein OneStep+ Rapid Dip	48	Negative	15	31.25%
Inverness Signify Strep A Dipstick	6	Negative	1	16.67%
McKesson Strep A Cas	4	Negative	1	25.00%
McKesson Strep A Dip	46	Negative	6	13.04%
McKesson Strep A Test - Twist	2	Negative	1	50.00%
Polymedco Poly Stat Strep A (W)	3	Negative	1	33.33%
Quidel QuickVue Dipstick Strep	97	Negative	4	4.12%
Quidel QuickVue In-Line - 25 Test	108	Negative	19	17.59%
Quidel QuickVue+ Strep A	34	Negative	10	29.41%
Select Diag Strep A Dip	4	Negative	2	50.00%
SP (Cardinal) Rapid Strep Cassette	6	Negative	2	33.33%
SP Brand (Cardinal) Rapid Strep A Dipstick	108	Negative	21	19.44%
Alere Clearview Exact II Dip	40	Negative	7	17.50%
Consult Diag (PSS) Strep A Dipstick	105	Negative	18	17.14%
MooreBrand Strep A Dipstick	5	Negative	2	40.00%
NDC ProAdvantage Strep Strips	2	Negative	1	50.00%
LifeSign Bio Strep A Test STRIP	1	Negative	1	100.00%
BD Veritor Strep A	48	Negative	1	2.08%
Alere Acceava DO NOT USE	10	Negative	3	30.00%
RS-2		* Denotes Correct Response		



AG: All Methods	36			
		*Negative	36	100.00%
Sekisui Diagnostics OSOM	2	*Negative	2	100.00%
Sekisui Diagnostics OSOM Ultra Strep	4	*Negative	4	100.00%
Innovacon One Step Strep Ag	1	*Negative	1	100.00%
McKesson Strep A Dip	1	*Negative	1	100.00%
Quidel QuickVue Dipstick Strep	3	*Negative	3	100.00%
Quidel QuickVue In-Line - 25 Test	1	*Negative	1	100.00%
Quidel QuickVue+ Strep A	1	*Negative	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	7	*Negative	7	100.00%
Consult Diag (PSS) Strep A Dipstick	7	*Negative	7	100.00%
Consult Diag (PSS) Strep A Twist	1	*Negative	1	100.00%
MooreBrand Strep A Dipstick	1	*Negative	1	100.00%
Jant Accutest Accustrip Strep A	3	*Negative	3	100.00%
Alere Aceava DO NOT USE	1	*Negative	1	100.00%
Alere i Strep A	3	*Negative	3	100.00%
RS-3		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Alere i Strep A	2	*Positive	2	100.00%
RS-4		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Alere i Strep A	2	*Negative	2	100.00%
RS-5		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Alere i Strep A	2	*Positive	2	100.00%
		Strep A Aq Det. (International)		
RS-1		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Alere Clearview Exact II Cas	1	*Positive	1	100.00%
RS-2		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Alere Clearview Exact II Cas	1	*Negative	1	100.00%
RS-3		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Alere Clearview Exact II Cas	1	*Positive	1	100.00%
RS-4		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Alere Clearview Exact II Cas	1	*Negative	1	100.00%
RS-5		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Alere Clearview Exact II Cas	1	*Positive	1	100.00%
997 - CVS Group A Strep - 2 samples	Labs	Response	Number	Percent
		Strep A Aq Det. (Waived)		
RS-1		* Denotes Correct Response		
AG: All Methods	38			
		*Positive	38	100.00%
Sekisui Diagnostics OSOM Ultra Strep	38	*Positive	38	100.00%
RS-2		* Denotes Correct Response		
AG: All Methods	38			
		*Negative	38	100.00%
Sekisui Diagnostics OSOM Ultra Strep	38	*Negative	38	100.00%
739 - Sickle Cell Screen	Labs	Response	Number	Percent
		Sickle Cell Screen		
SS-1		* Denotes Correct Response		
AG: All Methods	12			
		*Negative	12	100.00%
Pacific Hemostasis Sickle Screen	3	*Negative	3	100.00%
Streck SickleDex	8	*Negative	8	100.00%
ASI Sickle Cell Test	1	*Negative	1	100.00%
679 - Syphilis Serology	Labs	Response	Number	Percent



		RPR		
SY-1		* Denotes Correct Response		
AG: All Methods	10			
		*Nonreactive	10	100.00%
ASI RPR	2	*Nonreactive	2	100.00%
Becton Dickinson BBL	3	*Nonreactive	3	100.00%
New Horizons	1	*Nonreactive	1	100.00%
Stanbio Quicktest	1	*Nonreactive	1	100.00%
Wampole Impact	1	*Nonreactive	1	100.00%
Diagnostic Direct Syphilis Health Check	1	*Nonreactive	1	100.00%
Intl ASI RPR	1	*Nonreactive	1	100.00%
SY-2		* Denotes Correct Response		
AG: All Methods	10			
		*Reactive	9	90.00%
		*Weakly Reactive	1	10.00%
ASI RPR	2	*Reactive	2	100.00%
Becton Dickinson BBL	3	*Reactive	3	100.00%
New Horizons	1	*Reactive	1	100.00%
Stanbio Quicktest	1	*Reactive	1	100.00%
Diagnostic Direct Syphilis Health Check	1	*Reactive	1	100.00%
Intl ASI RPR	1	*Reactive	1	100.00%
Wampole Impact	1	*Weakly Reactive	1	100.00%
SY-3		* Denotes Correct Response		
AG: All Methods	10			
		*Reactive	9	90.00%
		*Weakly Reactive	1	10.00%
ASI RPR	2	*Reactive	1	50.00%
Becton Dickinson BBL	3	*Reactive	3	100.00%
New Horizons	1	*Reactive	1	100.00%
Stanbio Quicktest	1	*Reactive	1	100.00%
Wampole Impact	1	*Reactive	1	100.00%
Diagnostic Direct Syphilis Health Check	1	*Reactive	1	100.00%
Intl ASI RPR	1	*Reactive	1	100.00%
ASI RPR	2	*Weakly Reactive	1	50.00%
SY-4		* Denotes Correct Response		
AG: All Methods	10			
		*Nonreactive	10	100.00%
ASI RPR	2	*Nonreactive	2	100.00%
Becton Dickinson BBL	3	*Nonreactive	3	100.00%
New Horizons	1	*Nonreactive	1	100.00%
Stanbio Quicktest	1	*Nonreactive	1	100.00%
Wampole Impact	1	*Nonreactive	1	100.00%
Diagnostic Direct Syphilis Health Check	1	*Nonreactive	1	100.00%
Intl ASI RPR	1	*Nonreactive	1	100.00%
SY-5		* Denotes Correct Response		
AG: All Methods	10			
		*Reactive	10	100.00%
ASI RPR	2	*Reactive	2	100.00%
Becton Dickinson BBL	3	*Reactive	3	100.00%
New Horizons	1	*Reactive	1	100.00%
Stanbio Quicktest	1	*Reactive	1	100.00%
Wampole Impact	1	*Reactive	1	100.00%
Diagnostic Direct Syphilis Health Check	1	*Reactive	1	100.00%
Intl ASI RPR	1	*Reactive	1	100.00%
		RPR/VDRL (titer)		
SY-1		* Denotes Correct Response		
AG: All Methods	2			
		*Nonreactive	2	100.00%
ASI RPR	1	*Nonreactive	1	100.00%
New Horizons	1	*Nonreactive	1	100.00%
SY-2		* Denotes Correct Response		
AG: All Methods	2			
		*1:4	1	50.00%
		*1:8	1	50.00%
New Horizons	1	*1:4	1	100.00%
ASI RPR	1	*1:8	1	100.00%
SY-3		* Denotes Correct Response		
AG: All Methods	2			



		*1:2	1	50.00%
		*1:4	1	50.00%
New Horizons	1	*1:2	1	100.00%
ASI RPR	1	*1:4	1	100.00%
SY-4		* Denotes Correct Response		
AG: All Methods	2			
		*Nonreactive	2	100.00%
ASI RPR	1	*Nonreactive	1	100.00%
New Horizons	1	*Nonreactive	1	100.00%
SY-5		* Denotes Correct Response		
AG: All Methods	2			
		*1:2	1	50.00%
		*1:4	1	50.00%
New Horizons	1	*1:2	1	100.00%
ASI RPR	1	*1:4	1	100.00%
		RPR (International)		
SY-1		* Denotes Correct Response		
AG: All Methods	3			
		*Nonreactive	3	100.00%
BD Macro-Vue RPR	1	*Nonreactive	1	100.00%
Fisher HC Sure-Vue RPR	1	*Nonreactive	1	100.00%
Wampole Impact	1	*Nonreactive	1	100.00%
SY-2		* Denotes Correct Response		
AG: All Methods	3			
		*Reactive	2	66.67%
		*Weakly Reactive	1	33.33%
Fisher HC Sure-Vue RPR	1	*Reactive	1	100.00%
Wampole Impact	1	*Reactive	1	100.00%
BD Macro-Vue RPR	1	*Weakly Reactive	1	100.00%
SY-3		* Denotes Correct Response		
AG: All Methods	3			
		*Reactive	3	100.00%
BD Macro-Vue RPR	1	*Reactive	1	100.00%
Fisher HC Sure-Vue RPR	1	*Reactive	1	100.00%
Wampole Impact	1	*Reactive	1	100.00%
SY-4		* Denotes Correct Response		
AG: All Methods	3			
		*Nonreactive	3	100.00%
BD Macro-Vue RPR	1	*Nonreactive	1	100.00%
Fisher HC Sure-Vue RPR	1	*Nonreactive	1	100.00%
Wampole Impact	1	*Nonreactive	1	100.00%
SY-5		* Denotes Correct Response		
AG: All Methods	3			
		*Reactive	3	100.00%
BD Macro-Vue RPR	1	*Reactive	1	100.00%
Fisher HC Sure-Vue RPR	1	*Reactive	1	100.00%
Wampole Impact	1	*Reactive	1	100.00%
776 - Throat Culture;780 - Throat & Genital Cultures	Labs	Response	Number	Percent
		Throat Culture		
TC-1		* Denotes Correct Response		
AG: All Methods	77			
		*Beta hemolysis, referred for ID/confirmation	1	1.30%
		*Positive for Group A Streptococcus	38	49.35%
		*Presumptive Positive for Group A Streptococcus	34	44.16%
		*Presumptive Streptococcus species	1	1.30%
		Negative for Group A Streptococcus	3	3.90%
Cultures	77	*Beta hemolysis, referred for ID/confirmation	1	1.30%
Cultures		*Positive for Group A Streptococcus	38	49.35%
Cultures		*Presumptive Positive for Group A Streptococcus	34	44.16%
Cultures		*Presumptive Streptococcus species	1	1.30%
Cultures		Negative for Group A Streptococcus	3	3.90%
TC-2		* Denotes Correct Response		
AG: All Methods	77			
		*Beta hemolysis, referred for ID/confirmation	3	3.90%
		*Negative for Group A Streptococcus	60	77.92%
		*No Growth - Strep Selective medium	1	1.30%
		*Presumptive Streptococcus species	1	1.30%
		*Streptococcus, beta-hemolytic, not Group A	3	3.90%



		No beta hemolysis	1	1.30%	
		Positive for Group A Streptococcus	5	6.49%	
		Presumptive Positive for Group A Streptococcus	3	3.90%	
Cultures	77	*Beta hemolysis, referred for ID/confirmation	3	3.90%	
Cultures		*Negative for Group A Streptococcus	60	77.92%	
Cultures		*No Growth - Strep Selective medium	1	1.30%	
Cultures		*Presumptive Streptococcus species	1	1.30%	
Cultures		*Streptococcus, beta-hemolytic, not Group A	3	3.90%	
Cultures		No beta hemolysis	1	1.30%	
Cultures		Positive for Group A Streptococcus	5	6.49%	
Cultures		Presumptive Positive for Group A Streptococcus	3	3.90%	
TC-3		* Denotes Correct Response			
AG: All Methods	79				
		*Growth, referred for identification	3	3.80%	
		*Negative for Group A Streptococcus	55	69.62%	
		*No beta hemolysis	5	6.33%	
		*No Growth - Strep Selective medium	1	1.27%	
		Positive for Group A Streptococcus	5	6.33%	
		Presumptive Positive for Group A Streptococcus	9	11.39%	
		Streptococcus, beta-hemolytic, not Group A	1	1.27%	
Cultures	79	*Growth, referred for identification	3	3.80%	
Cultures		*Negative for Group A Streptococcus	55	69.62%	
Cultures		*No beta hemolysis	5	6.33%	
Cultures		*No Growth - Strep Selective medium	1	1.27%	
Cultures		Positive for Group A Streptococcus	5	6.33%	
Cultures		Presumptive Positive for Group A Streptococcus	9	11.39%	
Cultures		Streptococcus, beta-hemolytic, not Group A	1	1.27%	
TC-4		* Denotes Correct Response			
AG: All Methods	98				
		*Beta hemolysis, referred for ID/confirmation	3	3.06%	
		*Growth, referred for identification	2	2.04%	
		*Positive for Group A Streptococcus	47	47.96%	
		*Presumptive Positive for Group A Streptococcus	40	40.82%	
		*Presumptive Streptococcus species	1	1.02%	
		*Streptococcus, beta-hemolytic, Group A	1	1.02%	
		Negative for Group A Streptococcus	4	4.08%	
Cultures	98	*Beta hemolysis, referred for ID/confirmation	3	3.06%	
Cultures		*Growth, referred for identification	2	2.04%	
Cultures		*Positive for Group A Streptococcus	47	47.96%	
Cultures		*Presumptive Positive for Group A Streptococcus	40	40.82%	
Cultures		*Presumptive Streptococcus species	1	1.02%	
Cultures		*Streptococcus, beta-hemolytic, Group A	1	1.02%	
Cultures		Negative for Group A Streptococcus	4	4.08%	
TC-5		* Denotes Correct Response			
AG: All Methods	98				
		*Negative for Group A Streptococcus	63	64.29%	
		*No beta hemolysis	5	5.10%	
		*No Growth - Strep Selective medium	23	23.47%	
		Beta hemolysis, referred for ID/confirmation	2	2.04%	
		Positive for Group A Streptococcus	2	2.04%	
		Presumptive Positive for Group A Streptococcus	3	3.06%	
Cultures	98	*Negative for Group A Streptococcus	63	64.29%	
Cultures		*No beta hemolysis	5	5.10%	
Cultures		*No Growth - Strep Selective medium	23	23.47%	
Cultures		Beta hemolysis, referred for ID/confirmation	2	2.04%	
Cultures		Positive for Group A Streptococcus	2	2.04%	
Cultures		Presumptive Positive for Group A Streptococcus	3	3.06%	
977 - Trichomonas Ag Det (Waived)		Labs	Response	Number	Percent
TRA-1		Trichomonas Ag Det (Waived)			
AG: All Methods		* Denotes Correct Response			
	72				
		*Positive	68	94.44%	
		Negative	4	5.56%	
Sekisui Diagnostics OSOM Trichomonas	72	*Positive	68	94.44%	
Sekisui Diagnostics OSOM Trichomonas		Negative	4	5.56%	
TRA-2		* Denotes Correct Response			
AG: All Methods	72				
		*Negative	66	91.67%	
		Positive	6	8.33%	



Sekisui Diagnostics OSOM Trichomonas		72	*Negative	66	91.67%
Sekisui Diagnostics OSOM Trichomonas			Positive	6	8.33%
640 - Urinalysis;642 - Urinalysis/hCG;639 - Complete Urinalysis Package;601 - Waived/PPM Tests II;609 - Waived/PPM Tests I;647 - Mini-Waived Package;628 - AFC Waived/PPM;999 - CVS Complete Urinalysis		Labs	Response	Number	Percent
UA-1			Urine Sp Gravity		
PG: Siemens Clinitek 10 / 100		16	* Denotes Correct Response		
			*1.015	7	43.75%
			*1.020	2	12.50%
			*1.025	5	31.25%
			1.010	1	6.25%
			1.030	1	6.25%
UA-1			* Denotes Correct Response		
PG: Siemens Clinitek 50		105			
			*1.020	51	48.57%
			*1.025	44	41.90%
			1.015	9	8.57%
			1.030	1	0.95%
UA-1			* Denotes Correct Response		
PG: Siemens Clinitek 500		4			
			*1.020	2	50.00%
			*1.025	2	50.00%
UA-1			* Denotes Correct Response		
PG: Siemens Clinitek Atlas		1			
			*1.025	1	100.00%
UA-1			* Denotes Correct Response		
PG: Siemens Clinitek Status		622			
			*1.015	5	0.80%
			*1.020	157	25.24%
			*1.025	437	70.26%
			*1.030	20	3.22%
			1.000	1	0.16%
			1.010	1	0.16%
			1.035	1	0.16%
UA-1			* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick		16			
			*1.020	3	18.75%
			*1.025	12	75.00%
			>=1.030	1	6.25%
UA-1			* Denotes Correct Response		
PG: Siemens Multistix - Visual		286			
			*1.010	43	15.03%
			*1.015	142	49.65%
			*1.020	51	17.83%
			*1.025	43	15.03%
			1.000	3	1.05%
			1.005	1	0.35%
			1.030	3	1.05%
UA-1			* Denotes Correct Response		
PG: Cole Taylor Strips		1			
			*1.020	1	100.00%
UA-1			* Denotes Correct Response		
PG: Germaine Aimstick		4			
			*1.020	3	75.00%
			*1.025	1	25.00%
UA-1			* Denotes Correct Response		
PG: Henry Schein Urispec		5			
			*1.010	1	20.00%
			*1.015	4	80.00%
UA-1			* Denotes Correct Response		
PG: Hypoguard DiaScreen - Visual		2			
			*1.015	2	100.00%
UA-1			* Denotes Correct Response		
PG: Iris Diagnostics vChem Strips		1			
			*1.015	1	100.00%
UA-1			* Denotes Correct Response		



PG: Peer Group	2			
		*1.015	1	50.00%
		*1.025	1	50.00%
UA-1		* Denotes Correct Response		
PG: PSS Select Optima II Urine Analyzer	1			
		*1.025	1	100.00%
UA-1		* Denotes Correct Response		
PG: PSS Consult/Select Reagent Strips	5			
		*1.015	2	40.00%
		*1.020	2	40.00%
		*1.025	1	20.00%
UA-1		* Denotes Correct Response		
PG: Quidel QuickVue UrinChek - Visual	1			
		*1.015	1	100.00%
UA-1		* Denotes Correct Response		
PG: Diag Test Group Clarity Urochek	35			
		*1.015	4	11.43%
		*1.020	22	62.86%
		*1.025	6	17.14%
		1.010	3	8.57%
UA-1		* Denotes Correct Response		
PG: Roche Chemstrip 101 Analyzer	3			
		*1.010	2	66.67%
		*1.020	1	33.33%
UA-1		* Denotes Correct Response		
PG: Roche Chemstrips - Visual	65			
		*1.010	25	38.46%
		*1.015	37	56.92%
		1.005	2	3.08%
		1.020	1	1.54%
UA-1		* Denotes Correct Response		
PG: Roche Criterion Analyzer	1			
		*1.010	1	100.00%
UA-1		* Denotes Correct Response		
PG: Roche SuperUA/ChemstripUA	1			
		*1.010	1	100.00%
UA-1		* Denotes Correct Response		
PG: Roche Urisys	51			
		*1.010	35	68.63%
		*1.015	12	23.53%
		*1.020	2	3.92%
		1.005	1	1.96%
		1.030	1	1.96%
UA-1		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	29			
		*1.010	16	55.17%
		*1.015	12	41.38%
		1.020	1	3.45%
UA-1		* Denotes Correct Response		
PG: Teco (TC) Diag. URS-10	1			
		*1.025	1	100.00%
UA-1		* Denotes Correct Response		
PG: UriScan Reagent Strips	6			
		*1.020	1	16.67%
		*1.025	4	66.67%
		1.010	1	16.67%
UA-1		* Denotes Correct Response		
PG: Medi-Lab Performance Urine Strip Analyzer	1			
		*1.025	1	100.00%
UA-1		* Denotes Correct Response		
PG: Jant Pharm Accutest URS 10	1			
		*1.015	1	100.00%
UA-1		* Denotes Correct Response		
PG: Medi-Lab Performance Urine Strip - Visual	10			
		*1.015	4	40.00%
		*1.020	6	60.00%
UA-1		* Denotes Correct Response		
PG: Siemens Clinitek Advantus	16			



		*1.015	1	6.25%
		*1.020	13	81.25%
		*1.025	2	12.50%
UA-1		* Denotes Correct Response		
PG: Consult Diag (PSS) Urinalysis Test Strips	13			
		*1.015	6	46.15%
		*1.020	2	15.38%
		*1.025	4	30.77%
		1.010	1	7.69%
UA-1		* Denotes Correct Response		
PG: Henry Schein One Step Plus Strip Reader	2			
		*1.015	1	50.00%
		*1.020	1	50.00%
UA-1		* Denotes Correct Response		
PG: Consult Diag (PSS) Urine Analyzer	11			
		*1.015	4	36.36%
		*1.020	5	45.45%
		*1.025	2	18.18%
UA-1		* Denotes Correct Response		
PG: McKesson 120 Urine Analyzer	34			
		*1.020	28	82.35%
		*1.025	5	14.71%
		1.015	1	2.94%
UA-1		* Denotes Correct Response		
PG: Medline Urinalysis Reagent Strips	2			
		*1.015	1	50.00%
		*1.020	1	50.00%
UA-1		* Denotes Correct Response		
PG: NDC ProAdvantage Urine Strip Analyzer	1			
		*1.020	1	100.00%
UA-1		* Denotes Correct Response		
PG: Moore Medical Urine Reagent Strips	1			
		*1.020	1	100.00%
UA-1		* Denotes Correct Response		
PG: Roche Cobas u 411	1			
		*1.010	1	100.00%
UA-1		* Denotes Correct Response		
PG: McKesson 10 SG Urine Dipstick	1			
		*1.020	1	100.00%
UA-1		* Denotes Correct Response		
PG: Siemens Clinitek Status	5			
		*200 mg/dL	1	20.00%
		*300 mg/dL	4	80.00%
UA-1		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	16			
		*Negative	16	100.00%
UA-1		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	30			
		*Negative	30	100.00%
UA-1		* Denotes Correct Response		
MG: Method Group	1302			
		*Negative	1297	99.62%
		Moderate (2+)	3	0.23%
		Trace/Small (1+)	2	0.15%
UA-1		* Denotes Correct Response		
MG: Method Group	30			
		*Negative	30	100.00%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*Negative	16	100.00%
UA-1		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	16			
		*Negative	16	100.00%



UA-1		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	30			
		*Negative	30	100.00%
UA-1		* Denotes Correct Response		
MG: Method Group	1320			
		*Negative	1309	99.17%
		Large (3+) / 80 mg/dL	1	0.08%
		Small (1+) / 15mg/dL	3	0.23%
		Trace/5 mg/dL	7	0.53%
UA-1		* Denotes Correct Response		
MG: Method Group	30			
		*Negative	30	100.00%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*Negative	16	100.00%
UA-1		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	16			
		*7.0	16	100.00%
UA-1		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	30			
		*7.0	30	100.00%
UA-1		* Denotes Correct Response		
MG: Method Group	1319			
		*6.5	23	1.74%
		*7.0	1067	80.89%
		*7.5	217	16.45%
		5.0 or less	3	0.23%
		6.0	2	0.15%
		8.0 or greater	7	0.53%
UA-1		* Denotes Correct Response		
MG: Method Group	30			
		*7.0	30	100.00%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*7.0	16	100.00%
UA-1		* Denotes Correct Response		
PG: Siemens Clinitek Status	623			
		*Large (3+)	596	95.67%
		Moderate (2+)	26	4.17%
		Negative	1	0.16%
UA-1		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	30			
		*250 Ery/uL	30	100.00%
UA-1		* Denotes Correct Response		
MG: Method Group	1325			
		*Large (3+)	1250	94.34%
		Moderate (2+)	60	4.53%
		Negative	8	0.60%
		Small (1+)	6	0.45%
		Trace	1	0.08%
UA-1		* Denotes Correct Response		
MG: Method Group	30			
		*250 Ery/uL	30	100.00%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%
		2+	2	12.50%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*3+	14	87.50%</



PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	16			
		*Negative	16	100.00%
UA-1		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	30			
		*Normal	29	96.67%
		250 mg/dL	1	3.33%
UA-1		* Denotes Correct Response		
MG: Method Group	1332			
		*Negative	1326	99.55%
		100 mg/dL	5	0.38%
		500 mg/dL	1	0.08%
UA-1		* Denotes Correct Response		
MG: Method Group	30			
		*Normal	29	96.67%
		250 mg/dL	1	3.33%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*Negative	16	100.00%
UA-1		Urine Protein - Qual		
		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	16			
		*1+	1	6.25%
		*2+	8	50.00%
		*3+	7	43.75%
UA-1		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	30			
		*100 mg/dL	12	40.00%
		*30 mg/dL	18	60.00%
UA-1		* Denotes Correct Response		
MG: Method Group	1333			
		*100 (2+)	665	49.89%
		*300-500 (3+)	600	45.01%
		>/= 1000 (4+)	22	1.65%
		30 (1+)	34	2.55%
		Negative	11	0.83%
		Trace	1	0.08%
UA-1		* Denotes Correct Response		
MG: Method Group	30			
		*100 mg/dL	12	40.00%
		*30 mg/dL	18	60.00%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*1+	1	6.25%
		*2+	8	50.00%
		*3+	7	43.75%
UA-1		Urine Urobilinogen		
		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	8			
		*3.2 umol/L	8	100.00%
UA-1		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Urobilinogen	7			
		*0.2 EU/dL	7	100.00%
UA-1		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	30			
		*Normal	30	100.00%
UA-1		* Denotes Correct Response		
MG: Method Group	1297			
		*0.2/Normal	1281	98.77%
		1.0 mg/dL	7	0.54%
		2.0 mg/dL	8	0.62%
		4.0 mg/dL	1	0.08%
UA-1		* Denotes Correct Response		
MG: Method Group	30			
		*Normal	30	100.00%
UA-1		* Denotes Correct Response		



MG: Method Group	15			
		*0.2 EU/dL	7	46.67%
		*3.2 umol/L	8	53.33%
		Urine Leuk Esterase		
UA-1		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	16			
		*Negative	16	100.00%
UA-1		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	30			
		*Negative	30	100.00%
UA-1		* Denotes Correct Response		
MG: Method Group	1324			
		*Negative	1319	99.62%
		Small (1+)	2	0.15%
		Trace	3	0.23%
UA-1		* Denotes Correct Response		
MG: Method Group	30			
		*Negative	30	100.00%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*Negative	16	100.00%
		Urine Nitrite		
UA-1		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	16			
		*Negative	16	100.00%
UA-1		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	30			
		*Negative	30	100.00%
UA-1		* Denotes Correct Response		
MG: Method Group	1321			
		*Negative	1316	99.62%
		Positive	5	0.38%
UA-1		* Denotes Correct Response		
MG: Method Group	30			
		*Negative	30	100.00%
UA-1		* Denotes Correct Response		
MG: Method Group	16			
		*Negative	16	100.00%
		Urine Microalbumin		
UA-1		* Denotes Correct Response		
AG: All Methods	142			
		*100 mg/L (3+)	52	36.62%
		*150 mg/L	63	44.37%
		*50 mg/L (2+)	16	11.27%
		20/30 mg/L (Small)	4	2.82%
		Negative	7	4.93%
Siemens Clinitek Status	44	*100 mg/L (3+)	1	2.27%
Siemens Multistix - Visual	3	*100 mg/L (3+)	3	100.00%
Roche Chemstrip Micral	70	*100 mg/L (3+)	48	68.57%
Siemens Clinitek 10 / 100	2	*150 mg/L	2	100.00%
Siemens Clinitek 50	20	*150 mg/L	20	100.00%
Siemens Clinitek Status	44	*150 mg/L	39	88.64%
Roche Chemstrip Micral	70	*150 mg/L	1	1.43%
Hemocue Albumin 201	1	*150 mg/L	1	100.00%
Roche Chemstrip Micral	70	*50 mg/L (2+)	15	21.43%
Roche Chemstrips - Visual	2	*50 mg/L (2+)	1	50.00%
Siemens Clinitek Status	44	20/30 mg/L (Small)	3	6.82%
Roche Chemstrip Micral	70	20/30 mg/L (Small)	1	1.43%
Siemens Clinitek Status	44	Negative	1	2.27%
Roche Chemstrip Micral	70	Negative	5	7.14%
Roche Chemstrips - Visual	2	Negative	1	50.00%
		Fluid pH		
UA-1		* Denotes Correct Response		
PG: Nitrazine Paper	10			
		*6.5	1	10.00%
		*7.0	4	40.00%



		*7.5	5	50.00%
		Urine hCG		
UA-1		* Denotes Correct Response		
AG: All Methods	1192			
		*Positive	1179	98.91%
		Negative	13	1.09%
Abbott Signify hCG Combo	1	*Positive	1	100.00%
ASI ProPhase Plus	1	*Positive	1	100.00%
Siemens Clinitek Status	14	*Positive	14	100.00%
Beck Coulter ICON 20 HCG	7	*Positive	7	100.00%
Beck Coulter ICON 25 hCG	78	*Positive	78	100.00%
Alere Acceava HCG Basic II	23	*Positive	22	95.65%
Alere Acceava hCG Combo	5	*Positive	5	100.00%
Alere Acceava hCG-Urine	15	*Positive	15	100.00%
Conception Tech One Step hCG	2	*Positive	2	100.00%
Diag Test Group Clarity hCG Plus Urine Cassettes	7	*Positive	7	100.00%
Fisher HC Sure-Vue hCG STAT	5	*Positive	5	100.00%
Fisher HC Sure-Vue Urine hCG	16	*Positive	16	100.00%
Fisher HC Sure-Vue U/S hCG	6	*Positive	6	100.00%
Sekisui Diagnostics OSOM Urine	16	*Positive	16	100.00%
Sekisui Diagnostics OSOM Card Pregnancy	32	*Positive	31	96.88%
Sekisui Diagnostics OSOM hCG Combo	50	*Positive	49	98.00%
Henry Schein One Step + (hCG)	47	*Positive	47	100.00%
Immunostics Detector Combo	3	*Positive	2	66.67%
Immunostics hCG Detector-urine	7	*Positive	7	100.00%
Instant Tech. i Pregnancy	4	*Positive	4	100.00%
Inverness Clearview Combo II	4	*Positive	4	100.00%
Inv Clearview Urine - Cassette	3	*Positive	3	100.00%
Inverness Signify hCG combo	4	*Positive	4	100.00%
Inverness Signify hCG urine	5	*Positive	5	100.00%
LifeSign Status hCG	1	*Positive	1	100.00%
McKesson Medi-Lab hCG - Urine	41	*Positive	41	100.00%
Meridian Immunocard Stat!	1	*Positive	1	100.00%
Permaxim RediScreen	4	*Positive	4	100.00%
Polymedco Poly stat hCG Urine	1	*Positive	1	100.00%
Quidel QuickVue + hCG	30	*Positive	30	100.00%
Quidel QuickVue One-Step hCG Combo U/S	94	*Positive	93	98.94%
Quidel QuickVue One-Step Urine	192	*Positive	190	98.96%
Quidel RapidVue hCG - Urine	10	*Positive	10	100.00%
SAS Ultra HCG	3	*Positive	3	100.00%
Select Diag hCG Cassette (urine)	23	*Positive	22	95.65%
Select Diag hCG Combo	11	*Positive	11	100.00%
Select Diag hCG Dipstick	7	*Positive	7	100.00%
SP (Cardinal) Brand combo	16	*Positive	16	100.00%
SP (Cardinal) Brand-cassette	8	*Positive	8	100.00%
SP Brand (Cardinal) HCG Combo	32	*Positive	32	100.00%
Stanbio Qupid	5	*Positive	5	100.00%
Stanbio QuPID Plus	1	*Positive	1	100.00%
Stanbio QuStick	2	*Positive	2	100.00%
Stanbio TRUE hCG	4	*Positive	4	100.00%
MooreBrand hCG Cassette	5	*Positive	5	100.00%
Nova Scientific Century S/U hCG	1	*Positive	1	100.00%
Diag Test Group Clarity hCG Urine Strips	2	*Positive	2	100.00%
Jant AccuTest	7	*Positive	7	100.00%
Becton Dickinson Precise One-Step	1	*Positive	1	100.00%
Consult Diag (PSS) hCG Urine Cassette	203	*Positive	200	98.52%
CLIAwaived Inc Pregnancy Urine Dip-Strip	5	*Positive	4	80.00%
Germaine AIM Step Combo Pregnancy	1	*Positive	1	100.00%
NDC ProAdvantage Urine hCG Cassette	5	*Positive	5	100.00%
Consult Diag (PSS) hCG Urine Dipstick	34	*Positive	34	100.00%
SP (Cardinal) Brand Dipstick	9	*Positive	9	100.00%
Consult Diag hCG Combo Cassette (Urine)	10	*Positive	10	100.00%
McKesson Medi-Lab Performance hCG Urine Cassette	10	*Positive	9	90.00%
Alere hCG Combo Cassette (25 mIU/mL)	3	*Positive	3	100.00%
Alere hCG Cassette (25 mIU/mL)	12	*Positive	12	100.00%
Alere hCG Combo Cassette (20/10 mIU/mL)	1	*Positive	1	100.00%
Alere hCG Dipstick	5	*Positive	5	100.00%
Medline Pregnancy Urine Dip-Strip	2	*Positive	2	100.00%
Medline Pregnancy Urine Cassette Test	1	*Positive	1	100.00%



McKesson hCG Urine Test Dipstick	5	*Positive	5	100.00%
Quidel QuickVue hCG Urine Test	19	*Positive	19	100.00%
MooreBrand hCG Dipstick	3	*Positive	3	100.00%
Good Sense One Step Pregnancy Test	1	*Positive	1	100.00%
Henry Schein One Step + HCG Combo	1	*Positive	1	100.00%
Henry Schein One Step hCG Urine Cassette	4	*Positive	4	100.00%
Alere Aceava HCG Basic II	23	Negative	1	4.35%
Sekisui Diagnostics OSOM Card Pregnancy	32	Negative	1	3.13%
Sekisui Diagnostics OSOM hCG Combo	50	Negative	1	2.00%
Immunostics Detector Combo	3	Negative	1	33.33%
Quidel QuickVue One-Step hCG Combo U/S	94	Negative	1	1.06%
Quidel QuickVue One-Step Urine	192	Negative	2	1.04%
Select Diag hCG Cassette (urine)	23	Negative	1	4.35%
Consult Diag (PSS) hCG Urine Cassette	203	Negative	3	1.48%
CLIAwaived Inc Pregnancy Urine Dip-Strip	5	Negative	1	20.00%
McKesson Medi-Lab Performance hCG Urine Cassette	10	Negative	1	10.00%
661 - Urine Colony Count;784 - Presump ID/Colony Count - Urine;785 - Presump ID/Colony Count-Urine & Strep Combo;775 - Urine Culture;778 - Urine & Throat Cultures;779 - Urine & Genital Cultures;781 - Urine, Throat & Genital Cultures;782 - Complete Bacteri	Labs	Response	Number	Percent
UC-1		Susceptibility Cefazolin		
UC-1		* Denotes Correct Response		
AG: All Methods	2			
		*Resistant	2	100.00%
Susceptibility	2	*Resistant	2	100.00%
UC-1		Susceptibility Cefepime		
UC-1		* Denotes Correct Response		
AG: All Methods	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
UC-1		Susceptibility Cefotaxime		
UC-1		* Denotes Correct Response		
AG: All Methods	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
UC-1		Urine Culture		
UC-1		* Denotes Correct Response		
AG: All Methods	33			
		*Enterobacter aerogenes	1	3.03%
		*Enterobacter species, NOS	2	6.06%
		*Gram Negative organism	4	12.12%
		*Growth, referred for identification	17	51.52%
		*Presumptive Enterobacter species	1	3.03%
		*Presumptive Klebsiella species	2	6.06%
		*Uricult - Growth (Positive)	5	15.15%
		Uricult - No Growth (Negative)	1	3.03%
Cultures	33	*Enterobacter aerogenes	1	3.03%
Cultures		*Enterobacter species, NOS	2	6.06%
Cultures		*Gram Negative organism	4	12.12%
Cultures		*Growth, referred for identification	17	51.52%
Cultures		*Presumptive Enterobacter species	1	3.03%
Cultures		*Presumptive Klebsiella species	2	6.06%
Cultures		*Uricult - Growth (Positive)	5	15.15%
Cultures		Uricult - No Growth (Negative)	1	3.03%
UC-2		* Denotes Correct Response		
AG: All Methods	33			
		*Escherichia coli	4	12.12%
		*Gram Negative organism	4	12.12%
		*Growth, referred for identification	17	51.52%
		*Presumptive E. coli	2	6.06%
		*Uricult - Growth (Positive)	6	18.18%
Cultures	33	*Escherichia coli	4	12.12%
Cultures		*Gram Negative organism	4	12.12%
Cultures		*Growth, referred for identification	17	51.52%
Cultures		*Presumptive E. coli	2	6.06%
Cultures		*Uricult - Growth (Positive)	6	18.18%
UC-3		* Denotes Correct Response		
AG: All Methods	31			



		*Enterococcus faecium	2	6.45%
		*Gram Positive organism	2	6.45%
		*Growth, referred for identification	19	61.29%
		*Streptococcus species, NOS	1	3.23%
		*Uricult - No Growth (Negative)	4	12.90%
		No Growth	1	3.23%
		Uricult - Growth (Positive)	2	6.45%
Cultures	31	*Enterococcus faecium	2	6.45%
Cultures		*Gram Positive organism	2	6.45%
Cultures		*Growth, referred for identification	19	61.29%
Cultures		*Streptococcus species, NOS	1	3.23%
Cultures		*Uricult - No Growth (Negative)	4	12.90%
Cultures		No Growth	1	3.23%
Cultures		Uricult - Growth (Positive)	2	6.45%
UC-4		* Denotes Correct Response		
AG: All Methods	7			
		*Gram Negative organism	2	28.57%
		*Growth, referred for identification	1	14.29%
		*Presumptive Pseudomonas species	1	14.29%
		*Pseudomonas aeruginosa	2	28.57%
		*Uricult - Growth (Positive)	1	14.29%
Cultures	7	*Gram Negative organism	2	28.57%
Cultures		*Growth, referred for identification	1	14.29%
Cultures		*Presumptive Pseudomonas species	1	14.29%
Cultures		*Pseudomonas aeruginosa	2	28.57%
Cultures		*Uricult - Growth (Positive)	1	14.29%
UC-5		* Denotes Correct Response		
AG: All Methods	7			
		*Gram Negative organism	3	42.86%
		*Growth, referred for identification	1	14.29%
		*Presumptive Proteus species	1	14.29%
		*Proteus species NOS	1	14.29%
		*Uricult - Growth (Positive)	1	14.29%
Cultures	7	*Gram Negative organism	3	42.86%
Cultures		*Growth, referred for identification	1	14.29%
Cultures		*Presumptive Proteus species	1	14.29%
Cultures		*Proteus species NOS	1	14.29%
Cultures		*Uricult - Growth (Positive)	1	14.29%
UC-1		Susceptibility Cefpodoxime		
UC-1		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
UC-1		Susceptibility Cefazidime		
UC-1		* Denotes Correct Response		
AG: All Methods	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
UC-1		Susceptibility Ceftriaxone		
UC-1		* Denotes Correct Response		
AG: All Methods	6			
		*Sensitive	6	100.00%
Susceptibility	6	*Sensitive	6	100.00%
UC-1		Susceptibility Cephalothin		
UC-1		* Denotes Correct Response		
AG: All Methods	6			
		*Resistant	6	100.00%
Susceptibility	6	*Resistant	6	100.00%
UC-1		Susceptibility Ciprofloxacin		
UC-1		* Denotes Correct Response		
AG: All Methods	10			
		*Sensitive	10	100.00%
Susceptibility	10	*Sensitive	10	100.00%
UC-1		Susceptibility Gentamicin		
UC-1		* Denotes Correct Response		
AG: All Methods	6			
		*Sensitive	6	100.00%
Susceptibility	6	*Sensitive	6	100.00%



UC-1		Susceptibility Imipenem		
AG: All Methods		* Denotes Correct Response		
	4			
		*Sensitive	4	100.00%
Susceptibility	4	*Sensitive	4	100.00%
UC-1		Susceptibility Levofloxacin		
AG: All Methods		* Denotes Correct Response		
	6			
		*Sensitive	6	100.00%
Susceptibility	6	*Sensitive	6	100.00%
UC-1		Susceptibility Norfloxacin/Noroxin		
AG: All Methods		* Denotes Correct Response		
	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
UC-1		Susceptibility Piperacillin/Tazobactam		
AG: All Methods		* Denotes Correct Response		
	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
UC-1		Susceptibility Sulfamethoxazole		
AG: All Methods		* Denotes Correct Response		
	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
UC-1		Susceptibility Tetracycline		
AG: All Methods		* Denotes Correct Response		
	6			
		*Intermediate	2	33.33%
		*Sensitive	4	66.67%
Susceptibility	6	*Intermediate	2	33.33%
Susceptibility		*Sensitive	4	66.67%
UC-1		Susceptibility Ticarcillin/Clavulanate		
AG: All Methods		* Denotes Correct Response		
	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
UC-1		Susceptibility Tobramycin		
AG: All Methods		* Denotes Correct Response		
	4			
		*Sensitive	4	100.00%
Susceptibility	4	*Sensitive	4	100.00%
UC-1		Susceptibility Trimethoprim		
AG: All Methods		* Denotes Correct Response		
	2			
		*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
UC-1		Susceptibility Trimeth./Sulfameth.		
AG: All Methods		* Denotes Correct Response		
	8			
		*Sensitive	8	100.00%
Susceptibility	8	*Sensitive	8	100.00%
UC-1		Susceptibility Amikacin		
AG: All Methods		* Denotes Correct Response		
	2			
		*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
UC-1		Susceptibility Amox/Clavulanate		
AG: All Methods		* Denotes Correct Response		
	6			
		*Resistant	6	100.00%
Susceptibility	6	*Resistant	6	100.00%
UC-1		Susceptibility Ampicillin		
AG: All Methods		* Denotes Correct Response		
	8			
		*Resistant	7	87.50%
		Intermediate	1	12.50%



Susceptibility	8	*Resistant	7	87.50%
Susceptibility		Intermediate	1	12.50%
		Susceptibility Aztreonam		
UC-1		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
		Susceptibility Ertapenem		
UC-1		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
703 - Urine Drug Screen;713 - Urine Drug Screen (50 ml sample volume)	Labs	Response	Number	Percent
		Barbiturates		
UDS-1		* Denotes Correct Response		
AG: All Methods	127			
		*Negative	125	98.43%
		Positive	2	1.57%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand	1	*Negative	1	100.00%
E-Screen	5	*Negative	5	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	13	*Negative	13	100.00%
Instant Technologies i Screen	3	*Negative	3	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
BioTek	11	*Negative	11	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	2	*Negative	2	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	3	*Negative	3	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	1	*Negative	1	100.00%
Alere iCassette	34	*Negative	34	100.00%
ABMC Rapid Tox	1	*Negative	1	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	5	*Negative	5	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	2	66.67%
NexScreen Cup	1	*Negative	1	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
Discover Multi Panel Drug Test	2	*Negative	2	100.00%
Mindray 480	1	*Negative	1	100.00%
Discover Drug Test Card	1	Positive	1	100.00%
Noble Medical Insta-Screen Test Cup	3	Positive	1	33.33%
UDS-2		* Denotes Correct Response		
AG: All Methods	128			
		*Negative	126	98.44%
		Positive	2	1.56%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand	1	*Negative	1	100.00%
E-Screen	5	*Negative	5	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	13	*Negative	13	100.00%
Instant Technologies i Screen	3	*Negative	2	66.67%
MedTox Profile II	1	*Negative	1	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
BioTek	12	*Negative	12	100.00%



MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	2	*Negative	2	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	3	*Negative	3	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	1	*Negative	1	100.00%
Alere iCassette	34	*Negative	34	100.00%
ABMC Rapid Tox	1	*Negative	1	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	5	*Negative	5	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	2	66.67%
NexScreen Cup	1	*Negative	1	100.00%
Alere Immutest Multi-Drug Screen Card	1	*Negative	1	100.00%
Discover Multi Panel Drug Test	2	*Negative	2	100.00%
Mindray 480	1	*Negative	1	100.00%
Instant Technologies i Screen	3	Positive	1	33.33%
Noble Medical Insta-Screen Test Cup	3	Positive	1	33.33%
		Benzodiazepines		
UDS-1		* Denotes Correct Response		
AG: All Methods	133			
		*Negative	130	97.74%
		Positive	3	2.26%
American bioMedica Rapid Drug	1	*Negative	1	100.00%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand	2	*Negative	2	100.00%
E-Screen	5	*Negative	5	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	14	*Negative	14	100.00%
Instant Technologies i Screen	3	*Negative	3	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Noble Med Split Specimen Cup	2	*Negative	2	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
BioTek	11	*Negative	10	90.91%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	4	80.00%
ASC One Screen Drug Test Card	4	*Negative	4	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	3	*Negative	3	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	1	*Negative	1	100.00%
Alere iCassette	34	*Negative	34	100.00%
ABMC Rapid Tox	1	*Negative	1	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	5	*Negative	5	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	2	66.67%
NexScreen Cup	1	*Negative	1	100.00%
Alere Immutest Multi-Drug Screen Card	1	*Negative	1	100.00%
Discover Multi Panel Drug Test	2	*Negative	2	100.00%
Mindray 480	1	*Negative	1	100.00%
BioTek	11	Positive	1	9.09%
CLIAwaived Inc Rapid Drug Cup	5	Positive	1	20.00%
Noble Medical Insta-Screen Test Cup	3	Positive	1	33.33%
UDS-2		* Denotes Correct Response		



AG: All Methods	134			
		*Negative	133	99.25%
		Positive	1	0.75%
American bioMedica Rapid Drug	1	*Negative	1	100.00%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand	2	*Negative	2	100.00%
E-Screen	5	*Negative	5	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	14	*Negative	14	100.00%
Instant Technologies i Screen	3	*Negative	3	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Noble Med Split Specimen Cup	2	*Negative	2	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
BioTek	12	*Negative	12	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	4	*Negative	4	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	3	*Negative	3	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	1	*Negative	1	100.00%
Alere iCassette	34	*Negative	34	100.00%
ABMC Rapid Tox	1	*Negative	1	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	5	*Negative	5	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	2	66.67%
NexScreen Cup	1	*Negative	1	100.00%
Alere Immutest Multi-Drug Screen Card	1	*Negative	1	100.00%
Discover Multi Panel Drug Test	2	*Negative	2	100.00%
Mindray 480	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	Positive	1	33.33%
Amphetamines				
* Denotes Correct Response				
UDS-1				
AG: All Methods	136			
		*Negative	133	97.79%
		Positive	3	2.21%
American bioMedica Rapid Drug	1	*Negative	1	100.00%
AlereTriage	3	*Negative	2	66.67%
Siemens Dim/AR/ES/RxL/Xpand	3	*Negative	3	100.00%
E-Screen	2	*Negative	2	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	16	*Negative	16	100.00%
Instant Technologies i Screen	3	*Negative	3	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	2	*Negative	2	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	9	*Negative	9	100.00%
BioTek	11	*Negative	11	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	4	*Negative	4	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	3	*Negative	3	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	1	*Negative	1	100.00%
Alere iCassette	37	*Negative	37	100.00%
ABMC Rapid Tox	1	*Negative	1	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%



US Diagnostics UScreen Cup	5	*Negative	5	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	2	66.67%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
Discover Multi Panel Drug Test	2	*Negative	2	100.00%
Mindray 480	1	*Negative	1	100.00%
AlereTriage	3	Positive	1	33.33%
Noble Medical Insta-Screen Test Cup	3	Positive	1	33.33%
NexScreen Cup	1	Positive	1	100.00%
UDS-2		* Denotes Correct Response		
AG: All Methods	137			
		*Positive	129	94.16%
		Negative	8	5.84%
American bioMedica Rapid Drug	1	*Positive	1	100.00%
AlereTriage	3	*Positive	2	66.67%
Siemens Dim/AR/ES/RxL/Xpand	3	*Positive	3	100.00%
E-Screen	2	*Positive	2	100.00%
Express Diag. Express Check	1	*Positive	1	100.00%
Alere iCup Drugs of Abuse	16	*Positive	16	100.00%
Instant Technologies i Screen	3	*Positive	3	100.00%
MedTox Profile II	2	*Positive	2	100.00%
Noble Med Split Specimen Cup	2	*Positive	2	100.00%
Beckman (Olympus) AU	1	*Positive	1	100.00%
Innovacon Multi-Clin	9	*Positive	9	100.00%
BioTek	12	*Positive	11	91.67%
MedTox Profile III ER	1	*Positive	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Positive	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Positive	5	100.00%
ASC One Screen Drug Test Card	4	*Positive	4	100.00%
MedTox Profile V	2	*Positive	2	100.00%
Jant Accutest Multi Drug Screen	2	*Positive	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	3	*Positive	3	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Positive	4	100.00%
Mindray BS-200	1	*Positive	1	100.00%
Alere iCassette	37	*Positive	33	89.19%
ABMC Rapid Tox	1	*Positive	1	100.00%
Roche Hitachi 717	1	*Positive	1	100.00%
US Diagnostics UScreen Cup	5	*Positive	5	100.00%
Abbott Architect c4000	1	*Positive	1	100.00%
Discover Drug Test Card	1	*Positive	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Positive	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Positive	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Positive	2	66.67%
Alere Immatest Multi-Drug Screen Card	1	*Positive	1	100.00%
Discover Multi Panel Drug Test	2	*Positive	2	100.00%
Mindray 480	1	*Positive	1	100.00%
AlereTriage	3	Negative	1	33.33%
BioTek	12	Negative	1	8.33%
Alere iCassette	37	Negative	4	10.81%
Noble Medical Insta-Screen Test Cup	3	Negative	1	33.33%
NexScreen Cup	1	Negative	1	100.00%
UDS-1		* Denotes Correct Response		
AG: All Methods	147			
		*Negative	145	98.64%
		Positive	2	1.36%
American bioMedica Rapid Drug	1	*Negative	1	100.00%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand	3	*Negative	3	100.00%
E-Screen	11	*Negative	11	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	16	*Negative	16	100.00%
Instant Technologies i Screen	4	*Negative	4	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	2	*Negative	2	100.00%



Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
BioTek	11	*Negative	10	90.91%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	4	*Negative	4	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	3	*Negative	3	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	1	*Negative	1	100.00%
Alere iCassette	37	*Negative	37	100.00%
ABMC Rapid Tox	1	*Negative	1	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	5	*Negative	5	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	2	66.67%
NexScreen Cup	1	*Negative	1	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
Discover Multi Panel Drug Test	2	*Negative	2	100.00%
Mindray 480	1	*Negative	1	100.00%
BioTek	11	Positive	1	9.09%
Noble Medical Insta-Screen Test Cup	3	Positive	1	33.33%
UDS-2		* Denotes Correct Response		
AG: All Methods	148			
		*Negative	147	99.32%
		Positive	1	0.68%
American bioMedica Rapid Drug	1	*Negative	1	100.00%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand	3	*Negative	3	100.00%
E-Screen	11	*Negative	11	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	16	*Negative	16	100.00%
Instant Technologies i Screen	4	*Negative	4	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	2	*Negative	2	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
BioTek	12	*Negative	12	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	4	*Negative	4	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	3	*Negative	3	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	1	*Negative	1	100.00%
Alere iCassette	37	*Negative	37	100.00%
ABMC Rapid Tox	1	*Negative	1	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	5	*Negative	5	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	2	66.67%
NexScreen Cup	1	*Negative	1	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
Discover Multi Panel Drug Test	2	*Negative	2	100.00%
Mindray 480	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	Positive	1	33.33%
		Cocaine Metabolites		



UDS-1		* Denotes Correct Response		
AG: All Methods	146			
		*Positive	136	93.15%
		Negative	10	6.85%
American bioMedica Rapid Drug	1	*Positive	1	100.00%
AlereTriage	3	*Positive	2	66.67%
Siemens Dim/AR/ES/RxL/Xpand	3	*Positive	3	100.00%
E-Screen	11	*Positive	9	81.82%
Express Diag. Express Check	1	*Positive	1	100.00%
Alere iCup Drugs of Abuse	16	*Positive	16	100.00%
Instant Technologies i Screen	4	*Positive	3	75.00%
MedTox Profile II	2	*Positive	2	100.00%
Noble Med Split Specimen Cup	2	*Positive	2	100.00%
Beckman (Olympus) AU	1	*Positive	1	100.00%
Innovacon Multi-Clin	9	*Positive	9	100.00%
BioTek	11	*Positive	10	90.91%
MedTox Profile III ER	1	*Positive	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Positive	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Positive	4	80.00%
ASC One Screen Drug Test Card	4	*Positive	4	100.00%
MedTox Profile V	2	*Positive	2	100.00%
Jant Accutest Multi Drug Screen	2	*Positive	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	3	*Positive	3	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Positive	4	100.00%
Mindray BS-200	1	*Positive	1	100.00%
Alere iCassette	37	*Positive	35	94.59%
ABMC Rapid Tox	1	*Positive	1	100.00%
Roche Hitachi 717	1	*Positive	1	100.00%
US Diagnostics UScreen Cup	5	*Positive	5	100.00%
Abbott Architect c4000	1	*Positive	1	100.00%
Discover Drug Test Card	1	*Positive	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Positive	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Positive	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Positive	2	66.67%
Alere Immatest Multi-Drug Screen Card	1	*Positive	1	100.00%
Discover Multi Panel Drug Test	2	*Positive	2	100.00%
Mindray 480	1	*Positive	1	100.00%
AlereTriage	3	Negative	1	33.33%
E-Screen	11	Negative	2	18.18%
Instant Technologies i Screen	4	Negative	1	25.00%
BioTek	11	Negative	1	9.09%
CLIAwaived Inc Rapid Drug Cup	5	Negative	1	20.00%
Alere iCassette	37	Negative	2	5.41%
Noble Medical Insta-Screen Test Cup	3	Negative	1	33.33%
NexScreen Cup	1	Negative	1	100.00%
UDS-2		* Denotes Correct Response		
AG: All Methods	147			
		*Negative	141	95.92%
		Positive	6	4.08%
American bioMedica Rapid Drug	1	*Negative	1	100.00%
AlereTriage	3	*Negative	2	66.67%
Siemens Dim/AR/ES/RxL/Xpand	3	*Negative	3	100.00%
E-Screen	11	*Negative	11	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	16	*Negative	16	100.00%
Instant Technologies i Screen	4	*Negative	4	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	2	*Negative	2	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	9	*Negative	9	100.00%
BioTek	12	*Negative	11	91.67%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	4	*Negative	4	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	3	*Negative	3	100.00%



Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	1	*Negative	1	100.00%
Alere iCassette	37	*Negative	35	94.59%
ABMC Rapid Tox	1	*Negative	1	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	5	*Negative	5	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	2	66.67%
Alere Immutest Multi-Drug Screen Card	1	*Negative	1	100.00%
Discover Multi Panel Drug Test	2	*Negative	2	100.00%
Mindray 480	1	*Negative	1	100.00%
AlereTriage	3	Positive	1	33.33%
BioTek	12	Positive	1	8.33%
Alere iCassette	37	Positive	2	5.41%
Noble Medical Insta-Screen Test Cup	3	Positive	1	33.33%
NexScreen Cup	1	Positive	1	100.00%
		Ethanol		
UDS-1		* Denotes Correct Response		
AG: All Methods	5			
		*Negative	5	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
Mindray BS-200	1	*Negative	1	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
Mindray 480	1	*Negative	1	100.00%
UDS-2		* Denotes Correct Response		
AG: All Methods	5			
		*Positive	5	100.00%
Beckman (Olympus) AU	1	*Positive	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Positive	1	100.00%
Mindray BS-200	1	*Positive	1	100.00%
Roche Hitachi 717	1	*Positive	1	100.00%
Mindray 480	1	*Positive	1	100.00%
		Acetaminophen		
UDS-1		* Denotes Correct Response		
AG: All Methods	4			
		*Negative	4	100.00%
AlereTriage	3	*Negative	3	100.00%
E-Screen	1	*Negative	1	100.00%
UDS-2		* Denotes Correct Response		
AG: All Methods	4			
		*Negative	4	100.00%
AlereTriage	3	*Negative	3	100.00%
E-Screen	1	*Negative	1	100.00%
		MDMA (Ecstasy)		
UDS-1		* Denotes Correct Response		
AG: All Methods	101			
		*Negative	99	98.02%
		Positive	2	1.98%
American bioMedica Rapid Drug	1	*Negative	1	100.00%
E-Screen	2	*Negative	2	100.00%
Alere iCup Drugs of Abuse	13	*Negative	13	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
BioTek	11	*Negative	11	100.00%
CLIAwaived Inc Rapid Drug Cup	4	*Negative	4	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	3	*Negative	3	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	2	*Negative	2	100.00%
Alere iCassette	34	*Negative	34	100.00%
US Diagnostics UScreen Cup	5	*Negative	5	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%



Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	2	66.67%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
Discover Multi Panel Drug Test	2	*Negative	2	100.00%
Noble Medical Insta-Screen Test Cup	3	Positive	1	33.33%
NexScreen Cup	1	Positive	1	100.00%
UDS-2		* Denotes Correct Response		
AG: All Methods	102			
		*Positive	96	94.12%
		Negative	6	5.88%
American bioMedica Rapid Drug	1	*Positive	1	100.00%
E-Screen	2	*Positive	1	50.00%
Alere iCup Drugs of Abuse	13	*Positive	12	92.31%
Noble Med Split Specimen Cup	1	*Positive	1	100.00%
Innovacon Multi-Clin	10	*Positive	10	100.00%
BioTek	12	*Positive	11	91.67%
CLIAwaived Inc Rapid Drug Cup	4	*Positive	4	100.00%
Jant Accutest Multi Drug Screen	2	*Positive	1	50.00%
Amedica Biotech AmediCheck Instant Test Cup	3	*Positive	3	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	2	*Positive	2	100.00%
Alere iCassette	34	*Positive	34	100.00%
US Diagnostics UScreen Cup	5	*Positive	5	100.00%
Discover Drug Test Card	1	*Positive	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Positive	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Positive	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Positive	2	66.67%
Alere Immatest Multi-Drug Screen Card	1	*Positive	1	100.00%
Discover Multi Panel Drug Test	2	*Positive	2	100.00%
E-Screen	2	Negative	1	50.00%
Alere iCup Drugs of Abuse	13	Negative	1	7.69%
BioTek	12	Negative	1	8.33%
Jant Accutest Multi Drug Screen	2	Negative	1	50.00%
Noble Medical Insta-Screen Test Cup	3	Negative	1	33.33%
NexScreen Cup	1	Negative	1	100.00%
UDS-1		* Denotes Correct Response		
AG: All Methods	81			
		*Negative	80	98.77%
		Positive	1	1.23%
American bioMedica Rapid Drug	1	*Negative	1	100.00%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand	2	*Negative	2	100.00%
E-Screen	5	*Negative	5	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	12	*Negative	12	100.00%
Instant Technologies i Screen	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
BioTek	11	*Negative	11	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	2	*Negative	2	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	3	*Negative	3	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	1	*Negative	1	100.00%
ABMC Rapid Tox	1	*Negative	1	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	5	*Negative	5	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	2	66.67%
NexScreen Cup	1	*Negative	1	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%



Discover Multi Panel Drug Test	2	*Negative	2	100.00%
Mindray 480	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	Positive	1	33.33%
UDS-2		* Denotes Correct Response		
AG: All Methods	81			
		*Negative	79	97.53%
		Positive	2	2.47%
American bioMedica Rapid Drug	1	*Negative	1	100.00%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand	2	*Negative	2	100.00%
E-Screen	5	*Negative	5	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	12	*Negative	12	100.00%
Instant Technologies i Screen	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
BioTek	12	*Negative	11	91.67%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	2	*Negative	2	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	3	*Negative	3	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	3	*Negative	3	100.00%
Mindray BS-200	1	*Negative	1	100.00%
ABMC Rapid Tox	1	*Negative	1	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	5	*Negative	5	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	2	66.67%
NexScreen Cup	1	*Negative	1	100.00%
Alere Immutest Multi-Drug Screen Card	1	*Negative	1	100.00%
Discover Multi Panel Drug Test	2	*Negative	2	100.00%
Mindray 480	1	*Negative	1	100.00%
BioTek	12	Positive	1	8.33%
Noble Medical Insta-Screen Test Cup	3	Positive	1	33.33%
		Methamphetamine		
UDS-1		* Denotes Correct Response		
AG: All Methods	88			
		*Negative	84	95.45%
		Positive	4	4.55%
American bioMedica Rapid Drug	1	*Negative	1	100.00%
AlereTriage	3	*Negative	2	66.67%
E-Screen	11	*Negative	11	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	15	*Negative	15	100.00%
Instant Technologies i Screen	3	*Negative	3	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Noble Med Split Specimen Cup	2	*Negative	2	100.00%
BioTek	11	*Negative	10	90.91%
MedTox Profile III ER	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	3	*Negative	3	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	3	*Negative	3	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Alere iCassette	1	*Negative	1	100.00%
ABMC Rapid Tox	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	5	*Negative	5	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	2	66.67%



Alere Immutest Multi-Drug Screen Card	1	*Negative	1	100.00%
Discover Multi Panel Drug Test	2	*Negative	2	100.00%
AlereTriage	3	Positive	1	33.33%
BioTek	11	Positive	1	9.09%
Noble Medical Insta-Screen Test Cup	3	Positive	1	33.33%
NexScreen Cup	1	Positive	1	100.00%
UDS-2		* Denotes Correct Response		
AG: All Methods	89			
		*Positive	79	88.76%
		Negative	10	11.24%
AlereTriage	3	*Positive	2	66.67%
E-Screen	11	*Positive	9	81.82%
Express Diag. Express Check	1	*Positive	1	100.00%
Alere iCup Drugs of Abuse	15	*Positive	15	100.00%
Instant Technologies i Screen	3	*Positive	3	100.00%
MedTox Profile II	1	*Positive	1	100.00%
Noble Med Split Specimen Cup	2	*Positive	2	100.00%
BioTek	12	*Positive	10	83.33%
MedTox Profile III ER	1	*Positive	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Positive	5	100.00%
ASC One Screen Drug Test Card	3	*Positive	2	66.67%
MedTox Profile V	2	*Positive	2	100.00%
Jant Accutest Multi Drug Screen	2	*Positive	1	50.00%
Amedica Biotech AmediCheck Instant Test Cup	3	*Positive	3	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Positive	4	100.00%
Alere iCassette	1	*Positive	1	100.00%
ABMC Rapid Tox	1	*Positive	1	100.00%
US Diagnostics UScreen Cup	5	*Positive	5	100.00%
Discover Drug Test Card	1	*Positive	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Positive	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Positive	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Positive	2	66.67%
Alere Immutest Multi-Drug Screen Card	1	*Positive	1	100.00%
Discover Multi Panel Drug Test	2	*Positive	2	100.00%
American bioMedica Rapid Drug	1	Negative	1	100.00%
AlereTriage	3	Negative	1	33.33%
E-Screen	11	Negative	2	18.18%
BioTek	12	Negative	2	16.67%
ASC One Screen Drug Test Card	3	Negative	1	33.33%
Jant Accutest Multi Drug Screen	2	Negative	1	50.00%
Noble Medical Insta-Screen Test Cup	3	Negative	1	33.33%
NexScreen Cup	1	Negative	1	100.00%
Opiates				
UDS-1		* Denotes Correct Response		
AG: All Methods	146			
		*Positive	134	91.78%
		Negative	12	8.22%
American bioMedica Rapid Drug	1	*Positive	1	100.00%
AlereTriage	3	*Positive	2	66.67%
Siemens Dim/AR/ES/RxL/Xpand	3	*Positive	3	100.00%
E-Screen	11	*Positive	8	72.73%
Express Diag. Express Check	1	*Positive	1	100.00%
Alere iCup Drugs of Abuse	16	*Positive	16	100.00%
Instant Technologies i Screen	4	*Positive	4	100.00%
MedTox Profile II	1	*Positive	1	100.00%
Noble Med Split Specimen Cup	2	*Positive	2	100.00%
Beckman (Olympus) AU	1	*Positive	1	100.00%
Innovacon Multi-Clin	10	*Positive	10	100.00%
BioTek	11	*Positive	10	90.91%
MedTox Profile III ER	1	*Positive	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Positive	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Positive	5	100.00%
ASC One Screen Drug Test Card	4	*Positive	3	75.00%
MedTox Profile V	2	*Positive	2	100.00%
Jant Accutest Multi Drug Screen	2	*Positive	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	3	*Positive	2	66.67%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Positive	3	75.00%



Mindray BS-200	1	*Positive	1	100.00%
Alere iCassette	37	*Positive	35	94.59%
ABMC Rapid Tox	1	*Positive	1	100.00%
Roche Hitachi 717	1	*Positive	1	100.00%
US Diagnostics UScreen Cup	5	*Positive	5	100.00%
Abbott Architect c4000	1	*Positive	1	100.00%
Discover Drug Test Card	1	*Positive	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Positive	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Positive	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Positive	2	66.67%
Alere Immatest Multi-Drug Screen Card	1	*Positive	1	100.00%
Discover Multi Panel Drug Test	2	*Positive	2	100.00%
Mindray 480	1	*Positive	1	100.00%
AlereTriage	3	Negative	1	33.33%
E-Screen	11	Negative	3	27.27%
BioTek	11	Negative	1	9.09%
ASC One Screen Drug Test Card	4	Negative	1	25.00%
Amedica Biotech AmediCheck Instant Test Cup	3	Negative	1	33.33%
American Screening Corp Reveal Multi-Drug Test Cup	4	Negative	1	25.00%
Alere iCassette	37	Negative	2	5.41%
Noble Medical Insta-Screen Test Cup	3	Negative	1	33.33%
NexScreen Cup	1	Negative	1	100.00%
UDS-2		* Denotes Correct Response		
AG: All Methods	147			
		*Negative	143	97.28%
		Positive	4	2.72%
American bioMedica Rapid Drug	1	*Negative	1	100.00%
AlereTriage	3	*Negative	2	66.67%
Siemens Dim/AR/ES/RxL/Xpand	3	*Negative	3	100.00%
E-Screen	11	*Negative	11	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	16	*Negative	16	100.00%
Instant Technologies i Screen	4	*Negative	4	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Noble Med Split Specimen Cup	2	*Negative	2	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
BioTek	12	*Negative	11	91.67%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	4	*Negative	4	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	3	*Negative	3	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	1	*Negative	1	100.00%
Alere iCassette	37	*Negative	37	100.00%
ABMC Rapid Tox	1	*Negative	1	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	5	*Negative	5	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	2	66.67%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
Discover Multi Panel Drug Test	2	*Negative	2	100.00%
Mindray 480	1	*Negative	1	100.00%
AlereTriage	3	Positive	1	33.33%
BioTek	12	Positive	1	8.33%
Noble Medical Insta-Screen Test Cup	3	Positive	1	33.33%
NexScreen Cup	1	Positive	1	100.00%
UDS-1		* Denotes Correct Response		
AG: All Methods	116			
		*Negative	114	98.28%
		Positive	2	1.72%



American bioMedica Rapid Drug	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	13	*Negative	13	100.00%
Instant Technologies i Screen	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
BioTek	11	*Negative	11	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	3	*Negative	3	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	3	*Negative	3	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	3	75.00%
Mindray BS-200	1	*Negative	1	100.00%
Alere iCassette	34	*Negative	34	100.00%
ABMC Rapid Tox	1	*Negative	1	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	5	*Negative	5	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	2	66.67%
NexScreen Cup	1	*Negative	1	100.00%
Alere Immutest Multi-Drug Screen Card	1	*Negative	1	100.00%
Discover Multi Panel Drug Test	2	*Negative	2	100.00%
Mindray 480	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	Positive	1	25.00%
Noble Medical Insta-Screen Test Cup	3	Positive	1	33.33%
UDS-2		* Denotes Correct Response		
AG: All Methods	116			
		*Negative	115	99.14%
		Positive	1	0.86%
American bioMedica Rapid Drug	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	13	*Negative	13	100.00%
Instant Technologies i Screen	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
BioTek	12	*Negative	12	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	3	*Negative	3	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	3	*Negative	3	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	3	*Negative	3	100.00%
Mindray BS-200	1	*Negative	1	100.00%
Alere iCassette	34	*Negative	34	100.00%
ABMC Rapid Tox	1	*Negative	1	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	5	*Negative	5	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	2	66.67%
NexScreen Cup	1	*Negative	1	100.00%
Alere Immutest Multi-Drug Screen Card	1	*Negative	1	100.00%
Discover Multi Panel Drug Test	2	*Negative	2	100.00%
Mindray 480	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	Positive	1	33.33%
PCP (Phencyclidine)				
UDS-1		* Denotes Correct Response		
AG: All Methods	125			
		*Negative	124	99.20%



		Positive	1	0.80%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand	2	*Negative	2	100.00%
E-Screen	11	*Negative	11	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	11	*Negative	11	100.00%
Instant Technologies i Screen	2	*Negative	2	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
BioTek	11	*Negative	11	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	2	*Negative	2	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	1	*Negative	1	100.00%
Alere iCassette	36	*Negative	36	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	2	*Negative	2	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	2	66.67%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
Discover Multi Panel Drug Test	2	*Negative	2	100.00%
Noble Medical Insta-Screen Test Cup	3	Positive	1	33.33%
UDS-2		* Denotes Correct Response		
AG: All Methods	126			
		*Negative	125	99.21%
		Positive	1	0.79%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand	2	*Negative	2	100.00%
E-Screen	11	*Negative	11	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	11	*Negative	11	100.00%
Instant Technologies i Screen	2	*Negative	2	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
BioTek	12	*Negative	12	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	2	*Negative	2	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	1	*Negative	1	100.00%
Alere iCassette	36	*Negative	36	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	2	*Negative	2	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	2	66.67%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
Discover Multi Panel Drug Test	2	*Negative	2	100.00%
Noble Medical Insta-Screen Test Cup	3	Positive	1	33.33%
UDS-1		Propoxyphene		
		* Denotes Correct Response		



AG: All Methods	60			
		*Negative	59	98.33%
		Positive	1	1.67%
E-Screen	3	*Negative	3	100.00%
Alere iCup Drugs of Abuse	6	*Negative	6	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	8	*Negative	8	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Alere iCassette	34	*Negative	34	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	1	*Negative	1	100.00%
NexScreen Cup	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	1	Positive	1	100.00%
UDS-2		* Denotes Correct Response		
AG: All Methods	60			
		*Negative	60	100.00%
E-Screen	3	*Negative	3	100.00%
Alere iCup Drugs of Abuse	6	*Negative	6	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	8	*Negative	8	100.00%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Alere iCassette	34	*Negative	34	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	1	*Negative	1	100.00%
NexScreen Cup	1	*Negative	1	100.00%
		Tricyclic Antidepressants		
UDS-1		* Denotes Correct Response		
AG: All Methods	83			
		*Positive	70	84.34%
		Negative	13	15.66%
AlereTriage	3	*Positive	2	66.67%
Alere iCup Drugs of Abuse	11	*Positive	11	100.00%
Instant Technologies i Screen	2	*Positive	1	50.00%
Noble Med Split Specimen Cup	1	*Positive	1	100.00%
Innovacon Multi-Clin	9	*Positive	7	77.78%
CLIAwaived Inc Rapid Drug Cup	3	*Positive	3	100.00%
MedTox Profile V	2	*Positive	2	100.00%
Jant Accutest Multi Drug Screen	2	*Positive	2	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
Alere iCassette	34	*Positive	30	88.24%
US Diagnostics UScreen Cup	3	*Positive	3	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Positive	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Positive	2	66.67%
Noble Medical Insta-Screen Test Cup	3	*Positive	2	66.67%
Alere Immatest Multi-Drug Screen Card	1	*Positive	1	100.00%
Mindray 480	1	*Positive	1	100.00%
AlereTriage	3	Negative	1	33.33%
Instant Technologies i Screen	2	Negative	1	50.00%
Innovacon Multi-Clin	9	Negative	2	22.22%
Alere iCassette	34	Negative	4	11.76%
Discover Drug Test Card	1	Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	Negative	1	33.33%
Noble Medical Insta-Screen Test Cup	3	Negative	1	33.33%
NexScreen Cup	1	Negative	1	100.00%
Discover Multi Panel Drug Test	1	Negative	1	100.00%
UDS-2		* Denotes Correct Response		
AG: All Methods	83			
		*Negative	78	93.98%
		Positive	5	6.02%
AlereTriage	3	*Negative	2	66.67%
Alere iCup Drugs of Abuse	11	*Negative	11	100.00%
Instant Technologies i Screen	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Innovacon Multi-Clin	9	*Negative	8	88.89%



CLIAwaived Inc Rapid Drug Cup	3	*Negative	3	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
Alere iCassette	34	*Negative	33	97.06%
US Diagnostics UScreen Cup	3	*Negative	3	100.00%
Discover Drug Test Card	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	2	66.67%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
Discover Multi Panel Drug Test	1	*Negative	1	100.00%
Mindray 480	1	*Negative	1	100.00%
AlereTriage	3	Positive	1	33.33%
Innovacon Multi-Clin	9	Positive	1	11.11%
Alere iCassette	34	Positive	1	2.94%
Noble Medical Insta-Screen Test Cup	3	Positive	1	33.33%
NexScreen Cup	1	Positive	1	100.00%
		Buprenorphine		
UDS-1		* Denotes Correct Response		
AG: All Methods	35			
		*Positive	27	77.14%
		Negative	8	22.86%
Alere iCup Drugs of Abuse	6	*Positive	4	66.67%
BioTek	11	*Positive	10	90.91%
Tokyo/Carolina Biolis 24i	1	*Positive	1	100.00%
ASC One Screen Drug Test Card	2	*Positive	2	100.00%
MedTox Profile V	1	*Positive	1	100.00%
Jant Accutest Multi Drug Screen	2	*Positive	1	50.00%
Mindray BS-200	1	*Positive	1	100.00%
Diag Test Group Clarity DOA Screen	1	*Positive	1	100.00%
Alere iScreen Drugs of Abuse Card	1	*Positive	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Positive	2	66.67%
Discover Multi Panel Drug Test	2	*Positive	2	100.00%
Mindray 480	1	*Positive	1	100.00%
Alere iCup Drugs of Abuse	6	Negative	2	33.33%
BioTek	11	Negative	1	9.09%
CLIAwaived Inc Rapid Drug Cup	2	Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	Negative	1	50.00%
Noble Medical Insta-Screen Test Cup	3	Negative	1	33.33%
NexScreen Cup	1	Negative	1	100.00%
UDS-2		* Denotes Correct Response		
AG: All Methods	36			
		*Negative	33	91.67%
		Positive	3	8.33%
Alere iCup Drugs of Abuse	6	*Negative	6	100.00%
BioTek	12	*Negative	11	91.67%
Tokyo/Carolina Biolis 24i	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	2	*Negative	2	100.00%
ASC One Screen Drug Test Card	2	*Negative	2	100.00%
MedTox Profile V	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Mindray BS-200	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	1	*Negative	1	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	2	66.67%
Discover Multi Panel Drug Test	2	*Negative	2	100.00%
Mindray 480	1	*Negative	1	100.00%
BioTek	12	Positive	1	8.33%
Noble Medical Insta-Screen Test Cup	3	Positive	1	33.33%
NexScreen Cup	1	Positive	1	100.00%
756 - Viral Antibodies	Labs	Response	Number	Percent
		HSV Ab (Herpes Simplex Virus)		
VA-1		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Biokit HSV-II	2	*Positive	2	100.00%



770 - Viral Antigen Detection;792 - Viral Antigen Detection - Waived Methods;798 - Viral Antigen Detection (ROT);797 - Viral Antigen Detection II	Labs	Response	Number	Percent
Rotavirus - Ag Detect				
VAD-1				
* Denotes Correct Response				
AG: All Methods	9			
		*Negative	9	100.00%
Meridian Immucard Stat! Rotavirus	5	*Negative	5	100.00%
SAS Rota-Test	4	*Negative	4	100.00%
VAD-2				
* Denotes Correct Response				
AG: All Methods	9			
		*Positive	9	100.00%
Meridian Immucard Stat! Rotavirus	5	*Positive	5	100.00%
SAS Rota-Test	4	*Positive	4	100.00%
VAD-3				
* Denotes Correct Response				
AG: All Methods	9			
		*Negative	9	100.00%
Meridian Immucard Stat! Rotavirus	5	*Negative	5	100.00%
SAS Rota-Test	4	*Negative	4	100.00%
VAD-4				
* Denotes Correct Response				
AG: All Methods	9			
		*Negative	8	88.89%
		Positive	1	11.11%
Meridian Immucard Stat! Rotavirus	5	*Negative	4	80.00%
SAS Rota-Test	4	*Negative	4	100.00%
Meridian Immucard Stat! Rotavirus	5	Positive	1	20.00%
VAD-5				
* Denotes Correct Response				
AG: All Methods	9			
		*Negative	9	100.00%
Meridian Immucard Stat! Rotavirus	5	*Negative	5	100.00%
SAS Rota-Test	4	*Negative	4	100.00%
Influenza A - Ag Detect				
VAD-1				
* Denotes Correct Response				
AG: All Methods	533			
		*Positive	527	98.87%
		Negative	6	1.13%
Binax Now Flu A	19	*Positive	19	100.00%
Alere BinaxNOW Influenza A+B Card	128	*Positive	127	99.22%
Sekisui Diagnostics OSOM Influenza A & B	22	*Positive	22	100.00%
Quidel QuickVue A+B	91	*Positive	90	98.90%
Remel Xpect	1	*Positive	1	100.00%
Quidel Sofia Influenza A+B	53	*Positive	53	100.00%
Alere Influenza A & B	31	*Positive	31	100.00%
BD Veritor Flu A+B	160	*Positive	156	97.50%
Henry Schein OneStep + Influenza A&B (distinguish)	12	*Positive	12	100.00%
Cepheid GeneXpert - Xpert Flu	2	*Positive	2	100.00%
Alere i Influenza A & B	10	*Positive	10	100.00%
Roche Cobas Liat Influenza A/B	2	*Positive	2	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	1	*Positive	1	100.00%
Alere BinaxNOW Influenza A+B Card	128	Negative	1	0.78%
Quidel QuickVue A+B	91	Negative	1	1.10%
BD Veritor Flu A+B	160	Negative	4	2.50%
VAD-2				
* Denotes Correct Response				
AG: All Methods	533			
		*Negative	519	97.37%
		Positive	14	2.63%
Binax Now Flu A	19	*Negative	18	94.74%
Alere BinaxNOW Influenza A+B Card	128	*Negative	124	96.88%
Sekisui Diagnostics OSOM Influenza A & B	22	*Negative	22	100.00%
Quidel QuickVue A+B	90	*Negative	90	100.00%
Remel Xpect	1	*Negative	1	100.00%
Quidel Sofia Influenza A+B	53	*Negative	53	100.00%
Alere Influenza A & B	31	*Negative	31	100.00%
BD Veritor Flu A+B	160	*Negative	154	96.25%
Henry Schein OneStep + Influenza A&B (distinguish)	12	*Negative	10	83.33%
Cepheid GeneXpert - Xpert Flu	2	*Negative	2	100.00%
Alere i Influenza A & B	10	*Negative	10	100.00%
Roche Cobas Liat Influenza A/B	2	*Negative	2	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	1	*Negative	1	100.00%



Binax Now Flu A	19	Positive	1	5.26%
Alere BinaxNOW Influenza A+B Card	128	Positive	4	3.13%
BD Veritor Flu A+B	160	Positive	6	3.75%
Henry Schein OneStep + Influenza A&B (distinguish)	12	Positive	2	16.67%
VAD-3		* Denotes Correct Response		
AG: All Methods	66			
		*Negative	65	98.48%
		Positive	1	1.52%
Alere BinaxNOW Influenza A+B Card	8	*Negative	8	100.00%
Sekisui Diagnostics OSOM Influenza A & B	13	*Negative	13	100.00%
Quidel QuickVue A+B	22	*Negative	22	100.00%
Remel Xpect	1	*Negative	1	100.00%
Quidel Sofia Influenza A+B	4	*Negative	4	100.00%
Alere Influenza A & B	3	*Negative	3	100.00%
BD Veritor Flu A+B	12	*Negative	11	91.67%
Cepheid GeneXpert - Xpert Flu	2	*Negative	2	100.00%
Alere i Influenza A & B	1	*Negative	1	100.00%
BD Veritor Flu A+B	12	Positive	1	8.33%
VAD-4		* Denotes Correct Response		
AG: All Methods	66			
		*Positive	66	100.00%
Alere BinaxNOW Influenza A+B Card	8	*Positive	8	100.00%
Sekisui Diagnostics OSOM Influenza A & B	13	*Positive	13	100.00%
Quidel QuickVue A+B	22	*Positive	22	100.00%
Remel Xpect	1	*Positive	1	100.00%
Quidel Sofia Influenza A+B	4	*Positive	4	100.00%
Alere Influenza A & B	3	*Positive	3	100.00%
BD Veritor Flu A+B	12	*Positive	12	100.00%
Cepheid GeneXpert - Xpert Flu	2	*Positive	2	100.00%
Alere i Influenza A & B	1	*Positive	1	100.00%
VAD-5		* Denotes Correct Response		
AG: All Methods	66			
		*Negative	65	98.48%
		Positive	1	1.52%
Alere BinaxNOW Influenza A+B Card	8	*Negative	8	100.00%
Sekisui Diagnostics OSOM Influenza A & B	13	*Negative	13	100.00%
Quidel QuickVue A+B	22	*Negative	21	95.45%
Remel Xpect	1	*Negative	1	100.00%
Quidel Sofia Influenza A+B	4	*Negative	4	100.00%
Alere Influenza A & B	3	*Negative	3	100.00%
BD Veritor Flu A+B	12	*Negative	12	100.00%
Cepheid GeneXpert - Xpert Flu	2	*Negative	2	100.00%
Alere i Influenza A & B	1	*Negative	1	100.00%
Quidel QuickVue A+B	22	Positive	1	4.55%
		Influenza A/B - Ag Detect		
VAD-1		* Denotes Correct Response		
AG: All Methods	34			
		*Positive	32	94.12%
		Negative	2	5.88%
Alere BinaxNOW Influenza A+B Card	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Influenza A & B	1	*Positive	1	100.00%
Quidel QuickVue	28	*Positive	27	96.43%
Henry Schein OneStep + Influenza A/B	4	*Positive	3	75.00%
Quidel QuickVue	28	Negative	1	3.57%
Henry Schein OneStep + Influenza A/B	4	Negative	1	25.00%
VAD-2		* Denotes Correct Response		
AG: All Methods	34			
		*Negative	34	100.00%
Alere BinaxNOW Influenza A+B Card	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM Influenza A & B	1	*Negative	1	100.00%
Quidel QuickVue	28	*Negative	28	100.00%
Henry Schein OneStep + Influenza A/B	4	*Negative	4	100.00%
VAD-3		* Denotes Correct Response		
AG: All Methods	5			
		*Negative	5	100.00%
Quidel QuickVue	5	*Negative	5	100.00%
VAD-4		* Denotes Correct Response		
AG: All Methods	5			



		*Positive	5	100.00%
Quidel QuickVue	5	*Positive	5	100.00%
VAD-5		* Denotes Correct Response		
AG: All Methods	5			
		*Positive	5	100.00%
Quidel QuickVue	5	*Positive	5	100.00%
		Influenza B - Aq Detect		
VAD-1		* Denotes Correct Response		
AG: All Methods	528			
		*Negative	520	98.48%
		Positive	8	1.52%
Binax Now Flu B	18	*Negative	18	100.00%
Alere BinaxNOW Influenza A+B Card	126	*Negative	124	98.41%
Sekisui Diagnostics OSOM Influenza A & B	22	*Negative	21	95.45%
Quidel QuickVue A+B	90	*Negative	89	98.89%
Remel Xpect	1	*Negative	1	100.00%
Quidel Sofia Influenza A+B	53	*Negative	52	98.11%
Alere Influenza A & B	31	*Negative	30	96.77%
BD Veritor Flu A+B	160	*Negative	159	99.38%
Henry Schein OneStep + Influenza A&B (distinguish)	12	*Negative	11	91.67%
Cepheid GeneXpert - Xpert Flu	2	*Negative	2	100.00%
Alere i Influenza A & B	10	*Negative	10	100.00%
Roche Cobas Liat Influenza A/B	2	*Negative	2	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	1	*Negative	1	100.00%
Alere BinaxNOW Influenza A+B Card	126	Positive	2	1.59%
Sekisui Diagnostics OSOM Influenza A & B	22	Positive	1	4.55%
Quidel QuickVue A+B	90	Positive	1	1.11%
Quidel Sofia Influenza A+B	53	Positive	1	1.89%
Alere Influenza A & B	31	Positive	1	3.23%
BD Veritor Flu A+B	160	Positive	1	0.63%
Henry Schein OneStep + Influenza A&B (distinguish)	12	Positive	1	8.33%
VAD-2		* Denotes Correct Response		
AG: All Methods	529			
		*Negative	519	98.11%
		Positive	10	1.89%
Binax Now Flu B	18	*Negative	18	100.00%
Alere BinaxNOW Influenza A+B Card	126	*Negative	123	97.62%
Sekisui Diagnostics OSOM Influenza A & B	22	*Negative	22	100.00%
Quidel QuickVue A+B	91	*Negative	88	96.70%
Remel Xpect	1	*Negative	1	100.00%
Quidel Sofia Influenza A+B	53	*Negative	53	100.00%
Alere Influenza A & B	31	*Negative	30	96.77%
BD Veritor Flu A+B	160	*Negative	158	98.75%
Henry Schein OneStep + Influenza A&B (distinguish)	12	*Negative	11	91.67%
Cepheid GeneXpert - Xpert Flu	2	*Negative	2	100.00%
Alere i Influenza A & B	10	*Negative	10	100.00%
Roche Cobas Liat Influenza A/B	2	*Negative	2	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	1	*Negative	1	100.00%
Alere BinaxNOW Influenza A+B Card	126	Positive	3	2.38%
Quidel QuickVue A+B	91	Positive	3	3.30%
Alere Influenza A & B	31	Positive	1	3.23%
BD Veritor Flu A+B	160	Positive	2	1.25%
Henry Schein OneStep + Influenza A&B (distinguish)	12	Positive	1	8.33%
VAD-3		* Denotes Correct Response		
AG: All Methods	66			
		*Negative	65	98.48%
		Positive	1	1.52%
Alere BinaxNOW Influenza A+B Card	8	*Negative	8	100.00%
Sekisui Diagnostics OSOM Influenza A & B	13	*Negative	13	100.00%
Quidel QuickVue A+B	22	*Negative	22	100.00%
Remel Xpect	1	*Negative	1	100.00%
Quidel Sofia Influenza A+B	4	*Negative	4	100.00%
Alere Influenza A & B	3	*Negative	3	100.00%
BD Veritor Flu A+B	12	*Negative	11	91.67%
Cepheid GeneXpert - Xpert Flu	2	*Negative	2	100.00%
Alere i Influenza A & B	1	*Negative	1	100.00%
BD Veritor Flu A+B	12	Positive	1	8.33%
VAD-4		* Denotes Correct Response		
AG: All Methods	66			



		*Negative	66	100.00%
Alere BinaxNOW Influenza A+B Card	8	*Negative	8	100.00%
Sekisui Diagnostics OSOM Influenza A & B	13	*Negative	13	100.00%
Quidel QuickVue A+B	22	*Negative	22	100.00%
Remel Xpect	1	*Negative	1	100.00%
Quidel Sofia Influenza A+B	4	*Negative	4	100.00%
Alere Influenza A & B	3	*Negative	3	100.00%
BD Veritor Flu A+B	12	*Negative	12	100.00%
Cepheid GeneXpert - Xpert Flu	2	*Negative	2	100.00%
Alere i Influenza A & B	1	*Negative	1	100.00%
VAD-5		* Denotes Correct Response		
AG: All Methods	66			
		*Positive	65	98.48%
		Negative	1	1.52%
Alere BinaxNOW Influenza A+B Card	8	*Positive	8	100.00%
Sekisui Diagnostics OSOM Influenza A & B	13	*Positive	13	100.00%
Quidel QuickVue A+B	22	*Positive	22	100.00%
Remel Xpect	1	*Positive	1	100.00%
Quidel Sofia Influenza A+B	4	*Positive	4	100.00%
Alere Influenza A & B	3	*Positive	3	100.00%
BD Veritor Flu A+B	12	*Positive	11	91.67%
Cepheid GeneXpert - Xpert Flu	2	*Positive	2	100.00%
Alere i Influenza A & B	1	*Positive	1	100.00%
BD Veritor Flu A+B	12	Negative	1	8.33%
VAD-1		RSV - Ag Detect		
AG: All Methods	260	* Denotes Correct Response		
		*Positive	256	98.46%
		Negative	4	1.54%
BD Directigen RSV	1	*Positive	1	100.00%
BD EZ RSV	1	*Positive	1	100.00%
Alere BinaxNOW RSV	168	*Positive	165	98.21%
Quidel QuickVue RSV	24	*Positive	24	100.00%
Wampole Clearview RSV	5	*Positive	5	100.00%
SAS RSV Alert	1	*Positive	1	100.00%
Quidel QuickVue RSV 10	2	*Positive	2	100.00%
Quidel Sofia RSV	12	*Positive	12	100.00%
BD Veritor RSV	45	*Positive	44	97.78%
Cepheid GeneXpert - Xpert RSV	1	*Positive	1	100.00%
Alere BinaxNOW RSV	168	Negative	3	1.79%
BD Veritor RSV	45	Negative	1	2.22%
VAD-2		* Denotes Correct Response		
AG: All Methods	260			
		*Negative	256	98.46%
		Positive	4	1.54%
BD Directigen RSV	1	*Negative	1	100.00%
BD EZ RSV	1	*Negative	1	100.00%
Alere BinaxNOW RSV	168	*Negative	165	98.21%
Quidel QuickVue RSV	24	*Negative	24	100.00%
Wampole Clearview RSV	5	*Negative	5	100.00%
SAS RSV Alert	1	*Negative	1	100.00%
Quidel QuickVue RSV 10	2	*Negative	2	100.00%
Quidel Sofia RSV	12	*Negative	12	100.00%
BD Veritor RSV	45	*Negative	44	97.78%
Cepheid GeneXpert - Xpert RSV	1	*Negative	1	100.00%
Alere BinaxNOW RSV	168	Positive	3	1.79%
BD Veritor RSV	45	Positive	1	2.22%
VAD-3		* Denotes Correct Response		
AG: All Methods	27			
		*Negative	27	100.00%
BD EZ RSV	1	*Negative	1	100.00%
Alere BinaxNOW RSV	10	*Negative	10	100.00%
Quidel QuickVue RSV	3	*Negative	3	100.00%
Wampole Clearview RSV	4	*Negative	4	100.00%
SAS RSV Alert	1	*Negative	1	100.00%
Quidel QuickVue RSV 10	2	*Negative	2	100.00%
Quidel Sofia RSV	5	*Negative	5	100.00%
Cepheid GeneXpert - Xpert RSV	1	*Negative	1	100.00%
VAD-4		* Denotes Correct Response		



AG: All Methods	27			
		*Negative	27	100.00%
BD EZ RSV	1	*Negative	1	100.00%
Alere BinaxNOW RSV	10	*Negative	10	100.00%
Quidel QuickVue RSV	3	*Negative	3	100.00%
Wampole Clearview RSV	4	*Negative	4	100.00%
SAS RSV Alert	1	*Negative	1	100.00%
Quidel QuickVue RSV 10	2	*Negative	2	100.00%
Quidel Sofia RSV	5	*Negative	5	100.00%
Cepheid GeneXpert - Xpert RSV	1	*Negative	1	100.00%
VAD-5		* Denotes Correct Response		
AG: All Methods	27			
		*Negative	26	96.30%
		Positive	1	3.70%
BD EZ RSV	1	*Negative	1	100.00%
Alere BinaxNOW RSV	10	*Negative	10	100.00%
Quidel QuickVue RSV	3	*Negative	3	100.00%
Wampole Clearview RSV	4	*Negative	3	75.00%
SAS RSV Alert	1	*Negative	1	100.00%
Quidel QuickVue RSV 10	2	*Negative	2	100.00%
Quidel Sofia RSV	5	*Negative	5	100.00%
Cepheid GeneXpert - Xpert RSV	1	*Negative	1	100.00%
Wampole Clearview RSV	4	Positive	1	25.00%
		Adenovirus		
VAD-1		* Denotes Correct Response		
AG: All Methods	51			
		*Negative	51	100.00%
SAS Adeno Test	10	*Negative	10	100.00%
RPS AdenoPlus	41	*Negative	41	100.00%
VAD-2		* Denotes Correct Response		
AG: All Methods	51			
		*Negative	51	100.00%
SAS Adeno Test	10	*Negative	10	100.00%
RPS AdenoPlus	41	*Negative	41	100.00%
VAD-3		* Denotes Correct Response		
AG: All Methods	14			
		*Positive	14	100.00%
SAS Adeno Test	10	*Positive	10	100.00%
RPS AdenoPlus	4	*Positive	4	100.00%
VAD-4		* Denotes Correct Response		
AG: All Methods	14			
		*Positive	14	100.00%
SAS Adeno Test	10	*Positive	10	100.00%
RPS AdenoPlus	4	*Positive	4	100.00%
VAD-5		* Denotes Correct Response		
AG: All Methods	14			
		*Negative	14	100.00%
SAS Adeno Test	10	*Negative	10	100.00%
RPS AdenoPlus	4	*Negative	4	100.00%
932 - Hepatitis C Antibody (Waived);767 - Hepatitis Markers;750 - Viral Markers	Labs	Response	Number	Percent
		HCV (Waived)		
VM-1		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
OraSure QraQuick HCV Rapid Antibody Test	3	*Negative	3	100.00%
VM-2		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
OraSure QraQuick HCV Rapid Antibody Test	3	*Negative	3	100.00%
		HBsAg		
VM-1		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Siemens Centaur	1	*Positive	1	100.00%
VM-2		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%



Siemens Centaur	1	*Negative	1	100.00%	
767 - Hepatitis Markers;750 - Viral Markers		Response		Number	Percent
		HBsAg			
VM-1		* Denotes Correct Response			
AG: All Methods	1				
Siemens Centaur	1	*Positive	1	100.00%	
VM-2		* Denotes Correct Response			
AG: All Methods	1				
Siemens Centaur	1	*Negative	1	100.00%	
VM-3		* Denotes Correct Response			
AG: All Methods	1				
Siemens Centaur	1	*Positive	1	100.00%	
VM-4		* Denotes Correct Response			
AG: All Methods	1				
Siemens Centaur	1	*Positive	1	100.00%	
VM-5		* Denotes Correct Response			
AG: All Methods	1				
Siemens Centaur	1	*Positive	1	100.00%	
		HCV			
VM-1		* Denotes Correct Response			
AG: All Methods	1				
Siemens Centaur	1	*Negative	1	100.00%	
VM-2		* Denotes Correct Response			
AG: All Methods	1				
Siemens Centaur	1	*Negative	1	100.00%	
VM-3		* Denotes Correct Response			
AG: All Methods	1				
Siemens Centaur	1	*Positive	1	100.00%	
VM-4		* Denotes Correct Response			
AG: All Methods	1				
Siemens Centaur	1	*Negative	1	100.00%	
VM-5		* Denotes Correct Response			
AG: All Methods	1				
Siemens Centaur	1	*Negative	1	100.00%	
		Anti-HBc (International)			
VM-1		* Denotes Correct Response			
AG: All Methods	3				
Intl OCD Vitros ECI/ECIQ	3	*Positive	3	100.00%	
VM-2		* Denotes Correct Response			
AG: All Methods	3				
Intl OCD Vitros ECI/ECIQ	3	*Positive	3	100.00%	
VM-3		* Denotes Correct Response			
AG: All Methods	3				
Intl OCD Vitros ECI/ECIQ	3	*Positive	3	100.00%	
VM-4		* Denotes Correct Response			
AG: All Methods	2				
Intl OCD Vitros ECI/ECIQ	2	*Positive	2	100.00%	
VM-5		* Denotes Correct Response			
AG: All Methods	3				
Intl OCD Vitros ECI/ECIQ	3	*Positive	3	100.00%	
		HBsAb (International)			
VM-1		* Denotes Correct Response			



AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
VM-2		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Positive	3	100.00%
VM-3		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Positive	3	100.00%
VM-4		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Intl OCD Vitros ECI/ECIQ	2	*Positive	2	100.00%
VM-5		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
		HBsAg (International)		
VM-1		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Positive	3	100.00%
VM-2		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Intl OCD Vitros ECI/ECIQ	2	*Negative	2	100.00%
VM-3		* Denotes Correct Response		
AG: All Methods	3			
		Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	Negative	3	100.00%
VM-4		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Intl OCD Vitros ECI/ECIQ	2	*Negative	2	100.00%
VM-5		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Positive	3	100.00%
		HCV (International)		
VM-1		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
VM-2		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
VM-3		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Positive	3	100.00%
VM-4		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Intl OCD Vitros ECI/ECIQ	2	*Negative	2	100.00%
VM-5		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
771 - Affirm VP	Labs	Response	Number	Percent
		Candida sp. - Non Culture		
VP-1		* Denotes Correct Response		
AG: All Methods	39			
		*Negative	39	100.00%
Affirm VP	39	*Negative	39	100.00%



VP-2		* Denotes Correct Response		
AG: All Methods	39			
		*Negative	39	100.00%
Affirm VP	39	*Negative	39	100.00%
VP-3		* Denotes Correct Response		
AG: All Methods	39			
		*Negative	39	100.00%
Affirm VP	39	*Negative	39	100.00%
VP-4		* Denotes Correct Response		
AG: All Methods	39			
		*Positive	37	94.87%
		Negative	2	5.13%
Affirm VP	39	*Positive	37	94.87%
Affirm VP		Negative	2	5.13%
VP-5		* Denotes Correct Response		
AG: All Methods	39			
		*Negative	39	100.00%
Affirm VP	39	*Negative	39	100.00%
G. vaginalis - Non Culture Tech				
VP-1		* Denotes Correct Response		
AG: All Methods	39			
		*Positive	39	100.00%
Affirm VP	39	*Positive	39	100.00%
VP-2		* Denotes Correct Response		
AG: All Methods	39			
		*Negative	39	100.00%
Affirm VP	39	*Negative	39	100.00%
VP-3		* Denotes Correct Response		
AG: All Methods	39			
		*Negative	38	97.44%
		Positive	1	2.56%
Affirm VP	39	*Negative	38	97.44%
Affirm VP		Positive	1	2.56%
VP-4		* Denotes Correct Response		
AG: All Methods	39			
		*Positive	39	100.00%
Affirm VP	39	*Positive	39	100.00%
VP-5		* Denotes Correct Response		
AG: All Methods	39			
		*Negative	39	100.00%
Affirm VP	39	*Negative	39	100.00%
T. vaginalis - Non Culture				
VP-1		* Denotes Correct Response		
AG: All Methods	39			
		*Negative	39	100.00%
Affirm VP	39	*Negative	39	100.00%
VP-2		* Denotes Correct Response		
AG: All Methods	39			
		*Positive	39	100.00%
Affirm VP	39	*Positive	39	100.00%
VP-3		* Denotes Correct Response		
AG: All Methods	39			
		*Negative	39	100.00%
Affirm VP	39	*Negative	39	100.00%
VP-4		* Denotes Correct Response		
AG: All Methods	39			
		*Negative	39	100.00%
Affirm VP	39	*Negative	39	100.00%
VP-5		* Denotes Correct Response		
AG: All Methods	39			
		*Negative	39	100.00%
Affirm VP	39	*Negative	39	100.00%