



| 666 - Afinion Glycohemoglobin;998 - CVS Afinion Glycohemoglobin | AFF-2 | | | | | | | | | | | | | | | | | | | | | | | | |
|---|-------|--------|------|---------------|-------|-------|--------|------|---------------|-------|-------|--------|------|---------------|-------|-------|--------|------|---------------|-------|--------|--------|------|---------------|-------|
| Glycohemoglobin | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % |
| PG: Afinion Glycohemoglobin | 175 | 6.44 | 0.77 | 4.10-8.70 | 12.03 | | | | | | | | | | | | | | | | | | | | |
| 683 - ASO/CRP;759 - Complete Immunology | AS-6N | | | | | AS-7N | | | | | AS-8N | | | | | AS-9N | | | | | AS-10N | | | | |
| ASO (Semi-Quant) | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % |
| PG: Beckman (Olympus) AU | 5 | 100.00 | 0.00 | 100.00-100.00 | 0.00 | 5 | 418.00 | 0.00 | 418.00-418.00 | 0.00 | 5 | 100.00 | 0.00 | 100.00-100.00 | 0.00 | 5 | 100.00 | 0.00 | 100.00-100.00 | 0.00 | 5 | 534.00 | 0.00 | 534.00-534.00 | 0.00 |
| 747 - Auto Diff 7 (AcT 5 Diff & Horiba 5-Part Diff) | BCX-6 | | | | | BCX-7 | | | | | BCX-8 | | | | | BCX-9 | | | | | BCX-10 | | | | |
| White Blood Cell Count | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % |
| PG: Beckman Coulter ACT 5 diff (v 1.9) | 15 | 2.43 | 0.17 | 2.00-2.80 | 7.08 | 15 | 16.48 | 0.40 | 14.00-19.00 | 2.42 | 15 | 24.97 | 0.66 | 21.30-28.80 | 2.64 | 14 | 2.44 | 0.09 | 2.00-2.80 | 3.75 | 14 | 23.64 | 0.45 | 20.10-27.10 | 1.91 |
| PG: Horiba ABX Diag Pentra 60 | 16 | 2.47 | 0.10 | 2.10-2.90 | 3.60 | 16 | 16.41 | 0.56 | 13.90-18.90 | 3.41 | 16 | 24.87 | 1.09 | 21.20-28.60 | 4.38 | 16 | 2.44 | 0.10 | 2.00-2.80 | 4.17 | 16 | 23.44 | 0.90 | 19.90-26.90 | 3.85 |
| PG: Horiba ABX Diag Pentra 60C+ (v. 2.21 & up) | 6 | 2.42 | 0.12 | 2.00-2.80 | 5.00 | 6 | 16.58 | 0.40 | 14.10-19.10 | 2.41 | 6 | 25.02 | 0.65 | 21.30-28.80 | 2.60 | 6 | 2.40 | 0.09 | 2.00-2.80 | 3.75 | 6 | 23.45 | 0.55 | 20.00-27.00 | 2.34 |
| SG: Select Beckman Coulter AcT Instruments (5-part) | 19 | 2.44 | 0.15 | 2.00-2.80 | 6.25 | 19 | 16.45 | 0.39 | 13.90-18.90 | 2.38 | 19 | 24.96 | 0.59 | 21.30-28.80 | 2.36 | 18 | 2.45 | 0.09 | 2.10-2.90 | 3.60 | 18 | 23.62 | 0.39 | 20.10-27.10 | 1.65 |
| SG: Select Horiba ABX Instruments (5-part Diff) | 23 | 2.46 | 0.10 | 2.10-2.90 | 4.00 | 23 | 16.44 | 0.51 | 13.90-18.90 | 3.11 | 23 | 24.87 | 0.97 | 21.20-28.60 | 3.90 | 23 | 2.43 | 0.10 | 2.00-2.80 | 4.17 | 23 | 23.43 | 0.79 | 19.90-26.90 | 3.38 |
| SG: International Horiba Instruments | 26 | 2.46 | 0.09 | 2.10-2.90 | 3.60 | 26 | 16.47 | 0.53 | 14.00-19.00 | 3.21 | 26 | 24.93 | 1.00 | 21.20-28.60 | 4.02 | 26 | 2.44 | 0.10 | 2.00-2.80 | 4.17 | 26 | 23.49 | 0.85 | 20.00-27.00 | 3.62 |
| Lymphocyte | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % |
| PG: Beckman Coulter ACT 5 diff (v 1.9) | 14 | 29.69 | 2.67 | 21.70-37.70 | 8.99 | 15 | 47.57 | 3.06 | 38.40-56.80 | 6.43 | 15 | 50.85 | 2.85 | 42.30-59.50 | 5.60 | 14 | 20.71 | 3.02 | 11.60-29.80 | 14.59 | 14 | 44.36 | 1.77 | 39.10-49.70 | 3.99 |
| PG: Horiba ABX Diag Pentra 60 | 15 | 33.99 | 3.06 | 24.80-43.20 | 9.00 | 16 | 51.44 | 3.98 | 39.50-63.30 | 7.74 | 16 | 54.20 | 2.67 | 46.20-62.20 | 4.93 | 16 | 23.87 | 2.89 | 15.20-32.60 | 12.09 | 16 | 48.54 | 2.97 | 39.60-57.40 | 6.12 |
| PG: Horiba ABX Diag Pentra 60C+ (v. 2.21 & up) | 6 | 33.72 | 1.92 | 27.90-39.50 | 5.70 | 6 | 50.53 | 4.11 | 38.20-62.80 | 8.14 | 6 | 52.98 | 4.26 | 40.20-65.80 | 8.04 | 6 | 24.32 | 4.30 | 11.40-37.20 | 17.70 | 6 | 46.82 | 3.96 | 34.90-58.70 | 8.46 |
| SG: Select Beckman Coulter AcT Instruments (5-part) | 18 | 29.13 | 2.76 | 20.80-37.40 | 9.48 | 19 | 47.61 | 2.82 | 39.10-56.10 | 5.92 | 19 | 50.64 | 2.58 | 42.90-58.30 | 5.10 | 18 | 20.34 | 2.84 | 11.80-28.80 | 13.99 | 18 | 44.38 | 1.78 | 39.10-49.70 | 4.01 |
| SG: Select Horiba ABX Instruments (5-part Diff) | 22 | 33.68 | 2.89 | 25.00-42.40 | 8.58 | 23 | 51.48 | 4.09 | 39.20-63.80 | 7.94 | 23 | 54.03 | 3.14 | 44.60-63.40 | 5.81 | 23 | 24.06 | 3.17 | 14.60-33.60 | 13.15 | 23 | 48.47 | 3.72 | 37.30-59.70 | 7.67 |
| SG: International Horiba Instruments | 25 | 34.41 | 3.55 | 23.80-45.00 | 10.32 | 26 | 52.85 | 5.46 | 36.50-69.30 | 10.32 | 26 | 55.07 | 4.19 | 42.50-67.70 | 7.60 | 26 | 24.75 | 3.59 | 13.90-35.50 | 14.53 | 26 | 49.73 | 4.99 | 34.70-64.70 | 10.04 |
| Monocyte | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % |
| PG: Beckman Coulter ACT 5 diff (v 1.9) | 13 | 1.15 | 0.60 | 0.00-2.96 | 52.52 | 15 | 2.70 | 0.51 | 1.18-4.22 | 18.74 | 14 | 2.81 | 0.39 | 1.64-3.98 | 13.91 | 15 | 1.06 | 0.65 | 0.00-3.00 | 61.04 | 15 | 2.28 | 0.83 | 0.00-4.76 | 36.27 |
| PG: Horiba ABX Diag Pentra 60 | 15 | 1.32 | 0.48 | 0.00-2.76 | 36.44 | 15 | 2.39 | 0.35 | 1.33-3.45 | 14.81 | 15 | 2.49 | 0.46 | 1.12-3.86 | 18.39 | 16 | 0.81 | 0.48 | 0.00-2.25 | 59.26 | 15 | 2.32 | 0.68 | 0.29-4.35 | 29.14 |
| PG: Horiba ABX Diag Pentra 60C+ (v. 2.21 & up) | 6 | 1.25 | 0.68 | 0.00-3.28 | 54.00 | 6 | 2.10 | 0.78 | 0.00-4.43 | 36.90 | 6 | 2.32 | 0.34 | 1.29-3.35 | 14.78 | 5 | 0.60 | 0.22 | 0.00-1.27 | 37.33 | 6 | 2.37 | 0.70 | 0.26-4.48 | 29.66 |
| SG: Select Beckman Coulter AcT Instruments (5-part) | 17 | 1.12 | 0.60 | 0.00-2.92 | 53.66 | 19 | 2.72 | 0.54 | 1.09-4.35 | 19.93 | 18 | 2.78 | 0.36 | 1.69-3.87 | 13.02 | 19 | 0.97 | 0.61 | 0.00-2.80 | 62.99 | 19 | 2.32 | 0.77 | 0.02-4.62 | 33.02 |
| SG: Select Horiba ABX Instruments (5-part Diff) | 22 | 1.30 | 0.51 | 0.00-2.84 | 39.54 | 22 | 2.38 | 0.60 | 0.58-4.18 | 25.17 | 22 | 2.50 | 0.51 | 0.98-4.02 | 20.32 | 22 | 0.76 | 0.43 | 0.00-2.04 | 56.18 | 22 | 2.34 | 0.65 | 0.39-4.29 | 27.82 |
| SG: International Horiba Instruments | 25 | 1.26 | 0.50 | 0.00-2.75 | 39.29 | 25 | 2.28 | 0.63 | 0.40-4.16 | 27.50 | 25 | 2.47 | 0.49 | 1.01-3.93 | 19.72 | 25 | 0.86 | 0.48 | 0.00-2.31 | 56.05 | 25 | 2.29 | 0.66 | 0.32-4.26 | 28.73 |
| Granulocyte | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % |
| PG: Beckman Coulter ACT 5 diff (v 1.9) | 14 | 46.17 | 4.41 | 33.00-59.40 | 9.55 | 14 | 38.41 | 4.43 | 25.10-51.70 | 11.54 | 12 | 35.42 | 2.29 | 28.50-42.30 | 6.47 | 15 | 54.29 | 4.43 | 41.00-67.60 | 8.16 | 14 | 44.22 | 3.87 | 32.60-55.80 | 8.76 |
| PG: Horiba ABX Diag Pentra 60 | 16 | 51.36 | 4.86 | 36.80-66.00 | 9.46 | 16 | 41.51 | 3.36 | 31.40-51.60 | 8.10 | 16 | 39.70 | 2.67 | 31.70-47.70 | 6.73 | 16 | 62.12 | 3.46 | 51.70-72.50 | 5.57 | 16 | 45.39 | 2.64 | 37.50-53.30 | 5.81 |
| PG: Horiba ABX Diag Pentra 60C+ (v. 2.21 & up) | 6 | 51.58 | 5.27 | 35.80-67.40 | 10.21 | 6 | 41.95 | 1.84 | 36.50-47.50 | 4.36 | 6 | 39.60 | 3.44 | 29.30-49.90 | 8.69 | 6 | 60.65 | 5.15 | 45.30-76.20 | 8.48 | 6 | 45.97 | 2.89 | 37.30-54.70 | 6.28 |
| SG: Select Beckman Coulter AcT Instruments (5-part) | 17 | 46.56 | 4.09 | 34.30-58.90 | 8.80 | 17 | 38.53 | 4.07 | 26.30-50.70 | 10.57 | 15 | 35.81 | 2.21 | 29.20-42.40 | 6.17 | 18 | 54.80 | 4.32 | 41.80-67.80 | 7.88 | 16 | 44.20 | 3.62 | 33.30-55.10 | 8.19 |
| SG: Select Horiba ABX Instruments (5-part Diff) | 23 | 51.35 | 4.75 | 37.10-65.70 | 9.24 | 23 | 41.07 | 3.94 | 29.30-52.90 | 9.59 | 23 | 39.20 | 3.55 | 28.60-49.90 | 9.06 | 23 | 61.31 | 4.29 | 48.40-74.20 | 7.00 | 23 | 44.93 | 3.92 | 33.10-56.70 | 8.73 |
| SG: International Horiba Instruments | 26 | 51.27 | 4.58 | 37.60-65.00 | 8.91 | 26 | 40.03 | 4.76 | 25.70-54.30 | 11.90 | 26 | 38.35 | 4.14 | 25.90-50.70 | 10.81 | 26 | 61.21 | 4.25 | 48.50-74.00 | 6.94 | 26 | 43.90 | 4.71 | 29.80-58.00 | 10.73 |
| Eosinophil | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % |
| PG: Beckman Coulter ACT 5 diff (v 1.9) | 14 | 10.09 | 3.03 | 1.00-19.20 | 30.00 | 13 | 3.26 | 0.93 | 0.50-6.10 | 28.18 | 12 | 2.95 | 0.69 | 0.90-5.10 | 23.00 | 15 | 9.67 | 3.79 | 0.00-21.10 | 39.18 | 13 | 3.06 | 0.36 | 2.00-4.20 | 11.61 |
| PG: Horiba ABX Diag Pentra 60 | 16 | 11.53 | 3.40 | 1.30-21.70 | 29.57 | 15 | 4.17 | 0.93 | 1.40-7.00 | 22.14 | 15 | 3.14 | 0.64 | 1.20-5.00 | 20.65 | 15 | 11.86 | 2.54 | 4.30-19.50 | 21.34 | 16 | 3.26 | 0.49 | 1.80-4.80 | 14.85 |
| PG: Horiba ABX Diag Pentra 60C+ (v. 2.21 & up) | 6 | 10.38 | 1.84 | 4.90-15.90 | 17.69 | 6 | 3.95 | 1.15 | 0.60-7.40 | 28.75 | 6 | 3.48 | 0.71 | 1.40-5.60 | 20.29 | 6 | 10.93 | 2.87 | 2.30-19.50 | 26.33 | 6 | 3.38 | 0.39 | 2.20-4.60 | 11.47 |
| SG: Select Beckman Coulter AcT Instruments (5-part) | 18 | 10.02 | 2.74 | 1.80-18.20 | 27.40 | 17 | 3.28 | 0.84 | 0.80-5.80 | 25.45 | 15 | 2.85 | 0.67 | 0.80-4.80 | 23.93 | 19 | 9.52 | 3.53 | 0.00-20.10 | 37.16 | 16 | 3.14 | 0.42 | 1.80-4.40 | 13.55 |
| SG: Select Horiba ABX Instruments (5-part Diff) | 23 | 11.07 | 3.06 | 1.90-20.30 | 27.57 | 22 | 4.03 | 1.02 | 0.90-7.10 | 25.50 | 22 | 3.20 | 0.67 | 1.20-5.20 | 20.94 | 22 | 11.42 | 2.66 | 3.40-19.40 | 23.33 | 23 | 3.23 | 0.55 | 1.60-4.90 | 17.19 |
| SG: International Horiba Instruments | 26 | 10.72 | 3.12 | 1.30-20.10 | 29.16 | 25 | 3.88 | 1.05 | 0.70-7.10 | 26.92 | 25 | 3.13 | 0.69 | 1.00-5.20 | 22.26 | 25 | 10.93 | 2.97 | 2.00-19.80 | 27.25 | 26 | 3.14 | 0.61 | 1.30-4.90 | 19.68 |
| Basophil | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % |
| PG: Beckman Coulter ACT 5 diff (v 1.9) | 14 | 12.71 | 1.11 | 9.39-16.03 | 8.70 | 15 | 6.79 | 0.35 | 5.75-7.84 | 5.11 | 15 | 7.34 | 0.46 | 5.96-8.72 | 6.26 | 14 | 13.05 | 1.12 | 9.70-16.41 | 8.57 | 14 | 6.98 | 0.19 | 6.41-7.54 | 2.71 |
| PG: Horiba ABX Diag Pentra 60 | 16 | 0.65 | 0.06 | 0.46-0.84 | 9.72 | 16 | 0.35 | 0.05 | 0.20-0.51 | 14.74 | 16 | 0.39 | 0.03 | 0.32-0.47 | 6.35 | 16 | 0.73 | 0.07 | 0.52-0.93 | 9.42 | 16 | 0.37 | 0.05 | 0.23-0.51 | 12.98 |
| PG: Horiba ABX Diag Pentra 60C+ (v. 2.21 & up) | 5 | 0.68 | 0.08 | 0.43-0.93 | 12.31 | 5 | 0.32 | | | | | | | | | | | | | | | | | | |



| Test Name | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % |
|---|---------------|-------------|-----------|----------------|-------------|-------------|-------------|-----------|---------------|-------------|-------------|-------------|-----------|-----------------|-------------|-------------|-------------|-----------|---------------|-------------|-------------|-------------|-----------|-----------------|-------------|
| SG: Select Beckman Coulter ACT Instruments (5-part) | 19 | 16.92 | 0.34 | 15.90-17.90 | 2.01 | 19 | 48.62 | 0.68 | 45.70-51.50 | 1.40 | 19 | 41.85 | 0.57 | 39.30-44.30 | 1.36 | 18 | 18.49 | 0.27 | 17.40-19.60 | 1.46 | 18 | 51.67 | 0.76 | 48.60-54.80 | 1.47 |
| SG: Select Horiba ABX Instruments (5-part Diff) | 23 | 16.83 | 0.34 | 15.80-17.80 | 2.02 | 23 | 48.41 | 0.69 | 45.50-51.30 | 1.43 | 23 | 41.79 | 0.57 | 39.30-44.30 | 1.36 | 23 | 18.43 | 0.25 | 17.30-19.50 | 1.36 | 23 | 51.47 | 0.88 | 48.40-54.60 | 1.71 |
| SG: International Horiba Instruments | 26 | 16.90 | 0.40 | 15.90-17.90 | 2.37 | 26 | 48.34 | 0.71 | 45.40-51.20 | 1.47 | 26 | 41.80 | 0.60 | 39.30-44.30 | 1.44 | 26 | 18.50 | 0.33 | 17.40-19.60 | 1.78 | 26 | 51.45 | 0.87 | 48.30-54.50 | 1.69 |
| MCV | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % |
| PG: Beckman Coulter ACT 5 diff (v 1.9) | 15 | 70.60 | 1.50 | 66.10-75.10 | 2.12 | 15 | 87.73 | 0.96 | 84.80-90.60 | 1.09 | 15 | 87.33 | 1.11 | 84.00-90.60 | 1.27 | 15 | 92.53 | 2.00 | 86.50-98.50 | 2.16 | 15 | 88.07 | 1.49 | 83.60-92.60 | 1.69 |
| PG: Horiba ABX Diag Pentra 60 | 16 | 70.88 | 1.09 | 67.60-74.20 | 1.54 | 16 | 88.06 | 1.12 | 84.70-91.50 | 1.27 | 16 | 87.63 | 1.15 | 84.10-91.00 | 1.31 | 16 | 93.00 | 1.16 | 89.50-96.50 | 1.24 | 16 | 88.06 | 1.29 | 84.20-92.00 | 1.46 |
| PG: Horiba ABX Diag Pentra 60C+ (v. 2.21 & up) | 6 | 69.33 | 0.52 | 67.70-70.90 | 0.75 | 6 | 86.50 | 0.55 | 84.80-88.20 | 0.64 | 6 | 86.17 | 0.75 | 84.00-88.50 | 0.87 | 6 | 91.33 | 0.82 | 88.80-93.80 | 0.90 | 6 | 86.67 | 0.52 | 85.10-88.30 | 0.60 |
| SG: Select Beckman Coulter ACT Instruments (5-part) | 19 | 70.51 | 1.35 | 66.50-74.50 | 1.90 | 19 | 87.68 | 0.89 | 85.00-90.40 | 1.01 | 19 | 87.25 | 1.04 | 84.10-90.30 | 1.19 | 19 | 92.47 | 1.77 | 87.20-97.80 | 1.92 | 19 | 87.98 | 1.36 | 83.90-92.10 | 1.55 |
| SG: Select Horiba ABX Instruments (5-part Diff) | 23 | 70.61 | 1.37 | 66.50-74.70 | 1.94 | 23 | 87.61 | 1.20 | 84.00-91.20 | 1.37 | 23 | 87.26 | 1.21 | 83.70-90.90 | 1.39 | 23 | 92.74 | 1.57 | 88.00-97.40 | 1.69 | 23 | 87.65 | 1.26 | 83.90-91.50 | 1.45 |
| SG: International Horiba Instruments | 26 | 71.08 | 1.88 | 65.50-76.70 | 2.63 | 26 | 87.62 | 1.13 | 84.20-91.00 | 1.29 | 26 | 87.31 | 1.16 | 83.80-90.80 | 1.33 | 26 | 93.31 | 2.19 | 86.70-99.90 | 2.35 | 26 | 87.62 | 1.20 | 84.00-91.20 | 1.37 |
| RDW | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % |
| PG: Beckman Coulter ACT 5 diff (v 1.9) | 15 | 14.17 | 0.37 | 13.10-15.30 | 2.61 | 15 | 11.89 | 0.42 | 10.60-13.20 | 3.53 | 15 | 12.01 | 0.44 | 10.70-13.30 | 3.67 | 15 | 11.23 | 1.48 | 6.80-15.60 | 13.21 | 15 | 11.86 | 0.43 | 10.60-13.20 | 3.61 |
| PG: Horiba ABX Diag Pentra 60 | 16 | 14.09 | 0.43 | 12.80-15.40 | 3.05 | 16 | 11.76 | 0.25 | 11.10-12.60 | 2.12 | 16 | 11.84 | 0.33 | 10.80-12.80 | 2.80 | 16 | 10.69 | 0.28 | 9.90-11.50 | 2.62 | 16 | 11.76 | 0.28 | 11.00-12.60 | 2.37 |
| PG: Horiba ABX Diag Pentra 60C+ (v. 2.21 & up) | 6 | 14.23 | 0.32 | 13.20-15.20 | 2.25 | 6 | 12.08 | 0.31 | 11.20-13.00 | 2.56 | 6 | 12.27 | 0.27 | 11.50-13.10 | 2.20 | 6 | 11.07 | 0.23 | 10.40-11.80 | 2.07 | 6 | 12.12 | 0.36 | 11.00-13.20 | 2.98 |
| SG: Select Beckman Coulter ACT Instruments (5-part) | 19 | 14.17 | 0.37 | 13.10-15.30 | 2.61 | 19 | 11.88 | 0.38 | 10.80-13.00 | 3.19 | 19 | 11.95 | 0.41 | 10.80-13.20 | 3.42 | 19 | 11.11 | 1.33 | 7.10-15.10 | 11.98 | 19 | 11.85 | 0.40 | 10.70-13.10 | 3.28 |
| SG: Select Horiba ABX Instruments (5-part Diff) | 23 | 14.08 | 0.46 | 12.70-15.50 | 3.26 | 23 | 11.83 | 0.31 | 10.90-12.70 | 2.63 | 23 | 11.94 | 0.36 | 10.80-13.00 | 3.03 | 23 | 10.76 | 0.34 | 9.80-11.80 | 3.15 | 23 | 11.84 | 0.33 | 10.80-12.80 | 2.80 |
| SG: International Horiba Instruments | 26 | 14.15 | 0.51 | 12.60-15.60 | 3.62 | 26 | 11.92 | 0.42 | 10.60-13.20 | 3.53 | 26 | 12.02 | 0.41 | 10.80-13.20 | 3.42 | 26 | 10.81 | 0.36 | 9.70-11.90 | 3.33 | 26 | 11.91 | 0.37 | 10.80-13.00 | 3.11 |
| Platelet Count | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % |
| PG: Beckman Coulter ACT 5 diff (v 1.9) | 15 | 73.13 | 4.55 | 55.00-91.00 | 6.16 | 15 | 313.53 | 13.32 | 236.00-393.00 | 4.24 | 15 | 506.00 | 17.33 | 380.00-633.00 | 3.42 | 14 | 74.00 | 4.52 | 56.00-93.00 | 6.08 | 15 | 458.47 | 107.78 | 344.00-573.00 | 23.54 |
| PG: Horiba ABX Diag Pentra 60 | 16 | 75.81 | 3.69 | 57.00-95.00 | 4.87 | 16 | 320.25 | 12.12 | 240.00-400.00 | 3.78 | 16 | 513.94 | 22.40 | 386.00-643.00 | 4.36 | 16 | 71.88 | 5.02 | 54.00-90.00 | 6.94 | 16 | 497.25 | 18.32 | 373.00-621.00 | 3.68 |
| PG: Horiba ABX Diag Pentra 60C+ (v. 2.21 & up) | 6 | 71.17 | 4.67 | 53.00-89.00 | 6.62 | 6 | 314.50 | 16.49 | 236.00-394.00 | 5.24 | 6 | 508.50 | 22.08 | 382.00-636.00 | 4.34 | 6 | 71.33 | 5.71 | 53.00-89.00 | 8.03 | 6 | 491.67 | 26.40 | 369.00-615.00 | 5.37 |
| SG: Select Beckman Coulter ACT Instruments (5-part) | 19 | 72.95 | 4.13 | 55.00-91.00 | 5.62 | 19 | 315.37 | 12.71 | 236.00-394.00 | 4.03 | 19 | 509.11 | 17.04 | 382.00-636.00 | 3.34 | 18 | 74.33 | 4.10 | 56.00-93.00 | 5.54 | 19 | 466.32 | 96.58 | 350.00-583.00 | 20.73 |
| SG: Select Horiba ABX Instruments (5-part Diff) | 23 | 74.48 | 4.32 | 56.00-93.00 | 5.81 | 23 | 319.17 | 13.19 | 239.00-399.00 | 4.14 | 23 | 513.35 | 21.84 | 385.00-641.00 | 4.25 | 23 | 71.74 | 4.96 | 54.00-90.00 | 6.94 | 23 | 496.96 | 20.69 | 373.00-621.00 | 4.16 |
| SG: International Horiba Instruments | 26 | 76.19 | 10.07 | 57.00-95.00 | 13.29 | 26 | 319.31 | 14.91 | 239.00-399.00 | 4.67 | 26 | 517.19 | 27.15 | 388.00-646.00 | 5.24 | 26 | 71.65 | 4.87 | 54.00-90.00 | 6.81 | 26 | 497.96 | 21.86 | 374.00-623.00 | 4.40 |
| 915 - Body Fluid pH | BFPH-2 | | | | | | | | | | | | | | | | | | | | | | | | |
| pH - Body Fluid | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % |
| PG: Nitrazine Paper | 5 | 7.50 | 0.00 | 7.50-7.50 | 0.00 | | | | | | | | | | | | | | | | | | | | |
| 702 - Blood Gases | BG-6 | | | | | | | | | | | | | | | | | | | | | | | | |
| Blood Gas - pCO2 | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % |
| SG: Non - i-STAT Blood Gas Instruments | 5 | 78.06 | 6.79 | 72.00-84.00 | 8.72 | 5 | 21.60 | 0.73 | 17.00-27.00 | 3.18 | 5 | 34.26 | 2.75 | 29.00-39.00 | 8.24 | 5 | 29.08 | 2.84 | 24.00-34.00 | 9.66 | 5 | 59.74 | 4.02 | 55.00-65.00 | 6.67 |
| SG: International Blood Gas Inst | 6 | 78.35 | 5.65 | 72.00-84.00 | 7.18 | 6 | 20.47 | 2.44 | 15.00-25.00 | 12.00 | 6 | 33.00 | 3.91 | 28.00-38.00 | 11.82 | 6 | 28.50 | 2.87 | 24.00-34.00 | 10.00 | 5 | 56.20 | 3.48 | 51.00-61.00 | 6.25 |
| PG: Alere epoc Blood Analysis System | 5 | 68.60 | 0.00 | 63.00-75.00 | 0.00 | 5 | 20.50 | 0.00 | 16.00-26.00 | 0.00 | 5 | 29.70 | 0.00 | 25.00-35.00 | 0.00 | 5 | 24.10 | 0.00 | 19.00-29.00 | 0.00 | 5 | 53.60 | 0.00 | 49.00-59.00 | 0.00 |
| Blood Gas - pH | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % |
| SG: Non - i-STAT Blood Gas Instruments | 5 | 7.12 | 0.03 | 7.08-7.16 | 0.39 | 5 | 7.43 | 0.01 | 7.39-7.47 | 0.12 | 5 | 7.47 | 0.01 | 7.43-7.51 | 0.15 | 5 | 7.45 | 0.01 | 7.41-7.49 | 0.09 | 5 | 7.18 | 0.01 | 7.14-7.22 | 0.19 |
| SG: International Blood Gas Inst | 6 | 7.13 | 0.01 | 7.09-7.17 | 0.20 | 6 | 7.46 | 0.04 | 7.42-7.50 | 0.50 | 6 | 7.48 | 0.03 | 7.44-7.52 | 0.43 | 6 | 7.46 | 0.02 | 7.42-7.50 | 0.27 | 6 | 7.20 | 0.02 | 7.16-7.24 | 0.31 |
| Blood Gas - pO2 | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % |
| SG: Non - i-STAT Blood Gas Instruments | 5 | 79.58 | 7.61 | 57.00-103.00 | 9.50 | 5 | 175.52 | 19.50 | 118.00-235.00 | 11.08 | 5 | 108.04 | 11.36 | 74.00-142.00 | 10.56 | 5 | 144.68 | 13.53 | 105.00-186.00 | 9.31 | 5 | 200.88 | 21.84 | 136.00-266.00 | 10.85 |
| SG: International Blood Gas Inst | 6 | 83.13 | 10.39 | 52.00-114.00 | 12.53 | 6 | 178.70 | 12.38 | 142.00-216.00 | 6.93 | 6 | 112.77 | 3.98 | 101.00-125.00 | 3.54 | 6 | 147.05 | 8.64 | 121.00-173.00 | 5.85 | 6 | 202.74 | 15.27 | 157.00-249.00 | 7.54 |
| 714 - Complete Cardiac Markers;715 - Basic Cardiac Markers;711 - BNP;720 - Cardiac Markers I;740 - D-Dimer;689 - Cardiac Profile | CDM-6N | | | | | | | | | | | | | | | | | | | | | | | | |
| BNP | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % | Labs | Mean | SD | Range | CV % |
| PG: Abbott i-Stat | 41 | 835.59 | 86.33 | 576.60-1094.60 | 10.33 | 39 | 47.62 | 7.20 | 26.00-69.20 | 15.13 | 33 | 1576.36 | 132.91 | 1177.70-1975.10 | 8.43 | 34 | 416.62 | 43.44 | 286.30-546.90 | 10.43 | 33 | 2974.67 | 215.88 | 2327.10-3622.30 | 7.26 |
| PG: AlereTriage | 39 | 734.63 | 111.09 | 401.30-1067.90 | 15.12 | 13 | 57.26 | 7.05 | 36.10-78.40 | 12.30 | 9 | 1570.00 | 280.62 | 728.10-2411.90 | 17.87 | 9 | 434.00 | 49.38 | 285.90-582.10 | 11.38 | | | | | |



| Sodium (NA+) | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
|---|-------------|-------------|-----------|---------------|-------------|-------------|-------------|-----------|---------------|-------------|-------------|-------------|-----------|---------------|-------------|-------------|-------------|-----------|---------------|-------------|-------------|-------------|-----------|---------------|-------------|
| PG: Abaxis Piccolo (Non-Waived) | 12 | 155.25 | 2.34 | 151.00-159.00 | 1.48 | 6 | 116.00 | 3.35 | 112.00-120.00 | 2.84 | 12 | 125.25 | 2.45 | 121.00-129.00 | 2.00 | 12 | 150.33 | 1.92 | 146.00-154.00 | 1.27 | 12 | 165.67 | 2.61 | 162.00-170.00 | 1.57 |
| PG: Abbott Architect | 7 | 152.63 | 1.51 | 149.00-157.00 | 0.98 | 7 | 113.38 | 0.74 | 109.00-117.00 | 0.62 | 7 | 122.38 | 0.74 | 118.00-126.00 | 0.57 | 7 | 146.63 | 1.30 | 143.00-151.00 | 0.88 | 7 | 161.75 | 1.17 | 158.00-166.00 | 0.74 |
| PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent | 32 | 160.24 | 2.08 | 156.00-164.00 | 1.31 | 31 | 111.88 | 1.79 | 108.00-116.00 | 1.61 | 32 | 121.65 | 1.18 | 118.00-126.00 | 0.98 | 32 | 151.72 | 1.93 | 148.00-156.00 | 1.25 | 32 | 172.29 | 1.73 | 168.00-176.00 | 0.99 |
| PG: Clinical Data Envoy 500 | 7 | 151.00 | 2.27 | 147.00-155.00 | 1.52 | 7 | 114.88 | 1.96 | 111.00-119.00 | 1.74 | 7 | 123.50 | 2.00 | 120.00-128.00 | 1.61 | 7 | 145.38 | 1.92 | 141.00-149.00 | 1.31 | 6 | 161.29 | 2.81 | 157.00-165.00 | 1.74 |
| PG: Siemens Dimension/AR/ES/RxL/Xpand | 19 | 150.60 | 2.16 | 147.00-155.00 | 1.46 | 19 | 115.44 | 1.50 | 111.00-119.00 | 1.30 | 19 | 123.87 | 1.74 | 120.00-128.00 | 1.37 | 19 | 146.19 | 1.90 | 142.00-150.00 | 1.30 | 19 | 159.99 | 1.85 | 156.00-164.00 | 1.19 |
| PG: Horiba ABX Pentra 400 | 6 | 160.00 | 2.58 | 156.00-164.00 | 1.63 | 6 | 114.86 | 1.86 | 111.00-119.00 | 1.65 | 6 | 123.86 | 1.46 | 120.00-128.00 | 1.21 | 5 | 155.67 | 3.44 | 152.00-160.00 | 2.18 | 5 | 175.50 | 3.39 | 172.00-180.00 | 1.93 |
| PG: Beckman (Olympus) AU | 17 | 150.53 | 4.57 | 147.00-155.00 | 3.05 | 17 | 112.65 | 2.00 | 109.00-117.00 | 1.77 | 17 | 121.88 | 2.26 | 118.00-126.00 | 1.89 | 18 | 145.56 | 4.49 | 142.00-150.00 | 3.08 | 18 | 160.56 | 5.42 | 157.00-165.00 | 3.35 |
| PG: Roche Cobas Integra / Roche Reagent | 8 | 149.57 | 1.99 | 146.00-154.00 | 1.33 | 8 | 112.86 | 1.46 | 109.00-117.00 | 1.33 | 8 | 121.64 | 1.60 | 118.00-126.00 | 1.31 | 8 | 144.75 | 2.12 | 141.00-149.00 | 1.45 | 8 | 159.12 | 2.10 | 155.00-163.00 | 1.32 |
| PG: Roche Cobas Mira / DCL Reagent (Sekisui) | 5 | 165.00 | 0.00 | 161.00-169.00 | 0.00 | 5 | 109.00 | 0.00 | 105.00-113.00 | 0.00 | 5 | 121.00 | 0.00 | 117.00-125.00 | 0.00 | 5 | 156.00 | 0.00 | 152.00-160.00 | 0.00 | 5 | 180.00 | 0.00 | 176.00-184.00 | 0.00 |
| PG: OCD Vitros 250,350,400,500,750,950 | 11 | 166.00 | 2.49 | 162.00-170.00 | 1.51 | 11 | 114.73 | 2.24 | 111.00-119.00 | 1.91 | 11 | 126.91 | 2.30 | 123.00-131.00 | 1.81 | 11 | 160.00 | 2.19 | 156.00-164.00 | 1.38 | 11 | 181.82 | 2.44 | 178.00-186.00 | 1.32 |
| SG: Select Alfa Wasserman Instruments / Alfa Wasserman Reagent | 33 | 160.24 | 2.05 | 156.00-164.00 | 1.31 | 32 | 111.91 | 1.77 | 108.00-116.00 | 1.61 | 33 | 121.66 | 1.17 | 118.00-126.00 | 0.98 | 33 | 151.73 | 1.90 | 148.00-156.00 | 1.25 | 33 | 172.28 | 1.70 | 168.00-176.00 | 0.99 |
| SG: Select Beckman Coulter Instruments (Chem) | 18 | 150.61 | 4.45 | 147.00-155.00 | 2.91 | 18 | 112.56 | 1.98 | 109.00-117.00 | 1.77 | 18 | 121.89 | 2.19 | 118.00-126.00 | 1.80 | 19 | 145.63 | 4.37 | 142.00-150.00 | 3.01 | 19 | 160.63 | 5.27 | 157.00-165.00 | 3.29 |
| SG: Select Roche Instruments / Roche Reagent | 9 | 149.51 | 1.87 | 146.00-154.00 | 1.27 | 9 | 112.44 | 1.88 | 108.00-116.00 | 1.70 | 9 | 121.57 | 1.51 | 118.00-126.00 | 1.23 | 9 | 144.78 | 1.99 | 141.00-149.00 | 1.38 | 9 | 159.11 | 1.97 | 155.00-163.00 | 1.26 |
| SG: Select OCD Vitros Instruments | 12 | 166.08 | 2.39 | 162.00-170.00 | 1.45 | 12 | 114.67 | 2.15 | 111.00-119.00 | 1.83 | 12 | 126.92 | 2.19 | 123.00-131.00 | 1.73 | 12 | 160.00 | 2.09 | 156.00-164.00 | 1.31 | 12 | 181.83 | 2.33 | 178.00-186.00 | 1.26 |
| SG: Abaxis Piccolo Instruments | 75 | 155.03 | 2.12 | 151.00-159.00 | 1.35 | 7 | 116.71 | 3.59 | 113.00-121.00 | 3.08 | 13 | 125.31 | 2.36 | 121.00-129.00 | 1.92 | 13 | 150.38 | 1.85 | 146.00-154.00 | 1.27 | 13 | 165.77 | 2.52 | 162.00-170.00 | 1.51 |
| SG: Select Abbott Architect Instruments | 9 | 151.90 | 2.47 | 148.00-156.00 | 1.64 | 9 | 113.30 | 0.68 | 109.00-117.00 | 0.62 | 9 | 122.30 | 0.68 | 118.00-126.00 | 0.57 | 9 | 146.50 | 1.18 | 143.00-151.00 | 0.82 | 9 | 161.60 | 1.08 | 158.00-166.00 | 0.68 |
| SG: Select Electrolyte Instruments 1 | 7 | 159.75 | 2.49 | 156.00-164.00 | 1.56 | 7 | 114.75 | 1.75 | 111.00-119.00 | 1.57 | 7 | 123.88 | 1.36 | 120.00-128.00 | 1.13 | 6 | 155.00 | 3.61 | 151.00-159.00 | 2.32 | 6 | 174.57 | 3.95 | 171.00-179.00 | 2.29 |
| SG: International OCD Vitros Instruments | 14 | 166.50 | 2.77 | 163.00-171.00 | 1.68 | 14 | 115.07 | 2.30 | 111.00-119.00 | 2.00 | 14 | 127.14 | 2.38 | 123.00-131.00 | 1.89 | 14 | 160.36 | 2.47 | 156.00-164.00 | 1.56 | 14 | 182.29 | 2.81 | 178.00-186.00 | 1.54 |
| SG: International Blood Gas Inst | 5 | 176.50 | 0.71 | 173.00-181.00 | 0.40 | 5 | 111.50 | 0.71 | 108.00-116.00 | 0.63 | 5 | 124.50 | 0.71 | 121.00-129.00 | 0.56 | 5 | 158.00 | 2.83 | 154.00-162.00 | 1.77 | 5 | 175.50 | 0.71 | 172.00-180.00 | 0.40 |
| SG: International Abaxis Piccolo Instruments | 79 | 155.01 | 2.07 | 151.00-159.00 | 1.35 | 8 | 117.13 | 3.52 | 113.00-121.00 | 2.99 | 17 | 125.59 | 2.24 | 122.00-130.00 | 1.75 | 17 | 150.41 | 2.06 | 146.00-154.00 | 1.40 | 17 | 165.71 | 2.26 | 162.00-170.00 | 1.39 |
| Potassium (K+) | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| PG: Abaxis Piccolo (Non-Waived) | 11 | 6.53 | 0.21 | 6.00-7.00 | 3.23 | 11 | 2.27 | 0.14 | 1.80-2.80 | 6.09 | 12 | 3.34 | 0.23 | 2.80-3.80 | 6.97 | 12 | 6.03 | 0.24 | 5.50-6.50 | 4.00 | 12 | 7.56 | 0.27 | 7.10-8.10 | 3.55 |
| PG: Abbott Architect | 7 | 6.64 | 0.05 | 6.10-7.10 | 0.76 | 7 | 2.63 | 0.05 | 2.10-3.10 | 1.92 | 7 | 3.60 | 0.00 | 3.10-4.10 | 0.00 | 7 | 6.11 | 0.04 | 5.60-6.60 | 0.66 | 7 | 7.65 | 0.08 | 7.20-8.20 | 1.04 |
| PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent | 32 | 6.93 | 0.23 | 6.40-7.40 | 3.33 | 32 | 2.63 | 0.10 | 2.10-3.10 | 3.85 | 32 | 3.57 | 0.09 | 3.10-4.10 | 2.50 | 32 | 6.31 | 0.16 | 5.80-6.80 | 2.54 | 32 | 8.14 | 0.25 | 7.60-8.60 | 3.09 |
| PG: Clinical Data Envoy 500 | 7 | 6.56 | 0.29 | 6.10-7.10 | 4.39 | 7 | 2.68 | 0.12 | 2.20-3.20 | 4.44 | 7 | 3.64 | 0.13 | 3.10-4.10 | 3.61 | 7 | 6.06 | 0.17 | 5.60-6.60 | 2.79 | 7 | 7.45 | 0.30 | 7.00-8.00 | 4.00 |
| PG: Siemens Dimension/AR/ES/RxL/Xpand | 19 | 6.55 | 0.09 | 6.00-7.00 | 1.38 | 19 | 2.57 | 0.04 | 2.10-3.10 | 1.54 | 19 | 3.55 | 0.06 | 3.10-4.10 | 1.67 | 19 | 6.05 | 0.07 | 5.50-6.50 | 1.17 | 19 | 7.58 | 0.07 | 7.10-8.10 | 0.92 |
| PG: Horiba ABX Pentra 400 | 6 | 6.66 | 0.20 | 6.20-7.20 | 2.99 | 6 | 2.70 | 0.13 | 2.20-3.20 | 4.81 | 6 | 3.59 | 0.11 | 3.10-4.10 | 3.06 | 6 | 6.14 | 0.23 | 5.60-6.60 | 3.77 | 6 | 7.81 | 0.25 | 7.30-8.30 | 3.21 |
| PG: Beckman (Olympus) AU | 17 | 6.49 | 0.17 | 6.00-7.00 | 2.62 | 17 | 2.64 | 0.06 | 2.10-3.10 | 2.31 | 17 | 3.56 | 0.08 | 3.10-4.10 | 2.22 | 18 | 6.00 | 0.17 | 5.50-6.50 | 2.83 | 18 | 7.51 | 0.22 | 7.00-8.00 | 2.93 |
| PG: Roche Cobas Integra / Roche Reagent | 8 | 6.55 | 0.09 | 6.10-7.10 | 1.36 | 8 | 2.63 | 0.04 | 2.10-3.10 | 1.92 | 8 | 3.60 | 0.05 | 3.10-4.10 | 1.39 | 8 | 6.00 | 0.18 | 5.50-6.50 | 3.00 | 8 | 7.54 | 0.09 | 7.00-8.00 | 1.20 |
| PG: OCD Vitros 250,350,400,500,750,950 | 11 | 6.94 | 0.13 | 6.40-7.40 | 1.88 | 11 | 2.69 | 0.07 | 2.20-3.20 | 2.59 | 11 | 3.71 | 0.08 | 3.20-4.20 | 2.16 | 11 | 6.39 | 0.09 | 5.90-6.90 | 1.41 | 11 | 8.11 | 0.14 | 7.60-8.60 | 1.73 |
| SG: Select Alfa Wasserman Instruments / Alfa Wasserman Reagent | 33 | 6.92 | 0.23 | 6.40-7.40 | 3.33 | 33 | 2.63 | 0.11 | 2.10-3.10 | 3.85 | 33 | 3.57 | 0.09 | 3.10-4.10 | 2.50 | 33 | 6.30 | 0.16 | 5.80-6.80 | 2.54 | 33 | 8.14 | 0.25 | 7.60-8.60 | 3.09 |
| SG: Select Alfa Wasserman Instruments / Alfa Wasserman Reagent Groups | 33 | 6.92 | 0.23 | 6.40-7.40 | 3.33 | 33 | 2.63 | 0.11 | 2.10-3.10 | 3.85 | 33 | 3.57 | 0.09 | 3.10-4.10 | 2.50 | 33 | 6.30 | 0.16 | 5.80-6.80 | 2.54 | 33 | 8.14 | 0.25 | 7.60-8.60 | 3.09 |
| SG: Select Beckman Coulter Instruments (Chem) | 18 | 6.50 | 0.17 | 6.00-7.00 | 2.62 | 18 | 2.63 | 0.06 | 2.10-3.10 | 2.31 | 18 | 3.56 | 0.08 | 3.10-4.10 | 2.22 | 19 | 6.01 | 0.17 | 5.50-6.50 | 2.67 | 19 | 7.51 | 0.21 | 7.00-8.00 | 2.80 |
| SG: Select Roche Instruments / Roche Reagent | 9 | 6.55 | 0.09 | 6.00-7.00 | 1.38 | 9 | 2.59 | 0.12 | 2.10-3.10 | 4.62 | 9 | 3.59 | 0.06 | 3.10-4.10 | 1.67 | 9 | 6.00 | 0.17 | 5.50-6.50 | 2.83 | 9 | 7.52 | 0.10 | 7.00-8.00 | 1.33 |
| SG: Select Roche Instruments / Roche Reagents Groups | 10 | 6.52 | 0.11 | 6.00-7.00 | 1.69 | 10 | 2.60 | 0.12 | 2.10-3.10 | 4.62 | 10 | 3.59 | 0.06 | 3.10-4.10 | 1.67 | 10 | 5.99 | 0.16 | 5.50-6.50 | 2.67 | 10 | 7.50 | 0.12 | 7.00-8.00 | 1.60 |
| SG: Select OCD Vitros Instruments | 12 | 6.93 | 0.12 | 6.40-7.40 | 1.74 | 12 | 2.68 | 0.07 | 2.20-3.20 | 2.59 | 12 | 3.70 | 0.09 | 3.20-4.20 | 2.43 | 12 | 6.39 | 0.09 | 5.90-6.90 | 1.41 | 12 | 8.11 | 0.14 | 7.60-8.60 | 1.73 |
| SG: Abaxis Piccolo Instruments | 73 | 6.62 | 0.22 | 6.10-7.10 | 3.33 | 12 | 2.29 | 0.15 | 1.80-2.80 | 6.52 | 13 | 3.36 | 0.23 | 2.90-3.90 | 6.76 | 13 | 6.06 | 0.25 | 5.60-6.60 | 4.10 | 13 | 7.58 | 0.28 | 7.10-8.10 | 3.68 |
| SG: Select Abbott Architect Instruments | 9 | 6.61 | 0.09 | 6.10-7.10 | 1.36 | 9 | 2.62 | 0.04 | 2.10-3.10 | 1.54 | 9 | 3.60 | 0.00 | 3.10-4.10 | 0.00 | 9 | 6.11 | 0.03 | 5.60-6.60 | 0.49 | 9 | 7.65 | 0.07 | 7.20-8.20 | 0.91 |
| SG: International OCD Vitros Instruments | 14 | 6.95 | 0.12 | 6.50-7.50 | 1.71 | 14 | 2.69 | 0.06 | 2.20-3.20 | 2.22 | 14 | 3.71 | 0.08 | 3.20-4.20 | 2.16 | 14 | 6.40 | 0.09 | 5.90-6.90 | 1.41 | 14 | 8.14 | 0.14 | 7.60-8.60 | 1.73 |
| SG: International Siemens Instruments | 21 | 6.56 | 0.10 | 6.10-7.10 | 1.52 | 21 | 2.57 | 0.04 | 2. | | | | | | | | | | | | | | | | |



| PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent | 31 | 33.60 | 2.61 | 26.00-42.00 | 7.65 | 31 | 15.96 | 2.04 | 10.00-22.00 | 12.50 | 31 | 19.41 | 2.86 | 10.00-28.00 | 15.26 | 31 | 31.25 | 2.53 | 24.00-39.00 | 8.06 | 31 | 37.64 | 2.77 | 30.00-46.00 | 7.37 |
|--|-------------|-------------|-----------|--------------|------------|-------------|-------------|-----------|--------------|------------|-------------|-------------|-----------|--------------|------------|-------------|-------------|-----------|--------------|------------|-------------|-------------|-----------|--------------|------------|
| PG: Clinical Data Envoy 500 | 6 | 32.57 | 7.93 | 9.00-57.00 | 23.94 | 6 | 17.17 | 3.31 | 7.00-27.00 | 19.41 | 6 | 18.43 | 5.59 | 1.00-35.00 | 31.11 | 6 | 32.14 | 4.22 | 19.00-45.00 | 13.13 | 6 | 37.43 | 4.32 | 24.00-50.00 | 11.62 |
| PG: Siemens Dimension/AR/ES/RxL/Xpand | 19 | 34.44 | 2.51 | 27.00-42.00 | 7.35 | 19 | 17.28 | 1.72 | 12.00-22.00 | 10.00 | 19 | 20.62 | 1.74 | 16.00-26.00 | 8.10 | 19 | 32.30 | 1.29 | 28.00-36.00 | 4.06 | 19 | 37.58 | 2.32 | 31.00-45.00 | 6.05 |
| PG: Horiba ABX Pentra 400 | 5 | 34.18 | 0.81 | 32.00-36.00 | 2.35 | 5 | 17.20 | 1.30 | 13.00-21.00 | 7.65 | 5 | 21.20 | 1.92 | 15.00-27.00 | 9.05 | 5 | 32.40 | 1.52 | 28.00-37.00 | 4.69 | 6 | 35.29 | 4.54 | 22.00-49.00 | 12.86 |
| PG: Beckman (Olympus) AU | 17 | 32.24 | 3.98 | 20.00-44.00 | 12.50 | 16 | 14.44 | 2.10 | 8.00-20.00 | 15.00 | 15 | 18.73 | 2.05 | 13.00-25.00 | 11.05 | 18 | 29.67 | 4.52 | 17.00-44.00 | 15.00 | 18 | 36.83 | 4.57 | 23.00-51.00 | 12.43 |
| PG: Roche Cobas Integra / Roche Reagent | 8 | 31.96 | 2.41 | 25.00-39.00 | 7.50 | 8 | 15.56 | 1.57 | 11.00-21.00 | 10.00 | 8 | 18.94 | 1.18 | 15.00-23.00 | 6.32 | 8 | 29.12 | 1.55 | 25.00-34.00 | 5.17 | 8 | 35.68 | 2.10 | 30.00-42.00 | 5.83 |
| PG: OCD Vitros 250,350,400,500,750,950 | 11 | 33.91 | 1.30 | 30.00-38.00 | 3.82 | 11 | 14.64 | 1.21 | 11.00-19.00 | 8.00 | 11 | 18.73 | 1.19 | 15.00-23.00 | 6.32 | 11 | 30.36 | 1.12 | 27.00-33.00 | 3.67 | 11 | 36.55 | 1.70 | 32.00-42.00 | 4.59 |
| SG: Select Beckman Coulter Instruments (Chem) | 18 | 32.28 | 3.86 | 20.00-44.00 | 12.19 | 17 | 14.53 | 2.06 | 9.00-21.00 | 14.00 | 16 | 18.75 | 1.98 | 13.00-25.00 | 10.53 | 19 | 29.68 | 4.40 | 17.00-43.00 | 14.67 | 19 | 37.00 | 4.50 | 24.00-51.00 | 12.16 |
| SG: Select Roche Instruments / Roche Reagent | 9 | 31.63 | 2.46 | 25.00-40.00 | 7.81 | 9 | 15.16 | 1.88 | 9.00-21.00 | 12.67 | 9 | 18.83 | 1.15 | 16.00-22.00 | 5.79 | 9 | 29.22 | 1.48 | 25.00-34.00 | 5.17 | 9 | 35.72 | 1.97 | 30.00-42.00 | 5.56 |
| SG: Select Roche Instruments / Roche Reagents Groups | 10 | 30.77 | 3.58 | 20.00-42.00 | 11.61 | 10 | 14.65 | 2.41 | 8.00-22.00 | 16.00 | 10 | 18.35 | 1.87 | 12.00-24.00 | 10.56 | 10 | 28.70 | 2.16 | 22.00-36.00 | 7.59 | 10 | 35.34 | 2.20 | 28.00-42.00 | 6.29 |
| SG: Select OCD Vitros Instruments | 12 | 34.00 | 1.28 | 30.00-38.00 | 3.82 | 12 | 14.75 | 1.22 | 11.00-19.00 | 8.00 | 12 | 18.75 | 1.14 | 16.00-22.00 | 5.79 | 12 | 30.50 | 1.17 | 27.00-35.00 | 3.87 | 11 | 36.55 | 1.70 | 32.00-42.00 | 4.59 |
| SG: Abaxis Piccolo Instruments | 74 | 31.97 | 1.80 | 27.00-37.00 | 5.63 | 12 | 14.92 | 0.79 | 13.00-17.00 | 5.33 | 12 | 19.58 | 0.90 | 17.00-23.00 | 4.50 | 12 | 31.00 | 1.65 | 26.00-36.00 | 5.48 | 12 | 34.50 | 2.15 | 28.00-42.00 | 6.29 |
| SG: Select Abbott Architect Instruments | 9 | 33.30 | 2.26 | 26.00-40.00 | 6.97 | 9 | 15.80 | 2.39 | 9.00-23.00 | 15.00 | 9 | 19.20 | 2.20 | 12.00-26.00 | 11.58 | 9 | 31.10 | 2.77 | 23.00-39.00 | 9.03 | 9 | 38.40 | 2.12 | 32.00-44.00 | 5.53 |
| SG: International OCD Vitros Instruments | 13 | 34.00 | 1.23 | 30.00-38.00 | 3.53 | 13 | 14.69 | 1.18 | 11.00-19.00 | 8.00 | 13 | 18.69 | 1.25 | 15.00-23.00 | 6.84 | 13 | 30.15 | 1.21 | 26.00-34.00 | 4.00 | 13 | 36.31 | 1.65 | 31.00-41.00 | 4.72 |
| SG: International Abaxis Piccolo Instruments | 78 | 31.97 | 1.76 | 27.00-37.00 | 5.63 | 16 | 14.56 | 0.96 | 12.00-18.00 | 6.67 | 16 | 19.19 | 1.05 | 16.00-22.00 | 5.26 | 16 | 31.00 | 1.75 | 26.00-36.00 | 5.81 | 16 | 34.63 | 2.00 | 29.00-41.00 | 5.71 |
| AG: All Group | 184 | 32.81 | 2.91 | 24.00-42.00 | 8.79 | 121 | 15.81 | 2.16 | 9.00-23.00 | 13.75 | 120 | 19.44 | 2.46 | 12.00-27.00 | 13.16 | 123 | 31.02 | 2.78 | 23.00-39.00 | 9.03 | 123 | 36.92 | 3.13 | 28.00-46.00 | 8.38 |
| Calcium | Labs | Mean | SD | Range | CV% | Labs | Mean | SD | Range | CV% | Labs | Mean | SD | Range | CV% | Labs | Mean | SD | Range | CV% | Labs | Mean | SD | Range | CV% |
| PG: Abaxis Piccolo (Non-Waived) | 11 | 12.25 | 0.19 | 11.30-13.30 | 1.54 | 11 | 7.31 | 0.11 | 6.30-8.30 | 1.51 | 11 | 8.50 | 0.20 | 7.50-9.50 | 2.35 | 11 | 11.72 | 0.32 | 10.70-12.70 | 2.74 | 11 | 13.44 | 0.44 | 12.40-14.40 | 3.28 |
| PG: Abbott Architect | 7 | 12.28 | 0.24 | 11.30-13.30 | 1.95 | 7 | 7.06 | 0.16 | 6.10-8.10 | 2.25 | 7 | 8.28 | 0.20 | 7.30-9.30 | 2.53 | 7 | 11.64 | 0.25 | 10.60-12.60 | 2.16 | 7 | 13.63 | 0.21 | 12.60-14.60 | 1.54 |
| PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent | 32 | 11.93 | 0.35 | 10.90-12.90 | 2.94 | 32 | 7.34 | 0.23 | 6.30-8.30 | 3.15 | 32 | 8.57 | 0.20 | 7.60-9.60 | 2.33 | 32 | 11.63 | 0.27 | 10.60-12.60 | 2.33 | 32 | 13.31 | 0.30 | 12.30-14.30 | 2.26 |
| PG: Clinical Data Envoy 500 | 7 | 12.16 | 0.57 | 11.20-13.20 | 4.67 | 7 | 6.83 | 0.26 | 5.80-7.80 | 3.68 | 7 | 8.09 | 0.18 | 7.10-9.10 | 2.22 | 7 | 11.41 | 0.69 | 10.40-12.40 | 6.05 | 6 | 13.30 | 0.36 | 12.30-14.30 | 2.71 |
| PG: Siemens Dimension/AR/ES/RxL/Xpand | 18 | 12.21 | 0.27 | 11.20-13.20 | 2.21 | 18 | 7.08 | 0.27 | 6.10-8.10 | 3.80 | 18 | 8.28 | 0.26 | 7.30-9.30 | 3.13 | 18 | 11.49 | 0.22 | 10.50-12.50 | 1.91 | 18 | 13.59 | 0.23 | 12.60-14.60 | 1.69 |
| PG: Horiba ABX Pentra 400 | 6 | 12.30 | 0.35 | 11.30-13.30 | 2.85 | 6 | 7.03 | 0.28 | 6.00-8.00 | 4.00 | 6 | 8.33 | 0.17 | 7.30-9.30 | 2.05 | 6 | 11.69 | 0.35 | 10.70-12.70 | 2.99 | 6 | 13.61 | 0.45 | 12.60-14.60 | 3.31 |
| PG: Beckman (Olympus) AU | 17 | 12.17 | 0.35 | 11.20-13.20 | 2.87 | 17 | 7.06 | 0.33 | 6.10-8.10 | 4.65 | 17 | 8.45 | 0.14 | 7.40-9.40 | 1.67 | 18 | 11.66 | 0.22 | 10.70-12.70 | 1.88 | 18 | 13.54 | 0.28 | 12.50-14.50 | 2.07 |
| PG: Roche Cobas Integra / Roche Reagent | 8 | 12.37 | 0.40 | 11.40-13.40 | 3.23 | 8 | 7.21 | 0.36 | 6.20-8.20 | 5.00 | 8 | 8.51 | 0.37 | 7.50-9.50 | 4.35 | 8 | 11.79 | 0.36 | 10.80-12.80 | 3.05 | 8 | 13.71 | 0.39 | 12.70-14.70 | 2.85 |
| PG: OCD Vitros 250,350,400,500,750,950 | 11 | 12.14 | 0.30 | 11.10-13.10 | 2.48 | 11 | 6.97 | 0.17 | 6.00-8.00 | 2.43 | 11 | 8.40 | 0.17 | 7.40-9.40 | 2.02 | 11 | 11.64 | 0.23 | 10.60-12.60 | 1.98 | 11 | 13.29 | 0.32 | 12.30-14.30 | 2.41 |
| SG: Select Beckman Coulter Instruments (Chem) | 18 | 12.16 | 0.34 | 11.20-13.20 | 2.79 | 18 | 7.08 | 0.33 | 6.10-8.10 | 4.51 | 18 | 8.45 | 0.14 | 7.50-9.50 | 1.65 | 19 | 11.65 | 0.22 | 10.60-12.60 | 1.90 | 19 | 13.52 | 0.30 | 12.50-14.50 | 2.22 |
| SG: Select Roche Instruments / Roche Reagent | 9 | 12.36 | 0.38 | 11.40-13.40 | 3.06 | 9 | 7.16 | 0.38 | 6.20-8.20 | 5.28 | 9 | 8.51 | 0.35 | 7.50-9.50 | 4.12 | 9 | 11.79 | 0.33 | 10.80-12.80 | 2.80 | 9 | 13.71 | 0.36 | 12.70-14.70 | 2.63 |
| SG: Select Roche Instruments / Roche Reagents Groups | 10 | 12.30 | 0.41 | 11.30-13.30 | 3.33 | 10 | 7.11 | 0.38 | 6.10-8.10 | 5.35 | 10 | 8.51 | 0.33 | 7.50-9.50 | 3.88 | 10 | 11.73 | 0.37 | 10.70-12.70 | 3.16 | 10 | 13.60 | 0.49 | 12.60-14.60 | 3.60 |
| SG: Select OCD Vitros Instruments | 12 | 12.09 | 0.33 | 11.10-13.10 | 2.73 | 12 | 6.96 | 0.17 | 6.00-8.00 | 2.43 | 12 | 8.38 | 0.17 | 7.40-9.40 | 2.02 | 12 | 11.60 | 0.25 | 10.60-12.60 | 2.16 | 12 | 13.24 | 0.35 | 12.20-14.20 | 2.65 |
| SG: Abaxis Piccolo Instruments | 74 | 12.19 | 0.26 | 11.20-13.20 | 2.13 | 12 | 7.29 | 0.12 | 6.30-8.30 | 1.64 | 12 | 8.50 | 0.19 | 7.50-9.50 | 2.24 | 12 | 11.71 | 0.31 | 10.70-12.70 | 2.65 | 12 | 13.40 | 0.44 | 12.40-14.40 | 3.28 |
| SG: Select Abbott Architect Instruments | 9 | 12.32 | 0.23 | 11.30-13.30 | 1.87 | 9 | 7.07 | 0.14 | 6.10-8.10 | 1.97 | 9 | 8.30 | 0.19 | 7.30-9.30 | 2.29 | 9 | 11.64 | 0.22 | 10.60-12.60 | 1.90 | 9 | 13.64 | 0.20 | 12.60-14.60 | 1.47 |
| SG: International OCD Vitros Instruments | 14 | 12.15 | 0.28 | 11.20-13.20 | 2.30 | 14 | 6.96 | 0.17 | 6.00-8.00 | 2.43 | 14 | 8.39 | 0.17 | 7.40-9.40 | 2.02 | 14 | 11.66 | 0.21 | 10.70-12.70 | 1.79 | 14 | 13.34 | 0.32 | 12.30-14.30 | 2.41 |
| SG: International Abaxis Piccolo Instruments | 79 | 12.17 | 0.26 | 11.20-13.20 | 2.13 | 17 | 7.22 | 0.16 | 6.20-8.20 | 2.22 | 17 | 8.44 | 0.19 | 7.40-9.40 | 2.26 | 17 | 11.61 | 0.36 | 10.60-12.60 | 3.19 | 17 | 13.27 | 0.49 | 12.30-14.30 | 3.68 |
| AG: All Group | 186 | 12.15 | 0.33 | 11.20-13.20 | 2.70 | 124 | 7.14 | 0.29 | 6.10-8.10 | 4.08 | 124 | 8.41 | 0.25 | 7.40-9.40 | 2.98 | 125 | 11.61 | 0.31 | 10.60-12.60 | 2.67 | 124 | 13.45 | 0.35 | 12.50-14.50 | 2.59 |
| Creatinine | Labs | Mean | SD | Range | CV% | Labs | Mean | SD | Range | CV% | Labs | Mean | SD | Range | CV% | Labs | Mean | SD | Range | CV% | Labs | Mean | SD | Range | CV% |
| PG: Abaxis Piccolo (Non-Waived) | 12 | 4.50 | 0.15 | 3.83-5.17 | 3.42 | 12 | 0.62 | 0.16 | 0.30-0.90 | 26.67 | 12 | 1.52 | 0.10 | 1.22-1.82 | 6.78 | 12 | 3.94 | 0.23 | 3.30-4.50 | 6.15 | 12 | 5.44 | 0.16 | 4.62-6.26 | 2.98 |
| PG: Abbott Architect | 7 | 4.75 | 0.06 | 4.00-5.40 | 1.33 | 7 | 0.53 | 0.02 | 0.23-0.83 | 4.00 | 7 | 1.43 | 0.04 | 1.13-1.73 | 2.86 | 7 | 4.22 | 0.07 | 3.60-4.80 | 1.67 | 7 | 5.85 | 0.08 | 4.97-6.73 | 1.38 |
| PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent | 33 | 4.26 | 0.16 | 3.62-4.90 | 3.73 | 33 | 0.65 | 0.04 | 0.40-1.00 | 6.62 | 33 | 1.60 | 0.07 | 1.30-1.90 | 4.63 | 33 | 3.85 | 0.14 | 3.30-4.50 | 3.85 | 33 | 5.11 | 0.22 | 4.34-5.88 | 4.21 |
| PG: Clinical Data Envoy 500 | 7 | 4.25 | 0.13 | 3.60-4.80 | 3.11 | 7 | 0.56 | 0.05 | 0.26-0.86 | 8.93 | 7 | 1.50 | 0.01 | 1.20-1.80 | 0.73 | 7 | 3.78 | 0.14 | 3.20-4.40 | 3.70 | 6 | 5.05 | 0.11 | 4.30-5.90 | 2.22 |
| PG: Siemens Dimension/AR/ES/RxL/Xpand | 14 | 4.53 | 0.16 | 3.85-5.21 | 3.56 | 13 | 0.48 | 0.05 | 0.20-0.80 | 10.00 | 13 | 1.55 | 0.06 | 1.30-1.90 | 4 | | | | | | | | | | |



| Test Name | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
|--|-------------|-------------|-----------|---------------|-------------|-------------|-------------|-----------|--------------|-------------|-------------|-------------|-----------|--------------|-------------|-------------|-------------|-----------|---------------|-------------|-------------|-------------|-----------|---------------|-------------|
| SG: Select OCD Vitros Instruments | 12 | 219.58 | 3.78 | 198.00-242.00 | 1.73 | 12 | 52.00 | 1.76 | 46.00-58.00 | 3.46 | 12 | 92.25 | 1.60 | 83.00-101.00 | 1.74 | 12 | 196.83 | 2.79 | 177.00-217.00 | 1.42 | 12 | 267.17 | 3.88 | 240.00-294.00 | 1.46 |
| SG: Abaxis Piccolo Instruments | 75 | 219.84 | 2.30 | 198.00-242.00 | 1.05 | 13 | 55.23 | 0.44 | 49.00-61.00 | 0.73 | 13 | 96.08 | 0.86 | 86.00-106.00 | 0.94 | 13 | 199.31 | 1.84 | 179.00-219.00 | 0.90 | 13 | 262.08 | 1.80 | 236.00-288.00 | 0.69 |
| SG: Select Abbott Architect Instruments | 9 | 223.70 | 4.19 | 202.00-246.00 | 1.88 | 9 | 52.30 | 0.82 | 46.00-58.00 | 1.54 | 9 | 95.50 | 1.78 | 86.00-106.00 | 1.88 | 9 | 201.10 | 3.38 | 181.00-221.00 | 1.69 | 9 | 268.70 | 4.79 | 242.00-296.00 | 1.78 |
| SG: International OCD Vitros Instruments | 14 | 219.93 | 2.90 | 198.00-242.00 | 1.32 | 14 | 51.79 | 1.72 | 46.00-58.00 | 3.27 | 14 | 92.29 | 1.64 | 83.00-101.00 | 1.74 | 14 | 197.21 | 2.94 | 177.00-217.00 | 1.47 | 14 | 268.29 | 2.37 | 241.00-295.00 | 0.90 |
| SG: International Siemens Instruments | 21 | 229.06 | 12.91 | 206.00-252.00 | 5.63 | 21 | 56.35 | 2.39 | 50.00-62.00 | 4.29 | 21 | 99.63 | 4.25 | 90.00-110.00 | 4.30 | 21 | 206.79 | 10.53 | 186.00-228.00 | 5.07 | 21 | 273.24 | 17.26 | 246.00-300.00 | 6.34 |
| SG: International Abaxis Piccolo Instruments | 80 | 219.73 | 2.29 | 198.00-242.00 | 1.05 | 18 | 55.47 | 0.86 | 49.00-61.00 | 1.64 | 18 | 96.03 | 0.79 | 86.00-106.00 | 0.83 | 18 | 198.85 | 2.07 | 179.00-219.00 | 1.06 | 18 | 261.69 | 1.90 | 236.00-288.00 | 0.73 |
| AG: All Group | 189 | 226.17 | 8.51 | 203.00-249.00 | 3.76 | 128 | 56.18 | 3.05 | 50.00-62.00 | 5.54 | 127 | 99.43 | 4.69 | 89.00-109.00 | 4.75 | 127 | 207.59 | 9.20 | 187.00-229.00 | 4.42 | 127 | 273.69 | 9.72 | 247.00-301.00 | 3.54 |
| Urea Nitrogen (BUN) | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| PG: Abaxis Piccolo (Non-Waived) | 12 | 36.00 | 0.85 | 33.00-39.00 | 2.50 | 12 | 7.58 | 0.52 | 6.00-10.00 | 6.25 | 12 | 13.83 | 0.58 | 12.00-16.00 | 4.29 | 12 | 31.67 | 0.89 | 29.00-35.00 | 2.81 | 12 | 44.50 | 1.09 | 41.00-49.00 | 2.44 |
| PG: Abbott Architect | 7 | 40.00 | 0.54 | 36.00-44.00 | 1.25 | 7 | 6.38 | 0.52 | 4.00-8.00 | 8.33 | 7 | 14.88 | 0.35 | 13.00-17.00 | 2.67 | 7 | 35.75 | 0.71 | 33.00-39.00 | 1.94 | 7 | 48.88 | 0.99 | 45.00-53.00 | 2.04 |
| PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent | 32 | 37.67 | 1.37 | 35.00-41.00 | 3.68 | 31 | 6.39 | 0.55 | 4.00-8.00 | 8.33 | 32 | 14.25 | 0.74 | 12.00-16.00 | 5.00 | 32 | 34.06 | 1.30 | 31.00-37.00 | 3.82 | 32 | 46.04 | 1.96 | 42.00-50.00 | 4.35 |
| PG: Clinical Data Envoy 500 | 6 | 37.43 | 0.79 | 34.00-40.00 | 2.16 | 6 | 6.71 | 1.11 | 5.00-9.00 | 15.71 | 6 | 14.29 | 0.95 | 12.00-16.00 | 7.14 | 6 | 34.00 | 1.00 | 31.00-37.00 | 2.94 | 5 | 46.00 | 2.00 | 42.00-50.00 | 4.35 |
| PG: Siemens Dimension/AR/ES/RxL/Xpand | 19 | 40.28 | 1.49 | 36.00-44.00 | 3.75 | 18 | 6.13 | 0.55 | 4.00-8.00 | 8.33 | 19 | 14.60 | 1.04 | 13.00-17.00 | 6.67 | 19 | 35.92 | 1.41 | 33.00-39.00 | 3.89 | 19 | 48.68 | 1.95 | 45.00-53.00 | 4.08 |
| PG: Horiba ABX Pentra 400 | 5 | 38.83 | 1.47 | 35.00-43.00 | 3.85 | 5 | 5.67 | 0.52 | 4.00-8.00 | 8.33 | 5 | 14.00 | 0.00 | 12.00-16.00 | 0.00 | 5 | 35.17 | 0.98 | 32.00-38.00 | 2.86 | 5 | 46.33 | 1.51 | 42.00-50.00 | 3.26 |
| PG: Beckman (Olympus) AU | 17 | 38.71 | 1.93 | 35.00-43.00 | 4.87 | 17 | 6.65 | 0.70 | 5.00-9.00 | 10.00 | 17 | 14.88 | 0.93 | 13.00-17.00 | 6.00 | 18 | 35.00 | 1.57 | 32.00-38.00 | 4.57 | 17 | 47.06 | 2.38 | 43.00-51.00 | 5.11 |
| PG: Roche Cobas Integra / Roche Reagent | 8 | 38.26 | 1.49 | 35.00-41.00 | 3.95 | 8 | 6.26 | 0.46 | 4.00-8.00 | 8.33 | 8 | 13.75 | 0.46 | 12.00-16.00 | 3.57 | 8 | 33.77 | 1.04 | 31.00-37.00 | 2.94 | 8 | 46.18 | 1.86 | 42.00-50.00 | 4.13 |
| PG: OCD Vitros 250,350,400,500,750,950 | 11 | 34.00 | 0.63 | 31.00-37.00 | 1.76 | 11 | 5.00 | 0.45 | 3.00-7.00 | 8.00 | 11 | 12.00 | 0.45 | 10.00-14.00 | 3.33 | 11 | 30.36 | 0.81 | 27.00-33.00 | 2.67 | 11 | 41.91 | 0.70 | 38.00-46.00 | 1.67 |
| SG: Select Beckman Coulter Instruments (Chem) | 18 | 38.83 | 1.95 | 35.00-43.00 | 4.87 | 18 | 6.67 | 0.69 | 5.00-9.00 | 10.00 | 18 | 14.94 | 0.94 | 13.00-17.00 | 6.00 | 19 | 35.11 | 1.60 | 32.00-38.00 | 4.57 | 18 | 47.22 | 2.42 | 43.00-51.00 | 5.11 |
| SG: Select Roche Instruments / Roche Reagent | 9 | 38.45 | 1.51 | 35.00-41.00 | 3.95 | 9 | 6.23 | 0.44 | 4.00-8.00 | 6.67 | 9 | 13.88 | 0.60 | 12.00-16.00 | 4.29 | 9 | 34.02 | 1.23 | 31.00-37.00 | 3.53 | 9 | 46.38 | 1.84 | 42.00-50.00 | 3.91 |
| SG: Select Roche Instruments / Roche Reagents Groups | 10 | 38.51 | 1.43 | 35.00-43.00 | 3.59 | 10 | 6.31 | 0.48 | 4.00-8.00 | 8.33 | 10 | 14.10 | 0.88 | 12.00-16.00 | 6.43 | 10 | 34.02 | 1.16 | 31.00-37.00 | 3.53 | 10 | 46.34 | 1.74 | 42.00-50.00 | 3.70 |
| SG: Select OCD Vitros Instruments | 12 | 33.92 | 0.67 | 31.00-37.00 | 2.06 | 12 | 4.92 | 0.52 | 3.00-7.00 | 10.00 | 12 | 12.00 | 0.43 | 10.00-14.00 | 3.33 | 12 | 30.25 | 0.87 | 27.00-33.00 | 3.00 | 12 | 41.67 | 1.07 | 38.00-46.00 | 2.62 |
| SG: Abaxis Piccolo Instruments | 74 | 35.62 | 0.86 | 33.00-39.00 | 2.50 | 13 | 7.62 | 0.51 | 6.00-10.00 | 6.25 | 13 | 13.85 | 0.56 | 12.00-16.00 | 4.29 | 13 | 31.69 | 0.85 | 29.00-35.00 | 2.81 | 13 | 44.46 | 1.05 | 40.00-48.00 | 2.50 |
| PG: Polymedco Spot-Chem EZ | 5 | 19.80 | 0.00 | 18.00-22.00 | 0.00 | 5 | 6.60 | 0.00 | 5.00-9.00 | 0.00 | 5 | 11.80 | 0.00 | 10.00-14.00 | 0.00 | 5 | 20.90 | 0.00 | 19.00-23.00 | 0.00 | 5 | 16.40 | 0.00 | 14.00-18.00 | 0.00 |
| SG: Select Abbott Architect Instruments | 9 | 39.80 | 0.79 | 36.00-44.00 | 2.00 | 9 | 6.40 | 0.52 | 4.00-8.00 | 8.33 | 9 | 14.80 | 0.42 | 13.00-17.00 | 2.67 | 9 | 35.40 | 1.08 | 32.00-38.00 | 3.14 | 9 | 48.60 | 1.08 | 45.00-53.00 | 2.24 |
| SG: International OCD Vitros Instruments | 14 | 34.07 | 0.62 | 31.00-37.00 | 1.76 | 14 | 5.00 | 0.39 | 3.00-7.00 | 8.00 | 14 | 12.00 | 0.39 | 10.00-14.00 | 3.33 | 14 | 30.36 | 0.75 | 27.00-33.00 | 2.33 | 14 | 42.14 | 1.03 | 38.00-46.00 | 2.38 |
| SG: International Abaxis Piccolo Instruments | 79 | 35.63 | 0.85 | 33.00-39.00 | 2.22 | 18 | 7.64 | 0.48 | 6.00-10.00 | 6.25 | 18 | 13.91 | 0.54 | 12.00-16.00 | 3.57 | 18 | 31.75 | 0.81 | 29.00-35.00 | 2.50 | 18 | 44.53 | 0.97 | 41.00-49.00 | 2.22 |
| AG: All Group | 186 | 37.20 | 2.56 | 34.00-40.00 | 7.03 | 121 | 6.39 | 0.80 | 4.00-8.00 | 13.33 | 125 | 14.17 | 1.10 | 12.00-16.00 | 7.86 | 126 | 34.03 | 2.35 | 31.00-37.00 | 7.06 | 124 | 46.15 | 3.68 | 42.00-50.00 | 8.04 |
| Albumin | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| PG: Abaxis Piccolo (Non-Waived) / BCP | 11 | 2.95 | 0.07 | 2.60-3.20 | 2.41 | 11 | 1.96 | 0.05 | 1.80-2.20 | 2.50 | 11 | 2.21 | 0.05 | 2.00-2.40 | 2.27 | 11 | 2.80 | 0.08 | 2.50-3.10 | 2.86 | 11 | 3.15 | 0.11 | 2.90-3.50 | 3.44 |
| PG: Abaxis Piccolo (Waived) / BCP | 64 | 2.92 | 0.13 | 2.60-3.20 | 4.48 | | | | | | | | | | | | | | | | | | | | |
| PG: Abbott Architect / BCP | 7 | 2.91 | 0.04 | 2.60-3.20 | 1.38 | 7 | 1.90 | 0.00 | 1.70-2.10 | 0.00 | 7 | 2.18 | 0.05 | 2.00-2.40 | 2.27 | 7 | 2.80 | 0.00 | 2.50-3.10 | 0.00 | 7 | 3.20 | 0.00 | 2.90-3.50 | 0.00 |
| PG: Alfa Was. Ace/NexCT/Alera/Axcel / BCG / Alfa Wasserman Reagent | 33 | 3.87 | 0.09 | 3.50-4.30 | 2.31 | 33 | 2.07 | 0.05 | 1.90-2.30 | 2.38 | 33 | 2.53 | 0.06 | 2.30-2.80 | 2.40 | 33 | 3.68 | 0.08 | 3.30-4.10 | 2.16 | 33 | 4.35 | 0.11 | 4.00-4.80 | 2.50 |
| PG: Clinical Data Envoy 500 / BCG | 7 | 3.90 | 0.13 | 3.50-4.30 | 3.33 | 7 | 2.11 | 0.08 | 1.90-2.30 | 3.81 | 7 | 2.56 | 0.09 | 2.30-2.90 | 3.46 | 7 | 3.70 | 0.09 | 3.30-4.10 | 2.43 | 7 | 4.24 | 0.11 | 3.80-4.60 | 2.62 |
| PG: Siemens Dimension/AR/ES/RxL/Xpand / BCG | 11 | 3.09 | 0.11 | 2.80-3.40 | 3.55 | 11 | 1.83 | 0.07 | 1.60-2.00 | 3.89 | 11 | 2.13 | 0.07 | 1.90-2.30 | 3.33 | 11 | 2.96 | 0.11 | 2.70-3.30 | 3.67 | 11 | 3.41 | 0.11 | 3.10-3.70 | 3.24 |
| PG: Siemens Dimension/AR/ES/RxL/Xpand / BCP | 8 | 3.07 | 0.05 | 2.80-3.40 | 1.61 | 8 | 1.82 | 0.06 | 1.60-2.00 | 3.33 | 8 | 2.13 | 0.07 | 1.90-2.30 | 3.33 | 8 | 2.92 | 0.06 | 2.60-3.20 | 2.07 | 8 | 3.36 | 0.05 | 3.10-3.70 | 1.47 |
| PG: Horiba ABX Pentra 400 / BCG | 6 | 3.90 | 0.14 | 3.50-4.30 | 3.59 | 6 | 2.13 | 0.08 | 1.90-2.30 | 3.81 | 6 | 2.56 | 0.10 | 2.30-2.90 | 3.85 | 6 | 3.71 | 0.15 | 3.30-4.10 | 4.05 | 6 | 4.31 | 0.15 | 3.90-4.70 | 3.49 |
| PG: Beckman (Olympus) AU / BCG | 17 | 3.92 | 0.12 | 3.50-4.30 | 3.08 | 17 | 1.91 | 0.07 | 1.70-2.10 | 3.68 | 17 | 2.42 | 0.08 | 2.20-2.60 | 3.33 | 18 | 3.69 | 0.10 | 3.30-4.10 | 2.70 | 18 | 4.39 | 0.12 | 4.00-4.80 | 2.73 |
| PG: Roche Cobas Integra / BCG / Roche Reagent | 8 | 4.20 | 0.13 | 3.80-4.60 | 3.10 | 8 | 2.09 | 0.07 | 1.90-2.30 | 3.33 | 8 | 2.67 | 0.06 | 2.40-3.00 | 2.22 | 8 | 3.98 | 0.06 | 3.60-4.40 | 1.50 | 8 | 4.73 | 0.11 | 4.20-5.20 | 2.34 |
| PG: OCD Vitros 250,350,400,500,750,950 / BCG | 11 | 3.94 | 0.21 | 3.50-4.30 | 5.38 | 11 | 1.80 | 0.08 | 1.60-2.00 | 4.44 | 11 | 2.30 | 0.10 | 2.10-2.50 | 4.35 | 11 | 3.71 | 0.16 | 3.30-4.10 | 4.32 | 11 | 4.62 | 0.26 | 4.10-5.10 | 5.65 |
| SG: Select Roche Instruments / BCG / Roche Reagent | 9 | 4.20 | 0.12 | 3.80-4.60 | 2.86 | 9 | 2.08 | 0.07 | 1.90-2.30 | 3.33 | 9 | 2.67 | 0.06 | 2.40-3.00 | 2.22 | 9 | 3.98 | 0.06 | 3.60-4.40 | 1.50 | 9 | 4.73 | 0.10 | 4.20-5.20 | 2.13 |
| SG: Select Roche Instruments / BCG / Roche Reagents Groups | 10 | 4.17 | 0.15 | 3.80-4.60 | 3.57 | 1 | | | | | | | | | | | | | | | | | | | |



| PG: Abaxis Piccolo (Non-Waived) | 11 | 188.73 | 5.92 | 151.00-227.00 | 3.12 | 11 | 24.09 | 3.65 | 19.00-29.00 | 15.00 | 11 | 66.36 | 3.11 | 53.00-79.00 | 4.70 | 11 | 168.27 | 6.99 | 134.00-202.00 | 4.17 | 11 | 229.55 | 7.69 | 184.00-276.00 | 3.35 |
|---|-------------|-------------|-----------|---------------|-------------|-------------|-------------|-----------|--------------|-------------|-------------|-------------|-----------|--------------|-------------|-------------|-------------|-----------|---------------|-------------|-------------|-------------|-----------|---------------|-------------|
| PG: Abbott Architect | 7 | 218.50 | 1.60 | 175.00-263.00 | 0.73 | 7 | 16.38 | 0.92 | 13.00-19.00 | 5.63 | 7 | 68.63 | 1.06 | 55.00-83.00 | 1.59 | 7 | 194.63 | 1.69 | 156.00-234.00 | 0.87 | 7 | 269.00 | 1.85 | 215.00-323.00 | 0.71 |
| PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent | 32 | 176.09 | 5.97 | 141.00-211.00 | 3.41 | 28 | 10.77 | 1.65 | 9.00-13.00 | 14.55 | 32 | 53.35 | 5.95 | 42.00-64.00 | 11.32 | 32 | 158.32 | 3.64 | 126.00-190.00 | 2.28 | 32 | 218.65 | 5.46 | 175.00-263.00 | 2.51 |
| SG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent Groups | 33 | 174.67 | 10.35 | 140.00-210.00 | 5.89 | 29 | 11.03 | 2.18 | 9.00-13.00 | 20.00 | 33 | 53.80 | 6.43 | 43.00-65.00 | 11.85 | 33 | 157.66 | 5.33 | 126.00-190.00 | 3.35 | 33 | 215.77 | 17.84 | 173.00-259.00 | 8.24 |
| PG: Clinical Data Envoy 500 | 6 | 212.00 | 12.19 | 170.00-254.00 | 5.75 | 6 | 17.29 | 1.11 | 14.00-20.00 | 6.47 | 6 | 70.00 | 3.70 | 56.00-84.00 | 5.29 | 6 | 191.43 | 11.86 | 153.00-229.00 | 6.23 | 5 | 262.67 | 22.91 | 210.00-316.00 | 8.71 |
| PG: Siemens Dimension/AR/ES/RxL/Xpand | 8 | 211.30 | 8.46 | 169.00-253.00 | 4.03 | 8 | 18.90 | 1.91 | 15.00-23.00 | 10.00 | 8 | 69.10 | 2.56 | 55.00-83.00 | 3.77 | 8 | 191.00 | 3.30 | 153.00-229.00 | 1.73 | 8 | 261.30 | 3.89 | 209.00-313.00 | 1.49 |
| PG: Siemens Dimension/AR/ES/RxL/Xpand / IFCC | 11 | 218.80 | 4.49 | 175.00-263.00 | 2.05 | 11 | 20.40 | 1.12 | 16.00-24.00 | 5.50 | 11 | 72.47 | 2.33 | 58.00-86.00 | 3.19 | 11 | 194.47 | 3.36 | 155.00-233.00 | 1.75 | 11 | 265.67 | 4.59 | 213.00-319.00 | 1.73 |
| PG: Beckman (Olympus) AU | 17 | 196.76 | 7.78 | 158.00-236.00 | 3.96 | 16 | 15.69 | 0.95 | 13.00-19.00 | 5.63 | 17 | 63.59 | 5.81 | 51.00-77.00 | 9.06 | 18 | 177.28 | 10.67 | 142.00-212.00 | 6.05 | 18 | 240.61 | 9.21 | 193.00-289.00 | 3.82 |
| PG: Roche Cobas Integra / Roche Reagent | 8 | 208.30 | 3.48 | 166.00-250.00 | 1.68 | 8 | 16.70 | 0.61 | 14.00-20.00 | 3.53 | 8 | 66.68 | 1.17 | 54.00-80.00 | 1.79 | 8 | 185.80 | 2.36 | 149.00-223.00 | 1.29 | 8 | 255.31 | 4.67 | 204.00-306.00 | 1.84 |
| PG: OCD Vitros 250,350,400,500,750,950 | 10 | 197.10 | 5.71 | 158.00-236.00 | 2.89 | 9 | 35.67 | 4.80 | 29.00-43.00 | 13.33 | 10 | 76.00 | 6.04 | 61.00-91.00 | 7.89 | 10 | 177.30 | 6.46 | 142.00-212.00 | 3.67 | 10 | 235.00 | 8.21 | 188.00-282.00 | 3.49 |
| SG: Select Alfa Wasserman Instruments / Alfa Wasserman Reagent Groups | 33 | 174.67 | 10.35 | 140.00-210.00 | 5.89 | | | | | | 33 | 53.80 | 6.43 | 43.00-65.00 | 11.85 | 33 | 157.66 | 5.33 | 126.00-190.00 | 3.35 | 33 | 215.77 | 17.84 | 173.00-259.00 | 8.24 |
| SG: Select Beckman Coulter Instruments (Chem) | 18 | 197.67 | 8.46 | 158.00-238.00 | 4.29 | 17 | 16.12 | 2.00 | 13.00-19.00 | 12.50 | 18 | 64.11 | 6.06 | 51.00-77.00 | 9.53 | 19 | 177.79 | 10.61 | 142.00-214.00 | 5.96 | 19 | 241.21 | 9.32 | 193.00-289.00 | 3.86 |
| SG: Select Roche Instruments / Roche Reagent | 9 | 209.71 | 5.34 | 168.00-252.00 | 2.52 | 9 | 16.51 | 0.81 | 14.00-20.00 | 4.71 | 9 | 66.82 | 1.18 | 54.00-80.00 | 1.79 | 9 | 185.82 | 2.21 | 149.00-223.00 | 1.18 | 9 | 257.50 | 7.88 | 206.00-308.00 | 3.07 |
| SG: Select Roche Instruments / Roche Reagents Groups | 10 | 206.94 | 10.11 | 166.00-248.00 | 4.88 | 10 | 16.06 | 1.62 | 13.00-19.00 | 10.00 | 10 | 65.94 | 3.00 | 53.00-79.00 | 4.55 | 10 | 183.64 | 7.21 | 147.00-221.00 | 3.91 | 10 | 254.05 | 13.20 | 203.00-305.00 | 5.20 |
| SG: Select OCD Vitros Instruments | 11 | 197.27 | 5.44 | 158.00-236.00 | 2.74 | 10 | 35.40 | 4.60 | 28.00-42.00 | 13.14 | 11 | 76.09 | 5.74 | 61.00-91.00 | 7.50 | 11 | 177.64 | 6.23 | 142.00-214.00 | 3.48 | 11 | 235.45 | 7.93 | 188.00-282.00 | 3.36 |
| SG: Abaxis Piccolo Instruments | 73 | 186.47 | 3.94 | 149.00-223.00 | 2.10 | 12 | 23.92 | 3.53 | 19.00-29.00 | 14.58 | 12 | 66.08 | 3.12 | 53.00-79.00 | 4.70 | 12 | 167.75 | 6.90 | 134.00-202.00 | 4.11 | 12 | 229.00 | 7.57 | 183.00-275.00 | 3.32 |
| SG: Select Abbott Architect Instruments | 9 | 217.30 | 3.13 | 174.00-260.00 | 1.43 | 9 | 16.40 | 1.08 | 13.00-19.00 | 6.88 | 9 | 68.20 | 1.75 | 54.00-82.00 | 2.65 | 9 | 193.50 | 3.14 | 155.00-233.00 | 1.60 | 9 | 267.00 | 4.90 | 214.00-320.00 | 1.84 |
| SG: International OCD Vitros Instruments | 13 | 198.15 | 5.96 | 158.00-238.00 | 3.03 | 12 | 37.00 | 4.80 | 30.00-44.00 | 12.97 | 13 | 78.23 | 6.75 | 62.00-94.00 | 8.59 | 13 | 178.92 | 6.55 | 143.00-215.00 | 3.69 | 13 | 236.00 | 7.91 | 189.00-283.00 | 3.35 |
| SG: International Abaxis Piccolo Instruments | 75 | 186.43 | 3.93 | 149.00-223.00 | 2.10 | 14 | 24.00 | 3.26 | 19.00-29.00 | 13.75 | 14 | 65.71 | 3.02 | 53.00-79.00 | 4.55 | 14 | 167.50 | 6.49 | 134.00-202.00 | 3.87 | 14 | 228.71 | 7.04 | 183.00-275.00 | 3.06 |
| SG: Select Gen Chem Instruments | 37 | 209.05 | 14.59 | 167.00-251.00 | 6.99 | 35 | 16.49 | 1.35 | 13.00-19.00 | 8.13 | 37 | 67.31 | 5.63 | 54.00-80.00 | 8.36 | 38 | 187.23 | 13.62 | 150.00-224.00 | 7.27 | 37 | 255.69 | 19.31 | 205.00-307.00 | 7.54 |
| MG: IFCC | 13 | 223.00 | 9.80 | 178.00-268.00 | 4.39 | 13 | 21.00 | 1.00 | 17.00-25.00 | 4.76 | 13 | 74.00 | 3.50 | 59.00-89.00 | 4.73 | 13 | 198.00 | 9.70 | 158.00-238.00 | 4.90 | 13 | 271.00 | 12.60 | 217.00-325.00 | 4.65 |
| AST/SGOT | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| PG: Abaxis Piccolo (Non-Waived) | 11 | 252.73 | 5.08 | 202.00-304.00 | 2.02 | 11 | 23.00 | 1.55 | 18.00-28.00 | 6.52 | 11 | 81.18 | 2.44 | 65.00-97.00 | 2.96 | 11 | 224.55 | 5.24 | 180.00-270.00 | 2.31 | 11 | 310.82 | 7.45 | 249.00-373.00 | 2.41 |
| PG: Abbott Architect | 7 | 256.13 | 3.64 | 205.00-307.00 | 1.41 | 7 | 17.13 | 1.13 | 14.00-20.00 | 6.47 | 7 | 77.25 | 1.58 | 62.00-92.00 | 2.08 | 7 | 227.88 | 3.14 | 182.00-274.00 | 1.36 | 7 | 317.38 | 5.10 | 254.00-380.00 | 1.61 |
| PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent | 32 | 233.54 | 6.36 | 187.00-281.00 | 2.74 | 32 | 13.00 | 1.57 | 10.00-16.00 | 12.31 | 32 | 69.00 | 2.80 | 55.00-83.00 | 4.06 | 32 | 210.09 | 5.43 | 168.00-252.00 | 2.57 | 32 | 289.71 | 8.16 | 232.00-348.00 | 2.83 |
| SG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent Groups | 33 | 233.67 | 6.31 | 187.00-281.00 | 2.69 | 33 | 12.97 | 1.56 | 10.00-16.00 | 12.31 | 33 | 69.00 | 2.75 | 55.00-83.00 | 4.06 | 33 | 210.14 | 5.36 | 168.00-252.00 | 2.57 | 33 | 289.63 | 8.05 | 232.00-348.00 | 2.79 |
| PG: Clinical Data Envoy 500 | 6 | 262.57 | 10.49 | 210.00-316.00 | 3.99 | 6 | 18.14 | 1.35 | 14.00-22.00 | 7.22 | 6 | 82.29 | 3.40 | 66.00-98.00 | 4.15 | 6 | 236.00 | 11.49 | 189.00-283.00 | 4.87 | 5 | 324.83 | 15.12 | 260.00-390.00 | 4.65 |
| PG: Siemens Dimension/AR/ES/RxL/Xpand | 19 | 273.68 | 8.38 | 219.00-329.00 | 3.07 | 19 | 16.88 | 2.30 | 14.00-20.00 | 13.53 | 19 | 82.12 | 3.19 | 66.00-98.00 | 3.90 | 19 | 241.00 | 7.05 | 193.00-289.00 | 2.90 | 19 | 338.40 | 9.09 | 270.00-406.00 | 2.69 |
| PG: Beckman (Olympus) AU | 17 | 234.94 | 22.62 | 188.00-322.00 | 9.62 | 17 | 16.41 | 1.37 | 13.00-19.00 | 8.75 | 17 | 71.47 | 4.77 | 57.00-85.00 | 6.76 | 18 | 208.22 | 18.36 | 166.00-250.00 | 8.85 | 18 | 290.61 | 33.58 | 233.00-349.00 | 11.55 |
| PG: Roche Cobas Integra / Roche Reagent | 8 | 257.80 | 9.51 | 206.00-310.00 | 3.68 | 8 | 18.06 | 0.78 | 14.00-22.00 | 4.44 | 8 | 79.62 | 2.82 | 64.00-96.00 | 3.50 | 8 | 229.81 | 7.87 | 184.00-276.00 | 3.43 | 8 | 320.33 | 10.94 | 256.00-384.00 | 3.41 |
| PG: OCD Vitros 250,350,400,500,750,950 | 10 | 306.10 | 8.44 | 245.00-367.00 | 2.75 | 10 | 18.50 | 0.85 | 15.00-23.00 | 4.21 | 10 | 85.00 | 3.94 | 68.00-102.00 | 4.59 | 10 | 268.20 | 9.80 | 214.00-322.00 | 3.66 | 10 | 405.40 | 12.75 | 324.00-486.00 | 3.14 |
| SG: Select Beckman Coulter Instruments (Chem) | 18 | 236.33 | 22.73 | 189.00-283.00 | 9.62 | 18 | 16.78 | 2.04 | 14.00-20.00 | 11.76 | 18 | 72.11 | 5.37 | 58.00-86.00 | 7.50 | 19 | 209.21 | 18.35 | 167.00-251.00 | 8.80 | 19 | 291.84 | 33.07 | 234.00-350.00 | 11.34 |
| SG: Select Roche Instruments / Roche Reagent | 9 | 259.38 | 10.08 | 207.00-311.00 | 3.90 | 9 | 17.72 | 1.25 | 14.00-22.00 | 7.22 | 9 | 80.11 | 3.02 | 64.00-96.00 | 3.75 | 9 | 231.05 | 8.25 | 185.00-277.00 | 3.59 | 9 | 321.85 | 11.20 | 258.00-386.00 | 3.48 |
| SG: Select Roche Instruments / Roche Reagents Groups | 10 | 255.44 | 15.66 | 204.00-306.00 | 6.16 | 10 | 17.55 | 1.30 | 14.00-22.00 | 7.22 | 10 | 79.10 | 4.28 | 63.00-95.00 | 5.44 | 10 | 227.85 | 12.78 | 182.00-274.00 | 5.61 | 10 | 317.47 | 17.43 | 254.00-380.00 | 5.49 |
| SG: Select OCD Vitros Instruments | 11 | 305.91 | 8.03 | 245.00-367.00 | 2.61 | 11 | 18.36 | 0.92 | 14.00-22.00 | 5.00 | 11 | 84.45 | 4.16 | 67.00-101.00 | 5.00 | 11 | 267.09 | 9.99 | 214.00-320.00 | 3.75 | 11 | 404.64 | 12.36 | 324.00-486.00 | 3.06 |
| SG: Abaxis Piccolo Instruments | 74 | 249.07 | 7.43 | 199.00-299.00 | 2.97 | 12 | 23.00 | 1.48 | 18.00-28.00 | 6.52 | 12 | 81.42 | 2.47 | 65.00-97.00 | 3.09 | 12 | 225.25 | 5.66 | 180.00-270.00 | 2.49 | 12 | 311.92 | 8.06 | 250.00-374.00 | 2.60 |
| SG: Select Abbott Architect Instruments | 9 | 255.00 | 4.32 | 204.00-306.00 | 1.69 | 9 | 17.00 | 1.05 | 14.00-20.00 | 6.47 | 9 | 77.10 | 1.45 | 62.00-92.00 | 1.82 | 9 | 226.80 | 3.65 | 182.00-272.00 | 1.59 | 9 | 315.60 | 5.89 | 253.00-379.00 | 1.87 |
| SG: International OCD Vitros Instruments | 13 | 307.31 | 9.39 | 246.00-368.00 | 3.06 | 13 | 18.46 | 0.78 | 14.00-22.00 | 4.44 | 13 | 85.31 | 3.68 | 68.00-102.00 | 4.35 | 13 | 270.62 | 11.21 | 217.00-325.00 | 4.13 | 13 | 409.69 | 17.16 | 328.00-492.00 | 4.20 |
| SG: International Abaxis Piccolo Instruments | 76 | 249.08 | 7.36 | 199.00-299.00 | 2.97 | 14 | 22.93 | 1.69 | 18.00-28.00 | 7.39 | 14 | 81.14 | 2.38 | 65.00-97.00 | 2.96 | 14 | 224.36 | 5.61 | 179.00-269.00 | 2.50 | 14 | 310.57 | 8.17 | 249.00-373.00 | 2.64 |
| AG: All Group | 186 | 252.79 | 20.45 | 202.00-304.00 | 8.10 | 114 | 16.23 | 2.61 | 13.00-19.00 | 16.25 | 124 | 76.81 | 6.69 | 62.00-92.00 | 8.70 | 125 | 226.50 | 19.58 | 182.00-272.00 | 8.63 | 124 | 317.35 | 35.21 | 254.00-380.00 | 11.10 |
| Bilirubin Total | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| PG: Abaxis Piccolo (Non-Waived) | 11 | 3.25 | 0.16 | 2.60-3.80 | 5.00 | 11 | 0.40 | 0.06 | 0.00-0.80 | 15.00 | 11 | 1.06 | 0.08 | 0.70-1.50 | 7.27 | 11 | 2.91 | 0.15 | 2.30-3.50 | 5.17 | 11 | 4.07 | 0.20 | 3.30-4.90 | 5.12 |
| PG: Abbott Architect | 7 | 3.10 | 0.22 | 2.50-3.70 | 6.77 | 7 | 0.19 | 0.07 | 0.00-0.60 | 35.00 | 7 | 0.92 | 0.12 | 0.50-1.30 | 13.33 | 7 | 2.79 | | | | | | | | |



| | Labs | Mean | SD | Range | CV, % | Labs | Mean | SD | Range | CV, % | Labs | Mean | SD | Range | CV, % | Labs | Mean | SD | Range | CV, % | Labs | Mean | SD | Range | CV, % |
|---|------|--------|-------|---------------|-------|------|--------|------|---------------|-------|------|--------|-------|---------------|-------|------|--------|-------|---------------|-------|------|--------|-------|---------------|-------|
| PG: Horiba ABX Pentra 400 | 6 | 6.99 | 0.27 | 6.30-7.70 | 3.86 | 6 | 2.89 | 0.11 | 2.60-3.20 | 3.79 | 6 | 3.89 | 0.16 | 3.50-4.30 | 4.10 | 6 | 6.47 | 0.24 | 5.80-7.20 | 3.69 | 6 | 8.04 | 0.32 | 7.20-8.80 | 4.00 |
| PG: Beckman (Olympus) AU | 17 | 7.07 | 0.19 | 6.40-7.80 | 2.68 | 17 | 2.91 | 0.18 | 2.60-3.20 | 6.21 | 17 | 3.97 | 0.11 | 3.60-4.40 | 2.75 | 18 | 6.59 | 0.18 | 5.90-7.30 | 2.73 | 18 | 8.09 | 0.22 | 7.30-8.90 | 2.72 |
| PG: Roche Cobas Integra / Roche Reagent | 8 | 6.97 | 0.14 | 6.30-7.70 | 2.00 | 8 | 2.89 | 0.10 | 2.60-3.20 | 3.45 | 8 | 3.90 | 0.09 | 3.50-4.30 | 2.31 | 8 | 6.49 | 0.10 | 5.80-7.20 | 1.54 | 8 | 8.03 | 0.17 | 7.20-8.80 | 2.13 |
| PG: OCD Vitros 250,350,400,500,750,950 | 10 | 6.76 | 0.22 | 6.10-7.50 | 3.24 | 10 | 2.96 | 0.08 | 2.70-3.30 | 2.67 | 10 | 4.02 | 0.12 | 3.60-4.40 | 3.00 | 10 | 6.40 | 0.24 | 5.80-7.00 | 3.75 | 10 | 7.77 | 0.21 | 7.00-8.60 | 2.69 |
| SG: Select Beckman Coulter Instruments (Chem) | 18 | 7.07 | 0.18 | 6.40-7.80 | 2.54 | 18 | 2.91 | 0.17 | 2.60-3.20 | 5.86 | 18 | 3.97 | 0.11 | 3.60-4.40 | 2.75 | 19 | 6.59 | 0.18 | 5.90-7.30 | 2.73 | 19 | 8.08 | 0.22 | 7.30-8.90 | 2.72 |
| SG: Select Roche Instruments / Roche Reagent | 9 | 6.97 | 0.13 | 6.30-7.70 | 1.86 | 9 | 2.85 | 0.13 | 2.60-3.20 | 4.48 | 9 | 3.92 | 0.09 | 3.50-4.30 | 2.31 | 9 | 6.50 | 0.10 | 5.80-7.20 | 1.54 | 9 | 8.03 | 0.16 | 7.20-8.80 | 2.00 |
| SG: Select Roche Instruments / Roche Reagents Groups | 10 | 6.97 | 0.13 | 6.30-7.70 | 1.86 | 10 | 2.86 | 0.13 | 2.60-3.20 | 4.48 | 10 | 3.92 | 0.09 | 3.50-4.30 | 2.31 | 10 | 6.52 | 0.11 | 5.80-7.20 | 1.69 | 10 | 8.06 | 0.17 | 7.30-8.90 | 2.10 |
| SG: Select OCD Vitros Instruments | 11 | 6.76 | 0.21 | 6.10-7.50 | 3.09 | 11 | 2.96 | 0.08 | 2.70-3.30 | 2.67 | 11 | 4.02 | 0.12 | 3.60-4.40 | 3.00 | 11 | 6.39 | 0.23 | 5.80-7.00 | 3.59 | 11 | 7.76 | 0.20 | 7.00-8.60 | 2.56 |
| SG: Abaxis Piccolo Instruments | 72 | 7.14 | 0.11 | 6.40-7.80 | 1.55 | 11 | 3.12 | 0.06 | 2.80-3.40 | 1.94 | 11 | 4.11 | 0.07 | 3.70-4.50 | 1.71 | 11 | 6.63 | 0.14 | 5.90-7.30 | 2.12 | 11 | 8.25 | 0.19 | 7.40-9.00 | 2.32 |
| SG: Select Abbott Architect Instruments | 9 | 7.21 | 0.10 | 6.50-7.90 | 1.39 | 9 | 2.83 | 0.05 | 2.50-3.10 | 1.79 | 9 | 3.90 | 0.05 | 3.50-4.30 | 1.28 | 9 | 6.60 | 0.11 | 5.90-7.30 | 1.67 | 9 | 8.31 | 0.13 | 7.50-9.10 | 1.57 |
| SG: International OCD Vitros Instruments | 13 | 6.75 | 0.19 | 6.10-7.50 | 2.79 | 13 | 2.95 | 0.08 | 2.60-3.20 | 2.76 | 13 | 4.02 | 0.11 | 3.60-4.40 | 2.75 | 13 | 6.39 | 0.21 | 5.80-7.00 | 3.28 | 13 | 7.79 | 0.19 | 7.00-8.60 | 2.31 |
| SG: International Abaxis Piccolo Instruments | 73 | 7.14 | 0.11 | 6.40-7.80 | 1.55 | 12 | 3.12 | 0.06 | 2.80-3.40 | 1.94 | 12 | 4.12 | 0.07 | 3.70-4.50 | 1.71 | 12 | 6.64 | 0.14 | 5.90-7.30 | 2.12 | 12 | 8.26 | 0.18 | 7.50-9.10 | 2.17 |
| AG: All Group | 183 | 7.17 | 0.24 | 6.50-7.90 | 3.33 | 122 | 2.99 | 0.15 | 2.70-3.30 | 5.00 | 122 | 4.06 | 0.15 | 3.70-4.50 | 3.66 | 123 | 6.68 | 0.25 | 6.00-7.40 | 3.73 | 123 | 8.25 | 0.36 | 7.40-9.00 | 4.39 |
| Phosphorous | Labs | Mean | SD | Range | CV, % | Labs | Mean | SD | Range | CV, % | Labs | Mean | SD | Range | CV, % | Labs | Mean | SD | Range | CV, % | Labs | Mean | SD | Range | CV, % |
| PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent | 8 | 4.78 | 0.07 | 4.60-5.00 | 1.46 | 8 | 1.81 | 0.06 | 1.60-2.00 | 3.33 | 8 | 2.56 | 0.07 | 2.40-2.80 | 2.69 | 8 | 4.43 | 0.09 | 4.10-4.70 | 2.05 | 8 | 5.49 | 0.08 | 5.30-5.70 | 1.45 |
| PG: Siemens Dimension/AR/ES/RxL/Xpand | 6 | 5.29 | 0.14 | 4.90-5.70 | 2.64 | 6 | 1.86 | 0.05 | 1.80-2.00 | 2.63 | 6 | 2.77 | 0.13 | 2.40-3.20 | 4.64 | 6 | 4.81 | 0.08 | 4.60-5.00 | 1.67 | 6 | 6.11 | 0.09 | 5.80-6.40 | 1.48 |
| PG: Beckman (Olympus) AU | 12 | 4.96 | 0.14 | 4.60-5.40 | 2.80 | 12 | 1.65 | 0.09 | 1.40-2.00 | 5.29 | 12 | 2.50 | 0.10 | 2.20-2.80 | 4.00 | 13 | 4.58 | 0.09 | 4.30-4.90 | 1.96 | 13 | 5.81 | 0.11 | 5.50-6.10 | 1.90 |
| SG: Select Beckman Coulter Instruments (Chem) | 13 | 4.98 | 0.16 | 4.50-5.50 | 3.20 | 13 | 1.65 | 0.09 | 1.40-2.00 | 5.29 | 13 | 2.51 | 0.10 | 2.20-2.80 | 4.00 | 14 | 4.61 | 0.12 | 4.20-5.00 | 2.61 | 14 | 5.84 | 0.15 | 5.30-6.30 | 2.59 |
| AG: All Group | 43 | 5.07 | 0.23 | 4.40-5.80 | 4.51 | 41 | 1.77 | 0.12 | 1.40-2.20 | 6.67 | 41 | 2.60 | 0.14 | 2.20-3.00 | 5.38 | 42 | 4.65 | 0.17 | 4.10-5.10 | 3.70 | 42 | 5.87 | 0.24 | 5.20-6.60 | 4.07 |
| Cholesterol Total | Labs | Mean | SD | Range | CV, % | Labs | Mean | SD | Range | CV, % | Labs | Mean | SD | Range | CV, % | Labs | Mean | SD | Range | CV, % | Labs | Mean | SD | Range | CV, % |
| PG: Abaxis Piccolo (Non-Waived) | 5 | 192.40 | 1.52 | 173.00-211.00 | 0.78 | 5 | 105.60 | 0.89 | 95.00-117.00 | 0.85 | 5 | 129.80 | 1.30 | 117.00-143.00 | 1.00 | 5 | 184.00 | 3.08 | 166.00-202.00 | 1.68 | 5 | 210.60 | 3.36 | 190.00-232.00 | 1.61 |
| PG: Abbott Architect | 5 | 206.33 | 1.03 | 185.00-227.00 | 0.49 | 5 | 110.00 | 0.63 | 99.00-121.00 | 0.55 | 5 | 131.83 | 4.88 | 119.00-145.00 | 3.71 | 5 | 193.33 | 0.52 | 174.00-212.00 | 0.26 | 5 | 230.83 | 0.98 | 208.00-254.00 | 0.43 |
| PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent | 28 | 206.03 | 3.94 | 185.00-227.00 | 1.89 | 28 | 112.48 | 2.48 | 101.00-123.00 | 2.23 | 28 | 136.57 | 3.51 | 123.00-151.00 | 2.55 | 28 | 194.03 | 4.75 | 175.00-213.00 | 2.42 | 28 | 229.93 | 7.59 | 207.00-253.00 | 3.30 |
| PG: Clinical Data Envoy 500 | 5 | 210.50 | 15.50 | 190.00-232.00 | 7.35 | 5 | 108.00 | 5.97 | 97.00-119.00 | 5.56 | 5 | 140.17 | 12.61 | 126.00-154.00 | 9.00 | 5 | 197.50 | 10.67 | 178.00-218.00 | 5.40 | 5 | 232.33 | 10.46 | 209.00-255.00 | 4.53 |
| PG: Siemens Dimension/AR/ES/RxL/Xpand | 18 | 202.13 | 5.28 | 182.00-222.00 | 2.62 | 18 | 110.46 | 2.96 | 99.00-121.00 | 2.73 | 18 | 133.46 | 2.81 | 120.00-146.00 | 2.11 | 18 | 190.92 | 4.50 | 172.00-210.00 | 2.36 | 18 | 227.08 | 4.12 | 204.00-250.00 | 1.81 |
| PG: Horiba ABX Pentra 400 | 6 | 210.29 | 9.57 | 189.00-231.00 | 4.57 | 6 | 111.71 | 5.02 | 101.00-123.00 | 4.46 | 6 | 137.71 | 7.13 | 124.00-152.00 | 5.14 | 6 | 195.86 | 8.93 | 176.00-216.00 | 4.54 | 6 | 235.14 | 9.62 | 212.00-259.00 | 4.09 |
| PG: Beckman (Olympus) AU | 16 | 200.38 | 4.72 | 180.00-220.00 | 2.35 | 16 | 106.69 | 2.41 | 96.00-118.00 | 2.24 | 16 | 130.38 | 2.34 | 117.00-143.00 | 1.77 | 17 | 188.82 | 4.26 | 170.00-208.00 | 2.28 | 17 | 223.41 | 4.33 | 201.00-245.00 | 1.93 |
| PG: Roche Cobas Integra / Roche Reagent | 7 | 201.64 | 4.57 | 182.00-222.00 | 2.28 | 7 | 109.88 | 2.26 | 99.00-121.00 | 2.09 | 7 | 132.82 | 2.59 | 120.00-146.00 | 1.95 | 7 | 191.25 | 3.50 | 172.00-210.00 | 1.83 | 7 | 226.60 | 4.45 | 204.00-250.00 | 1.98 |
| PG: OCD Vitros 250,350,400,500,750,950 | 10 | 199.40 | 9.97 | 179.00-219.00 | 5.03 | 10 | 99.90 | 5.59 | 90.00-110.00 | 5.60 | 10 | 124.90 | 6.40 | 113.00-138.00 | 5.12 | 10 | 187.80 | 7.57 | 169.00-207.00 | 4.04 | 10 | 225.10 | 11.39 | 203.00-248.00 | 5.07 |
| SG: Select Beckman Coulter Instruments (Chem) | 17 | 201.18 | 5.64 | 181.00-221.00 | 2.79 | 17 | 106.88 | 2.47 | 96.00-118.00 | 2.34 | 17 | 130.71 | 2.64 | 118.00-144.00 | 1.98 | 18 | 189.56 | 5.17 | 171.00-209.00 | 2.74 | 18 | 224.06 | 5.01 | 202.00-246.00 | 2.23 |
| SG: Select Roche Instruments / Roche Reagent | 8 | 202.81 | 5.37 | 183.00-223.00 | 2.66 | 8 | 110.77 | 3.27 | 100.00-122.00 | 2.97 | 8 | 133.97 | 4.04 | 121.00-147.00 | 2.99 | 8 | 192.22 | 4.25 | 173.00-211.00 | 2.19 | 8 | 227.52 | 4.88 | 205.00-251.00 | 2.15 |
| SG: Select Roche Instruments / Roche Reagents Groups | 9 | 202.72 | 5.03 | 183.00-223.00 | 2.46 | 9 | 110.46 | 3.20 | 99.00-121.00 | 2.91 | 9 | 133.64 | 3.90 | 121.00-147.00 | 2.91 | 9 | 191.75 | 4.21 | 173.00-211.00 | 2.19 | 9 | 226.69 | 5.21 | 204.00-250.00 | 2.29 |
| SG: Select OCD Vitros Instruments | 11 | 199.09 | 9.51 | 179.00-219.00 | 4.77 | 11 | 99.00 | 6.08 | 89.00-109.00 | 6.16 | 11 | 124.00 | 6.77 | 112.00-136.00 | 5.48 | 11 | 187.18 | 7.47 | 168.00-206.00 | 4.01 | 11 | 224.64 | 10.91 | 203.00-248.00 | 4.84 |
| SG: Abaxis Piccolo Instruments | 14 | 192.64 | 2.53 | 174.00-212.00 | 1.30 | 6 | 105.50 | 0.84 | 95.00-117.00 | 0.75 | 6 | 130.00 | 1.26 | 117.00-143.00 | 1.00 | 6 | 183.83 | 2.79 | 166.00-202.00 | 1.52 | 6 | 211.00 | 3.16 | 190.00-232.00 | 1.52 |
| SG: Select Abbott Architect Instruments | 7 | 205.63 | 2.45 | 185.00-227.00 | 1.17 | 7 | 109.38 | 1.51 | 98.00-120.00 | 1.38 | 7 | 131.50 | 4.38 | 119.00-145.00 | 3.33 | 7 | 192.38 | 2.62 | 173.00-211.00 | 1.35 | 7 | 229.88 | 2.90 | 207.00-253.00 | 1.26 |
| SG: International OCD Vitros Instruments | 13 | 199.15 | 8.65 | 179.00-219.00 | 4.32 | 13 | 99.23 | 5.07 | 89.00-109.00 | 5.15 | 13 | 124.46 | 5.80 | 112.00-136.00 | 4.68 | 13 | 187.46 | 6.70 | 168.00-206.00 | 3.58 | 13 | 224.46 | 10.07 | 202.00-246.00 | 4.51 |
| SG: International Abaxis Piccolo Instruments | 15 | 192.27 | 2.84 | 173.00-211.00 | 1.46 | 7 | 105.57 | 0.79 | 95.00-117.00 | 0.75 | 7 | 130.14 | 1.22 | 117.00-143.00 | 0.92 | 7 | 183.71 | 2.56 | 166.00-202.00 | 1.41 | 7 | 210.86 | 2.91 | 190.00-232.00 | 1.37 |
| AG: All Group | 120 | 202.99 | 7.91 | 183.00-223.00 | 3.89 | 112 | 109.14 | 6.22 | 98.00-120.00 | 5.69 | 112 | 133.24 | 6.15 | 120.00-146.00 | 4.59 | 113 | 191.63 | 6.32 | 173.00-211.00 | 3.28 | 113 | 227.65 | 8.91 | 205.00-251.00 | 3.90 |
| Cholesterol HDL | Labs | Mean | SD | Range | CV, % | Labs | Mean | SD | Range | CV, % | Labs | Mean | SD | Range | CV, % | Labs | Mean | SD | Range | CV, % | Labs | Mean | SD | Range | CV, % |
| PG: Alfa Was. Ace/NexCT/Alera/Axcel / Direct Method / Alfa Wasserman Reagent | 27 | 85.33 | 4.16 | 60.00-111.00 | 4.94 | 27 | 47.90 | 2.80 | 34.00-62.00 | 5.83 | 27 | 57.45 | 3.36 | 40.00-74.00 | 5.96 | 27 | 80.72 | 3.97 | 57.00-105.00 | 4.94 | 27 | 94.24 | 5.83 | 66.00-122.00 | 6.17 |
| SG: Alfa Was. Ace/NexCT/Alera/Axcel / Direct Method / Alfa Wasserman Reagent Groups | 28 | 84.61 | 5.73 | 60.00-111.00 | 6.71 | 28 | 47.52 | 3.48 | 34.00-62.00 | 7.29 | 28 | 57.07 | 3.90 | 40.00-74.00 | 6.84 | 28 | 80.10 | 5.19 | 56.00-104.00 | 6.50 | 28 | 93.80 | 6.22 | 66.00-122.00 | 6.60 |
| PG: Clinical Data Envoy 500 / Direct Method | 5 | 95.33 | 9.07 | 67.00-124.00 | 9.58 | 5 | 54.83 | 7.05 | 39.00-72.00 | 12.91 | 5 | 63.67 | 8.76 | 45.00-83.00 | 13.75 | 5 | 88.83 | 10.42 | 62.00-116.00 | 11.69 | 5 | 98.67 | 23.70 | 69.00-129.00 | 23.94 |
| PG: Siemens Dimension/AR/ES/RxL/Xpand / Direct Method | 17 | 123.87 | 3.78 | 87.00-161.00 | 3.06 | 17 | 86.48 | 6.35 | 60.00-112.00 | 7.44 | 17 | 95.70 | 2.46 | 67.00-125.00 | 2.60 | 17 | 120.35 | 3.75 | 84.00-156.00 | 3.08 | 17 | 133.26 | 3.62 | 93.00-173.00 | 2.71 |
| PG: Horiba ABX Pentra 400 / Direct Method | 5 | 81.83 | 4.79 | 57.00-107.00 | 5.85 | 5 | 46.50 | 2.51 | 33.00-61.00 | 5.32 | 5 | 56.33 | 1.21 | 39.00-73.00 | 2.14 | 5 | 79.67 | 4.72 | 56.00-104.00 | 5.88 | 5 | 94.17 | 5.71 | 66.00-122.00 | 6.06 |
| PG: Beckman (Olympus) AU / Direct Method | 16 | 93.81 | 5.71 | 66.00-122.00 | 6.06 | 16 | 52.44 | 3.93 | | | | | | | | | | | | | | | | | |



| PG: Abaxis Piccolo (Non-Waived) | 5 | 217.80 | 3.63 | 164.00-273.00 | 1.65 | 5 | 93.60 | 0.55 | 71.00-118.00 | 0.53 | 5 | 124.80 | 2.77 | 94.00-156.00 | 2.24 | 5 | 203.00 | 3.54 | 152.00-254.00 | 1.72 | 5 | 250.20 | 4.92 | 188.00-313.00 | 1.96 |
|--|-------------|-------------|-----------|---------------|-------------|-------------|-------------|-----------|--------------|-------------|-------------|-------------|-----------|--------------|-------------|-------------|-------------|-----------|---------------|-------------|-------------|-------------|-----------|---------------|-------------|
| PG: Abaxis Piccolo (Waived) | 9 | 217.33 | 3.54 | 163.00-271.00 | 1.61 | | | | | | | | | | | | | | | | | | | | |
| PG: Abbott Architect | 5 | 198.33 | 3.44 | 149.00-248.00 | 1.72 | 5 | 83.83 | 1.60 | 63.00-105.00 | 1.90 | 5 | 112.17 | 1.33 | 84.00-140.00 | 1.16 | 5 | 182.83 | 2.14 | 137.00-229.00 | 1.15 | 5 | 225.67 | 1.97 | 170.00-283.00 | 0.88 |
| PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent | 28 | 201.03 | 6.01 | 151.00-251.00 | 2.99 | 28 | 92.13 | 3.67 | 69.00-115.00 | 4.02 | 28 | 120.40 | 4.19 | 90.00-150.00 | 3.50 | 28 | 188.80 | 6.28 | 142.00-236.00 | 3.33 | 28 | 231.13 | 10.68 | 173.00-289.00 | 4.63 |
| PG: Clinical Data Envoy 500 | 6 | 199.57 | 4.93 | 150.00-250.00 | 2.45 | 6 | 85.71 | 2.69 | 65.00-108.00 | 3.14 | 6 | 116.00 | 3.87 | 87.00-145.00 | 3.36 | 6 | 185.86 | 5.76 | 140.00-233.00 | 3.12 | 5 | 227.82 | 4.70 | 171.00-285.00 | 2.06 |
| PG: Siemens Dimension/AR/ES/RxL/Xpand | 18 | 192.25 | 4.75 | 144.00-240.00 | 2.45 | 18 | 78.75 | 2.72 | 59.00-99.00 | 3.42 | 18 | 107.50 | 2.21 | 81.00-135.00 | 2.04 | 18 | 178.88 | 3.92 | 134.00-224.00 | 2.18 | 18 | 222.13 | 5.44 | 167.00-278.00 | 2.43 |
| PG: Horiba ABX Pentra 400 | 6 | 199.71 | 6.90 | 150.00-250.00 | 3.45 | 6 | 89.00 | 5.75 | 67.00-111.00 | 6.40 | 6 | 117.57 | 5.16 | 89.00-148.00 | 4.41 | 6 | 187.86 | 8.28 | 141.00-235.00 | 4.41 | 6 | 227.86 | 7.58 | 171.00-285.00 | 3.33 |
| PG: Beckman (Olympus) AU | 15 | 200.73 | 4.64 | 151.00-251.00 | 2.29 | 15 | 86.73 | 2.25 | 65.00-109.00 | 2.64 | 15 | 115.07 | 2.66 | 86.00-144.00 | 2.35 | 16 | 186.75 | 4.43 | 140.00-234.00 | 2.35 | 16 | 229.00 | 5.66 | 172.00-286.00 | 2.49 |
| PG: Roche Cobas Integra / Roche Reagent | 7 | 196.86 | 3.72 | 148.00-246.00 | 1.88 | 7 | 89.38 | 1.86 | 67.00-111.00 | 2.13 | 7 | 117.41 | 2.37 | 88.00-146.00 | 2.05 | 7 | 185.54 | 3.86 | 140.00-233.00 | 2.10 | 7 | 224.57 | 5.86 | 169.00-281.00 | 2.62 |
| PG: OCD Vitros 250,350,400,500,750,950 | 10 | 221.20 | 14.72 | 166.00-276.00 | 6.65 | 10 | 88.80 | 6.25 | 67.00-111.00 | 7.08 | 10 | 120.80 | 9.10 | 91.00-151.00 | 7.52 | 10 | 204.80 | 13.67 | 154.00-256.00 | 6.68 | 10 | 259.10 | 15.71 | 194.00-324.00 | 6.06 |
| SG: Select Beckman Coulter Instruments (Chem) | 16 | 201.44 | 5.29 | 151.00-251.00 | 2.64 | 16 | 86.75 | 2.18 | 65.00-109.00 | 2.53 | 16 | 115.38 | 2.85 | 86.00-144.00 | 2.43 | 17 | 187.29 | 4.84 | 140.00-234.00 | 2.57 | 17 | 229.71 | 6.20 | 173.00-288.00 | 2.70 |
| SG: Select Roche Instruments / Roche Reagent | 8 | 198.26 | 5.23 | 149.00-248.00 | 2.63 | 8 | 89.58 | 1.81 | 68.00-113.00 | 2.00 | 8 | 118.11 | 2.95 | 89.00-148.00 | 2.54 | 8 | 186.60 | 4.66 | 140.00-234.00 | 2.51 | 8 | 225.50 | 6.03 | 170.00-283.00 | 2.65 |
| SG: Select Roche Instruments / Roche Reagents Groups | 9 | 197.89 | 5.01 | 149.00-248.00 | 2.53 | 9 | 89.41 | 1.78 | 67.00-111.00 | 2.02 | 9 | 118.10 | 2.76 | 89.00-148.00 | 2.37 | 9 | 186.20 | 4.52 | 140.00-233.00 | 2.42 | 9 | 224.89 | 5.93 | 169.00-281.00 | 2.62 |
| SG: Select OCD Vitros Instruments | 11 | 220.55 | 14.13 | 166.00-276.00 | 6.38 | 11 | 88.45 | 6.04 | 66.00-110.00 | 6.82 | 11 | 120.18 | 8.88 | 90.00-150.00 | 7.42 | 11 | 204.09 | 13.18 | 153.00-255.00 | 6.47 | 11 | 258.18 | 15.21 | 194.00-323.00 | 5.89 |
| SG: Abaxis Piccolo Instruments | 14 | 217.50 | 3.44 | 164.00-273.00 | 1.56 | 6 | 93.33 | 0.82 | 70.00-116.00 | 0.86 | 6 | 124.33 | 2.73 | 93.00-155.00 | 2.18 | 6 | 202.33 | 3.56 | 152.00-253.00 | 1.78 | 6 | 249.67 | 4.59 | 188.00-313.00 | 1.84 |
| SG: Select Abbott Architect Instruments | 7 | 198.75 | 3.20 | 149.00-249.00 | 1.61 | 7 | 84.13 | 1.46 | 63.00-105.00 | 1.79 | 7 | 112.63 | 1.41 | 85.00-141.00 | 1.24 | 7 | 183.50 | 2.33 | 138.00-230.00 | 1.25 | 7 | 226.88 | 2.90 | 170.00-284.00 | 1.28 |
| SG: International OCD Vitros Instruments | 13 | 221.23 | 12.75 | 166.00-276.00 | 5.75 | 13 | 88.85 | 5.46 | 67.00-111.00 | 6.18 | 13 | 121.08 | 7.91 | 91.00-151.00 | 6.53 | 13 | 205.00 | 11.91 | 154.00-256.00 | 5.80 | 13 | 259.31 | 13.69 | 194.00-324.00 | 5.29 |
| SG: International Abaxis Piccolo Instruments | 15 | 217.47 | 3.31 | 163.00-271.00 | 1.52 | 7 | 93.29 | 0.76 | 70.00-116.00 | 0.86 | 7 | 124.29 | 2.50 | 93.00-155.00 | 2.02 | 7 | 202.29 | 3.25 | 152.00-253.00 | 1.63 | 7 | 249.71 | 4.19 | 188.00-313.00 | 1.68 |
| AG: All Group | 118 | 202.43 | 10.69 | 152.00-253.00 | 5.30 | 110 | 87.10 | 5.88 | 65.00-109.00 | 6.78 | 110 | 116.12 | 6.45 | 87.00-145.00 | 5.52 | 111 | 187.78 | 9.50 | 141.00-235.00 | 5.05 | 110 | 231.45 | 13.25 | 173.00-289.00 | 5.76 |
| Amylase | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% |
| PG: Abaxis Piccolo (Waived) | 45 | 142.87 | 3.92 | 100.00-186.00 | 2.73 | | | | | | | | | | | | | | | | | | | | |
| PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent | 12 | 162.75 | 7.63 | 114.00-212.00 | 4.66 | 12 | 17.00 | 1.21 | 12.00-22.00 | 7.06 | 12 | 53.33 | 2.15 | 37.00-69.00 | 3.96 | 12 | 145.83 | 4.71 | 102.00-190.00 | 3.22 | 12 | 202.83 | 6.90 | 142.00-264.00 | 3.40 |
| PG: Siemens Dimension/AR/ES/RxL/Xpand | 7 | 177.09 | 1.64 | 124.00-230.00 | 0.90 | 7 | 15.73 | 1.19 | 11.00-21.00 | 7.50 | 7 | 56.36 | 0.92 | 39.00-73.00 | 1.61 | 7 | 157.45 | 1.21 | 110.00-204.00 | 0.76 | 7 | 218.00 | 3.26 | 153.00-283.00 | 1.51 |
| PG: Beckman (Olympus) AU | 7 | 125.86 | 3.72 | 88.00-164.00 | 2.94 | 7 | 11.43 | 0.54 | 8.00-14.00 | 4.55 | 7 | 41.14 | 1.57 | 29.00-53.00 | 3.90 | 8 | 112.38 | 4.69 | 78.00-146.00 | 4.20 | 8 | 154.50 | 6.09 | 109.00-202.00 | 3.94 |
| SG: Select Beckman Coulter Instruments (Chem) | 8 | 130.88 | 14.60 | 92.00-170.00 | 11.15 | 7 | 11.43 | 0.54 | 8.00-14.00 | 4.55 | 8 | 42.88 | 5.11 | 30.00-56.00 | 11.86 | 9 | 116.44 | 12.97 | 81.00-151.00 | 11.21 | 9 | 160.00 | 17.46 | 112.00-208.00 | 10.94 |
| SG: Abaxis Piccolo Instruments | 48 | 142.94 | 3.81 | 100.00-186.00 | 2.66 | | | | | | | | | | | | | | | | | | | | |
| SG: International OCD Vitros Instruments | 5 | 98.40 | 6.15 | 69.00-127.00 | 6.22 | 5 | 30.00 | 0.00 | 21.00-39.00 | 0.00 | 5 | 38.40 | 1.52 | 27.00-49.00 | 3.95 | 5 | 90.20 | 7.26 | 63.00-117.00 | 8.11 | 5 | 120.20 | 8.84 | 84.00-156.00 | 7.33 |
| SG: Select Gen Chem Instruments | 13 | 148.40 | 22.15 | 104.00-192.00 | 14.93 | 12 | 14.43 | 3.25 | 10.00-18.00 | 23.57 | 13 | 48.73 | 7.47 | 34.00-64.00 | 15.31 | 14 | 131.75 | 20.39 | 92.00-172.00 | 15.45 | 14 | 180.50 | 27.64 | 127.00-235.00 | 15.25 |
| AG: All Group | 86 | 148.82 | 18.40 | 104.00-194.00 | 12.35 | 36 | 15.90 | 2.43 | 11.00-21.00 | 15.00 | 40 | 51.30 | 6.22 | 36.00-66.00 | 12.16 | 42 | 138.06 | 20.71 | 97.00-179.00 | 15.00 | 39 | 191.82 | 26.68 | 134.00-250.00 | 13.91 |
| Creatine Kinase | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% |
| PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent | 9 | 281.89 | 10.47 | 197.00-367.00 | 3.72 | 9 | 25.44 | 1.74 | 18.00-33.00 | 6.80 | 9 | 92.22 | 4.32 | 64.00-120.00 | 4.67 | | | | | | 9 | 350.89 | 15.31 | 246.00-456.00 | 4.36 |
| PG: Siemens Dimension/AR/ES/RxL/Xpand / IFCC | 6 | 305.00 | 12.15 | 214.00-397.00 | 3.97 | 5 | 17.13 | 2.29 | 12.00-22.00 | 13.53 | 5 | 92.78 | 4.71 | 65.00-121.00 | 5.05 | 6 | 270.90 | 9.67 | 190.00-352.00 | 3.58 | 6 | 378.00 | 19.43 | 265.00-491.00 | 5.13 |
| PG: Beckman (Olympus) AU | 6 | 268.83 | 12.22 | 188.00-350.00 | 4.54 | 6 | 19.00 | 1.67 | 13.00-25.00 | 8.95 | 6 | 81.50 | 2.74 | 57.00-107.00 | 3.29 | 7 | 242.00 | 10.79 | 169.00-315.00 | 4.46 | 8 | 325.50 | 47.01 | 228.00-424.00 | 14.42 |
| PG: Roche Cobas Integra / Roche Reagent | 5 | 306.61 | 9.66 | 215.00-399.00 | 3.16 | 5 | 22.90 | 1.52 | 16.00-30.00 | 6.52 | 5 | 96.58 | 3.21 | 68.00-126.00 | 3.30 | 5 | 274.91 | 9.30 | 193.00-358.00 | 3.38 | 5 | 382.43 | 15.03 | 267.00-497.00 | 3.93 |
| SG: Select Beckman Coulter Instruments (Chem) | 7 | 272.71 | 15.16 | 191.00-355.00 | 5.57 | 7 | 19.57 | 2.15 | 14.00-26.00 | 10.50 | 7 | 83.29 | 5.34 | 58.00-108.00 | 6.39 | 8 | 244.75 | 12.66 | 172.00-319.00 | 5.18 | 9 | 329.44 | 45.54 | 230.00-428.00 | 13.83 |
| SG: Select Roche Instruments / Roche Reagent | 6 | 306.85 | 8.65 | 215.00-399.00 | 2.83 | 6 | 22.41 | 1.80 | 15.00-29.00 | 8.18 | 6 | 95.82 | 3.42 | 67.00-125.00 | 3.54 | 6 | 274.43 | 8.40 | 192.00-356.00 | 3.07 | 6 | 382.20 | 13.46 | 267.00-497.00 | 3.53 |
| SG: Abaxis Piccolo Instruments | 37 | 309.27 | 14.21 | 216.00-402.00 | 4.60 | | | | | | | | | | | | | | | | | | | | |
| AG: All Group | 77 | 299.52 | 21.33 | 210.00-390.00 | 7.10 | 38 | 21.10 | 3.42 | 15.00-27.00 | 16.19 | 40 | 91.64 | 6.53 | 64.00-120.00 | 7.07 | 41 | 260.64 | 18.87 | 183.00-339.00 | 7.24 | 43 | 358.39 | 38.12 | 251.00-465.00 | 10.64 |
| GGT | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% |
| PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent | 6 | 130.33 | 2.58 | 122.00-138.00 | 2.00 | 6 | 16.50 | 2.43 | 10.00-24.00 | 14.12 | 6 | 45.33 | 1.86 | 39.00-51.00 | 4.22 | 6 | 116.67 | 2.66 | 109.00-125.00 | 2.31 | 6 | 160.50 | 3.21 | 151.00-171.00 | 1.99 |
| PG: Siemens Dimension/AR/ES/RxL/Xpand | 6 | 208.50 | 6.09 | 191.00-227.00 | 2.92 | 6 | 26.50 | 0.84 | 25.00-29.00 | 2.96 | 6 | 72.50 | 3.15 | 64.00-82.00 | 4.25 | 6 | 186.33 | 5.39 | 170.00-202.00 | 2.90 | 6 | 254.50 | 6.32 | 236.00-274.00 | 2.47 |
| PG: Beckman (Olympus) AU | 5 | 145.40 | 6.54 | 126.00-165.00 | 4.48 | 5 | 14.80 | 0.84 | 13.00-17.00 | 5.33 | 5 | 47.80 | 2.59 | 40.00-56.00 | 5.42 | 6 | 129.17 | 6.08 | 111.00-147.00 | 4.73 | 6 | 177.33 | 7.26 | 155.00-199.00 | 4.12 |
| SG: Select Beckman Coulter Instruments (Chem) | 6 | 153.17 | 19.90 | 93.00-213.00 | 13.01 | 6 | 15.83 | 2.64 | 8.00-24.00 | 16.25 | 6 | 50.33 | 6.62 | 30.00-70.00 | 13.20 | 7 | 135.29 | 17.11 | 84.00-186.00 | 12.67 | 7 | 185.71 | 23.14 | 117.00-255.00 | 12.42 |
| SG: Abaxis Piccolo Instruments | 49 | 164.12 | 4.58 | 150.00-178.00 | 2.80 | | | | | | | | | | | | | | | | | | | | |
| AG: All Group | 69 | 164.87 | 18.65 | 109.00-221.00 | 11.33 | 26 | 19.96 | 5.24 | 4.00-36.00 | 26.00 | 23 | 56.25 | 11.46 | 22.00-91.00 | 20.54 | 24 | 147.12 | 27.96 | 63.00-231.00 | 19.05 | 24 | 201.64 | 37.44 | 90.00 | |



| Lactic Acid | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
|--|-------------|-------------|-----------|---------------|-------------|-------------|-------------|-----------|---------------|-------------|-------------|-------------|-----------|---------------|-------------|-------------|-------------|-----------|---------------|-------------|-------------|-------------|-----------|---------------|-------------|
| AG: All Group | 5 | 6.67 | 0.43 | 5.40-8.00 | 6.42 | 5 | 1.30 | 0.11 | 1.00-1.60 | 8.46 | 5 | 2.52 | 0.04 | 2.40-2.60 | 1.60 | 5 | 5.44 | 0.32 | 4.40-6.40 | 5.93 | 5 | 7.48 | 0.92 | 4.80-10.20 | 12.13 |
| Magnesium | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent | 12 | 4.01 | 0.35 | 3.00-5.00 | 8.75 | 12 | 1.24 | 0.18 | 0.90-1.50 | 15.00 | 11 | 1.85 | 0.21 | 1.40-2.40 | 11.05 | 12 | 3.68 | 0.37 | 2.80-4.60 | 10.00 | 12 | 4.56 | 0.24 | 3.40-5.80 | 5.22 |
| PG: Siemens Dimension/AR/ES/RxL/Xpand | 7 | 4.36 | 0.07 | 3.30-5.50 | 1.59 | 7 | 1.32 | 0.04 | 1.00-1.60 | 3.08 | 7 | 2.05 | 0.07 | 1.50-2.50 | 3.50 | 7 | 3.93 | 0.10 | 2.90-4.90 | 2.56 | 7 | 5.12 | 0.09 | 3.80-6.40 | 1.76 |
| PG: Beckman (Olympus) AU | 9 | 4.20 | 0.09 | 3.20-5.30 | 2.14 | 9 | 1.36 | 0.05 | 1.00-1.80 | 3.57 | 9 | 2.06 | 0.05 | 1.60-2.60 | 2.38 | 10 | 3.87 | 0.07 | 2.90-4.90 | 1.79 | 10 | 4.94 | 0.10 | 3.70-6.10 | 2.04 |
| SG: Select Beckman Coulter Instruments (Chem) | 10 | 4.19 | 0.09 | 3.20-5.30 | 2.14 | 10 | 1.35 | 0.05 | 1.00-1.80 | 3.57 | 10 | 2.05 | 0.05 | 1.60-2.60 | 2.38 | 11 | 3.86 | 0.07 | 2.90-4.90 | 1.79 | 11 | 4.93 | 0.10 | 3.70-6.10 | 2.04 |
| SG: Select Gen Chem Instruments | 16 | 4.16 | 0.18 | 3.20-5.30 | 4.29 | 16 | 1.31 | 0.12 | 1.00-1.60 | 9.23 | 17 | 1.95 | 0.24 | 1.40-2.40 | 12.63 | 18 | 3.79 | 0.18 | 2.80-4.80 | 4.74 | 18 | 4.88 | 0.20 | 3.70-6.10 | 4.08 |
| AG: All Group | 41 | 4.17 | 0.25 | 3.20-5.30 | 5.95 | 41 | 1.30 | 0.12 | 1.00-1.60 | 9.23 | 41 | 1.96 | 0.20 | 1.50-2.50 | 10.00 | 43 | 3.80 | 0.24 | 2.80-4.80 | 6.32 | 43 | 4.86 | 0.27 | 3.70-6.10 | 5.51 |
| Uric Acid | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent | 19 | 10.38 | 0.88 | 8.60-12.20 | 8.46 | 19 | 2.74 | 0.21 | 2.20-3.20 | 7.78 | 19 | 4.52 | 0.20 | 3.70-5.30 | 4.44 | 19 | 9.52 | 0.47 | 7.90-11.10 | 4.95 | 19 | 12.94 | 0.67 | 10.70-15.10 | 5.19 |
| PG: Siemens Dimension/AR/ES/RxL/Xpand | 14 | 10.14 | 0.47 | 8.40-11.80 | 4.65 | 14 | 1.79 | 0.10 | 1.50-2.10 | 5.56 | 14 | 3.89 | 0.13 | 3.20-4.60 | 3.33 | 14 | 9.15 | 0.36 | 7.60-10.80 | 3.91 | 14 | 12.40 | 0.36 | 10.30-14.50 | 2.90 |
| PG: Beckman (Olympus) AU | 11 | 10.85 | 0.46 | 9.00-12.80 | 4.22 | 11 | 1.85 | 0.13 | 1.60-2.20 | 6.84 | 11 | 4.13 | 0.17 | 3.40-4.80 | 4.15 | 12 | 9.79 | 0.34 | 8.10-11.50 | 3.47 | 12 | 13.19 | 0.52 | 11.00-15.40 | 3.94 |
| PG: Roche Cobas Integra / Roche Reagent | 5 | 10.68 | 0.36 | 8.90-12.50 | 3.36 | 5 | 1.74 | 0.06 | 1.40-2.00 | 2.94 | 5 | 3.98 | 0.11 | 3.30-4.70 | 2.75 | 5 | 9.54 | 0.26 | 7.90-11.10 | 2.74 | 5 | 12.80 | 0.29 | 10.60-15.00 | 2.27 |
| PG: OCD Vitros 250,350,400,500,750,950 | 8 | 10.53 | 0.20 | 8.70-12.30 | 1.90 | 8 | 1.70 | 0.05 | 1.40-2.00 | 2.94 | 8 | 3.94 | 0.09 | 3.20-4.60 | 2.31 | 8 | 9.33 | 0.16 | 7.70-10.90 | 1.72 | 8 | 12.78 | 0.27 | 10.60-15.00 | 2.11 |
| SG: Select Beckman Coulter Instruments (Chem) | 12 | 10.78 | 0.50 | 9.00-12.60 | 4.63 | 12 | 1.85 | 0.12 | 1.60-2.20 | 6.32 | 12 | 4.12 | 0.16 | 3.40-4.80 | 3.90 | 13 | 9.74 | 0.38 | 8.10-11.30 | 3.92 | 13 | 13.09 | 0.62 | 10.90-15.30 | 4.73 |
| SG: Select Roche Instruments / Roche Reagent | 6 | 10.65 | 0.33 | 8.90-12.50 | 3.08 | 6 | 1.75 | 0.06 | 1.50-2.10 | 2.78 | 6 | 4.00 | 0.11 | 3.30-4.70 | 2.75 | 6 | 9.55 | 0.23 | 8.00-11.20 | 2.40 | 6 | 12.80 | 0.26 | 10.60-15.00 | 2.03 |
| SG: Select OCD Vitros Instruments | 9 | 10.54 | 0.19 | 8.70-12.30 | 1.81 | 9 | 1.73 | 0.11 | 1.40-2.00 | 6.47 | 9 | 3.96 | 0.10 | 3.30-4.70 | 2.50 | 9 | 9.34 | 0.16 | 7.70-10.90 | 1.72 | 9 | 12.78 | 0.25 | 10.60-15.00 | 1.95 |
| SG: Abaxis Piccolo Instruments | 8 | 9.93 | 0.27 | 8.20-11.60 | 2.73 | | | | | | | | | | | | | | | | | | | | |
| SG: Select SpotChem Instruments | 9 | 10.57 | 0.22 | 8.80-12.40 | 2.08 | 9 | 1.72 | 0.08 | 1.40-2.00 | 4.71 | 9 | 3.98 | 0.15 | 3.30-4.70 | 3.75 | 9 | 9.37 | 0.19 | 7.80-11.00 | 2.02 | 9 | 12.71 | 0.32 | 10.50-14.90 | 2.52 |
| SG: Select Abbott Architect Instruments | 6 | 11.39 | 0.20 | 9.50-13.30 | 1.75 | 6 | 1.97 | 0.10 | 1.70-2.30 | 5.00 | 6 | 4.34 | 0.11 | 3.60-5.00 | 2.56 | 6 | 10.23 | 0.20 | 8.50-11.90 | 1.96 | 6 | 13.87 | 0.21 | 11.50-16.30 | 1.51 |
| SG: International OCD Vitros Instruments | 11 | 10.54 | 0.17 | 8.70-12.30 | 1.62 | 11 | 1.70 | 0.06 | 1.40-2.00 | 3.53 | 11 | 3.96 | 0.09 | 3.30-4.70 | 2.25 | 11 | 9.37 | 0.16 | 7.80-11.00 | 1.70 | 11 | 12.82 | 0.25 | 10.60-15.00 | 1.95 |
| SG: International Abaxis Piccolo Instruments | 9 | 9.87 | 0.31 | 8.20-11.60 | 3.13 | | | | | | | | | | | | | | | | | | | | |
| SG: Select Gen Chem Instruments | 25 | 10.93 | 0.46 | 9.00-12.80 | 4.22 | 25 | 1.83 | 0.16 | 1.50-2.10 | 8.89 | 25 | 4.11 | 0.21 | 3.40-4.80 | 5.12 | 26 | 9.83 | 0.39 | 8.10-11.50 | 3.98 | 25 | 13.23 | 0.63 | 11.00-15.40 | 4.70 |
| AG: All Group | 82 | 10.50 | 0.63 | 8.70-12.30 | 6.00 | 65 | 1.93 | 0.32 | 1.60-2.20 | 16.84 | 75 | 4.12 | 0.32 | 3.40-4.80 | 7.80 | 76 | 9.52 | 0.46 | 7.90-11.10 | 4.84 | 75 | 12.87 | 0.62 | 10.70-15.10 | 4.81 |
| Iron | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent | 9 | 173.44 | 8.38 | 138.00-208.00 | 4.86 | 9 | 30.11 | 3.44 | 24.00-36.00 | 11.33 | 9 | 65.56 | 2.79 | 53.00-79.00 | 4.24 | 9 | 157.67 | 4.53 | 126.00-190.00 | 2.85 | 9 | 211.22 | 7.51 | 169.00-253.00 | 3.55 |
| PG: Siemens Dimension/AR/ES/RxL/Xpand | 7 | 185.10 | 1.91 | 148.00-222.00 | 1.03 | 7 | 33.20 | 0.79 | 26.00-40.00 | 2.42 | 7 | 71.00 | 0.47 | 57.00-85.00 | 0.70 | 7 | 166.20 | 1.14 | 133.00-199.00 | 0.66 | 7 | 224.20 | 2.10 | 179.00-269.00 | 0.94 |
| PG: Beckman (Olympus) AU | 8 | 212.25 | 3.20 | 170.00-254.00 | 1.51 | 8 | 35.88 | 3.44 | 29.00-43.00 | 9.44 | 8 | 78.63 | 3.29 | 63.00-95.00 | 4.18 | 9 | 188.67 | 3.74 | 151.00-227.00 | 1.96 | 9 | 256.44 | 4.19 | 205.00-307.00 | 1.64 |
| SG: Select Beckman Coulter Instruments (Chem) | 9 | 209.89 | 7.69 | 168.00-252.00 | 3.67 | 9 | 35.33 | 3.61 | 28.00-42.00 | 10.29 | 9 | 77.67 | 4.21 | 62.00-94.00 | 5.38 | 10 | 187.10 | 6.08 | 150.00-224.00 | 3.26 | 10 | 254.00 | 8.68 | 203.00-305.00 | 3.43 |
| AG: All Group | 38 | 187.56 | 15.82 | 150.00-226.00 | 8.40 | 38 | 32.47 | 3.91 | 26.00-38.00 | 12.19 | 38 | 70.30 | 5.62 | 56.00-84.00 | 8.00 | 39 | 168.48 | 13.51 | 134.00-202.00 | 8.04 | 39 | 227.50 | 19.92 | 182.00-274.00 | 8.73 |
| TIBC | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent | 5 | 347.60 | 41.75 | 222.40-472.90 | 12.01 | 5 | 153.25 | 10.56 | 121.60-185.00 | 6.89 | 5 | 193.60 | 37.70 | 80.50-306.70 | 19.47 | 5 | 320.40 | 42.20 | 193.80-447.00 | 13.17 | 5 | 394.40 | 43.54 | 263.80-525.00 | 11.04 |
| PG: Siemens Dimension/AR/ES/RxL/Xpand | 6 | 321.67 | 11.26 | 287.90-355.50 | 3.50 | 6 | 75.44 | 9.82 | 45.90-104.90 | 13.02 | 6 | 138.89 | 6.17 | 120.40-157.40 | 4.44 | 6 | 291.89 | 6.90 | 271.20-312.60 | 2.36 | 6 | 382.44 | 8.28 | 357.60-407.20 | 2.17 |
| PG: Beckman (Olympus) AU | 7 | 408.57 | 78.60 | 172.80-644.40 | 19.24 | 7 | 111.57 | 15.15 | 66.10-157.10 | 13.58 | 7 | 187.29 | 26.06 | 109.10-265.50 | 13.91 | 6 | 395.33 | 13.29 | 355.40-435.20 | 3.36 | 7 | 480.57 | 93.93 | 198.80-762.40 | 19.54 |
| AG: All Group | 22 | 348.12 | 71.70 | 133.00-563.20 | 20.60 | 22 | 105.00 | 31.15 | 11.60-198.40 | 29.66 | 23 | 166.99 | 37.10 | 55.70-278.30 | 22.22 | 22 | 328.62 | 71.47 | 114.20-543.00 | 21.75 | 23 | 419.58 | 104.31 | 106.70-732.50 | 24.86 |
| TSH | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| PG: Abbott Architect | 8 | 7.98 | 0.47 | 6.59-9.38 | 5.83 | 8 | 1.54 | 0.12 | 1.18-1.90 | 7.79 | 8 | 3.34 | 0.22 | 2.68-4.00 | 6.56 | 8 | 7.31 | 0.45 | 5.97-8.65 | 6.13 | 8 | 9.60 | 0.56 | 7.91-11.29 | 5.88 |
| PG: Beckman ACCESS | 23 | 7.65 | 0.38 | 6.51-8.79 | 4.98 | 23 | 1.48 | 0.11 | 1.15-1.81 | 7.43 | 23 | 3.15 | 0.23 | 2.47-3.83 | 7.21 | 22 | 6.95 | 0.41 | 5.71-8.19 | 5.96 | 23 | 8.74 | 0.65 | 6.78-10.70 | 7.46 |
| PG: Siemens Dimension/AR/ES/RxL/Xpand | 14 | 5.44 | 0.55 | 3.78-7.10 | 10.17 | 15 | 1.17 | 0.11 | 0.84-1.50 | 9.49 | 15 | 2.32 | 0.21 | 1.70-2.94 | 8.88 | 15 | 4.96 | 0.45 | 3.61-6.31 | 9.07 | 15 | 6.42 | 0.60 | 4.61-8.23 | 9.39 |
| PG: Qualigen FastPak Thyroid | 27 | 7.30 | 0.65 | 5.36-9.24 | 8.86 | 25 | 1.28 | 0.12 | 0.92-1.64 | 9.45 | 26 | 3.00 | 0.28 | 2.17-3.83 | 9.20 | 26 | 6.64 | 0.59 | 4.88-8.40 | 8.84 | 27 | 8.43 | 0.66 | 6.44-10.42 | 7.88 |
| PG: TOSOH AIA | 25 | 11.51 | 0.59 | 9.74-13.28 | 5.12 | 25 | 2.23 | 0.13 | 1.83-2.63 | 5.96 | 25 | 4.78 | 0.27 | 3.96-5.60 | 5.71 | 25 | 10.33 | 0.63 | 8.44-12.22 | 6.09 | 25 | 13.26 | 0.55 | 11.60-14.92 | 4.16 |
| SG: Select Beckman Coulter Instruments (Chem) | 25 | 7.60 | 0.41 | 6.38-8.82 | 5.36 | 25 | 1.47 | 0.12 | 1.12-1.82 | 7.89 | 25 | 3.13 | 0.23 | 2.43-3.83 | 7.44 | 25 | 6.89 | 0.45 | 5.54-8.24 | 6.52 | 26 | 8.70 | 0.63 | 6.81-10.59 | 7.24 |
| SG: Select Roche Instruments | 6 | 8.33 | 0.48 | 6.89-9.77 | 5.76 | 6 | 1.90 | 0.07 | 1.67-2.13 | 3.95 | 6 | 3.85 | 0.14 | 3.44-4.26 | 3.53 | 6 | 8.01 | 1.13 | 4.62-11.40 | 14.09 | 6 | 9.12 | 0.71 | 6.99-11.25 | 7.79 |
| PG: Abbott Architect i1000 SR | 10 | 7.94 | 0.34 | 6.92-8.96 | 4.28 | 10 | 1.53 | 0.07 | 1.33-1.73 | 4.38 | 10 | 3.30 | 0.15 | 2.85-3.75 | 4.58 | 10 | 7.10 | 0.30 | 6.18-8.02 | 4.30 | 10 | 9.32 | 0.33 | 8.33-10.31 | 3.53 |
| SG: Select Abbott Architect Instruments | 20 | 7.95 | 0.38 | 6.82-9.08 | 4.73 | 20 | 1.54 | 0.09 | 1.27-1.81 | 5.91 | 20 | 3.33 | 0.19 | 2.76-3.90 | 5.68 | 20 | 7.21 | 0.38 | 6.07-8.35 | 5.26 | 20 | 9.44 | 0.48 | 7.99-10.89 | 5.11 |
| PG: NanoEnTek FrenD System | 5 | 6.24 | 0.74 | 4.01-8.47 | 11.92 | 5 | 1.26 | 0.48 | | | | | | | | | | | | | | | | | |



| PG: TOSOH AIA | 5 | 4.54 | 1.08 | 1.30-7.70 | 23.78 | 5 | 1.91 | 0.01 | 1.90-1.90 | 0.53 | 5 | 2.62 | 0.03 | 2.50-2.70 | 1.15 | 5 | 3.90 | 0.15 | 3.50-4.30 | 3.85 | 5 | 4.71 | 0.41 | 3.50-5.90 | 8.72 |
|--|-------|--------|-------|---------------|--------|-------|-------|------|--------------|-------|------|--------|-------|---------------|-------|------|-------|------|--------------|-------|-------|--------|-------|---------------|-------|
| SG: Select Beckman Coulter Instruments (Chem) | 6 | 2.08 | 0.23 | 1.40-2.80 | 10.95 | 6 | 0.87 | 0.04 | 0.80-1.00 | 4.44 | 6 | 1.17 | 0.08 | 1.00-1.40 | 6.67 | 6 | 1.89 | 0.17 | 1.40-2.40 | 8.95 | 6 | 2.43 | 0.23 | 1.70-3.10 | 9.58 |
| AG: All Group | 12 | 1.71 | 0.48 | 0.30-3.10 | 28.24 | 12 | 0.84 | 0.18 | 0.30-1.30 | 22.50 | 12 | 1.11 | 0.22 | 0.40-1.80 | 20.00 | 12 | 1.62 | 0.37 | 0.50-2.70 | 23.13 | 12 | 1.94 | 0.64 | 0.00-3.80 | 33.68 |
| T3 Uptake | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| AG: All Group | 10 | 0.48 | 0.05 | 0.33-0.63 | 10.58 | 10 | 0.53 | 0.09 | 0.27-0.79 | 16.33 | 10 | 0.52 | 0.08 | 0.29-0.75 | 14.63 | 10 | 0.47 | 0.05 | 0.31-0.63 | 11.14 | 10 | 0.46 | 0.04 | 0.34-0.58 | 8.70 |
| Free T3 | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| PG: Beckman ACCESS | 7 | 4.13 | 0.16 | 3.60-4.60 | 3.90 | 7 | 2.69 | 0.19 | 2.10-3.30 | 7.04 | 7 | 3.20 | 0.14 | 2.80-3.60 | 4.38 | 7 | 4.04 | 0.16 | 3.50-4.50 | 4.00 | 7 | 4.35 | 0.19 | 3.80-5.00 | 4.32 |
| SG: Select Beckman Coulter Instruments (Chem) | 9 | 4.12 | 0.15 | 3.60-4.50 | 3.66 | 9 | 2.72 | 0.19 | 2.10-3.30 | 7.04 | 9 | 3.25 | 0.17 | 2.80-3.80 | 5.15 | 10 | 4.04 | 0.13 | 3.60-4.40 | 3.25 | 10 | 4.37 | 0.18 | 3.90-4.90 | 4.09 |
| SG: International OCD Vitros Instruments | 5 | 16.67 | 1.48 | 12.30-21.10 | 8.86 | 5 | 10.06 | 1.26 | 6.30-13.90 | 12.48 | 5 | 12.57 | 1.17 | 9.10-16.10 | 9.29 | 5 | 16.30 | 1.45 | 12.00-20.60 | 8.90 | 5 | 17.87 | 1.20 | 14.30-21.50 | 6.70 |
| AG: All Group | 14 | 4.18 | 0.26 | 3.40-5.00 | 6.19 | 14 | 3.01 | 0.60 | 1.20-4.80 | 20.00 | 14 | 3.58 | 0.74 | 1.40-5.80 | 20.56 | 15 | 4.11 | 0.22 | 3.40-4.80 | 5.37 | 15 | 4.46 | 0.23 | 3.80-5.20 | 5.11 |
| 746 - Auto Diff 6 (Abbott 5-Part Diff) | CL-6 | | | | | CL-7 | | | | | CL-8 | | | | | CL-9 | | | | | CL-10 | | | | |
| White Blood Cell Count | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| SG: Select Abbott Cell Dyn Instruments (5-part Diff) | 6 | 8.17 | 0.56 | 7.00-9.40 | 6.83 | 6 | 3.14 | 0.35 | 2.60-3.60 | 11.29 | 6 | 21.47 | 1.20 | 18.30-24.70 | 5.58 | 6 | 3.21 | 0.41 | 2.70-3.70 | 12.81 | 6 | 8.14 | 0.63 | 6.90-9.30 | 7.78 |
| AG: All Group | 7 | 8.16 | 0.52 | 7.00-9.40 | 6.34 | 7 | 3.15 | 0.33 | 2.70-3.70 | 10.31 | 7 | 21.63 | 1.19 | 18.40-24.80 | 5.51 | 7 | 3.21 | 0.38 | 2.70-3.70 | 11.88 | 7 | 8.15 | 0.59 | 7.00-9.40 | 7.20 |
| Lymphocyte | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| SG: Select Abbott Cell Dyn Instruments (5-part Diff) | 5 | 24.60 | 2.36 | 17.50-31.70 | 9.59 | 5 | 38.10 | 1.90 | 32.40-43.80 | 4.99 | 5 | 16.80 | 0.76 | 14.50-19.10 | 4.52 | 5 | 36.92 | 1.35 | 32.80-41.00 | 3.66 | 5 | 27.33 | 1.16 | 23.80-30.80 | 4.25 |
| AG: All Group | 6 | 24.61 | 2.16 | 18.10-31.10 | 8.78 | 6 | 38.86 | 2.65 | 30.90-46.80 | 6.81 | 6 | 17.96 | 3.14 | 8.60-27.40 | 17.44 | 6 | 37.56 | 2.10 | 31.30-43.90 | 5.59 | 6 | 27.47 | 1.12 | 24.10-30.90 | 4.07 |
| Monocyte | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| SG: Select Abbott Cell Dyn Instruments (5-part Diff) | 5 | 1.96 | 1.29 | 0.00-5.82 | 65.71 | 5 | 6.66 | 0.98 | 3.72-9.60 | 14.73 | 5 | 2.94 | 0.11 | 2.61-3.27 | 3.71 | 5 | 5.98 | 1.11 | 2.66-9.30 | 18.51 | 5 | 3.13 | 0.93 | 0.34-5.92 | 29.71 |
| AG: All Group | 6 | 2.15 | 1.28 | 0.00-5.99 | 59.53 | 6 | 6.78 | 0.95 | 3.93-9.63 | 14.01 | 6 | 3.69 | 1.99 | 0.00-9.66 | 53.90 | 6 | 5.83 | 1.09 | 2.56-9.10 | 18.71 | 6 | 2.94 | 0.99 | 0.00-5.90 | 33.54 |
| Granulocyte | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| SG: Select Abbott Cell Dyn Instruments (5-part Diff) | 6 | 65.24 | 9.21 | 37.60-92.80 | 14.13 | 6 | 50.44 | 1.56 | 45.70-55.10 | 3.10 | 6 | 74.83 | 2.22 | 68.10-81.50 | 2.97 | 6 | 52.44 | 2.54 | 44.80-60.00 | 4.85 | 6 | 64.47 | 1.71 | 59.40-69.60 | 2.65 |
| AG: All Group | 7 | 65.71 | 8.63 | 39.80-91.60 | 13.14 | 7 | 49.94 | 2.04 | 43.80-56.00 | 4.07 | 7 | 73.49 | 4.32 | 60.60-86.40 | 5.86 | 7 | 52.26 | 2.40 | 45.10-59.50 | 4.59 | 7 | 64.72 | 1.74 | 59.50-69.90 | 2.69 |
| Eosinophil | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| SG: Select Abbott Cell Dyn Instruments (5-part Diff) | 5 | 4.67 | 0.31 | 3.80-5.60 | 6.60 | 5 | 3.78 | 0.27 | 3.00-4.60 | 7.11 | 6 | 4.09 | 0.50 | 2.60-5.60 | 12.20 | 6 | 3.33 | 0.34 | 2.30-4.30 | 10.30 | 6 | 4.66 | 0.65 | 2.80-6.70 | 13.83 |
| AG: All Group | 6 | 4.41 | 0.73 | 2.20-6.60 | 16.59 | 6 | 3.60 | 0.55 | 2.00-5.20 | 15.00 | 7 | 3.89 | 0.73 | 1.70-6.10 | 18.72 | 7 | 3.24 | 0.41 | 2.00-4.40 | 12.81 | 7 | 4.48 | 0.79 | 2.10-6.90 | 17.56 |
| Basophil | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| SG: Select Abbott Cell Dyn Instruments (5-part Diff) | 5 | 0.20 | 0.25 | 0.00-0.94 | 123.50 | 5 | 0.52 | 0.33 | 0.00-1.52 | 63.85 | 6 | 0.37 | 0.10 | 0.07-0.67 | 27.30 | 5 | 0.50 | 0.25 | 0.00-1.25 | 50.00 | 5 | 0.23 | 0.10 | 0.00-0.52 | 42.17 |
| AG: All Group | 5 | 0.18 | 0.23 | 0.00-0.86 | 125.00 | 6 | 0.47 | 0.33 | 0.00-1.45 | 69.15 | 7 | 0.36 | 0.10 | 0.07-0.65 | 26.94 | 6 | 0.44 | 0.27 | 0.00-1.26 | 62.05 | 5 | 0.23 | 0.09 | 0.00-0.49 | 38.26 |
| Red Blood Cell Count | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| SG: Select Abbott Cell Dyn Instruments (5-part Diff) | 6 | 4.47 | 0.08 | 4.20-4.74 | 1.72 | 6 | 2.11 | 0.04 | 1.98-2.24 | 2.09 | 6 | 5.10 | 0.13 | 4.79-5.41 | 2.57 | 6 | 2.11 | 0.06 | 1.98-2.24 | 3.03 | 6 | 4.47 | 0.09 | 4.20-4.74 | 1.95 |
| AG: All Group | 7 | 4.45 | 0.09 | 4.18-4.72 | 1.96 | 7 | 2.11 | 0.04 | 1.98-2.24 | 2.04 | 7 | 5.06 | 0.17 | 4.76-5.36 | 3.30 | 7 | 2.11 | 0.06 | 1.98-2.24 | 2.80 | 7 | 4.45 | 0.10 | 4.18-4.72 | 2.22 |
| Hemoglobin | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| SG: Select Abbott Cell Dyn Instruments (5-part Diff) | 6 | 13.26 | 0.92 | 12.40-14.20 | 6.92 | 5 | 5.58 | 0.22 | 5.20-6.00 | 3.93 | 6 | 16.33 | 0.74 | 15.20-17.40 | 4.54 | 5 | 5.57 | 0.15 | 5.20-6.00 | 2.68 | 6 | 13.20 | 0.74 | 12.30-14.10 | 5.61 |
| AG: All Group | 7 | 13.24 | 0.86 | 12.30-14.10 | 6.52 | 6 | 5.61 | 0.22 | 5.20-6.00 | 3.93 | 7 | 16.23 | 0.74 | 15.10-17.30 | 4.57 | 6 | 5.60 | 0.16 | 5.20-6.00 | 2.86 | 7 | 13.15 | 0.70 | 12.30-14.10 | 5.30 |
| Hematocrit | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| SG: Select Abbott Cell Dyn Instruments (5-part Diff) | 6 | 37.30 | 1.56 | 35.10-39.50 | 4.18 | 6 | 15.46 | 0.42 | 14.60-16.40 | 2.71 | 6 | 43.79 | 1.67 | 41.20-46.40 | 3.81 | 6 | 15.56 | 0.61 | 14.70-16.50 | 3.91 | 6 | 37.31 | 1.31 | 35.10-39.50 | 3.51 |
| PG: Siemens Advia 560 | 5 | 41.10 | 0.00 | 38.60-43.60 | 0.00 | 5 | 17.10 | 0.00 | 16.10-18.10 | 0.00 | 5 | 48.30 | 0.00 | 45.40-51.20 | 0.00 | 5 | 17.40 | 0.00 | 16.40-18.40 | 0.00 | 5 | 40.30 | 0.00 | 37.90-42.70 | 0.00 |
| MCV | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| SG: Select Abbott Cell Dyn Instruments (5-part Diff) | 6 | 83.54 | 3.97 | 71.60-95.40 | 4.75 | 6 | 73.34 | 2.51 | 65.80-80.80 | 3.42 | 6 | 86.03 | 4.34 | 73.00-99.00 | 5.05 | 6 | 73.76 | 2.55 | 66.10-81.50 | 3.46 | 6 | 83.40 | 3.43 | 73.10-93.70 | 4.11 |
| AG: All Group | 7 | 84.95 | 5.42 | 68.70-101.30 | 6.38 | 7 | 74.54 | 4.10 | 62.20-86.80 | 5.50 | 7 | 87.95 | 6.76 | 67.70-108.30 | 7.68 | 7 | 74.97 | 4.17 | 62.50-87.50 | 5.56 | 7 | 84.65 | 4.75 | 70.50-99.00 | 5.61 |
| RDW | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| SG: Select Abbott Cell Dyn Instruments (5-part Diff) | 6 | 12.61 | 1.05 | 9.40-15.80 | 8.33 | 6 | 14.73 | 1.33 | 10.70-18.70 | 9.05 | 6 | 10.86 | 1.35 | 6.80-15.00 | 12.39 | 6 | 14.76 | 1.35 | 10.80-18.90 | 9.12 | 6 | 12.57 | 1.22 | 9.00-16.20 | 9.60 |
| AG: All Group | 7 | 13.14 | 1.77 | 7.80-18.40 | 13.51 | 7 | 15.19 | 1.79 | 9.80-20.60 | 11.78 | 7 | 11.33 | 1.82 | 5.80-16.80 | 16.11 | 7 | 15.29 | 1.95 | 9.50-21.10 | 12.75 | 7 | 13.05 | 1.76 | 7.80-18.40 | 13.44 |
| Platelet Count | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| SG: Select Abbott Cell Dyn Instruments (5-part Diff) | 6 | 337.29 | 19.84 | 253.00-421.00 | 5.88 | 6 | 95.29 | 7.76 | 71.00-119.00 | 8.21 | 6 | 539.29 | 15.15 | 404.00-674.00 | 2.82 | 6 | 90.71 | 2.06 | 68.00-114.00 | 2.31 | 6 | 328.57 | 17.61 | 247.00-411.00 | 5.35 |
| AG: All Group | 7 | 330.50 | 26.57 | 248.00-414.00 | 8.04 | 7 | 94.25 | 7.76 | 71.00-118.00 | 8.30 | 7 | 527.50 | 36.17 | 396.00-660.00 | 6.86 | 7 | 89.00 | 5.21 | 67.00-111.00 | 5.84 | 7 | 321.63 | 25.53 | 242.00-403.00 | 7.92 |
| 683 - ASO/CRP:684 - CRP (Quant & Qual) | CR-2N | | | | | CR-4N | | | | | | | | | | | | | | | | | | | |
| CRP Quantitative | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| 600 - Cholestech LDX/CardioChek Pkg | CT-2 | | | | | | | | | | | | | | | | | | | | | | | | |
| CholesterolTotal (Waived) | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| PG: Cholestech LDX (Waived) | 136 | 207.27 | 19.51 | 149.00-266.00 | 9.42 | | | | | | | | | | | | | | | | | | | | |
| PG: Polymer CardioChek | 66 | 236.70 | 42.28 | 110.00-364.00 | 17.85 | | | | | | | | | | | | | | | | | | | | |
| SG: Select Chemistry Instr - Enzyme 3 (AST, GGT) | 136 | 207.27 | 19.51 | 149.00-266.00 | 9.42 | | | | | | | | | | | | | | | | | | | | |
| SG: Select Cholestech Instruments | 136 | 207.27 | 19.51 | 149.00-266.00 | 9.42 | | | | | | | | | | | | | | | | | | | | |
| Glucose (Waived) | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | L | | | | |



| PG: Abbott Cell-Dyn Emerald | 136 | 80.16 | 1.25 | 76.50-84.00 | 1.56 | 136 | 87.75 | 5.04 | 72.70-102.90 | 5.73 | 136 | 98.37 | 1.36 | 94.30-102.50 | 1.38 | 136 | 85.09 | 1.06 | 81.90-88.30 | 1.25 | 136 | 98.64 | 1.16 | 95.10-102.10 | 1.18 |
|--|------|-------|-------|-------------|-------|------|--------|-------|---------------|-------|------|--------|-------|---------------|-------|------|--------|-------|---------------|-------|------|--------|-------|---------------|-------|
| PG: CDS M Series | 144 | 72.22 | 1.25 | 68.40-76.00 | 1.75 | 145 | 79.76 | 5.65 | 62.80-96.80 | 7.08 | 145 | 91.92 | 1.88 | 86.30-97.50 | 2.03 | 145 | 77.35 | 1.58 | 72.60-82.00 | 2.03 | 145 | 92.68 | 2.38 | 85.60-99.80 | 2.57 |
| SG: CDS Hematology Instruments | 154 | 72.16 | 1.27 | 68.40-76.00 | 1.76 | 155 | 79.73 | 5.48 | 63.30-96.10 | 6.88 | 155 | 91.85 | 1.88 | 86.20-97.40 | 2.05 | 155 | 77.29 | 1.57 | 72.60-82.00 | 2.03 | 155 | 92.60 | 2.36 | 85.50-99.70 | 2.55 |
| SG: International Abbott Emerald Instruments | 138 | 80.11 | 1.32 | 76.10-84.10 | 1.65 | 138 | 87.31 | 6.97 | 66.40-108.20 | 7.98 | 138 | 98.33 | 1.42 | 94.00-102.60 | 1.44 | 138 | 85.05 | 1.14 | 81.60-88.40 | 1.34 | 138 | 98.60 | 1.23 | 94.90-102.30 | 1.25 |
| Monocyte/Mixed | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% |
| PG: Abbott Cell-Dyn 1700 | 17 | 14.36 | 1.85 | 8.80-20.00 | 12.85 | 17 | 10.20 | 0.83 | 7.70-12.70 | 8.14 | 17 | 6.07 | 0.61 | 4.20-8.00 | 10.16 | 17 | 10.76 | 0.96 | 7.90-13.70 | 8.89 | 17 | 6.04 | 0.57 | 4.30-7.70 | 9.50 |
| PG: Abbott Cell-Dyn 1800 | 33 | 17.68 | 1.41 | 13.50-21.90 | 7.97 | 33 | 11.72 | 1.17 | 8.20-15.20 | 10.00 | 33 | 7.05 | 0.82 | 4.50-9.50 | 11.71 | 33 | 13.11 | 1.01 | 10.10-16.10 | 7.71 | 33 | 6.90 | 0.37 | 5.80-8.00 | 5.36 |
| PG: CDS Medonic 620 | 9 | 9.67 | 1.39 | 5.50-13.90 | 14.33 | 10 | 10.13 | 1.34 | 6.10-14.10 | 13.27 | 10 | 7.77 | 1.07 | 4.60-11.00 | 13.72 | 10 | 11.96 | 1.25 | 8.30-15.80 | 10.42 | 10 | 7.95 | 1.22 | 4.30-11.70 | 15.25 |
| PG: Beckman Coulter AcT Diff/Diff 2 | 198 | 6.63 | 1.23 | 2.90-10.30 | 18.64 | 218 | 8.20 | 0.74 | 6.00-10.40 | 9.02 | 218 | 5.53 | 0.45 | 4.20-6.80 | 8.18 | 218 | 8.25 | 1.08 | 5.10-11.50 | 13.01 | 218 | 5.56 | 0.48 | 4.10-7.10 | 8.75 |
| PG: Horiba ABX Diag. Micros/45/60 | 73 | 16.83 | 2.25 | 10.10-23.60 | 13.39 | 74 | 9.01 | 0.76 | 6.80-11.30 | 8.33 | 73 | 3.98 | 0.25 | 3.30-4.80 | 6.25 | 74 | 10.56 | 1.12 | 7.20-14.00 | 10.57 | 73 | 4.03 | 0.25 | 3.30-4.70 | 6.00 |
| SG: Select Abbott Cell Dyn Instruments (3-part Diff) | 187 | 12.85 | 2.76 | 4.60-21.20 | 21.40 | 187 | 8.10 | 2.07 | 1.90-14.30 | 25.56 | 187 | 4.32 | 1.55 | 0.00-8.90 | 36.05 | 187 | 9.05 | 2.24 | 2.30-15.70 | 24.89 | 184 | 4.34 | 1.46 | 0.00-8.70 | 33.95 |
| SG: Select Beckman Coulter Instruments (3-part Diff) | 198 | 6.63 | 1.23 | 2.90-10.30 | 18.64 | 218 | 8.20 | 0.74 | 6.00-10.40 | 9.02 | 218 | 5.53 | 0.45 | 4.20-6.80 | 8.18 | 218 | 8.25 | 1.08 | 5.10-11.50 | 13.01 | 218 | 5.56 | 0.48 | 4.10-7.10 | 8.75 |
| PG: Drew Scientific D3 | 26 | 10.33 | 1.33 | 6.30-14.30 | 12.91 | 26 | 7.00 | 0.41 | 5.80-8.20 | 5.86 | 26 | 3.63 | 0.22 | 3.00-4.20 | 5.83 | 26 | 7.65 | 0.72 | 5.40-9.80 | 9.61 | 26 | 3.67 | 0.19 | 3.10-4.30 | 5.14 |
| PG: Abbott Cell-Dyn Emerald | 136 | 11.48 | 1.27 | 7.70-15.30 | 11.04 | 136 | 6.93 | 0.55 | 5.30-8.60 | 7.97 | 136 | 3.43 | 0.24 | 2.70-4.10 | 7.06 | 136 | 7.83 | 0.75 | 5.50-10.10 | 9.62 | 133 | 3.47 | 0.22 | 2.80-4.20 | 6.29 |
| PG: CDS M Series | 138 | 10.84 | 2.15 | 4.40-17.30 | 19.91 | 145 | 9.46 | 1.26 | 5.70-13.30 | 13.26 | 145 | 7.50 | 0.87 | 4.90-10.10 | 11.60 | 145 | 10.80 | 1.76 | 5.50-16.10 | 16.30 | 145 | 7.51 | 0.88 | 4.90-10.10 | 11.73 |
| SG: CDS Hematology Instruments | 147 | 10.77 | 2.13 | 4.40-17.20 | 19.72 | 155 | 9.51 | 1.27 | 5.70-13.30 | 13.47 | 155 | 7.52 | 0.88 | 4.90-10.10 | 11.73 | 155 | 10.88 | 1.75 | 5.70-16.10 | 16.06 | 155 | 7.54 | 0.91 | 4.80-10.20 | 12.13 |
| SG: International Abbott Emerald Instruments | 137 | 11.50 | 1.28 | 7.70-15.30 | 11.13 | 137 | 6.94 | 0.55 | 5.30-8.60 | 7.97 | 137 | 3.42 | 0.24 | 2.70-4.10 | 7.06 | 137 | 7.83 | 0.75 | 5.50-10.10 | 9.62 | 134 | 3.47 | 0.22 | 2.90-4.10 | 6.00 |
| Platelet Count | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% |
| PG: Abbott Cell-Dyn 1700 | 18 | 63.72 | 6.90 | 48.00-80.00 | 10.78 | 18 | 268.89 | 21.59 | 202.00-336.00 | 8.03 | 18 | 560.61 | 44.16 | 421.00-701.00 | 7.88 | 18 | 159.17 | 13.24 | 119.00-199.00 | 8.30 | 18 | 581.17 | 44.77 | 436.00-726.00 | 7.71 |
| PG: Abbott Cell-Dyn 1800 | 35 | 67.33 | 4.82 | 50.00-84.00 | 7.16 | 35 | 277.42 | 52.09 | 208.00-346.00 | 18.81 | 35 | 560.58 | 58.48 | 421.00-701.00 | 10.43 | 35 | 163.31 | 8.01 | 122.00-204.00 | 4.91 | 35 | 590.56 | 32.27 | 443.00-739.00 | 5.47 |
| PG: CDS Medonic 620 | 12 | 62.42 | 7.75 | 47.00-78.00 | 12.58 | 12 | 248.50 | 24.25 | 187.00-311.00 | 9.76 | 12 | 485.50 | 22.40 | 365.00-608.00 | 4.61 | 12 | 151.08 | 10.33 | 113.00-189.00 | 6.82 | 12 | 488.92 | 26.83 | 367.00-611.00 | 5.48 |
| PG: Beckman Coulter AcT Diff/Diff 2 | 223 | 69.28 | 6.49 | 52.00-86.00 | 9.42 | 226 | 261.11 | 12.54 | 196.00-326.00 | 4.79 | 226 | 534.77 | 37.98 | 401.00-669.00 | 7.10 | 226 | 164.71 | 27.80 | 124.00-206.00 | 16.85 | 226 | 543.16 | 31.31 | 407.00-679.00 | 5.76 |
| PG: Horiba ABX Diag. Micros/45/60 | 74 | 75.09 | 7.21 | 56.00-94.00 | 9.60 | 75 | 268.33 | 13.85 | 201.00-335.00 | 5.19 | 74 | 520.84 | 24.02 | 391.00-651.00 | 4.61 | 75 | 173.61 | 9.12 | 131.00-218.00 | 5.23 | 74 | 527.21 | 21.03 | 395.00-659.00 | 3.98 |
| SG: Select Abbott Cell Dyn Instruments (3-part Diff) | 189 | 69.52 | 13.40 | 53.00-88.00 | 19.14 | 191 | 267.47 | 35.97 | 200.00-334.00 | 13.48 | 191 | 530.54 | 44.66 | 398.00-664.00 | 8.42 | 191 | 174.74 | 21.55 | 131.00-219.00 | 12.29 | 191 | 542.74 | 65.31 | 407.00-679.00 | 12.03 |
| SG: Select Beckman Coulter Instruments (3-part Diff) | 223 | 69.28 | 6.49 | 52.00-86.00 | 9.42 | 226 | 261.11 | 12.54 | 196.00-326.00 | 4.79 | 226 | 534.77 | 37.98 | 401.00-669.00 | 7.10 | 226 | 164.71 | 27.80 | 124.00-206.00 | 16.85 | 226 | 543.16 | 31.31 | 407.00-679.00 | 5.76 |
| PG: Drew Scientific D3 | 26 | 76.62 | 8.07 | 58.00-96.00 | 10.52 | 26 | 273.58 | 13.87 | 206.00-343.00 | 5.07 | 26 | 549.46 | 77.46 | 412.00-686.00 | 14.12 | 26 | 181.96 | 38.66 | 137.00-228.00 | 21.26 | 26 | 544.31 | 28.65 | 408.00-680.00 | 5.28 |
| PG: Abbott Cell-Dyn Emerald | 135 | 70.85 | 15.27 | 53.00-89.00 | 21.55 | 137 | 264.52 | 31.97 | 199.00-331.00 | 12.08 | 137 | 518.20 | 33.62 | 389.00-648.00 | 6.49 | 137 | 179.80 | 22.83 | 135.00-225.00 | 12.67 | 137 | 524.61 | 65.60 | 394.00-656.00 | 12.50 |
| PG: CDS M Series | 145 | 62.41 | 7.91 | 47.00-78.00 | 12.74 | 146 | 234.18 | 19.17 | 176.00-293.00 | 8.21 | 146 | 465.15 | 34.40 | 349.00-581.00 | 7.40 | 146 | 146.26 | 25.47 | 110.00-183.00 | 17.47 | 146 | 480.26 | 41.31 | 360.00-600.00 | 8.60 |
| SG: CDS Hematology Instruments | 157 | 62.41 | 7.87 | 47.00-78.00 | 12.74 | 158 | 235.25 | 19.87 | 176.00-294.00 | 8.47 | 158 | 466.67 | 34.02 | 350.00-584.00 | 7.28 | 158 | 146.63 | 24.68 | 110.00-184.00 | 16.80 | 158 | 480.91 | 40.41 | 361.00-601.00 | 8.40 |
| SG: International Abbott Emerald Instruments | 137 | 70.99 | 15.30 | 53.00-89.00 | 21.55 | 139 | 264.80 | 32.15 | 199.00-331.00 | 12.15 | 139 | 520.36 | 37.90 | 390.00-650.00 | 7.29 | 139 | 180.20 | 22.91 | 135.00-225.00 | 12.72 | 139 | 525.93 | 66.39 | 395.00-658.00 | 12.62 |
| RDW | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% |
| PG: Abbott Cell-Dyn 1700 | 15 | 17.43 | 0.90 | 14.70-20.10 | 5.17 | 15 | 16.49 | 0.77 | 14.20-18.80 | 4.67 | 15 | 15.37 | 0.51 | 13.90-16.90 | 3.31 | 15 | 17.93 | 0.86 | 15.30-20.50 | 4.80 | 15 | 15.69 | 0.48 | 14.30-17.10 | 3.06 |
| PG: Abbott Cell-Dyn 1800 | 35 | 17.14 | 0.67 | 15.10-19.10 | 3.92 | 35 | 15.91 | 0.76 | 13.60-18.20 | 4.78 | 35 | 14.50 | 0.54 | 12.90-16.10 | 3.72 | 35 | 17.04 | 0.88 | 14.40-19.60 | 5.18 | 35 | 14.61 | 0.60 | 12.80-16.40 | 4.11 |
| PG: CDS Medonic 620 | 10 | 13.55 | 0.87 | 11.00-16.20 | 6.40 | 10 | 13.85 | 0.68 | 11.90-15.90 | 4.89 | 10 | 13.47 | 0.66 | 11.50-15.50 | 4.89 | 10 | 14.01 | 0.72 | 11.80-16.20 | 5.14 | 10 | 13.09 | 1.77 | 7.80-18.40 | 13.51 |
| PG: Beckman Coulter AcT Diff/Diff 2 | 216 | 15.42 | 0.33 | 14.40-16.40 | 2.14 | 217 | 14.34 | 0.35 | 13.30-15.40 | 2.45 | 217 | 13.40 | 0.65 | 11.40-15.40 | 4.85 | 217 | 15.44 | 0.74 | 13.20-17.60 | 4.81 | 217 | 13.48 | 1.68 | 8.50-18.50 | 12.44 |
| PG: Horiba ABX Diag. Micros/45/60 | 63 | 13.71 | 0.42 | 12.40-15.00 | 3.07 | 64 | 13.17 | 0.50 | 11.70-14.70 | 3.79 | 63 | 12.35 | 0.42 | 11.10-13.70 | 3.39 | 63 | 13.91 | 0.39 | 12.70-15.10 | 2.81 | 63 | 12.35 | 0.39 | 11.10-13.50 | 3.17 |
| SG: Select Abbott Cell Dyn Instruments (3-part Diff) | 185 | 15.35 | 1.29 | 11.50-19.30 | 8.38 | 185 | 14.56 | 1.22 | 10.90-18.30 | 8.36 | 185 | 13.35 | 1.15 | 9.90-16.80 | 8.65 | 185 | 15.62 | 1.34 | 11.60-19.60 | 8.59 | 185 | 13.40 | 1.16 | 9.90-16.90 | 8.66 |
| SG: Select Beckman Coulter Instruments (3-part Diff) | 216 | 15.42 | 0.33 | 14.40-16.40 | 2.14 | 217 | 14.34 | 0.35 | 13.30-15.40 | 2.45 | 217 | 13.40 | 0.65 | 11.40-15.40 | 4.85 | 217 | 15.44 | 0.74 | 13.20-17.60 | 4.81 | 217 | 13.48 | 1.68 | 8.50-18.50 | 12.44 |
| PG: Drew Scientific D3 | 26 | 14.78 | 0.86 | 12.20-17.40 | 5.81 | 26 | 14.27 | 0.69 | 12.20-16.40 | 4.83 | 26 | 13.09 | 0.67 | 11.10-15.10 | 5.11 | 26 | 15.33 | 0.84 | 12.80-17.80 | 5.49 | 26 | 13.13 | 0.80 | 10.70-15.50 | 6.11 |
| PG: Abbott Cell-Dyn Emerald | 135 | 14.65 | 0.53 | 13.10-16.30 | 3.61 | 135 | 13.99 | 0.77 | 11.70-16.30 | 5.50 | 135 | 12.81 | 0.79 | 10.40-15.20 | 6.17 | 135 | 14.99 | 0.82 | 12.50-17.50 | 5.47 | 135 | 12.82 | 0.65 | 10.90-14.80 | 5.08 |
| PG: CDS M Series | 143 | 13.95 | 0.79 | 11.70-16.30 | 5.57 | 144 | 13.94 | 0.98 | 11.00-16.80 | 7.05 | 144 | 13.62 | 0.86 | 11.00-16.20 | 6.32 | 144 | 14.15 | 0.80 | 11.70-16.50 | 5.67 | 144 | 13.70 | 0.85 | 11.10-16.30 | 6.20 |
| SG: CDS Hematology Instruments | 153 | 13.93 | 0.79 | 11.50-16.30 | 5.68 | 154 | 13.93 | 0.96 | 11.00-16.80 | 6.91 | 154 | 13.61 | 0.85 | 11.10-16.10 | 6.25 | 154 | 14.14 | 0.79 | 11.70-16.50 | 5.60 | 154 | 13.66 | 0.94 | 10.90-16.50 | 6.86 |
| SG: International Abbott Emerald Instruments | 137 | 14.66 | 0.54 | 13.10-16.30 | 3.67 | 137 | 13.99 | 0.77 | 11.70-16.30 | 5.50 | 137 | 12.82 | 0.79 | 10.50-15.10 | 6.09 | 137 | 14.98 | 0.81 | 12.60-17.40 | 5.40 | 137 | 12.83 | 0.65 | 10.90-14.80 | 5.08 |
| Red Blood Cell Count | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% | Labs | Mean | SD | Range | CV,% |
| PG: Abbott Cell-Dyn 1700 | 18 | 2.44 | 0.08 | 2.29-2.59 | 3.11 | | | | | | | | | | | | | | | | | | | | |



| Hematocrit (Waived) | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | | |
|--|-------|--------|------|---------------|-------|-------|--------|------|---------------|-------|-------|--------|------|---------------|------|-------|--------|------|---------------|------|--------|--------|-------|---------------|------|--|--|
| PG: Clay Adam Auto_Readacrit,Triac | 5 | 17.70 | 1.92 | 11.90-23.50 | 10.85 | | | | | | | | | | | | | | | | | | | | | | |
| PG: Microhematocrit, Spun | 22 | 15.89 | 1.60 | 11.10-20.70 | 10.06 | 12 | 33.75 | 0.96 | 30.90-36.70 | 2.87 | | | | | | | | | | | | | | | | | |
| AG: All Group | 31 | 16.48 | 1.80 | 11.10-21.90 | 10.91 | | | | | | | | | | | | | | | | | | | | | | |
| Hemoglobin (Waived) | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | | |
| PG: HemoCue | 54 | 6.09 | 0.59 | 4.30-7.90 | 9.67 | | | | | | | | | | | | | | | | | | | | | | |
| PG: HemoCue 201 | 109 | 6.20 | 0.82 | 3.70-8.70 | 13.23 | | | | | | | | | | | | | | | | | | | | | | |
| PG: Alere HemoPoint H2 | 38 | 6.38 | 0.21 | 5.80-7.00 | 3.28 | | | | | | | | | | | | | | | | | | | | | | |
| SG: Select Waived Hemoglobin Instruments | 202 | 6.21 | 0.69 | 4.10-8.30 | 11.13 | | | | | | | | | | | | | | | | | | | | | | |
| AG: All Group | 205 | 6.21 | 0.68 | 4.20-8.20 | 10.97 | 5 | 13.44 | 1.37 | 9.30-17.50 | 10.22 | | | | | | | | | | | | | | | | | |
| 608 - Hemocue Glucose/Hgb;628 - AFC Waived/PPM | HQ-2 | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Glucose Whole Blood (Hemocue) | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | | |
| PG: HemoCue | 41 | 48.46 | 5.70 | 31.00-65.00 | 11.88 | | | | | | | | | | | | | | | | | | | | | | |
| PG: HemoCue 201 | 167 | 46.62 | 6.12 | 29.00-65.00 | 12.98 | | | | | | | | | | | | | | | | | | | | | | |
| SG: Select Waived Hemoglobin Instruments | 208 | 46.99 | 6.07 | 29.00-65.00 | 12.98 | | | | | | | | | | | | | | | | | | | | | | |
| Hemoglobin (Hemocue) | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | | |
| PG: HemoCue | 30 | 4.96 | 0.48 | 3.60-6.40 | 9.60 | | | | | | | | | | | | | | | | | | | | | | |
| PG: HemoCue 201 | 83 | 5.08 | 0.25 | 4.30-5.80 | 4.90 | | | | | | | | | | | | | | | | | | | | | | |
| SG: Select Waived Hemoglobin Instruments | 113 | 5.05 | 0.33 | 4.00-6.00 | 6.60 | | | | | | | | | | | | | | | | | | | | | | |
| 606 - Hemocue WBC | HQW-6 | | | | | HQW-7 | | | | | HQW-8 | | | | | HQW-9 | | | | | HQW-10 | | | | | | |
| White Blood Cell Count | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | | |
| PG: Hemocue WBC System | 21 | 8.18 | 0.48 | 7.00-9.40 | 5.85 | 21 | 3.04 | 0.17 | 2.50-3.50 | 5.67 | 22 | 8.15 | 0.33 | 6.90-9.30 | 4.07 | 22 | 21.74 | 1.71 | 18.40-25.00 | 7.88 | 22 | 3.09 | 0.25 | 2.60-3.60 | 8.06 | | |
| 700 - i-Stat Chemistry;721 - i-Stat Chemistry & Blood Gases;722 - i-Stat Chemistry - Waived;723 - i-Stat Blood Gases | IST-6 | | | | | IST-7 | | | | | IST-8 | | | | | IST-9 | | | | | IST-10 | | | | | | |
| Blood Gas - pCO2 | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | | |
| PG: Abbott i-Stat | 14 | 69.67 | 2.64 | 64.00-76.00 | 3.71 | 14 | 20.41 | 1.08 | 15.00-25.00 | 5.50 | 14 | 29.25 | 1.25 | 24.00-34.00 | 4.14 | 13 | 26.42 | 0.98 | 21.00-31.00 | 3.85 | 13 | 50.31 | 1.98 | 45.00-55.00 | 4.00 | | |
| SG: International Abbott i-Stat Instruments | 18 | 69.47 | 2.50 | 63.00-75.00 | 3.62 | 18 | 20.54 | 0.97 | 16.00-26.00 | 4.76 | 18 | 29.25 | 1.16 | 24.00-34.00 | 4.14 | 17 | 26.67 | 1.00 | 22.00-32.00 | 3.70 | 17 | 50.35 | 1.81 | 45.00-55.00 | 3.60 | | |
| Blood Gas - pH | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | | |
| PG: Abbott i-Stat | 14 | 7.16 | 0.01 | 7.12-7.20 | 0.13 | 14 | 7.51 | 0.01 | 7.47-7.55 | 0.19 | 14 | 7.54 | 0.01 | 7.50-7.58 | 0.17 | 14 | 7.52 | 0.04 | 7.48-7.56 | 0.59 | 14 | 7.24 | 0.02 | 7.20-7.28 | 0.29 | | |
| SG: International Abbott i-Stat Instruments | 18 | 7.16 | 0.01 | 7.12-7.20 | 0.11 | 18 | 7.51 | 0.01 | 7.47-7.55 | 0.16 | 18 | 7.54 | 0.01 | 7.50-7.58 | 0.16 | 18 | 7.52 | 0.04 | 7.48-7.56 | 0.51 | 18 | 7.24 | 0.02 | 7.20-7.28 | 0.26 | | |
| Blood Gas - pO2 | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | | |
| PG: Abbott i-Stat | 14 | 91.43 | 5.27 | 75.00-107.00 | 5.82 | 14 | 162.36 | 8.42 | 137.00-187.00 | 5.19 | 14 | 106.36 | 6.56 | 86.00-126.00 | 6.23 | 13 | 138.46 | 7.32 | 116.00-160.00 | 5.29 | 13 | 176.23 | 10.28 | 145.00-207.00 | 5.85 | | |
| SG: International Abbott i-Stat Instruments | 18 | 89.94 | 5.45 | 74.00-107.00 | 6.11 | 18 | 163.06 | 7.57 | 140.00-186.00 | 4.66 | 18 | 106.00 | 5.81 | 89.00-123.00 | 5.47 | 17 | 138.53 | 6.42 | 120.00-158.00 | 4.60 | 17 | 178.29 | 9.80 | 149.00-207.00 | 5.51 | | |
| Chloride (CL-) | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | | |
| PG: Abbott i-Stat / Thermal | 8 | 86.75 | 0.46 | 83.00-91.00 | 0.57 | 8 | 104.75 | 1.03 | 100.00-110.00 | 0.99 | 8 | 107.50 | 0.54 | 103.00-113.00 | 0.49 | 8 | 80.25 | 0.46 | 76.00-84.00 | 0.63 | 8 | 117.00 | 1.07 | 111.00-123.00 | 0.94 | | |
| SG: International Abbott i-Stat Instruments / Thermal | 30 | 86.77 | 0.63 | 83.00-91.00 | 0.73 | 12 | 104.92 | 0.90 | 100.00-110.00 | 0.86 | 12 | 107.58 | 0.52 | 103.00-113.00 | 0.47 | 12 | 80.58 | 0.67 | 77.00-85.00 | 0.86 | 12 | 116.92 | 0.90 | 111.00-123.00 | 0.77 | | |
| CO2 (Total CO2 Bicarb) | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | | |
| PG: Abbott i-Stat / Thermal | 11 | 27.18 | 0.75 | 24.90-29.40 | 2.76 | 11 | 18.18 | 0.98 | 15.30-21.10 | 5.38 | 11 | 26.73 | 0.65 | 24.80-28.60 | 2.43 | 11 | 22.73 | 0.91 | 20.00-25.40 | 3.96 | 11 | 24.00 | 1.18 | 20.50-27.50 | 4.92 | | |
| SG: International Abbott i-Stat Instruments / Thermal | 33 | 27.15 | 0.80 | 24.80-29.60 | 2.94 | 15 | 18.47 | 0.99 | 15.50-21.50 | 5.35 | 15 | 26.87 | 0.64 | 25.00-28.80 | 2.38 | 15 | 22.87 | 0.83 | 20.40-25.40 | 3.62 | 15 | 24.07 | 1.03 | 21.00-27.20 | 4.27 | | |
| Creatinine | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | | |
| PG: Abbott i-Stat / Thermal | 15 | 5.87 | 0.22 | 5.00-6.80 | 3.73 | 14 | 1.12 | 0.04 | 0.80-1.40 | 3.64 | 14 | 0.76 | 0.05 | 0.50-1.10 | 6.25 | 14 | 1.60 | 0.04 | 1.30-1.90 | 2.50 | 14 | 2.57 | 0.08 | 2.20-3.00 | 3.08 | | |
| SG: International Abbott i-Stat Instruments / Thermal | 40 | 5.87 | 0.17 | 5.00-6.80 | 2.88 | 19 | 1.12 | 0.04 | 0.80-1.40 | 3.64 | 19 | 0.76 | 0.05 | 0.50-1.10 | 6.25 | 19 | 1.60 | 0.04 | 1.30-1.90 | 2.50 | 19 | 2.58 | 0.08 | 2.20-3.00 | 3.08 | | |
| Glucose | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | | |
| PG: Abbott i-Stat / Thermal | 11 | 140.27 | 1.19 | 126.00-154.00 | 0.86 | 11 | 81.45 | 0.69 | 73.30-89.70 | 0.86 | 11 | 65.91 | 0.94 | 59.30-72.50 | 1.43 | 11 | 123.27 | 1.42 | 111.00-135.60 | 1.15 | 11 | 232.18 | 2.04 | 209.00-255.40 | 0.88 | | |
| SG: International Abbott i-Stat Instruments / Thermal | 33 | 139.90 | 1.18 | 126.00-154.00 | 0.86 | 15 | 81.48 | 0.72 | 73.30-89.70 | 0.88 | 15 | 66.11 | 0.87 | 59.50-72.70 | 1.36 | 15 | 123.08 | 1.25 | 110.80-135.40 | 1.02 | 15 | 232.02 | 1.81 | 208.80-255.20 | 0.78 | | |
| Hematocrit | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | | |
| PG: Abbott i-Stat / Thermal | 7 | 26.14 | 0.38 | 24.50-27.70 | 1.46 | 7 | 28.71 | 0.49 | 27.00-30.40 | 1.71 | 7 | 37.71 | 0.49 | 35.40-40.00 | 1.30 | 7 | 36.00 | 0.00 | 33.80-38.20 | 0.00 | 7 | 30.43 | 0.54 | 28.60-32.20 | 1.74 | | |
| SG: International Abbott i-Stat Instruments / Thermal | 20 | 26.05 | 0.39 | 24.40-27.60 | 1.50 | 11 | 28.45 | 0.52 | 26.80-30.20 | 1.82 | 11 | 37.82 | 0.41 | 35.50-40.10 | 1.06 | 11 | 35.91 | 0.30 | 33.70-38.10 | 0.84 | 11 | 30.45 | 0.52 | 28.70-32.30 | 1.70 | | |
| Hemoglobin | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | | |
| PG: Abbott i-Stat / Thermal | 7 | 8.89 | 0.16 | 8.30-9.50 | 1.80 | 7 | 9.80 | 0.21 | 9.10-10.50 | 2.14 | 7 | 12.83 | 0.16 | 11.90-13.70 | 1.25 | 7 | 12.17 | 0.08 | 11.30-13.10 | 0.66 | 7 | 10.30 | 0.20 | 9.60-11.00 | 1.94 | | |
| SG: International Abbott i-Stat Instruments / Thermal | 20 | 8.84 | 0.15 | 8.20-9.40 | 1.70 | 11 | 9.69 | 0.22 | 9.00-10.40 | 2.27 | 11 | 12.85 | 0.13 | 12.00-13.80 | 1.01 | 11 | 12.15 | 0.10 | 11.30-13.10 | 0.82 | 11 | 10.32 | 0.18 | 9.60-11.00 | 1.75 | | |
| Ionized Calcium | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | | |
| PG: Abbott i-Stat / Thermal | 7 | 2.16 | 0.02 | 2.09-2.23 | 1.06 | 7 | 0.78 | 0.01 | 0.74-0.82 | 1.67 | 7 | 0.90 | 0.01 | 0.86-0.94 | 1.33 | 7 | 1.42 | 0.01 | 1.38-1.46 | 0.92 | 7 | 0.59 | 0.01 | 0.56-0.62 | 1.53 | | |
| SG: International Abbott i-Stat Instruments / Thermal | 26 | 2.17 | 0.03 | 2.08-2.26 | 1.43 | 11 | 0.78 | 0.01 | 0.75-0.81 | 1.28 | 11 | 0.90 | 0.01 | 0.87-0.93 | 1.11 | 11 | 1.42 | 0.01 | 1.38-1.46 | 0.92 | 11 | 0.59 | 0.01 | 0.56-0.62 | 1.53 | | |
| Potassium (K+) | Labs | Mean | | | | | | | | | | | | | | | | | | | | | | | | | |



| BEecf | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
|---|-------|--------|-------|---------------|--------|-------|--------|-------|---------------|--------|-------|--------|-------|---------------|-------|-------|--------|-------|-----------------|---------|--------|-------|------|--------------|--------|
| PG: Abbott i-Stat / Thermal | 6 | -4.50 | 1.05 | 0.00-2.00 | -20.00 | 6 | -6.50 | 1.05 | 0.00-4.00 | -14.29 | 6 | 3.17 | 0.41 | 2.00-4.00 | 13.33 | 5 | -1.20 | 1.48 | 0.00-4.00 | -150.00 | 6 | -6.17 | 0.75 | 0.00-4.00 | -13.33 |
| sO2 | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| PG: Abbott i-Stat | 6 | 93.17 | 0.98 | 90.00-96.00 | 1.08 | 6 | 100.00 | 0.00 | 100.00-100.00 | 0.00 | 6 | 99.00 | 0.00 | 99.00-99.00 | 0.00 | 6 | 99.33 | 0.52 | 98.00-101.00 | 0.51 | 6 | 99.17 | 0.41 | 98.00-100.00 | 0.40 |
| SG: International Abbott i-Stat Instruments | 6 | 93.17 | 0.98 | 90.00-96.00 | 1.08 | 6 | 100.00 | 0.00 | 100.00-100.00 | 0.00 | 6 | 99.00 | 0.00 | 99.00-99.00 | 0.00 | 6 | 99.33 | 0.52 | 98.00-101.00 | 0.51 | 6 | 99.17 | 0.41 | 98.00-100.00 | 0.40 |
| Lactate | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| PG: Abbott i-Stat | 5 | 3.55 | 0.03 | 3.46-3.64 | 0.87 | 5 | 1.15 | 0.11 | 0.82-1.48 | 9.48 | 5 | 0.93 | 0.08 | 0.68-1.18 | 8.92 | 5 | 1.44 | 0.07 | 1.24-1.64 | 4.65 | 5 | 3.25 | 0.08 | 3.02-3.48 | 2.34 |
| SG: International Abbott i-Stat Instruments | 9 | 3.56 | 0.03 | 3.46-3.66 | 0.93 | 9 | 1.13 | 0.08 | 0.89-1.37 | 6.99 | 9 | 0.91 | 0.06 | 0.73-1.09 | 6.70 | 9 | 1.44 | 0.05 | 1.30-1.58 | 3.33 | 9 | 3.26 | 0.06 | 3.09-3.43 | 1.78 |
| Anion Gap | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| SG: International Abbott i-Stat Instruments / Thermal | 8 | 30.75 | 1.39 | 27.00-35.00 | 4.52 | 5 | 35.00 | 0.00 | 35.00-35.00 | 0.00 | 5 | 41.00 | 0.00 | 41.00-41.00 | 0.00 | 5 | 47.00 | 0.00 | 47.00-47.00 | 0.00 | 5 | 36.00 | 0.00 | 36.00-36.00 | 0.00 |
| PG: Abbott i-Stat (Waived) / Thermal | 7 | 30.57 | 1.40 | 27.00-35.00 | 4.52 | | | | | | | | | | | | | | | | | | | | |
| 716 - Blood Lead - Waived Methods;709 - Blood Lead | LE-6 | | | | | LE-7 | | | | | LE-8 | | | | | LE-9 | | | | | LE-10 | | | | |
| Blood Lead | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| SG: Select ESA LeadCare Instruments | 5 | 16.20 | 1.16 | 12.20-20.20 | 7.16 | 5 | 5.00 | 1.11 | 1.00-9.00 | 22.20 | 5 | 43.80 | 2.04 | 39.40-48.20 | 4.66 | 5 | 14.63 | 1.98 | 10.60-18.60 | 13.56 | 5 | 5.48 | 1.62 | 1.50-9.50 | 29.45 |
| PG: Magellan Diag LeadCare Ultra | 5 | 14.00 | 0.00 | 10.00-18.00 | 0.00 | 5 | 4.00 | 0.00 | 0.00-8.00 | 0.00 | 5 | 41.00 | 0.00 | 36.90-45.10 | 0.00 | 5 | 14.00 | 0.00 | 10.00-18.00 | 0.00 | 5 | 5.00 | 0.00 | 1.00-9.00 | 0.00 |
| Blood Lead (Waived) | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| PG: ESA LeadCare II (Waived) | 39 | 16.00 | 1.52 | 11.40-20.60 | 9.50 | | | | | | | | | | | | | | | | | | | | |
| 662 - Neonatal Bilirubin (5 specimens);663 - Neonatal Bilirubin (2 specimens) | NB-6 | | | | | NB-7 | | | | | NB-8 | | | | | NB-9 | | | | | NB-10 | | | | |
| Neonatal Bilirubin (Total) | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| AG: All Group | 9 | 0.23 | 0.13 | 0.00-0.60 | 65.00 | 18 | 18.37 | 0.90 | 14.70-22.10 | 4.89 | 16 | 15.86 | 1.33 | 12.70-19.10 | 8.36 | 16 | 11.36 | 0.57 | 9.10-13.70 | 5.00 | 16 | 5.76 | 0.36 | 4.60-7.00 | 6.21 |
| 748 - i-Stat Prottime | PTI-6 | | | | | PTI-7 | | | | | PTI-8 | | | | | PTI-9 | | | | | PTI-10 | | | | |
| Prottime - Whole Blood | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| PG: Abbott i-Stat Prottime | 8 | 27.39 | 2.00 | 23.30-31.50 | 7.30 | 8 | 13.46 | 0.95 | 11.50-15.50 | 7.04 | 8 | 13.45 | 0.89 | 11.50-15.50 | 6.59 | 8 | 28.15 | 2.77 | 24.00-32.40 | 9.82 | 8 | 27.17 | 2.78 | 23.10-31.30 | 10.22 |
| PG: Abbott Coaguse | 8 | 39.91 | 3.41 | 33.90-45.90 | 8.55 | 8 | 14.13 | 2.68 | 12.00-16.20 | 19.01 | 8 | 15.13 | 2.50 | 12.80-17.40 | 16.56 | 8 | 36.09 | 4.36 | 30.70-41.50 | 12.08 | 8 | 36.96 | 2.19 | 31.40-42.50 | 5.92 |
| SG: International Abbott i-Stat Instruments | 9 | 27.39 | 1.87 | 23.30-31.50 | 6.82 | 9 | 13.46 | 0.89 | 11.50-15.50 | 6.59 | 9 | 13.39 | 0.85 | 11.40-15.40 | 6.34 | 9 | 27.99 | 2.64 | 23.80-32.20 | 9.43 | 9 | 27.02 | 2.64 | 22.90-31.10 | 9.78 |
| SG: International Abbott Coaguse Instruments | 9 | 39.51 | 3.41 | 33.60-45.40 | 8.63 | 9 | 13.96 | 2.56 | 11.90-16.10 | 18.29 | 9 | 14.71 | 2.64 | 12.50-16.90 | 17.96 | 9 | 36.04 | 4.08 | 30.60-41.40 | 11.33 | 9 | 36.72 | 2.17 | 31.20-42.20 | 5.91 |
| INR | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| PG: Abbott i-Stat Prottime | 9 | 2.39 | 0.19 | 1.80-3.00 | 7.92 | 9 | 1.11 | 0.09 | 0.80-1.40 | 8.18 | 9 | 1.12 | 0.07 | 0.90-1.30 | 6.36 | 9 | 2.44 | 0.26 | 1.60-3.20 | 10.83 | 9 | 2.38 | 0.24 | 1.70-3.10 | 10.00 |
| PG: Abbott Coaguse | 9 | 3.36 | 0.30 | 2.50-4.30 | 8.82 | 9 | 1.21 | 0.20 | 0.60-1.80 | 16.67 | 9 | 1.29 | 0.28 | 0.50-2.10 | 21.54 | 9 | 3.09 | 0.47 | 1.70-4.50 | 15.16 | 9 | 3.17 | 0.43 | 1.90-4.50 | 13.44 |
| SG: International Abbott i-Stat Instruments | 14 | 2.36 | 0.17 | 1.90-2.90 | 6.67 | 13 | 1.12 | 0.08 | 0.90-1.30 | 7.27 | 14 | 1.12 | 0.06 | 0.90-1.30 | 5.45 | 14 | 2.37 | 0.23 | 1.70-3.10 | 9.58 | 14 | 2.31 | 0.21 | 1.70-2.90 | 9.13 |
| SG: International Abbott Coaguse Instruments | 10 | 3.32 | 0.30 | 2.40-4.20 | 9.09 | 10 | 1.19 | 0.20 | 0.60-1.80 | 16.67 | 10 | 1.25 | 0.29 | 0.40-2.20 | 22.31 | 10 | 3.07 | 0.45 | 1.80-4.50 | 14.52 | 10 | 3.13 | 0.42 | 1.80-4.40 | 13.55 |
| 611 - QBC Hematology | QBC-6 | | | | | QBC-7 | | | | | QBC-8 | | | | | QBC-9 | | | | | QBC-10 | | | | |
| Hemoglobin | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| PG: QBC Star - QBC Star Tubes | 9 | 11.02 | 0.18 | 10.20-11.80 | 1.64 | 9 | 11.56 | 0.24 | 10.80-12.40 | 2.07 | 8 | 7.13 | 0.23 | 6.60-7.60 | 3.24 | 9 | 9.40 | 0.32 | 8.70-10.10 | 3.40 | 9 | 11.04 | 0.24 | 10.20-11.80 | 2.18 |
| AG: All Group | 16 | 10.79 | 0.39 | 10.00-11.60 | 3.61 | 16 | 11.29 | 0.46 | 10.50-12.10 | 4.07 | 15 | 7.15 | 0.26 | 6.70-7.70 | 3.61 | 16 | 9.33 | 0.33 | 8.60-10.00 | 3.55 | 16 | 10.81 | 0.43 | 10.00-11.60 | 3.98 |
| Hematocrit | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| PG: QBC Star - QBC Star Tubes | 9 | 33.22 | 0.68 | 31.20-35.20 | 2.05 | 9 | 34.93 | 0.97 | 32.80-37.00 | 2.78 | 9 | 21.79 | 0.96 | 20.50-23.10 | 4.40 | 9 | 28.58 | 1.09 | 26.90-30.30 | 3.81 | 9 | 33.31 | 0.83 | 31.30-35.30 | 2.49 |
| AG: All Group | 16 | 33.21 | 0.65 | 31.20-35.20 | 1.96 | 16 | 34.73 | 1.09 | 32.60-36.80 | 3.14 | 16 | 21.99 | 0.88 | 20.70-23.30 | 4.00 | 16 | 28.48 | 0.88 | 26.80-30.20 | 3.09 | 16 | 32.75 | 0.94 | 30.80-34.80 | 2.87 |
| White Blood Cell Count | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| PG: QBC Star - QBC Star Tubes | 9 | 18.41 | 1.23 | 15.60-21.20 | 6.68 | 7 | 28.70 | 2.98 | 24.40-33.00 | 10.38 | 7 | 6.29 | 0.33 | 5.40-7.20 | 5.24 | 9 | 10.19 | 0.72 | 8.70-11.70 | 7.06 | 9 | 18.57 | 0.91 | 15.80-21.40 | 4.89 |
| SG: Select QBC - Cap and Ven | 5 | 12.74 | 1.11 | 10.80-14.60 | 8.74 | 5 | 15.80 | 1.19 | 13.40-18.20 | 7.53 | 5 | 4.47 | 0.31 | 3.80-5.20 | 6.89 | 5 | 7.96 | 0.18 | 6.80-9.20 | 2.25 | 5 | 12.82 | 0.33 | 10.90-14.70 | 2.58 |
| SG: Select QBC Instruments | 11 | 18.50 | 1.13 | 15.70-21.30 | 6.11 | 8 | 28.79 | 2.77 | 24.50-33.10 | 9.62 | 8 | 6.40 | 0.44 | 5.40-7.40 | 6.88 | 11 | 10.16 | 0.65 | 8.70-11.70 | 6.37 | 11 | 18.62 | 1.02 | 15.80-21.40 | 5.48 |
| Granulocyte | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| PG: QBC Star - QBC Star Tubes | 9 | 51.22 | 2.95 | 42.00-60.00 | 5.69 | 9 | 32.89 | 2.76 | 25.00-41.00 | 8.48 | 9 | 57.67 | 6.38 | 39.00-77.00 | 11.03 | 9 | 56.67 | 4.12 | 45.00-69.00 | 7.19 | 9 | 50.78 | 1.48 | 47.00-56.00 | 2.94 |
| AG: All Group | 16 | 50.88 | 4.73 | 37.00-65.00 | 9.22 | 15 | 33.60 | 4.12 | 22.00-46.00 | 12.06 | 16 | 56.38 | 6.99 | 35.00-77.00 | 12.50 | 16 | 57.94 | 4.97 | 43.00-73.00 | 8.62 | 16 | 50.25 | 4.39 | 37.00-63.00 | 8.80 |
| Lymphocyte/Monocyte | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| PG: QBC Star - QBC Star Tubes | 9 | 48.78 | 2.95 | 40.00-58.00 | 5.92 | 9 | 67.11 | 2.76 | 59.00-75.00 | 4.18 | 9 | 42.33 | 6.38 | 23.00-61.00 | 15.24 | 9 | 43.33 | 4.12 | 31.00-55.00 | 9.53 | 9 | 49.22 | 1.48 | 45.00-54.00 | 3.06 |
| AG: All Group | 16 | 49.13 | 4.73 | 35.00-63.00 | 9.59 | 16 | 65.56 | 5.20 | 50.00-82.00 | 7.88 | 16 | 43.63 | 6.99 | 23.00-65.00 | 15.91 | 16 | 42.06 | 4.97 | 27.00-57.00 | 11.90 | 16 | 49.75 | 4.39 | 37.00-63.00 | 8.80 |
| Platelet Count | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| PG: QBC Star - QBC Star Tubes | 9 | 177.11 | 10.01 | 133.00-221.00 | 5.65 | 9 | 167.56 | 11.02 | 126.00-210.00 | 6.55 | 8 | 689.13 | 30.05 | 517.00-861.00 | 4.35 | 9 | 389.78 | 20.22 | 293.00-488.00</ | | | | | | |



| 708 - Tumor Markers | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
|--|--------|-------|---------------|-------|------|--------|-------|---------------|-------|------|-------|------|-------------|-------|------|--------|------|---------------|-------|------|-------|------|-------------|-------|-------|------|----|-------|------|--|--|--|--|--|--|--|--|--|--|
| CA 125 | | | | | TM-6 | | | | | TM-7 | | | | | TM-8 | | | | | TM-9 | | | | | TM-10 | | | | | | | | | | | | | | |
| Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | | | | | | | | | | |
| PG: TOSOH AIA | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 5 | 202.95 | 7.99 | 179.00-227.00 | 3.94 | 5 | 139.60 | 6.50 | 120.10-159.10 | 4.66 | 5 | 18.55 | 2.19 | 11.90-25.10 | 11.84 | 5 | 256.25 | 7.85 | 232.80-279.90 | 3.06 | 5 | 81.10 | 3.54 | 70.50-91.70 | 4.36 | | | | | | | | | | | | | | | |
| SG: Select OCD Vitros Instruments | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 5 | 110.33 | 3.21 | 100.70-119.90 | 2.91 | 5 | 78.13 | 4.10 | 65.80-90.40 | 5.25 | 5 | 10.97 | 0.71 | 8.90-13.10 | 6.45 | 5 | 145.00 | 4.36 | 131.90-158.10 | 3.01 | 5 | 44.53 | 1.97 | 38.60-50.40 | 4.43 | | | | | | | | | | | | | | | |
| SG: International OCD Vitros Instruments | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 5 | 110.50 | 2.65 | 102.50-118.50 | 2.40 | 5 | 78.13 | 3.35 | 68.00-88.10 | 4.29 | 5 | 10.95 | 0.58 | 9.30-12.70 | 5.27 | 5 | 145.25 | 3.59 | 134.50-156.10 | 2.47 | 5 | 44.22 | 1.73 | 39.00-49.40 | 3.91 | | | | | | | | | | | | | | | |
| PSA - Free | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| AG: All Group | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 6 | 4.13 | 0.67 | 2.10-6.10 | 16.34 | 6 | 2.71 | 0.49 | 1.20-4.20 | 18.15 | 5 | 0.01 | 0.00 | 0.00-0.00 | 0.00 | 6 | 5.59 | 1.02 | 2.50-8.70 | 18.21 | 6 | 1.36 | 0.27 | 0.60-2.20 | 19.29 | | | | | | | | | | | | | | | |
| AG: All Group | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 6 | 4.47 | 0.56 | 2.80-6.13 | 12.42 | 6 | 2.99 | 0.36 | 1.91-4.07 | 12.01 | 5 | 0.04 | 0.01 | 0.03-0.06 | 12.50 | 6 | 6.07 | 0.83 | 3.59-8.55 | 13.61 | 6 | 1.39 | 0.21 | 0.77-2.01 | 14.89 | | | | | | | | | | | | | | | |
| 705 - Urine Microalbumin & Creatinine | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| UM-2 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Urine Creatinine | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| PG: Siemens Clinitek 50 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 9 | 133.33 | 50.00 | 0.00-283.00 | 37.59 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| PG: Siemens Clinitek Status | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 29 | 101.72 | 50.85 | 0.00-255.00 | 49.90 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| PG: Siemens Dimension/AR/ES/RxL/Xpand | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 8 | 64.70 | 4.29 | 52.00-78.00 | 6.62 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| PG: Beckman (Olympus) AU | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 7 | 70.92 | 2.17 | 64.00-78.00 | 3.10 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| SG: Semi-quant Methods - Ur Microalbumin-Creat | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 43 | 108.29 | 51.59 | 0.00-263.00 | 47.78 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| PG: Siemens (Bayer) DCA Vantage | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 15 | 75.39 | 4.05 | 63.00-87.00 | 5.47 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| PG: Afinion ACR | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 8 | 71.66 | 3.42 | 62.00-82.00 | 4.72 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| AG: All Group | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 79 | 76.90 | 16.12 | 29.00-125.00 | 20.91 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Urine Microalbumin | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| PG: Siemens Clinitek 50 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 9 | 134.44 | 30.87 | 41.00-227.00 | 23.06 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| PG: Siemens Clinitek Status | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 30 | 122.00 | 34.88 | 17.00-227.00 | 28.61 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| PG: Siemens Dimension/AR/ES/RxL/Xpand | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 8 | 159.31 | 7.82 | 136.00-182.00 | 4.91 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| PG: Beckman (Olympus) AU | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 7 | 159.27 | 6.53 | 140.00-179.00 | 4.09 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| SG: Semi-quant Methods - Ur Microalbumin-Creat | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 44 | 127.01 | 34.31 | 24.00-230.00 | 27.01 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| PG: Siemens (Bayer) DCA Vantage | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 15 | 166.42 | 37.70 | 53.00-279.00 | 22.71 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| SG: Select Siemens (Bayer) DCA Instruments | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 19 | 164.98 | 33.38 | 65.00-265.00 | 20.24 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| PG: Afinion ACR | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 8 | 153.59 | 13.35 | 114.00-194.00 | 8.70 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| AG: All Group | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 92 | 140.11 | 31.26 | 46.00-234.00 | 22.36 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 728 - Urine Chemistry | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| URC-3 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Total Protein - Urine | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| AG: All Group | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 6 | 98.03 | 3.11 | 88.70-107.30 | 3.17 | 6 | 11.07 | 2.30 | 4.20-18.00 | 20.72 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Creatinine - Urine | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| AG: All Group | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 5 | 176.36 | 15.62 | 129.50-223.30 | 8.85 | 5 | 65.72 | 10.55 | 34.00-97.40 | 16.06 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |



| 601 - Waived/PPM Tests II;609 - Waived/PPM Tests I;664 - Whole Blood Glucose;669 - Whole Blood Glucose | WBG-6 | | | | | WBG-7 | | | | | WBG-8 | | | | | WBG-9 | | | | | WBG-10 | | | | |
|--|-------|--------|-------|---------------|-------|-------|--------|-------|---------------|-------|-------|--------|--------|---------------|-------|-------|--------|--------|---------------|-------|--------|--------|-------|---------------|-------|
| | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% | Labs | Mean | SD | Range | CV.% |
| PG: Abbott Freestyle Lite | 11 | 59.36 | 3.04 | 50.00-68.00 | 5.08 | | | | | | | | | | | | | | | | | | | | |
| PG: Bayer Ascensia Contour | 109 | 48.84 | 9.43 | 21.00-77.00 | 19.18 | 20 | 343.30 | 27.98 | 259.00-427.00 | 8.16 | | | | | | | | | | | | | | | |
| PG: Bayer Ascensia Elite XL | 9 | 45.89 | 3.02 | 37.00-55.00 | 6.52 | | | | | | | | | | | | | | | | | | | | |
| PG: Lifescan One Touch Ultra | 41 | 65.71 | 11.78 | 31.00-101.00 | 17.88 | | | | | | 5 | 294.50 | 128.63 | 0.00-681.00 | 43.59 | 5 | 347.25 | 136.60 | 0.00-757.00 | 39.37 | 5 | 205.75 | 92.37 | 0.00-483.00 | 44.85 |
| PG: Lifescan Sure Step, others | 5 | 108.60 | 12.03 | 73.00-145.00 | 11.01 | | | | | | | | | | | | | | | | | | | | |
| PG: McKesson/SunMark True Track | 52 | 154.29 | 58.82 | 0.00-330.00 | 38.18 | | | | | | 5 | 472.00 | 0.00 | 472.00-472.00 | 0.00 | 5 | 318.00 | 0.00 | 318.00-318.00 | 0.00 | 5 | 347.00 | 0.00 | 347.00-347.00 | 0.00 |
| PG: Roche Accu-Chek Advantage/Comp/GTS | 27 | 60.56 | 8.02 | 37.00-85.00 | 13.11 | | | | | | | | | | | | | | | | | | | | |
| PG: Roche AccuChek Aviva | 36 | 65.56 | 2.89 | 57.00-75.00 | 4.39 | | | | | | | | | | | | | | | | | | | | |
| PG: Roche Accu-Chek II | 5 | 60.80 | 3.27 | 51.00-71.00 | 5.41 | | | | | | | | | | | | | | | | | | | | |
| PG: Roche Accu-Chek, others | 18 | 64.50 | 12.61 | 27.00-103.00 | 19.38 | | | | | | | | | | | | | | | | | | | | |
| SG: Select Lifescan One Touch Instruments | 51 | 65.71 | 12.69 | 28.00-104.00 | 19.24 | | | | | | | | | | | | | | | | | | | | |
| SG: Select Abbott Freestyle Instruments | 19 | 57.21 | 5.94 | 39.00-75.00 | 10.35 | | | | | | | | | | | | | | | | | | | | |
| PG: Roche AccuChek Inform | 71 | 64.75 | 2.15 | 58.00-72.00 | 3.38 | 38 | 377.26 | 11.34 | 343.00-411.00 | 3.00 | 33 | 171.42 | 5.12 | 156.00-186.00 | 2.98 | 33 | 202.39 | 4.78 | 188.00-216.00 | 2.38 | 33 | 114.39 | 3.01 | 105.00-123.00 | 2.63 |
| PG: Consult Diag Quintet Blood Glucose | 29 | 60.41 | 11.65 | 25.00-95.00 | 19.50 | 10 | 489.70 | 28.75 | 404.00-576.00 | 5.86 | | | | | | | | | | | | | | | |
| PG: Medisense Precision Xceed Pro | 26 | 57.69 | 3.89 | 46.00-70.00 | 6.72 | 5 | 423.00 | 25.91 | 345.00-501.00 | 6.12 | | | | | | | | | | | | | | | |
| SG: Select Medisense Instruments | 34 | 58.62 | 5.91 | 41.00-77.00 | 10.00 | 5 | 423.00 | 25.91 | 345.00-501.00 | 6.12 | | | | | | | | | | | | | | | |
| SG: Select Roche WBG Meters | 274 | 64.56 | 5.25 | 49.00-81.00 | 8.00 | 56 | 374.75 | 14.01 | 333.00-417.00 | 3.73 | 35 | 170.60 | 6.14 | 153.00-189.00 | 3.57 | 35 | 201.80 | 5.26 | 186.00-218.00 | 2.62 | 35 | 114.23 | 3.06 | 105.00-123.00 | 2.72 |
| SG: Select Bayer WBG Meters | 134 | 49.33 | 9.82 | 20.00-78.00 | 20.00 | 26 | 354.88 | 41.12 | 232.00-478.00 | 11.58 | 7 | 137.86 | 18.23 | 83.00-193.00 | 13.19 | 7 | 165.86 | 21.09 | 103.00-229.00 | 12.71 | 7 | 89.29 | 13.39 | 49.00-129.00 | 15.06 |
| PG: Nova Biomedical StatStrip Glucose meter | 6 | 54.83 | 4.40 | 42.00-68.00 | 8.00 | | | | | | | | | | | | | | | | | | | | |
| SG: Select Nova Biomedical StatStrip Instruments | 9 | 55.22 | 3.67 | 44.00-66.00 | 6.73 | | | | | | | | | | | | | | | | | | | | |
| PG: McKesson TrueResult | 6 | 84.50 | 15.14 | 40.00-130.00 | 17.76 | | | | | | | | | | | | | | | | | | | | |
| PG: Bayer Contour Next | 7 | 47.43 | 6.85 | 26.00-68.00 | 14.68 | | | | | | | | | | | | | | | | | | | | |
| PG: Nipro TRUeTrack | 6 | 191.00 | 17.91 | 137.00-245.00 | 9.37 | | | | | | | | | | | | | | | | | | | | |
| PG: Roche AccuChek Performa | 115 | 65.57 | 3.17 | 56.00-76.00 | 4.85 | 6 | 377.17 | 9.62 | 348.00-406.00 | 2.55 | | | | | | | | | | | | | | | |
| PG: McKesson True Metrix Pro Blood Glucose Meter | 8 | 132.25 | 54.13 | 0.00-294.00 | 40.98 | | | | | | | | | | | | | | | | | | | | |
| SG: Select EvenCare Whole Blood Glucose | 5 | 75.00 | 2.92 | 66.00-84.00 | 3.87 | | | | | | | | | | | | | | | | | | | | |
| AG: All Group | 605 | 61.69 | 12.44 | 25.00-99.00 | 20.00 | 111 | 392.23 | 63.83 | 201.00-583.00 | 16.28 | 52 | 172.75 | 24.44 | 100.00-246.00 | 14.10 | 52 | 205.29 | 28.64 | 119.00-291.00 | 13.95 | 51 | 113.22 | 12.98 | 74.00-152.00 | 11.50 |



| 789 - Bacterial Antigen Detection | Labs | Response | Number | Percent |
|---|-------------|------------------------------------|---------------|----------------|
| | | C. difficile Toxin Detect | | |
| AD-6 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Positive | 2 | 100.00% |
| Alere C diff Quik Chek Complete | 2 | *Positive | 2 | 100.00% |
| AD-7 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Positive | 2 | 100.00% |
| Alere C diff Quik Chek Complete | 2 | *Positive | 2 | 100.00% |
| AD-8 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Negative | 2 | 100.00% |
| Alere C diff Quik Chek Complete | 2 | *Negative | 2 | 100.00% |
| AD-9 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Positive | 2 | 100.00% |
| Alere C diff Quik Chek Complete | 2 | *Positive | 2 | 100.00% |
| AD-10 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Negative | 2 | 100.00% |
| Alere C diff Quik Chek Complete | 2 | *Negative | 2 | 100.00% |
| 643 - ANA (Latex);760 - ANA (ELISA);757 - Special Immunology | Labs | Response | Number | Percent |
| | | ANA (ELISA) (International) | | |
| AE-6 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Positive | 2 | 100.00% |
| Dynex DS2 | 2 | *Positive | 2 | 100.00% |
| AE-7 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Positive | 2 | 100.00% |
| Dynex DS2 | 2 | *Positive | 2 | 100.00% |
| AE-8 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Negative | 2 | 100.00% |
| Dynex DS2 | 2 | *Negative | 2 | 100.00% |
| AE-9 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Negative | 2 | 100.00% |
| Dynex DS2 | 2 | *Negative | 2 | 100.00% |
| AE-10 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Positive | 1 | 50.00% |
| | | Negative | 1 | 50.00% |
| Dynex DS2 | 2 | *Positive | 1 | 50.00% |
| Dynex DS2 | | Negative | 1 | 50.00% |
| 671 - Total IgE;674 - Diagnostic Allergy;765 - Allergen Testing (1 specimen) | Labs | Response | Number | Percent |
| | | Aspergillus fumigatus Alrgn | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 5 | | | |
| | | *1 | 4 | 80.00% |
| | | 0 | 1 | 20.00% |
| | | Bermuda Grass Alrgn | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 7 | | | |
| | | *1 | 3 | 42.86% |
| | | *2 | 3 | 42.86% |
| | | 0 | 1 | 14.29% |
| | | Cat Epithelium Alrgn | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 5 | | | |
| | | *4 | 5 | 100.00% |
| | | Common Ragweed Alrgn | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 1 | | | |



| | | | | |
|-------------------|---|--|---|---------|
| | | *3 | 1 | 100.00% |
| | | Dog Dander Alrgn | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 5 | | | |
| | | *4 | 5 | 100.00% |
| | | Dog Epithelium Alrgn | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 3 | | | |
| | | *4 | 3 | 100.00% |
| | | English Plantain Alrgn | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 4 | | | |
| | | *0 | 1 | 25.00% |
| | | *1 | 2 | 50.00% |
| | | *1/0 | 1 | 25.00% |
| | | House Dust Mite Alrgn | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 6 | | | |
| | | *2 | 1 | 16.67% |
| | | *3 | 3 | 50.00% |
| | | *4 | 2 | 33.33% |
| | | Oak Alrgn | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 4 | | | |
| | | *2 | 2 | 50.00% |
| | | *3 | 2 | 50.00% |
| | | Peanut (Arachis hypogaea) Alrgn | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 5 | | | |
| | | *4 | 5 | 100.00% |
| | | Soybean Alrgn | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 5 | | | |
| | | *2 | 2 | 40.00% |
| | | *3 | 1 | 20.00% |
| | | *4 | 2 | 40.00% |
| | | Timothy Grass Alrgn | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 6 | | | |
| | | *4 | 6 | 100.00% |
| | | Alternaria alternata Alrgn | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 7 | | | |
| | | *1 | 5 | 71.43% |
| | | *1/0 | 1 | 14.29% |
| | | 0 | 1 | 14.29% |
| | | Almond Alrgn | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 3 | | | |
| | | *1 | 1 | 33.33% |
| | | *1/0 | 1 | 33.33% |
| | | *2 | 1 | 33.33% |
| | | Apple Alrgn | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 3 | | | |
| | | *1 | 1 | 33.33% |
| | | *1/0 | 2 | 66.67% |
| | | Ash - White Alrgn | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 4 | | | |
| | | *0 | 2 | 50.00% |
| | | *1/0 | 2 | 50.00% |
| | | Barley Alrgn | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 3 | | | |
| | | *1/0 | 3 | 100.00% |
| | | Elm - White Alrgn | | |



| AL-6(Q) | | * Denotes Correct Response | | |
|-----------------------------------|---|----------------------------|---|---------|
| PG: Hitachi CLA-1 | 5 | | | |
| | | *0 | 3 | 60.00% |
| | | *1/0 | 2 | 40.00% |
| Bean - White Alrgn | | | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 3 | | | |
| | | *1 | 1 | 33.33% |
| | | *1/0 | 2 | 66.67% |
| Birch/Alder Mix Alrgn | | | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 5 | | | |
| | | *1 | 1 | 20.00% |
| | | *1/0 | 1 | 20.00% |
| | | *2 | 3 | 60.00% |
| Cedar - Mountain Alrgn | | | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 4 | | | |
| | | *0 | 2 | 50.00% |
| | | *1 | 1 | 25.00% |
| | | *1/0 | 1 | 25.00% |
| Chicken Alrgn | | | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 3 | | | |
| | | *0 | 2 | 66.67% |
| | | *1/0 | 1 | 33.33% |
| Chocolate Alrgn | | | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 3 | | | |
| | | *0 | 2 | 66.67% |
| | | *1/0 | 1 | 33.33% |
| Cladosporium Alrgn | | | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 5 | | | |
| | | *0 | 1 | 20.00% |
| | | *1 | 1 | 20.00% |
| | | *1/0 | 3 | 60.00% |
| Cocklebur Alrgn | | | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 5 | | | |
| | | *0 | 3 | 60.00% |
| | | *1/0 | 2 | 40.00% |
| Oat Alrgn | | | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 3 | | | |
| | | *1 | 3 | 100.00% |
| Corn Alrgn | | | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 4 | | | |
| | | *1 | 2 | 50.00% |
| | | *1/0 | 2 | 50.00% |
| Cottonwood - Eastern Alrgn | | | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 4 | | | |
| | | *0 | 3 | 75.00% |
| | | *1/0 | 1 | 25.00% |
| Egg - whole Alrgn | | | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 4 | | | |
| | | *2 | 3 | 75.00% |
| | | *4 | 1 | 25.00% |
| Garlic Alrgn | | | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 3 | | | |
| | | *1 | 1 | 33.33% |
| | | *1/0 | 2 | 66.67% |
| Lambs Quarters Alrgn | | | | |



| | | | | |
|-------------------|---|----------------------------|---|---------|
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 5 | | | |
| | | *0 | 1 | 20.00% |
| | | *1 | 2 | 40.00% |
| | | *1/0 | 2 | 40.00% |
| AL-6(Q) | | Marshelder - Rough Alrqn | | |
| PG: Hitachi CLA-1 | 4 | * Denotes Correct Response | | |
| | | *0 | 1 | 25.00% |
| | | *1 | 2 | 50.00% |
| | | *1/0 | 1 | 25.00% |
| AL-6(Q) | | Milk Alrqn | | |
| PG: Hitachi CLA-1 | 4 | * Denotes Correct Response | | |
| | | *2 | 1 | 25.00% |
| | | *3 | 2 | 50.00% |
| | | *4 | 1 | 25.00% |
| AL-6(Q) | | Mite D Farinae Alrqn | | |
| PG: Hitachi CLA-1 | 6 | * Denotes Correct Response | | |
| | | *4 | 6 | 100.00% |
| AL-6(Q) | | Mulberry Mix Alrqn | | |
| PG: Hitachi CLA-1 | 5 | * Denotes Correct Response | | |
| | | *0 | 4 | 80.00% |
| | | *1 | 1 | 20.00% |
| AL-6(Q) | | Onion Mix Alrqn | | |
| PG: Hitachi CLA-1 | 3 | * Denotes Correct Response | | |
| | | *1 | 1 | 33.33% |
| | | *1/0 | 2 | 66.67% |
| AL-6(Q) | | Orange Alrqn | | |
| PG: Hitachi CLA-1 | 4 | * Denotes Correct Response | | |
| | | *1 | 1 | 25.00% |
| | | *2 | 2 | 50.00% |
| | | *3 | 1 | 25.00% |
| AL-6(Q) | | Pigweed Alrqn | | |
| PG: Hitachi CLA-1 | 5 | * Denotes Correct Response | | |
| | | *0 | 1 | 20.00% |
| | | *1 | 1 | 20.00% |
| | | *1/0 | 3 | 60.00% |
| AL-6(Q) | | Pine Mix Alrqn | | |
| PG: Hitachi CLA-1 | 5 | * Denotes Correct Response | | |
| | | *0 | 5 | 100.00% |
| AL-6(Q) | | Pork Alrqn | | |
| PG: Hitachi CLA-1 | 3 | * Denotes Correct Response | | |
| | | *1 | 1 | 33.33% |
| | | *3 | 2 | 66.67% |
| AL-6(Q) | | Potato Alrqn | | |
| PG: Hitachi CLA-1 | 3 | * Denotes Correct Response | | |
| | | *1 | 1 | 33.33% |
| | | *2 | 2 | 66.67% |
| AL-6(Q) | | Privet Alrqn | | |
| PG: Hitachi CLA-1 | 5 | * Denotes Correct Response | | |
| | | *0 | 5 | 100.00% |
| AL-6(Q) | | Ragweed - Short Alrqn | | |
| PG: Hitachi CLA-1 | 5 | * Denotes Correct Response | | |
| | | *1 | 2 | 40.00% |
| | | *2 | 2 | 40.00% |
| | | 0 | 1 | 20.00% |



| AL-6(Q) | | Rice Alrgn | | |
|-------------------|---|--------------------------------|---|---------|
| PG: Hitachi CLA-1 | | * Denotes Correct Response | | |
| | 4 | | | |
| | | *1 | 3 | 75.00% |
| | | *4 | 1 | 25.00% |
| AL-6(Q) | | Sheep Sorrel Alrgn | | |
| PG: Hitachi CLA-1 | | * Denotes Correct Response | | |
| | 5 | | | |
| | | *2 | 1 | 20.00% |
| | | *3 | 3 | 60.00% |
| | | 1/0 | 1 | 20.00% |
| AL-6(Q) | | Shellfish Mix Alrgn | | |
| PG: Hitachi CLA-1 | | * Denotes Correct Response | | |
| | 4 | | | |
| | | *3 | 1 | 25.00% |
| | | *4 | 3 | 75.00% |
| AL-6(Q) | | Sycamore - American Alrgn | | |
| PG: Hitachi CLA-1 | | * Denotes Correct Response | | |
| | 5 | | | |
| | | *0 | 4 | 80.00% |
| | | *1/0 | 1 | 20.00% |
| AL-6(Q) | | Tomato Alrgn | | |
| PG: Hitachi CLA-1 | | * Denotes Correct Response | | |
| | 4 | | | |
| | | *1 | 1 | 25.00% |
| | | *1/0 | 2 | 50.00% |
| | | *2 | 1 | 25.00% |
| AL-6(Q) | | Tuna Alrgn | | |
| PG: Hitachi CLA-1 | | * Denotes Correct Response | | |
| | 3 | | | |
| | | *0 | 1 | 33.33% |
| | | *1/0 | 2 | 66.67% |
| AL-6(Q) | | Vegetable Mix Alrgn | | |
| PG: Hitachi CLA-1 | | * Denotes Correct Response | | |
| | 3 | | | |
| | | *1 | 2 | 66.67% |
| | | *1/0 | 1 | 33.33% |
| AL-6(Q) | | Walnut/Hickory/Pecan Mix Alrgn | | |
| PG: Hitachi CLA-1 | | * Denotes Correct Response | | |
| | 5 | | | |
| | | *0 | 1 | 20.00% |
| | | *1 | 1 | 20.00% |
| | | *1/0 | 3 | 60.00% |
| AL-6(Q) | | Waterhemp Alrgn | | |
| PG: Hitachi CLA-1 | | * Denotes Correct Response | | |
| | 3 | | | |
| | | *2 | 1 | 33.33% |
| | | *3 | 2 | 66.67% |
| AL-6(Q) | | Wheat Alrgn | | |
| PG: Hitachi CLA-1 | | * Denotes Correct Response | | |
| | 3 | | | |
| | | *3 | 2 | 66.67% |
| | | *4 | 1 | 33.33% |
| AL-6(Q) | | Acacia Alrgn | | |
| PG: Hitachi CLA-1 | | * Denotes Correct Response | | |
| | 2 | | | |
| | | *1/0 | 2 | 100.00% |
| AL-6(Q) | | Bahia Grass Alrgn | | |
| PG: Hitachi CLA-1 | | * Denotes Correct Response | | |
| | 2 | | | |
| | | *2 | 1 | 50.00% |
| | | *3 | 1 | 50.00% |
| AL-6(Q) | | Beech, American Alrgn | | |
| PG: Hitachi CLA-1 | | * Denotes Correct Response | | |
| | 2 | | | |
| | | *0 | 2 | 100.00% |
| AL-6(Q) | | Candida Alrgn | | |



| AL-6(Q) | | * Denotes Correct Response | | |
|--|------|-------------------------------------|--------|---------|
| PG: Hitachi CLA-1 | 2 | | | |
| | | *1/0 | 2 | 100.00% |
| | | Cat Dander Alrgn | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 2 | | | |
| | | *4 | 2 | 100.00% |
| | | Melaleuca Alrgn | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 2 | | | |
| | | *1/0 | 2 | 100.00% |
| | | Oak, White Alrgn | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 2 | | | |
| | | *3 | 2 | 100.00% |
| | | Penicillium Alrgn | | |
| AL-6(Q) | | * Denotes Correct Response | | |
| PG: Hitachi CLA-1 | 2 | | | |
| | | *1 | 1 | 50.00% |
| | | *1/0 | 1 | 50.00% |
| 683 - ASO/CRP;759 - Complete Immunology;729 - AFC Immunology | Labs | Response | Number | Percent |
| | | ASO (Qual) (International) | | |
| AS-6 | | * Denotes Correct Response | | |
| PG: Wampole ColorCard ASO | 1 | | | |
| | | *Negative | 1 | 100.00% |
| AS-7 | | * Denotes Correct Response | | |
| PG: Wampole ColorCard ASO | 1 | | | |
| | | *Positive | 1 | 100.00% |
| AS-8 | | * Denotes Correct Response | | |
| PG: Wampole ColorCard ASO | 1 | | | |
| | | *Negative | 1 | 100.00% |
| AS-9 | | * Denotes Correct Response | | |
| PG: Wampole ColorCard ASO | 1 | | | |
| | | *Negative | 1 | 100.00% |
| AS-10 | | * Denotes Correct Response | | |
| PG: Wampole ColorCard ASO | 1 | | | |
| | | *Positive | 1 | 100.00% |
| 782 - Complete Bacteriology | Labs | Response | Number | Percent |
| | | Ear Culture | | |
| BA-4 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Haemophilus influenzae | 1 | 50.00% |
| | | *Haemophilus species | 1 | 50.00% |
| | | *Staphylococcus epidermidis | 1 | 50.00% |
| | | *Staphylococcus, coagulase negative | 1 | 50.00% |
| Cultures | 2 | | | |
| | | *Haemophilus influenzae | 1 | 50.00% |
| Cultures | | *Haemophilus species | 1 | 50.00% |
| Cultures | | *Staphylococcus epidermidis | 1 | 50.00% |
| Cultures | | *Staphylococcus, coagulase negative | 1 | 50.00% |
| | | CSF Culture | | |
| BA-6 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Streptococcus, Group B | 2 | 100.00% |
| Cultures | 2 | | | |
| | | *Streptococcus, Group B | 2 | 100.00% |
| 650 - Blood Bank;736 - ABO/Rh Only | Labs | Response | Number | Percent |
| | | ABO Group | | |
| BB-6 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Group AB | 1 | 100.00% |
| Blood Bank Testing | 1 | | | |
| | | *Group AB | 1 | 100.00% |
| BB-7 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Group A | 1 | 100.00% |
| Blood Bank Testing | 1 | | | |
| | | *Group A | 1 | 100.00% |
| BB-8 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |



| | | | | |
|--|-------------|---------------------------------------|---------------|----------------|
| | | *Group AB | 1 | 100.00% |
| Blood Bank Testing | 1 | *Group AB | 1 | 100.00% |
| BB-9 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Group A | 1 | 100.00% |
| Blood Bank Testing | 1 | *Group A | 1 | 100.00% |
| BB-10 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Group O | 1 | 100.00% |
| Blood Bank Testing | 1 | *Group O | 1 | 100.00% |
| | | Rh Type | | |
| BB-6 | | * Denotes Correct Response | | |
| AG: All Methods | 6 | | | |
| | | *Rh Positive | 6 | 100.00% |
| Blood Bank Testing | 6 | *Rh Positive | 6 | 100.00% |
| BB-7 | | * Denotes Correct Response | | |
| AG: All Methods | 6 | | | |
| | | *Rh Negative | 6 | 100.00% |
| Blood Bank Testing | 6 | *Rh Negative | 6 | 100.00% |
| BB-8 | | * Denotes Correct Response | | |
| AG: All Methods | 6 | | | |
| | | *Rh Positive | 6 | 100.00% |
| Blood Bank Testing | 6 | *Rh Positive | 6 | 100.00% |
| BB-9 | | * Denotes Correct Response | | |
| AG: All Methods | 6 | | | |
| | | *Rh Positive | 6 | 100.00% |
| Blood Bank Testing | 6 | *Rh Positive | 6 | 100.00% |
| BB-10 | | * Denotes Correct Response | | |
| AG: All Methods | 6 | | | |
| | | *Rh Positive | 6 | 100.00% |
| Blood Bank Testing | 6 | *Rh Positive | 6 | 100.00% |
| 975 - Bacterial Vaginosis | Labs | Response | Number | Percent |
| | | Bacterial Vaginosa (Sialidase) | | |
| BV-3 | | * Denotes Correct Response | | |
| AG: All Methods | 84 | | | |
| | | *Positive | 74 | 88.10% |
| | | Negative | 10 | 11.90% |
| Sekisui Diagnostics OSOM BVBlue | 84 | *Positive | 74 | 88.10% |
| Sekisui Diagnostics OSOM BVBlue | | Negative | 10 | 11.90% |
| BV-4 | | * Denotes Correct Response | | |
| AG: All Methods | 84 | | | |
| | | *Negative | 79 | 94.05% |
| | | Positive | 5 | 5.95% |
| Sekisui Diagnostics OSOM BVBlue | 84 | *Negative | 79 | 94.05% |
| Sekisui Diagnostics OSOM BVBlue | | Positive | 5 | 5.95% |
| 661 - Urine Colony Count;784 - Presump ID/Colony Count - Urine;785 - Presump ID/Colony Count-Urine & Strep Combo;775 - Urine Culture;778 - Urine & Throat Cultures;779 - Urine & Genital Cultures;781 - Urine, Throat & Genital Cultures;782 - Complete Bacteri | Labs | Response | Number | Percent |
| | | Colony Count | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 32 | | | |
| | | *10,000 - 100,000 | 15 | 46.88% |
| | | *Greater than 100,000 | 17 | 53.13% |
| Calibrated Loop - Manual | 9 | *10,000 - 100,000 | 4 | 44.44% |
| Uri-Chek (Troy Bio) | 5 | *10,000 - 100,000 | 1 | 20.00% |
| Uricult (Orion) | 16 | *10,000 - 100,000 | 9 | 56.25% |
| Uricult (Lifesign) | 1 | *10,000 - 100,000 | 1 | 100.00% |
| Calibrated Loop - Manual | 9 | *Greater than 100,000 | 5 | 55.56% |
| Uri-Chek (Troy Bio) | 5 | *Greater than 100,000 | 4 | 80.00% |
| Uricult (Orion) | 16 | *Greater than 100,000 | 7 | 43.75% |
| Uri-Three (Culture Kits | 1 | *Greater than 100,000 | 1 | 100.00% |
| 784 - Presump ID/Colony Count - Urine;785 - Presump ID/Colony Count-Urine & Strep Combo | Labs | Response | Number | Percent |
| | | Colony Count | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 32 | | | |
| | | *10,000 - 100,000 | 15 | 46.88% |



| | | | | |
|--------------------------|----|-----------------------------------|----|---------|
| | | *Greater than 100,000 | 17 | 53.13% |
| Calibrated Loop - Manual | 9 | *10,000 - 100,000 | 4 | 44.44% |
| Uri-Chek (Troy Bio) | 5 | *10,000 - 100,000 | 1 | 20.00% |
| Uricult (Orion) | 16 | *10,000 - 100,000 | 9 | 56.25% |
| Uricult (Lifesign) | 1 | *10,000 - 100,000 | 1 | 100.00% |
| Calibrated Loop - Manual | 9 | *Greater than 100,000 | 5 | 55.56% |
| Uri-Chek (Troy Bio) | 5 | *Greater than 100,000 | 4 | 80.00% |
| Uricult (Orion) | 16 | *Greater than 100,000 | 7 | 43.75% |
| Uri-Three (Culture Kits) | 1 | *Greater than 100,000 | 1 | 100.00% |
| CC-7 | | * Denotes Correct Response | | |
| AG: All Methods | 6 | | | |
| | | *Greater than 100,000 | 6 | 100.00% |
| Calibrated Loop - Manual | 4 | *Greater than 100,000 | 4 | 100.00% |
| Uri-Chek (Troy Bio) | 1 | *Greater than 100,000 | 1 | 100.00% |
| Uricult (Orion) | 1 | *Greater than 100,000 | 1 | 100.00% |
| CC-8 | | * Denotes Correct Response | | |
| AG: All Methods | 6 | | | |
| | | *Greater than 100,000 | 6 | 100.00% |
| Calibrated Loop - Manual | 4 | *Greater than 100,000 | 4 | 100.00% |
| Uri-Chek (Troy Bio) | 1 | *Greater than 100,000 | 1 | 100.00% |
| Uricult (Orion) | 1 | *Greater than 100,000 | 1 | 100.00% |
| CC-9 | | * Denotes Correct Response | | |
| AG: All Methods | 6 | | | |
| | | *No growth | 6 | 100.00% |
| Calibrated Loop - Manual | 4 | *No growth | 4 | 100.00% |
| Uri-Chek (Troy Bio) | 1 | *No growth | 1 | 100.00% |
| Uricult (Orion) | 1 | *No growth | 1 | 100.00% |
| CC-10 | | * Denotes Correct Response | | |
| AG: All Methods | 6 | | | |
| | | *Greater than 100,000 | 6 | 100.00% |
| Calibrated Loop - Manual | 4 | *Greater than 100,000 | 4 | 100.00% |
| Uri-Chek (Troy Bio) | 1 | *Greater than 100,000 | 1 | 100.00% |
| Uricult (Orion) | 1 | *Greater than 100,000 | 1 | 100.00% |
| | | Urine Presump Id | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Presumptive E. Coli | 3 | 100.00% |
| Calibrated Loop - Manual | 2 | *Presumptive E. Coli | 2 | 100.00% |
| Uricult (Orion) | 1 | *Presumptive E. Coli | 1 | 100.00% |
| CC-7 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Gram Positive Organism | 1 | 33.33% |
| | | *Presumptive Enterococcus sp. | 1 | 33.33% |
| | | Presumptive Proteus sp. | 1 | 33.33% |
| Uricult (Orion) | 1 | *Gram Positive Organism | 1 | 100.00% |
| Calibrated Loop - Manual | 2 | *Presumptive Enterococcus sp. | 1 | 50.00% |
| Calibrated Loop - Manual | | Presumptive Proteus sp. | 1 | 50.00% |
| CC-8 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Presumptive Klebsiella sp. | 3 | 100.00% |
| Calibrated Loop - Manual | 2 | *Presumptive Klebsiella sp. | 2 | 100.00% |
| Uricult (Orion) | 1 | *Presumptive Klebsiella sp. | 1 | 100.00% |
| CC-9 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *No Growth | 3 | 100.00% |
| Calibrated Loop - Manual | 2 | *No Growth | 2 | 100.00% |
| Uricult (Orion) | 1 | *No Growth | 1 | 100.00% |
| CC-10 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Gram Positive Organism | 1 | 33.33% |
| | | *Presumptive Staphylococcus sp. | 1 | 33.33% |
| | | Presumptive Proteus sp. | 1 | 33.33% |
| Uricult (Orion) | 1 | *Gram Positive Organism | 1 | 100.00% |
| Calibrated Loop - Manual | 2 | *Presumptive Staphylococcus sp. | 1 | 50.00% |
| Calibrated Loop - Manual | | Presumptive Proteus sp. | 1 | 50.00% |
| | | Susceptibility Amikacin | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |



| | | | | |
|-----------------|----|-------------------------------------|---|---------|
| | | *Sensitive | 2 | 100.00% |
| Susceptibility | 2 | *Sensitive | 2 | 100.00% |
| | | Susceptibility Amox/Clavulanate | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 7 | | | |
| | | *Sensitive | 7 | 100.00% |
| Susceptibility | 7 | *Sensitive | 7 | 100.00% |
| | | Susceptibility Ampicillin | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 9 | | | |
| | | *Resistant | 9 | 100.00% |
| Susceptibility | 9 | *Resistant | 9 | 100.00% |
| | | Susceptibility Ampicillin/Sulbactam | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Intermediate | 3 | 100.00% |
| Susceptibility | 3 | *Intermediate | 3 | 100.00% |
| | | Susceptibility Aztreonam | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Sensitive | 1 | 100.00% |
| Susceptibility | 1 | *Sensitive | 1 | 100.00% |
| | | Susceptibility Cefazolin | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Sensitive | 3 | 100.00% |
| Susceptibility | 3 | *Sensitive | 3 | 100.00% |
| | | Susceptibility Cefepime | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Sensitive | 3 | 100.00% |
| Susceptibility | 3 | *Sensitive | 3 | 100.00% |
| | | Susceptibility Cefotaxime | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Sensitive | 3 | 100.00% |
| Susceptibility | 3 | *Sensitive | 3 | 100.00% |
| | | Susceptibility Cefpodoxime | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Sensitive | 1 | 100.00% |
| Susceptibility | 1 | *Sensitive | 1 | 100.00% |
| | | Susceptibility Ceftazidime | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Sensitive | 3 | 100.00% |
| Susceptibility | 3 | *Sensitive | 3 | 100.00% |
| | | Susceptibility Ceftriaxone | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 6 | | | |
| | | *Sensitive | 6 | 100.00% |
| Susceptibility | 6 | *Sensitive | 6 | 100.00% |
| | | Susceptibility Cefuroxime | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 5 | | | |
| | | *Sensitive | 5 | 100.00% |
| Susceptibility | 5 | *Sensitive | 5 | 100.00% |
| | | Susceptibility Cephalothin | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 7 | | | |
| | | *Intermediate | 3 | 42.86% |
| | | *Sensitive | 4 | 57.14% |
| Susceptibility | 7 | *Intermediate | 3 | 42.86% |
| Susceptibility | | *Sensitive | 4 | 57.14% |
| | | Susceptibility Ciprofloxacin | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 11 | | | |



| | | | | |
|-----------------|----|---|----|---------|
| | | *Sensitive | 11 | 100.00% |
| Susceptibility | 11 | *Sensitive | 11 | 100.00% |
| | | Susceptibility Gentamicin | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 6 | | | |
| | | *Sensitive | 6 | 100.00% |
| Susceptibility | 6 | *Sensitive | 6 | 100.00% |
| | | Susceptibility Imipenem | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 4 | | | |
| | | *Sensitive | 4 | 100.00% |
| Susceptibility | 4 | *Sensitive | 4 | 100.00% |
| | | Susceptibility Levofloxacin | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 6 | | | |
| | | *Sensitive | 6 | 100.00% |
| Susceptibility | 6 | *Sensitive | 6 | 100.00% |
| | | Susceptibility Nitrofurantoin | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 10 | | | |
| | | *Sensitive | 10 | 100.00% |
| Susceptibility | 10 | *Sensitive | 10 | 100.00% |
| | | Susceptibility Norfloxacin/Noroxin | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Sensitive | 2 | 100.00% |
| Susceptibility | 2 | *Sensitive | 2 | 100.00% |
| | | Susceptibility Piperacillin/Tazobactam | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Sensitive | 3 | 100.00% |
| Susceptibility | 3 | *Sensitive | 3 | 100.00% |
| | | Susceptibility Sulfamethoxazole | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Sensitive | 1 | 100.00% |
| Susceptibility | 1 | *Sensitive | 1 | 100.00% |
| | | Susceptibility Tetracycline | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 6 | | | |
| | | *Sensitive | 6 | 100.00% |
| Susceptibility | 6 | *Sensitive | 6 | 100.00% |
| | | Susceptibility Ticarcillin/Clavulanate | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Sensitive | 1 | 100.00% |
| Susceptibility | 1 | *Sensitive | 1 | 100.00% |
| | | Susceptibility Tobramycin | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 4 | | | |
| | | *Sensitive | 4 | 100.00% |
| Susceptibility | 4 | *Sensitive | 4 | 100.00% |
| | | Susceptibility Trimethoprim | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Sensitive | 2 | 100.00% |
| Susceptibility | 2 | *Sensitive | 2 | 100.00% |
| | | Susceptibility Trimeth./Sulfameth. | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 9 | | | |
| | | *Sensitive | 9 | 100.00% |
| Susceptibility | 9 | *Sensitive | 9 | 100.00% |
| | | Susceptibility Ertapenem | | |
| CC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Sensitive | 1 | 100.00% |
| Susceptibility | 1 | *Sensitive | 1 | 100.00% |



| 714 - Complete Cardiac Markers;715 - Basic Cardiac Markers;720 - Cardiac Markers I;740 - D-Dimer | Labs | Response | Number | Percent |
|---|-------------|---------------------------------|---------------|----------------|
| | | Myoglobin (Qualitative) | | |
| CDM-6 | | * Denotes Correct Response | | |
| PG: Princeton Biomedical Corp. Lifesign MI | 2 | | | |
| | | *Positive | 2 | 100.00% |
| CDM-7 | | * Denotes Correct Response | | |
| PG: Princeton Biomedical Corp. Lifesign MI | 2 | | | |
| | | *Negative | 1 | 50.00% |
| | | Positive | 1 | 50.00% |
| CDM-8 | | * Denotes Correct Response | | |
| PG: Princeton Biomedical Corp. Lifesign MI | 2 | | | |
| | | *Positive | 2 | 100.00% |
| CDM-9 | | * Denotes Correct Response | | |
| PG: Princeton Biomedical Corp. Lifesign MI | 2 | | | |
| | | *Positive | 2 | 100.00% |
| CDM-10 | | * Denotes Correct Response | | |
| PG: Princeton Biomedical Corp. Lifesign MI | 2 | | | |
| | | *Positive | 2 | 100.00% |
| | | CK-MB (Qualitative) | | |
| CDM-6 | | * Denotes Correct Response | | |
| PG: Princeton Biomedical Corp. Lifesign MI | 2 | | | |
| | | *Positive | 2 | 100.00% |
| CDM-7 | | * Denotes Correct Response | | |
| PG: Princeton Biomedical Corp. Lifesign MI | 2 | | | |
| | | *Negative | 1 | 50.00% |
| | | Positive | 1 | 50.00% |
| CDM-8 | | * Denotes Correct Response | | |
| PG: Princeton Biomedical Corp. Lifesign MI | 2 | | | |
| | | *Positive | 2 | 100.00% |
| CDM-9 | | * Denotes Correct Response | | |
| PG: Princeton Biomedical Corp. Lifesign MI | 2 | | | |
| | | *Positive | 2 | 100.00% |
| CDM-10 | | * Denotes Correct Response | | |
| PG: Princeton Biomedical Corp. Lifesign MI | 2 | | | |
| | | *Positive | 2 | 100.00% |
| | | Troponin I (Qualitative) | | |
| CDM-6 | | * Denotes Correct Response | | |
| PG: Princeton Biomedical Corp. Lifesign MI | 2 | | | |
| | | *Positive | 2 | 100.00% |
| CDM-7 | | * Denotes Correct Response | | |
| PG: Princeton Biomedical Corp. Lifesign MI | 2 | | | |
| | | *Negative | 2 | 100.00% |
| CDM-8 | | * Denotes Correct Response | | |
| PG: Princeton Biomedical Corp. Lifesign MI | 2 | | | |
| | | *Positive | 2 | 100.00% |
| CDM-9 | | * Denotes Correct Response | | |
| PG: Princeton Biomedical Corp. Lifesign MI | 3 | | | |
| | | *Negative | 1 | 33.33% |
| | | Positive | 2 | 66.67% |
| CDM-10 | | * Denotes Correct Response | | |
| PG: Princeton Biomedical Corp. Lifesign MI | 2 | | | |
| | | *Positive | 2 | 100.00% |
| | | Response | Number | Percent |
| 612 - MultiChem 5;616 - MultiChem 10;617 - MultiChem 15;618 - MultiChem 20;619 - MultiChem 21+;676 - Endocrinology | Labs | | | |
| | | TSH (Semi-Quant) | | |
| CH-6 | | * Denotes Correct Response | | |
| PG: Biomedical Thyro Test | 1 | | | |
| | | *≤ 5 | 1 | 100.00% |
| CH-7 | | * Denotes Correct Response | | |
| PG: Biomedical Thyro Test | 1 | | | |
| | | *≤ 5 | 1 | 100.00% |
| CH-8 | | * Denotes Correct Response | | |
| PG: Biomedical Thyro Test | 1 | | | |
| | | *≤ 5 | 1 | 100.00% |
| CH-9 | | * Denotes Correct Response | | |



| PG: Biomedical Thyro Test | | 1 | | |
|--|-----|------|---|----------------|
| | | | * <= 5 | 1 100.00% |
| CH-10 | | | * Denotes Correct Response | |
| PG: Biomedical Thyro Test | | 1 | | |
| | | | * <= 5 | 1 100.00% |
| 639 - Complete Urinalysis Package;638 - Urine Sediment;641 - Clinical Microscopy;601 - Waived/PPM Tests II;609 - Waived/PPM Tests I;646 - Clinical Microscopy Extra;628 - AFC Waived/PPM | | Labs | Response | Number Percent |
| CM-10 | | | Urine Sediment Id | |
| | | | * Denotes Correct Response | |
| AG: All Methods | | 709 | | |
| | | | *Ident Unknown - Referred | 1 0.14% |
| | | | *Yeast/fungi | 636 89.70% |
| | | | Ammonium magnesium phosphate (Triple phosphate) | 1 0.14% |
| | | | Bacteria | 2 0.28% |
| | | | Calcium carbonate | 8 1.13% |
| | | | Cellular (RTE) cast | 1 0.14% |
| | | | Cellular (WBC) cast | 2 0.28% |
| | | | Fat droplets or globules | 15 2.12% |
| | | | Fatty (fat globules) cast | 6 0.85% |
| | | | Fiber (exogenous)/fecal contamination | 1 0.14% |
| | | | Granular cast | 1 0.14% |
| | | | RBC/Blood/Hgb cast | 2 0.28% |
| | | | Red blood cell (RBC) (Erythrocyte) | 7 0.99% |
| | | | Renal tubular epithelial (RTE) | 4 0.56% |
| | | | RTE with fat globules (oval fat bodies) | 11 1.55% |
| | | | Squamous epithelial cell | 1 0.14% |
| | | | Starch granules | 2 0.28% |
| | | | Transitional epithelial cell (urothelial) | 1 0.14% |
| | | | Trichomonas vaginalis | 4 0.56% |
| | | | White blood cell (WBC) (Leukocyte) | 3 0.42% |
| Microscopy | 709 | | *Ident Unknown - Referred | 1 0.14% |
| Microscopy | | | *Yeast/fungi | 636 89.70% |
| Microscopy | | | Ammonium magnesium phosphate (Triple phosphate) | 1 0.14% |
| Microscopy | | | Bacteria | 2 0.28% |
| Microscopy | | | Calcium carbonate | 8 1.13% |
| Microscopy | | | Cellular (RTE) cast | 1 0.14% |
| Microscopy | | | Cellular (WBC) cast | 2 0.28% |
| Microscopy | | | Fat droplets or globules | 15 2.12% |
| Microscopy | | | Fatty (fat globules) cast | 6 0.85% |
| Microscopy | | | Fiber (exogenous)/fecal contamination | 1 0.14% |
| Microscopy | | | Granular cast | 1 0.14% |
| Microscopy | | | RBC/Blood/Hgb cast | 2 0.28% |
| Microscopy | | | Red blood cell (RBC) (Erythrocyte) | 7 0.99% |
| Microscopy | | | Renal tubular epithelial (RTE) | 4 0.56% |
| Microscopy | | | RTE with fat globules (oval fat bodies) | 11 1.55% |
| Microscopy | | | Squamous epithelial cell | 1 0.14% |
| Microscopy | | | Starch granules | 2 0.28% |
| Microscopy | | | Transitional epithelial cell (urothelial) | 1 0.14% |
| Microscopy | | | Trichomonas vaginalis | 4 0.56% |
| Microscopy | | | White blood cell (WBC) (Leukocyte) | 3 0.42% |
| CM-11 | | | * Denotes Correct Response | |
| AG: All Methods | | 709 | | |
| | | | *Squamous epithelial cell | 665 93.79% |
| | | | *Transitional epithelial cell (urothelial) | 36 5.08% |
| | | | Ident Unknown - Referred | 1 0.14% |
| | | | Renal tubular epithelial (RTE) | 4 0.56% |
| | | | White blood cell (WBC) (Leukocyte) | 1 0.14% |
| | | | Yeast/fungi | 2 0.28% |
| Microscopy | 709 | | *Squamous epithelial cell | 665 93.79% |
| Microscopy | | | *Transitional epithelial cell (urothelial) | 36 5.08% |
| Microscopy | | | Ident Unknown - Referred | 1 0.14% |
| Microscopy | | | Renal tubular epithelial (RTE) | 4 0.56% |
| Microscopy | | | White blood cell (WBC) (Leukocyte) | 1 0.14% |
| Microscopy | | | Yeast/fungi | 2 0.28% |
| CM-12 | | | * Denotes Correct Response | |
| AG: All Methods | | 702 | | |
| | | | *Fatty (fat globules) cast | 218 31.05% |



| | | | | |
|-----------------|-----|---|-----|--------|
| | | *Ident Unknown - Referred | 11 | 1.57% |
| | | Amorphous urates/phosphates | 16 | 2.28% |
| | | Bacteria | 3 | 0.43% |
| | | Broad cast | 1 | 0.14% |
| | | Calcium carbonate | 1 | 0.14% |
| | | Cellular (RTE) cast | 7 | 1.00% |
| | | Cellular (WBC) cast | 19 | 2.71% |
| | | Fat droplets or globules | 4 | 0.57% |
| | | Fiber (exogenous)/fecal contamination | 5 | 0.71% |
| | | Granular cast | 288 | 41.03% |
| | | Hyaline cast | 30 | 4.27% |
| | | Mucous strands | 1 | 0.14% |
| | | RBC/Blood/Hgb cast | 85 | 12.11% |
| | | Red blood cell (RBC) (Erythrocyte) | 4 | 0.57% |
| | | Uric acid crystal | 1 | 0.14% |
| | | Waxy cast | 7 | 1.00% |
| | | White blood cell (WBC) (Leukocyte) | 1 | 0.14% |
| Microscopy | 702 | *Fatty (fat globules) cast | 218 | 31.05% |
| Microscopy | | *Ident Unknown - Referred | 11 | 1.57% |
| Microscopy | | Amorphous urates/phosphates | 16 | 2.28% |
| Microscopy | | Bacteria | 3 | 0.43% |
| Microscopy | | Broad cast | 1 | 0.14% |
| Microscopy | | Calcium carbonate | 1 | 0.14% |
| Microscopy | | Cellular (RTE) cast | 7 | 1.00% |
| Microscopy | | Cellular (WBC) cast | 19 | 2.71% |
| Microscopy | | Fat droplets or globules | 4 | 0.57% |
| Microscopy | | Fiber (exogenous)/fecal contamination | 5 | 0.71% |
| Microscopy | | Granular cast | 288 | 41.03% |
| Microscopy | | Hyaline cast | 30 | 4.27% |
| Microscopy | | Mucous strands | 1 | 0.14% |
| Microscopy | | RBC/Blood/Hgb cast | 85 | 12.11% |
| Microscopy | | Red blood cell (RBC) (Erythrocyte) | 4 | 0.57% |
| Microscopy | | Uric acid crystal | 1 | 0.14% |
| Microscopy | | Waxy cast | 7 | 1.00% |
| Microscopy | | White blood cell (WBC) (Leukocyte) | 1 | 0.14% |
| CM-13 | | * Denotes Correct Response | | |
| AG: All Methods | 707 | | | |
| | | *Amorphous urates/phosphates | 439 | 62.09% |
| | | *Ident Unknown - Referred | 55 | 7.78% |
| | | Ammonium biurate crystal | 1 | 0.14% |
| | | Bacteria | 70 | 9.90% |
| | | Broad cast | 3 | 0.42% |
| | | Calcium carbonate | 5 | 0.71% |
| | | Calcium phosphate crystal | 1 | 0.14% |
| | | Cellular (RTE) cast | 1 | 0.14% |
| | | Fat droplets or globules | 24 | 3.39% |
| | | Fatty (fat globules) cast | 8 | 1.13% |
| | | Fiber (exogenous)/fecal contamination | 13 | 1.84% |
| | | Fiber (exogenous)/fecal contamination | 1 | 0.14% |
| | | Granular cast | 15 | 2.12% |
| | | Hyaline cast | 9 | 1.27% |
| | | Mucous strand | 1 | 0.14% |
| | | Mucous strands | 29 | 4.10% |
| | | Pollen grain | 1 | 0.14% |
| | | RBC/Blood/Hgb cast | 3 | 0.42% |
| | | Red blood cell (RBC) (Erythrocyte) | 5 | 0.71% |
| | | Renal tubular epithelial (RTE) | 3 | 0.42% |
| | | RTE with fat globules (oval fat bodies) | 1 | 0.14% |
| | | Squamous epithelial cell | 3 | 0.42% |
| | | Squamous epithelial cell | 1 | 0.14% |
| | | Starch granules | 6 | 0.85% |
| | | Transitional epithelial cell (urothelial) | 2 | 0.28% |
| | | Uric acid crystal | 1 | 0.14% |
| | | Waxy cast | 2 | 0.28% |
| | | White blood cell (WBC) (Leukocyte) | 3 | 0.42% |
| | | Yeast/fungi | 1 | 0.14% |
| Microscopy | 707 | *Amorphous urates/phosphates | 439 | 62.09% |
| Microscopy | | *Ident Unknown - Referred | 55 | 7.78% |
| Microscopy | | Ammonium biurate crystal | 1 | 0.14% |



| | | | | |
|------------------------|-----|---|-----|--------|
| Microscopy | | Bacteria | 70 | 9.90% |
| Microscopy | | Broad cast | 3 | 0.42% |
| Microscopy | | Calcium carbonate | 5 | 0.71% |
| Microscopy | | Calcium phosphate crystal | 1 | 0.14% |
| Microscopy | | Cellular (RTE) cast | 1 | 0.14% |
| Microscopy | | Fat droplets or globules | 24 | 3.39% |
| Microscopy | | Fatty (fat globules) cast | 8 | 1.13% |
| Microscopy | | Fiber (exogenous)/fecal contamination | 13 | 1.84% |
| Microscopy | | Fiber (exogenous)/fecal contamination | 1 | 0.14% |
| Microscopy | | Granular cast | 15 | 2.12% |
| Microscopy | | Hyaline cast | 9 | 1.27% |
| Microscopy | | Mucous strand | 1 | 0.14% |
| Microscopy | | Mucous strands | 29 | 4.10% |
| Microscopy | | Pollen grain | 1 | 0.14% |
| Microscopy | | RBC/Blood/Hgb cast | 3 | 0.42% |
| Microscopy | | Red blood cell (RBC) (Erythrocyte) | 5 | 0.71% |
| Microscopy | | Renal tubular epithelial (RTE) | 3 | 0.42% |
| Microscopy | | RTE with fat globules (oval fat bodies) | 1 | 0.14% |
| Microscopy | | Squamous epithelial cell | 3 | 0.42% |
| Microscopy | | Squamous epithelial cell | 1 | 0.14% |
| Microscopy | | Starch granules | 6 | 0.85% |
| Microscopy | | Transitional epithelial cell (urothelial) | 2 | 0.28% |
| Microscopy | | Uric acid crystal | 1 | 0.14% |
| Microscopy | | Waxy cast | 2 | 0.28% |
| Microscopy | | White blood cell (WBC) (Leukocyte) | 3 | 0.42% |
| Microscopy | | Yeast/fungi | 1 | 0.14% |
| | | Vaginal Wet Prep | | |
| CM-14 | | * Denotes Correct Response | | |
| AG: All Methods | 985 | | | |
| | | *Trichomonas vaginalis | 932 | 94.62% |
| | | Clue cell | 7 | 0.71% |
| | | Red Blood Cell (RBC) (Erythrocyte) | 4 | 0.41% |
| | | Spermatozoa absent | 1 | 0.10% |
| | | Spermatozoa present | 5 | 0.51% |
| | | Squamous epithelial cell | 20 | 2.03% |
| | | White Blood Cell (WBC) (Leukocyte) | 14 | 1.42% |
| | | Yeast/fungi present | 2 | 0.20% |
| Microscopy | 985 | *Trichomonas vaginalis | 932 | 94.62% |
| Microscopy | | Clue cell | 7 | 0.71% |
| Microscopy | | Red Blood Cell (RBC) (Erythrocyte) | 4 | 0.41% |
| Microscopy | | Spermatozoa absent | 1 | 0.10% |
| Microscopy | | Spermatozoa present | 5 | 0.51% |
| Microscopy | | Squamous epithelial cell | 20 | 2.03% |
| Microscopy | | White Blood Cell (WBC) (Leukocyte) | 14 | 1.42% |
| Microscopy | | Yeast/fungi present | 2 | 0.20% |
| | | CM-15 | | |
| AG: All Methods | | * Denotes Correct Response | | |
| | 984 | | | |
| | | *Squamous epithelial cell | 609 | 61.89% |
| | | Bacteria present | 3 | 0.30% |
| | | Clue cell | 337 | 34.25% |
| | | Red Blood Cell (RBC) (Erythrocyte) | 1 | 0.10% |
| | | Renal tubular epithelial (RTE) | 1 | 0.10% |
| | | Squamous epithelial cell | 609 | 61.89% |
| | | Trichomonas vaginalis | 26 | 2.64% |
| | | White Blood Cell (WBC) (Leukocyte) | 7 | 0.71% |
| Microscopy | 984 | *Squamous epithelial cell | 609 | 61.89% |
| Microscopy | | Bacteria present | 3 | 0.30% |
| Microscopy | | Clue cell | 337 | 34.25% |
| Microscopy | | Red Blood Cell (RBC) (Erythrocyte) | 1 | 0.10% |
| Microscopy | | Renal tubular epithelial (RTE) | 1 | 0.10% |
| Microscopy | | Squamous epithelial cell | 609 | 61.89% |
| Microscopy | | Trichomonas vaginalis | 26 | 2.64% |
| Microscopy | | White Blood Cell (WBC) (Leukocyte) | 7 | 0.71% |
| | | Skin Scraping | | |
| CM-16 | | * Denotes Correct Response | | |
| AG: All Methods | 603 | | | |
| | | *Scabies present | 581 | 96.35% |
| | | *Yeast/fungi absent | 12 | 1.99% |
| | | No scabies present | 1 | 0.17% |



| | | | | |
|--|-----|-------------------------------------|---------------|----------------|
| | | Yeast/fungi present | 9 | 1.49% |
| Microscopy | 603 | *Scabies present | 581 | 96.35% |
| Microscopy | | *Yeast/fungi absent | 12 | 1.99% |
| Microscopy | | No scabies present | 1 | 0.17% |
| Microscopy | | Yeast/fungi present | 9 | 1.49% |
| | | Nasal Smear | | |
| CM-17 | | * Denotes Correct Response | | |
| AG: All Methods | 239 | | | |
| | | *Eosinophils are present | 193 | 80.75% |
| | | No eosinophils are present | 46 | 19.25% |
| Microscopy | 239 | *Eosinophils are present | 193 | 80.75% |
| Microscopy | | No eosinophils are present | 46 | 19.25% |
| 683 - ASO/CRP;684 - CRP (Quant & Qual) | | Response | Number | Percent |
| | | CRP Qualitative | | |
| 774 - GenProbe GC/Chlamydia;794 - GC & Chlamydia Antigen Detection (5 specimens);795 - GC & Chlamydia Antigen Detection (2 specimens) | | Response | Number | Percent |
| | | Chlamydia - Non Culture Tech | | |
| CY-6 | | * Denotes Correct Response | | |
| AG: All Methods | 7 | | | |
| | | *Negative | 7 | 100.00% |
| BD Probe Tec ET | 1 | *Negative | 1 | 100.00% |
| Gen-Probe | 1 | *Negative | 1 | 100.00% |
| Quidel QuickVue Chlamydia | 2 | *Negative | 2 | 100.00% |
| Cepheid GeneXpert | 2 | *Negative | 2 | 100.00% |
| BD Viper XTR CT/GC | 1 | *Negative | 1 | 100.00% |
| CY-7 | | * Denotes Correct Response | | |
| AG: All Methods | 7 | | | |
| | | *Positive | 7 | 100.00% |
| BD Probe Tec ET | 1 | *Positive | 1 | 100.00% |
| Gen-Probe | 1 | *Positive | 1 | 100.00% |
| Quidel QuickVue Chlamydia | 2 | *Positive | 2 | 100.00% |
| Cepheid GeneXpert | 2 | *Positive | 2 | 100.00% |
| BD Viper XTR CT/GC | 1 | *Positive | 1 | 100.00% |
| CY-8 | | * Denotes Correct Response | | |
| AG: All Methods | 7 | | | |
| | | *Negative | 6 | 85.71% |
| | | Positive | 1 | 14.29% |
| Gen-Probe | 1 | *Negative | 1 | 100.00% |
| Quidel QuickVue Chlamydia | 2 | *Negative | 2 | 100.00% |
| Cepheid GeneXpert | 2 | *Negative | 2 | 100.00% |
| BD Viper XTR CT/GC | 1 | *Negative | 1 | 100.00% |
| BD Probe Tec ET | 1 | Positive | 1 | 100.00% |
| CY-9 | | * Denotes Correct Response | | |
| AG: All Methods | 7 | | | |
| | | *Positive | 7 | 100.00% |
| BD Probe Tec ET | 1 | *Positive | 1 | 100.00% |
| Gen-Probe | 1 | *Positive | 1 | 100.00% |
| Quidel QuickVue Chlamydia | 2 | *Positive | 2 | 100.00% |
| Cepheid GeneXpert | 2 | *Positive | 2 | 100.00% |
| BD Viper XTR CT/GC | 1 | *Positive | 1 | 100.00% |
| CY-10 | | * Denotes Correct Response | | |
| AG: All Methods | 7 | | | |
| | | *Negative | 6 | 85.71% |
| | | Positive | 1 | 14.29% |
| BD Probe Tec ET | 1 | *Negative | 1 | 100.00% |
| Gen-Probe | 1 | *Negative | 1 | 100.00% |
| Quidel QuickVue Chlamydia | 2 | *Negative | 2 | 100.00% |
| Cepheid GeneXpert | 2 | *Negative | 1 | 50.00% |
| BD Viper XTR CT/GC | 1 | *Negative | 1 | 100.00% |
| Cepheid GeneXpert | 2 | Positive | 1 | 50.00% |
| | | GC - Non Culture Tech | | |
| CY-6 | | * Denotes Correct Response | | |
| AG: All Methods | 5 | | | |
| | | *Positive | 4 | 80.00% |
| | | Negative | 1 | 20.00% |
| BD Probe Tec ET | 1 | *Positive | 1 | 100.00% |
| Cepheid GeneXpert | 2 | *Positive | 2 | 100.00% |



| | | | | |
|--|-------------|-------------------------------------|---------------|----------------|
| BD Viper XTR CT/GC | 1 | *Positive | 1 | 100.00% |
| Gen-Probe | 1 | Negative | 1 | 100.00% |
| CY-7 | | * Denotes Correct Response | | |
| AG: All Methods | 5 | | | |
| | | *Negative | 3 | 60.00% |
| | | Positive | 2 | 40.00% |
| BD Probe Tec ET | 1 | *Negative | 1 | 100.00% |
| Gen-Probe | 1 | *Negative | 1 | 100.00% |
| BD Viper XTR CT/GC | 1 | *Negative | 1 | 100.00% |
| Cepheid GeneXpert | 2 | Positive | 2 | 100.00% |
| CY-8 | | * Denotes Correct Response | | |
| AG: All Methods | 5 | | | |
| | | *Positive | 4 | 80.00% |
| | | Negative | 1 | 20.00% |
| BD Probe Tec ET | 1 | *Positive | 1 | 100.00% |
| Cepheid GeneXpert | 2 | *Positive | 2 | 100.00% |
| BD Viper XTR CT/GC | 1 | *Positive | 1 | 100.00% |
| Gen-Probe | 1 | Negative | 1 | 100.00% |
| CY-9 | | * Denotes Correct Response | | |
| AG: All Methods | 5 | | | |
| | | *Positive | 4 | 80.00% |
| | | Negative | 1 | 20.00% |
| BD Probe Tec ET | 1 | *Positive | 1 | 100.00% |
| Cepheid GeneXpert | 2 | *Positive | 2 | 100.00% |
| BD Viper XTR CT/GC | 1 | *Positive | 1 | 100.00% |
| Gen-Probe | 1 | Negative | 1 | 100.00% |
| CY-10 | | * Denotes Correct Response | | |
| AG: All Methods | 5 | | | |
| | | *Negative | 4 | 80.00% |
| | | Positive | 1 | 20.00% |
| Gen-Probe | 1 | *Negative | 1 | 100.00% |
| Cepheid GeneXpert | 2 | *Negative | 2 | 100.00% |
| BD Viper XTR CT/GC | 1 | *Negative | 1 | 100.00% |
| BD Probe Tec ET | 1 | Positive | 1 | 100.00% |
| 772 - Dermatophyte Screen/Culture | Labs | Response | Number | Percent |
| | | Dermatophyte Culture | | |
| DM-6 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Trichophyton species | 1 | 50.00% |
| | | *Trichophyton tonsurans | 1 | 50.00% |
| Dermatophytes | 2 | *Trichophyton species | 1 | 50.00% |
| Dermatophytes | | *Trichophyton tonsurans | 1 | 50.00% |
| DM-7 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Negative, No Dermatophytes Present | 2 | 100.00% |
| Dermatophytes | 2 | *Negative, No Dermatophytes Present | 2 | 100.00% |
| DM-8 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Trichophyton mentagrophytes | 1 | 50.00% |
| | | *Trichophyton species | 1 | 50.00% |
| Dermatophytes | 2 | *Trichophyton mentagrophytes | 1 | 50.00% |
| Dermatophytes | | *Trichophyton species | 1 | 50.00% |
| DM-9 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Negative, No Dermatophytes Present | 2 | 100.00% |
| Dermatophytes | 2 | *Negative, No Dermatophytes Present | 2 | 100.00% |
| DM-10 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Trichophyton verrucosum | 1 | 50.00% |
| | | Negative, No Dermatophytes Present | 1 | 50.00% |
| Dermatophytes | 2 | *Trichophyton verrucosum | 1 | 50.00% |
| Dermatophytes | | Negative, No Dermatophytes Present | 1 | 50.00% |
| | | Dermatophyte Screen | | |
| DM-6 | | * Denotes Correct Response | | |
| AG: All Methods | 12 | | | |
| | | *Positive, Dermatophytes Present | 12 | 100.00% |
| Dermatophytes | 12 | *Positive, Dermatophytes Present | 12 | 100.00% |
| DM-7 | | * Denotes Correct Response | | |



| | | | | |
|--|-------------|---|---------------|----------------|
| AG: All Methods | 11 | | | |
| | | *Negative, No Dermatophytes Present | 9 | 81.82% |
| | | Positive, Dermatophytes Present | 2 | 18.18% |
| Dermatophytes | 11 | *Negative, No Dermatophytes Present | 9 | 81.82% |
| Dermatophytes | | Positive, Dermatophytes Present | 2 | 18.18% |
| DM-8 | | * Denotes Correct Response | | |
| AG: All Methods | 12 | | | |
| | | *Positive, Dermatophytes Present | 12 | 100.00% |
| Dermatophytes | 12 | *Positive, Dermatophytes Present | 12 | 100.00% |
| DM-9 | | * Denotes Correct Response | | |
| AG: All Methods | 11 | | | |
| | | *Negative, No Dermatophytes Present | 10 | 90.91% |
| | | Positive, Dermatophytes Present | 1 | 9.09% |
| Dermatophytes | 11 | *Negative, No Dermatophytes Present | 10 | 90.91% |
| Dermatophytes | | Positive, Dermatophytes Present | 1 | 9.09% |
| DM-10 | | * Denotes Correct Response | | |
| AG: All Methods | 12 | | | |
| | | *Positive, Dermatophytes Present | 5 | 41.67% |
| | | Negative, No Dermatophytes Present | 7 | 58.33% |
| Dermatophytes | 12 | *Positive, Dermatophytes Present | 5 | 41.67% |
| Dermatophytes | | Negative, No Dermatophytes Present | 7 | 58.33% |
| 777 - Genital Culture:796 - Genital Culture (2 specimens) | Labs | Response | Number | Percent |
| | | Genital Culture | | |
| GC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *No growth, GC selective media | 3 | 100.00% |
| Cultures | 3 | *No growth, GC selective media | 3 | 100.00% |
| GC-7 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *No Gram Negative diplococci present | 1 | 33.33% |
| | | *No growth, GC selective media | 2 | 66.67% |
| Cultures | 3 | *No Gram Negative diplococci present | 1 | 33.33% |
| Cultures | | *No growth, GC selective media | 2 | 66.67% |
| GC-8 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *No growth, GC selective media | 3 | 100.00% |
| Cultures | 3 | *No growth, GC selective media | 3 | 100.00% |
| GC-9 | | * Denotes Correct Response | | |
| AG: All Methods | 5 | | | |
| | | *Positive for GC | 1 | 20.00% |
| | | *Presumptive positive for GC, GR(-) diplococci, oxidase positive organism | 3 | 60.00% |
| | | Group B Streptococcus not present | 1 | 20.00% |
| Cultures | 5 | *Positive for GC | 1 | 20.00% |
| Cultures | | *Presumptive positive for GC, GR(-) diplococci, oxidase positive organism | 3 | 60.00% |
| Cultures | | Group B Streptococcus not present | 1 | 20.00% |
| GC-10 | | * Denotes Correct Response | | |
| AG: All Methods | 4 | | | |
| | | *Negative for GC | 1 | 25.00% |
| | | *No growth, GC selective media | 3 | 75.00% |
| Cultures | 4 | *Negative for GC | 1 | 25.00% |
| Cultures | | *No growth, GC selective media | 3 | 75.00% |
| 668 - Gram Stain | Labs | Response | Number | Percent |
| | | Gram Stain | | |
| GS-6 | | * Denotes Correct Response | | |
| AG: All Methods | 8 | | | |
| | | *Gram Negative | 8 | 100.00% |
| Gram Stain | 8 | *Gram Negative | 8 | 100.00% |
| GS-7 | | * Denotes Correct Response | | |
| AG: All Methods | 8 | | | |
| | | *Gram Negative | 8 | 100.00% |
| Gram Stain | 8 | *Gram Negative | 8 | 100.00% |
| GS-8 | | * Denotes Correct Response | | |
| AG: All Methods | 8 | | | |
| | | *Gram Positive | 8 | 100.00% |
| Gram Stain | 8 | *Gram Positive | 8 | 100.00% |
| GS-9 | | * Denotes Correct Response | | |



| | | | | |
|---|-------------|-----------------------------------|---------------|----------------|
| AG: All Methods | 8 | | | |
| | | *Gram Positive | 8 | 100.00% |
| Gram Stain | 8 | *Gram Positive | 8 | 100.00% |
| GS-10 | | * Denotes Correct Response | | |
| AG: All Methods | 8 | | | |
| | | *Gram Negative | 8 | 100.00% |
| Gram Stain | 8 | *Gram Negative | 8 | 100.00% |
| | | Gram Stain Morph | | |
| GS-6 | | * Denotes Correct Response | | |
| AG: All Methods | 7 | | | |
| | | *Rods/Bacilli | 7 | 100.00% |
| Gram Stain | 7 | *Rods/Bacilli | 7 | 100.00% |
| GS-7 | | * Denotes Correct Response | | |
| AG: All Methods | 7 | | | |
| | | *Cocco-Bacilli | 2 | 28.57% |
| | | *Rods/Bacilli | 4 | 57.14% |
| | | Cocci | 1 | 14.29% |
| Gram Stain | 7 | *Cocco-Bacilli | 2 | 28.57% |
| Gram Stain | | *Rods/Bacilli | 4 | 57.14% |
| Gram Stain | | Cocci | 1 | 14.29% |
| GS-8 | | * Denotes Correct Response | | |
| AG: All Methods | 7 | | | |
| | | *Cocci | 4 | 57.14% |
| | | *Cocci in Chains | 2 | 28.57% |
| | | *Cocci in Pairs | 1 | 14.29% |
| Gram Stain | 7 | *Cocci | 4 | 57.14% |
| Gram Stain | | *Cocci in Chains | 2 | 28.57% |
| Gram Stain | | *Cocci in Pairs | 1 | 14.29% |
| GS-9 | | * Denotes Correct Response | | |
| AG: All Methods | 7 | | | |
| | | *Cocci | 4 | 57.14% |
| | | *Cocci in Pairs | 3 | 42.86% |
| Gram Stain | 7 | *Cocci | 4 | 57.14% |
| Gram Stain | | *Cocci in Pairs | 3 | 42.86% |
| GS-10 | | * Denotes Correct Response | | |
| AG: All Methods | 7 | | | |
| | | *Rods/Bacilli | 7 | 100.00% |
| Gram Stain | 7 | *Rods/Bacilli | 7 | 100.00% |
| 678 - Serum hCG;759 - Complete Immunology;729 - AFC Immunology | Labs | Response | Number | Percent |
| | | HCG - Qual Serum | | |
| HCG-6 | | * Denotes Correct Response | | |
| AG: All Methods | 92 | | | |
| | | *Positive | 92 | 100.00% |
| Beck Coulter ICON 25 | 7 | *Positive | 7 | 100.00% |
| Fisher HC Sure-Vue hCG STAT | 2 | *Positive | 2 | 100.00% |
| Sekisui Diagnostics OSOM | 3 | *Positive | 3 | 100.00% |
| Henry Schein One Step + (hCG) | 5 | *Positive | 5 | 100.00% |
| Immunostics Immuno/hCG | 2 | *Positive | 2 | 100.00% |
| Inverness Signify HCG | 2 | *Positive | 2 | 100.00% |
| LifeSign Status hCG | 2 | *Positive | 2 | 100.00% |
| McKesson Medi-Lab hCG - Urine | 1 | *Positive | 1 | 100.00% |
| Quidel QuickVue + hCG | 1 | *Positive | 1 | 100.00% |
| Quidel QuickVue One-Step hCG Combo (U/S) | 33 | *Positive | 33 | 100.00% |
| Select Diag hCG Combo | 3 | *Positive | 3 | 100.00% |
| SP Brand (Cardinal) HCG Combo | 9 | *Positive | 9 | 100.00% |
| Wampole Clearview | 1 | *Positive | 1 | 100.00% |
| Polymedco Poly stat hCG Serum | 1 | *Positive | 1 | 100.00% |
| Stanbio True 20 Plus | 1 | *Positive | 1 | 100.00% |
| Consult Diag hCG Combo Cassette | 12 | *Positive | 12 | 100.00% |
| Alere hCG Combo Cassette (25 mIU/mL) | 1 | *Positive | 1 | 100.00% |
| Medline Pregnancy Combo (serum) | 1 | *Positive | 1 | 100.00% |
| Alere hCG Combo Cassette (20/10 mIU/mL) | 1 | *Positive | 1 | 100.00% |
| McKesson Medi-Lab hCG - Serum | 4 | *Positive | 4 | 100.00% |
| HCG-7 | | * Denotes Correct Response | | |
| AG: All Methods | 92 | | | |
| | | *Negative | 92 | 100.00% |
| Beck Coulter ICON 25 | 7 | *Negative | 7 | 100.00% |



| | | | | |
|--|----|-----------------------------------|----|---------|
| Fisher HC Sure-Vue hCG STAT | 2 | *Negative | 2 | 100.00% |
| Sekisui Diagnostics OSOM | 3 | *Negative | 3 | 100.00% |
| Henry Schein One Step + (hCG) | 5 | *Negative | 5 | 100.00% |
| Immunostics Immuno/hCG | 2 | *Negative | 2 | 100.00% |
| Inverness Signify HCG | 2 | *Negative | 2 | 100.00% |
| LifeSign Status hCG | 2 | *Negative | 2 | 100.00% |
| McKesson Medi-Lab hCG - Urine | 1 | *Negative | 1 | 100.00% |
| Quidel QuickVue + hCG | 1 | *Negative | 1 | 100.00% |
| Quidel QuickVue One-Step hCG Combo (U/S) | 33 | *Negative | 33 | 100.00% |
| Select Diag hCG Combo | 3 | *Negative | 3 | 100.00% |
| SP Brand (Cardinal) HCG Combo | 9 | *Negative | 9 | 100.00% |
| Wampole Clearview | 1 | *Negative | 1 | 100.00% |
| Polymedco Poly stat hCG Serum | 1 | *Negative | 1 | 100.00% |
| Stanbio True 20 Plus | 1 | *Negative | 1 | 100.00% |
| Consult Diag hCG Combo Cassette | 12 | *Negative | 12 | 100.00% |
| Alere hCG Combo Cassette (25 mIU/mL) | 1 | *Negative | 1 | 100.00% |
| Medline Pregnancy Combo (serum) | 1 | *Negative | 1 | 100.00% |
| Alere hCG Combo Cassette (20/10 mIU/mL) | 1 | *Negative | 1 | 100.00% |
| McKesson Medi-Lab hCG - Serum | 4 | *Negative | 4 | 100.00% |
| HCG-8 | | * Denotes Correct Response | | |
| AG: All Methods | 92 | | | |
| | | *Negative | 92 | 100.00% |
| Beck Coulter ICON 25 | 7 | *Negative | 7 | 100.00% |
| Fisher HC Sure-Vue hCG STAT | 2 | *Negative | 2 | 100.00% |
| Sekisui Diagnostics OSOM | 3 | *Negative | 3 | 100.00% |
| Henry Schein One Step + (hCG) | 5 | *Negative | 5 | 100.00% |
| Immunostics Immuno/hCG | 2 | *Negative | 2 | 100.00% |
| Inverness Signify HCG | 2 | *Negative | 2 | 100.00% |
| LifeSign Status hCG | 2 | *Negative | 2 | 100.00% |
| McKesson Medi-Lab hCG - Urine | 1 | *Negative | 1 | 100.00% |
| Quidel QuickVue + hCG | 1 | *Negative | 1 | 100.00% |
| Quidel QuickVue One-Step hCG Combo (U/S) | 33 | *Negative | 33 | 100.00% |
| Select Diag hCG Combo | 3 | *Negative | 3 | 100.00% |
| SP Brand (Cardinal) HCG Combo | 9 | *Negative | 9 | 100.00% |
| Wampole Clearview | 1 | *Negative | 1 | 100.00% |
| Polymedco Poly stat hCG Serum | 1 | *Negative | 1 | 100.00% |
| Stanbio True 20 Plus | 1 | *Negative | 1 | 100.00% |
| Consult Diag hCG Combo Cassette | 12 | *Negative | 12 | 100.00% |
| Alere hCG Combo Cassette (25 mIU/mL) | 1 | *Negative | 1 | 100.00% |
| Medline Pregnancy Combo (serum) | 1 | *Negative | 1 | 100.00% |
| Alere hCG Combo Cassette (20/10 mIU/mL) | 1 | *Negative | 1 | 100.00% |
| McKesson Medi-Lab hCG - Serum | 4 | *Negative | 4 | 100.00% |
| HCG-9 | | * Denotes Correct Response | | |
| AG: All Methods | 92 | | | |
| | | *Positive | 92 | 100.00% |
| Beck Coulter ICON 25 | 7 | *Positive | 7 | 100.00% |
| Fisher HC Sure-Vue hCG STAT | 2 | *Positive | 2 | 100.00% |
| Sekisui Diagnostics OSOM | 3 | *Positive | 3 | 100.00% |
| Henry Schein One Step + (hCG) | 5 | *Positive | 5 | 100.00% |
| Immunostics Immuno/hCG | 2 | *Positive | 2 | 100.00% |
| Inverness Signify HCG | 2 | *Positive | 2 | 100.00% |
| LifeSign Status hCG | 2 | *Positive | 2 | 100.00% |
| McKesson Medi-Lab hCG - Urine | 1 | *Positive | 1 | 100.00% |
| Quidel QuickVue + hCG | 1 | *Positive | 1 | 100.00% |
| Quidel QuickVue One-Step hCG Combo (U/S) | 33 | *Positive | 33 | 100.00% |
| Select Diag hCG Combo | 3 | *Positive | 3 | 100.00% |
| SP Brand (Cardinal) HCG Combo | 9 | *Positive | 9 | 100.00% |
| Wampole Clearview | 1 | *Positive | 1 | 100.00% |
| Polymedco Poly stat hCG Serum | 1 | *Positive | 1 | 100.00% |
| Stanbio True 20 Plus | 1 | *Positive | 1 | 100.00% |
| Consult Diag hCG Combo Cassette | 12 | *Positive | 12 | 100.00% |
| Alere hCG Combo Cassette (25 mIU/mL) | 1 | *Positive | 1 | 100.00% |
| Medline Pregnancy Combo (serum) | 1 | *Positive | 1 | 100.00% |
| Alere hCG Combo Cassette (20/10 mIU/mL) | 1 | *Positive | 1 | 100.00% |
| McKesson Medi-Lab hCG - Serum | 4 | *Positive | 4 | 100.00% |
| HCG-10 | | * Denotes Correct Response | | |
| AG: All Methods | 92 | | | |
| | | *Positive | 92 | 100.00% |
| Beck Coulter ICON 25 | 7 | *Positive | 7 | 100.00% |



| | | | | |
|--|------|-----------|--------|---------|
| Fisher HC Sure-Vue hCG STAT | 2 | *Positive | 2 | 100.00% |
| Sekisui Diagnostics OSOM | 3 | *Positive | 3 | 100.00% |
| Henry Schein One Step + (hCG) | 5 | *Positive | 5 | 100.00% |
| Immunostics Immuno/hCG | 2 | *Positive | 2 | 100.00% |
| Inverness Signify HCG | 2 | *Positive | 2 | 100.00% |
| LifeSign Status hCG | 2 | *Positive | 2 | 100.00% |
| McKesson Medi-Lab hCG - Urine | 1 | *Positive | 1 | 100.00% |
| Quidel QuickVue + hCG | 1 | *Positive | 1 | 100.00% |
| Quidel QuickVue One-Step hCG Combo (U/S) | 33 | *Positive | 33 | 100.00% |
| Select Diag hCG Combo | 3 | *Positive | 3 | 100.00% |
| SP Brand (Cardinal) HCG Combo | 9 | *Positive | 9 | 100.00% |
| Wampole Clearview | 1 | *Positive | 1 | 100.00% |
| Polymedco Poly stat hCG Serum | 1 | *Positive | 1 | 100.00% |
| Stanbio True 20 Plus | 1 | *Positive | 1 | 100.00% |
| Consult Diag hCG Combo Cassette | 12 | *Positive | 12 | 100.00% |
| Alere hCG Combo Cassette (25 mIU/mL) | 1 | *Positive | 1 | 100.00% |
| Medline Pregnancy Combo (serum) | 1 | *Positive | 1 | 100.00% |
| Alere hCG Combo Cassette (20/10 mIU/mL) | 1 | *Positive | 1 | 100.00% |
| McKesson Medi-Lab hCG - Serum | 4 | *Positive | 4 | 100.00% |
| HCG - Qual Serum (International) | | | | |
| HCG-6 | | | | |
| * Denotes Correct Response | | | | |
| AG: All Methods | 3 | | | |
| | | *Positive | 3 | 100.00% |
| Henry Schein One Step + (hCG) | 1 | *Positive | 1 | 100.00% |
| SAS Ultra HCG | 2 | *Positive | 2 | 100.00% |
| HCG-7 | | | | |
| * Denotes Correct Response | | | | |
| AG: All Methods | 3 | | | |
| | | *Negative | 3 | 100.00% |
| Henry Schein One Step + (hCG) | 1 | *Negative | 1 | 100.00% |
| SAS Ultra HCG | 2 | *Negative | 2 | 100.00% |
| HCG-8 | | | | |
| * Denotes Correct Response | | | | |
| AG: All Methods | 3 | | | |
| | | *Negative | 3 | 100.00% |
| Henry Schein One Step + (hCG) | 1 | *Negative | 1 | 100.00% |
| SAS Ultra HCG | 2 | *Negative | 2 | 100.00% |
| HCG-9 | | | | |
| * Denotes Correct Response | | | | |
| AG: All Methods | 3 | | | |
| | | *Positive | 3 | 100.00% |
| Henry Schein One Step + (hCG) | 1 | *Positive | 1 | 100.00% |
| SAS Ultra HCG | 2 | *Positive | 2 | 100.00% |
| HCG-10 | | | | |
| * Denotes Correct Response | | | | |
| AG: All Methods | 3 | | | |
| | | *Positive | 3 | 100.00% |
| Henry Schein One Step + (hCG) | 1 | *Positive | 1 | 100.00% |
| SAS Ultra HCG | 2 | *Positive | 2 | 100.00% |
| 762 - HIV Antibody (Waived);766 - HIV Antibody (5 specimens);750 - Viral Markers | Labs | Response | Number | Percent |
| HIV Antibody (Waived) | | | | |
| HIV-6 | | | | |
| * Denotes Correct Response | | | | |
| AG: All Methods | 82 | | | |
| | | *Negative | 82 | 100.00% |
| Alere Clearview Complete HIV 1/2 (mod) | 9 | *Negative | 9 | 100.00% |
| Alere Clearview HIV 1/2 STAT PAK (WB) | 48 | *Negative | 48 | 100.00% |
| Orasure OraQuick Anti-HIV | 10 | *Negative | 10 | 100.00% |
| Trinity Biotech Uni-Gold Recombigen HIV | 3 | *Negative | 3 | 100.00% |
| Chembio HIV 1/2 STAT-PAK Dipstick | 1 | *Negative | 1 | 100.00% |
| Chembio HIV 1/2 STAT-PAK Cassette | 8 | *Negative | 8 | 100.00% |
| bioLytical INSTI HIV Antibody | 2 | *Negative | 2 | 100.00% |
| HIV Antibody (International) | | | | |
| HIV-6 | | | | |
| * Denotes Correct Response | | | | |
| AG: All Methods | 4 | | | |
| | | *Negative | 4 | 100.00% |
| Intl OCD Vitros ECI/ECIQ | 3 | *Negative | 3 | 100.00% |
| Intl Chembio HIV 1/2 STAT-PAK Dipstick | 1 | *Negative | 1 | 100.00% |
| HIV Antibody | | | | |
| HIV-6 | | | | |
| * Denotes Correct Response | | | | |
| AG: All Methods | 3 | | | |
| | | *Negative | 2 | 66.67% |



| | Labs | Response | Number | Percent |
|---|------|-----------------------------------|--------|---------|
| | | Positive | 1 | 33.33% |
| Alere Clearview Complete HIV 1/2 (mod) | 2 | *Negative | 2 | 100.00% |
| Siemens Centaur | 1 | Positive | 1 | 100.00% |
| 766 - HIV Antibody (5 specimens);750 - Viral Markers | | | | |
| | | HIV Antibody | | |
| | | * Denotes Correct Response | | |
| HIV-6 | | | | |
| AG: All Methods | 3 | | | |
| | | *Negative | 2 | 66.67% |
| | | Positive | 1 | 33.33% |
| Alere Clearview Complete HIV 1/2 (mod) | 2 | *Negative | 2 | 100.00% |
| Siemens Centaur | 1 | Positive | 1 | 100.00% |
| | | * Denotes Correct Response | | |
| HIV-7 | | | | |
| AG: All Methods | 3 | | | |
| | | *Negative | 2 | 66.67% |
| | | Positive | 1 | 33.33% |
| Alere Clearview Complete HIV 1/2 (mod) | 2 | *Negative | 2 | 100.00% |
| Siemens Centaur | 1 | Positive | 1 | 100.00% |
| | | * Denotes Correct Response | | |
| HIV-8 | | | | |
| AG: All Methods | 3 | | | |
| | | *Positive | 3 | 100.00% |
| Siemens Centaur | 1 | *Positive | 1 | 100.00% |
| Alere Clearview Complete HIV 1/2 (mod) | 2 | *Positive | 2 | 100.00% |
| | | * Denotes Correct Response | | |
| HIV-9 | | | | |
| AG: All Methods | 3 | | | |
| | | *Positive | 3 | 100.00% |
| Siemens Centaur | 1 | *Positive | 1 | 100.00% |
| Alere Clearview Complete HIV 1/2 (mod) | 2 | *Positive | 2 | 100.00% |
| | | * Denotes Correct Response | | |
| HIV-10 | | | | |
| AG: All Methods | 3 | | | |
| | | *Positive | 3 | 100.00% |
| Siemens Centaur | 1 | *Positive | 1 | 100.00% |
| Alere Clearview Complete HIV 1/2 (mod) | 2 | *Positive | 2 | 100.00% |
| | | HIV Antibody (Waived) | | |
| | | * Denotes Correct Response | | |
| HIV-6 | | | | |
| AG: All Methods | 82 | | | |
| | | *Negative | 82 | 100.00% |
| Alere Clearview Complete HIV 1/2 (mod) | 9 | *Negative | 9 | 100.00% |
| Alere Clearview HIV 1/2 STAT PAK (WB) | 48 | *Negative | 48 | 100.00% |
| Orasure OraQuick Anti-HIV | 10 | *Negative | 10 | 100.00% |
| Trinity Biotech Uni-Gold Recombigen HIV | 3 | *Negative | 3 | 100.00% |
| Chembio HIV 1/2 STAT-PAK Dipstick | 1 | *Negative | 1 | 100.00% |
| Chembio HIV 1/2 STAT-PAK Cassette | 8 | *Negative | 8 | 100.00% |
| bioLytical INSTI HIV Antibody | 2 | *Negative | 2 | 100.00% |
| | | * Denotes Correct Response | | |
| HIV-7 | | | | |
| AG: All Methods | 7 | | | |
| | | *Negative | 7 | 100.00% |
| Alere Clearview HIV 1/2 STAT PAK (WB) | 4 | *Negative | 4 | 100.00% |
| Orasure OraQuick Anti-HIV | 1 | *Negative | 1 | 100.00% |
| Trinity Biotech Uni-Gold Recombigen HIV | 1 | *Negative | 1 | 100.00% |
| bioLytical INSTI HIV Antibody | 1 | *Negative | 1 | 100.00% |
| | | * Denotes Correct Response | | |
| HIV-8 | | | | |
| AG: All Methods | 7 | | | |
| | | *Positive | 7 | 100.00% |
| Alere Clearview HIV 1/2 STAT PAK (WB) | 4 | *Positive | 4 | 100.00% |
| Orasure OraQuick Anti-HIV | 1 | *Positive | 1 | 100.00% |
| Trinity Biotech Uni-Gold Recombigen HIV | 1 | *Positive | 1 | 100.00% |
| bioLytical INSTI HIV Antibody | 1 | *Positive | 1 | 100.00% |
| | | * Denotes Correct Response | | |
| HIV-9 | | | | |
| AG: All Methods | 7 | | | |
| | | *Positive | 7 | 100.00% |
| Alere Clearview HIV 1/2 STAT PAK (WB) | 4 | *Positive | 4 | 100.00% |
| Orasure OraQuick Anti-HIV | 1 | *Positive | 1 | 100.00% |
| Trinity Biotech Uni-Gold Recombigen HIV | 1 | *Positive | 1 | 100.00% |
| bioLytical INSTI HIV Antibody | 1 | *Positive | 1 | 100.00% |
| | | * Denotes Correct Response | | |
| HIV-10 | | | | |
| AG: All Methods | 7 | | | |
| | | *Positive | 7 | 100.00% |
| Alere Clearview HIV 1/2 STAT PAK (WB) | 4 | *Positive | 4 | 100.00% |



| | | | | |
|--|-------------|-------------------------------------|---------------|----------------|
| Orasure OraQuick Anti-HIV | 1 | *Positive | 1 | 100.00% |
| Trinity Biotech Uni-Gold Recombigen HIV | 1 | *Positive | 1 | 100.00% |
| bioLytical INSTI HIV Antibody | 1 | *Positive | 1 | 100.00% |
| | | HIV Antibody (International) | | |
| | | * Denotes Correct Response | | |
| HIV-6 | | | | |
| AG: All Methods | 4 | | | |
| | | *Negative | 4 | 100.00% |
| Intl OCD Vitros ECi/ECiQ | 3 | *Negative | 3 | 100.00% |
| Intl Chembio HIV 1/2 STAT-PAK Dipstick | 1 | *Negative | 1 | 100.00% |
| | | * Denotes Correct Response | | |
| HIV-7 | | | | |
| AG: All Methods | 3 | | | |
| | | *Negative | 3 | 100.00% |
| Intl OCD Vitros ECi/ECiQ | 3 | *Negative | 3 | 100.00% |
| | | * Denotes Correct Response | | |
| HIV-8 | | | | |
| AG: All Methods | 3 | | | |
| | | *Positive | 3 | 100.00% |
| Intl OCD Vitros ECi/ECiQ | 3 | *Positive | 3 | 100.00% |
| | | * Denotes Correct Response | | |
| HIV-9 | | | | |
| AG: All Methods | 3 | | | |
| | | *Positive | 3 | 100.00% |
| Intl OCD Vitros ECi/ECiQ | 3 | *Positive | 3 | 100.00% |
| | | * Denotes Correct Response | | |
| HIV-10 | | | | |
| AG: All Methods | 3 | | | |
| | | *Positive | 3 | 100.00% |
| Intl OCD Vitros ECi/ECiQ | 3 | *Positive | 3 | 100.00% |
| 644 - H. pylori;628 - AFC Waived/PPM | Labs | Response | Number | Percent |
| | | H. pylori Ab Detect | | |
| | | * Denotes Correct Response | | |
| HP-2 | | | | |
| AG: All Methods | 79 | | | |
| | | *Positive | 74 | 93.67% |
| | | Negative | 5 | 6.33% |
| Beck Coulter FlexSure | 3 | *Positive | 3 | 100.00% |
| Beck Coulter ICON HP | 2 | *Positive | 2 | 100.00% |
| Instant Technologies i-Screen | 1 | *Positive | 1 | 100.00% |
| McKesson Medi-Lab Perf H. pylori | 8 | *Positive | 6 | 75.00% |
| Polymedco Poly stat (WB - Waived) | 3 | *Positive | 3 | 100.00% |
| Henry Schein One Step+ (H. pylori) | 5 | *Positive | 5 | 100.00% |
| Select Diag H. pylori Cassette | 3 | *Positive | 3 | 100.00% |
| Quidel QuickVue H. pylori gII (S/P) | 5 | *Positive | 5 | 100.00% |
| Quidel QuickVue H. pylori gII (WB) | 11 | *Positive | 11 | 100.00% |
| SP Brand (Cardinal) HP Rapid Test | 10 | *Positive | 9 | 90.00% |
| Alere Clearview II S/P (mod) | 3 | *Positive | 3 | 100.00% |
| Alere Clearview II WB (waived) | 5 | *Positive | 4 | 80.00% |
| LifeSign Status H. pylori | 2 | *Positive | 2 | 100.00% |
| Instant-View H. pylori | 1 | *Positive | 1 | 100.00% |
| Consult Diag H pylori | 16 | *Positive | 15 | 93.75% |
| Jant Accutest H pylori | 1 | *Positive | 1 | 100.00% |
| McKesson Medi-Lab Perf H. pylori | 8 | Negative | 2 | 25.00% |
| SP Brand (Cardinal) HP Rapid Test | 10 | Negative | 1 | 10.00% |
| Alere Clearview II WB (waived) | 5 | Negative | 1 | 20.00% |
| Consult Diag H pylori | 16 | Negative | 1 | 6.25% |
| 933 - HIV Antibody/Antigen Combo (2 specimens);934 - HIV Antibody/Antigen Combo (5 specimens) | Labs | Response | Number | Percent |
| | | HIV Antibody (Waived) | | |
| | | * Denotes Correct Response | | |
| HVAA-6 | | | | |
| AG: All Methods | 27 | | | |
| | | *Negative | 27 | 100.00% |
| Alere Determine HIV Ag/Ab | 27 | *Negative | 27 | 100.00% |
| | | * Denotes Correct Response | | |
| HVAA-7 | | | | |
| AG: All Methods | 27 | | | |
| | | *Negative | 25 | 92.59% |
| | | Positive | 2 | 7.41% |
| Alere Determine HIV Ag/Ab | 27 | *Negative | 25 | 92.59% |
| Alere Determine HIV Ag/Ab | | Positive | 2 | 7.41% |
| | | HIV Antigen (p24) (Waived) | | |
| | | * Denotes Correct Response | | |
| HVAA-6 | | | | |
| AG: All Methods | 27 | | | |
| | | *Negative | 27 | 100.00% |



| | | | | |
|--|-------------|--|---------------|----------------|
| Alere Determine HIV Ag/Ab | 27 | *Negative | 27 | 100.00% |
| HVAA-7 | | * Denotes Correct Response | | |
| AG: All Methods | 27 | | | |
| | | *Negative | 27 | 100.00% |
| Alere Determine HIV Ag/Ab | 27 | *Negative | 27 | 100.00% |
| 603 - Blood Cell Identification;610 - Regular Hematology;749 - Auto Diff II w-Blood Cell Identification | Labs | Response | Number | Percent |
| | | Blood Cell Identification | | |
| ID-7 | | * Denotes Correct Response | | |
| AG: All Methods | 70 | | | |
| | | *Nucleated red cell | 69 | 98.57% |
| | | Red Blood Cell, normal (RBC) (Erythrocyte) | 1 | 1.43% |
| Blood Cell Microscopy | 70 | *Nucleated red cell | 69 | 98.57% |
| Blood Cell Microscopy | | Red Blood Cell, normal (RBC) (Erythrocyte) | 1 | 1.43% |
| ID-8 | | * Denotes Correct Response | | |
| AG: All Methods | 70 | | | |
| | | *Echinocyte (burr cell, crenated cell) | 66 | 94.29% |
| | | Acanthocyte (spur cell) | 3 | 4.29% |
| | | Fragmented cell (schist., helmet cell, keratocyte) | 1 | 1.43% |
| Blood Cell Microscopy | 70 | *Echinocyte (burr cell, crenated cell) | 66 | 94.29% |
| Blood Cell Microscopy | | Acanthocyte (spur cell) | 3 | 4.29% |
| Blood Cell Microscopy | | Fragmented cell (schist., helmet cell, keratocyte) | 1 | 1.43% |
| ID-9 | | * Denotes Correct Response | | |
| AG: All Methods | 70 | | | |
| | | *Sickle cell (drepanocyte) | 68 | 97.14% |
| | | Ovalocyte (elliptocyte) | 2 | 2.86% |
| Blood Cell Microscopy | 70 | *Sickle cell (drepanocyte) | 68 | 97.14% |
| Blood Cell Microscopy | | Ovalocyte (elliptocyte) | 2 | 2.86% |
| ID-10 | | * Denotes Correct Response | | |
| AG: All Methods | 70 | | | |
| | | *Platelet, normal | 67 | 95.71% |
| | | Identification Unknown - Referred | 1 | 1.43% |
| | | Platelet, abnormal morphology | 1 | 1.43% |
| | | Platelet, giant (macrothrombocyte) | 1 | 1.43% |
| Blood Cell Microscopy | 70 | *Platelet, normal | 67 | 95.71% |
| Blood Cell Microscopy | | Identification Unknown - Referred | 1 | 1.43% |
| Blood Cell Microscopy | | Platelet, abnormal morphology | 1 | 1.43% |
| Blood Cell Microscopy | | Platelet, giant (macrothrombocyte) | 1 | 1.43% |
| ID-11 | | * Denotes Correct Response | | |
| AG: All Methods | 70 | | | |
| | | *Stomatocyte | 69 | 98.57% |
| | | Ovalocyte (elliptocyte) | 1 | 1.43% |
| Blood Cell Microscopy | 70 | *Stomatocyte | 69 | 98.57% |
| Blood Cell Microscopy | | Ovalocyte (elliptocyte) | 1 | 1.43% |
| ID-12 | | * Denotes Correct Response | | |
| AG: All Methods | 70 | | | |
| | | *Target cell (codocyte) | 70 | 100.00% |
| Blood Cell Microscopy | 70 | *Target cell (codocyte) | 70 | 100.00% |
| | | Blood Cell Identification (International) | | |
| ID-7 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Nucleated red cell | 1 | 100.00% |
| Intl Blood Cell Microscopy | 1 | *Nucleated red cell | 1 | 100.00% |
| ID-8 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Echinocyte (burr cell, crenated cell) | 1 | 100.00% |
| Intl Blood Cell Microscopy | 1 | *Echinocyte (burr cell, crenated cell) | 1 | 100.00% |
| ID-9 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Sickle cell (drepanocyte) | 1 | 100.00% |
| Intl Blood Cell Microscopy | 1 | *Sickle cell (drepanocyte) | 1 | 100.00% |
| ID-10 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Platelet, normal | 1 | 100.00% |
| Intl Blood Cell Microscopy | 1 | *Platelet, normal | 1 | 100.00% |
| ID-11 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Stomatocyte | 1 | 100.00% |



| Test Name | Labs | Response | Number | Percent |
|--|------|-----------------------------------|--------|---------|
| Intl Blood Cell Microscopy | 1 | *Stomatocyte | 1 | 100.00% |
| ID-12 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Target cell (codocyte) | 1 | 100.00% |
| Intl Blood Cell Microscopy | 1 | *Target cell (codocyte) | 1 | 100.00% |
| 609 - Waived/PPM Tests ;675 - Infectious Mononucleosis;753 - Infectious Mononucleosis (Waived) (1 specimens);754 - Infectious Mono/RF Combo;759 - Complete Immunology;628 - AFC Waived/PPM;729 - AFC Immunology | Labs | Response | Number | Percent |
| | | Infectious Mono (Waived) | | |
| IM-6 | | * Denotes Correct Response | | |
| AG: All Methods | 506 | | | |
| | | *Positive | 500 | 98.81% |
| | | Negative | 6 | 1.19% |
| Alere Acceava (S) | 1 | *Positive | 1 | 100.00% |
| Alere Acceava (WB) | 57 | *Positive | 56 | 98.25% |
| Sekisui Diagnostics OSOM Mono (WB) | 214 | *Positive | 211 | 98.60% |
| Henry Schein One-Step+ Mono | 29 | *Positive | 29 | 100.00% |
| Instant Technologies i-Screen (WB) | 2 | *Positive | 2 | 100.00% |
| Inverness Signify Mono | 3 | *Positive | 3 | 100.00% |
| LifeSign Status Inf Mono | 33 | *Positive | 33 | 100.00% |
| McKesson Medi-Lab Perf - Inf Mono | 15 | *Positive | 15 | 100.00% |
| Polymedco Poly stat Mono - (WB) | 7 | *Positive | 7 | 100.00% |
| Quidel QuickVue+ (WB) | 16 | *Positive | 15 | 93.75% |
| SP Brand (Cardinal) Rapid Test | 9 | *Positive | 9 | 100.00% |
| Alere Clearview Mono Cassette | 27 | *Positive | 26 | 96.30% |
| Pulse Scientific | 1 | *Positive | 1 | 100.00% |
| Diag Test Group Clarity Mono | 3 | *Positive | 3 | 100.00% |
| Fisher HC Sure-View Mono | 4 | *Positive | 4 | 100.00% |
| Immunostics Inc. ImmunoDetector Mono (WB) | 1 | *Positive | 1 | 100.00% |
| Meridian Immunocard STAT! Mono (WB) | 5 | *Positive | 5 | 100.00% |
| NDC ProAdvantage Mono | 2 | *Positive | 2 | 100.00% |
| Consult Diag Mono Cassette | 68 | *Positive | 68 | 100.00% |
| Alere Acceava (WB) Mono Cassette | 8 | *Positive | 8 | 100.00% |
| Medline Mono Cassette (WB) | 1 | *Positive | 1 | 100.00% |
| Alere Acceava (WB) | 57 | Negative | 1 | 1.75% |
| Sekisui Diagnostics OSOM Mono (WB) | 214 | Negative | 3 | 1.40% |
| Quidel QuickVue+ (WB) | 16 | Negative | 1 | 6.25% |
| Alere Clearview Mono Cassette | 27 | Negative | 1 | 3.70% |
| | | Infectious Mononucleosis | | |
| IM-6 | | * Denotes Correct Response | | |
| AG: All Methods | 79 | | | |
| | | *Positive | 79 | 100.00% |
| Beck Coulter ICON (WB) | 1 | *Positive | 1 | 100.00% |
| Alere Acceava (WB) | 1 | *Positive | 1 | 100.00% |
| Sekisui Diagnostics OSOM Mono (WB) | 36 | *Positive | 36 | 100.00% |
| Henry Schein One-Step+ Mono | 1 | *Positive | 1 | 100.00% |
| LifeSign Status Inf Mono | 2 | *Positive | 2 | 100.00% |
| Lifesign Unistep | 1 | *Positive | 1 | 100.00% |
| McKesson Medi-Lab Perf - Inf Mono | 3 | *Positive | 3 | 100.00% |
| Polymedco Poly stat Mono - (WB) | 1 | *Positive | 1 | 100.00% |
| Polymedco Poly stat Mono (S/P) | 2 | *Positive | 2 | 100.00% |
| Quidel QuickVue+ (S/P) | 2 | *Positive | 2 | 100.00% |
| Quidel QuickVue+ (WB) | 8 | *Positive | 8 | 100.00% |
| Seradyn Color Slide II | 4 | *Positive | 4 | 100.00% |
| SP Brand (Cardinal) Rapid Test | 1 | *Positive | 1 | 100.00% |
| Alere Clearview Mono Cassette | 8 | *Positive | 8 | 100.00% |
| Sekisui Diagnostics OSOM Mono Non-Waived | 3 | *Positive | 3 | 100.00% |
| NDC ProAdvantage Mono | 1 | *Positive | 1 | 100.00% |
| Consult Diag Mono Cassette | 4 | *Positive | 4 | 100.00% |
| 675 - Infectious Mononucleosis;754 - Infectious Mono/RF Combo;759 - Complete Immunology;628 - AFC Waived/PPM;729 - AFC Immunology | Labs | Response | Number | Percent |
| | | Infectious Mononucleosis | | |
| IM-6 | | * Denotes Correct Response | | |
| AG: All Methods | 79 | | | |
| | | *Positive | 79 | 100.00% |
| Beck Coulter ICON (WB) | 1 | *Positive | 1 | 100.00% |
| Alere Acceava (WB) | 1 | *Positive | 1 | 100.00% |



| | | | | |
|--|----|-----------------------------------|----|---------|
| Sekisui Diagnostics OSOM Mono (WB) | 36 | *Positive | 36 | 100.00% |
| Henry Schein One-Step+ Mono | 1 | *Positive | 1 | 100.00% |
| LifeSign Status Inf Mono | 2 | *Positive | 2 | 100.00% |
| Lifesign Unistep | 1 | *Positive | 1 | 100.00% |
| McKesson Medi-Lab Perf - Inf Mono | 3 | *Positive | 3 | 100.00% |
| Polymedco Poly stat Mono - (WB) | 1 | *Positive | 1 | 100.00% |
| Polymedco Poly stat Mono (S/P) | 2 | *Positive | 2 | 100.00% |
| Quidel QuickVue+ (S/P) | 2 | *Positive | 2 | 100.00% |
| Quidel QuickVue+ (WB) | 8 | *Positive | 8 | 100.00% |
| Seradyn Color Slide II | 4 | *Positive | 4 | 100.00% |
| SP Brand (Cardinal) Rapid Test | 1 | *Positive | 1 | 100.00% |
| Alere Clearview Mono Cassette | 8 | *Positive | 8 | 100.00% |
| Sekisui Diagnostics OSOM Mono Non-Waived | 3 | *Positive | 3 | 100.00% |
| NDC ProAdvantage Mono | 1 | *Positive | 1 | 100.00% |
| Consult Diag Mono Cassette | 4 | *Positive | 4 | 100.00% |
| IM-7 | | * Denotes Correct Response | | |
| AG: All Methods | 79 | | | |
| | | *Positive | 77 | 97.47% |
| | | Negative | 2 | 2.53% |
| Beck Coulter ICON (WB) | 1 | *Positive | 1 | 100.00% |
| Alere Acceava (WB) | 1 | *Positive | 1 | 100.00% |
| Sekisui Diagnostics OSOM Mono (WB) | 36 | *Positive | 36 | 100.00% |
| Henry Schein One-Step+ Mono | 1 | *Positive | 1 | 100.00% |
| LifeSign Status Inf Mono | 2 | *Positive | 2 | 100.00% |
| Lifesign Unistep | 1 | *Positive | 1 | 100.00% |
| McKesson Medi-Lab Perf - Inf Mono | 3 | *Positive | 2 | 66.67% |
| Polymedco Poly stat Mono - (WB) | 1 | *Positive | 1 | 100.00% |
| Polymedco Poly stat Mono (S/P) | 2 | *Positive | 2 | 100.00% |
| Quidel QuickVue+ (S/P) | 2 | *Positive | 2 | 100.00% |
| Quidel QuickVue+ (WB) | 8 | *Positive | 8 | 100.00% |
| Seradyn Color Slide II | 4 | *Positive | 3 | 75.00% |
| SP Brand (Cardinal) Rapid Test | 1 | *Positive | 1 | 100.00% |
| Alere Clearview Mono Cassette | 8 | *Positive | 8 | 100.00% |
| Sekisui Diagnostics OSOM Mono Non-Waived | 3 | *Positive | 3 | 100.00% |
| NDC ProAdvantage Mono | 1 | *Positive | 1 | 100.00% |
| Consult Diag Mono Cassette | 4 | *Positive | 4 | 100.00% |
| McKesson Medi-Lab Perf - Inf Mono | 3 | Negative | 1 | 33.33% |
| Seradyn Color Slide II | 4 | Negative | 1 | 25.00% |
| IM-8 | | * Denotes Correct Response | | |
| AG: All Methods | 79 | | | |
| | | *Negative | 78 | 98.73% |
| | | Positive | 1 | 1.27% |
| Beck Coulter ICON (WB) | 1 | *Negative | 1 | 100.00% |
| Alere Acceava (WB) | 1 | *Negative | 1 | 100.00% |
| Sekisui Diagnostics OSOM Mono (WB) | 36 | *Negative | 36 | 100.00% |
| Henry Schein One-Step+ Mono | 1 | *Negative | 1 | 100.00% |
| LifeSign Status Inf Mono | 2 | *Negative | 2 | 100.00% |
| Lifesign Unistep | 1 | *Negative | 1 | 100.00% |
| McKesson Medi-Lab Perf - Inf Mono | 3 | *Negative | 3 | 100.00% |
| Polymedco Poly stat Mono - (WB) | 1 | *Negative | 1 | 100.00% |
| Polymedco Poly stat Mono (S/P) | 2 | *Negative | 2 | 100.00% |
| Quidel QuickVue+ (S/P) | 2 | *Negative | 2 | 100.00% |
| Quidel QuickVue+ (WB) | 8 | *Negative | 8 | 100.00% |
| Seradyn Color Slide II | 4 | *Negative | 3 | 75.00% |
| SP Brand (Cardinal) Rapid Test | 1 | *Negative | 1 | 100.00% |
| Alere Clearview Mono Cassette | 8 | *Negative | 8 | 100.00% |
| Sekisui Diagnostics OSOM Mono Non-Waived | 3 | *Negative | 3 | 100.00% |
| NDC ProAdvantage Mono | 1 | *Negative | 1 | 100.00% |
| Consult Diag Mono Cassette | 4 | *Negative | 4 | 100.00% |
| Seradyn Color Slide II | 4 | Positive | 1 | 25.00% |
| IM-9 | | * Denotes Correct Response | | |
| AG: All Methods | 79 | | | |
| | | *Positive | 78 | 98.73% |
| | | Negative | 1 | 1.27% |
| Beck Coulter ICON (WB) | 1 | *Positive | 1 | 100.00% |
| Alere Acceava (WB) | 1 | *Positive | 1 | 100.00% |
| Sekisui Diagnostics OSOM Mono (WB) | 36 | *Positive | 36 | 100.00% |
| Henry Schein One-Step+ Mono | 1 | *Positive | 1 | 100.00% |
| LifeSign Status Inf Mono | 2 | *Positive | 2 | 100.00% |



| | | | | |
|---|-----|---|-----|---------|
| Lifesign Unistep | 1 | *Positive | 1 | 100.00% |
| McKesson Medi-Lab Perf - Inf Mono | 3 | *Positive | 3 | 100.00% |
| Polymedco Poly stat Mono - (WB) | 1 | *Positive | 1 | 100.00% |
| Polymedco Poly stat Mono (S/P) | 2 | *Positive | 2 | 100.00% |
| Quidel QuickVue+ (S/P) | 2 | *Positive | 2 | 100.00% |
| Quidel QuickVue+ (WB) | 8 | *Positive | 8 | 100.00% |
| Seradyn Color Slide II | 4 | *Positive | 3 | 75.00% |
| SP Brand (Cardinal) Rapid Test | 1 | *Positive | 1 | 100.00% |
| Alere Clearview Mono Cassette | 8 | *Positive | 8 | 100.00% |
| Sekisui Diagnostics OSOM Mono Non-Waived | 3 | *Positive | 3 | 100.00% |
| NDC ProAdvantage Mono | 1 | *Positive | 1 | 100.00% |
| Consult Diag Mono Cassette | 4 | *Positive | 4 | 100.00% |
| Seradyn Color Slide II | 4 | Negative | 1 | 25.00% |
| IM-10 | | * Denotes Correct Response | | |
| AG: All Methods | 79 | | | |
| | | *Negative | 78 | 98.73% |
| | | Positive | 1 | 1.27% |
| Beck Coulter ICON (WB) | 1 | *Negative | 1 | 100.00% |
| Alere Acceava (WB) | 1 | *Negative | 1 | 100.00% |
| Sekisui Diagnostics OSOM Mono (WB) | 36 | *Negative | 36 | 100.00% |
| Henry Schein One-Step+ Mono | 1 | *Negative | 1 | 100.00% |
| LifeSign Status Inf Mono | 2 | *Negative | 2 | 100.00% |
| Lifesign Unistep | 1 | *Negative | 1 | 100.00% |
| McKesson Medi-Lab Perf - Inf Mono | 3 | *Negative | 3 | 100.00% |
| Polymedco Poly stat Mono - (WB) | 1 | *Negative | 1 | 100.00% |
| Polymedco Poly stat Mono (S/P) | 2 | *Negative | 2 | 100.00% |
| Quidel QuickVue+ (S/P) | 2 | *Negative | 2 | 100.00% |
| Quidel QuickVue+ (WB) | 8 | *Negative | 8 | 100.00% |
| Seradyn Color Slide II | 4 | *Negative | 3 | 75.00% |
| SP Brand (Cardinal) Rapid Test | 1 | *Negative | 1 | 100.00% |
| Alere Clearview Mono Cassette | 8 | *Negative | 8 | 100.00% |
| Sekisui Diagnostics OSOM Mono Non-Waived | 3 | *Negative | 3 | 100.00% |
| NDC ProAdvantage Mono | 1 | *Negative | 1 | 100.00% |
| Consult Diag Mono Cassette | 4 | *Negative | 4 | 100.00% |
| Seradyn Color Slide II | 4 | Positive | 1 | 25.00% |
| IM-6 | | Infectious Mono (Waived) | | |
| IM-6 | | * Denotes Correct Response | | |
| AG: All Methods | 506 | | | |
| | | *Positive | 500 | 98.81% |
| | | Negative | 6 | 1.19% |
| Alere Acceava (S) | 1 | *Positive | 1 | 100.00% |
| Alere Acceava (WB) | 57 | *Positive | 56 | 98.25% |
| Sekisui Diagnostics OSOM Mono (WB) | 214 | *Positive | 211 | 98.60% |
| Henry Schein One-Step+ Mono | 29 | *Positive | 29 | 100.00% |
| Instant Technologies i-Screen (WB) | 2 | *Positive | 2 | 100.00% |
| Inverness Signify Mono | 3 | *Positive | 3 | 100.00% |
| LifeSign Status Inf Mono | 33 | *Positive | 33 | 100.00% |
| McKesson Medi-Lab Perf - Inf Mono | 15 | *Positive | 15 | 100.00% |
| Polymedco Poly stat Mono - (WB) | 7 | *Positive | 7 | 100.00% |
| Quidel QuickVue+ (WB) | 16 | *Positive | 15 | 93.75% |
| SP Brand (Cardinal) Rapid Test | 9 | *Positive | 9 | 100.00% |
| Alere Clearview Mono Cassette | 27 | *Positive | 26 | 96.30% |
| Pulse Scientific | 1 | *Positive | 1 | 100.00% |
| Diag Test Group Clarity Mono | 3 | *Positive | 3 | 100.00% |
| Fisher HC Sure-Vue Mono | 4 | *Positive | 4 | 100.00% |
| ImmunoStics Inc. ImmunoDetector Mono (WB) | 1 | *Positive | 1 | 100.00% |
| Meridian Immunocard STAT! Mono (WB) | 5 | *Positive | 5 | 100.00% |
| NDC ProAdvantage Mono | 2 | *Positive | 2 | 100.00% |
| Consult Diag Mono Cassette | 68 | *Positive | 68 | 100.00% |
| Alere Acceava (WB) Mono Cassette | 8 | *Positive | 8 | 100.00% |
| Medline Mono Cassette (WB) | 1 | *Positive | 1 | 100.00% |
| Alere Acceava (WB) | 57 | Negative | 1 | 1.75% |
| Sekisui Diagnostics OSOM Mono (WB) | 214 | Negative | 3 | 1.40% |
| Quidel QuickVue+ (WB) | 16 | Negative | 1 | 6.25% |
| Alere Clearview Mono Cassette | 27 | Negative | 1 | 3.70% |
| IM-6 | | Infectious Mononucleosis (International) | | |
| IM-6 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Positive | 2 | 100.00% |



| | | | | |
|--|-------------|-------------------------------------|---------------|----------------|
| Alere Clearview Mono Cassette | 1 | *Positive | 1 | 100.00% |
| Fisher HC Sure-Vue Mono | 1 | *Positive | 1 | 100.00% |
| IM-7 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Positive | 2 | 100.00% |
| Alere Clearview Mono Cassette | 1 | *Positive | 1 | 100.00% |
| Fisher HC Sure-Vue Mono | 1 | *Positive | 1 | 100.00% |
| IM-8 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Negative | 2 | 100.00% |
| Alere Clearview Mono Cassette | 1 | *Negative | 1 | 100.00% |
| Fisher HC Sure-Vue Mono | 1 | *Negative | 1 | 100.00% |
| IM-9 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Positive | 2 | 100.00% |
| Alere Clearview Mono Cassette | 1 | *Positive | 1 | 100.00% |
| Fisher HC Sure-Vue Mono | 1 | *Positive | 1 | 100.00% |
| IM-10 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Negative | 2 | 100.00% |
| Alere Clearview Mono Cassette | 1 | *Negative | 1 | 100.00% |
| Fisher HC Sure-Vue Mono | 1 | *Negative | 1 | 100.00% |
| 637 - KOH - Glass Slides (Skin Hair Nail) | Labs | Response | Number | Percent |
| | | KOH Preparation | | |
| KOH-3 | | * Denotes Correct Response | | |
| AG: All Methods | 18 | | | |
| | | *Fungal elements/yeast observed | 18 | 100.00% |
| KOH Glass Slide | 18 | *Fungal elements/yeast observed | 18 | 100.00% |
| KOH-4 | | * Denotes Correct Response | | |
| AG: All Methods | 18 | | | |
| | | *No fungal elements/yeast observed | 16 | 88.89% |
| | | Fungal elements/yeast observed | 2 | 11.11% |
| KOH Glass Slide | 18 | *No fungal elements/yeast observed | 16 | 88.89% |
| KOH Glass Slide | | Fungal elements/yeast observed | 2 | 11.11% |
| 758 - Mycoplasma Antibody | Labs | Response | Number | Percent |
| | | Mycoplasma Antibody | | |
| MY-2 | | * Denotes Correct Response | | |
| AG: All Methods | 5 | | | |
| | | *Positive | 5 | 100.00% |
| Meridian ImmunoCard | 5 | *Positive | 5 | 100.00% |
| 791 - Fecal Occult Blood Add-On;601 - Waived/PPM Tests II;609 - Waived/PPM Tests I;647 - Mini-Waived Package;628 - AFC Waived/PPM | Labs | Response | Number | Percent |
| | | Fecal Occult Blood | | |
| OC-2 | | * Denotes Correct Response | | |
| MG: Guaiac methods | 746 | | | |
| | | *Negative | 694 | 93.03% |
| | | Positive | 52 | 6.97% |
| OC-2 | | * Denotes Correct Response | | |
| MG: Immunochemical methods | 117 | | | |
| | | *Negative | 112 | 95.73% |
| | | Positive | 5 | 4.27% |
| 950 - Post Vasectomy Sperm Analysis | Labs | Response | Number | Percent |
| | | Sperm Count (Post Vasectomy) | | |
| PVAS-3 | | * Denotes Correct Response | | |
| AG: All Methods | 16 | | | |
| | | *Negative for Sperm | 13 | 81.25% |
| | | Positive for Sperm | 3 | 18.75% |
| Hemocytometer | 4 | *Negative for Sperm | 3 | 75.00% |
| Post Vasectomy Microscopy | 12 | *Negative for Sperm | 10 | 83.33% |
| Hemocytometer | 4 | Positive for Sperm | 1 | 25.00% |
| Post Vasectomy Microscopy | 12 | Positive for Sperm | 2 | 16.67% |
| PVAS-4 | | * Denotes Correct Response | | |
| AG: All Methods | 16 | | | |
| | | *Positive for Sperm | 13 | 81.25% |
| | | Negative for Sperm | 3 | 18.75% |
| Hemocytometer | 4 | *Positive for Sperm | 3 | 75.00% |
| Post Vasectomy Microscopy | 12 | *Positive for Sperm | 10 | 83.33% |



| 754 - Infectious Mono/RF Combo;653 - Rheumatoid Factor;759 - Complete Immunology | Labs | Response | Number | Percent |
|--|------|--------------------|--------|---------|
| Hemocytometer | 4 | Negative for Sperm | 1 | 25.00% |
| Post Vasectomy Microscopy | 12 | Negative for Sperm | 2 | 16.67% |
| Rheumatoid Factor | | | | |
| RF-6 | | | | |
| * Denotes Correct Response | | | | |
| AG: All Methods | 12 | | | |
| | | *Positive | 12 | 100.00% |
| ASI | 1 | *Positive | 1 | 100.00% |
| Siemens RapiTex | 2 | *Positive | 2 | 100.00% |
| Seradyn Color Slide | 2 | *Positive | 2 | 100.00% |
| Wampole ColorCard RF | 2 | *Positive | 2 | 100.00% |
| Wampole Rheumatex | 2 | *Positive | 2 | 100.00% |
| Biokit Rhemajet RF | 2 | *Positive | 2 | 100.00% |
| Remel SeraTest RF-Latex | 1 | *Positive | 1 | 100.00% |
| RF-7 | | | | |
| * Denotes Correct Response | | | | |
| AG: All Methods | 12 | | | |
| | | *Positive | 12 | 100.00% |
| ASI | 1 | *Positive | 1 | 100.00% |
| Siemens RapiTex | 2 | *Positive | 2 | 100.00% |
| Seradyn Color Slide | 2 | *Positive | 2 | 100.00% |
| Wampole ColorCard RF | 2 | *Positive | 2 | 100.00% |
| Wampole Rheumatex | 2 | *Positive | 2 | 100.00% |
| Biokit Rhemajet RF | 2 | *Positive | 2 | 100.00% |
| Remel SeraTest RF-Latex | 1 | *Positive | 1 | 100.00% |
| RF-8 | | | | |
| * Denotes Correct Response | | | | |
| AG: All Methods | 11 | | | |
| | | *N/A (Negative) | 10 | 90.91% |
| | | Positive | 1 | 9.09% |
| ASI | 1 | *N/A (Negative) | 1 | 100.00% |
| Siemens RapiTex | 2 | *N/A (Negative) | 2 | 100.00% |
| Seradyn Color Slide | 2 | *N/A (Negative) | 2 | 100.00% |
| Wampole ColorCard RF | 2 | *N/A (Negative) | 2 | 100.00% |
| Wampole Rheumatex | 2 | *N/A (Negative) | 2 | 100.00% |
| Biokit Rhemajet RF | 2 | *N/A (Negative) | 1 | 50.00% |
| Biokit Rhemajet RF | | Positive | 1 | 50.00% |
| RF-9 | | | | |
| * Denotes Correct Response | | | | |
| AG: All Methods | 11 | | | |
| | | *N/A (Negative) | 10 | 90.91% |
| | | Positive | 1 | 9.09% |
| ASI | 1 | *N/A (Negative) | 1 | 100.00% |
| Siemens RapiTex | 2 | *N/A (Negative) | 2 | 100.00% |
| Seradyn Color Slide | 2 | *N/A (Negative) | 1 | 50.00% |
| Wampole ColorCard RF | 2 | *N/A (Negative) | 2 | 100.00% |
| Wampole Rheumatex | 2 | *N/A (Negative) | 2 | 100.00% |
| Biokit Rhemajet RF | 2 | *N/A (Negative) | 2 | 100.00% |
| Seradyn Color Slide | 2 | Positive | 1 | 50.00% |
| RF-10 | | | | |
| * Denotes Correct Response | | | | |
| AG: All Methods | 12 | | | |
| | | *Positive | 11 | 91.67% |
| | | N/A (Negative) | 1 | 8.33% |
| ASI | 1 | *Positive | 1 | 100.00% |
| Siemens RapiTex | 2 | *Positive | 2 | 100.00% |
| Seradyn Color Slide | 2 | *Positive | 1 | 50.00% |
| Wampole ColorCard RF | 2 | *Positive | 2 | 100.00% |
| Wampole Rheumatex | 2 | *Positive | 2 | 100.00% |
| Biokit Rhemajet RF | 2 | *Positive | 2 | 100.00% |
| Remel SeraTest RF-Latex | 1 | *Positive | 1 | 100.00% |
| Seradyn Color Slide | 2 | N/A (Negative) | 1 | 50.00% |
| Rheumatoid Factor (International) | | | | |
| RF-6 | | | | |
| * Denotes Correct Response | | | | |
| AG: All Methods | 3 | | | |
| | | *Positive | 3 | 100.00% |
| Wampole ColorCard RF | 1 | *Positive | 1 | 100.00% |
| Fisher HC Sure-Vue RF | 1 | *Positive | 1 | 100.00% |
| Immunostics Rhuemacol/LEX | 1 | *Positive | 1 | 100.00% |
| RF-7 | | | | |
| * Denotes Correct Response | | | | |
| AG: All Methods | 3 | | | |



| | | | | |
|--|-------------|---------------------------------------|---------------|----------------|
| | | *Positive | 3 | 100.00% |
| Wampole ColorCard RF | 1 | *Positive | 1 | 100.00% |
| Fisher HC Sure-Vue RF | 1 | *Positive | 1 | 100.00% |
| Immunostics Rhuemacol/LEX | 1 | *Positive | 1 | 100.00% |
| RF-8 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Negative | 3 | 100.00% |
| Wampole ColorCard RF | 1 | *Negative | 1 | 100.00% |
| Fisher HC Sure-Vue RF | 1 | *Negative | 1 | 100.00% |
| Immunostics Rhuemacol/LEX | 1 | *Negative | 1 | 100.00% |
| RF-9 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Negative | 3 | 100.00% |
| Wampole ColorCard RF | 1 | *Negative | 1 | 100.00% |
| Fisher HC Sure-Vue RF | 1 | *Negative | 1 | 100.00% |
| Immunostics Rhuemacol/LEX | 1 | *Negative | 1 | 100.00% |
| RF-10 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Positive | 3 | 100.00% |
| Wampole ColorCard RF | 1 | *Positive | 1 | 100.00% |
| Fisher HC Sure-Vue RF | 1 | *Positive | 1 | 100.00% |
| Immunostics Rhuemacol/LEX | 1 | *Positive | 1 | 100.00% |
| | | Rheumatoid Factor (Semi Quant) | | |
| RF-6 | | * Denotes Correct Response | | |
| PG: Fisher HC Sure-Vue RF (Semi Quant) | 1 | | | |
| | | *160 IU/mL | 1 | 100.00% |
| RF-7 | | * Denotes Correct Response | | |
| PG: Fisher HC Sure-Vue RF (Semi Quant) | 1 | | | |
| | | *320 IU/mL | 1 | 100.00% |
| RF-8 | | * Denotes Correct Response | | |
| PG: Fisher HC Sure-Vue RF (Semi Quant) | 1 | | | |
| | | *10 IU/mL | 1 | 100.00% |
| RF-9 | | * Denotes Correct Response | | |
| PG: Fisher HC Sure-Vue RF (Semi Quant) | 1 | | | |
| | | *20 IU/mL | 1 | 100.00% |
| RF-10 | | * Denotes Correct Response | | |
| PG: Fisher HC Sure-Vue RF (Semi Quant) | 1 | | | |
| | | *160 IU/mL | 1 | 100.00% |
| 645 - Fetal Membrane Rupture | Labs | Response | Number | Percent |
| | | Fetal Membranes Rupture PAMG-1 | | |
| ROM-3 | | * Denotes Correct Response | | |
| AG: All Methods | 17 | | | |
| | | *Negative | 17 | 100.00% |
| AmniSure ROM Test | 17 | *Negative | 17 | 100.00% |
| ROM-4 | | * Denotes Correct Response | | |
| AG: All Methods | 17 | | | |
| | | *Positive | 17 | 100.00% |
| AmniSure ROM Test | 17 | *Positive | 17 | 100.00% |
| 601 - Waived/PPM Tests II;609 - Waived/PPM Tests I;785 - Presump ID/Colony Count-Urine & Strep Combo;691 - Group A Strep Antigen (5 specimens);699 - Group A Strep Antigen (1 specimen);783 - Group A Strep Antigen/Add-on (2 specimens);665 - Group A Strep An | Labs | Response | Number | Percent |
| | | Strep A Ag Det. (Waived) | | |
| RS-6 | | * Denotes Correct Response | | |
| AG: All Methods | 888 | | | |
| | | *Negative | 878 | 98.87% |
| | | Positive | 10 | 1.13% |
| Abbott Signify Strep A - Waived | 6 | *Negative | 6 | 100.00% |
| Beck Coulter ICON DS Strep | 1 | *Negative | 1 | 100.00% |
| Beck Coulter ICON SC Strep | 1 | *Negative | 1 | 100.00% |
| Alere Acceava Strep A Dipstick | 114 | *Negative | 114 | 100.00% |
| Diag Test Group Clarity - Strip Strep Ag | 1 | *Negative | 1 | 100.00% |
| Fisher HC Sure-Vue Strep A (Waived) | 8 | *Negative | 8 | 100.00% |
| Sekisui Diagnostics OSOM | 45 | *Negative | 45 | 100.00% |
| Sekisui Diagnostics OSOM Ultra Strep | 51 | *Negative | 49 | 96.08% |
| Henry Schein OneStep + Rapid Strep Cass | 5 | *Negative | 5 | 100.00% |
| Henry Schein OneStep+ Rapid Strep Dip | 63 | *Negative | 62 | 98.41% |
| Innovacon One Step Strep Ag | 1 | *Negative | 1 | 100.00% |



| | | | | |
|--|-----|-----------------------------------|-----|---------|
| Instant Technologies i Strep | 1 | *Negative | 1 | 100.00% |
| Inverness Signify Strep A Dipstick | 7 | *Negative | 7 | 100.00% |
| McKesson Strep A Cas | 4 | *Negative | 4 | 100.00% |
| McKesson Strep A Dip | 43 | *Negative | 43 | 100.00% |
| McKesson Strep A Test - Twist | 2 | *Negative | 2 | 100.00% |
| Permaxim Rediscreen Strep A Rapid | 1 | *Negative | 1 | 100.00% |
| Polymedco Poly Stat Strep A (W) | 3 | *Negative | 3 | 100.00% |
| PSS Rapid Test Strip Strep A | 2 | *Negative | 2 | 100.00% |
| Quidel QuickVue Dipstick Strep | 95 | *Negative | 95 | 100.00% |
| Quidel QuickVue In-Line - 25 Test | 84 | *Negative | 82 | 97.62% |
| Quidel QuickVue+ Strep A | 31 | *Negative | 31 | 100.00% |
| Select Diag Strep A Dip | 3 | *Negative | 3 | 100.00% |
| SP (Cardinal) Rapid Strep Cassette | 6 | *Negative | 6 | 100.00% |
| SP Brand (Cardinal) Rapid Strep A Dipstick | 89 | *Negative | 89 | 100.00% |
| Stanbio QuStick Strep A | 1 | *Negative | 1 | 100.00% |
| Alere Clearview Exact II Dip | 30 | *Negative | 29 | 96.67% |
| Sekisui Diagnostics OSOM Ultra | 5 | *Negative | 5 | 100.00% |
| Consult Diag Strep A Dipstick | 126 | *Negative | 123 | 97.62% |
| Consult Diag Strep A Twist | 2 | *Negative | 2 | 100.00% |
| MooreBrand Strep A Dipstick | 5 | *Negative | 5 | 100.00% |
| NDC ProAdvantage Strep Strips | 2 | *Negative | 2 | 100.00% |
| Meridian ImmunoCard STAT Strep A | 2 | *Negative | 2 | 100.00% |
| LifeSign Bio Strep A Test STRip | 1 | *Negative | 1 | 100.00% |
| Quidel Sofia Strep A FIA | 7 | *Negative | 7 | 100.00% |
| BD Veritor Strep A | 33 | *Negative | 32 | 96.97% |
| z DO NOT USE Alere Acceava | 1 | *Negative | 1 | 100.00% |
| Alere i Strep A | 4 | *Negative | 4 | 100.00% |
| Roche Cobas Liat Strep A | 2 | *Negative | 2 | 100.00% |
| Sekisui Diagnostics OSOM Ultra Strep | 51 | Positive | 2 | 3.92% |
| Henry Schein OneStep+ Rapid Strep Dip | 63 | Positive | 1 | 1.59% |
| Quidel QuickVue In-Line - 25 Test | 84 | Positive | 2 | 2.38% |
| Alere Clearview Exact II Dip | 30 | Positive | 1 | 3.33% |
| Consult Diag Strep A Dipstick | 126 | Positive | 3 | 2.38% |
| BD Veritor Strep A | 33 | Positive | 1 | 3.03% |
| RS-7 | | * Denotes Correct Response | | |
| AG: All Methods | 37 | | | |
| | | *Positive | 29 | 78.38% |
| | | Negative | 8 | 21.62% |
| Alere Acceava Strep A Dipstick | 1 | *Positive | 1 | 100.00% |
| Sekisui Diagnostics OSOM | 4 | *Positive | 2 | 50.00% |
| Sekisui Diagnostics OSOM Ultra Strep | 5 | *Positive | 4 | 80.00% |
| Innovacon One Step Strep Ag | 1 | *Positive | 1 | 100.00% |
| McKesson Strep A Dip | 2 | *Positive | 2 | 100.00% |
| Quidel QuickVue Dipstick Strep | 3 | *Positive | 3 | 100.00% |
| Quidel QuickVue In-Line - 25 Test | 2 | *Positive | 1 | 50.00% |
| Quidel QuickVue+ Strep A | 1 | *Positive | 1 | 100.00% |
| SP Brand (Cardinal) Rapid Strep A Dipstick | 7 | *Positive | 3 | 42.86% |
| Consult Diag Strep A Dipstick | 8 | *Positive | 8 | 100.00% |
| Consult Diag Strep A Twist | 1 | *Positive | 1 | 100.00% |
| MooreBrand Strep A Dipstick | 1 | *Positive | 1 | 100.00% |
| Alere i Strep A | 1 | *Positive | 1 | 100.00% |
| Sekisui Diagnostics OSOM | 4 | Negative | 2 | 50.00% |
| Sekisui Diagnostics OSOM Ultra Strep | 5 | Negative | 1 | 20.00% |
| Quidel QuickVue In-Line - 25 Test | 2 | Negative | 1 | 50.00% |
| SP Brand (Cardinal) Rapid Strep A Dipstick | 7 | Negative | 4 | 57.14% |
| RS-6 | | * Denotes Correct Response | | |
| AG: All Methods | 209 | | | |
| | | *Negative | 208 | 99.52% |
| | | Positive | 1 | 0.48% |
| Abbott Signify Strep A - Waived | 2 | *Negative | 2 | 100.00% |
| BD Check Strep A (Dip) | 2 | *Negative | 2 | 100.00% |
| Beck Coulter ICON DS Strep | 2 | *Negative | 2 | 100.00% |
| Alere Acceava Strep A Dipstick | 16 | *Negative | 16 | 100.00% |
| Diag Test Group Clarity - Strip Strep Ag | 1 | *Negative | 1 | 100.00% |
| Fisher HC Sure-Vue Select Strep Ag | 1 | *Negative | 1 | 100.00% |
| Sekisui Diagnostics OSOM | 25 | *Negative | 25 | 100.00% |
| Sekisui Diagnostics OSOM Ultra Strep | 28 | *Negative | 28 | 100.00% |
| Henry Schein OneStep+ Rapid Strep Dip | 10 | *Negative | 10 | 100.00% |



| | | | | |
|--|-------------|--|---------------|----------------|
| Immunostics Immuno/Strep A Det | 2 | *Negative | 2 | 100.00% |
| Inverness Signify Strep A Dipstick | 1 | *Negative | 1 | 100.00% |
| McKesson Strep A Cas | 2 | *Negative | 2 | 100.00% |
| McKesson Strep A Dip | 10 | *Negative | 10 | 100.00% |
| Quidel QuickVue Dipstick Strep | 22 | *Negative | 22 | 100.00% |
| Quidel QuickVue In-Line - 25 Test | 14 | *Negative | 14 | 100.00% |
| Quidel QuickVue+ Strep A | 31 | *Negative | 30 | 96.77% |
| Select Diag Strep A Dip | 1 | *Negative | 1 | 100.00% |
| SP (Cardinal) Rapid Strep Cassette | 1 | *Negative | 1 | 100.00% |
| SP Brand (Cardinal) Rapid Strep A Dipstick | 3 | *Negative | 3 | 100.00% |
| Stanbio QuStick Strep A | 1 | *Negative | 1 | 100.00% |
| Alere Clearview Exact II Dip | 1 | *Negative | 1 | 100.00% |
| Sekisui Diagnostics OSOM Ultra | 1 | *Negative | 1 | 100.00% |
| Consult Diag Strep A Dipstick | 18 | *Negative | 18 | 100.00% |
| Meridian ImmunoCard STAT Strep A | 1 | *Negative | 1 | 100.00% |
| Quidel Sofia Strep A FIA | 4 | *Negative | 4 | 100.00% |
| BD Veritor Strep A | 6 | *Negative | 6 | 100.00% |
| Quidel Solana | 3 | *Negative | 3 | 100.00% |
| Quidel QuickVue+ Strep A | 31 | Positive | 1 | 3.23% |
| RS-7 | | * Denotes Correct Response | | |
| AG: All Methods | 209 | | | |
| | | *Positive | 201 | 96.17% |
| | | Negative | 8 | 3.83% |
| Abbott Signify Strep A - Waived | 2 | *Positive | 2 | 100.00% |
| BD Check Strep A (Dip) | 2 | *Positive | 2 | 100.00% |
| Beck Coulter ICON DS Strep | 2 | *Positive | 1 | 50.00% |
| Alere Acceava Strep A Dipstick | 16 | *Positive | 16 | 100.00% |
| Diag Test Group Clarity - Strip Strep Ag | 1 | *Positive | 1 | 100.00% |
| Fisher HC Sure-Vue Select Strep Ag | 1 | *Positive | 1 | 100.00% |
| Sekisui Diagnostics OSOM | 25 | *Positive | 25 | 100.00% |
| Sekisui Diagnostics OSOM Ultra Strep | 28 | *Positive | 28 | 100.00% |
| Henry Schein OneStep+ Rapid Strep Dip | 10 | *Positive | 8 | 80.00% |
| Immunostics Immuno/Strep A Det | 2 | *Positive | 2 | 100.00% |
| Inverness Signify Strep A Dipstick | 1 | *Positive | 1 | 100.00% |
| McKesson Strep A Cas | 2 | *Positive | 2 | 100.00% |
| McKesson Strep A Dip | 10 | *Positive | 9 | 90.00% |
| Quidel QuickVue Dipstick Strep | 22 | *Positive | 22 | 100.00% |
| Quidel QuickVue In-Line - 25 Test | 14 | *Positive | 14 | 100.00% |
| Quidel QuickVue+ Strep A | 31 | *Positive | 30 | 96.77% |
| Select Diag Strep A Dip | 1 | *Positive | 1 | 100.00% |
| SP (Cardinal) Rapid Strep Cassette | 1 | *Positive | 1 | 100.00% |
| SP Brand (Cardinal) Rapid Strep A Dipstick | 3 | *Positive | 2 | 66.67% |
| Stanbio QuStick Strep A | 1 | *Positive | 1 | 100.00% |
| Alere Clearview Exact II Dip | 1 | *Positive | 1 | 100.00% |
| Sekisui Diagnostics OSOM Ultra | 1 | *Positive | 1 | 100.00% |
| Consult Diag Strep A Dipstick | 18 | *Positive | 16 | 88.89% |
| Meridian ImmunoCard STAT Strep A | 1 | *Positive | 1 | 100.00% |
| Quidel Sofia Strep A FIA | 4 | *Positive | 4 | 100.00% |
| BD Veritor Strep A | 6 | *Positive | 6 | 100.00% |
| Quidel Solana | 3 | *Positive | 3 | 100.00% |
| Beck Coulter ICON DS Strep | 2 | Negative | 1 | 50.00% |
| Henry Schein OneStep+ Rapid Strep Dip | 10 | Negative | 2 | 20.00% |
| McKesson Strep A Dip | 10 | Negative | 1 | 10.00% |
| Quidel QuickVue+ Strep A | 31 | Negative | 1 | 3.23% |
| SP Brand (Cardinal) Rapid Strep A Dipstick | 3 | Negative | 1 | 33.33% |
| Consult Diag Strep A Dipstick | 18 | Negative | 2 | 11.11% |
| | | Strep A Ag Det. (International) | | |
| RS-6 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Negative | 1 | 100.00% |
| Alere Clearview Exact II Cas | 1 | *Negative | 1 | 100.00% |
| RS-7 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Positive | 1 | 100.00% |
| Alere Clearview Exact II Cas | 1 | *Positive | 1 | 100.00% |
| 691 - Group A Strep Antigen (5 specimens) | Labs | Response | Number | Percent |
| | | Strep A Ag Det. | | |
| RS-6 | | * Denotes Correct Response | | |
| AG: All Methods | 209 | | | |



| | | | | |
|--|-----|-----------------------------------|-----|---------|
| | | *Negative | 208 | 99.52% |
| | | Positive | 1 | 0.48% |
| Abbott Signify Strep A - Waived | 2 | *Negative | 2 | 100.00% |
| BD Check Strep A (Dip) | 2 | *Negative | 2 | 100.00% |
| Beck Coulter ICON DS Strep | 2 | *Negative | 2 | 100.00% |
| Alere Acceava Strep A Dipstick | 16 | *Negative | 16 | 100.00% |
| Diag Test Group Clarity - Strip Strep Ag | 1 | *Negative | 1 | 100.00% |
| Fisher HC Sure-Vue Select Strep Ag | 1 | *Negative | 1 | 100.00% |
| Sekisui Diagnostics OSOM | 25 | *Negative | 25 | 100.00% |
| Sekisui Diagnostics OSOM Ultra Strep | 28 | *Negative | 28 | 100.00% |
| Henry Schein OneStep+ Rapid Strep Dip | 10 | *Negative | 10 | 100.00% |
| Immunostics Immuno/Strep A Det | 2 | *Negative | 2 | 100.00% |
| Inverness Signify Strep A Dipstick | 1 | *Negative | 1 | 100.00% |
| McKesson Strep A Cas | 2 | *Negative | 2 | 100.00% |
| McKesson Strep A Dip | 10 | *Negative | 10 | 100.00% |
| Quidel QuickVue Dipstick Strep | 22 | *Negative | 22 | 100.00% |
| Quidel QuickVue In-Line - 25 Test | 14 | *Negative | 14 | 100.00% |
| Quidel QuickVue+ Strep A | 31 | *Negative | 30 | 96.77% |
| Select Diag Strep A Dip | 1 | *Negative | 1 | 100.00% |
| SP (Cardinal) Rapid Strep Cassette | 1 | *Negative | 1 | 100.00% |
| SP Brand (Cardinal) Rapid Strep A Dipstick | 3 | *Negative | 3 | 100.00% |
| Stanbio QuStick Strep A | 1 | *Negative | 1 | 100.00% |
| Alere Clearview Exact II Dip | 1 | *Negative | 1 | 100.00% |
| Sekisui Diagnostics OSOM Ultra | 1 | *Negative | 1 | 100.00% |
| Consult Diag Strep A Dipstick | 18 | *Negative | 18 | 100.00% |
| Meridian ImmunoCard STAT Strep A | 1 | *Negative | 1 | 100.00% |
| Quidel Sofia Strep A FIA | 4 | *Negative | 4 | 100.00% |
| BD Veritor Strep A | 6 | *Negative | 6 | 100.00% |
| Quidel Solana | 3 | *Negative | 3 | 100.00% |
| Quidel QuickVue+ Strep A | 31 | Positive | 1 | 3.23% |
| RS-7 | | * Denotes Correct Response | | |
| AG: All Methods | 209 | | | |
| | | *Positive | 201 | 96.17% |
| | | Negative | 8 | 3.83% |
| Abbott Signify Strep A - Waived | 2 | *Positive | 2 | 100.00% |
| BD Check Strep A (Dip) | 2 | *Positive | 2 | 100.00% |
| Beck Coulter ICON DS Strep | 2 | *Positive | 1 | 50.00% |
| Alere Acceava Strep A Dipstick | 16 | *Positive | 16 | 100.00% |
| Diag Test Group Clarity - Strip Strep Ag | 1 | *Positive | 1 | 100.00% |
| Fisher HC Sure-Vue Select Strep Ag | 1 | *Positive | 1 | 100.00% |
| Sekisui Diagnostics OSOM | 25 | *Positive | 25 | 100.00% |
| Sekisui Diagnostics OSOM Ultra Strep | 28 | *Positive | 28 | 100.00% |
| Henry Schein OneStep+ Rapid Strep Dip | 10 | *Positive | 8 | 80.00% |
| Immunostics Immuno/Strep A Det | 2 | *Positive | 2 | 100.00% |
| Inverness Signify Strep A Dipstick | 1 | *Positive | 1 | 100.00% |
| McKesson Strep A Cas | 2 | *Positive | 2 | 100.00% |
| McKesson Strep A Dip | 10 | *Positive | 9 | 90.00% |
| Quidel QuickVue Dipstick Strep | 22 | *Positive | 22 | 100.00% |
| Quidel QuickVue In-Line - 25 Test | 14 | *Positive | 14 | 100.00% |
| Quidel QuickVue+ Strep A | 31 | *Positive | 30 | 96.77% |
| Select Diag Strep A Dip | 1 | *Positive | 1 | 100.00% |
| SP (Cardinal) Rapid Strep Cassette | 1 | *Positive | 1 | 100.00% |
| SP Brand (Cardinal) Rapid Strep A Dipstick | 3 | *Positive | 2 | 66.67% |
| Stanbio QuStick Strep A | 1 | *Positive | 1 | 100.00% |
| Alere Clearview Exact II Dip | 1 | *Positive | 1 | 100.00% |
| Sekisui Diagnostics OSOM Ultra | 1 | *Positive | 1 | 100.00% |
| Consult Diag Strep A Dipstick | 18 | *Positive | 16 | 88.89% |
| Meridian ImmunoCard STAT Strep A | 1 | *Positive | 1 | 100.00% |
| Quidel Sofia Strep A FIA | 4 | *Positive | 4 | 100.00% |
| BD Veritor Strep A | 6 | *Positive | 6 | 100.00% |
| Quidel Solana | 3 | *Positive | 3 | 100.00% |
| Beck Coulter ICON DS Strep | 2 | Negative | 1 | 50.00% |
| Henry Schein OneStep+ Rapid Strep Dip | 10 | Negative | 2 | 20.00% |
| McKesson Strep A Dip | 10 | Negative | 1 | 10.00% |
| Quidel QuickVue+ Strep A | 31 | Negative | 1 | 3.23% |
| SP Brand (Cardinal) Rapid Strep A Dipstick | 3 | Negative | 1 | 33.33% |
| Consult Diag Strep A Dipstick | 18 | Negative | 2 | 11.11% |
| RS-8 | | * Denotes Correct Response | | |
| AG: All Methods | 176 | | | |



| | | | | |
|--|-----|-----------------------------------|-----|---------|
| | | *Negative | 175 | 99.43% |
| | | Positive | 1 | 0.57% |
| Abbott Signify Strep A - Waived | 2 | *Negative | 2 | 100.00% |
| BD Check Strep A (Dip) | 2 | *Negative | 2 | 100.00% |
| Beck Coulter ICON DS Strep | 2 | *Negative | 2 | 100.00% |
| Alere Acceava Strep A Dipstick | 16 | *Negative | 16 | 100.00% |
| Diag Test Group Clarity - Strip Strep Ag | 1 | *Negative | 1 | 100.00% |
| Fisher HC Sure-Vue Select Strep Ag | 1 | *Negative | 1 | 100.00% |
| Sekisui Diagnostics OSOM | 25 | *Negative | 25 | 100.00% |
| Sekisui Diagnostics OSOM Ultra Strep | 24 | *Negative | 24 | 100.00% |
| Henry Schein OneStep+ Rapid Strep Dip | 10 | *Negative | 10 | 100.00% |
| Immunostics Immuno/Strep A Det | 2 | *Negative | 2 | 100.00% |
| Inverness Signify Strep A Dipstick | 1 | *Negative | 1 | 100.00% |
| McKesson Strep A Cas | 1 | *Negative | 1 | 100.00% |
| McKesson Strep A Dip | 5 | *Negative | 5 | 100.00% |
| Quidel QuickVue Dipstick Strep | 18 | *Negative | 18 | 100.00% |
| Quidel QuickVue In-Line - 25 Test | 12 | *Negative | 11 | 91.67% |
| Quidel QuickVue+ Strep A | 29 | *Negative | 29 | 100.00% |
| SP (Cardinal) Rapid Strep Cassette | 1 | *Negative | 1 | 100.00% |
| SP Brand (Cardinal) Rapid Strep A Dipstick | 3 | *Negative | 3 | 100.00% |
| Alere Clearview Exact II Dip | 1 | *Negative | 1 | 100.00% |
| Sekisui Diagnostics OSOM Ultra | 1 | *Negative | 1 | 100.00% |
| Consult Diag Strep A Dipstick | 13 | *Negative | 13 | 100.00% |
| Meridian ImmunoCard STAT Strep A | 1 | *Negative | 1 | 100.00% |
| Quidel Sofia Strep A FIA | 3 | *Negative | 3 | 100.00% |
| BD Veritor Strep A | 1 | *Negative | 1 | 100.00% |
| Quidel Solana | 1 | *Negative | 1 | 100.00% |
| Quidel QuickVue In-Line - 25 Test | 12 | Positive | 1 | 8.33% |
| RS-9 | | * Denotes Correct Response | | |
| AG: All Methods | 176 | | | |
| | | *Positive | 170 | 96.59% |
| | | Negative | 6 | 3.41% |
| Abbott Signify Strep A - Waived | 2 | *Positive | 2 | 100.00% |
| BD Check Strep A (Dip) | 2 | *Positive | 2 | 100.00% |
| Beck Coulter ICON DS Strep | 2 | *Positive | 1 | 50.00% |
| Alere Acceava Strep A Dipstick | 16 | *Positive | 16 | 100.00% |
| Diag Test Group Clarity - Strip Strep Ag | 1 | *Positive | 1 | 100.00% |
| Fisher HC Sure-Vue Select Strep Ag | 1 | *Positive | 1 | 100.00% |
| Sekisui Diagnostics OSOM | 25 | *Positive | 24 | 96.00% |
| Sekisui Diagnostics OSOM Ultra Strep | 24 | *Positive | 24 | 100.00% |
| Henry Schein OneStep+ Rapid Strep Dip | 10 | *Positive | 8 | 80.00% |
| Immunostics Immuno/Strep A Det | 2 | *Positive | 2 | 100.00% |
| Inverness Signify Strep A Dipstick | 1 | *Positive | 1 | 100.00% |
| McKesson Strep A Cas | 1 | *Positive | 1 | 100.00% |
| McKesson Strep A Dip | 5 | *Positive | 5 | 100.00% |
| Quidel QuickVue Dipstick Strep | 18 | *Positive | 18 | 100.00% |
| Quidel QuickVue In-Line - 25 Test | 12 | *Positive | 12 | 100.00% |
| Quidel QuickVue+ Strep A | 29 | *Positive | 28 | 96.55% |
| SP (Cardinal) Rapid Strep Cassette | 1 | *Positive | 1 | 100.00% |
| SP Brand (Cardinal) Rapid Strep A Dipstick | 3 | *Positive | 3 | 100.00% |
| Alere Clearview Exact II Dip | 1 | *Positive | 1 | 100.00% |
| Sekisui Diagnostics OSOM Ultra | 1 | *Positive | 1 | 100.00% |
| Consult Diag Strep A Dipstick | 13 | *Positive | 12 | 92.31% |
| Meridian ImmunoCard STAT Strep A | 1 | *Positive | 1 | 100.00% |
| Quidel Sofia Strep A FIA | 3 | *Positive | 3 | 100.00% |
| BD Veritor Strep A | 1 | *Positive | 1 | 100.00% |
| Quidel Solana | 1 | *Positive | 1 | 100.00% |
| Beck Coulter ICON DS Strep | 2 | Negative | 1 | 50.00% |
| Sekisui Diagnostics OSOM | 25 | Negative | 1 | 4.00% |
| Henry Schein OneStep+ Rapid Strep Dip | 10 | Negative | 2 | 20.00% |
| Quidel QuickVue+ Strep A | 29 | Negative | 1 | 3.45% |
| Consult Diag Strep A Dipstick | 13 | Negative | 1 | 7.69% |
| RS-10 | | * Denotes Correct Response | | |
| AG: All Methods | 176 | | | |
| | | *Negative | 174 | 98.86% |
| | | Positive | 2 | 1.14% |
| Abbott Signify Strep A - Waived | 2 | *Negative | 2 | 100.00% |
| BD Check Strep A (Dip) | 2 | *Negative | 2 | 100.00% |
| Beck Coulter ICON DS Strep | 2 | *Negative | 2 | 100.00% |



| | | | | |
|--|-----|-----------------------------------|-----|---------|
| Alere Acceava Strep A Dipstick | 16 | *Negative | 16 | 100.00% |
| Diag Test Group Clarity - Strip Strep Ag | 1 | *Negative | 1 | 100.00% |
| Fisher HC Sure-Vue Select Strep Ag | 1 | *Negative | 1 | 100.00% |
| Sekisui Diagnostics OSOM | 25 | *Negative | 24 | 96.00% |
| Sekisui Diagnostics OSOM Ultra Strep | 24 | *Negative | 24 | 100.00% |
| Henry Schein OneStep+ Rapid Strep Dip | 10 | *Negative | 10 | 100.00% |
| Immunostics Immuno/Strep A Det | 2 | *Negative | 2 | 100.00% |
| Inverness Signify Strep A Dipstick | 1 | *Negative | 1 | 100.00% |
| McKesson Strep A Cas | 1 | *Negative | 1 | 100.00% |
| McKesson Strep A Dip | 5 | *Negative | 5 | 100.00% |
| Quidel QuickVue Dipstick Strep | 18 | *Negative | 18 | 100.00% |
| Quidel QuickVue In-Line - 25 Test | 12 | *Negative | 12 | 100.00% |
| Quidel QuickVue+ Strep A | 29 | *Negative | 28 | 96.55% |
| SP (Cardinal) Rapid Strep Cassette | 1 | *Negative | 1 | 100.00% |
| SP Brand (Cardinal) Rapid Strep A Dipstick | 3 | *Negative | 3 | 100.00% |
| Alere Clearview Exact II Dip | 1 | *Negative | 1 | 100.00% |
| Sekisui Diagnostics OSOM Ultra | 1 | *Negative | 1 | 100.00% |
| Consult Diag Strep A Dipstick | 13 | *Negative | 13 | 100.00% |
| Meridian ImmunoCard STAT Strep A | 1 | *Negative | 1 | 100.00% |
| Quidel Sofia Strep A FIA | 3 | *Negative | 3 | 100.00% |
| BD Veritor Strep A | 1 | *Negative | 1 | 100.00% |
| Quidel Solana | 1 | *Negative | 1 | 100.00% |
| Sekisui Diagnostics OSOM | 25 | Positive | 1 | 4.00% |
| Quidel QuickVue+ Strep A | 29 | Positive | 1 | 3.45% |
| | | Strep A Ag Det. (Waived) | | |
| RS-6 | | * Denotes Correct Response | | |
| AG: All Methods | 888 | | | |
| | | *Negative | 878 | 98.87% |
| | | Positive | 10 | 1.13% |
| Abbott Signify Strep A - Waived | 6 | *Negative | 6 | 100.00% |
| Beck Coulter ICON DS Strep | 1 | *Negative | 1 | 100.00% |
| Beck Coulter ICON SC Strep | 1 | *Negative | 1 | 100.00% |
| Alere Acceava Strep A Dipstick | 114 | *Negative | 114 | 100.00% |
| Diag Test Group Clarity - Strip Strep Ag | 1 | *Negative | 1 | 100.00% |
| Fisher HC Sure-Vue Strep A (Waived) | 8 | *Negative | 8 | 100.00% |
| Sekisui Diagnostics OSOM | 45 | *Negative | 45 | 100.00% |
| Sekisui Diagnostics OSOM Ultra Strep | 51 | *Negative | 49 | 96.08% |
| Henry Schein OneStep + Rapid Strep Cass | 5 | *Negative | 5 | 100.00% |
| Henry Schein OneStep+ Rapid Strep Dip | 63 | *Negative | 62 | 98.41% |
| Innovacon One Step Strep Ag | 1 | *Negative | 1 | 100.00% |
| Instant Technologies i Strep | 1 | *Negative | 1 | 100.00% |
| Inverness Signify Strep A Dipstick | 7 | *Negative | 7 | 100.00% |
| McKesson Strep A Cas | 4 | *Negative | 4 | 100.00% |
| McKesson Strep A Dip | 43 | *Negative | 43 | 100.00% |
| McKesson Strep A Test - Twist | 2 | *Negative | 2 | 100.00% |
| Permaxim Rediscreen Strep A Rapid | 1 | *Negative | 1 | 100.00% |
| Polymedco Poly Stat Strep A (W) | 3 | *Negative | 3 | 100.00% |
| PSS Rapid Test Strip Strep A | 2 | *Negative | 2 | 100.00% |
| Quidel QuickVue Dipstick Strep | 95 | *Negative | 95 | 100.00% |
| Quidel QuickVue In-Line - 25 Test | 84 | *Negative | 82 | 97.62% |
| Quidel QuickVue+ Strep A | 31 | *Negative | 31 | 100.00% |
| Select Diag Strep A Dip | 3 | *Negative | 3 | 100.00% |
| SP (Cardinal) Rapid Strep Cassette | 6 | *Negative | 6 | 100.00% |
| SP Brand (Cardinal) Rapid Strep A Dipstick | 89 | *Negative | 89 | 100.00% |
| Stanbio QuStick Strep A | 1 | *Negative | 1 | 100.00% |
| Alere Clearview Exact II Dip | 30 | *Negative | 29 | 96.67% |
| Sekisui Diagnostics OSOM Ultra | 5 | *Negative | 5 | 100.00% |
| Consult Diag Strep A Dipstick | 126 | *Negative | 123 | 97.62% |
| Consult Diag Strep A Twist | 2 | *Negative | 2 | 100.00% |
| MooreBrand Strep A Dipstick | 5 | *Negative | 5 | 100.00% |
| NDC ProAdvantage Strep Strips | 2 | *Negative | 2 | 100.00% |
| Meridian ImmunoCard STAT Strep A | 2 | *Negative | 2 | 100.00% |
| LifeSign Bio Strep A Test STRip | 1 | *Negative | 1 | 100.00% |
| Quidel Sofia Strep A FIA | 7 | *Negative | 7 | 100.00% |
| BD Veritor Strep A | 33 | *Negative | 32 | 96.97% |
| z DO NOT USE Alere Acceava | 1 | *Negative | 1 | 100.00% |
| Alere i Strep A | 4 | *Negative | 4 | 100.00% |
| Roche Cobas Liat Strep A | 2 | *Negative | 2 | 100.00% |
| Sekisui Diagnostics OSOM Ultra Strep | 51 | Positive | 2 | 3.92% |



| | | | | |
|--|------|--|--------|---------|
| Henry Schein OneStep+ Rapid Strep Dip | 63 | Positive | 1 | 1.59% |
| Quidel QuickVue In-Line - 25 Test | 84 | Positive | 2 | 2.38% |
| Alere Clearview Exact II Dip | 30 | Positive | 1 | 3.33% |
| Consult Diag Strep A Dipstick | 126 | Positive | 3 | 2.38% |
| BD Veritor Strep A | 33 | Positive | 1 | 3.03% |
| RS-7 | | * Denotes Correct Response | | |
| AG: All Methods | 37 | | | |
| | | *Positive | 29 | 78.38% |
| | | Negative | 8 | 21.62% |
| Alere Aceava Strep A Dipstick | 1 | *Positive | 1 | 100.00% |
| Sekisui Diagnostics OSOM | 4 | *Positive | 2 | 50.00% |
| Sekisui Diagnostics OSOM Ultra Strep | 5 | *Positive | 4 | 80.00% |
| Innovacon One Step Strep Ag | 1 | *Positive | 1 | 100.00% |
| McKesson Strep A Dip | 2 | *Positive | 2 | 100.00% |
| Quidel QuickVue Dipstick Strep | 3 | *Positive | 3 | 100.00% |
| Quidel QuickVue In-Line - 25 Test | 2 | *Positive | 1 | 50.00% |
| Quidel QuickVue+ Strep A | 1 | *Positive | 1 | 100.00% |
| SP Brand (Cardinal) Rapid Strep A Dipstick | 7 | *Positive | 3 | 42.86% |
| Consult Diag Strep A Dipstick | 8 | *Positive | 8 | 100.00% |
| Consult Diag Strep A Twist | 1 | *Positive | 1 | 100.00% |
| MooreBrand Strep A Dipstick | 1 | *Positive | 1 | 100.00% |
| Alere i Strep A | 1 | *Positive | 1 | 100.00% |
| Sekisui Diagnostics OSOM | 4 | Negative | 2 | 50.00% |
| Sekisui Diagnostics OSOM Ultra Strep | 5 | Negative | 1 | 20.00% |
| Quidel QuickVue In-Line - 25 Test | 2 | Negative | 1 | 50.00% |
| SP Brand (Cardinal) Rapid Strep A Dipstick | 7 | Negative | 4 | 57.14% |
| RS-8 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Negative | 1 | 100.00% |
| Quidel QuickVue In-Line - 25 Test | 1 | *Negative | 1 | 100.00% |
| RS-9 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Positive | 1 | 100.00% |
| Quidel QuickVue In-Line - 25 Test | 1 | *Positive | 1 | 100.00% |
| RS-10 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Negative | 1 | 100.00% |
| Quidel QuickVue In-Line - 25 Test | 1 | *Negative | 1 | 100.00% |
| | | Strep A Ag Det. (International) | | |
| RS-6 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Negative | 1 | 100.00% |
| Alere Clearview Exact II Cas | 1 | *Negative | 1 | 100.00% |
| RS-7 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Positive | 1 | 100.00% |
| Alere Clearview Exact II Cas | 1 | *Positive | 1 | 100.00% |
| RS-8 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Negative | 1 | 100.00% |
| Alere Clearview Exact II Cas | 1 | *Negative | 1 | 100.00% |
| RS-9 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Positive | 1 | 100.00% |
| Alere Clearview Exact II Cas | 1 | *Positive | 1 | 100.00% |
| RS-10 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Negative | 1 | 100.00% |
| Alere Clearview Exact II Cas | 1 | *Negative | 1 | 100.00% |
| 997 - CVS Group A Strep - 2 samples | Labs | Response | Number | Percent |
| | | Strep A Ag Det. (Waived) | | |
| RS-6 | | * Denotes Correct Response | | |
| AG: All Methods | 38 | | | |
| | | *Negative | 38 | 100.00% |
| Sekisui Diagnostics OSOM Ultra Strep | 38 | *Negative | 38 | 100.00% |
| RS-7 | | * Denotes Correct Response | | |
| AG: All Methods | 38 | | | |
| | | *Positive | 38 | 100.00% |



| Sekisui Diagnostics OSOM Ultra Strep | 38 | *Positive | 38 | 100.00% |
|--------------------------------------|------|------------------|--------|---------|
| 739 - Sickle Cell Screen | Labs | Response | Number | Percent |
| Sickle Cell Screen | | | | |
| SS-2 | | | | |
| * Denotes Correct Response | | | | |
| AG: All Methods | 12 | | | |
| | | *Positive | 11 | 91.67% |
| | | Negative | 1 | 8.33% |
| Pacific Hemostasis Sickle Screen | 3 | *Positive | 3 | 100.00% |
| Streck SickleDex | 8 | *Positive | 7 | 87.50% |
| ASI Sickle Cell Test | 1 | *Positive | 1 | 100.00% |
| Streck SickleDex | 8 | Negative | 1 | 12.50% |
| 679 - Syphilis Serology | | | | |
| Labs | | | | |
| Response | | | | |
| RPR | | | | |
| * Denotes Correct Response | | | | |
| SY-6 | | | | |
| AG: All Methods | 9 | | | |
| | | *Nonreactive | 9 | 100.00% |
| | | Nonreactive | 9 | 100.00% |
| ASI RPR | 2 | *Nonreactive | 2 | 100.00% |
| Becton Dickinson BBL | 3 | *Nonreactive | 3 | 100.00% |
| Stanbio Quicktest | 1 | *Nonreactive | 1 | 100.00% |
| Wampole Impact | 1 | *Nonreactive | 1 | 100.00% |
| Intl ASI RPR | 1 | *Nonreactive | 1 | 100.00% |
| New Horizons | 1 | Nonreactive | 1 | 100.00% |
| * Denotes Correct Response | | | | |
| SY-7 | | | | |
| AG: All Methods | 9 | | | |
| | | *Reactive | 8 | 88.89% |
| | | *Weakly Reactive | 1 | 11.11% |
| ASI RPR | 2 | *Reactive | 2 | 100.00% |
| Becton Dickinson BBL | 3 | *Reactive | 2 | 66.67% |
| New Horizons | 1 | *Reactive | 1 | 100.00% |
| Stanbio Quicktest | 1 | *Reactive | 1 | 100.00% |
| Wampole Impact | 1 | *Reactive | 1 | 100.00% |
| Intl ASI RPR | 1 | *Reactive | 1 | 100.00% |
| Becton Dickinson BBL | 3 | *Weakly Reactive | 1 | 33.33% |
| * Denotes Correct Response | | | | |
| SY-8 | | | | |
| AG: All Methods | 9 | | | |
| | | *Nonreactive | 9 | 100.00% |
| ASI RPR | 2 | *Nonreactive | 2 | 100.00% |
| Becton Dickinson BBL | 3 | *Nonreactive | 3 | 100.00% |
| New Horizons | 1 | *Nonreactive | 1 | 100.00% |
| Stanbio Quicktest | 1 | *Nonreactive | 1 | 100.00% |
| Wampole Impact | 1 | *Nonreactive | 1 | 100.00% |
| Intl ASI RPR | 1 | *Nonreactive | 1 | 100.00% |
| * Denotes Correct Response | | | | |
| SY-9 | | | | |
| AG: All Methods | 9 | | | |
| | | *Reactive | 7 | 77.78% |
| | | *Weakly Reactive | 2 | 22.22% |
| ASI RPR | 2 | *Reactive | 2 | 100.00% |
| Becton Dickinson BBL | 3 | *Reactive | 2 | 66.67% |
| New Horizons | 1 | *Reactive | 1 | 100.00% |
| Stanbio Quicktest | 1 | *Reactive | 1 | 100.00% |
| Intl ASI RPR | 1 | *Reactive | 1 | 100.00% |
| Becton Dickinson BBL | 3 | *Weakly Reactive | 1 | 33.33% |
| Wampole Impact | 1 | *Weakly Reactive | 1 | 100.00% |
| * Denotes Correct Response | | | | |
| SY-10 | | | | |
| AG: All Methods | 9 | | | |
| | | *Reactive | 9 | 100.00% |
| ASI RPR | 2 | *Reactive | 2 | 100.00% |
| Becton Dickinson BBL | 3 | *Reactive | 3 | 100.00% |
| New Horizons | 1 | *Reactive | 1 | 100.00% |
| Stanbio Quicktest | 1 | *Reactive | 1 | 100.00% |
| Wampole Impact | 1 | *Reactive | 1 | 100.00% |
| Intl ASI RPR | 1 | *Reactive | 1 | 100.00% |
| RPR/VDRL (titer) | | | | |
| * Denotes Correct Response | | | | |
| SY-6 | | | | |
| AG: All Methods | 2 | | | |
| | | *N/A (Negative) | 1 | 50.00% |
| | | *Nonreactive | 1 | 50.00% |



| | | | | |
|---|-------------|-------------------------------------|---------------|----------------|
| ASI RPR | 1 | *N/A (Negative) | 1 | 100.00% |
| New Horizons | 1 | *Nonreactive | 1 | 100.00% |
| SY-7 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *1:2 | 2 | 100.00% |
| ASI RPR | 1 | *1:2 | 1 | 100.00% |
| New Horizons | 1 | *1:2 | 1 | 100.00% |
| SY-8 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *N/A (Negative) | 1 | 50.00% |
| | | *Nonreactive | 1 | 50.00% |
| ASI RPR | 1 | *N/A (Negative) | 1 | 100.00% |
| New Horizons | 1 | *Nonreactive | 1 | 100.00% |
| SY-9 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *1:1 | 1 | 50.00% |
| | | *1:2 | 1 | 50.00% |
| New Horizons | 1 | *1:1 | 1 | 100.00% |
| ASI RPR | 1 | *1:2 | 1 | 100.00% |
| SY-10 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *1:2 | 1 | 50.00% |
| | | *1:4 | 1 | 50.00% |
| New Horizons | 1 | *1:2 | 1 | 100.00% |
| ASI RPR | 1 | *1:4 | 1 | 100.00% |
| | | RPR (International) | | |
| SY-6 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Nonreactive | 3 | 100.00% |
| BD Macro-Vue RPR | 1 | *Nonreactive | 1 | 100.00% |
| Fisher HC Sure-Vue RPR | 1 | *Nonreactive | 1 | 100.00% |
| Wampole Impact | 1 | *Nonreactive | 1 | 100.00% |
| SY-7 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Reactive | 3 | 100.00% |
| BD Macro-Vue RPR | 1 | *Reactive | 1 | 100.00% |
| Fisher HC Sure-Vue RPR | 1 | *Reactive | 1 | 100.00% |
| Wampole Impact | 1 | *Reactive | 1 | 100.00% |
| SY-8 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Nonreactive | 3 | 100.00% |
| BD Macro-Vue RPR | 1 | *Nonreactive | 1 | 100.00% |
| Fisher HC Sure-Vue RPR | 1 | *Nonreactive | 1 | 100.00% |
| Wampole Impact | 1 | *Nonreactive | 1 | 100.00% |
| SY-9 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Reactive | 3 | 100.00% |
| BD Macro-Vue RPR | 1 | *Reactive | 1 | 100.00% |
| Fisher HC Sure-Vue RPR | 1 | *Reactive | 1 | 100.00% |
| Wampole Impact | 1 | *Reactive | 1 | 100.00% |
| SY-10 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Reactive | 3 | 100.00% |
| BD Macro-Vue RPR | 1 | *Reactive | 1 | 100.00% |
| Fisher HC Sure-Vue RPR | 1 | *Reactive | 1 | 100.00% |
| Wampole Impact | 1 | *Reactive | 1 | 100.00% |
| 776 - Throat Culture;780 - Throat & Genital Cultures | Labs | Response | Number | Percent |
| TC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 78 | | | |
| | | *Negative for Group A Streptococcus | 67 | 85.90% |
| | | *No Growth - Strep Selective medium | 10 | 12.82% |
| | | Positive for Group A Streptococcus | 1 | 1.28% |
| Cultures | 78 | *Negative for Group A Streptococcus | 67 | 85.90% |
| Cultures | | *No Growth - Strep Selective medium | 10 | 12.82% |
| Cultures | | Positive for Group A Streptococcus | 1 | 1.28% |
| TC-7 | | * Denotes Correct Response | | |
| AG: All Methods | 78 | | | |



| | | | | | |
|---|----|---|-----------------|---------------|----------------|
| | | *Beta hemolysis, referred for ID/confirmation | 3 | 3.85% | |
| | | *Positive for Group A Streptococcus | 35 | 44.87% | |
| | | *Presumptive Positive for Group A Streptococcus | 36 | 46.15% | |
| | | *Presumptive Streptococcus species | 1 | 1.28% | |
| | | *Streptococcus, beta-hemolytic, Group A | 1 | 1.28% | |
| | | Negative for Group A Streptococcus | 2 | 2.56% | |
| Cultures | 78 | *Beta hemolysis, referred for ID/confirmation | 3 | 3.85% | |
| Cultures | | *Positive for Group A Streptococcus | 35 | 44.87% | |
| Cultures | | *Presumptive Positive for Group A Streptococcus | 36 | 46.15% | |
| Cultures | | *Presumptive Streptococcus species | 1 | 1.28% | |
| Cultures | | *Streptococcus, beta-hemolytic, Group A | 1 | 1.28% | |
| Cultures | | Negative for Group A Streptococcus | 2 | 2.56% | |
| TC-8 | | * Denotes Correct Response | | | |
| AG: All Methods | 80 | | | | |
| | | *Negative for Group A Streptococcus | 65 | 81.25% | |
| | | *No Growth - Strep Selective medium | 11 | 13.75% | |
| | | Positive for Group A Streptococcus | 2 | 2.50% | |
| | | Presumptive Positive for Group A Streptococcus | 2 | 2.50% | |
| Cultures | 80 | *Negative for Group A Streptococcus | 65 | 81.25% | |
| Cultures | | *No Growth - Strep Selective medium | 11 | 13.75% | |
| Cultures | | Positive for Group A Streptococcus | 2 | 2.50% | |
| Cultures | | Presumptive Positive for Group A Streptococcus | 2 | 2.50% | |
| TC-9 | | * Denotes Correct Response | | | |
| AG: All Methods | 96 | | | | |
| | | *Growth, referred for identification | 1 | 1.04% | |
| | | *Negative for Group A Streptococcus | 77 | 80.21% | |
| | | *No beta hemolysis | 1 | 1.04% | |
| | | *No Growth - Strep Selective medium | 14 | 14.58% | |
| | | Positive for Group A Streptococcus | 3 | 3.13% | |
| Cultures | 96 | *Growth, referred for identification | 1 | 1.04% | |
| Cultures | | *Negative for Group A Streptococcus | 77 | 80.21% | |
| Cultures | | *No beta hemolysis | 1 | 1.04% | |
| Cultures | | *No Growth - Strep Selective medium | 14 | 14.58% | |
| Cultures | | Positive for Group A Streptococcus | 3 | 3.13% | |
| TC-10 | | * Denotes Correct Response | | | |
| AG: All Methods | 96 | | | | |
| | | *Beta hemolysis, referred for ID/confirmation | 4 | 4.17% | |
| | | *Growth, referred for identification | 1 | 1.04% | |
| | | *Positive for Group A Streptococcus | 44 | 45.83% | |
| | | *Presumptive Positive for Group A Streptococcus | 41 | 42.71% | |
| | | *Presumptive Streptococcus species | 1 | 1.04% | |
| | | *Streptococcus, beta-hemolytic, Group A | 1 | 1.04% | |
| | | Negative for Group A Streptococcus | 4 | 4.17% | |
| Cultures | 96 | *Beta hemolysis, referred for ID/confirmation | 4 | 4.17% | |
| Cultures | | *Growth, referred for identification | 1 | 1.04% | |
| Cultures | | *Positive for Group A Streptococcus | 44 | 45.83% | |
| Cultures | | *Presumptive Positive for Group A Streptococcus | 41 | 42.71% | |
| Cultures | | *Presumptive Streptococcus species | 1 | 1.04% | |
| Cultures | | *Streptococcus, beta-hemolytic, Group A | 1 | 1.04% | |
| Cultures | | Negative for Group A Streptococcus | 4 | 4.17% | |
| 977 - Trichomonas Aq Det (Waived) | | Labs | Response | Number | Percent |
| TRA-3 | | Trichomonas Aq Det (Waived) | | | |
| TRA-3 | | * Denotes Correct Response | | | |
| AG: All Methods | 82 | | | | |
| | | *Negative | 74 | 90.24% | |
| | | Positive | 8 | 9.76% | |
| Sekisui Diagnostics OSOM Trichomonas | 82 | *Negative | 74 | 90.24% | |
| Sekisui Diagnostics OSOM Trichomonas | | Positive | 8 | 9.76% | |
| TRA-4 | | * Denotes Correct Response | | | |
| AG: All Methods | 83 | | | | |
| | | *Positive | 74 | 89.16% | |
| | | Negative | 9 | 10.84% | |
| Sekisui Diagnostics OSOM Trichomonas | 83 | *Positive | 74 | 89.16% | |
| Sekisui Diagnostics OSOM Trichomonas | | Negative | 9 | 10.84% | |
| 640 - Urinalysis;642 - Urinalysis/hCG;639 - Complete Urinalysis Package;601 - Waived/PPM Tests II;609 - Waived/PPM Tests I;647 - Mini-Waived Package;628 - AFC Waived/PPM;999 - CVS Complete Urinalysis | | Labs | Response | Number | Percent |
| | | Urine Sp Gravity | | | |



| UA-2 | | * Denotes Correct Response | | |
|---|-----|----------------------------|-----|---------|
| PG: Siemens Clinitek 10 / 100 | 14 | | | |
| | | *1.010 | 3 | 21.43% |
| | | *1.020 | 2 | 14.29% |
| | | *1.025 | 1 | 7.14% |
| | | *1.030 | 6 | 42.86% |
| | | 1.005 | 1 | 7.14% |
| | | 1.035 | 1 | 7.14% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Siemens Clinitek 50 | 94 | | | |
| | | *1.025 | 2 | 2.13% |
| | | *1.030 | 82 | 87.23% |
| | | *1.035 | 10 | 10.64% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Siemens Clinitek 500 | 3 | | | |
| | | *1.030 | 2 | 66.67% |
| | | *1.035 | 1 | 33.33% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Siemens Clinitek Atlas | 1 | | | |
| | | *1.030 | 1 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Siemens Clinitek Status | 626 | | | |
| | | *1.025 | 8 | 1.28% |
| | | *1.030 | 576 | 92.01% |
| | | *1.035 | 41 | 6.55% |
| | | 1.020 | 1 | 0.16% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Siemens Multistix - Visual | 261 | | | |
| | | *1.020 | 76 | 29.12% |
| | | *1.025 | 78 | 29.89% |
| | | *1.030 | 92 | 35.25% |
| | | 1.000 | 2 | 0.77% |
| | | 1.005 | 1 | 0.38% |
| | | 1.010 | 3 | 1.15% |
| | | 1.015 | 5 | 1.92% |
| | | 1.035 | 4 | 1.53% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Cole Taylor Strips | 1 | | | |
| | | *1.030 | 1 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Germaine Aimstick | 4 | | | |
| | | *1.030 | 4 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Henry Schein Urispec | 6 | | | |
| | | *1.015 | 1 | 16.67% |
| | | *1.020 | 4 | 66.67% |
| | | *1.025 | 1 | 16.67% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Hypoguard DiaScreen - Visual | 2 | | | |
| | | *1.020 | 1 | 50.00% |
| | | *1.025 | 1 | 50.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Iris Diagnostics vChem Strips | 1 | | | |
| | | *1.020 | 1 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: PSS Select Optima II Urine Analyzer | 1 | | | |
| | | *1.030 | 1 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: PSS Consult/Select Reagent Strips | 5 | | | |
| | | *1.020 | 2 | 40.00% |
| | | *1.025 | 2 | 40.00% |
| | | *1.030 | 1 | 20.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Quidel QuickVue UrinChek - Visual | 1 | | | |
| | | *1.020 | 1 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Diag Test Group Clarity Urochek | 33 | | | |
| | | *1.020 | 4 | 12.12% |



| | | | | |
|---|----|-----------------------------------|----|---------|
| | | *1.025 | 5 | 15.15% |
| | | *1.030 | 24 | 72.73% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Roche Chemstrip 101 Analyzer | 3 | | | |
| | | *1.010 | 2 | 66.67% |
| | | 1.030 | 1 | 33.33% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Roche Chemstrips - Visual | 61 | | | |
| | | *1.005 | 19 | 31.15% |
| | | *1.010 | 32 | 52.46% |
| | | *1.015 | 3 | 4.92% |
| | | 1.000 | 3 | 4.92% |
| | | 1.025 | 2 | 3.28% |
| | | 1.030 | 2 | 3.28% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Roche Criterion Analyzer | 1 | | | |
| | | *1.010 | 1 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Roche SuperUA/ChemstripUA | 3 | | | |
| | | *1.015 | 1 | 33.33% |
| | | *1.020 | 1 | 33.33% |
| | | 1.000 | 1 | 33.33% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Roche Urisys | 52 | | | |
| | | *1.010 | 40 | 76.92% |
| | | *1.015 | 8 | 15.38% |
| | | 1.005 | 1 | 1.92% |
| | | 1.030 | 3 | 5.77% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Teco (TC) Diag. URS-10 | 1 | | | |
| | | *1.020 | 1 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: UriScan Reagent Strips | 5 | | | |
| | | *1.020 | 1 | 20.00% |
| | | *1.025 | 1 | 20.00% |
| | | *1.030 | 2 | 40.00% |
| | | 1.010 | 1 | 20.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Medi-Lab Performance Urine Strip Analyzer | 1 | | | |
| | | *1.030 | 1 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Jant Pharm Accutest URS 10 | 2 | | | |
| | | *1.020 | 2 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Medi-Lab Performance Urine Strip - Visual | 2 | | | |
| | | *1.020 | 1 | 50.00% |
| | | *1.030 | 1 | 50.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Siemens Clinitek Advantus | 17 | | | |
| | | *1.025 | 4 | 23.53% |
| | | *1.030 | 12 | 70.59% |
| | | 1.020 | 1 | 5.88% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Consult Diag Urinalysis Test Strips | 14 | | | |
| | | *1.025 | 2 | 14.29% |
| | | *1.030 | 11 | 78.57% |
| | | 1.020 | 1 | 7.14% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Henry Schein One Step Plus Strip Reader | 2 | | | |
| | | *1.020 | 1 | 50.00% |
| | | *1.030 | 1 | 50.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Consult Diag Urine Analyzer | 11 | | | |
| | | *1.020 | 4 | 36.36% |
| | | *1.025 | 4 | 36.36% |
| | | *1.030 | 3 | 27.27% |
| UA-2 | | * Denotes Correct Response | | |
| PG: McKesson 120 Urine Analyzer | 36 | | | |



| | | | | |
|---|------|-------------------------------|------|---------|
| | | *1.025 | 5 | 13.89% |
| | | *1.030 | 30 | 83.33% |
| | | 1.020 | 1 | 2.78% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Medline Urinalysis Reagent Strips | 2 | | | |
| | | *1.015 | 1 | 50.00% |
| | | *1.025 | 1 | 50.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: NDC ProAdvantage Urine Strip Analyzer | 2 | | | |
| | | *1.025 | 1 | 50.00% |
| | | *1.030 | 1 | 50.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Roche Cobas u 411 | 1 | | | |
| | | *1.010 | 1 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: McKesson 10 SG Urine Dipstick | 2 | | | |
| | | *1.020 | 1 | 50.00% |
| | | *1.025 | 1 | 50.00% |
| UA-2 | | * Denotes Correct Response | | |
| MG: Conventional - Roche | 30 | | | |
| | | *1.005 | 1 | 3.33% |
| | | *1.010 | 26 | 86.67% |
| | | *1.015 | 3 | 10.00% |
| UA-2 | | * Denotes Correct Response | | |
| MG: Plus Sys - Siemens | 18 | | | |
| | | *>=1.030 | 18 | 100.00% |
| | | Urine Creatinine (semi-quant) | | |
| UA-2 | | * Denotes Correct Response | | |
| PG: Siemens Clinitek Status | 5 | | | |
| | | *200 mg/dL | 2 | 40.00% |
| | | *300 mg/dL | 3 | 60.00% |
| | | Urine Bilirubin | | |
| UA-2 | | * Denotes Correct Response | | |
| MG: Method Group | 1256 | | | |
| | | *Large (3+) | 564 | 44.90% |
| | | *Moderate (2+) | 545 | 43.39% |
| | | Negative | 20 | 1.59% |
| | | Trace/Small (1+) | 127 | 10.11% |
| UA-2 | | * Denotes Correct Response | | |
| MG: Conventional - Roche | 31 | | | |
| | | *6 mg/dL | 30 | 96.77% |
| | | Negative | 1 | 3.23% |
| UA-2 | | * Denotes Correct Response | | |
| MG: Plus Sys - Siemens | 18 | | | |
| | | *1+ | 1 | 5.56% |
| | | *2+ | 11 | 61.11% |
| | | *3+ | 6 | 33.33% |
| | | Urine Ketones | | |
| UA-2 | | * Denotes Correct Response | | |
| MG: Method Group | 1280 | | | |
| | | *Large (3+) / 80 mg/dL | 147 | 11.48% |
| | | *Moderate (2+) / 40 mg/dL | 1048 | 81.88% |
| | | *Small (1+) / 15mg/dL | 82 | 6.41% |
| | | >=160 mg/dL | 1 | 0.08% |
| | | Negative | 1 | 0.08% |
| | | Trace/5 mg/dL | 1 | 0.08% |
| UA-2 | | * Denotes Correct Response | | |
| MG: Conventional - Roche | 31 | | | |
| | | *15 mg/dL | 11 | 35.48% |
| | | *50 mg/dL | 19 | 61.29% |
| | | Negative | 1 | 3.23% |
| UA-2 | | * Denotes Correct Response | | |
| MG: Plus Sys - Siemens | 18 | | | |
| | | *1+ | 1 | 5.56% |
| | | *2+ | 16 | 88.89% |
| | | 4+ | 1 | 5.56% |
| | | Urine pH | | |
| UA-2 | | * Denotes Correct Response | | |



| | | | | |
|---|-----|-----------------------------------|-----|---------|
| PG: Siemens Clinitek 10 / 100 | 14 | | | |
| | | *7.0 | 9 | 64.29% |
| | | *7.5 | 3 | 21.43% |
| | | 6.0 | 1 | 7.14% |
| | | 8.0 or greater | 1 | 7.14% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Siemens Clinitek 50 | 94 | | | |
| | | *6.5 | 2 | 2.13% |
| | | *7.0 | 88 | 93.62% |
| | | *7.5 | 4 | 4.26% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Siemens Clinitek 500 | 3 | | | |
| | | *7.0 | 2 | 66.67% |
| | | *7.5 | 1 | 33.33% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Siemens Clinitek Atlas | 1 | | | |
| | | *7.0 | 1 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Siemens Clinitek Status | 627 | | | |
| | | *6.5 | 2 | 0.32% |
| | | *7.0 | 577 | 92.03% |
| | | *7.5 | 44 | 7.02% |
| | | 6.0 | 2 | 0.32% |
| | | 8.0 or greater | 2 | 0.32% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick | 17 | | | |
| | | *7.0 | 17 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Siemens Multistix - Visual | 262 | | | |
| | | *7.0 | 84 | 32.06% |
| | | *7.5 | 161 | 61.45% |
| | | *8.0 or greater | 13 | 4.96% |
| | | 6.0 | 1 | 0.38% |
| | | 6.5 | 3 | 1.15% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Cole Taylor Strips | 1 | | | |
| | | *7.0 | 1 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Germaine Aimstick | 4 | | | |
| | | *7.0 | 4 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Henry Schein Urispec | 6 | | | |
| | | *7.0 | 6 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Hypoguard DiaScreen - Visual | 2 | | | |
| | | *7.0 | 1 | 50.00% |
| | | *7.5 | 1 | 50.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Iris Diagnostics vChem Strips | 1 | | | |
| | | *7.5 | 1 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: PSS Select Optima II Urine Analyzer | 1 | | | |
| | | *7.0 | 1 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: PSS Consult/Select Reagent Strips | 5 | | | |
| | | *6.5 | 1 | 20.00% |
| | | *7.0 | 4 | 80.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Quidel QuickVue UrinChek - Visual | 2 | | | |
| | | *7.0 | 1 | 50.00% |
| | | 6.0 | 1 | 50.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Diag Test Group Clarity Urochek | 33 | | | |
| | | *6.5 | 4 | 12.12% |
| | | *7.0 | 24 | 72.73% |
| | | *7.5 | 4 | 12.12% |
| | | 6.0 | 1 | 3.03% |



| UA-2 | | * Denotes Correct Response | | |
|---|----|----------------------------|----|---------|
| PG: Roche Chemstrip 101 Analyzer | 3 | | | |
| | | *7.0 | 2 | 66.67% |
| | | *8.0 or greater | 1 | 33.33% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Roche Chemstrips - Visual | 64 | | | |
| | | *7.0 | 18 | 28.13% |
| | | *7.5 | 5 | 7.81% |
| | | *8.0 or greater | 41 | 64.06% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Roche Criterion Analyzer | 1 | | | |
| | | *8.0 or greater | 1 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Roche SuperUA/ChemstripUA | 1 | | | |
| | | *7.0 | 1 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Roche Urisys | 52 | | | |
| | | *7.0 | 17 | 32.69% |
| | | *8.0 or greater | 35 | 67.31% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Teco (TC) Diag. URS-10 | 1 | | | |
| | | *7.0 | 1 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: UriScan Reagent Strips | 7 | | | |
| | | *6.5 | 1 | 14.29% |
| | | *7.0 | 4 | 57.14% |
| | | *7.5 | 2 | 28.57% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Medi-Lab Performance Urine Strip Analyzer | 1 | | | |
| | | *7.0 | 1 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Jant Pharm Accutest URS 10 | 2 | | | |
| | | *7.0 | 2 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Medi-Lab Performance Urine Strip - Visual | 2 | | | |
| | | *7.0 | 2 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Siemens Clinitek Advantus | 17 | | | |
| | | *7.5 | 16 | 94.12% |
| | | 6.5 | 1 | 5.88% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Consult Diag Urinalysis Test Strips | 14 | | | |
| | | *7.0 | 12 | 85.71% |
| | | *7.5 | 2 | 14.29% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Henry Schein One Step Plus Strip Reader | 2 | | | |
| | | *7.0 | 2 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Consult Diag Urine Analyzer | 11 | | | |
| | | *6.5 | 6 | 54.55% |
| | | *7.0 | 5 | 45.45% |
| UA-2 | | * Denotes Correct Response | | |
| PG: McKesson 120 Urine Analyzer | 36 | | | |
| | | *6.5 | 5 | 13.89% |
| | | *7.0 | 26 | 72.22% |
| | | *7.5 | 5 | 13.89% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Medline Urinalysis Reagent Strips | 2 | | | |
| | | *7.0 | 1 | 50.00% |
| | | *7.5 | 1 | 50.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: NDC ProAdvantage Urine Strip Analyzer | 2 | | | |
| | | *7.0 | 2 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Roche Cobas u 411 | 1 | | | |
| | | *7.0 | 1 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |



| | | | | |
|-----------------------------------|------|------------------------------|------|---------|
| PG: McKesson 10 SG Urine Dipstick | 2 | | | |
| | | *7.0 | 1 | 50.00% |
| | | *7.5 | 1 | 50.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Siemens Hema Combistix | 1 | | | |
| | | *8.0 or greater | 1 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| MG: Conventional - Roche | 31 | | | |
| | | *7.0 | 7 | 22.58% |
| | | *8 | 24 | 77.42% |
| UA-2 | | * Denotes Correct Response | | |
| MG: Plus Sys - Siemens | 18 | | | |
| | | *7.0 | 17 | 94.44% |
| | | *7.5 | 1 | 5.56% |
| | | Urine Blood/Hemoglobin | | |
| UA-2 | | * Denotes Correct Response | | |
| MG: Method Group | 1285 | | | |
| | | *Large (3+) | 1196 | 93.07% |
| | | *Moderate (2+) | 79 | 6.15% |
| | | Negative | 5 | 0.39% |
| | | Small (1+) | 3 | 0.23% |
| | | Trace | 2 | 0.16% |
| UA-2 | | * Denotes Correct Response | | |
| MG: Conventional - Roche | 31 | | | |
| | | *250 Ery/uL | 31 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| MG: Plus Sys - Siemens | 18 | | | |
| | | *3+ | 18 | 100.00% |
| | | Urine Glucose (Reducing Sub) | | |
| UA-2 | | * Denotes Correct Response | | |
| MG: Method Group | 1292 | | | |
| | | *Negative | 1290 | 99.85% |
| | | <100 mg/dL | 2 | 0.15% |
| UA-2 | | * Denotes Correct Response | | |
| MG: Conventional - Roche | 31 | | | |
| | | *Normal | 31 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| MG: Plus Sys - Siemens | 18 | | | |
| | | *Negative | 18 | 100.00% |
| | | Urine Protein - Qual | | |
| UA-2 | | * Denotes Correct Response | | |
| MG: Method Group | 1294 | | | |
| | | *>= 1000 (4+) | 26 | 2.01% |
| | | *100 (2+) | 353 | 27.28% |
| | | *300-500 (3+) | 897 | 69.32% |
| | | 30 (1+) | 13 | 1.00% |
| | | Negative | 3 | 0.23% |
| | | Trace | 2 | 0.15% |
| UA-2 | | * Denotes Correct Response | | |
| MG: Conventional - Roche | 31 | | | |
| | | *100 mg/dL | 28 | 90.32% |
| | | *500 mg/dL | 3 | 9.68% |
| UA-2 | | * Denotes Correct Response | | |
| MG: Plus Sys - Siemens | 18 | | | |
| | | *2+ | 1 | 5.56% |
| | | *3+ | 16 | 88.89% |
| | | 1+ | 1 | 5.56% |
| | | Urine Urobilinogen | | |
| UA-2 | | * Denotes Correct Response | | |
| MG: Method Group | 1246 | | | |
| | | *1.0 mg/dL | 101 | 8.11% |
| | | *2.0 mg/dL | 694 | 55.70% |
| | | *4.0 mg/dL | 270 | 21.67% |
| | | >/= 8.0 mg/dL | 48 | 3.85% |
| | | 0.2/Normal | 133 | 10.67% |
| UA-2 | | * Denotes Correct Response | | |
| MG: Conventional - Roche | 31 | | | |
| | | *4.0 mg/dL | 26 | 83.87% |



| | | | | |
|--------------------------------|------|-----------------------------------|------|---------|
| | | *8 mg/dL | 3 | 9.68% |
| | | >=12 mg/dL | 1 | 3.23% |
| | | Normal | 1 | 3.23% |
| UA-2 | | * Denotes Correct Response | | |
| MG: Plus Sys - Siemens | 16 | | | |
| | | *16 umol/L | 1 | 6.25% |
| | | *2.0 EU/dL | 5 | 31.25% |
| | | *33 umol/L | 6 | 37.50% |
| | | *4.0 EU/dL | 2 | 12.50% |
| | | 3.2 umol/L | 2 | 12.50% |
| UA-2 | | Urine Leuk Esterase | | |
| UA-2 | | * Denotes Correct Response | | |
| MG: Method Group | 1283 | | | |
| | | *Negative | 1275 | 99.38% |
| | | Large (3+) | 3 | 0.23% |
| | | Small (1+) | 1 | 0.08% |
| | | Trace | 4 | 0.31% |
| UA-2 | | * Denotes Correct Response | | |
| MG: Conventional - Roche | 31 | | | |
| | | *Negative | 31 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| MG: Plus Sys - Siemens | 18 | | | |
| | | *Negative | 18 | 100.00% |
| UA-2 | | Urine Nitrite | | |
| UA-2 | | * Denotes Correct Response | | |
| MG: Method Group | 1280 | | | |
| | | *Negative | 1229 | 96.02% |
| | | Positive | 51 | 3.98% |
| UA-2 | | * Denotes Correct Response | | |
| MG: Conventional - Roche | 31 | | | |
| | | *Negative | 31 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| MG: Plus Sys - Siemens | 18 | | | |
| | | *Negative | 18 | 100.00% |
| UA-2 | | Urine Microalbumin | | |
| UA-2 | | * Denotes Correct Response | | |
| PG: Siemens Clinitek 10 / 100 | 2 | | | |
| | | *150 mg/L | 2 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Siemens Clinitek 50 | 11 | | | |
| | | *150 mg/L | 11 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Siemens Clinitek Status | 43 | | | |
| | | *150 mg/L | 41 | 95.35% |
| | | *80 mg/L | 1 | 2.33% |
| | | 20/30 mg/L (Small) | 1 | 2.33% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Siemens Multistix - Visual | 4 | | | |
| | | *100 mg/L (3+) | 2 | 50.00% |
| | | *50 mg/L (2+) | 2 | 50.00% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Roche Chemstrip Micral | 61 | | | |
| | | *100 mg/L (3+) | 53 | 86.89% |
| | | *150 mg/L | 1 | 1.64% |
| | | *50 mg/L (2+) | 2 | 3.28% |
| | | 20/30 mg/L (Small) | 3 | 4.92% |
| | | Negative | 2 | 3.28% |
| UA-2 | | * Denotes Correct Response | | |
| PG: Hemocue Albumin 201 | 1 | | | |
| | | *150 mg/L | 1 | 100.00% |
| UA-2 | | * Denotes Correct Response | | |
| MG: Method Group | 123 | | | |
| | | *100 mg/L (3+) | 55 | 44.72% |
| | | *150 mg/L | 56 | 45.53% |
| | | *50 mg/L (2+) | 4 | 3.25% |
| | | *80 mg/L | 1 | 0.81% |
| | | 20/30 mg/L (Small) | 5 | 4.07% |
| | | Negative | 2 | 1.63% |



| | | Fluid pH | | |
|--|------|----------------------------|------|---------|
| UA-2 | | * Denotes Correct Response | | |
| PG: Nitrazine Paper | 8 | | | |
| | | *7.0 | 2 | 25.00% |
| | | *7.5 | 6 | 75.00% |
| UA-2 | | Urine hCG | | |
| AG: All Methods | | * Denotes Correct Response | | |
| | 1142 | | | |
| | | *Negative | 1137 | 99.56% |
| | | Positive | 5 | 0.44% |
| ASI ProPhase Plus | 1 | *Negative | 1 | 100.00% |
| Siemens Clinitek Status | 15 | *Negative | 15 | 100.00% |
| Beck Coulter ICON 20 HCG | 7 | *Negative | 7 | 100.00% |
| Beck Coulter ICON 25 hCG | 78 | *Negative | 78 | 100.00% |
| Alere Aceava HCG Basic II | 23 | *Negative | 23 | 100.00% |
| Alere Aceava hCG Combo | 5 | *Negative | 5 | 100.00% |
| Alere Aceava hCG-Urine | 20 | *Negative | 20 | 100.00% |
| Conception Tech One Step hCG | 2 | *Negative | 2 | 100.00% |
| Diag Test Group Clarity hCG Plus Urine Cassettes | 7 | *Negative | 7 | 100.00% |
| Fisher HC Sure-Vue hCG STAT | 5 | *Negative | 5 | 100.00% |
| Fisher HC Sure-Vue Urine hCG | 16 | *Negative | 16 | 100.00% |
| Fisher HC Sure-Vue U/S hCG | 6 | *Negative | 6 | 100.00% |
| Sekisui Diagnostics OSOM Urine | 15 | *Negative | 15 | 100.00% |
| Sekisui Diagnostics OSOM Card Pregnancy | 31 | *Negative | 31 | 100.00% |
| Sekisui Diagnostics OSOM hCG Combo | 49 | *Negative | 49 | 100.00% |
| Henry Schein One Step + (hCG) | 47 | *Negative | 46 | 97.87% |
| Immunostics Detector Combo | 3 | *Negative | 3 | 100.00% |
| Immunostics hCG Detector-urine | 4 | *Negative | 4 | 100.00% |
| Instant Tech. i Pregnancy | 4 | *Negative | 4 | 100.00% |
| Inverness Clearview Combo II | 4 | *Negative | 4 | 100.00% |
| Inv Clearview Urine - Cassette | 4 | *Negative | 4 | 100.00% |
| Inverness Signify hCG combo | 4 | *Negative | 4 | 100.00% |
| Inverness Signify hCG urine | 4 | *Negative | 4 | 100.00% |
| LifeSign Status hCG | 1 | *Negative | 1 | 100.00% |
| McKesson Medi-Lab hCG - Urine | 41 | *Negative | 41 | 100.00% |
| Meridian Immunocard Stat! | 1 | *Negative | 1 | 100.00% |
| Permaxim RediScreen | 4 | *Negative | 4 | 100.00% |
| Polymedco Poly stat hCG Urine | 1 | *Negative | 1 | 100.00% |
| Quidel QuickVue + hCG | 28 | *Negative | 28 | 100.00% |
| Quidel QuickVue One-Step hCG Combo U/S | 87 | *Negative | 87 | 100.00% |
| Quidel QuickVue One-Step Urine | 177 | *Negative | 176 | 99.44% |
| Quidel RapidVue hCG - Urine | 8 | *Negative | 8 | 100.00% |
| SAS Ultra HCG | 3 | *Negative | 3 | 100.00% |
| Select Diag hCG Cassette (urine) | 24 | *Negative | 24 | 100.00% |
| Select Diag hCG Combo | 11 | *Negative | 11 | 100.00% |
| Select Diag hCG Dipstick | 7 | *Negative | 6 | 85.71% |
| SP (Cardinal) Brand combo | 16 | *Negative | 16 | 100.00% |
| SP (Cardinal) Brand-cassette | 7 | *Negative | 7 | 100.00% |
| SP Brand (Cardinal) HCG Combo | 37 | *Negative | 37 | 100.00% |
| Stanbio Qupid | 6 | *Negative | 6 | 100.00% |
| Stanbio QuPID Plus | 1 | *Negative | 1 | 100.00% |
| Stanbio QuStick | 2 | *Negative | 2 | 100.00% |
| Stanbio TRUE hCG | 4 | *Negative | 4 | 100.00% |
| MooreBrand hCG Cassette | 4 | *Negative | 4 | 100.00% |
| Nova Scientific Century S/U hCG | 1 | *Negative | 1 | 100.00% |
| Diag Test Group Clarity hCG Urine Strips | 2 | *Negative | 2 | 100.00% |
| Jant AccuTest | 7 | *Negative | 7 | 100.00% |
| Becton Dickinson Precise One-Step | 1 | *Negative | 1 | 100.00% |
| Consult Diag hCG Urine Cassette | 183 | *Negative | 181 | 98.91% |
| CLIAwaived Inc Pregnancy Urine Dip-Strip | 5 | *Negative | 5 | 100.00% |
| BTNX Rapid Response hCG Strip (urine) | 1 | *Negative | 1 | 100.00% |
| Germaine AIM Step Combo Pregnancy | 1 | *Negative | 1 | 100.00% |
| NDC ProAdvantage Urine hCG Cassette | 4 | *Negative | 4 | 100.00% |
| Consult Diag hCG Urine Dipstick | 31 | *Negative | 31 | 100.00% |
| SP (Cardinal) Brand Dipstick | 9 | *Negative | 9 | 100.00% |
| Consult Diag hCG Combo Cassette (Urine) | 10 | *Negative | 10 | 100.00% |
| McKesson Medi-Lab Performance hCG Urine Cassette | 8 | *Negative | 8 | 100.00% |
| Alere hCG Combo Cassette (25 mIU/mL) | 3 | *Negative | 3 | 100.00% |
| Alere hCG Cassette (25 mIU/mL) | 3 | *Negative | 3 | 100.00% |



| | | | | |
|---|------|--------------------------------------|--------|---------|
| Alere hCG Combo Cassette (20/10 mIU/mL) | 1 | *Negative | 1 | 100.00% |
| Alere hCG Dipstick | 5 | *Negative | 5 | 100.00% |
| Medline Pregnancy Urine Dip-Strip | 2 | *Negative | 2 | 100.00% |
| Medline Pregnancy Urine Cassette Test | 1 | *Negative | 1 | 100.00% |
| McKesson hCG Urine Test Dipstick | 5 | *Negative | 5 | 100.00% |
| Quidel QuickVue hCG Urine Test | 17 | *Negative | 17 | 100.00% |
| MooreBrand hCG Dipstick | 3 | *Negative | 3 | 100.00% |
| Siemens Clinitest hCG | 1 | *Negative | 1 | 100.00% |
| Henry Schein One Step + HCG Combo | 1 | *Negative | 1 | 100.00% |
| Henry Schein One Step hCG Urine Cassette | 12 | *Negative | 12 | 100.00% |
| NCS hCG One Step Pregnancy Test | 1 | *Negative | 1 | 100.00% |
| Henry Schein One Step + (hCG) | 47 | Positive | 1 | 2.13% |
| Quidel QuickVue One-Step Urine | 177 | Positive | 1 | 0.56% |
| Select Diag hCG Dipstick | 7 | Positive | 1 | 14.29% |
| Consult Diag hCG Urine Cassette | 183 | Positive | 2 | 1.09% |
| 661 - Urine Colony Count;784 - Presump ID/Colony Count - Urine;785 - Presump ID/Colony Count-Urine & Strep Combo;775 - Urine Culture;778 - Urine & Throat Cultures;779 - Urine & Genital Cultures;781 - Urine, Throat & Genital Cultures;782 - Complete Bacteri | Labs | Response | Number | Percent |
| | | Susceptibility Cefazolin | | |
| UC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Sensitive | 3 | 100.00% |
| Susceptibility | 3 | *Sensitive | 3 | 100.00% |
| | | Susceptibility Cefepime | | |
| UC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Sensitive | 3 | 100.00% |
| Susceptibility | 3 | *Sensitive | 3 | 100.00% |
| | | Susceptibility Cefotaxime | | |
| UC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Sensitive | 3 | 100.00% |
| Susceptibility | 3 | *Sensitive | 3 | 100.00% |
| | | Urine Culture | | |
| UC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 32 | | | |
| | | *Escherichia coli | 4 | 12.50% |
| | | *Gram Negative organism | 4 | 12.50% |
| | | *Growth, referred for identification | 14 | 43.75% |
| | | *Presumptive E. coli | 3 | 9.38% |
| | | *Uricult - Growth (Positive) | 7 | 21.88% |
| Cultures | 32 | *Escherichia coli | 4 | 12.50% |
| Cultures | | *Gram Negative organism | 4 | 12.50% |
| Cultures | | *Growth, referred for identification | 14 | 43.75% |
| Cultures | | *Presumptive E. coli | 3 | 9.38% |
| Cultures | | *Uricult - Growth (Positive) | 7 | 21.88% |
| | | * Denotes Correct Response | | |
| AG: All Methods | 32 | | | |
| | | *Enterococcus faecalis | 3 | 9.38% |
| | | *Gram Positive organism | 2 | 6.25% |
| | | *Growth, referred for identification | 17 | 53.13% |
| | | *Presumptive Enterococcus species | 1 | 3.13% |
| | | *Uricult - Growth (Positive) | 7 | 21.88% |
| | | No Growth | 1 | 3.13% |
| | | Staphylococcus, coagulase positive | 1 | 3.13% |
| Cultures | 32 | *Enterococcus faecalis | 3 | 9.38% |
| Cultures | | *Gram Positive organism | 2 | 6.25% |
| Cultures | | *Growth, referred for identification | 17 | 53.13% |
| Cultures | | *Presumptive Enterococcus species | 1 | 3.13% |
| Cultures | | *Uricult - Growth (Positive) | 7 | 21.88% |
| Cultures | | No Growth | 1 | 3.13% |
| Cultures | | Staphylococcus, coagulase positive | 1 | 3.13% |
| | | * Denotes Correct Response | | |
| AG: All Methods | 30 | | | |
| | | *Gram Negative organism | 4 | 13.33% |
| | | *Growth, referred for identification | 16 | 53.33% |
| | | *Klebsiella pneumoniae | 2 | 6.67% |



| | | | | |
|-----------------|----|--------------------------------------|----|---------|
| | | *Presumptive Klebsiella species | 1 | 3.33% |
| | | *Uricult - Growth (Positive) | 6 | 20.00% |
| | | Uricult - No Growth (Negative) | 1 | 3.33% |
| Cultures | 30 | *Gram Negative organism | 4 | 13.33% |
| Cultures | | *Growth, referred for identification | 16 | 53.33% |
| Cultures | | *Klebsiella pneumoniae | 2 | 6.67% |
| Cultures | | *Presumptive Klebsiella species | 1 | 3.33% |
| Cultures | | *Uricult - Growth (Positive) | 6 | 20.00% |
| Cultures | | Uricult - No Growth (Negative) | 1 | 3.33% |
| UC-9 | | * Denotes Correct Response | | |
| AG: All Methods | 7 | | | |
| | | *No Growth | 5 | 71.43% |
| | | *Uricult - No Growth (Negative) | 1 | 14.29% |
| | | Uricult - Growth (Positive) | 1 | 14.29% |
| Cultures | 7 | *No Growth | 5 | 71.43% |
| Cultures | | *Uricult - No Growth (Negative) | 1 | 14.29% |
| Cultures | | Uricult - Growth (Positive) | 1 | 14.29% |
| UC-10 | | * Denotes Correct Response | | |
| AG: All Methods | 7 | | | |
| | | *Gram Positive organism | 1 | 14.29% |
| | | *Presumptive Staphylococcus species | 1 | 14.29% |
| | | *Staphylococcus saprophyticus | 2 | 28.57% |
| | | *Uricult - Growth (Positive) | 2 | 28.57% |
| | | Gram Negative organism | 1 | 14.29% |
| Cultures | 7 | *Gram Positive organism | 1 | 14.29% |
| Cultures | | *Presumptive Staphylococcus species | 1 | 14.29% |
| Cultures | | *Staphylococcus saprophyticus | 2 | 28.57% |
| Cultures | | *Uricult - Growth (Positive) | 2 | 28.57% |
| Cultures | | Gram Negative organism | 1 | 14.29% |
| UC-6 | | Susceptibility Cefpodoxime | | |
| UC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Sensitive | 1 | 100.00% |
| Susceptibility | 1 | *Sensitive | 1 | 100.00% |
| UC-6 | | Susceptibility Cefazidime | | |
| UC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Sensitive | 3 | 100.00% |
| Susceptibility | 3 | *Sensitive | 3 | 100.00% |
| UC-6 | | Susceptibility Ceftriaxone | | |
| UC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 6 | | | |
| | | *Sensitive | 6 | 100.00% |
| Susceptibility | 6 | *Sensitive | 6 | 100.00% |
| UC-6 | | Susceptibility Cefuroxime | | |
| UC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 5 | | | |
| | | *Sensitive | 5 | 100.00% |
| Susceptibility | 5 | *Sensitive | 5 | 100.00% |
| UC-6 | | Susceptibility Cephalothin | | |
| UC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 7 | | | |
| | | *Intermediate | 3 | 42.86% |
| | | *Sensitive | 4 | 57.14% |
| Susceptibility | 7 | *Intermediate | 3 | 42.86% |
| Susceptibility | | *Sensitive | 4 | 57.14% |
| UC-6 | | Susceptibility Ciprofloxacin | | |
| UC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 11 | | | |
| | | *Sensitive | 11 | 100.00% |
| Susceptibility | 11 | *Sensitive | 11 | 100.00% |
| UC-6 | | Susceptibility Gentamicin | | |
| UC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 6 | | | |
| | | *Sensitive | 6 | 100.00% |
| Susceptibility | 6 | *Sensitive | 6 | 100.00% |
| UC-6 | | Susceptibility Imipenem | | |
| UC-6 | | * Denotes Correct Response | | |



| | | | | |
|-----------------|----|--|----|---------|
| AG: All Methods | 4 | | | |
| | | *Sensitive | 4 | 100.00% |
| Susceptibility | 4 | *Sensitive | 4 | 100.00% |
| | | Susceptibility Levofloxacin | | |
| UC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 6 | | | |
| | | *Sensitive | 6 | 100.00% |
| Susceptibility | 6 | *Sensitive | 6 | 100.00% |
| | | Susceptibility Nitrofurantoin | | |
| UC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 10 | | | |
| | | *Sensitive | 10 | 100.00% |
| Susceptibility | 10 | *Sensitive | 10 | 100.00% |
| | | Susceptibility Norfloxacin/Noroxin | | |
| UC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Sensitive | 2 | 100.00% |
| Susceptibility | 2 | *Sensitive | 2 | 100.00% |
| | | Susceptibility Piperacillin/Tazobactam | | |
| UC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Sensitive | 3 | 100.00% |
| Susceptibility | 3 | *Sensitive | 3 | 100.00% |
| | | Susceptibility Sulfamethoxazole | | |
| UC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Sensitive | 1 | 100.00% |
| Susceptibility | 1 | *Sensitive | 1 | 100.00% |
| | | Susceptibility Tetracycline | | |
| UC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 6 | | | |
| | | *Sensitive | 6 | 100.00% |
| Susceptibility | 6 | *Sensitive | 6 | 100.00% |
| | | Susceptibility Ticarcillin/Clavulanate | | |
| UC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Sensitive | 1 | 100.00% |
| Susceptibility | 1 | *Sensitive | 1 | 100.00% |
| | | Susceptibility Tobramycin | | |
| UC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 4 | | | |
| | | *Sensitive | 4 | 100.00% |
| Susceptibility | 4 | *Sensitive | 4 | 100.00% |
| | | Susceptibility Trimethoprim | | |
| UC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Sensitive | 2 | 100.00% |
| Susceptibility | 2 | *Sensitive | 2 | 100.00% |
| | | Susceptibility Trimeth./Sulfameth. | | |
| UC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 9 | | | |
| | | *Sensitive | 9 | 100.00% |
| Susceptibility | 9 | *Sensitive | 9 | 100.00% |
| | | Susceptibility Amikacin | | |
| UC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Sensitive | 2 | 100.00% |
| Susceptibility | 2 | *Sensitive | 2 | 100.00% |
| | | Susceptibility Amox/Clavulanate | | |
| UC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 7 | | | |
| | | *Sensitive | 7 | 100.00% |
| Susceptibility | 7 | *Sensitive | 7 | 100.00% |
| | | Susceptibility Ampicillin | | |
| UC-6 | | * Denotes Correct Response | | |
| AG: All Methods | 9 | | | |
| | | *Resistant | 9 | 100.00% |



| Susceptibility | | 9 | *Resistant | 9 | 100.00% |
|--|--|-------------|-----------------------------------|---------------|----------------|
| Susceptibility Ampicillin/Sulbactam | | | | | |
| UC-6 | | | * Denotes Correct Response | | |
| AG: All Methods | | 3 | | | |
| | | | *Intermediate | 3 | 100.00% |
| Susceptibility | | 3 | *Intermediate | 3 | 100.00% |
| Susceptibility Aztreonam | | | | | |
| UC-6 | | | * Denotes Correct Response | | |
| AG: All Methods | | 1 | | | |
| | | | *Sensitive | 1 | 100.00% |
| Susceptibility | | 1 | *Sensitive | 1 | 100.00% |
| Susceptibility Ertapenem | | | | | |
| UC-6 | | | * Denotes Correct Response | | |
| AG: All Methods | | 1 | | | |
| | | | *Sensitive | 1 | 100.00% |
| Susceptibility | | 1 | *Sensitive | 1 | 100.00% |
| 703 - Urine Drug Screen;713 - Urine Drug Screen (50 ml sample volume) | | Labs | Response | Number | Percent |
| UDS-3 | | | Barbiturates | | |
| UC-6 | | | * Denotes Correct Response | | |
| AG: All Methods | | 132 | | | |
| | | | *Negative | 131 | 99.24% |
| | | | Positive | 1 | 0.76% |
| AlereTriage | | 3 | *Negative | 3 | 100.00% |
| Siemens Dim/AR/ES/RxL/Xpand | | 1 | *Negative | 1 | 100.00% |
| E-Screen | | 8 | *Negative | 8 | 100.00% |
| Express Diag. Express Check | | 1 | *Negative | 1 | 100.00% |
| Innovacon One Step | | 1 | *Negative | 1 | 100.00% |
| Alere iCup Drugs of Abuse | | 13 | *Negative | 12 | 92.31% |
| Instant Technologies i Screen | | 3 | *Negative | 3 | 100.00% |
| MedTox Profile II | | 1 | *Negative | 1 | 100.00% |
| Noble Med Split Specimen Cup | | 1 | *Negative | 1 | 100.00% |
| Beckman (Olympus) AU | | 1 | *Negative | 1 | 100.00% |
| Innovacon Multi-Clin | | 10 | *Negative | 10 | 100.00% |
| BioTek | | 12 | *Negative | 12 | 100.00% |
| MedTox Profile III ER | | 1 | *Negative | 1 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | | 5 | *Negative | 5 | 100.00% |
| ASC One Screen Drug Test Card | | 2 | *Negative | 2 | 100.00% |
| MedTox Profile V | | 2 | *Negative | 2 | 100.00% |
| Jant Accutest Multi Drug Screen | | 2 | *Negative | 2 | 100.00% |
| Amedica Biotech AmediCheck Instant Test Cup | | 4 | *Negative | 4 | 100.00% |
| Nova Century Scientific Nova Screen DOA Cassette | | 1 | *Negative | 1 | 100.00% |
| American Screening Corp Reveal Multi-Drug Test Cup | | 4 | *Negative | 4 | 100.00% |
| Mindray BS-200 | | 1 | *Negative | 1 | 100.00% |
| Alere iCassette | | 24 | *Negative | 24 | 100.00% |
| Roche Hitachi 717 | | 1 | *Negative | 1 | 100.00% |
| US Diagnostics UScreen Cup | | 5 | *Negative | 5 | 100.00% |
| Abbott Architect c4000 | | 1 | *Negative | 1 | 100.00% |
| Discover Drug Test Card | | 1 | *Negative | 1 | 100.00% |
| Bio-Rad Tox/See Rapid Urine Drug Screen | | 1 | *Negative | 1 | 100.00% |
| Diag Test Group Clarity DOA Screen | | 6 | *Negative | 6 | 100.00% |
| Noble Medical Insta-Screen Test Cup | | 3 | *Negative | 3 | 100.00% |
| Alere Immatest Multi-Drug Screen Card | | 1 | *Negative | 1 | 100.00% |
| Mindray 480 | | 1 | *Negative | 1 | 100.00% |
| Hemosure First Sign UDS Dipstick | | 10 | *Negative | 10 | 100.00% |
| Hemosure First Sign UDS Cup | | 1 | *Negative | 1 | 100.00% |
| Alere iCup Drugs of Abuse | | 13 | Positive | 1 | 7.69% |
| UDS-4 | | | * Denotes Correct Response | | |
| AG: All Methods | | 132 | | | |
| | | | *Negative | 132 | 100.00% |
| AlereTriage | | 3 | *Negative | 3 | 100.00% |
| Siemens Dim/AR/ES/RxL/Xpand | | 1 | *Negative | 1 | 100.00% |
| E-Screen | | 8 | *Negative | 8 | 100.00% |
| Express Diag. Express Check | | 1 | *Negative | 1 | 100.00% |
| Innovacon One Step | | 1 | *Negative | 1 | 100.00% |
| Alere iCup Drugs of Abuse | | 13 | *Negative | 13 | 100.00% |
| Instant Technologies i Screen | | 3 | *Negative | 3 | 100.00% |
| MedTox Profile II | | 1 | *Negative | 1 | 100.00% |



| | | | | |
|--|-----|-----------------------------------|-----|---------|
| Noble Med Split Specimen Cup | 1 | *Negative | 1 | 100.00% |
| Beckman (Olympus) AU | 1 | *Negative | 1 | 100.00% |
| Innovacon Multi-Clin | 10 | *Negative | 10 | 100.00% |
| BioTek | 12 | *Negative | 12 | 100.00% |
| MedTox Profile III ER | 1 | *Negative | 1 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 5 | *Negative | 5 | 100.00% |
| ASC One Screen Drug Test Card | 2 | *Negative | 2 | 100.00% |
| MedTox Profile V | 2 | *Negative | 2 | 100.00% |
| Jant Accutest Multi Drug Screen | 2 | *Negative | 2 | 100.00% |
| Amedica Biotech AmediCheck Instant Test Cup | 4 | *Negative | 4 | 100.00% |
| Nova Century Scientific Nova Screen DOA Cassette | 1 | *Negative | 1 | 100.00% |
| American Screening Corp Reveal Multi-Drug Test Cup | 4 | *Negative | 4 | 100.00% |
| Mindray BS-200 | 1 | *Negative | 1 | 100.00% |
| Alere iCassette | 24 | *Negative | 24 | 100.00% |
| Roche Hitachi 717 | 1 | *Negative | 1 | 100.00% |
| US Diagnostics UScreen Cup | 5 | *Negative | 5 | 100.00% |
| Abbott Architect c4000 | 1 | *Negative | 1 | 100.00% |
| Discover Drug Test Card | 1 | *Negative | 1 | 100.00% |
| Bio-Rad Tox/See Rapid Urine Drug Screen | 1 | *Negative | 1 | 100.00% |
| Diag Test Group Clarity DOA Screen | 6 | *Negative | 6 | 100.00% |
| Noble Medical Insta-Screen Test Cup | 3 | *Negative | 3 | 100.00% |
| Alere Immatest Multi-Drug Screen Card | 1 | *Negative | 1 | 100.00% |
| Mindray 480 | 1 | *Negative | 1 | 100.00% |
| Hemosure First Sign UDS Dipstick | 10 | *Negative | 10 | 100.00% |
| Hemosure First Sign UDS Cup | 1 | *Negative | 1 | 100.00% |
| | | Benzodiazepines | | |
| UDS-3 | | * Denotes Correct Response | | |
| AG: All Methods | 135 | | | |
| | | *Positive | 133 | 98.52% |
| | | Negative | 2 | 1.48% |
| AlereTriage | 3 | *Positive | 3 | 100.00% |
| Siemens Dim/AR/ES/RxL/Xpand | 2 | *Positive | 2 | 100.00% |
| E-Screen | 8 | *Positive | 8 | 100.00% |
| Express Diag. Express Check | 1 | *Positive | 1 | 100.00% |
| Innovacon One Step | 1 | *Positive | 1 | 100.00% |
| Alere iCup Drugs of Abuse | 14 | *Positive | 13 | 92.86% |
| Instant Technologies i Screen | 3 | *Positive | 3 | 100.00% |
| MedTox Profile II | 1 | *Positive | 1 | 100.00% |
| Noble Med Split Specimen Cup | 2 | *Positive | 2 | 100.00% |
| Beckman (Olympus) AU | 1 | *Positive | 1 | 100.00% |
| Innovacon Multi-Clin | 10 | *Positive | 10 | 100.00% |
| BioTek | 12 | *Positive | 12 | 100.00% |
| MedTox Profile III ER | 1 | *Positive | 1 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 5 | *Positive | 5 | 100.00% |
| ASC One Screen Drug Test Card | 2 | *Positive | 2 | 100.00% |
| MedTox Profile V | 2 | *Positive | 2 | 100.00% |
| Jant Accutest Multi Drug Screen | 2 | *Positive | 1 | 50.00% |
| Amedica Biotech AmediCheck Instant Test Cup | 4 | *Positive | 4 | 100.00% |
| Nova Century Scientific Nova Screen DOA Cassette | 1 | *Positive | 1 | 100.00% |
| American Screening Corp Reveal Multi-Drug Test Cup | 4 | *Positive | 4 | 100.00% |
| Mindray BS-200 | 1 | *Positive | 1 | 100.00% |
| Alere iCassette | 24 | *Positive | 24 | 100.00% |
| Roche Hitachi 717 | 1 | *Positive | 1 | 100.00% |
| US Diagnostics UScreen Cup | 5 | *Positive | 5 | 100.00% |
| Abbott Architect c4000 | 1 | *Positive | 1 | 100.00% |
| Discover Drug Test Card | 1 | *Positive | 1 | 100.00% |
| Bio-Rad Tox/See Rapid Urine Drug Screen | 1 | *Positive | 1 | 100.00% |
| Diag Test Group Clarity DOA Screen | 6 | *Positive | 6 | 100.00% |
| Noble Medical Insta-Screen Test Cup | 3 | *Positive | 3 | 100.00% |
| Alere Immatest Multi-Drug Screen Card | 1 | *Positive | 1 | 100.00% |
| Mindray 480 | 1 | *Positive | 1 | 100.00% |
| Hemosure First Sign UDS Dipstick | 10 | *Positive | 10 | 100.00% |
| Hemosure First Sign UDS Cup | 1 | *Positive | 1 | 100.00% |
| Alere iCup Drugs of Abuse | 14 | Negative | 1 | 7.14% |
| Jant Accutest Multi Drug Screen | 2 | Negative | 1 | 50.00% |
| UDS-4 | | * Denotes Correct Response | | |
| AG: All Methods | 135 | | | |
| | | *Negative | 135 | 100.00% |
| AlereTriage | 3 | *Negative | 3 | 100.00% |



| | | | | |
|--|-----|-----------------------------------|-----|---------|
| Siemens Dim/AR/ES/RxL/Xpand | 2 | *Negative | 2 | 100.00% |
| E-Screen | 8 | *Negative | 8 | 100.00% |
| Express Diag. Express Check | 1 | *Negative | 1 | 100.00% |
| Innovacon One Step | 1 | *Negative | 1 | 100.00% |
| Alere iCup Drugs of Abuse | 14 | *Negative | 14 | 100.00% |
| Instant Technologies i Screen | 3 | *Negative | 3 | 100.00% |
| MedTox Profile II | 1 | *Negative | 1 | 100.00% |
| Noble Med Split Specimen Cup | 2 | *Negative | 2 | 100.00% |
| Beckman (Olympus) AU | 1 | *Negative | 1 | 100.00% |
| Innovacon Multi-Clin | 10 | *Negative | 10 | 100.00% |
| BioTek | 12 | *Negative | 12 | 100.00% |
| MedTox Profile III ER | 1 | *Negative | 1 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 5 | *Negative | 5 | 100.00% |
| ASC One Screen Drug Test Card | 2 | *Negative | 2 | 100.00% |
| MedTox Profile V | 2 | *Negative | 2 | 100.00% |
| Jant Accutest Multi Drug Screen | 2 | *Negative | 2 | 100.00% |
| Amedica Biotech AmediCheck Instant Test Cup | 4 | *Negative | 4 | 100.00% |
| Nova Century Scientific Nova Screen DOA Cassette | 1 | *Negative | 1 | 100.00% |
| American Screening Corp Reveal Multi-Drug Test Cup | 4 | *Negative | 4 | 100.00% |
| Mindray BS-200 | 1 | *Negative | 1 | 100.00% |
| Alere iCassette | 24 | *Negative | 24 | 100.00% |
| Roche Hitachi 717 | 1 | *Negative | 1 | 100.00% |
| US Diagnostics UScreen Cup | 5 | *Negative | 5 | 100.00% |
| Abbott Architect c4000 | 1 | *Negative | 1 | 100.00% |
| Discover Drug Test Card | 1 | *Negative | 1 | 100.00% |
| Bio-Rad Tox/See Rapid Urine Drug Screen | 1 | *Negative | 1 | 100.00% |
| Diag Test Group Clarity DOA Screen | 6 | *Negative | 6 | 100.00% |
| Noble Medical Insta-Screen Test Cup | 3 | *Negative | 3 | 100.00% |
| Alere Immatest Multi-Drug Screen Card | 1 | *Negative | 1 | 100.00% |
| Mindray 480 | 1 | *Negative | 1 | 100.00% |
| Hemosure First Sign UDS Dipstick | 10 | *Negative | 10 | 100.00% |
| Hemosure First Sign UDS Cup | 1 | *Negative | 1 | 100.00% |
| | | Amphetamines | | |
| UDS-3 | | * Denotes Correct Response | | |
| AG: All Methods | 140 | | | |
| | | *Negative | 140 | 100.00% |
| AlereTriage | 3 | *Negative | 3 | 100.00% |
| Siemens Dim/AR/ES/RxL/Xpand | 3 | *Negative | 3 | 100.00% |
| E-Screen | 6 | *Negative | 6 | 100.00% |
| Express Diag. Express Check | 1 | *Negative | 1 | 100.00% |
| Innovacon One Step | 1 | *Negative | 1 | 100.00% |
| Alere iCup Drugs of Abuse | 17 | *Negative | 17 | 100.00% |
| Instant Technologies i Screen | 3 | *Negative | 3 | 100.00% |
| MedTox Profile II | 2 | *Negative | 2 | 100.00% |
| Noble Med Split Specimen Cup | 2 | *Negative | 2 | 100.00% |
| Beckman (Olympus) AU | 1 | *Negative | 1 | 100.00% |
| Innovacon Multi-Clin | 9 | *Negative | 9 | 100.00% |
| BioTek | 12 | *Negative | 12 | 100.00% |
| MedTox Profile III ER | 1 | *Negative | 1 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 5 | *Negative | 5 | 100.00% |
| ASC One Screen Drug Test Card | 2 | *Negative | 2 | 100.00% |
| MedTox Profile V | 2 | *Negative | 2 | 100.00% |
| Jant Accutest Multi Drug Screen | 2 | *Negative | 2 | 100.00% |
| Amedica Biotech AmediCheck Instant Test Cup | 4 | *Negative | 4 | 100.00% |
| Nova Century Scientific Nova Screen DOA Cassette | 1 | *Negative | 1 | 100.00% |
| American Screening Corp Reveal Multi-Drug Test Cup | 4 | *Negative | 4 | 100.00% |
| Mindray BS-200 | 1 | *Negative | 1 | 100.00% |
| Alere iCassette | 27 | *Negative | 27 | 100.00% |
| Roche Hitachi 717 | 1 | *Negative | 1 | 100.00% |
| US Diagnostics UScreen Cup | 5 | *Negative | 5 | 100.00% |
| Abbott Architect c4000 | 1 | *Negative | 1 | 100.00% |
| Discover Drug Test Card | 1 | *Negative | 1 | 100.00% |
| Bio-Rad Tox/See Rapid Urine Drug Screen | 1 | *Negative | 1 | 100.00% |
| Diag Test Group Clarity DOA Screen | 6 | *Negative | 6 | 100.00% |
| Noble Medical Insta-Screen Test Cup | 3 | *Negative | 3 | 100.00% |
| Alere Immatest Multi-Drug Screen Card | 1 | *Negative | 1 | 100.00% |
| Mindray 480 | 1 | *Negative | 1 | 100.00% |
| Hemosure First Sign UDS Dipstick | 10 | *Negative | 10 | 100.00% |
| Hemosure First Sign UDS Cup | 1 | *Negative | 1 | 100.00% |



| UDS-4 | | * Denotes Correct Response | | |
|--|-----|----------------------------|-----|---------|
| AG: All Methods | 140 | | | |
| | | *Positive | 136 | 97.14% |
| | | Negative | 4 | 2.86% |
| AlereTriage | 3 | *Positive | 3 | 100.00% |
| Siemens Dim/AR/ES/RxL/Xpand | 3 | *Positive | 3 | 100.00% |
| E-Screen | 6 | *Positive | 6 | 100.00% |
| Express Diag. Express Check | 1 | *Positive | 1 | 100.00% |
| Innovacon One Step | 1 | *Positive | 1 | 100.00% |
| Alere iCup Drugs of Abuse | 17 | *Positive | 17 | 100.00% |
| Instant Technologies i Screen | 3 | *Positive | 3 | 100.00% |
| MedTox Profile II | 2 | *Positive | 2 | 100.00% |
| Noble Med Split Specimen Cup | 2 | *Positive | 2 | 100.00% |
| Beckman (Olympus) AU | 1 | *Positive | 1 | 100.00% |
| Innovacon Multi-Clin | 9 | *Positive | 7 | 77.78% |
| BioTek | 12 | *Positive | 12 | 100.00% |
| MedTox Profile III ER | 1 | *Positive | 1 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 5 | *Positive | 5 | 100.00% |
| ASC One Screen Drug Test Card | 2 | *Positive | 1 | 50.00% |
| MedTox Profile V | 2 | *Positive | 2 | 100.00% |
| Jant Accutest Multi Drug Screen | 2 | *Positive | 2 | 100.00% |
| Amedica Biotech AmediCheck Instant Test Cup | 4 | *Positive | 4 | 100.00% |
| Nova Century Scientific Nova Screen DOA Cassette | 1 | *Positive | 1 | 100.00% |
| American Screening Corp Reveal Multi-Drug Test Cup | 4 | *Positive | 4 | 100.00% |
| Mindray BS-200 | 1 | *Positive | 1 | 100.00% |
| Alere iCassette | 27 | *Positive | 26 | 96.30% |
| Roche Hitachi 717 | 1 | *Positive | 1 | 100.00% |
| US Diagnostics UScreen Cup | 5 | *Positive | 5 | 100.00% |
| Abbott Architect c4000 | 1 | *Positive | 1 | 100.00% |
| Discover Drug Test Card | 1 | *Positive | 1 | 100.00% |
| Bio-Rad Tox/See Rapid Urine Drug Screen | 1 | *Positive | 1 | 100.00% |
| Diag Test Group Clarity DOA Screen | 6 | *Positive | 6 | 100.00% |
| Noble Medical Insta-Screen Test Cup | 3 | *Positive | 3 | 100.00% |
| Alere Immutest Multi-Drug Screen Card | 1 | *Positive | 1 | 100.00% |
| Mindray 480 | 1 | *Positive | 1 | 100.00% |
| Hemosure First Sign UDS Dipstick | 10 | *Positive | 10 | 100.00% |
| Hemosure First Sign UDS Cup | 1 | *Positive | 1 | 100.00% |
| Innovacon Multi-Clin | 9 | Negative | 2 | 22.22% |
| ASC One Screen Drug Test Card | 2 | Negative | 1 | 50.00% |
| Alere iCassette | 27 | Negative | 1 | 3.70% |
| | | Cannabinoids (THC) | | |
| UDS-3 | | * Denotes Correct Response | | |
| AG: All Methods | 154 | | | |
| | | *Positive | 151 | 98.05% |
| | | Negative | 3 | 1.95% |
| AlereTriage | 3 | *Positive | 3 | 100.00% |
| Siemens Dim/AR/ES/RxL/Xpand | 3 | *Positive | 3 | 100.00% |
| E-Screen | 16 | *Positive | 14 | 87.50% |
| Express Diag. Express Check | 1 | *Positive | 1 | 100.00% |
| Innovacon One Step | 1 | *Positive | 1 | 100.00% |
| Alere iCup Drugs of Abuse | 17 | *Positive | 17 | 100.00% |
| Instant Technologies i Screen | 4 | *Positive | 4 | 100.00% |
| Integrated E-Z Split Key Cup | 1 | *Positive | 1 | 100.00% |
| MedTox Profile II | 2 | *Positive | 2 | 100.00% |
| Noble Med Split Specimen Cup | 2 | *Positive | 1 | 50.00% |
| Beckman (Olympus) AU | 1 | *Positive | 1 | 100.00% |
| Innovacon Multi-Clin | 10 | *Positive | 10 | 100.00% |
| BioTek | 12 | *Positive | 12 | 100.00% |
| MedTox Profile III ER | 1 | *Positive | 1 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 5 | *Positive | 5 | 100.00% |
| ASC One Screen Drug Test Card | 2 | *Positive | 2 | 100.00% |
| MedTox Profile V | 2 | *Positive | 2 | 100.00% |
| Jant Accutest Multi Drug Screen | 2 | *Positive | 2 | 100.00% |
| Amedica Biotech AmediCheck Instant Test Cup | 4 | *Positive | 4 | 100.00% |
| Nova Century Scientific Nova Screen DOA Cassette | 1 | *Positive | 1 | 100.00% |
| American Screening Corp Reveal Multi-Drug Test Cup | 4 | *Positive | 4 | 100.00% |
| Mindray BS-200 | 1 | *Positive | 1 | 100.00% |
| Alere iCassette | 27 | *Positive | 27 | 100.00% |
| Roche Hitachi 717 | 1 | *Positive | 1 | 100.00% |



| | | | | |
|--|-----|-----------------------------------|-----|---------|
| US Diagnostics UScreen Cup | 5 | *Positive | 5 | 100.00% |
| Abbott Architect c4000 | 1 | *Positive | 1 | 100.00% |
| Discover Drug Test Card | 1 | *Positive | 1 | 100.00% |
| Bio-Rad Tox/See Rapid Urine Drug Screen | 1 | *Positive | 1 | 100.00% |
| Diag Test Group Clarity DOA Screen | 6 | *Positive | 6 | 100.00% |
| Noble Medical Insta-Screen Test Cup | 3 | *Positive | 3 | 100.00% |
| Alere Immutest Multi-Drug Screen Card | 1 | *Positive | 1 | 100.00% |
| Mindray 480 | 1 | *Positive | 1 | 100.00% |
| CRL Stat Cup | 1 | *Positive | 1 | 100.00% |
| Hemosure First Sign UDS Dipstick | 10 | *Positive | 10 | 100.00% |
| Hemosure First Sign UDS Cup | 1 | *Positive | 1 | 100.00% |
| E-Screen | 16 | Negative | 2 | 12.50% |
| Noble Med Split Specimen Cup | 2 | Negative | 1 | 50.00% |
| UDS-4 | | * Denotes Correct Response | | |
| AG: All Methods | 154 | | | |
| | | *Negative | 154 | 100.00% |
| AlereTriage | 3 | *Negative | 3 | 100.00% |
| Siemens Dim/AR/ES/RxL/Xpand | 3 | *Negative | 3 | 100.00% |
| E-Screen | 16 | *Negative | 16 | 100.00% |
| Express Diag. Express Check | 1 | *Negative | 1 | 100.00% |
| Innovacon One Step | 1 | *Negative | 1 | 100.00% |
| Alere iCup Drugs of Abuse | 17 | *Negative | 17 | 100.00% |
| Instant Technologies i Screen | 4 | *Negative | 4 | 100.00% |
| Integrated E-Z Split Key Cup | 1 | *Negative | 1 | 100.00% |
| MedTox Profile II | 2 | *Negative | 2 | 100.00% |
| Noble Med Split Specimen Cup | 2 | *Negative | 2 | 100.00% |
| Beckman (Olympus) AU | 1 | *Negative | 1 | 100.00% |
| Innovacon Multi-Clin | 10 | *Negative | 10 | 100.00% |
| BioTek | 12 | *Negative | 12 | 100.00% |
| MedTox Profile III ER | 1 | *Negative | 1 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 5 | *Negative | 5 | 100.00% |
| ASC One Screen Drug Test Card | 2 | *Negative | 2 | 100.00% |
| MedTox Profile V | 2 | *Negative | 2 | 100.00% |
| Jant Accutest Multi Drug Screen | 2 | *Negative | 2 | 100.00% |
| Amedica Biotech AmediCheck Instant Test Cup | 4 | *Negative | 4 | 100.00% |
| Nova Century Scientific Nova Screen DOA Cassette | 1 | *Negative | 1 | 100.00% |
| American Screening Corp Reveal Multi-Drug Test Cup | 4 | *Negative | 4 | 100.00% |
| Mindray BS-200 | 1 | *Negative | 1 | 100.00% |
| Alere iCassette | 27 | *Negative | 27 | 100.00% |
| Roche Hitachi 717 | 1 | *Negative | 1 | 100.00% |
| US Diagnostics UScreen Cup | 5 | *Negative | 5 | 100.00% |
| Abbott Architect c4000 | 1 | *Negative | 1 | 100.00% |
| Discover Drug Test Card | 1 | *Negative | 1 | 100.00% |
| Bio-Rad Tox/See Rapid Urine Drug Screen | 1 | *Negative | 1 | 100.00% |
| Diag Test Group Clarity DOA Screen | 6 | *Negative | 6 | 100.00% |
| Noble Medical Insta-Screen Test Cup | 3 | *Negative | 3 | 100.00% |
| Alere Immutest Multi-Drug Screen Card | 1 | *Negative | 1 | 100.00% |
| Mindray 480 | 1 | *Negative | 1 | 100.00% |
| CRL Stat Cup | 1 | *Negative | 1 | 100.00% |
| Hemosure First Sign UDS Dipstick | 10 | *Negative | 10 | 100.00% |
| Hemosure First Sign UDS Cup | 1 | *Negative | 1 | 100.00% |
| | | Cocaine Metabolites | | |
| UDS-3 | | * Denotes Correct Response | | |
| AG: All Methods | 153 | | | |
| | | *Negative | 150 | 98.04% |
| | | Positive | 3 | 1.96% |
| AlereTriage | 3 | *Negative | 3 | 100.00% |
| Siemens Dim/AR/ES/RxL/Xpand | 3 | *Negative | 3 | 100.00% |
| E-Screen | 16 | *Negative | 14 | 87.50% |
| Express Diag. Express Check | 1 | *Negative | 1 | 100.00% |
| Innovacon One Step | 1 | *Negative | 1 | 100.00% |
| Alere iCup Drugs of Abuse | 17 | *Negative | 17 | 100.00% |
| Instant Technologies i Screen | 4 | *Negative | 4 | 100.00% |
| Integrated E-Z Split Key Cup | 1 | *Negative | 1 | 100.00% |
| MedTox Profile II | 2 | *Negative | 2 | 100.00% |
| Noble Med Split Specimen Cup | 2 | *Negative | 2 | 100.00% |
| Beckman (Olympus) AU | 1 | *Negative | 1 | 100.00% |
| Innovacon Multi-Clin | 9 | *Negative | 9 | 100.00% |
| BioTek | 12 | *Negative | 12 | 100.00% |



| | | | | |
|--|-----|-----------------------------------|-----|---------|
| MedTox Profile III ER | 1 | *Negative | 1 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 5 | *Negative | 5 | 100.00% |
| ASC One Screen Drug Test Card | 2 | *Negative | 2 | 100.00% |
| MedTox Profile V | 2 | *Negative | 2 | 100.00% |
| Jant Accutest Multi Drug Screen | 2 | *Negative | 2 | 100.00% |
| Amedica Biotech AmediCheck Instant Test Cup | 4 | *Negative | 4 | 100.00% |
| Nova Century Scientific Nova Screen DOA Cassette | 1 | *Negative | 1 | 100.00% |
| American Screening Corp Reveal Multi-Drug Test Cup | 4 | *Negative | 4 | 100.00% |
| Mindray BS-200 | 1 | *Negative | 1 | 100.00% |
| Alere iCassette | 27 | *Negative | 27 | 100.00% |
| Roche Hitachi 717 | 1 | *Negative | 1 | 100.00% |
| US Diagnostics UScreen Cup | 5 | *Negative | 5 | 100.00% |
| Abbott Architect c4000 | 1 | *Negative | 1 | 100.00% |
| Discover Drug Test Card | 1 | *Negative | 1 | 100.00% |
| Bio-Rad Tox/See Rapid Urine Drug Screen | 1 | *Negative | 1 | 100.00% |
| Diag Test Group Clarity DOA Screen | 6 | *Negative | 6 | 100.00% |
| Noble Medical Insta-Screen Test Cup | 3 | *Negative | 3 | 100.00% |
| Alere Immatest Multi-Drug Screen Card | 1 | *Negative | 1 | 100.00% |
| Mindray 480 | 1 | *Negative | 1 | 100.00% |
| CRL Stat Cup | 1 | *Negative | 1 | 100.00% |
| Hemosure First Sign UDS Dipstick | 10 | *Negative | 9 | 90.00% |
| Hemosure First Sign UDS Cup | 1 | *Negative | 1 | 100.00% |
| E-Screen | 16 | Positive | 2 | 12.50% |
| Hemosure First Sign UDS Dipstick | 10 | Positive | 1 | 10.00% |
| UDS-4 | | * Denotes Correct Response | | |
| AG: All Methods | 153 | | | |
| | | *Negative | 153 | 100.00% |
| AlereTriage | 3 | *Negative | 3 | 100.00% |
| Siemens Dim/AR/ES/RxL/Xpand | 3 | *Negative | 3 | 100.00% |
| E-Screen | 16 | *Negative | 16 | 100.00% |
| Express Diag. Express Check | 1 | *Negative | 1 | 100.00% |
| Innovacon One Step | 1 | *Negative | 1 | 100.00% |
| Alere iCup Drugs of Abuse | 17 | *Negative | 17 | 100.00% |
| Instant Technologies i Screen | 4 | *Negative | 4 | 100.00% |
| Integrated E-Z Split Key Cup | 1 | *Negative | 1 | 100.00% |
| MedTox Profile II | 2 | *Negative | 2 | 100.00% |
| Noble Med Split Specimen Cup | 2 | *Negative | 2 | 100.00% |
| Beckman (Olympus) AU | 1 | *Negative | 1 | 100.00% |
| Innovacon Multi-Clin | 9 | *Negative | 9 | 100.00% |
| BioTek | 12 | *Negative | 12 | 100.00% |
| MedTox Profile III ER | 1 | *Negative | 1 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 5 | *Negative | 5 | 100.00% |
| ASC One Screen Drug Test Card | 2 | *Negative | 2 | 100.00% |
| MedTox Profile V | 2 | *Negative | 2 | 100.00% |
| Jant Accutest Multi Drug Screen | 2 | *Negative | 2 | 100.00% |
| Amedica Biotech AmediCheck Instant Test Cup | 4 | *Negative | 4 | 100.00% |
| Nova Century Scientific Nova Screen DOA Cassette | 1 | *Negative | 1 | 100.00% |
| American Screening Corp Reveal Multi-Drug Test Cup | 4 | *Negative | 4 | 100.00% |
| Mindray BS-200 | 1 | *Negative | 1 | 100.00% |
| Alere iCassette | 27 | *Negative | 27 | 100.00% |
| Roche Hitachi 717 | 1 | *Negative | 1 | 100.00% |
| US Diagnostics UScreen Cup | 5 | *Negative | 5 | 100.00% |
| Abbott Architect c4000 | 1 | *Negative | 1 | 100.00% |
| Discover Drug Test Card | 1 | *Negative | 1 | 100.00% |
| Bio-Rad Tox/See Rapid Urine Drug Screen | 1 | *Negative | 1 | 100.00% |
| Diag Test Group Clarity DOA Screen | 6 | *Negative | 6 | 100.00% |
| Noble Medical Insta-Screen Test Cup | 3 | *Negative | 3 | 100.00% |
| Alere Immatest Multi-Drug Screen Card | 1 | *Negative | 1 | 100.00% |
| Mindray 480 | 1 | *Negative | 1 | 100.00% |
| CRL Stat Cup | 1 | *Negative | 1 | 100.00% |
| Hemosure First Sign UDS Dipstick | 10 | *Negative | 10 | 100.00% |
| Hemosure First Sign UDS Cup | 1 | *Negative | 1 | 100.00% |
| UDS-3 | | * Denotes Correct Response | | |
| AG: All Methods | 4 | | | |
| | | *Positive | 4 | 100.00% |
| Beckman (Olympus) AU | 1 | *Positive | 1 | 100.00% |
| Mindray BS-200 | 1 | *Positive | 1 | 100.00% |
| Roche Hitachi 717 | 1 | *Positive | 1 | 100.00% |



| | | | | |
|--|-----|-----------------------------------|-----|---------|
| Mindray 480 | 1 | *Positive | 1 | 100.00% |
| UDS-4 | | * Denotes Correct Response | | |
| AG: All Methods | 4 | | | |
| | | *Positive | 4 | 100.00% |
| Beckman (Olympus) AU | 1 | *Positive | 1 | 100.00% |
| Mindray BS-200 | 1 | *Positive | 1 | 100.00% |
| Roche Hitachi 717 | 1 | *Positive | 1 | 100.00% |
| Mindray 480 | 1 | *Positive | 1 | 100.00% |
| | | MDMA (Ecstasy) | | |
| UDS-3 | | * Denotes Correct Response | | |
| AG: All Methods | 104 | | | |
| | | *Negative | 104 | 100.00% |
| E-Screen | 4 | *Negative | 4 | 100.00% |
| Alere iCup Drugs of Abuse | 13 | *Negative | 13 | 100.00% |
| Noble Med Split Specimen Cup | 1 | *Negative | 1 | 100.00% |
| Innovacon Multi-Clin | 10 | *Negative | 10 | 100.00% |
| BioTek | 12 | *Negative | 12 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 4 | *Negative | 4 | 100.00% |
| Jant Accutest Multi Drug Screen | 2 | *Negative | 2 | 100.00% |
| Amedica Biotech AmediCheck Instant Test Cup | 4 | *Negative | 4 | 100.00% |
| Nova Century Scientific Nova Screen DOA Cassette | 1 | *Negative | 1 | 100.00% |
| American Screening Corp Reveal Multi-Drug Test Cup | 2 | *Negative | 2 | 100.00% |
| Alere iCassette | 24 | *Negative | 24 | 100.00% |
| US Diagnostics UScreen Cup | 5 | *Negative | 5 | 100.00% |
| Discover Drug Test Card | 1 | *Negative | 1 | 100.00% |
| Bio-Rad Tox/See Rapid Urine Drug Screen | 1 | *Negative | 1 | 100.00% |
| Diag Test Group Clarity DOA Screen | 5 | *Negative | 5 | 100.00% |
| Noble Medical Insta-Screen Test Cup | 3 | *Negative | 3 | 100.00% |
| Alere Immutest Multi-Drug Screen Card | 1 | *Negative | 1 | 100.00% |
| Hemosure First Sign UDS Dipstick | 10 | *Negative | 10 | 100.00% |
| Hemosure First Sign UDS Cup | 1 | *Negative | 1 | 100.00% |
| UDS-4 | | * Denotes Correct Response | | |
| AG: All Methods | 104 | | | |
| | | *Negative | 104 | 100.00% |
| E-Screen | 4 | *Negative | 4 | 100.00% |
| Alere iCup Drugs of Abuse | 13 | *Negative | 13 | 100.00% |
| Noble Med Split Specimen Cup | 1 | *Negative | 1 | 100.00% |
| Innovacon Multi-Clin | 10 | *Negative | 10 | 100.00% |
| BioTek | 12 | *Negative | 12 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 4 | *Negative | 4 | 100.00% |
| Jant Accutest Multi Drug Screen | 2 | *Negative | 2 | 100.00% |
| Amedica Biotech AmediCheck Instant Test Cup | 4 | *Negative | 4 | 100.00% |
| Nova Century Scientific Nova Screen DOA Cassette | 1 | *Negative | 1 | 100.00% |
| American Screening Corp Reveal Multi-Drug Test Cup | 2 | *Negative | 2 | 100.00% |
| Alere iCassette | 24 | *Negative | 24 | 100.00% |
| US Diagnostics UScreen Cup | 5 | *Negative | 5 | 100.00% |
| Discover Drug Test Card | 1 | *Negative | 1 | 100.00% |
| Bio-Rad Tox/See Rapid Urine Drug Screen | 1 | *Negative | 1 | 100.00% |
| Diag Test Group Clarity DOA Screen | 5 | *Negative | 5 | 100.00% |
| Noble Medical Insta-Screen Test Cup | 3 | *Negative | 3 | 100.00% |
| Alere Immutest Multi-Drug Screen Card | 1 | *Negative | 1 | 100.00% |
| Hemosure First Sign UDS Dipstick | 10 | *Negative | 10 | 100.00% |
| Hemosure First Sign UDS Cup | 1 | *Negative | 1 | 100.00% |
| | | Methadone | | |
| UDS-3 | | * Denotes Correct Response | | |
| AG: All Methods | 96 | | | |
| | | *Negative | 96 | 100.00% |
| AlereTriage | 3 | *Negative | 3 | 100.00% |
| Siemens Dim/AR/ES/RxL/Xpand | 2 | *Negative | 2 | 100.00% |
| E-Screen | 8 | *Negative | 8 | 100.00% |
| Express Diag. Express Check | 1 | *Negative | 1 | 100.00% |
| Alere iCup Drugs of Abuse | 12 | *Negative | 12 | 100.00% |
| Instant Technologies i Screen | 2 | *Negative | 2 | 100.00% |
| Noble Med Split Specimen Cup | 1 | *Negative | 1 | 100.00% |
| Beckman (Olympus) AU | 1 | *Negative | 1 | 100.00% |
| BioTek | 12 | *Negative | 12 | 100.00% |
| MedTox Profile III ER | 1 | *Negative | 1 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 5 | *Negative | 5 | 100.00% |
| ASC One Screen Drug Test Card | 2 | *Negative | 2 | 100.00% |



| | | | | |
|--|-----|-----------------------------------|-----|---------|
| MedTox Profile V | 2 | *Negative | 2 | 100.00% |
| Jant Accutest Multi Drug Screen | 2 | *Negative | 2 | 100.00% |
| Amedica Biotech AmediCheck Instant Test Cup | 4 | *Negative | 4 | 100.00% |
| Nova Century Scientific Nova Screen DOA Cassette | 1 | *Negative | 1 | 100.00% |
| American Screening Corp Reveal Multi-Drug Test Cup | 4 | *Negative | 4 | 100.00% |
| Mindray BS-200 | 1 | *Negative | 1 | 100.00% |
| Roche Hitachi 717 | 1 | *Negative | 1 | 100.00% |
| US Diagnostics UScreen Cup | 5 | *Negative | 5 | 100.00% |
| Abbott Architect c4000 | 1 | *Negative | 1 | 100.00% |
| Discover Drug Test Card | 1 | *Negative | 1 | 100.00% |
| Bio-Rad Tox/See Rapid Urine Drug Screen | 1 | *Negative | 1 | 100.00% |
| Diag Test Group Clarity DOA Screen | 6 | *Negative | 6 | 100.00% |
| Noble Medical Insta-Screen Test Cup | 3 | *Negative | 3 | 100.00% |
| Alere Immutest Multi-Drug Screen Card | 1 | *Negative | 1 | 100.00% |
| Mindray 480 | 1 | *Negative | 1 | 100.00% |
| CRL Stat Cup | 1 | *Negative | 1 | 100.00% |
| Hemosure First Sign UDS Dipstick | 10 | *Negative | 10 | 100.00% |
| Hemosure First Sign UDS Cup | 1 | *Negative | 1 | 100.00% |
| UDS-4 | | * Denotes Correct Response | | |
| AG: All Methods | 96 | | | |
| | | *Negative | 96 | 100.00% |
| AlereTriage | 3 | *Negative | 3 | 100.00% |
| Siemens Dim/AR/ES/RxL/Xpand | 2 | *Negative | 2 | 100.00% |
| E-Screen | 8 | *Negative | 8 | 100.00% |
| Express Diag. Express Check | 1 | *Negative | 1 | 100.00% |
| Alere iCup Drugs of Abuse | 12 | *Negative | 12 | 100.00% |
| Instant Technologies i Screen | 2 | *Negative | 2 | 100.00% |
| Noble Med Split Specimen Cup | 1 | *Negative | 1 | 100.00% |
| Beckman (Olympus) AU | 1 | *Negative | 1 | 100.00% |
| BioTek | 12 | *Negative | 12 | 100.00% |
| MedTox Profile III ER | 1 | *Negative | 1 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 5 | *Negative | 5 | 100.00% |
| ASC One Screen Drug Test Card | 2 | *Negative | 2 | 100.00% |
| MedTox Profile V | 2 | *Negative | 2 | 100.00% |
| Jant Accutest Multi Drug Screen | 2 | *Negative | 2 | 100.00% |
| Amedica Biotech AmediCheck Instant Test Cup | 4 | *Negative | 4 | 100.00% |
| Nova Century Scientific Nova Screen DOA Cassette | 1 | *Negative | 1 | 100.00% |
| American Screening Corp Reveal Multi-Drug Test Cup | 4 | *Negative | 4 | 100.00% |
| Mindray BS-200 | 1 | *Negative | 1 | 100.00% |
| Roche Hitachi 717 | 1 | *Negative | 1 | 100.00% |
| US Diagnostics UScreen Cup | 5 | *Negative | 5 | 100.00% |
| Abbott Architect c4000 | 1 | *Negative | 1 | 100.00% |
| Discover Drug Test Card | 1 | *Negative | 1 | 100.00% |
| Bio-Rad Tox/See Rapid Urine Drug Screen | 1 | *Negative | 1 | 100.00% |
| Diag Test Group Clarity DOA Screen | 6 | *Negative | 6 | 100.00% |
| Noble Medical Insta-Screen Test Cup | 3 | *Negative | 3 | 100.00% |
| Alere Immutest Multi-Drug Screen Card | 1 | *Negative | 1 | 100.00% |
| Mindray 480 | 1 | *Negative | 1 | 100.00% |
| CRL Stat Cup | 1 | *Negative | 1 | 100.00% |
| Hemosure First Sign UDS Dipstick | 10 | *Negative | 10 | 100.00% |
| Hemosure First Sign UDS Cup | 1 | *Negative | 1 | 100.00% |
| UDS-3 | | Methamphetamine | | |
| | | * Denotes Correct Response | | |
| AG: All Methods | 103 | | | |
| | | *Negative | 103 | 100.00% |
| AlereTriage | 3 | *Negative | 3 | 100.00% |
| E-Screen | 14 | *Negative | 14 | 100.00% |
| Express Diag. Express Check | 1 | *Negative | 1 | 100.00% |
| Innovacon One Step | 1 | *Negative | 1 | 100.00% |
| Alere iCup Drugs of Abuse | 16 | *Negative | 16 | 100.00% |
| Instant Technologies i Screen | 3 | *Negative | 3 | 100.00% |
| Integrated E-Z Split Key Cup | 1 | *Negative | 1 | 100.00% |
| MedTox Profile II | 1 | *Negative | 1 | 100.00% |
| Noble Med Split Specimen Cup | 2 | *Negative | 2 | 100.00% |
| BioTek | 12 | *Negative | 12 | 100.00% |
| MedTox Profile III ER | 1 | *Negative | 1 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 5 | *Negative | 5 | 100.00% |
| ASC One Screen Drug Test Card | 2 | *Negative | 2 | 100.00% |
| MedTox Profile V | 2 | *Negative | 2 | 100.00% |



| | | | | |
|--|-----|-----------------------------------|-----|---------|
| Jant Accutest Multi Drug Screen | 2 | *Negative | 2 | 100.00% |
| Amedica Biotech AmediCheck Instant Test Cup | 4 | *Negative | 4 | 100.00% |
| Nova Century Scientific Nova Screen DOA Cassette | 1 | *Negative | 1 | 100.00% |
| American Screening Corp Reveal Multi-Drug Test Cup | 4 | *Negative | 4 | 100.00% |
| Alere iCassette | 1 | *Negative | 1 | 100.00% |
| US Diagnostics UScreen Cup | 5 | *Negative | 5 | 100.00% |
| Bio-Rad Tox/See Rapid Urine Drug Screen | 1 | *Negative | 1 | 100.00% |
| Diag Test Group Clarity DOA Screen | 6 | *Negative | 6 | 100.00% |
| Noble Medical Insta-Screen Test Cup | 3 | *Negative | 3 | 100.00% |
| Alere Immutest Multi-Drug Screen Card | 1 | *Negative | 1 | 100.00% |
| Hemosure First Sign UDS Dipstick | 10 | *Negative | 10 | 100.00% |
| Hemosure First Sign UDS Cup | 1 | *Negative | 1 | 100.00% |
| UDS-4 | | * Denotes Correct Response | | |
| AG: All Methods | 103 | | | |
| | | *Negative | 103 | 100.00% |
| AlereTriage | 3 | *Negative | 3 | 100.00% |
| E-Screen | 14 | *Negative | 14 | 100.00% |
| Express Diag. Express Check | 1 | *Negative | 1 | 100.00% |
| Innovacon One Step | 1 | *Negative | 1 | 100.00% |
| Alere iCup Drugs of Abuse | 16 | *Negative | 16 | 100.00% |
| Instant Technologies i Screen | 3 | *Negative | 3 | 100.00% |
| Integrated E-Z Split Key Cup | 1 | *Negative | 1 | 100.00% |
| MedTox Profile II | 1 | *Negative | 1 | 100.00% |
| Noble Med Split Specimen Cup | 2 | *Negative | 2 | 100.00% |
| BioTek | 12 | *Negative | 12 | 100.00% |
| MedTox Profile III ER | 1 | *Negative | 1 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 5 | *Negative | 5 | 100.00% |
| ASC One Screen Drug Test Card | 2 | *Negative | 2 | 100.00% |
| MedTox Profile V | 2 | *Negative | 2 | 100.00% |
| Jant Accutest Multi Drug Screen | 2 | *Negative | 2 | 100.00% |
| Amedica Biotech AmediCheck Instant Test Cup | 4 | *Negative | 4 | 100.00% |
| Nova Century Scientific Nova Screen DOA Cassette | 1 | *Negative | 1 | 100.00% |
| American Screening Corp Reveal Multi-Drug Test Cup | 4 | *Negative | 4 | 100.00% |
| Alere iCassette | 1 | *Negative | 1 | 100.00% |
| US Diagnostics UScreen Cup | 5 | *Negative | 5 | 100.00% |
| Bio-Rad Tox/See Rapid Urine Drug Screen | 1 | *Negative | 1 | 100.00% |
| Diag Test Group Clarity DOA Screen | 6 | *Negative | 6 | 100.00% |
| Noble Medical Insta-Screen Test Cup | 3 | *Negative | 3 | 100.00% |
| Alere Immutest Multi-Drug Screen Card | 1 | *Negative | 1 | 100.00% |
| Hemosure First Sign UDS Dipstick | 10 | *Negative | 10 | 100.00% |
| Hemosure First Sign UDS Cup | 1 | *Negative | 1 | 100.00% |
| | | Opiates | | |
| UDS-3 | | * Denotes Correct Response | | |
| AG: All Methods | 152 | | | |
| | | *Negative | 152 | 100.00% |
| AlereTriage | 3 | *Negative | 3 | 100.00% |
| Siemens Dim/AR/ES/RxL/Xpand | 3 | *Negative | 3 | 100.00% |
| E-Screen | 16 | *Negative | 16 | 100.00% |
| Express Diag. Express Check | 1 | *Negative | 1 | 100.00% |
| Innovacon One Step | 1 | *Negative | 1 | 100.00% |
| Alere iCup Drugs of Abuse | 17 | *Negative | 17 | 100.00% |
| Instant Technologies i Screen | 4 | *Negative | 4 | 100.00% |
| Integrated E-Z Split Key Cup | 1 | *Negative | 1 | 100.00% |
| MedTox Profile II | 1 | *Negative | 1 | 100.00% |
| Noble Med Split Specimen Cup | 2 | *Negative | 2 | 100.00% |
| Beckman (Olympus) AU | 1 | *Negative | 1 | 100.00% |
| Innovacon Multi-Clin | 10 | *Negative | 10 | 100.00% |
| BioTek | 12 | *Negative | 12 | 100.00% |
| MedTox Profile III ER | 1 | *Negative | 1 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 5 | *Negative | 5 | 100.00% |
| ASC One Screen Drug Test Card | 2 | *Negative | 2 | 100.00% |
| MedTox Profile V | 2 | *Negative | 2 | 100.00% |
| Jant Accutest Multi Drug Screen | 2 | *Negative | 2 | 100.00% |
| Amedica Biotech AmediCheck Instant Test Cup | 4 | *Negative | 4 | 100.00% |
| Nova Century Scientific Nova Screen DOA Cassette | 1 | *Negative | 1 | 100.00% |
| American Screening Corp Reveal Multi-Drug Test Cup | 4 | *Negative | 4 | 100.00% |
| Mindray BS-200 | 1 | *Negative | 1 | 100.00% |
| Alere iCassette | 27 | *Negative | 27 | 100.00% |
| Roche Hitachi 717 | 1 | *Negative | 1 | 100.00% |



| | | | | |
|--|-----|-----------------------------------|-----|---------|
| US Diagnostics UScreen Cup | 5 | *Negative | 5 | 100.00% |
| Abbott Architect c4000 | 1 | *Negative | 1 | 100.00% |
| Discover Drug Test Card | 1 | *Negative | 1 | 100.00% |
| Bio-Rad Tox/See Rapid Urine Drug Screen | 1 | *Negative | 1 | 100.00% |
| Diag Test Group Clarity DOA Screen | 5 | *Negative | 5 | 100.00% |
| Noble Medical Insta-Screen Test Cup | 3 | *Negative | 3 | 100.00% |
| Alere Immatest Multi-Drug Screen Card | 1 | *Negative | 1 | 100.00% |
| Mindray 480 | 1 | *Negative | 1 | 100.00% |
| CRL Stat Cup | 1 | *Negative | 1 | 100.00% |
| Hemosure First Sign UDS Dipstick | 10 | *Negative | 10 | 100.00% |
| Hemosure First Sign UDS Cup | 1 | *Negative | 1 | 100.00% |
| UDS-4 | | * Denotes Correct Response | | |
| AG: All Methods | 152 | | | |
| | | *Negative | 149 | 98.03% |
| | | Positive | 3 | 1.97% |
| AlereTriage | 3 | *Negative | 3 | 100.00% |
| Siemens Dim/AR/ES/RxL/Xpand | 3 | *Negative | 3 | 100.00% |
| E-Screen | 16 | *Negative | 16 | 100.00% |
| Express Diag. Express Check | 1 | *Negative | 1 | 100.00% |
| Innovacon One Step | 1 | *Negative | 1 | 100.00% |
| Alere iCup Drugs of Abuse | 17 | *Negative | 17 | 100.00% |
| Instant Technologies i Screen | 4 | *Negative | 4 | 100.00% |
| Integrated E-Z Split Key Cup | 1 | *Negative | 1 | 100.00% |
| MedTox Profile II | 1 | *Negative | 1 | 100.00% |
| Noble Med Split Specimen Cup | 2 | *Negative | 2 | 100.00% |
| Beckman (Olympus) AU | 1 | *Negative | 1 | 100.00% |
| Innovacon Multi-Clin | 10 | *Negative | 10 | 100.00% |
| BioTek | 12 | *Negative | 12 | 100.00% |
| MedTox Profile III ER | 1 | *Negative | 1 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 5 | *Negative | 5 | 100.00% |
| ASC One Screen Drug Test Card | 2 | *Negative | 2 | 100.00% |
| MedTox Profile V | 2 | *Negative | 2 | 100.00% |
| Jant Accutest Multi Drug Screen | 2 | *Negative | 2 | 100.00% |
| Amedica Biotech AmediCheck Instant Test Cup | 4 | *Negative | 4 | 100.00% |
| Nova Century Scientific Nova Screen DOA Cassette | 1 | *Negative | 1 | 100.00% |
| American Screening Corp Reveal Multi-Drug Test Cup | 4 | *Negative | 4 | 100.00% |
| Mindray BS-200 | 1 | *Negative | 1 | 100.00% |
| Alere iCassette | 27 | *Negative | 26 | 96.30% |
| Roche Hitachi 717 | 1 | *Negative | 1 | 100.00% |
| US Diagnostics UScreen Cup | 5 | *Negative | 4 | 80.00% |
| Abbott Architect c4000 | 1 | *Negative | 1 | 100.00% |
| Discover Drug Test Card | 1 | *Negative | 1 | 100.00% |
| Bio-Rad Tox/See Rapid Urine Drug Screen | 1 | *Negative | 1 | 100.00% |
| Diag Test Group Clarity DOA Screen | 5 | *Negative | 5 | 100.00% |
| Noble Medical Insta-Screen Test Cup | 3 | *Negative | 3 | 100.00% |
| Alere Immatest Multi-Drug Screen Card | 1 | *Negative | 1 | 100.00% |
| Mindray 480 | 1 | *Negative | 1 | 100.00% |
| CRL Stat Cup | 1 | *Negative | 1 | 100.00% |
| Hemosure First Sign UDS Dipstick | 10 | *Negative | 9 | 90.00% |
| Hemosure First Sign UDS Cup | 1 | *Negative | 1 | 100.00% |
| Alere iCassette | 27 | Positive | 1 | 3.70% |
| US Diagnostics UScreen Cup | 5 | Positive | 1 | 20.00% |
| Hemosure First Sign UDS Dipstick | 10 | Positive | 1 | 10.00% |
| UDS-3 | | * Denotes Correct Response | | |
| AG: All Methods | 106 | | | |
| | | *Negative | 105 | 99.06% |
| | | Positive | 1 | 0.94% |
| Innovacon One Step | 1 | *Negative | 1 | 100.00% |
| Alere iCup Drugs of Abuse | 13 | *Negative | 13 | 100.00% |
| Instant Technologies i Screen | 2 | *Negative | 2 | 100.00% |
| Noble Med Split Specimen Cup | 1 | *Negative | 1 | 100.00% |
| Innovacon Multi-Clin | 10 | *Negative | 10 | 100.00% |
| BioTek | 12 | *Negative | 12 | 100.00% |
| MedTox Profile III ER | 1 | *Negative | 1 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 5 | *Negative | 5 | 100.00% |
| ASC One Screen Drug Test Card | 2 | *Negative | 2 | 100.00% |
| MedTox Profile V | 2 | *Negative | 2 | 100.00% |
| Jant Accutest Multi Drug Screen | 2 | *Negative | 2 | 100.00% |



| | | | | |
|--|-----|-----------------------------------|-----|---------|
| Amedica Biotech AmediCheck Instant Test Cup | 4 | *Negative | 4 | 100.00% |
| Nova Century Scientific Nova Screen DOA Cassette | 1 | *Negative | 1 | 100.00% |
| American Screening Corp Reveal Multi-Drug Test Cup | 4 | *Negative | 4 | 100.00% |
| Mindray BS-200 | 1 | *Negative | 1 | 100.00% |
| Alere iCassette | 24 | *Negative | 23 | 95.83% |
| Roche Hitachi 717 | 1 | *Negative | 1 | 100.00% |
| US Diagnostics UScreen Cup | 5 | *Negative | 5 | 100.00% |
| Discover Drug Test Card | 1 | *Negative | 1 | 100.00% |
| Bio-Rad Tox/See Rapid Urine Drug Screen | 2 | *Negative | 2 | 100.00% |
| Diag Test Group Clarity DOA Screen | 6 | *Negative | 6 | 100.00% |
| Alere iScreen Drugs of Abuse Card | 1 | *Negative | 1 | 100.00% |
| Noble Medical Insta-Screen Test Cup | 3 | *Negative | 3 | 100.00% |
| Alere Immatest Multi-Drug Screen Card | 1 | *Negative | 1 | 100.00% |
| Mindray 480 | 1 | *Negative | 1 | 100.00% |
| Alere iCassette | 24 | Positive | 1 | 4.17% |
| UDS-4 | | * Denotes Correct Response | | |
| AG: All Methods | 106 | | | |
| | | *Positive | 104 | 98.11% |
| | | Negative | 2 | 1.89% |
| Innovacon One Step | 1 | *Positive | 1 | 100.00% |
| Alere iCup Drugs of Abuse | 13 | *Positive | 13 | 100.00% |
| Instant Technologies i Screen | 2 | *Positive | 2 | 100.00% |
| Noble Med Split Specimen Cup | 1 | *Positive | 1 | 100.00% |
| Innovacon Multi-Clin | 10 | *Positive | 10 | 100.00% |
| BioTek | 12 | *Positive | 12 | 100.00% |
| MedTox Profile III ER | 1 | *Positive | 1 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 5 | *Positive | 5 | 100.00% |
| ASC One Screen Drug Test Card | 2 | *Positive | 2 | 100.00% |
| MedTox Profile V | 2 | *Positive | 2 | 100.00% |
| Jant Accutest Multi Drug Screen | 2 | *Positive | 2 | 100.00% |
| Amedica Biotech AmediCheck Instant Test Cup | 4 | *Positive | 4 | 100.00% |
| Nova Century Scientific Nova Screen DOA Cassette | 1 | *Positive | 1 | 100.00% |
| American Screening Corp Reveal Multi-Drug Test Cup | 4 | *Positive | 4 | 100.00% |
| Mindray BS-200 | 1 | *Positive | 1 | 100.00% |
| Alere iCassette | 24 | *Positive | 22 | 91.67% |
| Roche Hitachi 717 | 1 | *Positive | 1 | 100.00% |
| US Diagnostics UScreen Cup | 5 | *Positive | 5 | 100.00% |
| Discover Drug Test Card | 1 | *Positive | 1 | 100.00% |
| Bio-Rad Tox/See Rapid Urine Drug Screen | 2 | *Positive | 2 | 100.00% |
| Diag Test Group Clarity DOA Screen | 6 | *Positive | 6 | 100.00% |
| Alere iScreen Drugs of Abuse Card | 1 | *Positive | 1 | 100.00% |
| Noble Medical Insta-Screen Test Cup | 3 | *Positive | 3 | 100.00% |
| Alere Immatest Multi-Drug Screen Card | 1 | *Positive | 1 | 100.00% |
| Mindray 480 | 1 | *Positive | 1 | 100.00% |
| Alere iCassette | 24 | Negative | 2 | 8.33% |
| UDS-3 | | PCP (Phencyclidine) | | |
| | | * Denotes Correct Response | | |
| AG: All Methods | 136 | | | |
| | | *Positive | 134 | 98.53% |
| | | Negative | 2 | 1.47% |
| AlereTriage | 3 | *Positive | 3 | 100.00% |
| Siemens Dim/AR/ES/RxL/Xpand | 2 | *Positive | 2 | 100.00% |
| E-Screen | 16 | *Positive | 16 | 100.00% |
| Express Diag. Express Check | 1 | *Positive | 1 | 100.00% |
| Innovacon One Step | 1 | *Positive | 1 | 100.00% |
| Alere iCup Drugs of Abuse | 11 | *Positive | 11 | 100.00% |
| Instant Technologies i Screen | 2 | *Positive | 2 | 100.00% |
| Integrated E-Z Split Key Cup | 1 | *Positive | 1 | 100.00% |
| MedTox Profile II | 2 | *Positive | 2 | 100.00% |
| Noble Med Split Specimen Cup | 1 | *Positive | 1 | 100.00% |
| Beckman (Olympus) AU | 1 | *Positive | 1 | 100.00% |
| Innovacon Multi-Clin | 10 | *Positive | 10 | 100.00% |
| BioTek | 12 | *Positive | 12 | 100.00% |
| MedTox Profile III ER | 1 | *Positive | 1 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 5 | *Positive | 5 | 100.00% |
| ASC One Screen Drug Test Card | 2 | *Positive | 2 | 100.00% |
| MedTox Profile V | 2 | *Positive | 2 | 100.00% |
| Jant Accutest Multi Drug Screen | 2 | *Positive | 2 | 100.00% |
| Amedica Biotech AmediCheck Instant Test Cup | 1 | *Positive | 1 | 100.00% |



| | | | | |
|--|-----|-----------------------------------|-----|---------|
| Nova Century Scientific Nova Screen DOA Cassette | 1 | *Positive | 1 | 100.00% |
| American Screening Corp Reveal Multi-Drug Test Cup | 4 | *Positive | 4 | 100.00% |
| Mindray BS-200 | 1 | *Positive | 1 | 100.00% |
| Alere iCassette | 26 | *Positive | 25 | 96.15% |
| US Diagnostics UScreen Cup | 2 | *Positive | 2 | 100.00% |
| Abbott Architect c4000 | 1 | *Positive | 1 | 100.00% |
| Discover Drug Test Card | 1 | *Positive | 1 | 100.00% |
| Bio-Rad Tox/See Rapid Urine Drug Screen | 1 | *Positive | 1 | 100.00% |
| Diag Test Group Clarity DOA Screen | 6 | *Positive | 6 | 100.00% |
| Noble Medical Insta-Screen Test Cup | 3 | *Positive | 3 | 100.00% |
| Alere Immatest Multi-Drug Screen Card | 1 | *Positive | 1 | 100.00% |
| CRL Stat Cup | 1 | *Positive | 1 | 100.00% |
| Hemosure First Sign UDS Dipstick | 10 | *Positive | 10 | 100.00% |
| Hemosure First Sign UDS Cup | 1 | *Positive | 1 | 100.00% |
| Alere iCassette | 26 | Negative | 1 | 3.85% |
| Roche Hitachi 717 | 1 | Negative | 1 | 100.00% |
| UDS-4 | | * Denotes Correct Response | | |
| AG: All Methods | 136 | | | |
| | | *Negative | 134 | 98.53% |
| | | Positive | 2 | 1.47% |
| AlereTriage | 3 | *Negative | 3 | 100.00% |
| Siemens Dim/AR/ES/RxL/Xpand | 2 | *Negative | 2 | 100.00% |
| E-Screen | 16 | *Negative | 16 | 100.00% |
| Express Diag. Express Check | 1 | *Negative | 1 | 100.00% |
| Innovacon One Step | 1 | *Negative | 1 | 100.00% |
| Alere iCup Drugs of Abuse | 11 | *Negative | 11 | 100.00% |
| Instant Technologies i Screen | 2 | *Negative | 2 | 100.00% |
| Integrated E-Z Split Key Cup | 1 | *Negative | 1 | 100.00% |
| MedTox Profile II | 2 | *Negative | 2 | 100.00% |
| Noble Med Split Specimen Cup | 1 | *Negative | 1 | 100.00% |
| Beckman (Olympus) AU | 1 | *Negative | 1 | 100.00% |
| Innovacon Multi-Clin | 10 | *Negative | 10 | 100.00% |
| BioTek | 12 | *Negative | 12 | 100.00% |
| MedTox Profile III ER | 1 | *Negative | 1 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 5 | *Negative | 5 | 100.00% |
| ASC One Screen Drug Test Card | 2 | *Negative | 2 | 100.00% |
| MedTox Profile V | 2 | *Negative | 2 | 100.00% |
| Jant Accutest Multi Drug Screen | 2 | *Negative | 2 | 100.00% |
| Amedica Biotech AmediCheck Instant Test Cup | 1 | *Negative | 1 | 100.00% |
| Nova Century Scientific Nova Screen DOA Cassette | 1 | *Negative | 1 | 100.00% |
| American Screening Corp Reveal Multi-Drug Test Cup | 4 | *Negative | 3 | 75.00% |
| Mindray BS-200 | 1 | *Negative | 1 | 100.00% |
| Alere iCassette | 26 | *Negative | 26 | 100.00% |
| US Diagnostics UScreen Cup | 2 | *Negative | 2 | 100.00% |
| Abbott Architect c4000 | 1 | *Negative | 1 | 100.00% |
| Discover Drug Test Card | 1 | *Negative | 1 | 100.00% |
| Bio-Rad Tox/See Rapid Urine Drug Screen | 1 | *Negative | 1 | 100.00% |
| Diag Test Group Clarity DOA Screen | 6 | *Negative | 6 | 100.00% |
| Noble Medical Insta-Screen Test Cup | 3 | *Negative | 3 | 100.00% |
| Alere Immatest Multi-Drug Screen Card | 1 | *Negative | 1 | 100.00% |
| CRL Stat Cup | 1 | *Negative | 1 | 100.00% |
| Hemosure First Sign UDS Dipstick | 10 | *Negative | 10 | 100.00% |
| Hemosure First Sign UDS Cup | 1 | *Negative | 1 | 100.00% |
| American Screening Corp Reveal Multi-Drug Test Cup | 4 | Positive | 1 | 25.00% |
| Roche Hitachi 717 | 1 | Positive | 1 | 100.00% |
| | | Propoxyphene | | |
| UDS-3 | | * Denotes Correct Response | | |
| AG: All Methods | 49 | | | |
| | | *Negative | 49 | 100.00% |
| E-Screen | 4 | *Negative | 4 | 100.00% |
| Alere iCup Drugs of Abuse | 7 | *Negative | 7 | 100.00% |
| MedTox Profile II | 1 | *Negative | 1 | 100.00% |
| Beckman (Olympus) AU | 1 | *Negative | 1 | 100.00% |
| Innovacon Multi-Clin | 8 | *Negative | 8 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 1 | *Negative | 1 | 100.00% |
| MedTox Profile V | 2 | *Negative | 2 | 100.00% |
| Alere iCassette | 24 | *Negative | 24 | 100.00% |
| Diag Test Group Clarity DOA Screen | 1 | *Negative | 1 | 100.00% |
| UDS-4 | | * Denotes Correct Response | | |



| | | | | |
|--|----|-----------------------------------|----|---------|
| AG: All Methods | 49 | | | |
| | | *Negative | 49 | 100.00% |
| E-Screen | 4 | *Negative | 4 | 100.00% |
| Alere iCup Drugs of Abuse | 7 | *Negative | 7 | 100.00% |
| MedTox Profile II | 1 | *Negative | 1 | 100.00% |
| Beckman (Olympus) AU | 1 | *Negative | 1 | 100.00% |
| Innovacon Multi-Clin | 8 | *Negative | 8 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 1 | *Negative | 1 | 100.00% |
| MedTox Profile V | 2 | *Negative | 2 | 100.00% |
| Alere iCassette | 24 | *Negative | 24 | 100.00% |
| Diag Test Group Clarity DOA Screen | 1 | *Negative | 1 | 100.00% |
| | | Tricyclic Antidepressants | | |
| UDS-3 | | * Denotes Correct Response | | |
| AG: All Methods | 72 | | | |
| | | *Negative | 72 | 100.00% |
| AlereTriage | 3 | *Negative | 3 | 100.00% |
| Alere iCup Drugs of Abuse | 11 | *Negative | 11 | 100.00% |
| Instant Technologies i Screen | 2 | *Negative | 2 | 100.00% |
| Noble Med Split Specimen Cup | 1 | *Negative | 1 | 100.00% |
| Innovacon Multi-Clin | 8 | *Negative | 8 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 3 | *Negative | 3 | 100.00% |
| MedTox Profile V | 2 | *Negative | 2 | 100.00% |
| Jant Accutest Multi Drug Screen | 2 | *Negative | 2 | 100.00% |
| Amedica Biotech AmediCheck Instant Test Cup | 1 | *Negative | 1 | 100.00% |
| Nova Century Scientific Nova Screen DOA Cassette | 1 | *Negative | 1 | 100.00% |
| Alere iCassette | 24 | *Negative | 24 | 100.00% |
| US Diagnostics UScreen Cup | 3 | *Negative | 3 | 100.00% |
| Bio-Rad Tox/See Rapid Urine Drug Screen | 2 | *Negative | 2 | 100.00% |
| Diag Test Group Clarity DOA Screen | 4 | *Negative | 4 | 100.00% |
| Noble Medical Insta-Screen Test Cup | 3 | *Negative | 3 | 100.00% |
| Alere Immatest Multi-Drug Screen Card | 1 | *Negative | 1 | 100.00% |
| Mindray 480 | 1 | *Negative | 1 | 100.00% |
| UDS-4 | | * Denotes Correct Response | | |
| AG: All Methods | 72 | | | |
| | | *Negative | 72 | 100.00% |
| AlereTriage | 3 | *Negative | 3 | 100.00% |
| Alere iCup Drugs of Abuse | 11 | *Negative | 11 | 100.00% |
| Instant Technologies i Screen | 2 | *Negative | 2 | 100.00% |
| Noble Med Split Specimen Cup | 1 | *Negative | 1 | 100.00% |
| Innovacon Multi-Clin | 8 | *Negative | 8 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 3 | *Negative | 3 | 100.00% |
| MedTox Profile V | 2 | *Negative | 2 | 100.00% |
| Jant Accutest Multi Drug Screen | 2 | *Negative | 2 | 100.00% |
| Amedica Biotech AmediCheck Instant Test Cup | 1 | *Negative | 1 | 100.00% |
| Nova Century Scientific Nova Screen DOA Cassette | 1 | *Negative | 1 | 100.00% |
| Alere iCassette | 24 | *Negative | 24 | 100.00% |
| US Diagnostics UScreen Cup | 3 | *Negative | 3 | 100.00% |
| Bio-Rad Tox/See Rapid Urine Drug Screen | 2 | *Negative | 2 | 100.00% |
| Diag Test Group Clarity DOA Screen | 4 | *Negative | 4 | 100.00% |
| Noble Medical Insta-Screen Test Cup | 3 | *Negative | 3 | 100.00% |
| Alere Immatest Multi-Drug Screen Card | 1 | *Negative | 1 | 100.00% |
| Mindray 480 | 1 | *Negative | 1 | 100.00% |
| | | Buprenorphine | | |
| UDS-3 | | * Denotes Correct Response | | |
| AG: All Methods | 33 | | | |
| | | *Negative | 32 | 96.97% |
| | | Positive | 1 | 3.03% |
| Alere iCup Drugs of Abuse | 7 | *Negative | 7 | 100.00% |
| BioTek | 12 | *Negative | 12 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 2 | *Negative | 2 | 100.00% |
| ASC One Screen Drug Test Card | 2 | *Negative | 2 | 100.00% |
| MedTox Profile V | 1 | *Negative | 1 | 100.00% |
| Mindray BS-200 | 1 | *Negative | 1 | 100.00% |
| Discover Drug Test Card | 1 | *Negative | 1 | 100.00% |
| Diag Test Group Clarity DOA Screen | 1 | *Negative | 1 | 100.00% |
| Alere iScreen Drugs of Abuse Card | 1 | *Negative | 1 | 100.00% |
| Noble Medical Insta-Screen Test Cup | 3 | *Negative | 3 | 100.00% |
| Mindray 480 | 1 | *Negative | 1 | 100.00% |
| Jant Accutest Multi Drug Screen | 1 | Positive | 1 | 100.00% |



| UDS-4 | | * Denotes Correct Response | | |
|--|------|----------------------------|--------|---------|
| AG: All Methods | 33 | | | |
| | | *Negative | 33 | 100.00% |
| Alere iCup Drugs of Abuse | 7 | *Negative | 7 | 100.00% |
| BioTek | 12 | *Negative | 12 | 100.00% |
| CLIAwaived Inc Rapid Drug Cup | 2 | *Negative | 2 | 100.00% |
| ASC One Screen Drug Test Card | 2 | *Negative | 2 | 100.00% |
| MedTox Profile V | 1 | *Negative | 1 | 100.00% |
| Jant Accutest Multi Drug Screen | 1 | *Negative | 1 | 100.00% |
| Mindray BS-200 | 1 | *Negative | 1 | 100.00% |
| Discover Drug Test Card | 1 | *Negative | 1 | 100.00% |
| Diag Test Group Clarity DOA Screen | 1 | *Negative | 1 | 100.00% |
| Alere iScreen Drugs of Abuse Card | 1 | *Negative | 1 | 100.00% |
| Noble Medical Insta-Screen Test Cup | 3 | *Negative | 3 | 100.00% |
| Mindray 480 | 1 | *Negative | 1 | 100.00% |
| 756 - Viral Antibodies | Labs | Response | Number | Percent |
| HSV Ab (Herpes Simplex Virus) | | | | |
| VA-2 | | * Denotes Correct Response | | |
| AG: All Methods | 2 | | | |
| | | *Negative | 2 | 100.00% |
| Biokit HSV-II | 2 | *Negative | 2 | 100.00% |
| 770 - Viral Antigen Detection;798 - Viral Antigen Detection (ROT);797 - Viral Antigen Detection II | Labs | Response | Number | Percent |
| Rotavirus - Ag Detect | | | | |
| VAD-6 | | * Denotes Correct Response | | |
| AG: All Methods | 9 | | | |
| | | *Positive | 9 | 100.00% |
| Meridian Immunocard Stat! Rotavirus | 5 | *Positive | 5 | 100.00% |
| SAS Rota-Test | 4 | *Positive | 4 | 100.00% |
| VAD-7 | | * Denotes Correct Response | | |
| AG: All Methods | 9 | | | |
| | | *Negative | 9 | 100.00% |
| Meridian Immunocard Stat! Rotavirus | 5 | *Negative | 5 | 100.00% |
| SAS Rota-Test | 4 | *Negative | 4 | 100.00% |
| VAD-8 | | * Denotes Correct Response | | |
| AG: All Methods | 9 | | | |
| | | *Negative | 9 | 100.00% |
| Meridian Immunocard Stat! Rotavirus | 5 | *Negative | 5 | 100.00% |
| SAS Rota-Test | 4 | *Negative | 4 | 100.00% |
| VAD-9 | | * Denotes Correct Response | | |
| AG: All Methods | 9 | | | |
| | | *Negative | 9 | 100.00% |
| Meridian Immunocard Stat! Rotavirus | 5 | *Negative | 5 | 100.00% |
| SAS Rota-Test | 4 | *Negative | 4 | 100.00% |
| VAD-10 | | * Denotes Correct Response | | |
| AG: All Methods | 9 | | | |
| | | *Negative | 9 | 100.00% |
| Meridian Immunocard Stat! Rotavirus | 5 | *Negative | 5 | 100.00% |
| SAS Rota-Test | 4 | *Negative | 4 | 100.00% |
| Influenza A - Ag Detect | | | | |
| VAD-6 | | * Denotes Correct Response | | |
| AG: All Methods | 62 | | | |
| | | *Negative | 62 | 100.00% |
| Alere BinaxNOW Influenza A+B Card | 7 | *Negative | 7 | 100.00% |
| Sekisui Diagnostics OSOM Influenza A & B | 11 | *Negative | 11 | 100.00% |
| Quidel QuickVue A+B | 22 | *Negative | 22 | 100.00% |
| Remel Xpect | 1 | *Negative | 1 | 100.00% |
| Quidel Sofia Influenza A+B | 4 | *Negative | 4 | 100.00% |
| Alere Influenza A & B | 4 | *Negative | 4 | 100.00% |
| BD Veritor Flu A+B | 12 | *Negative | 12 | 100.00% |
| Cepheid GeneXpert - Xpert Flu | 1 | *Negative | 1 | 100.00% |
| VAD-7 | | * Denotes Correct Response | | |
| AG: All Methods | 62 | | | |
| | | *Negative | 61 | 98.39% |
| | | Positive | 1 | 1.61% |
| Alere BinaxNOW Influenza A+B Card | 7 | *Negative | 7 | 100.00% |
| Sekisui Diagnostics OSOM Influenza A & B | 11 | *Negative | 11 | 100.00% |
| Quidel QuickVue A+B | 22 | *Negative | 21 | 95.45% |



| | | | | |
|--|----|-----------------------------------|----|---------|
| Remel Xpect | 1 | *Negative | 1 | 100.00% |
| Quidel Sofia Influenza A+B | 4 | *Negative | 4 | 100.00% |
| Alere Influenza A & B | 4 | *Negative | 4 | 100.00% |
| BD Veritor Flu A+B | 12 | *Negative | 12 | 100.00% |
| Cepheid GeneXpert - Xpert Flu | 1 | *Negative | 1 | 100.00% |
| Quidel QuickVue A+B | 22 | Positive | 1 | 4.55% |
| VAD-8 | | * Denotes Correct Response | | |
| AG: All Methods | 62 | | | |
| | | *Positive | 62 | 100.00% |
| Alere BinaxNOW Influenza A+B Card | 7 | *Positive | 7 | 100.00% |
| Sekisui Diagnostics OSOM Influenza A & B | 11 | *Positive | 11 | 100.00% |
| Quidel QuickVue A+B | 22 | *Positive | 22 | 100.00% |
| Remel Xpect | 1 | *Positive | 1 | 100.00% |
| Quidel Sofia Influenza A+B | 4 | *Positive | 4 | 100.00% |
| Alere Influenza A & B | 4 | *Positive | 4 | 100.00% |
| BD Veritor Flu A+B | 12 | *Positive | 12 | 100.00% |
| Cepheid GeneXpert - Xpert Flu | 1 | *Positive | 1 | 100.00% |
| VAD-9 | | * Denotes Correct Response | | |
| AG: All Methods | 62 | | | |
| | | *Negative | 61 | 98.39% |
| | | Positive | 1 | 1.61% |
| Alere BinaxNOW Influenza A+B Card | 7 | *Negative | 7 | 100.00% |
| Sekisui Diagnostics OSOM Influenza A & B | 11 | *Negative | 11 | 100.00% |
| Quidel QuickVue A+B | 22 | *Negative | 21 | 95.45% |
| Remel Xpect | 1 | *Negative | 1 | 100.00% |
| Quidel Sofia Influenza A+B | 4 | *Negative | 4 | 100.00% |
| Alere Influenza A & B | 4 | *Negative | 4 | 100.00% |
| BD Veritor Flu A+B | 12 | *Negative | 12 | 100.00% |
| Cepheid GeneXpert - Xpert Flu | 1 | *Negative | 1 | 100.00% |
| Quidel QuickVue A+B | 22 | Positive | 1 | 4.55% |
| VAD-10 | | * Denotes Correct Response | | |
| AG: All Methods | 62 | | | |
| | | *Positive | 60 | 96.77% |
| | | Negative | 2 | 3.23% |
| Alere BinaxNOW Influenza A+B Card | 7 | *Positive | 7 | 100.00% |
| Sekisui Diagnostics OSOM Influenza A & B | 11 | *Positive | 11 | 100.00% |
| Quidel QuickVue A+B | 22 | *Positive | 20 | 90.91% |
| Remel Xpect | 1 | *Positive | 1 | 100.00% |
| Quidel Sofia Influenza A+B | 4 | *Positive | 4 | 100.00% |
| Alere Influenza A & B | 4 | *Positive | 4 | 100.00% |
| BD Veritor Flu A+B | 12 | *Positive | 12 | 100.00% |
| Cepheid GeneXpert - Xpert Flu | 1 | *Positive | 1 | 100.00% |
| Quidel QuickVue A+B | 22 | Negative | 2 | 9.09% |
| | | Influenza A/B - Ag Detect | | |
| VAD-6 | | * Denotes Correct Response | | |
| AG: All Methods | 4 | | | |
| | | *Positive | 4 | 100.00% |
| Quidel QuickVue | 4 | *Positive | 4 | 100.00% |
| VAD-7 | | * Denotes Correct Response | | |
| AG: All Methods | 4 | | | |
| | | *Negative | 4 | 100.00% |
| Quidel QuickVue | 4 | *Negative | 4 | 100.00% |
| VAD-8 | | * Denotes Correct Response | | |
| AG: All Methods | 4 | | | |
| | | *Positive | 4 | 100.00% |
| Quidel QuickVue | 4 | *Positive | 4 | 100.00% |
| VAD-9 | | * Denotes Correct Response | | |
| AG: All Methods | 4 | | | |
| | | *Positive | 4 | 100.00% |
| Quidel QuickVue | 4 | *Positive | 4 | 100.00% |
| VAD-10 | | * Denotes Correct Response | | |
| AG: All Methods | 4 | | | |
| | | *Positive | 4 | 100.00% |
| Quidel QuickVue | 4 | *Positive | 4 | 100.00% |
| | | Influenza B - Ag Detect | | |
| VAD-6 | | * Denotes Correct Response | | |
| AG: All Methods | 62 | | | |
| | | *Positive | 61 | 98.39% |



| | | | | |
|--|----|-----------------------------------|----|---------|
| | | Negative | 1 | 1.61% |
| Alere BinaxNOW Influenza A+B Card | 7 | *Positive | 7 | 100.00% |
| Sekisui Diagnostics OSOM Influenza A & B | 11 | *Positive | 11 | 100.00% |
| Quidel QuickVue A+B | 22 | *Positive | 21 | 95.45% |
| Remel Xpect | 1 | *Positive | 1 | 100.00% |
| Quidel Sofia Influenza A+B | 4 | *Positive | 4 | 100.00% |
| Alere Influenza A & B | 4 | *Positive | 4 | 100.00% |
| BD Veritor Flu A+B | 12 | *Positive | 12 | 100.00% |
| Cepheid GeneXpert - Xpert Flu | 1 | *Positive | 1 | 100.00% |
| Quidel QuickVue A+B | 22 | Negative | 1 | 4.55% |
| VAD-7 | | * Denotes Correct Response | | |
| AG: All Methods | 62 | | | |
| | | *Negative | 61 | 98.39% |
| | | Positive | 1 | 1.61% |
| Alere BinaxNOW Influenza A+B Card | 7 | *Negative | 7 | 100.00% |
| Sekisui Diagnostics OSOM Influenza A & B | 11 | *Negative | 11 | 100.00% |
| Quidel QuickVue A+B | 22 | *Negative | 21 | 95.45% |
| Remel Xpect | 1 | *Negative | 1 | 100.00% |
| Quidel Sofia Influenza A+B | 4 | *Negative | 4 | 100.00% |
| Alere Influenza A & B | 4 | *Negative | 4 | 100.00% |
| BD Veritor Flu A+B | 12 | *Negative | 12 | 100.00% |
| Cepheid GeneXpert - Xpert Flu | 1 | *Negative | 1 | 100.00% |
| Quidel QuickVue A+B | 22 | Positive | 1 | 4.55% |
| VAD-8 | | * Denotes Correct Response | | |
| AG: All Methods | 62 | | | |
| | | *Negative | 61 | 98.39% |
| | | Positive | 1 | 1.61% |
| Alere BinaxNOW Influenza A+B Card | 7 | *Negative | 7 | 100.00% |
| Sekisui Diagnostics OSOM Influenza A & B | 11 | *Negative | 11 | 100.00% |
| Quidel QuickVue A+B | 22 | *Negative | 21 | 95.45% |
| Remel Xpect | 1 | *Negative | 1 | 100.00% |
| Quidel Sofia Influenza A+B | 4 | *Negative | 4 | 100.00% |
| Alere Influenza A & B | 4 | *Negative | 4 | 100.00% |
| BD Veritor Flu A+B | 12 | *Negative | 12 | 100.00% |
| Cepheid GeneXpert - Xpert Flu | 1 | *Negative | 1 | 100.00% |
| Quidel QuickVue A+B | 22 | Positive | 1 | 4.55% |
| VAD-9 | | * Denotes Correct Response | | |
| AG: All Methods | 62 | | | |
| | | *Positive | 60 | 96.77% |
| | | Negative | 2 | 3.23% |
| Alere BinaxNOW Influenza A+B Card | 7 | *Positive | 7 | 100.00% |
| Sekisui Diagnostics OSOM Influenza A & B | 11 | *Positive | 11 | 100.00% |
| Quidel QuickVue A+B | 22 | *Positive | 21 | 95.45% |
| Remel Xpect | 1 | *Positive | 1 | 100.00% |
| Quidel Sofia Influenza A+B | 4 | *Positive | 4 | 100.00% |
| Alere Influenza A & B | 4 | *Positive | 4 | 100.00% |
| BD Veritor Flu A+B | 12 | *Positive | 11 | 91.67% |
| Cepheid GeneXpert - Xpert Flu | 1 | *Positive | 1 | 100.00% |
| Quidel QuickVue A+B | 22 | Negative | 1 | 4.55% |
| BD Veritor Flu A+B | 12 | Negative | 1 | 8.33% |
| VAD-10 | | * Denotes Correct Response | | |
| AG: All Methods | 62 | | | |
| | | *Negative | 60 | 96.77% |
| | | Positive | 2 | 3.23% |
| Alere BinaxNOW Influenza A+B Card | 7 | *Negative | 7 | 100.00% |
| Sekisui Diagnostics OSOM Influenza A & B | 11 | *Negative | 11 | 100.00% |
| Quidel QuickVue A+B | 22 | *Negative | 20 | 90.91% |
| Remel Xpect | 1 | *Negative | 1 | 100.00% |
| Quidel Sofia Influenza A+B | 4 | *Negative | 4 | 100.00% |
| Alere Influenza A & B | 4 | *Negative | 4 | 100.00% |
| BD Veritor Flu A+B | 12 | *Negative | 12 | 100.00% |
| Cepheid GeneXpert - Xpert Flu | 1 | *Negative | 1 | 100.00% |
| Quidel QuickVue A+B | 22 | Positive | 2 | 9.09% |
| | | RSV - Aq Detect | | |
| VAD-6 | | * Denotes Correct Response | | |
| AG: All Methods | 26 | | | |
| | | *Negative | 26 | 100.00% |
| BD EZ RSV | 1 | *Negative | 1 | 100.00% |
| Alere BinaxNOW RSV | 10 | *Negative | 10 | 100.00% |



| | | | | |
|------------------------|----|-----------------------------------|----|---------|
| Quidel QuickVue RSV | 3 | *Negative | 3 | 100.00% |
| Wampole Clearview RSV | 4 | *Negative | 4 | 100.00% |
| SAS RSV Alert | 1 | *Negative | 1 | 100.00% |
| Quidel QuickVue RSV 10 | 2 | *Negative | 2 | 100.00% |
| Quidel Sofia RSV | 5 | *Negative | 5 | 100.00% |
| VAD-7 | | * Denotes Correct Response | | |
| AG: All Methods | 26 | | | |
| | | *Negative | 25 | 96.15% |
| | | Positive | 1 | 3.85% |
| BD EZ RSV | 1 | *Negative | 1 | 100.00% |
| Alere BinaxNOW RSV | 10 | *Negative | 10 | 100.00% |
| Quidel QuickVue RSV | 3 | *Negative | 3 | 100.00% |
| Wampole Clearview RSV | 4 | *Negative | 3 | 75.00% |
| SAS RSV Alert | 1 | *Negative | 1 | 100.00% |
| Quidel QuickVue RSV 10 | 2 | *Negative | 2 | 100.00% |
| Quidel Sofia RSV | 5 | *Negative | 5 | 100.00% |
| Wampole Clearview RSV | 4 | Positive | 1 | 25.00% |
| VAD-8 | | * Denotes Correct Response | | |
| AG: All Methods | 26 | | | |
| | | *Negative | 25 | 96.15% |
| | | Positive | 1 | 3.85% |
| BD EZ RSV | 1 | *Negative | 1 | 100.00% |
| Alere BinaxNOW RSV | 10 | *Negative | 10 | 100.00% |
| Quidel QuickVue RSV | 3 | *Negative | 3 | 100.00% |
| Wampole Clearview RSV | 4 | *Negative | 3 | 75.00% |
| SAS RSV Alert | 1 | *Negative | 1 | 100.00% |
| Quidel QuickVue RSV 10 | 2 | *Negative | 2 | 100.00% |
| Quidel Sofia RSV | 5 | *Negative | 5 | 100.00% |
| Wampole Clearview RSV | 4 | Positive | 1 | 25.00% |
| VAD-9 | | * Denotes Correct Response | | |
| AG: All Methods | 26 | | | |
| | | *Positive | 26 | 100.00% |
| BD EZ RSV | 1 | *Positive | 1 | 100.00% |
| Alere BinaxNOW RSV | 10 | *Positive | 10 | 100.00% |
| Quidel QuickVue RSV | 3 | *Positive | 3 | 100.00% |
| Wampole Clearview RSV | 4 | *Positive | 4 | 100.00% |
| SAS RSV Alert | 1 | *Positive | 1 | 100.00% |
| Quidel QuickVue RSV 10 | 2 | *Positive | 2 | 100.00% |
| Quidel Sofia RSV | 5 | *Positive | 5 | 100.00% |
| VAD-10 | | * Denotes Correct Response | | |
| AG: All Methods | 26 | | | |
| | | *Positive | 25 | 96.15% |
| | | Negative | 1 | 3.85% |
| BD EZ RSV | 1 | *Positive | 1 | 100.00% |
| Alere BinaxNOW RSV | 10 | *Positive | 10 | 100.00% |
| Quidel QuickVue RSV | 3 | *Positive | 3 | 100.00% |
| Wampole Clearview RSV | 4 | *Positive | 3 | 75.00% |
| SAS RSV Alert | 1 | *Positive | 1 | 100.00% |
| Quidel QuickVue RSV 10 | 2 | *Positive | 2 | 100.00% |
| Quidel Sofia RSV | 5 | *Positive | 5 | 100.00% |
| Wampole Clearview RSV | 4 | Negative | 1 | 25.00% |
| Adenovirus | | | | |
| VAD-6 | | * Denotes Correct Response | | |
| AG: All Methods | 14 | | | |
| | | *Negative | 14 | 100.00% |
| SAS Adeno Test | 10 | *Negative | 10 | 100.00% |
| RPS AdenoPlus | 4 | *Negative | 4 | 100.00% |
| VAD-7 | | * Denotes Correct Response | | |
| AG: All Methods | 14 | | | |
| | | *Negative | 13 | 92.86% |
| | | Positive | 1 | 7.14% |
| SAS Adeno Test | 10 | *Negative | 9 | 90.00% |
| RPS AdenoPlus | 4 | *Negative | 4 | 100.00% |
| SAS Adeno Test | 10 | Positive | 1 | 10.00% |
| VAD-8 | | * Denotes Correct Response | | |
| AG: All Methods | 14 | | | |
| | | *Positive | 14 | 100.00% |
| SAS Adeno Test | 10 | *Positive | 10 | 100.00% |



| | | | | |
|--|-------------|-----------------------------------|---------------|----------------|
| RPS AdenoPlus | 4 | *Positive | 4 | 100.00% |
| VAD-9 | | * Denotes Correct Response | | |
| AG: All Methods | 14 | | | |
| | | *Negative | 14 | 100.00% |
| SAS Adeno Test | 10 | *Negative | 10 | 100.00% |
| RPS AdenoPlus | 4 | *Negative | 4 | 100.00% |
| VAD-10 | | * Denotes Correct Response | | |
| AG: All Methods | 14 | | | |
| | | *Negative | 14 | 100.00% |
| SAS Adeno Test | 10 | *Negative | 10 | 100.00% |
| RPS AdenoPlus | 4 | *Negative | 4 | 100.00% |
| 932 - Hepatitis C Antibody (Waived);767 - Hepatitis Markers;750 - Viral Markers | Labs | Response | Number | Percent |
| | | HCV (Waived) | | |
| VM-6 | | * Denotes Correct Response | | |
| AG: All Methods | 4 | | | |
| | | *Negative | 4 | 100.00% |
| OraSure OraQuick HCV Rapid Antibody Test | 4 | *Negative | 4 | 100.00% |
| VM-7 | | * Denotes Correct Response | | |
| AG: All Methods | 4 | | | |
| | | *Negative | 4 | 100.00% |
| OraSure OraQuick HCV Rapid Antibody Test | 4 | *Negative | 4 | 100.00% |
| | | HBsAg | | |
| VM-6 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Negative | 1 | 100.00% |
| Siemens Centaur | 1 | *Negative | 1 | 100.00% |
| VM-7 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Positive | 1 | 100.00% |
| Siemens Centaur | 1 | *Positive | 1 | 100.00% |
| 767 - Hepatitis Markers;750 - Viral Markers | Labs | Response | Number | Percent |
| | | HBsAg | | |
| VM-6 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Negative | 1 | 100.00% |
| Siemens Centaur | 1 | *Negative | 1 | 100.00% |
| VM-7 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Positive | 1 | 100.00% |
| Siemens Centaur | 1 | *Positive | 1 | 100.00% |
| VM-8 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Positive | 1 | 100.00% |
| Siemens Centaur | 1 | *Positive | 1 | 100.00% |
| VM-9 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Negative | 1 | 100.00% |
| Siemens Centaur | 1 | *Negative | 1 | 100.00% |
| VM-10 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Negative | 1 | 100.00% |
| Siemens Centaur | 1 | *Negative | 1 | 100.00% |
| | | HCV | | |
| VM-6 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Negative | 1 | 100.00% |
| Siemens Centaur | 1 | *Negative | 1 | 100.00% |
| VM-7 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Negative | 1 | 100.00% |
| Siemens Centaur | 1 | *Negative | 1 | 100.00% |
| VM-8 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Positive | 1 | 100.00% |
| Siemens Centaur | 1 | *Positive | 1 | 100.00% |
| VM-9 | | * Denotes Correct Response | | |



| | | | | |
|--------------------------|---|-----------------------------------|---|---------|
| AG: All Methods | 1 | | | |
| | | *Positive | 1 | 100.00% |
| Siemens Centaur | 1 | *Positive | 1 | 100.00% |
| VM-10 | | * Denotes Correct Response | | |
| AG: All Methods | 1 | | | |
| | | *Negative | 1 | 100.00% |
| Siemens Centaur | 1 | *Negative | 1 | 100.00% |
| | | Anti-HBc (International) | | |
| VM-6 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Negative | 3 | 100.00% |
| Intl OCD Vitros ECi/ECiQ | 3 | *Negative | 3 | 100.00% |
| VM-7 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Positive | 3 | 100.00% |
| Intl OCD Vitros ECi/ECiQ | 3 | *Positive | 3 | 100.00% |
| VM-8 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Positive | 3 | 100.00% |
| Intl OCD Vitros ECi/ECiQ | 3 | *Positive | 3 | 100.00% |
| VM-9 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Positive | 3 | 100.00% |
| Intl OCD Vitros ECi/ECiQ | 3 | *Positive | 3 | 100.00% |
| VM-10 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Positive | 3 | 100.00% |
| Intl OCD Vitros ECi/ECiQ | 3 | *Positive | 3 | 100.00% |
| | | HBsAb (International) | | |
| VM-6 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Negative | 3 | 100.00% |
| Intl OCD Vitros ECi/ECiQ | 3 | *Negative | 3 | 100.00% |
| VM-7 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Negative | 3 | 100.00% |
| Intl OCD Vitros ECi/ECiQ | 3 | *Negative | 3 | 100.00% |
| VM-8 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Negative | 3 | 100.00% |
| Intl OCD Vitros ECi/ECiQ | 3 | *Negative | 3 | 100.00% |
| VM-9 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Positive | 3 | 100.00% |
| Intl OCD Vitros ECi/ECiQ | 3 | *Positive | 3 | 100.00% |
| VM-10 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Positive | 3 | 100.00% |
| Intl OCD Vitros ECi/ECiQ | 3 | *Positive | 3 | 100.00% |
| | | HBsAg (International) | | |
| VM-6 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Negative | 3 | 100.00% |
| Intl OCD Vitros ECi/ECiQ | 3 | *Negative | 3 | 100.00% |
| VM-7 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Positive | 3 | 100.00% |
| Intl OCD Vitros ECi/ECiQ | 3 | *Positive | 3 | 100.00% |
| VM-8 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Positive | 3 | 100.00% |
| Intl OCD Vitros ECi/ECiQ | 3 | *Positive | 3 | 100.00% |
| VM-9 | | * Denotes Correct Response | | |
| AG: All Methods | 3 | | | |
| | | *Negative | 3 | 100.00% |
| Intl OCD Vitros ECi/ECiQ | 3 | *Negative | 3 | 100.00% |
| VM-10 | | * Denotes Correct Response | | |



| | | | | |
|--|------|-----------|--------|---------|
| AG: All Methods | 3 | | | |
| | | *Negative | 3 | 100.00% |
| Intl OCD Vitros ECi/ECiQ | 3 | *Negative | 3 | 100.00% |
| HCV (International) | | | | |
| * Denotes Correct Response | | | | |
| VM-6 | | | | |
| AG: All Methods | 3 | | | |
| | | *Negative | 3 | 100.00% |
| Intl OCD Vitros ECi/ECiQ | 3 | *Negative | 3 | 100.00% |
| * Denotes Correct Response | | | | |
| VM-7 | | | | |
| AG: All Methods | 3 | | | |
| | | *Negative | 3 | 100.00% |
| Intl OCD Vitros ECi/ECiQ | 3 | *Negative | 3 | 100.00% |
| * Denotes Correct Response | | | | |
| VM-8 | | | | |
| AG: All Methods | 3 | | | |
| | | *Positive | 3 | 100.00% |
| Intl OCD Vitros ECi/ECiQ | 3 | *Positive | 3 | 100.00% |
| * Denotes Correct Response | | | | |
| VM-9 | | | | |
| AG: All Methods | 3 | | | |
| | | *Positive | 3 | 100.00% |
| Intl OCD Vitros ECi/ECiQ | 3 | *Positive | 3 | 100.00% |
| * Denotes Correct Response | | | | |
| VM-10 | | | | |
| AG: All Methods | 3 | | | |
| | | *Negative | 3 | 100.00% |
| Intl OCD Vitros ECi/ECiQ | 3 | *Negative | 3 | 100.00% |
| 771 - Affirm VP | | | | |
| | Labs | Response | Number | Percent |
| Candida sp. - Non Culture | | | | |
| * Denotes Correct Response | | | | |
| VP-6 | | | | |
| AG: All Methods | 39 | | | |
| | | *Positive | 38 | 97.44% |
| | | Negative | 1 | 2.56% |
| Affirm VP | 39 | *Positive | 38 | 97.44% |
| Affirm VP | | Negative | 1 | 2.56% |
| * Denotes Correct Response | | | | |
| VP-7 | | | | |
| AG: All Methods | 39 | | | |
| | | *Negative | 39 | 100.00% |
| Affirm VP | 39 | *Negative | 39 | 100.00% |
| * Denotes Correct Response | | | | |
| VP-8 | | | | |
| AG: All Methods | 39 | | | |
| | | *Positive | 38 | 97.44% |
| | | Negative | 1 | 2.56% |
| Affirm VP | 39 | *Positive | 38 | 97.44% |
| Affirm VP | | Negative | 1 | 2.56% |
| * Denotes Correct Response | | | | |
| VP-9 | | | | |
| AG: All Methods | 39 | | | |
| | | *Positive | 37 | 94.87% |
| | | Negative | 2 | 5.13% |
| Affirm VP | 39 | *Positive | 37 | 94.87% |
| Affirm VP | | Negative | 2 | 5.13% |
| * Denotes Correct Response | | | | |
| VP-10 | | | | |
| AG: All Methods | 39 | | | |
| | | *Negative | 36 | 92.31% |
| | | Positive | 3 | 7.69% |
| Affirm VP | 39 | *Negative | 36 | 92.31% |
| Affirm VP | | Positive | 3 | 7.69% |
| G. vaginalis - Non Culture Tech | | | | |
| * Denotes Correct Response | | | | |
| VP-6 | | | | |
| AG: All Methods | 39 | | | |
| | | *Negative | 39 | 100.00% |
| Affirm VP | 39 | *Negative | 39 | 100.00% |
| * Denotes Correct Response | | | | |
| VP-7 | | | | |
| AG: All Methods | 39 | | | |
| | | *Positive | 39 | 100.00% |
| Affirm VP | 39 | *Positive | 39 | 100.00% |
| * Denotes Correct Response | | | | |
| VP-8 | | | | |
| AG: All Methods | 39 | | | |
| | | *Negative | 38 | 97.44% |
| | | Positive | 1 | 2.56% |
| Affirm VP | 39 | *Negative | 38 | 97.44% |



| | | | | |
|-----------------|----|-----------------------------------|----|---------|
| Affirm VP | | Positive | 1 | 2.56% |
| VP-9 | | * Denotes Correct Response | | |
| AG: All Methods | 39 | | | |
| | | *Negative | 39 | 100.00% |
| Affirm VP | 39 | *Negative | 39 | 100.00% |
| VP-10 | | * Denotes Correct Response | | |
| AG: All Methods | 39 | | | |
| | | *Negative | 37 | 94.87% |
| | | Positive | 2 | 5.13% |
| Affirm VP | 39 | *Negative | 37 | 94.87% |
| Affirm VP | | Positive | 2 | 5.13% |
| | | T. vaginalis - Non Culture | | |
| VP-6 | | * Denotes Correct Response | | |
| AG: All Methods | 39 | | | |
| | | *Negative | 39 | 100.00% |
| Affirm VP | 39 | *Negative | 39 | 100.00% |
| VP-7 | | * Denotes Correct Response | | |
| AG: All Methods | 39 | | | |
| | | *Negative | 39 | 100.00% |
| Affirm VP | 39 | *Negative | 39 | 100.00% |
| VP-8 | | * Denotes Correct Response | | |
| AG: All Methods | 39 | | | |
| | | *Negative | 39 | 100.00% |
| Affirm VP | 39 | *Negative | 39 | 100.00% |
| VP-9 | | * Denotes Correct Response | | |
| AG: All Methods | 39 | | | |
| | | *Negative | 38 | 97.44% |
| | | Positive | 1 | 2.56% |
| Affirm VP | 39 | *Negative | 38 | 97.44% |
| Affirm VP | | Positive | 1 | 2.56% |
| VP-10 | | * Denotes Correct Response | | |
| AG: All Methods | 39 | | | |
| | | *Positive | 36 | 92.31% |
| | | Negative | 3 | 7.69% |
| Affirm VP | 39 | *Positive | 36 | 92.31% |
| Affirm VP | | Negative | 3 | 7.69% |