



666 - Afinion Glycohemoglobin;998 - CVS Afinion Glycohemoglobin	AFF-3																								
Glycohemoglobin	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Afinion Glycohemoglobin	141	6.06	0.27	5.30-6.90	4.43																				
747 - Auto Diff 7 (AcT 5 Diff & Horiba 5-Part Diff)	BCX-11					BCX-12					BCX-13					BCX-14					BCX-15				
White Blood Cell Count	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Beckman Coulter ACT 5 diff (v 1.9)	15	4.84	0.29	4.10-5.50	6.04	15	27.35	0.48	23.20-31.40	1.76	15	6.53	0.17	5.50-7.50	2.62	15	6.96	0.14	6.00-8.10	2.00	15	3.86	0.10	3.30-4.50	2.56
PG: Beckman Coulter ACT 5 diff	6	4.93	0.15	4.20-5.60	3.06	6	27.15	0.60	23.10-31.30	2.21	6	6.53	0.17	5.50-7.50	2.77	6	6.88	0.25	5.90-7.90	3.62	6	3.82	0.07	3.20-4.40	2.11
PG: Horiba ABX Pentra 60	16	5.02	0.12	4.30-5.80	2.40	16	28.01	0.72	23.80-32.20	2.57	16	6.66	0.20	5.70-7.70	2.99	16	7.11	0.21	6.00-8.20	2.96	16	3.91	0.10	3.30-4.50	2.56
PG: Horiba ABX Pentra 60C+ (v. 2.21 & up)	6	4.87	0.10	4.20-5.60	2.04	6	27.28	0.56	23.20-31.40	2.05	6	6.48	0.15	5.50-7.50	2.31	6	7.00	0.16	6.00-8.10	2.14	6	3.90	0.20	3.30-4.50	5.13
SG: Select Beckman Coulter ACT Instruments (5-part)	21	4.87	0.26	4.20-5.60	5.31	21	27.29	0.51	23.20-31.40	1.87	21	6.53	0.17	5.50-7.50	2.62	21	6.94	0.17	5.90-7.90	2.46	21	3.85	0.09	3.20-4.40	2.37
SG: Select Horiba ABX Instruments (5-part Diff)	23	4.97	0.13	4.30-5.80	2.60	23	27.68	0.96	23.50-31.90	3.47	23	6.60	0.20	5.60-7.60	3.03	23	7.08	0.20	6.00-8.20	2.82	23	3.91	0.13	3.30-4.50	3.33
SG: International Horiba Instruments	25	4.97	0.13	4.30-5.80	2.60	25	27.56	1.01	23.50-31.70	3.66	25	6.59	0.21	5.60-7.60	3.18	25	7.07	0.19	6.00-8.20	2.68	25	3.91	0.12	3.30-4.50	3.08
Lymphocyte	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Beckman Coulter ACT 5 diff (v 1.9)	15	20.63	2.57	12.90-28.30	12.48	15	18.24	1.27	14.40-22.00	6.98	15	28.35	2.05	22.30-34.50	7.22	15	29.79	3.77	18.50-41.10	12.65	15	26.60	2.88	18.00-35.20	10.83
PG: Beckman Coulter ACT 5 diff	6	20.77	3.36	10.70-30.90	16.15	6	17.87	1.76	12.60-23.20	9.83	6	26.90	3.08	17.70-36.10	11.45	6	27.18	3.15	17.80-36.60	11.58	6	25.43	0.95	22.50-28.30	3.74
PG: Horiba ABX Pentra 60	16	25.68	2.37	18.60-32.80	9.22	16	20.77	1.36	16.70-24.90	6.54	16	31.16	2.85	22.60-39.80	9.13	16	31.88	2.90	23.20-40.60	9.09	16	30.63	3.27	20.80-40.40	10.69
PG: Horiba ABX Pentra 60C+ (v. 2.21 & up)	6	23.35	3.23	13.70-33.10	13.80	6	19.60	0.96	16.70-22.50	4.90	6	29.03	1.88	23.40-34.60	6.48	6	30.28	3.51	19.80-40.80	11.58	6	28.07	2.80	19.70-36.50	9.96
SG: Select Beckman Coulter ACT Instruments (5-part)	21	20.67	2.73	12.50-28.90	13.19	21	18.13	1.39	13.90-22.30	7.68	21	27.94	2.40	20.70-35.10	8.60	21	29.04	3.73	17.80-40.20	12.86	21	26.27	2.52	18.70-33.90	9.58
SG: Select Horiba ABX Instruments (5-part Diff)	23	25.16	2.75	16.90-33.50	10.91	23	20.73	1.90	15.00-26.40	9.18	23	30.93	3.17	21.40-40.40	10.26	23	31.57	3.08	22.40-40.80	9.75	23	29.88	3.23	20.20-39.60	10.80
SG: International Horiba Instruments	25	26.04	4.07	13.80-38.20	15.62	25	21.53	3.41	11.30-31.70	15.86	25	32.08	5.00	17.10-47.10	15.58	25	32.55	4.61	18.70-46.30	14.18	25	30.91	4.73	16.70-45.10	15.31
Monocyte	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Beckman Coulter ACT 5 diff (v 1.9)	14	0.84	0.32	0.00-1.80	38.10	15	0.93	0.39	0.00-2.11	42.37	14	1.29	0.41	0.05-2.53	32.09	15	1.09	0.51	0.00-2.61	46.51	13	0.62	0.21	0.00-1.25	33.71
PG: Beckman Coulter ACT 5 diff	6	1.30	0.49	0.00-2.76	37.38	6	0.80	0.36	0.00-1.87	44.75	6	1.58	0.65	0.00-3.54	41.27	6	1.12	0.46	0.00-2.49	40.89	6	0.78	0.37	0.00-1.88	46.92
PG: Horiba ABX Pentra 60	16	1.08	0.42	0.00-2.35	39.17	16	0.72	0.17	0.22-1.22	23.33	16	1.07	0.50	0.00-2.58	47.10	16	0.96	0.40	0.00-2.15	41.15	16	0.63	0.23	0.00-1.33	36.98
PG: Horiba ABX Pentra 60C+ (v. 2.21 & up)	6	0.98	0.45	0.00-2.34	46.33	6	0.95	0.44	0.00-2.26	46.00	6	1.37	0.80	0.00-3.76	58.18	6	1.03	0.52	0.00-2.58	50.10	6	0.72	0.29	0.00-1.58	39.72
SG: Select Beckman Coulter ACT Instruments (5-part)	20	0.98	0.42	0.00-2.25	43.27	21	0.90	0.38	0.00-2.04	42.22	20	1.38	0.50	0.00-2.87	36.09	21	1.10	0.48	0.00-2.55	43.82	19	0.67	0.27	0.00-1.47	40.00
SG: Select Horiba ABX Instruments (5-part Diff)	23	1.20	0.83	0.00-3.69	69.17	23	0.94	0.82	0.00-3.40	87.34	23	1.41	1.37	0.00-5.52	97.09	23	1.22	1.22	0.00-4.88	100.08	23	0.75	0.53	0.00-2.33	70.13
SG: International Horiba Instruments	25	1.24	0.81	0.00-3.66	65.08	25	0.93	0.79	0.00-3.30	84.84	25	1.46	1.35	0.00-5.50	92.19	25	1.19	1.18	0.00-4.73	99.24	25	0.73	0.51	0.00-2.25	69.59
Granulocyte	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Beckman Coulter ACT 5 diff (v 1.9)	15	60.75	2.99	51.70-69.70	4.93	15	67.16	1.40	63.00-71.40	2.08	15	57.06	1.67	52.10-62.10	2.92	15	56.05	2.94	47.20-64.80	5.23	15	54.87	3.85	43.30-66.50	7.01
PG: Beckman Coulter ACT 5 diff	5	60.36	3.77	49.10-71.70	6.24	5	67.40	2.43	60.10-74.70	3.61	5	57.52	2.41	50.30-64.70	4.19	5	57.94	1.28	54.10-61.70	2.21	5	56.04	1.24	52.30-59.70	2.21
PG: Horiba ABX Pentra 60	16	67.11	1.96	61.20-73.00	2.92	16	73.26	1.20	69.70-76.90	1.64	16	60.71	2.58	53.00-68.40	4.25	16	59.76	2.76	51.50-68.10	4.62	16	57.54	3.34	47.50-67.50	5.81
PG: Horiba ABX Pentra 60C+ (v. 2.21 & up)	6	69.37	2.56	61.70-77.10	3.69	6	74.15	0.76	71.90-76.50	1.02	6	63.28	2.04	57.20-69.40	3.22	6	60.55	3.52	49.90-71.10	5.83	6	59.20	3.50	48.70-69.70	5.91
SG: Select Beckman Coulter ACT Instruments (5-part)	20	60.65	3.10	51.40-70.00	5.11	20	67.22	1.64	62.30-72.10	2.44	20	57.17	1.82	51.70-62.70	3.18	20	56.52	2.72	48.30-64.70	4.81	20	55.16	3.40	45.00-65.40	6.16
SG: Select Horiba ABX Instruments (5-part Diff)	23	66.95	4.35	53.90-80.00	6.50	23	72.69	4.08	60.50-84.90	5.61	23	60.53	4.97	45.60-75.40	8.23	23	59.43	3.89	47.70-71.10	6.55	23	57.50	4.05	45.40-69.70	7.04
SG: International Horiba Instruments	25	66.17	4.98	51.30-81.10	7.52	25	72.02	4.60	58.20-85.80	6.39	25	59.49	5.97	41.60-77.40	10.03	25	58.56	4.96	43.70-73.50	8.46	25	56.73	4.75	42.50-71.00	8.38
Eosinophil	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Beckman Coulter ACT 5 diff (v 1.9)	14	6.10	1.61	1.30-10.90	26.39	14	5.41	0.64	3.50-7.30	11.85	15	6.87	0.94	4.10-9.70	13.62	15	8.05	1.32	4.00-12.00	16.50	15	11.01	2.00	5.00-17.00	18.27
PG: Beckman Coulter ACT 5 diff	6	4.75	0.72	2.60-7.00	15.00	6	5.05	0.67	3.10-7.10	13.14	6	6.17	1.29	2.30-10.10	20.81	6	7.70	2.23	1.00-14.40	28.96	6	10.60	1.45	6.30-14.90	13.68
PG: Horiba ABX Pentra 60	16	5.56	1.54	1.00-10.20	27.50	15	4.71	0.65	2.80-6.70	13.83	15	6.53	1.04	3.40-9.60	16.00	16	7.17	1.33	3.20-11.20	18.47	14	10.87	1.16	7.40-14.40	10.64
PG: Horiba ABX Pentra 60C+ (v. 2.21 & up)	6	5.67	0.81	3.30-8.10	14.21	6	5.05	0.68	3.10-7.10	13.33	6	5.98	1.25	2.30-9.80	20.83	6	7.90	1.49	3.40-12.40	18.86	6	11.65	2.12	5.30-18.10	18.12
SG: Select Beckman Coulter ACT Instruments (5-part)	20	5.70	1.52	1.10-10.30	26.67	20	5.31	0.65	3.30-7.30	12.26	21	6.67	1.07	3.50-9.90	15.97	21	7.95	1.58	3.20-12.60	20.00	21	10.89	1.84	5.40-16.40	16.88
SG: Select Horiba ABX Instruments (5-part Diff)	23	5.57	1.33	1.60-9.60	23.75	22	4.81	0.65	2.80-6.80	13.54	22	6.39	1.07	3.20-9.60	16.72	23	7.31	1.38	3.20-11.40	18.90	21	11.18	1.49	6.70-15.70	13.30
SG: International Horiba Instruments	25	5.47	1.34	1.50-9.50	24.36	24	4.73	0.68	2.70-6.70	14.47	24	6.26	1.11	3.00-9.60	17.62	25	7.25	1.38	3.20-11.40	18.90	23	10.95	1.63	6.10-15.90	14.82
Basophil	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %															



SG: International Horiba Instruments	25	12.57	0.21	11.70-13.50	1.67	25	7.67	0.13	7.20-8.20	1.69	25	13.44	0.22	12.50-14.30	1.64	25	13.42	0.23	12.50-14.30	1.72	25	9.71	0.17	9.00-10.40	1.75					
Hematocrit	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%					
PG: Beckman Coulter ACT 5 diff (v 1.9)	15	35.53	0.43	33.40-37.60	1.21	15	22.97	0.36	21.60-24.40	1.57	15	37.94	0.59	35.60-40.20	1.56	15	37.83	0.54	35.50-40.10	1.43	15	28.37	0.34	26.70-30.10	1.20					
PG: Beckman Coulter ACT 5 diff	6	35.60	0.35	33.50-37.70	0.98	6	22.97	0.33	21.60-24.40	1.43	6	37.75	0.41	35.50-40.10	1.08	6	38.25	0.95	36.00-40.60	2.48	6	28.53	0.48	26.80-30.20	1.68					
PG: Horiba ABX Pentra 60	16	35.47	0.71	33.40-37.60	2.00	16	22.96	0.52	21.60-24.40	2.26	16	37.94	0.59	35.60-40.20	1.56	16	37.92	0.79	35.60-40.20	2.08	16	28.50	0.55	26.80-30.20	1.93					
PG: Horiba ABX Pentra 60C+ (v. 2.21 & up)	6	35.37	0.51	33.30-37.50	1.44	6	22.83	0.43	21.40-24.20	1.89	6	37.85	0.82	35.50-40.10	2.17	6	37.78	0.77	35.50-40.10	2.04	6	28.28	0.54	26.60-30.00	1.91					
SG: Select Beckman Coulter ACT Instruments (5-part)	21	35.55	0.40	33.40-37.60	1.13	21	22.97	0.34	21.60-24.40	1.48	21	37.89	0.54	35.60-40.20	1.42	21	37.95	0.69	35.60-40.20	1.82	21	28.41	0.38	26.70-30.10	1.34					
SG: Select Horiba ABX Instruments (5-part Diff)	23	35.33	0.81	33.20-37.40	2.29	23	22.87	0.54	21.50-24.30	2.36	23	37.79	0.87	35.50-40.10	2.30	23	37.78	0.90	35.50-40.10	2.38	23	28.38	0.61	26.70-30.10	2.15					
SG: International Horiba Instruments	25	35.36	0.79	33.30-37.50	2.23	25	22.86	0.52	21.50-24.30	2.27	25	37.77	0.84	35.50-40.10	2.22	25	37.74	0.88	35.40-40.00	2.33	25	28.37	0.59	26.70-30.10	2.08					
MCV	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%					
PG: Beckman Coulter ACT 5 diff (v 1.9)	15	83.51	0.63	81.60-85.40	0.75	15	74.68	0.84	72.20-77.20	1.12	15	83.53	0.75	81.30-85.80	0.90	15	83.75	0.69	81.60-85.80	0.82	15	73.51	0.80	71.10-75.90	1.09					
PG: Beckman Coulter ACT 5 diff	6	82.92	0.80	80.50-85.30	0.97	6	74.22	0.75	72.00-76.50	1.01	6	83.13	0.33	82.10-84.10	0.40	6	83.50	0.55	81.80-85.20	0.66	6	73.32	0.80	70.90-75.70	1.09					
PG: Horiba ABX Pentra 60	16	83.81	0.66	81.80-85.80	0.79	16	74.94	0.77	72.60-77.20	1.03	16	84.00	0.73	81.80-86.20	0.87	16	83.94	0.85	81.40-86.50	1.01	15	74.00	0.66	72.00-76.00	0.88					
PG: Horiba ABX Pentra 60C+ (v. 2.21 & up)	6	83.50	1.05	80.30-86.70	1.26	6	74.67	0.82	72.20-77.20	1.10	6	84.00	1.10	80.70-87.30	1.31	6	83.83	0.98	80.90-86.70	1.17	6	73.67	0.82	71.20-76.20	1.11					
SG: Select Beckman Coulter ACT Instruments (5-part)	21	83.34	0.72	81.10-85.50	0.86	21	74.55	0.83	72.00-77.00	1.11	21	83.41	0.67	81.40-85.40	0.80	21	83.68	0.65	81.80-85.70	0.78	21	73.46	0.79	71.20-75.80	1.06					
SG: Select Horiba ABX Instruments (5-part Diff)	23	83.70	0.77	81.40-86.00	0.91	23	74.91	0.79	72.50-77.30	1.05	23	83.96	0.82	81.50-86.50	0.98	23	83.87	0.87	81.30-86.50	1.04	22	74.00	0.82	71.50-76.50	1.11					
SG: International Horiba Instruments	25	83.68	0.75	81.50-86.00	0.90	25	75.00	0.82	72.50-77.50	1.09	25	83.92	0.81	81.50-86.30	0.97	25	83.80	0.87	81.20-86.40	1.04	24	74.00	0.78	71.70-76.30	1.05					
RDW	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%					
PG: Beckman Coulter ACT 5 diff (v 1.9)	15	12.44	0.34	11.40-13.40	2.74	15	15.15	0.35	14.10-16.30	2.30	15	12.39	0.47	11.00-13.80	3.79	15	12.33	0.52	10.70-13.90	4.23	15	15.25	0.41	14.00-16.40	2.70					
PG: Beckman Coulter ACT 5 diff	6	12.87	0.70	10.80-15.00	5.43	6	15.63	0.68	13.60-17.60	4.36	6	12.70	0.81	10.30-15.10	6.38	6	12.62	0.70	10.50-14.70	5.56	6	15.68	0.48	14.30-17.10	3.06					
PG: Horiba ABX Pentra 60	16	12.23	0.32	11.20-13.20	2.62	16	15.06	0.43	13.80-16.40	2.85	16	12.29	0.34	11.30-13.40	2.85	16	12.19	0.36	11.10-13.30	2.95	16	15.19	0.45	13.90-16.50	2.89					
PG: Horiba ABX Pentra 60C+ (v. 2.21 & up)	6	12.40	0.40	11.20-13.60	3.15	6	15.22	0.66	13.20-17.20	4.34	6	12.27	0.42	11.00-13.60	3.41	6	12.42	0.49	10.90-13.90	3.95	6	15.15	0.58	13.50-16.90	3.82					
SG: Select Beckman Coulter ACT Instruments (5-part)	21	12.56	0.49	11.10-14.10	3.89	21	15.29	0.50	13.80-16.80	3.27	21	12.48	0.58	10.80-14.20	4.64	21	12.41	0.58	10.70-14.10	4.68	21	15.37	0.46	14.00-16.80	2.99					
SG: Select Horiba ABX Instruments (5-part Diff)	23	12.23	0.40	11.00-13.40	3.28	23	15.06	0.53	13.50-16.70	3.51	23	12.26	0.38	11.20-13.40	3.09	23	12.21	0.45	10.80-13.50	3.69	23	15.13	0.52	13.50-16.70	3.44					
SG: International Horiba Instruments	25	12.27	0.41	11.10-13.50	3.33	25	15.17	0.64	13.30-17.10	4.21	25	12.29	0.39	11.10-13.50	3.17	25	12.25	0.47	10.90-13.70	3.74	25	15.22	0.58	13.50-16.90	3.82					
Platelet Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%					
PG: Beckman Coulter ACT 5 diff (v 1.9)	15	257.53	9.91	194.00-323.00	3.84	15	411.47	16.11	308.00-514.00	3.92	15	260.27	11.68	195.00-325.00	4.50	15	250.60	8.05	188.00-314.00	3.23	15	113.00	3.78	85.00-141.00	3.36					
PG: Beckman Coulter ACT 5 diff	6	254.17	5.04	191.00-318.00	1.97	6	416.00	6.10	312.00-520.00	1.47	6	261.00	3.69	196.00-326.00	1.42	6	255.33	7.03	191.00-319.00	2.75	6	111.83	4.96	84.00-140.00	4.46					
PG: Horiba ABX Pentra 60	16	258.88	9.32	194.00-324.00	3.59	16	412.88	7.58	310.00-516.00	1.84	16	266.13	9.21	200.00-333.00	3.46	16	256.75	8.25	193.00-321.00	3.23	16	114.81	6.34	86.00-144.00	5.48					
PG: Horiba ABX Pentra 60C+ (v. 2.21 & up)	6	259.83	12.48	195.00-325.00	4.81	6	406.83	14.50	305.00-509.00	3.56	6	266.33	10.88	200.00-333.00	4.10	6	251.33	10.33	188.00-314.00	4.10	6	111.50	8.48	84.00-140.00	7.59					
SG: Select Beckman Coulter ACT Instruments (5-part)	21	256.57	8.80	193.00-321.00	3.42	21	412.76	13.98	310.00-516.00	3.39	21	260.48	9.95	195.00-325.00	3.85	21	251.95	7.91	189.00-315.00	3.13	21	112.67	4.05	85.00-141.00	3.63					
SG: Select Horiba ABX Instruments (5-part Diff)	23	259.22	9.74	194.00-324.00	3.75	23	410.52	10.28	308.00-514.00	2.51	23	265.17	10.40	199.00-331.00	3.92	23	255.04	8.81	191.00-319.00	3.45	23	113.91	6.78	86.00-143.00	5.96					
SG: International Horiba Instruments	25	259.32	9.42	194.00-324.00	3.63	25	410.36	9.87	308.00-513.00	2.41	25	264.84	10.28	199.00-331.00	3.89	25	255.16	8.80	191.00-319.00	3.45	25	113.52	6.65	86.00-143.00	5.88					
915 - Body Fluid pH	BFPH-3																													
pH - Body Fluid	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%					
PG: Nitrazine Paper	5	6.00	0.00	6.00-6.00	0.00																									
702 - Blood Gases	BG-11						BG-12						BG-13						BG-14						BG-15					
Blood Gas - pCO2	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%					
SG: Non - i-STAT Blood Gas Instruments	5	21.04	1.86	16.00-26.00	9.05	5	28.72	2.72	24.00-34.00	9.31	5	78.50	5.41	73.00-85.00	6.84	5	33.98	2.36	29.00-39.00	7.06	5	21.36	1.17	16.00-26.00	5.71					
AG: All Group	7	21.11	1.53	16.00-26.00	7.14	7	28.53	2.25	24.00-34.00	7.59	6	75.48	4.51	69.00-81.00	6.00	7	33.10	2.50	28.00-38.00	7.58	7	21.30	1.17	16.00-26.00	5.71					
Blood Gas - pH	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%					
SG: Non - i-STAT Blood Gas Instruments	5	7.44	0.01	7.40-7.48	0.13	5	7.46	0.01	7.42-7.50	0.12	5	7.11	0.02	7.07-7.15	0.27	5	7.47	0.01	7.43-7.51	0.13	5	7.45	0.01	7.41-7.49	0.20					
AG: All Group	7	7.46	0.03	7.42-7.50	0.40	7	7.46	0.02	7.42-7.50	0.21	7	7.12	0.02	7.08-7.16	0.22	7	7.49	0.03	7.45-7.53	0.41	7	7.46	0.03	7.42-7.50	0.38					
Blood Gas - pO2	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%					
SG: Non - i-STAT Blood Gas Instruments	5	175.00	13.23	135.00-215.00	7.54	5	143.80	14.03	102.00-186.00	9.72	5	76.60																		



PG: Intl Abbott i-Stat	5	20057.5	1324.53	16083.9-24031.1	6.60	5	4612.50	485.62	3155.65-6069.35	10.53	5	312.50	26.30	233.60-391.40	8.42	5	9637.50	1103.10	6328.21-12946.8	11.45	5	38290.0	3311.78	28354.7-48225.3	8.65
D-dimer (Quantitative)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: AlereTriage	70	1591.00	147.68	1148.00-2034.00	9.28	42	812.62	60.91	630.00-996.00	7.50	36	565.11	66.73	365.00-765.00	11.81	36	1076.50	94.33	794.00-1360.00	8.76	36	2580.56	236.93	1870.00-3292.00	9.18
PG: Beckman (Olympus) AU	5	860.00	0.00	860.00-860.00	0.00	5	290.00	0.00	290.00-290.00	0.00	5	250.00	0.00	250.00-250.00	0.00	5	470.00	0.00	470.00-470.00	0.00	5	1660.00	0.00	1660.00-1660.00	0.00
AG: All Group	77	1593.38	181.31	1049.50-2137.30	11.38	48	808.65	100.95	506.00-1112.00	12.48	42	567.38	87.93	303.00-831.00	15.50	42	1080.33	150.16	629.00-1531.00	13.91	42	2581.43	276.88	1750.00-3412.00	10.73
680 - Coagulation - Plasma	CG-11					CG-12					CG-13					CG-14					CG-15				
Protime - Plasma	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	5	11.10	0.32	9.40-12.80	2.88	5	19.98	0.66	17.00-23.00	3.30	5	11.45	0.30	9.80-13.20	2.61	5	26.50	1.11	22.50-30.50	4.19	5	21.63	1.51	18.40-24.80	6.99
INR	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	5	1.05	0.06	0.90-1.30	5.45	5	1.93	0.13	1.50-2.30	6.84	5	1.10	0.08	0.90-1.30	7.27	5	2.48	0.13	2.10-2.90	5.20	5	2.00	0.14	1.60-2.40	7.00
612 - MultiChem 5;616 - MultiChem 10;617 - MultiChem 15;618 - MultiChem 20;619 - MultiChem 21+;717 - Abaxis Piccolo (Waived);676 - Endocrinology;719 - Quantitative hCG (Serum);718 - Abaxis Piccolo (Non-Waived)	CH-11N					CH-12N					CH-13N					CH-14N					CH-15N				
HCG - Quant	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	19	951.13	168.23	446.00-1456.00	17.69	20	1215.13	217.13	564.00-1866.00	17.87	21	402.11	57.13	231.00-573.00	14.20	21	256.80	31.63	162.00-352.00	12.30	20	820.89	102.09	515.00-1127.00	12.44
Sodium (NA+)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	12	147.08	2.02	143.00-151.00	1.36	12	156.58	2.43	153.00-161.00	1.53	12	126.17	1.80	122.00-130.00	1.43	11	121.45	1.29	117.00-125.00	1.07	12	140.67	2.64	137.00-145.00	1.84
PG: Abaxis Piccolo (Waived)	57	145.61	2.67	142.00-150.00	1.85																				
PG: Abbott Architect	7	143.75	0.89	140.00-148.00	0.63	7	154.50	1.51	151.00-159.00	0.97	7	122.75	0.46	119.00-127.00	0.41	7	117.63	0.52	114.00-122.00	0.42	7	139.00	1.41	135.00-143.00	1.01
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	30	146.12	1.42	142.00-150.00	0.96	29	157.39	1.13	153.00-161.00	0.70	29	121.57	1.15	118.00-126.00	0.98	29	115.07	0.97	111.00-119.00	0.87	29	138.57	0.80	135.00-143.00	0.58
PG: Clinical Data Envoy 500	8	140.29	2.00	136.00-144.00	1.43	8	150.69	2.02	147.00-155.00	1.32	8	121.74	2.21	118.00-126.00	1.80	8	116.51	2.36	113.00-121.00	2.05	8	136.16	2.32	132.00-140.00	1.69
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	19	144.47	1.58	140.00-148.00	1.11	19	153.58	1.68	150.00-158.00	1.10	19	124.53	1.43	121.00-129.00	1.12	19	119.68	1.25	116.00-124.00	1.00	19	139.11	1.66	135.00-143.00	1.22
PG: Horiba ABX Pentra 400	6	145.17	2.99	141.00-149.00	2.07	6	156.17	2.14	152.00-160.00	1.35	6	122.83	1.33	119.00-127.00	1.06	6	117.50	2.07	114.00-122.00	1.78	6	139.17	1.47	135.00-143.00	1.08
PG: Beckman (Olympus) AU	17	142.25	1.21	138.00-146.00	0.85	17	152.45	1.43	148.00-156.00	0.92	17	122.10	1.02	118.00-126.00	0.82	17	116.70	1.13	113.00-121.00	0.94	17	137.70	0.80	134.00-142.00	0.58
PG: Roche Cobas Integra / Roche Reagent	8	143.38	1.27	139.00-147.00	0.91	8	153.26	1.25	149.00-157.00	0.85	8	122.41	0.70	118.00-126.00	0.57	8	117.32	0.83	113.00-121.00	0.68	8	138.15	0.60	134.00-142.00	0.43
PG: Roche Cobas Mira / DCL Reagent (Sekisui)	5	151.00	0.00	147.00-155.00	0.00	5	165.00	0.00	161.00-169.00	0.00	5	122.00	0.00	118.00-126.00	0.00	5	114.00	0.00	110.00-118.00	0.00	5	143.00	0.00	139.00-147.00	0.00
PG: OCD Vitros 250,350,400,500,750,950	10	149.90	1.79	146.00-154.00	1.20	10	161.60	1.78	158.00-166.00	1.11	10	125.80	1.48	122.00-130.00	1.19	10	119.70	1.57	116.00-124.00	1.33	10	143.80	2.20	140.00-148.00	1.53
SG: Select Alfa Wasserman Instruments / Alfa Wasserman Reagent	31	146.08	1.41	142.00-150.00	0.96	30	157.41	1.12	153.00-161.00	0.70	30	121.58	1.14	118.00-126.00	0.90	30	115.07	0.96	111.00-119.00	0.87	30	138.58	0.79	135.00-143.00	0.58
SG: Select Beckman Coulter Instruments (Chem)	18	142.29	1.19	138.00-146.00	0.85	18	152.52	1.44	149.00-157.00	0.92	18	122.14	1.01	118.00-126.00	0.82	18	116.76	1.14	113.00-121.00	0.94	18	137.76	0.83	134.00-142.00	0.58
SG: Select Roche Instruments / Roche Reagent	9	143.24	1.27	139.00-147.00	0.91	9	153.24	1.19	149.00-157.00	0.78	9	122.27	0.79	118.00-126.00	0.66	9	117.19	0.88	113.00-121.00	0.77	9	138.13	0.56	134.00-142.00	0.43
SG: Select Roche Instruments / Roche Reagents Groups	10	143.94	2.63	140.00-148.00	1.81	10	154.31	3.72	150.00-158.00	2.40	10	122.24	0.76	118.00-126.00	0.66	10	116.90	1.28	113.00-121.00	1.11	10	138.57	1.56	135.00-143.00	1.15
SG: Select OCD Vitros Instruments	12	150.31	1.80	146.00-154.00	1.20	12	162.38	2.26	158.00-166.00	1.42	12	125.92	1.32	122.00-130.00	1.03	12	119.92	1.50	116.00-124.00	1.25	12	144.38	2.22	140.00-148.00	1.53
SG: Select Abbott Architect Instruments	9	143.50	1.18	140.00-148.00	0.83	9	154.40	1.35	150.00-158.00	0.84	9	122.80	0.42	119.00-127.00	0.33	9	117.50	0.53	114.00-122.00	0.42	9	138.80	1.40	135.00-143.00	1.01
SG: International OCD Vitros Instruments	12	150.17	1.75	146.00-154.00	1.13	12	162.00	1.91	158.00-166.00	1.17	12	126.00	1.48	122.00-130.00	1.19	13	119.85	1.46	116.00-124.00	1.25	13	144.15	2.12	140.00-148.00	1.46
SG: International Abaxis Piccolo Instruments	73	146.00	2.63	142.00-150.00	1.78	16	157.00	2.34	153.00-161.00	1.46	16	126.56	1.75	123.00-131.00	1.42	15	121.80	1.26	118.00-126.00	1.07	16	141.19	2.46	137.00-145.00	1.77
SG: Select Gen Chem Instruments	42	142.57	2.14	139.00-147.00	1.47	42	153.11	2.29	149.00-157.00	1.50	42	122.28	1.28	118.00-126.00	1.07	42	116.96	1.47	113.00-121.00	1.28	42	137.88	1.65	134.00-142.00	1.23
AG: All Group	185	145.16	2.92	141.00-149.00	2.00	127	155.51	3.56	152.00-160.00	2.31	127	123.14	2.10	119.00-127.00	1.71	126	117.58	2.39	114.00-122.00	2.03	127	139.17	2.50	135.00-143.00	1.80
Potassium (K+)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	12	5.66	0.17	5.20-6.20	2.98	11	6.55	0.16	6.10-7.10	2.42	11	3.58	0.18	3.10-4.10	5.00	12	3.08	0.10	2.60-3.60	3.23	12	5.13	0.17	4.60-5.60	3.33
PG: Abaxis Piccolo (Waived)	56	5.65	0.20	5.10-6.10	3.57																				
PG: Abbott Architect	7	5.61	0.04	5.10-6.10	0.71	7	6.60	0.05	6.10-7.10	0.76	7	3.68	0.05	3.20-4.20	1.35	7	3.23	0.12	2.70-3.70	3.75	7	5.14	0.05	4.60-5.60	0.98
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	30	5.72	0.14	5.20-6.20	2.46	29	6.74	0.24	6.20-7.20	3.58	29	3.72	0.13	3.20-4.20	3.51	29	3.19	0.16	2.70-3.70	5.00	29	5.17	0.09	4.70-5.70	1.73
PG: Clinical Data Envoy 500	8	5.45	0.16	5.00-6.00	2.91	8	6.46	0.16	6.00-7.00	2.46	8	3.61	0.10	3.10-4.10	2.78	8	3.19	0.08	2.70-3.70	2.50	8	5.03	0.15	4.50-5.50	3.00
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	19	5.62	0.07	5.10-6.10	1.25	19	6.56	0.08	6.10-7.10	1.21	19	3.67	0.05	3.20-4.20	1.35	19	3.18	0.05	2.70-3.70	1.56	19	5.14	0.07	4.60-5.60	1.37
PG: Horiba ABX Pentra 400	6	5.47	0.12	5.00-6.00	2.18	6	6.38	0.12	5.90-6.90	1.88	6	3.67	0.08	3.20-4.20	2.16	6	3.23	0.05	2.70-3.70	1.56	6	5.00	0.11	4.50-5.50	2.20
PG: Beckman (Olympus)																									



PG: OCD Vitros 250,350,400,500,750,950	10	102.20	1.55	97.00-107.00	1.47	10	106.20	1.40	101.00-111.00	1.32	10	94.50	1.35	90.00-100.00	1.47	10	93.00	1.16	88.00-98.00	1.29	10	100.80	1.62	96.00-106.00	1.58
SG: Select Alfa Wasserman Instruments / Alfa Wasserman Reagent	30	105.40	0.99	100.00-110.00	0.95	29	110.16	0.88	105.00-116.00	0.82	29	94.28	1.41	89.00-99.00	1.49	29	91.26	1.51	86.00-96.00	1.65	29	102.04	1.17	97.00-107.00	1.18
SG: Select Beckman Coulter Instruments (Chem)	17	102.32	0.67	97.00-107.00	0.69	17	105.42	0.90	100.00-110.00	0.86	17	94.89	0.88	90.00-100.00	0.95	17	92.95	1.13	88.00-98.00	1.18	17	100.42	0.77	95.00-105.00	0.80
SG: Select Roche Instruments / Roche Reagent	9	104.14	3.42	99.00-109.00	3.27	9	109.27	4.09	104.00-114.00	3.76	9	94.41	2.37	89.00-99.00	2.55	9	92.16	2.62	87.00-97.00	2.83	9	101.48	3.20	96.00-106.00	3.17
SG: Select Roche Instruments / Roche Reagents Groups	10	104.58	3.56	100.00-110.00	3.43	10	109.70	4.14	105.00-116.00	3.73	10	94.73	2.50	90.00-100.00	2.63	10	92.51	2.74	88.00-98.00	2.90	10	101.89	3.33	97.00-107.00	3.24
SG: Select OCD Vitros Instruments	12	102.15	1.35	97.00-107.00	1.27	12	106.23	1.24	101.00-111.00	1.13	12	94.23	1.30	89.00-99.00	1.38	12	92.77	1.09	88.00-98.00	1.18	12	100.85	1.41	96.00-106.00	1.39
SG: Abaxis Piccolo Instruments	69	104.94	2.10	100.00-110.00	2.00	12	107.58	2.35	103.00-113.00	2.22	12	98.92	2.31	94.00-104.00	2.32	12	97.67	2.81	93.00-103.00	2.86	12	102.92	2.71	98.00-108.00	2.62
SG: Select Abbott Architect Instruments	9	103.10	0.88	98.00-108.00	0.87	9	106.60	1.26	102.00-112.00	1.21	9	96.50	0.85	92.00-102.00	0.82	9	94.80	0.92	90.00-100.00	0.95	9	102.00	0.94	97.00-107.00	0.88
SG: Select Electrolyte Instruments 1	7	104.07	2.05	99.00-109.00	1.92	7	109.00	4.00	104.00-114.00	3.67	7	92.36	0.63	87.00-97.00	0.65	7	90.00	1.83	86.00-95.00	2.00	7	100.79	2.34	96.00-106.00	2.28
SG: International OCD Vitros Instruments	12	102.42	1.50	97.00-107.00	1.47	12	106.50	1.51	102.00-112.00	1.40	12	94.75	1.42	90.00-100.00	1.47	13	93.08	1.12	88.00-98.00	1.18	13	101.08	1.66	96.00-106.00	1.68
SG: International Siemens Instruments	21	105.19	1.03	100.00-110.00	0.95	21	110.52	1.08	105.00-117.00	0.99	21	94.19	0.98	89.00-99.00	1.06	21	91.43	0.98	86.00-96.00	1.10	21	102.43	0.93	97.00-107.00	0.88
SG: International Abaxis Piccolo Instruments	73	104.95	2.09	100.00-110.00	2.00	16	107.69	2.24	103.00-113.00	2.04	16	98.88	2.13	94.00-104.00	2.12	16	97.19	2.71	92.00-102.00	2.78	16	102.44	2.61	97.00-107.00	2.55
AG: All Group	181	104.42	2.10	99.00-109.00	2.02	123	108.51	2.80	104.00-114.00	2.57	123	94.88	2.11	90.00-100.00	2.21	123	92.61	2.56	88.00-98.00	2.80	123	101.78	1.88	97.00-107.00	1.86
CO2 (Total CO2 Bicarb)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	11	26.55	1.70	22.00-32.00	6.30	11	29.82	1.17	26.00-34.00	4.00	11	16.82	0.98	14.00-20.00	5.88	11	14.09	1.14	11.00-17.00	7.86	11	23.73	0.91	21.00-27.00	3.75
PG: Abaxis Piccolo (Waived)	57	26.77	1.34	23.00-31.00	4.81																				
PG: Abbott Architect	7	24.88	1.96	19.00-31.00	8.00	7	29.50	1.93	24.00-36.00	6.33	7	17.13	1.36	13.00-21.00	8.24	7	14.88	1.36	11.00-19.00	9.33	7	22.88	1.96	17.00-29.00	8.70
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	30	26.62	2.63	19.00-35.00	9.63	29	30.43	2.55	23.00-38.00	8.33	29	17.36	2.22	10.00-24.00	12.94	29	15.43	1.73	10.00-20.00	11.33	29	23.26	2.63	15.00-31.00	11.30
PG: Clinical Data Envoy 500	8	26.95	2.09	21.00-33.00	7.78	8	31.33	3.04	22.00-40.00	9.68	8	16.04	2.17	9.00-23.00	13.75	8	14.16	1.77	9.00-19.00	12.86	8	24.14	2.94	15.00-33.00	12.08
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	19	28.82	2.04	23.00-35.00	6.90	19	32.16	1.80	27.00-37.00	5.63	19	19.29	1.48	15.00-24.00	7.89	19	17.19	1.45	13.00-22.00	8.82	19	25.35	1.73	20.00-30.00	6.80
PG: Horiba ABX Pentra 400	5	27.22	3.02	18.00-36.00	11.11	6	29.72	4.14	18.00-42.00	13.67	6	17.83	3.25	8.00-28.00	18.33	6	15.92	2.62	8.00-24.00	16.25	5	25.42	2.06	19.00-31.00	8.40
PG: Beckman (Olympus) AU	17	27.60	1.73	23.00-33.00	6.07	17	31.80	1.91	26.00-38.00	5.94	17	18.05	1.40	14.00-22.00	7.78	17	15.75	1.25	12.00-20.00	8.13	17	24.40	1.70	19.00-29.00	7.08
PG: Roche Cobas Integra / Roche Reagent	8	24.05	1.55	19.00-29.00	6.67	8	27.29	2.16	20.00-34.00	8.15	8	15.80	1.28	12.00-20.00	8.13	8	13.83	1.34	10.00-18.00	9.29	8	20.87	2.07	15.00-27.00	10.00
PG: OCD Vitros 250,350,400,500,750,950	10	25.40	2.01	19.00-31.00	8.00	10	29.50	2.37	23.00-37.00	8.00	10	16.30	1.06	13.00-19.00	6.88	10	13.60	0.84	12.00-16.00	5.71	10	21.50	1.90	16.00-28.00	8.64
SG: Select Beckman Coulter Instruments (Chem)	18	27.48	1.78	22.00-32.00	6.67	18	31.62	2.04	26.00-38.00	6.25	18	18.05	1.36	14.00-22.00	7.78	18	15.76	1.22	12.00-20.00	7.50	18	24.38	1.66	19.00-29.00	7.08
SG: Select Roche Instruments / Roche Reagent	9	24.15	1.49	20.00-29.00	6.25	9	27.46	2.10	21.00-33.00	7.78	9	15.92	1.27	12.00-20.00	8.13	9	13.95	1.31	10.00-18.00	9.29	9	21.08	2.06	15.00-27.00	10.00
SG: Select Roche Instruments / Roche Reagents Groups	10	23.95	1.56	19.00-29.00	6.67	10	27.09	2.34	20.00-34.00	8.52	10	15.67	1.47	12.00-21.00	9.38	10	13.86	1.28	10.00-18.00	9.29	10	20.89	2.06	15.00-27.00	10.00
SG: Select OCD Vitros Instruments	12	25.31	1.75	20.00-30.00	7.20	12	29.15	2.15	22.00-36.00	7.59	12	16.62	1.19	13.00-21.00	7.06	12	14.00	1.41	10.00-18.00	10.00	12	21.62	1.80	17.00-27.00	8.18
SG: Abaxis Piccolo Instruments	68	26.74	1.39	23.00-31.00	5.19																				
SG: Select Abbott Architect Instruments	9	25.10	2.02	19.00-31.00	8.00	9	29.50	1.84	25.00-35.00	6.00	9	16.90	1.60	12.00-22.00	9.41	9	14.90	1.29	11.00-19.00	8.67	9	23.00	1.89	17.00-29.00	8.26
SG: International OCD Vitros Instruments	11	24.91	2.51	18.00-33.00	10.00	11	28.82	3.19	19.00-39.00	11.03	11	16.00	1.41	12.00-20.00	8.75	12	13.42	1.38	9.00-17.00	10.77	12	21.17	2.41	14.00-28.00	11.43
SG: International Abaxis Piccolo Instruments	72	26.64	1.43	23.00-31.00	5.19	15	29.60	1.18	26.00-34.00	4.00	15	16.60	1.12	14.00-20.00	6.47	15	13.93	1.10	11.00-17.00	7.86	15	23.40	1.06	20.00-26.00	4.78
Calcium	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	11	10.91	0.31	9.90-11.90	2.84	11	12.14	0.20	11.10-13.10	1.65	11	8.58	0.14	7.60-9.60	1.63	11	7.80	0.37	6.80-8.80	4.74	11	10.46	0.17	9.50-11.50	1.62
PG: Abaxis Piccolo (Waived)	57	11.02	0.22	10.00-12.00	2.00																				
PG: Abbott Architect	7	11.09	0.16	10.10-12.10	1.44	7	12.45	0.24	11.50-13.50	1.92	7	8.35	0.17	7.40-9.40	2.02	7	7.70	0.15	6.70-8.70	1.95	7	10.43	0.21	9.40-11.40	2.02
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	30	10.83	0.30	9.80-11.80	2.78	29	12.09	0.31	11.10-13.10	2.56	29	8.56	0.22	7.60-9.60	2.56	29	7.88	0.24	6.90-8.90	3.04	29	10.41	0.26	9.40-11.40	2.50
PG: Clinical Data Envoy 500	8	10.68	0.28	9.70-11.70	2.62	8	11.83	0.32	10.80-12.80	2.71	8	8.07	0.15	7.10-9.10	1.85	8	7.48	0.40	6.50-8.50	5.33	8	10.07	0.29	9.10-11.10	2.87
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	18	10.90	0.27	9.90-11.90	2.48	18	12.26	0.28	11.30-13.30	2.28	18	8.31	0.20	7.30-9.30	2.41	18	7.62	0.23	6.60-8.60	3.03	18	10.26	0.28	9.30-11.30	2.62
PG: Horiba ABX Pentra 400	6	11.27	0.16	10.30-12.30	1.42	6	12.55	0.21	11.60-13.60	1.67	6	8.50	0.09	7.50-9.50	1.06	6	7.78	0.12	6.80-8.80	1.54	6	10.67	0.10	9.70-11.70	0.93
PG: Beckman (Olympus) AU	17	10.95	0.24	9.90-11.90	2.20	17	12.24	0.34	11.20-13.20	2.79	17	8.33	0.20	7.30-9.30	2.29	17	7.67	0.17	6.70-8.70	2.34	17	10.31	0.23	9.30-11.30	2.23
PG: Roche Cobas Integra / Roche Reagent																									



SG: Select SpotChem Instruments	12	3.00	0.25	2.50-3.50	8.40	12	3.78	0.40	3.20-4.40	10.53	12	1.39	0.06	1.09-1.69	4.53	12	1.01	0.06	0.70-1.30	6.00	12	2.64	0.17	2.24-3.04	6.59	
SG: Select Abbott Architect Instruments	9	3.48	0.08	3.00-4.00	2.39	9	4.48	0.09	3.80-5.20	2.08	9	1.47	0.03	1.17-1.77	2.24	9	0.95	0.02	0.70-1.30	2.53	9	2.99	0.08	2.54-3.44	2.67	
SG: International OCD Vitros Instruments	12	3.09	0.11	2.60-3.60	3.66	12	3.95	0.12	3.40-4.60	3.00	12	1.40	0.04	1.10-1.70	3.07	13	1.01	0.05	0.70-1.30	5.00	13	2.70	0.10	2.29-3.11	3.70	
SG: International Abaxis Piccolo Instruments	73	3.37	0.19	2.90-3.90	5.61	17	4.27	0.17	3.70-4.90	3.95	17	1.60	0.13	1.30-1.90	8.13	17	1.25	0.13	0.90-1.50	10.83	17	2.95	0.11	2.51-3.39	3.67	
Glucose	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	
PG: Abaxis Piccolo (Non-Waived)	12	183.08	1.31	165.00-201.00	0.71	12	224.67	1.37	203.00-248.00	0.62	12	98.42	1.08	88.00-108.00	1.12	12	77.08	1.17	69.00-85.00	1.56	12	161.92	1.38	146.00-178.00	0.86	
PG: Abaxis Piccolo (Waived)	56	182.39	1.76	164.00-200.00	0.99																					
PG: Abbott Architect	7	185.25	3.96	167.00-204.00	2.16	7	229.50	6.19	207.00-253.00	2.70	7	98.00	2.45	88.00-108.00	2.45	7	75.50	2.51	68.00-84.00	3.29	7	162.75	4.23	147.00-179.00	2.58	
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	31	193.84	6.87	175.00-213.00	3.56	30	237.47	8.61	213.00-261.00	3.63	30	104.67	3.83	95.00-116.00	3.62	30	82.33	3.54	74.00-90.00	4.27	30	172.03	6.16	155.00-189.00	3.60	
PG: Clinical Data Envoy 500	8	204.20	6.46	184.00-224.00	3.19	8	250.91	6.74	226.00-276.00	2.67	8	108.61	2.97	98.00-120.00	2.75	8	85.50	2.88	77.00-95.00	3.37	8	180.56	5.01	163.00-199.00	2.76	
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	19	189.11	3.77	170.00-208.00	2.01	19	232.11	5.13	209.00-255.00	2.20	19	101.42	2.50	91.00-111.00	2.48	19	79.00	2.36	71.00-87.00	3.04	19	167.47	3.66	150.00-184.00	2.22	
PG: Horiba ABX Pentra 400	6	186.50	7.34	168.00-206.00	3.90	6	229.33	7.50	206.00-252.00	3.28	6	97.17	2.79	87.00-107.00	2.89	6	75.50	1.87	68.00-84.00	2.50	6	166.83	6.85	150.00-184.00	4.13	
PG: Beckman (Olympus) AU	17	189.80	3.62	171.00-209.00	1.89	17	233.35	4.26	210.00-256.00	1.85	17	100.90	2.47	91.00-111.00	2.48	17	78.30	1.95	70.00-86.00	2.44	17	168.30	4.03	151.00-185.00	2.38	
PG: Roche Cobas Integra / Roche Reagent	8	187.19	2.35	168.00-206.00	1.28	8	230.29	3.15	207.00-253.00	1.35	8	100.23	1.19	90.00-110.00	1.20	8	77.58	1.42	70.00-86.00	1.79	8	166.15	2.40	149.00-183.00	1.45	
PG: OCD Vitros 250,350,400,500,750,950	10	178.80	3.79	161.00-197.00	2.12	10	222.50	4.38	201.00-245.00	1.97	10	93.10	2.47	84.00-102.00	2.69	10	72.90	1.97	66.00-80.00	2.74	10	156.50	3.10	141.00-173.00	1.97	
SG: Select Beckman Coulter Instruments (Chem)	18	189.52	3.75	171.00-209.00	1.95	18	233.14	4.26	210.00-256.00	1.85	18	100.86	2.41	91.00-111.00	2.38	18	78.24	1.92	70.00-86.00	2.44	18	168.05	4.09	151.00-185.00	2.44	
SG: Select Roche Instruments / Roche Reagent	9	186.97	2.32	168.00-206.00	1.23	9	230.16	2.99	207.00-253.00	1.30	9	100.11	1.19	90.00-110.00	1.20	9	77.32	1.56	69.00-85.00	2.08	9	165.94	2.36	149.00-183.00	1.45	
SG: Select Roche Instruments / Roche Reagents Groups	10	187.16	2.29	168.00-206.00	1.23	10	230.60	3.19	208.00-254.00	1.39	10	100.46	1.63	90.00-110.00	1.60	10	77.65	1.85	70.00-86.00	2.44	10	166.03	2.26	149.00-183.00	1.39	
SG: Select OCD Vitros Instruments	12	179.00	3.34	161.00-197.00	1.84	12	223.15	4.00	201.00-245.00	1.79	12	93.23	2.20	84.00-102.00	2.37	12	73.08	1.80	66.00-80.00	2.47	12	156.46	2.85	140.00-172.00	1.79	
SG: Abaxis Piccolo Instruments	68	182.51	1.71	165.00-201.00	0.93																					
SG: Select SpotChem Instruments	11	182.91	14.10	165.00-201.00	7.70	11	224.27	7.20	202.00-246.00	3.21	11	94.00	3.79	85.00-103.00	4.04	11	73.36	2.42	66.00-80.00	3.29	11	159.45	10.23	143.00-175.00	6.42	
SG: Select Abbott Architect Instruments	9	185.30	3.56	167.00-204.00	1.95	9	229.70	5.52	207.00-253.00	2.39	9	98.00	2.21	88.00-108.00	2.24	9	75.40	2.27	68.00-83.00	3.07	9	163.20	3.97	147.00-179.00	2.45	
SG: International OCD Vitros Instruments	12	179.00	3.49	161.00-197.00	1.96	12	223.08	4.21	201.00-245.00	1.88	12	93.00	2.26	84.00-102.00	2.47	13	72.85	1.82	66.00-80.00	2.47	13	156.69	2.72	141.00-173.00	1.72	
SG: International Siemens Instruments	21	187.58	6.03	169.00-207.00	3.19	21	229.80	8.96	207.00-253.00	3.91	21	100.69	3.36	91.00-111.00	3.37	21	78.24	3.29	70.00-86.00	4.23	21	165.94	6.06	149.00-183.00	3.67	
SG: International Abaxis Piccolo Instruments	73	182.56	1.77	165.00-201.00	0.98	17	224.63	2.10	203.00-248.00	0.93	17	98.62	1.10	89.00-109.00	1.11	17	77.22	1.16	69.00-85.00	1.56	17	162.22	1.64	146.00-178.00	0.99	
AG: All Group	182	187.28	7.54	168.00-206.00	4.01	125	232.72	8.67	210.00-256.00	3.73	125	100.90	4.65	91.00-111.00	4.65	125	78.72	4.06	71.00-87.00	5.19	125	167.50	7.35	150.00-184.00	4.37	
Urea Nitrogen (BUN)	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	
PG: Abaxis Piccolo (Non-Waived)	12	27.75	0.45	25.00-31.00	1.79	12	35.33	0.78	32.00-38.00	2.29	12	12.75	0.62	11.00-15.00	4.62	12	9.00	0.43	7.00-11.00	4.44	12	23.83	0.39	22.00-26.00	1.67	
PG: Abaxis Piccolo (Waived)	56	27.69	0.69	25.00-31.00	2.50																					
PG: Abbott Architect	7	31.13	1.25	28.00-34.00	3.87	7	39.50	2.62	36.00-44.00	6.50	7	14.38	0.74	12.00-16.00	5.00	7	10.88	0.99	9.00-13.00	9.09	7	26.75	1.39	25.00-29.00	5.19	
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	31	29.39	1.41	26.00-32.00	4.83	29	37.24	1.66	34.00-40.00	4.59	30	14.13	0.86	12.00-16.00	6.43	30	10.20	0.61	8.00-12.00	6.00	30	25.80	1.03	24.00-28.00	3.85	
PG: Clinical Data Envoy 500	7	30.71	1.38	28.00-34.00	4.52	6	37.50	1.76	35.00-41.00	4.74	7	14.71	0.76	13.00-17.00	5.33	7	10.57	0.98	9.00-13.00	9.09	7	26.83	1.55	25.00-29.00	5.93	
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	18	30.56	1.15	28.00-34.00	3.55	19	38.53	1.50	35.00-43.00	3.85	19	14.63	1.01	13.00-17.00	6.67	19	10.37	0.68	8.00-12.00	7.00	19	26.79	1.23	25.00-29.00	4.44	
PG: Horiba ABX Pentra 400	5	28.62	1.11	26.00-32.00	3.79	5	36.52	1.63	34.00-40.00	4.32	5	13.58	0.91	12.00-16.00	6.43	5	10.06	0.82	8.00-12.00	8.00	5	25.42	2.06	23.00-27.00	8.40	
PG: Beckman (Olympus) AU	17	30.05	0.89	27.00-33.00	3.00	17	37.90	1.07	35.00-41.00	2.89	17	14.35	0.49	12.00-16.00	3.57	17	10.50	0.61	9.00-13.00	5.45	17	26.25	1.16	24.00-28.00	4.62	
PG: Roche Cobas Integra / Roche Reagent	8	29.40	1.27	26.00-32.00	4.48	8	37.47	1.52	34.00-40.00	4.05	8	13.90	0.56	12.00-16.00	4.29	8	9.95	0.51	8.00-12.00	5.00	8	25.52	1.12	24.00-28.00	4.23	
PG: OCD Vitros 250,350,400,500,750,950	10	24.70	0.68	23.00-27.00	2.80	10	31.70	0.82	29.00-35.00	2.50	10	11.70	0.48	10.00-14.00	4.17	10	8.20	0.42	6.00-10.00	5.00	10	21.40	0.84	19.00-23.00	3.81	
SG: Select Beckman Coulter Instruments (Chem)	18	30.14	0.96	27.00-33.00	3.33	18	38.00	1.14	35.00-41.00	2.89	18	14.38	0.50	12.00-16.00	3.57	18	10.52	0.60	9.00-13.00	5.45	18	26.33	1.20	24.00-28.00	4.62	
SG: Select Roche Instruments / Roche Reagent	9	29.56	1.30	27.00-33.00	4.33	9	37.62	1.51	35.00-41.00	3.95	9	14.01	0.63	12.00-16.00	4.29	9	10.05	0.58	8.00-12.00	6.00	9	25.67	1.15	24.00-28.00	4.62	
SG: Select Roche Instruments / Roche Reagents Groups	10	29.69	1.31	27.00-33.00	4.33	10	37.75	1.49	35.00-41.00	3.95	10	14.01	0.60	12.00-16.00	4.29	10	10.14	0.62	8.00-12.00	6.00	10	25.79	1.17	24.00-28.00	4.62	
SG: Select OCD Vitros Instruments	12	24.62	0.65	23.00-27.00	2.80	12	31.85	0.80	29.00-35.00	2.50	12	11.54	0.52	10.00-14.00	4.17	12	8.15	0.38	6.00-10.00	5.00	12					



Alkaline Phosphatase	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Abaxis Piccolo (Non-Waived)	11	180.45	8.20	126.00-234.00	4.56	11	220.55	9.56	155.00-287.00	4.34	11	88.55	6.41	62.00-116.00	7.19	11	62.27	5.88	43.00-81.00	9.52	11	158.36	5.20	111.00-205.00	3.29
PG: Abaxis Piccolo (Waived)	57	178.70	9.03	125.00-233.00	5.03																				
PG: Abbott Architect	7	246.63	20.26	173.00-321.00	8.22	7	308.00	24.55	216.00-400.00	7.99	7	112.13	9.00	78.00-146.00	8.04	7	76.88	7.57	54.00-100.00	9.87	7	215.00	16.84	151.00-280.00	7.81
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	30	235.27	15.66	165.00-306.00	6.68	29	297.48	17.79	208.00-386.00	5.99	29	109.28	6.98	76.00-142.00	6.42	29	74.86	4.84	53.00-98.00	6.40	29	205.79	15.63	144.00-268.00	7.57
PG: Clinical Data Envoy 500	8	211.32	16.42	148.00-274.00	7.77	8	264.70	18.39	186.00-345.00	6.94	8	99.15	8.26	69.00-129.00	8.38	8	69.86	4.97	49.00-91.00	7.14	8	183.89	13.57	129.00-239.00	7.39
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	12	262.58	6.86	184.00-342.00	2.62	12	328.83	8.70	230.00-428.00	2.64	12	119.33	4.27	83.00-155.00	3.61	12	81.08	3.83	57.00-105.00	4.69	12	229.00	6.71	160.00-298.00	2.93
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL / IFCC	7	263.71	7.37	185.00-343.00	2.80	7	331.57	10.08	232.00-432.00	3.04	7	119.57	5.22	84.00-156.00	4.33	7	82.29	5.22	57.00-107.00	6.34	7	227.86	8.28	160.00-296.00	3.64
PG: Horiba ABX Pentra 400	6	261.17	10.32	183.00-339.00	3.95	6	323.83	11.89	227.00-421.00	3.67	6	119.17	4.75	83.00-155.00	4.03	6	82.67	5.68	58.00-108.00	6.87	6	227.00	11.98	159.00-295.00	5.29
PG: Beckman (Olympus) AU	17	215.40	11.84	151.00-280.00	5.49	17	268.70	14.77	188.00-350.00	5.50	17	98.35	5.72	69.00-127.00	5.82	17	66.25	7.09	46.00-86.00	10.76	17	186.70	10.39	131.00-243.00	5.56
PG: Roche Cobas Integra / Roche Reagent	8	232.67	4.68	163.00-303.00	2.02	8	293.17	6.75	205.00-381.00	2.29	8	106.24	2.00	74.00-138.00	1.89	8	73.21	2.24	51.00-95.00	3.01	8	200.41	6.30	140.00-260.00	3.15
PG: OCD Vitros 250,350,400,500,750,950	9	177.44	10.38	124.00-230.00	5.88	9	201.89	12.38	141.00-263.00	6.14	9	98.44	3.81	69.00-127.00	3.88	9	71.44	3.71	50.00-92.00	5.21	9	163.78	7.93	115.00-213.00	4.82
SG: Select Beckman Coulter Instruments (Chem)	18	215.05	11.65	151.00-280.00	5.44	18	268.43	14.45	188.00-348.00	5.37	18	98.29	5.59	69.00-127.00	5.71	18	66.14	6.92	46.00-86.00	10.45	18	186.38	10.23	130.00-242.00	5.48
SG: Select Roche Instruments / Roche Reagent	9	232.10	4.77	162.00-302.00	2.07	9	292.05	7.28	204.00-380.00	2.50	9	106.01	2.01	74.00-138.00	1.89	9	72.89	2.34	51.00-95.00	3.15	9	199.67	6.38	140.00-260.00	3.20
SG: Select Roche Instruments / Roche Reagents Groups	10	226.18	20.14	158.00-294.00	8.89	10	284.41	26.27	199.00-369.00	9.26	10	103.47	8.66	72.00-134.00	8.45	10	71.08	6.40	50.00-92.00	9.01	10	194.70	17.56	137.00-254.00	9.03
SG: Select OCD Vitros Instruments	11	180.25	10.25	126.00-234.00	5.72	11	207.67	15.15	146.00-270.00	7.26	11	98.92	3.50	69.00-129.00	3.54	11	72.00	3.33	50.00-94.00	4.58	11	165.58	7.55	116.00-216.00	4.52
SG: Select Abbott Architect Instruments	9	250.90	20.06	176.00-326.00	8.01	9	313.20	24.31	219.00-407.00	7.76	9	115.20	11.00	81.00-150.00	9.57	9	78.40	7.44	55.00-101.00	9.49	9	218.90	17.01	153.00-285.00	7.76
SG: International OCD Vitros Instruments	11	175.91	9.98	123.00-229.00	5.68	11	200.91	11.47	141.00-261.00	5.72	11	97.27	4.29	68.00-126.00	4.43	12	72.08	5.02	50.00-94.00	6.94	12	164.50	10.15	116.00-215.00	6.12
SG: International Abaxis Piccolo Instruments	70	178.86	9.11	125.00-233.00	5.08	13	217.00	12.43	152.00-282.00	5.71	13	88.31	6.54	62.00-114.00	7.39	13	62.38	5.38	43.00-81.00	8.71	13	157.85	6.66	111.00-205.00	4.24
AG: All Group	179	211.67	33.87	148.00-276.00	15.99	121	281.65	41.60	197.00-367.00	14.75	121	105.99	11.48	74.00-138.00	10.85	121	73.06	8.15	51.00-95.00	11.23	121	197.78	25.48	139.00-257.00	12.88
ALT/SGPT	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Abaxis Piccolo (Non-Waived)	11	163.91	5.38	131.00-197.00	3.29	11	209.91	7.11	168.00-252.00	3.38	11	71.45	3.59	57.00-85.00	5.07	11	47.73	2.57	38.00-58.00	5.42	11	139.27	4.78	111.00-167.00	3.45
PG: Abaxis Piccolo (Waived)	56	161.80	3.14	130.00-194.00	1.91																				
PG: Abbott Architect	7	186.25	3.73	149.00-223.00	1.99	7	242.50	3.63	194.00-292.00	1.48	7	73.00	2.14	58.00-88.00	2.88	7	44.75	1.49	36.00-54.00	3.33	7	157.13	3.64	126.00-188.00	2.29
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	30	151.93	5.98	122.00-182.00	3.95	29	198.83	5.46	159.00-239.00	2.76	29	60.59	2.37	49.00-73.00	3.93	29	36.24	2.57	29.00-43.00	7.22	29	130.93	4.28	105.00-157.00	3.28
PG: Clinical Data Envoy 500	8	185.81	7.11	149.00-223.00	3.82	8	239.04	8.18	191.00-287.00	3.43	8	77.00	2.27	62.00-92.00	2.99	8	48.61	1.43	39.00-59.00	2.86	8	162.66	5.86	130.00-196.00	3.62
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	7	191.29	4.89	153.00-229.00	2.57	7	247.86	5.27	198.00-298.00	2.14	7	79.29	2.69	63.00-95.00	3.42	7	50.57	2.44	41.00-61.00	4.71	7	164.43	2.57	131.00-197.00	1.59
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL / IFCC	12	188.67	3.45	151.00-227.00	1.80	12	243.58	4.60	195.00-293.00	1.89	12	77.92	2.15	62.00-94.00	2.82	12	48.75	1.76	39.00-59.00	3.67	12	162.33	2.19	130.00-194.00	1.36
PG: Beckman (Olympus) AU	17	163.65	3.79	131.00-197.00	2.32	17	212.00	4.96	170.00-254.00	2.36	17	66.15	2.76	53.00-79.00	4.24	17	40.90	1.25	33.00-49.00	3.17	17	139.40	3.44	111.00-167.00	2.45
PG: Roche Cobas Integra / Roche Reagent	8	180.72	3.01	145.00-217.00	1.66	8	233.73	4.09	187.00-281.00	1.75	8	72.52	0.99	58.00-88.00	1.37	8	44.76	0.80	36.00-54.00	1.78	8	154.16	2.85	123.00-185.00	1.88
PG: OCD Vitros 250,350,400,500,750,950	9	191.00	5.22	153.00-229.00	2.72	9	239.56	5.64	192.00-288.00	2.33	9	87.56	5.13	70.00-106.00	5.80	9	61.00	4.66	49.00-73.00	7.70	9	165.44	4.85	132.00-198.00	2.97
SG: Select Alfa Wasserman Instruments / Alfa Wasserman Reagent Groups	31	152.87	7.86	122.00-184.00	5.16	30	199.97	8.23	160.00-240.00	4.10	30	61.10	3.65	49.00-73.00	6.07	30	36.60	3.20	30.00-44.00	8.65	30	131.80	6.35	106.00-158.00	4.77
SG: Select Beckman Coulter Instruments (Chem)	18	164.52	5.45	132.00-198.00	3.27	18	213.10	6.97	170.00-256.00	3.29	18	66.76	3.88	54.00-80.00	5.82	18	41.43	2.71	33.00-49.00	6.59	18	140.24	5.10	112.00-168.00	3.64
SG: Select Roche Instruments / Roche Reagent	9	180.25	3.21	144.00-216.00	1.78	9	233.36	4.04	186.00-280.00	1.72	9	72.47	0.94	58.00-86.00	1.25	9	44.69	0.79	36.00-54.00	1.78	9	153.74	2.99	123.00-185.00	1.95
SG: Select Roche Instruments / Roche Reagents Groups	10	178.23	7.37	142.00-214.00	4.16	10	230.42	10.48	184.00-276.00	4.57	10	71.97	1.88	58.00-86.00	2.64	10	44.35	1.34	35.00-53.00	2.95	10	152.13	6.06	122.00-182.00	4.01
SG: Select OCD Vitros Instruments	11	189.83	5.31	152.00-228.00	2.79	11	238.33	5.57	190.00-286.00	2.35	11	86.08	5.13	69.00-103.00	5.93	11	60.50	4.34	49.00-73.00	7.05	11	164.33	4.83	131.00-197.00	2.93
SG: Abaxis Piccolo Instruments	67	162.15	3.63	130.00-194.00	2.22	11	209.91	7.11	168.00-252.00	3.38	11	71.45	3.59	57.00-85.00	5.07	11	47.73	2.57	38.00-58.00	5.42	11	139.27	4.78	111.00-167.00	3.45
SG: Select Abbott Architect Instruments	9	186.20	3.33	149.00-223.00	1.77	9	242.30	3.30	194.00-290.00	1.36	9	73.30	2.00	58.00-88.00	2.74	9	45.00	1.49	36.00-54.00	3.33	9	157.40	3.27	126.00-188.00	2.10
SG: International OCD Vitros Instruments	11	190.27	6.37	152.00-228.00	3.37	11	239.09	6.27	191.00-287.00	2.64	11	86.82	6.18	70.00-104.00	7.13	12	60.58	4.52	49.00-73.00	7.38	12	164.58	5.11	132.00-198.00	3.09
SG: International Abaxis Piccolo Instruments	69	162.23	3.61	130.00-194.00	2.22	13	208.77	7.54	167.00-251.00	3.59	13	70.54	4.21	57.00-85.00	5.92	13	47.00	2.97	38.00-56.00	6.38	13				



PG: Clinical Data Envoy 500	8	2.61	0.08	2.10-3.10	3.08	8	3.40	0.13	2.70-4.10	3.82	8	1.11	0.06	0.70-1.50	5.45	8	0.71	0.04	0.30-1.10	5.71	8	2.23	0.10	1.80-2.60	4.55
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	11	2.59	0.17	2.10-3.10	6.54	11	3.33	0.21	2.60-4.00	6.36	11	1.08	0.06	0.70-1.50	5.45	11	0.67	0.07	0.30-1.10	8.57	11	2.18	0.16	1.80-2.60	7.27
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL / IFCC	6	2.68	0.12	2.20-3.20	4.44	6	3.50	0.16	2.80-4.20	4.29	6	1.15	0.08	0.80-1.60	6.67	6	0.75	0.06	0.40-1.20	6.25	6	2.28	0.12	1.80-2.80	5.22
PG: Horiba ABX Pentra 400	6	2.82	0.22	2.20-3.40	7.86	6	3.63	0.28	2.90-4.30	7.78	6	1.18	0.10	0.80-1.60	8.33	6	0.78	0.10	0.40-1.20	12.50	6	2.42	0.21	1.90-2.90	8.75
PG: Beckman (Olympus) AU	17	2.82	0.10	2.20-3.40	3.57	17	3.63	0.15	2.90-4.30	4.17	17	1.26	0.07	0.90-1.70	6.15	17	0.86	0.05	0.50-1.30	5.56	17	2.41	0.10	1.90-2.90	4.17
PG: Roche Cobas Integra / Roche Reagent	8	2.51	0.12	2.00-3.00	4.80	8	3.24	0.14	2.60-3.80	4.38	8	1.04	0.06	0.60-1.40	6.00	8	0.72	0.05	0.30-1.10	7.14	8	2.13	0.08	1.70-2.50	3.81
PG: OCD Vitros 250,350,400,500,750,950	9	2.99	0.26	2.40-3.60	8.67	9	3.88	0.30	3.10-4.70	7.69	9	1.29	0.18	0.90-1.70	13.85	9	0.83	0.13	0.40-1.20	16.25	9	2.53	0.23	2.00-3.00	9.20
SG: Select Beckman Coulter Instruments (Chem)	18	2.83	0.10	2.20-3.40	3.57	18	3.64	0.16	2.90-4.30	4.44	18	1.27	0.08	0.90-1.70	6.15	18	0.85	0.05	0.50-1.30	5.56	18	2.42	0.11	1.90-2.90	4.58
SG: Select Roche Instruments / Roche Reagents Groups	10	2.54	0.14	2.00-3.00	5.60	10	3.27	0.19	2.60-4.00	5.76	10	1.08	0.12	0.70-1.50	10.91	10	0.75	0.12	0.40-1.20	15.00	10	2.17	0.13	1.80-2.60	5.91
SG: Select OCD Vitros Instruments	11	2.93	0.32	2.30-3.50	11.03	11	3.81	0.41	3.00-4.60	10.79	11	1.23	0.21	0.80-1.60	17.50	11	0.80	0.15	0.40-1.20	18.75	11	2.47	0.29	2.00-3.00	11.60
SG: Select SpotChem Instruments	10	2.94	0.29	2.30-3.50	10.00	10	3.84	0.31	3.00-4.60	8.16	10	1.25	0.21	0.90-1.70	16.15	10	0.81	0.14	0.40-1.20	17.50	10	2.50	0.25	2.00-3.00	9.60
SG: Select Abbott Architect Instruments	9	2.61	0.11	2.10-3.10	4.23	9	3.32	0.16	2.60-4.00	4.55	9	1.11	0.10	0.70-1.50	9.09	9	0.72	0.06	0.30-1.10	8.57	9	2.21	0.11	1.80-2.60	5.00
SG: International OCD Vitros Instruments	11	2.98	0.23	2.40-3.60	7.67	11	3.87	0.28	3.10-4.70	7.18	11	1.28	0.16	0.90-1.70	12.31	12	0.83	0.12	0.40-1.20	15.00	12	2.53	0.20	2.00-3.00	8.40
SG: International Abaxis Piccolo Instruments	70	2.48	0.18	2.00-3.00	7.20	13	3.25	0.33	2.60-4.00	10.00	13	1.18	0.12	0.80-1.60	10.00	13	0.85	0.10	0.40-1.20	12.50	13	2.21	0.20	1.80-2.60	9.09
AG: All Group	180	2.67	0.28	2.20-3.20	10.37	122	3.58	0.33	2.90-4.30	9.17	122	1.20	0.15	0.80-1.60	12.50	122	0.80	0.12	0.40-1.20	15.00	122	2.37	0.23	1.90-2.90	9.58
Total Protein	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	10	6.22	0.35	5.60-6.80	5.65	10	7.17	0.11	6.50-7.90	1.53	10	4.08	0.06	3.70-4.50	1.46	10	3.60	0.11	3.20-4.00	3.06	10	5.63	0.05	5.00-6.20	0.89
PG: Abaxis Piccolo (Waived)	56	6.16	0.09	5.60-6.80	1.45																				
PG: Abbott Architect	7	6.31	0.18	5.70-6.90	2.86	7	7.45	0.19	6.80-8.30	2.53	7	4.04	0.16	3.60-4.40	4.00	7	3.49	0.14	3.10-3.90	4.00	7	5.74	0.18	5.10-6.30	3.16
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	31	6.27	0.20	5.70-6.90	3.17	30	7.33	0.23	6.60-8.00	3.15	30	4.13	0.13	3.70-4.50	3.17	30	3.56	0.09	3.20-4.00	2.50	30	5.73	0.19	5.10-6.30	3.33
PG: Clinical Data Envoy 500	8	6.21	0.15	5.60-6.80	2.42	8	7.13	0.18	6.40-7.80	2.54	8	4.08	0.14	3.70-4.50	3.41	8	3.52	0.11	3.10-3.90	3.14	8	5.67	0.15	5.10-6.30	2.63
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	19	6.44	0.18	5.80-7.00	2.81	19	7.51	0.21	6.80-8.30	2.80	19	4.22	0.13	3.80-4.60	3.10	19	3.64	0.12	3.20-4.00	3.33	19	5.88	0.17	5.30-6.50	2.88
PG: Horiba ABX Pentra 400	6	6.17	0.12	5.60-6.80	1.94	6	7.20	0.16	6.50-7.90	2.08	6	4.02	0.04	3.60-4.40	1.00	6	3.50	0.06	3.10-3.90	1.71	6	5.62	0.10	5.00-6.20	1.79
PG: Beckman (Olympus) AU	17	6.04	0.16	5.40-6.60	2.67	17	7.04	0.15	6.30-7.70	2.14	17	3.94	0.08	3.50-4.30	2.05	17	3.42	0.06	3.10-3.70	1.76	17	5.54	0.14	5.00-6.00	2.55
PG: Roche Cobas Integra / Roche Reagent	8	6.11	0.08	5.50-6.70	1.31	8	7.14	0.16	6.40-7.80	2.25	8	4.00	0.07	3.60-4.40	1.75	8	3.47	0.07	3.10-3.90	2.00	8	5.61	0.09	5.00-6.20	1.61
PG: OCD Vitros 250,350,400,500,750,950	9	5.99	0.13	5.40-6.60	2.17	9	6.98	0.15	6.30-7.70	2.14	9	4.00	0.07	3.60-4.40	1.75	9	3.50	0.07	3.10-3.90	2.00	9	5.59	0.13	5.00-6.20	2.32
SG: Select Beckman Coulter Instruments (Chem)	18	6.04	0.16	5.40-6.60	2.67	18	7.04	0.15	6.30-7.70	2.14	18	3.94	0.08	3.50-4.30	2.05	18	3.41	0.07	3.10-3.70	2.06	18	5.54	0.13	5.00-6.00	2.36
SG: Select Roche Instruments / Roche Reagent	9	6.12	0.08	5.50-6.70	1.31	9	7.15	0.15	6.40-7.80	2.11	9	4.01	0.07	3.60-4.40	1.75	9	3.47	0.07	3.10-3.90	2.00	9	5.61	0.09	5.00-6.20	1.61
SG: Select Roche Instruments / Roche Reagents Groups	10	6.13	0.08	5.50-6.70	1.31	10	7.16	0.15	6.50-7.90	2.08	10	4.02	0.07	3.60-4.40	1.75	10	3.47	0.06	3.10-3.90	1.71	10	5.62	0.09	5.00-6.20	1.61
SG: Select OCD Vitros Instruments	11	5.98	0.12	5.40-6.60	2.00	11	6.98	0.13	6.30-7.70	1.86	11	4.00	0.06	3.60-4.40	1.50	11	3.50	0.06	3.10-3.90	1.71	11	5.57	0.12	5.00-6.20	2.14
SG: Abaxis Piccolo Instruments	66	6.17	0.15	5.60-6.80	2.42																				
SG: Select Abbott Architect Instruments	9	6.28	0.17	5.70-6.90	2.86	9	7.41	0.19	6.70-8.10	2.57	9	4.02	0.15	3.60-4.40	3.75	9	3.47	0.13	3.10-3.90	3.71	9	5.72	0.16	5.10-6.30	2.81
SG: International OCD Vitros Instruments	11	5.96	0.17	5.40-6.60	2.83	11	6.95	0.18	6.30-7.70	2.57	11	3.99	0.07	3.60-4.40	1.75	12	3.49	0.09	3.10-3.90	2.57	12	5.58	0.14	5.00-6.20	2.50
SG: International Abaxis Piccolo Instruments	67	6.17	0.15	5.60-6.80	2.42	11	7.18	0.11	6.50-7.90	1.53	11	4.09	0.07	3.70-4.50	1.71	11	3.61	0.10	3.20-4.00	2.78	11	5.65	0.07	5.00-6.20	1.25
SG: Select Gen Chem Instruments	41	6.14	0.18	5.50-6.70	2.95	41	7.16	0.22	6.50-7.90	3.06	41	3.99	0.12	3.60-4.40	3.00	41	3.46	0.10	3.10-3.90	2.86	41	5.61	0.16	5.00-6.20	2.68
AG: All Group	178	6.19	0.20	5.60-6.80	3.23	121	7.24	0.25	6.50-7.90	3.47	121	4.07	0.14	3.70-4.50	3.41	121	3.53	0.12	3.10-3.90	3.43	121	5.68	0.18	5.10-6.30	3.16
Phosphorous	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	9	3.73	0.19	3.10-4.30	5.14	8	4.28	0.22	3.60-5.00	5.12	8	2.43	0.14	2.00-2.80	5.83	8	2.10	0.11	1.80-2.40	5.24	8	3.38	0.19	2.80-4.00	5.59
PG: Clinical Data Envoy 500	5	3.94	0.11	3.60-4.20	2.82	5	4.72	0.19	4.10-5.30	4.04	5	2.56	0.11	2.30-2.90	4.23	5	2.14	0.09	1.80-2.40	4.29	5	3.66	0.17	3.20-4.20	4.59
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	6	4.10	0.21	3.50-4.70	5.12	6	4.75	0.18	4.30-5.30	3.75	6	2.62	0.15	2.20-3.00	5.77	6	2.22	0.15	1.80-2.70	6.82	6	3.73	0.19	3.10-4.30	5.14
PG: Beckman (Olympus) AU	11	3.84	0.23	3.10-4.50	6.05	11	4.54	0.25	3.80-5.30	5.56	11	2.38	0.20	1.80-3.00	8.33	11	2.01	0.13	1.60-2.40	6.50	11	3.45	0.15	3.00-3.80	4.41
SG: Select Beckman Coulter Instruments (Chem)	12	3.84	0.22	3.10-4.50	5.79	12	4.54	0.24	3.80-5.20	5.33	12	2.39	0.19	1.80-3.00	7.92	12	2.01	0.13	1.60-2.40	6.50	12	3.45	0.14	3.00-4.00	4.29
PG: Intl Abaxis Piccolo	5	4.45	0.05	4.30-4.70	1.11	5	5.06	0.18	4.60-5.60	3.53	5	2.86	0.17	2.40-3.40	5.86	5	2.66	0.21	2.10-3.30	7.78					



Method	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	
PG: Abbott Architect / Direct Method	5	81.00	2.68	57.00-105.00	3.33	5	93.33	4.41	65.00-121.00	4.73	5	60.00	2.83	42.00-78.00	4.67	5	54.67	2.88	39.00-72.00	5.27	5	76.33	3.27	53.00-99.00	4.34	
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Direct Method / Alfa Wasserman Reagent	25	75.44	4.41	53.00-98.00	5.87	24	84.58	5.71	60.00-111.00	6.71	24	56.42	3.63	39.00-73.00	6.43	24	51.50	3.01	36.00-68.00	5.77	24	70.63	4.20	50.00-92.00	5.92	
SG: Alfa Was. Ace/NexCT/Alera/Axcel / Direct Method / Alfa Wasserman Reagent Groups	26	75.04	4.78	53.00-98.00	6.40	25	84.00	6.30	59.00-109.00	7.50	25	55.80	4.71	39.00-73.00	8.39	25	51.20	3.30	36.00-66.00	6.47	25	70.24	4.54	49.00-91.00	6.43	
PG: Clinical Data Envoy 500 / Direct Method	7	87.67	7.56	62.00-114.00	8.64	7	100.09	7.51	70.00-130.00	7.50	7	63.93	5.33	45.00-83.00	8.28	7	58.34	4.29	41.00-75.00	7.41	7	85.04	6.35	60.00-111.00	7.41	
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL / Direct Method	17	117.24	18.00	82.00-152.00	15.38	17	125.35	3.48	88.00-163.00	2.80	17	84.18	2.27	59.00-109.00	2.74	17	77.47	2.00	54.00-100.00	2.60	17	106.29	3.58	74.00-138.00	3.40	
PG: Horiba ABX Pentra 400 / Direct Method	5	72.40	6.46	50.00-94.00	9.03	5	83.00	8.00	58.00-108.00	9.64	5	53.40	5.86	37.00-69.00	11.13	5	47.00	4.69	33.00-61.00	10.00	5	69.80	5.97	49.00-91.00	8.57	
PG: Beckman (Olympus) AU / Direct Method	17	85.56	5.29	60.00-112.00	6.16	17	97.18	6.45	68.00-126.00	6.60	17	61.53	3.66	43.00-81.00	5.97	17	55.97	3.15	39.00-73.00	5.71	17	78.62	4.47	55.00-103.00	5.70	
PG: Roche Cobas Integra / Direct Method / Roche Reagent	7	105.42	8.71	74.00-137.00	8.29	7	119.06	11.25	83.00-155.00	9.41	7	80.07	5.68	56.00-104.00	7.13	7	73.21	4.77	51.00-95.00	6.58	7	99.17	7.52	69.00-129.00	7.58	
PG: OCD Vitros 250,350,400,500,750,950 / Direct Method	9	93.78	3.38	66.00-122.00	3.62	9	107.67	3.71	76.00-140.00	3.43	9	67.56	3.78	48.00-88.00	5.59	9	62.11	2.26	43.00-81.00	3.71	9	90.56	3.00	64.00-118.00	3.30	
SG: Select Beckman Coulter Instruments (Chem) / Direct Method	18	85.24	5.35	60.00-111.00	6.24	18	96.93	6.38	68.00-126.00	6.60	18	61.50	3.57	43.00-81.00	5.81	18	55.87	3.10	39.00-73.00	5.54	18	78.59	4.36	55.00-103.00	5.57	
SG: Select Roche Instruments / Direct Method / Roche Reagent	8	103.82	9.46	73.00-135.00	9.13	8	117.05	12.12	82.00-152.00	10.34	8	79.06	6.12	55.00-103.00	7.72	8	72.52	4.92	51.00-95.00	6.71	8	97.59	8.47	69.00-127.00	8.67	
SG: Select Roche Instruments / Direct Method / Roche Reagents Groups	9	100.96	12.71	71.00-131.00	12.57	9	113.92	15.13	80.00-148.00	13.25	9	76.70	9.42	54.00-100.00	12.21	9	70.71	7.35	50.00-92.00	10.28	9	94.59	12.40	67.00-124.00	13.05	
SG: Select OCD Vitros Instruments / Direct Method	11	93.58	3.00	66.00-122.00	3.19	11	107.50	3.18	76.00-140.00	2.96	11	66.92	3.50	47.00-87.00	5.22	11	61.25	2.53	43.00-79.00	4.10	11	89.42	3.29	62.00-116.00	3.71	
SG: Abaxis Piccolo Instruments / Direct Method	13	95.77	2.42	67.00-125.00	2.50																					
SG: Select Abbott Architect Instruments / Direct Method	7	82.25	3.41	57.00-107.00	4.15	7	94.25	4.13	66.00-122.00	4.36	7	60.75	2.96	43.00-79.00	4.92	7	55.25	2.71	39.00-72.00	4.91	7	77.25	3.41	54.00-100.00	4.42	
SG: International OCD Vitros Instruments / Direct Method	11	94.00	3.19	66.00-122.00	3.40	11	108.91	5.21	76.00-142.00	4.77	11	67.91	3.48	48.00-88.00	5.15	12	62.17	2.04	43.00-81.00	3.23	12	90.67	3.00	64.00-118.00	3.30	
SG: International Abaxis Piccolo Instruments / Direct Method	14	95.64	2.37	67.00-125.00	2.50	5	102.33	4.04	71.00-133.00	3.92	6	60.83	8.35	43.00-79.00	13.77	6	52.33	5.89	36.00-68.00	11.35	6	84.83	10.17	60.00-111.00	12.00	
AG: All Group	120	89.90	16.38	63.00-117.00	18.22	109	100.32	15.76	70.00-130.00	15.80	111	64.85	11.24	46.00-85.00	17.23	111	58.97	10.62	41.00-77.00	17.97	111	83.54	13.73	59.00-109.00	16.31	
Cholesterol LDL	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	
PG: Abaxis Piccolo (Waived) / Calc	7	51.71	1.89	46.00-58.00	3.65																					
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Calc / Alfa Wasserman Reagent	15	81.99	8.65	56.00-108.00	10.49	15	89.44	10.21	58.00-120.00	11.46	15	61.91	5.57	45.00-79.00	9.03	15	56.63	6.13	39.00-75.00	10.70	15	76.69	5.47	61.00-94.00	7.14	
PG: Clinical Data Envoy 500 / Calc	6	62.52	9.82	34.00-92.00	15.56	6	72.10	11.73	37.00-107.00	16.25	6	47.82	9.15	21.00-75.00	18.96	6	46.55	9.18	19.00-75.00	19.57	6	55.68	6.60	36.00-76.00	11.79	
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL / Calc	9	49.33	8.66	23.00-75.00	17.76	9	54.67	11.28	21.00-89.00	20.55	9	37.33	4.90	22.00-52.00	13.24	9	34.22	3.15	24.00-44.00	9.41	9	45.78	7.03	25.00-67.00	15.22	
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL / Direct Method	7	57.43	4.83	43.00-71.00	8.42	7	65.29	5.22	49.00-81.00	8.00	7	41.86	3.72	31.00-53.00	8.81	7	37.43	2.99	28.00-46.00	8.11	7	54.14	4.84	40.00-68.00	8.89	
PG: Beckman (Olympus) AU / Calc	11	63.67	10.37	33.00-95.00	16.25	11	69.55	11.32	36.00-104.00	16.14	11	49.40	10.11	19.00-79.00	20.61	11	46.42	9.80	17.00-75.00	21.30	11	61.11	11.08	28.00-94.00	18.20	
PG: OCD Vitros 250,350,400,500,750,950 / Calc	6	55.83	6.65	36.00-76.00	11.79	6	59.83	5.78	43.00-77.00	9.67	6	44.00	3.52	34.00-55.00	7.95	6	38.67	4.55	26.00-53.00	11.54	6	50.83	5.27	35.00-67.00	10.39	
SG: Select Beckman Coulter Instruments (Chem) / Direct Method	5	43.60	2.51	37.00-52.00	5.68	5	50.40	2.88	41.00-59.00	5.80	5	31.60	2.19	25.00-39.00	6.88	5	27.80	2.17	21.00-35.00	7.86	5	41.20	2.86	32.00-50.00	7.07	
SG: Select Roche Instruments / Calc / Roche Reagents Groups	5	70.72	22.14	5.00-137.00	31.13	5	78.78	26.22	0.00-158.00	33.16	5	52.26	15.05	7.00-97.00	29.04	5	47.62	13.27	8.00-88.00	27.71	5	66.16	19.71	7.00-125.00	29.85	
SG: Select OCD Vitros Instruments / Calc	8	55.89	5.33	40.00-72.00	9.46	8	60.33	4.64	46.00-74.00	7.67	8	43.33	3.94	31.00-55.00	9.07	8	38.67	4.12	27.00-51.00	10.51	8	51.44	4.42	38.00-64.00	8.63	
SG: Abaxis Piccolo Instruments / Calc	11	51.09	2.07	45.00-57.00	4.12																					
SG: International OCD Vitros Instruments / Calc	8	56.00	7.23	34.00-78.00	12.86	8	61.50	8.05	38.00-86.00	13.06	8	44.13	3.00	35.00-53.00	6.82	9	38.22	4.58	24.00-52.00	12.11	9	51.44	4.88	36.00-66.00	9.61	
SG: Select Gen Chem Instruments / Calc	22	66.79	12.50	30.00-105.00	18.66	22	73.88	13.77	33.00-115.00	18.65	22	51.20	11.48	17.00-86.00	22.55	22	48.50	10.82	17.00-81.00	22.04	22	62.53	12.96	24.00-102.00	20.63	
SG: Select Gen Chem Instruments / Direct Method	9	46.11	4.81	32.00-60.00	10.43	9	53.33	5.83	36.00-70.00	10.94	9	33.56	4.30	21.00-47.00	12.65	9	29.33	2.87	20.00-38.00	10.00	9	43.33	4.66	29.00-57.00	10.93	
MG: Calc	77	65.00	16.10	17.00-113.00	24.77	63	74.00	18.00	20.00-128.00	24.32	70	52.00	12.00	16.00-88.00	23.08	71	48.00	11.90	12.00-84.00	24.79	71	62.00	14.70	18.00-106.00	23.71	
MG: Direct Method	20	55.00	10.80	23.00-87.00	19.64	20	63.00	12.30	26.00-100.00	19.52	20	40.00	8.00	16.00-64.00	20.00	20	35.00	7.50	13.00-58.00	21.43	20	51.00	10.20	20.00-82.00	20.00	
Triglycerides	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	Labs	Mean	SD	Range	CV%	
PG: Abaxis Piccolo (Non-Waived)	5	120.20	2.28	90.00-150.00	1.92	5	143.60	2.19	108.00-180.00	1.53	5	73.20	2.86	55.00-91.00	3.97	5	59.00	0.71	44.00-74.00	1.19	5	108.60	1.82	82.00-136.00	1.65	
PG: Abaxis Piccolo (Waived)	8	119.50	0.76	90.00-150.00	0.67																					
PG: Abbott Architect	5	110.67	2.73	83.00-139.00	2.43	5	132.83	2.86	100.00-166.00	2.18	5	66.17	2.23	50.00-83.00	3.33	5	54.17	2.14	41.00-68.00	3.89	5	100.00	2.76	75.00-125.00	2.80	
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	26	117.08	6.46	88.00-146.00	5.56	25	138.04	8.16	104.00-173.00	5.94	25	74.28	6.42	56.00-93.00	8.65	23	61.30	4.29	46.00-76.00	7.05	25	107.60	5.94	81.00-135.00	5.46	
PG: Clinical Data Envoy 500	8	112.18	3.97	84.00-140.																						



AG: All Group	84	113.86	15.28	80.00-148.00	13.42	42	151.45	23.07	106.00-196.00	15.30	40	49.92	6.49	35.00-65.00	13.00	44	31.90	3.95	22.00-42.00	12.19	45	98.81	17.30	69.00-129.00	17.47	
Creatine Kinase	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	10	215.60	10.65	151.00-281.00	4.91	10	278.90	10.97	195.00-363.00	3.94	10	91.80	4.80	64.00-120.00	5.22	10	57.70	3.53	41.00-75.00	6.03	10	187.00	7.05	131.00-243.00	3.80	
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL / IFCC	7	224.29	15.34	157.00-291.00	6.83	7	290.57	23.55	204.00-378.00	8.11	7	90.43	9.78	63.00-117.00	10.89	6	51.17	3.19	36.00-66.00	6.27	7	191.43	13.06	134.00-248.00	6.86	
PG: Beckman (Olympus) AU	8	208.40	6.93	146.00-270.00	3.32	8	272.70	9.07	191.00-355.00	3.33	8	81.50	2.95	57.00-107.00	3.66	8	49.10	1.79	34.00-64.00	3.67	8	177.80	5.77	125.00-231.00	3.26	
PG: Roche Cobas Integra / Roche Reagent	5	237.48	8.06	166.00-308.00	3.42	5	310.05	8.41	217.00-403.00	2.71	5	94.30	1.75	66.00-122.00	1.81	5	57.03	1.73	40.00-74.00	2.98	5	200.83	6.46	141.00-261.00	3.23	
SG: Select Beckman Coulter Instruments (Chem)	9	210.36	9.26	147.00-273.00	4.43	9	275.18	11.91	193.00-358.00	4.33	9	82.64	4.70	58.00-108.00	5.66	9	50.00	3.44	35.00-65.00	6.80	9	179.82	8.65	126.00-234.00	4.78	
SG: Select Roche Instruments / Roche Reagent	6	236.56	7.75	166.00-308.00	3.29	6	308.75	8.41	216.00-402.00	2.72	6	93.83	2.02	66.00-122.00	2.13	6	56.60	1.95	40.00-74.00	3.51	6	200.14	6.17	140.00-260.00	3.10	
SG: Abaxis Piccolo Instruments	30	232.10	10.77	162.00-302.00	4.66																					
SG: International Abaxis Piccolo Instruments	33	233.64	11.63	164.00-304.00	4.96																					
SG: Select Gen Chem Instruments	16	218.39	19.02	153.00-283.00	8.72	16	285.61	24.79	200.00-372.00	8.67	16	85.06	7.91	60.00-111.00	9.29	16	51.78	5.61	36.00-68.00	10.77	16	186.33	15.55	130.00-242.00	8.33	
AG: All Group	75	225.14	15.80	158.00-293.00	7.02	46	286.09	24.06	200.00-372.00	8.43	46	89.54	8.25	63.00-117.00	9.11	45	54.29	5.67	38.00-70.00	10.56	46	189.64	13.67	133.00-247.00	7.21	
GGT	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: Abaxis Piccolo (Waived)	39	125.82	2.65	118.00-134.00	2.14																					
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	6	102.17	5.23	86.00-118.00	5.10	5	130.40	4.28	117.00-143.00	3.31	5	46.00	1.00	43.00-49.00	2.17	5	30.80	2.49	24.00-39.00	8.06	5	89.00	3.32	79.00-99.00	3.71	
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	6	160.67	5.85	143.00-179.00	3.66	6	203.67	7.29	182.00-226.00	3.58	6	71.50	2.51	65.00-80.00	3.47	6	48.83	1.47	45.00-54.00	3.06	6	139.17	4.54	126.00-153.00	3.24	
PG: Beckman (Olympus) AU	5	109.43	3.55	98.00-120.00	3.30	5	140.29	4.19	127.00-153.00	3.00	5	45.57	1.62	41.00-51.00	3.48	5	29.57	1.13	27.00-33.00	3.67	5	94.43	2.99	85.00-103.00	3.19	
SG: Select Beckman Coulter Instruments (Chem)	6	112.88	10.29	82.00-144.00	9.12	6	145.13	14.23	102.00-188.00	9.79	6	47.25	4.98	32.00-62.00	10.64	6	30.63	3.16	21.00-41.00	10.32	6	97.50	9.12	71.00-125.00	9.29	
SG: Abaxis Piccolo Instruments	42	125.74	2.63	118.00-134.00	2.06																					
AG: All Group	65	127.73	19.69	69.00-187.00	15.39	25	167.85	37.69	55.00-281.00	22.44	25	56.81	13.73	16.00-98.00	24.04	26	38.07	10.06	8.00-68.00	26.58	26	116.18	29.27	28.00-204.00	25.26	
LDH	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: Intl OCD Vitros 250,350,400,500,750,950	5	1610.00	2.83	1288.00-1932.00	0.17	5	2147.50	43.14	1718.00-2578.00	2.01	5	666.00	14.14	533.00-799.00	2.12	5	417.67	12.66	334.00-502.00	3.04	5	1370.67	38.02	1097.00-1645.00	2.77	
SG: Select Gen Chem Instruments	8	486.50	48.07	390.00-584.00	9.88	8	621.10	60.15	497.00-745.00	9.69	8	199.10	19.69	159.00-239.00	9.90	8	126.50	14.56	102.00-152.00	11.50	8	416.30	41.24	333.00-499.00	9.90	
AG: All Group	14	464.44	62.13	371.00-557.00	13.38	14	592.94	82.64	474.00-712.00	13.93	14	190.44	24.93	152.00-228.00	13.11	14	120.88	15.21	97.00-145.00	12.56	13	388.53	56.84	311.00-467.00	14.60	
Lipase	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	6	191.00	11.51	157.00-226.00	6.02	6	227.00	13.94	185.00-269.00	6.12	6	119.50	6.86	99.00-141.00	5.75	6	99.83	7.33	78.00-122.00	7.30	6	175.00	10.84	143.00-207.00	6.17	
PG: Beckman (Olympus) AU	8	46.25	2.25	39.00-53.00	5.00	8	56.00	2.00	50.00-62.00	3.57	7	25.43	1.62	20.00-30.00	6.40	8	21.00	1.85	15.00-27.00	9.05	8	41.25	1.17	37.00-45.00	2.93	
PG: OCD Vitros 250,350,400,500,750,950	5	536.00	14.73	492.00-580.00	2.74	5	629.00	23.39	559.00-699.00	3.72	5	350.00	4.36	337.00-363.00	1.26	5	299.33	4.04	287.00-311.00	1.34	5	485.67	19.66	427.00-545.00	4.05	
SG: Select Beckman Coulter Instruments (Chem)	9	47.11	3.33	37.00-57.00	7.02	9	56.78	2.99	48.00-66.00	5.26	8	27.25	5.37	11.00-43.00	20.00	9	22.89	5.92	5.00-41.00	25.65	9	42.56	4.07	31.00-55.00	9.53	
SG: International OCD Vitros Instruments	5	541.50	16.30	493.00-591.00	3.01	5	637.25	25.24	561.00-713.00	3.96	5	351.50	4.66	338.00-366.00	1.34	5	300.00	3.56	289.00-311.00	1.20	5	491.75	20.14	432.00-552.00	4.09	
SG: Select Gen Chem Instruments	11	47.00	3.61	36.00-58.00	7.66	11	56.73	3.64	46.00-68.00	6.32	10	26.90	5.07	12.00-42.00	18.89	11	22.45	5.50	6.00-39.00	25.00	11	42.27	4.10	30.00-54.00	9.76	
AG: All Group	15	47.40	5.47	31.00-64.00	11.70	14	57.50	3.96	46.00-70.00	6.90	13	27.23	4.62	13.00-41.00	17.04	14	22.36	4.84	8.00-36.00	21.82	14	43.64	4.92	29.00-59.00	11.14	
Bilirubin Direct	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: Abbott Architect	6	1.12	0.08	0.89-1.35	6.88	6	1.35	0.08	1.11-1.59	5.85	6	0.56	0.04	0.44-0.68	6.96	6	0.38	0.04	0.27-0.49	10.00	6	0.97	0.05	0.81-1.13	5.46	
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	21	0.94	0.14	0.52-1.36	14.89	20	1.18	0.14	0.76-1.60	11.95	20	0.50	0.07	0.29-0.71	13.80	20	0.36	0.06	0.18-0.54	16.67	20	0.84	0.10	0.53-1.15	12.38	
PG: Clinical Data Envoy 500	5	1.02	0.04	0.89-1.16	4.41	5	1.28	0.04	1.15-1.42	3.52	5	0.48	0.04	0.35-0.62	9.38	5	0.36	0.06	0.19-0.53	15.28	5	0.88	0.04	0.75-1.02	5.11	
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	5	0.50	0.07	0.29-0.71	14.20	5	0.64	0.09	0.37-0.91	13.91	5	0.22	0.04	0.09-0.36	20.45	5	0.18	0.04	0.04-0.32	25.00	5	0.42	0.08	0.17-0.67	20.00	
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL / IFCC	7	0.53	0.11	0.19-0.87	21.51	6	0.72	0.07	0.50-0.94	10.00	7	0.23	0.07	0.02-0.44	30.87	7	0.18	0.05	0.02-0.34	30.00	7	0.43	0.12	0.08-0.78	26.74	
PG: Beckman (Olympus) AU	13	0.84	0.10	0.53-1.15	12.38	13	1.03	0.12	0.68-1.38	11.36	13	0.45	0.07	0.25-0.65	14.89	13	0.33	0.05	0.18-0.48	15.45	13	0.72	0.09	0.44-1.00	12.78	
PG: Roche Cobas Integra / Roche Reagent	7	0.60	0.06	0.43-0.77	9.50	7	0.78	0.06	0.59-0.97	8.21	7	0.26	0.04	0.16-0.37	13.46	7	0.19	0.01	0.15-0.23	6.84	7	0.51	0.04	0.39-0.63	7.65	
PG: OCD Vitros 250,350,400,500,750,950	5	1.02	0.19	0.44-1.60	18.82	5	1.22	0.15	0.78-1.66	12.13	5	0.48	0.10	0.19-0.77	20.00	5	0.46	0.15	0.00-0.92	33.04	5	0.84	0.11	0.50-1.18	13.57	
SG: Select Beckman Coul																										



UIBC	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	5	205.60	10.60	173.80-237.40	5.16	5	240.80	15.26	195.10-286.60	6.33	5	136.00	10.89	103.30-168.70	8.01	5	120.20	9.18	92.70-147.70	7.64	5	192.20	15.06	147.00-237.40	7.84
Digoxin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	5	1.73	0.10	1.40-2.00	5.88	5	2.25	0.14	1.80-2.80	6.09	5	1.04	0.13	0.80-1.20	13.00	5	0.90	0.12	0.70-1.10	13.33	5	1.56	0.09	1.30-1.90	5.63
746 - Auto Diff 6 (Abbott 5-Part Diff)	CL-11					CL-12					CL-13					CL-14					CL-15				
White Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn Ruby	5	3.16	0.14	2.70-3.70	4.38	5	7.93	0.18	6.70-9.10	2.28	5	20.82	0.22	17.70-23.90	1.06	5	7.87	0.17	6.70-9.10	2.15	5	3.16	0.15	2.70-3.70	4.69
AG: All Group	6	3.15	0.12	2.60-3.60	3.87	6	7.92	0.16	6.70-9.10	2.03	6	21.02	0.52	17.90-24.10	2.48	6	7.88	0.15	6.70-9.10	1.90	6	3.17	0.13	2.70-3.70	4.06
Lymphocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn Ruby	5	40.14	2.10	33.80-46.40	5.24	5	23.56	1.24	19.90-27.30	5.25	5	14.36	1.36	10.30-18.50	9.44	5	23.64	0.94	20.80-26.40	3.98	5	39.30	1.48	34.90-43.70	3.77
AG: All Group	6	40.98	2.79	32.60-49.40	6.80	6	23.90	1.39	19.70-28.10	5.82	6	15.78	3.69	4.70-26.90	23.35	6	24.42	2.08	18.20-30.60	8.52	6	40.75	3.79	29.30-52.10	9.31
Monocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn Ruby	5	6.54	0.72	4.37-8.71	11.09	5	5.36	0.42	4.10-6.62	7.84	5	3.04	0.53	1.45-4.63	17.47	5	5.23	0.37	4.11-6.35	7.11	5	6.26	0.55	4.60-7.92	8.83
AG: All Group	6	6.22	1.02	3.15-9.29	16.48	6	5.14	0.67	3.13-7.15	13.05	6	3.77	1.84	0.00-9.30	48.86	6	5.11	0.45	3.77-6.45	8.73	6	5.93	0.94	3.11-8.75	15.85
Granulocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn Ruby	5	48.76	1.44	44.50-53.10	2.95	5	66.56	0.85	64.00-69.10	1.28	5	77.44	0.61	75.60-79.20	0.79	5	66.52	1.21	62.90-70.10	1.80	5	50.34	1.27	46.50-54.10	2.52
AG: All Group	6	48.53	1.40	44.30-52.70	2.89	6	66.68	0.82	64.20-69.20	1.23	6	75.80	4.05	63.60-87.90	5.34	6	66.00	1.67	61.00-71.00	2.53	6	49.45	2.46	42.10-56.90	4.97
Eosinophil	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn Ruby	5	3.78	0.35	2.80-4.80	9.21	5	4.20	0.30	3.30-5.10	7.14	5	4.38	0.22	3.70-5.10	5.00	5	4.34	0.15	3.80-4.80	3.49	5	3.58	0.52	2.00-5.20	14.44
AG: All Group	6	3.60	0.54	2.00-5.20	15.00	6	3.99	0.60	2.20-5.80	15.00	6	3.97	1.03	0.90-7.10	25.75	6	4.21	0.33	3.20-5.20	7.86	6	3.42	0.62	1.50-5.30	18.24
Basophil	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn Ruby	5	0.77	0.45	0.00-2.12	58.44	5	0.31	0.22	0.00-0.97	70.65	5	0.81	0.31	0.00-1.75	38.64	5	0.26	0.12	0.00-0.63	46.92	5	0.56	0.15	0.11-1.01	26.61
AG: All Group	6	0.66	0.49	0.00-2.12	73.64	6	0.29	0.20	0.00-0.89	69.31	6	0.71	0.37	0.00-1.83	52.68	6	0.25	0.11	0.00-0.59	44.80	6	0.48	0.23	0.00-1.17	47.92
Red Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn Ruby	5	2.13	0.04	2.00-2.26	2.07	5	4.67	0.12	4.39-4.95	2.61	5	5.16	0.05	4.85-5.47	1.03	5	4.63	0.05	4.35-4.91	0.99	5	2.10	0.04	1.97-2.23	2.05
AG: All Group	6	2.15	0.06	2.02-2.28	2.60	6	4.67	0.11	4.39-4.95	2.36	6	5.16	0.05	4.85-5.47	0.93	6	4.63	0.04	4.35-4.91	0.91	6	2.12	0.06	1.99-2.25	2.59
Hemoglobin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn Ruby	5	5.53	0.14	5.10-5.90	2.73	5	13.78	0.29	12.80-14.80	2.10	5	16.80	0.17	15.60-18.00	1.01	5	13.70	0.16	12.70-14.70	1.17	5	5.56	0.15	5.20-6.00	2.68
AG: All Group	6	5.59	0.20	5.20-6.00	3.57	6	13.68	0.35	12.70-14.70	2.55	6	16.65	0.40	15.50-17.90	2.40	6	13.58	0.32	12.60-14.60	2.35	6	5.60	0.17	5.20-6.00	3.04
Hematocrit	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn Ruby	5	15.34	0.57	14.40-16.20	3.73	5	38.24	1.28	35.90-40.50	3.35	5	44.50	0.83	41.80-47.20	1.87	5	37.90	0.52	35.60-40.20	1.37	5	15.10	0.35	14.20-16.00	2.32
PG: Siemens Advia 560	5	18.50	0.00	17.40-19.60	0.00	5	43.60	0.00	41.00-46.20	0.00	5	53.20	0.00	50.00-56.40	0.00	5	43.70	0.00	41.10-46.30	0.00	5	18.30	0.00	17.20-19.40	0.00
MCV	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn Ruby	5	71.92	1.39	67.70-76.10	1.93	5	81.88	1.89	76.20-87.60	2.31	5	86.22	2.21	79.60-92.80	2.56	5	81.86	1.80	76.50-87.30	2.20	5	71.84	1.35	67.80-75.80	1.88
AG: All Group	6	73.78	4.73	59.60-88.00	6.41	6	83.87	5.15	68.50-99.40	6.14	6	89.05	7.21	67.40-110.60	8.10	6	84.00	5.48	67.60-100.40	6.52	6	73.73	4.79	59.30-88.10	6.50
RDW	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn Ruby	5	13.12	0.74	10.90-15.30	5.65	5	12.10	0.60	10.30-13.90	4.96	5	10.78	0.48	9.40-12.20	4.44	5	12.04	0.57	10.30-13.70	4.75	5	13.00	0.62	11.10-14.90	4.77
AG: All Group	6	13.92	2.06	7.70-20.10	14.82	6	12.70	1.56	8.00-17.40	12.28	6	11.52	1.85	5.90-17.10	16.09	6	12.68	1.66	7.70-17.70	13.07	6	13.70	1.80	8.30-19.10	13.14
Platelet Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn Ruby	5	80.84	7.02	61.00-101.00	8.64	5	300.20	12.87	225.00-375.00	4.30	5	557.20	31.41	418.00-696.00	5.64	5	305.60	12.97	230.00-383.00	4.25	5	79.80	5.07	60.00-100.00	6.38
AG: All Group	6	80.53	6.32	61.00-101.00	7.78	6	297.17	13.70	223.00-371.00	4.61	6	548.17	35.76	411.00-685.00	6.53	6	303.83	12.38	228.00-380.00	4.08	6	79.83	4.54	60.00-100.00	5.63
683 - ASO/CRP:684 - CRP (Quant & Qual)	CR-3N					CR-6N																			
CRP Quantitative	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
600 - Cholestech LDX/CardioChek Pkg	CT-3																								
Cholesterol Total (Waived)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Cholestech LDX (Waived)	101	179.93	18.21	125.00-235.00	10.11																				
PG: Polymer CardioChek	68	197.34	19.08	140.00-254.00	9.70																				
Glucose (Waived)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Cholestech LDX (Waived)	22	186.14	13.96	144.00-228.00	7.53																				
AG: All Group	23	187.70	15.56	141.00-235.00	8.30																				
Triglycerides (Waived)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Cholestech LDX (Waived)	90	115.56	10.85	83.00-149.00	9.40																				
PG: Polymer CardioChek	63	136.27	18.08	82.00-190.00	13.31																				
Cholesterol HDL (Waived)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Cholestech LDX (Waived)	94	92.34	11.20	58.00-126.00	12.17																				
PG: Polymer CardioChek / Direct Method	68	105.15	10.36	74.00-136.00	9.90																				
Cholesterol LDL (Waived)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Cholestech LDX (Waived)	40	64.08	10.55	33.00-96.00	16.41																				
PG: Polymer CardioChek / Direct Method	9	53.91	10.59	22.00-86.00	19.63																				
763 - CoaguChek XS Plus/XS Pro (2 specimens);764 - CoaguChek XS Plus/XS Pro (5 specimens)	CXP-11					CXP-12					CXP-13					CXP-14					CXP-15				
Protime - Whole Blood	Labs																								



PG: CoaguChek XS Plus	112	1.07	0.05	1.00-1.30	4.55	112	1.83	0.12	1.40-2.20	6.67	59	2.21	0.13	1.80-2.60	5.91	59	1.82	0.12	1.40-2.20	6.67	59	1.05	0.05	1.00-1.30	4.55	
PG: CoaguChek XS Pro	13	1.05	0.05	1.00-1.30	4.55	13	1.85	0.05	1.60-2.00	2.78	11	2.24	0.09	1.90-2.50	4.09	11	1.84	0.09	1.50-2.10	5.00	11	1.05	0.05	1.00-1.30	4.55	
745 - Auto Diff 5 (Beckman Coulter 5-Part Diff)																										
DIF-11					DIF-12					DIF-13					DIF-14					DIF-15						
White Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
SG: Select Beckman Coulter Instruments (5-part Diff)	6	3.88	0.15	3.30-4.50	3.85	6	10.35	0.27	8.80-12.00	2.60	6	22.98	0.75	19.60-26.40	3.26	6	10.48	0.19	8.90-12.10	1.81	6	3.87	0.08	3.30-4.50	2.05	
Lymphocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
SG: Select Beckman Coulter Instruments (5-part Diff)	6	34.80	0.64	32.90-36.70	1.84	6	30.00	2.01	24.00-36.00	6.70	6	18.67	1.09	15.40-22.00	5.83	6	28.92	2.86	20.30-37.50	9.90	6	34.83	0.94	32.00-37.60	2.70	
Monocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
SG: Select Beckman Coulter Instruments (5-part Diff)	6	2.43	0.37	1.33-3.53	15.10	6	4.17	2.26	0.00-10.95	54.17	6	8.40	1.45	4.04-12.76	17.29	6	5.30	1.77	0.00-10.62	33.45	6	2.25	0.32	1.29-3.21	14.27	
Neutrophil	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
SG: Select Beckman Coulter Instruments (5-part Diff)	6	53.67	0.79	51.30-56.10	1.47	6	58.87	2.95	50.00-67.80	5.01	6	67.27	2.13	60.90-73.70	3.16	6	58.75	2.20	52.20-65.40	3.74	6	53.70	0.81	51.30-56.10	1.51	
Eosinophil	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
SG: Select Beckman Coulter Instruments (5-part Diff)	6	8.53	0.92	5.70-11.30	10.82	6	6.83	0.32	5.80-7.80	4.71	6	5.60	0.36	4.50-6.70	6.43	6	6.97	0.76	4.70-9.30	10.86	6	8.60	0.34	7.60-9.60	3.84	
Red Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
SG: Select Beckman Coulter Instruments (5-part Diff)	6	2.67	0.04	2.51-2.83	1.46	6	4.16	0.11	3.91-4.41	2.64	6	5.46	0.14	5.13-5.79	2.64	6	4.14	0.11	3.89-4.39	2.56	6	2.64	0.04	2.48-2.80	1.59	
Hemoglobin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
SG: Select Beckman Coulter Instruments (5-part Diff)	6	6.63	0.10	6.10-7.10	1.52	6	11.63	0.16	10.80-12.40	1.38	6	16.97	0.31	15.80-18.20	1.82	6	11.67	0.23	10.90-12.50	1.97	6	6.57	0.08	6.10-7.10	1.21	
Hematocrit	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
SG: Select Beckman Coulter Instruments (5-part Diff)	6	20.38	0.40	19.20-21.60	1.96	6	35.90	0.98	33.70-38.10	2.73	6	53.22	1.15	50.00-56.40	2.16	6	35.80	0.92	33.70-37.90	2.57	6	20.10	0.32	18.90-21.30	1.59	
Platelet Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
SG: Select Beckman Coulter Instruments (5-part Diff)	6	86.50	2.59	65.00-109.00	2.99	6	270.00	6.07	203.00-338.00	2.26	6	449.50	22.28	338.00-563.00	4.96	6	259.00	11.80	194.00-324.00	4.56	6	85.17	5.49	64.00-106.00	6.47	
MCV	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
SG: Select Beckman Coulter Instruments (5-part Diff)	6	76.30	0.69	74.20-78.40	0.90	6	86.22	0.90	83.50-88.90	1.04	6	97.48	0.94	94.70-100.30	0.96	6	86.42	0.84	83.90-88.90	0.97	6	76.37	0.68	74.40-78.40	0.89	
RDW	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
SG: Select Beckman Coulter Instruments (5-part Diff)	6	15.60	0.20	15.00-16.20	1.28	6	14.32	0.26	13.50-15.10	1.82	6	12.97	0.23	12.30-13.70	1.77	6	14.15	0.18	13.70-14.70	1.27	6	15.73	0.23	15.00-16.40	1.46	
730 - Sed Rate (ESR)																										
ES-3																										
Sedimentation Rate	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: Clinical Data (Vital Diag) Excyte M/10	17	73.29	8.56	47.00-99.00	11.78																					
PG: Modified Westergren - Plastic	14	53.93	8.91	27.00-81.00	16.48																					
PG: Streck ESR-Auto Plus	8	79.88	7.26	58.00-102.00	9.13																					
PG: Westergren - Glass	11	63.82	10.94	31.00-97.00	17.03																					
PG: Westergren - Plastic	23	47.83	11.13	15.00-81.00	23.13																					
PG: Polymedco Sediplast Westergren ESR	13	45.77	9.48	18.00-75.00	20.65																					
PG: Hypoguard (ARKRAY) Dispette 2 (Labsco)	14	51.29	11.36	17.00-85.00	22.35																					
PG: Sedi-Rate (Globe Scientific)	5	55.00	10.03	25.00-85.00	18.18																					
PG: Streck ESR-10 Manual	16	64.13	5.19	48.00-80.00	8.13																					
PG: Alcor Scientific iSED	5	1.00	0.00	1.00-1.00	0.00																					
AG: All Group	130	58.38	13.93	16.00-100.00	23.97																					
672 - Glycohemoglobin																										
GH-5																										
Glycohemoglobin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: Bayer DCA 2000 series	24	9.82	0.38	8.70-10.90	3.88																					
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	15	10.83	0.26	10.00-11.60	2.41																					
PG: Chek Diagnostics (Bayer) A1C Now	46	8.07	1.20	4.50-11.70	14.81																					
PG: Roche Integra A1C	5	11.04	0.23	10.30-11.70	2.09																					
SG: Select Beckman Coulter Instruments (Chem)	5	11.04	0.56	9.30-12.70	5.09																					
SG: Select Roche Instruments	7	11.20	0.77	8.90-13.50	6.88																					
PG: Siemens (Bayer) DCA Vantage	112	10.01	0.38	8.90-11.10	3.80																					
PG: Bayer A1CNow+	32	7.65	1.04	4.60-10.80	13.51																					
PG: TOSOH G8	6	8.80	0.30	7.90-9.70	3.41																					
SG: TOSOH G Series GHB	6	8.80	0.30	7.90-9.70	3.41																					
SG: Select Abbott Architect Instruments	5	10.96	1.30	7.10-14.90	11.82																					
PG: PTS A1CNow	8	8.24	1.40	4.00-12.40	17.07																					
AG: All Group	266	9.40	1.33	5.40-13.40	14.15																					
761 - HS CRP																										
HCR-3																										
hsCRP	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
AG: All Group	7	0.08	0.02	0.02-0.14	25.00																					
602 - Hemoglobin and Hematocrit - Waived; 614 - Auto Diff II (All Other 3-part); 610 - Regular Hematology; 601 - Waived/PPM Tests II; 609 - Waived/PPM Tests I; 749 - Auto Diff II w-Blood Cell Identification																										
HD-11					HD-12					HD-13					HD-14					HD-15						
Granulocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	
PG: Abbott Cell-Dyn 1700	12	61.66	4.05	49.60-73.80	6.56	12	82.20	0.89	79.50-84.90	1.08	12	33.97	2.42	26.70-41.30	7.12	12	76.13	1.27	72.30-79.90	1.67	12	56.78	1.50	52.30-61.30	2.64	
PG: Abbott Cell-Dyn 1800	32	63.24	1.13	59.80-66.60	1.79	32	82.28																			



Instrument	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %
PG: Drew Scientific D3	24	63.60	1.40	59.40-67.80	2.20	24	82.10	2.41	74.90-89.30	2.94	24	34.06	4.28	21.30-46.90	12.55	24	77.03	1.06	73.80-80.20	1.38	24	57.52	3.36	47.40-67.60	5.84
PG: Abbott Cell-Dyn Emerald	130	63.88	7.47	41.50-86.30	11.69	130	82.50	6.11	64.20-100.80	7.41	131	34.87	3.27	25.10-44.70	9.37	131	76.62	8.35	51.50-101.60	10.90	131	57.66	5.32	41.80-73.60	9.20
PG: CDS M Series	140	60.99	7.01	40.00-82.00	11.49	140	79.99	1.03	76.90-83.10	1.29	140	30.74	3.70	19.60-41.80	12.05	140	74.77	1.38	70.70-78.90	1.84	140	54.50	6.63	34.60-74.40	12.17
SG: CDS Hematology Instruments	150	60.95	6.80	40.50-81.30	11.17	150	79.91	1.13	76.50-83.30	1.41	150	30.81	3.66	19.80-41.80	11.88	149	74.70	1.44	70.40-79.00	1.93	150	54.45	6.43	35.20-73.80	11.80
SG: International Abbott Emerald Instruments	131	63.84	7.45	41.40-86.20	11.68	131	82.42	6.16	63.90-100.90	7.48	132	34.85	3.27	25.00-44.60	9.40	132	76.53	8.39	51.30-101.70	10.97	132	57.65	5.30	41.70-73.50	9.20
AG: All Group	628	62.79	5.09	47.50-78.10	8.11	627	81.52	4.14	69.10-93.90	5.08	629	33.07	3.28	23.30-42.90	9.91	626	75.77	5.66	58.80-92.80	7.47	629	56.25	5.04	41.20-71.40	8.95
Hematocrit	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %
PG: Abbott Cell-Dyn 1700	13	39.97	1.04	37.60-42.40	2.60	13	53.20	1.54	50.00-56.40	2.89	13	17.47	0.56	16.40-18.60	3.20	13	46.94	1.56	44.10-49.70	3.33	13	28.58	0.87	26.90-30.30	3.04
PG: Abbott Cell-Dyn 1800	32	40.34	1.51	37.90-42.70	3.75	32	52.61	2.24	49.40-55.80	4.26	32	17.71	0.80	16.60-18.80	4.52	32	46.52	1.78	43.70-49.30	3.83	32	28.55	1.01	26.90-30.30	3.53
PG: CDS Medonic 620	12	37.14	1.20	34.90-39.30	3.23	10	50.83	1.99	47.80-53.80	3.92	12	15.67	0.54	14.80-16.60	3.44	12	44.14	1.57	41.50-46.70	3.56	12	25.58	0.79	24.10-27.10	3.09
PG: Beckman Coulter AcT Diff/Diff 2	214	39.54	1.14	37.10-41.90	2.89	214	52.77	2.28	49.60-56.00	4.32	214	17.14	0.50	16.10-18.10	2.92	213	45.79	2.18	43.10-48.50	4.76	211	27.79	0.77	26.10-29.50	2.77
PG: Horiba ABX Micros/45/60	74	37.63	0.72	35.30-39.90	1.91	74	50.14	1.30	47.10-53.10	2.59	74	15.44	0.34	14.50-16.30	2.21	74	43.95	0.86	41.40-46.60	1.95	74	25.67	0.55	24.20-27.20	2.14
SG: Select Abbott Cell Dyn Instruments (3-part Diff)	177	40.28	1.39	37.90-42.70	3.45	177	52.83	1.91	49.60-56.00	3.62	176	17.45	0.62	16.40-18.60	3.54	178	46.61	1.73	43.80-49.40	3.71	178	28.39	1.82	26.70-30.10	6.41
SG: Select Beckman Coulter Instruments (3-part Diff)	214	39.54	1.14	37.10-41.90	2.89	214	52.77	2.28	49.60-56.00	4.32	214	17.14	0.50	16.10-18.10	2.92	213	45.79	2.18	43.10-48.50	4.76	211	27.79	0.77	26.10-29.50	2.77
SG: Select Drew Scientific Instruments	23	39.59	1.17	37.20-42.00	2.95	24	53.04	3.13	49.80-56.20	5.89	23	17.30	0.49	16.30-18.30	2.83	24	46.82	3.61	44.00-49.60	7.71	22	28.05	0.82	26.30-29.70	2.93
PG: Drew Scientific D3	23	39.59	1.17	37.20-42.00	2.95	24	53.04	3.13	49.80-56.20	5.89	23	17.30	0.49	16.30-18.30	2.83	24	46.82	3.61	44.00-49.60	7.71	22	28.05	0.82	26.30-29.70	2.93
PG: Abbott Cell-Dyn Emerald	131	40.30	1.41	37.90-42.70	3.50	131	52.85	1.87	49.70-56.10	3.53	130	17.39	0.57	16.40-18.40	3.28	132	46.59	1.75	43.80-49.40	3.76	132	28.33	2.03	26.60-30.00	7.17
PG: CDS M Series	141	37.59	1.01	35.30-39.90	2.69	141	51.88	1.85	48.80-55.00	3.56	141	15.99	0.67	15.00-17.00	4.19	141	44.76	1.36	42.10-47.50	3.04	141	25.86	0.49	24.30-27.50	1.89
SG: CDS Hematology Instruments	153	37.55	1.02	35.30-39.90	2.71	151	51.82	1.87	48.70-54.90	3.61	153	15.97	0.66	15.00-17.00	4.13	153	44.71	1.38	42.00-47.40	3.09	153	25.84	0.53	24.30-27.30	2.05
SG: International Abbott Emerald Instruments	132	40.29	1.40	37.90-42.70	3.47	132	52.84	1.87	49.60-56.00	3.54	131	17.38	0.57	16.40-18.40	3.28	133	46.58	1.75	43.80-49.40	3.76	133	28.32	2.02	26.60-30.00	7.14
Hemoglobin	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %
PG: Abbott Cell-Dyn 1700	13	13.88	0.17	12.90-14.90	1.22	13	18.58	0.30	17.30-19.90	1.61	13	6.14	0.18	5.70-6.50	2.95	13	16.38	0.26	15.30-17.50	1.59	13	9.95	0.19	9.30-10.70	1.90
PG: Abbott Cell-Dyn 1800	32	13.90	0.36	12.90-14.90	2.59	32	18.66	0.54	17.40-20.00	2.89	32	6.03	0.20	5.60-6.40	3.33	32	16.30	0.39	15.20-17.40	2.39	32	9.84	0.24	9.10-10.50	2.45
PG: CDS Medonic 620	12	13.52	0.29	12.60-14.40	2.15	12	18.13	0.56	16.80-19.40	3.09	12	5.74	0.14	5.30-6.10	2.46	12	15.88	0.37	14.80-17.00	2.33	12	9.53	0.19	8.80-10.20	2.00
PG: Beckman Coulter AcT Diff/Diff 2	214	13.61	0.33	12.60-14.60	2.43	214	18.41	0.73	17.10-19.70	3.97	214	5.78	0.14	5.40-6.20	2.24	213	15.98	0.72	14.90-17.10	4.50	211	9.59	0.23	8.90-10.30	2.40
PG: Horiba ABX Micros/45/60	73	13.57	0.24	12.60-14.60	1.76	73	17.98	0.38	16.70-19.30	2.11	73	5.73	0.21	5.30-6.10	3.68	73	15.87	0.23	14.80-17.00	1.45	73	9.51	0.19	8.80-10.20	2.00
SG: Select Abbott Cell Dyn Instruments (3-part Diff)	177	13.69	0.37	12.70-14.70	2.70	177	18.37	0.48	17.10-19.70	2.61	178	5.81	0.22	5.40-6.20	3.79	178	16.09	0.52	15.00-17.20	3.23	178	9.68	0.60	9.00-10.40	6.19
SG: Select Drew Scientific Instruments	23	13.50	0.39	12.60-14.40	2.89	24	18.24	1.17	16.90-19.50	6.43	23	5.73	0.19	5.30-6.10	3.33	23	15.82	0.33	14.70-16.90	2.09	22	9.54	0.24	8.80-10.20	2.53
PG: Drew Scientific D3	23	13.50	0.39	12.60-14.40	2.89	24	18.24	1.17	16.90-19.50	6.43	23	5.73	0.19	5.30-6.10	3.33	23	15.82	0.33	14.70-16.90	2.09	22	9.54	0.24	8.80-10.20	2.53
PG: Abbott Cell-Dyn Emerald	131	13.61	0.36	12.60-14.60	2.65	131	18.27	0.44	17.00-19.60	2.40	132	5.71	0.14	5.30-6.10	2.46	132	16.00	0.53	14.90-17.10	3.31	132	9.62	0.67	8.90-10.30	6.98
PG: CDS M Series	141	13.63	0.28	12.60-14.60	2.06	141	18.40	0.53	17.10-19.70	2.88	141	5.84	0.22	5.40-6.20	3.62	141	16.06	0.35	15.00-17.20	2.17	141	9.58	0.14	8.90-10.30	1.46
SG: CDS Hematology Instruments	153	13.62	0.28	12.60-14.60	2.06	153	18.38	0.54	17.10-19.70	2.93	153	5.83	0.21	5.40-6.20	3.62	153	16.05	0.36	14.90-17.10	2.25	153	9.58	0.14	8.90-10.30	1.46
SG: International Abbott Emerald Instruments	132	13.61	0.36	12.60-14.60	2.65	132	18.27	0.44	17.00-19.60	2.40	133	5.71	0.14	5.30-6.10	2.46	133	16.00	0.53	14.90-17.10	3.31	133	9.61	0.67	8.90-10.30	6.98
Lymphocyte	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %
PG: Abbott Cell-Dyn 1700	10	28.44	1.32	24.48-32.40	4.64	12	12.24	0.97	9.32-15.16	7.95	11	52.45	3.37	42.33-62.57	6.43	11	17.35	1.10	14.04-20.66	6.35	11	33.51	1.95	27.67-39.35	5.81
PG: Abbott Cell-Dyn 1800	32	25.11	1.63	20.23-29.99	6.48	32	11.11	0.67	9.11-13.11	6.00	32	47.24	3.26	37.45-57.03	6.91	32	15.38	0.85	12.83-17.93	5.53	32	30.31	1.96	24.42-36.20	6.48
PG: CDS Medonic 620	10	31.42	1.15	27.99-34.86	3.64	10	14.75	0.49	13.29-16.21	3.29	10	58.72	2.67	50.70-66.74	4.55	10	19.36	0.88	16.72-22.00	4.55	10	38.06	1.23	34.39-41.73	3.22
PG: Beckman Coulter AcT Diff/Diff 2	205	32.28	0.88	29.63-34.93	2.74	204	14.28	0.62	12.41-16.15	4.36	205	60.27	4.45	46.93-73.61	7.38	201	19.53	0.67	17.51-21.55	3.45	205	38.66	2.24	31.94-45.38	5.80
PG: Horiba ABX Micros/45/60	69	24.26	1.51	19.74-28.78	6.21	73	11.50	0.81	9.08-13.92	7.02	70	43.22	3.08	33.99-52.45	7.12	73	15.61	1.18	12.06-19.16	7.58	71	28.31	2.01	22.28-34.34	7.10
SG: Select Abbott Cell Dyn Instruments (3-part Diff)	171	28.54	2.08	22.30-34.78	7.29</																				



Instrument	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
SG: Select Drew Scientific Instruments	23	5.89	0.62	4.00-7.80	10.51	23	3.23	0.20	2.60-3.80	6.25	23	10.33	1.14	6.90-13.70	11.07	23	3.97	0.32	3.00-5.00	8.00	23	6.82	0.78	4.50-9.10	11.47
PG: Drew Scientific D3	23	5.89	0.62	4.00-7.80	10.51	23	3.23	0.20	2.60-3.80	6.25	23	10.33	1.14	6.90-13.70	11.07	23	3.97	0.32	3.00-5.00	8.00	23	6.82	0.78	4.50-9.10	11.47
PG: Abbott Cell-Dyn Emerald	130	5.75	0.77	3.50-8.10	13.28	130	2.99	0.35	2.00-4.00	11.67	131	12.20	1.74	7.00-17.40	14.26	131	3.76	0.57	2.10-5.50	15.00	131	7.14	1.28	3.30-10.90	18.03
PG: CDS M Series	140	7.25	1.20	3.60-10.80	16.67	140	5.45	0.75	3.30-7.80	13.64	139	9.51	2.05	3.40-15.60	21.58	140	6.11	0.94	3.30-8.90	15.57	140	7.44	1.40	3.20-11.60	18.92
SG: International Abbott Emerald Instruments	131	5.76	0.77	3.50-8.10	13.28	131	2.99	0.35	2.00-4.00	11.67	132	12.19	1.73	7.00-17.40	14.18	132	3.78	0.59	2.00-5.60	15.53	132	7.15	1.28	3.30-10.90	18.03
Platelet Count	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Abbott Cell-Dyn 1700	13	262.15	21.00	197.00-328.00	8.02	13	601.77	39.29	452.00-753.00	6.53	13	60.23	4.17	45.00-75.00	7.00	13	436.92	32.55	328.00-546.00	7.44	13	158.92	14.79	119.00-199.00	9.31
PG: Abbott Cell-Dyn 1800	32	266.63	12.63	200.00-334.00	4.72	32	595.22	31.62	446.00-744.00	5.31	32	64.28	4.65	48.00-80.00	7.34	32	431.91	20.74	324.00-540.00	4.79	32	159.47	8.46	119.00-199.00	5.35
PG: CDS Medonic 620	12	238.00	13.28	179.00-298.00	5.59	12	508.08	26.17	381.00-635.00	5.16	12	59.75	5.43	45.00-75.00	9.00	12	375.25	21.30	281.00-469.00	5.68	12	147.75	9.50	111.00-185.00	6.42
PG: Beckman Coulter AcT Diff/Diff 2	214	261.18	15.63	196.00-326.00	5.98	214	571.90	29.02	429.00-715.00	5.07	214	66.00	6.66	50.00-83.00	10.15	213	420.15	33.26	315.00-525.00	7.93	214	163.75	28.15	123.00-205.00	17.13
PG: Horiba ABX Micros/45/60	74	263.84	12.33	198.00-330.00	4.66	74	548.72	25.26	412.00-686.00	4.61	74	72.45	6.83	54.00-90.00	9.44	74	415.15	18.35	311.00-519.00	4.43	74	170.99	10.40	128.00-214.00	6.08
SG: Select Abbott Cell Dyn Instruments (3-part Diff)	177	263.74	17.05	198.00-330.00	6.44	177	558.66	56.36	419.00-699.00	10.09	178	70.56	13.24	53.00-89.00	18.59	178	420.01	26.64	315.00-525.00	6.33	178	173.55	22.64	131.00-218.00	12.99
SG: Select Beckman Coulter Instruments (3-part Diff)	214	261.18	15.63	196.00-326.00	5.98	214	571.90	29.02	429.00-715.00	5.07	214	66.00	6.66	50.00-83.00	10.15	213	420.15	33.26	315.00-525.00	7.93	214	163.75	28.15	123.00-205.00	17.13
SG: Select Drew Scientific Instruments	24	269.38	17.73	202.00-336.00	6.58	24	557.88	24.16	419.00-698.00	4.34	24	72.54	11.55	55.00-91.00	15.75	24	414.29	18.70	311.00-518.00	4.52	24	181.67	29.39	137.00-228.00	16.15
PG: Drew Scientific D3	24	269.38	17.73	202.00-336.00	6.58	24	557.88	24.16	419.00-698.00	4.34	24	72.54	11.55	55.00-91.00	15.75	24	414.29	18.70	311.00-518.00	4.52	24	181.67	29.39	137.00-228.00	16.15
PG: Abbott Cell-Dyn Emerald	131	263.23	17.66	197.00-329.00	6.73	131	545.19	56.58	409.00-681.00	10.39	132	73.16	14.22	55.00-91.00	19.45	132	415.29	25.86	311.00-519.00	6.24	132	178.52	23.61	134.00-224.00	13.18
PG: CDS M Series	141	231.23	10.62	173.00-289.00	4.59	141	500.35	20.58	375.00-625.00	4.12	141	59.62	8.10	45.00-75.00	13.50	141	371.04	14.77	278.00-464.00	3.99	141	143.50	8.67	108.00-180.00	6.04
SG: International Abbott Emerald Instruments	132	262.75	18.43	197.00-329.00	7.00	132	544.69	56.66	409.00-681.00	10.40	133	73.07	14.20	55.00-91.00	19.45	133	415.22	25.78	311.00-519.00	6.22	133	178.43	23.55	134.00-223.00	13.20
RDW	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Abbott Cell-Dyn 1700	11	17.39	0.90	14.70-20.10	5.11	11	15.22	0.79	12.80-17.60	5.20	11	17.14	0.53	15.50-18.70	3.10	11	16.54	0.74	14.30-18.70	4.48	11	18.30	0.49	16.80-19.80	2.68
PG: Abbott Cell-Dyn 1800	32	16.75	0.65	14.80-18.60	3.89	32	14.66	0.69	12.60-16.80	4.69	32	17.35	0.55	15.70-18.90	3.18	32	16.01	0.64	14.10-17.90	4.00	32	17.87	0.71	15.70-20.10	4.02
PG: CDS Medonic 620	10	15.00	0.82	12.50-17.50	5.47	10	13.95	0.66	12.00-16.00	4.71	10	13.79	0.65	11.90-15.80	4.71	10	14.78	0.84	12.30-17.30	5.68	10	14.63	0.70	12.50-16.70	4.79
PG: Beckman Coulter AcT Diff/Diff 2	204	15.19	2.58	7.50-22.90	16.97	204	13.38	0.32	12.40-14.40	2.39	204	15.72	0.32	14.70-16.70	2.04	203	14.52	0.37	13.40-15.60	2.55	204	16.27	0.67	14.30-18.30	4.11
PG: Horiba ABX Micros/45/60	63	13.69	0.88	11.10-16.30	6.42	63	12.42	0.72	10.20-14.60	5.81	63	13.75	0.76	11.40-16.00	5.55	63	13.37	0.83	10.90-15.90	6.19	63	14.46	0.90	11.80-17.20	6.21
SG: Select Abbott Cell Dyn Instruments (3-part Diff)	172	15.09	1.22	11.40-18.80	8.08	172	13.26	1.03	10.20-16.40	7.74	173	15.42	1.21	11.80-19.00	7.86	173	14.42	1.22	10.70-18.10	8.47	173	16.21	1.17	12.70-19.70	7.22
SG: Select Beckman Coulter Instruments (3-part Diff)	204	15.19	2.58	7.50-22.90	16.97	204	13.38	0.32	12.40-14.40	2.39	204	15.72	0.32	14.70-16.70	2.04	203	14.52	0.37	13.40-15.60	2.55	204	16.27	0.67	14.30-18.30	4.11
SG: Select Drew Scientific Instruments	24	14.90	0.91	12.20-17.60	6.11	24	13.13	0.77	10.80-15.40	5.88	24	14.87	0.72	12.70-17.10	4.83	24	14.34	0.81	11.90-16.70	5.66	24	15.83	0.84	13.30-18.30	5.32
PG: Drew Scientific D3	24	14.90	0.91	12.20-17.60	6.11	24	13.13	0.77	10.80-15.40	5.88	24	14.87	0.72	12.70-17.10	4.83	24	14.34	0.81	11.90-16.70	5.66	24	15.83	0.84	13.30-18.30	5.32
PG: Abbott Cell-Dyn Emerald	129	14.49	0.57	12.80-16.20	3.93	129	12.75	0.44	11.50-14.10	3.44	130	14.80	0.55	13.20-16.40	3.72	130	13.85	0.72	11.70-16.10	5.18	130	15.63	0.55	13.90-17.30	3.53
PG: CDS M Series	140	14.92	2.98	6.00-23.80	20.00	140	13.78	1.04	10.70-16.90	7.54	140	13.90	1.15	10.50-17.40	8.27	140	14.40	1.16	10.90-17.90	8.06	140	14.43	1.15	11.00-17.90	7.99
SG: CDS Hematology Instruments	150	14.92	2.88	6.30-23.50	19.33	150	13.79	1.02	10.70-16.90	7.39	150	13.90	1.13	10.50-17.30	8.13	150	14.42	1.14	11.00-17.80	7.92	150	14.44	1.12	11.00-17.80	7.78
SG: International Abbott Emerald Instruments	130	14.50	0.56	12.80-16.20	3.93	130	12.75	0.44	11.50-14.10	3.44	131	14.80	0.55	13.20-16.40	3.72	131	13.86	0.72	11.70-16.10	5.18	131	15.64	0.55	13.90-17.30	3.53
AG: All Group	613	14.93	2.23	8.20-21.60	14.90	613	13.34	0.90	10.60-16.00	6.77	614	14.95	1.23	11.30-18.70	8.20	613	14.34	0.99	11.30-17.30	6.92	614	15.61	1.29	11.70-19.50	8.27
Red Blood Cell Count	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Abbott Cell-Dyn 1700	13	4.71	0.08	4.43-4.99	1.78	13	5.70	0.10	5.36-6.04	1.77	13	2.32	0.04	2.18-2.46	1.90	13	5.25	0.14	4.93-5.57	2.61	13	3.51	0.07	3.30-3.72	1.97
PG: Abbott Cell-Dyn 1800	32	4.61	0.13	4.33-4.89	2.78	32	5.48	0.19	5.15-5.81	3.38	32	2.31	0.08	2.17-2.45	3.29	32	5.05	0.14	4.75-5.35	2.69	32	3.44	0.09	3.23-3.65	2.67
PG: CDS Medonic 620	12	4.72	0.12	4.44-5.00	2.52	12	5.78	0.16	5.43-6.13	2.82	12	2.23	0.05	2.10-2.36	2.15	12	5.28	0.08	4.96-5.60	1.53	12	3.43	0.07	3.22-3.64	2.10
PG: Beckman Coulter AcT Diff/Diff 2	214	4.75	0.13	4.46-5.04	2.67	214	5.82	0.25	5.47-6.17	4.21	213	2.30	0.06	2.16-2.44	2.57	210	5.27	0.13	4.95-5.59	2.54	212	3.49	0.21	3.28-3.70	6.05
PG: Horiba ABX Micros/45/60	74	4.68	0.08	4.40-4.96	1.73	74	5.69	0.13	5.35-6.03	2.34	74	2.22	0.04	2.09-2.35	1.94	74	5.21	0.10	4.90-5.52	1.90	74</				



SG: International Abbott i-Stat Instruments	9	1.15	0.03	1.05-1.25	2.87	9	1.47	0.03	1.38-1.56	2.11	8	3.58	0.07	3.36-3.80	2.01	8	0.94	0.03	0.84-1.04	3.51	8	1.14	0.03	1.04-1.24	2.81
Anion Gap	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: International Abbott i-Stat Instruments / Thermal	8	33.25	1.49	29.00-38.00	4.55																				
PG: Abbott i-Stat (Waived) / Thermal	8	33.25	1.49	29.00-38.00	4.55																				
SG: Select Abbott i-Stat Instruments / Thermal	8	33.25	1.49	29.00-38.00	4.55																				
Creatinine	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat / Thermal	14	1.11	0.05	0.80-1.40	4.55	13	1.62	0.06	1.30-1.90	3.75	13	5.92	0.20	5.00-6.80	3.39	13	0.75	0.05	0.50-1.10	6.25	13	1.10	0.00	0.80-1.40	0.00
SG: International Abbott i-Stat Instruments / Thermal	45	1.12	0.05	0.80-1.40	4.55	18	1.64	0.06	1.30-1.90	3.75	18	5.96	0.18	5.10-6.90	3.00	18	0.76	0.05	0.50-1.10	6.25	18	1.11	0.02	0.80-1.40	1.82
PG: Abbott i-Stat (Waived) / Thermal	26	1.13	0.05	0.80-1.40	4.55																				
716 - Blood Lead - Waived Methods;709 - Blood Lead	LE-11					LE-12					LE-13					LE-14					LE-15				
Blood Lead	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select ESA LeadCare Instruments	6	13.63	2.42	9.60-17.60	17.87	5	38.00	2.64	34.00-42.00	6.95	6	4.60	1.05	0.60-8.60	22.83	6	14.77	1.39	10.80-18.80	9.39	6	39.70	3.12	35.70-43.70	7.86
Blood Lead (Waived)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: ESA LeadCare II (Waived)	37	15.29	1.65	10.40-20.20	10.72																				
SG: Select ESA LeadCare Instruments	37	15.29	1.65	10.40-20.20	10.72																				
662 - Neonatal Bilirubin (5 specimens);663 - Neonatal Bilirubin (2 specimens)	NB-11					NB-12					NB-13					NB-14					NB-15				
Neonatal Bilirubin (Total)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Reichert (Leica) Unistat Bilirubinometer	11	5.39	0.20	4.30-6.50	3.70	11	16.04	0.34	12.80-19.20	2.13	5	0.10	0.00	0.00-0.50	0.00	11	18.72	0.56	15.00-22.40	2.99	11	11.26	0.34	9.00-13.60	3.01
SG: Select Neonatal Bilirubin Instruments	19	5.53	0.34	4.40-6.60	6.18	19	15.41	0.89	12.30-18.50	5.78	7	0.27	0.13	0.00-0.70	43.33	17	18.27	0.91	14.60-22.00	4.97	17	11.09	0.52	8.90-13.30	4.68
AG: All Group	20	5.55	0.34	4.40-6.60	6.18	20	15.40	0.87	12.30-18.50	5.65	7	0.27	0.13	0.00-0.70	43.33	17	18.27	0.91	14.60-22.00	4.97	17	11.09	0.52	8.90-13.30	4.68
748 - i-Stat Prottime	PTI-11					PTI-12					PTI-13					PTI-14					PTI-15				
Prottime - Whole Blood	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat Prottime	6	14.27	0.44	12.20-16.40	3.08	6	26.23	1.43	22.30-30.10	5.46	6	13.90	0.41	11.80-16.00	2.88	6	13.73	0.75	11.60-15.80	5.47	7	26.56	1.97	22.60-30.60	7.41
PG: Abbott CoagSense	5	13.58	1.63	11.60-15.60	11.99	5	35.60	3.06	30.30-40.90	8.60	5	13.36	1.46	11.40-15.40	10.90	5	13.05	0.24	11.10-15.10	1.83	6	34.58	3.61	29.40-39.80	10.43
SG: International Abbott i-Stat Instruments	7	14.19	0.45	12.10-16.30	3.17	7	26.30	1.31	22.40-30.20	5.02	7	13.90	0.37	11.80-16.00	2.66	6	13.73	0.75	11.60-15.80	5.47	8	26.45	1.85	22.50-30.50	6.98
SG: International Abbott CoagSense Instruments	6	13.67	1.47	11.60-15.80	10.73	6	35.07	3.03	29.80-40.40	8.63	6	13.65	1.48	11.60-15.80	10.80	5	13.06	0.21	11.10-15.10	1.60	7	35.04	3.51	29.80-40.30	10.03
INR	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat Prottime	7	1.20	0.06	1.00-1.40	5.00	7	2.26	0.11	2.00-2.60	4.78	7	1.17	0.05	1.00-1.40	4.17	7	1.14	0.08	0.90-1.30	7.27	8	2.30	0.19	1.70-2.90	8.26
PG: Abbott CoagSense	6	1.25	0.22	0.60-2.00	16.92	6	3.48	0.66	1.50-5.50	18.86	6	1.10	0.09	0.80-1.40	8.18	6	1.37	0.30	0.50-2.30	21.43	6	3.12	0.46	1.70-4.50	14.84
SG: International Abbott i-Stat Instruments	12	1.18	0.06	1.00-1.40	5.00	12	2.27	0.09	2.00-2.60	3.91	12	1.17	0.05	1.00-1.40	4.17	11	1.16	0.07	1.00-1.40	5.83	13	2.28	0.16	1.80-2.80	6.96
SG: International Abbott CoagSense Instruments	7	1.24	0.20	0.60-1.80	16.67	7	3.37	0.67	1.40-5.40	19.71	7	1.11	0.09	0.80-1.40	8.18	7	1.33	0.29	0.40-2.20	22.31	7	3.11	0.42	1.80-4.40	13.55
611 - QBC Hematology	QBC-11					QBC-12					QBC-13					QBC-14					QBC-15				
Hemoglobin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: QBC Star - QBC Star Tubes	8	11.01	0.53	10.20-11.80	4.82	8	9.38	0.27	8.70-10.10	2.87	8	11.66	0.26	10.90-12.50	2.22	7	7.29	0.21	6.80-7.80	2.88	8	11.71	0.37	10.90-12.50	3.16
AG: All Group	14	10.75	0.56	10.00-11.60	5.19	14	9.36	0.29	8.70-10.10	3.19	14	11.46	0.51	10.70-12.30	4.43	12	7.36	0.28	6.90-7.90	3.78	14	11.52	0.49	10.70-12.30	4.26
Hematocrit	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: QBC Star - QBC Star Tubes	8	33.26	1.48	31.30-35.30	4.44	8	28.39	0.97	26.70-30.10	3.42	8	35.09	0.85	33.00-37.20	2.42	8	22.25	0.83	21.00-23.60	3.72	8	35.36	1.11	33.30-37.50	3.14
AG: All Group	14	32.71	1.39	30.70-34.70	4.25	14	28.27	1.04	26.60-30.00	3.67	14	34.80	0.97	32.70-36.90	2.79	13	22.37	0.80	21.10-23.70	3.57	14	35.04	1.08	32.90-37.10	3.06
White Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: QBC Autoread - Accutubes	5	23.00	0.00	19.60-26.40	0.00	5	10.80	0.00	9.20-12.40	0.00	5	19.50	0.00	16.60-22.40	0.00	5	4.20	0.00	3.60-4.80	0.00	5	20.50	0.00	17.40-23.60	0.00
PG: QBC Star - QBC Star Tubes	7	23.30	2.82	19.80-26.80	12.10	8	11.59	0.80	9.90-13.30	6.90	6	24.45	1.12	20.80-28.20	4.57	6	6.45	0.68	5.50-7.50	10.46	8	22.85	1.98	19.50-26.30	8.65
SG: Select QBC - Cap and Ven	5	14.96	1.36	12.80-17.30	9.07	5	8.46	0.86	7.20-9.80	10.12	5	14.18	1.53	12.10-16.30	10.77	5	4.38	0.73	3.70-5.10	16.59	5	14.52	1.58	12.30-16.70	10.90
Granulocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: QBC Star - QBC Star Tubes	7	40.57	5.68	24.00-58.00	13.90	8	54.25	1.83	49.00-59.00	3.33	8	39.75	3.28	30.00-50.00	8.25	8	57.25	7.85	33.00-81.00	13.86	8	42.13	5.57	25.00-59.00	13.33
AG: All Group	13	41.77	5.02	27.00-57.00	11.90	14	54.79	4.64	41.00-69.00	8.36	14	40.79	3.49	31.00-52.00	8.54	13	56.23	7.63	33.00-79.00	13.57	14	42.57	5.34	27.00-59.00	12.33
Lymphocyte/Monocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: QBC Star - QBC Star Tubes	8	57.00	8.65	31.00-83.00	15.26	8	45.75	1.83	41.00-51.00	3.91	8	60.25	3.28	50.00-70.00	5.50	8	42.75	7.85	19.00-67.00	18.37	8	57.88	5.57	41.00-75.00	9.66
AG: All Group	14	56.93	6.86	36.00-78.00	12.11	14	45.21	4.64	31.00-59.00	10.22	14	59.21	3.49	49.00-70.00	5.93	13	43.77	7.63	21.00-67.00	17.27	14	57.43	5.34	41.00-73.00	9.30
Platelet Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: QBC Star - QBC Star Tubes	8	173.50	17.07	131.00-218.00	9.83	8	414.38	42.77	311.00-518.00	10.34	8	178.63	20.50	134.00-224.00	11.45	7	748.00	67.43	561.00-935.00	9.01	8	180.88	20.90	136.00-226.00	11.55
AG: All Group	14	163.07	23.74	122.00-204.00	14.54	14	395.00	52.70	296.00-494.00	13.34	11	176.18													



PG: Alfa Was. Ace/NexCT/Alera/Axcel / Other Reagent	5	122.38	6.01	104.40-140.40	4.91																					
PG: Beckman ACCESS	8	115.90	6.42	96.60-135.20	5.54																					
SG: Select Beckman Coulter Instruments (Chem)	12	119.27	10.94	86.50-152.10	9.17																					
SG: Select Siemens/Dade Instruments	5	154.08	10.19	123.50-184.70	6.61																					
SG: Select OCD Vitros Instruments	5	111.40	4.39	98.20-124.60	3.94																					
PG: Abbott Architect i1000 SR	5	185.91	4.93	171.10-200.70	2.65																					
SG: Select Abbott Architect Instruments	10	185.72	6.95	164.80-206.50	3.74																					
AG: All Group	38	133.18	25.50	56.70-209.70	19.14																					
Folate	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	
PG: Beckman ACCESS	8	4.66	0.32	3.70-5.70	6.81																					
PG: TOSOH AIA	5	3.27	0.64	1.40-5.20	19.39																					
SG: Select Beckman Coulter Instruments (Chem)	11	4.67	0.32	3.70-5.70	6.81																					
SG: Select Roche Instruments	5	2.86	0.47	1.50-4.30	16.21																					
PG: Abbott Architect i1000 SR	5	4.52	0.52	2.90-6.10	11.56																					
SG: Select Abbott Architect Instruments	12	4.50	0.46	3.10-5.90	10.22																					
AG: All Group	40	3.93	0.89	1.20-6.60	22.82																					
FSH	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	
PG: TOSOH AIA	6	33.27	3.65	22.30-44.30	10.96																					
SG: Select Beckman Coulter Instruments (Chem)	5	28.54	2.73	20.30-36.70	9.54																					
AG: All Group	22	29.58	3.36	19.50-39.70	11.35																					
LH	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	
PG: TOSOH AIA	6	34.47	1.88	28.90-40.10	5.45																					
PG: OCD Vitros ECi/ECIQ	5	52.53	3.80	41.10-63.90	7.24																					
AG: All Group	16	33.39	3.73	22.20-44.60	11.17																					
Progesterone	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	
PG: TOSOH AIA	5	32.10	3.63	21.20-43.00	11.31																					
SG: Select Beckman Coulter Instruments (Chem)	5	25.88	3.91	14.20-37.60	15.10																					
AG: All Group	16	27.62	4.50	14.10-41.10	16.30																					
Prolactin	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	
AG: All Group	13	36.97	5.55	20.40-53.60	15.00																					
PSA	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	
PG: Abbott Architect	5	14.63	0.39	13.45-15.81	2.69																					
PG: Beckman ACCESS	25	19.16	1.32	15.19-23.13	6.90																					
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	16	15.96	0.75	13.71-18.21	4.69																					
PG: Qualigen FastPack PSA	16	20.81	3.96	8.92-32.70	19.05																					
PG: TOSOH AIA	21	15.09	0.88	12.46-17.72	5.82																					
SG: Select Beckman Coulter Instruments (Chem)	28	19.15	1.25	15.40-22.90	6.52																					
SG: Select Siemens/Dade Instruments	18	16.20	1.09	12.94-19.46	6.72																					
SG: Select Roche Instruments	5	18.89	1.51	14.36-23.42	8.00																					
SG: Select OCD Vitros Instruments	6	10.52	0.19	9.94-11.10	1.84																					
PG: Roche Cobas e 411	5	21.29	4.23	8.60-33.98	19.86																					
PG: Abbott Architect i1000 SR	8	14.35	0.78	12.00-16.70	5.46																					
SG: Select Abbott Architect Instruments	16	14.44	0.60	12.65-16.23	4.13																					
PG: NanoEnTek FrenD System	5	21.95	2.59	14.19-29.71	11.79																					
AG: All Group	115	17.18	3.38	7.05-27.30	19.64																					
Testosterone	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	
PG: Beckman ACCESS	16	8.20	0.35	7.10-9.30	4.27																					
PG: Qualigen FastPak Testosterone	9	7.10	0.95	4.20-10.00	13.52																					
PG: TOSOH AIA	8	11.64	1.01	8.60-14.60	8.71																					
SG: Select Beckman Coulter Instruments (Chem)	18	8.16	0.36	7.10-9.30	4.39																					
SG: Select Abbott Architect Instruments	6	10.47	0.52	8.90-12.10	4.95																					
AG: All Group	55	9.41	1.85	3.80-15.00	19.68																					
Vitamin B12	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	
PG: Beckman ACCESS	15	685.07	47.07	544.00-826.00	6.88																					
PG: TOSOH AIA	5	864.00	51.20	710.00-1018.00	5.93																					
SG: Select Beckman Coulter Instruments (Chem)	18	679.50	45.82	543.00-817.00	6.74																					
SG: Select Siemens/Dade Instruments	5	717.00	122.20	350.00-1084.00	17.04																					
SG: Select Roche Instruments	6	841.90	41.83	717.00-967.00	4.96																					
SG: Select OCD Vitros Instruments	5	825.80	25.20	750.00-902.00	3.05																					
PG: Roche Cobas e 411	5	839.44	46.28	700.00-978.00	5.52																					
PG: Abbott Architect i1000 SR	5	889.80	85.84	633.00-1147.00	9.64																					
SG: Select Abbott Architect Instruments	12	835.08	80.12	595.00-1075.00	9.59																					



PTH	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	11	23.73	4.53	10.10-37.30	19.11	9	9.51	1.61	4.70-14.30	16.95															
Vitamin D	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman ACCESS	18	57.96	6.93	37.16-78.76	11.96	18	35.67	4.50	22.16-49.18	12.62															
PG: TOSOH AIA	8	64.74	7.56	42.07-87.41	11.67	8	50.24	5.45	33.90-66.58	10.84															
SG: Select Beckman Coulter Instruments (Chem)	21	58.54	7.25	36.77-80.31	12.39	21	35.56	4.57	21.85-49.27	12.85															
SG: Select Roche Instruments	7	37.73	6.26	18.94-56.52	16.60	6	21.48	1.89	15.82-27.14	8.78															
PG: Roche Cobas e 411	6	35.74	3.74	24.51-46.97	10.47	6	21.48	1.89	15.82-27.14	8.78															
PG: Abbott Architect i1000 SR	10	50.50	3.12	41.15-59.85	6.17	10	32.24	2.51	24.72-39.76	7.78															
SG: Select Abbott Architect Instruments	16	49.86	2.97	40.96-58.76	5.95	16	31.81	2.26	25.03-38.59	7.10															
PG: Qualigen FastPack Vitamin D	8	50.45	8.62	24.60-76.30	17.08	8	35.39	5.78	18.06-52.72	16.32															
AG: All Group	66	54.87	10.79	22.50-87.24	19.67	64	35.18	7.73	11.98-58.38	21.98															
752 - Auto Diff 8 (Sysmex 5-part Diff)	SXT-11					SXT-12					SXT-13					SXT-14					SXT-15				
White Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex XS-1000i	13	4.03	0.11	3.40-4.60	2.75	13	7.27	0.59	6.20-8.40	8.08	13	21.08	0.80	17.90-24.30	3.79	13	7.22	0.56	6.10-8.30	7.78	13	4.01	0.12	3.40-4.60	3.00
SG: Select Sysmex Instruments (5-part Diff)	14	4.03	0.11	3.40-4.60	2.75	14	7.33	0.60	6.20-8.40	8.36	14	21.11	0.77	17.90-24.30	3.65	14	7.30	0.63	6.20-8.40	8.63	14	4.02	0.12	3.40-4.60	3.00
Lymphocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex XS-1000i	13	11.27	2.14	4.90-17.70	18.94	13	27.60	4.13	15.20-40.00	14.96	12	19.47	1.28	15.70-23.30	6.56	13	27.62	2.71	19.50-35.70	9.78	13	12.81	2.02	6.70-18.90	15.78
SG: Select Sysmex Instruments (5-part Diff)	13	11.27	2.14	4.90-17.70	18.94	14	28.01	4.25	15.30-40.80	15.18	13	19.65	1.39	15.40-23.80	7.09	14	27.99	2.93	19.20-36.80	10.46	14	13.10	2.22	6.40-19.80	16.95
Neutrophil	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex XS-1000i	13	62.58	1.68	57.60-67.60	2.68	13	49.53	3.41	39.30-59.70	6.89	11	55.33	1.37	51.20-59.40	2.48	13	49.45	2.37	42.30-56.50	4.80	13	60.71	2.31	53.80-67.60	3.81
SG: Select Sysmex Instruments (5-part Diff)	14	62.77	1.77	57.50-68.10	2.82	14	49.64	3.30	39.70-59.50	6.65	12	55.78	2.05	49.60-61.90	3.67	14	49.64	2.39	42.40-56.80	4.82	14	61.14	2.75	52.90-69.30	4.50
Mono/Mixed	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex XS-1000i	11	0.61	0.26	0.00-1.40	43.33	11	2.50	0.48	1.10-3.90	19.20	12	1.52	0.63	0.00-3.40	42.00	11	2.52	0.39	1.30-3.70	15.60	13	0.87	0.85	0.00-3.40	94.44
SG: Select Sysmex Instruments (5-part Diff)	12	0.68	0.36	0.00-1.80	51.43	12	2.60	0.58	0.90-4.30	22.31	13	1.62	0.71	0.00-3.70	44.38	12	2.58	0.42	1.30-3.90	16.15	14	0.91	0.83	0.00-3.40	92.22
Eosinophil	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex XS-1000i	11	16.42	0.44	15.10-17.70	2.68	11	12.75	0.89	10.00-15.40	7.01	11	14.22	0.57	12.50-15.90	4.01	11	12.68	0.87	10.10-15.30	6.85	11	15.87	0.59	14.10-17.70	3.71
SG: Select Sysmex Instruments (5-part Diff)	12	16.34	0.49	14.80-17.80	3.01	12	12.68	0.87	10.10-15.30	6.85	12	14.24	0.55	12.50-15.80	3.87	12	12.62	0.86	10.00-15.20	6.83	12	15.78	0.64	13.90-17.70	4.05
Basophil	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex XS-1000i	11	9.18	0.74	6.97-11.39	8.02	11	7.17	1.14	3.75-10.59	15.91	11	9.50	0.69	7.44-11.56	7.24	11	7.65	0.62	5.78-9.52	8.13	11	10.27	0.64	8.35-12.19	6.23
SG: Select Sysmex Instruments (5-part Diff)	11	9.18	0.74	6.97-11.39	8.02	11	7.17	1.14	3.75-10.59	15.91	11	9.50	0.69	7.44-11.56	7.24	11	7.65	0.62	5.78-9.52	8.13	11	10.27	0.64	8.35-12.19	6.23
Red Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex XS-1000i	13	2.35	0.02	2.21-2.49	0.85	13	4.72	0.06	4.44-5.00	1.19	12	5.63	0.18	5.29-5.97	3.13	13	4.70	0.04	4.42-4.98	0.89	13	2.36	0.04	2.22-2.50	1.69
SG: Select Sysmex Instruments (5-part Diff)	14	2.36	0.03	2.22-2.50	1.06	14	4.72	0.06	4.44-5.00	1.29	13	5.64	0.17	5.30-5.98	3.01	14	4.71	0.05	4.43-4.99	1.15	14	2.36	0.05	2.22-2.50	1.99
Hemoglobin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex XS-1000i	13	5.78	0.08	5.40-6.20	1.38	13	13.15	0.16	12.20-14.00	1.22	13	17.73	0.52	16.50-18.90	2.94	13	13.15	0.13	12.30-14.10	0.98	13	5.80	0.09	5.40-6.20	1.55
SG: Select Sysmex Instruments (5-part Diff)	14	5.79	0.08	5.40-6.20	1.38	14	13.13	0.16	12.20-14.00	1.22	14	17.70	0.51	16.50-18.90	2.88	14	13.14	0.14	12.20-14.00	1.07	14	5.80	0.09	5.40-6.20	1.55
Hematocrit	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex XS-1000i	13	18.41	0.32	17.30-19.50	1.68	13	39.99	0.67	37.60-42.40	1.68	13	53.53	1.55	50.30-56.70	2.90	13	39.82	0.61	37.40-42.20	1.53	13	18.42	0.31	17.30-19.50	1.68
SG: Select Sysmex Instruments (5-part Diff)	14	18.44	0.33	17.30-19.50	1.79	14	40.01	0.65	37.60-42.40	1.63	14	53.54	1.49	50.30-56.70	2.79	14	39.86	0.62	37.50-42.30	1.55	14	18.48	0.36	17.40-19.60	1.95
MCV	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex XS-1000i	13	78.28	1.18	74.80-81.80	1.51	13	84.81	0.99	81.80-87.80	1.17	13	95.19	1.37	91.10-99.30	1.44	13	84.81	1.10	81.50-88.10	1.30	13	78.18	1.23	74.50-81.90	1.57
SG: Select Sysmex Instruments (5-part Diff)	14	78.29	1.13	74.90-81.70	1.44	14	84.72	1.00	81.70-87.70	1.18	14	95.10	1.36	91.00-99.20	1.43	14	84.74	1.09	81.40-88.00	1.29	14	78.17	1.18	74.70-81.70	1.51
Platelet Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex XS-1000i	13	64.46	3.23	48.00-80.00	5.00	13	238.38	6.54	179.00-298.00	2.73	13	443.62	12.18	333.00-555.00	2.75	13	237.77	5.89	179.00-298.00	2.48	13	65.54	5.14	50.00-83.00	7.73
SG: Select Sysmex Instruments (5-part Diff)	14	64.71	3.25	49.00-81.00	4.92	14	238.86	6.53	179.00-299.00	2.72	14	444.93	12.69	334.00-556.00	2.85	14	238.00	5.72	179.00-298.00	2.39	14	66.07	5.33	50.00-83.00	8.03
RDW	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex XS-1000i	12	17.15	0.17	16.70-17.70	0.99	13	16.12	0.13	15.70-16.50	0.81	13	14.89	0.56	13.20-16.60	3.76	13	16.09	0.17	15.60-16.60	1.06	13	17.17	0.21	16.60-17.80	1.22
SG: Select Sysmex Instruments (5-part Diff)	13	17.15	0.17	16.60-17.60	0.99	14	16.11	0.13	15.70-16.50	0.81	14	14.88	0.54	13.30-16.50	3.62	1									



PG: Sysmex XP-300	70	2.45	0.03	2.30-2.60	1.35	69	4.20	0.05	3.95-4.45	1.10	71	5.74	0.37	5.40-6.08	6.43	68	4.20	0.05	3.95-4.45	1.19	70	2.45	0.03	2.30-2.60	1.18
SG: International Sysmex Instruments	33	2.48	0.05	2.33-2.63	2.02	33	4.28	0.09	4.02-4.54	2.03	33	5.81	0.31	5.46-6.16	5.35	32	4.29	0.10	4.03-4.55	2.35	33	2.49	0.06	2.34-2.64	2.37
Hemoglobin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex K-800,1000,4500,KX-21	17	6.15	0.14	5.70-6.50	2.30	17	12.18	0.15	11.30-13.10	1.23	17	18.24	0.66	16.90-19.50	3.63	17	12.21	0.19	11.30-13.10	1.56	17	6.16	0.16	5.80-6.60	2.58
PG: Sysmex pocH-100i	29	6.26	0.16	5.90-6.70	2.54	29	12.32	0.20	11.40-13.20	1.63	29	18.39	0.98	17.10-19.70	5.33	28	12.33	0.24	11.40-13.20	1.95	29	6.25	0.14	5.90-6.70	2.38
SG: Select Sysmex Instruments (3-part Diff)	116	6.14	0.15	5.70-6.50	2.46	115	12.27	0.18	11.40-13.20	1.46	117	18.30	1.00	17.00-19.60	5.46	114	12.30	0.39	11.40-13.20	3.17	116	6.18	0.14	5.80-6.60	2.26
PG: Sysmex XP-300	70	6.09	0.13	5.70-6.50	2.13	69	12.26	0.18	11.40-13.20	1.46	71	18.27	1.08	17.00-19.60	5.90	69	12.31	0.47	11.40-13.20	3.82	70	6.16	0.13	5.80-6.60	2.10
SG: International Sysmex Instruments	33	6.25	0.15	5.80-6.60	2.42	33	12.32	0.19	11.40-13.20	1.54	33	18.37	0.93	17.10-19.70	5.05	32	12.33	0.22	11.40-13.20	1.79	33	6.25	0.14	5.80-6.60	2.26
Lymphocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex K-800,1000,4500,KX-21	17	11.68	1.19	8.10-15.30	10.17	17	31.42	1.04	28.30-34.50	3.31	17	59.64	1.01	56.60-62.60	1.69	17	31.42	1.28	27.60-35.20	4.08	16	11.27	1.35	7.30-15.40	11.95
PG: Sysmex pocH-100i	28	10.81	1.01	7.80-13.80	9.35	28	31.04	0.85	28.40-33.50	2.74	28	60.22	0.56	58.50-61.90	0.93	28	31.39	1.21	27.80-35.00	3.85	27	10.64	0.94	7.80-13.40	8.87
SG: Select Sysmex Instruments (3-part Diff)	113	11.24	1.20	7.60-14.80	10.71	113	31.40	0.97	28.50-34.30	3.09	115	59.40	4.51	45.90-72.90	7.59	115	31.49	1.86	25.90-37.10	5.90	110	10.97	1.25	7.30-14.80	11.36
PG: Sysmex XP-300	68	11.31	1.24	7.60-15.00	10.97	68	31.54	0.98	28.60-34.40	3.11	70	59.01	5.73	41.80-76.20	9.71	70	31.54	2.17	25.00-38.00	6.89	67	11.03	1.32	7.00-15.00	12.00
SG: International Sysmex Instruments	32	10.73	0.98	7.80-13.60	9.16	32	31.00	0.83	28.50-33.50	2.68	32	60.22	0.53	58.60-61.80	0.88	32	31.29	1.21	27.70-34.90	3.83	31	10.69	0.89	8.00-13.40	8.32
Hematocrit	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex K-800,1000,4500,KX-21	17	17.22	0.47	16.20-18.20	2.73	17	32.55	0.74	30.60-34.50	2.28	17	49.03	2.26	46.10-51.90	4.61	17	32.65	0.93	30.70-34.70	2.84	17	17.20	0.45	16.20-18.20	2.62
PG: Sysmex pocH-100i	29	18.26	0.55	17.20-19.40	3.01	29	34.79	1.05	32.70-36.90	3.02	29	51.52	3.17	48.40-54.60	6.16	28	34.94	1.07	32.80-37.00	3.07	29	18.37	0.59	17.30-19.50	3.21
SG: Select Sysmex Instruments (3-part Diff)	116	17.68	0.56	16.60-18.80	3.16	115	33.48	1.08	31.50-35.50	3.22	117	50.11	3.25	47.10-53.10	6.49	113	33.53	1.13	31.50-35.50	3.37	116	17.71	0.58	16.60-18.80	3.28
PG: Sysmex XP-300	70	17.55	0.36	16.40-18.60	2.06	69	33.17	0.55	31.20-35.20	1.66	71	49.80	3.33	46.80-52.80	6.69	68	33.17	0.55	31.20-35.20	1.66	70	17.56	0.34	16.50-18.70	1.93
SG: International Sysmex Instruments	33	18.25	0.53	17.20-19.40	2.90	33	34.77	1.01	32.70-36.90	2.90	33	51.42	2.99	48.30-54.50	5.82	32	34.91	1.03	32.80-37.00	2.95	33	18.32	0.58	17.20-19.40	3.17
Mono/Mixed	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex K-800,1000,4500,KX-21	17	20.10	1.24	16.40-23.80	6.17	17	16.25	0.86	13.70-18.90	5.28	17	14.93	0.51	13.40-16.40	3.42	17	16.29	1.19	12.70-19.90	7.30	17	19.91	1.12	16.50-23.30	5.63
PG: Sysmex pocH-100i	28	18.29	1.45	14.00-22.60	7.92	28	14.60	0.96	11.70-17.50	6.58	28	13.84	0.63	11.90-15.70	4.57	28	14.84	1.18	11.30-18.30	7.97	28	18.32	1.70	13.20-23.40	9.29
SG: Select Sysmex Instruments (3-part Diff)	113	19.75	1.71	14.60-24.80	8.68	112	15.82	1.19	12.20-19.40	7.53	110	14.55	1.01	11.60-17.60	6.92	114	15.87	1.34	11.90-19.90	8.43	114	19.98	1.83	14.50-25.50	9.15
PG: Sysmex XP-300	68	20.26	1.59	15.50-25.10	7.83	67	16.22	0.99	13.20-19.20	6.11	65	14.76	1.09	11.50-18.10	7.36	69	16.18	1.23	12.50-19.90	7.59	69	20.67	1.58	16.00-25.40	7.63
PG: Intl Sysmex pocH-100i	5	18.75	1.29	14.90-22.70	6.86	5	14.43	0.57	12.70-16.10	3.96	5	13.50	0.26	12.70-14.30	1.93	5	15.68	1.46	11.30-20.10	9.30	5	19.75	0.93	17.00-22.60	4.70
SG: International Sysmex Instruments	32	18.35	1.42	14.00-22.60	7.76	32	14.58	0.92	11.80-17.40	6.30	32	13.79	0.60	12.00-15.60	4.35	32	14.94	1.22	11.20-18.60	8.19	32	18.50	1.68	13.50-23.50	9.08
MCV	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex K-800,1000,4500,KX-21	17	70.22	1.22	66.50-73.90	1.74	17	77.46	1.40	73.30-81.70	1.81	17	85.55	1.48	81.10-89.90	1.73	17	77.66	1.37	73.60-81.80	1.76	17	70.16	1.22	66.50-73.90	1.74
PG: Sysmex pocH-100i	26	73.88	1.15	70.50-77.40	1.56	26	81.60	1.23	77.90-85.30	1.51	26	88.70	1.33	84.70-92.70	1.50	26	81.67	1.24	78.00-85.40	1.52	26	73.87	1.19	70.30-77.50	1.61
SG: Select Sysmex Instruments (3-part Diff)	112	71.92	1.62	67.00-76.80	2.25	112	79.30	1.77	74.00-84.60	2.23	113	87.05	1.58	82.30-91.70	1.82	112	79.44	1.71	74.30-84.50	2.15	112	71.89	1.62	67.00-76.80	2.25
PG: Sysmex XP-300	69	71.59	1.08	68.40-74.80	1.49	69	78.89	1.03	75.80-82.00	1.31	70	86.80	1.11	83.50-90.10	1.28	69	79.04	0.96	76.10-81.90	1.22	69	71.57	1.03	68.50-74.70	1.44
SG: International Sysmex Instruments	30	73.75	1.19	70.20-77.40	1.61	30	81.50	1.24	77.80-85.20	1.52	30	88.62	1.32	84.60-92.60	1.49	30	81.55	1.23	77.90-85.30	1.51	30	73.76	1.19	70.20-77.40	1.61
Platelet Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex K-800,1000,4500,KX-21	17	57.47	8.91	43.00-71.00	15.61	17	189.53	6.33	143.00-238.00	3.32	17	395.35	14.27	296.00-494.00	3.62	17	189.94	6.30	143.00-238.00	3.32	17	59.00	3.48	44.00-74.00	5.93
PG: Sysmex pocH-100i	29	58.62	4.27	44.00-74.00	7.29	29	183.55	8.17	138.00-230.00	4.46	29	370.10	22.90	278.00-463.00	6.19	29	184.52	10.39	139.00-231.00	5.62	29	59.48	4.19	44.00-74.00	7.12
SG: Select Sysmex Instruments (3-part Diff)	115	58.87	5.30	44.00-74.00	8.98	115	188.76	9.90	142.00-236.00	5.24	115	389.10	25.18	292.00-486.00	6.48	115	189.27	9.38	142.00-236.00	4.97	115	59.26	3.92	44.00-74.00	6.61
PG: Sysmex XP-300	69	59.32	4.49	44.00-74.00	7.63	69	190.75	10.56	143.00-239.00	5.55	69	395.54	24.37	297.00-495.00	6.16	69	191.10	8.98	143.00-239.00	4.71	69	59.23	3.96	44.00-74.00	6.78
SG: International Sysmex Instruments	33	58.52	4.06	44.00-74.00	6.95	33	183.24	8.36	137.00-229.00	4.59	33	369.06	21.96	277.00-461.00	5.96	33	184.36	10.12	138.00-230.00	5.49	33	58.94	4.58	44.00-74.00	7.80
RDW	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD </							



PG: Siemens Clinitek 50	8	67.50	23.15	0.00-137.00	33.97																					
PG: Siemens Clinitek Status	25	72.00	18.71	16.00-128.00	25.97																					
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	8	84.21	2.03	78.00-90.00	2.38																					
PG: Beckman (Olympus) AU	7	82.73	4.43	70.00-96.00	5.30																					
SG: Select Beckman Coulter Instruments (Chem)	8	82.51	4.15	71.00-95.00	4.94																					
SG: Select Siemens/Dade Instruments	8	84.21	2.03	78.00-90.00	2.38																					
SG: Semi-quant Methods - Ur Microalbumin-Creat	37	72.23	18.90	15.00-129.00	26.25																					
PG: Siemens (Bayer) DCA Vantage	14	80.86	5.27	65.00-97.00	6.54																					
SG: Select Siemens (Bayer) DCA Instruments	17	81.44	5.16	65.00-97.00	6.42																					
PG: Afinion ACR	7	87.61	2.74	80.00-96.00	3.07																					
AG: All Group	86	76.47	17.53	24.00-129.00	23.03																					
728 - Urine Chemistry		URC-5					URC-6																			
Creatinine - Urine	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	
AG: All Group	5	127.02	10.94	94.20-159.80	8.61	5	186.86	13.78	145.60-228.20	7.37																
Total Protein - Urine	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	
AG: All Group	6	57.48	6.87	36.90-78.10	11.95	6	97.47	6.27	78.70-116.30	6.43																
601 - Waived/PPM Tests II;609 - Waived/PPM Tests I;664 - Whole Blood Glucose;669 - Whole Blood Glucose		WBG-11					WBG-12					WBG-13					WBG-14					WBG-15				
Glucose Wh Blood (Waived)	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	Labs	Mean	SD	Range	CV,%	
PG: Abbott Freestyle Lite	11	238.91	10.77	207.00-271.00	4.52																					
PG: Bayer Ascensia Contour	112	205.66	35.82	99.00-313.00	17.38	18	289.44	31.43	195.00-383.00	10.87																
PG: Bayer Ascensia Elite XL	10	192.80	26.80	113.00-273.00	13.89																					
PG: Lifescan One Touch Ultra	39	270.70	40.27	150.00-392.00	14.87																					
PG: Lifescan Sure Step, others	5	247.80	27.63	165.00-331.00	11.13																					
PG: McKesson/SunMark True Track	42	337.95	143.45	0.00-769.00	42.46																					
PG: Roche Accu-Chek Advantage/Comp/GTS	27	244.74	20.44	184.00-306.00	8.33																					
PG: Roche AccuChek Aviva	32	258.22	16.38	209.00-307.00	6.36	5	377.20	62.84	189.00-565.00	16.66																
PG: Roche Accu-Chek II	5	236.20	39.10	119.00-353.00	16.57																					
PG: Roche Accu-Chek, others	18	245.94	20.23	185.00-307.00	8.21																					
SG: Select Lifescan One Touch Instruments	48	266.43	40.29	145.00-387.00	15.15																					
SG: Select Lifescan SureStep Instruments	5	247.80	27.63	165.00-331.00	11.13																					
SG: Select Abbott Freestyle Instruments	18	237.33	12.40	200.00-274.00	5.23																					
PG: Roche AccuChek Inform	71	255.03	7.55	232.00-278.00	2.98	38	343.95	6.67	324.00-364.00	1.95	34	116.06	2.51	109.00-124.00	2.16	34	208.00	4.49	195.00-222.00	2.16	34	341.47	17.92	287.00-395.00	5.25	
PG: Consult Diag Quintet Blood Glucose	30	261.80	21.48	198.00-327.00	8.21	10	361.90	23.36	292.00-432.00	6.46																
PG: Medisense Precision Xceed Pro	27	236.70	20.87	174.00-300.00	8.82	5	311.40	24.99	236.00-386.00	8.04																
SG: Select Medisense Instruments	34	236.65	21.83	172.00-302.00	9.20	5	311.40	24.99	236.00-386.00	8.04																
SG: Select Roche WBG Meters	283	252.76	18.16	198.00-308.00	7.19	56	345.20	22.84	277.00-413.00	6.61	36	115.81	2.79	108.00-124.00	2.41	36	208.19	4.45	195.00-221.00	2.12	36	341.78	17.45	290.00-395.00	5.12	
SG: Select Bayer WBG Meters	137	205.42	36.28	96.00-314.00	17.71	24	294.75	41.11	172.00-418.00	13.93	7	92.00	19.77	33.00-151.00	21.52	7	177.29	25.88	99.00-255.00	14.63	7	294.14	42.25	167.00-421.00	14.35	
SG: Select True/True Track WBG Instruments	51	350.47	143.75	0.00-781.00	41.06																					
PG: Nova Biomedical StatStrip Glucose meter	6	232.17	9.60	203.00-261.00	4.14																					
SG: Select Nova Biomedical StatStrip Instruments	9	227.11	12.14	191.00-263.00	5.33																					
PG: McKesson TrueResult	7	228.00	33.99	126.00-330.00	14.91																					
SG: Select True Track/TRUE Balance WBG Instruments	42	337.95	143.45	0.00-769.00	42.46																					
PG: Bayer Contour Next	8	198.00	30.94	105.00-291.00	15.61																					
PG: Nipro TRUEtrack	6	499.67	27.11	419.00-581.00	5.42																					
PG: Roche AccuChek Performa	129	253.78	19.61	195.00-313.00	7.72	7	345.29	5.25	329.00-361.00	1.54																
SG: Select TrueResult Meters	10	227.80	27.97	144.00-312.00	12.28																					
PG: McKesson True Metrix Pro Blood Glucose Meter	20	223.10	14.54	180.00-267.00	6.50																					
PG: EvenCare G3 Blood Glucose System	5	248.40	12.60	210.00-286.00	5.08																					
SG: Select EvenCare Whole Blood Glucose	7	250.00	10.89	217.00-283.00	4.36																					
AG: All Group	669	249.60	59.22	72.00-428.00	23.68	108	338.82	55.19	173.00-505.00	16.28	55	110.13	12.16	73.00-147.00	11.09	56	206.29	37.04	95.00-317.00	17.96	56	336.91	48.06	193.00-481.00	14.27	



789 - Bacterial Antigen Detection	Labs	Response	Number	Percent
C. difficile Toxin Detect				
AD-11		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Alere C diff Quik Chek Complete	2	*Positive	2	100.00%
AD-12		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Alere C diff Quik Chek Complete	2	*Negative	2	100.00%
AD-13		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Alere C diff Quik Chek Complete	2	*Positive	2	100.00%
AD-14		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Alere C diff Quik Chek Complete	2	*Positive	2	100.00%
AD-15		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Alere C diff Quik Chek Complete	2	*Negative	2	100.00%
760 - ANA (ELISA);643 - ANA (Latex);757 - Special Immunology	Labs	Response	Number	Percent
ANA (ELISA) (International)				
AE-11		* Denotes Correct Response		
PG: Dynex DS2	1			
		*Negative	1	100.00%
AE-12		* Denotes Correct Response		
PG: Dynex DS2	1			
		*Positive	1	100.00%
AE-13		* Denotes Correct Response		
PG: Dynex DS2	2			
		*Positive	1	50.00%
		Negative	1	50.00%
AE-14		* Denotes Correct Response		
PG: Dynex DS2	1			
		*Negative	1	100.00%
AE-15		* Denotes Correct Response		
PG: Dynex DS2	1			
		*Positive	1	100.00%
765 - Allergen Testing (1 specimen);671 - Total IgE;674 - Diagnostic Allergy	Labs	Response	Number	Percent
Ash - White Alrqn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	2			
		*0	1	50.00%
		*1/0	1	50.00%
Hitachi CLA-1	2	*0	1	50.00%
Hitachi CLA-1		*1/0	1	50.00%
Birch/Alder Mix Alrqn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	3			
		*1	2	66.67%
		*2	1	33.33%
Hitachi CLA-1	3	*1	2	66.67%
Hitachi CLA-1		*2	1	33.33%
Cedar - Mountain Alrqn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	3			
		*0	2	66.67%
		*1/0	1	33.33%
Hitachi CLA-1	3	*0	2	66.67%
Hitachi CLA-1		*1/0	1	33.33%
Cottonwood - Eastern Alrqn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	2			



		*0	2	100.00%
Hitachi CLA-1	2	*0	2	100.00%
Elm - White Alrgn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	3			
		*1/0	3	100.00%
Hitachi CLA-1	3	*1/0	3	100.00%
Mulberry Mix Alrgn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	3			
		*0	1	33.33%
		*1/0	2	66.67%
Hitachi CLA-1	3	*0	1	33.33%
Hitachi CLA-1		*1/0	2	66.67%
Pine Mix Alrgn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	3			
		*0	2	66.67%
		*1/0	1	33.33%
Hitachi CLA-1	3	*0	2	66.67%
Hitachi CLA-1		*1/0	1	33.33%
Privet Alrgn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	3			
		*0	1	33.33%
		*1/0	2	66.67%
Hitachi CLA-1	3	*0	1	33.33%
Hitachi CLA-1		*1/0	2	66.67%
Sycamore - American Alrgn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	3			
		*0	1	33.33%
		*1/0	2	66.67%
Hitachi CLA-1	3	*0	1	33.33%
Hitachi CLA-1		*1/0	2	66.67%
Cocklebur Alrgn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	3			
		*0	1	33.33%
		*1/0	2	66.67%
Hitachi CLA-1	3	*0	1	33.33%
Hitachi CLA-1		*1/0	2	66.67%
Lambs Quarters Alrgn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	3			
		*1	3	100.00%
Hitachi CLA-1	3	*1	3	100.00%
Marshelder - Rough Alrgn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	2			
		*1/0	2	100.00%
Hitachi CLA-1	2	*1/0	2	100.00%
Pigweed Alrgn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	3			
		*0	2	66.67%
		*1/0	1	33.33%
Hitachi CLA-1	3	*0	2	66.67%
Hitachi CLA-1		*1/0	1	33.33%
Ragweed - Short Alrgn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	3			
		*1	1	33.33%
		*2	2	66.67%
Hitachi CLA-1	3	*1	1	33.33%
Hitachi CLA-1		*2	2	66.67%
Sheep Sorrel Alrgn				



AL-11(Q)		* Denotes Correct Response		
AG: All Methods	3			
		*2	2	66.67%
		*3	1	33.33%
Hitachi CLA-1	3	*2	2	66.67%
Hitachi CLA-1		*3	1	33.33%
Waterhemp Alrqn		* Denotes Correct Response		
AG: All Methods	2			
		*2	1	50.00%
		*3	1	50.00%
Hitachi CLA-1	2	*2	1	50.00%
Hitachi CLA-1		*3	1	50.00%
Mite D Farinae Alrqn		* Denotes Correct Response		
AG: All Methods	4			
		*4	4	100.00%
Hitachi CLA-1	4	*4	4	100.00%
Cladosporium Alrqn		* Denotes Correct Response		
AG: All Methods	3			
		*1	1	33.33%
		*1/0	2	66.67%
Hitachi CLA-1	3	*1	1	33.33%
Hitachi CLA-1		*1/0	2	66.67%
Tomato Alrqn		* Denotes Correct Response		
AG: All Methods	2			
		*1	2	100.00%
Hitachi CLA-1	2	*1	2	100.00%
Wheat Alrqn		* Denotes Correct Response		
AG: All Methods	2			
		*4	2	100.00%
Hitachi CLA-1	2	*4	2	100.00%
Shellfish Mix Alrqn		* Denotes Correct Response		
AG: All Methods	2			
		*4	2	100.00%
Hitachi CLA-1	2	*4	2	100.00%
Rice Alrqn		* Denotes Correct Response		
AG: All Methods	2			
		*2	1	50.00%
		*3	1	50.00%
Hitachi CLA-1	2	*2	1	50.00%
Hitachi CLA-1		*3	1	50.00%
Milk Alrqn		* Denotes Correct Response		
AG: All Methods	2			
		*4	2	100.00%
Hitachi CLA-1	2	*4	2	100.00%
Egg - whole Alrqn		* Denotes Correct Response		
AG: All Methods	2			
		*4	2	100.00%
Hitachi CLA-1	2	*4	2	100.00%
Corn Alrqn		* Denotes Correct Response		
AG: All Methods	2			
		*1	2	100.00%
Hitachi CLA-1	2	*1	2	100.00%
Alternaria alternata Alrqn		* Denotes Correct Response		
AG: All Methods	5			
		*1	4	80.00%
		*1/0	1	20.00%



Hitachi CLA-1	5	*1	4	80.00%
Hitachi CLA-1		*1/0	1	20.00%
Acacia Alrgn				
AL-11(Q)		* Denotes Correct Response		
PG: Hitachi CLA-1	2			
		*0	2	100.00%
Aspergillus fumigatus Alrgn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	3			
		*1	3	100.00%
Hitachi CLA-1	3	*1	3	100.00%
Bermuda Grass Alrgn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	5			
		*1	3	60.00%
		*2	1	20.00%
		3	1	20.00%
Hitachi CLA-1	5	*1	3	60.00%
Hitachi CLA-1		*2	1	20.00%
Hitachi CLA-1		3	1	20.00%
Cat Epithelium Alrgn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	3			
		*4	3	100.00%
Hitachi CLA-1	3	*4	3	100.00%
Dog Epithelium Alrgn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	2			
		*4	2	100.00%
Hitachi CLA-1	2	*4	2	100.00%
Dog Dander Alrgn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	4			
		*4	4	100.00%
Hitachi CLA-1	4	*4	4	100.00%
Bahia Grass Alrgn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	2			
		*3	2	100.00%
Hitachi CLA-1	2	*3	2	100.00%
English Plantain Alrgn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	4			
		*1	3	75.00%
		*1/0	1	25.00%
Hitachi CLA-1	4	*1	3	75.00%
Hitachi CLA-1		*1/0	1	25.00%
Oak Alrgn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	2			
		*3	2	100.00%
Hitachi CLA-1	2	*3	2	100.00%
Timothy Grass Alrgn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	4			
		*4	4	100.00%
Hitachi CLA-1	4	*4	4	100.00%
Beech, American Alrgn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	2			
		*0	2	100.00%
Hitachi CLA-1	2	*0	2	100.00%
Box Elder, Maple Alrgn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	3			
		*1	1	33.33%
		*1/0	2	66.67%



Hitachi CLA-1	3	*1	1	33.33%
Hitachi CLA-1		*1/0	2	66.67%
Candida Alrqn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	2			
		*0	1	50.00%
		*1/0	1	50.00%
Hitachi CLA-1	2	*0	1	50.00%
Hitachi CLA-1		*1/0	1	50.00%
Peanut (Arachis hypogaea) Alrqn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	3			
		*4	3	100.00%
Hitachi CLA-1	3	*4	3	100.00%
Cat Dander Alrqn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	2			
		*4	2	100.00%
Hitachi CLA-1	2	*4	2	100.00%
Soybean Alrqn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	3			
		*3	2	66.67%
		*4	1	33.33%
Hitachi CLA-1	3	*3	2	66.67%
Hitachi CLA-1		*4	1	33.33%
Cockroach Mix Alrqn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	3			
		*1	1	33.33%
		*1/0	1	33.33%
		*2	1	33.33%
Hitachi CLA-1	3	*1	1	33.33%
Hitachi CLA-1		*1/0	1	33.33%
Hitachi CLA-1		*2	1	33.33%
Melaleuca Alrqn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	2			
		*0	2	100.00%
Hitachi CLA-1	2	*0	2	100.00%
Oak, White Alrqn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	2			
		*3	2	100.00%
Hitachi CLA-1	2	*3	2	100.00%
Penicillium Alrqn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	2			
		*1	2	100.00%
Hitachi CLA-1	2	*1	2	100.00%
Walnut/Hickory/Pecan Mix Alrqn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	3			
		*1	3	100.00%
Hitachi CLA-1	3	*1	3	100.00%
Yeast - Bakers Alrqn				
AL-11(Q)		* Denotes Correct Response		
AG: All Methods	2			
		*0	1	50.00%
		*1	1	50.00%
Hitachi CLA-1	2	*0	1	50.00%
Hitachi CLA-1		*1	1	50.00%
683 - ASO/CRP;759 - Complete Immunology;729 - AFC Immunology	Labs	Response	Number	Percent
ASO (Qual) (International)				
AS-11		* Denotes Correct Response		
AG: All Methods	1			



		*Positive	1	100.00%
Wampole ColorCard ASO	1	*Positive	1	100.00%
AS-12		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Wampole ColorCard ASO	1	*Positive	1	100.00%
AS-13		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Wampole ColorCard ASO	1	*Negative	1	100.00%
AS-14		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Wampole ColorCard ASO	1	*Negative	1	100.00%
AS-15		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Wampole ColorCard ASO	1	*Positive	1	100.00%
782 - Complete Bacteriology	Labs	Response	Number	Percent
		Blood Culture		
BA-9		* Denotes Correct Response		
AG: All Methods	1			
		*Pseudomonas aeruginosa	1	100.00%
Cultures	1	*Pseudomonas aeruginosa	1	100.00%
		Wound Culture		
BA-8		* Denotes Correct Response		
AG: All Methods	2			
		*Staphylococcus aureus	1	50.00%
		*Staphylococcus, coagulase positive	1	50.00%
Cultures	2	*Staphylococcus aureus	1	50.00%
Cultures		*Staphylococcus, coagulase positive	1	50.00%
736 - ABO/Rh Only;650 - Blood Bank	Labs	Response	Number	Percent
		ABO Group		
BB-11		* Denotes Correct Response		
AG: All Methods	1			
		*Group B	1	100.00%
Blood Bank Testing	1	*Group B	1	100.00%
BB-12		* Denotes Correct Response		
AG: All Methods	1			
		*Group B	1	100.00%
Blood Bank Testing	1	*Group B	1	100.00%
BB-13		* Denotes Correct Response		
AG: All Methods	1			
		*Group O	1	100.00%
Blood Bank Testing	1	*Group O	1	100.00%
BB-14		* Denotes Correct Response		
AG: All Methods	1			
		*Group A	1	100.00%
Blood Bank Testing	1	*Group A	1	100.00%
BB-15		* Denotes Correct Response		
AG: All Methods	1			
		*Group A	1	100.00%
Blood Bank Testing	1	*Group A	1	100.00%
		Rh Type		
BB-11		* Denotes Correct Response		
AG: All Methods	6			
		*Rh Positive	6	100.00%
Blood Bank Testing	6	*Rh Positive	6	100.00%
BB-12		* Denotes Correct Response		
AG: All Methods	6			
		*Rh Positive	6	100.00%
Blood Bank Testing	6	*Rh Positive	6	100.00%
BB-13		* Denotes Correct Response		
AG: All Methods	6			
		*Rh Positive	6	100.00%
Blood Bank Testing	6	*Rh Positive	6	100.00%
BB-14		* Denotes Correct Response		



AG: All Methods	6			
		*Rh Negative	6	100.00%
Blood Bank Testing	6	*Rh Negative	6	100.00%
BB-15		* Denotes Correct Response		
AG: All Methods	6			
		*Rh Positive	4	66.67%
		Rh Negative	2	33.33%
Blood Bank Testing	6	*Rh Positive	4	66.67%
Blood Bank Testing	6	Rh Negative	2	33.33%
975 - Bacterial Vaginosis	Labs	Response	Number	Percent
		Bacterial Vaginosa (Sialidase)		
BV-5		* Denotes Correct Response		
AG: All Methods	92			
		*Negative	86	93.48%
		Positive	6	6.52%
Sekisui Diagnostics OSOM BVBlue	92	*Negative	86	93.48%
Sekisui Diagnostics OSOM BVBlue	92	Positive	6	6.52%
BV-6		* Denotes Correct Response		
AG: All Methods	92			
		*Positive	81	88.04%
		Negative	11	11.96%
Sekisui Diagnostics OSOM BVBlue	92	*Positive	81	88.04%
Sekisui Diagnostics OSOM BVBlue	92	Negative	11	11.96%
661 - Urine Colony Count;784 - Presump ID/Colony Count - Urine;785 - Presump ID/Colony Count-Urine & Strep Combo;775 - Urine Culture;778 - Urine & Throat Cultures;779 - Urine & Genital Cultures;781 - Urine, Throat & Genital Cultures;782 - Complete Bacteri	Labs	Response	Number	Percent
		Colony Count		
CC-11		* Denotes Correct Response		
AG: All Methods	33			
		*Greater than 100,000	27	81.82%
		10,000 - 100,000	5	15.15%
		No growth	1	3.03%
Calibrated Loop - Manual	10	*Greater than 100,000	5	50.00%
Uri-Chek (Troy Bio)	5	*Greater than 100,000	4	80.00%
Uricult (Orion)	16	*Greater than 100,000	16	100.00%
Uri-Three (Culture Kits	1	*Greater than 100,000	1	100.00%
Uricult (Lifesign)	1	*Greater than 100,000	1	100.00%
Calibrated Loop - Manual	10	10,000 - 100,000	4	40.00%
Uri-Chek (Troy Bio)	5	10,000 - 100,000	1	20.00%
Calibrated Loop - Manual	10	No growth	1	10.00%
784 - Presump ID/Colony Count - Urine;785 - Presump ID/Colony Count-Urine & Strep Combo	Labs	Response	Number	Percent
		Colony Count		
CC-11		* Denotes Correct Response		
AG: All Methods	33			
		*Greater than 100,000	27	81.82%
		10,000 - 100,000	5	15.15%
		No growth	1	3.03%
Calibrated Loop - Manual	10	*Greater than 100,000	5	50.00%
Uri-Chek (Troy Bio)	5	*Greater than 100,000	4	80.00%
Uricult (Orion)	16	*Greater than 100,000	16	100.00%
Uri-Three (Culture Kits	1	*Greater than 100,000	1	100.00%
Uricult (Lifesign)	1	*Greater than 100,000	1	100.00%
Calibrated Loop - Manual	10	10,000 - 100,000	4	40.00%
Uri-Chek (Troy Bio)	5	10,000 - 100,000	1	20.00%
Calibrated Loop - Manual	10	No growth	1	10.00%
CC-12		* Denotes Correct Response		
AG: All Methods	6			
		*Greater than 100,000	5	83.33%
		10,000 - 100,000	1	16.67%
Calibrated Loop - Manual	4	*Greater than 100,000	3	75.00%
Uri-Chek (Troy Bio)	1	*Greater than 100,000	1	100.00%
Uricult (Orion)	1	*Greater than 100,000	1	100.00%
Calibrated Loop - Manual	4	10,000 - 100,000	1	25.00%
CC-13		* Denotes Correct Response		
AG: All Methods	6			
		*10,000 - 100,000	1	16.67%



		*Greater than 100,000	4	66.67%
		Less than 10,000	1	16.67%
Uri-Chek (Troy Bio)	1	*10,000 - 100,000	1	100.00%
Calibrated Loop - Manual	4	*Greater than 100,000	3	75.00%
Uricult (Orion)	1	*Greater than 100,000	1	100.00%
Calibrated Loop - Manual	4	Less than 10,000	1	25.00%
CC-14		* Denotes Correct Response		
AG: All Methods	6			
		*Greater than 100,000	5	83.33%
		10,000 - 100,000	1	16.67%
Calibrated Loop - Manual	4	*Greater than 100,000	3	75.00%
Uri-Chek (Troy Bio)	1	*Greater than 100,000	1	100.00%
Uricult (Orion)	1	*Greater than 100,000	1	100.00%
Calibrated Loop - Manual	4	10,000 - 100,000	1	25.00%
CC-15		* Denotes Correct Response		
AG: All Methods	6			
		*Greater than 100,000	6	100.00%
Calibrated Loop - Manual	4	*Greater than 100,000	4	100.00%
Uri-Chek (Troy Bio)	1	*Greater than 100,000	1	100.00%
Uricult (Orion)	1	*Greater than 100,000	1	100.00%
		Urine Presump Id		
CC-11		* Denotes Correct Response		
AG: All Methods	3			
		*Presumptive Pseudomonas sp.	3	100.00%
Calibrated Loop - Manual	2	*Presumptive Pseudomonas sp.	2	100.00%
Uricult (Orion)	1	*Presumptive Pseudomonas sp.	1	100.00%
CC-12		* Denotes Correct Response		
AG: All Methods	3			
		*Presumptive Klebsiella sp.	3	100.00%
Calibrated Loop - Manual	2	*Presumptive Klebsiella sp.	2	100.00%
Uricult (Orion)	1	*Presumptive Klebsiella sp.	1	100.00%
CC-13		* Denotes Correct Response		
AG: All Methods	3			
		*Presumptive Proteus sp.	3	100.00%
Calibrated Loop - Manual	2	*Presumptive Proteus sp.	2	100.00%
Uricult (Orion)	1	*Presumptive Proteus sp.	1	100.00%
CC-14		* Denotes Correct Response		
AG: All Methods	3			
		*Gram Positive Organism	1	33.33%
		*Presumptive Staphylococcus sp.	1	33.33%
		Presumptive Proteus sp.	1	33.33%
Uricult (Orion)	1	*Gram Positive Organism	1	100.00%
Calibrated Loop - Manual	2	*Presumptive Staphylococcus sp.	1	50.00%
Calibrated Loop - Manual		Presumptive Proteus sp.	1	50.00%
CC-15		* Denotes Correct Response		
AG: All Methods	3			
		*Presumptive E. Coli	3	100.00%
Calibrated Loop - Manual	2	*Presumptive E. Coli	2	100.00%
Uricult (Orion)	1	*Presumptive E. Coli	1	100.00%
		Susceptibility Amikacin		
CC-11		* Denotes Correct Response		
AG: All Methods	2			
		*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
		Susceptibility Aztreonam		
CC-11		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
		Susceptibility Cefepime		
CC-11		* Denotes Correct Response		
AG: All Methods	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
		Susceptibility Cefazidime		
CC-11		* Denotes Correct Response		
AG: All Methods	3			
		*Sensitive	3	100.00%



Susceptibility	3	*Sensitive	3	100.00%
CC-11		Susceptibility Levofloxacin		
AG: All Methods		* Denotes Correct Response		
	5			
		*Intermediate	4	80.00%
		*Resistant	1	20.00%
Susceptibility	5	*Intermediate	4	80.00%
Susceptibility		*Resistant	1	20.00%
CC-11		Susceptibility Piperacillin		
AG: All Methods		* Denotes Correct Response		
	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
CC-11		Susceptibility Piperacillin/Tazobactam		
AG: All Methods		* Denotes Correct Response		
	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
CC-11		Susceptibility Tobramycin		
AG: All Methods		* Denotes Correct Response		
	4			
		*Sensitive	4	100.00%
Susceptibility	4	*Sensitive	4	100.00%
CC-11		Susceptibility Ertapenem		
AG: All Methods		* Denotes Correct Response		
	1			
		*Resistant	1	100.00%
Susceptibility	1	*Resistant	1	100.00%
714 - Complete Cardiac Markers;715 - Basic Cardiac Markers;720 - Cardiac Markers I;740 - D-Dimer	Labs	Response	Number	Percent
CDM-11		Myoglobin (Qualitative)		
PG: Princeton Biomedical Corp. Lifesign MI		* Denotes Correct Response		
	2			
		*Positive	2	100.00%
CDM-12		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI				
	2			
		*Positive	1	50.00%
		Negative	1	50.00%
CDM-13		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI				
	2			
		*Negative	1	50.00%
		Positive	1	50.00%
CDM-14		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI				
	2			
		*Positive	2	100.00%
CDM-15		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI				
	2			
		*Positive	2	100.00%
CDM-11		CK-MB (Qualitative)		
PG: Princeton Biomedical Corp. Lifesign MI		* Denotes Correct Response		
	2			
		*Positive	2	100.00%
CDM-12		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI				
	2			
		*Positive	2	100.00%
CDM-13		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI				
	2			
		Negative	1	50.00%
		Positive	1	50.00%
CDM-14		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI				
	2			
		*Positive	2	100.00%
CDM-15		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI				
	2			
		*Positive	2	100.00%
CDM-11		Troponin I (Qualitative)		
PG: Princeton Biomedical Corp. Lifesign MI		* Denotes Correct Response		
	2			
		*Positive	2	100.00%



PG: Princeton Biomedical Corp. Lifesign MI	2			
		*Positive	2	100.00%
CDM-12		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI	2			
		*Negative	1	50.00%
		Positive	1	50.00%
CDM-13		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI	2			
		*Negative	2	100.00%
CDM-14		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI	2			
		*Positive	2	100.00%
CDM-15		* Denotes Correct Response		
PG: Princeton Biomedical Corp. Lifesign MI	2			
		*Positive	2	100.00%
612 - MultiChem 5;616 - MultiChem 10;617 - MultiChem 15;618 - MultiChem 20;619 - MultiChem 21+;676 - Endocrinology	Labs	Response	Number	Percent
		TSH (Semi-Quant)		
CH-11		* Denotes Correct Response		
PG: Biomedical Thyro Test	1			
		<= 5	1	100.00%
CH-12		* Denotes Correct Response		
PG: Biomedical Thyro Test	1			
		<= 5	1	100.00%
CH-13		* Denotes Correct Response		
PG: Biomedical Thyro Test	1			
		*<= 5	1	100.00%
CH-14		* Denotes Correct Response		
PG: Biomedical Thyro Test	1			
		<= 5	1	100.00%
CH-15		* Denotes Correct Response		
PG: Biomedical Thyro Test	1			
		<= 5	1	100.00%
639 - Complete Urinalysis Package;638 - Urine Sediment;641 - Clinical Microscopy;601 - Waived/PPM Tests I;609 - Waived/PPM Tests I;628 - AFC Waived/PPM;646 - Clinical Microscopy Extra	Labs	Response	Number	Percent
		Urine Sediment Id		
CM-19		* Denotes Correct Response		
AG: All Methods	690			
		*Red blood cell (RBC) (Erythrocyte)	680	98.55%
		Amorphous urates/phosphates	1	0.14%
		RBC/Blood/Hgb cast	3	0.43%
		Renal tubular epithelial (RTE)	1	0.14%
		Transitional epithelial cell (urothelial)	2	0.29%
		White blood cell (WBC) (Leukocyte)	3	0.43%
Microscopy	690	*Red blood cell (RBC) (Erythrocyte)	680	98.55%
Microscopy		Amorphous urates/phosphates	1	0.14%
Microscopy		RBC/Blood/Hgb cast	3	0.43%
Microscopy		Renal tubular epithelial (RTE)	1	0.14%
Microscopy		Transitional epithelial cell (urothelial)	2	0.29%
Microscopy		White blood cell (WBC) (Leukocyte)	3	0.43%
CM-20		* Denotes Correct Response		
AG: All Methods	687			
		*Ident Unknown - Referred	1	0.15%
		*Renal tubular epithelial (RTE)	281	40.90%
		*Transitional epithelial cell (urothelial)	305	44.40%
		Cellular (RTE) cast	1	0.15%
		Cellular (WBC) cast	1	0.15%
		Clue cell	1	0.15%
		Red blood cell (RBC) (Erythrocyte)	4	0.58%
		RTE with fat globules (oval fat bodies)	5	0.73%
		Squamous epithelial cell	79	11.50%
		White blood cell (WBC) (Leukocyte)	8	1.16%
		Yeast/fungi	1	0.15%
Microscopy	687	*Ident Unknown - Referred	1	0.15%
Microscopy		*Renal tubular epithelial (RTE)	281	40.90%
Microscopy		*Transitional epithelial cell (urothelial)	305	44.40%



Microscopy		Cellular (RTE) cast	1	0.15%
Microscopy		Cellular (WBC) cast	1	0.15%
Microscopy		Clue cell	1	0.15%
Microscopy		Red blood cell (RBC) (Erythrocyte)	4	0.58%
Microscopy		RTE with fat globules (oval fat bodies)	5	0.73%
Microscopy		Squamous epithelial cell	79	11.50%
Microscopy		White blood cell (WBC) (Leukocyte)	8	1.16%
Microscopy		Yeast/fungi	1	0.15%
CM-21		* Denotes Correct Response		
AG: All Methods	689			
		*Bacteria	672	97.53%
		Broad cast	1	0.15%
		Hyaline cast	1	0.15%
		RBC/Blood/Hgb cast	1	0.15%
		Red blood cell (RBC) (Erythrocyte)	1	0.15%
		Renal tubular epithelial (RTE)	2	0.29%
		RTE with fat globules (oval fat bodies)	1	0.15%
		White blood cell (WBC) (Leukocyte)	5	0.73%
		Yeast/fungi	5	0.73%
Microscopy	689	*Bacteria	672	97.53%
Microscopy		Broad cast	1	0.15%
Microscopy		Hyaline cast	1	0.15%
Microscopy		RBC/Blood/Hgb cast	1	0.15%
Microscopy		Red blood cell (RBC) (Erythrocyte)	1	0.15%
Microscopy		Renal tubular epithelial (RTE)	2	0.29%
Microscopy		RTE with fat globules (oval fat bodies)	1	0.15%
Microscopy		White blood cell (WBC) (Leukocyte)	5	0.73%
Microscopy		Yeast/fungi	5	0.73%
CM-22		* Denotes Correct Response		
AG: All Methods	690			
		*White blood cell (WBC) (Leukocyte)	678	98.26%
		Bacteria	4	0.58%
		Bacteria present	1	0.14%
		Cellular (WBC) cast	5	0.72%
		Fat droplets or globules	1	0.14%
		Renal tubular epithelial (RTE)	1	0.14%
Microscopy	690	*White blood cell (WBC) (Leukocyte)	678	98.26%
Microscopy		Bacteria	4	0.58%
Microscopy		Bacteria present	1	0.14%
Microscopy		Cellular (WBC) cast	5	0.72%
Microscopy		Fat droplets or globules	1	0.14%
Microscopy		Renal tubular epithelial (RTE)	1	0.14%
CM-23		* Denotes Correct Response		
AG: All Methods	947			
		*White Blood Cell (WBC) (Leukocyte)	870	91.87%
		Bacteria	1	0.11%
		Bacteria present	8	0.84%
		Clue cell	11	1.16%
		Pollen grain	3	0.32%
		Red blood cell (RBC) (Erythrocyte)	1	0.11%
		Red Blood Cell (RBC) (Erythrocyte)	36	3.80%
		Squamous epithelial cell	4	0.42%
		Starch granule	2	0.21%
		Trichomonas vaginalis	4	0.42%
		Yeast/fungi present	6	0.63%
		Yeast/fungi present	1	0.11%
Microscopy	947	*White Blood Cell (WBC) (Leukocyte)	870	91.87%
Microscopy		Bacteria	1	0.11%
Microscopy		Bacteria present	8	0.84%
Microscopy		Clue cell	11	1.16%
Microscopy		Pollen grain	3	0.32%
Microscopy		Red blood cell (RBC) (Erythrocyte)	1	0.11%
Microscopy		Red Blood Cell (RBC) (Erythrocyte)	36	3.80%
Microscopy		Squamous epithelial cell	4	0.42%
Microscopy		Starch granule	2	0.21%
Microscopy		Trichomonas vaginalis	4	0.42%
Microscopy		Yeast/fungi present	6	0.63%
Microscopy		Yeast/fungi present	1	0.11%



CM-24		* Denotes Correct Response		
AG: All Methods	947			
		*Clue cell	889	93.88%
		Bacteria present	6	0.63%
		Mucous strand	3	0.32%
		Red Blood Cell (RBC) (Erythrocyte)	1	0.11%
		Squamous epithelial cell	39	4.12%
		Trichomonas vaginalis	1	0.11%
		White Blood Cell (WBC) (Leukocyte)	5	0.53%
		Yeast/fungi present	2	0.21%
		Yeast/fungi present	1	0.11%
Microscopy	947	*Clue cell	889	93.88%
Microscopy		Bacteria present	6	0.63%
Microscopy		Mucous strand	3	0.32%
Microscopy		Red Blood Cell (RBC) (Erythrocyte)	1	0.11%
Microscopy		Squamous epithelial cell	39	4.12%
Microscopy		Trichomonas vaginalis	1	0.11%
Microscopy		White Blood Cell (WBC) (Leukocyte)	5	0.53%
Microscopy		Yeast/fungi present	2	0.21%
Microscopy		Yeast/fungi present	1	0.11%
CM-25		* Denotes Correct Response		
AG: All Methods	612			
		*No scabies present	7	1.14%
		*Yeast/fungi present	598	97.71%
		Yeast/fungi absent	7	1.14%
Microscopy	612	*No scabies present	7	1.14%
Microscopy		*Yeast/fungi present	598	97.71%
Microscopy		Yeast/fungi absent	7	1.14%
CM-26		* Denotes Correct Response		
AG: All Methods	415			
		*Negative Fern Test	406	97.83%
		Bacteria absent	1	0.24%
		Positive Fern Test	3	0.72%
		Red Blood Cell (RBC) (Erythrocyte)	1	0.24%
		Spermatozoa absent	1	0.24%
		Squamous epithelial cell	3	0.72%
Microscopy	415	*Negative Fern Test	406	97.83%
Microscopy		Bacteria absent	1	0.24%
Microscopy		Positive Fern Test	3	0.72%
Microscopy		Red Blood Cell (RBC) (Erythrocyte)	1	0.24%
Microscopy		Spermatozoa absent	1	0.24%
Microscopy		Squamous epithelial cell	3	0.72%
683 - ASO/CRP;684 - CRP (Quant & Qual)	Labs	Response	Number	Percent
		CRP Qualitative		
794 - GC & Chlamydia Antigen Detection (5 specimens);795 - GC & Chlamydia Antigen Detection (2 specimens);774 - GenProbe GC/Chlamydia	Labs	Response	Number	Percent
		Chlamydia - Non Culture Tech		
CY-11		* Denotes Correct Response		
PG: BD Probe Tec ET	1			
		*Positive	1	100.00%
CY-12		* Denotes Correct Response		
PG: BD Probe Tec ET	1			
		*Positive	1	100.00%
CY-13		* Denotes Correct Response		
PG: BD Probe Tec ET	1			
		*Negative	1	100.00%
CY-14		* Denotes Correct Response		
PG: BD Probe Tec ET	1			
		*Positive	1	100.00%
CY-15		* Denotes Correct Response		
PG: BD Probe Tec ET	1			
		*Negative	1	100.00%
CY-11		* Denotes Correct Response		
PG: Gen-Probe	1			
		*Positive	1	100.00%



CY-12		* Denotes Correct Response		
PG: Gen-Probe	1			
		*Positive	1	100.00%
CY-13		* Denotes Correct Response		
PG: Gen-Probe	1			
		*Negative	1	100.00%
CY-14		* Denotes Correct Response		
PG: Gen-Probe	1			
		*Positive	1	100.00%
CY-15		* Denotes Correct Response		
PG: Gen-Probe	1			
		*Negative	1	100.00%
CY-11		* Denotes Correct Response		
PG: Quidel QuickVue Chlamydia	2			
		*Positive	2	100.00%
CY-13		* Denotes Correct Response		
PG: Quidel QuickVue Chlamydia	2			
		*Negative	2	100.00%
CY-14		* Denotes Correct Response		
PG: Quidel QuickVue Chlamydia	3			
		*Positive	1	33.33%
		Negative	2	66.67%
CY-15		* Denotes Correct Response		
PG: Quidel QuickVue Chlamydia	2			
		*Negative	2	100.00%
CY-11		* Denotes Correct Response		
PG: Cepheid GeneXpert	4			
		*Positive	4	100.00%
CY-12		* Denotes Correct Response		
PG: Cepheid GeneXpert	5			
		*Positive	5	100.00%
CY-13		* Denotes Correct Response		
PG: Cepheid GeneXpert	4			
		*Negative	4	100.00%
CY-14		* Denotes Correct Response		
PG: Cepheid GeneXpert	5			
		*Positive	4	80.00%
		Negative	1	20.00%
CY-15		* Denotes Correct Response		
PG: Cepheid GeneXpert	4			
		*Negative	4	100.00%
CY-11		* Denotes Correct Response		
PG: BD Viper XTR CT/GC	1			
		*Positive	1	100.00%
CY-12		* Denotes Correct Response		
PG: BD Viper XTR CT/GC	1			
		*Positive	1	100.00%
CY-13		* Denotes Correct Response		
PG: BD Viper XTR CT/GC	1			
		*Negative	1	100.00%
CY-14		* Denotes Correct Response		
PG: BD Viper XTR CT/GC	1			
		*Positive	1	100.00%
CY-15		* Denotes Correct Response		
PG: BD Viper XTR CT/GC	1			
		*Negative	1	100.00%
CY-12		* Denotes Correct Response		
PG: BD Viper XTR CT/GC	11			
		*Positive	9	81.82%
		Negative	2	18.18%
BD Probe Tec ET	1	*Positive	1	100.00%
Gen-Probe	1	*Positive	1	100.00%
Quidel QuickVue Chlamydia	3	*Positive	1	33.33%
Cepheid GeneXpert	5	*Positive	5	100.00%
BD Viper XTR CT/GC	1	*Positive	1	100.00%
Quidel QuickVue Chlamydia	3	Negative	2	66.67%
		GC - Non Culture Tech		



CY-11		* Denotes Correct Response			
AG: All Methods	7				
		*Negative	7	100.00%	
BD Probe Tec ET	1	*Negative	1	100.00%	
Gen-Probe	1	*Negative	1	100.00%	
Cepheid GeneXpert	4	*Negative	4	100.00%	
BD Viper XTR CT/GC	1	*Negative	1	100.00%	
CY-12		* Denotes Correct Response			
AG: All Methods	7				
		*Positive	7	100.00%	
BD Probe Tec ET	1	*Positive	1	100.00%	
Gen-Probe	1	*Positive	1	100.00%	
Cepheid GeneXpert	4	*Positive	4	100.00%	
BD Viper XTR CT/GC	1	*Positive	1	100.00%	
CY-13		* Denotes Correct Response			
AG: All Methods	7				
		*Positive	7	100.00%	
BD Probe Tec ET	1	*Positive	1	100.00%	
Gen-Probe	1	*Positive	1	100.00%	
Cepheid GeneXpert	4	*Positive	4	100.00%	
BD Viper XTR CT/GC	1	*Positive	1	100.00%	
CY-14		* Denotes Correct Response			
AG: All Methods	7				
		*Positive	6	85.71%	
		Negative	1	14.29%	
BD Probe Tec ET	1	*Positive	1	100.00%	
Gen-Probe	1	*Positive	1	100.00%	
Cepheid GeneXpert	4	*Positive	3	75.00%	
BD Viper XTR CT/GC	1	*Positive	1	100.00%	
Cepheid GeneXpert	4	Negative	1	25.00%	
CY-15		* Denotes Correct Response			
AG: All Methods	7				
		*Negative	7	100.00%	
BD Probe Tec ET	1	*Negative	1	100.00%	
Gen-Probe	1	*Negative	1	100.00%	
Cepheid GeneXpert	4	*Negative	4	100.00%	
BD Viper XTR CT/GC	1	*Negative	1	100.00%	
772 - Dermatophyte Screen/Culture		Labs	Response	Number	Percent
DM-11		* Denotes Correct Response			
AG: All Methods	15				
		*Positive, Dermatophytes Present	14	93.33%	
		Negative, No Dermatophytes Present	1	6.67%	
Dermatophytes	15	*Positive, Dermatophytes Present	14	93.33%	
Dermatophytes		Negative, No Dermatophytes Present	1	6.67%	
DM-12		* Denotes Correct Response			
AG: All Methods	15				
		*Negative, No Dermatophytes Present	5	33.33%	
		Positive, Dermatophytes Present	10	66.67%	
Dermatophytes	15	*Negative, No Dermatophytes Present	5	33.33%	
Dermatophytes		Positive, Dermatophytes Present	10	66.67%	
DM-13		* Denotes Correct Response			
AG: All Methods	15				
		*Positive, Dermatophytes Present	12	80.00%	
		Negative, No Dermatophytes Present	3	20.00%	
Dermatophytes	15	*Positive, Dermatophytes Present	12	80.00%	
Dermatophytes		Negative, No Dermatophytes Present	3	20.00%	
DM-14		* Denotes Correct Response			
AG: All Methods	15				
		*Negative, No Dermatophytes Present	15	100.00%	
Dermatophytes	15	*Negative, No Dermatophytes Present	15	100.00%	
DM-15		* Denotes Correct Response			
AG: All Methods	15				
		*Positive, Dermatophytes Present	14	93.33%	
		Negative, No Dermatophytes Present	1	6.67%	
Dermatophytes	15	*Positive, Dermatophytes Present	14	93.33%	
Dermatophytes		Negative, No Dermatophytes Present	1	6.67%	
777 - Genital Culture; 796 - Genital Culture (2 specimens)		Labs	Response	Number	Percent



GC-11		Genital Culture			
AG: All Methods		* Denotes Correct Response			
	3				
		*Presumptive positive for GC, GR(-) diplococci, oxidase positive organism	2	66.67%	
		*Presumptive positive for GC, growth referred for ID/confirmation	1	33.33%	
Cultures	3	*Presumptive positive for GC, GR(-) diplococci, oxidase positive organism	2	66.67%	
Cultures		*Presumptive positive for GC, growth referred for ID/confirmation	1	33.33%	
GC-12		* Denotes Correct Response			
AG: All Methods					
	3	*No growth, GC selective media	3	100.00%	
Cultures	3	*No growth, GC selective media	3	100.00%	
GC-13		* Denotes Correct Response			
AG: All Methods					
	3	*No growth, GC selective media	3	100.00%	
Cultures	3	*No growth, GC selective media	3	100.00%	
GC-14		* Denotes Correct Response			
AG: All Methods					
	6	*Group B Streptococcus not present	1	16.67%	
		*Growth, referred for identification	1	16.67%	
		*Negative for GC	2	33.33%	
		*No growth, GC selective media	2	33.33%	
Cultures	6	*Group B Streptococcus not present	1	16.67%	
Cultures		*Growth, referred for identification	1	16.67%	
Cultures		*Negative for GC	2	33.33%	
Cultures		*No growth, GC selective media	2	33.33%	
GC-15		* Denotes Correct Response			
AG: All Methods					
	4	*Negative for GC	1	25.00%	
		*No growth, GC selective media	3	75.00%	
Cultures	4	*Negative for GC	1	25.00%	
Cultures		*No growth, GC selective media	3	75.00%	
668 - Gram Stain		Labs	Response	Number	Percent
GS-11		* Denotes Correct Response			
AG: All Methods					
	7				
		*Gram Negative	7	100.00%	
Gram Stain	7	*Gram Negative	7	100.00%	
GS-12		* Denotes Correct Response			
AG: All Methods					
	8				
		*Gram Negative	8	100.00%	
Gram Stain	8	*Gram Negative	8	100.00%	
GS-13		* Denotes Correct Response			
AG: All Methods					
	8				
		*Gram Positive	8	100.00%	
Gram Stain	8	*Gram Positive	8	100.00%	
GS-14		* Denotes Correct Response			
AG: All Methods					
	8				
		*Gram Negative	8	100.00%	
Gram Stain	8	*Gram Negative	8	100.00%	
GS-15		* Denotes Correct Response			
AG: All Methods					
	8				
		*Gram Positive	8	100.00%	
Gram Stain	8	*Gram Positive	8	100.00%	
GS-11		* Denotes Correct Response			
AG: All Methods					
	6				
		*Cocci	4	66.67%	
		Cocco-Bacilli	2	33.33%	
Gram Stain	6	*Cocci	4	66.67%	
Gram Stain		Cocco-Bacilli	2	33.33%	
GS-12		* Denotes Correct Response			
AG: All Methods					
	7				
		*Rods/Bacilli	6	85.71%	
		Cocco-Bacilli	1	14.29%	
Gram Stain	7	*Rods/Bacilli	6	85.71%	



Gram Stain		Cocco-Bacilli	1	14.29%
GS-13		* Denotes Correct Response		
AG: All Methods	7			
		*Cocci	4	57.14%
		*Cocci in Pairs	3	42.86%
Gram Stain	7	*Cocci	4	57.14%
Gram Stain		*Cocci in Pairs	3	42.86%
GS-14		* Denotes Correct Response		
AG: All Methods	7			
		*Rods/Bacilli	7	100.00%
Gram Stain	7	*Rods/Bacilli	7	100.00%
GS-15		* Denotes Correct Response		
AG: All Methods	7			
		*Cocci	5	71.43%
		*Cocci in Chains	1	14.29%
		*Cocci in Pairs	1	14.29%
Gram Stain	7	*Cocci	5	71.43%
Gram Stain		*Cocci in Chains	1	14.29%
Gram Stain		*Cocci in Pairs	1	14.29%
678 - Serum hCG;759 - Complete Immunology;729 - AFC Immunology	Labs	Response	Number	Percent
		HCG - Qual Serum		
HCG-11		* Denotes Correct Response		
AG: All Methods	92			
		*Negative	92	100.00%
Beck Coulter ICON 25	6	*Negative	6	100.00%
Fisher HC Sure-Vue hCG STAT	2	*Negative	2	100.00%
Sekisui Diagnostics OSOM	3	*Negative	3	100.00%
Henry Schein One Step + (hCG)	5	*Negative	5	100.00%
Immunostics Immuno/hCG	2	*Negative	2	100.00%
Inverness Signify HCG	2	*Negative	2	100.00%
LifeSign Status hCG	3	*Negative	3	100.00%
McKesson Medi-Lab hCG - Urine	1	*Negative	1	100.00%
Quidel QuickVue + hCG	2	*Negative	2	100.00%
Quidel QuickVue One-Step hCG Combo (U/S)	32	*Negative	32	100.00%
Select Diag hCG Combo	3	*Negative	3	100.00%
SP Brand (Cardinal) HCG Combo	9	*Negative	9	100.00%
Wampole Clearview	1	*Negative	1	100.00%
Polymedco Poly stat hCG Serum	1	*Negative	1	100.00%
Stanbio True 20 Plus	1	*Negative	1	100.00%
Consult Diag hCG Combo Cassette	12	*Negative	12	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	1	*Negative	1	100.00%
Medline Pregnancy Combo (serum)	1	*Negative	1	100.00%
Alere hCG Combo Cassette (20/10 mIU/mL)	1	*Negative	1	100.00%
McKesson Medi-Lab hCG - Serum	4	*Negative	4	100.00%
HCG-12		* Denotes Correct Response		
AG: All Methods	92			
		*Positive	92	100.00%
Beck Coulter ICON 25	6	*Positive	6	100.00%
Fisher HC Sure-Vue hCG STAT	2	*Positive	2	100.00%
Sekisui Diagnostics OSOM	3	*Positive	3	100.00%
Henry Schein One Step + (hCG)	5	*Positive	5	100.00%
Immunostics Immuno/hCG	2	*Positive	2	100.00%
Inverness Signify HCG	2	*Positive	2	100.00%
LifeSign Status hCG	3	*Positive	3	100.00%
McKesson Medi-Lab hCG - Urine	1	*Positive	1	100.00%
Quidel QuickVue + hCG	2	*Positive	2	100.00%
Quidel QuickVue One-Step hCG Combo (U/S)	32	*Positive	32	100.00%
Select Diag hCG Combo	3	*Positive	3	100.00%
SP Brand (Cardinal) HCG Combo	9	*Positive	9	100.00%
Wampole Clearview	1	*Positive	1	100.00%
Polymedco Poly stat hCG Serum	1	*Positive	1	100.00%
Stanbio True 20 Plus	1	*Positive	1	100.00%
Consult Diag hCG Combo Cassette	12	*Positive	12	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	1	*Positive	1	100.00%
Medline Pregnancy Combo (serum)	1	*Positive	1	100.00%
Alere hCG Combo Cassette (20/10 mIU/mL)	1	*Positive	1	100.00%
McKesson Medi-Lab hCG - Serum	4	*Positive	4	100.00%
HCG-13		* Denotes Correct Response		



AG: All Methods	92			
		*Positive	92	100.00%
Beck Coulter ICON 25	6	*Positive	6	100.00%
Fisher HC Sure-Vue hCG STAT	2	*Positive	2	100.00%
Sekisui Diagnostics OSOM	3	*Positive	3	100.00%
Henry Schein One Step + (hCG)	5	*Positive	5	100.00%
Immunostics Immuno/hCG	2	*Positive	2	100.00%
Inverness Signify HCG	2	*Positive	2	100.00%
LifeSign Status hCG	3	*Positive	3	100.00%
McKesson Medi-Lab hCG - Urine	1	*Positive	1	100.00%
Quidel QuickVue + hCG	2	*Positive	2	100.00%
Quidel QuickVue One-Step hCG Combo (U/S)	32	*Positive	32	100.00%
Select Diag hCG Combo	3	*Positive	3	100.00%
SP Brand (Cardinal) HCG Combo	9	*Positive	9	100.00%
Wampole Clearview	1	*Positive	1	100.00%
Polymedco Poly stat hCG Serum	1	*Positive	1	100.00%
Stanbio True 20 Plus	1	*Positive	1	100.00%
Consult Diag hCG Combo Cassette	12	*Positive	12	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	1	*Positive	1	100.00%
Medline Pregnancy Combo (serum)	1	*Positive	1	100.00%
Alere hCG Combo Cassette (20/10 mIU/mL)	1	*Positive	1	100.00%
McKesson Medi-Lab hCG - Serum	4	*Positive	4	100.00%
HCG-14		* Denotes Correct Response		
AG: All Methods	92			
		*Positive	91	98.91%
		Negative	1	1.09%
Beck Coulter ICON 25	6	*Positive	6	100.00%
Fisher HC Sure-Vue hCG STAT	2	*Positive	2	100.00%
Sekisui Diagnostics OSOM	3	*Positive	3	100.00%
Henry Schein One Step + (hCG)	5	*Positive	5	100.00%
Immunostics Immuno/hCG	2	*Positive	2	100.00%
Inverness Signify HCG	2	*Positive	2	100.00%
LifeSign Status hCG	3	*Positive	3	100.00%
McKesson Medi-Lab hCG - Urine	1	*Positive	1	100.00%
Quidel QuickVue + hCG	2	*Positive	2	100.00%
Quidel QuickVue One-Step hCG Combo (U/S)	32	*Positive	32	100.00%
Select Diag hCG Combo	3	*Positive	3	100.00%
SP Brand (Cardinal) HCG Combo	9	*Positive	8	88.89%
Wampole Clearview	1	*Positive	1	100.00%
Polymedco Poly stat hCG Serum	1	*Positive	1	100.00%
Stanbio True 20 Plus	1	*Positive	1	100.00%
Consult Diag hCG Combo Cassette	12	*Positive	12	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	1	*Positive	1	100.00%
Medline Pregnancy Combo (serum)	1	*Positive	1	100.00%
Alere hCG Combo Cassette (20/10 mIU/mL)	1	*Positive	1	100.00%
McKesson Medi-Lab hCG - Serum	4	*Positive	4	100.00%
SP Brand (Cardinal) HCG Combo	9	Negative	1	11.11%
HCG-15		* Denotes Correct Response		
AG: All Methods	92			
		*Positive	90	97.83%
		Negative	2	2.17%
Beck Coulter ICON 25	6	*Positive	6	100.00%
Fisher HC Sure-Vue hCG STAT	2	*Positive	2	100.00%
Sekisui Diagnostics OSOM	3	*Positive	3	100.00%
Henry Schein One Step + (hCG)	5	*Positive	5	100.00%
Immunostics Immuno/hCG	2	*Positive	2	100.00%
Inverness Signify HCG	2	*Positive	2	100.00%
LifeSign Status hCG	3	*Positive	3	100.00%
McKesson Medi-Lab hCG - Urine	1	*Positive	1	100.00%
Quidel QuickVue + hCG	2	*Positive	2	100.00%
Quidel QuickVue One-Step hCG Combo (U/S)	32	*Positive	31	96.88%
Select Diag hCG Combo	3	*Positive	3	100.00%
SP Brand (Cardinal) HCG Combo	9	*Positive	9	100.00%
Polymedco Poly stat hCG Serum	1	*Positive	1	100.00%
Stanbio True 20 Plus	1	*Positive	1	100.00%
Consult Diag hCG Combo Cassette	12	*Positive	12	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	1	*Positive	1	100.00%
Medline Pregnancy Combo (serum)	1	*Positive	1	100.00%
Alere hCG Combo Cassette (20/10 mIU/mL)	1	*Positive	1	100.00%



McKesson Medi-Lab hCG - Serum	4	*Positive	4	100.00%
Quidel QuickVue One-Step hCG Combo (U/S)	32	Negative	1	3.13%
Wampole Clearview	1	Negative	1	100.00%
		HCG - Qual Serum (International)		
HCG-11		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Henry Schein One Step + (hCG)	1	*Negative	1	100.00%
SAS Ultra HCG	2	*Negative	2	100.00%
HCG-12		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Henry Schein One Step + (hCG)	1	*Positive	1	100.00%
SAS Ultra HCG	2	*Positive	2	100.00%
HCG-13		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Henry Schein One Step + (hCG)	1	*Positive	1	100.00%
SAS Ultra HCG	2	*Positive	2	100.00%
HCG-14		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Henry Schein One Step + (hCG)	1	*Positive	1	100.00%
SAS Ultra HCG	2	*Positive	2	100.00%
HCG-15		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Henry Schein One Step + (hCG)	1	*Positive	1	100.00%
SAS Ultra HCG	2	*Positive	2	100.00%
762 - HIV Antibody (Waived);766 - HIV Antibody (5 specimens);750 - Viral Markers	Labs	Response	Number	Percent
		HIV Antibody (Waived)		
HIV-11		* Denotes Correct Response		
AG: All Methods	81			
		*Negative	80	98.77%
		Positive	1	1.23%
Alere Clearview Complete HIV 1/2 (mod)	9	*Negative	9	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	49	*Negative	48	97.96%
Orasure OraQuick Anti-HIV	9	*Negative	9	100.00%
Trinity Biotech Uni-Gold Recombigen HIV	3	*Negative	3	100.00%
Chembio HIV 1/2 STAT-PAK Dipstick	1	*Negative	1	100.00%
Chembio HIV 1/2 STAT-PAK Cassette	8	*Negative	8	100.00%
bioLytical INSTI HIV Antibody	2	*Negative	2	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	49	Positive	1	2.04%
		HIV Antibody (International)		
HIV-11		* Denotes Correct Response		
AG: All Methods	4			
		*Negative	4	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
Intl Chembio HIV 1/2 STAT-PAK Dipstick	1	*Negative	1	100.00%
		HIV Antibody		
HIV-11		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Siemens Centaur	1	*Negative	1	100.00%
Alere Clearview Complete HIV 1/2 (mod)	2	*Negative	2	100.00%
766 - HIV Antibody (5 specimens);750 - Viral Markers	Labs	Response	Number	Percent
		HIV Antibody		
HIV-11		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Siemens Centaur	1	*Negative	1	100.00%
Alere Clearview Complete HIV 1/2 (mod)	2	*Negative	2	100.00%
HIV-12		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Siemens Centaur	1	*Positive	1	100.00%
Alere Clearview Complete HIV 1/2 (mod)	2	*Positive	2	100.00%



HIV-13		* Denotes Correct Response	
AG: All Methods	3		
		*Positive	3 100.00%
Siemens Centaur	1	*Positive	1 100.00%
Alere Clearview Complete HIV 1/2 (mod)	2	*Positive	2 100.00%
HIV-14		* Denotes Correct Response	
AG: All Methods	3		
		*Negative	3 100.00%
Siemens Centaur	1	*Negative	1 100.00%
Alere Clearview Complete HIV 1/2 (mod)	2	*Negative	2 100.00%
HIV-15		* Denotes Correct Response	
AG: All Methods	3		
		*Positive	3 100.00%
Siemens Centaur	1	*Positive	1 100.00%
Alere Clearview Complete HIV 1/2 (mod)	2	*Positive	2 100.00%
HIV-11		HIV Antibody (Waived)	
HIV-11		* Denotes Correct Response	
AG: All Methods	81		
		*Negative	80 98.77%
		Positive	1 1.23%
Alere Clearview Complete HIV 1/2 (mod)	9	*Negative	9 100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	49	*Negative	48 97.96%
Orasure OraQuick Anti-HIV	9	*Negative	9 100.00%
Trinity Biotech Uni-Gold Recombigen HIV	3	*Negative	3 100.00%
Chembio HIV 1/2 STAT-PAK Dipstick	1	*Negative	1 100.00%
Chembio HIV 1/2 STAT-PAK Cassette	8	*Negative	8 100.00%
bioLytical INSTI HIV Antibody	2	*Negative	2 100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	49	Positive	1 2.04%
HIV-12		* Denotes Correct Response	
AG: All Methods	7		
		*Positive	7 100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	4	*Positive	4 100.00%
Orasure OraQuick Anti-HIV	1	*Positive	1 100.00%
Trinity Biotech Uni-Gold Recombigen HIV	1	*Positive	1 100.00%
bioLytical INSTI HIV Antibody	1	*Positive	1 100.00%
HIV-13		* Denotes Correct Response	
AG: All Methods	7		
		*Positive	7 100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	4	*Positive	4 100.00%
Orasure OraQuick Anti-HIV	1	*Positive	1 100.00%
Trinity Biotech Uni-Gold Recombigen HIV	1	*Positive	1 100.00%
bioLytical INSTI HIV Antibody	1	*Positive	1 100.00%
HIV-14		* Denotes Correct Response	
AG: All Methods	7		
		*Negative	7 100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	4	*Negative	4 100.00%
Orasure OraQuick Anti-HIV	1	*Negative	1 100.00%
Trinity Biotech Uni-Gold Recombigen HIV	1	*Negative	1 100.00%
bioLytical INSTI HIV Antibody	1	*Negative	1 100.00%
HIV-15		* Denotes Correct Response	
AG: All Methods	7		
		*Positive	7 100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	4	*Positive	4 100.00%
Orasure OraQuick Anti-HIV	1	*Positive	1 100.00%
Trinity Biotech Uni-Gold Recombigen HIV	1	*Positive	1 100.00%
bioLytical INSTI HIV Antibody	1	*Positive	1 100.00%
HIV-11		HIV Antibody (International)	
HIV-11		* Denotes Correct Response	
AG: All Methods	4		
		*Negative	4 100.00%
Intl OCD Vitros ECI/ECiQ	3	*Negative	3 100.00%
Intl Chembio HIV 1/2 STAT-PAK Dipstick	1	*Negative	1 100.00%
HIV-12		* Denotes Correct Response	
AG: All Methods	3		
		*Positive	3 100.00%
Intl OCD Vitros ECI/ECiQ	3	*Positive	3 100.00%
HIV-13		* Denotes Correct Response	
AG: All Methods	3		



		*Positive	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Positive	3	100.00%
HIV-14		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
HIV-15		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Positive	3	100.00%
644 - H. pylori;628 - AFC Waived/PPM	Labs	Response	Number	Percent
		H. pylori Ab Detect		
HP-3		* Denotes Correct Response		
AG: All Methods	79			
		*Negative	75	94.94%
		Positive	4	5.06%
Beck Coulter FlexSure	3	*Negative	3	100.00%
Beck Coulter ICON HP	2	*Negative	2	100.00%
Instant Technologies i-Screen	1	*Negative	1	100.00%
McKesson Medi-Lab Perf H. pylori	8	*Negative	7	87.50%
Polymedco Poly stat (WB - Waived)	3	*Negative	3	100.00%
Henry Schein One Step+ (H. pylori)	5	*Negative	5	100.00%
Select Diag H. pylori Cassette	3	*Negative	3	100.00%
Quidel QuickVue H. pylori gII (S/P)	4	*Negative	4	100.00%
Quidel QuickVue H. pylori gII (WB)	11	*Negative	11	100.00%
SP Brand (Cardinal) HP Rapid Test	10	*Negative	9	90.00%
Alere Clearview II S/P (mod)	4	*Negative	4	100.00%
Alere Clearview II WB (waived)	4	*Negative	4	100.00%
LifeSign Status H. pylori	2	*Negative	2	100.00%
Consult Diag H pylori	17	*Negative	16	94.12%
Jant Accutest H pylori	1	*Negative	1	100.00%
McKesson Medi-Lab Perf H. pylori	8	Positive	1	12.50%
SP Brand (Cardinal) HP Rapid Test	10	Positive	1	10.00%
Instant-View H. pylori	1	Positive	1	100.00%
Consult Diag H pylori	17	Positive	1	5.88%
933 - HIV Antibody/Antigen Combo (2 specimens);934 - HIV Antibody/Antigen Combo (5 specimens)	Labs	Response	Number	Percent
		HIV Antibody (Waived)		
HVAA-11		* Denotes Correct Response		
AG: All Methods	27			
		*Negative	26	96.30%
		Positive	1	3.70%
Alere Determine HIV Ag/Ab	26	*Negative	25	96.15%
Chembio HIV 1/2 STAT-PAK Dipstick	1	*Negative	1	100.00%
Alere Determine HIV Ag/Ab	26	Positive	1	3.85%
HVAA-12		* Denotes Correct Response		
AG: All Methods	27			
		*Positive	26	96.30%
		Negative	1	3.70%
Alere Determine HIV Ag/Ab	26	*Positive	25	96.15%
Chembio HIV 1/2 STAT-PAK Dipstick	1	*Positive	1	100.00%
Alere Determine HIV Ag/Ab	26	Negative	1	3.85%
		HIV Antigen (p24) (Waived)		
HVAA-11		* Denotes Correct Response		
AG: All Methods	26			
		*Negative	26	100.00%
Alere Determine HIV Ag/Ab	26	*Negative	26	100.00%
HVAA-12		* Denotes Correct Response		
AG: All Methods	26			
		*Negative	25	96.15%
		Positive	1	3.85%
Alere Determine HIV Ag/Ab	26	*Negative	25	96.15%
Alere Determine HIV Ag/Ab		Positive	1	3.85%
603 - Blood Cell Identification;610 - Regular Hematology;749 - Auto Diff II w-Blood Cell Identification	Labs	Response	Number	Percent
		Blood Cell Identification		
ID-13		* Denotes Correct Response		
AG: All Methods	66			



		*Howell-Jolly body (Wright stain)	66	100.00%
Blood Cell Microscopy	66	*Howell-Jolly body (Wright stain)	66	100.00%
ID-14		* Denotes Correct Response		
AG: All Methods	66			
		*Target cell (codocyte)	64	96.97%
		Hemoglobin C Crystal	1	1.52%
		Other, not listed	1	1.52%
Blood Cell Microscopy	66	*Target cell (codocyte)	64	96.97%
Blood Cell Microscopy		Hemoglobin C Crystal	1	1.52%
Blood Cell Microscopy		Other, not listed	1	1.52%
ID-15		* Denotes Correct Response		
AG: All Methods	66			
		*Metarubricyte	1	1.52%
		*Nucleated red cell	64	96.97%
		Poikilocytosis	1	1.52%
Blood Cell Microscopy	66	*Metarubricyte	1	1.52%
Blood Cell Microscopy		*Nucleated red cell	64	96.97%
Blood Cell Microscopy		Poikilocytosis	1	1.52%
ID-16		* Denotes Correct Response		
AG: All Methods	66			
		*Platelet, normal	66	100.00%
Blood Cell Microscopy	66	*Platelet, normal	66	100.00%
ID-17		* Denotes Correct Response		
AG: All Methods	66			
		*Hypersegmented neutrophil	2	3.03%
		*Segmented or band neutrophil	64	96.97%
Blood Cell Microscopy	66	*Hypersegmented neutrophil	2	3.03%
Blood Cell Microscopy		*Segmented or band neutrophil	64	96.97%
ID-18		* Denotes Correct Response		
AG: All Methods	66			
		*Acanthocyte (spur cell)	37	56.06%
		*Echinocyte (burr cell, crenated cell)	23	34.85%
		Fragmented cell (schist., helmet cell, keratocyte)	2	3.03%
		Heinz body (supravital stain)	1	1.52%
		Other, not listed	1	1.52%
		Poikilocytosis	2	3.03%
Blood Cell Microscopy	66	*Acanthocyte (spur cell)	37	56.06%
Blood Cell Microscopy		*Echinocyte (burr cell, crenated cell)	23	34.85%
Blood Cell Microscopy		Fragmented cell (schist., helmet cell, keratocyte)	2	3.03%
Blood Cell Microscopy		Heinz body (supravital stain)	1	1.52%
Blood Cell Microscopy		Other, not listed	1	1.52%
Blood Cell Microscopy		Poikilocytosis	2	3.03%
Blood Cell Identification (International)				
ID-13		* Denotes Correct Response		
AG: All Methods	1			
		*Howell-Jolly body (Wright stain)	1	100.00%
Intl Blood Cell Microscopy	1	*Howell-Jolly body (Wright stain)	1	100.00%
ID-14		* Denotes Correct Response		
AG: All Methods	1			
		*Target cell (codocyte)	1	100.00%
Intl Blood Cell Microscopy	1	*Target cell (codocyte)	1	100.00%
ID-15		* Denotes Correct Response		
AG: All Methods	1			
		*Nucleated red cell	1	100.00%
Intl Blood Cell Microscopy	1	*Nucleated red cell	1	100.00%
ID-16		* Denotes Correct Response		
AG: All Methods	1			
		*Platelet, normal	1	100.00%
Intl Blood Cell Microscopy	1	*Platelet, normal	1	100.00%
ID-17		* Denotes Correct Response		
AG: All Methods	1			
		*Segmented or band neutrophil	1	100.00%
Intl Blood Cell Microscopy	1	*Segmented or band neutrophil	1	100.00%
ID-18		* Denotes Correct Response		
AG: All Methods	1			
		*Acanthocyte (spur cell)	1	100.00%
Intl Blood Cell Microscopy	1	*Acanthocyte (spur cell)	1	100.00%



609 - Waived/PPM Tests I;675 - Infectious Mononucleosis;753 - Infectious Mononucleosis (Waived) (1 specimens);754 - Infectious Mono/RF Combo;759 - Complete Immunology;628 - AFC Waived/PPM;729 - AFC Immunology	Labs	Response	Number	Percent
		Infectious Mono (Waived)		
		* Denotes Correct Response		
IM-11				
AG: All Methods	466			
		*Positive	456	97.85%
		Negative	10	2.15%
Alere Acceava (S)	1	*Positive	1	100.00%
Alere Acceava (WB)	58	*Positive	58	100.00%
Sekisui Diagnostics OSOM Mono (WB)	182	*Positive	179	98.35%
Henry Schein One-Step+ Mono	30	*Positive	30	100.00%
Instant Technologies i-Screen (WB)	1	*Positive	1	100.00%
Inverness Signify Mono	3	*Positive	3	100.00%
LifeSign Status Inf Mono	31	*Positive	30	96.77%
McKesson Medi-Lab Perf - Inf Mono	12	*Positive	12	100.00%
Polymedco Poly stat Mono - (WB)	6	*Positive	5	83.33%
Quidel QuickVue+ (WB)	15	*Positive	14	93.33%
Select Diag Mono Cas (WB)	1	*Positive	1	100.00%
SP Brand (Cardinal) Rapid Test	9	*Positive	9	100.00%
Alere Clearview Mono Cassette	26	*Positive	25	96.15%
Diag Test Group Clarity Mono	3	*Positive	3	100.00%
Fisher HC Sure-View Mono	4	*Positive	4	100.00%
Immunostics Inc. ImmunoDetector Mono (WB)	2	*Positive	2	100.00%
Meridian Immunocard STAT! Mono (WB)	4	*Positive	4	100.00%
NDC ProAdvantage Mono	2	*Positive	2	100.00%
Consult Diag Mono Cassette	67	*Positive	65	97.01%
Alere Acceava (WB) Mono Cassette	7	*Positive	7	100.00%
Medline Mono Cassette (WB)	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	182	Negative	3	1.65%
LifeSign Status Inf Mono	31	Negative	1	3.23%
Polymedco Poly stat Mono - (WB)	6	Negative	1	16.67%
Quidel QuickVue+ (WB)	15	Negative	1	6.67%
Alere Clearview Mono Cassette	26	Negative	1	3.85%
Pulse Scientific	1	Negative	1	100.00%
Consult Diag Mono Cassette	67	Negative	2	2.99%
		Infectious Mononucleosis		
		* Denotes Correct Response		
IM-11				
AG: All Methods	78			
		*Positive	78	100.00%
Beck Coulter ICON (WB)	1	*Positive	1	100.00%
Alere Acceava (WB)	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	35	*Positive	35	100.00%
Henry Schein One-Step+ Mono	1	*Positive	1	100.00%
LifeSign Status Inf Mono	2	*Positive	2	100.00%
Lifesign Unistep	1	*Positive	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	3	*Positive	3	100.00%
Polymedco Poly stat Mono - (WB)	1	*Positive	1	100.00%
Polymedco Poly stat Mono (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (WB)	9	*Positive	9	100.00%
Seradyn Color Slide II	4	*Positive	4	100.00%
SP Brand (Cardinal) Rapid Test	1	*Positive	1	100.00%
Alere Clearview Mono Cassette	8	*Positive	8	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	3	*Positive	3	100.00%
Consult Diag Mono Cassette	4	*Positive	4	100.00%
675 - Infectious Mononucleosis;754 - Infectious Mono/RF Combo;759 - Complete Immunology;628 - AFC Waived/PPM;729 - AFC Immunology	Labs	Response	Number	Percent
		Infectious Mononucleosis		
		* Denotes Correct Response		
IM-11				
AG: All Methods	78			
		*Positive	78	100.00%
Beck Coulter ICON (WB)	1	*Positive	1	100.00%
Alere Acceava (WB)	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	35	*Positive	35	100.00%
Henry Schein One-Step+ Mono	1	*Positive	1	100.00%
LifeSign Status Inf Mono	2	*Positive	2	100.00%



Lifesign Unistep	1	*Positive	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	3	*Positive	3	100.00%
Polymedco Poly stat Mono - (WB)	1	*Positive	1	100.00%
Polymedco Poly stat Mono (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (WB)	9	*Positive	9	100.00%
Seradyn Color Slide II	4	*Positive	4	100.00%
SP Brand (Cardinal) Rapid Test	1	*Positive	1	100.00%
Alere Clearview Mono Cassette	8	*Positive	8	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	3	*Positive	3	100.00%
Consult Diag Mono Cassette	4	*Positive	4	100.00%
IM-12		* Denotes Correct Response		
AG: All Methods	78			
		*Negative	78	100.00%
Beck Coulter ICON (WB)	1	*Negative	1	100.00%
Alere Acceava (WB)	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	35	*Negative	35	100.00%
Henry Schein One-Step+ Mono	1	*Negative	1	100.00%
LifeSign Status Inf Mono	2	*Negative	2	100.00%
Lifesign Unistep	1	*Negative	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	3	*Negative	3	100.00%
Polymedco Poly stat Mono - (WB)	1	*Negative	1	100.00%
Polymedco Poly stat Mono (S/P)	2	*Negative	2	100.00%
Quidel QuickVue+ (S/P)	2	*Negative	2	100.00%
Quidel QuickVue+ (WB)	9	*Negative	9	100.00%
Seradyn Color Slide II	4	*Negative	4	100.00%
SP Brand (Cardinal) Rapid Test	1	*Negative	1	100.00%
Alere Clearview Mono Cassette	8	*Negative	8	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	3	*Negative	3	100.00%
Consult Diag Mono Cassette	4	*Negative	4	100.00%
IM-13		* Denotes Correct Response		
AG: All Methods	78			
		*Positive	78	100.00%
Beck Coulter ICON (WB)	1	*Positive	1	100.00%
Alere Acceava (WB)	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	35	*Positive	35	100.00%
Henry Schein One-Step+ Mono	1	*Positive	1	100.00%
LifeSign Status Inf Mono	2	*Positive	2	100.00%
Lifesign Unistep	1	*Positive	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	3	*Positive	3	100.00%
Polymedco Poly stat Mono - (WB)	1	*Positive	1	100.00%
Polymedco Poly stat Mono (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (WB)	9	*Positive	9	100.00%
Seradyn Color Slide II	4	*Positive	4	100.00%
SP Brand (Cardinal) Rapid Test	1	*Positive	1	100.00%
Alere Clearview Mono Cassette	8	*Positive	8	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	3	*Positive	3	100.00%
Consult Diag Mono Cassette	4	*Positive	4	100.00%
IM-14		* Denotes Correct Response		
AG: All Methods	78			
		*Positive	77	98.72%
		Negative	1	1.28%
Beck Coulter ICON (WB)	1	*Positive	1	100.00%
Alere Acceava (WB)	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	35	*Positive	35	100.00%
Henry Schein One-Step+ Mono	1	*Positive	1	100.00%
LifeSign Status Inf Mono	2	*Positive	2	100.00%
Lifesign Unistep	1	*Positive	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	3	*Positive	3	100.00%
Polymedco Poly stat Mono - (WB)	1	*Positive	1	100.00%
Polymedco Poly stat Mono (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (WB)	9	*Positive	9	100.00%
Seradyn Color Slide II	4	*Positive	4	100.00%
SP Brand (Cardinal) Rapid Test	1	*Positive	1	100.00%
Alere Clearview Mono Cassette	8	*Positive	8	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	3	*Positive	2	66.67%
Consult Diag Mono Cassette	4	*Positive	4	100.00%



Sekisui Diagnostics OSOM Mono Non-Waived	3	Negative	1	33.33%
IM-15		* Denotes Correct Response		
AG: All Methods	78			
		*Negative	77	98.72%
		Positive	1	1.28%
Beck Coulter ICON (WB)	1	*Negative	1	100.00%
Alere Acceava (WB)	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	35	*Negative	34	97.14%
Henry Schein One-Step+ Mono	1	*Negative	1	100.00%
LifeSign Status Inf Mono	2	*Negative	2	100.00%
Lifesign Unistep	1	*Negative	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	3	*Negative	3	100.00%
Polymedco Poly stat Mono - (WB)	1	*Negative	1	100.00%
Polymedco Poly stat Mono (S/P)	2	*Negative	2	100.00%
Quidel QuickVue+ (S/P)	2	*Negative	2	100.00%
Quidel QuickVue+ (WB)	9	*Negative	9	100.00%
Seradyn Color Slide II	4	*Negative	4	100.00%
SP Brand (Cardinal) Rapid Test	1	*Negative	1	100.00%
Alere Clearview Mono Cassette	8	*Negative	8	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	3	*Negative	3	100.00%
Consult Diag Mono Cassette	4	*Negative	4	100.00%
Sekisui Diagnostics OSOM Mono (WB)	35	Positive	1	2.86%
		Infectious Mono (Waived)		
IM-11		* Denotes Correct Response		
AG: All Methods	466			
		*Positive	456	97.85%
		Negative	10	2.15%
Alere Acceava (S)	1	*Positive	1	100.00%
Alere Acceava (WB)	58	*Positive	58	100.00%
Sekisui Diagnostics OSOM Mono (WB)	182	*Positive	179	98.35%
Henry Schein One-Step+ Mono	30	*Positive	30	100.00%
Instant Technologies i-Screen (WB)	1	*Positive	1	100.00%
Inverness Signify Mono	3	*Positive	3	100.00%
LifeSign Status Inf Mono	31	*Positive	30	96.77%
McKesson Medi-Lab Perf - Inf Mono	12	*Positive	12	100.00%
Polymedco Poly stat Mono - (WB)	6	*Positive	5	83.33%
Quidel QuickVue+ (WB)	15	*Positive	14	93.33%
Select Diag Mono Cas (WB)	1	*Positive	1	100.00%
SP Brand (Cardinal) Rapid Test	9	*Positive	9	100.00%
Alere Clearview Mono Cassette	26	*Positive	25	96.15%
Diag Test Group Clarity Mono	3	*Positive	3	100.00%
Fisher HC Sure-Vue Mono	4	*Positive	4	100.00%
Immunostics Inc. ImmunoDetector Mono (WB)	2	*Positive	2	100.00%
Meridian Immunocard STAT! Mono (WB)	4	*Positive	4	100.00%
NDC ProAdvantage Mono	2	*Positive	2	100.00%
Consult Diag Mono Cassette	67	*Positive	65	97.01%
Alere Acceava (WB) Mono Cassette	7	*Positive	7	100.00%
Medline Mono Cassette (WB)	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	182	Negative	3	1.65%
LifeSign Status Inf Mono	31	Negative	1	3.23%
Polymedco Poly stat Mono - (WB)	6	Negative	1	16.67%
Quidel QuickVue+ (WB)	15	Negative	1	6.67%
Alere Clearview Mono Cassette	26	Negative	1	3.85%
Pulse Scientific	1	Negative	1	100.00%
Consult Diag Mono Cassette	67	Negative	2	2.99%
		Infectious Mononucleosis (International)		
IM-11		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Alere Clearview Mono Cassette	1	*Positive	1	100.00%
Fisher HC Sure-Vue Mono	1	*Positive	1	100.00%
IM-12		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Alere Clearview Mono Cassette	1	*Negative	1	100.00%
Fisher HC Sure-Vue Mono	1	*Negative	1	100.00%
IM-13		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%



Alere Clearview Mono Cassette	1	*Positive	1	100.00%
Fisher HC Sure-Vue Mono	1	*Positive	1	100.00%
IM-14		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Alere Clearview Mono Cassette	1	*Positive	1	100.00%
Fisher HC Sure-Vue Mono	1	*Positive	1	100.00%
IM-15		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Alere Clearview Mono Cassette	1	*Negative	1	100.00%
Fisher HC Sure-Vue Mono	1	*Negative	1	100.00%
637 - KOH - Glass Slides (Skin Hair Nail)	Labs	Response	Number	Percent
		KOH Preparation		
KOH-5		* Denotes Correct Response		
AG: All Methods	18			
		*No fungal elements/yeast observed	16	88.89%
		Fungal elements/yeast observed	2	11.11%
KOH Glass Slide	18	*No fungal elements/yeast observed	16	88.89%
KOH Glass Slide		Fungal elements/yeast observed	2	11.11%
KOH-6		* Denotes Correct Response		
AG: All Methods	18			
		*Fungal elements/yeast observed	18	100.00%
KOH Glass Slide	18	*Fungal elements/yeast observed	18	100.00%
758 - Mycoplasma Antibody	Labs	Response	Number	Percent
		Mycoplasma Antibody		
MY-3		* Denotes Correct Response		
AG: All Methods	5			
		*Positive	4	80.00%
		Negative	1	20.00%
Meridian ImmunoCard	5	*Positive	4	80.00%
Meridian ImmunoCard		Negative	1	20.00%
791 - Fecal Occult Blood Add-On;601 - Waived/PPM Tests II;609 - Waived/PPM Tests I;647 - Mini-Waived Package;628 - AFC Waived/PPM	Labs	Response	Number	Percent
		Fecal Occult Blood		
OC-3		* Denotes Correct Response		
MG: Guaiac methods	727			
		*Positive	706	97.11%
		Negative	21	2.89%
OC-3		* Denotes Correct Response		
MG: Immunochemical methods	118			
		*Positive	116	98.31%
		Negative	2	1.69%
950 - Post Vasectomy Sperm Analysis	Labs	Response	Number	Percent
		Sperm Count (Post Vasectomy)		
PVAS-5		* Denotes Correct Response		
AG: All Methods	16			
		*Positive for Sperm	16	100.00%
Hemocytometer	4	*Positive for Sperm	4	100.00%
Post Vasectomy Microscopy	12	*Positive for Sperm	12	100.00%
PVAS-6		* Denotes Correct Response		
AG: All Methods	16			
		*Negative for Sperm	15	93.75%
		Positive for Sperm	1	6.25%
Hemocytometer	4	*Negative for Sperm	4	100.00%
Post Vasectomy Microscopy	12	*Negative for Sperm	11	91.67%
Post Vasectomy Microscopy		Positive for Sperm	1	8.33%
754 - Infectious Mono/RF Combo;653 - Rheumatoid Factor;759 - Complete Immunology	Labs	Response	Number	Percent
		Rheumatoid Factor		
RF-11		* Denotes Correct Response		
AG: All Methods	12			
		*N/A (Negative)	11	91.67%
		Positive	1	8.33%
ASI	1	*N/A (Negative)	1	100.00%
Siemens RapiTex	2	*N/A (Negative)	2	100.00%
Seradyn Color Slide	2	*N/A (Negative)	2	100.00%



Wampole ColorCard RF	2	*N/A (Negative)	2	100.00%
Wampole Rheumatex	2	*N/A (Negative)	1	50.00%
Biokit Rhemajet RF	2	*N/A (Negative)	2	100.00%
Remel SeraTest RF-Latex	1	*N/A (Negative)	1	100.00%
Wampole Rheumatex	2	Positive	1	50.00%
RF-12		* Denotes Correct Response		
AG: All Methods	12			
		*N/A (Negative)	11	91.67%
		Positive	1	8.33%
ASI	1	*N/A (Negative)	1	100.00%
Siemens RapiTex	2	*N/A (Negative)	2	100.00%
Seradyn Color Slide	2	*N/A (Negative)	2	100.00%
Wampole ColorCard RF	2	*N/A (Negative)	2	100.00%
Wampole Rheumatex	2	*N/A (Negative)	1	50.00%
Biokit Rhemajet RF	2	*N/A (Negative)	2	100.00%
Remel SeraTest RF-Latex	1	*N/A (Negative)	1	100.00%
Wampole Rheumatex	2	Positive	1	50.00%
RF-13		* Denotes Correct Response		
AG: All Methods	12			
		*Positive	12	100.00%
ASI	1	*Positive	1	100.00%
Siemens RapiTex	2	*Positive	2	100.00%
Seradyn Color Slide	2	*Positive	2	100.00%
Wampole ColorCard RF	2	*Positive	2	100.00%
Wampole Rheumatex	2	*Positive	2	100.00%
Biokit Rhemajet RF	2	*Positive	2	100.00%
Remel SeraTest RF-Latex	1	*Positive	1	100.00%
RF-14		* Denotes Correct Response		
AG: All Methods	12			
		*N/A (Negative)	12	100.00%
ASI	1	*N/A (Negative)	1	100.00%
Siemens RapiTex	2	*N/A (Negative)	2	100.00%
Seradyn Color Slide	2	*N/A (Negative)	2	100.00%
Wampole ColorCard RF	2	*N/A (Negative)	2	100.00%
Wampole Rheumatex	2	*N/A (Negative)	2	100.00%
Biokit Rhemajet RF	2	*N/A (Negative)	2	100.00%
Remel SeraTest RF-Latex	1	*N/A (Negative)	1	100.00%
RF-15		* Denotes Correct Response		
AG: All Methods	12			
		*Positive	12	100.00%
ASI	1	*Positive	1	100.00%
Siemens RapiTex	2	*Positive	2	100.00%
Seradyn Color Slide	2	*Positive	2	100.00%
Wampole ColorCard RF	2	*Positive	2	100.00%
Wampole Rheumatex	2	*Positive	2	100.00%
Biokit Rhemajet RF	2	*Positive	2	100.00%
Remel SeraTest RF-Latex	1	*Positive	1	100.00%
		Rheumatoid Factor (International)		
RF-11		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Wampole ColorCard RF	1	*Negative	1	100.00%
Fisher HC Sure-Vue RF	1	*Negative	1	100.00%
Immunostics Rhuemacol/LEX	1	*Negative	1	100.00%
RF-12		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Wampole ColorCard RF	1	*Negative	1	100.00%
Fisher HC Sure-Vue RF	1	*Negative	1	100.00%
Immunostics Rhuemacol/LEX	1	*Negative	1	100.00%
RF-13		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Wampole ColorCard RF	1	*Positive	1	100.00%
Fisher HC Sure-Vue RF	1	*Positive	1	100.00%
Immunostics Rhuemacol/LEX	1	*Positive	1	100.00%
RF-14		* Denotes Correct Response		
AG: All Methods	3			



		*Negative	3	100.00%
Wampole ColorCard RF	1	*Negative	1	100.00%
Fisher HC Sure-Vue RF	1	*Negative	1	100.00%
Immunostics Rhuemacol/LEX	1	*Negative	1	100.00%
RF-15		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Wampole ColorCard RF	1	*Positive	1	100.00%
Fisher HC Sure-Vue RF	1	*Positive	1	100.00%
Immunostics Rhuemacol/LEX	1	*Positive	1	100.00%
		Rheumatoid Factor (Semi Quant)		
RF-11		* Denotes Correct Response		
PG: Fisher HC Sure-Vue RF (Semi Quant)	1			
		*Negative	1	100.00%
RF-12		* Denotes Correct Response		
PG: Fisher HC Sure-Vue RF (Semi Quant)	1			
		*Negative	1	100.00%
RF-13		* Denotes Correct Response		
PG: Fisher HC Sure-Vue RF (Semi Quant)	1			
		*320 IU/mL	1	100.00%
RF-14		* Denotes Correct Response		
PG: Fisher HC Sure-Vue RF (Semi Quant)	1			
		*Negative	1	100.00%
RF-15		* Denotes Correct Response		
PG: Fisher HC Sure-Vue RF (Semi Quant)	1			
		*320 IU/mL	1	100.00%
645 - Fetal Membrane Rupture	Labs	Response	Number	Percent
		Fetal Membranes Rupture PAMG-1		
ROM-5		* Denotes Correct Response		
AG: All Methods	17			
		*Positive	17	100.00%
AmniSure ROM Test	17	*Positive	17	100.00%
ROM-6		* Denotes Correct Response		
AG: All Methods	17			
		*Negative	17	100.00%
AmniSure ROM Test	17	*Negative	17	100.00%
601 - Waived/PPM Tests II;785 - Presump ID/Colony Count-Urine & Strep Combo;691 - Group A Strep Antigen (5 specimens);783 - Group A Strep Antigen/Add-on (2 specimens);665 - Group A Strep Antigen (Waived) (2 specimens);628 - AFC Waived/PPM	Labs	Response	Number	Percent
		Strep A Ag Det. (Waived)		
RS-11		* Denotes Correct Response		
AG: All Methods	889			
		*Positive	682	76.72%
		Negative	207	23.28%
Abbott Signify Strep A - Waived	6	*Positive	2	33.33%
Beck Coulter ICON SC Strep	1	*Positive	1	100.00%
Alere Acceava Strep A Dipstick	117	*Positive	96	82.05%
Diag Test Group Clarity - Strip Strep Ag	1	*Positive	1	100.00%
Fisher HC Sure-Vue Strep A (Waived)	7	*Positive	4	57.14%
Sekisui Diagnostics OSOM	41	*Positive	35	85.37%
Sekisui Diagnostics OSOM Ultra Strep	51	*Positive	47	92.16%
Henry Schein OneStep + Rapid Strep Cass	3	*Positive	2	66.67%
Henry Schein OneStep+ Rapid Strep Dip	65	*Positive	48	73.85%
Innovacon One Step Strep Ag	1	*Positive	1	100.00%
Instant Technologies I Strep	1	*Positive	1	100.00%
Inverness Signify Strep A Dipstick	7	*Positive	7	100.00%
McKesson Strep A Cas	4	*Positive	3	75.00%
McKesson Strep A Dip	46	*Positive	38	82.61%
McKesson Strep A Test - Twist	2	*Positive	1	50.00%
Permaxim Rediscreen Strep A Rapid	1	*Positive	1	100.00%
Polymedco Poly Stat Strep A (W)	2	*Positive	1	50.00%
PSS Rapid Test Strip Strep A	2	*Positive	1	50.00%
Quidel QuickVue Dipstick Strep	98	*Positive	88	89.80%
Quidel QuickVue In-Line - 25 Test	82	*Positive	60	73.17%
Quidel QuickVue+ Strep A	29	*Positive	23	79.31%
Select Diag Strep A Dip	4	*Positive	4	100.00%
SP (Cardinal) Rapid Strep Cassette	6	*Positive	5	83.33%



SP Brand (Cardinal) Rapid Strep A Dipstick	87	*Positive	53	60.92%
Stanbio QuStick Strep A	1	*Positive	1	100.00%
Alere Clearview Exact II Dip	37	*Positive	23	62.16%
Sekisui Diagnostics OSOM Ultra	5	*Positive	5	100.00%
Consult Diag Strep A Dipstick	134	*Positive	94	70.15%
Consult Diag Strep A Twist	2	*Positive	2	100.00%
MooreBrand Strep A Dipstick	5	*Positive	2	40.00%
NDC ProAdvantage Strep Strips	2	*Positive	1	50.00%
Meridian ImmunoCard STAT Strep A	2	*Positive	1	50.00%
Quidel Sofia Strep A FIA	8	*Positive	7	87.50%
BD Veritor Strep A	18	*Positive	15	83.33%
Alere i Strep A	6	*Positive	5	83.33%
Roche Cobas Liat Strep A	3	*Positive	3	100.00%
Abbott Signify Strep A - Waived	6	Negative	4	66.67%
Beck Coulter ICON DS Strep	1	Negative	1	100.00%
Alere Acceava Strep A Dipstick	117	Negative	21	17.95%
Fisher HC Sure-Vue Strep A (Waived)	7	Negative	3	42.86%
Sekisui Diagnostics OSOM	41	Negative	6	14.63%
Sekisui Diagnostics OSOM Ultra Strep	51	Negative	4	7.84%
Henry Schein OneStep + Rapid Strep Cass	3	Negative	1	33.33%
Henry Schein OneStep+ Rapid Strep Dip	65	Negative	17	26.15%
McKesson Strep A Cas	4	Negative	1	25.00%
McKesson Strep A Dip	46	Negative	8	17.39%
McKesson Strep A Test - Twist	2	Negative	1	50.00%
Polymedco Poly Stat Strep A (W)	2	Negative	1	50.00%
PSS Rapid Test Strip Strep A	2	Negative	1	50.00%
Quidel QuickVue Dipstick Strep	98	Negative	10	10.20%
Quidel QuickVue In-Line - 25 Test	82	Negative	22	26.83%
Quidel QuickVue+ Strep A	29	Negative	6	20.69%
SP (Cardinal) Rapid Strep Cassette	6	Negative	1	16.67%
SP Brand (Cardinal) Rapid Strep A Dipstick	87	Negative	34	39.08%
Alere Clearview Exact II Dip	37	Negative	14	37.84%
Consult Diag Strep A Dipstick	134	Negative	40	29.85%
MooreBrand Strep A Dipstick	5	Negative	3	60.00%
NDC ProAdvantage Strep Strips	2	Negative	1	50.00%
Meridian ImmunoCard STAT Strep A	2	Negative	1	50.00%
LifeSign Bio Strep A Test STRIP	1	Negative	1	100.00%
Quidel Sofia Strep A FIA	8	Negative	1	12.50%
BD Veritor Strep A	18	Negative	3	16.67%
Alere i Strep A	6	Negative	1	16.67%
RS-12		* Denotes Correct Response		
AG: All Methods	36			
		*Negative	34	94.44%
		Positive	2	5.56%
Alere Acceava Strep A Dipstick	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM	4	*Negative	3	75.00%
Sekisui Diagnostics OSOM Ultra Strep	4	*Negative	4	100.00%
Innovacon One Step Strep Ag	1	*Negative	1	100.00%
McKesson Strep A Dip	2	*Negative	2	100.00%
Quidel QuickVue Dipstick Strep	3	*Negative	3	100.00%
Quidel QuickVue In-Line - 25 Test	2	*Negative	1	50.00%
Quidel QuickVue+ Strep A	1	*Negative	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	7	*Negative	7	100.00%
Consult Diag Strep A Dipstick	8	*Negative	8	100.00%
Consult Diag Strep A Twist	1	*Negative	1	100.00%
MooreBrand Strep A Dipstick	1	*Negative	1	100.00%
Alere i Strep A	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM	4	Positive	1	25.00%
Quidel QuickVue In-Line - 25 Test	2	Positive	1	50.00%
RS-11		* Denotes Correct Response		
AG: All Methods	200			
		*Positive	184	92.00%
		Negative	16	8.00%
Abbott Signify Strep A - Waived	2	*Positive	2	100.00%
BD Check Strep A (Dip)	2	*Positive	2	100.00%
Beck Coulter ICON DS Strep	1	*Positive	1	100.00%
Alere Acceava Strep A Dipstick	17	*Positive	15	88.24%
Diag Test Group Clarity - Strip Strep Ag	1	*Positive	1	100.00%



Fisher HC Sure-Vue Select Strep Aq	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM	23	*Positive	22	95.65%
Sekisui Diagnostics OSOM Ultra Strep	28	*Positive	27	96.43%
Henry Schein OneStep+ Rapid Strep Dip	10	*Positive	7	70.00%
Immunostics Immuno/Strep A Det	2	*Positive	2	100.00%
Inverness Signify Strep A Dipstick	1	*Positive	1	100.00%
McKesson Strep A Cas	2	*Positive	1	50.00%
McKesson Strep A Dip	5	*Positive	4	80.00%
Quidel QuickVue Dipstick Strep	20	*Positive	20	100.00%
Quidel QuickVue In-Line - 25 Test	15	*Positive	14	93.33%
Quidel QuickVue+ Strep A	31	*Positive	31	100.00%
Select Diag Strep A Dip	1	*Positive	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	*Positive	1	33.33%
Stanbio QuStick Strep A	1	*Positive	1	100.00%
Consult Diag Strep A Dipstick	18	*Positive	15	83.33%
Meridian ImmunoCard STAT Strep A	1	*Positive	1	100.00%
Quidel Sofia Strep A FIA	4	*Positive	4	100.00%
BD Veritor Strep A	6	*Positive	6	100.00%
Quidel Solana	4	*Positive	4	100.00%
Alere Acceava Strep A Dipstick	17	Negative	2	11.76%
Sekisui Diagnostics OSOM	23	Negative	1	4.35%
Sekisui Diagnostics OSOM Ultra Strep	28	Negative	1	3.57%
Henry Schein OneStep+ Rapid Strep Dip	10	Negative	3	30.00%
McKesson Strep A Cas	2	Negative	1	50.00%
McKesson Strep A Dip	5	Negative	1	20.00%
Quidel QuickVue In-Line - 25 Test	15	Negative	1	6.67%
SP (Cardinal) Rapid Strep Cassette	1	Negative	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	Negative	2	66.67%
Consult Diag Strep A Dipstick	18	Negative	3	16.67%
RS-12		* Denotes Correct Response		
AG: All Methods	200			
		*Negative	200	100.00%
Abbott Signify Strep A - Waived	2	*Negative	2	100.00%
BD Check Strep A (Dip)	2	*Negative	2	100.00%
Beck Coulter ICON DS Strep	1	*Negative	1	100.00%
Alere Acceava Strep A Dipstick	17	*Negative	17	100.00%
Diag Test Group Clarity - Strip Strep Aq	1	*Negative	1	100.00%
Fisher HC Sure-Vue Select Strep Aq	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM	23	*Negative	23	100.00%
Sekisui Diagnostics OSOM Ultra Strep	28	*Negative	28	100.00%
Henry Schein OneStep+ Rapid Strep Dip	10	*Negative	10	100.00%
Immunostics Immuno/Strep A Det	2	*Negative	2	100.00%
Inverness Signify Strep A Dipstick	1	*Negative	1	100.00%
McKesson Strep A Cas	2	*Negative	2	100.00%
McKesson Strep A Dip	5	*Negative	5	100.00%
Quidel QuickVue Dipstick Strep	20	*Negative	20	100.00%
Quidel QuickVue In-Line - 25 Test	15	*Negative	15	100.00%
Quidel QuickVue+ Strep A	31	*Negative	31	100.00%
Select Diag Strep A Dip	1	*Negative	1	100.00%
SP (Cardinal) Rapid Strep Cassette	1	*Negative	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	*Negative	3	100.00%
Stanbio QuStick Strep A	1	*Negative	1	100.00%
Consult Diag Strep A Dipstick	18	*Negative	18	100.00%
Meridian ImmunoCard STAT Strep A	1	*Negative	1	100.00%
Quidel Sofia Strep A FIA	4	*Negative	4	100.00%
BD Veritor Strep A	6	*Negative	6	100.00%
Quidel Solana	4	*Negative	4	100.00%
		Strep A Aq Det. (International)		
RS-11		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Alere Clearview Exact II Cas	1	*Positive	1	100.00%
RS-12		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Alere Clearview Exact II Cas	1	*Negative	1	100.00%
691 - Group A Strep Antigen (5 specimens)	Labs	Response	Number	Percent
		Strep A Aq Det.		
RS-11		* Denotes Correct Response		



AG: All Methods	200			
		*Positive	184	92.00%
		Negative	16	8.00%
Abbott Signify Strep A - Waived	2	*Positive	2	100.00%
BD Check Strep A (Dip)	2	*Positive	2	100.00%
Beck Coulter ICON DS Strep	1	*Positive	1	100.00%
Alere Acceava Strep A Dipstick	17	*Positive	15	88.24%
Diag Test Group Clarity - Strip Strep Ag	1	*Positive	1	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM	23	*Positive	22	95.65%
Sekisui Diagnostics OSOM Ultra Strep	28	*Positive	27	96.43%
Henry Schein OneStep+ Rapid Strep Dip	10	*Positive	7	70.00%
Immunostics Immuno/Strep A Det	2	*Positive	2	100.00%
Inverness Signify Strep A Dipstick	1	*Positive	1	100.00%
McKesson Strep A Cas	2	*Positive	1	50.00%
McKesson Strep A Dip	5	*Positive	4	80.00%
Quidel QuickVue Dipstick Strep	20	*Positive	20	100.00%
Quidel QuickVue In-Line - 25 Test	15	*Positive	14	93.33%
Quidel QuickVue+ Strep A	31	*Positive	31	100.00%
Select Diag Strep A Dip	1	*Positive	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	*Positive	1	33.33%
Stanbio QuStick Strep A	1	*Positive	1	100.00%
Consult Diag Strep A Dipstick	18	*Positive	15	83.33%
Meridian ImmunoCard STAT Strep A	1	*Positive	1	100.00%
Quidel Sofia Strep A FIA	4	*Positive	4	100.00%
BD Veritor Strep A	6	*Positive	6	100.00%
Quidel Solana	4	*Positive	4	100.00%
Alere Acceava Strep A Dipstick	17	Negative	2	11.76%
Sekisui Diagnostics OSOM	23	Negative	1	4.35%
Sekisui Diagnostics OSOM Ultra Strep	28	Negative	1	3.57%
Henry Schein OneStep+ Rapid Strep Dip	10	Negative	3	30.00%
McKesson Strep A Cas	2	Negative	1	50.00%
McKesson Strep A Dip	5	Negative	1	20.00%
Quidel QuickVue In-Line - 25 Test	15	Negative	1	6.67%
SP (Cardinal) Rapid Strep Cassette	1	Negative	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	Negative	2	66.67%
Consult Diag Strep A Dipstick	18	Negative	3	16.67%
RS-12		* Denotes Correct Response		
AG: All Methods	200			
		*Negative	200	100.00%
Abbott Signify Strep A - Waived	2	*Negative	2	100.00%
BD Check Strep A (Dip)	2	*Negative	2	100.00%
Beck Coulter ICON DS Strep	1	*Negative	1	100.00%
Alere Acceava Strep A Dipstick	17	*Negative	17	100.00%
Diag Test Group Clarity - Strip Strep Ag	1	*Negative	1	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM	23	*Negative	23	100.00%
Sekisui Diagnostics OSOM Ultra Strep	28	*Negative	28	100.00%
Henry Schein OneStep+ Rapid Strep Dip	10	*Negative	10	100.00%
Immunostics Immuno/Strep A Det	2	*Negative	2	100.00%
Inverness Signify Strep A Dipstick	1	*Negative	1	100.00%
McKesson Strep A Cas	2	*Negative	2	100.00%
McKesson Strep A Dip	5	*Negative	5	100.00%
Quidel QuickVue Dipstick Strep	20	*Negative	20	100.00%
Quidel QuickVue In-Line - 25 Test	15	*Negative	15	100.00%
Quidel QuickVue+ Strep A	31	*Negative	31	100.00%
Select Diag Strep A Dip	1	*Negative	1	100.00%
SP (Cardinal) Rapid Strep Cassette	1	*Negative	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	*Negative	3	100.00%
Stanbio QuStick Strep A	1	*Negative	1	100.00%
Consult Diag Strep A Dipstick	18	*Negative	18	100.00%
Meridian ImmunoCard STAT Strep A	1	*Negative	1	100.00%
Quidel Sofia Strep A FIA	4	*Negative	4	100.00%
BD Veritor Strep A	6	*Negative	6	100.00%
Quidel Solana	4	*Negative	4	100.00%
RS-13		* Denotes Correct Response		
AG: All Methods	172			
		*Positive	154	89.53%
		Negative	18	10.47%



Abbott Signify Strep A - Waived	2	*Positive	2	100.00%
BD Check Strep A (Dip)	2	*Positive	2	100.00%
Beck Coulter ICON DS Strep	1	*Positive	1	100.00%
Alere Aceava Strep A Dipstick	17	*Positive	15	88.24%
Diag Test Group Clarity - Strip Strep Ag	1	*Positive	1	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM	23	*Positive	21	91.30%
Sekisui Diagnostics OSOM Ultra Strep	24	*Positive	23	95.83%
Henry Schein OneStep+ Rapid Strep Dip	10	*Positive	6	60.00%
Immunostics Immuno/Strep A Det	2	*Positive	2	100.00%
Inverness Signify Strep A Dipstick	1	*Positive	1	100.00%
McKesson Strep A Cas	1	*Positive	1	100.00%
McKesson Strep A Dip	5	*Positive	2	40.00%
Quidel QuickVue Dipstick Strep	17	*Positive	17	100.00%
Quidel QuickVue In-Line - 25 Test	13	*Positive	11	84.62%
Quidel QuickVue+ Strep A	29	*Positive	29	100.00%
SP (Cardinal) Rapid Strep Cassette	1	*Positive	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	*Positive	2	66.67%
Consult Diag Strep A Dipstick	12	*Positive	9	75.00%
Meridian ImmunoCard STAT Strep A	1	*Positive	1	100.00%
Quidel Sofia Strep A FIA	3	*Positive	3	100.00%
BD Veritor Strep A	1	*Positive	1	100.00%
Quidel Solana	2	*Positive	2	100.00%
Alere Aceava Strep A Dipstick	17	Negative	2	11.76%
Sekisui Diagnostics OSOM	23	Negative	2	8.70%
Sekisui Diagnostics OSOM Ultra Strep	24	Negative	1	4.17%
Henry Schein OneStep+ Rapid Strep Dip	10	Negative	4	40.00%
McKesson Strep A Dip	5	Negative	3	60.00%
Quidel QuickVue In-Line - 25 Test	13	Negative	2	15.38%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	Negative	1	33.33%
Consult Diag Strep A Dipstick	12	Negative	3	25.00%
RS-14		* Denotes Correct Response		
AG: All Methods	172			
		*Positive	163	94.77%
		Negative	9	5.23%
Abbott Signify Strep A - Waived	2	*Positive	2	100.00%
BD Check Strep A (Dip)	2	*Positive	2	100.00%
Beck Coulter ICON DS Strep	1	*Positive	1	100.00%
Alere Aceava Strep A Dipstick	17	*Positive	17	100.00%
Diag Test Group Clarity - Strip Strep Ag	1	*Positive	1	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM	23	*Positive	22	95.65%
Sekisui Diagnostics OSOM Ultra Strep	24	*Positive	24	100.00%
Henry Schein OneStep+ Rapid Strep Dip	10	*Positive	7	70.00%
Immunostics Immuno/Strep A Det	2	*Positive	2	100.00%
Inverness Signify Strep A Dipstick	1	*Positive	1	100.00%
McKesson Strep A Dip	5	*Positive	4	80.00%
Quidel QuickVue Dipstick Strep	17	*Positive	17	100.00%
Quidel QuickVue In-Line - 25 Test	13	*Positive	12	92.31%
Quidel QuickVue+ Strep A	29	*Positive	28	96.55%
SP (Cardinal) Rapid Strep Cassette	1	*Positive	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	*Positive	2	66.67%
Consult Diag Strep A Dipstick	12	*Positive	12	100.00%
Meridian ImmunoCard STAT Strep A	1	*Positive	1	100.00%
Quidel Sofia Strep A FIA	3	*Positive	3	100.00%
BD Veritor Strep A	1	*Positive	1	100.00%
Quidel Solana	2	*Positive	2	100.00%
Sekisui Diagnostics OSOM	23	Negative	1	4.35%
Henry Schein OneStep+ Rapid Strep Dip	10	Negative	3	30.00%
McKesson Strep A Cas	1	Negative	1	100.00%
McKesson Strep A Dip	5	Negative	1	20.00%
Quidel QuickVue In-Line - 25 Test	13	Negative	1	7.69%
Quidel QuickVue+ Strep A	29	Negative	1	3.45%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	Negative	1	33.33%
RS-15		* Denotes Correct Response		
AG: All Methods	172			
		*Negative	170	98.84%
		Positive	2	1.16%
Abbott Signify Strep A - Waived	2	*Negative	2	100.00%



BD Check Strep A (Dip)	2	*Negative	2	100.00%
Beck Coulter ICON DS Strep	1	*Negative	1	100.00%
Alere Acceava Strep A Dipstick	17	*Negative	17	100.00%
Diag Test Group Clarity - Strip Strep Ag	1	*Negative	1	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM	23	*Negative	22	95.65%
Sekisui Diagnostics OSOM Ultra Strep	24	*Negative	24	100.00%
Henry Schein OneStep+ Rapid Strep Dip	10	*Negative	10	100.00%
Immunostics Immuno/Strep A Det	2	*Negative	2	100.00%
Inverness Signify Strep A Dipstick	1	*Negative	1	100.00%
McKesson Strep A Cas	1	*Negative	1	100.00%
McKesson Strep A Dip	5	*Negative	5	100.00%
Quidel QuickVue Dipstick Strep	17	*Negative	17	100.00%
Quidel QuickVue In-Line - 25 Test	13	*Negative	12	92.31%
Quidel QuickVue+ Strep A	29	*Negative	29	100.00%
SP (Cardinal) Rapid Strep Cassette	1	*Negative	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	3	*Negative	3	100.00%
Consult Diag Strep A Dipstick	12	*Negative	12	100.00%
Meridian ImmunoCard STAT Strep A	1	*Negative	1	100.00%
Quidel Sofia Strep A FIA	3	*Negative	3	100.00%
BD Veritor Strep A	1	*Negative	1	100.00%
Quidel Solana	2	*Negative	2	100.00%
Sekisui Diagnostics OSOM	23	Positive	1	4.35%
Quidel QuickVue In-Line - 25 Test	13	Positive	1	7.69%
		Strep A Ag Det. (Waived)		
RS-11		* Denotes Correct Response		
AG: All Methods	889			
		*Positive	682	76.72%
		Negative	207	23.28%
Abbott Signify Strep A - Waived	6	*Positive	2	33.33%
Beck Coulter ICON SC Strep	1	*Positive	1	100.00%
Alere Acceava Strep A Dipstick	117	*Positive	96	82.05%
Diag Test Group Clarity - Strip Strep Ag	1	*Positive	1	100.00%
Fisher HC Sure-Vue Strep A (Waived)	7	*Positive	4	57.14%
Sekisui Diagnostics OSOM	41	*Positive	35	85.37%
Sekisui Diagnostics OSOM Ultra Strep	51	*Positive	47	92.16%
Henry Schein OneStep + Rapid Strep Cass	3	*Positive	2	66.67%
Henry Schein OneStep+ Rapid Strep Dip	65	*Positive	48	73.85%
Innovacon One Step Strep Ag	1	*Positive	1	100.00%
Instant Technologies i Strep	1	*Positive	1	100.00%
Inverness Signify Strep A Dipstick	7	*Positive	7	100.00%
McKesson Strep A Cas	4	*Positive	3	75.00%
McKesson Strep A Dip	46	*Positive	38	82.61%
McKesson Strep A Test - Twist	2	*Positive	1	50.00%
Permaxim Rediscreen Strep A Rapid	1	*Positive	1	100.00%
Polymedco Poly Stat Strep A (W)	2	*Positive	1	50.00%
PSS Rapid Test Strip Strep A	2	*Positive	1	50.00%
Quidel QuickVue Dipstick Strep	98	*Positive	88	89.80%
Quidel QuickVue In-Line - 25 Test	82	*Positive	60	73.17%
Quidel QuickVue+ Strep A	29	*Positive	23	79.31%
Select Diag Strep A Dip	4	*Positive	4	100.00%
SP (Cardinal) Rapid Strep Cassette	6	*Positive	5	83.33%
SP Brand (Cardinal) Rapid Strep A Dipstick	87	*Positive	53	60.92%
Stanbio QuStick Strep A	1	*Positive	1	100.00%
Alere Clearview Exact II Dip	37	*Positive	23	62.16%
Sekisui Diagnostics OSOM Ultra	5	*Positive	5	100.00%
Consult Diag Strep A Dipstick	134	*Positive	94	70.15%
Consult Diag Strep A Twist	2	*Positive	2	100.00%
MooreBrand Strep A Dipstick	5	*Positive	2	40.00%
NDC ProAdvantage Strep Strips	2	*Positive	1	50.00%
Meridian ImmunoCard STAT Strep A	2	*Positive	1	50.00%
Quidel Sofia Strep A FIA	8	*Positive	7	87.50%
BD Veritor Strep A	18	*Positive	15	83.33%
Alere i Strep A	6	*Positive	5	83.33%
Roche Cobas Liat Strep A	3	*Positive	3	100.00%
Abbott Signify Strep A - Waived	6	Negative	4	66.67%
Beck Coulter ICON DS Strep	1	Negative	1	100.00%
Alere Acceava Strep A Dipstick	117	Negative	21	17.95%
Fisher HC Sure-Vue Strep A (Waived)	7	Negative	3	42.86%



Sekisui Diagnostics OSOM	41	Negative	6	14.63%
Sekisui Diagnostics OSOM Ultra Strep	51	Negative	4	7.84%
Henry Schein OneStep + Rapid Strep Cass	3	Negative	1	33.33%
Henry Schein OneStep+ Rapid Strep Dip	65	Negative	17	26.15%
McKesson Strep A Cas	4	Negative	1	25.00%
McKesson Strep A Dip	46	Negative	8	17.39%
McKesson Strep A Test - Twist	2	Negative	1	50.00%
Polymedco Poly Stat Strep A (W)	2	Negative	1	50.00%
PSS Rapid Test Strip Strep A	2	Negative	1	50.00%
Quidel QuickVue Dipstick Strep	98	Negative	10	10.20%
Quidel QuickVue In-Line - 25 Test	82	Negative	22	26.83%
Quidel QuickVue+ Strep A	29	Negative	6	20.69%
SP (Cardinal) Rapid Strep Cassette	6	Negative	1	16.67%
SP Brand (Cardinal) Rapid Strep A Dipstick	87	Negative	34	39.08%
Alere Clearview Exact II Dip	37	Negative	14	37.84%
Consult Diag Strep A Dipstick	134	Negative	40	29.85%
MooreBrand Strep A Dipstick	5	Negative	3	60.00%
NDC ProAdvantage Strep Strips	2	Negative	1	50.00%
Meridian ImmunoCard STAT Strep A	2	Negative	1	50.00%
LifeSign Bio Strep A Test Strip	1	Negative	1	100.00%
Quidel Sofia Strep A FIA	8	Negative	1	12.50%
BD Veritor Strep A	18	Negative	3	16.67%
Alere i Strep A	6	Negative	1	16.67%
RS-12		* Denotes Correct Response		
AG: All Methods	36			
		*Negative	34	94.44%
		Positive	2	5.56%
Alere Aceava Strep A Dipstick	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM	4	*Negative	3	75.00%
Sekisui Diagnostics OSOM Ultra Strep	4	*Negative	4	100.00%
Innovacon One Step Strep Ag	1	*Negative	1	100.00%
McKesson Strep A Dip	2	*Negative	2	100.00%
Quidel QuickVue Dipstick Strep	3	*Negative	3	100.00%
Quidel QuickVue In-Line - 25 Test	2	*Negative	1	50.00%
Quidel QuickVue+ Strep A	1	*Negative	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	7	*Negative	7	100.00%
Consult Diag Strep A Dipstick	8	*Negative	8	100.00%
Consult Diag Strep A Twist	1	*Negative	1	100.00%
MooreBrand Strep A Dipstick	1	*Negative	1	100.00%
Alere i Strep A	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM	4	Positive	1	25.00%
Quidel QuickVue In-Line - 25 Test	2	Positive	1	50.00%
RS-13		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Quidel QuickVue In-Line - 25 Test	1	*Positive	1	100.00%
RS-14		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Quidel QuickVue In-Line - 25 Test	1	*Positive	1	100.00%
RS-15		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	1	50.00%
		Positive	1	50.00%
Quidel QuickVue In-Line - 25 Test	2	*Negative	1	50.00%
Quidel QuickVue In-Line - 25 Test		Positive	1	50.00%
		Strep A Ag Det. (International)		
RS-11		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Alere Clearview Exact II Cas	1	*Positive	1	100.00%
RS-12		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Alere Clearview Exact II Cas	1	*Negative	1	100.00%
RS-13		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%



Alere Clearview Exact II Cas	1	*Positive	1	100.00%
RS-14		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Alere Clearview Exact II Cas	1	*Positive	1	100.00%
RS-15		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Alere Clearview Exact II Cas	1	*Negative	1	100.00%
739 - Sickle Cell Screen	Labs	Response	Number	Percent
		Sickle Cell Screen		
SS-3		* Denotes Correct Response		
AG: All Methods	12			
		*Negative	12	100.00%
Pacific Hemostasis Sickle Screen	3	*Negative	3	100.00%
Streck SickleDex	8	*Negative	8	100.00%
ASI Sickle Cell Test	1	*Negative	1	100.00%
679 - Syphilis Serology	Labs	Response	Number	Percent
		RPR		
SY-11		* Denotes Correct Response		
AG: All Methods	9			
		*Reactive	6	66.67%
		*Weakly Reactive	3	33.33%
ASI RPR	2	*Reactive	1	50.00%
Becton Dickinson BBL	3	*Reactive	2	66.67%
New Horizons	1	*Reactive	1	100.00%
Stanbio Quicktest	1	*Reactive	1	100.00%
Intl ASI RPR	1	*Reactive	1	100.00%
ASI RPR	2	*Weakly Reactive	1	50.00%
Becton Dickinson BBL	3	*Weakly Reactive	1	33.33%
Wampole Impact	1	*Weakly Reactive	1	100.00%
SY-12		* Denotes Correct Response		
AG: All Methods	9			
		*Nonreactive	9	100.00%
ASI RPR	2	*Nonreactive	2	100.00%
Becton Dickinson BBL	3	*Nonreactive	3	100.00%
New Horizons	1	*Nonreactive	1	100.00%
Stanbio Quicktest	1	*Nonreactive	1	100.00%
Wampole Impact	1	*Nonreactive	1	100.00%
Intl ASI RPR	1	*Nonreactive	1	100.00%
SY-13		* Denotes Correct Response		
AG: All Methods	9			
		*Reactive	9	100.00%
ASI RPR	2	*Reactive	2	100.00%
Becton Dickinson BBL	3	*Reactive	3	100.00%
New Horizons	1	*Reactive	1	100.00%
Stanbio Quicktest	1	*Reactive	1	100.00%
Wampole Impact	1	*Reactive	1	100.00%
Intl ASI RPR	1	*Reactive	1	100.00%
SY-14		* Denotes Correct Response		
AG: All Methods	9			
		*Nonreactive	9	100.00%
ASI RPR	2	*Nonreactive	2	100.00%
Becton Dickinson BBL	3	*Nonreactive	3	100.00%
New Horizons	1	*Nonreactive	1	100.00%
Stanbio Quicktest	1	*Nonreactive	1	100.00%
Wampole Impact	1	*Nonreactive	1	100.00%
Intl ASI RPR	1	*Nonreactive	1	100.00%
SY-15		* Denotes Correct Response		
AG: All Methods	9			
		*Reactive	6	66.67%
		*Weakly Reactive	3	33.33%
ASI RPR	2	*Reactive	1	50.00%
Becton Dickinson BBL	3	*Reactive	2	66.67%
New Horizons	1	*Reactive	1	100.00%
Stanbio Quicktest	1	*Reactive	1	100.00%
Intl ASI RPR	1	*Reactive	1	100.00%
ASI RPR	2	*Weakly Reactive	1	50.00%



Becton Dickinson BBL	3	*Weakly Reactive	1	33.33%
Wampole Impact	1	*Weakly Reactive	1	100.00%
		RPR (International)		
SY-11		* Denotes Correct Response		
AG: All Methods	3			
		*Reactive	3	100.00%
BD Macro-Vue RPR	1	*Reactive	1	100.00%
Fisher HC Sure-Vue RPR	1	*Reactive	1	100.00%
Wampole Impact	1	*Reactive	1	100.00%
SY-12		* Denotes Correct Response		
AG: All Methods	3			
		*Nonreactive	3	100.00%
BD Macro-Vue RPR	1	*Nonreactive	1	100.00%
Fisher HC Sure-Vue RPR	1	*Nonreactive	1	100.00%
Wampole Impact	1	*Nonreactive	1	100.00%
SY-13		* Denotes Correct Response		
AG: All Methods	3			
		*Reactive	3	100.00%
BD Macro-Vue RPR	1	*Reactive	1	100.00%
Fisher HC Sure-Vue RPR	1	*Reactive	1	100.00%
Wampole Impact	1	*Reactive	1	100.00%
SY-14		* Denotes Correct Response		
AG: All Methods	3			
		*Nonreactive	2	66.67%
		Weakly Reactive	1	33.33%
BD Macro-Vue RPR	1	*Nonreactive	1	100.00%
Fisher HC Sure-Vue RPR	1	*Nonreactive	1	100.00%
Wampole Impact	1	Weakly Reactive	1	100.00%
SY-15		* Denotes Correct Response		
AG: All Methods	3			
		*Reactive	2	66.67%
		Nonreactive	1	33.33%
BD Macro-Vue RPR	1	*Reactive	1	100.00%
Fisher HC Sure-Vue RPR	1	*Reactive	1	100.00%
Wampole Impact	1	Nonreactive	1	100.00%
776 - Throat Culture;780 - Throat & Genital Cultures	Labs	Response	Number	Percent
		Throat Culture		
TC-11		* Denotes Correct Response		
AG: All Methods	71			
		*Beta hemolysis, referred for ID/confirmation	1	1.41%
		*Gram Positive organism	1	1.41%
		*Positive for Group A Streptococcus	38	53.52%
		*Presumptive Positive for Group A Streptococcus	29	40.85%
		*Presumptive Streptococcus species	1	1.41%
		Negative for Group A Streptococcus	1	1.41%
Cultures	71	*Beta hemolysis, referred for ID/confirmation	1	1.41%
Cultures		*Gram Positive organism	1	1.41%
Cultures		*Positive for Group A Streptococcus	38	53.52%
Cultures		*Presumptive Positive for Group A Streptococcus	29	40.85%
Cultures		*Presumptive Streptococcus species	1	1.41%
Cultures		Negative for Group A Streptococcus	1	1.41%
TC-12		* Denotes Correct Response		
AG: All Methods	71			
		*Growth, referred for identification	2	2.82%
		*Negative for Group A Streptococcus	58	81.69%
		*No beta hemolysis	3	4.23%
		*No Growth - Strep Selective medium	8	11.27%
Cultures	71	*Growth, referred for identification	2	2.82%
Cultures		*Negative for Group A Streptococcus	58	81.69%
Cultures		*No beta hemolysis	3	4.23%
Cultures		*No Growth - Strep Selective medium	8	11.27%
TC-13		* Denotes Correct Response		
AG: All Methods	73			
		*Growth, referred for identification	1	1.37%
		*Negative for Group A Streptococcus	58	79.45%
		*No beta hemolysis	3	4.11%
		*No Growth - Strep Selective medium	3	4.11%
		Positive for Group A Streptococcus	5	6.85%



		Presumptive Positive for Group A Streptococcus	2	2.74%
		Streptococcus, beta-hemolytic, not Group A	1	1.37%
Cultures	73	*Growth, referred for identification	1	1.37%
Cultures		*Negative for Group A Streptococcus	58	79.45%
Cultures		*No beta hemolysis	3	4.11%
Cultures		*No Growth - Strep Selective medium	3	4.11%
Cultures		Positive for Group A Streptococcus	5	6.85%
Cultures		Presumptive Positive for Group A Streptococcus	2	2.74%
Cultures		Streptococcus, beta-hemolytic, not Group A	1	1.37%
TC-14		* Denotes Correct Response		
AG: All Methods	90			
		*Growth, referred for identification	1	1.11%
		*Negative for Group A Streptococcus	67	74.44%
		*No beta hemolysis	5	5.56%
		*No Growth - Strep Selective medium	15	16.67%
		Presumptive Positive for Group A Streptococcus	2	2.22%
Cultures	90	*Growth, referred for identification	1	1.11%
Cultures		*Negative for Group A Streptococcus	67	74.44%
Cultures		*No beta hemolysis	5	5.56%
Cultures		*No Growth - Strep Selective medium	15	16.67%
Cultures		Presumptive Positive for Group A Streptococcus	2	2.22%
TC-15		* Denotes Correct Response		
AG: All Methods	90			
		*Beta hemolysis, referred for ID/confirmation	8	8.89%
		*Gram Positive organism	1	1.11%
		*Growth, referred for identification	1	1.11%
		*Negative for Group A Streptococcus	54	60.00%
		*No Growth - Strep Selective medium	4	4.44%
		*Presumptive Positive for Group A Streptococcus	5	5.56%
		*Presumptive Streptococcus species	1	1.11%
		*Streptococcus, beta-hemolytic, not Group A	9	10.00%
		Positive for Group A Streptococcus	7	7.78%
Cultures	90	*Beta hemolysis, referred for ID/confirmation	8	8.89%
Cultures		*Gram Positive organism	1	1.11%
Cultures		*Growth, referred for identification	1	1.11%
Cultures		*Negative for Group A Streptococcus	54	60.00%
Cultures		*No Growth - Strep Selective medium	4	4.44%
Cultures		*Presumptive Positive for Group A Streptococcus	5	5.56%
Cultures		*Presumptive Streptococcus species	1	1.11%
Cultures		*Streptococcus, beta-hemolytic, not Group A	9	10.00%
Cultures		Positive for Group A Streptococcus	7	7.78%
977 - Trichomonas Ag Det (Waived)	Labs	Response	Number	Percent
		Trichomonas Ag Det (Waived)		
TRA-5		* Denotes Correct Response		
AG: All Methods	87			
		*Positive	81	93.10%
		Negative	6	6.90%
Sekisui Diagnostics OSOM Trichomonas	87	*Positive	81	93.10%
Sekisui Diagnostics OSOM Trichomonas		Negative	6	6.90%
TRA-6		* Denotes Correct Response		
AG: All Methods	87			
		*Positive	78	89.66%
		Negative	9	10.34%
Sekisui Diagnostics OSOM Trichomonas	87	*Positive	78	89.66%
Sekisui Diagnostics OSOM Trichomonas		Negative	9	10.34%
640 - Urinalysis;642 - Urinalysis/hCG;639 - Complete Urinalysis Package;601 - Waived/PPM Tests II;609 - Waived/PPM Tests I;647 - Mini-Waived Package;999 - CVS Complete Urinalysis;628 - AFC Waived/PPM	Labs	Response	Number	Percent
		Urine Sp Gravity		
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	15			
		*1.015	6	40.00%
		*1.020	6	40.00%
		*1.025	3	20.00%
UA-3		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	32			
		*1.015	22	68.75%



		*1.020	10	31.25%
UA-3		* Denotes Correct Response		
PG: Roche Urisys / Roche - Arbitrary / Roche Urine Arbitrary Result	1			
		*1.015	1	100.00%
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek Advantus / Plus Sys - Siemens / Siemens Urine Dipstick	1			
		*1.020	1	100.00%
UA-3		* Denotes Correct Response		
MG: Method Group	1256			
		*1.015	337	26.83%
		*1.020	653	51.99%
		*1.025	174	13.85%
		*1.030	83	6.61%
		1.000	1	0.08%
		1.010	8	0.64%
UA-3		* Denotes Correct Response		
MG: Method Group	32			
		*1.015	22	68.75%
		*1.020	10	31.25%
UA-3		* Denotes Correct Response		
MG: Method Group	16			
		*1.015	6	37.50%
		*1.020	7	43.75%
		*1.025	3	18.75%
UA-3		* Denotes Correct Response		
MG: Method Group	1			
		*1.015	1	100.00%
		Urine Bilirubin		
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	15			
		*Negative	15	100.00%
UA-3		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	33			
		*Negative	33	100.00%
UA-3		* Denotes Correct Response		
PG: Roche Urisys / Roche - Arbitrary / Roche Urine Arbitrary Result	1			
		*Negative	1	100.00%
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek Advantus / Plus Sys - Siemens / Siemens Urine Dipstick	1			
		*Negative	1	100.00%
UA-3		* Denotes Correct Response		
MG: Method Group	1244			
		*Negative	1232	99.04%
		Moderate (2+)	8	0.64%
		Trace/Small (1+)	4	0.32%
UA-3		* Denotes Correct Response		
MG: Method Group	33			
		*Negative	33	100.00%
UA-3		* Denotes Correct Response		
MG: Method Group	16			
		*Negative	16	100.00%
UA-3		* Denotes Correct Response		
MG: Method Group	1			
		*Negative	1	100.00%
		Urine Ketones		
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	15			
		*1+	2	13.33%
		*2+	13	86.67%
UA-3		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	33			
		*15 mg/dL	13	39.39%
		*50 mg/dL	20	60.61%



UA-3		* Denotes Correct Response		
PG: Roche Urisys / Roche - Arbitrary / Roche Urine Arbitrary Result	1			
		***	1	100.00%
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek Advantus / Plus Sys - Siemens / Siemens Urine Dipstick	1			
		*2+	1	100.00%
UA-3		* Denotes Correct Response		
MG: Method Group	1264			
		*Large (3+) / 80 mg/dL	133	10.52%
		*Moderate (2+) / 40 mg/dL	997	78.88%
		*Small (1+) / 15mg/dL	116	9.18%
		>=160 mg/dL	1	0.08%
		Negative	10	0.79%
		Trace/5 mg/dL	7	0.55%
UA-3		* Denotes Correct Response		
MG: Method Group	33			
		*15 mg/dL	13	39.39%
		*50 mg/dL	20	60.61%
UA-3		* Denotes Correct Response		
MG: Method Group	16			
		*1+	2	12.50%
		*2+	14	87.50%
UA-3		* Denotes Correct Response		
MG: Method Group	1			
		***	1	100.00%
UA-3		Urine pH		
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	15			
		*5.0	6	40.00%
		*5.5	9	60.00%
UA-3		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	33			
		*5	33	100.00%
UA-3		* Denotes Correct Response		
PG: Roche Urisys / Roche - Arbitrary / Roche Urine Arbitrary Result	1			
		*5.0	1	100.00%
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek Advantus / Plus Sys - Siemens / Siemens Urine Dipstick	1			
		*5.5	1	100.00%
UA-3		* Denotes Correct Response		
MG: Method Group	1266			
		*5.0 or less	592	46.76%
		*5.5	501	39.57%
		*6.0	164	12.95%
		6.5	5	0.39%
		7.0	4	0.32%
UA-3		* Denotes Correct Response		
MG: Method Group	33			
		*5	33	100.00%
UA-3		* Denotes Correct Response		
MG: Method Group	16			
		*5.0	6	37.50%
		*5.5	10	62.50%
UA-3		* Denotes Correct Response		
MG: Method Group	1			
		*5.0	1	100.00%
UA-3		Urine Blood/Hemoglobin		
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	15			
		*Negative	15	100.00%
UA-3		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	33			
		*Negative	33	100.00%



UA-3		* Denotes Correct Response		
PG: Roche Urisys / Roche - Arbitrary / Roche Urine Arbitrary Result	1			
		*Negative	1	100.00%
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek Advantus / Plus Sys - Siemens / Siemens Urine Dipstick	1			
		*Negative	1	100.00%
UA-3		* Denotes Correct Response		
MG: Method Group	1269			
		*Negative	1211	95.43%
		Large (3+)	3	0.24%
		Moderate (2+)	2	0.16%
		Small (1+)	3	0.24%
		Trace	50	3.94%
UA-3		* Denotes Correct Response		
MG: Method Group	33			
		*Negative	33	100.00%
UA-3		* Denotes Correct Response		
MG: Method Group	16			
		*Negative	16	100.00%
UA-3		* Denotes Correct Response		
MG: Method Group	1			
		*Negative	1	100.00%
		Urine Protein - Qual		
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	15			
		*Negative	15	100.00%
UA-3		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	33			
		*Negative	31	93.94%
		500 mg/dL	1	3.03%
		Trace	1	3.03%
UA-3		* Denotes Correct Response		
PG: Roche Urisys / Roche - Arbitrary / Roche Urine Arbitrary Result	1			
		*Negative	1	100.00%
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek Advantus / Plus Sys - Siemens / Siemens Urine Dipstick	1			
		*Negative	1	100.00%
UA-3		* Denotes Correct Response		
MG: Method Group	1278			
		*Negative	1230	96.24%
		100 (2+)	2	0.16%
		30 (1+)	5	0.39%
		300-500 (3+)	4	0.31%
		Trace	37	2.90%
UA-3		* Denotes Correct Response		
MG: Method Group	33			
		*Negative	31	93.94%
		500 mg/dL	1	3.03%
		Trace	1	3.03%
UA-3		* Denotes Correct Response		
MG: Method Group	16			
		*Negative	16	100.00%
UA-3		* Denotes Correct Response		
MG: Method Group	1			
		*Negative	1	100.00%
		Urine Glucose (Reducing Sub)		
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek 10 / 100	13			
		*≥ 1000 mg/dL	3	23.08%
		*100 mg/dL	1	7.69%
		*250 mg/dL	5	38.46%
		*500 mg/dL	2	15.38%
		<100 mg/dL	2	15.38%
UA-3		* Denotes Correct Response		



PG: Siemens Clinitek 50	95			
		*100 mg/dL	15	15.79%
		*250 mg/dL	78	82.11%
		*500 mg/dL	1	1.05%
		<100 mg/dL	1	1.05%
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek 500	5			
		*100 mg/dL	2	40.00%
		*250 mg/dL	3	60.00%
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek Atlas	1			
		*250 mg/dL	1	100.00%
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek Status	624			
		*100 mg/dL	68	10.90%
		*250 mg/dL	535	85.74%
		*500 mg/dL	3	0.48%
		<100 mg/dL	6	0.96%
		>= 1000 mg/dL	1	0.16%
		Negative	11	1.76%
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	15			
		*1+	14	93.33%
		*Trace	1	6.67%
UA-3		* Denotes Correct Response		
PG: Siemens Multistix - Visual	252			
		*100 mg/dL	33	13.10%
		*250 mg/dL	184	73.02%
		*500 mg/dL	8	3.17%
		<100 mg/dL	13	5.16%
		Negative	14	5.56%
UA-3		* Denotes Correct Response		
PG: Cole Taylor Strips	1			
		*250 mg/dL	1	100.00%
UA-3		* Denotes Correct Response		
PG: Germaine Aimstick	4			
		*100 mg/dL	2	50.00%
		*250 mg/dL	2	50.00%
UA-3		* Denotes Correct Response		
PG: Henry Schein Urispec	9			
		*100 mg/dL	1	11.11%
		*250 mg/dL	3	33.33%
		*500 mg/dL	5	55.56%
UA-3		* Denotes Correct Response		
PG: Hypoguard DiaScreen - Visual	2			
		*250 mg/dL	2	100.00%
UA-3		* Denotes Correct Response		
PG: Iris Diagnostics vChem Strips	1			
		*250 mg/dL	1	100.00%
UA-3		* Denotes Correct Response		
PG: Consult Diagnostics Urine Reagent Strips	6			
		*250 mg/dL	3	50.00%
		*500 mg/dL	3	50.00%
UA-3		* Denotes Correct Response		
PG: Quidel QuickVue UrinChek - Visual	1			
		*250 mg/dL	1	100.00%
UA-3		* Denotes Correct Response		
PG: Diag Test Group Clarity Urochek	34			
		*100 mg/dL	6	17.65%
		*250 mg/dL	19	55.88%
		*500 mg/dL	6	17.65%
		<100 mg/dL	2	5.88%
		Negative	1	2.94%
UA-3		* Denotes Correct Response		
PG: Roche Chemstrips - Visual	65			
		*>= 1000 mg/dL	54	83.08%
		*250 mg/dL	5	7.69%



		*500 mg/dL	3	4.62%
		100 mg/dL	2	3.08%
		Negative	1	1.54%
UA-3		* Denotes Correct Response		
PG: Roche SuperUA/ChemstripUA	1			
		*≥/ 1000 mg/dL	1	100.00%
UA-3		* Denotes Correct Response		
PG: Roche Urisys	51			
		*≥/ 1000 mg/dL	44	86.27%
		*250 mg/dL	5	9.80%
		*500 mg/dL	2	3.92%
UA-3		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	33			
		*≥/ 1000 mg/dL	32	96.97%
		250 mg/dL	1	3.03%
UA-3		* Denotes Correct Response		
PG: Roche Urisys / Roche - Arbitrary / Roche Urine Arbitrary Result	1			
		*+++	1	100.00%
UA-3		* Denotes Correct Response		
PG: UriScan Reagent Strips	6			
		*≥/ 1000 mg/dL	2	33.33%
		*100 mg/dL	1	16.67%
		*250 mg/dL	2	33.33%
		Negative	1	16.67%
UA-3		* Denotes Correct Response		
PG: Siemens Uristix	2			
		*250 mg/dL	2	100.00%
UA-3		* Denotes Correct Response		
PG: Medi-Lab Performance Urine Strip Analyzer	1			
		*250 mg/dL	1	100.00%
UA-3		* Denotes Correct Response		
PG: Jant Pharm Accutest URS 10	2			
		*250 mg/dL	1	50.00%
		*500 mg/dL	1	50.00%
UA-3		* Denotes Correct Response		
PG: Medi-Lab Performance Urine Strip - Visual	7			
		*250 mg/dL	5	71.43%
		*500 mg/dL	1	14.29%
		Negative	1	14.29%
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek Advantus	17			
		*100 mg/dL	2	11.76%
		*250 mg/dL	14	82.35%
		<100 mg/dL	1	5.88%
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek Advantus / Plus Sys - Siemens / Siemens Urine Dipstick	1			
		*1+	1	100.00%
UA-3		* Denotes Correct Response		
PG: Consult Diag Urinalysis Test Strips	13			
		*100 mg/dL	1	7.69%
		*250 mg/dL	8	61.54%
		*500 mg/dL	4	30.77%
UA-3		* Denotes Correct Response		
PG: Consult Diag Urine Analyzer	10			
		*100 mg/dL	1	10.00%
		*250 mg/dL	5	50.00%
		*500 mg/dL	4	40.00%
UA-3		* Denotes Correct Response		
PG: McKesson 120 Urine Analyzer	35			
		*100 mg/dL	4	11.43%
		*250 mg/dL	16	45.71%
		*500 mg/dL	13	37.14%
		<100 mg/dL	2	5.71%
UA-3		* Denotes Correct Response		
PG: Medline Urinalysis Reagent Strips	2			
		*100 mg/dL	2	100.00%



UA-3		* Denotes Correct Response		
PG: NDC ProAdvantage Urine Strip Analyzer	2			
		*250 mg/dL	1	50.00%
		*500 mg/dL	1	50.00%
UA-3		* Denotes Correct Response		
PG: McKesson 10 SG Urine Dipstick	2			
		*500 mg/dL	2	100.00%
UA-3		* Denotes Correct Response		
PG: Siemens Hema Combistix	1			
		*250 mg/dL	1	100.00%
UA-3		* Denotes Correct Response		
PG: Teco Diag. URS-3	1			
		*100 mg/dL	1	100.00%
UA-3		* Denotes Correct Response		
PG: LW Scientific Urine Reagent Strip	1			
		*100 mg/dL	1	100.00%
UA-3		* Denotes Correct Response		
MG: Method Group	1277			
		*>= 1000 mg/dL	110	8.61%
		*100 mg/dL	144	11.28%
		*250 mg/dL	905	70.87%
		*500 mg/dL	60	4.70%
		<100 mg/dL	29	2.27%
		>= 1000 mg/dL	110	8.61%
		100 mg/dL	144	11.28%
		Negative	29	2.27%
UA-3		* Denotes Correct Response		
MG: Method Group	33			
		*>= 1000 mg/dL	32	96.97%
		250 mg/dL	1	3.03%
UA-3		* Denotes Correct Response		
MG: Method Group	16			
		*1+	15	93.75%
		*Trace	1	6.25%
UA-3		* Denotes Correct Response		
MG: Method Group	1			
		*+++	1	100.00%
		Urine Urobilinogen		
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	7			
		*3.2 umol/L	7	100.00%
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Urobilinogen	7			
		*0.2 EU/dL	7	100.00%
UA-3		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	33			
		*Normal	33	100.00%
UA-3		* Denotes Correct Response		
PG: Roche Urisys / Roche - Arbitrary / Roche Urine Arbitrary Result	1			
		*Normal	1	100.00%
UA-3		* Denotes Correct Response		
MG: Method Group	1238			
		*0.2/Normal	1223	98.79%
		>= 8.0 mg/dL	1	0.08%
		2.0 mg/dL	9	0.73%
		4.0 mg/dL	5	0.40%
UA-3		* Denotes Correct Response		
MG: Method Group	33			
		*Normal	33	100.00%
UA-3		* Denotes Correct Response		
MG: Plus Sys - Siemens	15			
		*0.2 EU/dL	7	46.67%
		*3.2 umol/L	8	53.33%
UA-3		* Denotes Correct Response		
MG: Plus Sys - Siemens	1			



		*Normal	1	100.00%
		Urine Leuk Esterase		
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek 10 / 100	13			
		*Large (3+)	2	15.38%
		*Moderate (2+)	6	46.15%
		*Small (1+)	5	38.46%
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek 50	95			
		*Moderate (2+)	25	26.32%
		*Small (1+)	63	66.32%
		*Trace	5	5.26%
		Large (3+)	1	1.05%
		Negative	1	1.05%
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek 500	5			
		*Moderate (2+)	2	40.00%
		*Small (1+)	3	60.00%
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek Status	621			
		*Moderate (2+)	204	32.85%
		*Small (1+)	354	57.00%
		*Trace	43	6.92%
		Large (3+)	16	2.58%
		Negative	4	0.64%
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	15			
		*1+	7	46.67%
		*2+	6	40.00%
		*Trace	2	13.33%
UA-3		* Denotes Correct Response		
PG: Siemens Multistix - Visual	250			
		*Moderate (2+)	138	55.20%
		*Small (1+)	91	36.40%
		*Trace	7	2.80%
		Large (3+)	9	3.60%
		Negative	5	2.00%
UA-3		* Denotes Correct Response		
PG: Germaine Aimstick	4			
		*Moderate (2+)	1	25.00%
		*Small (1+)	2	50.00%
		*Trace	1	25.00%
UA-3		* Denotes Correct Response		
PG: Henry Schein Urispec	7			
		*Large (3+)	6	85.71%
		*Moderate (2+)	1	14.29%
UA-3		* Denotes Correct Response		
PG: Hypoguard DiaScreen - Visual	2			
		*Moderate (2+)	1	50.00%
		*Small (1+)	1	50.00%
UA-3		* Denotes Correct Response		
PG: Biosys Uriscan Optima II Urine Analyzer	3			
		*Moderate (2+)	2	66.67%
		*Small (1+)	1	33.33%
UA-3		* Denotes Correct Response		
PG: Consult Diagnostics Urine Reagent Strips	6			
		*Moderate (2+)	4	66.67%
		*Small (1+)	2	33.33%
UA-3		* Denotes Correct Response		
PG: Diag Test Group Clarity Urochek	34			
		*Large (3+)	4	11.76%
		*Moderate (2+)	19	55.88%
		*Small (1+)	10	29.41%
		Trace	1	2.94%
UA-3		* Denotes Correct Response		
PG: Roche Chemstrip 101 Analyzer	3			
		*Large (3+)	1	33.33%



		*Moderate (2+)	1	33.33%
		*Small (1+)	1	33.33%
UA-3		* Denotes Correct Response		
PG: Roche Chemstrips - Visual	65			
		*Large (3+)	9	13.85%
		*Moderate (2+)	50	76.92%
		*Small (1+)	6	9.23%
UA-3		* Denotes Correct Response		
PG: Roche SuperUA/ChemstripUA	1			
		*Moderate (2+)	1	100.00%
UA-3		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	33			
		*500 Leu/uL	33	100.00%
UA-3		* Denotes Correct Response		
PG: Roche Urisys / Roche - Arbitrary / Roche Urine Arbitrary Result	1			
		*++	1	100.00%
UA-3		* Denotes Correct Response		
PG: UriScan Reagent Strips	5			
		*Large (3+)	2	40.00%
		*Moderate (2+)	2	40.00%
		*Small (1+)	1	20.00%
UA-3		* Denotes Correct Response		
PG: Siemens Uristix	2			
		*Moderate (2+)	2	100.00%
UA-3		* Denotes Correct Response		
PG: Medi-Lab Performance Urine Strip Analyzer	1			
		*Small (1+)	1	100.00%
UA-3		* Denotes Correct Response		
PG: Jant Pharm Accutest URS 10	2			
		*Large (3+)	2	100.00%
UA-3		* Denotes Correct Response		
PG: Medi-Lab Performance Urine Strip - Visual	7			
		*Moderate (2+)	5	71.43%
		*Small (1+)	2	28.57%
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek Advantus	17			
		*Moderate (2+)	7	41.18%
		*Small (1+)	10	58.82%
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek Advantus / Plus Sys - Siemens / Siemens Urine Dipstick	1			
		*1+	1	100.00%
UA-3		* Denotes Correct Response		
PG: Consult Diag Urinalysis Test Strips	13			
		*Moderate (2+)	9	69.23%
		*Small (1+)	4	30.77%
UA-3		* Denotes Correct Response		
PG: Consult Diag Urine Analyzer	10			
		*Large (3+)	3	30.00%
		*Moderate (2+)	5	50.00%
		*Small (1+)	1	10.00%
		Trace	1	10.00%
UA-3		* Denotes Correct Response		
PG: McKesson 120 Urine Analyzer	36			
		*Moderate (2+)	12	33.33%
		*Small (1+)	20	55.56%
		*Trace	3	8.33%
		Large (3+)	1	2.78%
UA-3		* Denotes Correct Response		
PG: Medline Urinalysis Reagent Strips	2			
		*Moderate (2+)	1	50.00%
		*Small (1+)	1	50.00%
UA-3		* Denotes Correct Response		
PG: NDC ProAdvantage Urine Strip Analyzer	2			
		*Large (3+)	1	50.00%
		*Moderate (2+)	1	50.00%
UA-3		* Denotes Correct Response		



PG: Roche Cobas u 411	2			
		*Large (3+)	2	100.00%
UA-3		* Denotes Correct Response		
PG: McKesson 10 SG Urine Dipstick	2			
		*Moderate (2+)	1	50.00%
		*Small (1+)	1	50.00%
UA-3		* Denotes Correct Response		
PG: LW Scientific Urine Reagent Strip	1			
		*Small (1+)	1	100.00%
UA-3		* Denotes Correct Response		
MG: Method Group	1268			
		*Large (3+)	100	7.89%
		*Moderate (2+)	510	40.22%
		*Small (1+)	585	46.14%
		*Trace	63	4.97%
		Large (3+)	100	7.89%
		Moderate (2+)	510	40.22%
		Negative	10	0.79%
		Small (1+)	585	46.14%
		Trace	63	4.97%
UA-3		* Denotes Correct Response		
MG: Method Group	33			
		*500 Leu/uL	33	100.00%
UA-3		* Denotes Correct Response		
MG: Method Group	16			
		*1+	8	50.00%
		*2+	6	37.50%
		*Trace	2	12.50%
UA-3		* Denotes Correct Response		
MG: Method Group	1			
		*++	1	100.00%
UA-3		Urine Nitrite		
		* Denotes Correct Response		
PG: Siemens Clinitek Status / Plus Sys - Siemens / Siemens Urine Dipstick	15			
		*Positive	15	100.00%
UA-3		* Denotes Correct Response		
PG: Roche Urisys / Conventional - Roche / Roche Urine Dipstick	33			
		*Positive	33	100.00%
UA-3		* Denotes Correct Response		
PG: Roche Urisys / Roche - Arbitrary / Roche Urine Arbitrary Result	1			
		*+	1	100.00%
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek Advantus / Plus Sys - Siemens / Siemens Urine Dipstick	1			
		*Positive	1	100.00%
UA-3		* Denotes Correct Response		
MG: Method Group	1267			
		*Positive	1253	98.90%
		Negative	14	1.10%
UA-3		* Denotes Correct Response		
MG: Method Group	33			
		*Positive	33	100.00%
UA-3		* Denotes Correct Response		
MG: Method Group	16			
		*Positive	16	100.00%
UA-3		* Denotes Correct Response		
MG: Method Group	1			
		*+	1	100.00%
UA-3		Urine Microalbumin		
		* Denotes Correct Response		
PG: Siemens Clinitek 10 / 100	2			
		*10 mg/L (Trace)	1	50.00%
		*Negative	1	50.00%
UA-3		* Denotes Correct Response		
PG: Siemens Clinitek 50	19			
		*10 mg/L (Trace)	19	100.00%



UA-3		* Denotes Correct Response	
PG: Siemens Clinitek Status	45		
		*10 mg/L (Trace)	39 86.67%
		*Negative	5 11.11%
		20/30 mg/L (Small)	1 2.22%
UA-3		* Denotes Correct Response	
PG: Siemens Multistix - Visual	5		
		*Negative	3 60.00%
		20/30 mg/L (Small)	2 40.00%
UA-3		* Denotes Correct Response	
PG: Roche Chemstrip Micral	31		
		*10 mg/L (Trace)	4 12.90%
		*20/30 mg/L (Small)	14 45.16%
		*Negative	10 32.26%
		50 mg/L (2+)	3 9.68%
UA-3		* Denotes Correct Response	
PG: Roche Chemstrips - Visual	1		
		*20/30 mg/L (Small)	1 100.00%
UA-3		* Denotes Correct Response	
PG: Hemocue Albumin 201	1		
		*10 mg/L (Trace)	1 100.00%
UA-3		* Denotes Correct Response	
PG: Nitrazine Paper	9		
		*5.0 or less	1 11.11%
		*5.5	5 55.56%
		*6.0	2 22.22%
		6.5	1 11.11%
Urine Creatinine (semi-quant)		* Denotes Correct Response	
PG: Siemens Clinitek Status	4		
		*10 mg/dL	3 75.00%
		100 mg/dL	1 25.00%
UA-3		* Denotes Correct Response	
SG: Semi-quant Methods - Ur Microalbumin-Creat	0		
Urine hCG		* Denotes Correct Response	
AG: All Methods	1130		
		*Negative	1122 99.29%
		Positive	8 0.71%
ASI ProPhase Plus	1	*Negative	1 100.00%
Siemens Clinitek Status	13	*Negative	13 100.00%
Beck Coulter ICON 20 HCG	6	*Negative	6 100.00%
Beck Coulter ICON 25 hCG	80	*Negative	79 98.75%
Alere Acceava HCG Basic II	24	*Negative	22 91.67%
Alere Acceava hCG Combo	5	*Negative	5 100.00%
Alere Acceava hCG-Urine	20	*Negative	20 100.00%
Conception Tech One Step hCG	2	*Negative	2 100.00%
Diag Test Group Clarity hCG Plus Urine Cassettes	7	*Negative	7 100.00%
Fisher HC Sure-Vue hCG STAT	6	*Negative	6 100.00%
Fisher HC Sure-Vue Urine hCG	16	*Negative	16 100.00%
Fisher HC Sure-Vue U/S hCG	5	*Negative	5 100.00%
Sekisui Diagnostics OSOM Urine	12	*Negative	12 100.00%
Sekisui Diagnostics OSOM Card Pregnancy	33	*Negative	33 100.00%
Sekisui Diagnostics OSOM hCG Combo	46	*Negative	45 97.83%
Henry Schein One Step + (hCG)	46	*Negative	46 100.00%
Immunostics Detector Combo	3	*Negative	3 100.00%
Immunostics hCG Detector-urine	4	*Negative	4 100.00%
Instant Tech. i Pregnancy	4	*Negative	4 100.00%
Inverness Clearview Combo II	3	*Negative	2 66.67%
Inv Clearview Urine - Cassette	3	*Negative	3 100.00%
Inverness Signify hCG combo	4	*Negative	4 100.00%
Inverness Signify hCG urine	4	*Negative	4 100.00%
LifeSign Status hCG	1	*Negative	1 100.00%
McKesson Medi-Lab hCG - Urine	40	*Negative	40 100.00%
Meridian Immunocard Stat!	1	*Negative	1 100.00%
Permaxim RediScreen	4	*Negative	4 100.00%
Polymedco Poly stat hCG Urine	1	*Negative	1 100.00%



Quidel QuickVue + hCG	26	*Negative	26	100.00%
Quidel QuickVue One-Step hCG Combo U/S	86	*Negative	86	100.00%
Quidel QuickVue One-Step Urine	173	*Negative	172	99.42%
Quidel RapidVue hCG - Urine	8	*Negative	8	100.00%
SAS Ultra HCG	3	*Negative	3	100.00%
Select Diag hCG Cassette (urine)	25	*Negative	25	100.00%
Select Diag hCG Combo	10	*Negative	10	100.00%
Select Diag hCG Dipstick	7	*Negative	7	100.00%
SP (Cardinal) Brand combo	16	*Negative	16	100.00%
SP (Cardinal) Brand-cassette	8	*Negative	8	100.00%
SP Brand (Cardinal) HCG Combo	35	*Negative	35	100.00%
Stanbio Qupid	5	*Negative	5	100.00%
Stanbio QuPID Plus	1	*Negative	1	100.00%
Stanbio QuStick	2	*Negative	2	100.00%
Stanbio TRUE hCG	4	*Negative	4	100.00%
MooreBrand hCG Cassette	3	*Negative	3	100.00%
Nova Scientific Century S/U hCG	1	*Negative	1	100.00%
Diag Test Group Clarity hCG Urine Strips	2	*Negative	2	100.00%
Jant AccuTest	6	*Negative	6	100.00%
Becton Dickinson Precise One-Step	1	*Negative	1	100.00%
Consult Diag hCG Urine Cassette	174	*Negative	173	99.43%
CLIAwaived Inc Pregnancy Urine Dip-Strip	5	*Negative	5	100.00%
BTNX Rapid Response hCG Strip (urine)	1	*Negative	1	100.00%
Germaine AIM Step Combo Pregnancy	1	*Negative	1	100.00%
NDC ProAdvantage Urine hCG Cassette	4	*Negative	4	100.00%
Consult Diag hCG Urine Dipstick	36	*Negative	36	100.00%
SP (Cardinal) Brand Dipstick	9	*Negative	9	100.00%
Consult Diag hCG Combo Cassette (Urine)	12	*Negative	12	100.00%
McKesson Medi-Lab Performance hCG Urine Cassette	8	*Negative	8	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	3	*Negative	3	100.00%
Alere hCG Cassette (25 mIU/mL)	12	*Negative	12	100.00%
Alere hCG Dipstick	6	*Negative	6	100.00%
Medline Pregnancy Urine Cassette Test	1	*Negative	1	100.00%
Quidel QuickVue hCG Urine Test	17	*Negative	16	94.12%
MooreBrand hCG Dipstick	2	*Negative	2	100.00%
Siemens Clinitest hCG	1	*Negative	1	100.00%
Henry Schein One Step + HCG Combo	2	*Negative	2	100.00%
Henry Schein One Step hCG Urine Cassette	17	*Negative	17	100.00%
Medline hCG Test Strip 25 mIU/mL	1	*Negative	1	100.00%
NCS hCG One Step Pregnancy Test	1	*Negative	1	100.00%
MediChoice Urine hCG Cassette	1	*Negative	1	100.00%
Beck Coulter ICON 25 hCG	80	Positive	1	1.25%
Alere Aceava HCG Basic II	24	Positive	2	8.33%
Sekisui Diagnostics OSOM hCG Combo	46	Positive	1	2.17%
Inverness Clearview Combo II	3	Positive	1	33.33%
Quidel QuickVue One-Step Urine	173	Positive	1	0.58%
Consult Diag hCG Urine Cassette	174	Positive	1	0.57%
Quidel QuickVue hCG Urine Test	17	Positive	1	5.88%
661 - Urine Colony Count;784 - Presump ID/Colony Count - Urine;785 - Presump ID/Colony Count-Urine & Strep Combo;775 - Urine Culture;778 - Urine & Throat Cultures;779 - Urine & Genital Cultures;781 - Urine, Throat & Genital Cultures;782 - Complete Bacteri	Labs	Response	Number	Percent
		Susceptibility Cefepime		
UC-11		* Denotes Correct Response		
AG: All Methods	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
		Urine Culture		
UC-11		* Denotes Correct Response		
AG: All Methods	33			
		*Gram Negative organism	3	9.09%
		*Growth, referred for identification	15	45.45%
		*Presumptive Pseudomonas species	4	12.12%
		*Pseudomonas aeruginosa	3	9.09%
		*Uricult - Growth (Positive)	8	24.24%
Cultures	33	*Gram Negative organism	3	9.09%
Cultures		*Growth, referred for identification	15	45.45%
Cultures		*Presumptive Pseudomonas species	4	12.12%
Cultures		*Pseudomonas aeruginosa	3	9.09%



Cultures		*Uricult - Growth (Positive)	8	24.24%
UC-12		* Denotes Correct Response		
AG: All Methods	33			
		*Gram Negative organism	4	12.12%
		*Growth, referred for identification	16	48.48%
		*Klebsiella pneumoniae	3	9.09%
		*Presumptive Klebsiella species	1	3.03%
		*Uricult - Growth (Positive)	4	12.12%
		Klebsiella oxytoca	1	3.03%
		Uricult - No Growth (Negative)	4	12.12%
Cultures	33	*Gram Negative organism	4	12.12%
Cultures		*Growth, referred for identification	16	48.48%
Cultures		*Klebsiella pneumoniae	3	9.09%
Cultures		*Presumptive Klebsiella species	1	3.03%
Cultures		*Uricult - Growth (Positive)	4	12.12%
Cultures		Klebsiella oxytoca	1	3.03%
Cultures		Uricult - No Growth (Negative)	4	12.12%
UC-13		* Denotes Correct Response		
AG: All Methods	31			
		*Gram Negative organism	2	6.45%
		*Growth, referred for identification	17	54.84%
		*Presumptive Proteus species	2	6.45%
		*Proteus mirabilis	2	6.45%
		*Uricult - Growth (Positive)	7	22.58%
		Uricult - No Growth (Negative)	1	3.23%
Cultures	31	*Gram Negative organism	2	6.45%
Cultures		*Growth, referred for identification	17	54.84%
Cultures		*Presumptive Proteus species	2	6.45%
Cultures		*Proteus mirabilis	2	6.45%
Cultures		*Uricult - Growth (Positive)	7	22.58%
Cultures		Uricult - No Growth (Negative)	1	3.23%
UC-14		* Denotes Correct Response		
AG: All Methods	8			
		*Gram Positive organism	1	12.50%
		*Presumptive Staphylococcus species	1	12.50%
		*Staphylococcus saprophyticus	2	25.00%
		*Uricult - Growth (Positive)	3	37.50%
		Streptococcus species, NOS	1	12.50%
Cultures	8	*Gram Positive organism	1	12.50%
Cultures		*Presumptive Staphylococcus species	1	12.50%
Cultures		*Staphylococcus saprophyticus	2	25.00%
Cultures		*Uricult - Growth (Positive)	3	37.50%
Cultures		Streptococcus species, NOS	1	12.50%
UC-15		* Denotes Correct Response		
AG: All Methods	8			
		*Escherichia coli	2	25.00%
		*Gram Negative organism	2	25.00%
		*Presumptive E. coli	1	12.50%
		*Uricult - Growth (Positive)	3	37.50%
Cultures	8	*Escherichia coli	2	25.00%
Cultures		*Gram Negative organism	2	25.00%
Cultures		*Presumptive E. coli	1	12.50%
Cultures		*Uricult - Growth (Positive)	3	37.50%
		Susceptibility Cefazidime		
UC-11		* Denotes Correct Response		
AG: All Methods	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
		Susceptibility Levofloxacin		
UC-11		* Denotes Correct Response		
AG: All Methods	5			
		*Intermediate	4	80.00%
		*Resistant	1	20.00%
Susceptibility	5	*Intermediate	4	80.00%
Susceptibility		*Resistant	1	20.00%
		Susceptibility Piperacillin		
UC-11		* Denotes Correct Response		
AG: All Methods	1			



		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
		Susceptibility Piperacillin/Tazobactam		
UC-11		* Denotes Correct Response		
AG: All Methods	3			
Susceptibility	3	*Sensitive	3	100.00%
		*Sensitive	3	100.00%
		Susceptibility Tobramycin		
UC-11		* Denotes Correct Response		
AG: All Methods	4			
Susceptibility	4	*Sensitive	4	100.00%
		*Sensitive	4	100.00%
		Susceptibility Amikacin		
UC-11		* Denotes Correct Response		
AG: All Methods	2			
Susceptibility	2	*Sensitive	2	100.00%
		*Sensitive	2	100.00%
		Susceptibility Aztreonam		
UC-11		* Denotes Correct Response		
AG: All Methods	1			
Susceptibility	1	*Sensitive	1	100.00%
		*Sensitive	1	100.00%
		Susceptibility Ertapenem		
UC-11		* Denotes Correct Response		
AG: All Methods	1			
		*Resistant	1	100.00%
Susceptibility	1	*Resistant	1	100.00%
703 - Urine Drug Screen;713 - Urine Drug Screen (50 ml sample volume)	Labs	Response	Number	Percent
		Barbiturates		
UDS-5		* Denotes Correct Response		
AG: All Methods	123			
		*Negative	122	99.19%
		Positive	1	0.81%
AlereTriage	3	*Negative	3	100.00%
E-Screen	7	*Negative	7	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Innovacon One Step	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	14	*Negative	14	100.00%
Instant Technologies i Screen	3	*Negative	3	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman (Olympus) AU	2	*Negative	2	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
BioTek	12	*Negative	12	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	3	75.00%
Mindray BS-200	2	*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Alere iCassette	3	*Negative	3	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	6	*Negative	6	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immutest Multi-Drug Screen Card	1	*Negative	1	100.00%
Mindray 480	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	24	*Negative	24	100.00%
Hemosure First Sign UDS Cup	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	Positive	1	25.00%
UDS-6		* Denotes Correct Response		
AG: All Methods	139			



		*Positive	135	97.12%
		Negative	4	2.88%
AlereTriage	3	*Positive	3	100.00%
E-Screen	23	*Positive	21	91.30%
Express Diag. Express Check	1	*Positive	1	100.00%
Innovacon One Step	1	*Positive	1	100.00%
Alere iCup Drugs of Abuse	14	*Positive	13	92.86%
Instant Technologies i Screen	3	*Positive	3	100.00%
MedTox Profile II	1	*Positive	1	100.00%
Noble Med Split Specimen Cup	1	*Positive	1	100.00%
Beckman (Olympus) AU	2	*Positive	2	100.00%
Innovacon Multi-Clin	10	*Positive	10	100.00%
BioTek	12	*Positive	12	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Positive	5	100.00%
ASC One Screen Drug Test Card	1	*Positive	1	100.00%
MedTox Profile V	2	*Positive	2	100.00%
Jant Accutest Multi Drug Screen	2	*Positive	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Positive	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Positive	4	100.00%
Mindray BS-200	2	*Positive	2	100.00%
Abbott Architect Ci4100	1	*Positive	1	100.00%
Alere iCassette	3	*Positive	3	100.00%
Roche Hitachi 717	1	*Positive	1	100.00%
US Diagnostics UScreen Cup	4	*Positive	4	100.00%
Abbott Architect c4000	1	*Positive	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Positive	1	100.00%
Diag Test Group Clarity DOA Screen	6	*Positive	5	83.33%
Noble Medical Insta-Screen Test Cup	3	*Positive	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Positive	1	100.00%
Mindray 480	1	*Positive	1	100.00%
Hemosure First Sign UDS Dipstick	24	*Positive	24	100.00%
Hemosure First Sign UDS Cup	1	*Positive	1	100.00%
E-Screen	23	Negative	2	8.70%
Alere iCup Drugs of Abuse	14	Negative	1	7.14%
Diag Test Group Clarity DOA Screen	6	Negative	1	16.67%
Benzodiazepines				
UDS-5				
* Denotes Correct Response				
AG: All Methods	126			
		*Negative	126	100.00%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	1	*Negative	1	100.00%
E-Screen	7	*Negative	7	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Innovacon One Step	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	15	*Negative	15	100.00%
Instant Technologies i Screen	3	*Negative	3	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Noble Med Split Specimen Cup	2	*Negative	2	100.00%
Beckman (Olympus) AU	2	*Negative	2	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
BioTek	12	*Negative	12	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Alere iCassette	3	*Negative	3	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	6	*Negative	6	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%



Mindray 480	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	24	*Negative	24	100.00%
Hemosure First Sign UDS Cup	1	*Negative	1	100.00%
UDS-6		* Denotes Correct Response		
AG: All Methods	142			
		*Negative	139	97.89%
		Positive	3	2.11%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	1	*Negative	1	100.00%
E-Screen	23	*Negative	20	86.96%
Express Diag. Express Check	1	*Negative	1	100.00%
Innovacon One Step	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	15	*Negative	15	100.00%
Instant Technologies i Screen	3	*Negative	3	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Noble Med Split Specimen Cup	2	*Negative	2	100.00%
Beckman (Olympus) AU	2	*Negative	2	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
BioTek	12	*Negative	12	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Alere iCassette	3	*Negative	3	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	6	*Negative	6	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
Mindray 480	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	24	*Negative	24	100.00%
Hemosure First Sign UDS Cup	1	*Negative	1	100.00%
E-Screen	23	Positive	3	13.04%
UDS-5		Amphetamines		
		* Denotes Correct Response		
AG: All Methods	130			
		*Negative	130	100.00%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	2	*Negative	2	100.00%
E-Screen	4	*Negative	4	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Innovacon One Step	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	18	*Negative	18	100.00%
Instant Technologies i Screen	3	*Negative	3	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	2	*Negative	2	100.00%
Beckman (Olympus) AU	2	*Negative	2	100.00%
Innovacon Multi-Clin	9	*Negative	9	100.00%
BioTek	12	*Negative	12	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Alere iCassette	6	*Negative	6	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%



Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	6	*Negative	6	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
Mindray 480	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	24	*Negative	24	100.00%
Hemosure First Sign UDS Cup	1	*Negative	1	100.00%
UDS-6		* Denotes Correct Response		
AG: All Methods	146			
		*Negative	144	98.63%
		Positive	2	1.37%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	2	*Negative	2	100.00%
E-Screen	20	*Negative	18	90.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Innovacon One Step	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	18	*Negative	18	100.00%
Instant Technologies i Screen	3	*Negative	3	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	2	*Negative	2	100.00%
Beckman (Olympus) AU	2	*Negative	2	100.00%
Innovacon Multi-Clin	9	*Negative	9	100.00%
BioTek	12	*Negative	12	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Alere iCassette	6	*Negative	6	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	6	*Negative	6	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
Mindray 480	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	24	*Negative	24	100.00%
Hemosure First Sign UDS Cup	1	*Negative	1	100.00%
E-Screen	20	Positive	2	10.00%
UDS-5		Cannabinoids (THC)		
		* Denotes Correct Response		
AG: All Methods	160			
		*Positive	152	95.00%
		Negative	8	5.00%
AlereTriage	3	*Positive	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	2	*Positive	2	100.00%
E-Screen	30	*Positive	23	76.67%
Express Diag. Express Check	1	*Positive	1	100.00%
Innovacon One Step	1	*Positive	1	100.00%
Alere iCup Drugs of Abuse	18	*Positive	17	94.44%
Instant Technologies i Screen	4	*Positive	4	100.00%
Integrated E-Z Split Key Cup	1	*Positive	1	100.00%
MedTox Profile II	2	*Positive	2	100.00%
Noble Med Split Specimen Cup	2	*Positive	2	100.00%
Beckman (Olympus) AU	2	*Positive	2	100.00%
Innovacon Multi-Clin	10	*Positive	10	100.00%
BioTek	12	*Positive	12	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Positive	5	100.00%
ASC One Screen Drug Test Card	1	*Positive	1	100.00%
MedTox Profile V	2	*Positive	2	100.00%
Jant Accutest Multi Drug Screen	2	*Positive	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Positive	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Positive	4	100.00%



Mindray BS-200	2	*Positive	2	100.00%
Abbott Architect Ci4100	1	*Positive	1	100.00%
Alere iCassette	6	*Positive	6	100.00%
Roche Hitachi 717	1	*Positive	1	100.00%
US Diagnostics UScreen Cup	4	*Positive	4	100.00%
Abbott Architect c4000	1	*Positive	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Positive	1	100.00%
Diag Test Group Clarity DOA Screen	6	*Positive	6	100.00%
Noble Medical Insta-Screen Test Cup	3	*Positive	3	100.00%
Alere Immutest Multi-Drug Screen Card	1	*Positive	1	100.00%
Mindray 480	1	*Positive	1	100.00%
CRL Stat Cup	1	*Positive	1	100.00%
Hemosure First Sign UDS Dipstick	24	*Positive	24	100.00%
Hemosure First Sign UDS Cup	1	*Positive	1	100.00%
E-Screen	30	Negative	7	23.33%
Alere iCup Drugs of Abuse	18	Negative	1	5.56%
UDS-6		* Denotes Correct Response		
AG: All Methods	159			
		*Negative	156	98.11%
		Positive	3	1.89%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	2	*Negative	2	100.00%
E-Screen	29	*Negative	26	89.66%
Express Diag. Express Check	1	*Negative	1	100.00%
Innovacon One Step	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	18	*Negative	18	100.00%
Instant Technologies i Screen	4	*Negative	4	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	2	*Negative	2	100.00%
Beckman (Olympus) AU	2	*Negative	2	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
BioTek	12	*Negative	12	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Alere iCassette	6	*Negative	6	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	6	*Negative	6	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immutest Multi-Drug Screen Card	1	*Negative	1	100.00%
Mindray 480	1	*Negative	1	100.00%
CRL Stat Cup	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	24	*Negative	24	100.00%
Hemosure First Sign UDS Cup	1	*Negative	1	100.00%
E-Screen	29	Positive	3	10.34%
UDS-5		* Denotes Correct Response		
AG: All Methods	159			
		*Negative	153	96.23%
		Positive	6	3.77%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	2	*Negative	2	100.00%
E-Screen	30	*Negative	24	80.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Innovacon One Step	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	18	*Negative	18	100.00%
Instant Technologies i Screen	4	*Negative	4	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	2	*Negative	2	100.00%



Noble Med Split Specimen Cup	2	*Negative	2	100.00%
Beckman (Olympus) AU	2	*Negative	2	100.00%
Innovacon Multi-Clin	9	*Negative	9	100.00%
BioTek	12	*Negative	12	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Alere iCassette	6	*Negative	6	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	6	*Negative	6	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
Mindray 480	1	*Negative	1	100.00%
CRL Stat Cup	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	24	*Negative	24	100.00%
Hemosure First Sign UDS Cup	1	*Negative	1	100.00%
E-Screen	30	Positive	6	20.00%
UDS-6		* Denotes Correct Response		
AG: All Methods	159			
		*Positive	150	94.34%
		Negative	9	5.66%
AlereTriage	3	*Positive	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	2	*Positive	2	100.00%
E-Screen	30	*Positive	23	76.67%
Express Diag. Express Check	1	*Positive	1	100.00%
Innovacon One Step	1	*Positive	1	100.00%
Alere iCup Drugs of Abuse	18	*Positive	18	100.00%
Instant Technologies i Screen	4	*Positive	3	75.00%
Integrated E-Z Split Key Cup	1	*Positive	1	100.00%
MedTox Profile II	2	*Positive	2	100.00%
Noble Med Split Specimen Cup	2	*Positive	2	100.00%
Beckman (Olympus) AU	2	*Positive	2	100.00%
Innovacon Multi-Clin	9	*Positive	9	100.00%
BioTek	12	*Positive	11	91.67%
CLIAwaived Inc Rapid Drug Cup	5	*Positive	5	100.00%
ASC One Screen Drug Test Card	1	*Positive	1	100.00%
MedTox Profile V	2	*Positive	2	100.00%
Jant Accutest Multi Drug Screen	2	*Positive	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Positive	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Positive	4	100.00%
Mindray BS-200	2	*Positive	2	100.00%
Abbott Architect Ci4100	1	*Positive	1	100.00%
Alere iCassette	6	*Positive	6	100.00%
Roche Hitachi 717	1	*Positive	1	100.00%
US Diagnostics UScreen Cup	4	*Positive	4	100.00%
Abbott Architect c4000	1	*Positive	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Positive	1	100.00%
Diag Test Group Clarity DOA Screen	6	*Positive	6	100.00%
Noble Medical Insta-Screen Test Cup	3	*Positive	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Positive	1	100.00%
Mindray 480	1	*Positive	1	100.00%
CRL Stat Cup	1	*Positive	1	100.00%
Hemosure First Sign UDS Dipstick	24	*Positive	24	100.00%
Hemosure First Sign UDS Cup	1	*Positive	1	100.00%
E-Screen	30	Negative	7	23.33%
Instant Technologies i Screen	4	Negative	1	25.00%
BioTek	12	Negative	1	8.33%
		Ethanol		
UDS-5		* Denotes Correct Response		



AG: All Methods	7			
		*Positive	5	71.43%
		Negative	2	28.57%
Beckman (Olympus) AU	1	*Positive	1	100.00%
Mindray BS-200	2	*Positive	1	50.00%
Abbott Architect Ci4100	1	*Positive	1	100.00%
Roche Hitachi 717	1	*Positive	1	100.00%
Mindray 480	1	*Positive	1	100.00%
CLIAwaived Inc Rapid Drug Cup	1	Negative	1	100.00%
Mindray BS-200	2	Negative	1	50.00%
UDS-6		* Denotes Correct Response		
AG: All Methods	7			
		*Negative	6	85.71%
		Positive	1	14.29%
CLIAwaived Inc Rapid Drug Cup	1	*Negative	1	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
Mindray 480	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	Positive	1	100.00%
UDS-5		Acetaminophen		
		* Denotes Correct Response		
AG: All Methods	7			
		*Negative	7	100.00%
AlereTriage	3	*Negative	3	100.00%
E-Screen	1	*Negative	1	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	1	*Negative	1	100.00%
UDS-6		* Denotes Correct Response		
AG: All Methods	7			
		*Negative	6	85.71%
		Positive	1	14.29%
AlereTriage	3	*Negative	3	100.00%
E-Screen	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	1	*Negative	1	100.00%
Noble Med Split Specimen Cup	1	Positive	1	100.00%
UDS-5		MDMA (Ecstasy)		
		* Denotes Correct Response		
AG: All Methods	94			
		*Negative	94	100.00%
E-Screen	3	*Negative	3	100.00%
Alere iCup Drugs of Abuse	13	*Negative	13	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
BioTek	12	*Negative	12	100.00%
CLIAwaived Inc Rapid Drug Cup	4	*Negative	4	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	2	*Negative	2	100.00%
Alere iCassette	3	*Negative	3	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	5	*Negative	5	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	24	*Negative	24	100.00%
Hemosure First Sign UDS Cup	1	*Negative	1	100.00%
UDS-6		* Denotes Correct Response		
AG: All Methods	94			
		*Negative	93	98.94%
		Positive	1	1.06%
E-Screen	3	*Negative	3	100.00%
Alere iCup Drugs of Abuse	13	*Negative	12	92.31%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
BioTek	12	*Negative	12	100.00%



CLIAwaived Inc Rapid Drug Cup	4	*Negative	4	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	2	*Negative	2	100.00%
Alere iCassette	3	*Negative	3	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	5	*Negative	5	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	24	*Negative	24	100.00%
Hemosure First Sign UDS Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	13	Positive	1	7.69%
		Methadone		
UDS-5		* Denotes Correct Response		
AG: All Methods	113			
		*Negative	101	89.38%
		Positive	12	10.62%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	1	*Negative	1	100.00%
E-Screen	13	*Negative	13	100.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	13	*Negative	13	100.00%
Instant Technologies i Screen	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	6	*Negative	6	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
Mindray 480	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	25	*Negative	25	100.00%
Hemosure First Sign UDS Cup	1	*Negative	1	100.00%
BioTek	12	Positive	12	100.00%
UDS-6		* Denotes Correct Response		
AG: All Methods	123			
		*Positive	105	85.37%
		Negative	18	14.63%
AlereTriage	3	*Positive	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	1	*Positive	1	100.00%
E-Screen	23	*Positive	20	86.96%
Express Diag. Express Check	1	*Positive	1	100.00%
Alere iCup Drugs of Abuse	13	*Positive	13	100.00%
Instant Technologies i Screen	2	*Positive	2	100.00%
Noble Med Split Specimen Cup	1	*Positive	1	100.00%
Beckman (Olympus) AU	1	*Positive	1	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Positive	5	100.00%
ASC One Screen Drug Test Card	1	*Positive	1	100.00%
MedTox Profile V	2	*Positive	2	100.00%
Jant Accutest Multi Drug Screen	2	*Positive	1	50.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Positive	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Positive	4	100.00%
Mindray BS-200	2	*Positive	2	100.00%
Abbott Architect Ci4100	1	*Positive	1	100.00%
Roche Hitachi 717	1	*Positive	1	100.00%



US Diagnostics UScreen Cup	4	*Positive	4	100.00%
Abbott Architect c4000	1	*Positive	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Positive	1	100.00%
Diag Test Group Clarity DOA Screen	6	*Positive	6	100.00%
Noble Medical Insta-Screen Test Cup	3	*Positive	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Positive	1	100.00%
Mindray 480	1	*Positive	1	100.00%
Hemosure First Sign UDS Dipstick	25	*Positive	23	92.00%
Hemosure First Sign UDS Cup	1	*Positive	1	100.00%
E-Screen	23	Negative	3	13.04%
BioTek	12	Negative	12	100.00%
Jant Accutest Multi Drug Screen	2	Negative	1	50.00%
Hemosure First Sign UDS Dipstick	25	Negative	2	8.00%
		Methamphetamine		
UDS-5		* Denotes Correct Response		
AG: All Methods	125			
		*Negative	121	96.80%
		Positive	4	3.20%
AlereTriage	3	*Negative	3	100.00%
E-Screen	22	*Negative	18	81.82%
Express Diag. Express Check	1	*Negative	1	100.00%
Innovacon One Step	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	17	*Negative	17	100.00%
Instant Technologies i Screen	3	*Negative	3	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Noble Med Split Specimen Cup	2	*Negative	2	100.00%
BioTek	12	*Negative	12	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Alere iCassette	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	6	*Negative	6	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
CRL Stat Cup	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	25	*Negative	25	100.00%
Hemosure First Sign UDS Cup	1	*Negative	1	100.00%
E-Screen	22	Positive	4	18.18%
		UDS-6		
AG: All Methods		* Denotes Correct Response		
	131			
		*Negative	125	95.42%
		Positive	6	4.58%
AlereTriage	3	*Negative	3	100.00%
E-Screen	28	*Negative	23	82.14%
Express Diag. Express Check	1	*Negative	1	100.00%
Innovacon One Step	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	17	*Negative	17	100.00%
Instant Technologies i Screen	3	*Negative	3	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Noble Med Split Specimen Cup	2	*Negative	2	100.00%
BioTek	12	*Negative	12	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Alere iCassette	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%



Diag Test Group Clarity DOA Screen	6	*Negative	6	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
CRL Stat Cup	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	25	*Negative	24	96.00%
Hemosure First Sign UDS Cup	1	*Negative	1	100.00%
E-Screen	28	Positive	5	17.86%
Hemosure First Sign UDS Dipstick	25	Positive	1	4.00%
Methaqualone				
UDS-5 * Denotes Correct Response				
AG: All Methods	2			
		*Negative	2	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	1	*Negative	1	100.00%
UDS-6 * Denotes Correct Response				
AG: All Methods	2			
		*Negative	2	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	1	*Negative	1	100.00%
Opiates				
UDS-5 * Denotes Correct Response				
AG: All Methods	158			
		*Negative	154	97.47%
		Positive	4	2.53%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	2	*Negative	2	100.00%
E-Screen	30	*Negative	26	86.67%
Express Diag. Express Check	1	*Negative	1	100.00%
Innovacon One Step	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	18	*Negative	18	100.00%
Instant Technologies i Screen	4	*Negative	4	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Noble Med Split Specimen Cup	2	*Negative	2	100.00%
Beckman (Olympus) AU	2	*Negative	2	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
BioTek	12	*Negative	12	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Alere iCassette	6	*Negative	6	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	5	*Negative	5	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
Mindray 480	1	*Negative	1	100.00%
CRL Stat Cup	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	24	*Negative	24	100.00%
Hemosure First Sign UDS Cup	1	*Negative	1	100.00%
E-Screen	30	Positive	4	13.33%
UDS-6 * Denotes Correct Response				
AG: All Methods	158			
		*Positive	149	94.30%
		Negative	9	5.70%
AlereTriage	3	*Positive	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	2	*Positive	2	100.00%
E-Screen	30	*Positive	24	80.00%
Express Diag. Express Check	1	*Positive	1	100.00%
Innovacon One Step	1	*Positive	1	100.00%
Alere iCup Drugs of Abuse	18	*Positive	18	100.00%



Instant Technologies i Screen	4	*Positive	4	100.00%
Integrated E-Z Split Key Cup	1	*Positive	1	100.00%
MedTox Profile II	1	*Positive	1	100.00%
Noble Med Split Specimen Cup	2	*Positive	2	100.00%
Beckman (Olympus) AU	2	*Positive	2	100.00%
Innovacon Multi-Clin	10	*Positive	10	100.00%
BioTek	12	*Positive	11	91.67%
CLIAwaived Inc Rapid Drug Cup	5	*Positive	5	100.00%
ASC One Screen Drug Test Card	1	*Positive	1	100.00%
MedTox Profile V	2	*Positive	2	100.00%
Jant Accutest Multi Drug Screen	2	*Positive	1	50.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Positive	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Positive	4	100.00%
Mindray BS-200	2	*Positive	2	100.00%
Abbott Architect Ci4100	1	*Positive	1	100.00%
Alere iCassette	6	*Positive	6	100.00%
Roche Hitachi 717	1	*Positive	1	100.00%
US Diagnostics UScreen Cup	4	*Positive	4	100.00%
Abbott Architect c4000	1	*Positive	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Positive	1	100.00%
Diag Test Group Clarity DOA Screen	5	*Positive	4	80.00%
Noble Medical Insta-Screen Test Cup	3	*Positive	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Positive	1	100.00%
Mindray 480	1	*Positive	1	100.00%
CRL Stat Cup	1	*Positive	1	100.00%
Hemosure First Sign UDS Dipstick	24	*Positive	24	100.00%
Hemosure First Sign UDS Cup	1	*Positive	1	100.00%
E-Screen	30	Negative	6	20.00%
BioTek	12	Negative	1	8.33%
Jant Accutest Multi Drug Screen	2	Negative	1	50.00%
Diag Test Group Clarity DOA Screen	5	Negative	1	20.00%
		Oxycodone		
UDS-5		* Denotes Correct Response		
AG: All Methods	85			
		*Negative	85	100.00%
Innovacon One Step	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	14	*Negative	14	100.00%
Instant Technologies i Screen	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
BioTek	12	*Negative	12	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Alere iCassette	3	*Negative	3	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	2	*Negative	2	100.00%
Diag Test Group Clarity DOA Screen	6	*Negative	6	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
Mindray 480	1	*Negative	1	100.00%
Hemosure First Sign UDS Cup	1	*Negative	1	100.00%
UDS-6		* Denotes Correct Response		
AG: All Methods	101			
		*Negative	97	96.04%
		Positive	4	3.96%
E-Screen	16	*Negative	14	87.50%
Innovacon One Step	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	14	*Negative	14	100.00%
Instant Technologies i Screen	2	*Negative	2	100.00%



Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
BioTek	12	*Negative	12	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	2	50.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Alere iCassette	3	*Negative	3	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	2	*Negative	2	100.00%
Diag Test Group Clarity DOA Screen	6	*Negative	6	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immutest Multi-Drug Screen Card	1	*Negative	1	100.00%
Mindray 480	1	*Negative	1	100.00%
Hemosure First Sign UDS Cup	1	*Negative	1	100.00%
E-Screen	16	Positive	2	12.50%
Amedica Biotech AmediCheck Instant Test Cup	4	Positive	2	50.00%
		PCP (Phencyclidine)		
UDS-5		* Denotes Correct Response		
AG: All Methods	139			
		*Positive	134	96.40%
		Negative	5	3.60%
AlereTriage	3	*Positive	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	1	*Positive	1	100.00%
E-Screen	30	*Positive	25	83.33%
Express Diag. Express Check	1	*Positive	1	100.00%
Innovacon One Step	1	*Positive	1	100.00%
Alere iCup Drugs of Abuse	11	*Positive	11	100.00%
Instant Technologies i Screen	1	*Positive	1	100.00%
Integrated E-Z Split Key Cup	1	*Positive	1	100.00%
MedTox Profile II	2	*Positive	2	100.00%
Noble Med Split Specimen Cup	1	*Positive	1	100.00%
Beckman (Olympus) AU	2	*Positive	2	100.00%
Innovacon Multi-Clin	10	*Positive	10	100.00%
BioTek	12	*Positive	12	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Positive	5	100.00%
ASC One Screen Drug Test Card	1	*Positive	1	100.00%
MedTox Profile V	2	*Positive	2	100.00%
Jant Accutest Multi Drug Screen	2	*Positive	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	1	*Positive	1	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Positive	4	100.00%
Mindray BS-200	1	*Positive	1	100.00%
Abbott Architect Ci4100	1	*Positive	1	100.00%
Alere iCassette	5	*Positive	5	100.00%
Roche Hitachi 717	1	*Positive	1	100.00%
US Diagnostics UScreen Cup	1	*Positive	1	100.00%
Abbott Architect c4000	1	*Positive	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Positive	1	100.00%
Diag Test Group Clarity DOA Screen	6	*Positive	6	100.00%
Noble Medical Insta-Screen Test Cup	3	*Positive	3	100.00%
Alere Immutest Multi-Drug Screen Card	1	*Positive	1	100.00%
CRL Stat Cup	1	*Positive	1	100.00%
Hemosure First Sign UDS Dipstick	24	*Positive	24	100.00%
Hemosure First Sign UDS Cup	1	*Positive	1	100.00%
E-Screen	30	Negative	5	16.67%
UDS-6		* Denotes Correct Response		
AG: All Methods	139			
		*Negative	136	97.84%
		Positive	3	2.16%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	1	*Negative	1	100.00%



E-Screen	30	*Negative	27	90.00%
Express Diag. Express Check	1	*Negative	1	100.00%
Innovacon One Step	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	11	*Negative	11	100.00%
Instant Technologies i Screen	1	*Negative	1	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman (Olympus) AU	2	*Negative	2	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
BioTek	12	*Negative	12	100.00%
CLIAwaived Inc Rapid Drug Cup	5	*Negative	5	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	1	*Negative	1	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	1	*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Alere iCassette	5	*Negative	5	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	1	*Negative	1	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	6	*Negative	6	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
CRL Stat Cup	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	24	*Negative	24	100.00%
Hemosure First Sign UDS Cup	1	*Negative	1	100.00%
E-Screen	30	Positive	3	10.00%
		Propoxyphene		
UDS-5		* Denotes Correct Response		
AG: All Methods	30			
		*Positive	27	90.00%
		Negative	3	10.00%
E-Screen	4	*Positive	4	100.00%
Alere iCup Drugs of Abuse	6	*Positive	6	100.00%
MedTox Profile II	1	*Positive	1	100.00%
Beckman (Olympus) AU	2	*Positive	2	100.00%
Innovacon Multi-Clin	7	*Positive	7	100.00%
CLIAwaived Inc Rapid Drug Cup	2	*Positive	1	50.00%
MedTox Profile V	2	*Positive	2	100.00%
Abbott Architect Ci4100	1	*Positive	1	100.00%
Alere iCassette	3	*Positive	3	100.00%
CLIAwaived Inc Rapid Drug Cup	2	Negative	1	50.00%
Jant Accutest Multi Drug Screen	1	Negative	1	100.00%
Hemosure First Sign UDS Cup	1	Negative	1	100.00%
UDS-6		* Denotes Correct Response		
AG: All Methods	30			
		*Negative	30	100.00%
E-Screen	4	*Negative	4	100.00%
Alere iCup Drugs of Abuse	6	*Negative	6	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Beckman (Olympus) AU	2	*Negative	2	100.00%
Innovacon Multi-Clin	7	*Negative	7	100.00%
CLIAwaived Inc Rapid Drug Cup	2	*Negative	2	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	1	*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Alere iCassette	3	*Negative	3	100.00%
Hemosure First Sign UDS Cup	1	*Negative	1	100.00%
		Tricyclic Antidepressants		
UDS-5		* Denotes Correct Response		
AG: All Methods	54			
		*Positive	48	88.89%
		Negative	6	11.11%
AlereTriage	3	*Positive	3	100.00%



Alere iCup Drugs of Abuse	11	*Positive	8	72.73%
Instant Technologies i Screen	2	*Positive	2	100.00%
Noble Med Split Specimen Cup	1	*Positive	1	100.00%
Innovacon Multi-Clin	9	*Positive	9	100.00%
CLIAwaived Inc Rapid Drug Cup	4	*Positive	3	75.00%
MedTox Profile V	2	*Positive	2	100.00%
Jant Accutest Multi Drug Screen	2	*Positive	2	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
Alere iCassette	3	*Positive	3	100.00%
US Diagnostics UScreen Cup	3	*Positive	3	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	2	*Positive	2	100.00%
Diag Test Group Clarity DOA Screen	4	*Positive	4	100.00%
Noble Medical Insta-Screen Test Cup	3	*Positive	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Positive	1	100.00%
Mindray 480	1	*Positive	1	100.00%
Alere iCup Drugs of Abuse	11	Negative	3	27.27%
CLIAwaived Inc Rapid Drug Cup	4	Negative	1	25.00%
Amedica Biotech AmediCheck Instant Test Cup	1	Negative	1	100.00%
Hemosure First Sign UDS Cup	1	Negative	1	100.00%
UDS-6		* Denotes Correct Response		
AG: All Methods	54			
		*Negative	54	100.00%
AlereTriage	3	*Negative	3	100.00%
Alere iCup Drugs of Abuse	11	*Negative	11	100.00%
Instant Technologies i Screen	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Innovacon Multi-Clin	9	*Negative	9	100.00%
CLIAwaived Inc Rapid Drug Cup	4	*Negative	4	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	1	*Negative	1	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
Alere iCassette	3	*Negative	3	100.00%
US Diagnostics UScreen Cup	3	*Negative	3	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	2	*Negative	2	100.00%
Diag Test Group Clarity DOA Screen	4	*Negative	4	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Alere Immatest Multi-Drug Screen Card	1	*Negative	1	100.00%
Mindray 480	1	*Negative	1	100.00%
Hemosure First Sign UDS Cup	1	*Negative	1	100.00%
		Buprenorphine		
UDS-5		* Denotes Correct Response		
AG: All Methods	32			
		*Negative	32	100.00%
Alere iCup Drugs of Abuse	7	*Negative	7	100.00%
BioTek	12	*Negative	12	100.00%
CLIAwaived Inc Rapid Drug Cup	2	*Negative	2	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	1	*Negative	1	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Diag Test Group Clarity DOA Screen	1	*Negative	1	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Mindray 480	1	*Negative	1	100.00%
UDS-6		* Denotes Correct Response		
AG: All Methods	32			
		*Negative	32	100.00%
Alere iCup Drugs of Abuse	7	*Negative	7	100.00%
BioTek	12	*Negative	12	100.00%
CLIAwaived Inc Rapid Drug Cup	2	*Negative	2	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	1	*Negative	1	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Diag Test Group Clarity DOA Screen	1	*Negative	1	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Mindray 480	1	*Negative	1	100.00%



756 - Viral Antibodies	Labs	Response	Number	Percent
		HSV Ab (Herpes Simplex Virus)		
VA-3		* Denotes Correct Response		
PG: Biokit HSV-II	2			
		*Positive	2	100.00%
770 - Viral Antigen Detection;792 - Viral Antigen Detection - Waived Methods;797 - Viral Antigen Detection II;798 - Viral Antigen Detection (ROT)	Labs	Response	Number	Percent
		Rotavirus - Aq Detect		
VAD-11		* Denotes Correct Response		
AG: All Methods	9			
		*Positive	8	88.89%
		Negative	1	11.11%
Meridian Immunocard Stat! Rotavirus	5	*Positive	5	100.00%
SAS Rota-Test	4	*Positive	3	75.00%
SAS Rota-Test		Negative	1	25.00%
VAD-12		* Denotes Correct Response		
AG: All Methods	9			
		*Negative	9	100.00%
Meridian Immunocard Stat! Rotavirus	5	*Negative	5	100.00%
SAS Rota-Test	4	*Negative	4	100.00%
VAD-13		* Denotes Correct Response		
AG: All Methods	9			
		*Positive	9	100.00%
Meridian Immunocard Stat! Rotavirus	5	*Positive	5	100.00%
SAS Rota-Test	4	*Positive	4	100.00%
VAD-14		* Denotes Correct Response		
AG: All Methods	9			
		*Negative	9	100.00%
Meridian Immunocard Stat! Rotavirus	5	*Negative	5	100.00%
SAS Rota-Test	4	*Negative	4	100.00%
VAD-15		* Denotes Correct Response		
AG: All Methods	9			
		*Negative	9	100.00%
Meridian Immunocard Stat! Rotavirus	5	*Negative	5	100.00%
SAS Rota-Test	4	*Negative	4	100.00%
		Influenza A - Aq Detect		
VAD-11		* Denotes Correct Response		
AG: All Methods	489			
		*Negative	483	98.77%
		Positive	6	1.23%
Binax Now Flu A	19	*Negative	19	100.00%
Alere BinaxNOW Influenza A+B Card	121	*Negative	120	99.17%
Sekisui Diagnostics OSOM Influenza A & B	20	*Negative	20	100.00%
Quidel QuickVue A+B	89	*Negative	87	97.75%
Remel Xpect	1	*Negative	1	100.00%
Quidel Sofia Influenza A+B	52	*Negative	52	100.00%
Alere Influenza A & B	29	*Negative	29	100.00%
BD Veritor Flu A+B	127	*Negative	124	97.64%
Henry Schein OneStep + Influenza A&B (distinguish)	12	*Negative	12	100.00%
Cepheid GeneXpert - Xpert Flu	2	*Negative	2	100.00%
Alere i Influenza A & B	11	*Negative	11	100.00%
Roche Cobas Liat Influenza A/B	3	*Negative	3	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	2	*Negative	2	100.00%
Alere BinaxNOW Influenza A+B Card	121	Positive	1	0.83%
Quidel QuickVue A+B	89	Positive	2	2.25%
BD Veritor Flu A+B	127	Positive	3	2.36%
VAD-12		* Denotes Correct Response		
AG: All Methods	487			
		*Negative	476	97.74%
		Positive	11	2.26%
Binax Now Flu A	19	*Negative	19	100.00%
Alere BinaxNOW Influenza A+B Card	121	*Negative	117	96.69%
Sekisui Diagnostics OSOM Influenza A & B	20	*Negative	20	100.00%
Quidel QuickVue A+B	89	*Negative	88	98.88%
Remel Xpect	1	*Negative	1	100.00%
Quidel Sofia Influenza A+B	52	*Negative	52	100.00%
Alere Influenza A & B	29	*Negative	28	96.55%



BD Veritor Flu A+B	127	*Negative	123	96.85%
Henry Schein OneStep + Influenza A&B (distinguish)	12	*Negative	12	100.00%
Cepheid GeneXpert - Xpert Flu	2	*Negative	2	100.00%
Alere i Influenza A & B	10	*Negative	10	100.00%
Roche Cobas Liat Influenza A/B	2	*Negative	2	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	2	*Negative	2	100.00%
Alere BinaxNOW Influenza A+B Card	121	Positive	4	3.31%
Quidel QuickVue A+B	89	Positive	1	1.12%
Alere Influenza A & B	29	Positive	1	3.45%
BD Veritor Flu A+B	127	Positive	4	3.15%
VAD-13		* Denotes Correct Response		
AG: All Methods	66			
		*Negative	65	98.48%
		Positive	1	1.52%
Alere BinaxNOW Influenza A+B Card	8	*Negative	8	100.00%
Sekisui Diagnostics OSOM Influenza A & B	12	*Negative	12	100.00%
Quidel QuickVue A+B	19	*Negative	19	100.00%
Remel Xpect	1	*Negative	1	100.00%
Quidel Sofia Influenza A+B	5	*Negative	5	100.00%
Alere Influenza A & B	4	*Negative	4	100.00%
BD Veritor Flu A+B	12	*Negative	11	91.67%
Cepheid GeneXpert - Xpert Flu	2	*Negative	2	100.00%
Alere i Influenza A & B	2	*Negative	2	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	1	*Negative	1	100.00%
BD Veritor Flu A+B	12	Positive	1	8.33%
VAD-14		* Denotes Correct Response		
AG: All Methods	66			
		*Positive	66	100.00%
Alere BinaxNOW Influenza A+B Card	8	*Positive	8	100.00%
Sekisui Diagnostics OSOM Influenza A & B	12	*Positive	12	100.00%
Quidel QuickVue A+B	19	*Positive	19	100.00%
Remel Xpect	1	*Positive	1	100.00%
Quidel Sofia Influenza A+B	5	*Positive	5	100.00%
Alere Influenza A & B	4	*Positive	4	100.00%
BD Veritor Flu A+B	12	*Positive	12	100.00%
Cepheid GeneXpert - Xpert Flu	2	*Positive	2	100.00%
Alere i Influenza A & B	2	*Positive	2	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	1	*Positive	1	100.00%
VAD-15		* Denotes Correct Response		
AG: All Methods	66			
		*Positive	66	100.00%
Alere BinaxNOW Influenza A+B Card	8	*Positive	8	100.00%
Sekisui Diagnostics OSOM Influenza A & B	12	*Positive	12	100.00%
Quidel QuickVue A+B	19	*Positive	19	100.00%
Remel Xpect	1	*Positive	1	100.00%
Quidel Sofia Influenza A+B	5	*Positive	5	100.00%
Alere Influenza A & B	4	*Positive	4	100.00%
BD Veritor Flu A+B	12	*Positive	12	100.00%
Cepheid GeneXpert - Xpert Flu	2	*Positive	2	100.00%
Alere i Influenza A & B	2	*Positive	2	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	1	*Positive	1	100.00%
Influenza A/B - Aq Detect				
VAD-11		* Denotes Correct Response		
AG: All Methods	34			
		*Negative	33	97.06%
		Positive	1	2.94%
Alere BinaxNOW Influenza A+B Card	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM Influenza A & B	1	*Negative	1	100.00%
Quidel QuickVue	28	*Negative	27	96.43%
Henry Schein OneStep + Influenza A/B	4	*Negative	4	100.00%
Quidel QuickVue	28	Positive	1	3.57%
VAD-12		* Denotes Correct Response		
AG: All Methods	34			
		*Positive	31	91.18%
		Negative	3	8.82%
Alere BinaxNOW Influenza A+B Card	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Influenza A & B	1	*Positive	1	100.00%
Quidel QuickVue	28	*Positive	26	92.86%
Henry Schein OneStep + Influenza A/B	4	*Positive	3	75.00%



Quidel QuickVue	28	Negative	2	7.14%
Henry Schein OneStep + Influenza A/B	4	Negative	1	25.00%
VAD-13		* Denotes Correct Response		
AG: All Methods	5			
		*Negative	5	100.00%
Quidel QuickVue	5	*Negative	5	100.00%
VAD-14		* Denotes Correct Response		
AG: All Methods	5			
		*Positive	5	100.00%
Quidel QuickVue	5	*Positive	5	100.00%
VAD-15		* Denotes Correct Response		
AG: All Methods	5			
		*Positive	5	100.00%
Quidel QuickVue	5	*Positive	5	100.00%
		Influenza B - Ag Detect		
VAD-11		* Denotes Correct Response		
AG: All Methods	486			
		*Negative	477	98.15%
		Positive	9	1.85%
Binax Now Flu B	19	*Negative	19	100.00%
Alere BinaxNOW Influenza A+B Card	119	*Negative	117	98.32%
Sekisui Diagnostics OSOM Influenza A & B	20	*Negative	20	100.00%
Quidel QuickVue A+B	89	*Negative	87	97.75%
Remel Xpect	1	*Negative	1	100.00%
Quidel Sofia Influenza A+B	52	*Negative	50	96.15%
Alere Influenza A & B	29	*Negative	29	100.00%
BD Veritor Flu A+B	127	*Negative	124	97.64%
Henry Schein OneStep + Influenza A&B (distinguish)	12	*Negative	12	100.00%
Cepheid GeneXpert - Xpert Flu	2	*Negative	2	100.00%
Alere i Influenza A & B	11	*Negative	11	100.00%
Roche Cobas Liat Influenza A/B	3	*Negative	3	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	2	*Negative	2	100.00%
Alere BinaxNOW Influenza A+B Card	119	Positive	2	1.68%
Quidel QuickVue A+B	89	Positive	2	2.25%
Quidel Sofia Influenza A+B	52	Positive	2	3.85%
BD Veritor Flu A+B	127	Positive	3	2.36%
VAD-12		* Denotes Correct Response		
AG: All Methods	484			
		*Positive	480	99.17%
		Negative	4	0.83%
Binax Now Flu B	19	*Positive	19	100.00%
Alere BinaxNOW Influenza A+B Card	119	*Positive	119	100.00%
Sekisui Diagnostics OSOM Influenza A & B	20	*Positive	20	100.00%
Quidel QuickVue A+B	89	*Positive	88	98.88%
Remel Xpect	1	*Positive	1	100.00%
Quidel Sofia Influenza A+B	52	*Positive	51	98.08%
Alere Influenza A & B	29	*Positive	28	96.55%
BD Veritor Flu A+B	127	*Positive	126	99.21%
Henry Schein OneStep + Influenza A&B (distinguish)	12	*Positive	12	100.00%
Cepheid GeneXpert - Xpert Flu	2	*Positive	2	100.00%
Alere i Influenza A & B	10	*Positive	10	100.00%
Roche Cobas Liat Influenza A/B	2	*Positive	2	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	2	*Positive	2	100.00%
Quidel QuickVue A+B	89	Negative	1	1.12%
Quidel Sofia Influenza A+B	52	Negative	1	1.92%
Alere Influenza A & B	29	Negative	1	3.45%
BD Veritor Flu A+B	127	Negative	1	0.79%
VAD-13		* Denotes Correct Response		
AG: All Methods	66			
		*Negative	65	98.48%
		Positive	1	1.52%
Alere BinaxNOW Influenza A+B Card	8	*Negative	8	100.00%
Sekisui Diagnostics OSOM Influenza A & B	12	*Negative	12	100.00%
Quidel QuickVue A+B	19	*Negative	18	94.74%
Remel Xpect	1	*Negative	1	100.00%
Quidel Sofia Influenza A+B	5	*Negative	5	100.00%
Alere Influenza A & B	4	*Negative	4	100.00%
BD Veritor Flu A+B	12	*Negative	12	100.00%
Cepheid GeneXpert - Xpert Flu	2	*Negative	2	100.00%



Alere i Influenza A & B	2	*Negative	2	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	1	*Negative	1	100.00%
Quidel QuickVue A+B	19	Positive	1	5.26%
VAD-14		* Denotes Correct Response		
AG: All Methods	66			
		*Negative	66	100.00%
Alere BinaxNOW Influenza A+B Card	8	*Negative	8	100.00%
Sekisui Diagnostics OSOM Influenza A & B	12	*Negative	12	100.00%
Quidel QuickVue A+B	19	*Negative	19	100.00%
Remel Xpect	1	*Negative	1	100.00%
Quidel Sofia Influenza A+B	5	*Negative	5	100.00%
Alere Influenza A & B	4	*Negative	4	100.00%
BD Veritor Flu A+B	12	*Negative	12	100.00%
Cepheid GeneXpert - Xpert Flu	2	*Negative	2	100.00%
Alere i Influenza A & B	2	*Negative	2	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	1	*Negative	1	100.00%
VAD-15		* Denotes Correct Response		
AG: All Methods	66			
		*Negative	65	98.48%
		Positive	1	1.52%
Alere BinaxNOW Influenza A+B Card	8	*Negative	8	100.00%
Sekisui Diagnostics OSOM Influenza A & B	12	*Negative	12	100.00%
Quidel QuickVue A+B	19	*Negative	19	100.00%
Remel Xpect	1	*Negative	1	100.00%
Quidel Sofia Influenza A+B	5	*Negative	5	100.00%
Alere Influenza A & B	4	*Negative	4	100.00%
BD Veritor Flu A+B	12	*Negative	12	100.00%
Cepheid GeneXpert - Xpert Flu	2	*Negative	2	100.00%
Alere i Influenza A & B	2	*Negative	1	50.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	1	*Negative	1	100.00%
Alere i Influenza A & B	2	Positive	1	50.00%
		RSV - Ag Detect		
VAD-11		* Denotes Correct Response		
AG: All Methods	265			
		*Positive	251	94.72%
		Negative	14	5.28%
BD Directigen RSV	1	*Positive	1	100.00%
BD EZ RSV	1	*Positive	1	100.00%
Alere BinaxNOW RSV	175	*Positive	163	93.14%
Quidel QuickVue RSV	27	*Positive	26	96.30%
Wampole Clearview RSV	5	*Positive	5	100.00%
SAS RSV Alert	1	*Positive	1	100.00%
Quidel QuickVue RSV 10	2	*Positive	2	100.00%
Quidel Sofia RSV	10	*Positive	10	100.00%
BD Veritor RSV	42	*Positive	41	97.62%
Cepheid GeneXpert - Xpert RSV	1	*Positive	1	100.00%
Alere BinaxNOW RSV	175	Negative	12	6.86%
Quidel QuickVue RSV	27	Negative	1	3.70%
BD Veritor RSV	42	Negative	1	2.38%
VAD-12		* Denotes Correct Response		
AG: All Methods	264			
		*Positive	260	98.48%
		Negative	4	1.52%
BD Directigen RSV	1	*Positive	1	100.00%
BD EZ RSV	1	*Positive	1	100.00%
Alere BinaxNOW RSV	174	*Positive	172	98.85%
Quidel QuickVue RSV	27	*Positive	25	92.59%
Wampole Clearview RSV	5	*Positive	5	100.00%
SAS RSV Alert	1	*Positive	1	100.00%
Quidel QuickVue RSV 10	2	*Positive	2	100.00%
Quidel Sofia RSV	10	*Positive	10	100.00%
BD Veritor RSV	42	*Positive	42	100.00%
Cepheid GeneXpert - Xpert RSV	1	*Positive	1	100.00%
Alere BinaxNOW RSV	174	Negative	2	1.15%
Quidel QuickVue RSV	27	Negative	2	7.41%
VAD-13		* Denotes Correct Response		
AG: All Methods	26			
		*Negative	26	100.00%
BD EZ RSV	1	*Negative	1	100.00%



Alere BinaxNOW RSV	10	*Negative	10	100.00%
Quidel QuickVue RSV	3	*Negative	3	100.00%
Wampole Clearview RSV	4	*Negative	4	100.00%
SAS RSV Alert	1	*Negative	1	100.00%
Quidel QuickVue RSV 10	2	*Negative	2	100.00%
Quidel Sofia RSV	4	*Negative	4	100.00%
Cepheid GeneXpert - Xpert RSV	1	*Negative	1	100.00%
VAD-14		* Denotes Correct Response		
AG: All Methods	26			
		*Negative	25	96.15%
		Positive	1	3.85%
BD EZ RSV	1	*Negative	1	100.00%
Alere BinaxNOW RSV	10	*Negative	9	90.00%
Quidel QuickVue RSV	3	*Negative	3	100.00%
Wampole Clearview RSV	4	*Negative	4	100.00%
SAS RSV Alert	1	*Negative	1	100.00%
Quidel QuickVue RSV 10	2	*Negative	2	100.00%
Quidel Sofia RSV	4	*Negative	4	100.00%
Cepheid GeneXpert - Xpert RSV	1	*Negative	1	100.00%
Alere BinaxNOW RSV	10	Positive	1	10.00%
VAD-15		* Denotes Correct Response		
AG: All Methods	26			
		*Negative	25	96.15%
		Positive	1	3.85%
BD EZ RSV	1	*Negative	1	100.00%
Alere BinaxNOW RSV	10	*Negative	9	90.00%
Quidel QuickVue RSV	3	*Negative	3	100.00%
Wampole Clearview RSV	4	*Negative	4	100.00%
SAS RSV Alert	1	*Negative	1	100.00%
Quidel QuickVue RSV 10	2	*Negative	2	100.00%
Quidel Sofia RSV	4	*Negative	4	100.00%
Cepheid GeneXpert - Xpert RSV	1	*Negative	1	100.00%
Alere BinaxNOW RSV	10	Positive	1	10.00%
		Adenovirus		
VAD-11		* Denotes Correct Response		
AG: All Methods	13			
		*Negative	13	100.00%
SAS Adeno Test	10	*Negative	10	100.00%
Quidel AdenoPlus	3	*Negative	3	100.00%
VAD-12		* Denotes Correct Response		
AG: All Methods	13			
		*Negative	13	100.00%
SAS Adeno Test	10	*Negative	10	100.00%
Quidel AdenoPlus	3	*Negative	3	100.00%
VAD-13		* Denotes Correct Response		
AG: All Methods	13			
		*Negative	12	92.31%
		Positive	1	7.69%
SAS Adeno Test	10	*Negative	9	90.00%
Quidel AdenoPlus	3	*Negative	3	100.00%
SAS Adeno Test	10	Positive	1	10.00%
VAD-14		* Denotes Correct Response		
AG: All Methods	13			
		*Positive	13	100.00%
SAS Adeno Test	10	*Positive	10	100.00%
Quidel AdenoPlus	3	*Positive	3	100.00%
VAD-15		* Denotes Correct Response		
AG: All Methods	13			
		*Negative	13	100.00%
SAS Adeno Test	10	*Negative	10	100.00%
Quidel AdenoPlus	3	*Negative	3	100.00%
932 - Hepatitis C Antibody (Waived);767 - Hepatitis Markers;750 - Viral Markers	Labs	Response	Number	Percent
		HCV (Waived)		
VM-11		* Denotes Correct Response		
PG: OraSure QraQuick HCV Rapid Antibody Test	4			
		*Negative	4	100.00%
VM-12		* Denotes Correct Response		



PG: OraSure QraQuick HCV Rapid Antibody Test	4			
		*Negative	4	100.00%
		HBsAg		
VM-11		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Siemens Centaur	1	*Positive	1	100.00%
Abbott Architect Ci4100	1	*Positive	1	100.00%
VM-12		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Siemens Centaur	1	*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
767 - Hepatitis Markers;750 - Viral Markers	Labs	Response	Number	Percent
		HBsAg		
VM-11		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Siemens Centaur	1	*Positive	1	100.00%
Abbott Architect Ci4100	1	*Positive	1	100.00%
VM-12		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Siemens Centaur	1	*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
VM-13		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Siemens Centaur	1	*Positive	1	100.00%
Abbott Architect Ci4100	1	*Positive	1	100.00%
VM-14		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	1	50.00%
		Negative	1	50.00%
Abbott Architect Ci4100	1	*Positive	1	100.00%
Siemens Centaur	1	Negative	1	100.00%
VM-15		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Siemens Centaur	1	*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
		HCV		
VM-11		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Siemens Centaur	1	*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
VM-12		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Siemens Centaur	1	*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
VM-13		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Siemens Centaur	1	*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
VM-14		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Siemens Centaur	1	*Positive	1	100.00%
Abbott Architect Ci4100	1	*Positive	1	100.00%
VM-15		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Siemens Centaur	1	*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
		Anti-HBc		



VM-11		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Abbott Architect Ci4100	1	*Positive	1	100.00%
VM-12		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
VM-13		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
VM-14		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
VM-15		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
HAV				
VM-11		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
VM-12		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
VM-13		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
VM-14		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
VM-15		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Anti-HBc (International)				
VM-11		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Positive	3	100.00%
VM-12		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Negative	3	100.00%
VM-13		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Positive	3	100.00%
VM-14		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Positive	3	100.00%
VM-15		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Negative	3	100.00%
HBsAb (International)				
VM-11		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECi/ECiQ	3	*Negative	3	100.00%



VM-12		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Positive	3	100.00%
VM-13		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Intl OCD Vitros ECI/ECIQ	2	*Negative	2	100.00%
VM-14		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Positive	3	100.00%
VM-15		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
HBSAg (International)		* Denotes Correct Response		
VM-11		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Positive	3	100.00%
VM-12		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
VM-13		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Positive	3	100.00%
VM-14		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	1	33.33%
		Negative	2	66.67%
Intl OCD Vitros ECI/ECIQ	3	*Positive	1	33.33%
Intl OCD Vitros ECI/ECIQ		Negative	2	66.67%
VM-15		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
HCV (International)		* Denotes Correct Response		
VM-11		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
VM-12		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
VM-13		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
VM-14		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Positive	3	100.00%
VM-15		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Intl OCD Vitros ECI/ECIQ	2	*Negative	2	100.00%
771 - Affirm VP		Response		
	Labs		Number	Percent
VP-11		* Denotes Correct Response		
AG: All Methods	35			
		*Negative	35	100.00%
Affirm VP	35	*Negative	35	100.00%
VP-12		* Denotes Correct Response		



AG: All Methods	35			
		*Negative	35	100.00%
Affirm VP	35	*Negative	35	100.00%
VP-13		* Denotes Correct Response		
AG: All Methods	35			
		*Negative	35	100.00%
Affirm VP	35	*Negative	35	100.00%
VP-14		* Denotes Correct Response		
AG: All Methods	35			
		*Negative	35	100.00%
Affirm VP	35	*Negative	35	100.00%
VP-15		* Denotes Correct Response		
AG: All Methods	35			
		*Positive	35	100.00%
Affirm VP	35	*Positive	35	100.00%
VP-11		G. vaginalis - Non Culture Tech		
		* Denotes Correct Response		
AG: All Methods	35			
		*Negative	35	100.00%
Affirm VP	35	*Negative	35	100.00%
VP-12		* Denotes Correct Response		
AG: All Methods	35			
		*Positive	35	100.00%
Affirm VP	35	*Positive	35	100.00%
VP-13		* Denotes Correct Response		
AG: All Methods	35			
		*Positive	35	100.00%
Affirm VP	35	*Positive	35	100.00%
VP-14		* Denotes Correct Response		
AG: All Methods	35			
		*Negative	35	100.00%
Affirm VP	35	*Negative	35	100.00%
VP-15		* Denotes Correct Response		
AG: All Methods	35			
		*Negative	35	100.00%
Affirm VP	35	*Negative	35	100.00%
VP-11		T. vaginalis - Non Culture		
		* Denotes Correct Response		
AG: All Methods	35			
		*Positive	35	100.00%
Affirm VP	35	*Positive	35	100.00%
VP-12		* Denotes Correct Response		
AG: All Methods	35			
		*Negative	35	100.00%
Affirm VP	35	*Negative	35	100.00%
VP-13		* Denotes Correct Response		
AG: All Methods	35			
		*Positive	34	97.14%
		Negative	1	2.86%
Affirm VP	35	*Positive	34	97.14%
Affirm VP		Negative	1	2.86%
VP-14		* Denotes Correct Response		
AG: All Methods	35			
		*Negative	35	100.00%
Affirm VP	35	*Negative	35	100.00%
VP-15		* Denotes Correct Response		
AG: All Methods	35			
		*Positive	35	100.00%
Affirm VP	35	*Positive	35	100.00%