



666 - Afinion Glycohemoglobin		AFF-1																								
Glycohemoglobin		Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %
PG: Afinion Glycohemoglobin		154	6.15	0.61	4.40-8.00	9.84																				
747 - Auto Diff 7 (Act 5 Diff & Horiba 5-Part Diff)		BCX-1					BCX-2					BCX-3					BCX-4					BCX-5				
White Blood Cell Count		Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %
PG: Beckman Coulter Act 5 diff		23	28.35	0.84	24.10-32.70	2.96	23	7.71	0.23	6.50-8.90	2.99	23	2.63	0.10	2.20-3.00	3.85	23	5.38	0.15	4.60-6.20	2.78	23	17.16	0.54	14.60-19.80	3.14
PG: Horiba Pentra 60		16	28.94	0.83	24.60-33.20	2.87	16	7.75	0.20	6.50-8.90	2.60	16	2.64	0.10	2.20-3.00	3.85	16	5.49	0.13	4.70-6.30	2.18	16	17.38	0.46	14.80-20.00	2.64
PG: Horiba Pentra 60C+		6	28.25	0.73	24.10-32.50	2.58	6	7.52	0.29	6.40-8.60	3.87	6	2.65	0.06	2.30-3.10	1.85	6	5.37	0.14	4.60-6.20	2.59	6	16.97	0.63	14.40-19.60	3.71
SG: Select Horiba Instruments (5-part Diff)		22	28.76	0.84	24.50-33.10	2.95	22	7.69	0.24	6.50-8.90	3.12	22	2.64	0.09	2.20-3.00	3.46	22	5.46	0.14	4.70-6.30	2.55	22	17.27	0.53	14.70-19.90	3.06
SG: International Horiba Instruments		25	28.57	0.97	24.30-32.90	3.39	25	7.70	0.23	6.50-8.90	2.99	25	2.65	0.10	2.30-3.10	3.33	25	5.47	0.13	4.70-6.30	2.36	25	17.23	0.51	14.60-19.80	2.97
Lymphocyte		Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %
PG: Beckman Coulter Act 5 diff		23	19.65	0.95	16.90-22.50	4.82	23	28.22	2.09	21.90-34.50	7.41	23	19.75	2.51	12.30-27.30	12.68	23	24.90	3.09	15.60-34.20	12.41	23	37.87	2.32	30.90-44.90	6.12
PG: Horiba Pentra 60		16	23.64	1.80	18.20-29.00	7.63	16	34.18	2.64	26.30-42.10	7.72	16	23.84	2.87	15.20-32.40	12.06	16	30.19	3.75	18.90-41.50	12.42	16	40.72	3.13	31.30-50.10	7.69
PG: Horiba Pentra 60C+		6	22.67	0.90	20.00-25.40	3.96	6	31.77	1.65	26.90-36.80	5.19	6	23.53	4.87	8.90-38.10	20.72	6	27.48	5.66	10.50-44.50	20.58	6	38.32	3.20	28.70-47.90	8.36
SG: Select Horiba Instruments (5-part Diff)		22	23.38	1.65	18.40-28.30	7.05	22	33.55	2.62	25.60-41.40	7.82	22	23.76	3.37	13.70-33.90	14.16	22	29.48	4.36	16.40-42.60	14.78	22	40.10	3.26	30.30-49.90	8.13
SG: International Horiba Instruments		25	24.30	3.07	15.10-33.50	12.63	25	34.64	3.97	22.70-46.50	11.47	25	24.58	3.99	12.60-36.60	16.22	25	30.57	5.17	15.10-46.10	16.93	25	41.65	5.48	25.30-58.10	13.14
Monocyte		Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %
PG: Beckman Coulter Act 5 diff		23	1.51	0.38	0.38-2.64	25.03	23	2.00	0.62	0.13-3.87	31.20	20	0.48	0.26	0.00-1.27	54.58	23	0.96	0.44	0.00-2.27	45.42	23	1.33	0.37	0.23-2.43	27.67
PG: Horiba Pentra 60		16	1.35	0.27	0.53-2.17	20.15	16	1.75	0.46	0.36-3.14	26.40	14	0.48	0.21	0.00-1.12	44.58	16	0.87	0.33	0.00-1.86	38.05	16	1.32	0.52	0.00-2.87	39.02
PG: Horiba Pentra 60C+		6	1.57	0.84	0.00-4.08	53.38	6	2.37	1.31	0.00-6.29	55.19	6	0.60	0.31	0.00-1.53	51.67	6	0.97	0.42	0.00-2.24	43.61	6	1.83	0.95	0.00-4.68	51.91
SG: Select Horiba Instruments (5-part Diff)		22	1.41	0.47	0.00-2.83	33.48	22	1.91	0.79	0.00-4.27	41.26	20	0.51	0.24	0.00-1.24	47.65	22	0.90	0.35	0.00-1.95	38.89	22	1.46	0.67	0.00-3.47	45.96
SG: International Horiba Instruments		25	1.39	0.45	0.04-2.74	32.37	25	1.89	0.80	0.00-4.28	42.22	22	0.54	0.27	0.00-1.36	50.74	25	0.89	0.34	0.00-1.91	38.09	25	1.42	0.65	0.00-3.36	45.56
Granulocyte		Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %
PG: Beckman Coulter Act 5 diff		22	63.10	1.20	59.50-66.70	1.90	22	56.42	1.97	50.50-62.30	3.49	22	61.49	3.28	51.70-71.30	5.33	22	59.80	7.59	37.00-82.60	12.69	22	48.86	2.54	41.30-56.50	5.19
PG: Horiba Pentra 60		16	70.86	1.67	65.90-75.90	2.36	16	57.81	7.13	36.40-79.20	12.34	16	64.83	8.23	40.10-89.50	12.70	16	62.94	3.59	52.10-73.70	5.71	16	52.41	2.31	45.50-59.30	4.41
PG: Horiba Pentra 60C+		6	71.40	0.83	68.90-73.90	1.16	6	60.63	1.73	55.40-65.80	2.85	6	66.65	4.88	52.10-81.30	7.32	6	65.68	5.95	47.90-83.60	9.06	6	53.88	3.19	44.30-63.50	5.92
SG: Select Horiba Instruments (5-part Diff)		22	71.00	1.50	66.50-75.50	2.11	22	58.54	6.27	39.70-77.30	10.72	22	65.30	7.44	43.00-87.60	11.39	22	63.65	4.35	50.70-76.80	6.83	22	52.80	2.57	45.10-60.50	4.87
SG: International Horiba Instruments		25	70.07	3.02	61.00-79.20	4.31	25	57.64	6.43	38.30-76.90	11.16	25	64.60	7.27	42.80-86.40	11.25	25	62.63	5.10	47.30-77.90	8.15	25	51.36	4.84	36.90-65.90	9.42
Eosinophil		Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %
PG: Beckman Coulter Act 5 diff		23	3.98	0.63	2.10-5.90	15.75	23	4.86	0.82	2.40-7.40	16.73	21	8.73	1.82	3.20-14.20	20.92	23	5.56	1.19	2.10-9.10	21.07	23	6.32	0.88	3.70-8.90	13.97
PG: Horiba Pentra 60		16	3.56	0.53	2.00-5.20	14.72	16	4.10	0.98	1.20-7.00	23.90	16	8.63	1.83	3.10-14.10	21.28	16	5.61	1.09	2.30-8.90	19.46	16	5.24	1.10	1.90-8.50	21.15
PG: Horiba Pentra 60C+		6	3.77	0.52	2.20-5.40	13.68	6	4.80	0.98	1.90-7.70	20.42	6	8.75	1.83	3.30-14.30	20.80	6	5.48	1.29	1.60-9.40	23.45	6	5.65	0.83	3.20-8.20	14.56
SG: Select Horiba Instruments (5-part Diff)		22	3.62	0.52	2.00-5.20	14.44	22	4.28	1.01	1.30-7.30	23.49	22	8.66	1.79	3.30-14.10	20.57	22	5.58	1.12	2.20-9.00	20.00	22	5.35	1.03	2.20-8.40	19.43
SG: International Horiba Instruments		25	3.65	0.51	2.20-5.20	13.78	25	4.27	1.02	1.20-7.40	23.72	25	8.70	1.80	3.30-14.10	20.69	25	5.51	1.13	2.10-8.90	20.55	25	5.27	1.01	2.30-8.30	19.06
Basophil		Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %
PG: Beckman Coulter Act 5 diff		22	11.60	0.85	9.07-14.15	7.29	22	8.50	0.81	6.07-10.93	9.52	22	9.07	0.49	7.60-10.54	5.40	22	7.16	0.45	5.81-8.52	6.31	22	5.57	0.52	4.02-7.12	9.28
PG: Horiba Pentra 60		16	0.58	0.04	0.45-0.71	7.59	16	0.40	0.00	0.40-0.40	0.03	16	0.43	0.05	0.29-0.57	10.96	16	0.37	0.05	0.23-0.51	12.67	16	0.30	0.00	0.30-0.30	0.00
PG: Horiba Pentra 60C+		6	0.60	0.00	0.60-0.60	0.00	6	0.43	0.05	0.28-0.59	11.92	5	0.46	0.06	0.30-0.62	11.91	6	0.38	0.04	0.26-0.51	10.65	6	0.32	0.04	0.20-0.44	12.87
SG: Select Horiba Instruments (5-part Diff)		22	0.58	0.04	0.47-0.70	6.66	22	0.41	0.03	0.32-0.50	7.04	21	0.44	0.05	0.29-0.58	11.28	22	0.37	0.04	0.24-0.51	12.01	22	0.30	0.02	0.24-0.37	6.88
SG: International Horiba Instruments		25	0.58	0.04	0.45-0.71	7.45	25	0.41	0.03	0.33-0.49	6.67	24	0.43	0.05	0.29-0.57	11.02	25	0.38	0.04	0.25-0.51	11.41	25	0.30	0.03	0.22-0.39	9.43
Red Blood Cell Count		Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %
PG: Beckman Coulter Act 5 diff		23	2.97	0.05	2.79-3.15	1.58	22	4.48	0.06	4.21-4.75	1.43	23	2.06	0.03	1.94-2.18	1.50	23	4.16	0.06	3.91-4.41	1.47	23	5.59	0.06	5.25-5.93	1.06
PG: Horiba Pentra 60		16	2.94	0.06	2.76-3.12	2.11	16	4.44	0.07	4.17-4.71	1.51	16	2.04	0.04	1.92-2.16	2.11	16	4.13	0.06	3.88-4.38	1.48	16	5.55	0.07	5.22-5.88	1.35
PG: Horiba Pentra 60C+		6	2.95	0.06	2.77-3.13	1.97	6	4.47	0.06	4.20-4.74	1.34	6	2.06	0.02	1.94-2.18	0.97	6	4.12	0.03	3.87-4.37	0.68	6	5.60	0.08	5.26-5.94	1.38
SG: Select Horiba Instruments (5-part Diff)		22	2.94	0.06	2.76-3.12	2.04	22	4.45	0.07	4.18-4.72	1.46	22	2.04	0.04	1.92-2.16	1.91	22	4.12	0.05	3.87-4.37	1.31	22	5.56	0.08	5.23-5.89	1.38
SG: International Horiba Instruments		25	2.94	0																						

PG: Horiba Pentra 60C+	6	14.30	0.83	11.80-16.80	5.80	6	12.60	0.57	10.90-14.30	4.52	6	11.13	0.87	8.50-13.70	7.84	6	13.18	0.63	11.30-15.10	4.77	6	13.08	0.75	10.80-15.30	5.73
SG: Select Horiba Instruments (5-part Diff)	22	14.20	0.64	12.30-16.10	4.51	22	12.43	0.53	10.80-14.00	4.27	22	11.13	0.55	9.40-12.80	4.95	22	12.98	0.63	11.10-14.90	4.85	22	12.84	0.58	11.10-14.50	4.53
SG: International Horiba Instruments	25	14.25	0.67	12.20-16.20	4.72	25	12.49	0.54	10.90-14.10	4.32	25	11.14	0.54	9.50-12.70	4.86	25	13.02	0.63	11.10-14.90	4.85	25	12.92	0.60	11.10-14.70	4.65
Platelet Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman Coulter Act 5 diff	23	380.48	12.12	285.00-475.00	3.18	23	243.91	9.50	183.00-305.00	3.89	23	76.57	4.48	58.00-96.00	5.84	23	249.65	7.13	188.00-313.00	2.84	23	294.96	6.81	221.00-369.00	2.31
PG: Horiba Pentra 60	16	378.41	15.59	284.00-473.00	4.13	16	247.24	9.92	185.00-309.00	4.01	16	76.59	4.26	58.00-96.00	5.58	16	250.24	10.38	188.00-313.00	4.16	16	296.71	10.20	223.00-371.00	3.43
PG: Horiba Pentra 60C+	6	389.50	9.16	293.00-488.00	2.36	6	247.67	15.24	186.00-310.00	6.13	6	77.17	5.04	58.00-96.00	6.49	6	257.00	4.52	193.00-321.00	1.75	6	306.00	17.98	230.00-383.00	5.88
SG: Select Horiba Instruments (5-part Diff)	22	381.30	14.86	286.00-476.00	3.91	22	247.35	11.15	185.00-309.00	4.53	22	76.74	4.36	58.00-96.00	5.71	22	252.00	9.60	189.00-315.00	3.81	22	299.13	12.91	224.00-374.00	4.31
SG: International Horiba Instruments	25	381.42	14.17	286.00-476.00	3.73	25	247.50	10.63	186.00-310.00	4.27	25	76.23	4.34	57.00-95.00	5.66	25	251.15	9.38	188.00-314.00	3.75	25	298.42	12.69	224.00-373.00	4.26
702 - Blood Gases	BG-1					BG-2					BG-3					BG-4					BG-5				
Blood Gas - pH	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Radiometer ABL80 FLEX CO-OX	5	7.43	0.01	7.39-7.47	0.11	5	7.12	0.03	7.08-7.16	0.38	5	7.27	0.01	7.23-7.31	0.14	5	7.27	0.01	7.23-7.31	0.14	5	7.12	0.03	7.08-7.16	0.41
AG: All Group	6	7.45	0.04	7.41-7.49	0.50	6	7.12	0.02	7.08-7.16	0.29	6	7.27	0.01	7.23-7.31	0.11	6	7.28	0.02	7.24-7.32	0.26	6	7.12	0.02	7.08-7.16	0.32
Blood Gas - pO2	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Radiometer ABL80 FLEX CO-OX	5	178.50	11.73	144.00-214.00	6.54	5	79.50	5.07	65.00-95.00	6.38	5	59.25	1.71	54.00-64.00	2.88	5	187.50	6.03	170.00-206.00	3.19	5	76.25	1.50	72.00-81.00	1.97
AG: All Group	6	175.85	9.98	146.00-206.00	5.68	6	82.68	6.30	64.00-102.00	7.59	6	63.57	6.82	44.00-84.00	10.63	6	183.33	7.99	159.00-207.00	4.37	6	80.55	6.76	61.00-101.00	8.40
714 - Complete Cardiac Markers;715 - Basic Cardiac Markers;711 - BNP;720 - Cardiac Markers I;740 - D-Dimer	CDM-1N					CDM-2N					CDM-3N					CDM-4N					CDM-5N				
BNP	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat	30	59.90	6.89	39.20-80.60	11.52	30	1914.40	184.52	1360.80-2468.00	9.64	23	942.04	87.60	679.20-1204.80	9.30	22	444.05	52.59	286.20-601.80	11.84	23	3870.43	323.43	2900.10-4840.70	8.36
PG: AlereTriage	35	51.63	7.25	29.90-73.30	14.05	34	1229.14	170.51	717.60-1740.60	13.87	31	659.28	88.33	394.30-924.30	13.40	31	344.56	38.52	229.00-460.20	11.18	29	2431.33	360.52	1349.70-3512.90	14.83
AG: All Group	66	55.22	8.22	30.50-79.90	14.89	65	1550.22	385.52	393.60-2706.80	24.87	54	769.96	156.19	301.40-1238.60	20.28	53	380.31	63.37	190.20-570.40	16.66	53	3068.94	795.31	683.00-5454.80	25.92
CK-MB (Quantitative)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: AlereTriage	36	1.63	0.40	0.40-2.80	25.00	37	18.84	3.20	9.20-28.40	17.02	37	9.79	1.76	4.50-15.10	17.96	37	5.59	0.86	3.00-8.20	15.36	36	33.69	7.40	11.50-55.90	21.96
PG: Intl OCD Vitros 250,350,400,500,750,950	5	4.60	0.00	4.60-4.60	0.00	5	35.20	0.00	35.20-35.20	0.00	5	20.40	0.00	20.40-20.40	0.00	5	11.80	0.00	11.80-11.80	0.00	5	66.80	0.00	66.80-66.80	0.00
AG: All Group	37	1.67	0.45	0.30-3.00	26.47	38	19.02	3.35	8.90-29.10	17.63	38	9.95	1.99	3.90-15.90	20.10	38	5.73	1.18	2.20-9.20	20.70	37	34.23	8.01	10.20-58.20	23.42
Myoglobin (Quantitative)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: AlereTriage	30	32.54	7.58	9.80-55.20	23.32	31	212.34	35.32	106.30-318.30	16.64	31	132.44	22.93	63.60-201.20	17.32	31	85.37	15.30	39.50-131.30	17.92	31	336.72	63.23	147.00-526.40	18.78
AG: All Group	32	32.45	7.45	10.10-54.90	22.92	33	214.57	36.49	105.10-324.10	17.00	33	132.92	22.82	64.40-201.40	17.17	33	85.28	15.13	39.90-130.70	17.74	33	346.60	74.47	123.20-570.00	21.48
Troponin I (Quantitative)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat	35	0.30	0.03	0.20-0.39	10.37	36	15.07	3.13	5.69-24.45	20.75	28	7.31	0.66	5.32-9.30	9.07	27	3.47	0.34	2.46-4.48	9.72	28	30.97	3.01	21.95-39.99	9.71
PG: AlereTriage	39	0.05	0.00	0.05-0.05	0.00	39	3.68	0.80	1.30-6.07	21.60	37	1.12	0.23	0.41-1.82	20.99	38	0.39	0.10	0.10-0.67	24.62	37	9.12	1.75	3.87-14.37	19.19
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	5	0.02	0.00	0.02-0.02	0.00	5	0.67	0.00	0.67-0.67	0.00	5	0.27	0.00	0.27-0.27	0.00	5	0.12	0.00	0.12-0.12	0.00	5	1.49	0.00	1.49-1.49	0.00
PG: TOSOH AIA	5	0.85	0.00	0.85-0.85	0.00	5	42.19	0.00	42.19-42.19	0.00	5	22.65	0.00	22.65-22.65	0.00	5	11.47	0.00	11.47-11.47	0.00	5	77.50	0.00	77.50-77.50	0.00
PG: Intl Abbott i-Stat	5	290.00	11.55	255.36-324.64	3.98	5	13902.5	274.38	13079.4-14725.6	1.97	5	6885.00	287.35	6022.95-7747.05	4.17	5	3337.50	179.88	2797.84-3877.16	5.39	5	28187.5	1716.98	23036.6-33338.4	6.09
D-dimer (Quantitative)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: AlereTriage	60	578.05	65.55	382.00-775.00	11.34	59	1321.17	129.60	932.40-1710.00	9.81	57	964.34	108.01	640.00-1288.00	11.20	57	775.57	73.77	555.00-997.00	9.51	57	2046.72	195.93	1459.00-2635.00	9.57
SG: Select Cardiac Markers II Insts	64	578.17	65.81	381.00-775.00	11.38	63	1304.06	144.82	869.60-1738.60	11.10	61	952.94	115.71	606.00-1300.00	12.14	61	771.34	76.71	541.00-1001.00	9.95	61	2021.13	215.28	1375.00-2667.00	10.65
AG: All Group	65	578.50	65.36	383.00-775.00	11.30																				
680 - Coagulation - Plasma	CG-1					CG-2					CG-3					CG-4					CG-5				
INR	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Diagnostica Stago STA Satellite	18	3.37	0.67	1.40-5.40	19.71	18	2.36	0.21	1.80-3.00	8.75	18	1.17	0.09	0.90-1.50	7.50	18	1.08	0.06	1.00-1.30	4.55	18	3.65	0.61	1.90-5.50	16.49
AG: All Group	19	3.39	0.51	1.90-4.90	15.00	20	2.26	0.39	1.10-3.50	16.96	20	1.15	0.10	0.90-1.50	8.33	20	1.12	0.20	0.50-1.70	17.27	19	3.60	0.64	1.70-5.50	17.78
612 - MultiChem 5;616 - MultiChem 10;617 - MultiChem 15;618 - MultiChem 20;619 - MultiChem 21+;717 - Abaxis Piccolo (Waived);676 - Endocrinology;719 - Quantitative hCG (Serum)	CH-1N					CH-2N					CH-3N					CH-4N					CH-5N				
HCG - Quant	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Intl Abbott i-Stat	5	48.00	1.00	45.00-51.00	2.08	5	59.00	3.61	48.00-70.00	6.10	5	1192.67	82.50	946.00-1441.00	6.92	5	632.00	25.71	555.00-709.00	4.07	5	844.00	69.09	637.00-1051.00	8.19
AG: All Group	17	55.06	5.48	39.00-72.00	10.00	17	66.10	15.09	21.00-111.00	22.88	17	1821.91	401.82	617.00-3027.00	22.05	17	796.76	196.02	209.00-1385.00	24.59	17	1148.47	291.89	272.0	



Free T3	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman ACCESS	7	2.27	0.21	1.70-2.90	9.13	7	2.41	0.34	1.40-3.40	14.17	7	5.09	0.20	4.50-5.70	3.92	7	4.46	0.16	4.00-5.00	3.56	7	4.70	0.19	4.10-5.30	4.04
PG: TOSOH AIA	6	2.44	0.29	1.50-3.30	12.50	6	5.79	0.56	4.10-7.50	9.66	6	13.12	0.88	10.50-15.70	6.72	6	9.74	0.78	7.40-12.00	8.04	6	11.06	0.63	9.20-13.00	5.68
PG: OCD Vitros 5,1 FS	5	4.20	0.00	4.20-4.20	0.00	5	2.70	0.00	2.70-2.70	0.00	5	5.40	0.00	5.40-5.40	0.00	5	4.70	0.00	4.70-4.70	0.00	5	4.40	0.00	4.40-4.40	0.00
SG: Select Beckman Coulter Instruments (Chem)	10	2.39	0.26	1.60-3.20	10.83	10	2.59	0.48	1.20-4.00	18.46	10	5.10	0.22	4.40-5.80	4.31	10	4.46	0.21	3.90-5.10	4.67	10	4.69	0.20	4.10-5.30	4.26
PG: Intl OCD Vitros ECI/ECIQ	5	2.87	1.07	0.00-6.10	36.90	5	10.92	2.72	2.70-19.10	24.95	5	19.37	2.37	12.30-26.50	12.22	5	17.97	1.82	12.50-23.50	10.11	5	18.43	1.97	12.50-24.30	10.71
AG: All Group	22	2.29	0.37	1.20-3.40	16.09	16	2.61	0.56	0.90-4.30	21.54	16	5.07	0.56	3.40-6.80	10.98	16	4.44	0.64	2.50-6.30	14.55	16	4.64	0.54	3.00-6.20	11.74
Potassium (K+)																									
PG: Abaxis Piccolo (Non-Waived)	9	2.43	0.15	1.90-2.90	6.25	10	2.23	0.34	1.70-2.70	15.45	10	6.71	0.37	6.20-7.20	5.52	10	4.12	0.25	3.60-4.60	6.10	10	5.01	0.30	4.50-5.50	6.00
PG: Abaxis Piccolo (Waived)	40	2.41	0.15	1.90-2.90	6.25																				
PG: Abbott Architect	6	2.53	0.05	2.00-3.00	2.00	6	2.43	0.08	1.90-2.90	3.33	6	6.63	0.05	6.10-7.10	0.76	6	4.10	0.00	3.60-4.60	0.00	6	4.91	0.07	4.40-5.40	1.43
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	32	2.66	0.12	2.20-3.20	4.44	32	2.36	0.08	1.90-2.90	3.33	32	6.87	0.13	6.40-7.40	1.88	32	4.12	0.06	3.60-4.60	1.46	32	4.98	0.06	4.50-5.50	1.20
PG: Clinical Data Envoy 500	6	2.51	0.08	2.00-3.00	3.20	6	2.38	0.09	1.90-2.90	3.75	6	6.40	0.14	5.90-6.90	2.19	6	3.95	0.05	3.50-4.50	1.25	6	4.71	0.28	4.20-5.20	5.74
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	18	2.43	0.04	1.90-2.90	2.08	18	2.39	0.02	1.90-2.90	0.83	18	6.55	0.05	6.10-7.10	0.76	18	4.06	0.05	3.60-4.60	1.22	18	4.88	0.05	4.40-5.40	1.02
PG: Beckman (Olympus) AU	17	2.51	0.03	2.00-3.00	1.20	16	2.42	0.04	1.90-2.90	1.67	16	6.43	0.11	5.90-6.90	1.72	16	4.02	0.04	3.50-4.50	1.00	16	4.80	0.05	4.30-5.30	1.04
PG: Roche Cobas Integra / Roche Reagent	6	2.47	0.05	2.00-3.00	2.00	6	2.38	0.04	1.90-2.90	1.67	6	6.48	0.10	6.00-7.00	1.54	6	4.18	0.26	3.70-4.70	6.19	6	4.87	0.08	4.40-5.40	1.63
PG: OCD Vitros 250,350,400,500,750,950	10	2.69	0.05	2.20-3.20	1.85	10	2.48	0.06	2.00-3.00	2.40	10	6.92	0.13	6.40-7.40	1.88	10	4.26	0.08	3.80-4.80	1.86	10	5.13	0.09	4.60-5.60	1.76
SG: Select Beckman Coulter Instruments (Chem)	19	2.50	0.05	2.00-3.00	2.00	18	2.41	0.05	1.90-2.90	2.08	18	6.44	0.11	5.90-6.90	1.72	18	4.02	0.04	3.50-4.50	1.00	18	4.80	0.05	4.30-5.30	1.04
SG: Select Roche Instruments / Roche Reagents Groups	7	2.50	0.10	2.00-3.00	4.00	7	2.40	0.06	1.90-2.90	2.50	7	6.46	0.11	6.00-7.00	1.69	7	4.17	0.24	3.70-4.70	5.71	7	4.86	0.08	4.40-5.40	1.63
SG: Select OCD Vitros Instruments	13	2.69	0.04	2.20-3.20	1.48	13	2.49	0.05	2.00-3.00	2.00	13	6.90	0.12	6.40-7.40	1.74	13	4.30	0.20	3.80-4.80	4.65	13	5.06	0.27	4.60-5.60	5.29
SG: Select Abbott Architect Instruments	8	2.53	0.05	2.00-3.00	2.00	8	2.43	0.07	1.90-2.90	2.92	8	6.63	0.05	6.10-7.10	0.76	8	4.10	0.00	3.60-4.60	0.00	8	4.92	0.07	4.40-5.40	1.43
SG: International OCD Vitros Instruments	14	2.71	0.05	2.20-3.20	1.85	14	2.50	0.06	2.00-3.00	2.40	14	6.92	0.12	6.40-7.40	1.74	14	4.26	0.07	3.80-4.80	1.63	14	5.13	0.10	4.60-5.60	1.96
SG: International Siemens Instruments	20	2.45	0.08	1.90-2.90	3.33	20	2.39	0.04	1.90-2.90	1.67	20	6.54	0.07	6.00-7.00	1.08	20	4.05	0.06	3.50-4.50	1.50	20	4.88	0.05	4.40-5.40	1.02
SG: International Abaxis Piccolo Instruments	53	2.41	0.15	1.90-2.90	6.25	14	2.23	0.29	1.70-2.70	13.18	14	6.69	0.35	6.20-7.20	5.22	14	4.15	0.22	3.60-4.60	5.37	14	4.97	0.27	4.50-5.50	5.40
SG: Select Gen Chem Instruments	37	2.53	0.09	2.00-3.00	3.60	36	2.42	0.09	1.90-2.90	3.75	36	6.49	0.14	6.00-7.00	2.15	36	4.03	0.06	3.50-4.50	1.50	36	4.82	0.14	4.30-5.30	2.92
AG: All Group	156	2.52	0.15	2.00-3.00	6.00	116	2.39	0.13	1.90-2.90	5.42	116	6.67	0.23	6.20-7.20	3.43	116	4.11	0.15	3.60-4.60	3.66	116	4.92	0.18	4.40-5.40	3.67
Chloride (CL-)																									
PG: Abaxis Piccolo (Non-Waived)	10	93.27	2.24	88.00-98.00	2.37	10	89.70	1.95	86.00-95.00	2.11	10	116.80	2.49	111.00-123.00	2.14	10	100.91	2.95	96.00-106.00	2.87	10	106.45	2.50	101.00-111.00	2.36
PG: Abaxis Piccolo (Waived)	41	93.85	1.70	89.00-99.00	1.81																				
PG: Abbott Architect	6	90.00	1.00	86.00-95.00	1.11	6	86.14	0.69	82.00-90.00	0.81	6	114.86	0.90	109.00-121.00	0.78	6	97.86	0.90	93.00-103.00	0.92	6	103.43	0.79	98.00-108.00	0.78
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	30	88.09	1.94	84.00-92.00	2.16	30	82.22	1.37	78.00-86.00	1.71	30	120.83	1.45	115.00-127.00	1.16	30	98.15	1.17	93.00-103.00	1.22	30	105.34	0.93	100.00-110.00	0.86
PG: Clinical Data Envoy 500	6	91.01	2.13	86.00-96.00	2.31	6	83.88	1.25	80.00-88.00	1.43	6	122.46	3.91	116.00-128.00	3.20	6	101.13	1.46	96.00-106.00	1.49	6	108.76	1.92	104.00-114.00	1.74
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	18	82.89	0.66	79.00-87.00	0.84	18	82.05	0.85	78.00-86.00	0.98	18	122.00	1.76	116.00-128.00	1.48	18	98.00	0.94	93.00-103.00	0.92	18	105.84	0.77	101.00-111.00	0.75
PG: Beckman (Olympus) AU	16	87.41	1.12	83.00-91.00	1.26	15	85.06	1.12	81.00-89.00	1.29	15	115.70	1.80	110.00-122.00	1.55	15	97.54	1.34	93.00-103.00	1.33	15	103.61	1.50	99.00-109.00	1.44
PG: Roche Cobas Integra / Roche Reagent	6	84.17	1.17	80.00-88.00	1.43	6	84.83	1.72	81.00-89.00	2.00	6	121.33	4.97	115.00-127.00	4.13	6	99.83	3.43	95.00-105.00	3.40	6	106.67	4.50	102.00-112.00	4.21
PG: Roche Cobas Mira / DCL Reagent (Sekisui)	5	82.00	0.00	78.00-86.00	0.00	5	79.00	0.00	75.00-83.00	0.00	5	119.00	0.00	113.00-125.00	0.00	5	94.00	0.00	89.00-99.00	0.00	5	103.00	0.00	98.00-108.00	0.00
PG: OCD Vitros 250,350,400,500,750,950	10	88.08	0.90	84.00-92.00	1.02	10	84.75	1.05	81.00-89.00	1.29	10	115.42	1.83	109.00-121.00	1.57	10	97.00	1.60	92.00-102.00	1.65	10	103.25	1.22	98.00-108.00	1.17
SG: Select Beckman Coulter Instruments (Chem)	18	87.53	1.12	84.00-92.00	1.25	17	85.11	1.08	81.00-89.00	1.29	17	115.46	1.85	109.00-121.00	1.57	17	97.48	1.27	92.00-102.00	1.34	17	103.43	1.54	98.00-108.00	1.46
SG: Select OCD Vitros Instruments	13	87.75	1.34	84.00-92.00	1.48	13	84.31	1.58	80.00-88.00	1.90	13	115.00	1.97	109.00-121.00	1.74	13	96.88	2.53	92.00-102.00	2.58	13	102.31	2.27	97.00-107.00	2.25
SG: Select Abbott Architect Instruments	8	90.00	0.87	86.00-95.00	1.00	8	86.22	0.67	82.00-90.00	0.81	8	114.67	0.87	109.00-121.00	0.78	8	97.78	0.83	93.00-103.00	0.82	8	103.33	0.71	98.00-108.00	0.68
SG: International OCD Vitros Instruments	14	87.95	1.51	84.00-92.00	1.70	14	84.68	1.64	81.00-89.00	1.88	14	115.11	1.88	109.00-121.00	1.65	14	96.78	1.93	92.00-102.00	1.96	14	103.06	1.89	98.00-108.00	1.84
SG: International Siemens Instruments	20	83.38	1.69	79.00-87.00	2.05	20	82.38	1.32	78.00-86.00	1.59	20	121.95	1.69	116.00-128.00	1.39	20	98.14	1.01	93.00-103.00	1.02	20	106.00	0.89	101.00-111.00	0.85
SG: International Abaxis Piccolo Instruments	55	93.68	1.79	89.00-99.00	1.91	14	89.50	1.70	86.00-95.00	1.89	14	117.00	2.39	111.00-123.00	2.05	14	101.20	2.78	96.00-106.00	2.77	14	106.53	2.13	102.00-112.00	1.96
SG: Select Gen Chem Instruments	36	88.64	2.24	85.00-93.00	2.47	35	84.95	1.38	81.00-89.00	1.65	35	117.13	4.09	111.00-123.00	3.50	35	98.35	1.95	93.00-103.00	2.04	35	104.66	2.71	100.00-110.00	2.57
Free T4																									
PG: Abbott Architect	7	1.35	0.04	1.22-1.48	3.31	7	3.22	0.26	2.45-3.98	7.93	7	4.77	0.50	3.28-6.26	10.40	7	3.92	0.20	3.34-4.51	4.98	7	4.10	0.31	3.16-5.03	7.62
PG: Beckman ACCESS	17	0.87	0.05	0.73-1.01	5.26	17	3.52	0.17	3.02-4.02	4.75	17	2.94	0.20	2.35-3.53	6.71	17	3.16	0.12	2.80-3.53	3.85	17	3.07	0.16	2.60-3.54	5.10
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	12	0.98	0.15	0.53-1.43	15.25	12	4.62	0.67	2.61-6.63	14.49	12	4.16	0.34	3.16-5.17	8.05	12	4.32	0.46	2.93-5.70	10.68	12	4.23	0.35	3.19-5.28	8.24



	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %
SG: Select OCD Vitros Instruments	13	10.38	1.15	7.00-13.00	11.00	13	11.13	1.02	8.00-14.00	9.09	13	34.81	2.43	28.00-42.00	6.86	12	18.93	2.12	13.00-25.00	11.05	13	23.25	1.98	17.00-29.00	8.70
SG: Select Abbott Architect Instruments	8	11.00	2.50	4.00-19.00	22.73	7	11.00	1.20	7.00-15.00	10.91	8	36.89	2.85	29.00-45.00	7.57	8	20.44	2.51	13.00-28.00	12.50	8	25.11	2.52	18.00-33.00	10.00
SG: International OCD Vitros Instruments	13	10.29	1.16	6.00-14.00	12.00	13	10.94	1.03	8.00-14.00	9.09	13	34.53	2.40	28.00-42.00	6.86	13	18.76	1.99	13.00-25.00	10.53	13	23.06	1.82	18.00-28.00	7.83
SG: International Abaxis Piccolo Instruments	54	12.67	1.20	9.00-17.00	9.23	13	12.08	1.04	9.00-15.00	8.33	13	33.69	2.06	28.00-40.00	6.18	13	22.71	0.82	21.00-25.00	3.48	13	26.64	1.08	24.00-30.00	4.07
AG: All Group	150	12.87	2.07	7.00-19.00	16.15	109	12.46	1.83	7.00-17.00	15.00	112	37.17	4.04	25.00-49.00	10.81	111	22.40	2.85	13.00-31.00	13.18	112	26.80	3.38	17.00-37.00	12.59
Calcium	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %
PG: Abaxis Piccolo (Non-Waived)	9	7.50	0.14	6.50-8.50	1.87	9	7.28	0.36	6.30-8.30	4.93	9	13.18	0.39	12.20-14.20	2.95	9	9.88	0.27	8.90-10.90	2.73	9	11.11	0.20	10.10-12.10	1.80
PG: Abaxis Piccolo (Waived)	41	7.39	0.15	6.40-8.40	2.03																				
PG: Abbott Architect	6	7.56	0.14	6.60-8.60	1.84	6	6.99	0.16	6.00-8.00	2.29	6	13.66	0.24	12.70-14.70	1.75	6	9.61	0.16	8.60-10.60	1.67	6	10.93	0.18	9.90-11.90	1.65
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	31	7.63	0.20	6.60-8.60	2.63	31	7.36	0.17	6.40-8.40	2.30	31	13.45	0.23	12.50-14.50	1.70	31	9.98	0.17	9.00-11.00	1.70	31	11.18	0.20	10.20-12.20	1.79
PG: Clinical Data Envoy 500	6	8.04	0.34	7.00-9.00	4.25	6	6.72	0.27	5.70-7.70	4.03	6	13.43	0.38	12.40-14.40	2.84	6	9.54	0.35	8.50-10.50	3.68	6	10.80	0.34	9.80-11.80	3.15
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	17	6.92	0.15	5.90-7.90	2.17	17	6.94	0.15	5.90-7.90	2.17	17	13.60	0.27	12.60-14.60	1.99	17	9.57	0.19	8.60-10.60	1.98	17	10.83	0.21	9.80-11.80	1.94
PG: Beckman (Olympus) AU	17	7.67	0.21	6.70-8.70	2.73	16	6.95	0.36	6.00-8.00	5.14	16	13.44	0.47	12.40-14.40	3.51	16	9.55	0.21	8.50-10.50	2.21	16	10.83	0.27	9.80-11.80	2.50
PG: Roche Cobas Integra / Roche Reagent	6	7.33	0.23	6.30-8.30	3.15	5	7.08	0.26	6.10-8.10	3.66	6	13.65	0.45	12.70-14.70	3.28	6	9.78	0.30	8.80-10.80	3.06	6	11.02	0.33	10.00-12.00	3.00
PG: OCD Vitros 250,350,400,500,750,950	10	7.34	0.19	6.30-8.30	2.60	10	6.98	0.23	6.00-8.00	3.29	10	13.29	0.33	12.30-14.30	2.48	10	9.82	0.32	8.80-10.80	3.27	10	11.07	0.31	10.10-12.10	2.79
SG: Select Beckman Coulter Instruments (Chem)	19	7.64	0.23	6.60-8.60	3.03	18	6.97	0.34	6.00-8.00	4.86	18	13.43	0.45	12.40-14.40	3.36	18	9.56	0.21	8.60-10.60	2.19	18	10.84	0.26	9.80-11.80	2.41
SG: Select Roche Instruments / Roche Reagents Groups	7	7.41	0.30	6.40-8.40	4.05	6	7.02	0.28	6.00-8.00	4.00	7	13.43	0.71	12.40-14.40	5.30	7	9.74	0.29	8.70-10.70	2.99	7	10.97	0.32	10.00-12.00	2.91
SG: Select OCD Vitros Instruments	13	7.34	0.16	6.30-8.30	2.19	13	6.96	0.21	6.00-8.00	3.00	13	13.25	0.30	12.30-14.30	2.26	13	9.88	0.39	8.90-10.90	3.94	13	10.96	0.43	10.00-12.00	3.91
SG: Select Abbott Architect Instruments	8	7.54	0.13	6.50-8.50	1.73	8	6.99	0.14	6.00-8.00	2.00	8	13.66	0.21	12.70-14.70	1.53	8	9.62	0.14	8.60-10.60	1.46	8	10.93	0.16	9.90-11.90	1.47
SG: International OCD Vitros Instruments	14	7.38	0.17	6.40-8.40	2.43	14	6.97	0.20	6.00-8.00	2.71	14	13.32	0.31	12.30-14.30	2.33	14	9.82	0.28	8.80-10.80	2.86	14	11.09	0.30	10.10-12.10	2.70
SG: International Abaxis Piccolo Instruments	55	7.41	0.17	6.40-8.40	2.30	14	7.32	0.35	6.30-8.30	4.79	14	13.17	0.36	12.20-14.20	2.73	14	9.91	0.26	8.90-10.90	2.63	14	11.13	0.23	10.10-12.10	2.16
SG: Select Gen Chem Instruments	36	7.69	0.32	6.70-8.70	4.16	35	6.91	0.29	5.90-7.90	4.20	35	13.46	0.39	12.50-14.50	2.89	35	9.57	0.23	8.60-10.60	2.40	35	10.84	0.26	9.80-11.80	2.41
T3 Uptake	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %
AG: All Group	5	0.26	0.05	0.11-0.40	18.95	5	0.52	0.03	0.45-0.60	4.79	5	0.40	0.04	0.28-0.53	10.30	5	0.48	0.03	0.38-0.57	6.68	5	0.45	0.03	0.35-0.54	7.19
Creatinine	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %
PG: Abaxis Piccolo (Non-Waived)	9	0.51	0.10	0.21-0.81	20.00	10	0.68	0.15	0.38-0.98	21.76	10	5.06	0.18	4.30-5.82	3.53	10	2.38	0.09	2.00-2.80	3.75	10	3.33	0.18	2.83-3.83	5.45
PG: Abaxis Piccolo (Waived)	38	0.55	0.08	0.25-0.85	15.27																				
PG: Abbott Architect	6	0.71	0.01	0.41-1.01	1.55	6	0.49	0.01	0.19-0.79	2.00	6	5.52	0.11	4.70-6.30	1.90	6	2.41	0.05	2.00-2.80	2.08	6	3.44	0.05	2.92-3.96	1.54
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	31	0.60	0.06	0.30-0.90	10.33	32	0.62	0.05	0.32-0.92	8.39	32	4.79	0.28	4.07-5.51	5.83	32	2.44	0.18	2.00-2.80	7.50	32	3.23	0.20	2.75-3.71	6.25
PG: Clinical Data Envoy 500	5	0.48	0.02	0.18-0.78	4.00	6	0.56	0.07	0.26-0.86	11.96	6	4.79	0.23	4.10-5.50	4.84	6	2.32	0.11	2.00-2.60	4.78	6	3.24	0.20	2.75-3.73	6.25
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	12	0.53	0.04	0.23-0.83	8.11	12	0.43	0.04	0.13-0.73	10.00	13	5.20	0.13	4.40-6.00	2.50	13	2.37	0.09	2.00-2.80	3.75	13	3.32	0.09	2.82-3.82	2.80
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL / IFCC	5	0.53	0.06	0.23-0.83	12.00	5	0.45	0.07	0.15-0.75	17.50	5	5.08	0.39	4.30-5.90	7.65	5	2.33	0.15	1.98-2.68	6.61	5	3.21	0.20	2.73-3.69	6.26
PG: Beckman (Olympus) AU	17	0.38	0.02	0.08-0.68	6.32	16	0.47	0.04	0.17-0.77	8.00	16	4.83	0.16	4.10-5.50	3.37	16	2.30	0.07	2.00-2.60	3.04	16	3.16	0.10	2.69-3.63	3.16
PG: Roche Cobas Integra / Roche Reagent	6	0.50	0.06	0.20-0.80	12.60	6	0.47	0.05	0.17-0.77	11.06	6	4.47	0.19	3.80-5.14	4.34	6	2.26	0.06	2.00-2.60	2.70	6	3.04	0.13	2.58-3.50	4.33
PG: Roche Cobas Mira / DCL Reagent (Sekisui)	5	0.80	0.00	0.50-1.10	0.00	5	0.80	0.00	0.50-1.10	0.00	5	5.40	0.00	4.59-6.21	0.00	5	2.70	0.00	2.30-3.10	0.00	5	3.60	0.00	3.06-4.14	0.00
PG: OCD Vitros 250,350,400,500,750,950	9	0.58	0.05	0.28-0.88	8.62	10	0.51	0.05	0.21-0.81	10.00	10	4.85	0.16	4.10-5.50	3.34	10	2.29	0.07	2.00-2.60	3.04	10	3.20	0.07	2.72-3.68	2.19
SG: Select Beckman Coulter Instruments (Chem)	19	0.39	0.04	0.09-0.69	10.00	18	0.47	0.04	0.17-0.77	8.00	18	4.85	0.16	4.10-5.50	3.38	18	2.31	0.06	2.00-2.60	2.77	18	3.17	0.10	2.69-3.65	3.19
SG: Select OCD Vitros Instruments	12	0.57	0.05	0.27-0.87	8.95	13	0.51	0.05	0.21-0.81	10.00	13	4.84	0.14	4.10-5.50	3.13	13	2.37	0.23	2.00-2.80	9.79	13	3.17	0.22	2.69-3.65	6.88
PG: Polymedco Spot-Chem EZ	5	0.74	0.00	0.44-1.04	0.00	5	0.66	0.00	0.36-0.96	0.00	5	2.45	0.00	2.10-2.90	0.00	5	1.97	0.00	1.70-2.30	0.00	5	2.28	0.00	1.94-2.62	0.00
SG: Select Abbott Architect Instruments	8	0.72	0.03	0.42-1.02	4.72	8	0.49	0.01	0.19-0.79	2.00	8	5.52	0.12	4.70-6.30	2.25	8	2.41	0.06	2.00-2.80	2.66	8	3.45	0.08	2.93-3.97	2.38
SG: International OCD Vitros Instruments	13	0.57	0.05	0.27-0.87	8.42	14	0.52	0.05	0.22-0.82	10.00	14	4.85	0.14	4.10-5.50	2.92	14	2.32	0.07	2.00-2.60	3.04	14	3.21	0.07	2.73-3.69	2.50
SG: International Abaxis Piccolo Instruments	51	0.54	0.09	0.24-0.84	18.00	15	0.69	0.12	0.39-0.99	17.97	15	5.12	0.18	4.35-5.89	3.53	15	2.41	0.14	2.00-2.80	5.83	15	3.38	0.18	2.87-3.89	5.33
MG: Method Group	144	0.54	0.10	0.24-0.84	18.70	107	0.53	0.09	0.23-0.83	16.98	111	4.90	0.38	4.17-5.64	7.82	111	2.40	0.16	2.00-2.80	6.67	111	3.23	0.21	2.75-3.71	6.38
MG: Method Group	144	0.54	0.10	0.24-0.84	18.70	107	0.53	0.09	0.23-0.83	16.98	111	4.90	0.38	4.17-5.64	7.82	111	2.40	0.16	2.00-2.80	6.67	111	3.23	0.21	2.75-3.71	6.38
Thyroxine (T4)	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %	Labs	Mean	SD	Range	CV. %
PG: Beckman ACCESS	7	7.08	0.69	5.70-8.50	9.72	7	7.81	0.81	6.20-9.40	10.38	7	15.06	0.88	12.10-18.10	5.83	7	12.05	0.67	9.70-14.50	5.54	7	13.01	0.60	10.40-15.60	4.62
PG: TOSOH AIA	5	8.20	0.80	6.60-9.80	9.76	5	7.10	0.69	5.70-8.50	9.72	5	15.86	1.64	12.70-19.10	10.31	5	10.88	1.41	8.70-13.10	12.94	5	12.82	0.94	10.20-15.40	7.34
SG: Select Beckman Coulter Instruments (Chem)	8	7.03	0.65	5.60-8.40	9.29	8	7.89	0.78																	



Triiodothyronine (T3)		Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: TOSOH AIA		5	1.05	0.08	0.80-1.20	8.00	5	2.31	0.13	1.90-2.70	5.65	5	5.40	0.01	5.40-5.40	0.19	5	4.09	0.01	4.10-4.10	0.24	5	4.48	0.17	4.00-5.00	3.78
SG: Select Beckman Coulter Instruments (Chem)		6	1.34	0.13	0.90-1.70	9.23	6	1.04	0.12	0.60-1.40	12.00	6	2.69	0.24	2.00-3.40	8.89	6	1.61	0.12	1.20-2.00	7.50	6	1.92	0.14	1.50-2.30	7.37
PG: OCD Vitros 5600		5	1.30	0.00	1.30-1.30	0.00	5	4.80	0.00	4.80-4.80	0.00	5	7.80	0.00	7.80-7.80	0.00	5	7.80	0.00	7.80-7.80	0.00	5	7.81	0.00	7.80-7.80	0.00
AG: All Group		8	1.03	0.38	0.00-2.10	38.00	9	1.00	0.20	0.40-1.60	20.00	9	2.38	0.60	0.60-4.20	25.00	9	1.55	0.22	0.80-2.20	14.67	9	1.79	0.34	0.80-2.80	18.89
TSH		Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Architect		7	2.31	0.12	1.94-2.68	5.37	7	0.51	0.01	0.47-0.55	2.75	7	8.02	0.18	7.47-8.57	2.28	7	3.74	0.15	3.28-4.20	4.06	7	5.24	0.15	4.80-5.68	2.82
PG: Beckman ACCESS		19	2.51	0.14	2.10-2.92	5.46	19	0.51	0.04	0.39-0.63	8.04	18	8.08	0.78	5.73-10.43	9.70	20	3.81	0.37	2.69-4.93	9.82	20	5.55	0.50	4.05-7.05	8.99
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL		15	2.60	0.14	2.17-3.03	5.50	15	0.49	0.03	0.39-0.59	6.94	15	7.90	0.61	6.06-9.74	7.77	15	3.77	0.24	3.05-4.49	6.34	15	5.18	0.37	4.07-6.29	7.14
PG: Qualigen FastPak Thyroid		19	2.46	0.15	2.01-2.91	6.10	22	0.48	0.08	0.23-0.73	17.08	23	8.33	0.67	6.33-10.33	8.02	23	4.07	0.32	3.12-5.02	7.76	22	5.80	0.45	4.46-7.14	7.69
PG: TOSOH AIA		20	3.03	0.17	2.51-3.55	5.74	20	0.74	0.03	0.64-0.84	4.59	20	10.10	0.61	8.27-11.93	6.04	20	4.97	0.33	3.98-5.96	6.66	20	6.72	0.49	5.25-8.19	7.31
SG: Select Beckman Coulter Instruments (Chem)		23	2.54	0.14	2.11-2.97	5.63	23	0.50	0.04	0.38-0.62	8.00	22	8.17	0.74	5.96-10.38	9.02	24	3.80	0.34	2.76-4.84	9.08	24	5.50	0.47	4.09-6.91	8.53
PG: Roche Cobas e 411		6	2.73	0.13	2.35-3.11	4.69	6	0.73	0.04	0.61-0.85	5.34	6	7.52	0.55	5.88-9.16	7.26	6	4.12	0.22	3.47-4.77	5.27	6	5.37	0.28	4.52-6.22	5.29
PG: Abbott Architect i1000 SR		6	2.31	0.09	2.05-2.57	3.81	6	0.51	0.02	0.44-0.58	4.31	6	7.91	0.23	7.22-8.60	2.92	6	3.63	0.11	3.31-3.95	2.98	6	5.12	0.17	4.60-5.64	3.38
SG: Select Abbott Architect Instruments		15	2.32	0.10	2.01-2.63	4.40	15	0.51	0.02	0.46-0.56	3.53	15	8.00	0.21	7.36-8.64	2.68	15	3.69	0.13	3.30-4.08	3.55	15	5.20	0.17	4.70-5.70	3.19
PG: NanoEnTek FrenD System		7	2.32	0.22	1.65-2.99	9.66	7	0.34	0.09	0.06-0.62	27.06	7	4.44	0.60	2.65-6.23	13.47	7	1.65	0.27	0.85-2.45	16.24	7	2.56	0.34	1.54-3.58	13.24
PG: Intl OCD Vitros ECI/ECIQ		5	2.49	0.06	2.30-2.68	2.57	5	0.53	0.02	0.47-0.59	3.77	5	10.43	0.15	9.97-10.89	1.47	5	5.18	0.09	4.91-5.45	1.76	5	7.28	0.19	6.72-7.84	2.55
AG: All Group		111	2.62	0.28	1.77-3.47	10.80	110	0.55	0.11	0.21-0.89	20.73	114	8.30	1.41	4.07-12.53	16.98	110	4.12	0.61	2.29-5.95	14.78	109	5.71	0.71	3.58-7.84	12.45
Urea Nitrogen (BUN)		Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)		10	4.82	0.41	3.00-7.00	8.00	10	4.40	0.52	2.00-6.00	12.50	10	41.50	0.85	38.00-46.00	1.90	10	18.64	0.67	17.00-21.00	3.68	10	26.09	0.54	24.00-28.00	1.92
PG: Abaxis Piccolo (Waived)		40	4.87	0.42	3.00-7.00	8.00																				
PG: Abbott Architect		6	6.43	0.54	4.00-8.00	8.33	6	6.00	0.00	4.00-8.00	0.00	6	45.29	1.25	41.00-49.00	2.89	6	21.29	0.49	19.00-23.00	2.38	6	29.14	0.69	26.00-32.00	2.41
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent		32	5.89	0.53	4.00-8.00	8.33	32	6.03	0.51	4.00-8.00	8.33	32	43.26	2.39	39.00-47.00	5.58	32	20.94	1.24	19.00-23.00	5.71	32	28.31	1.47	25.00-31.00	5.36
PG: Clinical Data Envoy 500		5	5.22	0.44	3.00-7.00	8.00	5	5.50	0.76	4.00-8.00	13.33	5	44.13	0.99	40.00-48.00	2.27	5	21.38	0.52	19.00-23.00	2.38	5	29.00	1.41	26.00-32.00	4.83
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL		17	6.06	0.72	4.00-8.00	11.67	18	6.00	0.75	4.00-8.00	11.67	18	44.32	2.03	40.00-48.00	4.55	18	21.47	1.07	19.00-23.00	5.24	18	29.26	1.49	26.00-32.00	5.17
PG: Beckman (Olympus) AU		17	6.44	0.51	4.00-8.00	8.33	16	6.12	0.33	4.00-8.00	5.00	16	43.65	1.17	40.00-48.00	2.73	16	21.06	0.83	19.00-23.00	3.81	15	28.38	0.72	25.00-31.00	2.50
PG: Roche Cobas Integra / Roche Reagent		6	5.83	0.75	4.00-8.00	13.33	6	5.67	0.52	4.00-8.00	8.33	6	42.00	1.90	38.00-46.00	4.52	6	19.83	1.17	18.00-22.00	6.00	6	27.33	1.21	25.00-29.00	4.44
PG: Roche Cobas Mira / DCL Reagent (Sekisui)		5	7.00	0.00	5.00-9.00	0.00	5	8.00	0.00	6.00-10.00	0.00	5	46.00	0.00	42.00-50.00	0.00	5	23.00	0.00	21.00-25.00	0.00	5	32.00	0.00	29.00-35.00	0.00
PG: OCD Vitros 250,350,400,500,750,950		10	5.00	0.60	3.00-7.00	12.00	10	4.58	0.52	3.00-7.00	10.00	10	37.33	0.65	34.00-40.00	1.89	10	17.58	0.52	16.00-20.00	2.78	10	24.08	0.67	22.00-26.00	2.92
SG: Select Beckman Coulter Instruments (Chem)		19	6.50	0.51	5.00-9.00	7.14	18	6.11	0.32	4.00-8.00	5.00	18	43.74	1.15	40.00-48.00	2.50	18	21.16	0.83	19.00-23.00	3.81	17	28.44	0.78	25.00-31.00	2.86
SG: Select OCD Vitros Instruments		13	4.94	0.57	3.00-7.00	12.00	13	4.44	0.51	2.00-6.00	12.50	13	37.19	0.66	34.00-40.00	1.89	13	17.75	1.24	16.00-20.00	6.67	13	23.38	2.09	21.00-25.00	9.13
PG: Polymedco Spot-Chem EZ		5	6.60	0.00	5.00-9.00	0.00	5	6.60	0.00	5.00-9.00	0.00	5	19.00	0.00	17.00-21.00	0.00	5	18.80	0.00	17.00-21.00	0.00	5	23.80	0.00	22.00-26.00	0.00
SG: Select Abbott Architect Instruments		8	6.33	0.50	4.00-8.00	8.33	8	6.00	0.00	4.00-8.00	0.00	8	45.22	1.09	41.00-49.00	2.44	8	21.22	0.44	19.00-23.00	1.90	8	29.11	0.60	26.00-32.00	2.07
SG: International OCD Vitros Instruments		14	5.00	0.47	3.00-7.00	10.00	14	4.42	0.51	2.00-6.00	12.50	14	37.39	0.61	34.00-40.00	1.62	14	17.56	0.51	16.00-20.00	2.78	14	23.94	0.64	22.00-26.00	2.50
SG: International Abaxis Piccolo Instruments		55	4.86	0.40	3.00-7.00	8.00	15	4.43	0.47	2.00-6.00	12.50	15	41.80	0.88	38.00-46.00	2.14	15	18.76	0.59	17.00-21.00	3.16	15	26.24	0.64	24.00-28.00	2.31
SG: Select Gen Chem Instruments		35	6.17	0.74	4.00-8.00	11.67	34	5.90	0.50	4.00-8.00	8.33	34	43.97	1.42	40.00-48.00	3.18	34	21.15	0.71	19.00-23.00	3.33	33	28.61	1.05	26.00-32.00	3.79
AG: All Group		155	5.58	0.83	4.00-8.00	13.33	115	5.66	0.85	4.00-8.00	15.00	115	42.50	3.53	39.00-47.00	8.14	115	20.44	1.63	18.00-22.00	8.00	114	27.68	2.33	25.00-31.00	8.21
Albumin		Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived) / BCP		8	5.48	0.15	5.00-6.00	2.73	8	1.81	0.06	1.60-2.00	3.33	8	3.53	0.12	3.10-3.90	3.43	8	2.54	0.10	2.30-2.80	4.00	8	2.87	0.09	2.60-3.20	3.10
PG: Abaxis Piccolo (Waived) / BCP		43	5.39	0.10	4.90-5.90	1.85																				
PG: Abbott Architect / BCP		6	5.44	0.17	4.90-5.90	3.15	6	1.73	0.05	1.50-1.90	2.94	6	3.26	0.14	3.00-3.60	4.24	6	2.37	0.13	2.20-2.60	5.42	6	2.66	0.10	2.40-3.00	3.70
PG: Alfa Was. Ace/NexCT/Alera/Axcel / BCG / Alfa Wasserman Reagent		32	5.99	0.14	5.40-6.60	2.33	32	1.97	0.06	1.80-2.20	3.00	32	4.53	0.54	4.00-5.00	12.00	32	3.09	0.08	2.80-3.40	2.58	32	3.62	0.07	3.20-4.00	1.94
PG: Clinical Data Envoy 500 / BCG		6	5.82	0.12	5.20-6.40	2.07	6	1.98	0.07	1.80-2.20	3.50	6	4.61	0.14	4.10-5.10	3.04	6	3.07	0.13	2.80-3.40	4.19	6	3.57	0.12	3.20-4.00	3.33
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL / BCG		11	5.88	0.3																						



Lab Name	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Beckman Coulter Instruments (Chem)	19	77.75	7.24	55.00-101.00	9.23	18	31.95	3.26	22.00-42.00	10.31	18	292.00	24.51	204.00-380.00	8.39	18	143.05	12.23	100.00-186.00	8.53	18	194.32	16.51	136.00-252.00	8.51
SG: Select OCD Vitros Instruments	12	97.40	8.56	68.00-126.00	8.87	12	39.73	4.93	28.00-52.00	12.25	12	156.20	16.90	109.00-203.00	10.83	12	126.80	13.22	89.00-165.00	10.39	12	143.13	13.62	100.00-186.00	9.51
SG: Select Abbott Architect Instruments	8	94.33	3.87	66.00-122.00	4.15	8	39.22	1.72	27.00-51.00	4.36	8	341.22	10.51	239.00-443.00	3.08	8	170.22	4.94	119.00-221.00	2.88	8	230.11	7.15	161.00-299.00	3.09
SG: International OCD Vitros Instruments	13	98.06	10.42	69.00-127.00	10.61	13	40.06	4.61	28.00-52.00	11.50	13	158.29	20.15	111.00-205.00	12.72	13	128.06	15.59	90.00-166.00	12.19	13	148.18	16.23	104.00-192.00	10.95
MG: Method Group	144	84.00	10.70	59.00-109.00	12.74	102	37.00	4.20	26.00-48.00	11.35	96	301.00	58.40	211.00-391.00	19.40	102	153.00	19.30	107.00-199.00	12.61	102	202.00	32.30	141.00-263.00	15.99
MG: Method Group	144	84.00	10.70	59.00-109.00	12.74	102	37.00	4.20	26.00-48.00	11.35	96	301.00	58.40	211.00-391.00	19.40	102	153.00	19.30	107.00-199.00	12.61	102	202.00	32.30	141.00-263.00	15.99
ALT/SGPT	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	8	30.44	1.94	24.00-36.00	6.33	8	19.13	3.23	15.00-23.00	16.84	8	224.00	6.68	179.00-269.00	2.99	8	100.67	1.87	81.00-121.00	1.88	8	141.89	3.14	114.00-170.00	2.18
PG: Abaxis Piccolo (Waived)	42	29.55	3.14	24.00-36.00	10.33																				
PG: Abbott Architect	6	26.29	0.95	21.00-31.00	3.85	6	15.00	1.16	12.00-18.00	8.00	6	265.57	5.53	213.00-319.00	2.07	6	114.71	3.30	92.00-138.00	2.87	6	163.86	3.34	131.00-197.00	2.01
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	32	19.34	2.40	15.00-23.00	12.63	32	11.51	1.69	10.00-14.00	14.17	32	213.26	4.31	170.00-256.00	2.02	32	90.11	9.67	72.00-108.00	10.78	32	133.09	4.04	106.00-160.00	3.01
PG: Clinical Data Envoy 500	6	30.40	1.65	24.00-36.00	5.33	6	16.37	0.48	13.00-19.00	3.13	6	252.44	8.66	202.00-302.00	3.45	6	115.11	3.37	92.00-138.00	2.96	6	165.10	5.11	132.00-198.00	3.09
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	8	33.75	2.31	27.00-41.00	6.76	8	18.88	1.73	15.00-23.00	8.95	8	265.63	6.63	213.00-319.00	2.48	8	118.88	2.90	95.00-143.00	2.44	8	167.25	4.23	134.00-200.00	2.51
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL / IFCC	10	32.91	1.81	26.00-40.00	5.45	10	18.00	1.34	14.00-22.00	7.22	10	266.73	5.04	214.00-320.00	1.87	10	118.36	2.42	94.00-142.00	2.03	10	166.36	3.14	133.00-199.00	1.87
PG: Beckman (Olympus) AU	17	28.17	1.50	22.00-34.00	5.36	16	13.94	0.83	11.00-17.00	5.71	16	232.53	10.12	186.00-280.00	4.33	16	100.88	3.10	81.00-121.00	3.07	16	143.24	4.25	114.00-172.00	3.01
PG: Roche Cobas Integra / Roche Reagent	6	24.17	0.75	19.00-29.00	3.33	6	15.00	1.10	12.00-18.00	7.33	6	245.83	13.18	197.00-295.00	5.37	6	107.17	4.17	86.00-128.00	3.93	6	154.00	7.43	123.00-185.00	4.81
PG: OCD Vitros 250,350,400,500,750,950	9	33.45	1.70	26.00-40.00	5.15	9	31.73	4.29	26.00-38.00	13.44	9	241.73	4.84	194.00-290.00	1.98	9	118.91	3.67	95.00-143.00	3.11	9	162.09	3.94	130.00-194.00	2.41
SG: Select Beckman Coulter Instruments (Chem)	19	29.00	2.94	23.00-35.00	10.00	18	14.74	2.51	12.00-18.00	16.67	18	234.63	11.43	188.00-282.00	4.85	18	102.32	5.20	82.00-122.00	5.10	18	145.00	6.64	116.00-174.00	4.55
SG: Select Roche Instruments / Roche Reagents Groups	7	24.14	0.69	19.00-29.00	2.92	7	14.86	1.07	12.00-18.00	7.33	7	241.86	15.98	194.00-290.00	6.61	7	105.71	5.41	85.00-127.00	5.09	7	152.00	8.60	122.00-182.00	5.66
SG: Select OCD Vitros Instruments	12	33.00	4.41	26.00-40.00	13.33	12	30.67	4.56	25.00-37.00	14.84	12	241.60	4.12	194.00-290.00	1.69	12	121.53	11.64	98.00-146.00	9.51	12	158.80	11.03	127.00-191.00	6.92
SG: Select Abbott Architect Instruments	8	26.11	0.93	21.00-31.00	3.46	8	15.00	1.12	12.00-18.00	7.33	8	265.44	5.03	212.00-318.00	1.89	8	114.78	2.91	92.00-138.00	2.52	8	164.00	2.92	131.00-197.00	1.77
SG: International OCD Vitros Instruments	13	33.94	3.17	27.00-41.00	9.41	13	31.94	5.01	26.00-38.00	15.63	13	242.12	5.61	194.00-290.00	2.31	13	119.94	3.94	96.00-144.00	3.25	13	162.82	4.28	130.00-196.00	2.64
SG: International Abaxis Piccolo Instruments	52	29.77	2.97	24.00-36.00	10.00	10	19.80	3.39	16.00-24.00	17.00	10	224.80	6.29	180.00-270.00	2.80	10	101.09	2.12	81.00-121.00	2.08	10	141.91	3.11	114.00-170.00	2.18
SG: Select Gen Chem Instruments	35	28.78	2.76	23.00-35.00	9.66	34	15.32	2.00	12.00-18.00	13.33	34	247.69	16.62	198.00-298.00	6.69	34	109.10	7.98	87.00-131.00	7.34	34	155.48	11.81	124.00-186.00	7.61
MG: Method Group	116	29.00	3.70	23.00-35.00	12.76	80	14.00	2.50	11.00-17.00	17.86	101	235.00	20.30	188.00-282.00	8.64	101	105.00	14.00	84.00-126.00	13.33	101	149.00	14.40	119.00-179.00	9.66
MG: IFCC	11	33.00	1.80	26.00-40.00	5.45	11	18.00	1.30	14.00-22.00	7.22	11	269.00	8.70	215.00-323.00	3.23	11	119.00	3.60	95.00-143.00	3.03	11	168.00	5.90	134.00-202.00	3.51
AST/SGOT	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Non-Waived)	8	40.00	1.80	32.00-48.00	4.50	8	106.13	5.41	85.00-127.00	5.09	8	265.63	14.34	213.00-319.00	5.38	8	172.78	6.57	138.00-208.00	3.82	8	202.11	8.04	162.00-242.00	3.96
PG: Abaxis Piccolo (Waived)	42	39.57	1.68	32.00-48.00	4.25																				
PG: Abbott Architect	6	36.71	0.95	30.00-44.00	2.70	6	104.57	2.64	84.00-126.00	2.48	6	272.86	6.44	218.00-328.00	2.34	6	173.14	3.24	138.00-208.00	1.85	6	206.57	3.78	166.00-248.00	1.84
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	32	29.71	2.11	24.00-36.00	7.00	32	95.17	5.15	76.00-114.00	5.47	32	244.34	9.22	195.00-293.00	3.77	32	157.29	5.46	126.00-188.00	3.50	32	187.40	5.89	150.00-224.00	3.16
PG: Clinical Data Envoy 500	6	36.29	1.87	29.00-43.00	5.28	6	106.99	2.18	86.00-128.00	2.06	6	270.74	4.98	217.00-325.00	1.85	6	173.89	3.69	139.00-209.00	2.13	6	207.98	3.94	166.00-250.00	1.88
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	18	31.95	1.90	26.00-38.00	5.94	18	105.37	3.44	84.00-126.00	3.24	18	298.21	8.34	238.00-358.00	2.79	18	184.37	5.17	147.00-221.00	2.83	18	223.05	6.67	178.00-268.00	3.00
PG: Beckman (Olympus) AU	16	33.35	1.17	26.00-40.00	3.64	16	93.35	3.41	74.00-112.00	3.66	16	242.41	9.76	194.00-290.00	4.05	16	153.82	5.76	123.00-185.00	3.77	16	183.59	7.12	147.00-221.00	3.86
PG: Roche Cobas Integra / Roche Reagent	6	31.67	1.37	26.00-38.00	4.38	6	103.17	3.97	82.00-124.00	3.88	6	264.83	11.96	212.00-318.00	4.53										
PG: OCD Vitros 250,350,400,500,750,950	9	37.91	1.22	30.00-46.00	3.16	9	97.82	2.96	78.00-118.00	3.06	9	324.55	13.10	260.00-390.00	4.03	9	178.55	5.48	143.00-215.00	3.07	9	226.18	8.86	181.00-271.00	3.94
SG: Select Beckman Coulter Instruments (Chem)	18	34.26	2.96	27.00-41.00	8.82	18	94.79	5.38	76.00-114.00	5.68	18	245.05	12.19	196.00-294.00	4.98	18	155.74	7.90	125.00-187.00	5.06	18	185.79	9.43	149.00-223.00	5.05
SG: Select Roche Instruments / Roche Reagents Groups	7	32.14	1.77	26.00-38.00	5.63	7	101.57	5.56	82.00-122.00	5.49	7	261.00	14.90	209.00-313.00	5.71	7	167.43	9.07	134.00-200.00	5.45	7	200.14	11.35	160.00-240.00	5.65
SG: Select OCD Vitros Instruments	12	38.33	1.50	30.00-46.00	3.95	12	98.33	2.79	78.00-118.00	2.86	12	323.93	11.36	259.00-389.00	3.52	12	182.13	12.07	146.00-218.00	6.65	12	222.60	15.16	178.00-268.00	6.82
SG: Select Abbott Architect Instruments	8	36.44	1.24	29.00-43.00	3.33	8</																			



PG: OCD Vitros 250,350,400,500,750,950																								
SG: Select Beckman Coulter Instruments (Chem)																								
SG: Select Roche Instruments / Roche Reagents Groups																								
SG: Select Abbott Architect Instruments																								
SG: International OCD Vitros Instruments																								
SG: Select Gen Chem Instruments																								
AG: All Group																								
Phosphorous					Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent																								
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL																								
PG: Beckman (Olympus) AU																								
SG: Select Beckman Coulter Instruments (Chem)																								
SG: Select Gen Chem Instruments																								
AG: All Group																								
Cholesterol Total					Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Waived)																								
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent																								
PG: Clinical Data Envoy 500																								
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL																								
PG: Beckman (Olympus) AU																								
PG: Roche Cobas Integra / Roche Reagent																								
PG: OCD Vitros 250,350,400,500,750,950																								
SG: Select Beckman Coulter Instruments (Chem)																								
SG: Select Roche Instruments / Roche Reagents Groups																								
SG: Select Abbott Architect Instruments																								
SG: International OCD Vitros Instruments																								
AG: All Group																								
Cholesterol HDL					Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Waived) / Direct Method																								
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Direct Method / Alfa Wasserman Reagent																								
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL / Direct Method																								
PG: Beckman (Olympus) AU / Direct Method																								
PG: Roche Cobas Integra / Direct Method / Roche Reagent																								
PG: OCD Vitros 250,350,400,500,750,950 / Direct Method																								
SG: Select Beckman Coulter Instruments (Chem) / Direct Method																								
SG: Select Roche Instruments / Direct Method / Roche Reagents Groups																								
SG: Select Abbott Architect Instruments / Direct Method																								
SG: International OCD Vitros Instruments / Direct Method																								
MG: Direct Method																								
MG: Direct Method																								
AG: All Group																								
Cholesterol LDL					Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Waived) / Calc																								
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Calc / Alfa Wasserman Reagent																								
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL / Calc																								
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL / Direct Method																								
PG: Beckman (Olympus) AU / Calc																								
SG: Select Beckman Coulter Instruments (Chem) / Calc																								
SG: Select Roche Instruments (Chem) / Direct Method																								
SG: Select Abbott Architect Instruments / Calc																								
SG: International OCD Vitros Instruments / Calc																								
MG: Calc																								
MG: Direct Method																								
Triglycerides					Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Waived)																								
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent																								
PG: Clinical Data Envoy 500																								
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL																								
PG: Beckman (Olympus) AU																								
PG: Roche Cobas Integra / Roche Reagent																								
PG: OCD Vitros 250,350,400,500,750,950																								
SG: Select Beckman Coulter Instruments (Chem)																								
SG: Select Roche Instruments / Roche Reagents Groups																								



Test Name	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
SG: Select OCD Vitros Instruments	12	186.67	6.82	140.00-234.00	3.64	12	101.73	3.75	77.00-128.00	3.73	12	282.87	10.99	212.00-354.00	3.89	12	177.20	8.59	133.00-221.00	4.86	12	208.53	14.81	157.00-261.00	7.08
SG: Select Abbott Architect Instruments	6	138.17	1.94	104.00-173.00	1.38	6	91.50	1.52	69.00-115.00	1.63	6	241.50	6.89	182.00-303.00	2.85	6	153.67	4.41	116.00-193.00	2.86	6	182.83	4.92	137.00-229.00	2.68
SG: International OCD Vitros Instruments	13	189.06	5.02	142.00-236.00	2.65	13	102.00	2.95	77.00-128.00	2.94	13	287.24	9.89	215.00-359.00	3.45	13	177.24	5.20	133.00-221.00	2.94	13	214.00	6.58	161.00-268.00	3.08
SG: Select Gen Chem Instruments	31	145.14	10.07	109.00-181.00	6.97	31	97.58	4.43	74.00-123.00	4.49	31	252.59	9.42	190.00-316.00	3.72	31	162.09	6.89	122.00-203.00	4.26	31	192.60	8.70	145.00-241.00	4.51
AG: All Group	102	148.14	18.08	111.00-185.00	12.23	97	98.94	6.47	74.00-124.00	6.57	97	255.57	16.67	192.00-320.00	6.52	97	164.17	10.37	123.00-205.00	6.34	97	194.97	12.26	146.00-244.00	6.31
Amylase	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Abaxis Piccolo (Waived)	31	68.35	1.40	48.00-88.00	2.06																				
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	9	76.55	2.88	54.00-100.00	3.77	9	15.36	1.03	11.00-20.00	6.67	8	191.30	8.39	134.00-248.00	4.40	9	86.64	4.18	61.00-113.00	4.83	9	122.64	5.09	86.00-160.00	4.15
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	8	78.50	1.60	55.00-103.00	2.03	8	14.75	1.98	11.00-20.00	13.33	8	207.25	1.83	145.00-269.00	0.87	8	93.00	1.51	65.00-121.00	1.61	8	131.38	2.20	92.00-170.00	1.68
PG: Beckman (Olympus) AU	8	60.22	3.23	42.00-78.00	5.33	8	11.89	0.33	8.00-16.00	2.50	8	153.89	3.14	108.00-200.00	2.01	8	68.89	1.76	48.00-90.00	2.61	8	97.44	1.94	68.00-126.00	1.96
SG: Select Beckman Coulter Instruments (Chem)	9	62.70	8.41	44.00-82.00	13.33	8	11.89	0.33	8.00-16.00	2.50	9	158.00	13.33	111.00-205.00	8.42	9	70.80	6.27	50.00-92.00	8.87	9	100.10	8.60	70.00-130.00	8.60
SG: Select OCD Vitros Instruments	5	85.50	9.33	60.00-112.00	10.81	5	30.00	0.00	21.00-39.00	0.00	5	118.50	4.04	83.00-155.00	3.36	5	48.25	4.50	34.00-62.00	9.38	5	74.50	5.20	53.00-98.00	6.93
SG: Abaxis Piccolo Instruments	33	68.42	1.39	48.00-88.00	2.06																				
SG: Select Chemistry Instr - Enzyme 2 (AMY) / Alfa Wasserman Reagent	9	76.55	2.88	54.00-100.00	3.77																				
SG: Select Chemistry Instr - Enzyme 3 (AST, GGT) / Alfa Wasserman Reagent	9	76.55	2.88	54.00-100.00	3.77																				
SG: Select Electrolyte Instruments 1 / Alfa Wasserman Reagent	9	76.55	2.88	54.00-100.00	3.77																				
SG: International OCD Vitros Instruments	5	84.33	5.79	59.00-109.00	6.90	5	30.00	0.00	21.00-39.00	0.00	5	122.33	4.72	85.00-159.00	3.85	5	52.33	2.66	36.00-68.00	5.19	5	77.17	2.32	54.00-100.00	2.99
SG: International Abaxis Piccolo Instruments	35	68.51	1.42	48.00-90.00	2.03	5	15.00	0.82	11.00-20.00	5.33	5	168.00	4.40	118.00-218.00	2.62	5	76.50	2.38	54.00-100.00	3.12	5	107.75	3.20	76.00-140.00	2.96
SG: Select Gen Chem Instruments	15	71.13	13.22	50.00-92.00	18.59	14	13.85	2.64	10.00-18.00	18.57	15	172.41	22.27	120.00-224.00	12.97	15	77.38	10.14	54.00-100.00	13.12	15	108.78	13.61	76.00-142.00	12.48
SG: Select Gen Chem Instruments / Alfa Wasserman Reagent	9	76.55	2.88	54.00-100.00	3.77																				
SG: Select Cardiac Markers II Insts / Alfa Wasserman Reagent	9	76.55	2.88	54.00-100.00	3.77																				
AG: All Group	74	73.01	8.51	51.00-95.00	11.64	42	16.76	5.55	12.00-22.00	32.94	43	172.86	31.88	121.00-225.00	18.44	43	78.32	14.52	55.00-101.00	18.59	43	111.05	19.12	78.00-144.00	17.21
Creatine Kinase	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Abaxis Piccolo (Waived)	22	183.18	10.55	128.00-238.00	5.74																				
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	9	152.92	6.26	107.00-199.00	4.12	9	27.17	1.64	19.00-35.00	5.93	9	276.25	36.99	193.00-359.00	13.41	9	137.00	11.08	96.00-178.00	8.10	9	183.58	22.38	129.00-239.00	12.17
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL / IFCC	5	162.80	12.03	114.00-212.00	7.36	5	20.40	3.71	14.00-26.00	18.50	5	298.60	32.50	209.00-389.00	10.87	5	140.80	12.60	99.00-183.00	8.94	5	195.60	19.15	137.00-255.00	9.80
PG: Beckman (Olympus) AU	7	145.25	3.65	102.00-189.00	2.55	7	18.75	0.71	13.00-25.00	3.68	7	291.13	15.63	204.00-378.00	5.36	7	131.88	5.87	92.00-172.00	4.47	7	185.75	8.33	130.00-242.00	4.46
SG: Select Beckman Coulter Instruments (Chem)	8	147.44	7.42	103.00-191.00	5.03	8	19.11	1.27	13.00-25.00	6.84	8	295.56	19.76	207.00-385.00	6.69	8	134.00	8.41	94.00-174.00	6.27	8	188.67	11.72	132.00-246.00	6.19
SG: Select OCD Vitros Instruments	5	173.20	12.22	121.00-225.00	7.05	5	20.25	0.50	14.00-26.00	2.50	5	197.80	22.36	139.00-257.00	11.31	5	128.00	19.95	90.00-166.00	15.55	5	152.40	25.20	106.00-198.00	16.58
SG: Abaxis Piccolo Instruments	26	182.77	9.97	128.00-238.00	5.46																				
SG: International OCD Vitros Instruments	6	166.13	8.53	116.00-216.00	5.12	6	21.25	2.05	15.00-27.00	10.00	6	185.00	20.51	130.00-241.00	11.08	6	110.43	11.80	77.00-143.00	10.73	6	147.43	18.46	103.00-191.00	12.59
SG: International Abaxis Piccolo Instruments	29	183.90	10.63	129.00-239.00	5.76																				
AG: All Group	65	166.38	17.09	116.00-216.00	10.30	38	21.56	4.20	15.00-29.00	19.09	41	283.06	40.60	198.00-368.00	14.35	43	136.68	18.24	96.00-178.00	13.28	43	185.68	26.96	130.00-242.00	14.52
GGT	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Abaxis Piccolo (Waived)	31	33.19	1.14	30.00-36.00	3.33																				
SG: Select Beckman Coulter Instruments (Chem)	5	29.80	1.30	26.00-34.00	4.33	5	13.80	2.49	7.00-22.00	17.86	5	183.60	23.78	113.00-255.00	12.93	5	82.20	11.35	48.00-116.00	13.78	5	116.60	15.85	69.00-165.00	13.59
SG: Abaxis Piccolo Instruments	33	33.03	1.43	29.00-37.00	4.24																				
AG: All Group	49	34.18	5.27	18.00-50.00	15.59	18	17.00	4.67	3.00-31.00	27.65	18	192.45	35.79	85.00-299.00	18.65	18	89.20	17.10	38.00-140.00	19.21	18	123.55	23.08	55.00-193.00	18.63
LDH	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Intl OCD Vitros 250,350,400,500,750,950	5	462.20	20.85	370.00-554.00	4.50	5	477.80	9.09	382.00-574.00	1.90	5	2193.25	123.70	1754.00-2632.00	5.64	5	1227.25	22.41	982.00-1472.00	1.83	5	1631.50	35.67	1306.00-1958.00	2.19
SG: Select Gen Chem Instruments	7	164.75	23.20	132.00-198.00	14.06	7	149.00	19.20	119.00-179.00	12.89	7	714.00	77.59	571.00-857.00	10.87	7	381.00	41.89	305.00-457.00	11.00	7	490.50	52.52	393.00-589.00	10.69
AG: All Group	10	150.91	23.21	121.00-181.00	15.36	11	139.83	22.03	112.00-168.00	15.71	11	672.83	101.25	538.00-808.00	15.05	10	349.73	52.66	280.00-420.00	15.06	11	458.33	73.38	366.00-550.00	16.03
Lipase	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %	Labs	Mean	SD	Range	CV %
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	6	208.50	9.91	179.00-239.00	4.74	6	63.50	6.22	45.00-83.00	9.69	6	235.67	10.44	205.00-267.00	4.41	6	135.50	7.23	114.00-158.00	5.29	6	168.67	8.04	145.00-193.00	4.73
PG: Beckman (Olympus) AU	7	45.50	3.89	34.00-58.00	8.48	7	14.38	0.92	11.00-17.00	6.43	7	70.00	4.87	55.00-85.00	7.00	7	36.13	1.73	31.00-41.00	4.72	7	47.63	2.97	39.00-57.00	6.25
SG: Select Beckman Coulter Instruments (Chem)	9	42.70	6.87	22.00-64.00	16.05	9	15.00	1.56	10.00-20.00	10.67	9	65.20	11.03	32.00-98.00	16.92	9	35.10	2.69	27.00-43.00	7.71	9	45.40	5.38	29.00-61.00	12.00
PG: Intl OCD Vitros 250,350,400,500,750,950	5	192.00	0.00	192.00-192.00	0.00	5	154.00	0.00	154.00-154.00	0.00	5	587.00	0.00	587.00-587.00	0.00	5	349.00	0.00	349.00-349.00	0.00	5	437.00	0.00	437.00-437.00	0.00
SG: International OCD Vitros Instruments	5	187.00	6.63	167.00-207.00	3.53	5	152.25	2.87	143.00-161.00	1.91	5	858.50													



PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	7	1.31	0.07	1.00-1.60	5.38	7	1.03	0.08	0.80-1.30	8.00	7	4.43	0.10	3.30-5.50	2.27	7	2.40	0.08	1.80-3.00	3.33	7	3.10	0.08	2.30-3.90	2.58
PG: Beckman (Olympus) AU	7	1.30	0.06	1.00-1.60	4.62	7	1.03	0.05	0.80-1.30	5.00	7	4.37	0.16	3.30-5.50	3.64	7	2.34	0.11	1.70-2.90	4.78	7	3.00	0.12	2.30-3.80	4.00
SG: Select Beckman Coulter Instruments (Chem)	9	1.30	0.05	1.00-1.60	3.85	9	1.00	0.07	0.80-1.30	7.00	9	4.31	0.18	3.20-5.40	4.19	9	2.32	0.11	1.70-2.90	4.78	9	2.98	0.11	2.30-3.80	3.67
SG: International Abaxis Piccolo Instruments	9	1.28	0.04	1.00-1.60	3.08	9	1.07	0.04	0.80-1.40	3.64	9	4.52	0.09	3.40-5.60	2.00	9	2.32	0.06	1.70-2.90	2.61	9	2.96	0.05	2.30-3.80	1.67
SG: Select Gen Chem Instruments	16	1.42	0.19	1.00-1.80	13.57	16	1.00	0.09	0.80-1.30	9.00	16	4.33	0.23	3.20-5.40	5.35	16	2.34	0.14	1.70-2.90	6.09	16	2.97	0.14	2.30-3.80	4.67
AG: All Group	44	1.35	0.17	1.00-1.60	13.08	43	1.00	0.10	0.80-1.30	10.00	44	4.35	0.23	3.20-5.40	5.35	44	2.38	0.18	1.80-3.00	7.50	44	3.00	0.18	2.30-3.80	6.00
Uric Acid	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abaxis Piccolo (Waived)	8	1.03	0.06	0.80-1.20	6.00																				
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	14	2.96	0.18	2.50-3.50	6.00	13	2.80	0.25	2.30-3.30	8.57	14	10.94	0.52	9.00-12.80	4.77	14	5.81	0.30	4.80-6.80	5.17	14	7.44	0.28	6.10-8.70	3.78
PG: Clinical Data Envoy 500	5	1.94	0.16	1.60-2.20	8.42	5	1.88	0.14	1.60-2.20	7.37	5	12.54	1.16	10.40-14.60	9.28	5	6.18	0.47	5.10-7.30	7.58	5	8.29	0.68	6.90-9.70	8.19
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	14	1.86	0.13	1.60-2.20	6.84	14	1.73	0.11	1.40-2.00	6.47	14	10.87	0.58	9.00-12.80	5.32	14	5.36	0.23	4.50-6.30	4.26	14	7.16	0.32	6.00-8.40	4.44
PG: Beckman (Olympus) AU	9	1.59	0.06	1.30-1.90	3.75	9	1.79	0.06	1.50-2.10	3.33	9	11.37	0.23	9.50-13.30	2.02	9	5.62	0.14	4.60-6.60	2.50	9	7.51	0.14	6.20-8.80	1.87
PG: OCD Vitros 250,350,400,500,750,950	7	1.85	0.08	1.60-2.20	4.21	7	1.66	0.07	1.40-2.00	4.12	7	11.06	0.30	9.20-13.00	2.70	7	5.49	0.13	4.60-6.40	2.18	7	7.29	0.19	6.10-8.50	2.60
SG: Select Beckman Coulter Instruments (Chem)	11	1.58	0.06	1.30-1.90	3.75	11	1.78	0.06	1.50-2.10	3.33	11	11.23	0.40	9.30-13.10	3.57	11	5.58	0.17	4.60-6.60	3.04	11	7.43	0.22	6.10-8.70	2.97
SG: Select OCD Vitros Instruments	9	1.86	0.07	1.60-2.20	3.68	9	1.67	0.07	1.40-2.00	4.12	9	11.10	0.28	9.20-13.00	2.52	9	5.66	0.52	4.70-6.70	9.12	9	7.10	0.62	5.90-8.30	8.73
SG: Select SpotChem Instruments	9	1.82	0.09	1.50-2.10	5.00	9	1.68	0.07	1.40-2.00	4.12	9	11.00	0.34	9.10-12.90	3.09	9	5.50	0.11	4.60-6.40	2.00	9	7.33	0.19	6.10-8.50	2.60
SG: Select Abbott Architect Instruments	5	1.82	0.04	1.50-2.10	2.22	5	1.86	0.06	1.60-2.20	2.63	5	11.96	0.25	10.00-14.00	2.08	5	5.92	0.11	4.90-6.90	1.86	5	7.92	0.16	6.60-9.20	2.03
SG: International OCD Vitros Instruments	10	1.86	0.07	1.60-2.20	3.68	10	1.69	0.08	1.40-2.00	4.71	10	11.08	0.25	9.20-13.00	2.25	10	5.51	0.12	4.60-6.40	2.18	10	7.30	0.16	6.10-8.50	2.19
SG: Select Gen Chem Instruments	23	1.74	0.20	1.40-2.00	11.76	23	1.81	0.11	1.50-2.10	6.11	23	11.72	0.91	9.70-13.70	7.78	23	5.80	0.40	4.80-6.80	6.72	23	7.76	0.56	6.50-9.10	7.18
AG: All Group	60	1.69	0.29	1.40-2.00	17.06	58	1.86	0.31	1.60-2.20	16.32	66	11.24	0.77	9.30-13.10	6.88	66	5.68	0.39	4.70-6.70	6.84	66	7.45	0.52	6.20-8.80	6.93
Iron	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	6	143.56	8.69	115.00-173.00	6.04	6	28.11	1.36	22.00-34.00	5.00	6	194.22	4.49	155.00-233.00	2.32	6	96.78	3.42	78.00-116.00	3.51	6	129.78	2.33	104.00-156.00	1.77
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	6	149.83	1.72	120.00-180.00	1.13	6	31.50	0.84	26.00-38.00	2.50	6	211.17	2.23	169.00-253.00	1.04	6	104.00	1.10	83.00-125.00	1.06	6	139.67	1.03	112.00-168.00	0.71
PG: Beckman (Olympus) AU	7	161.13	3.00	129.00-193.00	1.86	7	35.00	2.73	28.00-42.00	7.71	7	240.38	4.31	192.00-288.00	1.79	7	115.88	2.75	93.00-139.00	2.33	7	161.25	12.02	129.00-193.00	7.45
SG: Select Beckman Coulter Instruments (Chem)	8	160.11	4.14	128.00-192.00	2.56	8	34.56	2.88	28.00-42.00	8.29	8	237.67	9.07	190.00-286.00	3.82	8	114.78	4.18	92.00-138.00	3.65	8	159.11	12.95	127.00-191.00	8.11
SG: Select OCD Vitros Instruments	5	190.00	5.66	152.00-228.00	3.00	5	31.00	9.90	25.00-37.00	31.94	5	257.50	16.26	206.00-310.00	6.32	5	147.00	41.01	118.00-176.00	27.89	5	143.50	38.89	115.00-173.00	27.01
SG: Select Abbott Architect Instruments	5	164.20	2.59	131.00-197.00	1.59	5	33.00	1.58	26.00-40.00	4.85	5	220.20	0.45	176.00-264.00	0.18	5	108.60	1.34	87.00-131.00	1.19	5	145.20	2.28	116.00-174.00	1.59
PG: Intl OCD Vitros 250,350,400,500,750,950	5	177.20	5.02	142.00-212.00	2.82	5	19.00	4.18	15.00-23.00	22.11	5	251.00	3.83	201.00-301.00	1.51	5	114.25	2.22	91.00-137.00	1.93	5	158.00	3.37	126.00-190.00	2.15
AG: All Group	31	153.46	9.48	122.00-184.00	6.21	31	32.04	3.31	26.00-38.00	10.31	31	214.40	16.68	171.00-257.00	7.80	31	105.84	7.29	85.00-127.00	6.89	31	142.41	12.60	114.00-170.00	8.87
TIBC	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Alfa Was. Ace/NexCT/Alera/Axcel / Alfa Wasserman Reagent	5	504.50	9.47	476.10-532.90	1.88	5	149.50	11.73	114.30-184.70	7.85	5	419.75	7.80	396.40-443.20	1.86	5	273.75	7.46	251.40-296.20	2.72	5	326.25	3.30	316.40-336.20	1.01
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	5	481.60	8.41	456.40-506.80	1.75	5	78.80	11.39	44.60-113.00	14.45	5	389.80	6.80	369.40-410.20	1.74	5	203.60	9.42	175.30-231.90	4.63	5	266.20	10.73	234.00-298.40	4.03
PG: OCD Vitros 250,350,400,500,750,950	5	586.00	0.00	586.00-586.00	0.00	5	120.00	0.00	120.00-120.00	0.00	5	727.00	0.00	727.00-727.00	0.00	5	325.00	0.00	325.00-325.00	0.00	5	445.00	0.00	445.00-445.00	0.00
PG: Intl OCD Vitros 250,350,400,500,750,950	5	743.40	122.77	375.10-1111.70	16.51	5	135.20	10.90	102.50-167.90	8.05	5	691.00	95.91	403.30-978.70	13.88	5	320.25	12.26	283.50-357.10	3.83	5	487.25	23.41	417.10-557.50	4.80
PG: OCD Vitros 5600	5	452.00	0.00	452.00-452.00	0.00	5	133.00	0.00	133.00-133.00	0.00	5	343.00	0.00	343.00-343.00	0.00	5	273.00	0.00	273.00-273.00	0.00	5	236.00	0.00	236.00-236.00	0.00
AG: All Group	15	504.75	23.17	435.30-574.30	4.59	14	96.40	23.93	24.60-168.20	24.82	15	422.30	89.36	154.20-690.40	21.16	15	234.81	48.33	89.80-379.80	20.58	15	299.59	60.98	116.70-482.50	20.35
UIBC	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	6	369.29	15.90	321.60-417.00	4.31	6	75.71	3.82	64.20-87.20	5.05	6	295.00	9.16	267.50-322.50	3.11	6	164.00	6.58	144.30-183.70	4.01	6	209.29	8.38	184.20-234.40	4.00
746 - Auto Diff 6 (Abbott 5-Part Diff)																									
White Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Abbott Cell Dyn Instruments (5-part Diff)	7	20.13	0.44	17.10-23.10	2.19	7	3.15	0.19	2.70-3.70	5.94	7	7.47	0.35	6.40-8.60	4.67	7	20.27	0.86	17.30-23.30	4.24	7	3.09	0.04	2.60-3.60	1.29
PG: Abbott Cell-Dyn Ruby	6	20.17	0.47	17.20-23.20	2.33	6	3.20	0.17	2.70-3.70	5.31	6	7.53	0.34	6.40-8.60	4.53	6	20.52	0.62	17.40-23.60	3.02	6	3.08	0.05	2.60-3.60	1.61
AG: All Group	8	20.31	0.66	17.30-23.30	3.25	8	3.14	0.19	2.60-3.60	6.13	8	7.49	0.33	6.40-8.60	4.40	8	20.46	0.96	17.40-23.60	4.73	8	3.08	0.05	2.60-3.60	1.61
Lymphocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Abbott Cell Dyn Instruments (5-part Diff)	7	17.57	1.14	14.20-21.00	6.48	7	40.23	1.82	34.70-45.70	4.53	7	26.33	1.85	20.80-31.90	7.03	7	17.26	1.65	12.40-22.30	9.54	7	39.51	1.26	35.70-43.30	3.19
PG: Abbott Cell-Dyn Ruby	6	17.90	0.81	15.50-20.30	4.53	6	40.32	1.98	34.40-46.20	4.91	6	26.60	1.87	21.0											



SG: Select Abbott Cell Dyn Instruments (5-part Diff)	7	16.91	0.28	15.70-18.10	1.66	7	5.71	0.18	5.30-6.10	3.16	7	13.63	0.29	12.60-14.60	2.13	7	16.74	0.42	15.50-17.90	2.51	7	5.64	0.14	5.20-6.00	2.50
PG: Abbott Cell-Dyn Ruby	6	16.90	0.30	15.70-18.10	1.78	6	5.66	0.14	5.30-6.10	2.46	6	13.60	0.30	12.60-14.60	2.21	6	16.75	0.46	15.60-18.00	2.74	6	5.60	0.09	5.20-6.00	1.61
AG: All Group	8	16.80	0.41	15.60-18.00	2.44	8	5.73	0.18	5.30-6.10	3.16	8	13.59	0.29	12.60-14.60	2.13	8	16.66	0.45	15.50-17.90	2.69	8	5.69	0.18	5.30-6.10	3.16
Hematocrit	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Abbott Cell Dyn Instruments (5-part Diff)	7	44.00	0.78	41.40-46.60	1.77	7	15.51	0.39	14.60-16.40	2.45	7	37.74	0.91	35.40-40.00	2.41	7	43.74	1.33	41.10-46.30	3.04	7	15.43	0.33	14.50-16.30	2.14
PG: Abbott Cell-Dyn Ruby	6	44.15	0.73	41.50-46.90	1.65	6	15.53	0.42	14.60-16.40	2.71	6	37.92	0.85	35.60-40.20	2.24	6	43.93	1.35	41.30-46.50	3.08	6	15.43	0.36	14.50-16.30	2.34
AG: All Group	8	44.99	2.88	42.30-47.70	6.40	8	15.89	1.11	14.90-16.90	6.98	8	38.39	2.01	36.10-40.70	5.23	8	44.68	2.91	42.00-47.40	6.51	8	15.81	1.13	14.90-16.70	7.15
MCV	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Abbott Cell Dyn Instruments (5-part Diff)	7	81.30	1.75	76.00-86.50	2.15	7	71.27	1.06	68.10-74.50	1.49	7	81.77	1.68	76.80-86.80	2.05	7	81.54	1.81	76.10-86.90	2.22	7	71.36	1.24	67.70-75.10	1.74
PG: Abbott Cell-Dyn Ruby	6	81.13	1.86	75.50-86.70	2.29	6	71.22	1.16	67.80-74.70	1.62	6	81.63	1.79	76.20-87.00	2.19	6	81.32	1.88	75.70-86.90	2.31	6	71.25	1.32	67.30-75.30	1.85
AG: All Group	8	83.45	6.29	64.60-102.40	7.53	8	72.79	4.40	59.60-86.00	6.04	8	83.37	4.79	69.00-97.80	5.74	8	83.56	5.95	65.80-101.40	7.12	8	72.85	4.38	59.80-86.00	6.01
RDW	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Abbott Cell Dyn Instruments (5-part Diff)	7	10.21	0.63	8.30-12.10	6.18	7	12.67	0.91	10.00-15.40	7.17	7	11.99	0.61	10.20-13.80	5.08	7	10.16	0.49	8.70-11.70	4.80	7	12.69	0.85	10.10-15.30	6.69
PG: Abbott Cell-Dyn Ruby	6	10.13	0.65	8.10-12.10	6.44	6	12.58	0.96	9.70-15.50	7.62	6	11.92	0.64	10.00-13.80	5.38	6	10.07	0.46	8.70-11.50	4.55	6	12.63	0.92	9.80-15.40	7.30
AG: All Group	8	10.75	1.62	5.90-15.70	15.00	8	13.21	1.75	8.00-18.40	13.18	8	12.44	1.40	8.20-16.60	11.29	8	10.71	1.63	5.80-15.60	15.23	8	13.19	1.62	8.30-18.10	12.27
Platelet Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Abbott Cell Dyn Instruments (5-part Diff)	7	489.57	13.89	368.00-613.00	2.84	7	80.50	4.46	61.00-101.00	5.56	7	257.71	12.00	194.00-323.00	4.65	7	483.14	30.18	362.00-604.00	6.25	7	79.54	3.59	60.00-100.00	4.50
PG: Abbott Cell-Dyn Ruby	6	491.67	13.95	369.00-615.00	2.85	6	80.58	4.88	61.00-101.00	6.05	6	255.50	11.47	192.00-320.00	4.49	6	484.17	32.93	363.00-605.00	6.80	6	78.80	3.29	59.00-99.00	4.18
AG: All Group	8	489.63	12.86	368.00-613.00	2.63	8	82.19	6.31	62.00-103.00	7.68	8	258.63	11.40	194.00-324.00	4.40	8	481.25	28.45	361.00-601.00	5.93	8	80.97	5.24	61.00-101.00	6.42
683 - Antistreptolysin O (ASO);684 - C-Reactive Protein	CR-1N					CR-2N																			
CRP Quantitative	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	11	1.52	0.33	0.53-2.51	21.78	11	6.24	0.43	4.94-7.54	6.94															
600 - Cholestech LDX/CardioChek Pkg	CT-1																								
CholesterolTotal (Waived)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Cholestech LDX (Waived)	92	252.48	26.13	174.00-330.00	10.36																				
PG: Polymer CardioChek	76	254.26	44.48	121.00-388.00	17.52																				
Glucose (Waived)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Cholestech LDX (Waived)	18	145.83	7.00	125.00-167.00	4.79																				
AG: All Group	19	146.79	7.98	123.00-171.00	5.44																				
Triglycerides (Waived)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Cholestech LDX (Waived)	80	146.45	12.43	109.00-183.00	8.49																				
PG: Polymer CardioChek	73	153.45	22.84	85.00-221.00	14.90																				
SG: Select Chemistry Instr - Enzyme 3 (AST, GGT)	80	146.45	12.43	109.00-183.00	8.49																				
SG: Select Cholestech Instruments	80	146.45	12.43	109.00-183.00	8.49																				
Cholesterol HDL (Waived)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Cholestech LDX (Waived)	80	49.88	7.74	27.00-73.00	15.40																				
PG: Polymer CardioChek / Direct Method	70	106.85	15.00	62.00-152.00	14.02																				
Cholesterol LDL (Waived)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Cholestech LDX (Waived)	51	170.60	32.33	74.00-268.00	18.89																				
PG: Polymer CardioChek / Direct Method	10	129.36	41.01	6.00-252.00	31.78																				
763 - CoaguChek XS Plus/XS Pro (2 specimens);764 - CoaguChek XS Plus/XS Pro (5 specimens)	CXP-1					CXP-2					CXP-3					CXP-4					CXP-5				
Protime - Whole Blood	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: CoaguChek XS Plus	18	27.70	1.22	23.50-31.90	4.40	18	12.85	0.48	11.00-14.80	3.72	7	12.70	0.34	10.80-14.60	2.68	7	26.44	1.69	22.40-30.40	6.40	7	12.49	0.60	10.60-14.40	4.80
PG: CoaguChek XS Pro	10	28.03	0.45	23.80-32.20	1.61	10	13.02	0.17	11.10-14.90	1.38	10	12.98	0.23	11.10-14.90	1.77	10	27.42	1.31	23.30-31.50	4.78	10	12.99	0.19	11.10-14.90	1.46
Protime W Blood - Waived	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: CoaguChek XS Plus	20	27.89	1.45	23.50-32.30	5.20	20	13.07	0.44	11.80-14.40	3.36															
INR	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: CoaguChek XS Plus	112	2.30	0.21	1.70-2.90	9.13	112	1.12	0.17	0.60-1.60	15.45	48	1.09	0.02	1.00-1.20	1.82	48	2.29	0.13	1.90-2.70	5.65	48	1.08	0.04	1.00-1.20	3.64
PG: CoaguChek XS Pro	12	2.33	0.07	2.10-2.50	3.04	12	1.10	0.00	1.10-1.10	0.00	10	1.10	0.00	1.10-1.10	0.00	10	2.28	0.11	2.00-2.60	4.78	10	1.10	0.00	1.10-1.10	0.00
745 - Auto Diff 5 (Beckman Coulter 5-Part Diff)	DIF-1					DIF-2					DIF-3					DIF-4					DIF-5				
White Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Beckman Coulter Instruments (5-part Diff)	5	21.90	0.34	18.60-25.20	1.55	5	4.55	0.21	3.90-5.30	4.57	5	9.80	0.20	8.30-11.30	2.04	5	21.60	0.60	18.40-24.80	2.78	5	4.63	0.26	3.90-5.30	5.65
Lymphocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range						



Hematocrit	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Beckman Coulter Instruments (5-part Diff)	5	52.75	0.53	49.60-56.00	1.00	5	19.42	0.24	18.20-20.60	1.24	5	37.38	0.83	35.20-39.60	2.22	5	52.22	0.90	49.10-55.30	1.72	5	19.20	0.22	18.00-20.40	1.15
Platelet Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Beckman Coulter Instruments (5-part Diff)	5	510.00	33.85	383.00-638.00	6.65	5	78.75	2.06	59.00-99.00	2.66	5	237.00	7.07	178.00-296.00	3.00	5	495.75	25.20	372.00-620.00	5.08	5	78.00	2.00	59.00-98.00	2.56
MCV	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Beckman Coulter Instruments (5-part Diff)	5	96.00	0.75	93.80-98.30	0.78	5	74.57	0.51	73.10-76.10	0.68	5	89.90	0.39	88.70-91.10	0.43	5	95.90	0.47	94.50-97.30	0.49	5	74.50	0.58	72.80-76.20	0.78
RDW	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
SG: Select Beckman Coulter Instruments (5-part Diff)	5	13.00	0.14	12.60-13.40	1.08	5	16.33	0.29	15.40-17.20	1.78	5	13.78	0.13	13.40-14.20	0.94	5	13.10	0.08	12.90-13.30	0.61	5	16.33	0.15	15.90-16.80	0.92
730 - Sed Rate (ESR)	ES-1																								
Sedimentation Rate	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Clinical Data (Vital Diag) Excyte M/10	15	6.13	1.41	2.00-10.00	23.33																				
PG: Modified Westergren - Plastic	13	7.46	2.07	1.00-13.00	30.00																				
PG: Streck ESR-Auto Plus	6	10.00	0.89	7.00-13.00	9.00																				
PG: Westergren - Glass	9	6.00	1.58	1.00-11.00	26.67																				
PG: Westergren - Plastic	15	5.53	1.92	0.00-12.00	31.67																				
PG: Polymedco Sediplast Westergren ESR	9	4.67	1.58	0.00-10.00	32.00																				
PG: Hypoguard (ARKRAY) Dispette 2 (Labsco)	12	6.83	2.13	1.00-13.00	30.00																				
PG: Streck ESR-10 Manual	12	7.25	2.09	1.00-13.00	30.00																				
AG: All Group	105	6.57	2.14	1.00-13.00	30.00																				
672 - Glycohemoglobin	GH-1																								
Glycohemoglobin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Bayer DCA 2000 series	19	7.48	0.37	6.40-8.60	4.93																				
PG: Siemens Dimension/AR/ES/RxLXpand/EXL	13	9.45	0.46	8.10-10.90	4.84																				
PG: Chex Diagnostics (Bayer) A1C Now	43	6.89	0.75	4.70-9.20	10.87																				
PG: None	5	5.00	0.00	5.00-5.00	0.00																				
SG: Select Beckman Coulter Instruments (Chem)	5	10.10	0.88	7.50-12.70	8.71																				
SG: Select Roche Instruments	5	10.00	0.39	8.80-11.20	3.90																				
PG: Siemens (Bayer) DCA Vantage	111	7.65	0.66	5.60-9.60	8.68																				
SG: Select Siemens (Bayer) DCA Instruments	130	7.62	0.63	5.70-9.50	8.29																				
PG: Bayer A1CNow+	36	6.61	0.96	3.70-9.50	14.55																				
SG: Select Bayer (Metrika) Instruments	46	7.04	0.95	4.20-9.80	13.57																				
PG: TOSOH G8	6	9.39	0.39	8.20-10.60	4.15																				
SG: TOSOH G Series GHB	6	9.39	0.39	8.20-10.60	4.15																				
SG: Select Gen Chem Instruments	11	10.48	1.51	6.00-15.00	14.38																				
PG: PTS A1CNow	12	6.60	1.08	3.40-9.80	16.36																				
AG: All Group	251	7.50	0.99	4.50-10.50	13.20																				
761 - High Sensitivity CRP	HCR-1																								
hsCRP	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Roche Integra	5	0.80	0.00	0.80-0.80	0.00																				
AG: All Group	5	0.10	0.02	0.05-0.15	17.00																				
602 - Hemoglobin and Hematocrit - Waived;614 - Auto Diff II (All Other 3-part);610 - Regular Hematology;601 - Waived/PPM Tests II;609 - Waived/PPM Tests I;	HD-1																								
Granulocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn 1700	13	82.80	1.48	78.40-87.20	1.79	13	33.73	4.43	20.40-47.00	13.15	13	62.31	1.71	57.20-67.40	2.74	13	76.90	1.59	72.10-81.70	2.07	13	56.76	2.44	49.50-64.10	4.30
PG: Abbott Cell-Dyn 1800	23	82.56	1.63	77.70-87.50	1.97	22	33.57	1.97	27.70-39.50	5.83	23	61.60	6.73	41.40-81.80	10.93	23	77.31	1.70	72.20-82.40	2.20	23	57.43	1.92	51.60-63.20	3.34
PG: CDS Medonic 620	10	79.32	1.44	75.00-83.60	1.82	10	29.47	3.06	20.30-38.70	10.37	10	59.20	1.32	55.20-63.20	2.23	10	73.77	1.75	68.50-79.00	2.37	10	54.19	1.98	48.30-60.10	3.65
PG: Beckman Coulter AcT Diff/Diff 2	190	81.06	6.60	61.30-100.90	8.14	190	31.46	1.56	26.80-36.20	4.95	190	60.68	1.65	55.80-65.70	2.72	190	75.65	2.13	69.20-82.00	2.82	190	55.02	4.26	42.20-67.80	7.75
PG: Horiba Micros/45/60	60	84.59	0.79	82.20-87.00	0.93	60	34.04	4.84	19.50-48.50	14.24	60	65.67	1.40	61.50-69.90	2.13	60	78.55	1.90	72.90-84.30	2.42	60	59.26	2.12	53.00-65.60	3.56
SG: Select Abbott Cell Dyn Instruments (3-part Diff)	142	82.76	5.74	65.60-100.00	6.93	141	33.94	3.44	23.60-44.20	10.15	142	63.18	5.75	46.00-80.50	9.10	142	76.92	5.75	59.70-94.20	7.48	142	57.32	4.81	42.90-71.70	8.39
PG: Drew Scientific D3	18	83.27	0.84	80.80-85.80	1.01	17	32.58	0.95	29.80-35.50	2.91	18	62.87	0.85	60.40-65.50	1.35	18	77.32	0.73	75.10-79.50	0.94	18	56.96	0.94	54.20-59.80	1.65
PG: Abbott Cell-Dyn Emerald	105	82.80	6.65	62.80-102.80	8.03	105	34.06	3.58	23.40-44.80	10.50	105	63.67	5.84	46.20-81.20	9.17	105	76.86	6.65	57.00-96.90	8.65	105	57.37	5.49	40.90-73.90	9.56
PG: CDS M Series	138	78.84	7.92	55.00-102.60	10.05	138	29.78	3.51	19.30-40.30	11.78	138	59.24	4.84	44.70-73.70	8.18	138	74.03	5.52	57.40-90.60	7.46	138	54.45	4.67	40.40-68.40	8.60
SG: CDS Hematology Instruments	148	78.87	7.66	55.90-101.90	9.71	148	29.76	3.48	19.40-40.20	11.68	148	59.24	4.68	45.20-73.20	7.91	148	74.01	5.34	58.00-90.00	7.22	148	54.43	4.54	40.80-68.00	8.35
SG: International Abbott Emerald Instruments	106	82.81	6.62	62.90-102.70	8.00	106	34.05	3.56	23.40-44.80	10.44	106	63.67	5.82	46.20-81.20	9.14	106	76.86	6.62	57.00-96.80	8.61	106	57.36	5.46	41.00-73.80	9.51
AG: All Group	559	81.39	6.47	62.00-100.80	7.95	557	31.98	3.53	21.40-42.60	11.03	559	61.57	4.43	48.30-74.90	7.19	559	75.92	4.46	62.50-89.30	5.88	559	55.99	4.53	42.40-69.60	8.09
Hematocrit	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn 1700	14	52.33	1.58	49.20-55.40	3.02	11	17.73	0.53	16.60-18.80	2.99	14	39.21	1.59	36.80-41.60	4.06	14	45.80	1.75	43.10-48.50	3.82	14	29.05	0.98	27.40-30.80	3.37
PG: Abbott Cell-Dyn 1800	23	52.52	1.93	49.40-55.60	3.68	22	18.30	0.57	17.20-19.40	3.11	22	39.25	1.40	36.80-41.60	3.57	23	46.51	1.89	43.70-49.30	4.06	23	29.25	0.72	27.40-31.00	2.47
PG: CDS Medonic 620	12	50.77	2.12	47.80-53.80	4.17	12	15.96	0.48	15.00-17.00	3.00	12	36.55	0.94	34.30-38.70	2.58	12	44.05	1.76	41.40-46.60	4.00	11	26.45	0.82	24.80-28.00	3.11
PG: Beckman Coulter AcT Diff/Diff 2	203	52.57	2.01	49.40-55.80	3.82	203	17.57	0.51	16.50-18.70	2.90	203	38.78	1.05	36.50-41.10	2.71	203	45.66	1.78	43.00-48.40	3.89	203	28.64	1.41	26.90-30.30	4.93
PG: Horiba Micros/45/60	61	50.93	2.42	47.80-54.00	4.75	61	15.95	0.97	14.90-16.90	6.10	59	36.73	0.93	34.50-38.90	2.53	61	43.43	3.32	40.80-46.00	7.65	61	26.73	2.01	25.10-28.30	7.53
SG: Select Abbott Cell Dyn Instruments (3-part Diff)	144	52.64	1.87	49.40-55.80	3.56	140	17.89	0.85	16.80-19.00	4.75	143	39.16	1.49	36.80-41.60	3.80	144	46.16	2.46	43.40-49.00	5.32	144	28.91	0.90	27.20-30.60	3.11
PG: Drew Scientific D3	18	52.92	1.61	49.70-56.10	3.02	17	17.76	0.53	16.70-18.90	2.98	18	38.98	1.03	36.70-41.30	2.64	18	45.89	1.38	43.10-48.70	3.01	18	28.83	0.86	27.10-30.50	2.99
PG: Abbott Cell-Dyn Emerald	106	52.71	1.91	49.50-55.90	3.62	106	17.81	0.91	16.70-18.90	5.11	106	39.12													



AG: All Group	578	52.32	2.28	49.20-55.40	4.36	508	17.22	0.77	16.20-18.20	4.48	575	38.24	1.45	35.90-40.50	3.77	578	45.35	2.54	42.70-48.10	5.59	577	28.05	1.52	26.30-29.70	5.43
Hemoglobin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn 1700	14	18.17	0.42	16.90-19.50	2.36	14	6.21	0.28	5.80-6.60	4.52	14	13.57	0.46	12.60-14.60	3.38	14	15.93	0.48	14.80-17.00	3.02	14	10.07	0.29	9.40-10.80	2.87
PG: Abbott Cell-Dyn 1800	23	18.43	0.75	17.10-19.70	4.08	22	6.16	0.23	5.80-6.60	3.71	23	13.24	1.64	12.30-14.10	12.42	23	16.20	0.69	15.10-17.30	4.26	23	9.93	0.32	9.20-10.60	3.23
PG: CDS Medonic 620	12	18.03	0.57	16.70-19.30	3.17	12	5.96	0.16	5.60-6.40	2.67	12	13.34	0.22	12.40-14.20	1.65	12	15.74	0.32	14.60-16.80	2.04	11	9.83	0.20	9.10-10.50	2.04
PG: Beckman Coulter AcT Diff/Diff 2	203	18.24	0.54	16.90-19.50	2.97	203	5.97	0.17	5.60-6.40	2.83	203	13.47	0.32	12.60-14.40	2.37	203	15.89	0.56	14.80-17.00	3.52	203	9.92	0.47	9.20-10.60	4.75
PG: Horiba Micros/45/60	60	18.15	0.74	16.90-19.50	4.07	60	5.89	0.38	5.50-6.30	6.44	60	13.43	0.44	12.50-14.30	3.28	58	15.76	0.35	14.70-16.90	2.22	60	9.80	0.47	9.10-10.50	4.80
SG: Select Abbott Cell Dyn Instruments (3-part Diff)	144	18.14	0.47	16.80-19.40	2.60	143	5.96	0.23	5.60-6.40	3.83	143	13.44	0.43	12.50-14.30	3.21	144	15.94	0.79	14.80-17.00	4.97	144	9.85	0.24	9.20-10.60	2.42
PG: Drew Scientific D3	18	18.09	0.55	16.80-19.40	3.04	17	5.79	0.19	5.40-6.20	3.28	18	13.27	0.44	12.40-14.20	3.23	18	15.65	0.45	14.60-16.80	2.87	18	9.70	0.32	9.00-10.40	3.30
PG: Abbott Cell-Dyn Emerald	106	18.07	0.37	16.80-19.40	2.04	106	5.88	0.15	5.50-6.30	2.54	106	13.39	0.43	12.50-14.30	3.21	106	15.87	0.84	14.80-17.00	5.28	106	9.80	0.19	9.10-10.50	1.94
PG: CDS M Series	139	18.14	0.87	16.80-19.40	4.81	139	6.01	0.19	5.60-6.40	3.17	139	13.39	0.17	12.50-14.30	1.27	139	15.79	0.98	14.70-16.90	6.20	139	9.83	0.13	9.10-10.50	1.33
SG: CDS Hematology Instruments	151	18.13	0.85	16.80-19.40	4.70	151	6.01	0.19	5.60-6.40	3.17	151	13.38	0.17	12.50-14.30	1.27	151	15.79	0.94	14.70-16.90	5.95	150	9.83	0.14	9.10-10.50	1.43
SG: International Abbott Emerald Instruments	107	18.07	0.37	16.80-19.40	2.04	107	5.88	0.15	5.50-6.30	2.54	107	13.39	0.43	12.50-14.30	3.21	107	15.87	0.84	14.80-17.00	5.28	107	9.80	0.19	9.10-10.50	1.94
Lymphocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn 1700	13	12.04	1.50	7.55-16.52	12.42	13	52.56	5.72	35.39-69.73	10.89	13	27.70	2.40	20.50-34.90	8.66	13	16.82	1.67	11.81-21.83	9.93	13	33.29	3.08	24.05-42.53	9.26
PG: Abbott Cell-Dyn 1800	23	11.15	1.41	6.91-15.39	12.68	23	47.16	6.17	28.66-65.66	13.08	22	26.37	5.96	8.48-44.26	22.62	23	15.12	1.96	9.23-21.01	12.99	23	30.18	2.88	21.55-38.81	9.53
PG: CDS Medonic 620	10	15.09	0.51	13.56-16.62	3.37	10	61.57	2.25	54.83-68.31	3.65	10	33.34	1.02	30.27-36.42	3.07	10	20.06	0.81	17.63-22.49	4.03	10	38.17	1.16	34.69-41.65	3.04
PG: Beckman Coulter AcT Diff/Diff 2	191	14.20	0.49	12.72-15.68	3.48	190	62.02	1.70	56.91-67.13	2.74	191	32.63	1.49	28.16-37.10	4.57	187	19.72	0.67	17.70-21.74	3.42	191	39.13	1.76	33.84-44.42	4.50
PG: Horiba Micros/45/60	60	11.51	0.81	9.09-13.93	6.99	60	44.90	3.94	33.07-56.73	8.78	60	24.37	1.79	19.01-29.73	7.34	60	15.86	1.79	10.49-21.23	11.29	60	28.49	4.49	15.02-41.96	15.76
SG: Select Abbott Cell Dyn Instruments (3-part Diff)	142	13.20	1.88	7.56-18.84	14.23	142	52.34	6.05	34.18-70.50	11.56	141	28.65	3.64	17.72-39.58	12.72	142	17.90	2.45	10.56-25.24	13.66	142	33.66	3.67	22.64-44.68	10.91
PG: Drew Scientific D3	18	13.88	0.70	11.78-15.98	5.04	17	57.89	1.94	52.06-63.72	3.36	18	31.05	0.96	28.16-33.94	3.10	18	18.96	0.61	17.13-20.79	3.21	18	36.62	1.21	32.99-40.25	3.30
PG: Abbott Cell-Dyn Emerald	99	13.74	0.93	10.96-16.52	6.74	104	53.91	2.11	47.59-60.23	3.91	104	29.48	1.16	26.01-32.95	3.93	101	18.62	1.06	15.43-21.82	5.72	104	34.79	1.32	30.82-38.76	3.80
PG: CDS M Series	138	14.78	2.15	8.31-21.24	14.58	138	60.25	5.43	43.95-76.55	9.02	138	32.97	2.77	24.65-41.29	8.41	138	19.69	1.60	14.89-24.49	8.13	138	37.51	3.28	27.66-47.36	8.75
SG: CDS Hematology Instruments	148	14.81	2.09	8.55-21.07	14.08	148	60.34	5.29	44.48-76.20	8.76	148	33.00	2.69	24.93-41.07	8.15	148	19.72	1.56	15.04-24.40	7.92	148	37.56	3.19	28.00-47.12	8.49
SG: International Abbott Emerald Instruments	100	13.73	0.92	10.96-16.50	6.72	105	53.93	2.10	47.62-60.24	3.90	105	29.50	1.16	26.01-32.99	3.94	102	18.63	1.06	15.44-21.82	5.71	105	34.82	1.33	30.82-38.82	3.83
AG: All Group	560	13.79	1.79	8.43-19.15	12.96	558	57.05	7.26	35.26-78.84	12.73	559	30.74	3.81	19.30-42.18	12.40	556	18.79	2.09	12.52-25.06	11.12	560	36.06	4.58	22.34-49.79	12.69
MCV	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn 1700	14	93.74	1.89	88.00-99.40	2.02	14	75.53	1.77	70.20-80.80	2.34	14	85.48	1.79	80.10-90.90	2.09	14	89.97	1.99	84.00-96.00	2.21	14	82.19	1.80	76.80-87.60	2.19
PG: Abbott Cell-Dyn 1800	23	96.07	2.05	89.90-102.30	2.13	22	77.47	2.25	70.80-84.30	2.90	23	87.05	2.85	78.50-95.50	3.28	23	92.30	1.69	87.20-97.40	1.83	23	83.70	1.55	79.00-88.40	1.85
PG: CDS Medonic 620	10	88.49	1.98	82.60-94.40	2.24	10	69.96	1.12	66.60-73.40	1.60	10	78.93	1.32	74.90-82.90	1.67	10	84.57	1.86	79.00-90.20	2.20	10	75.59	1.26	71.80-79.40	1.67
PG: Beckman Coulter AcT Diff/Diff 2	203	90.61	5.85	73.00-108.10	6.46	203	74.87	0.86	72.30-77.50	1.15	202	82.97	5.37	66.90-99.10	6.47	203	87.40	1.24	83.70-91.10	1.42	203	80.40	1.04	77.30-83.50	1.29
PG: Horiba Micros/45/60	52	88.80	1.02	85.70-91.90	1.15	52	70.65	0.85	68.00-73.10	1.20	52	80.78	1.04	77.70-83.90	1.29	52	84.93	0.99	81.90-87.90	1.17	50	76.60	2.25	69.80-83.30	2.94
SG: Select Abbott Cell Dyn Instruments (3-part Diff)	143	96.77	1.91	91.10-102.50	1.97	142	79.10	2.08	72.90-85.30	2.63	143	88.30	2.12	81.90-94.70	2.40	143	93.00	1.78	87.70-98.30	1.91	143	85.15	1.84	79.60-90.60	2.16
PG: Drew Scientific D3	18	93.87	0.82	91.40-96.40	0.87	18	78.55	3.81	67.20-90.00	4.85	18	85.92	0.59	84.10-87.70	0.69	18	90.23	0.71	88.10-92.30	0.79	18	83.74	2.50	76.20-91.20	2.97
PG: Abbott Cell-Dyn Emerald	105	97.39	1.31	93.50-101.30	1.34	105	79.97	1.12	76.60-83.40	1.40	105	88.99	1.41	84.80-93.20	1.58	105	93.60	1.18	90.10-97.10	1.26	105	85.92	1.15	82.50-89.40	1.34
PG: CDS M Series	138	90.16	1.80	84.80-95.60	2.00	138	71.17	0.99	68.20-74.20	1.39	138	80.26	1.14	76.90-83.70	1.42	138	86.07	1.19	82.50-89.70	1.38	138	76.84	1.08	73.60-80.00	1.41
SG: CDS Hematology Instruments	148	90.04	1.85	84.50-95.50	2.06	148	71.09	1.04	68.00-74.20	1.46	148	80.17	1.20	76.60-83.80	1.50	148	85.97	1.29	82.10-89.90	1.51	148	76.76	1.14	73.40-80.20	1.48
SG: International Abbott Emerald Instruments	106	97.38	1.31	93.50-101.30	1.34	106	79.96	1.12	76.60-83.40	1.40	106	88.99	1.41	84.80-93.20	1.58	106	93.60	1.18	90.10-97.10	1.25	106	85.91	1.15	82.50-89.40	1.34
AG: All Group	565	91.99	4.77	77.70-106.30	5.18	564	74.69	3.54	64.10-85.30	4.75	564	83.51	4.65	69.50-97.50	5.57	565	88.34	3.25	78.60-98.00	3.67	563	80.44	3.64	69.50-91.30	4.53
Monocyte/Mixed	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Cell-Dyn 1700	13	5.16	0.58	3.50-6.90	11.15	13	13.71	2.85	5.10-22.30	20.80	13	9.99	1.32	6.00-14.00	13.20	13	6.30	0.67	4.30-8.30	10.63	13	9.96	1.29	6.10-13.90	12.90
PG: Abbott Cell-Dyn 1800	22	6.34	0.55	4.60-8.00	8.73	22	18.30	2.48	10.90-25.70	13.55	22	12.47	1.51	8.00-17.00	12.08	22	7.81	0.69	5.70-9.90	8.85	22	12.60	1.02	9.50-15.70	8.10
PG: CDS Medonic 620	10	5.60	1.05	2.40-8.80	18.75	10	8.96	2.04	2.90-15.10	22.67	10	7.46	1.11	4.20-10.80	14.80	10	6.17	1.15	2.80-9.70	18.55	10	7.64	1.31	3.70-11.50	17.24
PG: Beckman Coulter AcT Diff/Diff 2	191	4.07	0.41	2.80-5.40	10.24	191	6.52	1.25	2.80-10.30	19.23	191	6.70	0.64	4.80-8.60	9.55	191	4.45	0.52	2.80-6.00	11.82	190	5.58	0.69	3.50-7.70	12.32
PG: Horiba Micros/45/60	60	3.89	0.26	3.10-4.70	6.67	60	20.49	2.39	13.30-27.70	11.66	60	9.82	1.52	5.20-14.40	15.51	60	5.59	0.50	4.10-7.10	8.93	60	11.81	1.33	7.80-15.80	11.27
SG: Select Abbott Cell Dyn Instruments (3-part Diff)	134	3.37	1.23	0.00-7.10	36.18	135	12.61	2.76	4.30-20.90	21.90	135	7.32	2.18	0.80-13.8											



SG: International Abbott i-Stat Instruments	12	165.08	10.27	134.00-196.00	6.24	12	91.69	8.19	67.00-117.00	8.91	11	75.75	6.54	57.00-96.00	8.55	12	167.00	8.23	142.00-192.00	4.91	12	90.77	7.22	69.00-113.00	7.91
Chloride (CL-)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat / Thermal	8	105.63	0.52	100.30-110.90	0.49	6	86.83	0.75	83.00-91.00	0.92	6	79.83	0.41	75.80-83.80	0.51	6	96.83	0.75	92.00-102.00	0.82	6	86.50	0.84	83.00-91.00	0.97
SG: International Abbott i-Stat Instruments / Thermal	36	105.11	1.29	99.80-110.40	1.24	10	86.82	0.60	83.00-91.00	0.69	10	79.45	0.69	75.50-83.50	0.89	10	96.64	0.67	92.00-102.00	0.72	10	86.73	0.65	83.00-91.00	0.75
PG: Abbott i-Stat (Waived) / Thermal	24	105.08	1.50	99.80-110.40	1.43																				
CO2 (Total CO2 Bicarb)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat / Thermal	9	18.78	1.48	14.40-23.20	7.87	7	26.71	0.49	25.20-28.20	1.84	7	23.86	1.86	18.30-29.50	7.78	7	25.29	2.36	18.20-32.40	9.33	7	26.00	1.00	23.00-29.00	3.85
SG: International Abbott i-Stat Instruments / Thermal	37	18.61	1.05	15.50-21.80	5.65	11	26.33	0.65	24.40-28.30	2.47	11	23.75	1.42	19.50-28.10	5.97	11	25.42	1.78	20.10-30.70	7.01	11	25.92	0.90	23.20-28.60	3.47
PG: Abbott i-Stat (Waived) / Thermal	24	18.46	0.98	15.60-21.40	5.30																				
Glucose	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat / Thermal	11	84.91	1.97	77.00-94.00	2.35	9	143.67	3.43	129.30-158.10	2.39	9	51.56	0.88	45.60-57.60	1.73	9	194.44	4.22	175.00-213.00	2.17	9	143.22	2.54	128.90-157.50	1.77
SG: International Abbott i-Stat Instruments / Thermal	39	84.74	1.80	77.00-94.00	2.13	13	144.49	2.92	130.10-158.90	2.02	13	51.16	0.88	45.20-57.20	1.76	13	195.92	3.98	176.00-216.00	2.04	13	144.07	2.36	129.70-158.50	1.67
PG: Abbott i-Stat (Waived) / Thermal	24	84.29	1.71	76.00-92.00	2.03																				
Hematocrit	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat / Thermal	6	28.33	0.82	26.60-30.00	2.90	6	25.50	0.84	24.00-27.00	3.29	6	15.67	0.52	14.80-16.60	3.31	6	33.17	0.41	31.20-35.20	1.23	6	25.33	0.82	23.80-26.80	3.24
SG: International Abbott i-Stat Instruments / Thermal	20	27.81	0.68	26.10-29.50	2.45	10	25.27	0.65	23.80-26.80	2.57	10	15.55	0.52	14.60-16.40	3.35	10	33.09	0.30	31.10-35.10	0.91	10	25.18	0.60	23.70-26.70	2.38
PG: Abbott i-Stat (Waived) / Thermal	10	27.40	0.52	25.80-29.00	1.90																				
Hemoglobin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat / Thermal	6	9.58	0.20	8.90-10.30	2.08	6	8.63	0.22	8.00-9.20	2.56	6	5.23	0.19	4.80-5.60	3.65	6	11.33	0.33	10.50-12.10	2.92	6	8.58	0.20	8.00-9.20	2.33
SG: International Abbott i-Stat Instruments / Thermal	20	9.44	0.19	8.70-10.10	2.02	10	8.57	0.17	8.00-9.20	1.98	10	5.23	0.17	4.80-5.60	3.27	10	11.27	0.24	10.50-12.10	2.12	10	8.55	0.15	7.90-9.10	1.76
PG: Abbott i-Stat (Waived) / Thermal	10	9.32	0.16	8.60-10.00	1.61																				
Ionized Calcium	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat / Thermal	6	0.79	0.01	0.76-0.82	1.27	6	2.20	0.02	2.13-2.27	1.00	6	0.60	0.01	0.57-0.63	1.67	6	0.97	0.02	0.92-1.02	1.86	6	2.21	0.02	2.15-2.27	0.86
SG: International Abbott i-Stat Instruments / Thermal	31	0.79	0.01	0.76-0.82	1.27	10	2.21	0.02	2.15-2.27	0.90	10	0.60	0.01	0.57-0.63	1.50	10	0.97	0.01	0.93-1.01	1.34	10	2.21	0.02	2.16-2.26	0.72
PG: Abbott i-Stat (Waived) / Thermal	21	0.79	0.01	0.75-0.83	1.52																				
Potassium (K+)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat / Thermal	14	6.19	0.03	5.70-6.70	0.48	12	3.80	0.04	3.30-4.30	1.05	12	2.25	0.05	1.80-2.80	2.17	12	4.59	0.03	4.10-5.10	0.65	12	3.80	0.00	3.30-4.30	0.00
SG: International Abbott i-Stat Instruments / Thermal	42	6.18	0.05	5.70-6.70	0.81	16	3.80	0.04	3.30-4.30	1.05	16	2.25	0.05	1.70-2.70	2.27	16	4.59	0.02	4.10-5.10	0.43	16	3.80	0.00	3.30-4.30	0.00
PG: Abbott i-Stat (Waived) / Thermal	24	6.16	0.07	5.70-6.70	0.97																				
Sodium (NA+)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat / Thermal	11	149.64	0.51	146.00-154.00	0.33	9	140.56	0.53	137.00-145.00	0.38	9	126.22	0.44	122.20-130.20	0.35	9	155.78	0.67	152.00-160.00	0.45	9	140.78	0.83	137.00-145.00	0.59
SG: International Abbott i-Stat Instruments / Thermal	39	149.77	0.73	146.00-154.00	0.49	13	140.50	0.52	137.00-145.00	0.37	13	126.21	0.43	122.20-130.20	0.34	13	155.86	0.54	152.00-160.00	0.34	13	140.79	0.70	137.00-145.00	0.50
PG: Abbott i-Stat (Waived) / Thermal	24	149.88	0.85	146.00-154.00	0.60																				
Urea Nitrogen (BUN)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat / Thermal	8	22.75	0.46	20.70-24.90	2.02	6	61.67	1.37	56.10-67.30	2.22	6	9.50	0.55	7.50-11.50	5.79	6	58.50	1.64	53.20-63.80	2.80	6	62.17	1.33	56.60-67.80	2.14
SG: International Abbott i-Stat Instruments / Thermal	36	22.68	1.05	20.70-24.70	4.63	10	61.67	1.04	56.10-67.30	1.69	10	9.51	0.41	7.50-11.50	4.32	10	58.26	1.24	53.10-63.50	2.13	10	61.89	1.00	56.30-67.50	1.62
PG: Abbott i-Stat (Waived) / Thermal	24	22.63	1.28	20.60-24.60	5.66																				
HCO3	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat	5	15.40	0.79	13.00-17.80	5.13	5	23.50	1.01	20.50-26.50	4.30	5	20.50	1.87	14.90-26.10	9.12	5	21.70	2.36	14.60-28.80	10.88	5	23.07	1.29	19.20-27.00	5.58
SG: International Abbott i-Stat Instruments	7	15.76	0.63	13.90-17.70	3.99	7	23.90	0.70	21.80-26.00	2.93	7	20.54	1.19	16.90-24.10	5.80	7	22.54	1.61	17.70-27.30	7.16	7	23.61	0.91	20.90-26.30	3.86
Lactate	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat	5	1.18	0.00	1.18-1.18	0.00	5	3.57	0.00	3.57-3.57	0.00	5	6.35	0.00	6.35-6.35	0.00	5	2.21	0.00	2.21-2.21	0.00	5	3.54	0.00	3.54-3.54	0.00
SG: International Abbott i-Stat Instruments	5	1.23	0.03	1.14-1.32	2.52	5	3.69	0.07	3.49-3.89	1.82	5	6.56	0.13	6.16-6.96	2.04	5	2.30	0.07	2.08-2.52	3.17	5	3.71	0.09	3.46-3.96	2.29
PG: Abbott i-Stat (Waived) / Thermal	8	33.75	3.01	25.00-43.00	8.82																				
Creatinine	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott i-Stat / Thermal	15	1.08	0.04	0.80-1.40	3.64	12	5.74	0.23	4.80-6.60	4.04	11	0.43	0.05	0.10-0.70	12.50	12	4.84	0.24	4.10-5.50	5.00	12	5.74	0.26	4.80-6.60	4.56
SG: International Abbott i-Stat Instruments / Thermal	46	1.10	0.04	0.80-1.40	3.64	17	5.76	0.22	4.90-6.70	3.79	16	0.43	0.04	0.10-0.70	10.00	17	4.86	0.21	4.20-5.60	4.29	17	5.79	0.25	4.90-6.70	4.31
PG: Abbott i-Stat (Waived) / Thermal	26	1.11	0.03	0.80-1.40	2.73																				
716 - Blood Lead - Waived Methods;709 - Blood Lead	LE-1					LE-2					LE-3					LE-4					LE-5				
Blood Lead (Waived)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: ESA LeadCare II (Waived)	40	7.98	1.20	4.40-11.60	15.00																				



SG: International Abbott i-Stat Instruments	12	1.12	0.09	0.80-1.40	8.18	12	2.28	0.10	2.00-2.60	4.35	12	1.16	0.11	0.90-1.50	9.17	12	2.19	0.12	1.80-2.60	5.45	12	2.19	0.11	1.90-2.50	5.00
SG: International Abbott CoaguseNSE Instruments	7	1.37	0.36	0.30-2.50	25.71	7	3.23	0.73	1.00-5.40	22.81	7	1.26	0.23	0.60-2.00	17.69	7	3.03	0.62	1.10-4.90	20.67	7	3.29	0.71	1.10-5.50	21.82
611 - QBC Hematology																									
QBC-1					QBC-2					QBC-3					QBC-4					QBC-5					
Hemoglobin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: QBC Star - QBC Star Tubes	6	9.12	0.10	8.50-9.70	1.10	6	10.53	0.24	9.80-11.20	2.29	6	11.80	0.53	11.00-12.60	4.49	5	7.42	0.20	6.90-7.90	2.70	6	10.40	0.31	9.70-11.10	2.98
SG: Select QBC - Cap and Ven	5	8.74	0.22	8.10-9.30	2.53	5	9.96	0.30	9.30-10.70	3.00	5	10.95	0.44	10.20-11.80	4.00	5	7.24	0.21	6.70-7.70	2.92	5	10.02	0.45	9.30-10.70	4.50
Hematocrit	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: QBC Star - QBC Star Tubes	6	27.82	0.73	26.10-29.50	2.63	6	32.08	0.87	30.20-34.00	2.71	6	35.55	2.02	33.40-37.60	5.69	5	22.98	0.68	21.60-24.40	2.96	6	31.42	1.19	29.50-33.30	3.79
SG: Select QBC - Cap and Ven	5	26.90	1.09	25.30-28.50	4.05	5	30.46	1.12	28.70-32.30	3.64	5	33.74	0.98	31.70-35.70	2.91	5	21.88	0.68	20.60-23.20	3.11	5	30.66	1.01	28.90-32.50	3.29
White Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: QBC Star - QBC Star Tubes	6	11.93	0.77	10.10-13.70	6.47	6	24.72	2.52	21.00-28.40	10.20	6	28.22	4.50	24.00-32.40	15.96	6	6.83	0.34	5.80-7.80	5.00	6	24.18	2.63	20.60-27.80	10.87
SG: Select QBC - Cap and Ven	5	8.20	0.42	7.00-9.40	5.12	5	14.24	0.23	12.10-16.30	1.62	5	14.52	1.39	12.30-16.70	9.59	5	5.12	0.30	4.30-5.90	5.88	5	14.62	1.30	12.40-16.80	8.90
Granulocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: QBC Star - QBC Star Tubes	5	60.60	3.85	50.00-72.00	6.23	5	45.00	3.54	35.00-56.00	7.78	5	37.60	5.86	20.00-56.00	15.53	5	50.40	3.98	38.00-62.00	8.00	5	43.00	4.06	31.00-55.00	9.53
SG: Select QBC - Cap and Ven	5	57.00	5.94	39.00-75.00	10.35	5	42.25	3.30	32.00-52.00	7.86	5	37.50	1.00	35.00-41.00	2.63	5	44.75	3.10	36.00-54.00	6.89	5	43.25	3.30	33.00-53.00	7.67
Lymphocyte/Monocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: QBC Star - QBC Star Tubes	5	39.40	3.85	28.00-50.00	9.74	5	55.00	3.54	45.00-66.00	6.36	5	62.40	5.86	44.00-80.00	9.52	5	49.60	3.98	38.00-62.00	8.00	5	57.00	4.06	45.00-69.00	7.19
SG: Select QBC - Cap and Ven	5	43.00	5.94	25.00-61.00	13.72	5	57.75	3.30	48.00-68.00	5.69	5	62.50	1.00	60.00-66.00	1.59	5	55.25	3.10	46.00-64.00	5.64	5	56.75	3.30	47.00-67.00	5.79
Platelet Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: QBC Star - QBC Star Tubes	6	345.17	16.93	259.00-431.00	4.90	6	155.00	21.08	116.00-194.00	13.61	6	161.17	33.96	121.00-201.00	21.12	6	607.00	37.13	455.00-759.00	6.11	6	144.33	16.02	108.00-180.00	11.11
SG: Select QBC - Cap and Ven	5	292.00	34.73	219.00-365.00	11.88	5	115.00	18.70	86.00-144.00	16.26	5	122.25	22.53	92.00-153.00	18.44	5	601.00	27.54	451.00-751.00	4.58	5	109.20	20.02	82.00-136.00	18.35
754 - Infectious Mono/RF Combo;653 - Rheumatoid Factor																									
RF-1 N					RF-2 N					RF-3 N					RF-4 N					RF-5 N					
Rheumatoid Factor (Quant)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	5	7.41	7.13	0.00-28.80	96.35	5	118.85	13.84	77.40-160.40	11.64	5	67.29	9.07	40.10-94.50	13.48	5	117.44	12.46	80.00-154.80	10.61	5	7.96	6.69	0.00-28.10	83.63
686 - Special Chemistry;667 - PSA Add-on																									
SC-1																									
CEA	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	6	30.01	1.86	24.42-35.60	6.21																				
DHEA-S	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	5	0.54	0.07	0.30-0.70	14.00																				
Estradiol	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: TOSOH AIA	7	790.07	19.83	730.60-849.60	2.51																				
SG: Select Beckman Coulter Instruments (Chem)	5	495.60	52.09	339.30-651.90	10.51																				
AG: All Group	12	429.37	80.59	187.60-671.20	18.77																				
Ferritin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman ACCESS	7	228.09	13.26	188.30-267.90	5.81																				
SG: Select Beckman Coulter Instruments (Chem)	11	226.41	11.46	192.00-260.80	5.06																				
SG: Select OCD Vitros Instruments	5	201.20	9.98	171.30-231.10	4.96																				
SG: Select Abbott Architect Instruments	10	347.81	12.64	309.90-385.70	3.63																				
AG: All Group	29	230.81	28.33	145.80-315.80	12.27																				
Folate	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman ACCESS	8	10.16	0.92	7.40-13.00	9.02																				
PG: TOSOH AIA	5	5.28	0.34	4.30-6.30	6.42																				
SG: Select Beckman Coulter Instruments (Chem)	10	10.16	0.93	7.40-13.00	9.12																				
SG: Select Abbott Architect Instruments	11	10.45	1.25	6.70-14.20	12.02																				
AG: All Group	35	8.85	2.29	2.00-15.80	25.73																				
FSH	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: TOSOH AIA	6	61.75	3.32	51.80-71.80	5.37																				
AG: All Group	16	63.89	5.89	46.20-81.60	9.22																				
LH	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: TOSOH AIA	6	71.93	3.12	62.60-81.20	4.33																				
PG: OCD Vitros ECI/ECIQ	5	139.00	1.41	134.80-143.20	1.01																				
AG: All Group	11	69.79	13.22	30.20-109.40	18.93																				
Progesterone	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	15	38.57	6.41	19.40-57.80	16.61																				
Prolactin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: TOSOH AIA	5	58.66	2.12	52.30-65.10	3.61																				
AG: All Group	10	60.09	6.42	40.90-79.30	10.67																				
PSA	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Beckman ACCESS</																									



Test	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%					
SG: Select Abbott Architect Instruments	13	20.01	1.01	16.98-23.04	5.05																									
PG: NanoEnTek FrenD System	7	23.92	2.05	17.77-30.07	8.57																									
Testosterone																														
PG: Beckman ACCESS	15	11.99	0.79	9.60-14.40	6.58																									
PG: Qualigen FastPak Testosterone	8	9.22	0.47	7.80-10.60	5.11																									
PG: TOSOH AIA	11	18.78	0.89	16.20-21.40	4.68																									
SG: Select Beckman Coulter Instruments (Chem)	18	11.89	0.77	9.60-14.20	6.47																									
SG: Select Abbott Architect Instruments	5	11.08	2.19	4.50-17.70	19.73																									
AG: All Group	41	11.77	2.25	5.10-18.60	19.07																									
Vitamin B12																														
PG: Beckman ACCESS	13	1105.54	71.41	892.00-1320.00	6.46																									
PG: TOSOH AIA	5	1272.60	50.73	1121.00-1425.00	3.98																									
SG: Select Beckman Coulter Instruments (Chem)	16	1098.00	70.16	887.00-1309.00	6.39																									
SG: Select OCD Vitros Instruments	5	995.40	10.29	964.00-1026.00	1.04																									
PG: Roche Cobas e 411	5	1610.80	102.48	1304.00-1919.00	6.36																									
SG: Select Abbott Architect Instruments	11	1579.36	159.31	1101.00-2057.00	10.09																									
AG: All Group	31	1285.90	175.61	759.00-1813.00	13.65																									
688 - SHBG and Testosterone																														
		SHT-1						SHT-2																						
SHBG																														
AG: All Group	5	21.50	0.72	19.34-23.66	3.35	5	66.40	3.18	56.86-75.94	4.79																				
685 - Special Chemistry II																														
		SPC-1						SPC-2																						
Insulin																														
SG: Select Beckman Coulter Instruments (Chem)	5	16.39	0.55	14.70-18.00	3.35	5	107.02	10.70	74.90-139.10	10.00																				
SG: Select Siemens Immulite Instruments	5	7.55	0.64	5.70-9.50	8.42	5	41.75	4.60	28.00-55.60	11.00																				
AG: All Group	7	15.45	1.50	10.90-19.90	9.81	7	104.20	8.50	78.70-129.70	8.16																				
PTH																														
SG: Select Siemens Immulite Instruments	5	10.95	1.91	5.30-16.70	17.36	5	24.60	17.82	0.00-78.10	72.44																				
PG: Intl OCD Vitros ECI/ECIQ	5	10.70	0.00	10.70-10.70	0.00	5	38.50	0.00	38.50-38.50	0.00																				
AG: All Group	8	5.42	2.25	0.00-12.20	41.67	9	19.17	10.81	0.00-51.60	56.30																				
Vitamin D																														
PG: Beckman ACCESS	18	12.32	1.62	7.45-17.19	13.17	18	83.13	10.01	53.09-113.17	12.04																				
PG: TOSOH AIA	8	25.38	2.23	18.70-32.06	8.78	8	100.09	3.67	89.07-111.11	3.67																				
SG: Select Beckman Coulter Instruments (Chem)	21	12.13	1.75	6.89-17.37	14.41	21	84.77	10.55	53.12-116.42	12.44																				
PG: Roche Cobas e 411	5	5.62	1.07	2.41-8.83	19.04	6	41.81	6.54	22.21-61.42	15.63																				
PG: Abbott Architect i1000 SR	7	9.64	0.70	7.53-11.75	7.28	7	76.49	2.04	70.38-82.60	2.66																				
SG: Select Abbott Architect Instruments	13	9.69	0.77	7.39-11.98	7.89	13	77.63	3.35	67.59-87.67	4.31																				
AG: All Group	51	11.54	3.53	0.96-22.12	30.56	52	73.87	16.94	23.06-124.68	22.93																				
752 - Auto Diff 8 (Sysmex 5-part Diff)																														
		SXT-1						SXT-2						SXT-3						SXT-4						SXT-5				
White Blood Cell Count																														
PG: Sysmex XS-1000i	11	20.66	0.41	17.60-23.80	1.98	11	3.95	0.26	3.30-4.50	6.67	11	7.08	0.28	6.00-8.20	3.94	11	20.68	0.44	17.60-23.80	2.08	11	3.97	0.27	3.40-4.60	6.75					
AG: All Group	12	20.72	0.44	17.60-23.80	2.13	12	3.93	0.26	3.30-4.50	6.67	12	7.17	0.42	6.10-8.30	5.83	12	20.70	0.42	17.60-23.80	2.03	12	3.93	0.29	3.30-4.50	7.44					
Lymphocyte																														
PG: Sysmex XS-1000i	11	21.63	2.91	12.90-30.30	13.47	11	12.67	1.34	8.70-16.70	10.47	11	26.35	3.16	16.90-35.90	11.97	11	21.47	2.52	13.90-29.10	11.72	11	13.09	1.16	9.60-16.60	8.85					
AG: All Group	12	21.69	2.79	13.30-30.10	12.86	12	13.15	2.09	6.90-19.50	15.83	12	26.54	3.08	17.30-35.70	11.62	12	21.90	2.82	13.40-30.40	12.88	12	13.47	1.71	8.40-18.60	12.67					
Neutrophil																														
PG: Sysmex XS-1000i	10	53.99	2.00	48.00-60.00	3.70	11	61.37	1.42	57.10-65.70	2.31	10	52.27	3.04	43.20-61.40	5.81	11	55.47	4.39	42.30-68.70	7.91	11	61.54	1.27	57.70-65.30	2.07					
AG: All Group	11	54.53	2.60	46.70-62.30	4.77	12	61.81	2.03	55.70-67.90	3.28	11	52.61	3.10	43.30-61.90	5.89	12	55.66	4.24	43.00-68.40	7.61	12	61.99	1.99	56.00-68.00	3.21					
Mono/Mixed																														
PG: Sysmex XS-1000i	10	1.75	0.26	1.00-2.60	14.44	10	0.44	0.16	0.00-0.90	40.00	9	0.94	0.32	0.00-1.90	35.56	10	1.67	0.49	0.20-3.20	28.82	10	0.37	0.17	0.00-0.90	42.50					
AG: All Group	10	1.75	0.26	1.00-2.60	14.44	11	0.43	0.16	0.00-0.90	40.00	10	1.03	0.41	0.00-2.20	41.00	11	1.80	0.63	0.00-3.70	35.00	11	0.36	0.16	0.00-0.90	40.00					
Eosinophil																														
PG: Sysmex XS-1000i	10	13.70	0.86	11.10-16.30	6.28	10	15.93	0.73	13.70-18.10	4.59	10	13.22	0.81	10.80-15.60	6.14	10	13.64	0.94	10.80-16.40	6.91	10	16.00	0.57	14.30-17.70	3.56					
AG: All Group	11	13.58	0.91	10.90-16.30	6.69	11	15.82	0.79	13.40-18.20	5.00	11	13.25	0.78	11.00-15.60	5.86	11	13.55	0.95	10.70-16.40	7.04	11	15.92	0.60	14.10-17.70	3.77					
Basophil																														
PG: Sysmex XS-1000i	10	8.53	0.78	6.19-10.87	9.13	10	9.62	0.94	6.78-12.45	9.82	10	7.62	1.35	3.56-11.68	17.74	10	8.67	0.62	6.80-10.54	7.20	10	9.05	0.81	6.62-11.48	8.96					
AG: All Group	10	8.53	0.78	6.19-10.87	9.13	10	9.62	0.94	6.78-12.45	9.82	10	7.62	1.35	3.56-11.68	17.74	10	8.67	0.62	6.80-10.54	7.20	10	9.05	0.81	6.62-11.48	8.96					
Red Blood Cell Count																														
PG: Sysmex XS-1000i	11	5.59	0.03	5.25-5.93	0.52	11	2.17	0.01	2.04-2.30	0.46	11	4.69	0.04	4.41-4.97	0.90	11														



AG: All Group	12	93.54	1.21	89.90-97.10	1.29	12	80.23	0.66	78.20-82.20	0.82	12	85.96	0.91	83.30-88.70	1.06	12	93.38	1.30	89.50-97.30	1.39	12	80.26	0.65	78.30-82.30	0.81
Platelet Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex XS-1000i	11	422.18	10.74	317.00-528.00	2.54	11	62.09	4.18	47.00-78.00	6.77	11	215.00	5.42	161.00-269.00	2.51	11	421.82	10.28	317.00-528.00	2.44	11	63.09	1.87	47.00-79.00	3.02
AG: All Group	12	424.42	12.84	318.00-530.00	3.02	12	62.08	3.99	47.00-78.00	6.45	12	216.00	6.22	162.00-270.00	2.87	12	423.92	12.21	318.00-530.00	2.88	12	63.08	1.78	47.00-79.00	2.86
RDW	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex XS-1000i	11	14.08	0.13	13.70-14.50	0.92	11	16.12	0.24	15.40-16.80	1.49	11	15.19	0.12	14.80-15.60	0.79	11	14.07	0.13	13.70-14.50	0.92	11	16.22	0.27	15.40-17.00	1.67
AG: All Group	12	14.08	0.13	13.70-14.50	0.92	12	16.12	0.23	15.40-16.80	1.43	12	15.18	0.12	14.80-15.60	0.79	12	14.07	0.12	13.70-14.50	0.85	12	16.21	0.26	15.40-17.00	1.60
613 - Auto Diff I (Sysmex 3-part)	SYX-1					SYX-2					SYX-3					SYX-4					SYX-5				
White Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex K-800,1000,4500,KX-21	17	19.68	0.44	16.70-22.70	2.23	17	2.64	0.11	2.20-3.00	4.23	17	7.90	0.12	6.70-9.10	1.52	17	19.56	0.41	16.70-22.50	2.04	17	2.69	0.10	2.30-3.10	3.70
PG: Sysmex poch-100i	28	19.93	0.74	16.90-22.90	3.72	28	2.63	0.07	2.20-3.00	2.69	28	7.90	0.22	6.70-9.10	2.78	28	19.78	0.48	16.80-22.80	2.42	28	2.60	0.12	2.20-3.00	4.62
SG: Select Sysmex Instruments (3-part Diff)	121	20.30	1.84	17.30-23.30	9.06	120	2.73	0.13	2.30-3.10	4.81	120	8.13	0.29	6.90-9.30	3.58	121	20.14	1.21	17.10-23.10	6.02	121	2.73	0.14	2.30-3.10	5.19
PG: Sysmex XP-300	76	20.57	2.23	17.50-23.70	10.83	75	2.79	0.12	2.40-3.20	4.29	75	8.27	0.25	7.10-9.50	3.01	76	20.40	1.43	17.30-23.50	7.01	76	2.79	0.12	2.40-3.20	4.29
SG: International Sysmex Instruments	32	20.00	0.72	17.00-23.00	3.60	32	2.63	0.07	2.20-3.00	2.69	32	7.91	0.21	6.70-9.10	2.66	32	19.83	0.46	16.80-22.80	2.32	32	2.61	0.12	2.20-3.00	4.62
Neutrophil	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex K-800,1000,4500,KX-21	17	24.97	0.90	22.30-27.70	3.60	17	72.04	1.52	67.40-76.60	2.11	17	52.76	0.55	51.10-54.40	1.04	17	24.95	0.58	23.20-26.80	2.36	17	71.12	1.52	66.50-75.70	2.14
PG: Sysmex poch-100i	26	25.21	0.73	23.00-27.40	2.94	27	74.16	1.22	70.50-77.90	1.64	27	54.63	1.21	51.00-58.20	2.22	27	25.37	0.58	23.70-27.10	2.28	27	74.15	1.43	69.80-78.40	1.93
SG: Select Sysmex Instruments (3-part Diff)	116	24.85	0.87	22.30-27.50	3.49	118	71.90	2.75	63.70-80.10	3.81	118	53.41	2.66	45.40-61.40	4.98	117	24.91	0.70	22.80-27.00	2.81	118	71.76	2.03	65.70-77.90	2.83
PG: Sysmex XP-300	73	24.70	0.88	22.10-27.30	3.56	74	71.04	2.90	62.30-79.70	4.08	74	53.12	3.16	43.60-62.60	5.95	73	24.73	0.69	22.60-26.80	2.79	74	71.04	1.63	66.10-75.90	2.30
SG: International Sysmex Instruments	30	25.24	0.73	23.00-27.40	2.94	31	74.09	1.23	70.40-77.80	1.66	31	54.61	1.27	50.80-58.40	2.33	31	25.43	0.62	23.50-27.30	2.44	31	74.20	1.39	70.00-78.40	1.87
Red Blood Cell Count	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex K-800,1000,4500,KX-21	17	5.65	0.07	5.31-5.99	1.22	17	2.41	0.03	2.27-2.55	1.24	17	4.20	0.04	3.95-4.45	1.00	17	5.64	0.06	5.30-5.98	1.08	17	2.41	0.03	2.27-2.55	1.20
PG: Sysmex poch-100i	28	5.80	0.20	5.45-6.15	3.50	28	2.46	0.05	2.31-2.61	1.91	28	4.33	0.07	4.07-4.59	1.52	28	5.79	0.11	5.44-6.14	1.88	28	2.45	0.06	2.30-2.60	2.57
SG: Select Sysmex Instruments (3-part Diff)	121	5.67	0.35	5.33-6.01	6.14	120	2.42	0.04	2.27-2.57	1.65	121	4.24	0.17	3.99-4.49	4.08	121	5.67	0.16	5.33-6.01	2.89	121	2.42	0.05	2.27-2.57	2.07
PG: Sysmex XP-300	76	5.63	0.41	5.29-5.97	7.32	75	2.41	0.03	2.27-2.55	1.24	76	4.22	0.20	3.97-4.47	4.86	76	5.63	0.17	5.29-5.97	3.11	76	2.41	0.04	2.27-2.55	1.70
SG: International Sysmex Instruments	32	5.80	0.20	5.45-6.15	3.40	32	2.45	0.05	2.30-2.60	1.88	32	4.33	0.07	4.07-4.59	1.59	32	5.78	0.11	5.43-6.13	1.85	32	2.45	0.06	2.30-2.60	2.45
Hemoglobin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex K-800,1000,4500,KX-21	17	17.96	0.23	16.70-19.30	1.28	17	6.11	0.12	5.70-6.50	1.97	17	12.01	0.11	11.20-12.80	0.92	17	17.96	0.23	16.70-19.30	1.28	17	6.14	0.11	5.70-6.50	1.80
PG: Sysmex poch-100i	28	18.21	0.66	16.90-19.50	3.63	28	6.13	0.13	5.70-6.50	2.13	28	12.05	0.19	11.20-12.80	1.58	28	18.03	0.38	16.70-19.30	2.11	28	6.12	0.17	5.70-6.50	2.79
SG: Select Sysmex Instruments (3-part Diff)	119	18.04	0.43	16.70-19.30	2.39	120	6.12	0.15	5.70-6.50	2.46	121	12.10	0.58	11.30-12.90	4.79	121	17.97	0.58	16.70-19.30	3.22	121	6.12	0.16	5.70-6.50	2.62
PG: Sysmex XP-300	74	17.99	0.34	16.70-19.30	1.89	75	6.11	0.16	5.70-6.50	2.62	76	12.13	0.72	11.30-12.90	5.95	76	17.95	0.69	16.60-19.20	3.85	76	6.12	0.17	5.70-6.50	2.79
SG: International Sysmex Instruments	32	18.24	0.64	16.90-19.50	3.52	32	6.13	0.12	5.70-6.50	1.97	32	12.06	0.19	11.30-12.90	1.57	32	18.03	0.36	16.70-19.30	2.00	32	6.13	0.16	5.70-6.50	2.62
Lymphocyte	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex K-800,1000,4500,KX-21	17	61.64	0.86	59.00-64.20	1.40	17	10.87	1.73	5.70-16.10	15.87	17	30.49	0.76	28.20-32.80	2.49	17	61.78	0.73	59.60-64.00	1.18	17	11.87	1.37	7.80-16.00	11.51
PG: Sysmex poch-100i	27	62.12	0.77	59.80-64.40	1.24	25	10.71	0.99	7.70-13.70	9.25	27	30.03	1.02	26.90-33.10	3.40	27	62.09	0.63	60.20-64.00	1.01	27	10.50	1.34	6.50-14.50	12.76
SG: Select Sysmex Instruments (3-part Diff)	117	61.74	0.93	58.90-64.50	1.49	112	11.22	1.26	7.40-15.00	11.34	118	29.89	0.97	27.00-32.80	3.24	116	61.75	0.81	59.40-64.20	1.31	114	11.27	1.36	7.30-15.40	11.95
PG: Sysmex XP-300	73	61.62	0.96	58.70-64.50	1.57	70	11.48	1.16	8.00-15.00	10.09	74	29.71	0.94	26.90-32.50	3.16	72	61.62	0.86	59.00-64.20	1.40	70	11.42	1.24	7.70-15.10	10.88
SG: International Sysmex Instruments	31	62.10	0.74	59.90-64.30	1.19	29	10.68	0.97	7.80-13.60	9.07	31	30.04	0.99	27.00-33.00	3.30	31	62.03	0.63	60.10-63.90	1.02	31	10.46	1.32	6.50-14.50	12.57
Hematocrit	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Sysmex K-800,1000,4500,KX-21	17	48.10	0.71	45.20-51.00	1.48	17	16.85	0.37	15.80-17.80	2.20	17	32.16	0.56	30.30-34.10	1.74	17	48.09	0.69	45.20-51.00	1.43	17	16.82	0.38	15.80-17.80	2.26
PG: Sysmex poch-100i	28	51.00	1.87	47.90-54.10	3.67	28	18.01	0.52	16.90-19.10	2.89	28	34.60	0.84	32.50-36.70	2.43	28	50.92	1.37	47.80-54.00	2.69	28	18.00	0.53	16.90-19.10	2.94
SG: Select Sysmex Instruments (3-part Diff)	121	48.86	3.18	46.00-51.80	6.50	120	17.22	0.57	16.20-18.20	3.31	120	32.88	1.22	30.90-34.90	3.71	121	48.82	1.73	45.90-51.70	3.55	121	17.25	0.75	16.30-18.30	4.34
PG: Sysmex XP-300	76	48.24	3.54	45.30-51.10	7.34	75	17.01	0.29	16.00-18.00	1.71	75	32.40	0.76	30.50-34.30	2.35	76	48.21	1.37	45.30-51.10	2.84	76	17.07	0.69	16.10-18.10	4.04
SG: International Sysmex Instruments	32	50.96	1.86	47.90-54.10	3.65	32	17.98	0.50	16.90-19.10	2.78	32	34.52	0.85	32.40-36.60	2.										



MPV	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Intl Sysmex poch-100i	5	10.13	0.31	9.20-11.00	3.07	5	10.37	0.46	9.00-11.80	4.42	5	10.03	0.31	9.10-10.90	3.10	5	9.93	0.38	8.80-11.00	3.84	5	10.27	0.23	9.60-11.00	2.23
708 - Tumor Markers	TM-1					TM-2					TM-3					TM-4					TM-5				
CA 125	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	6	99.63	27.33	17.60-181.60	27.44	6	144.28	35.27	38.50-250.10	24.44	6	11.55	2.59	3.80-19.40	22.33	6	54.40	13.43	14.10-94.70	24.67	6	188.88	44.09	56.60-321.20	23.34
PSA - Free	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	5	4.15	0.44	2.84-5.46	10.52	5	5.38	0.66	3.42-7.35	12.17	5	1.79	0.20	1.21-2.38	10.90	5	2.87	0.38	1.73-4.00	13.16	5	6.62	0.70	4.51-8.71	10.58
PSA	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	6	4.79	1.75	0.00-10.04	36.51	6	6.14	2.44	0.00-13.46	39.72	6	2.13	0.40	0.94-3.32	18.69	6	3.32	0.98	0.39-6.25	29.40	6	7.57	3.21	0.00-17.21	42.47
705 - Urine Microalbumin & Creatinine	UM-1					UM-2					UM-3					UM-4					UM-5				
Urine Creatinine	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Siemens Clinitek 50	7	271.43	48.80	125.00-417.00	18.01																				
PG: Siemens Clinitek Status	24	268.00	47.61	125.00-411.00	17.76																				
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	8	203.34	12.93	164.00-242.00	6.35																				
PG: Beckman (Olympus) AU	7	220.54	6.52	202.00-241.00	2.94																				
SG: Select Beckman Coulter Instruments (Chem)	8	220.60	6.04	203.00-239.00	2.71																				
SG: Semi-quant Methods - Ur Microalbumin-Creat	34	266.90	46.72	127.00-407.00	17.49																				
PG: Siemens (Bayer) DCA Vantage	17	224.80	11.28	191.00-259.00	5.02																				
SG: Select Siemens (Bayer) DCA Instruments	19	224.38	10.71	192.00-256.00	4.78																				
PG: Afinion ACR	7	213.20	6.65	193.00-233.00	3.10																				
AG: All Group	79	239.00	40.59	117.00-361.00	16.99																				
Urine Microalbumin	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Siemens Clinitek 50	7	24.29	9.76	0.00-53.00	40.83																				
PG: Siemens Clinitek Status	25	26.92	7.36	5.00-49.00	27.41																				
PG: Siemens Dimension/AR/ES/RxL/Xpand/EXL	8	33.94	1.74	29.00-39.00	5.00																				
PG: Beckman (Olympus) AU	6	33.75	3.74	23.00-45.00	10.88																				
SG: Select Beckman Coulter Instruments (Chem)	7	34.07	3.52	24.00-45.00	10.29																				
PG: Siemens (Bayer) DCA Vantage	17	33.27	3.25	23.00-43.00	9.70																				
SG: Select Siemens (Bayer) DCA Instruments	19	33.25	3.10	24.00-42.00	9.39																				
PG: Afinion ACR	7	39.96	3.70	29.00-51.00	9.25																				
SG: Select Gen Chem Instruments	12	32.78	3.77	22.00-44.00	11.52																				
SG: Select Siemens Clinitek/Visual	33	26.47	7.74	3.00-49.00	29.62																				
AG: All Group	77	32.49	5.03	17.00-47.00	15.63																				
728 - Urine Chemistry	URC-1					URC-2					URC-3					URC-4					URC-5				
Total Protein - Urine	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
AG: All Group	5	57.70	1.66	52.70-62.70	2.88	5	14.35	0.82	11.90-16.90	5.69															
601 - Waived/PPM Tests II;609 - Waived/PPM Tests I;664 - Whole Blood Glucose;669 - Whole Blood Glucose	WBG-1					WBG-2					WBG-3					WBG-4					WBG-5				
Glucose Wb Blood (Waived)	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%	Labs	Mean	SD	Range	CV.%
PG: Abbott Freestyle Lite	12	168.67	8.28	144.00-194.00	4.91																				
PG: Bayer Ascensia Contour	112	148.37	13.53	108.00-189.00	9.12	15	175.40	10.74	143.00-207.00	6.11															
PG: Bayer Ascensia Elite XL	9	143.56	14.34	101.00-187.00	9.93																				
PG: Lifescan One Touch Ultra	32	187.78	26.51	109.00-268.00	14.10																				
PG: Lifescan Sure Step, others	5	200.00	11.47	166.00-235.00	5.75																				
PG: McKesson/SunMark True Track	14	396.43	26.70	316.00-476.00	6.74	5	401.00	0.00	401.00-401.00	0.00															
PG: Roche Accu-Chek Advantage/Comp/GTS	24	173.81	18.87	117.00-231.00	10.86																				
PG: Roche AccuChek Aviva	32	183.31	10.88	150.00-216.00	5.96																				
PG: Roche Accu-Chek, others	17	177.82	19.78	119.00-237.00	11.12																				
SG: Select Lifescan One Touch Instruments	41	185.27	26.02	107.00-263.00	14.05																				
SG: Select Abbott Freestyle Instruments	15	169.87	8.39	145.00-195.00	4.94																				
PG: Roche AccuChek Inform	66	185.29	10.74	153.00-217.00	5.78	36	205.97	4.25	193.00-219.00	2.04															
PG: Consult Diag Quintet Blood Glucose	25	181.40	17.86	127.00-235.00	9.89	8	211.00	8.04	187.00-235.00	3.79															
PG: Medisense Precision Xceed Pro	26	157.46	30.09	67.00-247.00	19.17																				
SG: Select Medisense Instruments	34	160.59	32.09	65.00-257.00	19.94																				
SG: Select Roche WBG Meters	291	182.87	13.51	143.00-224.00	7.38	52	205.50	8.22	181.00-231.00	3.98	6	74.50	1.38	71.00-79.00	1.87	6	112.50	2.59	105.00-121.00	2.30	6	335.67	5.57	319.00-353.00	1.67
SG: Select Bayer WBG Meters	141	149.50	16.26	101.00-199.00	10.87	20	177.50	14.77	134.00-222.00	8.31															
SG: Select True/True Track WBG Instruments	18	402.28	26.34	323.00-481.00	6.54																				
PG: Nova Biomedical StatStrip Glucose meter	5	160.20	6.76	140.00-180.00	4.25																				
SG: Select Nova Biomedical StatStrip Instruments	9	158.33	6.29	139.00-177.00	3.99																				
PG: McKesson TrueResult	6	170.50	17.16	119.00-223.00	10.06																				
PG: Bayer Contour Next	12	148.33	21.22	84.00-212.00	14.32																				
PG: Roche AccuChek Performa	146	184.17	12.54	147.00-222.00	6.79	6	203.00	8.85	176.00-230.00	4.38															
SG: Select TrueResult Meters	9	168.89	15.04	124.00-214.00	8.88																				
PG: True Metrix Blood Glucose Meter	6	162.83	9.79	134.00-192.00	6.01																				
PG: McKesson True Metrix Pro Blood Glucose Meter</																									



789 - Bacterial Antigen Detection	Labs	Response	Number	Percent
C. difficile Toxin Detect				
AD-1				
* Denotes Correct Response				
AG: All Methods	2			
		*Negative	2	100.00%
Alere C diff Quik Chek Complete	2	*Negative	2	100.00%
AD-2				
* Denotes Correct Response				
AG: All Methods	2			
		*Negative	2	100.00%
Alere C diff Quik Chek Complete	2	*Negative	2	100.00%
AD-3				
* Denotes Correct Response				
AG: All Methods	2			
		*Positive	2	100.00%
Alere C diff Quik Chek Complete	2	*Positive	2	100.00%
AD-4				
* Denotes Correct Response				
AG: All Methods	2			
		*Negative	2	100.00%
Alere C diff Quik Chek Complete	2	*Negative	2	100.00%
AD-5				
* Denotes Correct Response				
AG: All Methods	2			
		*Positive	2	100.00%
Alere C diff Quik Chek Complete	2	*Positive	2	100.00%
Strep B Ag Detect				
AD-1				
* Denotes Correct Response				
AG: All Methods	1			
		*Negative	1	100.00%
Cepheid GeneXpert	1	*Negative	1	100.00%
AD-2				
* Denotes Correct Response				
AG: All Methods	1			
		*Positive	1	100.00%
Cepheid GeneXpert	1	*Positive	1	100.00%
AD-3				
* Denotes Correct Response				
AG: All Methods	1			
		*Negative	1	100.00%
Cepheid GeneXpert	1	*Negative	1	100.00%
AD-4				
* Denotes Correct Response				
AG: All Methods	1			
		*Positive	1	100.00%
Cepheid GeneXpert	1	*Positive	1	100.00%
AD-5				
* Denotes Correct Response				
AG: All Methods	1			
		*Negative	1	100.00%
Cepheid GeneXpert	1	*Negative	1	100.00%
760 - ANA (ELISA)	Labs	Response	Number	Percent
ANA (ELISA) (International)				
AE-1				
* Denotes Correct Response				
AG: All Methods	1			
		*Positive	1	100.00%
Dynex DS2	1	*Positive	1	100.00%
AE-2				
* Denotes Correct Response				
AG: All Methods	1			
		*Negative	1	100.00%
Dynex DS2	1	*Negative	1	100.00%
AE-3				
* Denotes Correct Response				
AG: All Methods	1			
		*Negative	1	100.00%
Dynex DS2	1	*Negative	1	100.00%
AE-4				
* Denotes Correct Response				
AG: All Methods	1			
		*Positive	1	100.00%
Dynex DS2	1	*Positive	1	100.00%
AE-5				
* Denotes Correct Response				
AG: All Methods	1			
		*Negative	1	100.00%
Dynex DS2	1	*Negative	1	100.00%
765 - Allergen Testing	Labs	Response	Number	Percent



		Ash - White Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	2			
		*0	1	50.00%
		*1/0	1	50.00%
Hitachi CLA-1	2	*0	1	50.00%
Hitachi CLA-1		*1/0	1	50.00%
		Birch/Alder Mix Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	3			
		*4	3	100.00%
Hitachi CLA-1	3	*4	3	100.00%
		Cedar - Mountain Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	3			
		*0	1	33.33%
		*1/0	2	66.67%
Hitachi CLA-1	3	*0	1	33.33%
Hitachi CLA-1		*1/0	2	66.67%
		Cottonwood - Eastern Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	3			
		*0	2	66.67%
		*1/0	1	33.33%
Hitachi CLA-1	3	*0	2	66.67%
Hitachi CLA-1		*1/0	1	33.33%
		Eim - White Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	3			
		*0	3	100.00%
Hitachi CLA-1	3	*0	3	100.00%
		Mulberry Mix Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	3			
		*0	3	100.00%
Hitachi CLA-1	3	*0	3	100.00%
		Pine Mix Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	2			
		*0	1	50.00%
		*1/0	1	50.00%
Hitachi CLA-1	2	*0	1	50.00%
Hitachi CLA-1		*1/0	1	50.00%
		Privet Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	2			
		*0	2	100.00%
Hitachi CLA-1	2	*0	2	100.00%
		Sycamore - American Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	2			
		*0	2	100.00%
Hitachi CLA-1	2	*0	2	100.00%
		Cocklebur Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	2			
		*1	1	50.00%
		*1/0	1	50.00%
Hitachi CLA-1	2	*1	1	50.00%
Hitachi CLA-1		*1/0	1	50.00%
		Lambs Quarters Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	2			
		*0	1	50.00%
		*1	1	50.00%
Hitachi CLA-1	2	*0	1	50.00%
Hitachi CLA-1		*1	1	50.00%



		Marshelder - Rough Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	2			
		*1	1	50.00%
		*1/0	1	50.00%
Hitachi CLA-1	2	*1	1	50.00%
Hitachi CLA-1		*1/0	1	50.00%
AL-1		Pigweed Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	3			
		*0	3	100.00%
Hitachi CLA-1	3	*0	3	100.00%
AL-1		Ragweed - Short Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	3			
		*1	2	66.67%
		*1/0	1	33.33%
Hitachi CLA-1	3	*1	2	66.67%
Hitachi CLA-1		*1/0	1	33.33%
AL-1		Sheep Sorrel Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	3			
		*1	3	100.00%
Hitachi CLA-1	3	*1	3	100.00%
AL-1		Waterhemp Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	2			
		*1	2	100.00%
Hitachi CLA-1	2	*1	2	100.00%
AL-1		Mite D Farinae Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	4			
		*4	4	100.00%
Hitachi CLA-1	4	*4	4	100.00%
AL-1		Cladosporium Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	3			
		*0	2	66.67%
		*1/0	1	33.33%
Hitachi CLA-1	3	*0	2	66.67%
Hitachi CLA-1		*1/0	1	33.33%
AL-1		Tomato Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	3			
		*1/0	3	100.00%
Hitachi CLA-1	3	*1/0	3	100.00%
AL-1		Wheat Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	3			
		*0	1	33.33%
		*1/0	2	66.67%
Hitachi CLA-1	3	*0	1	33.33%
Hitachi CLA-1		*1/0	2	66.67%
AL-1		Vegetable Mix Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	2			
		*1	1	50.00%
		*1/0	1	50.00%
Hitachi CLA-1	2	*1	1	50.00%
Hitachi CLA-1		*1/0	1	50.00%
AL-1		Tuna Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	2			
		*0	2	100.00%
Hitachi CLA-1	2	*0	2	100.00%
AL-1		Shellfish Mix Alrqn		
AL-1		* Denotes Correct Response		



AG: All Methods	3			
		*1	2	66.67%
		*1/0	1	33.33%
Hitachi CLA-1	3	*1	2	66.67%
Hitachi CLA-1		*1/0	1	33.33%
Rice Alrgn				
AL-1		* Denotes Correct Response		
AG: All Methods	3			
		*0	1	33.33%
		*1/0	2	66.67%
Hitachi CLA-1	3	*0	1	33.33%
Hitachi CLA-1		*1/0	2	66.67%
Potato Alrgn				
AL-1		* Denotes Correct Response		
AG: All Methods	2			
		*1	1	50.00%
		*1/0	1	50.00%
Hitachi CLA-1	2	*1	1	50.00%
Hitachi CLA-1		*1/0	1	50.00%
Pork Alrgn				
AL-1		* Denotes Correct Response		
AG: All Methods	2			
		*0	1	50.00%
		*1/0	1	50.00%
Hitachi CLA-1	2	*0	1	50.00%
Hitachi CLA-1		*1/0	1	50.00%
Orange Alrgn				
AL-1		* Denotes Correct Response		
AG: All Methods	3			
		*1/0	3	100.00%
Hitachi CLA-1	3	*1/0	3	100.00%
Onion Mix Alrgn				
AL-1		* Denotes Correct Response		
AG: All Methods	2			
		*0	2	100.00%
Hitachi CLA-1	2	*0	2	100.00%
Oat Alrgn				
AL-1		* Denotes Correct Response		
AG: All Methods	2			
		*0	1	50.00%
		*1/0	1	50.00%
Hitachi CLA-1	2	*0	1	50.00%
Hitachi CLA-1		*1/0	1	50.00%
Milk Alrgn				
AL-1		* Denotes Correct Response		
AG: All Methods	3			
		*0	3	100.00%
Hitachi CLA-1	3	*0	3	100.00%
Garlic Alrgn				
AL-1		* Denotes Correct Response		
AG: All Methods	2			
		*1	1	50.00%
		*2	1	50.00%
Hitachi CLA-1	2	*1	1	50.00%
Hitachi CLA-1		*2	1	50.00%
Egg - whole Alrgn				
AL-1		* Denotes Correct Response		
AG: All Methods	3			
		*0	3	100.00%
Hitachi CLA-1	3	*0	3	100.00%
Corn Alrgn				
AL-1		* Denotes Correct Response		
AG: All Methods	3			
		*1	1	33.33%
		*1/0	2	66.67%
Hitachi CLA-1	3	*1	1	33.33%
Hitachi CLA-1		*1/0	2	66.67%



		Chocolate Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	2			
		*0	2	100.00%
Hitachi CLA-1	2	*0	2	100.00%
		Chicken Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	2			
		*0	2	100.00%
Hitachi CLA-1	2	*0	2	100.00%
		Beef Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	2			
		*0	1	50.00%
		*1/0	1	50.00%
Hitachi CLA-1	2	*0	1	50.00%
Hitachi CLA-1		*1/0	1	50.00%
		Bean - White Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	2			
		*1	1	50.00%
		*1/0	1	50.00%
Hitachi CLA-1	2	*1	1	50.00%
Hitachi CLA-1		*1/0	1	50.00%
		Barley Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	2			
		*0	1	50.00%
		*1/0	1	50.00%
Hitachi CLA-1	2	*0	1	50.00%
Hitachi CLA-1		*1/0	1	50.00%
		Apple Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	2			
		*0	2	100.00%
Hitachi CLA-1	2	*0	2	100.00%
		Almond Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	2			
		*0	1	50.00%
		*1/0	1	50.00%
Hitachi CLA-1	2	*0	1	50.00%
Hitachi CLA-1		*1/0	1	50.00%
		Alternaria alternata Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	5			
		*0	1	20.00%
		*1/0	4	80.00%
Hitachi CLA-1	5	*0	1	20.00%
Hitachi CLA-1		*1/0	4	80.00%
		Aspergillus fumigatus Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	4			
		*0	1	25.00%
		*1/0	3	75.00%
Hitachi CLA-1	4	*0	1	25.00%
Hitachi CLA-1		*1/0	3	75.00%
		Bermuda Grass Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	5			
		*3	2	40.00%
		*4	3	60.00%
Hitachi CLA-1	5	*3	2	40.00%
Hitachi CLA-1		*4	3	60.00%
		Cat Epithelium Alrqn		
AL-1		* Denotes Correct Response		
AG: All Methods	4			



		*2	1	25.00%
		*3	1	25.00%
		*4	2	50.00%
Hitachi CLA-1	4	*2	1	25.00%
Hitachi CLA-1		*3	1	25.00%
Hitachi CLA-1		*4	2	50.00%
Dog Epithelium Alrgn				
AL-1		* Denotes Correct Response		
AG: All Methods	3			
		*4	3	100.00%
Hitachi CLA-1	3	*4	3	100.00%
Dog Dander Alrgn				
AL-1		* Denotes Correct Response		
AG: All Methods	3			
		*4	3	100.00%
Hitachi CLA-1	3	*4	3	100.00%
House Dust Mite Alrgn				
AL-1		* Denotes Correct Response		
AG: All Methods	5			
		*2	3	60.00%
		*3	1	20.00%
		*4	1	20.00%
Hitachi CLA-1	5	*2	3	60.00%
Hitachi CLA-1		*3	1	20.00%
Hitachi CLA-1		*4	1	20.00%
English Plantain Alrgn				
AL-1		* Denotes Correct Response		
AG: All Methods	2			
		*1/0	1	50.00%
		*2	1	50.00%
Hitachi CLA-1	2	*1/0	1	50.00%
Hitachi CLA-1		*2	1	50.00%
Oak Alrgn				
AL-1		* Denotes Correct Response		
AG: All Methods	2			
		*3	1	50.00%
		*4	1	50.00%
Hitachi CLA-1	2	*3	1	50.00%
Hitachi CLA-1		*4	1	50.00%
Timothy Grass Alrgn				
AL-1		* Denotes Correct Response		
AG: All Methods	4			
		*4	4	100.00%
Hitachi CLA-1	4	*4	4	100.00%
Peanut (Arachis hyogaeea) Alrgn				
AL-1		* Denotes Correct Response		
PG: Hitachi CLA-1	4			
		*0	1	25.00%
		*1/0	3	75.00%
Cat Dander Alrgn				
AL-1		* Denotes Correct Response		
AG: All Methods	2			
		*4	2	100.00%
Hitachi CLA-1	2	*4	2	100.00%
Soybean Alrgn				
AL-1		* Denotes Correct Response		
AG: All Methods	4			
		*1	1	25.00%
		*1/0	3	75.00%
Hitachi CLA-1	4	*1	1	25.00%
Hitachi CLA-1		*1/0	3	75.00%
Cockroach Mix Alrgn				
AL-1		* Denotes Correct Response		
AG: All Methods	3			
		*0	1	33.33%
		*1	1	33.33%
		*1/0	1	33.33%



Hitachi CLA-1	3	*0	1	33.33%
Hitachi CLA-1		*1	1	33.33%
Hitachi CLA-1		*1/0	1	33.33%
Oak, White Alrqn				
AL-1		* Denotes Correct Response		
AG: All Methods	3			
		*2	1	33.33%
		*3	1	33.33%
		*4	1	33.33%
Hitachi CLA-1	3	*2	1	33.33%
Hitachi CLA-1		*3	1	33.33%
Hitachi CLA-1		*4	1	33.33%
Penicillium Alrqn				
AL-1		* Denotes Correct Response		
AG: All Methods	2			
		*0	1	50.00%
		*1/0	1	50.00%
Hitachi CLA-1	2	*0	1	50.00%
Hitachi CLA-1		*1/0	1	50.00%
Walnut/Hickory/Pecan Mix Alrqn				
AL-1		* Denotes Correct Response		
AG: All Methods	3			
		*0	1	33.33%
		*1	1	33.33%
		*1/0	1	33.33%
Hitachi CLA-1	3	*0	1	33.33%
Hitachi CLA-1		*1	1	33.33%
Hitachi CLA-1		*1/0	1	33.33%
Yeast - Bakers Alrqn				
AL-1		* Denotes Correct Response		
AG: All Methods	3			
		*0	3	100.00%
Hitachi CLA-1	3	*0	3	100.00%
683 - Antistreptolysin O (ASO)				
	Labs	Response	Number	Percent
ASO (Qual) (International)				
AS-1		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Wampole ColorCard ASO	1	*Positive	1	100.00%
AS-2				
AS-2		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Wampole ColorCard ASO	1	*Positive	1	100.00%
AS-3				
AS-3		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Wampole ColorCard ASO	1	*Negative	1	100.00%
AS-4				
AS-4		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Wampole ColorCard ASO	1	*Positive	1	100.00%
AS-5				
AS-5		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Wampole ColorCard ASO	1	*Negative	1	100.00%
975 - Bacterial Vaginosis				
	Labs	Response	Number	Percent
Bacterial Vaginosa (Sialidase)				
BV-1		* Denotes Correct Response		
AG: All Methods	96			
		*Negative	94	97.92%
		Positive	2	2.08%
Sekisui Diagnostics OSOM BVBlue	96	*Negative	94	97.92%
Sekisui Diagnostics OSOM BVBlue		Positive	2	2.08%
BV-2				
BV-2		* Denotes Correct Response		
AG: All Methods	96			
		*Negative	92	95.83%
		Positive	4	4.17%



Sekisui Diagnostics OSOM BVBlue		96	*Negative	92	95.83%
Sekisui Diagnostics OSOM BVBlue			Positive	4	4.17%
661 - Urine Colony Count;784 - Presump ID/Colony Count - Urine;785 - Presump ID/Colony Count-Urine & Strep Combo;775 - Urine Culture;778 - Urine & Throat Cultures;779 - Urine & Genital Cultures;781 - Urine, Throat & Genital Cultures;782 - Complete Bacteri		Labs	Response	Number	Percent
CC-1			Colony Count		
AG: All Methods		33	* Denotes Correct Response		
			*10,000 - 100,000	9	27.27%
			*Greater than 100,000	22	66.67%
			Less than 10,000	1	3.03%
			No growth	1	3.03%
Calibrated Loop - Manual	11		*10,000 - 100,000	2	18.18%
Uricult (Orion)	14		*10,000 - 100,000	6	42.86%
Uricult (Lifesign)	1		*10,000 - 100,000	1	100.00%
Calibrated Loop - Manual	11		*Greater than 100,000	8	72.73%
Uri-Chek (Troy Bio)	6		*Greater than 100,000	6	100.00%
Uricult (Orion)	14		*Greater than 100,000	7	50.00%
Uri-Three (Culture Kits)	1		*Greater than 100,000	1	100.00%
Uricult (Orion)	14		Less than 10,000	1	7.14%
Calibrated Loop - Manual	11		No growth	1	9.09%
784 - Presump ID/Colony Count - Urine;785 - Presump ID/Colony Count-Urine & Strep Combo		Labs	Response	Number	Percent
CC-1			Colony Count		
AG: All Methods		33	* Denotes Correct Response		
			*10,000 - 100,000	9	27.27%
			*Greater than 100,000	22	66.67%
			Less than 10,000	1	3.03%
			No growth	1	3.03%
Calibrated Loop - Manual	11		*10,000 - 100,000	2	18.18%
Uricult (Orion)	14		*10,000 - 100,000	6	42.86%
Uricult (Lifesign)	1		*10,000 - 100,000	1	100.00%
Calibrated Loop - Manual	11		*Greater than 100,000	8	72.73%
Uri-Chek (Troy Bio)	6		*Greater than 100,000	6	100.00%
Uricult (Orion)	14		*Greater than 100,000	7	50.00%
Uri-Three (Culture Kits)	1		*Greater than 100,000	1	100.00%
Uricult (Orion)	14		Less than 10,000	1	7.14%
Calibrated Loop - Manual	11		No growth	1	9.09%
CC-2			* Denotes Correct Response		
AG: All Methods		6			
			*10,000 - 100,000	3	50.00%
			*Less than 10,000	2	33.33%
			No growth	1	16.67%
Calibrated Loop - Manual	5		*10,000 - 100,000	3	60.00%
Calibrated Loop - Manual	5		*Less than 10,000	1	20.00%
Uri-Chek (Troy Bio)	1		*Less than 10,000	1	100.00%
Calibrated Loop - Manual	5		No growth	1	20.00%
CC-3			* Denotes Correct Response		
AG: All Methods		6			
			*Greater than 100,000	6	100.00%
Calibrated Loop - Manual	5		*Greater than 100,000	5	100.00%
Uri-Chek (Troy Bio)	1		*Greater than 100,000	1	100.00%
CC-4			* Denotes Correct Response		
AG: All Methods		6			
			*No growth	6	100.00%
Calibrated Loop - Manual	5		*No growth	5	100.00%
Uri-Chek (Troy Bio)	1		*No growth	1	100.00%
CC-5			* Denotes Correct Response		
AG: All Methods		6			
			*Greater than 100,000	6	100.00%
Calibrated Loop - Manual	5		*Greater than 100,000	5	100.00%
Uri-Chek (Troy Bio)	1		*Greater than 100,000	1	100.00%
CC-1			Urine Presump Id		
AG: All Methods		2	* Denotes Correct Response		
			*Gram Negative Organsim	1	50.00%



		*Presumptive E. Coli	1	50.00%
Calibrated Loop - Manual	2	*Gram Negative Organsim	1	50.00%
Calibrated Loop - Manual		*Presumptive E. Coli	1	50.00%
CC-2		* Denotes Correct Response		
AG: All Methods	2			
		*Gram Positive Organism	1	50.00%
		*Presumptive Proteus sp.	1	50.00%
Calibrated Loop - Manual	2	*Gram Positive Organism	1	50.00%
Calibrated Loop - Manual		*Presumptive Proteus sp.	1	50.00%
CC-3		* Denotes Correct Response		
AG: All Methods	2			
		*Gram Positive Organism	1	50.00%
		Presumptive Proteus sp.	1	50.00%
Calibrated Loop - Manual	2	*Gram Positive Organism	1	50.00%
Calibrated Loop - Manual		Presumptive Proteus sp.	1	50.00%
CC-4		* Denotes Correct Response		
AG: All Methods	2			
		*No Growth	2	100.00%
Calibrated Loop - Manual	2	*No Growth	2	100.00%
CC-5		* Denotes Correct Response		
AG: All Methods	2			
		*Gram Negative Organsim	1	50.00%
		*Presumptive Enterobacter sp.	1	50.00%
Calibrated Loop - Manual	2	*Gram Negative Organsim	1	50.00%
Calibrated Loop - Manual		*Presumptive Enterobacter sp.	1	50.00%
CC-1		Susceptibility Amikacin		
		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
CC-1		Susceptibility Amox/Clavulanate		
		* Denotes Correct Response		
AG: All Methods	5			
		*Intermediate	2	40.00%
		*Sensitive	3	60.00%
Susceptibility	5	*Intermediate	2	40.00%
Susceptibility		*Sensitive	3	60.00%
CC-1		Susceptibility Ampicillin		
		* Denotes Correct Response		
AG: All Methods	6			
		*Resistant	6	100.00%
Susceptibility	6	*Resistant	6	100.00%
CC-1		Susceptibility Ampicillin/Sulbactam		
		* Denotes Correct Response		
AG: All Methods	2			
		*Resistant	2	100.00%
Susceptibility	2	*Resistant	2	100.00%
CC-1		Susceptibility Cefazolin		
		* Denotes Correct Response		
AG: All Methods	2			
		*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
CC-1		Susceptibility Cefdinir		
		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
CC-1		Susceptibility Cefepime		
		* Denotes Correct Response		
AG: All Methods	2			
		*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
CC-1		Susceptibility Cefotaxime		
		* Denotes Correct Response		
AG: All Methods	2			
		*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%



Susceptibility Cefpodoxime				
* Denotes Correct Response				
CC-1				
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
Susceptibility Cefazidime				
* Denotes Correct Response				
CC-1				
AG: All Methods	2			
		*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
Susceptibility Ceftriaxone				
* Denotes Correct Response				
CC-1				
AG: All Methods	4			
		*Sensitive	4	100.00%
Susceptibility	4	*Sensitive	4	100.00%
Susceptibility Cefuroxime				
* Denotes Correct Response				
CC-1				
AG: All Methods	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
Susceptibility Cephalothin				
* Denotes Correct Response				
CC-1				
AG: All Methods	5			
		*Intermediate	2	40.00%
		*Sensitive	3	60.00%
Susceptibility	5	*Intermediate	2	40.00%
Susceptibility		*Sensitive	3	60.00%
Susceptibility Ciprofloxacin				
* Denotes Correct Response				
CC-1				
AG: All Methods	8			
		*Sensitive	8	100.00%
Susceptibility	8	*Sensitive	8	100.00%
Susceptibility Gentamicin				
* Denotes Correct Response				
CC-1				
AG: All Methods	5			
		*Sensitive	5	100.00%
Susceptibility	5	*Sensitive	5	100.00%
Susceptibility Imipenem				
* Denotes Correct Response				
CC-1				
AG: All Methods	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
Susceptibility Levofloxacin				
* Denotes Correct Response				
CC-1				
AG: All Methods	4			
		*Sensitive	4	100.00%
Susceptibility	4	*Sensitive	4	100.00%
Susceptibility Minocycline				
* Denotes Correct Response				
CC-1				
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
Susceptibility Nitrofurantoin				
* Denotes Correct Response				
CC-1				
AG: All Methods	7			
		*Sensitive	7	100.00%
Susceptibility	7	*Sensitive	7	100.00%
Susceptibility Norfloxacin/Noroxin				
* Denotes Correct Response				
CC-1				
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
Susceptibility Piperacillin				
* Denotes Correct Response				
CC-1				
AG: All Methods	1			
		*Resistant	1	100.00%
Susceptibility	1	*Resistant	1	100.00%



Susceptibility Piperacillin/Tazobactam				
CC-1		* Denotes Correct Response		
AG: All Methods	2			
		*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
Susceptibility Sulfamethoxazole				
CC-1		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
Susceptibility Tetracycline				
CC-1		* Denotes Correct Response		
AG: All Methods	5			
		*Sensitive	5	100.00%
Susceptibility	5	*Sensitive	5	100.00%
Susceptibility Tobramycin				
CC-1		* Denotes Correct Response		
AG: All Methods	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
Susceptibility Trimethoprim				
CC-1		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
Susceptibility Trimeth./Sulfameth.				
CC-1		* Denotes Correct Response		
AG: All Methods	6			
		*Sensitive	6	100.00%
Susceptibility	6	*Sensitive	6	100.00%
Susceptibility Ertapenem				
CC-1		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
963 - C. difficile Toxin/Ag (5 specimens);964 - C. difficile Toxin/Ag (2 specimens)	Labs	Response	Number	Percent
CDF-1		C. difficile Antigen Detection		
CDF-1		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Cepheid GeneXpert - Xpert C. diff	1	*Negative	1	100.00%
CDF-2		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Cepheid GeneXpert - Xpert C. diff	1	*Negative	1	100.00%
794 - Vaginal Pathogens II (5 specimens);795 - Vaginal Pathogens II (2 specimens);977 - Trichomonas Ag Det (Wavied)	Labs	Response	Number	Percent
CGT-1		Chlamydia - Non Culture Tech		
CGT-1		* Denotes Correct Response		
AG: All Methods	16			
		*Positive	16	100.00%
BD Probe Tec ET	1	*Positive	1	100.00%
Gen-Probe	1	*Positive	1	100.00%
Quidel QuickVue Chlamydia	2	*Positive	2	100.00%
Cepheid GeneXpert	11	*Positive	11	100.00%
BD Viper XTR	1	*Positive	1	100.00%
CGT-2		* Denotes Correct Response		
AG: All Methods	16			
		*Negative	16	100.00%
BD Probe Tec ET	1	*Negative	1	100.00%
Gen-Probe	1	*Negative	1	100.00%
Quidel QuickVue Chlamydia	2	*Negative	2	100.00%
Cepheid GeneXpert	11	*Negative	11	100.00%
BD Viper XTR	1	*Negative	1	100.00%
CGT-3		* Denotes Correct Response		
AG: All Methods	16			



		*Negative	16	100.00%
BD Probe Tec ET	1	*Negative	1	100.00%
Gen-Probe	1	*Negative	1	100.00%
Quidel QuickVue Chlamydia	2	*Negative	2	100.00%
Cepheid GeneXpert	11	*Negative	11	100.00%
BD Viper XTR	1	*Negative	1	100.00%
CGT-4		* Denotes Correct Response		
AG: All Methods	16			
		*Positive	16	100.00%
BD Probe Tec ET	1	*Positive	1	100.00%
Gen-Probe	1	*Positive	1	100.00%
Quidel QuickVue Chlamydia	2	*Positive	2	100.00%
Cepheid GeneXpert	11	*Positive	11	100.00%
BD Viper XTR	1	*Positive	1	100.00%
CGT-5		* Denotes Correct Response		
AG: All Methods	16			
		*Negative	16	100.00%
BD Probe Tec ET	1	*Negative	1	100.00%
Gen-Probe	1	*Negative	1	100.00%
Quidel QuickVue Chlamydia	2	*Negative	2	100.00%
Cepheid GeneXpert	11	*Negative	11	100.00%
BD Viper XTR	1	*Negative	1	100.00%
		GC - Non Culture Tech		
CGT-1		* Denotes Correct Response		
AG: All Methods	14			
		*Negative	14	100.00%
BD Probe Tec ET	1	*Negative	1	100.00%
Gen-Probe	1	*Negative	1	100.00%
Cepheid GeneXpert	11	*Negative	11	100.00%
BD Viper XTR	1	*Negative	1	100.00%
CGT-2		* Denotes Correct Response		
AG: All Methods	14			
		*Positive	14	100.00%
BD Probe Tec ET	1	*Positive	1	100.00%
Gen-Probe	1	*Positive	1	100.00%
Cepheid GeneXpert	11	*Positive	11	100.00%
BD Viper XTR	1	*Positive	1	100.00%
CGT-3		* Denotes Correct Response		
AG: All Methods	14			
		*Negative	14	100.00%
BD Probe Tec ET	1	*Negative	1	100.00%
Gen-Probe	1	*Negative	1	100.00%
Cepheid GeneXpert	11	*Negative	11	100.00%
BD Viper XTR	1	*Negative	1	100.00%
CGT-4		* Denotes Correct Response		
AG: All Methods	14			
		*Negative	14	100.00%
BD Probe Tec ET	1	*Negative	1	100.00%
Gen-Probe	1	*Negative	1	100.00%
Cepheid GeneXpert	11	*Negative	11	100.00%
BD Viper XTR	1	*Negative	1	100.00%
CGT-5		* Denotes Correct Response		
AG: All Methods	14			
		*Positive	14	100.00%
BD Probe Tec ET	1	*Positive	1	100.00%
Gen-Probe	1	*Positive	1	100.00%
Cepheid GeneXpert	11	*Positive	11	100.00%
BD Viper XTR	1	*Positive	1	100.00%
		T. vaginalis - Non Culture Tech		
CGT-1		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Cepheid GeneXpert	1	*Positive	1	100.00%
CGT-2		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Cepheid GeneXpert	1	*Positive	1	100.00%
CGT-3		* Denotes Correct Response		



AG: All Methods	1			
		*Negative	1	100.00%
Cepheid GeneXpert	1	*Negative	1	100.00%
CGT-4		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Cepheid GeneXpert	1	*Positive	1	100.00%
CGT-5		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Cepheid GeneXpert	1	*Negative	1	100.00%
		Trichomonas Ag Det (Waived)		
CGT-1		* Denotes Correct Response		
AG: All Methods	95			
		*Positive	86	90.53%
		Negative	9	9.47%
Sekisui Diagnostics OSOM Trichomonas	95	*Positive	86	90.53%
Sekisui Diagnostics OSOM Trichomonas		Negative	9	9.47%
CGT-2		* Denotes Correct Response		
AG: All Methods	95			
		*Positive	88	92.63%
		Negative	7	7.37%
Sekisui Diagnostics OSOM Trichomonas	95	*Positive	88	92.63%
Sekisui Diagnostics OSOM Trichomonas		Negative	7	7.37%
639 - Complete Urinalysis Package;638 - Urine Sediment;641 - Clinical Microscopy;601 - Waived/PPM Tests II;609 - Waived/PPM Tests I	Labs	Response	Number	Percent
CM-1		* Denotes Correct Response		
AG: All Methods	658			
		*Red blood cell (RBC) (Erythrocyte)	645	98.02%
		Granular cast	2	0.30%
		RBC/Blood/Hgb cast	5	0.76%
		Renal tubular epithelial (RTE)	2	0.30%
		White blood cell (WBC) (Leukocyte)	4	0.61%
Microscopy	658	*Red blood cell (RBC) (Erythrocyte)	645	98.02%
Microscopy		Granular cast	2	0.30%
Microscopy		RBC/Blood/Hgb cast	5	0.76%
Microscopy		Renal tubular epithelial (RTE)	2	0.30%
Microscopy		White blood cell (WBC) (Leukocyte)	4	0.61%
CM-2		* Denotes Correct Response		
AG: All Methods	652			
		*RBC/Blood/Hgb cast	574	88.04%
		Ammonium biurate crystal	1	0.15%
		Cellular (RTE) cast	14	2.15%
		Cellular (WBC) cast	7	1.07%
		Fatty (fat globules) cast	8	1.23%
		Granular cast	31	4.75%
		Hyaline cast	6	0.92%
		Ident Unknown - Referred	1	0.15%
		Red blood cell (RBC) (Erythrocyte)	7	1.07%
		Waxy cast	1	0.15%
		White blood cell (WBC) (Leukocyte)	2	0.31%
Microscopy	652	*RBC/Blood/Hgb cast	574	88.04%
Microscopy		Ammonium biurate crystal	1	0.15%
Microscopy		Cellular (RTE) cast	14	2.15%
Microscopy		Cellular (WBC) cast	7	1.07%
Microscopy		Fatty (fat globules) cast	8	1.23%
Microscopy		Granular cast	31	4.75%
Microscopy		Hyaline cast	6	0.92%
Microscopy		Ident Unknown - Referred	1	0.15%
Microscopy		Red blood cell (RBC) (Erythrocyte)	7	1.07%
Microscopy		Waxy cast	1	0.15%
Microscopy		White blood cell (WBC) (Leukocyte)	2	0.31%
CM-3		* Denotes Correct Response		
AG: All Methods	652			
		*Ammonium biurate crystal	584	89.57%
		*Ident Unknown - Referred	12	1.84%



		Ammonium magnesium phosphate (Triple phosphate)	1	0.15%
		Amorphous urates/phosphates	10	1.53%
		Bacteria	1	0.15%
		Bilirubin crystal	8	1.23%
		Calcium oxalate crystal	3	0.46%
		Calcium phosphate crystal	4	0.61%
		Fat droplets or globules	1	0.15%
		Fatty (fat globules) cast	1	0.15%
		Fiber (exogenous)/fecal contamination	7	1.07%
		Leucine crystal	1	0.15%
		Pollen grain	7	1.07%
		Spermatozoa present	1	0.15%
		Starch granules	1	0.15%
		Sulfonamide crystal	2	0.31%
		Tyrosine crystal	5	0.77%
		Uric acid crystal	2	0.31%
		Yeast/fungi present	1	0.15%
Microscopy	652	*Ammonium biurate crystal	584	89.57%
Microscopy		*Ident Unknown - Referred	12	1.84%
Microscopy		Ammonium magnesium phosphate (Triple phosphate)	1	0.15%
Microscopy		Amorphous urates/phosphates	10	1.53%
Microscopy		Bacteria	1	0.15%
Microscopy		Bilirubin crystal	8	1.23%
Microscopy		Calcium oxalate crystal	3	0.46%
Microscopy		Calcium phosphate crystal	4	0.61%
Microscopy		Fat droplets or globules	1	0.15%
Microscopy		Fatty (fat globules) cast	1	0.15%
Microscopy		Fiber (exogenous)/fecal contamination	7	1.07%
Microscopy		Leucine crystal	1	0.15%
Microscopy		Pollen grain	7	1.07%
Microscopy		Spermatozoa present	1	0.15%
Microscopy		Starch granules	1	0.15%
Microscopy		Sulfonamide crystal	2	0.31%
Microscopy		Tyrosine crystal	5	0.77%
Microscopy		Uric acid crystal	2	0.31%
Microscopy		Yeast/fungi present	1	0.15%
CM-4		* Denotes Correct Response		
AG: All Methods	652			
		*Ammonium magnesium phosphate (Triple phosphate)	626	96.01%
		Ammonium biurate crystal	1	0.15%
		Amorphous urates/phosphates	2	0.31%
		Calcium oxalate crystal	11	1.69%
		Calcium phosphate crystal	4	0.61%
		Cystine crystal	1	0.15%
		Fat droplets or globules	1	0.15%
		Hippuric acid	2	0.31%
		Uric acid crystal	4	0.61%
Microscopy	652	*Ammonium magnesium phosphate (Triple phosphate)	626	96.01%
Microscopy		Ammonium biurate crystal	1	0.15%
Microscopy		Amorphous urates/phosphates	2	0.31%
Microscopy		Calcium oxalate crystal	11	1.69%
Microscopy		Calcium phosphate crystal	4	0.61%
Microscopy		Cystine crystal	1	0.15%
Microscopy		Fat droplets or globules	1	0.15%
Microscopy		Hippuric acid	2	0.31%
Microscopy		Uric acid crystal	4	0.61%
CM-5		* Denotes Correct Response		
AG: All Methods	941			
		*Trichomonas vaginalis	923	98.09%
		Clue cell	4	0.43%
		Red Blood Cell (RBC) (Erythrocyte)	1	0.11%
		Spermatozoa present	4	0.43%
		Squamous epithelial cell	7	0.74%
		White Blood Cell (WBC) (Leukocyte)	1	0.11%
		Yeast/fungi present	1	0.11%
Microscopy	941	*Trichomonas vaginalis	923	98.09%
Microscopy		Clue cell	4	0.43%
Microscopy		Red Blood Cell (RBC) (Erythrocyte)	1	0.11%



Microscopy		Spermatozoa present	4	0.43%
Microscopy		Squamous epithelial cell	7	0.74%
Microscopy		White Blood Cell (WBC) (Leukocyte)	1	0.11%
Microscopy		Yeast/fungi present	1	0.11%
CM-6		* Denotes Correct Response		
AG: All Methods	941			
		*Squamous epithelial cell	662	70.35%
		Bacteria absent	1	0.11%
		Bacteria present	4	0.43%
		Clue cell	263	27.95%
		Trichomonas vaginalis	6	0.64%
		White Blood Cell (WBC) (Leukocyte)	4	0.43%
		Yeast/fungi present	1	0.11%
Microscopy	941	*Squamous epithelial cell	662	70.35%
Microscopy		Bacteria absent	1	0.11%
Microscopy		Bacteria present	4	0.43%
Microscopy		Clue cell	263	27.95%
Microscopy		Trichomonas vaginalis	6	0.64%
Microscopy		White Blood Cell (WBC) (Leukocyte)	4	0.43%
Microscopy		Yeast/fungi present	1	0.11%
CM-7		* Denotes Correct Response		
AG: All Methods	602			
		*No scabies present	34	5.65%
		*Yeast/fungi absent	523	86.88%
		Scabies present	12	1.99%
		Yeast/fungi present	33	5.48%
Microscopy	602	*No scabies present	34	5.65%
Microscopy		*Yeast/fungi absent	523	86.88%
Microscopy		Scabies present	12	1.99%
Microscopy		Yeast/fungi present	33	5.48%
CM-8		* Denotes Correct Response		
AG: All Methods	301			
		*No pinworm/pinworm eggs present	276	91.69%
		Pinworm/pinworm eggs present	25	8.31%
Microscopy	301	*No pinworm/pinworm eggs present	276	91.69%
Microscopy		Pinworm/pinworm eggs present	25	8.31%
CM-9		* Denotes Correct Response		
AG: All Methods	378			
		*Positive Fern Test	374	98.94%
		Negative Fern Test	4	1.06%
Microscopy	378	*Positive Fern Test	374	98.94%
Microscopy		Negative Fern Test	4	1.06%
683 - Antistreptolysin O (ASO);684 - C-Reactive Protein	Labs	Response	Number	Percent
CR-1		CRP Qualitative		
CR-1		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Inverness Immunex	1	*Positive	1	100.00%
684 - C-Reactive Protein	Labs	Response	Number	Percent
CR-1		CRP Qualitative		
CR-1		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Inverness Immunex	1	*Positive	1	100.00%
CR-2		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Inverness Immunex	1	*Positive	1	100.00%
772 - Dermatophyte Screen	Labs	Response	Number	Percent
DM-1		Dermatophyte Screen		
DM-1		* Denotes Correct Response		
AG: All Methods	13			
		*Positive, Dermatophytes Present	2	15.38%
		Negative, No Dermatophytes Present	11	84.62%
Dermatophytes	13	*Positive, Dermatophytes Present	2	15.38%



777 - Genital Culture; 796 - Genital Culture (2 specimens)	Labs	Response	Number	Percent
Dermatophytes				
		Negative, No Dermatophytes Present	11	84.62%
DM-2				
* Denotes Correct Response				
AG: All Methods			13	
		*Negative, No Dermatophytes Present	8	61.54%
		Positive, Dermatophytes Present	5	38.46%
Dermatophytes		*Negative, No Dermatophytes Present	8	61.54%
Dermatophytes		Positive, Dermatophytes Present	5	38.46%
DM-3				
* Denotes Correct Response				
AG: All Methods			13	
		*Negative, No Dermatophytes Present	8	61.54%
		Positive, Dermatophytes Present	5	38.46%
Dermatophytes		*Negative, No Dermatophytes Present	8	61.54%
Dermatophytes		Positive, Dermatophytes Present	5	38.46%
DM-4				
* Denotes Correct Response				
AG: All Methods			13	
		*Positive, Dermatophytes Present	11	84.62%
		Negative, No Dermatophytes Present	2	15.38%
Dermatophytes		*Positive, Dermatophytes Present	11	84.62%
Dermatophytes		Negative, No Dermatophytes Present	2	15.38%
DM-5				
* Denotes Correct Response				
AG: All Methods			13	
		*Positive, Dermatophytes Present	10	76.92%
		Negative, No Dermatophytes Present	3	23.08%
Dermatophytes		*Positive, Dermatophytes Present	10	76.92%
Dermatophytes		Negative, No Dermatophytes Present	3	23.08%
777 - Genital Culture; 796 - Genital Culture (2 specimens)				
GC-1				
* Denotes Correct Response				
AG: All Methods			3	
		*No growth, GC selective media	3	100.00%
Cultures		*No growth, GC selective media	3	100.00%
GC-2				
* Denotes Correct Response				
AG: All Methods			3	
		*No growth, GC selective media	3	100.00%
Cultures		*No growth, GC selective media	3	100.00%
GC-3				
* Denotes Correct Response				
AG: All Methods			3	
		*Negative for GC	1	33.33%
		*No growth, GC selective media	2	66.67%
Cultures		*Negative for GC	1	33.33%
Cultures		*No growth, GC selective media	2	66.67%
GC-4				
* Denotes Correct Response				
AG: All Methods			5	
		*Positive for GC	1	20.00%
		*Presumptive positive for GC, GR(-) diplococci, oxidase positive organism	1	20.00%
		*Presumptive positive for GC, growth referred for ID/confirmation	2	40.00%
		Group B Streptococcus present	1	20.00%
Cultures		*Positive for GC	1	20.00%
Cultures		*Presumptive positive for GC, GR(-) diplococci, oxidase positive organism	1	20.00%
Cultures		*Presumptive positive for GC, growth referred for ID/confirmation	2	40.00%
Cultures		Group B Streptococcus present	1	20.00%
GC-5				
* Denotes Correct Response				
AG: All Methods			4	
		*Negative for GC	1	25.00%
		*No growth, GC selective media	3	75.00%
Cultures		*Negative for GC	1	25.00%
Cultures		*No growth, GC selective media	3	75.00%
668 - Gram Stain				
GS-1				
* Denotes Correct Response				
AG: All Methods			8	
		*Gram Positive	8	100.00%
Gram Stain		*Gram Positive	8	100.00%
GS-2				
* Denotes Correct Response				
AG: All Methods			8	



		*Gram Negative	7	87.50%
		Gram Positive	1	12.50%
Gram Stain	8	*Gram Negative	7	87.50%
Gram Stain		Gram Positive	1	12.50%
GS-3		* Denotes Correct Response		
AG: All Methods	8			
		*Gram Negative	7	87.50%
		Gram Positive	1	12.50%
Gram Stain	8	*Gram Negative	7	87.50%
Gram Stain		Gram Positive	1	12.50%
GS-4		* Denotes Correct Response		
AG: All Methods	8			
		*Gram Positive	8	100.00%
Gram Stain	8	*Gram Positive	8	100.00%
GS-5		* Denotes Correct Response		
AG: All Methods	8			
		*Gram Negative	8	100.00%
Gram Stain	8	*Gram Negative	8	100.00%
		Gram Stain Morph		
GS-1		* Denotes Correct Response		
AG: All Methods	7			
		*Cocci	6	85.71%
		*Cocci in Chains	1	14.29%
Gram Stain	7	*Cocci	6	85.71%
Gram Stain		*Cocci in Chains	1	14.29%
GS-2		* Denotes Correct Response		
AG: All Methods	7			
		*Cocci	2	28.57%
		*Cocci in Pairs	1	14.29%
		*Diplococci	3	42.86%
		Cocco-Bacilli	1	14.29%
Gram Stain	7	*Cocci	2	28.57%
Gram Stain		*Cocci in Pairs	1	14.29%
Gram Stain		*Diplococci	3	42.86%
Gram Stain		Cocco-Bacilli	1	14.29%
GS-3		* Denotes Correct Response		
AG: All Methods	7			
		*Rods/Bacilli	7	100.00%
Gram Stain	7	*Rods/Bacilli	7	100.00%
GS-4		* Denotes Correct Response		
AG: All Methods	7			
		*Cocci	4	57.14%
		*Cocci in Chains	2	28.57%
		*Cocci in Pairs	1	14.29%
Gram Stain	7	*Cocci	4	57.14%
Gram Stain		*Cocci in Chains	2	28.57%
Gram Stain		*Cocci in Pairs	1	14.29%
GS-5		* Denotes Correct Response		
AG: All Methods	7			
		*Rods/Bacilli	7	100.00%
Gram Stain	7	*Rods/Bacilli	7	100.00%
678 - Serum hCG	Labs	Response	Number	Percent
		HCG - Qual Serum		
HCG-1		* Denotes Correct Response		
AG: All Methods	86			
		*Positive	86	100.00%
Beck Coulter ICON 25	7	*Positive	7	100.00%
Fisher HC Sure-Vue hCG STAT	2	*Positive	2	100.00%
Sekisui Diagnostics OSOM	2	*Positive	2	100.00%
Henry Schein One Step + (hCG)	5	*Positive	5	100.00%
Immunostics Immuno/hCG	2	*Positive	2	100.00%
Inverness Signify HCG	2	*Positive	2	100.00%
LifeSign Status hCG	1	*Positive	1	100.00%
McKesson Medi-Lab hCG - Urine	1	*Positive	1	100.00%
Quidel QuickVue + hCG	2	*Positive	2	100.00%
Quidel QuickVue One-Step hCG Combo (U/S)	32	*Positive	32	100.00%
Select Diag hCG Combo	3	*Positive	3	100.00%



SP Brand (Cardinal) HCG Combo	9	*Positive	9	100.00%
Wampole Clearview	1	*Positive	1	100.00%
Polymedco Poly stat hCG Serum	1	*Positive	1	100.00%
Stanbio True 20 Plus	1	*Positive	1	100.00%
Consult Diag hCG Combo Cassette	8	*Positive	8	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	1	*Positive	1	100.00%
Medline Pregnancy Combo (serum)	1	*Positive	1	100.00%
Alere hCG Combo Cassette (20/10 mIU/mL)	1	*Positive	1	100.00%
McKesson Medi-Lab hCG - Serum	4	*Positive	4	100.00%
HCG-2		* Denotes Correct Response		
AG: All Methods	86			
		*Positive	86	100.00%
Beck Coulter ICON 25	7	*Positive	7	100.00%
Fisher HC Sure-Vue hCG STAT	2	*Positive	2	100.00%
Sekisui Diagnostics OSOM	2	*Positive	2	100.00%
Henry Schein One Step + (hCG)	5	*Positive	5	100.00%
Immunostics Immuno/hCG	2	*Positive	2	100.00%
Inverness Signify HCG	2	*Positive	2	100.00%
LifeSign Status hCG	1	*Positive	1	100.00%
McKesson Medi-Lab hCG - Urine	1	*Positive	1	100.00%
Quidel QuickVue + hCG	2	*Positive	2	100.00%
Quidel QuickVue One-Step hCG Combo (U/S)	32	*Positive	32	100.00%
Select Diag hCG Combo	3	*Positive	3	100.00%
SP Brand (Cardinal) HCG Combo	9	*Positive	9	100.00%
Wampole Clearview	1	*Positive	1	100.00%
Polymedco Poly stat hCG Serum	1	*Positive	1	100.00%
Stanbio True 20 Plus	1	*Positive	1	100.00%
Consult Diag hCG Combo Cassette	8	*Positive	8	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	1	*Positive	1	100.00%
Medline Pregnancy Combo (serum)	1	*Positive	1	100.00%
Alere hCG Combo Cassette (20/10 mIU/mL)	1	*Positive	1	100.00%
McKesson Medi-Lab hCG - Serum	4	*Positive	4	100.00%
HCG-3		* Denotes Correct Response		
AG: All Methods	86			
		*Negative	86	100.00%
Beck Coulter ICON 25	7	*Negative	7	100.00%
Fisher HC Sure-Vue hCG STAT	2	*Negative	2	100.00%
Sekisui Diagnostics OSOM	2	*Negative	2	100.00%
Henry Schein One Step + (hCG)	5	*Negative	5	100.00%
Immunostics Immuno/hCG	2	*Negative	2	100.00%
Inverness Signify HCG	2	*Negative	2	100.00%
LifeSign Status hCG	1	*Negative	1	100.00%
McKesson Medi-Lab hCG - Urine	1	*Negative	1	100.00%
Quidel QuickVue + hCG	2	*Negative	2	100.00%
Quidel QuickVue One-Step hCG Combo (U/S)	32	*Negative	32	100.00%
Select Diag hCG Combo	3	*Negative	3	100.00%
SP Brand (Cardinal) HCG Combo	9	*Negative	9	100.00%
Wampole Clearview	1	*Negative	1	100.00%
Polymedco Poly stat hCG Serum	1	*Negative	1	100.00%
Stanbio True 20 Plus	1	*Negative	1	100.00%
Consult Diag hCG Combo Cassette	8	*Negative	8	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	1	*Negative	1	100.00%
Medline Pregnancy Combo (serum)	1	*Negative	1	100.00%
Alere hCG Combo Cassette (20/10 mIU/mL)	1	*Negative	1	100.00%
McKesson Medi-Lab hCG - Serum	4	*Negative	4	100.00%
HCG-4		* Denotes Correct Response		
AG: All Methods	86			
		*Negative	86	100.00%
Beck Coulter ICON 25	7	*Negative	7	100.00%
Fisher HC Sure-Vue hCG STAT	2	*Negative	2	100.00%
Sekisui Diagnostics OSOM	2	*Negative	2	100.00%
Henry Schein One Step + (hCG)	5	*Negative	5	100.00%
Immunostics Immuno/hCG	2	*Negative	2	100.00%
Inverness Signify HCG	2	*Negative	2	100.00%
LifeSign Status hCG	1	*Negative	1	100.00%
McKesson Medi-Lab hCG - Urine	1	*Negative	1	100.00%
Quidel QuickVue + hCG	2	*Negative	2	100.00%
Quidel QuickVue One-Step hCG Combo (U/S)	32	*Negative	32	100.00%
Select Diag hCG Combo	3	*Negative	3	100.00%



SP Brand (Cardinal) HCG Combo	9	*Negative	9	100.00%
Wampole Clearview	1	*Negative	1	100.00%
Polymedco Poly stat hCG Serum	1	*Negative	1	100.00%
Stanbio True 20 Plus	1	*Negative	1	100.00%
Consult Diag hCG Combo Cassette	8	*Negative	8	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	1	*Negative	1	100.00%
Medline Pregnancy Combo (serum)	1	*Negative	1	100.00%
Alere hCG Combo Cassette (20/10 mIU/mL)	1	*Negative	1	100.00%
McKesson Medi-Lab hCG - Serum	4	*Negative	4	100.00%
HCG-5		* Denotes Correct Response		
AG: All Methods	86			
		*Positive	86	100.00%
Beck Coulter ICON 25	7	*Positive	7	100.00%
Fisher HC Sure-Vue hCG STAT	2	*Positive	2	100.00%
Sekisui Diagnostics OSOM	2	*Positive	2	100.00%
Henry Schein One Step + (hCG)	5	*Positive	5	100.00%
Immunostics Immuno/hCG	2	*Positive	2	100.00%
Inverness Signify HCG	2	*Positive	2	100.00%
LifeSign Status hCG	1	*Positive	1	100.00%
McKesson Medi-Lab hCG - Urine	1	*Positive	1	100.00%
Quidel QuickVue + hCG	2	*Positive	2	100.00%
Quidel QuickVue One-Step hCG Combo (U/S)	32	*Positive	32	100.00%
Select Diag hCG Combo	3	*Positive	3	100.00%
SP Brand (Cardinal) HCG Combo	9	*Positive	9	100.00%
Wampole Clearview	1	*Positive	1	100.00%
Polymedco Poly stat hCG Serum	1	*Positive	1	100.00%
Stanbio True 20 Plus	1	*Positive	1	100.00%
Consult Diag hCG Combo Cassette	8	*Positive	8	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	1	*Positive	1	100.00%
Medline Pregnancy Combo (serum)	1	*Positive	1	100.00%
Alere hCG Combo Cassette (20/10 mIU/mL)	1	*Positive	1	100.00%
McKesson Medi-Lab hCG - Serum	4	*Positive	4	100.00%
		HCG - Qual Serum (International)		
HCG-1		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Henry Schein One Step + (hCG)	1	*Positive	1	100.00%
SAS Ultra HCG	2	*Positive	2	100.00%
HCG-2		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Henry Schein One Step + (hCG)	1	*Positive	1	100.00%
SAS Ultra HCG	2	*Positive	2	100.00%
HCG-3		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Henry Schein One Step + (hCG)	1	*Negative	1	100.00%
SAS Ultra HCG	2	*Negative	2	100.00%
HCG-4		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Henry Schein One Step + (hCG)	1	*Negative	1	100.00%
SAS Ultra HCG	2	*Negative	2	100.00%
HCG-5		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Henry Schein One Step + (hCG)	1	*Positive	1	100.00%
SAS Ultra HCG	2	*Positive	2	100.00%
933 - HIV Antibody/Antigen Combo (2 specimens);934 - HIV Antibody/Antigen Combo (5 specimens)	Labs	Response	Number	Percent
		HIV Antibody (Waived)		
HIV-1		* Denotes Correct Response		
AG: All Methods	105			
		*Positive	100	95.24%
		Negative	5	4.76%
Alere Clearview Complete HIV 1/2 (mod)	2	*Positive	2	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	48	*Positive	44	91.67%
Orasure OraQuick Anti-HIV	9	*Positive	9	100.00%



Trinity Biotech Uni-Gold Recombigen HIV	3	*Positive	3	100.00%
Alere Determine HIV Ag/Ab	31	*Positive	30	96.77%
Chembio HIV 1/2 STAT-PAK Dipstick	1	*Positive	1	100.00%
Chembio HIV 1/2 STAT-PAK Cassette	8	*Positive	8	100.00%
bioLytical INSTI HIV Antibody	3	*Positive	3	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	48	Negative	4	8.33%
Alere Determine HIV Ag/Ab	31	Negative	1	3.23%
HIV-2		* Denotes Correct Response		
AG: All Methods	105			
		*Negative	92	87.62%
		Positive	13	12.38%
Alere Clearview HIV 1/2 STAT PAK (WB)	48	*Negative	43	89.58%
Orasure OraQuick Anti-HIV	9	*Negative	7	77.78%
Trinity Biotech Uni-Gold Recombigen HIV	3	*Negative	3	100.00%
Alere Determine HIV Ag/Ab	31	*Negative	27	87.10%
Chembio HIV 1/2 STAT-PAK Dipstick	1	*Negative	1	100.00%
Chembio HIV 1/2 STAT-PAK Cassette	8	*Negative	8	100.00%
bioLytical INSTI HIV Antibody	3	*Negative	3	100.00%
Alere Clearview Complete HIV 1/2 (mod)	2	Positive	2	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	48	Positive	5	10.42%
Orasure OraQuick Anti-HIV	9	Positive	2	22.22%
Alere Determine HIV Ag/Ab	31	Positive	4	12.90%
		HIV Antibody (International)		
HIV-1		* Denotes Correct Response		
AG: All Methods	5			
		*Positive	5	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	1	*Positive	1	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Positive	3	100.00%
Intl Chembio HIV 1/2 STAT-PAK Dipstick	1	*Positive	1	100.00%
HIV-2		* Denotes Correct Response		
AG: All Methods	5			
		*Negative	5	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	1	*Negative	1	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
Intl Chembio HIV 1/2 STAT-PAK Dipstick	1	*Negative	1	100.00%
		HIV Antigen (p24) (Waived)		
HIV-1		* Denotes Correct Response		
AG: All Methods	30			
		*Negative	27	90.00%
		Positive	3	10.00%
Alere Determine HIV Ag/Ab	30	*Negative	27	90.00%
Alere Determine HIV Ag/Ab		Positive	3	10.00%
HIV-2		* Denotes Correct Response		
AG: All Methods	30			
		*Positive	27	90.00%
		Negative	3	10.00%
Alere Determine HIV Ag/Ab	30	*Positive	27	90.00%
Alere Determine HIV Ag/Ab		Negative	3	10.00%
934 - HIV Antibody/Antigen Combo (5 specimens)	Labs	Response	Number	Percent
		HIV Antibody (Waived)		
HIV-1		* Denotes Correct Response		
AG: All Methods	105			
		*Positive	100	95.24%
		Negative	5	4.76%
Alere Clearview Complete HIV 1/2 (mod)	2	*Positive	2	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	48	*Positive	44	91.67%
Orasure OraQuick Anti-HIV	9	*Positive	9	100.00%
Trinity Biotech Uni-Gold Recombigen HIV	3	*Positive	3	100.00%
Alere Determine HIV Ag/Ab	31	*Positive	30	96.77%
Chembio HIV 1/2 STAT-PAK Dipstick	1	*Positive	1	100.00%
Chembio HIV 1/2 STAT-PAK Cassette	8	*Positive	8	100.00%
bioLytical INSTI HIV Antibody	3	*Positive	3	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	48	Negative	4	8.33%
Alere Determine HIV Ag/Ab	31	Negative	1	3.23%
HIV-2		* Denotes Correct Response		
AG: All Methods	105			
		*Negative	92	87.62%
		Positive	13	12.38%



Alere Clearview HIV 1/2 STAT PAK (WB)	48	*Negative	43	89.58%
Orasure OraQuick Anti-HIV	9	*Negative	7	77.78%
Trinity Biotech Uni-Gold Recombigen HIV	3	*Negative	3	100.00%
Alere Determine HIV Ag/Ab	31	*Negative	27	87.10%
Chembio HIV 1/2 STAT-PAK Dipstick	1	*Negative	1	100.00%
Chembio HIV 1/2 STAT-PAK Cassette	8	*Negative	8	100.00%
bioLytical INSTI HIV Antibody	3	*Negative	3	100.00%
Alere Clearview Complete HIV 1/2 (mod)	2	Positive	2	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	48	Positive	5	10.42%
Orasure OraQuick Anti-HIV	9	Positive	2	22.22%
Alere Determine HIV Ag/Ab	31	Positive	4	12.90%
HIV-3		* Denotes Correct Response		
AG: All Methods	8			
		*Negative	8	100.00%
Alere Clearview Complete HIV 1/2 (mod)	2	*Negative	2	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	4	*Negative	4	100.00%
Trinity Biotech Uni-Gold Recombigen HIV	1	*Negative	1	100.00%
bioLytical INSTI HIV Antibody	1	*Negative	1	100.00%
HIV-4		* Denotes Correct Response		
AG: All Methods	8			
		*Positive	8	100.00%
Alere Clearview Complete HIV 1/2 (mod)	2	*Positive	2	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	4	*Positive	4	100.00%
Trinity Biotech Uni-Gold Recombigen HIV	1	*Positive	1	100.00%
bioLytical INSTI HIV Antibody	1	*Positive	1	100.00%
HIV-5		* Denotes Correct Response		
AG: All Methods	8			
		*Negative	8	100.00%
Alere Clearview Complete HIV 1/2 (mod)	2	*Negative	2	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	4	*Negative	4	100.00%
Trinity Biotech Uni-Gold Recombigen HIV	1	*Negative	1	100.00%
bioLytical INSTI HIV Antibody	1	*Negative	1	100.00%
		HIV Antibody (International)		
HIV-1		* Denotes Correct Response		
AG: All Methods	5			
		*Positive	5	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	1	*Positive	1	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Positive	3	100.00%
Intl Chembio HIV 1/2 STAT-PAK Dipstick	1	*Positive	1	100.00%
HIV-2		* Denotes Correct Response		
AG: All Methods	5			
		*Negative	5	100.00%
Alere Clearview HIV 1/2 STAT PAK (WB)	1	*Negative	1	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
Intl Chembio HIV 1/2 STAT-PAK Dipstick	1	*Negative	1	100.00%
HIV-3		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
HIV-4		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Positive	3	100.00%
HIV-5		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
		HIV Antigen (p24) (Waived)		
HIV-1		* Denotes Correct Response		
AG: All Methods	30			
		*Negative	27	90.00%
		Positive	3	10.00%
Alere Determine HIV Ag/Ab	30	*Negative	27	90.00%
Alere Determine HIV Ag/Ab		Positive	3	10.00%
HIV-2		* Denotes Correct Response		
AG: All Methods	30			
		*Positive	27	90.00%
		Negative	3	10.00%



Test Name	Labs	Response	Number	Percent
Alere Determine HIV Ag/Ab	30	*Positive	27	90.00%
Alere Determine HIV Ag/Ab		Negative	3	10.00%
644 - H. pylori	Labs	Response	Number	Percent
		H. pylori Ab Detect		
HP-1		* Denotes Correct Response		
AG: All Methods	66			
		*Positive	66	100.00%
Beck Coulter FlexSure	3	*Positive	3	100.00%
Beck Coulter ICON HP	2	*Positive	2	100.00%
McKesson Medi-Lab Perf H. pylori	7	*Positive	7	100.00%
Polymedco Poly stat (WB - Waived)	1	*Positive	1	100.00%
Henry Schein One Step+ (H. pylori)	6	*Positive	6	100.00%
Select Diag H. pylori Cassette	2	*Positive	2	100.00%
Quidel QuickVue H. pylori gII (S/P)	3	*Positive	3	100.00%
Quidel QuickVue H. pylori gII (WB)	9	*Positive	9	100.00%
SP Brand (Cardinal) HP Rapid Test	9	*Positive	9	100.00%
Alere Clearview II S/P (mod)	3	*Positive	3	100.00%
Alere Clearview II WB (waived)	3	*Positive	3	100.00%
LifeSign Status H. pylori	2	*Positive	2	100.00%
Instant-View H. pylori	1	*Positive	1	100.00%
Consult Diag H pylori	14	*Positive	14	100.00%
Jant Accutest H pylori	1	*Positive	1	100.00%
603 - Blood Cell Identification;610 - Regular Hematology	Labs	Response	Number	Percent
		Blood Cell Identification		
ID-1		* Denotes Correct Response		
AG: All Methods	63			
		*Lymphocyte	63	100.00%
Blood Cell Microscopy	63	*Lymphocyte	63	100.00%
ID-2		* Denotes Correct Response		
AG: All Methods	63			
		*Monocyte	62	98.41%
		Segmented or band neutrophil	1	1.59%
Blood Cell Microscopy	63	*Monocyte	62	98.41%
Blood Cell Microscopy		Segmented or band neutrophil	1	1.59%
ID-3		* Denotes Correct Response		
AG: All Methods	63			
		*Eosinophil, any stage	63	100.00%
Blood Cell Microscopy	63	*Eosinophil, any stage	63	100.00%
ID-4		* Denotes Correct Response		
AG: All Methods	63			
		*Basophil, any stage	63	100.00%
Blood Cell Microscopy	63	*Basophil, any stage	63	100.00%
ID-5		* Denotes Correct Response		
AG: All Methods	63			
		*Red Blood Cell, normal (RBC) (Erythrocyte)	61	96.83%
		Macrocyte (oval or round)	1	1.59%
		Polychromatophilic (non-nucleated) red cell	1	1.59%
Blood Cell Microscopy	63	*Red Blood Cell, normal (RBC) (Erythrocyte)	61	96.83%
Blood Cell Microscopy		Macrocyte (oval or round)	1	1.59%
Blood Cell Microscopy		Polychromatophilic (non-nucleated) red cell	1	1.59%
ID-6		* Denotes Correct Response		
AG: All Methods	63			
		*Platelet, normal	62	98.41%
		Other, not listed	1	1.59%
Blood Cell Microscopy	63	*Platelet, normal	62	98.41%
Blood Cell Microscopy		Other, not listed	1	1.59%
		Blood Cell Identification (International)		
ID-1		* Denotes Correct Response		
AG: All Methods	1			
		*Lymphocyte	1	100.00%
Intl Blood Cell Microscopy	1	*Lymphocyte	1	100.00%
ID-2		* Denotes Correct Response		
AG: All Methods	1			
		*Monocyte	1	100.00%
Intl Blood Cell Microscopy	1	*Monocyte	1	100.00%
ID-3		* Denotes Correct Response		



AG: All Methods	1			
		*Eosinophil, any stage	1	100.00%
Intl Blood Cell Microscopy	1	*Eosinophil, any stage	1	100.00%
ID-4		* Denotes Correct Response		
AG: All Methods	1			
		*Basophil, any stage	1	100.00%
Intl Blood Cell Microscopy	1	*Basophil, any stage	1	100.00%
ID-5		* Denotes Correct Response		
AG: All Methods	1			
		*Red Blood Cell, normal (RBC) (Erythrocyte)	1	100.00%
Intl Blood Cell Microscopy	1	*Red Blood Cell, normal (RBC) (Erythrocyte)	1	100.00%
ID-6		* Denotes Correct Response		
AG: All Methods	1			
		*Platelet, normal	1	100.00%
Intl Blood Cell Microscopy	1	*Platelet, normal	1	100.00%
609 - Waived/PPM Tests 1;675 - Infectious Mononucleosis;753 - Infectious Mononucleosis (Waived) (1 specimens);754 - Infectious Mono/RF Combo	Labs	Response	Number	Percent
		Infectious Mono (Waived)		
IM-1		* Denotes Correct Response		
PG: Alere Aceava (WB) Mono Cassette	65			
		*Negative	65	100.00%
IM-1		* Denotes Correct Response		
PG: Sekisui Diagnostics OSOM Mono (WB)	178			
		*Negative	177	99.44%
		Positive	1	0.56%
IM-1		* Denotes Correct Response		
PG: Henry Schein One-Step+ Mono	41			
		*Negative	40	97.56%
		Positive	1	2.44%
IM-1		* Denotes Correct Response		
PG: LifeSign Status Inf Mono	32			
		*Negative	32	100.00%
IM-1		* Denotes Correct Response		
PG: McKesson Medi-Lab Perf - Inf Mono	12			
		*Negative	12	100.00%
IM-1		* Denotes Correct Response		
PG: Quidel QuickVue+ (WB)	13			
		*Negative	13	100.00%
IM-1		* Denotes Correct Response		
PG: SP Brand (Cardinal) Rapid Test	9			
		*Negative	9	100.00%
IM-1		* Denotes Correct Response		
PG: Alere Clearview Mono Cassette	23			
		*Negative	22	95.65%
		Positive	1	4.35%
IM-1		* Denotes Correct Response		
PG: Fisher HC Sure-View Mono	5			
		*Negative	5	100.00%
IM-1		* Denotes Correct Response		
PG: Meridian Immunocard STAT! Mono (WB)	6			
		*Negative	5	83.33%
		Positive	1	16.67%
IM-1		* Denotes Correct Response		
PG: Consult Diag Mono Cassette	50			
		*Negative	49	98.00%
		Positive	1	2.00%
IM-1		* Denotes Correct Response		
AG: All Methods	451			
		*Negative	446	98.89%
		Positive	5	1.11%
Alere Aceava (S)	1	*Negative	1	100.00%
Alere Aceava (WB)	53	*Negative	53	100.00%
Sekisui Diagnostics OSOM Mono (WB)	178	*Negative	177	99.44%
Henry Schein One-Step+ Mono	41	*Negative	40	97.56%
Instant Technologies i-Screen (WB)	1	*Negative	1	100.00%



Inverness Signify Mono	4	*Negative	4	100.00%
LifeSign Status Inf Mono	32	*Negative	32	100.00%
McKesson Medi-Lab Perf - Inf Mono	12	*Negative	12	100.00%
Polymedco Poly stat Mono - (WB)	3	*Negative	3	100.00%
Quidel QuickVue+ (WB)	13	*Negative	13	100.00%
SP Brand (Cardinal) Rapid Test	9	*Negative	9	100.00%
Alere Clearview Mono Cassette	23	*Negative	22	95.65%
Pulse Scientific	1	*Negative	1	100.00%
Diag Test Group Clarity Mono	3	*Negative	3	100.00%
Fisher HC Sure-View Mono	5	*Negative	5	100.00%
Immunostics Inc. ImmunoDetector Mono (WB)	2	*Negative	2	100.00%
Meridian Immunocard STAT! Mono (WB)	6	*Negative	5	83.33%
NDC ProAdvantage Mono	2	*Negative	2	100.00%
Consult Diag Mono Cassette	50	*Negative	49	98.00%
Alere Aceava (WB) Mono Cassette	12	*Negative	12	100.00%
Sekisui Diagnostics OSOM Mono (WB)	178	Positive	1	0.56%
Henry Schein One-Step+ Mono	41	Positive	1	2.44%
Alere Clearview Mono Cassette	23	Positive	1	4.35%
Meridian Immunocard STAT! Mono (WB)	6	Positive	1	16.67%
Consult Diag Mono Cassette	50	Positive	1	2.00%
Infectious Mononucleosis				
IM-1				
* Denotes Correct Response				
AG: All Methods	73			
		*Negative	73	100.00%
Beck Coulter ICON (WB)	1	*Negative	1	100.00%
Alere Aceava (WB)	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	36	*Negative	36	100.00%
LifeSign Status Inf Mono	2	*Negative	2	100.00%
Lifesign Unistep	1	*Negative	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	2	*Negative	2	100.00%
Polymedco Poly stat Mono - (WB)	1	*Negative	1	100.00%
Polymedco Poly stat Mono (S/P)	2	*Negative	2	100.00%
Quidel QuickVue+ (S/P)	2	*Negative	2	100.00%
Quidel QuickVue+ (WB)	8	*Negative	8	100.00%
Seradyn Color Slide II	3	*Negative	3	100.00%
SP Brand (Cardinal) Rapid Test	1	*Negative	1	100.00%
Alere Clearview Mono Cassette	8	*Negative	8	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	1	*Negative	1	100.00%
Consult Diag Mono Cassette	4	*Negative	4	100.00%
675 - Infectious Mononucleosis;754 - Infectious Mono/RF Combo	Labs	Response	Number	Percent
Infectious Mononucleosis				
IM-1				
* Denotes Correct Response				
AG: All Methods	73			
		*Negative	73	100.00%
Beck Coulter ICON (WB)	1	*Negative	1	100.00%
Alere Aceava (WB)	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	36	*Negative	36	100.00%
LifeSign Status Inf Mono	2	*Negative	2	100.00%
Lifesign Unistep	1	*Negative	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	2	*Negative	2	100.00%
Polymedco Poly stat Mono - (WB)	1	*Negative	1	100.00%
Polymedco Poly stat Mono (S/P)	2	*Negative	2	100.00%
Quidel QuickVue+ (S/P)	2	*Negative	2	100.00%
Quidel QuickVue+ (WB)	8	*Negative	8	100.00%
Seradyn Color Slide II	3	*Negative	3	100.00%
SP Brand (Cardinal) Rapid Test	1	*Negative	1	100.00%
Alere Clearview Mono Cassette	8	*Negative	8	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	1	*Negative	1	100.00%
Consult Diag Mono Cassette	4	*Negative	4	100.00%
IM-2				
* Denotes Correct Response				
AG: All Methods	73			
		*Negative	72	98.63%
		Positive	1	1.37%
Beck Coulter ICON (WB)	1	*Negative	1	100.00%
Alere Aceava (WB)	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	36	*Negative	35	97.22%
LifeSign Status Inf Mono	2	*Negative	2	100.00%



Lifesign Unistep	1	*Negative	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	2	*Negative	2	100.00%
Polymedco Poly stat Mono - (WB)	1	*Negative	1	100.00%
Polymedco Poly stat Mono (S/P)	2	*Negative	2	100.00%
Quidel QuickVue+ (S/P)	2	*Negative	2	100.00%
Quidel QuickVue+ (WB)	8	*Negative	8	100.00%
Seradyn Color Slide II	3	*Negative	3	100.00%
SP Brand (Cardinal) Rapid Test	1	*Negative	1	100.00%
Alere Clearview Mono Cassette	8	*Negative	8	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	1	*Negative	1	100.00%
Consult Diag Mono Cassette	4	*Negative	4	100.00%
Sekisui Diagnostics OSOM Mono (WB)	36	Positive	1	2.78%
IM-3		* Denotes Correct Response		
AG: All Methods	73			
		*Positive	73	100.00%
Beck Coulter ICON (WB)	1	*Positive	1	100.00%
Alere Aceava (WB)	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	36	*Positive	36	100.00%
LifeSign Status Inf Mono	2	*Positive	2	100.00%
Lifesign Unistep	1	*Positive	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	2	*Positive	2	100.00%
Polymedco Poly stat Mono - (WB)	1	*Positive	1	100.00%
Polymedco Poly stat Mono (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (WB)	8	*Positive	8	100.00%
Seradyn Color Slide II	3	*Positive	3	100.00%
SP Brand (Cardinal) Rapid Test	1	*Positive	1	100.00%
Alere Clearview Mono Cassette	8	*Positive	8	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	1	*Positive	1	100.00%
Consult Diag Mono Cassette	4	*Positive	4	100.00%
IM-4		* Denotes Correct Response		
AG: All Methods	73			
		*Positive	71	97.26%
		Negative	2	2.74%
Beck Coulter ICON (WB)	1	*Positive	1	100.00%
Alere Aceava (WB)	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	36	*Positive	35	97.22%
LifeSign Status Inf Mono	2	*Positive	2	100.00%
Lifesign Unistep	1	*Positive	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	2	*Positive	2	100.00%
Polymedco Poly stat Mono - (WB)	1	*Positive	1	100.00%
Polymedco Poly stat Mono (S/P)	2	*Positive	1	50.00%
Quidel QuickVue+ (S/P)	2	*Positive	2	100.00%
Quidel QuickVue+ (WB)	8	*Positive	8	100.00%
Seradyn Color Slide II	3	*Positive	3	100.00%
SP Brand (Cardinal) Rapid Test	1	*Positive	1	100.00%
Alere Clearview Mono Cassette	8	*Positive	8	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	1	*Positive	1	100.00%
Consult Diag Mono Cassette	4	*Positive	4	100.00%
Sekisui Diagnostics OSOM Mono (WB)	36	Negative	1	2.78%
Polymedco Poly stat Mono (S/P)	2	Negative	1	50.00%
IM-5		* Denotes Correct Response		
AG: All Methods	73			
		*Negative	73	100.00%
Beck Coulter ICON (WB)	1	*Negative	1	100.00%
Alere Aceava (WB)	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM Mono (WB)	36	*Negative	36	100.00%
LifeSign Status Inf Mono	2	*Negative	2	100.00%
Lifesign Unistep	1	*Negative	1	100.00%
McKesson Medi-Lab Perf - Inf Mono	2	*Negative	2	100.00%
Polymedco Poly stat Mono - (WB)	1	*Negative	1	100.00%
Polymedco Poly stat Mono (S/P)	2	*Negative	2	100.00%
Quidel QuickVue+ (S/P)	2	*Negative	2	100.00%
Quidel QuickVue+ (WB)	8	*Negative	8	100.00%
Seradyn Color Slide II	3	*Negative	3	100.00%
SP Brand (Cardinal) Rapid Test	1	*Negative	1	100.00%
Alere Clearview Mono Cassette	8	*Negative	8	100.00%
Sekisui Diagnostics OSOM Mono Non-Waived	1	*Negative	1	100.00%
Consult Diag Mono Cassette	4	*Negative	4	100.00%



		Infectious Mono (Waived)		
IM-1		* Denotes Correct Response		
PG: Alere Acceava (WB) Mono Cassette	65			
		*Negative	65	100.00%
IM-1		* Denotes Correct Response		
PG: Sekisui Diagnostics OSOM Mono (WB)	178			
		*Negative	177	99.44%
		Positive	1	0.56%
IM-1		* Denotes Correct Response		
PG: Henry Schein One-Step+ Mono	41			
		*Negative	40	97.56%
		Positive	1	2.44%
IM-1		* Denotes Correct Response		
PG: LifeSign Status Inf Mono	32			
		*Negative	32	100.00%
IM-1		* Denotes Correct Response		
PG: McKesson Medi-Lab Perf - Inf Mono	12			
		*Negative	12	100.00%
IM-1		* Denotes Correct Response		
PG: Quidel QuickVue+ (WB)	13			
		*Negative	13	100.00%
IM-1		* Denotes Correct Response		
PG: SP Brand (Cardinal) Rapid Test	9			
		*Negative	9	100.00%
IM-1		* Denotes Correct Response		
PG: Alere Clearview Mono Cassette	23			
		*Negative	22	95.65%
		Positive	1	4.35%
IM-1		* Denotes Correct Response		
PG: Fisher HC Sure-View Mono	5			
		*Negative	5	100.00%
IM-1		* Denotes Correct Response		
PG: Meridian Immunocard STAT! Mono (WB)	6			
		*Negative	5	83.33%
		Positive	1	16.67%
IM-1		* Denotes Correct Response		
PG: Consult Diag Mono Cassette	50			
		*Negative	49	98.00%
		Positive	1	2.00%
IM-1		* Denotes Correct Response		
AG: All Methods	451			
		*Negative	446	98.89%
		Positive	5	1.11%
Alere Acceava (S)	1	*Negative	1	100.00%
Alere Acceava (WB)	53	*Negative	53	100.00%
Sekisui Diagnostics OSOM Mono (WB)	178	*Negative	177	99.44%
Henry Schein One-Step+ Mono	41	*Negative	40	97.56%
Instant Technologies i-Screen (WB)	1	*Negative	1	100.00%
Inverness Signify Mono	4	*Negative	4	100.00%
LifeSign Status Inf Mono	32	*Negative	32	100.00%
McKesson Medi-Lab Perf - Inf Mono	12	*Negative	12	100.00%
Polymedco Poly stat Mono - (WB)	3	*Negative	3	100.00%
Quidel QuickVue+ (WB)	13	*Negative	13	100.00%
SP Brand (Cardinal) Rapid Test	9	*Negative	9	100.00%
Alere Clearview Mono Cassette	23	*Negative	22	95.65%
Pulse Scientific	1	*Negative	1	100.00%
Diag Test Group Clarity Mono	3	*Negative	3	100.00%
Fisher HC Sure-View Mono	5	*Negative	5	100.00%
Immunostics Inc. ImmunoDetector Mono (WB)	2	*Negative	2	100.00%
Meridian Immunocard STAT! Mono (WB)	6	*Negative	5	83.33%
NDC ProAdvantage Mono	2	*Negative	2	100.00%
Consult Diag Mono Cassette	50	*Negative	49	98.00%
Alere Acceava (WB) Mono Cassette	12	*Negative	12	100.00%
Sekisui Diagnostics OSOM Mono (WB)	178	Positive	1	0.56%
Henry Schein One-Step+ Mono	41	Positive	1	2.44%
Alere Clearview Mono Cassette	23	Positive	1	4.35%
Meridian Immunocard STAT! Mono (WB)	6	Positive	1	16.67%
Consult Diag Mono Cassette	50	Positive	1	2.00%



Infectious Mononucleosis (International)				
IM-1				
* Denotes Correct Response				
AG: All Methods	2			
		*Negative	2	100.00%
Alere Clearview Mono Cassette	1	*Negative	1	100.00%
Fisher HC Sure-Vue Mono	1	*Negative	1	100.00%
IM-2				
* Denotes Correct Response				
AG: All Methods	2			
		*Negative	2	100.00%
Alere Clearview Mono Cassette	1	*Negative	1	100.00%
Fisher HC Sure-Vue Mono	1	*Negative	1	100.00%
IM-3				
* Denotes Correct Response				
AG: All Methods	2			
		*Positive	2	100.00%
Alere Clearview Mono Cassette	1	*Positive	1	100.00%
Fisher HC Sure-Vue Mono	1	*Positive	1	100.00%
IM-4				
* Denotes Correct Response				
AG: All Methods	2			
		*Positive	2	100.00%
Alere Clearview Mono Cassette	1	*Positive	1	100.00%
Fisher HC Sure-Vue Mono	1	*Positive	1	100.00%
IM-5				
* Denotes Correct Response				
AG: All Methods	2			
		*Negative	2	100.00%
Alere Clearview Mono Cassette	1	*Negative	1	100.00%
Fisher HC Sure-Vue Mono	1	*Negative	1	100.00%
637 - KOH - Glass Slides (Skin Hair Nail)	Labs	Response	Number	Percent
KOH Preparation				
KOH-1				
* Denotes Correct Response				
AG: All Methods	17			
		*Fungal elements/yeast observed	17	100.00%
KOH Glass Slide	17	*Fungal elements/yeast observed	17	100.00%
KOH-2				
* Denotes Correct Response				
AG: All Methods	17			
		*No fungal elements/yeast observed	16	94.12%
		Fungal elements/yeast observed	1	5.88%
KOH Glass Slide	17	*No fungal elements/yeast observed	16	94.12%
KOH Glass Slide		Fungal elements/yeast observed	1	5.88%
768 - MRSA Molecular	Labs	Response	Number	Percent
MRSA (Molecular)				
MRSA-1				
* Denotes Correct Response				
AG: All Methods	1			
		*Positive	1	100.00%
Cepheid GeneXpert - Xpert MRSA	1	*Positive	1	100.00%
MRSA-2				
* Denotes Correct Response				
AG: All Methods	1			
		*Negative	1	100.00%
Cepheid GeneXpert - Xpert MRSA	1	*Negative	1	100.00%
758 - Mycoplasma Antibody	Labs	Response	Number	Percent
Mycoplasma Antibody				
MY-1				
* Denotes Correct Response				
AG: All Methods	3			
		*Positive	3	100.00%
Meridian ImmunoCard	3	*Positive	3	100.00%
791 - Fecal Occult Blood Add-On;601 - Waived/PPM Tests I;609 - Waived/PPM Tests I;647 - Mini-Waived Package	Labs	Response	Number	Percent
Fecal Occult Blood				
OC-1				
* Denotes Correct Response				
MG: Guaiac methods	717			
		*Negative	653	91.07%
		Positive	64	8.93%
OC-1				
* Denotes Correct Response				
MG: Immunochemical methods	96			
		*Negative	92	95.83%
		Positive	4	4.17%
950 - Post Vasectomy Sperm Analysis	Labs	Response	Number	Percent



		Sperm Count (Post Vasectomy)		
PVAS-1		* Denotes Correct Response		
AG: All Methods	12			
		*Positive for Sperm	12	100.00%
Hemocytometer	3	*Positive for Sperm	3	100.00%
Post Vasectomy Microscopy	9	*Positive for Sperm	9	100.00%
PVAS-2		* Denotes Correct Response		
AG: All Methods	12			
		*Positive for Sperm	7	58.33%
		Negative for Sperm	5	41.67%
Hemocytometer	3	*Positive for Sperm	2	66.67%
Post Vasectomy Microscopy	9	*Positive for Sperm	5	55.56%
Hemocytometer	3	Negative for Sperm	1	33.33%
Post Vasectomy Microscopy	9	Negative for Sperm	4	44.44%
754 - Infectious Mono/RF Combo;653 - Rheumatoid Factor	Labs	Response	Number	Percent
RF-1		* Denotes Correct Response		
AG: All Methods	10			
		*N/A (Negative)	10	100.00%
Siemens RapiTex	2	*N/A (Negative)	2	100.00%
Seradyn Color Slide	2	*N/A (Negative)	2	100.00%
Wampole ColorCard RF	2	*N/A (Negative)	2	100.00%
Wampole Rheumatex	2	*N/A (Negative)	2	100.00%
Biokit Rhemajet RF	1	*N/A (Negative)	1	100.00%
Remel SeraTest RF-Latex	1	*N/A (Negative)	1	100.00%
RF-2		* Denotes Correct Response		
AG: All Methods	10			
		*Positive	10	100.00%
Siemens RapiTex	2	*Positive	2	100.00%
Seradyn Color Slide	2	*Positive	2	100.00%
Wampole ColorCard RF	2	*Positive	2	100.00%
Wampole Rheumatex	2	*Positive	2	100.00%
Biokit Rhemajet RF	1	*Positive	1	100.00%
Remel SeraTest RF-Latex	1	*Positive	1	100.00%
RF-3		* Denotes Correct Response		
AG: All Methods	10			
		*Positive	10	100.00%
Siemens RapiTex	2	*Positive	2	100.00%
Seradyn Color Slide	2	*Positive	2	100.00%
Wampole ColorCard RF	2	*Positive	2	100.00%
Wampole Rheumatex	2	*Positive	2	100.00%
Biokit Rhemajet RF	1	*Positive	1	100.00%
Remel SeraTest RF-Latex	1	*Positive	1	100.00%
RF-4		* Denotes Correct Response		
AG: All Methods	10			
		*Positive	10	100.00%
Siemens RapiTex	2	*Positive	2	100.00%
Seradyn Color Slide	2	*Positive	2	100.00%
Wampole ColorCard RF	2	*Positive	2	100.00%
Wampole Rheumatex	2	*Positive	2	100.00%
Biokit Rhemajet RF	1	*Positive	1	100.00%
Remel SeraTest RF-Latex	1	*Positive	1	100.00%
RF-5		* Denotes Correct Response		
AG: All Methods	10			
		*N/A (Negative)	10	100.00%
Siemens RapiTex	2	*N/A (Negative)	2	100.00%
Seradyn Color Slide	2	*N/A (Negative)	2	100.00%
Wampole ColorCard RF	2	*N/A (Negative)	2	100.00%
Wampole Rheumatex	2	*N/A (Negative)	2	100.00%
Biokit Rhemajet RF	1	*N/A (Negative)	1	100.00%
Remel SeraTest RF-Latex	1	*N/A (Negative)	1	100.00%
645 - Fetal Membrane Rupture	Labs	Response	Number	Percent
ROM-1		* Denotes Correct Response		
AG: All Methods	18			
		*Negative	18	100.00%
AmniSure ROM Test	18	*Negative	18	100.00%



ROM-2		* Denotes Correct Response		
AG: All Methods	18			
		*Positive	18	100.00%
AmniSure ROM Test	18	*Positive	18	100.00%
601 - Waived/PPM Tests II;609 - Waived/PPM Tests I;785 - Presump ID/Colony Count-Urine & Strep Combo;691 - Group A Strep Antigen (5 specimens);699 - Group A Strep Antigen (1 specimen);783 - Group A Strep Antigen/Add-on (2 specimens);665 - Group A Strep An	Labs	Response	Number	Percent
		Strep A Ag Det. (Waived)		
RS-1		* Denotes Correct Response		
AG: All Methods	874			
		*Negative	865	98.97%
		Positive	9	1.03%
Abbott Signify Strep A - Waived	6	*Negative	6	100.00%
Beck Coulter ICON DS Strep	1	*Negative	1	100.00%
Beck Coulter ICON SC Strep	1	*Negative	1	100.00%
Alere Aceava Strep A Dipstick	112	*Negative	111	99.11%
Fisher HC Sure-Vue Strep A (Waived)	8	*Negative	8	100.00%
Sekisui Diagnostics OSOM	42	*Negative	42	100.00%
Sekisui Diagnostics OSOM Ultra Strep	47	*Negative	47	100.00%
Henry Schein OneStep + Rapid Strep Cass	4	*Negative	4	100.00%
Henry Schein OneStep+ Rapid Strep Dip	70	*Negative	70	100.00%
Instant Technologies I Strep	1	*Negative	1	100.00%
Inverness Signify Strep A Dipstick	7	*Negative	7	100.00%
McKesson Strep A Cas	4	*Negative	4	100.00%
McKesson Strep A Dip	34	*Negative	34	100.00%
McKesson Strep A Test - Twist	2	*Negative	2	100.00%
Permaxim Rediscreen Strep A Rapid	1	*Negative	1	100.00%
Polymedco Poly Stat Strep A (W)	1	*Negative	1	100.00%
PSS Rapid Test Strip Strep A	2	*Negative	2	100.00%
Quidel QuickVue Dipstick Strep	90	*Negative	90	100.00%
Quidel QuickVue In-Line - 25 Test	69	*Negative	68	98.55%
Quidel QuickVue+ Strep A	27	*Negative	27	100.00%
Select Diag Strep A Dip	4	*Negative	3	75.00%
SP (Cardinal) Rapid Strep Cassette	5	*Negative	5	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	87	*Negative	87	100.00%
Stanbio QuStick Strep A	1	*Negative	1	100.00%
Alere Clearview Exact II Dip	36	*Negative	35	97.22%
Sekisui Diagnostics OSOM Ultra	5	*Negative	5	100.00%
Consult Diag Strep A Dipstick	116	*Negative	114	98.28%
MooreBrand Strep A Dipstick	5	*Negative	5	100.00%
NDC ProAdvantage Strep Strips	2	*Negative	2	100.00%
Meridian ImmunoCard STAT Strep A	2	*Negative	2	100.00%
LifeSign Bio Strep A Test STRIP	1	*Negative	1	100.00%
Quidel Sofia/Sofia 2 Strep A FIA	15	*Negative	14	93.33%
BD Veritor Strep A	32	*Negative	32	100.00%
Alere i Strep A	8	*Negative	8	100.00%
Medline Strep A Dipstick	2	*Negative	2	100.00%
Roche Cobas Liat Strep A	22	*Negative	22	100.00%
Alere Aceava Strep A Dipstick	112	Positive	1	0.89%
Innovacon One Step Strep Ag	1	Positive	1	100.00%
Quidel QuickVue In-Line - 25 Test	69	Positive	1	1.45%
Select Diag Strep A Dip	4	Positive	1	25.00%
Alere Clearview Exact II Dip	36	Positive	1	2.78%
Consult Diag Strep A Dipstick	116	Positive	2	1.72%
Consult Diag Strep A Twist	1	Positive	1	100.00%
Quidel Sofia/Sofia 2 Strep A FIA	15	Positive	1	6.67%
RS-2		* Denotes Correct Response		
AG: All Methods	57			
		*Negative	57	100.00%
Alere Aceava Strep A Dipstick	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM	10	*Negative	10	100.00%
Sekisui Diagnostics OSOM Ultra Strep	4	*Negative	4	100.00%
Innovacon One Step Strep Ag	1	*Negative	1	100.00%
McKesson Strep A Dip	2	*Negative	2	100.00%
Quidel QuickVue Dipstick Strep	1	*Negative	1	100.00%
Quidel QuickVue In-Line - 25 Test	4	*Negative	4	100.00%
Quidel QuickVue+ Strep A	1	*Negative	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	4	*Negative	4	100.00%



Consult Diag Strep A Dipstick	6	*Negative	6	100.00%
MooreBrand Strep A Dipstick	1	*Negative	1	100.00%
BD Veritor Strep A	1	*Negative	1	100.00%
Alere i Strep A	1	*Negative	1	100.00%
Roche Cobas Liat Strep A	20	*Negative	20	100.00%
		Strep A Aq Det.		
RS-1		* Denotes Correct Response		
AG: All Methods	162			
		*Negative	160	98.77%
		Positive	2	1.23%
Abbott Signify Strep A - Waived	2	*Negative	2	100.00%
BD Check Strep A (Dip)	2	*Negative	2	100.00%
Beck Coulter ICON DS Strep	1	*Negative	1	100.00%
Alere Aceava Strep A Dipstick	19	*Negative	19	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM	15	*Negative	15	100.00%
Sekisui Diagnostics OSOM Ultra Strep	15	*Negative	15	100.00%
Henry Schein OneStep+ Rapid Strep Dip	10	*Negative	10	100.00%
Immunostics Immuno/Strep A Det	2	*Negative	2	100.00%
Inverness Signify Strep A Dipstick	1	*Negative	1	100.00%
McKesson Strep A Cas	2	*Negative	1	50.00%
McKesson Strep A Dip	10	*Negative	10	100.00%
Quidel QuickVue Dipstick Strep	11	*Negative	11	100.00%
Quidel QuickVue In-Line - 25 Test	14	*Negative	14	100.00%
Quidel QuickVue+ Strep A	19	*Negative	19	100.00%
Select Diag Strep A Dip	1	*Negative	1	100.00%
SP (Cardinal) Rapid Strep Cassette	1	*Negative	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	2	*Negative	2	100.00%
Stanbio QuStick Strep A	1	*Negative	1	100.00%
Consult Diag Strep A Dipstick	17	*Negative	17	100.00%
Meridian ImmunoCard STAT Strep A	1	*Negative	1	100.00%
Quidel Sofia/Sofia 2 Strep A FIA	5	*Negative	4	80.00%
BD Veritor Strep A	6	*Negative	6	100.00%
Quidel Solana	3	*Negative	3	100.00%
Cepheid GeneXpert - Xpert Strep A	1	*Negative	1	100.00%
McKesson Strep A Cas	2	Positive	1	50.00%
Quidel Sofia/Sofia 2 Strep A FIA	5	Positive	1	20.00%
RS-2		* Denotes Correct Response		
AG: All Methods	162			
		*Negative	161	99.38%
		Positive	1	0.62%
Abbott Signify Strep A - Waived	2	*Negative	2	100.00%
BD Check Strep A (Dip)	2	*Negative	2	100.00%
Beck Coulter ICON DS Strep	1	*Negative	1	100.00%
Alere Aceava Strep A Dipstick	19	*Negative	19	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM	15	*Negative	15	100.00%
Sekisui Diagnostics OSOM Ultra Strep	15	*Negative	15	100.00%
Henry Schein OneStep+ Rapid Strep Dip	10	*Negative	10	100.00%
Immunostics Immuno/Strep A Det	2	*Negative	2	100.00%
Inverness Signify Strep A Dipstick	1	*Negative	1	100.00%
McKesson Strep A Cas	2	*Negative	2	100.00%
McKesson Strep A Dip	10	*Negative	10	100.00%
Quidel QuickVue Dipstick Strep	11	*Negative	11	100.00%
Quidel QuickVue In-Line - 25 Test	14	*Negative	13	92.86%
Quidel QuickVue+ Strep A	19	*Negative	19	100.00%
Select Diag Strep A Dip	1	*Negative	1	100.00%
SP (Cardinal) Rapid Strep Cassette	1	*Negative	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	2	*Negative	2	100.00%
Stanbio QuStick Strep A	1	*Negative	1	100.00%
Consult Diag Strep A Dipstick	17	*Negative	17	100.00%
Meridian ImmunoCard STAT Strep A	1	*Negative	1	100.00%
Quidel Sofia/Sofia 2 Strep A FIA	5	*Negative	5	100.00%
BD Veritor Strep A	6	*Negative	6	100.00%
Quidel Solana	3	*Negative	3	100.00%
Cepheid GeneXpert - Xpert Strep A	1	*Negative	1	100.00%
Quidel QuickVue In-Line - 25 Test	14	Positive	1	7.14%
		Strep A Aq Det. (International)		
RS-1		* Denotes Correct Response		



691 - Group A Strep Antigen (5 specimens)	Labs	Response	Number	Percent
RS-1				
* Denotes Correct Response				
AG: All Methods	162			
		*Negative	160	98.77%
		Positive	2	1.23%
Abbott Signify Strep A - Waived	2	*Negative	2	100.00%
BD Check Strep A (Dip)	2	*Negative	2	100.00%
Beck Coulter ICON DS Strep	1	*Negative	1	100.00%
Alere Acceava Strep A Dipstick	19	*Negative	19	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM	15	*Negative	15	100.00%
Sekisui Diagnostics OSOM Ultra Strep	15	*Negative	15	100.00%
Henry Schein OneStep+ Rapid Strep Dip	10	*Negative	10	100.00%
Immunostics Immuno/Strep A Det	2	*Negative	2	100.00%
Inverness Signify Strep A Dipstick	1	*Negative	1	100.00%
McKesson Strep A Cas	2	*Negative	1	50.00%
McKesson Strep A Dip	10	*Negative	10	100.00%
Quidel QuickVue Dipstick Strep	11	*Negative	11	100.00%
Quidel QuickVue In-Line - 25 Test	14	*Negative	14	100.00%
Quidel QuickVue+ Strep A	19	*Negative	19	100.00%
Select Diag Strep A Dip	1	*Negative	1	100.00%
SP (Cardinal) Rapid Strep Cassette	1	*Negative	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	2	*Negative	2	100.00%
Stanbio QuStick Strep A	1	*Negative	1	100.00%
Consult Diag Strep A Dipstick	17	*Negative	17	100.00%
Meridian ImmunoCard STAT Strep A	1	*Negative	1	100.00%
Quidel Sofia/Sofia 2 Strep A FIA	5	*Negative	4	80.00%
BD Veritor Strep A	6	*Negative	6	100.00%
Quidel Solana	3	*Negative	3	100.00%
Cepheid GeneXpert - Xpert Strep A	1	*Negative	1	100.00%
McKesson Strep A Cas	2	Positive	1	50.00%
Quidel Sofia/Sofia 2 Strep A FIA	5	Positive	1	20.00%
RS-2				
* Denotes Correct Response				
AG: All Methods	162			
		*Negative	161	99.38%
		Positive	1	0.62%
Abbott Signify Strep A - Waived	2	*Negative	2	100.00%
BD Check Strep A (Dip)	2	*Negative	2	100.00%
Beck Coulter ICON DS Strep	1	*Negative	1	100.00%
Alere Acceava Strep A Dipstick	19	*Negative	19	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM	15	*Negative	15	100.00%
Sekisui Diagnostics OSOM Ultra Strep	15	*Negative	15	100.00%
Henry Schein OneStep+ Rapid Strep Dip	10	*Negative	10	100.00%
Immunostics Immuno/Strep A Det	2	*Negative	2	100.00%
Inverness Signify Strep A Dipstick	1	*Negative	1	100.00%
McKesson Strep A Cas	2	*Negative	2	100.00%
McKesson Strep A Dip	10	*Negative	10	100.00%
Quidel QuickVue Dipstick Strep	11	*Negative	11	100.00%
Quidel QuickVue In-Line - 25 Test	14	*Negative	13	92.86%
Quidel QuickVue+ Strep A	19	*Negative	19	100.00%
Select Diag Strep A Dip	1	*Negative	1	100.00%
SP (Cardinal) Rapid Strep Cassette	1	*Negative	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	2	*Negative	2	100.00%
Stanbio QuStick Strep A	1	*Negative	1	100.00%
Consult Diag Strep A Dipstick	17	*Negative	17	100.00%
Meridian ImmunoCard STAT Strep A	1	*Negative	1	100.00%
Quidel Sofia/Sofia 2 Strep A FIA	5	*Negative	5	100.00%
BD Veritor Strep A	6	*Negative	6	100.00%
Quidel Solana	3	*Negative	3	100.00%
Cepheid GeneXpert - Xpert Strep A	1	*Negative	1	100.00%



Method	Count	Result	Count	Percentage
Quidel QuickVue In-Line - 25 Test	14	Positive	1	7.14%
RS-3		* Denotes Correct Response		
AG: All Methods	134			
		*Positive	133	99.25%
		Negative	1	0.75%
Abbott Signify Strep A - Waived	2	*Positive	2	100.00%
BD Check Strep A (Dip)	2	*Positive	2	100.00%
Beck Coulter ICON DS Strep	1	*Positive	1	100.00%
Alere Acceava Strep A Dipstick	19	*Positive	19	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM	15	*Positive	15	100.00%
Sekisui Diagnostics OSOM Ultra Strep	13	*Positive	13	100.00%
Henry Schein OneStep+ Rapid Strep Dip	10	*Positive	10	100.00%
Immunostics Immuno/Strep A Det	2	*Positive	2	100.00%
Inverness Signify Strep A Dipstick	1	*Positive	1	100.00%
McKesson Strep A Dip	5	*Positive	5	100.00%
Quidel QuickVue Dipstick Strep	9	*Positive	9	100.00%
Quidel QuickVue In-Line - 25 Test	13	*Positive	13	100.00%
Quidel QuickVue+ Strep A	17	*Positive	17	100.00%
SP (Cardinal) Rapid Strep Cassette	1	*Positive	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	2	*Positive	2	100.00%
Consult Diag Strep A Dipstick	13	*Positive	13	100.00%
Meridian ImmunoCard STAT Strep A	1	*Positive	1	100.00%
Quidel Sofia/Sofia 2 Strep A FIA	4	*Positive	4	100.00%
BD Veritor Strep A	1	*Positive	1	100.00%
Quidel Solana	1	*Positive	1	100.00%
McKesson Strep A Cas	1	Negative	1	100.00%
RS-4		* Denotes Correct Response		
AG: All Methods	134			
		*Positive	134	100.00%
Abbott Signify Strep A - Waived	2	*Positive	2	100.00%
BD Check Strep A (Dip)	2	*Positive	2	100.00%
Beck Coulter ICON DS Strep	1	*Positive	1	100.00%
Alere Acceava Strep A Dipstick	19	*Positive	19	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM	15	*Positive	15	100.00%
Sekisui Diagnostics OSOM Ultra Strep	13	*Positive	13	100.00%
Henry Schein OneStep+ Rapid Strep Dip	10	*Positive	10	100.00%
Immunostics Immuno/Strep A Det	2	*Positive	2	100.00%
Inverness Signify Strep A Dipstick	1	*Positive	1	100.00%
McKesson Strep A Cas	1	*Positive	1	100.00%
McKesson Strep A Dip	5	*Positive	5	100.00%
Quidel QuickVue Dipstick Strep	9	*Positive	9	100.00%
Quidel QuickVue In-Line - 25 Test	13	*Positive	13	100.00%
Quidel QuickVue+ Strep A	17	*Positive	17	100.00%
SP (Cardinal) Rapid Strep Cassette	1	*Positive	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	2	*Positive	2	100.00%
Consult Diag Strep A Dipstick	13	*Positive	13	100.00%
Meridian ImmunoCard STAT Strep A	1	*Positive	1	100.00%
Quidel Sofia/Sofia 2 Strep A FIA	4	*Positive	4	100.00%
BD Veritor Strep A	1	*Positive	1	100.00%
Quidel Solana	1	*Positive	1	100.00%
RS-5		* Denotes Correct Response		
AG: All Methods	134			
		*Positive	133	99.25%
		Negative	1	0.75%
Abbott Signify Strep A - Waived	2	*Positive	2	100.00%
BD Check Strep A (Dip)	2	*Positive	2	100.00%
Beck Coulter ICON DS Strep	1	*Positive	1	100.00%
Alere Acceava Strep A Dipstick	19	*Positive	19	100.00%
Fisher HC Sure-Vue Select Strep Ag	1	*Positive	1	100.00%
Sekisui Diagnostics OSOM	15	*Positive	15	100.00%
Sekisui Diagnostics OSOM Ultra Strep	13	*Positive	13	100.00%
Henry Schein OneStep+ Rapid Strep Dip	10	*Positive	10	100.00%
Immunostics Immuno/Strep A Det	2	*Positive	2	100.00%
Inverness Signify Strep A Dipstick	1	*Positive	1	100.00%
McKesson Strep A Cas	1	*Positive	1	100.00%
McKesson Strep A Dip	5	*Positive	5	100.00%
Quidel QuickVue Dipstick Strep	9	*Positive	9	100.00%



Quidel QuickVue In-Line - 25 Test	13	*Positive	13	100.00%
Quidel QuickVue+ Strep A	17	*Positive	16	94.12%
SP (Cardinal) Rapid Strep Cassette	1	*Positive	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	2	*Positive	2	100.00%
Consult Diag Strep A Dipstick	13	*Positive	13	100.00%
Meridian ImmunoCard STAT Strep A	1	*Positive	1	100.00%
Quidel Sofia/Sofia 2 Strep A FIA	4	*Positive	4	100.00%
BD Veritor Strep A	1	*Positive	1	100.00%
Quidel Solana	1	*Positive	1	100.00%
Quidel QuickVue+ Strep A	17	Negative	1	5.88%
		Strep A Ag Det. (Waived)		
RS-1		* Denotes Correct Response		
AG: All Methods	874			
		*Negative	865	98.97%
		Positive	9	1.03%
Abbott Signify Strep A - Waived	6	*Negative	6	100.00%
Beck Coulter ICON DS Strep	1	*Negative	1	100.00%
Beck Coulter ICON SC Strep	1	*Negative	1	100.00%
Alere Acceava Strep A Dipstick	112	*Negative	111	99.11%
Fisher HC Sure-Vue Strep A (Waived)	8	*Negative	8	100.00%
Sekisui Diagnostics OSOM	42	*Negative	42	100.00%
Sekisui Diagnostics OSOM Ultra Strep	47	*Negative	47	100.00%
Henry Schein OneStep + Rapid Strep Cass	4	*Negative	4	100.00%
Henry Schein OneStep+ Rapid Strep Dip	70	*Negative	70	100.00%
Instant Technologies i Strep	1	*Negative	1	100.00%
Inverness Signify Strep A Dipstick	7	*Negative	7	100.00%
McKesson Strep A Cas	4	*Negative	4	100.00%
McKesson Strep A Dip	34	*Negative	34	100.00%
McKesson Strep A Test - Twist	2	*Negative	2	100.00%
Permaxim Rediscreeen Strep A Rapid	1	*Negative	1	100.00%
Polymedco Poly Stat Strep A (W)	1	*Negative	1	100.00%
PSS Rapid Test Strip Strep A	2	*Negative	2	100.00%
Quidel QuickVue Dipstick Strep	90	*Negative	90	100.00%
Quidel QuickVue In-Line - 25 Test	69	*Negative	68	98.55%
Quidel QuickVue+ Strep A	27	*Negative	27	100.00%
Select Diag Strep A Dip	4	*Negative	3	75.00%
SP (Cardinal) Rapid Strep Cassette	5	*Negative	5	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	87	*Negative	87	100.00%
Stanbio QuStick Strep A	1	*Negative	1	100.00%
Alere Clearview Exact II Dip	36	*Negative	35	97.22%
Sekisui Diagnostics OSOM Ultra	5	*Negative	5	100.00%
Consult Diag Strep A Dipstick	116	*Negative	114	98.28%
MooreBrand Strep A Dipstick	5	*Negative	5	100.00%
NDC ProAdvantage Strep Strips	2	*Negative	2	100.00%
Meridian ImmunoCard STAT Strep A	2	*Negative	2	100.00%
LifeSign Bio Strep A Test Strip	1	*Negative	1	100.00%
Quidel Sofia/Sofia 2 Strep A FIA	15	*Negative	14	93.33%
BD Veritor Strep A	32	*Negative	32	100.00%
Alere i Strep A	8	*Negative	8	100.00%
Medline Strep A Dipstick	2	*Negative	2	100.00%
Roche Cobas Liat Strep A	22	*Negative	22	100.00%
Alere Acceava Strep A Dipstick	112	Positive	1	0.89%
Innovacon One Step Strep Ag	1	Positive	1	100.00%
Quidel QuickVue In-Line - 25 Test	69	Positive	1	1.45%
Select Diag Strep A Dip	4	Positive	1	25.00%
Alere Clearview Exact II Dip	36	Positive	1	2.78%
Consult Diag Strep A Dipstick	116	Positive	2	1.72%
Consult Diag Strep A Twist	1	Positive	1	100.00%
Quidel Sofia/Sofia 2 Strep A FIA	15	Positive	1	6.67%
RS-2		* Denotes Correct Response		
AG: All Methods	57			
		*Negative	57	100.00%
Alere Acceava Strep A Dipstick	1	*Negative	1	100.00%
Sekisui Diagnostics OSOM	10	*Negative	10	100.00%
Sekisui Diagnostics OSOM Ultra Strep	4	*Negative	4	100.00%
Innovacon One Step Strep Ag	1	*Negative	1	100.00%
McKesson Strep A Dip	2	*Negative	2	100.00%
Quidel QuickVue Dipstick Strep	1	*Negative	1	100.00%
Quidel QuickVue In-Line - 25 Test	4	*Negative	4	100.00%



Quidel QuickVue+ Strep A	1	*Negative	1	100.00%
SP Brand (Cardinal) Rapid Strep A Dipstick	4	*Negative	4	100.00%
Consult Diag Strep A Dipstick	6	*Negative	6	100.00%
MooreBrand Strep A Dipstick	1	*Negative	1	100.00%
BD Veritor Strep A	1	*Negative	1	100.00%
Alere i Strep A	1	*Negative	1	100.00%
Roche Cobas Liat Strep A	20	*Negative	20	100.00%
RS-3		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Quidel QuickVue In-Line - 25 Test	1	*Positive	1	100.00%
RS-4		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Quidel QuickVue In-Line - 25 Test	1	*Positive	1	100.00%
RS-5		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Quidel QuickVue In-Line - 25 Test	1	*Positive	1	100.00%
		Strep A Ag Det. (International)		
RS-1		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Alere Clearview Exact II Cas	1	*Negative	1	100.00%
RS-2		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Alere Clearview Exact II Cas	1	*Negative	1	100.00%
RS-3		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Alere Clearview Exact II Cas	1	*Positive	1	100.00%
RS-4		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Alere Clearview Exact II Cas	1	*Positive	1	100.00%
RS-5		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Alere Clearview Exact II Cas	1	*Positive	1	100.00%
739 - Sickle Cell Screen	Labs	Response	Number	Percent
		Sickle Cell Screen		
SS-1		* Denotes Correct Response		
AG: All Methods	8			
		*Positive	7	87.50%
		Negative	1	12.50%
Pacific Hemostasis Sickle Screen	1	*Positive	1	100.00%
Streck SickleDex	7	*Positive	6	85.71%
Streck SickleDex		Negative	1	14.29%
679 - Syphilis Serology	Labs	Response	Number	Percent
		RPR		
SY-1		* Denotes Correct Response		
AG: All Methods	9			
		*Reactive	9	100.00%
ASI RPR	2	*Reactive	2	100.00%
Becton Dickinson BBL	3	*Reactive	3	100.00%
New Horizons	1	*Reactive	1	100.00%
Stanbio Quicktest	1	*Reactive	1	100.00%
Wampole Impact	1	*Reactive	1	100.00%
Intl ASI RPR	1	*Reactive	1	100.00%
SY-2		* Denotes Correct Response		
AG: All Methods	9			
		*Nonreactive	9	100.00%
ASI RPR	2	*Nonreactive	2	100.00%
Becton Dickinson BBL	3	*Nonreactive	3	100.00%
New Horizons	1	*Nonreactive	1	100.00%
Stanbio Quicktest	1	*Nonreactive	1	100.00%
Wampole Impact	1	*Nonreactive	1	100.00%



776 - Throat Culture;780 - Throat & Genital Cultures	Labs	Response	Number	Percent
SY-3				
* Denotes Correct Response				
AG: All Methods	9			
		*Nonreactive	1	100.00%
		*Reactive	8	88.89%
		*Weakly Reactive	1	11.11%
ASI RPR	2	*Reactive	2	100.00%
Becton Dickinson BBL	3	*Reactive	3	100.00%
New Horizons	1	*Reactive	1	100.00%
Stanbio Quicktest	1	*Reactive	1	100.00%
Intl ASI RPR	1	*Reactive	1	100.00%
Wampole Impact	1	*Weakly Reactive	1	100.00%
SY-4				
* Denotes Correct Response				
AG: All Methods	9			
		*Reactive	9	100.00%
ASI RPR	2	*Reactive	2	100.00%
Becton Dickinson BBL	3	*Reactive	3	100.00%
New Horizons	1	*Reactive	1	100.00%
Stanbio Quicktest	1	*Reactive	1	100.00%
Wampole Impact	1	*Reactive	1	100.00%
Intl ASI RPR	1	*Reactive	1	100.00%
SY-5				
* Denotes Correct Response				
AG: All Methods	9			
		*Nonreactive	9	100.00%
ASI RPR	2	*Nonreactive	2	100.00%
Becton Dickinson BBL	3	*Nonreactive	3	100.00%
New Horizons	1	*Nonreactive	1	100.00%
Stanbio Quicktest	1	*Nonreactive	1	100.00%
Wampole Impact	1	*Nonreactive	1	100.00%
Intl ASI RPR	1	*Nonreactive	1	100.00%
TC-1				
* Denotes Correct Response				
AG: All Methods	64			
		*Beta hemolysis, referred for ID/confirmation	5	7.81%
		*Growth, referred for identification	2	3.13%
		*Negative for Group A Streptococcus	49	76.56%
		*Streptococcus, beta-hemolytic, not Group A	3	4.69%
		No Growth - Strep Selective medium	1	1.56%
		Positive for Group A Streptococcus	3	4.69%
		Presumptive Positive for Group A Streptococcus	1	1.56%
Cultures	64	*Beta hemolysis, referred for ID/confirmation	5	7.81%
Cultures		*Growth, referred for identification	2	3.13%
Cultures		*Negative for Group A Streptococcus	49	76.56%
Cultures		*Streptococcus, beta-hemolytic, not Group A	3	4.69%
Cultures		No Growth - Strep Selective medium	1	1.56%
Cultures		Positive for Group A Streptococcus	3	4.69%
Cultures		Presumptive Positive for Group A Streptococcus	1	1.56%
TC-2				
* Denotes Correct Response				
AG: All Methods	64			
		*Beta hemolysis, referred for ID/confirmation	1	1.56%
		*Positive for Group A Streptococcus	31	48.44%
		*Presumptive Positive for Group A Streptococcus	26	40.63%
		*Presumptive Streptococcus species	1	1.56%
		*Streptococcus, beta-hemolytic, Group A	1	1.56%
		Negative for Group A Streptococcus	4	6.25%
Cultures	64	*Beta hemolysis, referred for ID/confirmation	1	1.56%
Cultures		*Positive for Group A Streptococcus	31	48.44%
Cultures		*Presumptive Positive for Group A Streptococcus	26	40.63%
Cultures		*Presumptive Streptococcus species	1	1.56%
Cultures		*Streptococcus, beta-hemolytic, Group A	1	1.56%
Cultures		Negative for Group A Streptococcus	4	6.25%
TC-3				
* Denotes Correct Response				
AG: All Methods	65			
		*Beta hemolysis, referred for ID/confirmation	1	1.54%
		*Positive for Group A Streptococcus	33	50.77%
		*Presumptive Positive for Group A Streptococcus	27	41.54%
		*Presumptive Streptococcus species	2	3.08%
		Negative for Group A Streptococcus	2	3.08%



Cultures	65	*Beta hemolysis, referred for ID/confirmation	1	1.54%
Cultures		*Positive for Group A Streptococcus	33	50.77%
Cultures		*Presumptive Positive for Group A Streptococcus	27	41.54%
Cultures		*Presumptive Streptococcus species	2	3.08%
Cultures		Negative for Group A Streptococcus	2	3.08%
TC-4		* Denotes Correct Response		
AG: All Methods	81			
		*Negative for Group A Streptococcus	58	71.60%
		*No beta hemolysis	4	4.94%
		*No Growth - Strep Selective medium	16	19.75%
		Beta hemolysis, referred for ID/confirmation	1	1.23%
		Positive for Group A Streptococcus	2	2.47%
Cultures	81	*Negative for Group A Streptococcus	58	71.60%
Cultures		*No beta hemolysis	4	4.94%
Cultures		*No Growth - Strep Selective medium	16	19.75%
Cultures		Beta hemolysis, referred for ID/confirmation	1	1.23%
Cultures		Positive for Group A Streptococcus	2	2.47%
TC-5		* Denotes Correct Response		
AG: All Methods	81			
		*Beta hemolysis, referred for ID/confirmation	6	7.41%
		*Positive for Group A Streptococcus	41	50.62%
		*Presumptive Positive for Group A Streptococcus	26	32.10%
		*Presumptive Streptococcus species	2	2.47%
		*Streptococcus, beta-hemolytic, Group A	1	1.23%
		Negative for Group A Streptococcus	5	6.17%
Cultures	81	*Beta hemolysis, referred for ID/confirmation	6	7.41%
Cultures		*Positive for Group A Streptococcus	41	50.62%
Cultures		*Presumptive Positive for Group A Streptococcus	26	32.10%
Cultures		*Presumptive Streptococcus species	2	2.47%
Cultures		*Streptococcus, beta-hemolytic, Group A	1	1.23%
Cultures		Negative for Group A Streptococcus	5	6.17%
640 - Urinalysis;642 - Urinalysis/hCG;639 - Complete Urinalysis Package;601 - Waived/PPM Tests II;609 - Waived/PPM Tests I;647 - Mini-Waived Package	Labs	Response	Number	Percent
		Urine Sp Gravity		
UA-1		* Denotes Correct Response		
SG: Select Siemens Clinitek/Visual	947			
		*1.015	210	22.18%
		*1.020	596	62.94%
		*1.025	116	12.25%
		1.000	1	0.11%
		1.010	5	0.53%
		1.030	19	2.01%
UA-1		* Denotes Correct Response		
SG: Select Roche Urinalysis Inst/Visual	102			
		*1.015	67	65.69%
		*1.020	30	29.41%
		1.010	1	0.98%
		1.025	2	1.96%
		1.030	2	1.96%
UA-1		* Denotes Correct Response		
SG: Select Consult Diag Urinalysis Inst/Visual	28			
		*1.020	6	21.43%
		*1.025	10	35.71%
		*1.030	9	32.14%
		1.015	3	10.71%
UA-1		* Denotes Correct Response		
SG: Select McKesson Urinalysis Inst/Visual	39			
		*1.025	22	56.41%
		*1.030	15	38.46%
		1.020	2	5.13%
UA-1		* Denotes Correct Response		
MG: Method Group	1188			
		*1.015	293	24.66%
		*1.020	652	54.88%
		*1.025	171	14.39%
		*1.030	65	5.47%
		1.000	1	0.08%



		1.010	6	0.51%
		1.015	293	24.66%
		1.020	652	54.88%
		1.025	171	14.39%
		1.030	65	5.47%
UA-1		* Denotes Correct Response		
MG: Roche - Conventional	32			
		*1.015	23	71.88%
		*1.020	9	28.13%
UA-1		* Denotes Correct Response		
MG: Siemens - Plus Sys	17			
		*1.015	1	5.88%
		*1.020	15	88.24%
		*1.025	1	5.88%
UA-1		* Denotes Correct Response		
MG: Roche - Arbitrary	1			
		*1.020	1	100.00%
UA-1		* Denotes Correct Response		
MG: Method Group	1179			
		*Negative	1177	99.83%
		Trace/Small (1+)	2	0.17%
UA-1		* Denotes Correct Response		
MG: Roche - Conventional	33			
		*Negative	33	100.00%
UA-1		* Denotes Correct Response		
MG: Siemens - Plus Sys	17			
		*Negative	17	100.00%
UA-1		* Denotes Correct Response		
MG: Roche - Arbitrary	1			
		*Negative	1	100.00%
UA-1		* Denotes Correct Response		
MG: Method Group	1195			
		*Large (3+) / 80 mg/dL	504	42.18%
		*Moderate (2+) / 40 mg/dL	649	54.31%
		>=160 mg/dL	1	0.08%
		Negative	7	0.59%
		Small (1+) / 15mg/dL	28	2.34%
		Trace/5 mg/dL	6	0.50%
UA-1		* Denotes Correct Response		
MG: Roche - Conventional	33			
		*50 mg/dL	31	93.94%
		15 mg/dL	2	6.06%
UA-1		* Denotes Correct Response		
MG: Siemens - Plus Sys	17			
		*2+	11	64.71%
		*3+	5	29.41%
		*Trace	1	5.88%
UA-1		* Denotes Correct Response		
MG: Roche - Arbitrary	1			
		*++	1	100.00%
UA-1		* Denotes Correct Response		
MG: Method Group	1195			
		*5 or less	589	49.29%
		*5.5	478	40.00%
		*6	123	10.29%
		6.5	5	0.42%
UA-1		* Denotes Correct Response		
MG: Roche - Conventional	33			
		*5	33	100.00%
UA-1		* Denotes Correct Response		
MG: Siemens - Plus Sys	17			
		*5.0	6	35.29%
		*5.5	11	64.71%
UA-1		* Denotes Correct Response		
MG: Roche - Arbitrary	1			



		*5.0	1	100.00%
		Urine Blood/Hemoglobin		
UA-1		* Denotes Correct Response		
MG: Method Group	1198			
		*Negative	1161	96.91%
		Large (3+)	1	0.08%
		Moderate (2+)	3	0.25%
		Small (1+)	2	0.17%
		Trace	31	2.59%
UA-1		* Denotes Correct Response		
MG: Roche - Conventional	33			
		*Negative	31	93.94%
		50 Ery/uL	1	3.03%
		Trace	1	3.03%
UA-1		* Denotes Correct Response		
MG: Siemens - Plus Sys	17			
		*Negative	15	88.24%
		Trace-intact	2	11.76%
UA-1		* Denotes Correct Response		
MG: Roche - Arbitrary	1			
		*Negative	1	100.00%
		Urine Protein - Qual		
UA-1		* Denotes Correct Response		
MG: Method Group	1207			
		*Negative	1171	97.02%
		100 (2+)	1	0.08%
		30 (1+)	3	0.25%
		300-500 (3+)	1	0.08%
		Trace	31	2.57%
UA-1		* Denotes Correct Response		
MG: Roche - Conventional	33			
		*Negative	32	96.97%
		30 mg/dL	1	3.03%
UA-1		* Denotes Correct Response		
MG: Siemens - Plus Sys	17			
		*Negative	17	100.00%
UA-1		* Denotes Correct Response		
MG: Roche - Arbitrary	1			
		*Negative	1	100.00%
		Urine Glucose (Reducing Sub)		
UA-1		* Denotes Correct Response		
SG: Select Siemens Clinitek/Visual	953			
		*<100 mg/dL	53	5.56%
		*100 mg/dL	324	34.00%
		*250 mg/dL	514	53.93%
		>= 1000 mg/dL	6	0.63%
		500 mg/dL	4	0.42%
		Negative	52	5.46%
UA-1		* Denotes Correct Response		
SG: Select Roche Urinalysis Inst/Visual	110			
		*>= 1000 mg/dL	97	88.18%
		*500 mg/dL	4	3.64%
		100 mg/dL	3	2.73%
		250 mg/dL	3	2.73%
		Negative	3	2.73%
UA-1		* Denotes Correct Response		
SG: Select Consult Diag Urinalysis Inst/Visual	28			
		*100 mg/dL	5	17.86%
		*250 mg/dL	19	67.86%
		<100 mg/dL	2	7.14%
		500 mg/dL	2	7.14%
UA-1		* Denotes Correct Response		
SG: Select McKesson Urinalysis Inst/Visual	39			
		*100 mg/dL	9	23.08%
		*250 mg/dL	28	71.79%
		<100 mg/dL	2	5.13%
UA-1		* Denotes Correct Response		
MG: Method Group	1206			



		*<100 mg/dL	66	5.47%
		*>= 1000 mg/dL	104	8.62%
		*100 mg/dL	360	29.85%
		*250 mg/dL	590	48.92%
		*500 mg/dL	20	1.66%
		<100 mg/dL	66	5.47%
		>= 1000 mg/dL	104	8.62%
		100 mg/dL	360	29.85%
		250 mg/dL	590	48.92%
		500 mg/dL	20	1.66%
		Negative	66	5.47%
UA-1		* Denotes Correct Response		
MG: Roche - Conventional	33			
		*>= 1000 mg/dL	32	96.97%
		250 mg/dL	1	3.03%
UA-1		* Denotes Correct Response		
MG: Siemens - Plus Sys	17			
		*1+	11	64.71%
		*2+	1	5.88%
		*Trace	5	29.41%
UA-1		* Denotes Correct Response		
MG: Roche - Arbitrary	1			
		*+++	1	100.00%
UA-1		Urine Urobilinogen		
UA-1		* Denotes Correct Response		
PG: Siemens Clinitek Status / Siemens - Plus Sys / Siemens Urine Dipstick	6			
		*3.2 umol/L	6	100.00%
UA-1		* Denotes Correct Response		
PG: Siemens Clinitek Status / Siemens - Plus Sys / Siemens Urine Urobilinogen	9			
		*0.2 EU/dL	9	100.00%
UA-1		* Denotes Correct Response		
PG: Siemens Clinitek Advantus / Siemens - Plus Sys / Siemens Urine Dipstick	1			
		*3.2 umol/L	1	100.00%
UA-1		* Denotes Correct Response		
MG: Method Group	1172			
		*0.2/Normal	1168	99.66%
		2.0 mg/dL	3	0.26%
		4 mg/dL	1	0.09%
UA-1		* Denotes Correct Response		
MG: Roche - Conventional	33			
		*Normal	33	100.00%
UA-1		* Denotes Correct Response		
MG: Roche - Conventional	16			
		*0.2 EU/dL	9	56.25%
		*3.2 umol/L	7	43.75%
UA-1		* Denotes Correct Response		
MG: Roche - Arbitrary	1			
		*Normal	1	100.00%
UA-1		Urine Leuk Esterase		
UA-1		* Denotes Correct Response		
PG: McKesson 120 Urine Analyzer	37			
		*Moderate (2+)	5	13.51%
		*Small (1+)	25	67.57%
		*Trace	2	5.41%
		Negative	5	13.51%
UA-1		* Denotes Correct Response		
SG: Select Siemens Clinitek/Visual	948			
		*Moderate (2+)	229	24.16%
		*Small (1+)	635	66.98%
		*Trace	61	6.43%
		Large (3+)	12	1.27%
		Negative	11	1.16%
UA-1		* Denotes Correct Response		
SG: Select Roche Urinalysis Inst/Visual	109			
		*Large (3+)	46	42.20%
		*Moderate (2+)	53	48.62%



		Small (1+)	10	9.17%
UA-1		* Denotes Correct Response		
SG: Select Consult Diag Urinalysis Inst/Visual	28			
		*Moderate (2+)	13	46.43%
		*Small (1+)	12	42.86%
		Negative	3	10.71%
UA-1		* Denotes Correct Response		
MG: Method Group	1197			
		*Large (3+)	72	6.02%
		*Moderate (2+)	322	26.90%
		*Small (1+)	717	59.90%
		*Trace	65	5.43%
		Large (3+)	72	6.02%
		Negative	21	1.75%
		Small (1+)	717	59.90%
UA-1		* Denotes Correct Response		
MG: Roche - Conventional	33			
		*500 Leu/uL	33	100.00%
UA-1		* Denotes Correct Response		
MG: Siemens - Plus Sys	17			
		*1+	14	82.35%
		*2+	3	17.65%
UA-1		* Denotes Correct Response		
MG: Roche - Arbitrary	1			
		*++	1	100.00%
		Urine Microalbumin		
UA-1		* Denotes Correct Response		
PG: Roche Chemstrip Micral	26			
		10 mg/L (Trace)	5	19.23%
		20/30 mg/L (Small)	6	23.08%
		50 mg/L (2+)	4	15.38%
		Negative	11	42.31%
UA-1		* Denotes Correct Response		
PG: Roche Chemstrips - Visual	1			
		50 mg/L (2+)	1	100.00%
UA-1		* Denotes Correct Response		
PG: Hemocue Albumin 201	1			
		50 mg/L (2+)	1	100.00%
UA-1		* Denotes Correct Response		
SG: Select Siemens Clinitek/Visual	57			
		*10 mg/L (Trace)	51	89.47%
		*20/30 mg/L (Small)	2	3.51%
		Negative	4	7.02%
		Fluid pH		
UA-1		* Denotes Correct Response		
PG: Nitrazine Paper	11			
		*5.5	6	54.55%
		*6	4	36.36%
		5 or less	1	9.09%
		Urine hCG		
UA-1		* Denotes Correct Response		
AG: All Methods	1062			
		*Negative	1060	99.81%
		Positive	2	0.19%
ASI ProPhase Plus	1	*Negative	1	100.00%
Siemens Clinitek Status	12	*Negative	12	100.00%
Beck Coulter ICON 20 hCG	7	*Negative	7	100.00%
Beck Coulter ICON 25 hCG	65	*Negative	65	100.00%
Alere Acceava HCG Basic II	24	*Negative	24	100.00%
Alere Acceava hCG Combo	5	*Negative	5	100.00%
Alere Acceava hCG-Urine	20	*Negative	20	100.00%
Conception Tech One Step hCG	2	*Negative	2	100.00%
Diag Test Group Clarity hCG Plus Urine Cassettes	4	*Negative	4	100.00%
Fisher HC Sure-Vue hCG STAT	5	*Negative	5	100.00%
Fisher HC Sure-Vue Urine hCG	11	*Negative	11	100.00%
Fisher HC Sure-Vue U/S hCG	5	*Negative	5	100.00%
Sekisui Diagnostics OSOM Urine	11	*Negative	11	100.00%
Sekisui Diagnostics OSOM Card Pregnancy	27	*Negative	27	100.00%



Sekisui Diagnostics OSOM hCG Combo	42	*Negative	42	100.00%
Henry Schein One Step + (hCG)	50	*Negative	50	100.00%
Immunostics Detector Combo	3	*Negative	3	100.00%
Immunostics hCG Detector-urine	4	*Negative	4	100.00%
Instant Tech. i Pregnancy	3	*Negative	3	100.00%
Inverness Clearview Combo II	2	*Negative	2	100.00%
Inv Clearview Urine - Cassette	4	*Negative	4	100.00%
Inverness Signify hCG combo	3	*Negative	3	100.00%
Inverness Signify hCG urine	4	*Negative	4	100.00%
LifeSign Status hCG	1	*Negative	1	100.00%
McKesson Medi-Lab hCG - Urine	38	*Negative	38	100.00%
Meridian Immunocard Stat!	1	*Negative	1	100.00%
Permaxim RediScreen	3	*Negative	3	100.00%
Polymedco Poly stat hCG Urine	1	*Negative	1	100.00%
Quidel QuickVue + hCG	23	*Negative	23	100.00%
Quidel QuickVue One-Step hCG Combo U/S	77	*Negative	77	100.00%
Quidel QuickVue One-Step Urine	151	*Negative	151	100.00%
Quidel RapidVue hCG - Urine	7	*Negative	7	100.00%
SAS Ultra HCG	2	*Negative	2	100.00%
Select Diag hCG Cassette (urine)	21	*Negative	21	100.00%
Select Diag hCG Combo	10	*Negative	10	100.00%
Select Diag hCG Dipstick	7	*Negative	6	85.71%
SP (Cardinal) Brand combo	16	*Negative	16	100.00%
SP (Cardinal) Brand-cassette	29	*Negative	29	100.00%
SP Brand (Cardinal) HCG Combo	39	*Negative	39	100.00%
Stanbio Qupid	5	*Negative	5	100.00%
Stanbio QuStick	2	*Negative	2	100.00%
Stanbio TRUE hCG	5	*Negative	5	100.00%
MooreBrand hCG Cassette	3	*Negative	3	100.00%
Nova Scientific Century S/U hCG	1	*Negative	1	100.00%
Diag Test Group Clarity hCG Urine Strips	1	*Negative	1	100.00%
Jant AccuTest	5	*Negative	5	100.00%
Becton Dickinson Precise One-Step	1	*Negative	1	100.00%
Consult Diag hCG Urine Cassette	161	*Negative	160	99.38%
CLIAwaived Inc Pregnancy Urine Dip-Strip	5	*Negative	5	100.00%
Germaine AIM Step Combo Pregnancy	1	*Negative	1	100.00%
NDC ProAdvantage Urine hCG Cassette	5	*Negative	5	100.00%
Consult Diag hCG Urine Dipstick	28	*Negative	28	100.00%
SP (Cardinal) Brand Dipstick	9	*Negative	9	100.00%
Consult Diag hCG Combo Cassette (Urine)	12	*Negative	12	100.00%
McKesson Medi-Lab Performance hCG Urine Cassette	7	*Negative	7	100.00%
Alere hCG Combo Cassette (25 mIU/mL)	3	*Negative	3	100.00%
Alere hCG Cassette (25 mIU/mL)	10	*Negative	10	100.00%
Alere hCG Dipstick	6	*Negative	6	100.00%
Medline Pregnancy Urine Dip-Strip	2	*Negative	2	100.00%
Medline Pregnancy Urine Cassette Test	2	*Negative	2	100.00%
McKesson hCG Urine Test Dipstick	5	*Negative	5	100.00%
Quidel QuickVue hCG Urine Test	17	*Negative	17	100.00%
MooreBrand hCG Dipstick	2	*Negative	2	100.00%
Siemens Clinitest hCG	1	*Negative	1	100.00%
Henry Schein One Step + HCG Combo	2	*Negative	2	100.00%
Henry Schein One Step hCG Urine Cassette	18	*Negative	18	100.00%
Medline hCG Test Strip 25 mIU/mL	1	*Negative	1	100.00%
NCS hCG One Step Pregnancy Test	1	*Negative	1	100.00%
MediChoice Urine hCG Cassette	1	*Negative	1	100.00%
Select Diag hCG Dipstick	7	Positive	1	14.29%
Consult Diag hCG Urine Cassette	161	Positive	1	0.62%
661 - Urine Colony Count;784 - Presump ID/Colony Count - Urine;785 - Presump ID/Colony Count-Urine & Strep Combo;775 - Urine Culture;778 - Urine & Throat Cultures;779 - Urine & Genital Cultures;781 - Urine, Throat & Genital Cultures;782 - Complete Bacteri	Labs	Response	Number	Percent
		Susceptibility Cefazolin		
UC-1		* Denotes Correct Response		
AG: All Methods	2			
		*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
		Susceptibility Cefdinir		
UC-1		* Denotes Correct Response		
AG: All Methods	1			



		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
		Susceptibility Cefepime		
UC-1		* Denotes Correct Response		
AG: All Methods	2			
		*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
		Susceptibility Cefotaxime		
UC-1		* Denotes Correct Response		
AG: All Methods	2			
		*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
		Urine Culture		
UC-1		* Denotes Correct Response		
AG: All Methods	30			
		*Escherichia coli	3	10.00%
		*Gram Negative organism	4	13.33%
		*Growth, referred for identification	13	43.33%
		*Presumptive E. coli	3	10.00%
		*Uricult - Growth (Positive)	5	16.67%
		No Growth	2	6.67%
Cultures	30	*Escherichia coli	3	10.00%
Cultures		*Gram Negative organism	4	13.33%
Cultures		*Growth, referred for identification	13	43.33%
Cultures		*Presumptive E. coli	3	10.00%
Cultures		*Uricult - Growth (Positive)	5	16.67%
Cultures		No Growth	2	6.67%
UC-2		* Denotes Correct Response		
AG: All Methods	30			
		*Gram Negative organism	3	10.00%
		*Growth, referred for identification	15	50.00%
		*Presumptive Proteus species	2	6.67%
		*Proteus mirabilis	2	6.67%
		*Proteus species NOS	1	3.33%
		*Uricult - Growth (Positive)	4	13.33%
		No Growth	2	6.67%
		Uricult - No Growth (Negative)	1	3.33%
Cultures	30	*Gram Negative organism	3	10.00%
Cultures		*Growth, referred for identification	15	50.00%
Cultures		*Presumptive Proteus species	2	6.67%
Cultures		*Proteus mirabilis	2	6.67%
Cultures		*Proteus species NOS	1	3.33%
Cultures		*Uricult - Growth (Positive)	4	13.33%
Cultures		No Growth	2	6.67%
Cultures		Uricult - No Growth (Negative)	1	3.33%
UC-3		* Denotes Correct Response		
AG: All Methods	29			
		*Enterococcus faecalis	2	6.90%
		*Gram Positive organism	2	6.90%
		*Growth, referred for identification	16	55.17%
		*Uricult - Growth (Positive)	4	13.79%
		No Growth	2	6.90%
		Presumptive Streptococcus species	1	3.45%
		Uricult - No Growth (Negative)	2	6.90%
Cultures	29	*Enterococcus faecalis	2	6.90%
Cultures		*Gram Positive organism	2	6.90%
Cultures		*Growth, referred for identification	16	55.17%
Cultures		*Uricult - Growth (Positive)	4	13.79%
Cultures		No Growth	2	6.90%
Cultures		Presumptive Streptococcus species	1	3.45%
Cultures		Uricult - No Growth (Negative)	2	6.90%
UC-4		* Denotes Correct Response		
AG: All Methods	10			
		*No Growth	8	80.00%
		Uricult - Growth (Positive)	2	20.00%
Cultures	10	*No Growth	8	80.00%
Cultures		Uricult - Growth (Positive)	2	20.00%
UC-5		* Denotes Correct Response		



AG: All Methods	10			
		*Enterobacter cloacae	2	20.00%
		*Gram Negative organism	2	20.00%
		*Growth, referred for identification	2	20.00%
		*Presumptive Enterobacter species	1	10.00%
		*Uricult - Growth (Positive)	2	20.00%
		No Growth	1	10.00%
Cultures	10	*Enterobacter cloacae	2	20.00%
Cultures		*Gram Negative organism	2	20.00%
Cultures		*Growth, referred for identification	2	20.00%
Cultures		*Presumptive Enterobacter species	1	10.00%
Cultures		*Uricult - Growth (Positive)	2	20.00%
Cultures		No Growth	1	10.00%
		Susceptibility Cefpodoxime		
UC-1		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
		Susceptibility Cefazidime		
UC-1		* Denotes Correct Response		
AG: All Methods	2			
		*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
		Susceptibility Ceftriaxone		
UC-1		* Denotes Correct Response		
AG: All Methods	4			
		*Sensitive	4	100.00%
Susceptibility	4	*Sensitive	4	100.00%
		Susceptibility Cefuroxime		
UC-1		* Denotes Correct Response		
AG: All Methods	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
		Susceptibility Cephalothin		
UC-1		* Denotes Correct Response		
AG: All Methods	5			
		*Intermediate	2	40.00%
		*Sensitive	3	60.00%
Susceptibility	5	*Intermediate	2	40.00%
Susceptibility		*Sensitive	3	60.00%
		Susceptibility Ciprofloxacin		
UC-1		* Denotes Correct Response		
AG: All Methods	8			
		*Sensitive	8	100.00%
Susceptibility	8	*Sensitive	8	100.00%
		Susceptibility Gentamicin		
UC-1		* Denotes Correct Response		
AG: All Methods	5			
		*Sensitive	5	100.00%
Susceptibility	5	*Sensitive	5	100.00%
		Susceptibility Imipenem		
UC-1		* Denotes Correct Response		
AG: All Methods	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
		Susceptibility Levofloxacin		
UC-1		* Denotes Correct Response		
AG: All Methods	4			
		*Sensitive	4	100.00%
Susceptibility	4	*Sensitive	4	100.00%
		Susceptibility Minocycline		
UC-1		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
		Susceptibility Nitrofurantoin		
UC-1		* Denotes Correct Response		



AG: All Methods	7			
		*Sensitive	7	100.00%
Susceptibility	7	*Sensitive	7	100.00%
		Susceptibility Norfloxacin/Noroxin		
UC-1		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
		Susceptibility Piperacillin		
UC-1		* Denotes Correct Response		
AG: All Methods	1			
		*Resistant	1	100.00%
Susceptibility	1	*Resistant	1	100.00%
		Susceptibility Piperacillin/Tazobactam		
UC-1		* Denotes Correct Response		
AG: All Methods	2			
		*Sensitive	2	100.00%
Susceptibility	2	*Sensitive	2	100.00%
		Susceptibility Sulfamethoxazole		
UC-1		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
		Susceptibility Tetracycline		
UC-1		* Denotes Correct Response		
AG: All Methods	5			
		*Sensitive	5	100.00%
Susceptibility	5	*Sensitive	5	100.00%
		Susceptibility Tobramycin		
UC-1		* Denotes Correct Response		
AG: All Methods	3			
		*Sensitive	3	100.00%
Susceptibility	3	*Sensitive	3	100.00%
		Susceptibility Trimethoprim		
UC-1		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
		Susceptibility Trimeth./Sulfameth.		
UC-1		* Denotes Correct Response		
AG: All Methods	6			
		*Sensitive	6	100.00%
Susceptibility	6	*Sensitive	6	100.00%
		Susceptibility Amikacin		
UC-1		* Denotes Correct Response		
AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
		Susceptibility Amox/Clavulanate		
UC-1		* Denotes Correct Response		
AG: All Methods	5			
		*Intermediate	2	40.00%
		*Sensitive	3	60.00%
Susceptibility	5	*Intermediate	2	40.00%
Susceptibility		*Sensitive	3	60.00%
		Susceptibility Ampicillin		
UC-1		* Denotes Correct Response		
AG: All Methods	6			
		*Resistant	6	100.00%
Susceptibility	6	*Resistant	6	100.00%
		Susceptibility Ampicillin/Sulbactam		
UC-1		* Denotes Correct Response		
AG: All Methods	2			
		*Resistant	2	100.00%
Susceptibility	2	*Resistant	2	100.00%
		Susceptibility Ertapenem		
UC-1		* Denotes Correct Response		



AG: All Methods	1			
		*Sensitive	1	100.00%
Susceptibility	1	*Sensitive	1	100.00%
703 - Urine Drug Screen;713 - Urine Drug Screen (50 ml sample volume)	Labs	Response	Number	Percent
		Barbiturates		
UDS-1		* Denotes Correct Response		
AG: All Methods	137			
		*Negative	137	100.00%
AlereTriage	3	*Negative	3	100.00%
E-Screen/Alere	24	*Negative	24	100.00%
Express Diag. Express Check	9	*Negative	9	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Innovacon One Step	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	11	*Negative	11	100.00%
Instant Technologies i Screen	1	*Negative	1	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	3	*Negative	3	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	3	*Negative	3	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Alere iCassette	2	*Negative	2	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	4	*Negative	4	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Mindray 480	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	22	*Negative	22	100.00%
Premier Biotech Bio-Cup	13	*Negative	13	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
UDS-2		* Denotes Correct Response		
AG: All Methods	137			
		*Negative	137	100.00%
AlereTriage	3	*Negative	3	100.00%
E-Screen/Alere	24	*Negative	24	100.00%
Express Diag. Express Check	9	*Negative	9	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Innovacon One Step	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	11	*Negative	11	100.00%
Instant Technologies i Screen	1	*Negative	1	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	3	*Negative	3	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	3	*Negative	3	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Alere iCassette	2	*Negative	2	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	4	*Negative	4	100.00%



Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Mindray 480	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	22	*Negative	22	100.00%
Premier Biotech Bio-Cup	13	*Negative	13	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Benzodiazepines				
UDS-1		* Denotes Correct Response		
AG: All Methods	150			
		*Negative	150	100.00%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	1	*Negative	1	100.00%
E-Screen/Alere	24	*Negative	24	100.00%
Express Diag. Express Check	9	*Negative	9	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Innovacon One Step	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	18	*Negative	18	100.00%
Instant Technologies i Screen	1	*Negative	1	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	4	*Negative	4	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	3	*Negative	3	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Alere iCassette	2	*Negative	2	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	4	*Negative	4	100.00%
ThermoScientific Indiko Plus	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Mindray 480	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	22	*Negative	22	100.00%
Siemens Viva/ProE System	1	*Negative	1	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	13	*Negative	13	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
UDS-2		* Denotes Correct Response		
AG: All Methods	151			
		*Positive	147	97.35%
		Negative	4	2.65%
AlereTriage	3	*Positive	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	1	*Positive	1	100.00%
E-Screen/Alere	24	*Positive	24	100.00%
Express Diag. Express Check	9	*Positive	9	100.00%
Innovacon One Step	1	*Positive	1	100.00%
Alere iCup Drugs of Abuse	19	*Positive	18	94.74%
Instant Technologies i Screen	1	*Positive	1	100.00%
MedTox Profile II	1	*Positive	1	100.00%
Noble Med Split Specimen Cup	1	*Positive	1	100.00%
Beckman (Olympus) AU	1	*Positive	1	100.00%
Innovacon Multi-Clin	10	*Positive	10	100.00%
MedTox Profile III ER	1	*Positive	1	100.00%
CLIAwaived Inc Rapid Drug Cup	4	*Positive	3	75.00%
ASC One Screen Drug Test Card	1	*Positive	1	100.00%
MedTox Profile V	2	*Positive	2	100.00%
Jant Accutest Multi Drug Screen	3	*Positive	3	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Positive	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Positive	4	100.00%
Mindray BS-200	2	*Positive	2	100.00%



Abbott Architect Ci4100	1	*Positive	1	100.00%
Alere iCassette	2	*Positive	2	100.00%
Roche Hitachi 717	1	*Positive	1	100.00%
US Diagnostics UScreen Cup	4	*Positive	4	100.00%
Abbott Architect c4000	1	*Positive	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Positive	1	100.00%
Diag Test Group Clarity DOA Screen	4	*Positive	4	100.00%
ThermoScientific Indiko Plus	1	*Positive	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Positive	3	100.00%
Mindray 480	1	*Positive	1	100.00%
Hemosure First Sign UDS Dipstick	22	*Positive	22	100.00%
Siemens Viva/ProE System	1	*Positive	1	100.00%
TFS DRI Assay	1	*Positive	1	100.00%
Premier Biotech Bio-Cup	13	*Positive	12	92.31%
UCP Biosciences Home Drug Screening Test Cup	1	*Positive	1	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	Negative	1	100.00%
Alere iCup Drugs of Abuse	19	Negative	1	5.26%
CLIAwaived Inc Rapid Drug Cup	4	Negative	1	25.00%
Premier Biotech Bio-Cup	13	Negative	1	7.69%
		Amphetamines		
UDS-1		* Denotes Correct Response		
PG: Hemosure First Sign UDS Dipstick	22			
		*Positive	22	100.00%
UDS-2		* Denotes Correct Response		
PG: Hemosure First Sign UDS Dipstick	22			
		*Negative	22	100.00%
UDS-1		* Denotes Correct Response		
AG: All Methods	156			
		*Positive	154	98.72%
		Negative	2	1.28%
AlereTriage	3	*Positive	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	2	*Positive	2	100.00%
E-Screen/Alere	21	*Positive	21	100.00%
Express Diag. Express Check	9	*Positive	9	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Positive	1	100.00%
Innovacon One Step	1	*Positive	1	100.00%
Alere iCup Drugs of Abuse	22	*Positive	22	100.00%
Instant Technologies i Screen	1	*Positive	1	100.00%
MedTox Profile II	2	*Positive	2	100.00%
Noble Med Split Specimen Cup	1	*Positive	1	100.00%
Beckman (Olympus) AU	1	*Positive	1	100.00%
Innovacon Multi-Clin	9	*Positive	9	100.00%
MedTox Profile III ER	1	*Positive	1	100.00%
CLIAwaived Inc Rapid Drug Cup	4	*Positive	4	100.00%
ASC One Screen Drug Test Card	1	*Positive	1	100.00%
MedTox Profile V	2	*Positive	2	100.00%
Jant Accutest Multi Drug Screen	3	*Positive	3	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Positive	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Positive	4	100.00%
Mindray BS-200	2	*Positive	2	100.00%
Abbott Architect Ci4100	1	*Positive	1	100.00%
Alere iCassette	5	*Positive	4	80.00%
Roche Hitachi 717	1	*Positive	1	100.00%
US Diagnostics UScreen Cup	4	*Positive	4	100.00%
Abbott Architect c4000	1	*Positive	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Positive	1	100.00%
Diag Test Group Clarity DOA Screen	4	*Positive	4	100.00%
ThermoScientific Indiko Plus	1	*Positive	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Positive	3	100.00%
Mindray 480	1	*Positive	1	100.00%
Hemosure First Sign UDS Dipstick	22	*Positive	22	100.00%
Siemens Viva/ProE System	2	*Positive	2	100.00%
TFS DRI Assay	1	*Positive	1	100.00%
Premier Biotech Bio-Cup	13	*Positive	12	92.31%
UCP Biosciences Home Drug Screening Test Cup	1	*Positive	1	100.00%
Alere iCassette	5	Negative	1	20.00%
Premier Biotech Bio-Cup	13	Negative	1	7.69%
UDS-2		* Denotes Correct Response		



AG: All Methods	155			
		*Negative	153	98.71%
		Positive	2	1.29%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	2	*Negative	2	100.00%
E-Screen/Alere	21	*Negative	21	100.00%
Express Diag. Express Check	9	*Negative	9	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Innovacon One Step	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	21	*Negative	19	90.48%
Instant Technologies i Screen	1	*Negative	1	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	9	*Negative	9	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	4	*Negative	4	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	3	*Negative	3	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Alere iCassette	5	*Negative	5	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	4	*Negative	4	100.00%
ThermoScientific Indiko Plus	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Mindray 480	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	22	*Negative	22	100.00%
Siemens Viva/ProE System	2	*Negative	2	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	13	*Negative	13	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	21	Positive	2	9.52%
Cannabinoids (THC)				
UDS-1		* Denotes Correct Response		
AG: All Methods	185			
		*Negative	185	100.00%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	2	*Negative	2	100.00%
E-Screen/Alere	48	*Negative	48	100.00%
Express Diag. Express Check	9	*Negative	9	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Innovacon One Step	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	20	*Negative	20	100.00%
Instant Technologies i Screen	2	*Negative	2	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	4	*Negative	4	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	3	*Negative	3	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Alere iCassette	5	*Negative	5	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%



US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	4	*Negative	4	100.00%
ThermoScientific Indiko Plus	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Mindray 480	1	*Negative	1	100.00%
CRL Stat Cup	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	22	*Negative	22	100.00%
Siemens Viva/ProE System	2	*Negative	2	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	13	*Negative	13	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
UDS-2		* Denotes Correct Response		
AG: All Methods	185			
		*Negative	185	100.00%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	2	*Negative	2	100.00%
E-Screen/Alere	48	*Negative	48	100.00%
Express Diag. Express Check	9	*Negative	9	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Innovacon One Step	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	20	*Negative	20	100.00%
Instant Technologies i Screen	2	*Negative	2	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	4	*Negative	4	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	3	*Negative	3	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Alere iCassette	5	*Negative	5	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	4	*Negative	4	100.00%
ThermoScientific Indiko Plus	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Mindray 480	1	*Negative	1	100.00%
CRL Stat Cup	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	22	*Negative	22	100.00%
Siemens Viva/ProE System	2	*Negative	2	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	13	*Negative	13	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
UDS-1		* Denotes Correct Response		
AG: All Methods	184			
		*Negative	184	100.00%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	2	*Negative	2	100.00%
E-Screen/Alere	48	*Negative	48	100.00%
Express Diag. Express Check	9	*Negative	9	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Innovacon One Step	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	21	*Negative	21	100.00%
Instant Technologies i Screen	2	*Negative	2	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%



Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	9	*Negative	9	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	4	*Negative	4	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	3	*Negative	3	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Alere iCassette	5	*Negative	5	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	4	*Negative	4	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Mindray 480	1	*Negative	1	100.00%
CRL Stat Cup	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	22	*Negative	22	100.00%
Siemens Viva/ProE System	2	*Negative	2	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	13	*Negative	13	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
UDS-2		* Denotes Correct Response		
AG: All Methods	184			
		*Negative	184	100.00%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	2	*Negative	2	100.00%
E-Screen/Alere	48	*Negative	48	100.00%
Express Diag. Express Check	9	*Negative	9	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Innovacon One Step	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	21	*Negative	21	100.00%
Instant Technologies i Screen	2	*Negative	2	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	9	*Negative	9	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	4	*Negative	4	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	3	*Negative	3	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Alere iCassette	5	*Negative	5	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	4	*Negative	4	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Mindray 480	1	*Negative	1	100.00%
CRL Stat Cup	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	22	*Negative	22	100.00%
Siemens Viva/ProE System	2	*Negative	2	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	13	*Negative	13	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
		Ethanol		
UDS-1		* Denotes Correct Response		
AG: All Methods	6			



		*Negative	6	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
Mindray 480	1	*Negative	1	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
UDS-2		* Denotes Correct Response		
AG: All Methods	6			
		*Negative	6	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
Mindray 480	1	*Negative	1	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
		Acetaminophen		
UDS-1		* Denotes Correct Response		
PG: E-Screen/Alere	1			
		*Negative	1	100.00%
UDS-2		* Denotes Correct Response		
PG: E-Screen/Alere	1			
		*Negative	1	100.00%
UDS-1		* Denotes Correct Response		
AG: All Methods	4			
		*Negative	4	100.00%
AlereTriage	3	*Negative	3	100.00%
E-Screen/Alere	1	*Negative	1	100.00%
UDS-2		* Denotes Correct Response		
AG: All Methods	4			
		*Negative	4	100.00%
AlereTriage	3	*Negative	3	100.00%
E-Screen/Alere	1	*Negative	1	100.00%
		MDMA (Ecstasy)		
UDS-1		* Denotes Correct Response		
AG: All Methods	97			
		*Positive	95	97.94%
		Negative	2	2.06%
E-Screen/Alere	4	*Positive	4	100.00%
Express Diag. Express Check	8	*Positive	8	100.00%
Alere iCup Drugs of Abuse	11	*Positive	11	100.00%
Innovacon Multi-Clin	10	*Positive	10	100.00%
CLIAwaived Inc Rapid Drug Cup	3	*Positive	3	100.00%
Jant Accutest Multi Drug Screen	3	*Positive	3	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Positive	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	2	*Positive	2	100.00%
Alere iCassette	2	*Positive	2	100.00%
US Diagnostics UScreen Cup	4	*Positive	4	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Positive	1	100.00%
Diag Test Group Clarity DOA Screen	4	*Positive	4	100.00%
Noble Medical Insta-Screen Test Cup	3	*Positive	3	100.00%
Hemasure First Sign UDS Dipstick	22	*Positive	22	100.00%
Premier Biotech Bio-Cup	13	*Positive	12	92.31%
UCP Biosciences Home Drug Screening Test Cup	1	*Positive	1	100.00%
Beckman (Olympus) AU	1	Negative	1	100.00%
Premier Biotech Bio-Cup	13	Negative	1	7.69%
UDS-2		* Denotes Correct Response		
AG: All Methods	96			
		*Negative	93	96.88%
		Positive	3	3.13%
E-Screen/Alere	4	*Negative	4	100.00%
Express Diag. Express Check	8	*Negative	8	100.00%
Alere iCup Drugs of Abuse	10	*Negative	9	90.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
CLIAwaived Inc Rapid Drug Cup	3	*Negative	3	100.00%
Jant Accutest Multi Drug Screen	3	*Negative	3	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	2	*Negative	2	100.00%



Alere iCassette	2	*Negative	2	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	4	*Negative	4	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Hemosure First Sign UDS Dipstick	22	*Negative	22	100.00%
Premier Biotech Bio-Cup	13	*Negative	12	92.31%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	10	Positive	1	10.00%
Beckman (Olympus) AU	1	Positive	1	100.00%
Premier Biotech Bio-Cup	13	Positive	1	7.69%
Methadone				
UDS-1				
* Denotes Correct Response				
AG: All Methods	112			
		*Negative	108	96.43%
		Positive	4	3.57%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	1	*Negative	1	100.00%
E-Screen/Alere	24	*Negative	24	100.00%
Express Diag. Express Check	9	*Negative	9	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	12	*Negative	10	83.33%
MedTox Profile III ER	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	4	*Negative	3	75.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	3	*Negative	3	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	4	*Negative	4	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	2	66.67%
Mindray 480	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	22	*Negative	22	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	12	Positive	2	16.67%
CLIAwaived Inc Rapid Drug Cup	4	Positive	1	25.00%
Noble Medical Insta-Screen Test Cup	3	Positive	1	33.33%
UDS-2				
* Denotes Correct Response				
AG: All Methods	112			
		*Negative	111	99.11%
		Positive	1	0.89%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	1	*Negative	1	100.00%
E-Screen/Alere	24	*Negative	24	100.00%
Express Diag. Express Check	9	*Negative	9	100.00%
Alere iCup Drugs of Abuse	12	*Negative	12	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	4	*Negative	4	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	3	*Negative	3	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	4	*Negative	4	100.00%



Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Mindray 480	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	22	*Negative	22	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	Positive	1	100.00%
Methamphetamine				
UDS-1				
* Denotes Correct Response				
AG: All Methods	157			
		*Positive	148	94.27%
		Negative	9	5.73%
AlereTriage	3	*Positive	3	100.00%
E-Screen/Alere	47	*Positive	46	97.87%
Express Diag. Express Check	9	*Positive	9	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Positive	1	100.00%
Innovacon One Step	1	*Positive	1	100.00%
Alere iCup Drugs of Abuse	21	*Positive	20	95.24%
Instant Technologies i Screen	1	*Positive	1	100.00%
Integrated E-Z Split Key Cup	1	*Positive	1	100.00%
MedTox Profile II	1	*Positive	1	100.00%
MedTox Profile III ER	1	*Positive	1	100.00%
CLIAwaived Inc Rapid Drug Cup	4	*Positive	3	75.00%
ASC One Screen Drug Test Card	1	*Positive	1	100.00%
MedTox Profile V	2	*Positive	2	100.00%
Jant Accutest Multi Drug Screen	3	*Positive	3	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Positive	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Positive	4	100.00%
Alere iCassette	1	*Positive	1	100.00%
US Diagnostics UScreen Cup	4	*Positive	4	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Positive	1	100.00%
Diag Test Group Clarity DOA Screen	4	*Positive	4	100.00%
Noble Medical Insta-Screen Test Cup	3	*Positive	2	66.67%
CRL Stat Cup	1	*Positive	1	100.00%
Hemosure First Sign UDS Dipstick	22	*Positive	20	90.91%
Premier Biotech Bio-Cup	13	*Positive	12	92.31%
UCP Biosciences Home Drug Screening Test Cup	1	*Positive	1	100.00%
E-Screen/Alere	47	Negative	1	2.13%
Alere iCup Drugs of Abuse	21	Negative	1	4.76%
Noble Med Split Specimen Cup	1	Negative	1	100.00%
Beckman (Olympus) AU	1	Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	4	Negative	1	25.00%
Noble Medical Insta-Screen Test Cup	3	Negative	1	33.33%
Hemosure First Sign UDS Dipstick	22	Negative	2	9.09%
Premier Biotech Bio-Cup	13	Negative	1	7.69%
UDS-2				
* Denotes Correct Response				
AG: All Methods	156			
		*Negative	153	98.08%
		Positive	3	1.92%
AlereTriage	3	*Negative	3	100.00%
E-Screen/Alere	47	*Negative	46	97.87%
Express Diag. Express Check	9	*Negative	9	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Innovacon One Step	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	20	*Negative	19	95.00%
Instant Technologies i Screen	1	*Negative	1	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	4	*Negative	4	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	3	*Negative	3	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Alere iCassette	1	*Negative	1	100.00%



US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	4	*Negative	4	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
CRL Stat Cup	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	22	*Negative	22	100.00%
Premier Biotech Bio-Cup	13	*Negative	12	92.31%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
E-Screen/Alere	47	Positive	1	2.13%
Alere iCup Drugs of Abuse	20	Positive	1	5.00%
Premier Biotech Bio-Cup	13	Positive	1	7.69%
Methaqualone				
UDS-1				
* Denotes Correct Response				
AG: All Methods	1			
		*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
UDS-2				
* Denotes Correct Response				
AG: All Methods	1			
		*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Opiates				
UDS-1				
* Denotes Correct Response				
AG: All Methods	184			
		*Negative	183	99.46%
		Positive	1	0.54%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	2	*Negative	2	100.00%
E-Screen/Alere	48	*Negative	48	100.00%
Express Diag. Express Check	9	*Negative	9	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	21	*Negative	21	100.00%
Instant Technologies i Screen	2	*Negative	2	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Noble Med Split Specimen Cup	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	4	*Negative	4	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	3	*Negative	3	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Alere iCassette	5	*Negative	5	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
ThermoScientific Indiko Plus	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Mindray 480	1	*Negative	1	100.00%
CRL Stat Cup	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	22	*Negative	22	100.00%
Siemens Viva/ProE System	2	*Negative	2	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	13	*Negative	13	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Innovacon One Step	1	Positive	1	100.00%
UDS-2				
* Denotes Correct Response				
AG: All Methods	184			
		Negative	146	79.35%
		Positive	38	20.65%
AlereTriage	3	Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	2	Negative	1	50.00%



E-Screen/Alere	48	Negative	48	100.00%
Express Diag. Express Check	9	Negative	4	44.44%
Exp Results Integ Multi-Drug Screen Cup	1	Negative	1	100.00%
Innovacon One Step	1	Negative	1	100.00%
Alere iCup Drugs of Abuse	21	Negative	15	71.43%
Instant Technologies i Screen	2	Negative	2	100.00%
Integrated E-Z Split Key Cup	1	Negative	1	100.00%
Beckman (Olympus) AU	1	Negative	1	100.00%
Innovacon Multi-Clin	10	Negative	3	30.00%
CLIAwaived Inc Rapid Drug Cup	4	Negative	1	25.00%
ASC One Screen Drug Test Card	1	Negative	1	100.00%
Jant Accutest Multi Drug Screen	3	Negative	3	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	Negative	3	75.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	Negative	4	100.00%
Mindray BS-200	2	Negative	2	100.00%
Abbott Architect Ci4100	1	Negative	1	100.00%
Alere iCassette	5	Negative	3	60.00%
US Diagnostics UScreen Cup	4	Negative	4	100.00%
Abbott Architect c4000	1	Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	Negative	3	100.00%
ThermoScientific Indiko Plus	1	Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	Negative	3	100.00%
Mindray 480	1	Negative	1	100.00%
CRL Stat Cup	1	Negative	1	100.00%
Hemosure First Sign UDS Dipstick	22	Negative	22	100.00%
TFS DRI Assay	1	Negative	1	100.00%
Premier Biotech Bio-Cup	13	Negative	10	76.92%
UCP Biosciences Home Drug Screening Test Cup	1	Negative	1	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	2	Positive	1	50.00%
Express Diag. Express Check	9	Positive	5	55.56%
Alere iCup Drugs of Abuse	21	Positive	6	28.57%
MedTox Profile II	1	Positive	1	100.00%
Noble Med Split Specimen Cup	1	Positive	1	100.00%
Innovacon Multi-Clin	10	Positive	7	70.00%
MedTox Profile III ER	1	Positive	1	100.00%
CLIAwaived Inc Rapid Drug Cup	4	Positive	3	75.00%
MedTox Profile V	2	Positive	2	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	Positive	1	25.00%
Nova Century Scientific Nova Screen DOA Cassette	1	Positive	1	100.00%
Alere iCassette	5	Positive	2	40.00%
Roche Hitachi 717	1	Positive	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	Positive	1	100.00%
Siemens Viva/ProE System	2	Positive	2	100.00%
Premier Biotech Bio-Cup	13	Positive	3	23.08%
		Oxycodone		
UDS-1		* Denotes Correct Response		
AG: All Methods	109			
		*Negative	109	100.00%
E-Screen/Alere	16	*Negative	16	100.00%
Express Diag. Express Check	8	*Negative	8	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Innovacon One Step	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	17	*Negative	17	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	4	*Negative	4	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	3	*Negative	3	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	4	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Alere iCassette	2	*Negative	2	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	4	*Negative	4	100.00%



Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Mindray 480	1	*Negative	1	100.00%
Siemens Viva/ProE System	1	*Negative	1	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	13	*Negative	13	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
UDS-2		* Denotes Correct Response		
AG: All Methods	109			
		*Negative	101	92.66%
		Positive	8	7.34%
E-Screen/Alere	16	*Negative	16	100.00%
Express Diag. Express Check	8	*Negative	8	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Innovacon One Step	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	17	*Negative	14	82.35%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	9	90.00%
MedTox Profile III ER	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	4	*Negative	3	75.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	3	*Negative	3	100.00%
Amedica Biotech AmediCheck Instant Test Cup	4	*Negative	3	75.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	2	50.00%
Mindray BS-200	2	*Negative	2	100.00%
Alere iCassette	2	*Negative	2	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	4	*Negative	4	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	4	*Negative	4	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Mindray 480	1	*Negative	1	100.00%
Siemens Viva/ProE System	1	*Negative	1	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	13	*Negative	13	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	17	Positive	3	17.65%
Innovacon Multi-Clin	10	Positive	1	10.00%
CLIAwaived Inc Rapid Drug Cup	4	Positive	1	25.00%
Amedica Biotech AmediCheck Instant Test Cup	4	Positive	1	25.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	Positive	2	50.00%
UDS-1		* Denotes Correct Response		
AG: All Methods	160			
		*Negative	159	99.38%
		Positive	1	0.63%
AlereTriage	3	*Negative	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	1	*Negative	1	100.00%
E-Screen/Alere	48	*Negative	48	100.00%
Express Diag. Express Check	9	*Negative	9	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Innovacon One Step	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	9	*Negative	8	88.89%
Instant Technologies i Screen	2	*Negative	2	100.00%
Integrated E-Z Split Key Cup	1	*Negative	1	100.00%
MedTox Profile II	2	*Negative	2	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	10	*Negative	10	100.00%
MedTox Profile III ER	1	*Negative	1	100.00%
CLIAwaived Inc Rapid Drug Cup	4	*Negative	4	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	3	*Negative	3	100.00%
Amedica Biotech AmediCheck Instant Test Cup	1	*Negative	1	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Negative	4	100.00%



Mindray BS-200	1	*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Alere iCassette	4	*Negative	4	100.00%
Roche Hitachi 717	1	*Negative	1	100.00%
US Diagnostics UScreen Cup	1	*Negative	1	100.00%
Abbott Architect c4000	1	*Negative	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	4	*Negative	4	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
CRL Stat Cup	1	*Negative	1	100.00%
Hemosure First Sign UDS Dipstick	22	*Negative	22	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	13	*Negative	13	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	9	Positive	1	11.11%
UDS-2		* Denotes Correct Response		
AG: All Methods	161			
		*Positive	155	96.27%
		Negative	6	3.73%
AlereTriage	3	*Positive	3	100.00%
Siemens Dim/AR/ES/RxL/Xpand/EXL	1	*Positive	1	100.00%
E-Screen/Alere	48	*Positive	48	100.00%
Express Diag. Express Check	9	*Positive	9	100.00%
Innovacon One Step	1	*Positive	1	100.00%
Alere iCup Drugs of Abuse	10	*Positive	7	70.00%
Instant Technologies i Screen	2	*Positive	2	100.00%
Integrated E-Z Split Key Cup	1	*Positive	1	100.00%
MedTox Profile II	2	*Positive	2	100.00%
Beckman (Olympus) AU	1	*Positive	1	100.00%
Innovacon Multi-Clin	10	*Positive	10	100.00%
MedTox Profile III ER	1	*Positive	1	100.00%
CLIAwaived Inc Rapid Drug Cup	4	*Positive	4	100.00%
ASC One Screen Drug Test Card	1	*Positive	1	100.00%
MedTox Profile V	2	*Positive	2	100.00%
Jant Accutest Multi Drug Screen	3	*Positive	3	100.00%
Amedica Biotech AmediCheck Instant Test Cup	1	*Positive	1	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Positive	1	100.00%
American Screening Corp Reveal Multi-Drug Test Cup	4	*Positive	4	100.00%
Mindray BS-200	1	*Positive	1	100.00%
Abbott Architect Ci4100	1	*Positive	1	100.00%
Alere iCassette	4	*Positive	3	75.00%
Roche Hitachi 717	1	*Positive	1	100.00%
US Diagnostics UScreen Cup	1	*Positive	1	100.00%
Abbott Architect c4000	1	*Positive	1	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Positive	1	100.00%
Diag Test Group Clarity DOA Screen	4	*Positive	4	100.00%
Noble Medical Insta-Screen Test Cup	3	*Positive	3	100.00%
CRL Stat Cup	1	*Positive	1	100.00%
Hemosure First Sign UDS Dipstick	22	*Positive	22	100.00%
TFS DRI Assay	1	*Positive	1	100.00%
Premier Biotech Bio-Cup	13	*Positive	12	92.31%
UCP Biosciences Home Drug Screening Test Cup	1	*Positive	1	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	Negative	1	100.00%
Alere iCup Drugs of Abuse	10	Negative	3	30.00%
Alere iCassette	4	Negative	1	25.00%
Premier Biotech Bio-Cup	13	Negative	1	7.69%
		Propoxyphene		
UDS-1		* Denotes Correct Response		
AG: All Methods	26			
		*Negative	26	100.00%
E-Screen/Alere	4	*Negative	4	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	6	*Negative	6	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	7	*Negative	7	100.00%
CLIAwaived Inc Rapid Drug Cup	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%



Alere iCassette	2	*Negative	2	100.00%
UDS-2		* Denotes Correct Response		
AG: All Methods	26			
		*Negative	26	100.00%
E-Screen/Alere	4	*Negative	4	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	6	*Negative	6	100.00%
MedTox Profile II	1	*Negative	1	100.00%
Beckman (Olympus) AU	1	*Negative	1	100.00%
Innovacon Multi-Clin	7	*Negative	7	100.00%
CLIAwaived Inc Rapid Drug Cup	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
Alere iCassette	2	*Negative	2	100.00%
		Tricyclic Antidepressants		
UDS-1		* Denotes Correct Response		
AG: All Methods	51			
		*Negative	51	100.00%
AlereTriage	3	*Negative	3	100.00%
Express Diag. Express Check	8	*Negative	8	100.00%
Alere iCup Drugs of Abuse	9	*Negative	9	100.00%
Innovacon Multi-Clin	9	*Negative	9	100.00%
CLIAwaived Inc Rapid Drug Cup	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	3	*Negative	3	100.00%
Amedica Biotech AmediCheck Instant Test Cup	1	*Negative	1	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
Alere iCassette	2	*Negative	2	100.00%
US Diagnostics UScreen Cup	3	*Negative	3	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Mindray 480	1	*Negative	1	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
UDS-2		* Denotes Correct Response		
AG: All Methods	51			
		*Negative	50	98.04%
		Positive	1	1.96%
AlereTriage	3	*Negative	3	100.00%
Express Diag. Express Check	8	*Negative	8	100.00%
Alere iCup Drugs of Abuse	9	*Negative	8	88.89%
Innovacon Multi-Clin	9	*Negative	9	100.00%
CLIAwaived Inc Rapid Drug Cup	1	*Negative	1	100.00%
MedTox Profile V	2	*Negative	2	100.00%
Jant Accutest Multi Drug Screen	3	*Negative	3	100.00%
Amedica Biotech AmediCheck Instant Test Cup	1	*Negative	1	100.00%
Nova Century Scientific Nova Screen DOA Cassette	1	*Negative	1	100.00%
Alere iCassette	2	*Negative	2	100.00%
US Diagnostics UScreen Cup	3	*Negative	3	100.00%
Bio-Rad Tox/See Rapid Urine Drug Screen	1	*Negative	1	100.00%
Diag Test Group Clarity DOA Screen	3	*Negative	3	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Mindray 480	1	*Negative	1	100.00%
UCP Biosciences Home Drug Screening Test Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	9	Positive	1	11.11%
		Buprenorphine		
UDS-1		* Denotes Correct Response		
AG: All Methods	42			
		*Negative	42	100.00%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	10	*Negative	10	100.00%
CLIAwaived Inc Rapid Drug Cup	3	*Negative	3	100.00%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Diag Test Group Clarity DOA Screen	1	*Negative	1	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%



Mindray 480	1	*Negative	1	100.00%
Siemens Viva/ProE System	2	*Negative	2	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	13	*Negative	13	100.00%
UDS-2		* Denotes Correct Response		
AG: All Methods	42			
		*Negative	41	97.62%
		Positive	1	2.38%
Exp Results Integ Multi-Drug Screen Cup	1	*Negative	1	100.00%
Alere iCup Drugs of Abuse	10	*Negative	10	100.00%
CLIAwaived Inc Rapid Drug Cup	3	*Negative	2	66.67%
ASC One Screen Drug Test Card	1	*Negative	1	100.00%
MedTox Profile V	1	*Negative	1	100.00%
Jant Accutest Multi Drug Screen	2	*Negative	2	100.00%
Mindray BS-200	2	*Negative	2	100.00%
Diag Test Group Clarity DOA Screen	1	*Negative	1	100.00%
Alere iScreen Drugs of Abuse Card	1	*Negative	1	100.00%
Noble Medical Insta-Screen Test Cup	3	*Negative	3	100.00%
Mindray 480	1	*Negative	1	100.00%
Siemens Viva/ProE System	2	*Negative	2	100.00%
TFS DRI Assay	1	*Negative	1	100.00%
Premier Biotech Bio-Cup	13	*Negative	13	100.00%
CLIAwaived Inc Rapid Drug Cup	3	Positive	1	33.33%
		6-Monoacetylmorphine (6 AM)		
UDS-1		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Siemens Viva/ProE System	2	*Negative	2	100.00%
UDS-2		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Siemens Viva/ProE System	2	*Negative	2	100.00%
		EDDP		
UDS-1		* Denotes Correct Response		
AG: All Methods	15			
		*Negative	15	100.00%
Siemens Viva/ProE System	2	*Negative	2	100.00%
Premier Biotech Bio-Cup	13	*Negative	13	100.00%
UDS-2		* Denotes Correct Response		
AG: All Methods	15			
		*Negative	15	100.00%
Siemens Viva/ProE System	2	*Negative	2	100.00%
Premier Biotech Bio-Cup	13	*Negative	13	100.00%
756 - Viral Antibodies	Labs	Response	Number	Percent
		HSV Ab (Herpes Simplex Virus)		
VA-1		* Denotes Correct Response		
PG: Biokit HSV-II	2			
		*Negative	2	100.00%
770 - Viral Antigen Detection;792 - Viral Antigen Detection - Waived Methods;797 - Viral Antigen Detection II	Labs	Response	Number	Percent
		Rotavirus - Ag Detect		
VAD-1		* Denotes Correct Response		
AG: All Methods	8			
		*Positive	8	100.00%
Meridian Immunocard Stat! Rotavirus	4	*Positive	4	100.00%
SAS Rota-Test	4	*Positive	4	100.00%
VAD-2		* Denotes Correct Response		
AG: All Methods	8			
		*Negative	8	100.00%
Meridian Immunocard Stat! Rotavirus	4	*Negative	4	100.00%
SAS Rota-Test	4	*Negative	4	100.00%
VAD-3		* Denotes Correct Response		
AG: All Methods	8			
		*Negative	8	100.00%
Meridian Immunocard Stat! Rotavirus	4	*Negative	4	100.00%
SAS Rota-Test	4	*Negative	4	100.00%
VAD-4		* Denotes Correct Response		



AG: All Methods	8			
		*Negative	7	87.50%
		Positive	1	12.50%
Meridian Immunocard Stat! Rotavirus	4	*Negative	3	75.00%
SAS Rota-Test	4	*Negative	4	100.00%
Meridian Immunocard Stat! Rotavirus	4	Positive	1	25.00%
VAD-5		* Denotes Correct Response		
AG: All Methods	8			
		*Positive	8	100.00%
Meridian Immunocard Stat! Rotavirus	4	*Positive	4	100.00%
SAS Rota-Test	4	*Positive	4	100.00%
		Influenza A - Ag Detect		
VAD-1		* Denotes Correct Response		
AG: All Methods	495			
		*Positive	478	96.57%
		Negative	17	3.43%
Binax Now Flu A	10	*Positive	9	90.00%
Alere BinaxNOW Influenza A+B Card	60	*Positive	57	95.00%
Sekisui Diagnostics OSOM Influenza A & B	10	*Positive	9	90.00%
Quidel QuickVue A+B	42	*Positive	40	95.24%
Quidel Sofia/Sofia 2 Influenza A+B	148	*Positive	144	97.30%
Alere Influenza A & B	18	*Positive	18	100.00%
BD Veritor Flu A+B	81	*Positive	81	100.00%
Henry Schein OneStep + Influenza A&B (distinguish)	11	*Positive	10	90.91%
Medline Influenza A and B	1	*Positive	1	100.00%
Cepheid GeneXpert - Xpert Flu	3	*Positive	3	100.00%
Alere i Influenza A & B	17	*Positive	17	100.00%
Roche Cobas Liat Influenza A/B	8	*Positive	8	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	57	*Positive	52	91.23%
Consult Diag Influenza A & B	14	*Positive	14	100.00%
Lifesign Status Influenza A & B	15	*Positive	15	100.00%
Binax Now Flu A	10	Negative	1	10.00%
Alere BinaxNOW Influenza A+B Card	60	Negative	3	5.00%
Sekisui Diagnostics OSOM Influenza A & B	10	Negative	1	10.00%
Quidel QuickVue A+B	42	Negative	2	4.76%
Quidel Sofia/Sofia 2 Influenza A+B	148	Negative	4	2.70%
Henry Schein OneStep + Influenza A&B (distinguish)	11	Negative	1	9.09%
Sekisui Diagnostics OSOM Ultra Influenza A & B	57	Negative	5	8.77%
VAD-2		* Denotes Correct Response		
AG: All Methods	488			
		*Negative	469	96.11%
		Positive	19	3.89%
Binax Now Flu A	10	*Negative	10	100.00%
Alere BinaxNOW Influenza A+B Card	60	*Negative	57	95.00%
Sekisui Diagnostics OSOM Influenza A & B	10	*Negative	10	100.00%
Quidel QuickVue A+B	42	*Negative	39	92.86%
Quidel Sofia/Sofia 2 Influenza A+B	147	*Negative	144	97.96%
Alere Influenza A & B	18	*Negative	17	94.44%
BD Veritor Flu A+B	78	*Negative	77	98.72%
Henry Schein OneStep + Influenza A&B (distinguish)	10	*Negative	9	90.00%
Medline Influenza A and B	1	*Negative	1	100.00%
Cepheid GeneXpert - Xpert Flu	3	*Negative	3	100.00%
Alere i Influenza A & B	17	*Negative	17	100.00%
Roche Cobas Liat Influenza A/B	7	*Negative	7	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	56	*Negative	51	91.07%
Consult Diag Influenza A & B	14	*Negative	14	100.00%
Lifesign Status Influenza A & B	15	*Negative	13	86.67%
Alere BinaxNOW Influenza A+B Card	60	Positive	3	5.00%
Quidel QuickVue A+B	42	Positive	3	7.14%
Quidel Sofia/Sofia 2 Influenza A+B	147	Positive	3	2.04%
Alere Influenza A & B	18	Positive	1	5.56%
BD Veritor Flu A+B	78	Positive	1	1.28%
Henry Schein OneStep + Influenza A&B (distinguish)	10	Positive	1	10.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	56	Positive	5	8.93%
Lifesign Status Influenza A & B	15	Positive	2	13.33%
VAD-3		* Denotes Correct Response		
AG: All Methods	53			
		*Positive	50	94.34%
		Negative	3	5.66%



Alere BinaxNOW Influenza A+B Card	5	*Positive	5	100.00%
Sekisui Diagnostics OSOM Influenza A & B	4	*Positive	4	100.00%
Quidel QuickVue A+B	12	*Positive	11	91.67%
Quidel Sofia/Sofia 2 Influenza A+B	18	*Positive	16	88.89%
Alere Influenza A & B	2	*Positive	2	100.00%
BD Veritor Flu A+B	6	*Positive	6	100.00%
Cepheid GeneXpert - Xpert Flu	3	*Positive	3	100.00%
Alere i Influenza A & B	2	*Positive	2	100.00%
Consult Diag Influenza A & B	1	*Positive	1	100.00%
Quidel QuickVue A+B	12	Negative	1	8.33%
Quidel Sofia/Sofia 2 Influenza A+B	18	Negative	2	11.11%
VAD-4		* Denotes Correct Response		
AG: All Methods	53			
		*Negative	52	98.11%
		Positive	1	1.89%
Alere BinaxNOW Influenza A+B Card	5	*Negative	5	100.00%
Sekisui Diagnostics OSOM Influenza A & B	4	*Negative	4	100.00%
Quidel QuickVue A+B	12	*Negative	11	91.67%
Quidel Sofia/Sofia 2 Influenza A+B	18	*Negative	18	100.00%
Alere Influenza A & B	2	*Negative	2	100.00%
BD Veritor Flu A+B	6	*Negative	6	100.00%
Cepheid GeneXpert - Xpert Flu	3	*Negative	3	100.00%
Alere i Influenza A & B	2	*Negative	2	100.00%
Consult Diag Influenza A & B	1	*Negative	1	100.00%
Quidel QuickVue A+B	12	Positive	1	8.33%
VAD-5		* Denotes Correct Response		
AG: All Methods	53			
		*Negative	50	94.34%
		Positive	3	5.66%
Alere BinaxNOW Influenza A+B Card	5	*Negative	5	100.00%
Sekisui Diagnostics OSOM Influenza A & B	4	*Negative	4	100.00%
Quidel QuickVue A+B	12	*Negative	11	91.67%
Quidel Sofia/Sofia 2 Influenza A+B	18	*Negative	16	88.89%
Alere Influenza A & B	2	*Negative	2	100.00%
BD Veritor Flu A+B	6	*Negative	6	100.00%
Cepheid GeneXpert - Xpert Flu	3	*Negative	3	100.00%
Alere i Influenza A & B	2	*Negative	2	100.00%
Consult Diag Influenza A & B	1	*Negative	1	100.00%
Quidel QuickVue A+B	12	Positive	1	8.33%
Quidel Sofia/Sofia 2 Influenza A+B	18	Positive	2	11.11%
		Influenza A/B - Aq Detect		
VAD-1		* Denotes Correct Response		
AG: All Methods	28			
		*Positive	28	100.00%
Alere BinaxNOW Influenza A+B Card	1	*Positive	1	100.00%
Quidel QuickVue	17	*Positive	17	100.00%
Henry Schein OneStep + Influenza A/B	10	*Positive	10	100.00%
VAD-2		* Denotes Correct Response		
AG: All Methods	28			
		*Positive	28	100.00%
Alere BinaxNOW Influenza A+B Card	1	*Positive	1	100.00%
Quidel QuickVue	17	*Positive	17	100.00%
Henry Schein OneStep + Influenza A/B	10	*Positive	10	100.00%
VAD-3		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Quidel QuickVue	2	*Positive	2	100.00%
VAD-4		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Quidel QuickVue	2	*Negative	2	100.00%
VAD-5		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Quidel QuickVue	2	*Positive	2	100.00%
		Influenza B - Aq Detect		
VAD-1		* Denotes Correct Response		
AG: All Methods	490			



		*Negative	475	96.94%
		Positive	15	3.06%
Binax Now Flu B	10	*Negative	10	100.00%
Alere BinaxNOW Influenza A+B Card	60	*Negative	58	96.67%
Sekisui Diagnostics OSOM Influenza A & B	10	*Negative	9	90.00%
Quidel QuickVue A+B	42	*Negative	39	92.86%
Quidel Sofia/Sofia 2 Influenza A+B	147	*Negative	146	99.32%
Alere Influenza A & B	18	*Negative	18	100.00%
BD Veritor Flu A+B	80	*Negative	80	100.00%
Henry Schein OneStep + Influenza A&B (distinguish)	10	*Negative	9	90.00%
Medline Influenza A and B	1	*Negative	1	100.00%
Cepheid GeneXpert - Xpert Flu	3	*Negative	3	100.00%
Alere i Influenza A & B	17	*Negative	17	100.00%
Roche Cobas Liat Influenza A/B	7	*Negative	7	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	56	*Negative	51	91.07%
Consult Diag Influenza A & B	14	*Negative	14	100.00%
Lifesign Status Influenza A & B	15	*Negative	13	86.67%
Alere BinaxNOW Influenza A+B Card	60	Positive	2	3.33%
Sekisui Diagnostics OSOM Influenza A & B	10	Positive	1	10.00%
Quidel QuickVue A+B	42	Positive	3	7.14%
Quidel Sofia/Sofia 2 Influenza A+B	147	Positive	1	0.68%
Henry Schein OneStep + Influenza A&B (distinguish)	10	Positive	1	10.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	56	Positive	5	8.93%
Lifesign Status Influenza A & B	15	Positive	2	13.33%
VAD-2		* Denotes Correct Response		
AG: All Methods	493			
		*Positive	478	96.96%
		Negative	15	3.04%
Binax Now Flu B	10	*Positive	9	90.00%
Alere BinaxNOW Influenza A+B Card	60	*Positive	56	93.33%
Sekisui Diagnostics OSOM Influenza A & B	10	*Positive	10	100.00%
Quidel QuickVue A+B	42	*Positive	41	97.62%
Quidel Sofia/Sofia 2 Influenza A+B	148	*Positive	145	97.97%
Alere Influenza A & B	18	*Positive	17	94.44%
BD Veritor Flu A+B	79	*Positive	79	100.00%
Henry Schein OneStep + Influenza A&B (distinguish)	11	*Positive	10	90.91%
Medline Influenza A and B	1	*Positive	1	100.00%
Cepheid GeneXpert - Xpert Flu	3	*Positive	3	100.00%
Alere i Influenza A & B	17	*Positive	17	100.00%
Roche Cobas Liat Influenza A/B	8	*Positive	8	100.00%
Sekisui Diagnostics OSOM Ultra Influenza A & B	57	*Positive	53	92.98%
Consult Diag Influenza A & B	14	*Positive	14	100.00%
Lifesign Status Influenza A & B	15	*Positive	15	100.00%
Binax Now Flu B	10	Negative	1	10.00%
Alere BinaxNOW Influenza A+B Card	60	Negative	4	6.67%
Quidel QuickVue A+B	42	Negative	1	2.38%
Quidel Sofia/Sofia 2 Influenza A+B	148	Negative	3	2.03%
Alere Influenza A & B	18	Negative	1	5.56%
Henry Schein OneStep + Influenza A&B (distinguish)	11	Negative	1	9.09%
Sekisui Diagnostics OSOM Ultra Influenza A & B	57	Negative	4	7.02%
VAD-3		* Denotes Correct Response		
AG: All Methods	53			
		*Negative	52	98.11%
		Positive	1	1.89%
Alere BinaxNOW Influenza A+B Card	5	*Negative	5	100.00%
Sekisui Diagnostics OSOM Influenza A & B	4	*Negative	4	100.00%
Quidel QuickVue A+B	12	*Negative	12	100.00%
Quidel Sofia/Sofia 2 Influenza A+B	18	*Negative	17	94.44%
Alere Influenza A & B	2	*Negative	2	100.00%
BD Veritor Flu A+B	6	*Negative	6	100.00%
Cepheid GeneXpert - Xpert Flu	3	*Negative	3	100.00%
Alere i Influenza A & B	2	*Negative	2	100.00%
Consult Diag Influenza A & B	1	*Negative	1	100.00%
Quidel Sofia/Sofia 2 Influenza A+B	18	Positive	1	5.56%
VAD-4		* Denotes Correct Response		
AG: All Methods	53			
		*Negative	52	98.11%
		Positive	1	1.89%
Alere BinaxNOW Influenza A+B Card	5	*Negative	5	100.00%



Sekisui Diagnostics OSOM Influenza A & B	4	*Negative	4	100.00%
Quidel QuickVue A+B	12	*Negative	12	100.00%
Quidel Sofia/Sofia 2 Influenza A+B	18	*Negative	17	94.44%
Alere Influenza A & B	2	*Negative	2	100.00%
BD Veritor Flu A+B	6	*Negative	6	100.00%
Cepheid GeneXpert - Xpert Flu	3	*Negative	3	100.00%
Alere i Influenza A & B	2	*Negative	2	100.00%
Consult Diag Influenza A & B	1	*Negative	1	100.00%
Quidel Sofia/Sofia 2 Influenza A+B	18	Positive	1	5.56%
VAD-5		* Denotes Correct Response		
AG: All Methods	53			
		*Positive	49	92.45%
		Negative	4	7.55%
Alere BinaxNOW Influenza A+B Card	5	*Positive	5	100.00%
Sekisui Diagnostics OSOM Influenza A & B	4	*Positive	4	100.00%
Quidel QuickVue A+B	12	*Positive	11	91.67%
Quidel Sofia/Sofia 2 Influenza A+B	18	*Positive	15	83.33%
Alere Influenza A & B	2	*Positive	2	100.00%
BD Veritor Flu A+B	6	*Positive	6	100.00%
Cepheid GeneXpert - Xpert Flu	3	*Positive	3	100.00%
Alere i Influenza A & B	2	*Positive	2	100.00%
Consult Diag Influenza A & B	1	*Positive	1	100.00%
Quidel QuickVue A+B	12	Negative	1	8.33%
Quidel Sofia/Sofia 2 Influenza A+B	18	Negative	3	16.67%
VAD-1		RSV - Ag Detect		
		* Denotes Correct Response		
AG: All Methods	269			
		*Negative	260	96.65%
		Positive	9	3.35%
BD Directigen RSV	1	*Negative	1	100.00%
BD EZ RSV	1	*Negative	1	100.00%
Alere BinaxNOW RSV	163	*Negative	156	95.71%
Quidel QuickVue RSV	26	*Negative	26	100.00%
Wampole Clearview RSV	4	*Negative	4	100.00%
SAS RSV Alert	1	*Negative	1	100.00%
Quidel QuickVue RSV 10	2	*Negative	2	100.00%
Quidel Sofia/Sofia 2 RSV	32	*Negative	32	100.00%
BD Veritor RSV	36	*Negative	34	94.44%
Cepheid GeneXpert - Xpert RSV	1	*Negative	1	100.00%
Alere i RSV	2	*Negative	2	100.00%
Alere BinaxNOW RSV	163	Positive	7	4.29%
BD Veritor RSV	36	Positive	2	5.56%
VAD-2		* Denotes Correct Response		
AG: All Methods	269			
		*Negative	262	97.40%
		Positive	7	2.60%
BD Directigen RSV	1	*Negative	1	100.00%
BD EZ RSV	1	*Negative	1	100.00%
Alere BinaxNOW RSV	163	*Negative	157	96.32%
Quidel QuickVue RSV	26	*Negative	26	100.00%
Wampole Clearview RSV	4	*Negative	4	100.00%
SAS RSV Alert	1	*Negative	1	100.00%
Quidel QuickVue RSV 10	2	*Negative	2	100.00%
Quidel Sofia/Sofia 2 RSV	32	*Negative	32	100.00%
BD Veritor RSV	36	*Negative	35	97.22%
Cepheid GeneXpert - Xpert RSV	1	*Negative	1	100.00%
Alere i RSV	2	*Negative	2	100.00%
Alere BinaxNOW RSV	163	Positive	6	3.68%
BD Veritor RSV	36	Positive	1	2.78%
VAD-3		* Denotes Correct Response		
AG: All Methods	22			
		*Positive	22	100.00%
BD EZ RSV	1	*Positive	1	100.00%
Alere BinaxNOW RSV	8	*Positive	8	100.00%
Quidel QuickVue RSV	3	*Positive	3	100.00%
Wampole Clearview RSV	3	*Positive	3	100.00%
SAS RSV Alert	1	*Positive	1	100.00%
Quidel QuickVue RSV 10	2	*Positive	2	100.00%
Quidel Sofia/Sofia 2 RSV	3	*Positive	3	100.00%



Method	Count	Response	Count	Percent
Cepheid GeneXpert - Xpert RSV	1	*Positive	1	100.00%
VAD-4				
* Denotes Correct Response				
AG: All Methods	22			
		*Positive	21	95.45%
		Negative	1	4.55%
Alere BinaxNOW RSV	8	*Positive	8	100.00%
Quidel QuickVue RSV	3	*Positive	3	100.00%
Wampole Clearview RSV	3	*Positive	3	100.00%
SAS RSV Alert	1	*Positive	1	100.00%
Quidel QuickVue RSV 10	2	*Positive	2	100.00%
Quidel Sofia/Sofia 2 RSV	3	*Positive	3	100.00%
Cepheid GeneXpert - Xpert RSV	1	*Positive	1	100.00%
BD EZ RSV	1	Negative	1	100.00%
VAD-5				
* Denotes Correct Response				
AG: All Methods	22			
		*Negative	22	100.00%
BD EZ RSV	1	*Negative	1	100.00%
Alere BinaxNOW RSV	8	*Negative	8	100.00%
Quidel QuickVue RSV	3	*Negative	3	100.00%
Wampole Clearview RSV	3	*Negative	3	100.00%
SAS RSV Alert	1	*Negative	1	100.00%
Quidel QuickVue RSV 10	2	*Negative	2	100.00%
Quidel Sofia/Sofia 2 RSV	3	*Negative	3	100.00%
Cepheid GeneXpert - Xpert RSV	1	*Negative	1	100.00%
Adenovirus				
VAD-1				
* Denotes Correct Response				
AG: All Methods	13			
		*Negative	13	100.00%
SAS Adeno Test	9	*Negative	9	100.00%
Quidel AdenoPlus	4	*Negative	4	100.00%
VAD-2				
* Denotes Correct Response				
AG: All Methods	13			
		*Positive	13	100.00%
SAS Adeno Test	9	*Positive	9	100.00%
Quidel AdenoPlus	4	*Positive	4	100.00%
VAD-3				
* Denotes Correct Response				
AG: All Methods	13			
		*Negative	13	100.00%
SAS Adeno Test	9	*Negative	9	100.00%
Quidel AdenoPlus	4	*Negative	4	100.00%
VAD-4				
* Denotes Correct Response				
AG: All Methods	13			
		*Positive	13	100.00%
SAS Adeno Test	9	*Positive	9	100.00%
Quidel AdenoPlus	4	*Positive	4	100.00%
VAD-5				
* Denotes Correct Response				
AG: All Methods	13			
		*Negative	13	100.00%
SAS Adeno Test	9	*Negative	9	100.00%
Quidel AdenoPlus	4	*Negative	4	100.00%
932 - Hepatitis C Antibody (Waived); 767 - Hepatitis Markers				
	Labs	Response	Number	Percent
HCV (Waived)				
VM-1				
* Denotes Correct Response				
AG: All Methods	8			
		*Negative	8	100.00%
OraSure QraQuick HCV Rapid Antibody Test	8	*Negative	8	100.00%
VM-2				
* Denotes Correct Response				
AG: All Methods	8			
		*Positive	8	100.00%
OraSure QraQuick HCV Rapid Antibody Test	8	*Positive	8	100.00%
HBsAg				
VM-1				
* Denotes Correct Response				
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
VM-2				
* Denotes Correct Response				
AG: All Methods	1			



		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
		HCV		
VM-1		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
VM-2		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Abbott Architect Ci4100	1	*Positive	1	100.00%
767 - Hepatitis Markers	Labs	Response	Number	Percent
		HCV		
VM-1		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
VM-2		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Abbott Architect Ci4100	1	*Positive	1	100.00%
VM-3		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
VM-4		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
VM-5		* Denotes Correct Response		
AG: All Methods	1			
		*Positive	1	100.00%
Abbott Architect Ci4100	1	*Positive	1	100.00%
		Anti-HBc		
VM-1		* Denotes Correct Response		
PG: Abbott Architect Ci4100	1			
		*Negative	1	100.00%
VM-2		* Denotes Correct Response		
PG: Abbott Architect Ci4100	1			
		*Negative	1	100.00%
VM-3		* Denotes Correct Response		
PG: Abbott Architect Ci4100	1			
		*Positive	1	100.00%
VM-4		* Denotes Correct Response		
PG: Abbott Architect Ci4100	1			
		*Negative	1	100.00%
VM-5		* Denotes Correct Response		
PG: Abbott Architect Ci4100	1			
		*Negative	1	100.00%
		HAV		
VM-1		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
VM-2		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
VM-3		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
VM-4		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
VM-5		* Denotes Correct Response		



AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
		Anti-HBc (International)		
VM-1		* Denotes Correct Response		
PG: Intl OCD Vitros ECI/ECIQ	3			
		*Negative	3	100.00%
VM-2		* Denotes Correct Response		
PG: Intl OCD Vitros ECI/ECIQ	3			
		*Negative	3	100.00%
VM-3		* Denotes Correct Response		
PG: Intl OCD Vitros ECI/ECIQ	3			
		*Positive	3	100.00%
VM-4		* Denotes Correct Response		
PG: Intl OCD Vitros ECI/ECIQ	3			
		*Positive	3	100.00%
VM-5		* Denotes Correct Response		
PG: Intl OCD Vitros ECI/ECIQ	3			
		*Positive	3	100.00%
		HBsAb (International)		
VM-1		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Intl OCD Vitros ECI/ECIQ	2	*Positive	2	100.00%
VM-2		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Intl OCD Vitros ECI/ECIQ	2	*Negative	2	100.00%
VM-3		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Intl OCD Vitros ECI/ECIQ	2	*Negative	2	100.00%
VM-4		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	1	50.00%
		Positive	1	50.00%
Intl OCD Vitros ECI/ECIQ	2	*Negative	1	50.00%
Intl OCD Vitros ECI/ECIQ		Positive	1	50.00%
VM-5		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	1	50.00%
		Negative	1	50.00%
Intl OCD Vitros ECI/ECIQ	2	*Positive	1	50.00%
Intl OCD Vitros ECI/ECIQ		Negative	1	50.00%
		HBsAg (International)		
VM-1		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
VM-2		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
VM-3		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Positive	3	100.00%
VM-4		* Denotes Correct Response		
AG: All Methods	3			
		*Negative	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Negative	3	100.00%
VM-5		* Denotes Correct Response		
AG: All Methods	3			
		*Positive	3	100.00%
Intl OCD Vitros ECI/ECIQ	3	*Positive	3	100.00%
		HCV (International)		
VM-1		* Denotes Correct Response		



AG: All Methods	2			
		*Negative	2	100.00%
Intl OCD Vitros ECI/ECIQ	2	*Negative	2	100.00%
VM-2		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Intl OCD Vitros ECI/ECIQ	2	*Positive	2	100.00%
VM-3		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Intl OCD Vitros ECI/ECIQ	2	*Negative	2	100.00%
VM-4		* Denotes Correct Response		
AG: All Methods	2			
		*Negative	2	100.00%
Intl OCD Vitros ECI/ECIQ	2	*Negative	2	100.00%
VM-5		* Denotes Correct Response		
AG: All Methods	2			
		*Positive	2	100.00%
Intl OCD Vitros ECI/ECIQ	2	*Positive	2	100.00%
		HBsAg		
VM-1		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
VM-2		* Denotes Correct Response		
AG: All Methods	1			
		*Negative	1	100.00%
Abbott Architect Ci4100	1	*Negative	1	100.00%
771 - Vaginal Pathogens I - Affirm VP	Labs	Response	Number	Percent
		Candida sp. - Non Culture		
VP-1		* Denotes Correct Response		
AG: All Methods	33			
		*Positive	33	100.00%
Affirm VP	33	*Positive	33	100.00%
VP-2		* Denotes Correct Response		
AG: All Methods	33			
		*Negative	33	100.00%
Affirm VP	33	*Negative	33	100.00%
VP-3		* Denotes Correct Response		
AG: All Methods	33			
		*Positive	32	96.97%
		Negative	1	3.03%
Affirm VP	33	*Positive	32	96.97%
Affirm VP		Negative	1	3.03%
VP-4		* Denotes Correct Response		
AG: All Methods	33			
		*Positive	33	100.00%
Affirm VP	33	*Positive	33	100.00%
VP-5		* Denotes Correct Response		
AG: All Methods	33			
		*Negative	33	100.00%
Affirm VP	33	*Negative	33	100.00%
		G. vaginalis - Non Culture		
VP-1		* Denotes Correct Response		
AG: All Methods	33			
		*Negative	33	100.00%
Affirm VP	33	*Negative	33	100.00%
VP-2		* Denotes Correct Response		
AG: All Methods	33			
		*Positive	33	100.00%
Affirm VP	33	*Positive	33	100.00%
VP-3		* Denotes Correct Response		
AG: All Methods	33			
		*Negative	32	96.97%
		Positive	1	3.03%
Affirm VP	33	*Negative	32	96.97%
Affirm VP		Positive	1	3.03%



VP-4		* Denotes Correct Response	
AG: All Methods	33		
		*Negative	32 96.97%
		Positive	1 3.03%
Affirm VP	33	*Negative	32 96.97%
Affirm VP		Positive	1 3.03%
VP-5		* Denotes Correct Response	
AG: All Methods	33		
		*Negative	33 100.00%
Affirm VP	33	*Negative	33 100.00%
		T. vaginalis - Non Culture Tech	
VP-1		* Denotes Correct Response	
AG: All Methods	33		
		*Negative	33 100.00%
Affirm VP	33	*Negative	33 100.00%
VP-2		* Denotes Correct Response	
AG: All Methods	33		
		*Negative	33 100.00%
Affirm VP	33	*Negative	33 100.00%
VP-3		* Denotes Correct Response	
AG: All Methods	33		
		*Positive	33 100.00%
Affirm VP	33	*Positive	33 100.00%
VP-4		* Denotes Correct Response	
AG: All Methods	33		
		*Negative	33 100.00%
Affirm VP	33	*Negative	33 100.00%
VP-5		* Denotes Correct Response	
AG: All Methods	33		
		*Positive	32 96.97%
		Negative	1 3.03%
Affirm VP	33	*Positive	32 96.97%
Affirm VP		Negative	1 3.03%